

DEPARTMENT OF HEALTH SERVICES
TOXIC SUBSTANCES CONTROL PROGRAM
700 HEINZ AVE., BLDG. F, SUITE 200
BERKELEY, CA 94710-2737
(415) 540-3724

N60028_000029
TREASURE ISLAND
SSIC NO. 5090.3.A



RECEIVED
IN CODE 181

'91 JUL 18 P4:18

July 16, 1991

Commanding Officer
Western Division
Attn: Mr. Ernesto Galang
Naval Facilities Engineering Command
P.O. Box 727
San, Bruno, CA 94066-0720

Dear Mr. Galang:

DEPARTMENT OF HEALTH SERVICES COMMENTS ON THE DRAFT FINAL QUALITY ASSURANCE PROJECT PLAN, TREASURE ISLAND, CALIFORNIA

On June 21, 1991, you submitted to the Department of Health Services (DHS) a copy of Draft Final Quality Assurance Project Plan (QAPP) for the Remedial Investigation/Feasibility Study (RI/FS) at Treasure Island (TI). Your submission of the Draft Final QAPP is for DHS review and comment.

Enclosed you will find DHS' comments on the submitted report. Please ensure that these comments are addressed when revising the QAPP.

You may call me at (415) 540-3815 if you have questions or comments.

Sincerely,

A handwritten signature in cursive script that reads "Romy F. Fuentes".

Romy F. Fuentes
Waste Management Engineer
Region 2
Toxic Substances Control Program

Enclosure

cc: Commanding Officer
Building One, Code 82.2
Attn.: Mr. Eddie Sarmiento
Staff Civil Engineers Office
NAVSTA Treasure Island
San Francisco, CA 94103

Mr. Tom Gandesbery
S.F. Bay, RWQCB
2101 Webster Street, Suite 500
Oakland, CA 94611

ADMIN RECORD

D/N 1 29

DHS COMMENTS ON THE QUALITY ASSURANCE PROJECT PLAN
TREASURE ISLAND NAVAL WEAPON STATION

1. The QAPP is generally satisfactory for its intended purpose.
2. The QAPP does not fully conform to the page control format outlined in the EPA QAMS-005/80: Interim Guidelines and Specifications for Preparing Quality Assurance Project Plans, as follows:

Section No. _____
Revision No. _____
Date: _____
Page _____ of _____

The Section No. was not included in the QAPP. The Section No. should appear on individual pages as shown above.

3. On page 8 of 31, the detection or quantitation limits were not included in the process of development of the Data Quality Objectives (DQO). Quantitative goals for detection or quantitation limits to be achieved should be specified.
4. On page 16 of 31, it indicated that the calibration of laboratory equipment will be performed by a Navy-Certified laboratory. Please note that environmental analyses for regulatory purposes should be done by laboratories certified/credited by the Environmental Laboratory Accreditation Program of the California Department of Health Services (telephone: (415) 540-2800).