Navy_CLEAN_PFAS_FOIA_request_covers.indd

NAVFAC
Naval Facilities Engineering Command

Groundwater Sample Results, Level 2 Laboratory Report, Level 4 Laboratory Report, Electronic Data Deliverable, Data Validation Report, and the Sample Location Report, SDG 18-0602

Naval Air Station Point Mugu
Point Mugu, California

July 2019

APPROVED FOR PUBLIC RELEASE; DISTRIBUTION UNLIMITED

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

AQ, GW

Batch 18-0602

Package DP-18-0303

Submitted to:
CH2M
1100 NE Circle Blvd Suite 300
Corvallis, OR 97330 USA

Submitted by:
Battelle Norwell Operations
141 Longwater Drive Suite 202
Norwell, MA 02061

BATTELLE
It can be done

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

AQ, GW

Batch 18-0602

Package DP-18-0303

Submitted to:
CH2M
1100 NE Circle Blvd Suite 300
Corvallis, OR 97330 USA

NELAP Accreditation Number: E87856 (Florida Department of Health)
DoD-ELAP Accreditation Number: 91667

Submitted by:
Battelle Norwell Operations
141 Longwater Drive Suite 202
Norwell, MA 02061

Analyst Approval: Lauren Griffith 2018.10.25 14:21:15 -04'00'

QC Chemist Approval: fitche@battelle.org 2018.11.01 15:39:03 -04'00'

Project Manager Approval: Digitally signed by Jonathan Thorn Date: 2018.11.02 05:55:02 -04'00'

BATTELLE
It can be done

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

SD

Batch 18-0602

Package DP-18-0303

1	***Miscellaneous Documentation*** Case Narrative, Laboratory Accreditations.	1
2	***Chain of Custody Records*** Sample Receipt Records	11
3	***Tables*** Analytical Data Tables, Qualifier Definitions.	20

BATTELLE

It can be done

BATTELLE

Signature Page

Battelle 2018 (1 of 2)
Signature Page

Name (Printed)	Signature	Initials	Date
Jonathan Thorn		JRT	4/4/2018
Robert Lizotte, Jr.		BL	4.4 2018
FRANCO PALA			4-4-2018
Carla Devine		CRD	4/4/18
Denise Schumitz		DMS	4/4/18
Candice Penn McCoy		CPM	4/4/2018
Rich Rostucci		RR	4/4/2018
Michael Mears			4/4/2018
Christie Usher		cu	4/4/18
Kevin McInerney			4/4/18
Stephanie Schultz		SAS	4/4/18
Jordan Tower		JCT	4/4/18
KRISTEN NICHOLS		KN	4/4/18
Clinton H Brown		CB	4/4/18
Matt Schumitz		MS	4-4-18
Sam Guimaraes		SG	4-4-18
Lauren Griffith		LMG	4.4.18

BATTELLE

Signature Page

Battelle 2018 (2 of 2)
Signature Page

Name (Printed)	Signature	Initials	Date
KAVITHA DASU		KD	04/04/18
Kayla Lamarre		KAL	04/04/18
Weidong Li		W.L	04/04/18
Tracy W Stenner		TWS	04/04/18
Ellyn M Fitch		EMF	12-April-2018
Gail DeRuzzo		GD	4/18/18
Zachary Willenberg		ZJW	4/20/18

Sample Summary

Client: CH2M
SDG: 18-0602
Project/Site: Naval Base Ventura County
CTO: 4164

Lab Sample ID	Client Sample ID	Matrix	Collection Date	Receipt Date
CR939PB-FS	Procedural Blank	WATER	10/10/2018	10/10/2018
CR940LCS-FS	Laboratory Control Sample	WATER	10/10/2018	10/10/2018
J8613-FS	VC-PM553-GW0601A-1018	GW	10/1/2018	10/5/2018
J8614-FS	VC-AQ-FB09-10012018	AQ	10/1/2018	10/5/2018
J8615-FS	VC-AQ-EB09-10012018	AQ	10/1/2018	10/5/2018
J8616-FS	VC-PM13-GW21-1018	GW	10/1/2018	10/5/2018
J8617-FS	VC-KCHA16-GW01-1018	GW	10/2/2018	10/5/2018
J8618-FS	VC-KCHA16-GW01P-1018	GW	10/2/2018	10/5/2018
J8619-FS	VC-KCHA16-GW02-1018	GW	10/2/2018	10/5/2018
J8620-FS	VC-KCHA16-GW03-1018	GW	10/2/2018	10/5/2018
J8621-FS	VC-AQ-FB10-10022018	AQ	10/2/2018	10/5/2018
J8622-FS	VC-AQ-EB10-10022018	AQ	10/2/2018	10/5/2018
J8623-FS	VC-S09GW01-1018	GW	10/3/2018	10/5/2018
J8624-FS	VC-S09GW04-1018	GW	10/2/2018	10/5/2018
J8625-FS	VC-S09GW04P-1018	GW	10/2/2018	10/5/2018
J8626-FS	VC-S09GW06-1018	GW	10/3/2018	10/5/2018
J8627-FS	VC-S14GW01-1018	GW	10/3/2018	10/5/2018
J8627-FS	VC-S14GW01-1018	GW	10/3/2018	10/5/2018

Miscellaneous Documentation

BATTELLE
It can be done

QA/QC Summary
Batch 18-0602

Project:	CTO-4164 Naval Base Ventura County, California
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	GW and AQ
Data Set:	DP-18-0303
Analytical SOP:	5-369
Method Reference:	PFAS to QSM 5.1 Table B-15

Sample Custody		
Collection Date	Receipt Date	Temp (°C)
10/ 1-3 /2018	10/5/2018	1.0
Corrective Actions	None.	
Sample Storage	The water samples were stored refrigerated until extraction.	
Related samples	NA	

	METHOD SUMMARIES
Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with 0.4% NH_3 in methanol. Extracts were concentrated to dryness under nitrogen with a water bath set between 35 °C and 45 °C, reconstituted with 80:20 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.
Prep comments	All samples were pre-screened prior to initial extraction.
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.
Analysis Comments	Samples analyzed on Sciex 5500 LC-MS/MS. When detected, PFHxS and PFOS contain both linear and branched isomers. There are instances of extracted internal standards (surrogates) that would fall outside of the 50%-150% recovery criteria if reported from the non-diluted extract. This is due to the higher levels of native PFOA and PFOS found in the samples interfering with the 13C2-PFOA and 13C4-PFOS internal standards used to quantify the surrogates. In these cases, the surrogates are reported from dilutions, however, the native compounds are reported from the non-diluted extracts as the interferences are with the internal standards and not the surrogates used to quantify the native compounds.

QA/QC Summary
Batch 18-0602

Holding Times	Extraction Date(s)	Analysis Date(s)
	10/10/2018	10/17,23,24 /2018

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
≤ ½ the LOQ Samples >10x PB	Two exceedances noted.
	PFOA was detected in two authentic samples at less than 10 times the amount detected in the PB, and is B qualified in these samples.

Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
Laboratory derived control limits for recovery	No exceedances noted.
	No comments.

Matrix Spike and Matrix Spike Duplicate (MS/MSD)	A MS/MSD was not prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
Laboratory derived control limits for recovery and <30% RPD	NA
	None prepared with this batch.

Extracted Internal Standard Analytes	Labelled analog compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
50-150% of true value	No exceedances noted.
	No comments.

Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
+/- 50% of the area of the L5 calibration point.	No exceedances noted.
	There are several instances where 13C2-PFOA and 13C4-PFOS would fail IS area criteria. These failures are due to the amount of native PFOA and PFOS found in the samples interfering with the internal standards in the non-diluted extracts. In these cases, the extracted internal standards that would use these two internal standards are reported from a dilution to mitigate the interferences from the high concentrations in the non-diluted samples. No data is reported from extracts where the IS areas would have failed criteria.

Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadradic curve fitting.
+/- 30% of true value, $R^2 \geq 0.99$	No exceedances noted.
	No comments.

QA/QC Summary
Batch 18-0602

Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
+/- 30% of true value	No exceedances noted.
	No comments.

Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.
+/- 30% of true value	No exceedances noted.
	No comments.

Instrument Blank (IB)	Immediately following the highest standard analyzed and daily prior to sample analysis.
≤ ½ the LOQ	No exceedances noted.
	No comments.

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project Number: 100110125-01
Preparation Batch: 18-0602
Data Set: DP-18-0303
Test Code: Master_369

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	2	PFOA was detected in two authentic samples at less than 10 times the amount detected in the PB, and is B qualified in these samples.
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	0	None
Matrix Spike / Matrix Spike Duplicate Precision	0	None
Extracted Internal Standard Analytes (Surrogates)	0	None
Instrument Calibration	0	None
Instrument Blank	0	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
MISCELLANEOUS DOCUMENTATION FORM

Project Title:	CTO-4164 Naval Base Ventura County,	**Data Set Number:**	DP-18-0303
Project Number:	100110125-01	**Prep Batch Number:**	18-0602
Entered By:	Lauren Griffith	**Entered On:**	10/25/2018
Test Code (Matrix Type):	Master_369(L)		

Samples that were manually integrated are noted on the quant reports with the comment (TRUE).
LMG 10/25/18

PFOS appears to exhibit a retention time shift because it is a branched analyte and the retention time is defined by whichever branched peak has the largest area. The entire branched peak is picked in all cases.
LMG 10/25/18

In cases were native PFAS compounds were reported from dilutions (above calibration in non-diluted extracts), the extracted internal standard (surrogate) used to quantify the native compound was also reported from the dilution.
LMG 10/25/18

There are instances of extracted internal standards (surrogates) that would fall outside of the 50%-150% recovery criteria if reported from the non-diluted extract. This is due to the higher levels of native PFOA and PFOS found in the samples interfering with the 13C2-PFOA and 13C4-PFOS internal standards used to quantify the surrogates. In these cases, the surrogates are reported from dilutions, however, the native compounds are reported from the non-diluted extracts as the interferences are with the internal standards and not the surrogates used to quantify the native compounds.
LMG 10/25/18

Task Leader Approval:		
SupervisorApproval:		Robert Lizotte, Jr. 2018.10.26 11:06:41 -04'00'
PM Approval:		Robert Lizotte, Jr. 2018.10.26 11:06:55 -04'00'

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batch: 18-0602
Data Set: DP-18-0303

		CR939PB-FS (Procedural Blank)	CR940LCS-FS (Laboratory Control Sample)	J8613-FS (VC-PM553-GW0601A-1018)	J8614-FS (VC-AQ-FB09-10012018)	J8615-FS (VC-AQ-EB09-10012018)	J8616-FS (VC-PM13-GW21-1018)
PFHxA	307-24-4	-	L	L	-	-	L
PFHpA	375-85-9	-	L	L	-	-	L
PFOA	335-67-1	-	L	L	L	L	L
PFNA	375-95-1	-	L	L	-	-	L
PFDA	335-76-2	-	L	-	-	-	-
PFUnA	2058-94-8	-	L	-	-	-	-
PFDoA	307-55-1	-	L	-	-	-	-
PFTrDA	72629-94-8	-	L	-	-	-	-
PFTeDA	376-06-7	-	L	-	-	-	-
NMeFOSAA	2355-31-9	-	L	-	-	-	-
NEtFOSAA	2991-50-6	-	L	-	-	-	-
PFBS	375-73-5	-	L	L	-	-	L
PFHxS	355-46-4	-	L/Br	L/Br	-	-	L/Br
PFOS	1763-23-1	-	L/Br	L/Br	-	-	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Branched
"-": Not detected

BATTE
It can

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batcl
Data Set: DP-18-

	J8617-FS (VC-KCHA16-GW01-1018)	J8618-FS (VC-KCHA16-GW01P-1018)	J8619-FS (VC-KCHA16-GW02-1018)	J8620-FS (VC-KCHA16-GW03-1018)	J8621-FS (VC-AQ-FB10-10022018)	J8622-FS (VC-AQ-EB10-10022018)	J8623-FS (VC-S09GW01-1018)	J8624-FS (VC-S09GW04-1018)
PFHxA	L	L	L	L	-	-	L	L
PFHpA	L	L	L	L	-	-	L	L
PFOA	L	L	L	L	L	L	L	L
PFNA	-	-	-	-	-	-	L	L
PFDA	-	-	-	-	-	-	L	L
PFUnA	-	-	-	-	-	-	-	-
PFDoA	-	-	-	-	-	-	-	-
PFTrDA	-	-	-	-	-	-	-	-
PFTeDA	-	-	-	-	-	-	-	-
NMeFOSAA	-	-	-	-	-	-	-	-
NEtFOSAA	-	-	-	-	-	-	-	-
PFBS	L	L	L	L	-	-	L	L
PFHxS	L/Br	L/Br	L/Br	L/Br	-	-	L/Br	L/Br
PFOS	L/Br	L/Br	L/Br	L/Br	-	L/Br	L/Br	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Bra
"-": Not detected

BATTE
It can

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batcl
Data Set: DP-18-

	J8625-FS (VC-S09GW04P-1018)	J8626-FS (VC-S09GW06-1018)	J8627-FS (VC-S14GW01-1018)	J8628-FS (VC-S14GW03-1018)
PFHxA	L	L	L	-
PFHpA	L	L	L	-
PFOA	L	L	L	L
PFNA	L	L	L	-
PFDA	L	L	-	-
PFUnA	-	-	-	-
PFDoA	-	-	-	-
PFTrDA	-	-	-	-
PFTeDA	-	-	-	-
NMeFOSAA	-	-	-	-
NEtFOSAA	-	-	-	-
PFBS	L	L	L	-
PFHxS	L/Br	L/Br	L/Br	L/Br
PFOS	L/Br	L/Br	L/Br	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Bra
"-": Not detected

ACCREDITATIONS

Accrediting Authority	Laboratory ID
U.S. Department of Defense Environmental Laboratory Accreditation Program (DoD-ELAP)	91667
State of Florida Department of Health	E87856
State of New York Department of Health	12105
Commonwealth of Pennsylvania Department of Environmental Protection	68-05687
State of Washington Department of Ecology	C1050
State of California	3045
Commonwealth of Massachusetts	E87856

Current certificates and lists of accredited parameters are available upon request.

Custody Records

BATTELLE
It can be done

ShpNo SHP-181005-03

Battelle Project No: 0110125-01

Sample Receipt Form

Approved: ☑ Authorized ☐

Project Number: 695803 **Client:** CH2M

Received by: Schumitz, Matt **Date/Time Received:** Friday, October 05, 2018 10:30 AM

No. of Shipping Containers: 1

SHIPMENT

Method of Delivery: Commercial Carrier **Tracking Number:** Fed Ex

COC Forms: ☑ **Shipped with samples** ☐ **No Forms**

Cooler(s)/Box(es)

Cntr	Type	Tracking No.	Seal	Seal	Container	Therm.	Temp C	Smps
1 of 1	Cooler	7830 8987 6121	Custody Seals	Intact	Intact	Therm_1	1.0	16

Samples

Sample Labels:
☑ Sample labels agree with COC forms
☐ Discrepancies (see Sample Custody Corrective Action Form)

Container Seals:
☐ Tape ☐ Custody Seals ☐ Other Seals (See sample Log)
☑ Seals intact for each shipping container
☐ Seals broken (See sample log for impacted samples)

Condition of Samples:
☑ Sample containers intact
☐ Sample containers broken/leaking (See Custody Corrective Action Form)

Temperature upon receipt (°C): 1 Temperature Blank used ☑ Yes ☐ No
(Note: If temperature upon receipt differs from required conditions, see sample log comment field)

Samples Acidified: ☐ Yes ☐ No ☑ Unknown

Initial pH 5-9?: ☐ Yes ☐ No ☑ NA
If no, individual sample adjustments on the Auxiliary Sample Receipt Form

Total Residual Chlorine Present?: ☐ Yes ☐ No ☑ NA
If yes, individual sample adjustments on the Auxiliary Sample Receipt Form

Head Space <1% in samples for water VOC analysis: ☐ Yes ☐ No ☑ NA
Individual sample deviations noted on sample log

Samples Containers:
Samples returned in PC-grade jars: ☐ Yes ☐ No ☑ Unknown /Lot No.: Unknown

Storage Location: Custody: Refrigerator - R0119 (NA) **BDO IDs Assigned:** J8613 - J8628

Samples logged in by: Schumitz, Matt **Date/Time:** 10/05/2018 10:30 AM

Approved By: Fitch, Ellyn **Approved On:** 10/9/2018 8:26:00 AM

Authorized By: **Authorized On:**

BATTELLE
It can be done

ShpNo SHP-181005-03

Battelle Project No:)110125-01

Sample Receipt Form Details

Approved: ☐ Authorized ☐

Project Number: 695803 **Client:** CH2M

Received by: Schumitz, Matt **Date/Time Received:** Friday, October 05, 2018 10:30 AM

No. of Shipping Containers: 1

BDO Id:	Client Sample ID:	Collection Date:	Login Date:	Ctrs:	Matrix:	Temp:	pH:	TRC:	VOC:	Stored In:	Loc:	No:	Comments:
J8613	VC-PM553-GW0601A-1018	10/01/18 11:47	10/05/18 11:22	2	GW	1	NA	NA	NA	R0119 (NA)			
J8614	VC-AQ-FB09-10012018	10/01/18 15:57	10/05/18 11:22	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8615	VC-AQ-EB09-10012018	10/01/18 16:00	10/05/18 11:22	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8616	VC-PM13-GW21-1018	10/01/18 14:33	10/05/18 11:23	2	GW	1	NA	NA	NA	R0119 (NA)			
J8617	VC-KCHA16-GW01-1018	10/02/18 11:50	10/05/18 11:25	2	GW	1	NA	NA	NA	R0119 (NA)			
J8618	VC-KCHA16-GW01P-1018	10/02/18 11:54	10/05/18 11:28	2	GW	1	NA	NA	NA	R0119 (NA)			
J8619	VC-KCHA16-GW02-1018	10/02/18 10:15	10/05/18 11:30	2	GW	1	NA	NA	NA	R0119 (NA)			
J8620	VC-KCHA16-GW03-1018	10/02/18 13:18	10/05/18 11:30	2	GW	1	NA	NA	NA	R0119 (NA)			
J8621	VC-AQ-FB10-10022018	10/02/18 16:00	10/05/18 11:30	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8622	VC-AQ-EB10-10022018	10/02/18 16:10	10/05/18 11:31	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8623	VC-S09GW01-1018	10/03/18 12:22	10/05/18 11:31	2	GW	1	NA	NA	NA	R0119 (NA)			
J8624	VC-S09GW04-1018	10/02/18 16:25	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8625	VC-S09GW04P-1018	10/02/18 16:29	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8626	VC-S09GW06-1018	10/03/18 11:22	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8627	VC-S14GW01-1018	10/03/18 14:39	10/05/18 11:38	2	GW	1	NA	NA	NA	R0119 (NA)			
J8628	VC-S14GW03-1018	10/03/18 15:35	10/05/18 11:38	2	GW	1	NA	NA	NA	R0119 (NA)			

Total Samples: 16

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: NBVC Point Mugu PM553	Site Information: Hangar PM553
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3629 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal X Priority ☐ RUSH ☐ Time Zone: Pacific	Analysis: PFAS by Method 537 Mod	Page# 1 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
J8613 VC-PM553-GW0601A-1018	10/1/18	1147	Grab	GW	2	X
~~VC-PM553-DW01~~ KDR			Grab	GW		X
~~VC-PM553-DW01P~~ KDR			Grab	GW		X
~~VC-PM553-DW02~~ KDR			Grab	GW		X
~~VC-PM553-DW03~~ KDR			Grab	GW		X
~~VC-PM553-DW04~~ KDR			Grab	GW		X
J8614 VC ~~FDT~~-AQ-FB 09-1001208	10/1/18	1557	Grab	AQ	2	X
J8615 VC ~~FDT~~-AQ-EB 09-1001208	10/1/18	1600	Grab	AQ	2	X

On water meter

VK

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200		Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PM372 NBVC Point Mugu	Site Information: Hangar PM372
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal X Priority ☐ RUSH ☐ Time Zone: Pacific	Analysis: PFAS by Method 537 Mod	Page# 2 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
J8616 VC-PM13-GW21-1018	10/1/18	1433	Grab	GW	2	X
~~VC-PM372-DW01-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW02-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW02P-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW03-~~			~~Grab~~	~~GW~~		X
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~		X
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~		X

VKJ

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	NA	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: ADC16/ KCHA16	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724.977.3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal ☒ Priority ☐ RUSH ☐ Time Zone: PST	Analysis: PFAS by Method 537 Mod	Page# 3 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod	
VC-KCHA16-GW01-101[8] J8617	10/2/18	1150	Grab	GW	2	X	
VC-KCHA16-GW01P-101[8] J8618	10/2/18	1154	Grab	GW	2	X	
VC-KCHA16-GW02-101[8] J8619	10/2/18	1015	Grab	GW	2	X	
VC-KCHA16-GW03-101[8] J8620	10/2/18	1318	Grab	GW	2	X	
VC ~~FDT~~-AQ-FB 10-10022018 J8621	10/2/18	1600	Grab	AQ	2	X	
VC ~~FDT~~-AQ-EB 10-10022018 J8622	10/2/18	1618	Grab	AQ	2	X	on pump tubing

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200		Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PH S09	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3028 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal ☒ Priority ☐ RUSH ☐ Time Zone: PST	Analysis: PFAS by Method 537 Mod	Page# 4 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
VC-S09GW01-1018 J8623	10/3/18	1222	Grab	GW	2	X
VC-S09GW04-1018 J8624	10/2/18	1625	Grab	GW	2	X
VC-S09GW04P-1018 J8625	10/2/18	1629	Grab	GW	2	X
~~VC-HS09-DW01-~~			~~Grab~~	~~GW~~		X
~~VC-HS09-DW02-~~			~~Grab~~	~~GW~~		X
~~VC-HS09-DW03-~~			~~Grab~~	~~GW~~		X
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~		X
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~		X
J8626 VC-S09GW06-1018	10/3/18	1122	Grab	GW	2	X

VB

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	MS	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PH S14	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:		
Project Name: NBVC Basewide SI Project No.: 655803	Normal ☒ Priority ☐ RUSH ☐ Time Zone: PST		COC # 5 Page# 5 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	Preservative: NA	Analysis: PFAS by Method 537 Mod
J8627 VC-S14GW01-1018	10/3/18	1439	Grab	GW	2		X
~~VC-S14GW02-~~			~~Grab~~	~~GW~~			X
~~VC-S14GW02P-~~			~~Grab~~	~~GW~~			X
J8628 VC-S14GW03-1018	10/3/18	1535	Grab	GW	2		X
~~VC-S14GW19-~~			~~Grab~~	~~GW~~			X
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~			X
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~			X

VK

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	NB	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:



ORIGIN ID:OXRA (724) 977-3628
VICTORIA KILBERTS

402 W BROADWAY STE 1450

SAN DIEGO, CA 92101
UNITED STATES US

SHIP DATE: 040CT18
ACTWGT: 51.80 LB
CAD: 6997686/SSF01922
DIMS: 25x14x13 IN

BILL THIRD PARTY

TO JOHNATHON THORNE
BATELLE LABS
141 LONGWATER DR
SUITE 202
NORWELL MA 02061

(781) 681-5665 REF: 695803
INV:
PO: DEPT:

1.0 TB
Therm 1
10/5/18
10:30

FedEx Express

REL# 3785346

TRK# 0201 7830 8987 6121

FRI - 05 OCT 10:30A
PRIORITY OVERNIGHT

XE XPUA

02061
MA-US BOS

Data Tables

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM553-GW0601A-1018				
Battelle ID		J8613-FS				
Sample Type		SA				
Collection Date		10/01/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	7602.64	D	29.32	77.16	771.61
PFHpA	375-85-9	3916.20	D	24.69	77.16	771.61
PFOA	335-67-1	6727.83	D	27.78	77.16	771.61
PFNA	375-95-1	5.37		0.24	0.93	4.63
PFDA	335-76-2	0.46	U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	305.91	D	6.02	23.15	231.48
PFHxS	355-46-4	2350.63	D	5.09	18.52	231.48
PFOS	1763-23-1	110.92	D	1.76	4.63	46.30

Surrogate Recoveries (%)

13C5-PFHxA	109	D
13C4-PFHpA	109	D
13C8-PFOA	109	D
13C9-PFNA	106	D
13C6-PFDA	84	
13C7-PFUnA	93	
13C2-PFDoA	87	
13C2-PFTeDA	86	
d3-MeFOSAA	89	
d5-EtFOSAA	82	
13C3-PFBS	120	D
13C3-PFHxS	97	D
13C8-PFOS	105	D

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB09-10012018			
Battelle ID		J8614-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.290			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.43 U	0.16	0.43	4.31
PFHpA	375-85-9	0.43 U	0.14	0.43	4.31
PFOA	335-67-1	0.91 J	0.16	0.43	4.31
PFNA	375-95-1	0.86 U	0.22	0.86	4.31
PFDA	335-76-2	0.43 U	0.14	0.43	4.31
PFUnA	2058-94-8	0.86 U	0.25	0.86	4.31
PFDoA	307-55-1	0.43 U	0.16	0.43	4.31
PFTrDA	72629-94-8	0.43 U	0.13	0.43	4.31
PFTeDA	376-06-7	0.86 U	0.22	0.86	4.31
NMeFOSAA	2355-31-9	1.72 U	0.48	1.72	4.31
NEtFOSAA	2991-50-6	0.86 U	0.42	0.86	4.31
PFBS	375-73-5	0.43 U	0.11	0.43	4.31
PFHxS	355-46-4	0.34 U	0.09	0.34	4.31
PFOS	1763-23-1	0.43 U	0.16	0.43	4.31

Surrogate Recoveries (%)

13C5-PFHxA	85
13C4-PFHpA	83
13C8-PFOA	86
13C9-PFNA	74
13C6-PFDA	94
13C7-PFUnA	96
13C2-PFDoA	86
13C2-PFTeDA	93
d3-MeFOSAA	75
d5-EtFOSAA	83
13C3-PFBS	89
13C3-PFHxS	90
13C8-PFOS	88

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB09-10012018			
Battelle ID		J8615-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.45 U	0.17	0.45	4.55
PFHpA	375-85-9	0.45 U	0.15	0.45	4.55
PFOA	335-67-1	0.95 J	0.16	0.45	4.55
PFNA	375-95-1	0.91 U	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	0.45 U	0.12	0.45	4.55
PFHxS	355-46-4	0.36 U	0.10	0.36	4.55
PFOS	1763-23-1	0.45 U	0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	95
13C4-PFHpA	83
13C8-PFOA	91
13C9-PFNA	82
13C6-PFDA	84
13C7-PFUnA	87
13C2-PFDoA	85
13C2-PFTeDA	85
d3-MeFOSAA	71
d5-EtFOSAA	58
13C3-PFBS	73
13C3-PFHxS	78
13C8-PFOS	77

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM13-GW21-1018				
Battelle ID		J8616-FS				
Sample Type		SA				
Collection Date		10/01/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	245.61	D	2.88	7.58	75.76
PFHpA	375-85-9	48.74		0.15	0.45	4.55
PFOA	335-67-1	63.52		0.16	0.45	4.55
PFNA	375-95-1	1.00	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	397.55	D	1.97	7.58	75.76
PFHxS	355-46-4	755.99	D	1.67	6.06	75.76
PFOS	1763-23-1	94.92	D	2.88	7.58	75.76
Surrogate Recoveries (%)						
13C5-PFHxA		117	D			
13C4-PFHpA		99				
13C8-PFOA		79				
13C9-PFNA		74				
13C6-PFDA		90				
13C7-PFUnA		96				
13C2-PFDoA		93				
13C2-PFTeDA		99				
d3-MeFOSAA		99				
d5-EtFOSAA		102				
13C3-PFBS		128	D			
13C3-PFHxS		99	D			
13C8-PFOS		99	D			

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01-1018				
Battelle ID		J8617-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	338.20	D	1.73	4.55	45.45
PFHpA	375-85-9	13.04		0.15	0.45	4.55
PFOA	335-67-1	6.64	B	0.16	0.45	4.55
PFNA	375-95-1	0.91	U	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	153.20	D	1.18	4.55	45.45
PFHxS	355-46-4	177.76	D	1.00	3.64	45.45
PFOS	1763-23-1	8.94		0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	114	D
13C4-PFHpA	121	
13C8-PFOA	89	
13C9-PFNA	82	
13C6-PFDA	88	
13C7-PFUnA	96	
13C2-PFDoA	91	
13C2-PFTeDA	99	
d3-MeFOSAA	91	
d5-EtFOSAA	71	
13C3-PFBS	139	D
13C3-PFHxS	98	D
13C8-PFOS	100	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01P-1018				
Battelle ID		J8618-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.265				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	304.87	D	1.79	4.72	47.17
PFHpA	375-85-9	13.41		0.15	0.47	4.72
PFOA	335-67-1	5.94	B	0.17	0.47	4.72
PFNA	375-95-1	0.94	U	0.25	0.94	4.72
PFDA	335-76-2	0.47	U	0.15	0.47	4.72
PFUnA	2058-94-8	0.94	U	0.27	0.94	4.72
PFDoA	307-55-1	0.47	U	0.17	0.47	4.72
PFTrDA	72629-94-8	0.47	U	0.14	0.47	4.72
PFTeDA	376-06-7	0.94	U	0.24	0.94	4.72
NMeFOSAA	2355-31-9	1.89	U	0.53	1.89	4.72
NEtFOSAA	2991-50-6	0.94	U	0.46	0.94	4.72
PFBS	375-73-5	124.26	D	1.23	4.72	47.17
PFHxS	355-46-4	147.29	D	1.04	3.77	47.17
PFOS	1763-23-1	7.50		0.18	0.47	4.72

Surrogate Recoveries (%)

13C5-PFHxA	103	D
13C4-PFHpA	117	
13C8-PFOA	93	
13C9-PFNA	76	
13C6-PFDA	84	
13C7-PFUnA	87	
13C2-PFDoA	75	
13C2-PFTeDA	84	
d3-MeFOSAA	76	
d5-EtFOSAA	62	
13C3-PFBS	126	D
13C3-PFHxS	85	D
13C8-PFOS	94	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW02-1018				
Battelle ID		J8619-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	163.76	D	1.76	4.63	46.30
PFHpA	375-85-9	39.13		0.15	0.46	4.63
PFOA	335-67-1	39.85		0.17	0.46	4.63
PFNA	375-95-1	0.93	U	0.24	0.93	4.63
PFDA	335-76-2	0.46	U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	49.17		0.12	0.46	4.63
PFHxS	355-46-4	250.99	D	1.02	3.70	46.30
PFOS	1763-23-1	17.94		0.18	0.46	4.63

Surrogate Recoveries (%)

13C5-PFHxA	111	D
13C4-PFHpA	122	
13C8-PFOA	98	
13C9-PFNA	82	
13C6-PFDA	77	
13C7-PFUnA	82	
13C2-PFDoA	76	
13C2-PFTeDA	80	
d3-MeFOSAA	72	
d5-EtFOSAA	71	
13C3-PFBS	101	
13C3-PFHxS	87	D
13C8-PFOS	79	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW03-1018				
Battelle ID		J8620-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	229.43	D	1.73	4.55	45.45
PFHpA	375-85-9	32.09		0.15	0.45	4.55
PFOA	335-67-1	66.04	D	1.64	4.55	45.45
PFNA	375-95-1	0.91	U	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	69.78	D	1.18	4.55	45.45
PFHxS	355-46-4	136.86	D	1.00	3.64	45.45
PFOS	1763-23-1	10.97		0.17	0.45	4.55
Surrogate Recoveries (%)						
13C5-PFHxA		103	D			
13C4-PFHpA		104				
13C8-PFOA		110	D			
13C9-PFNA		75				
13C6-PFDA		78				
13C7-PFUnA		82				
13C2-PFDoA		84				
13C2-PFTeDA		94				
d3-MeFOSAA		72				
d5-EtFOSAA		73				
13C3-PFBS		126	D			
13C3-PFHxS		89	D			
13C8-PFOS		92				

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB10-10022018			
Battelle ID		J8621-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	375-85-9	0.44 U	0.14	0.44	4.39
PFOA	335-67-1	1.88 J	0.16	0.44	4.39
PFNA	375-95-1	0.88 U	0.23	0.88	4.39
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	2058-94-8	0.88 U	0.25	0.88	4.39
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	72629-94-8	0.44 U	0.13	0.44	4.39
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	2355-31-9	1.75 U	0.49	1.75	4.39
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	375-73-5	0.44 U	0.11	0.44	4.39
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	1763-23-1	0.44 U	0.17	0.44	4.39

Surrogate Recoveries (%)

13C5-PFHxA	90
13C4-PFHpA	88
13C8-PFOA	93
13C9-PFNA	84
13C6-PFDA	84
13C7-PFUnA	85
13C2-PFDoA	92
13C2-PFTeDA	94
d3-MeFOSAA	68
d5-EtFOSAA	74
13C3-PFBS	80
13C3-PFHxS	77
13C8-PFOS	75

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB10-10022018			
Battelle ID		J8622-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	375-85-9	0.44 U	0.14	0.44	4.39
PFOA	335-67-1	1.11 J	0.16	0.44	4.39
PFNA	375-95-1	0.88 U	0.23	0.88	4.39
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	2058-94-8	0.88 U	0.25	0.88	4.39
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	72629-94-8	0.44 U	0.13	0.44	4.39
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	2355-31-9	1.75 U	0.49	1.75	4.39
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	375-73-5	0.44 U	0.11	0.44	4.39
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	1763-23-1	0.24 J	0.17	0.44	4.39

Surrogate Recoveries (%)

13C5-PFHxA	94
13C4-PFHpA	87
13C8-PFOA	94
13C9-PFNA	85
13C6-PFDA	83
13C7-PFUnA	86
13C2-PFDoA	85
13C2-PFTeDA	82
d3-MeFOSAA	64
d5-EtFOSAA	63
13C3-PFBS	86
13C3-PFHxS	84
13C8-PFOS	83

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW01-1018				
Battelle ID		J8623-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	59.15		0.18	0.46	4.63
PFHpA	375-85-9	18.77		0.15	0.46	4.63
PFOA	335-67-1	220.43	D	1.67	4.63	46.30
PFNA	375-95-1	2.52	J	0.24	0.93	4.63
PFDA	335-76-2	0.16	J	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	27.51		0.12	0.46	4.63
PFHxS	355-46-4	78.91	D	1.02	3.70	46.30
PFOS	1763-23-1	70.73		0.18	0.46	4.63

Surrogate Recoveries (%)

13C5-PFHxA	91	
13C4-PFHpA	103	
13C8-PFOA	115	D
13C9-PFNA	67	
13C6-PFDA	77	
13C7-PFUnA	82	
13C2-PFDoA	71	
13C2-PFTeDA	77	
d3-MeFOSAA	83	
d5-EtFOSAA	81	
13C3-PFBS	72	
13C3-PFHxS	99	D
13C8-PFOS	81	

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04-1018				
Battelle ID		J8624-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	153.03	D	2.88	7.58	75.76
PFHpA	375-85-9	114.46	D	2.42	7.58	75.76
PFOA	335-67-1	381.85	D	2.73	7.58	75.76
PFNA	375-95-1	9.77		0.24	0.91	4.55
PFDA	335-76-2	1.08	J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	51.02		0.12	0.45	4.55
PFHxS	355-46-4	717.18	D	1.67	6.06	75.76
PFOS	1763-23-1	345.23	D	2.88	7.58	75.76
Surrogate Recoveries (%)						
13C5-PFHxA		131	D			
13C4-PFHpA		129	D			
13C8-PFOA		104	D			
13C9-PFNA		62				
13C6-PFDA		91				
13C7-PFUnA		97				
13C2-PFDoA		98				
13C2-PFTeDA		100				
d3-MeFOSAA		109				
d5-EtFOSAA		110				
13C3-PFBS		92				
13C3-PFHxS		92	D			
13C8-PFOS		94	D			

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04P-1018				
Battelle ID		J8625-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	180.39	D	2.88	7.58	75.76
PFHpA	375-85-9	128.79	D	2.42	7.58	75.76
PFOA	335-67-1	405.69	D	2.73	7.58	75.76
PFNA	375-95-1	9.85		0.24	0.91	4.55
PFDA	335-76-2	0.99	J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	51.95		0.12	0.45	4.55
PFHxS	355-46-4	693.31	D	1.67	6.06	75.76
PFOS	1763-23-1	382.48	D	2.88	7.58	75.76
Surrogate Recoveries (%)						
13C5-PFHxA		126	D			
13C4-PFHpA		125	D			
13C8-PFOA		102	D			
13C9-PFNA		64				
13C6-PFDA		91				
13C7-PFUnA		94				
13C2-PFDoA		86				
13C2-PFTeDA		98				
d3-MeFOSAA		99				
d5-EtFOSAA		109				
13C3-PFBS		86				
13C3-PFHxS		109	D			
13C8-PFOS		98	D			

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW06-1018				
Battelle ID		J8626-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	124.65	D	1.73	4.55	45.45
PFHpA	375-85-9	17.96		0.15	0.45	4.55
PFOA	335-67-1	156.31	D	1.64	4.55	45.45
PFNA	375-95-1	0.61	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	91.23		0.12	0.45	4.55
PFHxS	355-46-4	66.27		0.10	0.36	4.55
PFOS	1763-23-1	2.04	J	0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	128	D
13C4-PFHpA	109	
13C8-PFOA	109	D
13C9-PFNA	72	
13C6-PFDA	81	
13C7-PFUnA	79	
13C2-PFDoA	71	
13C2-PFTeDA	74	
d3-MeFOSAA	103	
d5-EtFOSAA	89	
13C3-PFBS	83	
13C3-PFHxS	108	
13C8-PFOS	103	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW01-1018				
Battelle ID		J8627-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	303.39	D	4.32	11.36	113.64
PFHpA	375-85-9	45.61		0.15	0.45	4.55
PFOA	335-67-1	171.14	D	4.09	11.36	113.64
PFNA	375-95-1	1.83	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	107.88		0.12	0.45	4.55
PFHxS	355-46-4	1798.48	D	12.50	45.45	568.18
PFOS	1763-23-1	1808.55	D	4.32	11.36	113.64

Surrogate Recoveries (%)

13C5-PFHxA	118	D
13C4-PFHpA	91	
13C8-PFOA	109	D
13C9-PFNA	67	
13C6-PFDA	95	
13C7-PFUnA	96	
13C2-PFDoA	90	
13C2-PFTeDA	91	
d3-MeFOSAA	113	
d5-EtFOSAA	102	
13C3-PFBS	86	
13C3-PFHxS	112	D
13C8-PFOS	80	D

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW03-1018				
Battelle ID		J8628-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.280				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	0.45	U	0.17	0.45	4.46
PFHpA	375-85-9	0.45	U	0.14	0.45	4.46
PFOA	335-67-1	4.43	J	0.16	0.45	4.46
PFNA	375-95-1	0.89	U	0.23	0.89	4.46
PFDA	335-76-2	0.45	U	0.14	0.45	4.46
PFUnA	2058-94-8	0.89	U	0.26	0.89	4.46
PFDoA	307-55-1	0.45	U	0.16	0.45	4.46
PFTrDA	72629-94-8	0.45	U	0.13	0.45	4.46
PFTeDA	376-06-7	0.89	U	0.22	0.89	4.46
NMeFOSAA	2355-31-9	1.79	U	0.50	1.79	4.46
NEtFOSAA	2991-50-6	0.89	U	0.44	0.89	4.46
PFBS	375-73-5	0.45	U	0.12	0.45	4.46
PFHxS	355-46-4	1.58	J	0.10	0.36	4.46
PFOS	1763-23-1	2.19	J	0.17	0.45	4.46

Surrogate Recoveries (%)

13C5-PFHxA	69
13C4-PFHpA	96
13C8-PFOA	85
13C9-PFNA	84
13C6-PFDA	83
13C7-PFUnA	96
13C2-PFDoA	89
13C2-PFTeDA	76
d3-MeFOSAA	112
d5-EtFOSAA	97
13C3-PFBS	60
13C3-PFHxS	98
13C8-PFOS	83

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB			
Battelle ID		KB80 IB_10/17/2018			
Sample Type		IB			
Collection Date		NA			
Extraction Date		NA			
Analysis Date		10/17/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		Water			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.50 U	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	98
13C4-PFHpA	98
13C8-PFOA	101
13C9-PFNA	100
13C6-PFDA	103
13C7-PFUnA	102
13C2-PFDoA	98
13C2-PFTeDA	93
d3-MeFOSAA	104
d5-EtFOSAA	96
13C3-PFBS	95
13C3-PFHxS	103
13C8-PFOS	96

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB			
Battelle ID		KB80 IB_10/22/2018			
Sample Type		IB			
Collection Date		NA			
Extraction Date		NA			
Analysis Date		10/22/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		Water			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.50 U	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	113
13C4-PFHpA	110
13C8-PFOA	107
13C9-PFNA	101
13C6-PFDA	103
13C7-PFUnA	104
13C2-PFDoA	109
13C2-PFTeDA	123
d3-MeFOSAA	113
d5-EtFOSAA	98
13C3-PFBS	96
13C3-PFHxS	95
13C8-PFOS	97

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB			
Battelle ID		KB80 IB_10/24/2018			
Sample Type		IB			
Collection Date		NA			
Extraction Date		NA			
Analysis Date		10/24/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		Water			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.50 U	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	106
13C4-PFHpA	106
13C8-PFOA	98
13C9-PFNA	96
13C6-PFDA	98
13C7-PFUnA	108
13C2-PFDoA	114
13C2-PFTeDA	121
d3-MeFOSAA	107
d5-EtFOSAA	88
13C3-PFBS	103
13C3-PFHxS	99
13C8-PFOS	100

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		Procedural Blank			
Battelle ID		CR939PB-FS			
Sample Type		PB			
Collection Date		10/10/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		WATER			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.96 J	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	92
13C4-PFHpA	96
13C8-PFOA	93
13C9-PFNA	88
13C6-PFDA	90
13C7-PFUnA	86
13C2-PFDoA	88
13C2-PFTeDA	94
d3-MeFOSAA	91
d5-EtFOSAA	75
13C3-PFBS	94
13C3-PFHxS	88
13C8-PFOS	82

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		Laboratory Control Sample					
Battelle ID		CR940LCS-FS					
Sample Type		LCS					
Collection Date		10/10/2018					
Extraction Date		10/10/2018					
Analysis Date		10/24/2018					
Analytical Instrument		Sciex 5500 LC/MS/MS					
% Moisture		NA					
Matrix		WATER					
Sample Size		0.250					
Size Unit-Basis		L				Control Limits	
Units		ng/L	Target	Recovery	Qual	Lower	Upper
PFHxA	307-24-4	35.34	30.30	117		51	137
PFHpA	375-85-9	32.54	30.00	108		48	136
PFOA	335-67-1	34.14	30.00	114		49	141
PFNA	375-95-1	33.30	30.00	111		58	122
PFDA	335-76-2	32.09	30.00	107		59	135
PFUnA	2058-94-8	35.16	30.00	117		64	134
PFDoA	307-55-1	32.59	30.00	109		75	131
PFTrDA	72629-94-8	33.59	30.00	112		42	148
PFTeDA	376-06-7	34.57	30.00	115		42	158
NMeFOSAA	2355-31-9	38.71	30.00	129		50	146
NEtFOSAA	2991-50-6	32.80	30.00	109		51	131
PFBS	375-73-5	29.43	30.30	97		56	134
PFHxS	355-46-4	31.94	30.30	105		52	128
PFOS	1763-23-1	29.90	30.00	100		40	144

Surrogate Recoveries (%)

13C5-PFHxA	97
13C4-PFHpA	92
13C8-PFOA	86
13C9-PFNA	92
13C6-PFDA	90
13C7-PFUnA	88
13C2-PFDoA	106
13C2-PFTeDA	107
d3-MeFOSAA	94
d5-EtFOSAA	87
13C3-PFBS	117
13C3-PFHxS	96
13C8-PFOS	98

BATTELLE
It can be done

Glossary of Data Qualifiers

Flag:	Application:
B	Analyte found in the sample at a concentration <10x the level found in the procedural blank
D	Dilution Run. Initial run outside the initial calibration range of the instrument
E	Estimate, result is greater than the highers concentration level in the calibration
H	Surrogate diluted out. Used when surrogate recovery is affected by excessive dilution of the sample extract.
J	Analyte detected below the Limit of Quantitation (LOQ)
ME	Significant Matrix Interference - Estimated value.
MI	Significant Matrix Interference - value could not be determined.
n	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO), but meets secondary criteria
N	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO)
NA	Not Applicable
T	Holding Time (HT) exceeded
U	Analyte not detected or detected below the Method detection limit (MDL) value, Limit of Detection (LOD) reported

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

AQ, GW

Batch 18-0602

Package DP-18-0303

Submitted to:
CH2M
1100 NE Circle Blvd Suite 300
Corvallis, OR 97330 USA

Submitted by:
Battelle Norwell Operations
141 Longwater Drive Suite 202
Norwell, MA 02061

BATTELLE
It can be done

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

AQ, GW

Batch 18-0602

Package DP-18-0303

Submitted to:
CH2M
1100 NE Circle Blvd Suite 300
Corvallis, OR 97330 USA

NELAP Accreditation Number: E87856 (Florida Department of Health)
DoD-ELAP Accreditation Number: 91667

Submitted by:
Battelle Norwell Operations
141 Longwater Drive Suite 202
Norwell, MA 02061

Analyst Approval: Lauren Griffith 2018.10.25 14:21:15 -04'00'

QC Chemist Approval: fitche@battelle.org 2018.11.01 15:39:03 -04'00'

Project Manager Approval: Digitally signed by Jonathan Thorn Date: 2018.11.02 05:55:02 -04'00'

BATTELLE
It can be done

CTO-4164 Naval Base Ventura County, California
Project No 100110125-01
PFAS by DoD QSM 5.1 Table B-15

AQ, GW

Batch 18-0602

Package DP-18-0303

1	***Work Plan*** Laboratory Work Plan, Addendums To Work Plan, Memos From Project Manager, Special Instructions, Chain-of-Custody Reports.	1
2	***Tables*** Analytical Data Tables, Qualifier Definitions.	24
3	***Miscellaneous Documentation*** Case Narrative, Miscellaneous Documentation Form, Quality Control Summary, Example Calculations, Internal Standard Recovery Report, Retention Time Window Report.	47
4	***Sample Preparation Records*** Sample Preparation Records, Dilution Worksheets, Standard Preparation Records, Certificates Of Analysis, GPC Check Report.	198
5	***Analytical Calibrations*** Analytical Sequence, Analytical Method, Tune Report, Initial Calibration, Pesticide Degradation Report, RF Summary, Calibration Verifications, Independent Calibration Verification Check.	217
6	***Analytical Data*** Raw Data Quantification Reports.	357
7	***Chromatograms*** Sample And Standard Chromatograms.	533
8	***Unused Data***	784

BATTELLE

It can be done

BATTELLE

Signature Page

Battelle 2018 (1 of 2)
Signature Page

Name (Printed)	Signature	Initials	Date
Jonathan Thorn		JRT	4/4/2018
Robert Lizotte, Jr.		BL	4.4 2018
FRANCO PALA			4-4-2018
Carla Devine		CRD	4/4/18
Denise Schumitz		DMS	4/4/18
Candice Penn McCoy		CPM	4/4/2018
Rich Rostucci		RR	4/4/2018
Michael Mears			4/4/2018
Christie Usher		cu	4/4/18
Kevin McInerney			4/4/18
Stephanie Schultz		SAS	4/4/18
Jordan Tower		JCT	4/4/18
KRISTEN NICHOLS		KN	4/4/18
Clinton H Brown		CB	4/4/18
Matt Schumitz		MS	4-4-18
Sam Guimaraes		SG	4-4-18
Lauren Griffith		LMG	4.4.18

BATTELLE

Signature Page

Battelle 2018 (2 of 2)
Signature Page

Name (Printed)	Signature	Initials	Date
KAVITHA DASU		KD	04/04/18
Kayla Lamarre		KAL	04/04/18
Weidong Li		W.L	04/04/18
Tracy W Stenner		TWS	04/04/18
Ellyn M Fitch		EMF	12-April-2018
Gail DeRuzzo		GD	4/18/18
Zachary Willenberg		ZJW	4/20/18

Sample Summary

Client: CH2M
SDG: 18-0602
Project/Site: Naval Base Ventura County
CTO: 4164

Lab Sample ID	Client Sample ID	Matrix	Collection Date	Receipt Date
CR939PB-FS	Procedural Blank	WATER	10/10/2018	10/10/2018
CR940LCS-FS	Laboratory Control Sample	WATER	10/10/2018	10/10/2018
J8613-FS	VC-PM553-GW0601A-1018	GW	10/1/2018	10/5/2018
J8614-FS	VC-AQ-FB09-10012018	AQ	10/1/2018	10/5/2018
J8615-FS	VC-AQ-EB09-10012018	AQ	10/1/2018	10/5/2018
J8616-FS	VC-PM13-GW21-1018	GW	10/1/2018	10/5/2018
J8617-FS	VC-KCHA16-GW01-1018	GW	10/2/2018	10/5/2018
J8618-FS	VC-KCHA16-GW01P-1018	GW	10/2/2018	10/5/2018
J8619-FS	VC-KCHA16-GW02-1018	GW	10/2/2018	10/5/2018
J8620-FS	VC-KCHA16-GW03-1018	GW	10/2/2018	10/5/2018
J8621-FS	VC-AQ-FB10-10022018	AQ	10/2/2018	10/5/2018
J8622-FS	VC-AQ-EB10-10022018	AQ	10/2/2018	10/5/2018
J8623-FS	VC-S09GW01-1018	GW	10/3/2018	10/5/2018
J8624-FS	VC-S09GW04-1018	GW	10/2/2018	10/5/2018
J8625-FS	VC-S09GW04P-1018	GW	10/2/2018	10/5/2018
J8626-FS	VC-S09GW06-1018	GW	10/3/2018	10/5/2018
J8627-FS	VC-S14GW01-1018	GW	10/3/2018	10/5/2018
J8627-FS	VC-S14GW01-1018	GW	10/3/2018	10/5/2018

Work Plan

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

1.0 GENERAL PROJECT INFORMATION

Project Title:	CTO-4164: Analysis of Non-Potable Waters
Project Number:	100110125-01
Client:	CH2M 1100 NE Circle Blvd Suite 300 Corvallis, OR 97330 USA
Client Contact Information:	Tiffany Hill Project Chemist (541) 768-3109(V) NA tiffany.hill@jacobs.com
Effective Date of QAPP:	9/11/2018
Version Number:	100110125-01(L)-02
Project Manager:	Thorn, Jonathan
Laboratory Task Manager:	Thorn, Jonathan
Deliverable Due Date:	10/19/2018

2.0 SCOPE OF WORK

Overview:	Analysis of non-potable water samples for PFAS.
Matrix:	Water

2.1 TECHNICAL APPROACH

2.1.1 Sample Receipt, Storage, and Handling

The list of samples for this project plan are presented in Attachment 1.

Storage Directions:	Store refrigerated.
Sub_Sampling:	None
Procedures:	NA
Contact:	NA
Comment:	None
Archiving:	Store remaining samples for six months after submission of final reports. Notify client prior to sample disposal.
Disposal:	Dispose of any remaining samples in the appropriate waste stream.

2.1.2 Sample Preparation

Samples will be batched as they arrive, with a 28-day TAT from receipt of samples.

Samples Expected:	Samples Per Batch:	Batches Expected:
150	20	8

Batch quality control samples are defined in Table 1.

Target samples are presented in Attachment 1.

Table 1: Quality Control Samples

Type:	*Description:*	*Count:*	*Rgt:*	*Reference:*	*Comment:*
PB	Laboratory control reagent blank.	1 per batch	--	NA	
LCS	Laboratory Control Sample	1 per batch	No	NA	
MS	Spiked field sample for determining method accuracy in the presence of matrix.	1 per batch	--	NA	Background for MS/MSD will be indicated on COC.
MSD	Spiked field sample for determining method accuracy and precision in the presence of matrix.	1 per batch	--	NA	Background for MS/MSD will be indicated on COC.

2.1.3 Extraction/Preparation

2.1.3.1 Extraction

SOP No.-Rev: **5-370-06**

SOP Title: *Extraction of Poly and Perfluoroalkyl Substances from Environmental Matrices*

Sample Size: 250 ml

SIS and LCS/MS Compounds: Defined in Table 2.

Deviations:
- Do not split samples post extract, PIV adjusted to compensate for removal of split.

Comments:
- All waters will be pre-screened per draft SOP 5-370-07.

Table 2: SIS and LCS/MS Spiking Level

Standard Type	Standard Contents	Spike Amount (ng)	Volume (uL)	Comment
PFAS - DoD Low Level Labelled Extracted Internal Standards (SIS)	JY28 SIS	~ 0.250 ng	50 uL	NA

Standard Type	Standard Contents		Spike Amount (ng)	Volume (uL)	Comment
PFAS - DoD Second Source LCS/MS Solution	JZ88	LCS/MS	~ 7.5 ng	150 uL	MS/MSD only - vary spike by batch 100, 150, 200, 250, 300 µL (pre-screening data will indicate if background sample is high level and needs a higher level fortification).
PFAS - DoD Second Source LCS/MS Solution	JZ88	LCS/MS	~ 2.50 ng	50 uL	LCS sample only - vary spike for each batch 50, 100, and 150 µL.

2.1.3.2 Cleanup

None.

RIS spiking levels are presented in Table 3.

Extract PIV (uL): 1000

Table 3: RIS Spiking Level

Standard Type	Standard Contents		Spike Amount (ng)	Volume (uL)	Comment
PFAS - DoD Internal Standard Spiking Solution	JY26	RIS	~ 0.250 ng	50 uL	NA

2.1.4 Instrumental Analysis

The list of analytes along with data quality criteria are presented in Attachment 2.

1) SOP_No-Rev: **5-369-06**

SOP_Title: *Analysis of Perfluoroalkyl Substances in Environmental Samples by Liquid Chromatography and Tandem Mass Spectrometry (LC-MS/MS)*

Deviations: None

Comments: None

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

2.2. DELIVERABLES

Deliverables Due:	10/19/2018
LIMS Reports:	*No*
Histograms:	*No*
Excel Tables:	*Yes*
EICs:	*No*
Chromatograms:	*No*
EDDs:	*Yes*
Comments:	• Individual data sets will be due 28 days after receipt of each sample set. • Full Level 4 data package (QSM 5.1 Table B-15 compliant) required. • SEDD file required. • weekly updates to client on status required.

3.0 QUALITY

The Method Quality Objectives are defined in Attachment 3.

4.0 ORGANIZATION AND COMMUNICATION

4.1 ORGANIZATION

The project team is defined in Table 4. Supervisors may make substitutions with Project Manager concurrence.

Table 4: Project Team and Roles

Staff Member	Role	Comment
Jonathan R. Thorn	Project Manager	NA
Stephanie A. Schultz	Sample Preparation	NA
Denise M. Schumitz	LC-MS/MS Analysis	NA
Matt D. Schumitz	Sample Custody	NA
Carla R. Devine	Quality Control Officer	NA
Zachary J. Willenberg	Quality Assurance Officer	NA

4.2 COMMUNICATION

A kick-off meeting will be held to discuss project scope and goals.

5.0 SCHEDULE

The project schedule is presented in Table 5.

Table 5. Schedule of Laboratory Activities

Activity:	Start Date:	End Date:	TAT (days):	Comment:
Sample Receipt	09/21/2018	09/21/2018	0	Timeline provided is an example based on samples arriving on 9/21. Schedule will shift for each delivery.
Sample Preparation	09/21/2018	09/25/2018	4	NA
Instrument Analysis	09/25/2018	10/05/2018	10	NA
Quality Control Review	10/05/2018	10/10/2018	5	NA
Quality Assurance Review	10/10/2018	10/15/2018	5	NA

6.0 BUDGET

The labor budget for the analytical task is presented in Table 6.

Table 6. Labor Budget (Laboratory Analytical Task)

Labor Activity:	Hours/ Batch:	Batches:	Total Hours:	Comment:
Sample Receipt	2	10	20	Hours are based on batches of 20 samples.
Sample Preparation	8	10	80	NA
Instrument Analysis	8	10	80	NA
Quality Control Review	3	10	30	NA
Quality Assurance Review	1	10	10	NA

7.0 STAFF DEVELOPMENT

None anticipated

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 1: Target Samples

Shipment: **SHP-180921-01**

Status: Pending

Description: NBVC Basewide SI

Range: J8210-J8277

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility: Location:	No:	Comments:
1	J8210	VC-PM3009-DW01-0918	09/17/2018 3:10 pm	GW	R0119 (NA)		
2	J8211	VC-PM3009-DW02-0918	09/17/2018 4:23 pm	GW	R0119 (NA)		
3	J8212	VC-PM3009-DW02P-0918	09/17/2018 4:27 pm	GW	R0119 (NA)		
4	J8213	VC-PM3009-DW03-0918	09/17/2018 4:38 pm	GW	R0119 (NA)		
5	J8214	VC-PM372-DW01-0918	09/18/2018 10:07 am	GW	R0119 (NA)		
6	J8215	VC-PM372-DW02-0918	09/18/2018 9:25 am	GW	R0119 (NA)		
7	J8216	VC-PM372-DW02P-0918	09/18/2018 9:27 am	GW	R0119 (NA)		
8	J8217	VC-PM372-DW03-0918	09/18/2018 11:49 am	GW	R0119 (NA)		
9	J8241	VC-PM649-DW01-0918	09/18/2018 12:50 pm	GW	R0119 (NA)		
10	J8242	VC-PM649-DW01P-0918	09/18/2018 12:55 pm	GW	R0119 (NA)		
11	J8243	VC-PM649-DW02-0918	09/18/2018 3:35 pm	GW	R0119 (NA)		
12	J8244	VC-PM649-DW03-0918	09/18/2018 2:02 pm	GW	R0119 (NA)		
13	J8245	VC-PM649-DW04-0918	09/18/2018 2:02 pm	GW	R0119 (NA)		
14	J8246	VC-AQ-FB01-0918	09/18/2018 1:30 pm	AQ	R0119 (NA)		
15	J8247	VC-AQ-EB01-0918	09/18/2018 1:40 pm	AQ	R0119 (NA)		
16	J8259	VC-PM365-DW01-0918	09/19/2018 11:10 am	GW	R0119 (NA)		
17	J8260	VC-PM365-DW02-0918	09/19/2018 12:10 pm	GW	R0119 (NA)		
18	J8261	VC-PM365-DW02P-0918	09/19/2018 12:18 pm	GW	R0119 (NA)		
19	J8262	VC-PM365-DW03-0918	09/19/2018 11:10 am	GW	R0119 (NA)		
20	J8272	VC-PM553-DW01-0918	09/19/2018 2:30 pm	GW	R0119 (NA)		
21	J8273	VC-PM553-DW01P-0918	09/19/2018 2:45 pm	GW	R0119 (NA)		
22	J8274	VC-PM553-DW02-0918	09/19/2018 3:15 pm	GW	R0119 (NA)		
23	J8275	VC-PM553-DW03-0918	09/19/2018 12:06 pm	GW	R0119 (NA)		
24	J8276	VC-SO-FB02-0918	09/19/2018 2:35 pm	AQ	R0119 (NA)		
25	J8277	VC-SO-EB02-0918	09/19/2018 2:30 pm	AQ	R0119 (NA)		

Shipment: **SHP-180925-02**

Status: Approved

Description: NBVC Basewide SI

Range: J8285-J8337

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility: Location:	No:	Comments:
1	J8285	VC-CS94-DW01-0918	09/21/2018 10:54 am	GW	R0119 (NA)		
2	J8286	VC-CS94-DW01P-0918	09/21/2018 10:58 am	GW	R0119 (NA)		
3	J8287	VC-CS94-DW02-0918	09/21/2018 11:50 am	GW	R0119 (NA)		
4	J8288	VC-CS94-DW03-0918	09/21/2018 11:08 am	GW	R0119 (NA)		
5	J8289	VC-CS94-DW04-0918	09/21/2018 12:14 pm	GW	R0119 (NA)		
6	J8290	VC-CS94-DW05-0918	09/21/2018 12:23 pm	GW	R0119 (NA)		
7	J8308	VC-SO-FB04-092118	09/21/2018 2:20 pm	AQ	R0119 (NA)		

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Shipment: **SHP-180925-02**

Status: Approved

Description: NBVC Basewide SI

Range: J8285-J8337

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
8	J8309	VC-SO-EB04-092118	09/21/2018 2:20 pm	AQ	R0119 (NA)			
9	J8319	VC-PM323-324-DW01-0918	09/20/2018 2:38 pm	GW	R0119 (NA)			
10	J8320	VC-PM323-324-DW01P-0918	09/20/2018 2:45 pm	GW	R0119 (NA)			
11	J8321	VC-PM323-324-DW02-0918	09/20/2018 11:41 am	GW	R0119 (NA)			
12	J8322	VC-PM323-324-DW03-0918	09/20/2018 10:35 am	GW	R0119 (NA)			
13	J8323	VC-AQ-FB03-092018	09/20/2018 8:00 am	AQ	R0119 (NA)			
14	J8324	VC-AQ-EB03-092018	09/20/2018 8:00 am	AQ	R0119 (NA)			
15	J8330	VC-PM323-DW02-0918	09/20/2018 3:13 pm	GW	R0119 (NA)			
16	J8331	VC-PM323-DW02P-0918	09/20/2018 4:02 pm	GW	R0119 (NA)			
17	J8337	VC-PM324-DW02-0918	09/20/2018 2:00 pm	GW	R0119 (NA)			

Shipment: **SHP-180927-02**

Status: Approved

Description: NBVC Basewide SI

Range: J8358-J8400

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J8358	VC-PM323-DW01-0918	09/24/2018 10:21 am	GW	R0119 (NA)			
2	J8359	VC-PM323-DW03-0918	09/24/2018 11:10 am	GW	R0119 (NA)			
3	J8368	VC-SO-FB05-092418	09/24/2018 1:30 pm	AQ	R0119 (NA)			
4	J8369	VC-SO-EB05-092418	09/24/2018 1:30 pm	AQ	R0119 (NA)			
5	J8370	VC-PM324-DW01-0918	09/24/2018 1:45 pm	GW	R0119 (NA)			
6	J8371	VC-PM324-DW01P-0918	09/24/2018 2:02 pm	GW	R0119 (NA)			
7	J8372	VC-PM324-DW03-0918	09/24/2018 11:15 am	GW	R0119 (NA)			
8	J8382	VC-PM64B-DW01-0918	09/24/2018 3:00 pm	GW	R0119 (NA)			
9	J8383	VC-PM64B-DW02-0918	09/24/2018 3:31 pm	GW	R0119 (NA)			
10	J8384	VC-PM64B-DW03-0918	09/24/2018 2:28 pm	GW	R0119 (NA)			
11	J8385	VC-PM64B-DW03P-0918	09/24/2018 2:35 pm	GW	R0119 (NA)			
12	J8395	VC-HS09-DW01-0918	09/25/2018 2:25 pm	GW	R0119 (NA)			
13	J8396	VC-HS09-DW02-0918	09/25/2018 3:03 pm	GW	R0119 (NA)			
14	J8397	VC-HS09-DW03-0918	09/25/2018 1:19 pm	GW	R0119 (NA)			
15	J8398	VC-AQ-FB-092518	09/25/2018 2:35 pm	AQ	R0119 (NA)			
16	J8399	VC-AQ-EB-092518	09/25/2018 2:30 pm	AQ	R0119 (NA)			
17	J8400	VC-HS09-DW03P-0918	09/25/2018 1:56 pm	GW	R0119 (NA)			

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Shipment: **SHP-180928-03**

Status: Pending

Description: NBVC Basewide SI

Range: J8455-J8483

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J8455	VC-SO-FB07-09262018	09/26/2018 11:43 am	AQ	R0118 (NA)			
2	J8456	VC-SO-EB07-09262018	09/26/2018 11:45 am	AQ	R0118 (NA)			
3	J8457	VC-MS09-DW01-0918	09/26/2018 10:16 am	GW	R0118 (NA)			
4	J8458	VC-MS09-DW02-0918	09/26/2018 10:05 am	GW	R0118 (NA)			
5	J8459	VC-MS09-DW03-0918	09/26/2018 11:10 am	GW	R0118 (NA)			
6	J8460	VC-MS09-DW04-0918	09/26/2018 12:22 pm	GW	R0118 (NA)			
7	J8461	VC-MS09-DW04P-0918	09/26/2018 12:20 pm	GW	R0118 (NA)			
8	J8462	VC-MS09-DW05-0918	09/26/2018 11:35 am	GW	R0118 (NA)			
9	J8477	VC-PM367-DW01-0918	09/27/2018 10:36 am	GW	R0119 (NA)			
10	J8478	VC-PM367-DW02-0918	09/27/2018 9:50 am	GW	R0119 (NA)			
11	J8479	VC-PM367-DW03-0918	09/27/2018 10:43 am	GW	R0119 (NA)			
12	J8480	VC-PM367-DW03P-0918	09/27/2018 10:46 am	GW	R0119 (NA)			
13	J8481	VC-PM367-DW04-0918	09/27/2018 10:04 am	GW	R0119 (NA)			
14	J8482	VC-AQ-FB08-09272018	09/27/2018 10:00 am	AQ	R0119 (NA)			
15	J8483	VC-AQ-EB08-09272018	09/27/2018 12:00 pm	AQ	R0119 (NA)			

Shipment: **SHP-181005-03**

Status: Approved

Description: NBVC Basewide SI

Range: J8613-J8628

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J8613	VC-PM553-GW0601A-1018	10/01/2018 11:47 am	GW	R0119 (NA)			
2	J8614	VC-AQ-FB09-10012018	10/01/2018 3:57 pm	AQ	R0119 (NA)			
3	J8615	VC-AQ-EB09-10012018	10/01/2018 4:00 pm	AQ	R0119 (NA)			
4	J8616	VC-PM13-GW21-1018	10/01/2018 2:33 pm	GW	R0119 (NA)			
5	J8617	VC-KCHA16-GW01-1018	10/02/2018 11:50 am	GW	R0119 (NA)			
6	J8618	VC-KCHA16-GW01P-1018	10/02/2018 11:54 am	GW	R0119 (NA)			
7	J8619	VC-KCHA16-GW02-1018	10/02/2018 10:15 am	GW	R0119 (NA)			
8	J8620	VC-KCHA16-GW03-1018	10/02/2018 1:18 pm	GW	R0119 (NA)			
9	J8621	VC-AQ-FB10-10022018	10/02/2018 4:00 pm	AQ	R0119 (NA)			
10	J8622	VC-AQ-EB10-10022018	10/02/2018 4:10 pm	AQ	R0119 (NA)			
11	J8623	VC-S09GW01-1018	10/03/2018 12:22 pm	GW	R0119 (NA)			
12	J8624	VC-S09GW04-1018	10/02/2018 4:25 pm	GW	R0119 (NA)			
13	J8625	VC-S09GW04P-1018	10/02/2018 4:29 pm	GW	R0119 (NA)			
14	J8626	VC-S09GW06-1018	10/03/2018 11:22 am	GW	R0119 (NA)			
15	J8627	VC-S14GW01-1018	10/03/2018 2:39 pm	GW	R0119 (NA)			
16	J8628	VC-S14GW03-1018	10/03/2018 3:35 pm	GW	R0119 (NA)			

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Shipment: **SHP-181009-01**

Status: Pending

Description: NBVC Basewide SI

Range: J8667-J8722

Comment: NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J8667	VC-CS12-DW01-1018	10/06/2018 8:44 am	GW	R0119 (NA)			
2	J8668	VC-CS12-DW01P-1018	10/06/2018 8:48 am	GW	R0119 (NA)			
3	J8669	VC-CS12-DW02-1018	10/06/2018 11:22 am	GW	R0119 (NA)			
4	J8670	VC-CS12-DW03-1018	10/06/2018 10:30 am	GW	R0119 (NA)			
5	J8683	VC-CS10-DW01-1018	10/06/2018 1:38 pm	GW	R0119 (NA)			
6	J8684	VC-CS10-DW02-1018	10/06/2018 10:30 am	GW	R0119 (NA)			
7	J8685	VC-CS10-DW02P-1018	10/06/2018 10:40 am	GW	R0119 (NA)			
8	J8686	VC-CS10-DW03-1018	10/06/2018 12:07 pm	GW	R0119 (NA)			
9	J8687	VC-CS10-DW04-1018	10/06/2018 10:55 am	GW	R0119 (NA)			
10	J8688	VC-CS18-DW01-1018	10/06/2018 3:18 pm	GW	R0119 (NA)			
11	J8714	VC-FB11-10062018	10/06/2018 6:34 pm	AQ	R0119 (NA)			
12	J8715	VC-EB11-10062018	10/06/2018 6:36 pm	AQ	R0119 (NA)			
13	J8716	VC-CS00-DW01-1018	10/06/2018 2:32 pm	GW	R0119 (NA)			
14	J8717	VC-CS00-DW02-1018	10/06/2018 4:25 pm	GW	R0119 (NA)			
15	J8718	VC-CS00-DW02P-1018	10/06/2018 4:27 pm	GW	R0119 (NA)			
16	J8719	VC-CS00-DW03-1018	10/06/2018 5:05 pm	GW	R0119 (NA)			
17	J8720	VC-CS00-DW04-1018	10/06/2018 1:30 pm	GW	R0119 (NA)			
18	J8721	VC-CS00-DW05-1018	10/06/2018 6:02 pm	GW	R0119 (NA)			
19	J8722	VC-CS00-DW06-1018	10/06/2018 5:40 pm	GW	R0119 (NA)			

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 2: Test Codes

Project Test Code Name:	Master_369
SOP Reference:	5-369 - Analysis of Perfluoroalkyl Substances in Environmental Samples by Liquid Chromatography and Tandem Mass Spectrometry (LC-MS/MS)
Description:	PFAS by DoD QSM 5.1 Table B-15
Matrix:	L - Liquid Samples, like water or sea water, prepared and anlyzed under the same class of detection limits.
Detection Limit Study:	5-369
Instrument:	LC-MS/MS
MQO Criteria	Universal_LC
Standard Report:	Standard Result Report

Method Specific Reporting				Holding Times (days)		Data Flags	
Result Units:	ng/L	***Unit Conversion:***	(none)	***Sample:***	14	***DL_Flag:***	U
Weight Basis:	LIQUID	***Result Format:***	Fixed Digits	***Frozen:***	14	***RL_Flag:***	J
Standard Basis:	SIS	***# of Figures/Digits:***	2	***Extract:***	28	***PB_Flag:***	B
Oil Weight Basis:	No	***Oil Weight Source:***	Oil Weight			***DIL_Flag:***	D
U-Value Substition:	U-Flag=MD	***Histograms:***	No			***HT_Flag:***	T
ECD_Reporting:	No						

No:	Analyte:	Report Name:	Type	RIS	SIS	Hidden:	Graph:
1	Perfluoro-n-hexanoic acid	PFHxA	T		13C5-PFHxA	No	No
2	Perfluoro-n-heptanoic Acid	PFHpA	T		13C4-PFHpA	No	No
3	Perfluoro-n-octanoic Acid	PFOA	T		13C8-PFOA	No	No
4	Perfluorononanoic Acid	PFNA	T		13C9-PFNA	No	No
5	Perfluoro-n-decanoic Acid	PFDA	T		13C6-PFDA	No	No
6	Perfluoro-n-undecanoic acid	PFUnA	T		13C7-PFUnA	No	No
7	Perfluoro-n-dodecanoic acid	PFDoA	T		13C2-PFDoA	No	No
8	Perfluoro-n-tridecanoic acid	PFTrDA	T		13C2-PFTeDA	No	No
9	Perfluoro-n-tetradecanoic acid	PFTeDA	T		13C2-PFTeDA	No	No
10	N-methylperfluoro-1-octanesulfonamidoacetic acid	NMeFOSAA	T		d3-MeFOSAA	No	No
11	N-ethylperfluoro-octanesulfonamidoacetic acid	NEtFOSAA	T		d5-EtFOSAA	No	No
12	Perfluoro-1-butanesulfonate	PFBS	T		13C3-PFBS	No	No
13	Perfluoro-1-hexanesulfonate	PFHxS	T		13C3-PFHxS	No	No
14	Perfluoro-1-octanesulfonate	PFOS	T		13C8-PFOS	No	No
1	13C5-PFHxA	13C5-PFHxA	SIS	13C2-PFOA		No	No
2	13C4-PFHpA	13C4-PFHpA	SIS	13C2-PFOA		No	No
3	13C8-PFOA	13C8-PFOA	SIS	13C2-PFOA		No	No
4	13C9-PFNA	13C9-PFNA	SIS	13C2-PFOA		No	No

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 2: Test Codes

Project Test Code Name: Master_369

No:	Analyte:	Report Name:	Type	RIS	SIS	Hidden:	Graph:
5	13C6-PFDA	13C6-PFDA	SIS	13C2-PFDA		No	No
6	13C7-PFUnA	13C7-PFUnA	SIS	13C2-PFDA		No	No
7	13C2-PFDoA	13C2-PFDoA	SIS	13C2-PFDA		No	No
8	13C2-PFTeDA	13C2-PFTeDA	SIS	13C2-PFDA		No	No
9	d3-MeFOSAA	d3-MeFOSAA	SIS	13C4-PFOS		No	No
10	d5-EtFOSAA	d5-EtFOSAA	SIS	13C4-PFOS		No	No
11	13C3-PFBS	13C3-PFBS	SIS	13C4-PFOS		No	No
12	13C3-PFHxS	13C3-PFHxS	SIS	13C4-PFOS		No	No
13	13C8-PFOS	13C8-PFOS	SIS	13C4-PFOS		No	No

Total Analytes: 27

Subtract Peaks:

None

Sum Peaks:

None

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 2: Test Codes

Project Test Code Name: Master_369

ICAL Acceptance Criteria:

Curve Fit:	Limit Mean(%):	Mean Qual:	Limit Ind.:	Ind. Qual:	Min Points:	Points Qual:	Comments:
Linear	NA	NA	0.99	N	5	N	y = Bx + C
Quadratic	NA	NA	0.99	N	6	N	y = Ax^2 + Bx + C

Continuing Calibration Verification Criteria:

CCV Name: 5-369

Frequency Hrs:	Mean PD(%):	Individual PD(%):	RIS/SIS RT Window (min):	Area Limit Low(%):	Area Limit High(%):	Comment:
12 (N)	30 (N)	30 (N)	0.04 (N)	-50	100 (N)	NA

Independent Calibration Verification:

ICC Name: 5-369

Mean PD Limit(%):	Ind. PD Limit(%):	RIS/SIS Window Limit (Secs):	Area Limit High(%):	Area Limit Low(%):	Comment:
30 (N)	30 (N)	0.04 (N)	-50	100 (N)	NA

Mass Discrimination Criteria:

None

Degredation Check Criteria:

None

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 3: Method Quality Objectives

MQO Application	***Universal_LC***		
MQO:	**Acceptance Criteria**	**Qual:**	**Corrective Action:**
Procedural Blank	Samples must be greater than five times the blank concentration (>5xPB).	B	Review with Project Manager; re-analyze or justify results in project records.
PB Measurement Quality Objective	Organic results in the Procedural Blank are less than 1/2 times the LOQ (<1/2xLOQ)	N	Review with Project Manager; re-analyze or justify results in project records.
Laboratory Control Sample	Recovery values 70-130%.	N	Review with project manager; re-analyze or justify reporting the results in project records.
Matrix Spike / Matrix Spike Duplicate Recovery	Organics 70-130%. Analyte concentration in MS/MSD must be greater than five times reported background concentration.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
	Organics Results in the Target is less than 5 times the Original	n	
Matrix Spike/Spike Duplicate Precision	Organics results less than 30% Relative Percent Difference (RPD). Analyte concentration in MS/MSD must be greater than five times reported background concentration.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
	Organics Results in the Target is less than 5 times the Original	n	
Standard Reference Material Accuracy	Organics Percent Difference less than 30% from a range of certified values on average. Analyte concentration must be greater than five times the Method Detection Limit (>5xMDL).	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
	Organics Results in the Target is less than 5 times the MDL	n	
Analytical Duplicate Precision	Organics results less than 30% Relative Percent Difference (RPD). Analyte concentration must be > 5x MDL.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
	Organics Results in the Original is less than 5 times the MDL	n	

BATTELLE
It can be done

WORK/QUALITY ASSURANCE PROJECT PLAN

Attachment 3: Method Quality Objectives

MQO Application	***Universal_LC***		
MQO:	**Acceptance Criteria**	**Qual:**	**Corrective Action:**
Analytical Triplicate Precision	Organics results less than 30% Relative Standard Deviation (RSD). Analyte concentration must be > 5x MDL.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
	Organics Results in the Original is less than 5 times the MDL	n	
Surrogate Compound Recovery	Recovery results between 50% and 150%.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
Control Oil	RPD < 30% for at least 90% of analytes	N	Results examined by project manager, task leader, or subcontractor lab manager. Reextraction, reanalysis, or justification documented.
Instrument Calibration	5-369-6: R-squared greater than or equal to 0.990		Results examined by project manager, task leader, or subcontractor lab manager. Reextraction, reanalysis, or justification documented.
Independent Calibration Check Solution	5-369-6: Individual PD less than or equal to 30%. Mean Percent Difference less than or equal to 30%.	N	Review with Project Manager; re-analyze or justify in project records.
Continuing Calibration Verification	5-369-6: Individual PD less than or equal to 30%. Mean Percent Difference less than or equal to 30%.	N	Review with Project Manager; re-analyze or justify in project records.

BATTELLE
It can be done

ShpNo SHP-181005-03

Battelle Project No: 0110125-01

Sample Receipt Form

Approved: ☑ Authorized ☐

Project Number: 695803 **Client:** CH2M

Received by: Schumitz, Matt **Date/Time Received:** Friday, October 05, 2018 10:30 AM

No. of Shipping Containers: 1

SHIPMENT

Method of Delivery: Commercial Carrier **Tracking Number:** Fed Ex

COC Forms: ☑ **Shipped with samples** ☐ **No Forms**

Cooler(s)/Box(es)

Cntr	Type	Tracking No.	Seal	Seal	Container	Therm.	Temp C	Smps
1 of 1	Cooler	7830 8987 6121	Custody Seals	Intact	Intact	Therm_1	1.0	16

Samples

Sample Labels:
☑ Sample labels agree with COC forms
☐ Discrepancies (see Sample Custody Corrective Action Form)

Container Seals:
☐ Tape ☐ Custody Seals ☐ Other Seals (See sample Log)
☑ Seals intact for each shipping container
☐ Seals broken (See sample log for impacted samples)

Condition of Samples:
☑ Sample containers intact
☐ Sample containers broken/leaking (See Custody Corrective Action Form)

Temperature upon receipt (°C): 1 Temperature Blank used ☑ Yes ☐ No
(Note: If temperature upon receipt differs from required conditions, see sample log comment field)

Samples Acidified: ☐ Yes ☐ No ☑ Unknown

Initial pH 5-9?: ☐ Yes ☐ No ☑ NA
If no, individual sample adjustments on the Auxiliary Sample Receipt Form

Total Residual Chlorine Present?: ☐ Yes ☐ No ☑ NA
If yes, individual sample adjustments on the Auxiliary Sample Receipt Form

Head Space <1% in samples for water VOC analysis: ☐ Yes ☐ No ☑ NA
Individual sample deviations noted on sample log

Samples Containers:
Samples returned in PC-grade jars: ☐ Yes ☐ No ☑ Unknown /Lot No.: Unknown

Storage Location: Custody: Refrigerator - R0119 (NA) **BDO IDs Assigned:** J8613 - J8628

Samples logged in by: Schumitz, Matt **Date/Time:** 10/05/2018 10:30 AM

Approved By: Fitch, Ellyn **Approved On:** 10/9/2018 8:26:00 AM

Authorized By: **Authorized On:**

BATTELLE
It can be done

ShpNo SHP-181005-03

Battelle Project No:)110125-01

Sample Receipt Form Details

Approved: ☐ Authorized ☐

Project Number: 695803 **Client:** CH2M

Received by: Schumitz, Matt **Date/Time Received:** Friday, October 05, 2018 10:30 AM

No. of Shipping Containers: 1

BDO Id:	Client Sample ID:	Collection Date:	Login Date:	Ctrs:	Matrix:	Temp:	pH:	TRC:	VOC:	Stored In:	Loc:	No:	Comments:
J8613	VC-PM553-GW0601A-1018	10/01/18 11:47	10/05/18 11:22	2	GW	1	NA	NA	NA	R0119 (NA)			
J8614	VC-AQ-FB09-10012018	10/01/18 15:57	10/05/18 11:22	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8615	VC-AQ-EB09-10012018	10/01/18 16:00	10/05/18 11:22	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8616	VC-PM13-GW21-1018	10/01/18 14:33	10/05/18 11:23	2	GW	1	NA	NA	NA	R0119 (NA)			
J8617	VC-KCHA16-GW01-1018	10/02/18 11:50	10/05/18 11:25	2	GW	1	NA	NA	NA	R0119 (NA)			
J8618	VC-KCHA16-GW01P-1018	10/02/18 11:54	10/05/18 11:28	2	GW	1	NA	NA	NA	R0119 (NA)			
J8619	VC-KCHA16-GW02-1018	10/02/18 10:15	10/05/18 11:30	2	GW	1	NA	NA	NA	R0119 (NA)			
J8620	VC-KCHA16-GW03-1018	10/02/18 13:18	10/05/18 11:30	2	GW	1	NA	NA	NA	R0119 (NA)			
J8621	VC-AQ-FB10-10022018	10/02/18 16:00	10/05/18 11:30	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8622	VC-AQ-EB10-10022018	10/02/18 16:10	10/05/18 11:31	2	AQ	1	NA	NA	NA	R0119 (NA)			
J8623	VC-S09GW01-1018	10/03/18 12:22	10/05/18 11:31	2	GW	1	NA	NA	NA	R0119 (NA)			
J8624	VC-S09GW04-1018	10/02/18 16:25	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8625	VC-S09GW04P-1018	10/02/18 16:29	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8626	VC-S09GW06-1018	10/03/18 11:22	10/05/18 11:32	2	GW	1	NA	NA	NA	R0119 (NA)			
J8627	VC-S14GW01-1018	10/03/18 14:39	10/05/18 11:38	2	GW	1	NA	NA	NA	R0119 (NA)			
J8628	VC-S14GW03-1018	10/03/18 15:35	10/05/18 11:38	2	GW	1	NA	NA	NA	R0119 (NA)			

Total Samples: 16

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: NBVC Point Mugu PM553	Site Information: Hangar PM553
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3629 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal X Priority ☐ RUSH ☐ Time Zone: Pacific	Analysis: PFAS by Method 537 Mod	Page# 1 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
J8613 VC-PM553-GW0601A-1018	10/1/18	1147	Grab	GW	2	X
~~VC-PM553-DW01~~ KDR			Grab	GW		X
~~VC-PM553-DW01P~~ KDR			Grab	GW		X
~~VC-PM553-DW02~~ KDR			Grab	GW		X
~~VC-PM553-DW03~~ KDR			Grab	GW		X
~~VC-PM553-DW04~~ KDR			Grab	GW		X
J8614 VC ~~FDT~~-AQ-FB 09-1001208	10/1/18	1557	Grab	AQ	2	X
J8615 VC ~~FDT~~-AQ-EB 09-1001208	10/1/18	1600	Grab	AQ	2	X

On water meter

VK

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200		Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PM372 NBVC Point Mugu	Site Information: Hangar PM372
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal X Priority ☐ RUSH ☐ Time Zone: Pacific	Analysis: PFAS by Method 537 Mod	Page# 2 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
J8616 VC-PM13-GW21-1018	10/1/18	1433	Grab	GW	2	X
~~VC-PM372-DW01-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW02-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW02P-~~			~~Grab~~	~~GW~~		X
~~VC-PM372-DW03-~~			~~Grab~~	~~GW~~		X
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~		X
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~		X

VKJ

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	NA	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: ACIC16/ KCHA16	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724.977.3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI	Normal ☒ Priority ☐ RUSH ☐	Analysis: PFAS by Method 537 Mod	Page# 3 of 5
Project No.: 695803	Time Zone: PST		

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod	
VC-KCHA16-GW01-101 8 J8617	10/2/18	1150	Grab	GW	2	X	
VC-KCHA16-GW01P-101 8 J8618	10/2/18	1154	Grab	GW	2	X	
VC-KCHA16-GW02-101 8 J8619	10/2/18	1015	Grab	GW	2	X	
VC-KCHA16-GW03-101 8 J8620	10/2/18	1318	Grab	GW	2	X	
VC ~~FDT~~-AQ-FB 10-10022018 J8621	10/2/18	1600	Grab	AQ	2	X	
VC ~~FDT~~-AQ-EB 10-10022018 J8622	10/2/18	1618	Grab	AQ	2	X	on pump tubing

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200		Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PH S09	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3028 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:	Preservative: NA	COC # 5
Project Name: NBVC Basewide SI Project No.: 695803	Normal ☒ Priority ☐ RUSH ☐ Time Zone: PST	Analysis: PFAS by Method 537 Mod	Page# 4 of 5

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	PFAS by Method 537 Mod
VC-S09GW01-1018 J8623	10/3/18	1222	Grab	GW	2	X
VC-S09GW04-1018 J8624	10/2/18	1625	Grab	GW	2	X
VC-S09GW04P-1018 J8625	10/2/18	1629	Grab	GW	2	X
~~VC-HS09-DW01-~~			~~Grab~~	~~GW~~		X
~~VC-HS09-DW02-~~			~~Grab~~	~~GW~~		X
~~VC-HS09-DW03-~~			~~Grab~~	~~GW~~		X
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~		X
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~		X
J8626 VC-S09GW06-1018	10/3/18	1122	Grab	GW	2	X

VB

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	MS	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:

BATTELLE It can be done

Chain-of-Custody

Client Contact Information	Project Manager: Eric Davis	Sampling Site: PH S14	Site Information:
Tiffany Hill 1100 NE Circle Blvd, Suite 300 Corvallis, OR 97330	Sampler Information (print name): V. Kilbert Phone: 724-977-3628 Email: victoria.kilbert@jacobs.com Turnaround Time (TAT) Requested:		
Project Name: NBVC Basewide SI Project No.: 655803	Normal ☒ Priority ☐ RUSH ☐ Time Zone: PST		

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.	Preservative: NA	Analysis: PFAS by Method 537 Mod	COC # / Page#
J8627 VC-S14GW01-1018	10/3/18	1439	Grab	GW	2		X	COC # 5
~~VC-S14GW02-~~			~~Grab~~	~~GW~~			X	Page# 5 of 5
~~VC-S14GW02P-~~			~~Grab~~	~~GW~~			X	
J8628 VC-S14GW03-1018	10/3/18	1535	Grab	GW	2		X	
~~VC-S14GW19-~~			~~Grab~~	~~GW~~			X	
~~FDT-AQ-FB-~~			~~Grab~~	~~AQ~~			X	
~~FDT-AQ-EB-~~			~~Grab~~	~~AQ~~			X	

VK

Receipt Temperature:(°C)	Samples Intact: Yes - No	Samples on Ice: Yes - No	Receipt Comments:

Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
V. Kilbert	Jacobs	10/4/18 1200	NB	Battelle	10-5-18 1030
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:
Relinquished by (Print/Sign):	Company:	Date/Time:	Received by (Print/Sign):	Company:	Date/Time:

Comments:



ORIGIN ID:OXRA (724) 977-3628
VICTORIA KILBERTS

402 W BROADWAY STE 1450

SAN DIEGO, CA 92101
UNITED STATES US

SHIP DATE: 040CT18
ACTWGT: 51.80 LB
CAD: 6997686/SSF01922
DIMS: 25x14x13 IN

BILL THIRD PARTY

TO JOHNATHON THORNE
BATELLE LABS
141 LONGWATER DR
SUITE 202
NORWELL MA 02061

(781) 681-5665 REF: 695803
INV:
PO: DEPT:

1.0 TB
Therm 1
10/5/18
10:30

FedEx Express

REL# 3785346

TRK# 0201 7830 8987 6121

FRI - 05 OCT 10:30A
PRIORITY OVERNIGHT

XE XPUA

02061
MA-US BOS

Data Tables

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM553-GW0601A-1018				
Battelle ID		J8613-FS				
Sample Type		SA				
Collection Date		10/01/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	7602.64	D	29.32	77.16	771.61
PFHpA	375-85-9	3916.20	D	24.69	77.16	771.61
PFOA	335-67-1	6727.83	D	27.78	77.16	771.61
PFNA	375-95-1	5.37		0.24	0.93	4.63
PFDA	335-76-2	0.46	U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	305.91	D	6.02	23.15	231.48
PFHxS	355-46-4	2350.63	D	5.09	18.52	231.48
PFOS	1763-23-1	110.92	D	1.76	4.63	46.30
Surrogate Recoveries (%)						
13C5-PFHxA		109	D			
13C4-PFHpA		109	D			
13C8-PFOA		109	D			
13C9-PFNA		106	D			
13C6-PFDA		84				
13C7-PFUnA		93				
13C2-PFDoA		87				
13C2-PFTeDA		86				
d3-MeFOSAA		89				
d5-EtFOSAA		82				
13C3-PFBS		120	D			
13C3-PFHxS		97	D			
13C8-PFOS		105	D			

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB09-10012018			
Battelle ID		J8614-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.290			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.43 U	0.16	0.43	4.31
PFHpA	375-85-9	0.43 U	0.14	0.43	4.31
PFOA	335-67-1	0.91 J	0.16	0.43	4.31
PFNA	375-95-1	0.86 U	0.22	0.86	4.31
PFDA	335-76-2	0.43 U	0.14	0.43	4.31
PFUnA	2058-94-8	0.86 U	0.25	0.86	4.31
PFDoA	307-55-1	0.43 U	0.16	0.43	4.31
PFTrDA	72629-94-8	0.43 U	0.13	0.43	4.31
PFTeDA	376-06-7	0.86 U	0.22	0.86	4.31
NMeFOSAA	2355-31-9	1.72 U	0.48	1.72	4.31
NEtFOSAA	2991-50-6	0.86 U	0.42	0.86	4.31
PFBS	375-73-5	0.43 U	0.11	0.43	4.31
PFHxS	355-46-4	0.34 U	0.09	0.34	4.31
PFOS	1763-23-1	0.43 U	0.16	0.43	4.31

Surrogate Recoveries (%)

13C5-PFHxA	85
13C4-PFHpA	83
13C8-PFOA	86
13C9-PFNA	74
13C6-PFDA	94
13C7-PFUnA	96
13C2-PFDoA	86
13C2-PFTeDA	93
d3-MeFOSAA	75
d5-EtFOSAA	83
13C3-PFBS	89
13C3-PFHxS	90
13C8-PFOS	88

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB09-10012018				
Battelle ID		J8615-FS				
Sample Type		SA				
Collection Date		10/01/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		AQ				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	0.45	U	0.17	0.45	4.55
PFHpA	375-85-9	0.45	U	0.15	0.45	4.55
PFOA	335-67-1	0.95	J	0.16	0.45	4.55
PFNA	375-95-1	0.91	U	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	0.45	U	0.12	0.45	4.55
PFHxS	355-46-4	0.36	U	0.10	0.36	4.55
PFOS	1763-23-1	0.45	U	0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	95
13C4-PFHpA	83
13C8-PFOA	91
13C9-PFNA	82
13C6-PFDA	84
13C7-PFUnA	87
13C2-PFDoA	85
13C2-PFTeDA	85
d3-MeFOSAA	71
d5-EtFOSAA	58
13C3-PFBS	73
13C3-PFHxS	78
13C8-PFOS	77

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM13-GW21-1018				
Battelle ID		J8616-FS				
Sample Type		SA				
Collection Date		10/01/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	245.61	D	2.88	7.58	75.76
PFHpA	375-85-9	48.74		0.15	0.45	4.55
PFOA	335-67-1	63.52		0.16	0.45	4.55
PFNA	375-95-1	1.00	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	397.55	D	1.97	7.58	75.76
PFHxS	355-46-4	755.99	D	1.67	6.06	75.76
PFOS	1763-23-1	94.92	D	2.88	7.58	75.76

Surrogate Recoveries (%)

13C5-PFHxA	117	D
13C4-PFHpA	99	
13C8-PFOA	79	
13C9-PFNA	74	
13C6-PFDA	90	
13C7-PFUnA	96	
13C2-PFDoA	93	
13C2-PFTeDA	99	
d3-MeFOSAA	99	
d5-EtFOSAA	102	
13C3-PFBS	128	D
13C3-PFHxS	99	D
13C8-PFOS	99	D

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01-1018				
Battelle ID		J8617-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	338.20	D	1.73	4.55	45.45
PFHpA	375-85-9	13.04		0.15	0.45	4.55
PFOA	335-67-1	6.64	B	0.16	0.45	4.55
PFNA	375-95-1	0.91	U	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	153.20	D	1.18	4.55	45.45
PFHxS	355-46-4	177.76	D	1.00	3.64	45.45
PFOS	1763-23-1	8.94		0.17	0.45	4.55
Surrogate Recoveries (%)						
13C5-PFHxA		114	D			
13C4-PFHpA		121				
13C8-PFOA		89				
13C9-PFNA		82				
13C6-PFDA		88				
13C7-PFUnA		96				
13C2-PFDoA		91				
13C2-PFTeDA		99				
d3-MeFOSAA		91				
d5-EtFOSAA		71				
13C3-PFBS		139	D			
13C3-PFHxS		98	D			
13C8-PFOS		100				

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01P-1018				
Battelle ID		J8618-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.265				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	304.87	D	1.79	4.72	47.17
PFHpA	375-85-9	13.41		0.15	0.47	4.72
PFOA	335-67-1	5.94	B	0.17	0.47	4.72
PFNA	375-95-1	0.94	U	0.25	0.94	4.72
PFDA	335-76-2	0.47	U	0.15	0.47	4.72
PFUnA	2058-94-8	0.94	U	0.27	0.94	4.72
PFDoA	307-55-1	0.47	U	0.17	0.47	4.72
PFTrDA	72629-94-8	0.47	U	0.14	0.47	4.72
PFTeDA	376-06-7	0.94	U	0.24	0.94	4.72
NMeFOSAA	2355-31-9	1.89	U	0.53	1.89	4.72
NEtFOSAA	2991-50-6	0.94	U	0.46	0.94	4.72
PFBS	375-73-5	124.26	D	1.23	4.72	47.17
PFHxS	355-46-4	147.29	D	1.04	3.77	47.17
PFOS	1763-23-1	7.50		0.18	0.47	4.72
Surrogate Recoveries (%)						
13C5-PFHxA		103	D			
13C4-PFHpA		117				
13C8-PFOA		93				
13C9-PFNA		76				
13C6-PFDA		84				
13C7-PFUnA		87				
13C2-PFDoA		75				
13C2-PFTeDA		84				
d3-MeFOSAA		76				
d5-EtFOSAA		62				
13C3-PFBS		126	D			
13C3-PFHxS		85	D			
13C8-PFOS		94				

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW02-1018				
Battelle ID		J8619-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	163.76	D	1.76	4.63	46.30
PFHpA	375-85-9	39.13		0.15	0.46	4.63
PFOA	335-67-1	39.85		0.17	0.46	4.63
PFNA	375-95-1	0.93	U	0.24	0.93	4.63
PFDA	335-76-2	0.46	U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	49.17		0.12	0.46	4.63
PFHxS	355-46-4	250.99	D	1.02	3.70	46.30
PFOS	1763-23-1	17.94		0.18	0.46	4.63

Surrogate Recoveries (%)

13C5-PFHxA	111	D
13C4-PFHpA	122	
13C8-PFOA	98	
13C9-PFNA	82	
13C6-PFDA	77	
13C7-PFUnA	82	
13C2-PFDoA	76	
13C2-PFTeDA	80	
d3-MeFOSAA	72	
d5-EtFOSAA	71	
13C3-PFBS	101	
13C3-PFHxS	87	D
13C8-PFOS	79	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW03-1018				
Battelle ID		J8620-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	229.43	D	1.73	4.55	45.45
PFHpA	375-85-9	32.09		0.15	0.45	4.55
PFOA	335-67-1	66.04	D	1.64	4.55	45.45
PFNA	375-95-1	0.91	U	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	69.78	D	1.18	4.55	45.45
PFHxS	355-46-4	136.86	D	1.00	3.64	45.45
PFOS	1763-23-1	10.97		0.17	0.45	4.55
Surrogate Recoveries (%)						
13C5-PFHxA		103	D			
13C4-PFHpA		104				
13C8-PFOA		110	D			
13C9-PFNA		75				
13C6-PFDA		78				
13C7-PFUnA		82				
13C2-PFDoA		84				
13C2-PFTeDA		94				
d3-MeFOSAA		72				
d5-EtFOSAA		73				
13C3-PFBS		126	D			
13C3-PFHxS		89	D			
13C8-PFOS		92				

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB10-10022018			
Battelle ID		J8621-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	375-85-9	0.44 U	0.14	0.44	4.39
PFOA	335-67-1	1.88 J	0.16	0.44	4.39
PFNA	375-95-1	0.88 U	0.23	0.88	4.39
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	2058-94-8	0.88 U	0.25	0.88	4.39
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	72629-94-8	0.44 U	0.13	0.44	4.39
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	2355-31-9	1.75 U	0.49	1.75	4.39
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	375-73-5	0.44 U	0.11	0.44	4.39
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	1763-23-1	0.44 U	0.17	0.44	4.39

Surrogate Recoveries (%)

13C5-PFHxA	90
13C4-PFHpA	88
13C8-PFOA	93
13C9-PFNA	84
13C6-PFDA	84
13C7-PFUnA	85
13C2-PFDoA	92
13C2-PFTeDA	94
d3-MeFOSAA	68
d5-EtFOSAA	74
13C3-PFBS	80
13C3-PFHxS	77
13C8-PFOS	75

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB10-10022018			
Battelle ID		J8622-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	375-85-9	0.44 U	0.14	0.44	4.39
PFOA	335-67-1	1.11 J	0.16	0.44	4.39
PFNA	375-95-1	0.88 U	0.23	0.88	4.39
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	2058-94-8	0.88 U	0.25	0.88	4.39
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	72629-94-8	0.44 U	0.13	0.44	4.39
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	2355-31-9	1.75 U	0.49	1.75	4.39
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	375-73-5	0.44 U	0.11	0.44	4.39
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	1763-23-1	0.24 J	0.17	0.44	4.39

Surrogate Recoveries (%)

13C5-PFHxA	94
13C4-PFHpA	87
13C8-PFOA	94
13C9-PFNA	85
13C6-PFDA	83
13C7-PFUnA	86
13C2-PFDoA	85
13C2-PFTeDA	82
d3-MeFOSAA	64
d5-EtFOSAA	63
13C3-PFBS	86
13C3-PFHxS	84
13C8-PFOS	83

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW01-1018				
Battelle ID		J8623-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.270				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	59.15		0.18	0.46	4.63
PFHpA	375-85-9	18.77		0.15	0.46	4.63
PFOA	335-67-1	220.43	D	1.67	4.63	46.30
PFNA	375-95-1	2.52	J	0.24	0.93	4.63
PFDA	335-76-2	0.16	J	0.15	0.46	4.63
PFUnA	2058-94-8	0.93	U	0.27	0.93	4.63
PFDoA	307-55-1	0.46	U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46	U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93	U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85	U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93	U	0.45	0.93	4.63
PFBS	375-73-5	27.51		0.12	0.46	4.63
PFHxS	355-46-4	78.91	D	1.02	3.70	46.30
PFOS	1763-23-1	70.73		0.18	0.46	4.63

Surrogate Recoveries (%)

13C5-PFHxA	91	
13C4-PFHpA	103	
13C8-PFOA	115	D
13C9-PFNA	67	
13C6-PFDA	77	
13C7-PFUnA	82	
13C2-PFDoA	71	
13C2-PFTeDA	77	
d3-MeFOSAA	83	
d5-EtFOSAA	81	
13C3-PFBS	72	
13C3-PFHxS	99	D
13C8-PFOS	81	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04-1018				
Battelle ID		J8624-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	153.03	D	2.88	7.58	75.76
PFHpA	375-85-9	114.46	D	2.42	7.58	75.76
PFOA	335-67-1	381.85	D	2.73	7.58	75.76
PFNA	375-95-1	9.77		0.24	0.91	4.55
PFDA	335-76-2	1.08	J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	51.02		0.12	0.45	4.55
PFHxS	355-46-4	717.18	D	1.67	6.06	75.76
PFOS	1763-23-1	345.23	D	2.88	7.58	75.76
Surrogate Recoveries (%)						
13C5-PFHxA		131	D			
13C4-PFHpA		129	D			
13C8-PFOA		104	D			
13C9-PFNA		62				
13C6-PFDA		91				
13C7-PFUnA		97				
13C2-PFDoA		98				
13C2-PFTeDA		100				
d3-MeFOSAA		109				
d5-EtFOSAA		110				
13C3-PFBS		92				
13C3-PFHxS		92	D			
13C8-PFOS		94	D			

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04P-1018				
Battelle ID		J8625-FS				
Sample Type		SA				
Collection Date		10/02/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	180.39	D	2.88	7.58	75.76
PFHpA	375-85-9	128.79	D	2.42	7.58	75.76
PFOA	335-67-1	405.69	D	2.73	7.58	75.76
PFNA	375-95-1	9.85		0.24	0.91	4.55
PFDA	335-76-2	0.99	J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	51.95		0.12	0.45	4.55
PFHxS	355-46-4	693.31	D	1.67	6.06	75.76
PFOS	1763-23-1	382.48	D	2.88	7.58	75.76
Surrogate Recoveries (%)						
13C5-PFHxA		126	D			
13C4-PFHpA		125	D			
13C8-PFOA		102	D			
13C9-PFNA		64				
13C6-PFDA		91				
13C7-PFUnA		94				
13C2-PFDoA		86				
13C2-PFTeDA		98				
d3-MeFOSAA		99				
d5-EtFOSAA		109				
13C3-PFBS		86				
13C3-PFHxS		109	D			
13C8-PFOS		98	D			

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW06-1018				
Battelle ID		J8626-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	124.65	D	1.73	4.55	45.45
PFHpA	375-85-9	17.96		0.15	0.45	4.55
PFOA	335-67-1	156.31	D	1.64	4.55	45.45
PFNA	375-95-1	0.61	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	91.23		0.12	0.45	4.55
PFHxS	355-46-4	66.27		0.10	0.36	4.55
PFOS	1763-23-1	2.04	J	0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	128	D
13C4-PFHpA	109	
13C8-PFOA	109	D
13C9-PFNA	72	
13C6-PFDA	81	
13C7-PFUnA	79	
13C2-PFDoA	71	
13C2-PFTeDA	74	
d3-MeFOSAA	103	
d5-EtFOSAA	89	
13C3-PFBS	83	
13C3-PFHxS	108	
13C8-PFOS	103	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW01-1018				
Battelle ID		J8627-FS				
Sample Type		SA				
Collection Date		10/03/2018				
Extraction Date		10/10/2018				
Analysis Date		10/23/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		GW				
Sample Size		0.275				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	303.39	D	4.32	11.36	113.64
PFHpA	375-85-9	45.61		0.15	0.45	4.55
PFOA	335-67-1	171.14	D	4.09	11.36	113.64
PFNA	375-95-1	1.83	J	0.24	0.91	4.55
PFDA	335-76-2	0.45	U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91	U	0.26	0.91	4.55
PFDoA	307-55-1	0.45	U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45	U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91	U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82	U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91	U	0.45	0.91	4.55
PFBS	375-73-5	107.88		0.12	0.45	4.55
PFHxS	355-46-4	1798.48	D	12.50	45.45	568.18
PFOS	1763-23-1	1808.55	D	4.32	11.36	113.64
Surrogate Recoveries (%)						
13C5-PFHxA		118	D			
13C4-PFHpA		91				
13C8-PFOA		109	D			
13C9-PFNA		67				
13C6-PFDA		95				
13C7-PFUnA		96				
13C2-PFDoA		90				
13C2-PFTeDA		91				
d3-MeFOSAA		113				
d5-EtFOSAA		102				
13C3-PFBS		86				
13C3-PFHxS		112	D			
13C8-PFOS		80	D			

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW03-1018			
Battelle ID		J8628-FS			
Sample Type		SA			
Collection Date		10/03/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.280			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.45 U	0.17	0.45	4.46
PFHpA	375-85-9	0.45 U	0.14	0.45	4.46
PFOA	335-67-1	4.43 J	0.16	0.45	4.46
PFNA	375-95-1	0.89 U	0.23	0.89	4.46
PFDA	335-76-2	0.45 U	0.14	0.45	4.46
PFUnA	2058-94-8	0.89 U	0.26	0.89	4.46
PFDoA	307-55-1	0.45 U	0.16	0.45	4.46
PFTrDA	72629-94-8	0.45 U	0.13	0.45	4.46
PFTeDA	376-06-7	0.89 U	0.22	0.89	4.46
NMeFOSAA	2355-31-9	1.79 U	0.50	1.79	4.46
NEtFOSAA	2991-50-6	0.89 U	0.44	0.89	4.46
PFBS	375-73-5	0.45 U	0.12	0.45	4.46
PFHxS	355-46-4	1.58 J	0.10	0.36	4.46
PFOS	1763-23-1	2.19 J	0.17	0.45	4.46

Surrogate Recoveries (%)

13C5-PFHxA	69
13C4-PFHpA	96
13C8-PFOA	85
13C9-PFNA	84
13C6-PFDA	83
13C7-PFUnA	96
13C2-PFDoA	89
13C2-PFTeDA	76
d3-MeFOSAA	112
d5-EtFOSAA	97
13C3-PFBS	60
13C3-PFHxS	98
13C8-PFOS	83

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB				
Battelle ID		KB80 IB_10/17/2018				
Sample Type		IB				
Collection Date		NA				
Extraction Date		NA				
Analysis Date		10/17/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		Water				
Sample Size		0.250				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	0.50	U	0.19	0.50	5.00
PFHpA	375-85-9	0.50	U	0.16	0.50	5.00
PFOA	335-67-1	0.50	U	0.18	0.50	5.00
PFNA	375-95-1	1.00	U	0.26	1.00	5.00
PFDA	335-76-2	0.50	U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00	U	0.29	1.00	5.00
PFDoA	307-55-1	0.50	U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50	U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00	U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00	U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00	U	0.49	1.00	5.00
PFBS	375-73-5	0.50	U	0.13	0.50	5.00
PFHxS	355-46-4	0.40	U	0.11	0.40	5.00
PFOS	1763-23-1	0.50	U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	98
13C4-PFHpA	98
13C8-PFOA	101
13C9-PFNA	100
13C6-PFDA	103
13C7-PFUnA	102
13C2-PFDoA	98
13C2-PFTeDA	93
d3-MeFOSAA	104
d5-EtFOSAA	96
13C3-PFBS	95
13C3-PFHxS	103
13C8-PFOS	96

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB				
Battelle ID		KB80 IB_10/22/2018				
Sample Type		IB				
Collection Date		NA				
Extraction Date		NA				
Analysis Date		10/22/2018				
Analytical Instrument		Sciex 5500 LC/MS/MS				
% Moisture		NA				
Matrix		Water				
Sample Size		0.250				
Size Unit-Basis		L				
Units		ng/L		MDL	LOD	LOQ
PFHxA	307-24-4	0.50	U	0.19	0.50	5.00
PFHpA	375-85-9	0.50	U	0.16	0.50	5.00
PFOA	335-67-1	0.50	U	0.18	0.50	5.00
PFNA	375-95-1	1.00	U	0.26	1.00	5.00
PFDA	335-76-2	0.50	U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00	U	0.29	1.00	5.00
PFDoA	307-55-1	0.50	U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50	U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00	U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00	U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00	U	0.49	1.00	5.00
PFBS	375-73-5	0.50	U	0.13	0.50	5.00
PFHxS	355-46-4	0.40	U	0.11	0.40	5.00
PFOS	1763-23-1	0.50	U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	113
13C4-PFHpA	110
13C8-PFOA	107
13C9-PFNA	101
13C6-PFDA	103
13C7-PFUnA	104
13C2-PFDoA	109
13C2-PFTeDA	123
d3-MeFOSAA	113
d5-EtFOSAA	98
13C3-PFBS	96
13C3-PFHxS	95
13C8-PFOS	97

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		KB80 IB			
Battelle ID		KB80 IB_10/24/2018			
Sample Type		IB			
Collection Date		NA			
Extraction Date		NA			
Analysis Date		10/24/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		Water			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.50 U	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	106
13C4-PFHpA	106
13C8-PFOA	98
13C9-PFNA	96
13C6-PFDA	98
13C7-PFUnA	108
13C2-PFDoA	114
13C2-PFTeDA	121
d3-MeFOSAA	107
d5-EtFOSAA	88
13C3-PFBS	103
13C3-PFHxS	99
13C8-PFOS	100

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		Procedural Blank			
Battelle ID		CR939PB-FS			
Sample Type		PB			
Collection Date		10/10/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		WATER			
Sample Size		0.250			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.50 U	0.19	0.50	5.00
PFHpA	375-85-9	0.50 U	0.16	0.50	5.00
PFOA	335-67-1	0.96 J	0.18	0.50	5.00
PFNA	375-95-1	1.00 U	0.26	1.00	5.00
PFDA	335-76-2	0.50 U	0.16	0.50	5.00
PFUnA	2058-94-8	1.00 U	0.29	1.00	5.00
PFDoA	307-55-1	0.50 U	0.18	0.50	5.00
PFTrDA	72629-94-8	0.50 U	0.15	0.50	5.00
PFTeDA	376-06-7	1.00 U	0.25	1.00	5.00
NMeFOSAA	2355-31-9	2.00 U	0.56	2.00	5.00
NEtFOSAA	2991-50-6	1.00 U	0.49	1.00	5.00
PFBS	375-73-5	0.50 U	0.13	0.50	5.00
PFHxS	355-46-4	0.40 U	0.11	0.40	5.00
PFOS	1763-23-1	0.50 U	0.19	0.50	5.00

Surrogate Recoveries (%)

13C5-PFHxA	92
13C4-PFHpA	96
13C8-PFOA	93
13C9-PFNA	88
13C6-PFDA	90
13C7-PFUnA	86
13C2-PFDoA	88
13C2-PFTeDA	94
d3-MeFOSAA	91
d5-EtFOSAA	75
13C3-PFBS	94
13C3-PFHxS	88
13C8-PFOS	82

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		Laboratory Control Sample					
Battelle ID		CR940LCS-FS					
Sample Type		LCS					
Collection Date		10/10/2018					
Extraction Date		10/10/2018					
Analysis Date		10/24/2018					
Analytical Instrument		Sciex 5500 LC/MS/MS					
% Moisture		NA					
Matrix		WATER					
Sample Size		0.250					
Size Unit-Basis		L				Control Limits	
Units		ng/L	Target	Recovery	Qual	Lower	Upper
PFHxA	307-24-4	35.34	30.30	117		51	137
PFHpA	375-85-9	32.54	30.00	108		48	136
PFOA	335-67-1	34.14	30.00	114		49	141
PFNA	375-95-1	33.30	30.00	111		58	122
PFDA	335-76-2	32.09	30.00	107		59	135
PFUnA	2058-94-8	35.16	30.00	117		64	134
PFDoA	307-55-1	32.59	30.00	109		75	131
PFTrDA	72629-94-8	33.59	30.00	112		42	148
PFTeDA	376-06-7	34.57	30.00	115		42	158
NMeFOSAA	2355-31-9	38.71	30.00	129		50	146
NEtFOSAA	2991-50-6	32.80	30.00	109		51	131
PFBS	375-73-5	29.43	30.30	97		56	134
PFHxS	355-46-4	31.94	30.30	105		52	128
PFOS	1763-23-1	29.90	30.00	100		40	144

Surrogate Recoveries (%)

13C5-PFHxA	97
13C4-PFHpA	92
13C8-PFOA	86
13C9-PFNA	92
13C6-PFDA	90
13C7-PFUnA	88
13C2-PFDoA	106
13C2-PFTeDA	107
d3-MeFOSAA	94
d5-EtFOSAA	87
13C3-PFBS	117
13C3-PFHxS	96
13C8-PFOS	98

BATTELLE
It can be done

Glossary of Data Qualifiers

Flag:	Application:
B	Analyte found in the sample at a concentration <10x the level found in the procedural blank
D	Dilution Run. Initial run outside the initial calibration range of the instrument
E	Estimate, result is greater than the highers concentration level in the calibration
H	Surrogate diluted out. Used when surrogate recovery is affected by excessive dilution of the sample extract.
J	Analyte detected below the Limit of Quantitation (LOQ)
ME	Significant Matrix Interference - Estimated value.
MI	Significant Matrix Interference - value could not be determined.
n	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO), but meets secondary criteria
N	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO)
NA	Not Applicable
T	Holding Time (HT) exceeded
U	Analyte not detected or detected below the Method detection limit (MDL) value, Limit of Detection (LOD) reported

Miscellaneous Documentation

QA/QC Summary
Batch 18-0602

Project:	CTO-4164 Naval Base Ventura County, California
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	GW and AQ
Data Set:	DP-18-0303
Analytical SOP:	5-369
Method Reference:	PFAS to QSM 5.1 Table B-15

Sample Custody		
Collection Date	Receipt Date	Temp (°C)
10/ 1-3 /2018	10/5/2018	1.0
Corrective Actions	None.	
Sample Storage	The water samples were stored refrigerated until extraction.	
Related samples	NA	

	METHOD SUMMARIES
Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with 0.4% NH_3 in methanol. Extracts were concentrated to dryness under nitrogen with a water bath set between 35 °C and 45 °C, reconstituted with 80:20 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.
Prep comments	All samples were pre-screened prior to initial extraction.
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.
Analysis Comments	Samples analyzed on Sciex 5500 LC-MS/MS. When detected, PFHxS and PFOS contain both linear and branched isomers. There are instances of extracted internal standards (surrogates) that would fall outside of the 50%-150% recovery criteria if reported from the non-diluted extract. This is due to the higher levels of native PFOA and PFOS found in the samples interfering with the 13C2-PFOA and 13C4-PFOS internal standards used to quantify the surrogates. In these cases, the surrogates are reported from dilutions, however, the native compounds are reported from the non-diluted extracts as the interferences are with the internal standards and not the surrogates used to quantify the native compounds.

QA/QC Summary
Batch 18-0602

Holding Times	Extraction Date(s)	Analysis Date(s)
	10/10/2018	10/17,23,24 /2018

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
≤ ½ the LOQ Samples >10x PB	Two exceedances noted.
	PFOA was detected in two authentic samples at less than 10 times the amount detected in the PB, and is B qualified in these samples.

Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
Laboratory derived control limits for recovery	No exceedances noted.
	No comments.

Matrix Spike and Matrix Spike Duplicate (MS/MSD)	A MS/MSD was not prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
Laboratory derived control limits for recovery and <30% RPD	NA
	None prepared with this batch.

Extracted Internal Standard Analytes	Labelled analog compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
50-150% of true value	No exceedances noted.
	No comments.

Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
+/- 50% of the area of the L5 calibration point.	No exceedances noted.
	There are several instances where 13C2-PFOA and 13C4-PFOS would fail IS area criteria. These failures are due to the amount of native PFOA and PFOS found in the samples interfering with the internal standards in the non-diluted extracts. In these cases, the extracted internal standards that would use these two internal standards are reported from a dilution to mitigate the interferences from the high concentrations in the non-diluted samples. No data is reported from extracts where the IS areas would have failed criteria.

Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadradic curve fitting.
+/- 30% of true value, $R^2 \geq 0.99$	No exceedances noted.
	No comments.

QA/QC Summary
Batch 18-0602

Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
+/- 30% of true value	No exceedances noted.
	No comments.

Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.
+/- 30% of true value	No exceedances noted.
	No comments.

Instrument Blank (IB)	Immediately following the highest standard analyzed and daily prior to sample analysis.
≤ ½ the LOQ	No exceedances noted.
	No comments.

BATTELLE
It can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project Number: 100110125-01
Preparation Batch: 18-0602
Data Set: DP-18-0303
Test Code: Master_369

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	2	PFOA was detected in two authentic samples at less than 10 times the amount detected in the PB, and is B qualified in these samples.
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	0	None
Matrix Spike / Matrix Spike Duplicate Precision	0	None
Extracted Internal Standard Analytes (Surrogates)	0	None
Instrument Calibration	0	None
Instrument Blank	0	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
MISCELLANEOUS DOCUMENTATION FORM

Project Title:	CTO-4164 Naval Base Ventura County,	**Data Set Number:**	DP-18-0303
Project Number:	100110125-01	**Prep Batch Number:**	18-0602
Entered By:	Lauren Griffith	**Entered On:**	10/25/2018
Test Code (Matrix Type):	Master_369(L)		

Samples that were manually integrated are noted on the quant reports with the comment (TRUE).
LMG 10/25/18

PFOS appears to exhibit a retention time shift because it is a branched analyte and the retention time is defined by whichever branched peak has the largest area. The entire branched peak is picked in all cases.
LMG 10/25/18

In cases were native PFAS compounds were reported from dilutions (above calibration in non-diluted extracts), the extracted internal standard (surrogate) used to quantify the native compound was also reported from the dilution.
LMG 10/25/18

There are instances of extracted internal standards (surrogates) that would fall outside of the 50%-150% recovery criteria if reported from the non-diluted extract. This is due to the higher levels of native PFOA and PFOS found in the samples interfering with the 13C2-PFOA and 13C4-PFOS internal standards used to quantify the surrogates. In these cases, the surrogates are reported from dilutions, however, the native compounds are reported from the non-diluted extracts as the interferences are with the internal standards and not the surrogates used to quantify the native compounds.
LMG 10/25/18

Task Leader Approval:		
SupervisorApproval:		Robert Lizotte, Jr. 2018.10.26 11:06:41 -04'00'
PM Approval:		Robert Lizotte, Jr. 2018.10.26 11:06:55 -04'00'

BATTELLE

Example Calculation for PFAS

Calculation of final concentration from area:

$$Concentration = \left[\frac{PA - b}{m}\right] * C_{IS} * PIV * DF/S$$

Where:

PA = Area of target / area of internal standard
b = y intercept from calibration curve
CIS = concentration of internal standard (ng/L)
m = slope of calibration
DF = dilution factor
S = Sample Size
PIV = Pre-injection volume (L)

Sample ID:	J8613-FS-D(7)
Client Sample ID:	VC-PM553-GW0601A-1018
Sample Size:	0.27
Units:	L
Dilution Factor:	166.667
PIV (L):	0.001
Target Analyte:	PFHxA
MRM Transition:	313.0 / 269.0
Data file:	10222018_5500.wiff
Result table:	18-0602_BASE
Area:	3,031,716.32
IS Name:	13C5-PFHxA
IS Area:	62,304.44
IS Amount (ng/L):	250
y-intercept:	0.13333
slope:	0.98501

$$\text{Concentration} = \frac{[(3031716.32/62304.44) - 0.13333]}{0.98501} * 250 * 0.001 * 166.667 / 0.27$$

ng/L = 7,602.62

*Final concentration may vary based on rounding.

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batch: 18-0602
Data Set: DP-18-0303

		CR939PB-FS (Procedural Blank)	CR940LCS-FS (Laboratory Control Sample)	J8613-FS (VC-PM553-GW0601A-1018)	J8614-FS (VC-AQ-FB09-10012018)	J8615-FS (VC-AQ-EB09-10012018)	J8616-FS (VC-PM13-GW21-1018)
PFHxA	307-24-4	-	L	L	-	-	L
PFHpA	375-85-9	-	L	L	-	-	L
PFOA	335-67-1	-	L	L	L	L	L
PFNA	375-95-1	-	L	L	-	-	L
PFDA	335-76-2	-	L	-	-	-	-
PFUnA	2058-94-8	-	L	-	-	-	-
PFDoA	307-55-1	-	L	-	-	-	-
PFTrDA	72629-94-8	-	L	-	-	-	-
PFTeDA	376-06-7	-	L	-	-	-	-
NMeFOSAA	2355-31-9	-	L	-	-	-	-
NEtFOSAA	2991-50-6	-	L	-	-	-	-
PFBS	375-73-5	-	L	L	-	-	L
PFHxS	355-46-4	-	L/Br	L/Br	-	-	L/Br
PFOS	1763-23-1	-	L/Br	L/Br	-	-	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Branched
"-": Not detected

BATTE
It can

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batcl
Data Set: DP-18-

	J8617-FS (VC-KCHA16-GW01-1018)	J8618-FS (VC-KCHA16-GW01P-1018)	J8619-FS (VC-KCHA16-GW02-1018)	J8620-FS (VC-KCHA16-GW03-1018)	J8621-FS (VC-AQ-FB10-10022018)	J8622-FS (VC-AQ-EB10-10022018)	J8623-FS (VC-S09GW01-1018)	J8624-FS (VC-S09GW04-1018)
PFHxA	L	L	L	L	-	-	L	L
PFHpA	L	L	L	L	-	-	L	L
PFOA	L	L	L	L	L	L	L	L
PFNA	-	-	-	-	-	-	L	L
PFDA	-	-	-	-	-	-	L	L
PFUnA	-	-	-	-	-	-	-	-
PFDoA	-	-	-	-	-	-	-	-
PFTrDA	-	-	-	-	-	-	-	-
PFTeDA	-	-	-	-	-	-	-	-
NMeFOSAA	-	-	-	-	-	-	-	-
NEtFOSAA	-	-	-	-	-	-	-	-
PFBS	L	L	L	L	-	-	L	L
PFHxS	L/Br	L/Br	L/Br	L/Br	-	-	L/Br	L/Br
PFOS	L/Br	L/Br	L/Br	L/Br	-	L/Br	L/Br	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Bra
"-": Not detected

BATTE
It can

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01
Preparation Batcl
Data Set: DP-18-

	J8625-FS (VC-S09GW04P-1018)	J8626-FS (VC-S09GW06-1018)	J8627-FS (VC-S14GW01-1018)	J8628-FS (VC-S14GW03-1018)
PFHxA	L	L	L	-
PFHpA	L	L	L	-
PFOA	L	L	L	L
PFNA	L	L	L	-
PFDA	L	L	-	-
PFUnA	-	-	-	-
PFDoA	-	-	-	-
PFTrDA	-	-	-	-
PFTeDA	-	-	-	-
NMeFOSAA	-	-	-	-
NEtFOSAA	-	-	-	-
PFBS	L	L	L	-
PFHxS	L/Br	L/Br	L/Br	L/Br
PFOS	L/Br	L/Br	L/Br	L/Br

"L" :Linear
"Br": branched
"L/Br": Linear/Bra
"-": Not detected

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C2-PFOA	80,369.12	40,184.56	120,553.68

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C2-PFOA	79,095.63	40,184.56	120,553.68	
KB74	L2	10/17/18 19:57	13C2-PFOA	89,971.31	40,184.56	120,553.68	
KB75	L3	10/17/18 20:08	13C2-PFOA	87,799.30	40,184.56	120,553.68	
KB76	L4	10/17/18 20:19	13C2-PFOA	84,567.91	40,184.56	120,553.68	
KB77	L5	10/17/18 20:30	13C2-PFOA	80,369.12	40,184.56	120,553.68	
KB78	L6	10/17/18 20:41	13C2-PFOA	85,964.25	40,184.56	120,553.68	
KB79	L7	10/17/18 20:52	13C2-PFOA	86,636.81	40,184.56	120,553.68	
KB80 IB	IB	10/17/18 21:02	13C2-PFOA	85,730.94	40,184.56	120,553.68	
KB81 ICC	ICC	10/17/18 21:13	13C2-PFOA	85,242.44	40,184.56	120,553.68	
KB76 CCV	CCV	10/23/18 3:41	13C2-PFOA	73,507.58	40,184.56	120,553.68	
CR939PB-FS(0)	Procedural Blank	10/23/18 4:03	13C2-PFOA	82,215.13	40,184.56	120,553.68	
~~CR940LCS-FS(0)~~	~~Laboratory Control Sample~~	~~10/23/18 4:14~~	~~13C2-PFOA~~	~~88,182.45~~	~~40,184.56~~	~~120,553.68~~	
~~J8613-FS(0)~~	~~VC-PM553-GW0601A-1018~~	~~10/23/18 4:25~~	~~13C2-PFOA~~	~~266,029.78~~	~~40,184.56~~	~~120,553.68~~	~~N~~
J8613-FS-D(3)	VC-PM553-GW0601A-1018	10/23/18 4:36	13C2-PFOA	108,889.36	40,184.56	120,553.68	
J8613-FS-D(5)	VC-PM553-GW0601A-1018	10/23/18 4:46	13C2-PFOA	87,385.04	40,184.56	120,553.68	
J8613-FS-D(7)	VC-PM553-GW0601A-1018	10/23/18 4:57	13C2-PFOA	77,259.53	40,184.56	120,553.68	
J8614-FS(0)	VC-AQ-FB09-10012018	10/23/18 5:08	13C2-PFOA	81,605.91	40,184.56	120,553.68	
J8615-FS(0)	VC-AQ-EB09-10012018	10/23/18 5:19	13C2-PFOA	77,824.78	40,184.56	120,553.68	
KB77 CCV	CCV	10/23/18 5:30	13C2-PFOA	74,339.52	40,184.56	120,553.68	
J8616-FS(0)	VC-PM13-GW21-1018	10/23/18 5:52	13C2-PFOA	85,784.84	40,184.56	120,553.68	
~~J8616-FS-D(3)~~	~~VC-PM13-GW21-1018~~	~~10/23/18 6:02~~	~~13C2-PFOA~~	~~94,464.39~~	~~40,184.56~~	~~120,553.68~~	
J8617-FS(0)	VC-KCHA16-GW01-1018	10/23/18 6:13	13C2-PFOA	95,284.80	40,184.56	120,553.68	
J8617-FS-D(3)	VC-KCHA16-GW01-1018	10/23/18 6:24	13C2-PFOA	78,203.75	40,184.56	120,553.68	
J8618-FS(0)	VC-KCHA16-GW01P-1018	10/23/18 6:35	13C2-PFOA	99,109.90	40,184.56	120,553.68	
J8618-FS-D(3)	VC-KCHA16-GW01P-1018	10/23/18 6:46	13C2-PFOA	83,507.54	40,184.56	120,553.68	
J8619-FS(0)	VC-KCHA16-GW02-1018	10/23/18 6:57	13C2-PFOA	78,450.85	40,184.56	120,553.68	
J8619-FS-D(3)	VC-KCHA16-GW02-1018	10/23/18 7:08	13C2-PFOA	76,829.11	40,184.56	120,553.68	
J8620-FS(0)	VC-KCHA16-GW03-1018	10/23/18 7:19	13C2-PFOA	82,880.95	40,184.56	120,553.68	
J8620-FS-D(3)	VC-KCHA16-GW03-1018	10/23/18 7:30	13C2-PFOA	80,118.83	40,184.56	120,553.68	
KB76 CCV	CCV	10/23/18 7:40	13C2-PFOA	74,313.12	40,184.56	120,553.68	
J8621-FS(0)	VC-AQ-FB10-10022018	10/23/18 8:02	13C2-PFOA	85,941.52	40,184.56	120,553.68	
J8622-FS(0)	VC-AQ-EB10-10022018	10/23/18 8:13	13C2-PFOA	79,902.67	40,184.56	120,553.68	
J8623-FS(0)	VC-S09GW01-1018	10/23/18 8:24	13C2-PFOA	88,712.19	40,184.56	120,553.68	
J8623-FS-D(3)	VC-S09GW01-1018	10/23/18 8:35	13C2-PFOA	76,513.66	40,184.56	120,553.68	
J8624-FS(0)	VC-S09GW04-1018	10/23/18 8:46	13C2-PFOA	93,151.25	40,184.56	120,553.68	
J8624-FS-D(3)	VC-S09GW04-1018	10/23/18 9:19	13C2-PFOA	81,244.05	40,184.56	120,553.68	
J8625-FS(0)	VC-S09GW04P-1018	10/23/18 9:30	13C2-PFOA	85,332.38	40,184.56	120,553.68	
J8625-FS-D(3)	VC-S09GW04P-1018	10/23/18 9:40	13C2-PFOA	83,619.59	40,184.56	120,553.68	
J8626-FS(0)	VC-S09GW06-1018	10/23/18 9:51	13C2-PFOA	84,304.23	40,184.56	120,553.68	
J8626-FS-D(3)	VC-S09GW06-1018	10/23/18 10:02	13C2-PFOA	84,872.05	40,184.56	120,553.68	
KB77 CCV	CCV	10/23/18 10:13	13C2-PFOA	82,580.58	40,184.56	120,553.68	
J8627-FS(0)	VC-S14GW01-1018	10/23/18 10:35	13C2-PFOA	78,988.28	40,184.56	120,553.68	
J8627-FS-D(3)	VC-S14GW01-1018	10/23/18 10:46	13C2-PFOA	82,763.70	40,184.56	120,553.68	
J8628-FS(0)	VC-S14GW03-1018	10/23/18 10:57	13C2-PFOA	41,307.84	40,184.56	120,553.68	
KB76 CCV	CCV	10/23/18 11:08	13C2-PFOA	47,380.33	40,184.56	120,553.68	
KB80 IB	IB	10/22/18 19:43	13C2-PFOA	70,816.22	40,184.56	120,553.68	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C2-PFDA	90,129.70	45,064.85	135,194.55

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C2-PFDA	100,139.46	45,064.85	135,194.55	
KB74	L2	10/17/18 19:57	13C2-PFDA	103,883.83	45,064.85	135,194.55	
KB75	L3	10/17/18 20:08	13C2-PFDA	98,052.33	45,064.85	135,194.55	
KB76	L4	10/17/18 20:19	13C2-PFDA	99,978.99	45,064.85	135,194.55	
KB77	L5	10/17/18 20:30	13C2-PFDA	90,129.70	45,064.85	135,194.55	
KB78	L6	10/17/18 20:41	13C2-PFDA	104,169.70	45,064.85	135,194.55	
KB79	L7	10/17/18 20:52	13C2-PFDA	100,765.46	45,064.85	135,194.55	
KB80 IB	IB	10/17/18 21:02	13C2-PFDA	96,280.28	45,064.85	135,194.55	
KB81 ICC	ICC	10/17/18 21:13	13C2-PFDA	103,402.36	45,064.85	135,194.55	
KB76 CCV	CCV	10/23/18 3:41	13C2-PFDA	91,954.41	45,064.85	135,194.55	
CR939PB-FS(0)	Procedural Blank	10/23/18 4:03	13C2-PFDA	102,442.20	45,064.85	135,194.55	
~~CR940LCS-FS(0)~~	~~Laboratory Control Sample~~	~~10/23/18 4:14~~	~~13C2-PFDA~~	~~107,338.53~~	~~45,064.85~~	~~135,194.55~~	
J8613-FS(0)	VC-PM553-GW0601A-1018	10/23/18 4:25	13C2-PFDA	96,413.31	45,064.85	135,194.55	
J8613-FS-D(3)	VC-PM553-GW0601A-1018	10/23/18 4:36	13C2-PFDA	102,089.87	45,064.85	135,194.55	
J8613-FS-D(5)	VC-PM553-GW0601A-1018	10/23/18 4:46	13C2-PFDA	99,786.05	45,064.85	135,194.55	
J8613-FS-D(7)	VC-PM553-GW0601A-1018	10/23/18 4:57	13C2-PFDA	88,204.29	45,064.85	135,194.55	
J8614-FS(0)	VC-AQ-FB09-10012018	10/23/18 5:08	13C2-PFDA	92,253.71	45,064.85	135,194.55	
J8615-FS(0)	VC-AQ-EB09-10012018	10/23/18 5:19	13C2-PFDA	100,021.36	45,064.85	135,194.55	
KB77 CCV	CCV	10/23/18 5:30	13C2-PFDA	85,518.46	45,064.85	135,194.55	
J8616-FS(0)	VC-PM13-GW21-1018	10/23/18 5:52	13C2-PFDA	97,426.57	45,064.85	135,194.55	
~~J8616-FS-D(3)~~	~~VC-PM13-GW21-1018~~	~~10/23/18 6:02~~	~~13C2-PFDA~~	~~124,675.49~~	~~45,064.85~~	~~135,194.55~~	
J8617-FS(0)	VC-KCHA16-GW01-1018	10/23/18 6:13	13C2-PFDA	115,359.89	45,064.85	135,194.55	
J8617-FS-D(3)	VC-KCHA16-GW01-1018	10/23/18 6:24	13C2-PFDA	92,237.23	45,064.85	135,194.55	
J8618-FS(0)	VC-KCHA16-GW01P-1018	10/23/18 6:35	13C2-PFDA	120,367.15	45,064.85	135,194.55	
J8618-FS-D(3)	VC-KCHA16-GW01P-1018	10/23/18 6:46	13C2-PFDA	96,628.38	45,064.85	135,194.55	
J8619-FS(0)	VC-KCHA16-GW02-1018	10/23/18 6:57	13C2-PFDA	99,465.21	45,064.85	135,194.55	
J8619-FS-D(3)	VC-KCHA16-GW02-1018	10/23/18 7:08	13C2-PFDA	95,242.23	45,064.85	135,194.55	
J8620-FS(0)	VC-KCHA16-GW03-1018	10/23/18 7:19	13C2-PFDA	100,808.73	45,064.85	135,194.55	
J8620-FS-D(3)	VC-KCHA16-GW03-1018	10/23/18 7:30	13C2-PFDA	94,373.84	45,064.85	135,194.55	
KB76 CCV	CCV	10/23/18 7:40	13C2-PFDA	91,947.72	45,064.85	135,194.55	
J8621-FS(0)	VC-AQ-FB10-10022018	10/23/18 8:02	13C2-PFDA	104,909.04	45,064.85	135,194.55	
J8622-FS(0)	VC-AQ-EB10-10022018	10/23/18 8:13	13C2-PFDA	101,274.22	45,064.85	135,194.55	
J8623-FS(0)	VC-S09GW01-1018	10/23/18 8:24	13C2-PFDA	107,653.70	45,064.85	135,194.55	
J8623-FS-D(3)	VC-S09GW01-1018	10/23/18 8:35	13C2-PFDA	94,042.15	45,064.85	135,194.55	
J8624-FS(0)	VC-S09GW04-1018	10/23/18 8:46	13C2-PFDA	88,819.82	45,064.85	135,194.55	
J8624-FS-D(3)	VC-S09GW04-1018	10/23/18 9:19	13C2-PFDA	95,495.07	45,064.85	135,194.55	
J8625-FS(0)	VC-S09GW04P-1018	10/23/18 9:30	13C2-PFDA	89,409.45	45,064.85	135,194.55	
J8625-FS-D(3)	VC-S09GW04P-1018	10/23/18 9:40	13C2-PFDA	98,089.53	45,064.85	135,194.55	
J8626-FS(0)	VC-S09GW06-1018	10/23/18 9:51	13C2-PFDA	106,485.79	45,064.85	135,194.55	
J8626-FS-D(3)	VC-S09GW06-1018	10/23/18 10:02	13C2-PFDA	98,829.36	45,064.85	135,194.55	
KB77 CCV	CCV	10/23/18 10:13	13C2-PFDA	93,847.21	45,064.85	135,194.55	
J8627-FS(0)	VC-S14GW01-1018	10/23/18 10:35	13C2-PFDA	92,228.61	45,064.85	135,194.55	
J8627-FS-D(3)	VC-S14GW01-1018	10/23/18 10:46	13C2-PFDA	96,491.53	45,064.85	135,194.55	
J8628-FS(0)	VC-S14GW03-1018	10/23/18 10:57	13C2-PFDA	57,803.16	45,064.85	135,194.55	
KB76 CCV	CCV	10/23/18 11:08	13C2-PFDA	60,520.54	45,064.85	135,194.55	
KB80 IB	IB	10/22/18 19:43	13C2-PFDA	91,293.54	45,064.85	135,194.55	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C4-PFOS	31,518.26	15,759.13	47,277.39

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C4-PFOS	29,846.55	15,759.13	47,277.39	
KB74	L2	10/17/18 19:57	13C4-PFOS	34,856.85	15,759.13	47,277.39	
KB75	L3	10/17/18 20:08	13C4-PFOS	30,684.77	15,759.13	47,277.39	
KB76	L4	10/17/18 20:19	13C4-PFOS	30,962.75	15,759.13	47,277.39	
KB77	L5	10/17/18 20:30	13C4-PFOS	31,518.26	15,759.13	47,277.39	
KB78	L6	10/17/18 20:41	13C4-PFOS	29,589.19	15,759.13	47,277.39	
KB79	L7	10/17/18 20:52	13C4-PFOS	26,984.04	15,759.13	47,277.39	
KB80 IB	IB	10/17/18 21:02	13C4-PFOS	29,371.12	15,759.13	47,277.39	
KB81 ICC	ICC	10/17/18 21:13	13C4-PFOS	33,690.55	15,759.13	47,277.39	
KB76 CCV	CCV	10/23/18 3:41	13C4-PFOS	31,073.45	15,759.13	47,277.39	
CR939PB-FS(0)	Procedural Blank	10/23/18 4:03	13C4-PFOS	29,773.88	15,759.13	47,277.39	
~~CR940LCS-FS(0)~~	~~Laboratory Control Sample~~	~~10/23/18 4:14~~	~~13C4-PFOS~~	~~32,288.43~~	~~15,759.13~~	~~47,277.39~~	
J8613-FS(0)	VC-PM553-GW0601A-1018	10/23/18 4:25	13C4-PFOS	24,859.22	15,759.13	47,277.39	
J8613-FS-D(3)	VC-PM553-GW0601A-1018	10/23/18 4:36	13C4-PFOS	28,050.74	15,759.13	47,277.39	
J8613-FS-D(5)	VC-PM553-GW0601A-1018	10/23/18 4:46	13C4-PFOS	30,703.24	15,759.13	47,277.39	
J8613-FS-D(7)	VC-PM553-GW0601A-1018	10/23/18 4:57	13C4-PFOS	30,804.31	15,759.13	47,277.39	
J8614-FS(0)	VC-AQ-FB09-10012018	10/23/18 5:08	13C4-PFOS	28,857.59	15,759.13	47,277.39	
J8615-FS(0)	VC-AQ-EB09-10012018	10/23/18 5:19	13C4-PFOS	32,731.62	15,759.13	47,277.39	
KB77 CCV	CCV	10/23/18 5:30	13C4-PFOS	25,806.64	15,759.13	47,277.39	
J8616-FS(0)	VC-PM13-GW21-1018	10/23/18 5:52	13C4-PFOS	26,773.90	15,759.13	47,277.39	
~~J8616-FS-D(3)~~	~~VC-PM13-GW21-1018~~	~~10/23/18 6:02~~	~~13C4-PFOS~~	~~30,037.87~~	~~15,759.13~~	~~47,277.39~~	
J8617-FS(0)	VC-KCHA16-GW01-1018	10/23/18 6:13	13C4-PFOS	29,782.00	15,759.13	47,277.39	
J8617-FS-D(3)	VC-KCHA16-GW01-1018	10/23/18 6:24	13C4-PFOS	27,075.35	15,759.13	47,277.39	
J8618-FS(0)	VC-KCHA16-GW01P-1018	10/23/18 6:35	13C4-PFOS	34,948.68	15,759.13	47,277.39	
J8618-FS-D(3)	VC-KCHA16-GW01P-1018	10/23/18 6:46	13C4-PFOS	31,106.92	15,759.13	47,277.39	
J8619-FS(0)	VC-KCHA16-GW02-1018	10/23/18 6:57	13C4-PFOS	30,593.68	15,759.13	47,277.39	
J8619-FS-D(3)	VC-KCHA16-GW02-1018	10/23/18 7:08	13C4-PFOS	30,626.59	15,759.13	47,277.39	
J8620-FS(0)	VC-KCHA16-GW03-1018	10/23/18 7:19	13C4-PFOS	26,558.92	15,759.13	47,277.39	
J8620-FS-D(3)	VC-KCHA16-GW03-1018	10/23/18 7:30	13C4-PFOS	30,284.94	15,759.13	47,277.39	
KB76 CCV	CCV	10/23/18 7:40	13C4-PFOS	27,928.87	15,759.13	47,277.39	
J8621-FS(0)	VC-AQ-FB10-10022018	10/23/18 8:02	13C4-PFOS	34,544.23	15,759.13	47,277.39	
J8622-FS(0)	VC-AQ-EB10-10022018	10/23/18 8:13	13C4-PFOS	32,365.60	15,759.13	47,277.39	
J8623-FS(0)	VC-S09GW01-1018	10/23/18 8:24	13C4-PFOS	27,775.17	15,759.13	47,277.39	
J8623-FS-D(3)	VC-S09GW01-1018	10/23/18 8:35	13C4-PFOS	27,638.21	15,759.13	47,277.39	
J8624-FS(0)	VC-S09GW04-1018	10/23/18 8:46	13C4-PFOS	21,821.01	15,759.13	47,277.39	
J8624-FS-D(3)	VC-S09GW04-1018	10/23/18 9:19	13C4-PFOS	32,802.29	15,759.13	47,277.39	
J8625-FS(0)	VC-S09GW04P-1018	10/23/18 9:30	13C4-PFOS	22,601.04	15,759.13	47,277.39	
J8625-FS-D(3)	VC-S09GW04P-1018	10/23/18 9:40	13C4-PFOS	29,612.80	15,759.13	47,277.39	
J8626-FS(0)	VC-S09GW06-1018	10/23/18 9:51	13C4-PFOS	23,461.57	15,759.13	47,277.39	
J8626-FS-D(3)	VC-S09GW06-1018	10/23/18 10:02	13C4-PFOS	28,724.47	15,759.13	47,277.39	
KB77 CCV	CCV	10/23/18 10:13	13C4-PFOS	28,300.31	15,759.13	47,277.39	
J8627-FS(0)	VC-S14GW01-1018	10/23/18 10:35	13C4-PFOS	23,155.03	15,759.13	47,277.39	
J8627-FS-D(3)	VC-S14GW01-1018	10/23/18 10:46	13C4-PFOS	33,386.85	15,759.13	47,277.39	
J8628-FS(0)	VC-S14GW03-1018	10/23/18 10:57	13C4-PFOS	16,042.16	15,759.13	47,277.39	
KB76 CCV	CCV	10/23/18 11:08	13C4-PFOS	15,858.09	15,759.13	47,277.39	
KB80 IB	IB	10/22/18 19:43	13C4-PFOS	27,153.34	15,759.13	47,277.39	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C2-PFOA	80,369.12	40,184.56	120,553.68

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C2-PFOA	79,095.63	40,184.56	120,553.68	
KB74	L2	10/17/18 19:57	13C2-PFOA	89,971.31	40,184.56	120,553.68	
KB75	L3	10/17/18 20:08	13C2-PFOA	87,799.30	40,184.56	120,553.68	
KB76	L4	10/17/18 20:19	13C2-PFOA	84,567.91	40,184.56	120,553.68	
KB77	L5	10/17/18 20:30	13C2-PFOA	80,369.12	40,184.56	120,553.68	
KB78	L6	10/17/18 20:41	13C2-PFOA	85,964.25	40,184.56	120,553.68	
KB79	L7	10/17/18 20:52	13C2-PFOA	86,636.81	40,184.56	120,553.68	
KB80 IB	IB	10/17/18 21:02	13C2-PFOA	85,730.94	40,184.56	120,553.68	
KB81 ICC	ICC	10/17/18 21:13	13C2-PFOA	85,242.44	40,184.56	120,553.68	
KB75 ISC	ISC	10/24/18 11:51	13C2-PFOA	77,879.25	40,184.56	120,553.68	
KB80 IB	IB	10/24/18 12:02	13C2-PFOA	98,912.42	40,184.56	120,553.68	
CR940LCS-FS(0)	Laboratory Control Sample	10/24/18 12:13	13C2-PFOA	83,273.49	40,184.56	120,553.68	
J8616-FS-D(3)	VC-PM13-GW21-1018	10/24/18 12:23	13C2-PFOA	94,531.08	40,184.56	120,553.68	
J8627-FS-D(5)	VC-S14GW01-1018	10/24/18 12:34	13C2-PFOA	79,401.49	40,184.56	120,553.68	
KB76	L4	10/24/18 12:56	13C2-PFOA	82,184.98	40,184.56	120,553.68	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C2-PFDA	90,129.70	45,064.85	135,194.55

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C2-PFDA	100,139.46	45,064.85	135,194.55	
KB74	L2	10/17/18 19:57	13C2-PFDA	103,883.83	45,064.85	135,194.55	
KB75	L3	10/17/18 20:08	13C2-PFDA	98,052.33	45,064.85	135,194.55	
KB76	L4	10/17/18 20:19	13C2-PFDA	99,978.99	45,064.85	135,194.55	
KB77	L5	10/17/18 20:30	13C2-PFDA	90,129.70	45,064.85	135,194.55	
KB78	L6	10/17/18 20:41	13C2-PFDA	104,169.70	45,064.85	135,194.55	
KB79	L7	10/17/18 20:52	13C2-PFDA	100,765.46	45,064.85	135,194.55	
KB80 IB	IB	10/17/18 21:02	13C2-PFDA	96,280.28	45,064.85	135,194.55	
KB81 ICC	ICC	10/17/18 21:13	13C2-PFDA	103,402.36	45,064.85	135,194.55	
KB75 ISC	ISC	10/24/18 11:51	13C2-PFDA	101,356.83	45,064.85	135,194.55	
KB80 IB	IB	10/24/18 12:02	13C2-PFDA	112,231.68	45,064.85	135,194.55	
CR940LCS-FS(0)	Laboratory Control Sample	10/24/18 12:13	13C2-PFDA	99,145.40	45,064.85	135,194.55	
J8616-FS-D(3)	VC-PM13-GW21-1018	10/24/18 12:23	13C2-PFDA	116,159.80	45,064.85	135,194.55	
J8627-FS-D(5)	VC-S14GW01-1018	10/24/18 12:34	13C2-PFDA	90,498.88	45,064.85	135,194.55	
KB76	L4	10/24/18 12:56	13C2-PFDA	92,873.70	45,064.85	135,194.55	

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

BATTELLE
It can be done

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper
KB77	L5	10/17/18 20:30	13C4-PFOS	31,518.26	15,759.13	47,277.39

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier
KB73	L1	10/17/18 19:46	13C4-PFOS	29,846.55	15,759.13	47,277.39	
KB74	L2	10/17/18 19:57	13C4-PFOS	34,856.85	15,759.13	47,277.39	
KB75	L3	10/17/18 20:08	13C4-PFOS	30,684.77	15,759.13	47,277.39	
KB76	L4	10/17/18 20:19	13C4-PFOS	30,962.75	15,759.13	47,277.39	
KB77	L5	10/17/18 20:30	13C4-PFOS	31,518.26	15,759.13	47,277.39	
KB78	L6	10/17/18 20:41	13C4-PFOS	29,589.19	15,759.13	47,277.39	
KB79	L7	10/17/18 20:52	13C4-PFOS	26,984.04	15,759.13	47,277.39	
KB80 IB	IB	10/17/18 21:02	13C4-PFOS	29,371.12	15,759.13	47,277.39	
KB81 ICC	ICC	10/17/18 21:13	13C4-PFOS	33,690.55	15,759.13	47,277.39	
KB75 ISC	ISC	10/24/18 11:51	13C4-PFOS	30,407.25	15,759.13	47,277.39	
KB80 IB	IB	10/24/18 12:02	13C4-PFOS	35,418.03	15,759.13	47,277.39	
CR940LCS-FS(0)	Laboratory Control Sample	10/24/18 12:13	13C4-PFOS	28,665.62	15,759.13	47,277.39	
J8616-FS-D(3)	VC-PM13-GW21-1018	10/24/18 12:23	13C4-PFOS	35,168.89	15,759.13	47,277.39	
J8627-FS-D(5)	VC-S14GW01-1018	10/24/18 12:34	13C4-PFOS	27,170.88	15,759.13	47,277.39	
KB76	L4	10/24/18 12:56	13C4-PFOS	29,631.33	15,759.13	47,277.39	

BATTELLE

Mass Spectral Acquisition Rate Report

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 8:52:06 PM	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
PFBS_1	298.9 / 80.0	1.56	22	>10
PFBS_2	298.9 / 99.0	1.56	22	>10
PFHxA_1	313.0 / 269.0	1.89	21	>10
PFHxA_2	313.0 / 119.0	1.88	22	>10
PFHpA_1	363.0 / 319.0	2.29	30	>10
PFHpA_2	363.0 / 169.0	2.30	23	>10
PFHxS_1	399.0 / 80.0	2.32	33	>10
PFHxS_2	399.0 / 99.0	2.32	35	>10
PFOA_1	413.0 / 369.0	2.71	32	>10
PFOA_2	413.0 / 169.0	2.71	30	>10
PFNA_1	463.0 / 419.0	3.11	29	>10
PFNA_2	463.0 / 219.0	3.11	27	>10
PFOS_1	499.0 / 80.0	3.10	43	>10
PFOS_2	499.0 / 99.0	3.10	45	>10
PFDA_1	513.0 / 469.0	3.47	31	>10
PFDA_2	513.0 / 219.0	3.47	34	>10
PFUnA_1	563.0 / 519.0	3.79	30	>10
PFUnA_2	563.0 / 269.0	3.79	43	>10
PFDoA_1	613.0 / 569.0	4.07	39	>10
PFDoA_2	613.0 / 319.0	4.07	46	>10
PFTrDA_1	663.0 / 619.0	4.32	56	>10
PFTrDA_2	663.0 / 169.0	4.32	40	>10
PFTeDA_1	713.0 / 669.0	4.54	59	>10
PFTeDA_2	713.0 / 169.0	4.54	64	>10
NMeFOSAA_1	570.0 / 419.0	3.62	29	>10
NMeFOSAA_2	570.0 / 512.0	3.62	39	>10
NEtFOSAA_1	584.0 / 419.0	3.78	31	>10
NEtFOSAA_2	584.0 / 483.0	3.78	19	>10

Mass Spectral Acquisition Rate Report

Created with Analyst Reporter
Printed: 24/10/2018 1:06:05 PM

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 8:52:06 PM	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
13C2-PFDoA	615.0 / 570.0	4.07	23	>10
d3-MeFOSAA	573.0 / 419.0	3.62	18	>10
d5-EtFOSAA	589.0 / 419.0	3.78	27	>10
13C5-PFHxA	318.0 / 273.0	1.87	23	>10
13C4-PFHpA	367.0 / 322.0	2.28	26	>10
13C8-PFOA	421.0 / 376.0	2.70	47	>10
13C9-PFNA	472.0 / 427.0	3.09	24	>10
13C6-PFDA	519.0 / 474.0	3.45	29	>10
13C7-PFUnA	570.0 / 525.0	3.78	36	>10
13C2-PFTeDA	715.0 / 670.0	4.54	38	>10
13C3-PFBS	302.0 / 99.0	1.54	24	>10
13C3-PFHxS	402.0 / 99.0	2.31	27	>10
13C8-PFOS	507.0 / 99.0	3.09	33	>10

BATTELLE
It can be done

Precision and Bias at the LOQ for PFAS in non-potable Water

Analyte	CAS No.	Average (ng/L)	ST DEV	2 Sigma	n
PFBA	375-22-4	12.29	2.02	4.04	13
PFPeA	2706-90-3	10.73	1.51	3.02	9
PFHxA	307-24-4	9.93	1.30	2.60	39
PFHpA	375-85-9	9.42	1.57	3.14	39
PFOA	335-67-1	10.16	1.46	2.92	41
PFNA	375-95-1	9.64	1.15	2.30	39
PFDA	335-76-2	9.89	1.32	2.64	39
PFUnA	2058-94-8	9.86	1.31	2.62	39
PFDoA	307-55-1	10.75	1.29	2.58	39
PFTrDA	72629-94-8	11.18	1.54	3.08	39
PFTeDA	376-06-7	10.70	1.91	3.82	39
NMeFOSAA	2355-31-9	10.26	1.87	3.74	39
NEtFOSAA	2991-50-6	9.63	1.54	3.08	39
PFOSA	754-91-6	9.74	1.14	2.28	4
PFBS	375-73-5	10.05	1.44	2.88	40
PFPeS	BDO-2114	9.80	0.96	1.92	5
PFHxS	355-46-4	9.76	1.40	2.80	39
PFHpS	375-99-6	10.96	0.96	1.92	10
PFOS	1763-23-1	10.06	1.36	2.72	39
PFNS	98789-57-2	9.34	1.10	2.20	4
PFDS	2806-15-7	10.13	1.88	3.76	9
4:2FTS	BDO-2205	11.03	1.26	2.52	9
6:2FTS	27619-97-2	12.52	2.91	5.82	9
8:2FTS	39108-34-4	12.11	2.54	5.08	9

BATTELLE DETECTION LIMITS FOR PFAS IN NON-POTABLE WATER

Analytical SOP 5-369
Extraction SOP 5-370

PFAS by LC-MS/MS Compliant with QSM 5.1 Compliant Table B-15

Analyte	CAS No.	MDL (ng/L)	LOD (ng/L)	LOQ (ng/L)
PFBA	375-22-4	0.14	0.5	5.0
PFPeA	2706-90-3	0.31	1.0	5.0
PFHxA	307-24-4	0.19	0.5	5.0
PFHpA	375-85-9	0.16	0.5	5.0
PFOA	335-67-1	0.18	0.5	5.0
PFNA	375-95-1	0.26	1.0	5.0
PFDA	335-76-2	0.16	0.5	5.0
PFUnA	2058-94-8	0.29	1.0	5.0
PFDoA	307-55-1	0.18	0.5	5.0
PFTrDA	72629-94-8	0.15	0.5	5.0
PFTeDA	376-06-7	0.25	1.0	5.0
NMeFOSAA	2355-31-9	0.56	2.0	5.0
NEtFOSAA	2991-50-6	0.49	1.0	5.0
PFOSA	754-91-6	TBD	TBD	TBD
PFBS	375-73-5	0.13	0.5	5.0
PFPeS	BDO-2114	0.67	2.5	5.0
PFHxS	355-46-4	0.11	0.4	5.0
PFHpS	375-99-6	0.20	0.5	5.0
PFOS	1763-23-1	0.19	0.5	5.0
PFNS	98789-57-2	0.46	1.0	5.0
PFDS	2806-15-7	0.17	0.5	5.0
4:2FTS	BDO-2205	0.14	0.5	5.0
6:2FTS	27619-97-2	1.36	2.5	5.0
8:2FTS	39108-34-4	0.22	0.5	5.0

Analytes on NELAP and ELAP QSM 5.1 Scope of accreditation

Battelle and its logos are registered trademarks of Battelle Memorial Institute. © Battelle Memorial Institute 2018. All Rights Reserved.

ID 596 04/18

Analytical Transitions for PFAS in non-potable water, solid, and tissue

EPA 537 MOD DoD QSM 5.1 compliant with Table B-15 requirements

Analyte	CAS No.	Type	Primary Transition	Secondary Transition
PFBA	375-22-4	Target	213.0 / 169.0	NA
PFPeA	2706-90-3	Target	263.0 / 219.0	NA
PFHxA	307-24-4	Target	313.0 / 269.0	313.0 / 119.0
PFHpA	375-85-9	Target	363.0 / 319.0	363.0 / 169.0
PFOA	335-67-1	Target	413.0 / 369.0	413.0 / 169.0
PFNA	375-95-1	Target	463.0 / 419.0	463.0 / 219.0
PFDA	335-76-2	Target	513.0 / 469.0	513.0 / 219.0
PFUnA	2058-94-8	Target	563.0 / 519.0	563.0 / 269.0
PFDoA	307-55-1	Target	613.0 / 569.0	613.0 / 319.0
PFTrDA	72629-94-8	Target	663.0 / 619.0	663.0 / 169.0
PFTeDA	376-06-7	Target	713.0 / 669.0	713.0 / 169.0
NMeFOSAA	2355-31-9	Target	570.0 / 419.0	570.0 / 512.0
NEtFOSAA	2991-50-6	Target	584.0 / 419.0	584.0 / 483.0
PFOSA	754-91-6	Target	498.0 / 78.0	498.0 / 83.0
PFBS	375-73-5	Target	299.0 / 80.0	299.0 / 99.0
PFPeS	BDO-2114	Target	349.0 / 99.0	249.0 / 80.0
PFHxS	355-46-4	Target	399.0 / 80.0	399.0 / 99.0
PFHpS	375-99-6	Target	449.0 / 80.0	449.0 / 99.0
PFOS	1763-23-1	Target	499.0 / 80.0	499.0 / 99.0
PFNS	98789-57-2	Target	549.0 / 99.0	549.0 / 80.0
PFDS	2806-15-7	Target	599.0 / 80.0	599.0 / 99.0
4:2FTS	BDO-2205	Target	327.0 / 307.0	327.0 / 80.0
6:2FTS	27619-97-2	Target	427.0 / 407.0	427.0 / 81.0
8:2FTS	39108-34-4	Target	527.0 / 507.0	527.0 / 487.0
13C4-PFBA	BDO-2105	SIS[1]	217.0 / 172.0	NA
13C5-PFPeA	BDO-2216	SIS[1]	268.0 / 223.0	NA
13C5-PFHxA	BDO-2217	SIS[1]	318.0 / 273.0	NA

Analyte	CAS No.	Type	Primary Transition	Secondary Transition
13C4-PFHpA	BDO-2218	SIS[1]	367.0 / 322.0	NA
13C8-PFOA	BDO-2219	SIS[1]	421.0 / 376.0	NA
13C9-PFNA	BDO-2221	SIS[1]	472.0 / 427.0	NA
13C6-PFDA	BDO-2222	SIS[1]	519.0 / 474.0	NA
13C7-PFUnA	BDO-2223	SIS[1]	570.0 / 525.0	NA
13C2-PFDoA	BDO-2112	SIS[1]	615.0 / 570.0	NA
13C2-PFTeDA	BDO-2224	SIS[1]	715.0 / 670.0	NA
d3-MeFOSAA	BDO-1838	SIS[1]	573.0 / 419.0	NA
d5-EtFOSAA	BDO-1839	SIS[1]	589.0 / 419.0	NA
13C8-FOSA	BDO-2225	SIS[1]	506.0 / 78.0	NA
13C3-PFBS	BDO-2226	SIS[1]	302.0 / 99.0	NA
13C3-PFHxS	BDO-2227	SIS[1]	402.0 / 99.0	NA
13C8-PFOS	BDO-2228	SIS[1]	507.0 / 99.0	NA
13C2-4:2FTS	BDO-2229	SIS[1]	329.0 / 81.0	NA
13C2-6:2FTS	BDO-2230	SIS[1]	429.0 / 81.0	NA
13C2-8:2FTS	BDO-2220	SIS[1]	529.0 / 81.0	NA
13C3-PFBA	BDO-2231	IS[2]	216.0 / 172.0	NA
13C2-PFOA	BDO-2107	IS[2]	415.0 / 370.0	NA
13C2-PFDA	BDO-2110	IS[2]	515.0 / 470.0	NA
13C4-PFOS	BDO-2121	IS[2]	503.0 / 99.0	NA

[1] – extracted internal standard (surrogate)
[2] – injection internal standard

BATTELLE

Non-Potable Water Calibration to Sample Equivalents

ICAL (ng/L)	PIV (mL)	DF[1]	Sample Size (L)	Sample Equivalent (ng/L)[2]
25	1	1	0.250	0.1
50	1	1	0.250	0.2
100	1	1	0.250	0.4
250	1	1	0.250	1.0
500	1	1	0.250	2.0
1,000	1	1	0.250	4.0
2,500	1	1	0.250	10.0
10,000	1	1	0.250	40.0
20,000	1	1	0.250	80.0

[1] - base level dilution as part of the extraction procedure

[2] - calculated equivalent of a sample based on the ICAL concentration

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

QTRAP 5500
Preventive Maintenance Checklist

Preventive Maintenance Date:	12-June-2018

Request ID:	9749
Company Name:	Battelle Memorial Institute
Instrument ID:	X60666
Instrument Model:	QTRAP 5500
Instrument Serial Number:	AU23051004

☑ **PASS** ☐ **FAIL**

Any failure will lead to an automatic Service Call being open to investigate fault.

Preventive Maintenance is performed twice every year unless specified in the Service Contract. It is designed to help maintain optimum system performance and to help diagnose any system deficiencies.

Engineer is required the assigned Request ID for this PM otherwise making this job invalid.

Comments: Suspected issue with pulse gas manifold. TRAP testing in POSITIVE mode couldn't be finished because of pulse gas issue. The same issue will be taken care in separate service call.

Performed By: Kaustubh Dhayagude **Date:** 12-June-2018

Approved By :_____ **Date:** _____

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

PRE PM PPG PERFORMANCE EVALUATION:

☑ Consult Customer concerning the unit overall performance.
☑ Check Logbook for Services recently performed.
☑ Check Vacuum Pressure:

CAD Settings	Vacuum Reading (x 10^{-5} Torr)	Acceptance Criteria
☑ CAD 0	0.6	0.4 to 1.1 x10^{-5} Torr
☑ CAD Low	1.3	Read Only
☑ CAD Medium	2.7	Read Only
☑ CAD High	3.7	Read Only
☑ CAD 12	3.7	2.4 to 4.5 x10^{-5} Torr

☑ Check for Front end contamination symptoms. Run Q1 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification

☑ No degradation or Sensitivity drop

☑ Check for Q3 contamination symptoms. Run Q3 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification

☑ No degradation or Sensitivity drop

Pre PM PPG Test: Perform each of the following tests. Optimize ion source position only. The specifications listed for these Pre PM tests are guidelines only, not required to be met.

☑ Perform Q1 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 175.133	4.01 e6	Read Only	0.6998	Read Only
Q1 500.380	2.81 e7	Read Only	0.7038	Read Only
Q1 906.673	4.21 e7	Read Only	0.7071	Read Only

☑ Perform Q3 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 175.133	5.45 e6	Read Only	0.6873	Read Only
Q3 500.380	2.69 e7	Read Only	0.7591	Read Only
Q3 906.673	4.50 e7	Read Only	0.7843	Read Only

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

☐ Perform MSMS POS in Product Ion scan with 609.3 parent and record daughter 195.1 using Reserpine 0.167 pmol/ul at the scan rate of 10 Da/s for 10 MCA. Calculate transmission efficiency comparing Q1POS 609 intensity. Transmission Efficiency: :__28.87%__ (Read Only)

Mass	MSMS Intensity		MSMS Width Value	Width Specs
	Value	Spec		
Q1 609.3	4.26 e7	Read Only	0.7011	Read Only
MS/MS 195.1	1.23 e7	Read Only	0.7069	Read Only

☑ Perform Q1 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 933.636	1.42 e7	Read Only	0.7686	Read Only

☑ Perform Q3 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 933.636	2.24 e7	Read Only	0.7243	Read Only

☑ Perform Product Ion scan using NEG PPG 3e-5M. Record10mca.

Mass	Scan Rate	MCA	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.31 e6	Read Only	0.6746	Read Only

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

PREVENTIVE MAINTENANCE CHECKLIST:

☑ Check Cooling Fans for Turbo Pumps while MS is ON.
☑ Check QJet and QPS tuning voltage for reference.
☐ Record AC input Voltage while MS is OFF: _____(200-240VAC).
If Out-of-Range, notify customer.

☑ Clean Interface
- ☑ Curtain Plate
- ☑ Orifice Plate
- ☑ QJet
- ☑ Q0 Rods.

☑ Replace Roughing Pump Oil.
☑ Inspect Oil Exhaust Filter, if Applicable. ☐ N/A
☑ Clean and inspect built-in divert valve if used. ☐ N/A
☑ Check Multiplier Voltage, optimize if necessary.
☑ Replace four Air Filters at the bottom of the mass spectrometer.

☑Pump down overnight if possible. ☐ N/A

☑Perform Maintenance on Turbo V source.

☑Replace Electrode, if necessary. ☐ N/A
☑Check Turbo heaters resistances.
☑Check if Temperature is reached at 500C with TIS Probe installed.
☐Check if Temperature is reached at 500C with APCI Probe installed. ☑ N/A

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System
Appendix ZEFPM003-2L

POST PM PPG PERFORMANCE TESTS:

☑ Set-up Sample for Infusion.
☑ Check spray and adjust sprayer's position of the TIS source.
☑ Check Vacuum Pressure:

CAD Settings	Vacuum Reading (x 10^{-5} Torr)	Acceptance Criteria
☑ CAD 0	0.7	0.4 to 1.1 x10^{-5} Torr
☑ CAD Low	1.3	Read Only
☑ CAD Medium	2.7	Read Only
☑ CAD High	3.7	Read Only
☑ CAD 12	3.7	2.4 to 4.5 x10^{-5} Torr

☑ Perform Q1 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q1 Intensity		Q1 Width	Width Specs
	Value	Spec	Value	
Scan Rate 10 Da/s Record 10 mca				
Q1 175.133	5.04 e6	≥1.2^{e}6	0.6737	0.6 to 0.8
Q1 500.380	1.60 e7	≥9.0^{e}6	0.6961	0.6 to 0.8
Q1 906.673	2.84 e7	≥1.4^{e}7	0.7179	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q1 906.673	1.33 e8	≥6.8^{e}7	0.7465	0.6 to 0.8

☑ Perform Q3 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q3 Intensity		Q3 Width	Width Specs
	Value	Spec	Value	
Scan Rate 10 Da/s Record 10 mca				
Q3 175.133	5.02 e6	≥1.2^{e}6	0.6719	0.6 to 0.8
Q3 500.380	1.72 e7	≥9.0^{e}6	0.7443	0.6 to 0.8
Q3 906.673	3.00 e7	≥1.4^{e}7	0.7504	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q3 906.673	1.46 e8	≥6.8^{e}7	0.7202	0.6 to 0.8

☑ Perform "Product of 609.3" POS and record product ion 195.1 using Reserpine 0.167pmol/uL. Record 10 mca. Calculate Transmission efficiency comparing Q1POS 609 intensity.
Transmission Efficiency: 21.10% (≥ 10.0%)

Mass	MSMS Intensity		Width	Width Specs
	Value	Spec	Value	
Q1 609.3	5.78 e7	N/A	0.6888	Read Only
MS/MS 195.1	1.22 e7	N/A	0.7003	Read Only

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

☑ Perform Q1 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q1 Intensity		Q1 Width Value	Width Specs
			Value	Spec		
Q1 933.636	10	10	1.35 e7	≥1.0^e7	0.7486	0.6 to 0.8
Q1 933.636	1000	50	7.52 e7	≥4.0^e7	0.7206	0.6 to 0.8

☑ Perform Q3 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q3 Intensity		Q3 Width Value	Width Specs
			Value	Spec		
Q3 933.636	10	10	2.15 e7	≥8.0^e6	0.7492	0.6 to 0.8
Q3 933.636	1000	50	8.33 e7	≥4.0^e7	0.7299	0.6 to 0.8

☑ Perform Product Ion scan using NEG PPG 3e-5M.

Mass	Scan Rate	Mca	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.33 e6	Read Only	0.6387	Read Only

☑ Perform ER POS 118.087 and 922.01 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 118.087	0.05	8.54 e6	≥7.2 ^e6	0.1473	<0.35
ER 922.010	0.05	4.96 e7	≥2.8 ^e6	0.2434	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 118.087	0.05		≥2.4 ^e7		<0.65
ER 922.010	0.05		≥6.8 ^e7		<0.65

☑ Perform ER NEG 431.982 and 601.978 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 431.982	0.05	1.81 e8	≥4.4 ^e7	0.1862	<0.35
ER 601.978	0.05	1.70 e8	≥5.6 ^e7	0.1809	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 431.982	0.05	5.72 e8	≥1.2 ^e8	0.5102	<0.65
ER 601.978	0.05	4.52 e8	≥1.6 ^e8	0.6187	<0.65

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.

Zef Scientific Inc.

12707 High Bluff Dr.
Suite 200
San Diego, CA
USA 92130

1975 Hymus Blvd.
Suite 230
Dorval, QC
Canada H9P 1J8

Phone: 1.866.854.7988

QTRAP 5500

LC/MS/MS Detector System

Appendix ZEFPM003-2L

☑ Perform EPI POS 397.2 using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Q0 Trapping OFF		Q0 Trapping ON	
		Intensity	Spec	Intensity	Spec
EPI 397.2	10000	> 3.0 e6	≥2.0 e6	> 7.0 e6	≥6.4 e6

☑ Perform MS3 POS full scan Fragmentation ON & OFF using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Fragamentation OFF		Fragmentation ON	
		Intensity	Spec	Intensity	Spec
MS3 397.2	1000	Yes	Contains only 397.2	N/A	N/A
☐ 236 OR ☑ 365	1000	Yes	Fragment Intensity	> 2.0 e6	≥1.6x $10^{e}6$

REVIEW:


☑ Attach all spectrums printouts to this procedure.
☑ If any parameter setting access modes were changed during the PM, ensure they are returned to their normal access mode and that their offsets are adjusted to match optimized values from the post-PM acquisition files.
☐ Empty tuning cache folder, if necessary. ☑ N/A
☑ Update Service Work Order status
☑ Fill and replace PM Label.


END OF PREVENTIVE MAINTENANCE CHECKLIST


Document history:


06 OCT 2016: Appendix ZEFPM003-2L: Removed requirements to fit Manufacturer's testing criteria.

© This document is the property of Zef Scientific and intended for Zef Scientific Inc trained engineers use only. No copying, all or partial, is permitted without prior authorization.









Peak List for " +ER: 50 MCA scans from Sample 4 (ER POS_10000 Da) of TRAP ER with NEW Pulse Manifold.wiff (Turbo Spray)"

	Target Mass (Da)	Found At (Da)	Intensity (cps)	Width (Da)	Mass Shift (Da)
1	118.0870	118.0702	3.0667e7	0.4146	0.0168
2	322.0490	322.0509	4.3500e7	0.4945	-1.9159e-3
3	622.0290	622.0370	4.6717e8	0.5757	-8.0044e-3
4	922.0100	922.0101	4.0400e8	0.5732	-1.4148e-4

Battelle Standard ID	Description	Intermediate Solutions	Battelle Reagent ID (purchased solutions)
KB72	PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)	KB71	180726-05
KC19	PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)	KB71	180726-05
JZ88	PFAS - DoD Second Source LCS/MS Solution	-	170724-01
KC03	PFAS - DoD Internal Standard Spiking Solution	JY25	180726-04
KB73	PFAS - DoD Calibration L1	KB71	180726-05
KB73	PFAS - DoD Calibration L1	JY25	180726-04
KB73	PFAS - DoD Calibration L1	JY23	180705-02
KB74	PFAS - DoD Calibration L2	JY25	180726-04
KB74	PFAS - DoD Calibration L2	KB71	180726-05
KB74	PFAS - DoD Calibration L2	JY23	180705-02
KB75	PFAS - DoD Calibration L3	KB71	180726-05
KB75	PFAS - DoD Calibration L3	JY25	180726-04
KB75	PFAS - DoD Calibration L3	KB70	180705-02
KB76	PFAS - DoD Calibration L4	KB70	180705-02
KB76	PFAS - DoD Calibration L4	JY25	180726-04
KB76	PFAS - DoD Calibration L4	KB71	180726-05
KB77	PFAS - DoD Calibration L5	KB71	180726-05
KB77	PFAS - DoD Calibration L5	KB70	180705-02
KB77	PFAS - DoD Calibration L5	JY25	180726-04
KB78	PFAS - DoD Calibraiton L6	JY25	180726-04
KB78	PFAS - DoD Calibraiton L6	KB70	180705-02
KB78	PFAS - DoD Calibraiton L6	KB71	180726-05
KB79	PFAS - DoD Calibration L7	KB71	180726-05
KB79	PFAS - DoD Calibration L7	KB70	180705-02
KB79	PFAS - DoD Calibration L7	JY25	180726-04
KB80	PFAS - DoD Instrument Blank	JY25	180726-04
KB80	PFAS - DoD Instrument Blank	KB71	180726-05
KB81	PFAS - DoD ICC	KB82	170724-01
KB81	PFAS - DoD ICC	KB71	180726-05
KB81	PFAS - DoD ICC	JY25	180726-04
KB89	PFAS Branched Solution (~5,000 ng/L)	JX28	180618-02
KB89	PFAS Branched Solution (~5,000 ng/L)	JX28	180618-03
KB89	PFAS Branched Solution (~5,000 ng/L)	JX28	180618-04
KB89	PFAS Branched Solution (~5,000 ng/L)	JX28	180618-06
KB89	PFAS Branched Solution (~5,000 ng/L)	JX28	180618-07

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: JX28

Description: PFAS Branched Standard Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180618-02	Branched NEtFOSAA Standard (50 µg/mL)	Neat	~50.000000	01/17/23	---	---	100 uL	1	10	~0.5000
180618-03	Branched NMeFOSAA Standard (50 µg/mL)	Neat	~50.000000	01/17/23	---	---	100 uL	1	10	~0.5000
180618-04	PFOA - Technical Mix	Neat	~50.000000	02/16/22	---	---	100 uL	1	10	~0.5000
180618-06	Branched PFHxS Standard (50 µg/mL)	Neat	~50.000000	01/04/22	---	---	100 uL	1	10	~0.5000
180618-07	Branched PFOS Standard (50 µg/mL)	Neat	~50.000000	01/12/22	---	---	100 uL	1	10	~0.5000

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 6/18/2018 **Expiration Date:** 6/18/2019

Solution Volume 25 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Balance ID:

Comment:

Solvent:	Lot:
Methanol (HPLC)	179315

Approved By: Thorn, Jonathan **Date:** 7/3/2018 8:10:00 AM

BATTELLE

It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JX28

Description: PFAS Branched Standard Stock

Stock Id: 180618-02

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	100	50.00	1	100.000	1	10	0.50000

Stock Id: 180618-03

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-methylperfluoro-1-octanesulfonamidoacetic acid	100	50.00	1	100.000	1	10	0.50000

Stock Id: 180618-04

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
Perfluoro-n-octanoic Acid	100	50.00	1	100.000	1	10	0.50000

Stock Id: 180618-06

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
Perfluoro-1-hexanesulfonate	100	50.00	1	100.000	1	10	0.50000

Stock Id: 180618-07

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
Perfluoro-1-octanesulfonate	100	50.00	1	100.000	1	10	0.50000

Final Concentrations:

Analyte:	Conc (ug/mL):
N-ethylperfluoro-octanesulfonamidoacetic acid	.50000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.50000
Perfluoro-1-hexanesulfonate	.50000
Perfluoro-1-octanesulfonate	.50000
Perfluoro-n-octanoic Acid	.50000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180618-02	Pipette	I0793912B
180618-03	Pipette	I0793912B
180618-04	Pipette	I0793912B
180618-06	Pipette	I0793912B
180618-07	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 6/18/2018 **Expiration Date:** 6/18/2019

Solution Volume 25 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment:

Approved By: Thorn, Jonathan **Date:** 7/3/2018 8:10:00 AM

BATTELLE

It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: JY23

Description: PFAS - DoD Low ICAL Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180705-02	PFOA - DOD	Neat	~1.000000	06/19/23	---	---	500 uL	1	100	~0.0050

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 7/16/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 4 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Balance ID: _____

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 8/29/2018 10:10:00 AM

BATTELLE

It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JY23

Description: PFAS - DoD Low ICAL Stock

Stock Id: 180705-02

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	500	1.01	1	100.000	1	100	0.00505
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	500	1.01	1	100.000	1	100	0.00505
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	500	1.00	1	100.000	1	100	0.00500
(Na) Perfluoro-1-decanesulfonate	500	1.01	1	100.000	1	100	0.00505
(NA) Perfluoro-1-heptanesulfonate	500	1.00	1	100.000	1	100	0.00500
(Na) Perfluoro-1-nonanesulfonate	500	1.01	1	100.000	1	100	0.00505
N-ethylperfluoro-octanesulfonamidoacetic acid	500	1.00	1	100.000	1	100	0.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-1-butanesulfonate	500	1.01	1	100.000	1	100	0.00505
Perfluoro-1-hexanesulfonate	500	1.01	1	100.000	1	100	0.00505
Perfluoro-1-octanesulfonamide	500	1.00	1	100.000	1	100	0.00500
Perfluoro-1-octanesulfonate	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-butanoic Acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-decanoic Acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-dodecanoic acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-heptanoic Acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-hexanoic acid	500	1.01	1	100.000	1	100	0.00505
Perfluoro-n-octanoic Acid	500	1.00	1	100.000	1	100	0.00500
Perfluorononanoic Acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-pentanoic acid	500	1.01	1	100.000	1	100	0.00505
Perfluoro-n-tetradecanoic acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-tridecanoic acid	500	1.00	1	100.000	1	100	0.00500
Perfluoro-n-undecanoic acid	500	1.00	1	100.000	1	100	0.00500
Sodium perfluoro-1-pentanesulfonate	500	1.00	1	100.000	1	100	0.00500

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00505
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00505
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00500
(Na) Perfluoro-1-decanesulfonate	.00505
(NA) Perfluoro-1-heptanesulfonate	.00500
(Na) Perfluoro-1-nonanesulfonate	.00505
N-ethylperfluoro-octanesulfonamidoacetic acid	.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00500
Perfluoro-1-butanesulfonate	.00505

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 7/16/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 4 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 8/29/2018 10:10:00 AM

BATTELLE

It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JY23

Description: PFAS - DoD Low ICAL Stock

Perfluoro-1-hexanesulfonate	.00505
Perfluoro-1-octanesulfonamide	.00500
Perfluoro-1-octanesulfonate	.00500
Perfluoro-n-butanoic Acid	.00500
Perfluoro-n-decanoic Acid	.00500
Perfluoro-n-dodecanoic acid	.00500
Perfluoro-n-heptanoic Acid	.00500
Perfluoro-n-hexanoic acid	.00505
Perfluoro-n-octanoic Acid	.00500
Perfluorononanoic Acid	.00500
Perfluoro-n-pentanoic acid	.00505
Perfluoro-n-tetradecanoic acid	.00500
Perfluoro-n-tridecanoic acid	.00500
Perfluoro-n-undecanoic acid	.00500
Sodium perfluoro-1-pentanesulfonate	.00500

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180705-02	Pipette	B820865811

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 7/16/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 4 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 8/29/2018 10:10:00 AM

BATTELLE

It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: JY25

Description: PFAS - DoD Internal Standard Stock Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180726-04	Mass-labelled PFAS injection standards	Neat	~2.000000	05/02/22	---	---	625 uL	1	25	~0.0500

Solution Prepared By: Schultz, Stephanie	**Date Prepared:** 7/16/2018	**Expiration Date:**	7/16/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 96/4 Methanol/Milli-q water (RP-180803-1)

Approved By: Thorn, Jonathan **Date:** 8/29/2018 10:09:00 AM

BATTELLE

It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JY25

Description: PFAS - DoD Internal Standard Stock Solution

Stock Id: 180726-04

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	625	2.00	1	100.000	1	25	0.05000
13C2-PFOA	625	2.00	1	100.000	1	25	0.05000
13C3-PFBA	625	2.00	1	100.000	1	25	0.05000
13C4-PFOS	625	1.91	1	100.000	1	25	0.04785

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.05000
13C2-PFOA	.05000
13C3-PFBA	.05000
13C4-PFOS	.04785

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180726-04	Pipette	B820865811

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 7/16/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/Milli-q water (RP-180803-1)

Approved By: Thorn, Jonathan **Date:** 8/29/2018 10:09:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **JZ88**

Description: PFAS - DoD Second Source LCS/MS Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
170724-01	PFOA- 2nd Source	Neat	~1.000000	03/22/22	---	---	1000 uL	1	20	~0.0500

Solution Prepared By: Schultz, Stephanie	**Date Prepared:** 8/20/2018	**Expiration Date:**	8/20/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 8/21/2018 7:17:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JZ88

Description: PFAS - DoD Second Source LCS/MS Solution

Stock Id: 170724-01

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	1000	1.01	1	100.000	1	20	0.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-decanesulfonate	1000	1.01	1	100.000	1	20	0.05050
(NA) Perfluoro-1-heptanesulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-nonanesulfonate	1000	1.01	1	100.000	1	20	0.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-butanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-hexanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-octanesulfonamide	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-octanesulfonate	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-butanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-decanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-dodecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-heptanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-hexanoic acid	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-n-octanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluorononanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-pentanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-tetradecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-tridecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-undecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Sodium perfluoro-1-pentanesulfonate	1000	1.00	1	100.000	1	20	0.05000

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.05000
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.05000
(Na) Perfluoro-1-decanesulfonate	.05050
(NA) Perfluoro-1-heptanesulfonate	.05000
(Na) Perfluoro-1-nonanesulfonate	.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.05000
Perfluoro-1-butanesulfonate	.05050

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 8/20/2018 **Expiration Date:** 8/20/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 8/21/2018 7:17:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: JZ88

Description: PFAS - DoD Second Source LCS/MS Solution

Perfluoro-1-hexanesulfonate	.05050
Perfluoro-1-octanesulfonamide	.05000
Perfluoro-1-octanesulfonate	.05000
Perfluoro-n-butanoic Acid	.05000
Perfluoro-n-decanoic Acid	.05000
Perfluoro-n-dodecanoic acid	.05000
Perfluoro-n-heptanoic Acid	.05000
Perfluoro-n-hexanoic acid	.05050
Perfluoro-n-octanoic Acid	.05000
Perfluorononanoic Acid	.05000
Perfluoro-n-pentanoic acid	.05000
Perfluoro-n-tetradecanoic acid	.05000
Perfluoro-n-tridecanoic acid	.05000
Perfluoro-n-undecanoic acid	.05000
Sodium perfluoro-1-pentanesulfonate	.05000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
170724-01	Pipette	C0982448K

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 8/20/2018 **Expiration Date:** 8/20/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 8/21/2018 7:17:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB70**

Description: PFAS - DoD High ICAL Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180705-02	PFOA - ICAL Mix	Neat	~1.000000	06/19/23	---	---	1000 uL	1	20	~0.0500

Solution Prepared By: Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date:**	10/1/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:03:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB70

Description: PFAS - DoD High ICAL Stock

Stock Id: 180705-02

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	1000	1.01	1	100.000	1	20	0.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	1000	1.01	1	100.000	1	20	0.05050
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-decanesulfonate	1000	1.01	1	100.000	1	20	0.05050
(NA) Perfluoro-1-heptanesulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-nonanesulfonate	1000	1.01	1	100.000	1	20	0.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-butanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-hexanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-octanesulfonamide	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-octanesulfonate	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-butanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-decanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-dodecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-heptanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-hexanoic acid	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-n-octanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluorononanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-pentanoic acid	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-n-tetradecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-tridecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-undecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Sodium perfluoro-1-pentanesulfonate	1000	1.00	1	100.000	1	20	0.05000

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.05050
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.05000
(Na) Perfluoro-1-decanesulfonate	.05050
(NA) Perfluoro-1-heptanesulfonate	.05000
(Na) Perfluoro-1-nonanesulfonate	.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.05000
Perfluoro-1-butanesulfonate	.05050

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:03:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB70**

Description: PFAS - DoD High ICAL Stock

Perfluoro-1-hexanesulfonate	.05050
Perfluoro-1-octanesulfonamide	.05000
Perfluoro-1-octanesulfonate	.05000
Perfluoro-n-butanoic Acid	.05000
Perfluoro-n-decanoic Acid	.05000
Perfluoro-n-dodecanoic acid	.05000
Perfluoro-n-heptanoic Acid	.05000
Perfluoro-n-hexanoic acid	.05050
Perfluoro-n-octanoic Acid	.05000
Perfluorononanoic Acid	.05000
Perfluoro-n-pentanoic acid	.05050
Perfluoro-n-tetradecanoic acid	.05000
Perfluoro-n-tridecanoic acid	.05000
Perfluoro-n-undecanoic acid	.05000
Sodium perfluoro-1-pentanesulfonate	.05000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180705-02	Pipette	B820865811

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:03:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB71**

Description: PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180726-05	Mass-labelled PFAS Extraction Standard Solution	Neat	~1.000000	02/07/23	---	---	1000 uL	1	20	~0.0500

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	10/1/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 96/4 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB71

Description: PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)

Stock Id: 180726-05

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	1000	0.94	1	100.000	1	20	0.04675
13C2-6:2FTS	1000	0.95	1	100.000	1	20	0.04745
13C2-8:2FTS	1000	0.96	1	100.000	1	20	0.04790
13C2-PFDoA	1000	1.00	1	100.000	1	20	0.05000
13C2-PFTeDA	1000	1.00	1	100.000	1	20	0.05000
13C3-PFBS	1000	0.93	1	100.000	1	20	0.04645
13C3-PFHxS	1000	0.95	1	100.000	1	20	0.04730
13C4-PFBA	1000	1.00	1	100.000	1	20	0.05000
13C4-PFHpA	1000	1.00	1	100.000	1	20	0.05000
13C5-PFHxA	1000	1.00	1	100.000	1	20	0.05000
13C5-PFPeA	1000	1.00	1	100.000	1	20	0.05000
13C6-PFDA	1000	1.00	1	100.000	1	20	0.05000
13C7-PFUnA	1000	1.00	1	100.000	1	20	0.05000
13C8-FOSA	1000	1.00	1	100.000	1	20	0.05000
13C8-PFOA	1000	1.00	1	100.000	1	20	0.05000
13C8-PFOS	1000	0.96	1	100.000	1	20	0.04785
13C9-PFNA	1000	1.00	1	100.000	1	20	0.05000
d3-MeFOSAA	1000	1.00	1	100.000	1	20	0.05000
d5-EtFOSAA	1000	1.00	1	100.000	1	20	0.05000

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-4:2FTS	.04675
13C2-6:2FTS	.04745
13C2-8:2FTS	.04790
13C2-PFDoA	.05000
13C2-PFTeDA	.05000
13C3-PFBS	.04645
13C3-PFHxS	.04730
13C4-PFBA	.05000
13C4-PFHpA	.05000
13C5-PFHxA	.05000
13C5-PFPeA	.05000
13C6-PFDA	.05000
13C7-PFUnA	.05000
13C8-FOSA	.05000

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB71

Description: PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)

13C8-PFOA	.05000
13C8-PFOS	.04785
13C9-PFNA	.05000
d3-MeFOSAA	.05000
d5-EtFOSAA	.05000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180726-05	Pipette	B820865811

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB72**

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	2500 uL	1	25	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	10/1/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB72

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	2500	0.05	---	---	1	25	0.00468
13C2-6:2FTS	2500	0.05	---	---	1	25	0.00475
13C2-8:2FTS	2500	0.05	---	---	1	25	0.00479
13C2-PFDoA	2500	0.05	---	---	1	25	0.00500
13C2-PFTeDA	2500	0.05	---	---	1	25	0.00500
13C3-PFBS	2500	0.05	---	---	1	25	0.00465
13C3-PFHxS	2500	0.05	---	---	1	25	0.00473
13C4-PFBA	2500	0.05	---	---	1	25	0.00500
13C4-PFHpA	2500	0.05	---	---	1	25	0.00500
13C5-PFHxA	2500	0.05	---	---	1	25	0.00500
13C5-PFPeA	2500	0.05	---	---	1	25	0.00500
13C6-PFDA	2500	0.05	---	---	1	25	0.00500
13C7-PFUnA	2500	0.05	---	---	1	25	0.00500
13C8-FOSA	2500	0.05	---	---	1	25	0.00500
13C8-PFOA	2500	0.05	---	---	1	25	0.00500
13C8-PFOS	2500	0.05	---	---	1	25	0.00478
13C9-PFNA	2500	0.05	---	---	1	25	0.00500
d3-MeFOSAA	2500	0.05	---	---	1	25	0.00500
d5-EtFOSAA	2500	0.05	---	---	1	25	0.00500

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-4:2FTS	.00468
13C2-6:2FTS	.00475
13C2-8:2FTS	.00479
13C2-PFDoA	.00500
13C2-PFTeDA	.00500
13C3-PFBS	.00465
13C3-PFHxS	.00473
13C4-PFBA	.00500
13C4-PFHpA	.00500
13C5-PFHxA	.00500
13C5-PFPeA	.00500
13C6-PFDA	.00500
13C7-PFUnA	.00500
13C8-FOSA	.00500

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB72

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

13C8-PFOA	.00500
13C8-PFOS	.00478
13C9-PFNA	.00500
d3-MeFOSAA	.00500

Syringes/Pipettes:

Solution Prepared By:	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date:**	10/1/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/4/2018 2:44:00 PM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB73**

Description: PFAS - DoD Calibration L1

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JY23	PFAS - DoD Low ICAL Stock	Solution	~0	07/16/19	---	---	200 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date**	7/16/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB73

Description: PFAS - DoD Calibration L1

Stock Id: JY23

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	200	0.01	---	---	1	10	0.00010
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	200	0.01	---	---	1	10	0.00010
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	200	0.01	---	---	1	10	0.00010
(Na) Perfluoro-1-decanesulfonate	200	0.01	---	---	1	10	0.00010
(NA) Perfluoro-1-heptanesulfonate	200	0.01	---	---	1	10	0.00010
(Na) Perfluoro-1-nonanesulfonate	200	0.01	---	---	1	10	0.00010
N-ethylperfluoro-octanesulfonamidoacetic acid	200	0.01	---	---	1	10	0.00010
N-methylperfluoro-1-octanesulfonamidoacetic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-1-butanesulfonate	200	0.01	---	---	1	10	0.00010
Perfluoro-1-hexanesulfonate	200	0.01	---	---	1	10	0.00010
Perfluoro-1-octanesulfonamide	200	0.01	---	---	1	10	0.00010
Perfluoro-1-octanesulfonate	200	0.01	---	---	1	10	0.00010
Perfluoro-n-butanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-decanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-dodecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-heptanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-hexanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-octanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluorononanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-pentanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-tetradecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-tridecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-undecanoic acid	200	0.01	---	---	1	10	0.00010
Sodium perfluoro-1-pentanesulfonate	200	0.01	---	---	1	10	0.00010

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB73**

Description: PFAS - DoD Calibration L1

13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00010
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00010
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00010
(Na) Perfluoro-1-decanesulfonate	.00010
(NA) Perfluoro-1-heptanesulfonate	.00010
(Na) Perfluoro-1-nonanesulfonate	.00010
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB73

Description: PFAS - DoD Calibration L1

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00010
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00010
Perfluoro-1-butanesulfonate	.00010
Perfluoro-1-hexanesulfonate	.00010
Perfluoro-1-octanesulfonamide	.00010
Perfluoro-1-octanesulfonate	.00010
Perfluoro-n-butanoic Acid	.00010
Perfluoro-n-decanoic Acid	.00010
Perfluoro-n-dodecanoic acid	.00010
Perfluoro-n-heptanoic Acid	.00010
Perfluoro-n-hexanoic acid	.00010
Perfluoro-n-octanoic Acid	.00010
Perfluorononanoic Acid	.00010
Perfluoro-n-pentanoic acid	.00010
Perfluoro-n-tetradecanoic acid	.00010
Perfluoro-n-tridecanoic acid	.00010
Perfluoro-n-undecanoic acid	.00010
Sodium perfluoro-1-pentanesulfonate	.00010

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY23	Pipette	B814657482
JY25	Pipette	B814659662
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB74**

Description: PFAS - DoD Calibration L2

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JY23	PFAS - DoD Low ICAL Stock	Solution	~0	07/16/19	---	---	500 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date**	7/16/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB74

Description: PFAS - DoD Calibration L2

Stock Id: JY23

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	500	0.01	---	---	1	10	0.00025
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	500	0.01	---	---	1	10	0.00025
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	500	0.01	---	---	1	10	0.00025
(Na) Perfluoro-1-decanesulfonate	500	0.01	---	---	1	10	0.00025
(NA) Perfluoro-1-heptanesulfonate	500	0.01	---	---	1	10	0.00025
(Na) Perfluoro-1-nonanesulfonate	500	0.01	---	---	1	10	0.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	500	0.01	---	---	1	10	0.00025
N-methylperfluoro-1-octanesulfonamidoacetic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-1-butanesulfonate	500	0.01	---	---	1	10	0.00025
Perfluoro-1-hexanesulfonate	500	0.01	---	---	1	10	0.00025
Perfluoro-1-octanesulfonamide	500	0.01	---	---	1	10	0.00025
Perfluoro-1-octanesulfonate	500	0.01	---	---	1	10	0.00025
Perfluoro-n-butanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-decanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-dodecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-heptanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-hexanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-octanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluorononanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-pentanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-tetradecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-tridecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-undecanoic acid	500	0.01	---	---	1	10	0.00025
Sodium perfluoro-1-pentanesulfonate	500	0.01	---	---	1	10	0.00025

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB74

Description: PFAS - DoD Calibration L2

13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00025
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00025
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00025
(Na) Perfluoro-1-decanesulfonate	.00025
(NA) Perfluoro-1-heptanesulfonate	.00025
(Na) Perfluoro-1-nonanesulfonate	.00025
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB74**

Description: PFAS - DoD Calibration L2

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00025
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00025
Perfluoro-1-butanesulfonate	.00025
Perfluoro-1-hexanesulfonate	.00025
Perfluoro-1-octanesulfonamide	.00025
Perfluoro-1-octanesulfonate	.00025
Perfluoro-n-butanoic Acid	.00025
Perfluoro-n-decanoic Acid	.00025
Perfluoro-n-dodecanoic acid	.00025
Perfluoro-n-heptanoic Acid	.00025
Perfluoro-n-hexanoic acid	.00025
Perfluoro-n-octanoic Acid	.00025
Perfluorononanoic Acid	.00025
Perfluoro-n-pentanoic acid	.00025
Perfluoro-n-tetradecanoic acid	.00025
Perfluoro-n-tridecanoic acid	.00025
Perfluoro-n-undecanoic acid	.00025
Sodium perfluoro-1-pentanesulfonate	.00025

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY23	Pipette	B820865811
JY25	Pipette	B814659662
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB75**

Description: PFAS - DoD Calibration L3

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB70	PFAS - DoD High ICAL Stock	Solution	~0	10/01/19	---	---	100 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB75

Description: PFAS - DoD Calibration L3

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB70

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	100	0.05	---	---	1	10	0.00051
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	100	0.05	---	---	1	10	0.00051
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	100	0.05	---	---	1	10	0.00050
(Na) Perfluoro-1-decanesulfonate	100	0.05	---	---	1	10	0.00051
(NA) Perfluoro-1-heptanesulfonate	100	0.05	---	---	1	10	0.00050
(Na) Perfluoro-1-nonanesulfonate	100	0.05	---	---	1	10	0.00051
N-ethylperfluoro-octanesulfonamidoacetic acid	100	0.05	---	---	1	10	0.00050
N-methylperfluoro-1-octanesulfonamidoacetic acid	100	0.05	---	---	1	10	0.00050
Perfluoro-1-butanesulfonate	100	0.05	---	---	1	10	0.00051
Perfluoro-1-hexanesulfonate	100	0.05	---	---	1	10	0.00051
Perfluoro-1-octanesulfonamide	100	0.05	---	---	1	10	0.00050
Perfluoro-1-octanesulfonate	100	0.05	---	---	1	10	0.00050
Perfluoro-n-butanoic Acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-decanoic Acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-dodecanoic acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-heptanoic Acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-hexanoic acid	100	0.05	---	---	1	10	0.00051
Perfluoro-n-octanoic Acid	100	0.05	---	---	1	10	0.00050
Perfluorononanoic Acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-pentanoic acid	100	0.05	---	---	1	10	0.00051
Perfluoro-n-tetradecanoic acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-tridecanoic acid	100	0.05	---	---	1	10	0.00050
Perfluoro-n-undecanoic acid	100	0.05	---	---	1	10	0.00050
Sodium perfluoro-1-pentanesulfonate	100	0.05	---	---	1	10	0.00050

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB75

Description: PFAS - DoD Calibration L3

13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00051
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00051
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00050
(Na) Perfluoro-1-decanesulfonate	.00051
(NA) Perfluoro-1-heptanesulfonate	.00050
(Na) Perfluoro-1-nonanesulfonate	.00051
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB75**

Description: PFAS - DoD Calibration L3

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00050
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00050
Perfluoro-1-butanesulfonate	.00051
Perfluoro-1-hexanesulfonate	.00051
Perfluoro-1-octanesulfonamide	.00050
Perfluoro-1-octanesulfonate	.00050
Perfluoro-n-butanoic Acid	.00050
Perfluoro-n-decanoic Acid	.00050
Perfluoro-n-dodecanoic acid	.00050
Perfluoro-n-heptanoic Acid	.00050
Perfluoro-n-hexanoic acid	.00051
Perfluoro-n-octanoic Acid	.00050
Perfluorononanoic Acid	.00050
Perfluoro-n-pentanoic acid	.00051
Perfluoro-n-tetradecanoic acid	.00050
Perfluoro-n-tridecanoic acid	.00050
Perfluoro-n-undecanoic acid	.00050
Sodium perfluoro-1-pentanesulfonate	.00050

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814659662
KB70	Pipette	B814659662
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB76**

Description: PFAS - DoD Calibration L4

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB70	PFAS - DoD High ICAL Stock	Solution	~0	10/01/19	---	---	1000 uL	1	50	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	250 uL	1	50	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	250 uL	1	50	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 2 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB76

Description: PFAS - DoD Calibration L4

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	250	0.05	---	---	1	50	0.00025
13C2-PFOA	250	0.05	---	---	1	50	0.00025
13C3-PFBA	250	0.05	---	---	1	50	0.00025
13C4-PFOS	250	0.05	---	---	1	50	0.00024

Stock Id: KB70

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	1000	0.05	---	---	1	50	0.00101
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	1000	0.05	---	---	1	50	0.00101
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	1000	0.05	---	---	1	50	0.00100
(Na) Perfluoro-1-decanesulfonate	1000	0.05	---	---	1	50	0.00101
(NA) Perfluoro-1-heptanesulfonate	1000	0.05	---	---	1	50	0.00100
(Na) Perfluoro-1-nonanesulfonate	1000	0.05	---	---	1	50	0.00101
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	0.05	---	---	1	50	0.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-1-butanesulfonate	1000	0.05	---	---	1	50	0.00101
Perfluoro-1-hexanesulfonate	1000	0.05	---	---	1	50	0.00101
Perfluoro-1-octanesulfonamide	1000	0.05	---	---	1	50	0.00100
Perfluoro-1-octanesulfonate	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-butanoic Acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-decanoic Acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-dodecanoic acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-heptanoic Acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-hexanoic acid	1000	0.05	---	---	1	50	0.00101
Perfluoro-n-octanoic Acid	1000	0.05	---	---	1	50	0.00100
Perfluorononanoic Acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-pentanoic acid	1000	0.05	---	---	1	50	0.00101
Perfluoro-n-tetradecanoic acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-tridecanoic acid	1000	0.05	---	---	1	50	0.00100
Perfluoro-n-undecanoic acid	1000	0.05	---	---	1	50	0.00100
Sodium perfluoro-1-pentanesulfonate	1000	0.05	---	---	1	50	0.00100

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	250	0.05	---	---	1	50	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB76**

Description: PFAS - DoD Calibration L4

13C2-6:2FTS	250	0.05	---	---	1	50	0.00024
13C2-8:2FTS	250	0.05	---	---	1	50	0.00024
13C2-PFDoA	250	0.05	---	---	1	50	0.00025
13C2-PFTeDA	250	0.05	---	---	1	50	0.00025
13C3-PFBS	250	0.05	---	---	1	50	0.00023
13C3-PFHxS	250	0.05	---	---	1	50	0.00024
13C4-PFBA	250	0.05	---	---	1	50	0.00025
13C4-PFHpA	250	0.05	---	---	1	50	0.00025
13C5-PFHxA	250	0.05	---	---	1	50	0.00025
13C5-PFPeA	250	0.05	---	---	1	50	0.00025
13C6-PFDA	250	0.05	---	---	1	50	0.00025
13C7-PFUnA	250	0.05	---	---	1	50	0.00025
13C8-FOSA	250	0.05	---	---	1	50	0.00025
13C8-PFOA	250	0.05	---	---	1	50	0.00025
13C8-PFOS	250	0.05	---	---	1	50	0.00024
13C9-PFNA	250	0.05	---	---	1	50	0.00025
d3-MeFOSAA	250	0.05	---	---	1	50	0.00025
d5-EtFOSAA	250	0.05	---	---	1	50	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00101
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00101
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00100
(Na) Perfluoro-1-decanesulfonate	.00101
(NA) Perfluoro-1-heptanesulfonate	.00100
(Na) Perfluoro-1-nonanesulfonate	.00101
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB76**

Description: PFAS - DoD Calibration L4

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00100
Perfluoro-1-butanesulfonate	.00101
Perfluoro-1-hexanesulfonate	.00101
Perfluoro-1-octanesulfonamide	.00100
Perfluoro-1-octanesulfonate	.00100
Perfluoro-n-butanoic Acid	.00100
Perfluoro-n-decanoic Acid	.00100
Perfluoro-n-dodecanoic acid	.00100
Perfluoro-n-heptanoic Acid	.00100
Perfluoro-n-hexanoic acid	.00101
Perfluoro-n-octanoic Acid	.00100
Perfluorononanoic Acid	.00100
Perfluoro-n-pentanoic acid	.00101
Perfluoro-n-tetradecanoic acid	.00100
Perfluoro-n-tridecanoic acid	.00100
Perfluoro-n-undecanoic acid	.00100
Sodium perfluoro-1-pentanesulfonate	.00100

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814657482
KB70	Pipette	B820865811
KB71	Pipette	B814657482

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:40:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB77**

Description: PFAS - DoD Calibration L5

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB70	PFAS - DoD High ICAL Stock	Solution	~0	10/01/19	---	---	2500 uL	1	50	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	250 uL	1	50	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	250 uL	1	50	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 2 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB77

Description: PFAS - DoD Calibration L5

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	250	0.05	---	---	1	50	0.00025
13C2-PFOA	250	0.05	---	---	1	50	0.00025
13C3-PFBA	250	0.05	---	---	1	50	0.00025
13C4-PFOS	250	0.05	---	---	1	50	0.00024

Stock Id: KB70

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	2500	0.05	---	---	1	50	0.00253
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	2500	0.05	---	---	1	50	0.00253
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	2500	0.05	---	---	1	50	0.00250
(Na) Perfluoro-1-decanesulfonate	2500	0.05	---	---	1	50	0.00253
(NA) Perfluoro-1-heptanesulfonate	2500	0.05	---	---	1	50	0.00250
(Na) Perfluoro-1-nonanesulfonate	2500	0.05	---	---	1	50	0.00253
N-ethylperfluoro-octanesulfonamidoacetic acid	2500	0.05	---	---	1	50	0.00250
N-methylperfluoro-1-octanesulfonamidoacetic acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-1-butanesulfonate	2500	0.05	---	---	1	50	0.00253
Perfluoro-1-hexanesulfonate	2500	0.05	---	---	1	50	0.00253
Perfluoro-1-octanesulfonamide	2500	0.05	---	---	1	50	0.00250
Perfluoro-1-octanesulfonate	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-butanoic Acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-decanoic Acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-dodecanoic acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-heptanoic Acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-hexanoic acid	2500	0.05	---	---	1	50	0.00253
Perfluoro-n-octanoic Acid	2500	0.05	---	---	1	50	0.00250
Perfluorononanoic Acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-pentanoic acid	2500	0.05	---	---	1	50	0.00253
Perfluoro-n-tetradecanoic acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-tridecanoic acid	2500	0.05	---	---	1	50	0.00250
Perfluoro-n-undecanoic acid	2500	0.05	---	---	1	50	0.00250
Sodium perfluoro-1-pentanesulfonate	2500	0.05	---	---	1	50	0.00250

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	250	0.05	---	---	1	50	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB77

Description: PFAS - DoD Calibration L5

13C2-6:2FTS	250	0.05	---	---	1	50	0.00024
13C2-8:2FTS	250	0.05	---	---	1	50	0.00024
13C2-PFDoA	250	0.05	---	---	1	50	0.00025
13C2-PFTeDA	250	0.05	---	---	1	50	0.00025
13C3-PFBS	250	0.05	---	---	1	50	0.00023
13C3-PFHxS	250	0.05	---	---	1	50	0.00024
13C4-PFBA	250	0.05	---	---	1	50	0.00025
13C4-PFHpA	250	0.05	---	---	1	50	0.00025
13C5-PFHxA	250	0.05	---	---	1	50	0.00025
13C5-PFPeA	250	0.05	---	---	1	50	0.00025
13C6-PFDA	250	0.05	---	---	1	50	0.00025
13C7-PFUnA	250	0.05	---	---	1	50	0.00025
13C8-FOSA	250	0.05	---	---	1	50	0.00025
13C8-PFOA	250	0.05	---	---	1	50	0.00025
13C8-PFOS	250	0.05	---	---	1	50	0.00024
13C9-PFNA	250	0.05	---	---	1	50	0.00025
d3-MeFOSAA	250	0.05	---	---	1	50	0.00025
d5-EtFOSAA	250	0.05	---	---	1	50	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00253
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00253
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00250
(Na) Perfluoro-1-decanesulfonate	.00253
(NA) Perfluoro-1-heptanesulfonate	.00250
(Na) Perfluoro-1-nonanesulfonate	.00253
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB77**

Description: PFAS - DoD Calibration L5

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00250
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00250
Perfluoro-1-butanesulfonate	.00253
Perfluoro-1-hexanesulfonate	.00253
Perfluoro-1-octanesulfonamide	.00250
Perfluoro-1-octanesulfonate	.00250
Perfluoro-n-butanoic Acid	.00250
Perfluoro-n-decanoic Acid	.00250
Perfluoro-n-dodecanoic acid	.00250
Perfluoro-n-heptanoic Acid	.00250
Perfluoro-n-hexanoic acid	.00253
Perfluoro-n-octanoic Acid	.00250
Perfluorononanoic Acid	.00250
Perfluoro-n-pentanoic acid	.00253
Perfluoro-n-tetradecanoic acid	.00250
Perfluoro-n-tridecanoic acid	.00250
Perfluoro-n-undecanoic acid	.00250
Sodium perfluoro-1-pentanesulfonate	.00250

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814657482
KB70	Pipette	OU16914
KB71	Pipette	B814657482

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 2 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB78**

Description: PFAS - DoD Calibraiton L6

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB70	PFAS - DoD High ICAL Stock	Solution	~0	10/01/19	---	---	2000 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date**	7/16/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB78

Description: PFAS - DoD Calibraiton L6

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB70

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	2000	0.05	---	---	1	10	0.01010
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	2000	0.05	---	---	1	10	0.01010
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	2000	0.05	---	---	1	10	0.01000
(Na) Perfluoro-1-decanesulfonate	2000	0.05	---	---	1	10	0.01010
(NA) Perfluoro-1-heptanesulfonate	2000	0.05	---	---	1	10	0.01000
(Na) Perfluoro-1-nonanesulfonate	2000	0.05	---	---	1	10	0.01010
N-ethylperfluoro-octanesulfonamidoacetic acid	2000	0.05	---	---	1	10	0.01000
N-methylperfluoro-1-octanesulfonamidoacetic acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-1-butanesulfonate	2000	0.05	---	---	1	10	0.01010
Perfluoro-1-hexanesulfonate	2000	0.05	---	---	1	10	0.01010
Perfluoro-1-octanesulfonamide	2000	0.05	---	---	1	10	0.01000
Perfluoro-1-octanesulfonate	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-butanoic Acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-decanoic Acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-dodecanoic acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-heptanoic Acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-hexanoic acid	2000	0.05	---	---	1	10	0.01010
Perfluoro-n-octanoic Acid	2000	0.05	---	---	1	10	0.01000
Perfluorononanoic Acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-pentanoic acid	2000	0.05	---	---	1	10	0.01010
Perfluoro-n-tetradecanoic acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-tridecanoic acid	2000	0.05	---	---	1	10	0.01000
Perfluoro-n-undecanoic acid	2000	0.05	---	---	1	10	0.01000
Sodium perfluoro-1-pentanesulfonate	2000	0.05	---	---	1	10	0.01000

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB78**

Description: PFAS - DoD Calibraiton L6

13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.01010
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.01010
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.01000
(Na) Perfluoro-1-decanesulfonate	.01010
(NA) Perfluoro-1-heptanesulfonate	.01000
(Na) Perfluoro-1-nonanesulfonate	.01010
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB78**

Description: PFAS - DoD Calibraiton L6

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.01000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.01000
Perfluoro-1-butanesulfonate	.01010
Perfluoro-1-hexanesulfonate	.01010
Perfluoro-1-octanesulfonamide	.01000
Perfluoro-1-octanesulfonate	.01000
Perfluoro-n-butanoic Acid	.01000
Perfluoro-n-decanoic Acid	.01000
Perfluoro-n-dodecanoic acid	.01000
Perfluoro-n-heptanoic Acid	.01000
Perfluoro-n-hexanoic acid	.01010
Perfluoro-n-octanoic Acid	.01000
Perfluorononanoic Acid	.01000
Perfluoro-n-pentanoic acid	.01010
Perfluoro-n-tetradecanoic acid	.01000
Perfluoro-n-tridecanoic acid	.01000
Perfluoro-n-undecanoic acid	.01000
Sodium perfluoro-1-pentanesulfonate	.01000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814659662
KB70	Pipette	OU16914
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB79**

Description: PFAS - DoD Calibration L7

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB70	PFAS - DoD High ICAL Stock	Solution	~0	10/01/19	---	---	4000 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB79

Description: PFAS - DoD Calibration L7

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB70

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	4000	0.05	---	---	1	10	0.02020
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	4000	0.05	---	---	1	10	0.02020
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	4000	0.05	---	---	1	10	0.02000
(Na) Perfluoro-1-decanesulfonate	4000	0.05	---	---	1	10	0.02020
(NA) Perfluoro-1-heptanesulfonate	4000	0.05	---	---	1	10	0.02000
(Na) Perfluoro-1-nonanesulfonate	4000	0.05	---	---	1	10	0.02020
N-ethylperfluoro-octanesulfonamidoacetic acid	4000	0.05	---	---	1	10	0.02000
N-methylperfluoro-1-octanesulfonamidoacetic acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-1-butanesulfonate	4000	0.05	---	---	1	10	0.02020
Perfluoro-1-hexanesulfonate	4000	0.05	---	---	1	10	0.02020
Perfluoro-1-octanesulfonamide	4000	0.05	---	---	1	10	0.02000
Perfluoro-1-octanesulfonate	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-butanoic Acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-decanoic Acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-dodecanoic acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-heptanoic Acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-hexanoic acid	4000	0.05	---	---	1	10	0.02020
Perfluoro-n-octanoic Acid	4000	0.05	---	---	1	10	0.02000
Perfluorononanoic Acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-pentanoic acid	4000	0.05	---	---	1	10	0.02020
Perfluoro-n-tetradecanoic acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-tridecanoic acid	4000	0.05	---	---	1	10	0.02000
Perfluoro-n-undecanoic acid	4000	0.05	---	---	1	10	0.02000
Sodium perfluoro-1-pentanesulfonate	4000	0.05	---	---	1	10	0.02000

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB79**

Description: PFAS - DoD Calibration L7

13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.02020
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.02020
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.02000
(Na) Perfluoro-1-decanesulfonate	.02020
(NA) Perfluoro-1-heptanesulfonate	.02000
(Na) Perfluoro-1-nonanesulfonate	.02020
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB79**

Description: PFAS - DoD Calibration L7

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.02000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.02000
Perfluoro-1-butanesulfonate	.02020
Perfluoro-1-hexanesulfonate	.02020
Perfluoro-1-octanesulfonamide	.02000
Perfluoro-1-octanesulfonate	.02000
Perfluoro-n-butanoic Acid	.02000
Perfluoro-n-decanoic Acid	.02000
Perfluoro-n-dodecanoic acid	.02000
Perfluoro-n-heptanoic Acid	.02000
Perfluoro-n-hexanoic acid	.02020
Perfluoro-n-octanoic Acid	.02000
Perfluorononanoic Acid	.02000
Perfluoro-n-pentanoic acid	.02020
Perfluoro-n-tetradecanoic acid	.02000
Perfluoro-n-tridecanoic acid	.02000
Perfluoro-n-undecanoic acid	.02000
Sodium perfluoro-1-pentanesulfonate	.02000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814659662
KB70	Pipette	OU16914
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB80**

Description: PFAS - DoD Instrument Blank

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/1/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB80

Description: PFAS - DoD Instrument Blank

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023
13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB80**

Description: PFAS - DoD Instrument Blank

13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025
13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814659662
KB71	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB81**

Description: PFAS - DoD ICC

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB82	PFAS - DoD Second Source LCS/MS Solution	Solution	~0	10/01/19	---	---	200 uL	1	10	~0.0000
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	50 uL	1	10	~0.0000
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date**	7/16/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB81

Description: PFAS - DoD ICC

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.05	---	---	1	10	0.00025
13C2-PFOA	50	0.05	---	---	1	10	0.00025
13C3-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFOS	50	0.05	---	---	1	10	0.00024

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	50	0.05	---	---	1	10	0.00023
13C2-6:2FTS	50	0.05	---	---	1	10	0.00024
13C2-8:2FTS	50	0.05	---	---	1	10	0.00024
13C2-PFDoA	50	0.05	---	---	1	10	0.00025
13C2-PFTeDA	50	0.05	---	---	1	10	0.00025
13C3-PFBS	50	0.05	---	---	1	10	0.00023
13C3-PFHxS	50	0.05	---	---	1	10	0.00024
13C4-PFBA	50	0.05	---	---	1	10	0.00025
13C4-PFHpA	50	0.05	---	---	1	10	0.00025
13C5-PFHxA	50	0.05	---	---	1	10	0.00025
13C5-PFPeA	50	0.05	---	---	1	10	0.00025
13C6-PFDA	50	0.05	---	---	1	10	0.00025
13C7-PFUnA	50	0.05	---	---	1	10	0.00025
13C8-FOSA	50	0.05	---	---	1	10	0.00025
13C8-PFOA	50	0.05	---	---	1	10	0.00025
13C8-PFOS	50	0.05	---	---	1	10	0.00024
13C9-PFNA	50	0.05	---	---	1	10	0.00025
d3-MeFOSAA	50	0.05	---	---	1	10	0.00025
d5-EtFOSAA	50	0.05	---	---	1	10	0.00025

Stock Id: KB82

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	200	0.05	---	---	1	10	0.00101
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	200	0.05	---	---	1	10	0.00100
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	200	0.05	---	---	1	10	0.00100
(Na) Perfluoro-1-decanesulfonate	200	0.05	---	---	1	10	0.00101
(NA) Perfluoro-1-heptanesulfonate	200	0.05	---	---	1	10	0.00100
(Na) Perfluoro-1-nonanesulfonate	200	0.05	---	---	1	10	0.00101

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB81**

Description: PFAS - DoD ICC

N-ethylperfluoro-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-1-butanesulfonate	200	0.05	---	---	1	10	0.00101
Perfluoro-1-hexanesulfonate	200	0.05	---	---	1	10	0.00101
Perfluoro-1-octanesulfonamide	200	0.05	---	---	1	10	0.00100
Perfluoro-1-octanesulfonate	200	0.05	---	---	1	10	0.00100
Perfluoro-n-butanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-decanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-dodecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-heptanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-hexanoic acid	200	0.05	---	---	1	10	0.00101
Perfluoro-n-octanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluorononanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-pentanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tetradecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tridecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-undecanoic acid	200	0.05	---	---	1	10	0.00100
Sodium perfluoro-1-pentanesulfonate	200	0.05	---	---	1	10	0.00100

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.00101
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.00100
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.00100
(Na) Perfluoro-1-decanesulfonate	.00101
(NA) Perfluoro-1-heptanesulfonate	.00100
(Na) Perfluoro-1-nonanesulfonate	.00101
13C2-4:2FTS	.00023
13C2-6:2FTS	.00024
13C2-8:2FTS	.00024
13C2-PFDA	.00025
13C2-PFDoA	.00025
13C2-PFOA	.00025
13C2-PFTeDA	.00025
13C3-PFBA	.00025
13C3-PFBS	.00023
13C3-PFHxS	.00024
13C4-PFBA	.00025
13C4-PFHpA	.00025

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KB81**

Description: PFAS - DoD ICC

13C4-PFOS	.00024
13C5-PFHxA	.00025
13C5-PFPeA	.00025
13C6-PFDA	.00025
13C7-PFUnA	.00025
13C8-FOSA	.00025
13C8-PFOA	.00025
13C8-PFOS	.00024
13C9-PFNA	.00025
d3-MeFOSAA	.00025
d5-EtFOSAA	.00025
N-ethylperfluoro-octanesulfonamidoacetic acid	.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00100
Perfluoro-1-butanesulfonate	.00101
Perfluoro-1-hexanesulfonate	.00101
Perfluoro-1-octanesulfonamide	.00100
Perfluoro-1-octanesulfonate	.00100
Perfluoro-n-butanoic Acid	.00100
Perfluoro-n-decanoic Acid	.00100
Perfluoro-n-dodecanoic acid	.00100
Perfluoro-n-heptanoic Acid	.00100
Perfluoro-n-hexanoic acid	.00101
Perfluoro-n-octanoic Acid	.00100
Perfluorononanoic Acid	.00100
Perfluoro-n-pentanoic acid	.00100
Perfluoro-n-tetradecanoic acid	.00100
Perfluoro-n-tridecanoic acid	.00100
Perfluoro-n-undecanoic acid	.00100
Sodium perfluoro-1-pentanesulfonate	.00100

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	B814659662
KB71	Pipette	B814659662
KB82	Pipette	B814657482

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:41:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB82**

Description: PFAS - DoD Second Source LCS/MS Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
170724-01	PFOA - 2nd Source	Neat	~1.000000	03/22/22	---	---	1000 uL	1	20	~0.0500

Solution Prepared By: Schultz, Stephanie	**Date Prepared:** 10/1/2018	**Expiration Date:**	10/1/2019
Solution Volume 40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:05:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB82

Description: PFAS - DoD Second Source LCS/MS Solution

Stock Id: 170724-01

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	1000	1.01	1	100.000	1	20	0.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-decanesulfonate	1000	1.01	1	100.000	1	20	0.05050
(NA) Perfluoro-1-heptanesulfonate	1000	1.00	1	100.000	1	20	0.05000
(Na) Perfluoro-1-nonanesulfonate	1000	1.01	1	100.000	1	20	0.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-butanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-hexanesulfonate	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-1-octanesulfonamide	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-1-octanesulfonate	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-butanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-decanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-dodecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-heptanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-hexanoic acid	1000	1.01	1	100.000	1	20	0.05050
Perfluoro-n-octanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluorononanoic Acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-pentanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-tetradecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-tridecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Perfluoro-n-undecanoic acid	1000	1.00	1	100.000	1	20	0.05000
Sodium perfluoro-1-pentanesulfonate	1000	1.00	1	100.000	1	20	0.05000

Final Concentrations:

Analyte:	Conc (ug/mL):
(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate	.05050
(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate	.05000
(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate	.05000
(Na) Perfluoro-1-decanesulfonate	.05050
(NA) Perfluoro-1-heptanesulfonate	.05000
(Na) Perfluoro-1-nonanesulfonate	.05050
N-ethylperfluoro-octanesulfonamidoacetic acid	.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.05000
Perfluoro-1-butanesulfonate	.05050

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:05:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB82

Description: PFAS - DoD Second Source LCS/MS Solution

Perfluoro-1-hexanesulfonate	.05050
Perfluoro-1-octanesulfonamide	.05000
Perfluoro-1-octanesulfonate	.05000
Perfluoro-n-butanoic Acid	.05000
Perfluoro-n-decanoic Acid	.05000
Perfluoro-n-dodecanoic acid	.05000
Perfluoro-n-heptanoic Acid	.05000
Perfluoro-n-hexanoic acid	.05050
Perfluoro-n-octanoic Acid	.05000
Perfluorononanoic Acid	.05000
Perfluoro-n-pentanoic acid	.05000
Perfluoro-n-tetradecanoic acid	.05000
Perfluoro-n-tridecanoic acid	.05000
Perfluoro-n-undecanoic acid	.05000
Sodium perfluoro-1-pentanesulfonate	.05000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
170724-01	Pipette	B820865811

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/1/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/Milli-q water

Approved By: Thorn, Jonathan **Date:** 10/12/2018 8:05:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KB89

Description: PFAS Branched Solution (~5,000 ng/L)

Stock Id: JX28

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	100	0.50	---	---	1	10	0.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	100	0.50	---	---	1	10	0.00500
Perfluoro-1-hexanesulfonate	100	0.50	---	---	1	10	0.00500
Perfluoro-1-octanesulfonate	100	0.50	---	---	1	10	0.00500
Perfluoro-n-octanoic Acid	100	0.50	---	---	1	10	0.00500

Final Concentrations:

Analyte:	Conc (ug/mL):
N-ethylperfluoro-octanesulfonamidoacetic acid	.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00500
Perfluoro-1-hexanesulfonate	.00500
Perfluoro-1-octanesulfonate	.00500
Perfluoro-n-octanoic Acid	.00500

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JX28	Pipette	B814659662

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/3/2018 **Expiration Date:** 6/18/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:43:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KB89**

Description: PFAS Branched Solution (~5,000 ng/L)

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JX28	PFAS Branched Standard Stock	Solution	~0	06/18/19	---	---	100 uL	1	10	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/3/2018	**Expiration Date**	6/18/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 80/20 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 9:43:00 AM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KC03**

Description: PFAS - DoD Internal Standard Spiking Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JY25	PFAS - DoD Internal Standard Stock Solution	Solution	~0	07/16/19	---	---	2500 uL	1	25	~0.0000

Solution Prepared By	Schultz, Stephanie	**Date Prepared:**	10/9/2018	**Expiration Date**	7/16/2019
Solution Volume	40 **mL X** 1 **Vials**	**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID: _____

Comment: 96/4 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 4:15:00 PM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KC03

Description: PFAS - DoD Internal Standard Spiking Solution

Stock Id: JY25

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	2500	0.05	---	---	1	25	0.00500
13C2-PFOA	2500	0.05	---	---	1	25	0.00500
13C3-PFBA	2500	0.05	---	---	1	25	0.00500
13C4-PFOS	2500	0.05	---	---	1	25	0.00479

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00500
13C2-PFOA	.00500
13C3-PFBA	.00500
13C4-PFOS	.00479

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JY25	Pipette	OU16914

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/9/2018 **Expiration Date:** 7/16/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/Milli-q water

Approved By: Schumitz, Denise **Date:** 10/9/2018 4:15:00 PM

BATTELLE
It can be done

Standard Solution Prep Form II

Approved: ☑

Standard Laboratory ID Number: **KC19**

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
KB71	PFAS - DoD High Level Labelled Extracted Internal Standards (SIS)	Solution	~0	10/01/19	---	---	2500 uL	1	25	~0.0000

Solution Prepared By: Schultz, Stephanie		**Date Prepared:**	10/16/2018	**Expiration Date:**	10/1/2019
Solution Volume 40 **mL X** 1 **Vials**		**Refrigerator/Freezer No:**	LC Laboratory: Refrigerator - R0107		

Balance ID:

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/18/2018 11:55:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: KC19

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

Stock Id: KB71

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-4:2FTS	2500	0.05	---	---	1	25	0.00468
13C2-6:2FTS	2500	0.05	---	---	1	25	0.00475
13C2-8:2FTS	2500	0.05	---	---	1	25	0.00479
13C2-PFDoA	2500	0.05	---	---	1	25	0.00500
13C2-PFTeDA	2500	0.05	---	---	1	25	0.00500
13C3-PFBS	2500	0.05	---	---	1	25	0.00465
13C3-PFHxS	2500	0.05	---	---	1	25	0.00473
13C4-PFBA	2500	0.05	---	---	1	25	0.00500
13C4-PFHpA	2500	0.05	---	---	1	25	0.00500
13C5-PFHxA	2500	0.05	---	---	1	25	0.00500
13C5-PFPeA	2500	0.05	---	---	1	25	0.00500
13C6-PFDA	2500	0.05	---	---	1	25	0.00500
13C7-PFUnA	2500	0.05	---	---	1	25	0.00500
13C8-FOSA	2500	0.05	---	---	1	25	0.00500
13C8-PFOA	2500	0.05	---	---	1	25	0.00500
13C8-PFOS	2500	0.05	---	---	1	25	0.00478
13C9-PFNA	2500	0.05	---	---	1	25	0.00500
d3-MeFOSAA	2500	0.05	---	---	1	25	0.00500
d5-EtFOSAA	2500	0.05	---	---	1	25	0.00500

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-4:2FTS	.00468
13C2-6:2FTS	.00475
13C2-8:2FTS	.00479
13C2-PFDoA	.00500
13C2-PFTeDA	.00500
13C3-PFBS	.00465
13C3-PFHxS	.00473
13C4-PFBA	.00500
13C4-PFHpA	.00500
13C5-PFHxA	.00500
13C5-PFPeA	.00500
13C6-PFDA	.00500
13C7-PFUnA	.00500
13C8-FOSA	.00500

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/16/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/18/2018 11:55:00 AM

BATTELLE
It can be done

Standard Solution Concentrations

Approved: ☑

Standard Laboratory ID Number: **KC19**

Description: PFAS - DoD Low Level Labelled Extracted Internal Standard (SIS)

13C8-PFOA	.00500
13C8-PFOS	.00478
13C9-PFNA	.00500
d3-MeFOSAA	.00500
d5-EtFOSAA	.00500

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
KB71	Pipette	OU16914

Solution Prepared By: Schultz, Stephanie **Date Prepared:** 10/16/2018 **Expiration Date:** 10/1/2019

Solution Volume 40 **mL X** 1 **Vials** **Refrigerator/Freezer No:** LC Laboratory: Refrigerator - R0107

Comment: 96/4 Methanol/milli-q water

Approved By: Schumitz, Denise **Date:** 10/18/2018 11:55:00 AM

BATTELLE
It can be done

BDO Id: 170724-01

Reagent Receipt Report

Approved: ☐ **Authorized** ☐

Name: PFOA- 2nd Source
Vendor: ABSOLUTE STANDARDS
Catalogue No: 99207
Type: Solution
Lot No: 032217
Quantity: 5 ea mL **% Moisture:**
Description: PFOA - 2nd Source

Received: 7/24/2017
Custodian: Schumitz, Matt
Expires: 3/22/2022
Consumed:
Stored In: LC Laboratory - F0111

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
(Na) 1H,1H,2H,2H-Perfluorodecane	39108-34-4	1.0100	100.00	--	--	☐			
(Na) 1H,1H,2H,2H-Perfluorohexane s	414911-30-1	1.0000	100.00	--	--	☐			
(Na) 1H,1H,2H,2H-Perfluorooctane s	27619-97-2	1.0000	100.00	--	--	☐			
(Na) Perfluoro-1-decanesulfonate	2806-15-7	1.0100	100.00	--	--	☐			
(NA) Perfluoro-1-heptanesulfonate	375-92-8	1.0000	100.00	--	--	☐			
(Na) Perfluoro-1-nonanesulfonate	98789-57-2	1.0100	100.00	--	--	☐			
N-ethylperfluoro-octanesulfonamidoa	2991-50-6	1.0000	100.00	--	--	☐			
N-methylperfluoro-1-octanesulfonami	2355-31-9	1.0000	100.00	--	--	☐			
Perfluoro-1-butanesulfonate	375-73-5	1.0100	100.00	--	--	☐			
Perfluoro-1-hexanesulfonate	355-46-4	1.0100	100.00	--	--	☐			
Perfluoro-1-octanesulfonamide	754-91-6	1.0000	100.00	--	--	☐			
Perfluoro-1-octanesulfonate	1763-23-1	1.0000	100.00	--	--	☐			
Perfluoro-n-butanoic Acid	375-22-4	1.0000	100.00	--	--	☐			
Perfluoro-n-decanoic Acid	335-76-2	1.0000	100.00	--	--	☐			
Perfluoro-n-dodecanoic acid	307-55-1	1.0000	100.00	--	--	☐			
Perfluoro-n-heptanoic Acid	375-85-9	1.0000	100.00	--	--	☐			
Perfluoro-n-hexanoic acid	307-24-4	1.0100	100.00	--	--	☐			
Perfluoro-n-octanoic Acid	335-67-1	1.0000	100.00	--	--	☐			
Perfluorononanoic Acid	375-95-1	1.0000	100.00	--	--	☐			
Perfluoro-n-pentanoic acid	2706-90-3	1.0000	100.00	--	--	☐			
Perfluoro-n-tetradecanoic acid	376-06-7	1.0000	100.00	--	--	☐			
Perfluoro-n-tridecanoic acid	72629-94-8	1.0000	100.00	--	--	☐			
Perfluoro-n-undecanoic acid	2058-94-8	1.0000	100.00	--	--	☐			
Sodium perfluoro-1-pentanesulfonate	2706-91-4	1.0000	100.00	--	--	☐			

Total Analytes: 24

Notes:

Approved by: _____ **Approved on:** _____

Authorized by: _____ **Authorized on:** _____

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ISO 9001 QS Registered
ISO 17025-34-35-43 Accredited
Scopes: http://AbsoluteStandards.com

CERTIFIED WEIGHT REPORT

17024-01

Part Number: 99207
Lot Number: 032217
Description: PFOA - DOD
24 components
Expiration Date: 032222
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 1.0
NIST Test ID#: 822-275872-11

Solvent(s):	Lot#	
Methanol (1 mM KOH)	031317	(98%)
2-Propanol	23214	(2%)

5E-05 Balance Uncertainty
0.007 Flask Uncertainty

Volume(s) shown below were combined and diluted to (mL): 50.0

Paul Barron		032217
Formulated By:	Paul Barron	DATE
Pedro Rentas		032217
Reviewed By:	Pedro L. Rentas	DATE


Note: All assigned values are anion concentrations.

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Perfluoro-n-butanoic acid	3670	PFBA0516	0.02	1.00	0.004	50.0	1.00	0.01	375-22-4	N/A	N/A
2. Perfluoro-n-pentanoic acid	3669	PFPeA0516	0.02	1.00	0.004	50.0	1.00	0.01	2706-90-3	N/A	N/A
3. Perfluorohexanoic acid	99199	030617	0.02	1.00	0.004	50.3	1.01	0.01	307-24-4	N/A	N/A
4. Perfluoroheptanoic acid	99197	030517	0.02	1.00	0.004	50.1	1.00	0.01	375-85-9	N/A	N/A
5. Perfluorooctanoic acid	99202	030617	0.02	1.00	0.004	50.2	1.00	0.01	335-67-1	N/A	ipr-rat 189mg/kg
6. Perfluorononanoic acid	99200	030617	0.02	1.00	0.004	50.1	1.00	0.01	375-95-1	N/A	N/A
7. Perfluorodecanoic acid	99195	030617	0.02	1.00	0.004	50.1	1.00	0.01	335-76-2	N/A	orl-rat 57mg/kg
8. Perfluoroundecanoic acid	99205	030617	0.02	1.00	0.004	50.1	1.00	0.01	2058-94-8	N/A	N/A
9. Tricosafluorododecanoic acid	99196	030617	0.02	1.00	0.004	50.1	1.00	0.01	307-55-1	N/A	N/A
10. Perfluorotridecanoic acid	99204	030617	0.02	1.00	0.004	50.1	1.00	0.01	72629-94-8	N/A	N/A
11. Perfluorotetradecanoic acid	99203	030617	0.02	1.00	0.004	50.1	1.00	0.01	376-06-7	N/A	N/A
12. Perfluoro-1-octanesulfonamide	3677	FOSA0916I	0.02	1.00	0.004	50.0	1.00	0.01	754-91-6	N/A	N/A
13. N-Methylperfluoro-1-octanesulfonamidoacetic acid	3667	NMeFOSAA0117	0.02	1.00	0.004	50.0	1.00	0.01	2355-31-9	N/A	N/A
14. N-Ethylperfluoro-1-octanesulfonamidoacetic acid	3664	NEtFOSAA0117	0.02	1.00	0.004	50.0	1.00	0.01	2991-50-6	N/A	N/A
15. Perfluorobutanesulfonic acid	99194	031017	0.02	1.00	0.004	50.7	1.01	0.01	375-73-5	N/A	N/A
16. Perfluoro-1-pentanesulfonic acid	3956	LPFPeS0117	0.0214	1.07	0.004	46.9	1.00	0.01	00-00-0	N/A	N/A
17. Perfluorohexanesulfonic acid (branched)	99198	030617	0.02	1.00	0.004	50.6	1.01	0.01	3871-99-6	N/A	N/A
18. Perfluoro-1-heptanesulfonic acid	3672	LPFHpS1016	0.021	1.05	0.004	47.6	1.00	0.01	375-92-8	N/A	N/A
19. Heptadecafluorooctanesulfonic acid (branched)	99201	030617	0.02	1.00	0.004	50.2	1.00	0.01	1763-23-1	N/A	N/A
20. Perfluoro-1-nonanesulfonic acid	3957	LPFNS0516	0.021	1.05	0.004	48.0	1.01	0.01	98789-57-2	N/A	N/A
21. Perfluoro-1-decanesulfonic acid	3671	LPFDS0217	0.021	1.05	0.004	48.2	1.01	0.01	2806-15-7	N/A	N/A
22. 1H,1H,2H,2H-Perfluorohexane sulfonic acid	3955	42FTS1216	0.0214	1.07	0.004	46.7	1.00	0.01	00-00-0	N/A	N/A
23. 1H,1H,2H,2H-Perfluorooctane sulfonic acid	3661	62FTS0616	0.021	1.05	0.004	47.4	1.00	0.01	27619-97-2	N/A	N/A
24. 1H,1H,2H,2H-Perfluorodecane sulfonic acid	3662	82FTS1216	0.021	1.05	0.004	47.9	1.01	0.01	39108-34-4	N/A	N/A

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Peak No.	Analyte	MSD RT (min.)
1	Perfluoro-n-butanoic acid (PFBA)	3.09
2	Perfluoro-n-pentanoic acid (PFPeA)	6.41
3	Perfluorobutanesulfonate (PFBS)	7.62
4	(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate (4:2FTS)	8.99
5	Perfluorohexanoic acid (PFHxA)	9.44
6	(Na) Perfluoro-1-pentanesulfonate (L-PFPeS)	10.24
7	Perfluoroheptanoic acid (PFHpA)	11.99
8	Perfluorohexanesulfonic acid (branched)	12.09
8	Perfluorohexanesulfonic acid (branched)	12.65
9	(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate (6:2FTS)	14.01
10	Perfluorooctanoic acid (PFOA)	14.35
11	(Na) Perfluoro-1-heptanesulfonate (L-PFHpS)	14.81
12	Heptadecafluorooctanesulfonic acid (branched)	16.19
13	Perfluorononanoic acid (PFNA)	16.40
12	Heptadecafluorooctanesulfonic acid (branched)	16.70
14	(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate (8:2FTS)	17.74
15	Perfluorodecanoic acid (PFDA)	17.86
16	(Na) Perfluoro-1-nonanesulfonate (L-PFNS)	18.00
17	N-Methylperfluoro-1-octanesulfonamidoacetic acid (N-MeFOSAA)	18.38
18	Perfluoro-1-octanesulfonamide (FOSA-1)	18.70
19	Perfluoroundecanoic acid (PFUdA)	18.86
20	N-Ethylperfluoro-1-octanesulfonamidoacetic acid (N-EtFOSAA)	18.87
21	(Na) Perfluoro-1-decanesulfonate (L-PFDS)	18.97
22	Tricosafluorododecanoic acid (PFDoA)	19.66
23	Perfluorotridecanoic acid (PFTrDA)	20.34
24	Perfluorotetradecanoic acid (PFTeDA)	20.91

Method: 537M2
Column: Advantage High Load (150mm X 2.1mm ID X 5µm df)
Column Temp.=40°C
Flowrate=0.3ml/min.
Injection Volume=5µL
Mobile Phase:
Solvent A=10mM Ammonium Acetate in Water
Solvent B=10mM Ammonium Acetate in Methanol:Acetonitrile (80:20)
Mobile Phase Profile:
Time=00 60% A:40% B
Time=04 45% A:55% B
Time=11 30% A:70% B
Time=16 10% A:90% B
Time=17 10% A:90% B
Time=27 60% A:40% B
Detector: MSD (API-ES, Negative Polarity)
Analyst: Pedro Rentas



BATTELLE

It can be done

BDO Id: 180618-02

Reagent Receipt Report

Approved: ☐ Authorized ☐

Name:	Branched NEtFOSAA Standard (50 µ	Received:	6/18/2018
Vendor:	Wellington Laboratories	Custodian:	Thorn, Jonathan
Catalogue No:	br-NEtFOSAAA	Expires:	1/17/2023
Type:	Solution	Consumed:	
Lot No:	brNEtFOSAA0118	Stored In:	Sample Preparation - C0103
Quantity:	1 ea mL % Moisture: 0		
Description:	Branched NEtFOSAA Standard (50 µg/mL)		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:

Notes:

Approved by: _____ Approved on: _____

Authorized by: _____ Authorized on: _____

180618-02

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

br-NEtFOSAA

N-Ethylperfluorooctanesulfonamidoacetic Acid Solution/Mixture of Linear and Branched Isomers

PRODUCT CODE:	br-NEtFOSAA
LOT NUMBER:	brNEtFOSAA0118
CONCENTRATION:	50.0 ± 2.5 µg/ml
SOLVENT(S):	Methanol/Water (<1%)
DATE PREPARED: (mm/dd/yyyy)	01/10/2018
LAST TESTED: (mm/dd/yyyy)	01/17/2018
EXPIRY DATE: (mm/dd/yyyy)	01/17/2023
RECOMMENDED STORAGE:	Refrigerate ampoule

DESCRIPTION:

The chemical purity has been determined to be ≥98% N-ethylperfluorooctanesulfonamidoacetic acid (linear and branched isomers). The full name, structure and percent composition for each of the identified isomeric components are given in Table A.

DOCUMENTATION/ DATA ATTACHED:

Table A: Isomeric Components and Percent Composition by ^{19}F-NMR
Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS Data (SIR)
Figure 3: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the acetic acid moiety to its respective methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HANDLING:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Our products are synthesized using single-product unambiguous routes whenever possible. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS, and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner. This further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters x_1, x_2,...x_n on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly calibrated by an external ISO/IEC 17025 accredited laboratory. In addition, their calibration is verified prior to each weighing using calibrated external weights traceable to an ISO/IEC 17025 accredited laboratory. All volumetric glassware used is calibrated, of Class A tolerance, and traceable to an ISO/IEC 17025 accredited laboratory. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO 17034 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED 17034 REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Table A: **br-NEtFOSAA; Isomeric Components and Percent Composition (by ^{19}F-NMR)***


Isomer	Name	Structure	Percent Composition by ^{19}F-NMR
1	N-ethylperfluoro-1-octanesulfonamidoacetic acid	$CF_3(CF_2)_7SO_2N(C_2H_5)CH_2CO_2H$	77.5
2	N-ethylperfluoro-3-methylheptanesulfonamidoacetic acid	$CF_3(CF_2)_3CF(CF_3)(CF_2)_2SO_2N(C_2H_5)CH_2CO_2H$	2.3
3	N-ethylperfluoro-4-methylheptanesulfonamidoacetic acid	$CF_3(CF_2)_2CF(CF_3)(CF_2)_3SO_2N(C_2H_5)CH_2CO_2H$	2.2
4	N-ethylperfluoro-5-methylheptanesulfonamidoacetic acid	$CF_3CF_2CF(CF_3)(CF_2)_4SO_2N(C_2H_5)CH_2CO_2H$	5.4
5	N-ethylperfluoro-6-methylheptanesulfonamidoacetic acid	$CF_3CF(CF_3)(CF_2)_5SO_2N(C_2H_5)CH_2CO_2H$	10.4
6	N-ethylperfluoro-5,5-dimethylhexanesulfonamidoacetic acid	$CF_3C(CF_3)_2(CF_2)_4SO_2N(C_2H_5)CH_2CO_2H$	0.3
7	N-ethylperfluoro-4,5-dimethylhexanesulfonamidoacetic acid	$CF_3CF(CF_3)CF(CF_2)_3SO_2N(C_2H_5)CH_2CO_2H$	0.3
8	N-ethylperfluoro-3,5-dimethylhexanesulfonamidoacetic acid	$CF_3CF(CF_3)CF_2CF(CF_2)_2SO_2N(C_2H_5)CH_2CO_2H$	0.3
9	Other Unidentified Isomers		1.3


* Percent of total N-ethylperfluorooctanesulfonamidoacetic acid isomers only.

Certified By: _____ **Date:** 03/22/2018 (mm/dd/yyyy)

B.G. Chittim, General Manager

Figure 1: **br-NEtFOSAA; LC/MS Data (TIC and Mass Spectrum)**





Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 55% (80:20 MeOH:ACN) / 45% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 7 min and hold for
2 min before returning to initial conditions in 0.5 min.

Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 35.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: **br-NEtFOSAA; LC/MS Data (SIR)**



Conditions for Figure 2:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 55% (80:20 MeOH:ACN) / 45% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 7 min and hold for
2 min before returning to initial conditions in 0.5 min.

Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: SIR (7 channels)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 15-60
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 3: **br-NEtFOSAA; LC/MS/MS Data (Selected MRM Transitions)**



*Note: N-EtFOSA is formed by in-source fragmentation.

Conditions for Figure 3:

Injection: On-column

Mobile phase: Same as Figure 2

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.39e-3
Collision Energy (eV) = 11-40 (variable)

BATTELLE
It can be done

BDO Id: 180618-03

Reagent Receipt Report

Approved: ☐ Authorized ☐

Name:	Branched NMeFOSAA Standard (50	Received:	6/18/2018
Vendor:	Wellington Laboratories	Custodian:	Thorn, Jonathan
Catalogue No:	brNMeFOSAA	Expires:	1/17/2023
Type:	Solution	Consumed:	
Lot No:	brNMeFOSAA0118	Stored In:	Sample Preparation - C0103
Quantity:	1 ea mL % Moisture: 0		
Description:	Branched NMeFOSAA Standard (50 µg/mL)		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:

Notes:

Approved by: _____ Approved on: _____

Authorized by: _____ Authorized on: _____

180618-03

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS
DOCUMENTATION

br-NMeFOSAA

N-Methylperfluorooctanesulfonamidoacetic Acid Solution/Mixture of Linear and Branched Isomers

PRODUCT CODE:	br-NMeFOSAA
LOT NUMBER:	brNMeFOSAA0118
CONCENTRATION:	50.0 ± 2.5 µg/ml
SOLVENT(S):	Methanol/Water (<1%)
DATE PREPARED: (mm/dd/yyyy)	01/10/2018
LAST TESTED: (mm/dd/yyyy)	01/17/2018
EXPIRY DATE: (mm/dd/yyyy)	01/17/2023
RECOMMENDED STORAGE:	Refrigerate ampoule

DESCRIPTION:

The chemical purity has been determined to be ≥98% N-methylperfluorooctanesulfonamidoacetic acid (linear and branched isomers). The full name, structure and percent composition for each of the identified isomeric components are given in Table A.

DOCUMENTATION/ DATA ATTACHED:

Table A: Isomeric Components and Percent Composition by ^{19}F-NMR
Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS Data (SIR)
Figure 3: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the acetic acid moiety to its respective methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HANDLING:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Our products are synthesized using single-product unambiguous routes whenever possible. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS, and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner. This further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \ldots x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly calibrated by an external ISO/IEC 17025 accredited laboratory. In addition, their calibration is verified prior to each weighing using calibrated external weights traceable to an ISO/IEC 17025 accredited laboratory. All volumetric glassware used is calibrated, of Class A tolerance, and traceable to an ISO/IEC 17025 accredited laboratory. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO 17034 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED ISO17034 REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Table A: **br-NMeFOSAA; Isomeric Components and Percent Composition (by ^{19}F-NMR)***


Isomer	Name	Structure	Percent Composition by ^{19}F-NMR
1	N-methylperfluoro-1-octanesulfonamidoacetic acid	$CF_3(CF_2)_7SO_2N(CH_3)CH_2CO_2H$	76.0
2	N-methylperfluoro-3-methylheptanesulfonamidoacetic acid	$CF_3(CF_2)_3CF(CF_3)(CF_2)_2SO_2N(CH_3)CH_2CO_2H$	0.7
3	N-methylperfluoro-4-methylheptanesulfonamidoacetic acid	$CF_3(CF_2)_2CF(CF_3)(CF_2)_3SO_2N(CH_3)CH_2CO_2H$	2.0
4	N-methylperfluoro-5-methylheptanesulfonamidoacetic acid	$CF_3CF_2CF(CF_3)(CF_2)_4SO_2N(CH_3)CH_2CO_2H$	6.0
5	N-methylperfluoro-6-methylheptanesulfonamidoacetic acid	$CF_3CF(CF_3)(CF_2)_5SO_2N(CH_3)CH_2CO_2H$	14.0
6	N-methylperfluoro-5,5-dimethylhexanesulfonamidoacetic acid	$CF_3C(CF_3)_2(CF_2)_4SO_2N(CH_3)CH_2CO_2H$	0.2
7	Other Unidentified Isomers		1.1


* Percent of total N-methylperfluorooctanesulfonamidoacetic acid isomers only.

Certified By: _____ **Date:** 03/22/2018 (mm/dd/yyyy)

B.G. Chittim, General Manager

Figure 1: **br-NMeFOSAA; LC/MS Data (TIC and Mass Spectrum)**





Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 55% (80:20 MeOH:ACN) / 45% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 7 min and hold for
2 min before returning to initial conditions in 0.5 min.

Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 35.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: **br-NMeFOSAA; LC/MS Data (SIR)**



Conditions for Figure 2:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 55% (80:20 MeOH:ACN) / 45% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 7 min and hold for
2 min before returning to initial conditions in 0.5 min.

Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: SIR (7 channels)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 15-60
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 3: **br-NMeFOSAA; LC/MS/MS Data (Selected MRM Transitions)**



*__Note__: N-MeFOSA is formed by in-source fragmentation.

Conditions for Figure 3:

Injection: On-column

Mobile phase: Same as Figure 2

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.39e-3
Collision Energy (eV) = 11-40 (variable)

BATTELLE

It can be done

BDO Id: 180618-04

Reagent Receipt Report

Approved: ☐ Authorized ☐

Name:	PFOA - Technical Mix	Received:	6/18/2018
Vendor:	Wellington Laboratories	Custodian:	Thorn, Jonathan
Catalogue No:	T-PFOA	Expires:	2/16/2022
Type:	Solution	Consumed:	
Lot No:	TPFOA0217	Stored In:	Sample Preparation - C0103
Quantity:	1 ea mL % Moisture: 0		
Description:	PFOA - Technical Mix		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:

Notes:

Approved by: _____ Approved on: _____

Authorized by: _____ Authorized on: _____

180618-04

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:	T-PFOA	**LOT NUMBER:**	TPFOA0217
COMPOUND:	Technical Ammonium Perfluorooctanoate		
STRUCTURE:	(see Table A)	**CAS #:**	95328-99-7 (for linear ammonium perfluorooctanoate)

MOLECULAR FORMULA:	$C_8F_{15}O_2NH_4$
CONCENTRATION:	50 ± 2.5 µg/ml (gravimetric)
CHEMICAL PURITY:	Technical material
SOLVENT(S):	Methanol/Water (<1%)
LAST TESTED: (mm/dd/yyyy)	02/16/2017
EXPIRY DATE: (mm/dd/yyyy)	02/16/2022
RECOMMENDED STORAGE:	Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Table A: Isomeric Components and Percent Composition
Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS Data (SIR)
Figure 3: LC/MS/MS Data (Selected MRM Transitions)
Figure 4: LC/MS Elution Profile of the Perfluorooctanoic Acid Isomers

ADDITIONAL INFORMATION:

- See page 2 for further details.
- This technical mixture is >97% ammonium perfluorooctanoate (branched and linear isomers). The remaining 3% consists of common impurities such as the perfluoroheptanoic and perfluorohexanoic acids.
- It is recommended that this solution be used as a *qualitative or semi-quantitative standard only*.
- Contains 4 mole eq. of NaOH to prevent conversion of any carboxylic acids to their corresponding methyl esters.
- The molecular weight of perfluoro-n-octanoic acid is 414.07 g/mol.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By: B.G. Chittim **Date:** 02/22/2017 (mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used for the identification and/or semi-quantitative determination of the specific chemical compound(s) it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

CHARACTERIZATION / HOMOGENEITY:

This product is a technical mixture obtained from an industrial manufacturer. It has been characterized as to its content and components using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Testing of samples in solution has shown it to be homogeneous. As this product is a technical mixture, it should not be used to quantitate any of the listed components.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters x_1, x_2,...x_n on which it depends is:

$$u_c(y(x_1, x_2, ...x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:










This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A 1226

ANAB ACCREDITED REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com


Table A: **T-PFOA; Isomeric Components and Percent Composition***


Isomer	Name	Structure	Percent Composition by ^{19}F-NMR
1	Linear ammonium perfluoro-n-octanoate		79
2	Ammonium 6-trifluoromethylperfluoroheptanoate		9
3	Ammonium 5-trifluoromethylperfluoroheptanoate		4.5
4	Ammonium 4-trifluoromethylperfluoroheptanoate		4
5	Ammonium 3-trifluoromethylperfluoroheptanoate		3
6[a]	Ammonium 2-trifluoromethylperfluoroheptanoate		0.5
7	Ammonium 5,5-bis(trifluoromethyl)perfluorohexanoate		
8	Ammonium 4,4-bis(trifluoromethyl)perfluorohexanoate		
9[a]	Ammonium 4,5-bis(trifluoromethyl)perfluorohexanoate		
10	Ammonium 3,5-bis(trifluoromethyl)perfluorohexanoate		

* Percent Composition was determined by ^{19}F-NMR. The percentages displayed are of total ammonium perfluorooctanoate isomers only (isomers are labelled in Figure 4).

[a] Presence of this isomer could not be verified by LC/MS due to co-elution.

Figure 1: **T-PFOA; LC/MS Data (TIC and Mass Spectrum)**





Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions:

Column: Kinetex PFP
2.6 µm, 4.6 x 100 mm

Mobile phase: Gradient
Start: 30% (80:20 MeOH:ACN) / 70% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 50% organic over 5 min. Ramp to
90% organic over 5 min and hold for 1.5 min.
Return to initial conditions over 0.5 min.
Time: 13 min


Flow: 1.0 ml/min

MS Parameters:

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: **T-PFOA; LC/MS Data (SIR)**



Conditions for Figure 2:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions:

Column: Kinetex PFP
2.6 µm, 4.6 x 100 mm

Mobile phase: Gradient
Start: 30% (80:20 MeOH:ACN) / 70% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 50% organic over 5 min. Ramp to 90% organic over 5 min and hold for 1.5 min.
Return to initial conditions over 0.5 min.
Time: 13 min


Flow: 1.0 ml/min

MS Parameters:

Experiment: SIR (ES^-)

Source conditions: see Figure 1
Source Temperature = 110 °C
Desolvation Temperature = 325 °C
Cone Voltage = 15V

Figure 3: **T-PFOA; LC/MS/MS Data (Selected MRM Transitions)**



Conditions for Figure 3:

Injection: Direct loop injection
10 µl (500 ng/ml T-PFOA)


Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.39e-3
Collision Energy (eV) = variable (9-40)

Figure 4: **T-PFOA; LC/MS Elution Profile of the Perfluorooctanoic Acid Isomers**



Conditions for Figure 4:

Same as Figure 2.

BATTELLE

It can be done

BDO Id: 180618-06

Reagent Receipt Report

Approved: ☐ Authorized ☐

Name:	Branched PFHxS Standard (50 µg/m	Received:	6/18/2018
Vendor:	Wellington Laboratories	Custodian:	Thorn, Jonathan
Catalogue No:	br-PFHxSK	Expires:	1/4/2022
Type:	Solution	Consumed:	
Lot No:	brPFHxSK0117	Stored In:	Sample Preparation - C0103
Quantity:	1 ea mL % Moisture: 0		
Description:	Branched PFHxS Standard (50 µg/mL)		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:

Notes:

Approved by: _____ Approved on: _____

Authorized by: _____ Authorized on: _____

180618-06

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

br-PFHxSK

Potassium Perfluorohexanesulfonate Solution/Mixture of Linear and Branched Isomers

PRODUCT CODE:	br-PFHxSK
LOT NUMBER:	brPFHxSK0117
CONCENTRATION:	50.0 ± 2.5 µg/ml (total potassium salt)
	45.5 ± 2.3 µg/ml (total PFHxS anion)
SOLVENT(S):	Methanol
DATE PREPARED: (mm/dd/yyyy)	01/03/2017
LAST TESTED: (mm/dd/yyyy)	01/04/2017
EXPIRY DATE: (mm/dd/yyyy)	01/04/2022
RECOMMENDED STORAGE:	Store ampoule in a cool, dark place

DESCRIPTION:

The chemical purity has been determined to be ≥98% perfluorohexanesulfonate linear and branched isomers. The full name, structure and percent composition for each of the identified isomeric components are given in Table A.

DOCUMENTATION/ DATA ATTACHED:

Table A: Isomeric Components and Percent Composition by ^{19}F-NMR
Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS Data (SIR)
Figure 3: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains ~ 0.5% of perfluoro-1-pentanesulfonate and ~ 0.2% of perfluoro-1-octanesulfonate.
- CAS#: 3871-99-6 (for linear isomer; potassium salt).

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters x_1, x_2,...x_n on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED ISO 17025 ISO 17034 REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com


Table A: **br-PFHxSK; Isomeric Components and Percent Composition (by ^{19}F-NMR)***


Isomer	Name	Structure	Percent Composition by ^{19}F-NMR
1	Potassium perfluoro-1-hexanesulfonate	$CF_3CF_2CF_2CF_2CF_2CF_2SO_3^-K^+$	81.1
2	Potassium 1-trifluoromethylperfluoropentanesulfonate**	$CF_3CF_2CF_2CF_2CF(CF_3)SO_3^-K^+$	2.9
3	Potassium 2-trifluoromethylperfluoropentanesulfonate	$CF_3CF_2CF_2CF(CF_3)CF_2SO_3^-K^+$	1.4
4	Potassium 3-trifluoromethylperfluoropentanesulfonate	$CF_3CF_2CF(CF_3)CF_2CF_2SO_3^-K^+$	5.0
5	Potassium 4-trifluoromethylperfluoropentanesulfonate	$CF_3CF(CF_3)CF_2CF_2CF_2SO_3^-K^+$	8.9
6	Potassium 3,3-di(trifluoromethyl)perfluorobutanesulfonate	$CF_3C(CF_3)_2CF_2CF_2SO_3^-K^+$	0.2
7	Other Unidentified Isomers		0.5

* Percent of total perfluorohexanesulfonate isomers only.
** Systematic Name: Potassium perfluorohexane-2-sulfonate.

Certified By: B.G. Chittim **Date:** 01/20/2017 (mm/dd/yyyy)

Figure 1: **br-PFHxSK; LC/MS Data (TIC and Mass Spectrum)**





Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 20% (80:20 MeOH:ACN) / 80% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 50% organic over 14 min. Ramp to 90% organic over 3 min and hold for 1.5 min before returning to initial conditions in 0.5 min.
Time: 20 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 50.00
Cone Gas Flow (l/hr) = 60
Desolvation Gas Flow (l/hr) = 750

Figure 2: **br-PFHxSK; LC/MS Data (SIR)**



Conditions for Figure 2:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 20% (80:20 MeOH:ACN) / 80% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 50% organic over 14 min. Ramp to 90% organic over 3 min and hold for 1.5 min before returning to initial conditions in 0.5 min.
Time: 20 min


Flow: 300 µl/min

MS Parameters

Experiment: SIR (6 channels)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = variable (15-62)
Cone Gas Flow (l/hr) = 60
Desolvation Gas Flow (l/hr) = 750

Figure 3: **br-PFHxSK; LC/MS/MS Data (Selected MRM Transitions)**



Conditions for Figure 3:

Injection: Direct loop injection
10 µl (500 ng/ml br-PFHxSK)

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.35e-3
Collision Energy (eV) = 30

BATTELLE

It can be done

BDO Id: 180618-07

Reagent Receipt Report

Approved: ☐ Authorized ☐

Name:	Branched PFOS Standard (50 µg/mL	**Received:**	6/18/2018
Vendor:	Wellington Laboratories	**Custodian:**	Thorn, Jonathan
Catalogue No:	br-PFOSK	**Expires:**	1/12/2022
Type:	Solution	**Consumed:**	
Lot No:	brPFOSK0117	**Stored In:**	Sample Preparation - C0103
Quantity:	1 ea mL **% Moisture:** 0		
Description:	Branched PFOS Standard (50 µg/mL)		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:

Notes:

Approved by: _____ **Approved on:** _____

Authorized by: _____ **Authorized on:** _____

180618-07

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

br-PFOSK

Potassium Perfluorooctanesulfonate Solution/Mixture of Linear and Branched Isomers

PRODUCT CODE:	br-PFOSK
LOT NUMBER:	brPFOSK0117
CONCENTRATION:	50 ± 2.5 µg/ml (total potassium salt)
	46.4 ± 2.3 µg/ml (total PFOS anion)
SOLVENT(S):	Methanol
DATE PREPARED: (mm/dd/yyyy)	01/09/2017
LAST TESTED: (mm/dd/yyyy)	01/12/2017
EXPIRY DATE: (mm/dd/yyyy)	01/12/2022
RECOMMENDED STORAGE:	Store ampoule in a cool, dark place

DESCRIPTION:

The chemical purity has been determined to be ≥98% perfluorooctanesulfonate linear and branched isomers. The full name, structure and percent composition for each of the isomeric components are given in Table A.

DOCUMENTATION/ DATA ATTACHED:

Table A: Isomeric Components and Percent Composition by ^{19}F-NMR
Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS Data (SIR)
Figure 3: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- A 5-point calibration curve was generated using linear PFOS (potassium salt) and mass-labelled PFOS as an internal standard to enable quantitation of br-PFOSK using isotopic dilution.
- CAS#: 2795-39-3 (for linear isomer; potassium salt).

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters x_1, x_2,...x_n on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com


Table A: **br-PFOSK; Isomeric Components and Percent Composition (by ^{19}F-NMR)***


Isomer	Name	Structure	Percent Composition by ^{19}F-NMR
1	Potassium perfluoro-1-octanesulfonate	$CF_3CF_2CF_2CF_2CF_2CF_2CF_2CF_2SO_3^-K^+$	78.8
2	Potassium 1-trifluoromethylperfluoroheptanesulfonate**	$CF_3CF_2CF_2CF_2CF_2CF_2CF(CF_3)SO_3^-K^+$	1.2
3	Potassium 2-trifluoromethylperfluoroheptanesulfonate	$CF_3CF_2CF_2CF_2CF_2CF(CF_3)CF_2SO_3^-K^+$	0.6
4	Potassium 3-trifluoromethylperfluoroheptanesulfonate	$CF_3CF_2CF_2CF_2CF(CF_3)CF_2CF_2SO_3^-K^+$	1.9
5	Potassium 4-trifluoromethylperfluoroheptanesulfonate	$CF_3CF_2CF_2CF(CF_3)CF_2CF_2CF_2SO_3^-K^+$	2.2
6	Potassium 5-trifluoromethylperfluoroheptanesulfonate	$CF_3CF_2CF(CF_3)CF_2CF_2CF_2CF_2SO_3^-K^+$	4.5
7	Potassium 6-trifluoromethylperfluoroheptanesulfonate	$CF_3CF(CF_3)CF_2CF_2CF_2CF_2CF_2SO_3^-K^+$	10.0
8	Potassium 5,5-di(trifluoromethyl)perfluorohexanesulfonate	$CF_3C(CF_3)CF_2CF_2CF_2CF_2SO_3^-K^+$	0.2
9	Potassium 4,4-di(trifluoromethyl)perfluorohexanesulfonate	$CF_3CF_2C(CF_3)CF_2CF_2CF_2SO_3^-K^+$	0.03
10	Potassium 4,5-di(trifluoromethyl)perfluorohexanesulfonate	$CF_3CF(CF_3)CF_2CF_2CF_2SO_3^-K^+$	0.4
11	Potassium 3,5-di(trifluoromethyl)perfluorohexanesulfonate	$CF_3CF(CF_3)CF_2CF(CF_3)CF_2CF_2SO_3^-K^+$	0.07

* Percent of total perfluorooctanesulfonate isomers only. Isomers are labelled in Figure 2.
** Systematic Name: Potassium perfluorooctane-2-sulfonate.

Certified By: B.G. Chittim **Date:** 01/20/2017 (mm/dd/yyyy)

Figure 1: **br-PFOSK; LC/MS Data (TIC and Mass Spectrum)**





Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 45% (80:20 MeOH:ACN) / 55% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 12 min and hold for 2 min.
Return to initial conditions over 0.5 min.
Time: 16 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 60.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: **br-PFOSK; LC/MS Data (SIR)**



Conditions for Figure 2:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**


Chromatographic Conditions:

Column:	Acquity UPLC BEH Shield RP_{18} (1.7 µm, 2.1 x 100 mm)
Injection:	1.0 µg/ml of br-PFOSK
Mobile Phase:	Gradient 45% (80:20 MeOH:ACN) / 55% H_2O (both with 10 mM NH_4OAc buffer) Ramp to 90% organic over 15 min and hold for 3 min. Return to initial conditions over 1 min. Time: 20 min
Flow:	300 µl/min

MS Conditions:

SIR (ES^-)
Source = 110 °C
Desolvation = 325 °C
Cone Voltage = 60V

Figure 3: **br-PFOSK; LC/MS/MS Data (Selected MRM Transitions)**



Conditions for Figure 3:

Injection: On-column

Mobile phase: Same as Figure 2

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.31e-3
Collision Energy (eV) = 11-50 (variable)

CERTIFIED WEIGHT REPORT

180705-02

Part Number:	99207
Lot Number:	061918
Description:	PFOA - DOD
	24 components
Expiration Date:	061923
Recommended Storage:	Freezer (0 °C)
Nominal Concentration (µg/mL):	1.0
NIST Test ID#:	2684186

Solvent(s):	Lot#	
Methanol (1 mM KOH)	061918	(98%)
2-Propanol	23214	(2%)

5E-05 Balance Uncertainty
0.007 Flask Uncertainty

Volume(s) shown below were combined and diluted to (mL): 50.0

Note: All assigned values are anion concentrations.


Formulated By:	Mario Luis	061918 DATE
Reviewed By:	Pedro L. Rentas	061918 DATE

	Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Perfluoro-n-butanoic acid (linear)	99542	110317	0.02	1.00	0.004	50.2	1.00	0.01	375-22-4	N/A	N/A
2.	Perfluoro-n-pentanoic acid	99543	110317	0.02	1.00	0.004	50.7	1.01	0.01	2706-90-3	N/A	N/A
3.	Perfluorohexanoic acid	99199	030617	0.02	1.00	0.004	50.3	1.01	0.01	307-24-4	N/A	N/A
4.	Perfluoroheptanoic acid	99197	030517	0.02	1.00	0.004	50.1	1.00	0.01	375-85-9	N/A	N/A
5.	Perfluorooctanoic acid	99202	030617	0.02	1.00	0.004	50.2	1.00	0.01	335-67-1	N/A	ipr-rat 189mg/kg
6.	Perfluorononanoic acid	99200	030617	0.02	1.00	0.004	50.1	1.00	0.01	375-95-1	N/A	N/A
7.	Perfluorodecanoic acid	99195	030617	0.02	1.00	0.004	50.1	1.00	0.01	335-76-2	N/A	orl-rat 57mg/kg
8.	Perfluoroundecanoic acid	99205	030617	0.02	1.00	0.004	50.1	1.00	0.01	2058-94-8	N/A	N/A
9.	Tricosafluorododecanoic acid	99196	030617	0.02	1.00	0.004	50.1	1.00	0.01	307-55-1	N/A	N/A
10.	Perfluorotridecanoic acid	99204	030617	0.02	1.00	0.004	50.1	1.00	0.01	72629-94-8	N/A	N/A
11.	Perfluorotetradecanoic acid	99203	030617	0.02	1.00	0.004	50.1	1.00	0.01	376-06-7	N/A	N/A
12.	Perfluoro-1-octanesulfonamide	3677	FOSA0817I	0.02	1.00	0.004	50.0	1.00	0.01	754-91-6	N/A	N/A
13.	N-Methylperfluoro-1-octanesulfonamidoacetic acid	3667	NMeFOSAA0118	0.02	1.00	0.004	50.0	1.00	0.01	2355-31-9	N/A	N/A
14.	N-Ethylperfluoro-1-octanesulfonamidoacetic acid	3664	NEtFOSAA0118	0.02	1.00	0.004	50.0	1.00	0.01	2991-50-6	N/A	N/A
15.	Perfluorobutanesulfonic acid	99194	031017	0.02	1.00	0.004	50.7	1.01	0.01	375-73-5	N/A	N/A
16.	Perfluoro-1-pentanesulfonate	99544	111017	0.02	0.98	0.004	51.3	1.00	0.01	630402-22-1	N/A	N/A
17.	Perfluorohexanesulfonic acid (branched)	99198	030617	0.02	1.00	0.004	50.6	1.01	0.01	3871-99-6	N/A	N/A
18.	Perfluoro-1-heptanesulfonic acid	3672	LPFHpS0817	0.021	1.05	0.004	47.6	1.00	0.01	375-92-8	N/A	N/A
19.	Heptadecafluorooctanesulfonic acid (branched)	99201	030617	0.02	1.00	0.004	50.2	1.00	0.01	1763-23-1	N/A	N/A
20.	Perfluoro-1-nonanesulfonic acid	3957	LPFNS0917	0.021	1.05	0.004	48.0	1.01	0.01	98789-57-2	N/A	N/A
21.	Perfluoro-1-decanesulfonic acid	3671	LPFDS0217	0.021	1.05	0.004	48.2	1.01	0.01	2806-15-7	N/A	N/A
22.	1H,1H,2H,2H-Perfluorohexane sulfonic acid	3955	42FTS1216	0.0214	1.07	0.004	46.7	1.00	0.01	00-00-0	N/A	N/A
23.	1H,1H,2H,2H-Perfluorooctane sulfonic acid	3661	62FTS0616	0.021	1.05	0.004	47.4	1.00	0.01	27619-97-2	N/A	N/A
24.	1H,1H,2H,2H-Perfluorodecane sulfonic acid	3662	82FTS1216	0.021	1.05	0.004	47.9	1.01	0.01	39108-34-4	N/A	N/A

Peak No.	Analyte	MSD RT (min.)
1	Perfluoro-n-butanoic acid (PFBA)	5.08
2	Perfluoro-n-pentanoic acid (PFPeA)	6.42
3	Perfluorobutanesulfonate (PFBS)	7.59
4	(Na) 1H,1H,2H,2H-Perfluorohexane sulfonate (4:2FTS)	8.96
5	Perfluorohexanoic acid (PFHxA)	9.42
6	(Na) Perfluoro-1-pentanesulfonate (L-PFPeS)	10.17
7	Perfluoroheptanoic acid (PFHpA)	11.89
8	Perfluorohexanesulfonic acid (branched) (br-PFHxS)	11.96
9	Perfluorohexanesulfonic acid (branched) (br-PFHxS)	12.46
10	(Na) 1H,1H,2H,2H-Perfluorooctane sulfonate (6:2FTS)	13.89
11	Perfluorooctanoic acid (PFOA)	14.19
12	(Na) Perfluoro-1-heptanesulfonate (L-PFHpS)	14.63
13	Heptadecafluorooctanesulfonic acid (branched)	15.97
14	Perfluorononanoic acid (PFNA)	16.21
15	Heptadecafluorooctanesulfonic acid (branched)	16.49
16	(Na) 1H,1H,2H,2H-Perfluorodecane sulfonate (8:2FTS)	17.55
17	Perfluorodecanoic acid (PFDA)	17.65
18	(Na) Perfluoro-1-nonanesulfonate (L-PFNS)	17.79
19	N-Methylperfluoro-1-octanesulfonamidoacetic acid (N-MeFOSAA)	18.17
20	Perfluoro-1-octanesulfonamide (FOSA-1)	18.43
21	N-Ethylperfluoro-1-octanesulfonamidoacetic acid (N-EtFOSAA)	18.64
22	Perfluoroundecanoic acid (PFUdA)	18.65
23	(Na) Perfluoro-1-decanesulfonate (L-PFDS)	18.72
24	Tricosafluorododecanoic acid (PFDoA)	19.42
25	Perfluorotridecanoic acid (PFTrDA)	20.06
26	Perfluorotetradecanoic acid (PFTeDA)	20.60

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Method: 537M2
Column: Advantage High Load (150mm X 2.1mm ID X 5µm df)
Column Temp.=40°C
Flowrate=0.3mL/min.
Injection Volume=1µL
Mobile Phase:
Solvent A=10mM Ammonium Acetate in Water
Solvent B=10mM Ammonium Acetate in Methanol:Acetonitrile (80:20)
Mobile Phase Profile:
Time=00 60%A:40%B
Time=04 45%A:55%B
Time=11 30%A:70%B
Time=16 10%A:90%B
Time=17 10%A:90%B
Time=27 60%A:40%B
Detector: MSD (API-ES, Negative Polarity, Scan mode)
Analyst: Pedro Rentas



BATTELLE

It can be done

BDO Id: 180726-04

Reagent Receipt Report

Approved: ☑ Authorized ☐

Name:	Mass-labelled PFAS injection standar	**Received:**	7/26/2018
Vendor:	Wellington Laboratories	**Custodian:**	Thorn, Jonathan
Catalogue No:	MPFAC-C-IS	**Expires:**	5/2/2022
Type:	Solution	**Consumed:**	
Lot No:	MPFACCIS0516	**Stored In:**	LC Laboratory - R0107
Quantity:	2 ea 1.2 mL **% Moisture:** 0		
Description:	Mass-labelled PFAS injection standards		

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
13C2-PFDA	BDO-2110	2.0000	100.00	--	--	☐			
13C2-PFOA	BDO-2107	2.0000	100.00	--	--	☐			
13C3-PFBA	BDO-2231	2.0000	100.00	--	--	☐			
13C4-PFOS	BDO-2121	1.9140	100.00	--	--	☐			

Total Analytes: 4

Notes:

Approved by:	Lizotte Jr, Robert	**Approved on:**	7/27/2018 11:10:00 AM
Authorized by:		**Authorized on:**	

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS
DOCUMENTATION

MPFAC-C-IS

Mass-Labelled Perfluorinated Compound Injection Standards Solution

PRODUCT CODE:	MPFAC-C-IS
LOT NUMBER:	MPFACCIS0516
SOLVENT(S):	Methanol / Water (<1%)
DATE PREPARED: (mm/dd/yyyy)	05/24/2016
LAST TESTED: (mm/dd/yyyy)	05/02/2017
EXPIRY DATE: (mm/dd/yyyy)	05/02/2022
RECOMMENDED STORAGE:	Store ampoule in a cool, dark place

DESCRIPTION:

MPFAC-C-IS is a solution/mixture of mass-labelled (^{13}C) perfluoroalkylcarboxylic acids and a mass-labelled (^{13}C) perfluoroalkylsulfonate. The components and their concentrations are given in Table A.

MPFAC-C-IS was designed for, and prepared to be used with, PFC-CVS-C.

The individual mass-labelled perfluoroalkylcarboxylic acids and mass-labelled perfluoroalkylsulfonate all have chemical purities of >98% and isotopic purities of ≥99%.

DOCUMENTATION/ DATA ATTACHED:

Table A: Components and Concentrations of the Solution/Mixture
Figure 1: LC/MS Data (SIR)
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- The mass-labelled perfluoroalkylsulfonate compound concentration is reported as the salt.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acids to their respective methyl esters.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \ldots x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using calibrated NIST and/or NRC traceable external weights. All volumetric glassware used is calibrated, of Class A tolerance, and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED ISO 17025 / 17034 REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com


Table A: **MPFAC-C-IS; Components and Concentrations (ng/ml; ± 5% in Methanol / Water (<1%))**

Compound	Abbreviation	Concentration (ng/ml)	Peak Assignment in Figure 1
Perfluoro-n-[2,3,4-$^{13}C_3$]butanoic acid	M3PFBA	2000	A
Perfluoro-n-[1,2-$^{13}C_2$]octanoic acid	M2PFOA	2000	B
Perfluoro-n-[1,2-$^{13}C_2$]decanoic acid	MPFDA	2000	D
Sodium perfluoro-1-[1,2,3,4-$^{13}C_4$]octanesulfonate	MPFOS	2000 ①	C

① 1914 when corrected for sodium
JMG 7/26/2017

Certified By: _____ **Date:** 05/04/2017
(mm/dd/yyyy)
B.G. Chittim, General Manager

Figure 1: **MPFAC-C-IS; LC/MS Data (Total Ion Current Chromatogram; SIR)**



Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 90% organic over 8 min
and hold for 2 min before returning
to initial conditions in 1 min.
Time: 12 min


Flow: 300 µl/min

MS Parameters

Experiment: SIR

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = variable (10-80)
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: **MPFAC-C-IS; LC/MS/MS Data (Selected MRM Transitions)**



Conditions for Figure 2:

Injection: On-column (MPFAC-C-IS)

Mobile phase: Same as Figure 1

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.50e-3

Collision Energy (eV) = 8-50 (variable)

BATTELLE

It can be done

BDO Id: 180726-05

Reagent Receipt Report

Approved: ☑ Authorized ☐

Name: Mass-labelled PFAS Extraction Stand
Vendor: Wellington Laboratories
Catalogue No: MPFAC-24ES
Type: Solution
Lot No: MPFAC24ES0218
Quantity: 2 ea 1.2 mL **% Moisture:** 0
Description: Mass-labelled PFAS Extraction Standard Solution

Received: 7/26/2018
Custodian: Thorn, Jonathan
Expires: 2/7/2023
Consumed:
Stored In: LC Laboratory - R0107

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
13C2-4:2FTS	BDO-2229	0.9350	100.00	--	--	☐			
13C2-6:2FTS	BDO-2230	0.9490	100.00	--	--	☐			
13C2-8:2FTS	BDO-2220	0.9580	100.00	--	--	☐			
13C2-PFDoA	BDO-2112	1.0000	100.00	--	--	☐			
13C2-PFTeDA	BDO-2224	1.0000	100.00	--	--	☐			
13C3-PFBS	BDO-2226	0.9290	100.00	--	--	☐			
13C3-PFHxS	BDO-2227	0.9460	100.00	--	--	☐			
13C4-PFBA	BDO-2105	1.0000	100.00	--	--	☐			
13C4-PFHpA	BDO-2218	1.0000	100.00	--	--	☐			
13C5-PFHxA	BDO-2217	1.0000	100.00	--	--	☐			
13C5-PFPeA	BDO-2216	1.0000	100.00	--	--	☐			
13C6-PFDA	BDO-2222	1.0000	100.00	--	--	☐			
13C7-PFUnA	BDO-2223	1.0000	100.00	--	--	☐			
13C8-FOSA	BDO-2225	1.0000	100.00	--	--	☐			
13C8-PFOA	BDO-2219	1.0000	100.00	--	--	☐			
13C8-PFOS	BDO-2228	0.9570	100.00	--	--	☐			
13C9-PFNA	BDO-2221	1.0000	100.00	--	--	☐			
d3-MeFOSAA	BDO-1838	1.0000	100.00	--	--	☐			
d5-EtFOSAA	BDO-1839	1.0000	100.00	--	--	☐			

Total Analytes: 19

Notes:

Approved by: Lizotte Jr, Robert **Approved on:** 7/27/2018 11:10:00 AM
Authorized by: **Authorized on:**

WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

MPFAC-24ES

Mass-Labelled Per- and Poly-fluoroalkyl Substance Extraction Standard Solution

PRODUCT CODE:	MPFAC-24ES
LOT NUMBER:	MPFAC24ES0218
SOLVENT(S):	Methanol / Isopropanol (2%) / Water (<1%)
DATE PREPARED: (mm/dd/yyyy)	02/07/2018
LAST TESTED: (mm/dd/yyyy)	02/07/2018
EXPIRY DATE: (mm/dd/yyyy)	02/07/2023
RECOMMENDED STORAGE:	Refrigerate ampoule

DESCRIPTION:

MPFAC-24ES is a solution/mixture of ten mass-labelled (^{13}C) perfluoroalkylcarboxylic acids (C_4-C_{12} and C_{14}), three mass-labelled (^{13}C) perfluoroalkylsulfonates (C_4, C_6, and C_8), three mass-labelled (^{13}C) telomer sulfonates (4:2, 6:2, and 8:2), two mass-labelled (^{2}H) perfluorooctanesulfonamidoacetic acids, and perfluoro-1-[$^{13}C_8$]octanesulfonamide. The components and their concentrations are given in Table A.

The individual mass-labelled perfluoroalkylcarboxylic acids, mass-labelled perfluoroalkylsulfonates, mass-labelled telomer sulfonates, and perfluoro-1-[$^{13}C_8$]octanesulfonamide all have chemical purities of >98% and isotopic purities of ≥99%. The individual mass-labelled perfluorooctanesulfonamidoacetic acids all have chemical purities of >98% and isotopic purities of ≥98%.

DOCUMENTATION/ DATA ATTACHED:

Table A: Components and Concentrations of the Solution/Mixture
Figure 1: LC/MS Data (SIR)
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acids to their respective methyl esters.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

HANDLING:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Our products are synthesized using single-product unambiguous routes whenever possible. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS, and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner. This further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \ldots x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \ldots x_n)) = \sqrt{\sum_{i=1}^{n} u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of ±5% (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly calibrated by an external ISO/IEC 17025 accredited laboratory. In addition, their calibration is verified prior to each weighing using calibrated external weights traceable to an ISO/IEC 17025 accredited laboratory. All volumetric glassware used is calibrated, of Class A tolerance, and traceable to an ISO/IEC 17025 accredited laboratory. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO 17034 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).

CALA Testing Accreditation No. A1226

ANAB ACCREDITED 9017034 REFERENCE MATERIAL PRODUCER

For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com


Table A: **MPFAC-24ES; Components and Concentrations (ng/ml, ± 5% in Methanol / Isopropanol (2%) / Water (<1%))**

Compound	Abbreviation	Concentration (ng/ml)		Peak Assignment in Figure 1
Perfluoro-n-[$^{13}C_4$]butanoic acid	MPFBA	1000		A
Perfluoro-n-[$^{13}C_5$]pentanoic acid	M5PFPeA	1000		B
Perfluoro-n-[1,2,3,4,6-$^{13}C_5$]hexanoic acid	M5PFHxA	1000		E
Perfluoro-n-[1,2,3,4-$^{13}C_4$]heptanoic acid	M4PFHpA	1000		F
Perfluoro-n-[$^{13}C_8$]octanoic acid	M8PFOA	1000		I
Perfluoro-n-[$^{13}C_9$]nonanoic acid	M9PFNA	1000		J
Perfluoro-n-[1,2,3,4,5,6-$^{13}C_6$]decanoic acid	M6PFDA	1000		M
Perfluoro-n-[1,2,3,4,5,6,7-$^{13}C_7$]undecanoic acid	M7PFUdA	1000		Q
Perfluoro-n-[1,2-$^{13}C_2$]dodecanoic acid	MPFDoA	1000		R
Perfluoro-n-[1,2-$^{13}C_2$]tetradecanoic acid	M2PFTeDA	1000		S
Perfluoro-1-[$^{13}C_8$]octanesulfonamide	M8FOSA	1000		O
N-methyl-d_3-perfluoro-1-octanesulfonamidoacetic acid	d3-N-MeFOSAA	1000		N
N-ethyl-d_5-perfluoro-1-octanesulfonamidoacetic acid	d5-N-EtFOSAA	1000		P
Compound	**Abbreviation**	**Concentration (ng/ml)**		**Peak Assignment in Figure 1**
		as the salt	**as the anion**	
Sodium perfluoro-1-[2,3,4-$^{13}C_3$]butanesulfonate	M3PFBS	1000	929	C
Sodium perfluoro-1-[1,2,3-$^{13}C_3$]hexanesulfonate	M3PFHxS	1000	946	G
Sodium perfluoro-1-[$^{13}C_8$]octanesulfonate	M8PFOS	1000	957	K
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-$^{13}C_2$]hexanesulfonate	M2-4:2FTS	1000	935	D
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-$^{13}C_2$]octanesulfonate	M2-6:2FTS	1000	949	H
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-$^{13}C_2$]decanesulfonate	M2-8:2FTS	1000	958	L

Certified By: _____ **Date:** 02/09/2018 (mm/dd/yyyy)

B.G. Chittim, General Manager

Figure 1: **MPFAC-24ES; LC/MS Data (Total Ion Current Chromatogram; SIR)**



Conditions for Figure 1:

LC: **Waters Acquity Ultra Performance LC**
MS: **Micromass Quattro *micro* API MS**

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP_{18}
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 40% (80:20 MeOH:ACN) / 60% H_2O
(both with 10 mM NH_4OAc buffer)
Ramp to 55% organic over 3.5 min.
Ramp to 70% organic over 6.5 min.
Ramp to 85% organic over 5 min and hold for
1 min before returning to initial conditions in 0.5 min.
Time: 17 min

Flow: 300 µl/min

MS Parameters

Experiment: SIR

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = variable (10-60)
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: **MPFAC-24ES; LC/MS/MS Data (Selected MRM Transitions)**




Figure 2: **MPFAC-24ES; LC/MS/MS Data (Selected MRM Transitions)**





Figure 2: **MPFAC-24ES; LC/MS/MS Data (Selected MRM Transitions)**



Conditions for Figure 2:

Injection: On-column (MPFAC-24ES)

Mobile phase: Same as Figure 1

Flow: 300 µl/min

MS Parameters

Collision Gas (mbar) = 3.28e-3

Collision Energy (eV) = 8-40 (variable)

Sample Preparation

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SAMPLE PREPARATION RECORDS

Project Title(s)	**Project No.(s)**
CTO-4164 Naval Base Ventura County, California	100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

SOP Numbers (see workplan for modifications)

ExtractionSOP No. 5-370

This Batch Contains The Following Samples:		
CR939PB-FS	J8617-FS	J8623-FS
CR940LCS-FS	J8618-FS	J8624-FS
J8613-FS	J8619-FS	J8625-FS
J8614-FS	J8620-FS	J8626-FS
J8615-FS	J8621-FS	J8627-FS
J8616-FS	J8622-FS	J8628-FS

Laboratory Preparation Records

COMPLETE AND VALIDATED

Prep Task Leader: Stephanie Schultz

Approved By:	Date	Initials
Jordan Tower	10/30/2018	JCT

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SAMPLE IDENTIFICATION PAGE

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Sample ID	Description
CR939PB-FS	Procedural Blank
CR940LCS-FS	Laboratory Control Sample
J8613-FS	VC-PM553-GW0601A-1018
J8614-FS	VC-AQ-FB09-10012018
J8615-FS	VC-AQ-EB09-10012018
J8616-FS	VC-PM13-GW21-1018
J8617-FS	VC-KCHA16-GW01-1018
J8618-FS	VC-KCHA16-GW01P-1018
J8619-FS	VC-KCHA16-GW02-1018
J8620-FS	VC-KCHA16-GW03-1018
J8621-FS	VC-AQ-FB10-10022018
J8622-FS	VC-AQ-EB10-10022018
J8623-FS	VC-S09GW01-1018
J8624-FS	VC-S09GW04-1018
J8625-FS	VC-S09GW04P-1018
J8626-FS	VC-S09GW06-1018
J8627-FS	VC-S14GW01-1018
J8628-FS	VC-S14GW03-1018

Samples Assigned By: Jonathan Thorn Date : October 5, 2018

Comments:

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SAMPLE CUSTODY LOG

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Requested On/By:	10/10/2018 SAS	**Purpose:**	Sample Preparation
Relinquished On/By:	10/10/2018 MDS	**Last Activity:**	Transfer

Accepted On/By:	10/10/2018 SAS	**Returned On/To:**	
Stored In Facility:	Sample Preparation	**Returned To Facility:**	
Stored Until	10/10/2018		
Stored Comment:	NA	**Returned Comment:**	NA

No.	BDO-ID:	Ctrs	*	Condition:	Custody Comment:
1	J8613	1	C	Consumed	NA
2	J8614	1	C	Consumed	NA
3	J8615	1	C	Consumed	NA
4	J8616	1	C	Consumed	NA
5	J8617	1	C	Consumed	NA
6	J8618	1	C	Consumed	NA
7	J8619	1	C	Consumed	NA
8	J8620	1	C	Consumed	NA
9	J8621	1	C	Consumed	NA
10	J8622	1	C	Consumed	NA
11	J8623	1	C	Consumed	NA
12	J8624	1	C	Consumed	NA
13	J8625	1	C	Consumed	NA
14	J8626	1	C	Consumed	NA
15	J8627	1	C	Consumed	NA
16	J8628	1	C	Consumed	NA

Total Samples 16 * "C" = Consumed Container

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
LIQUID SAMPLE ID FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Sample ID	Description	Volume (mL)	Bottles	*	Date Initials
CR939PB-FS	Procedural Blank	250.0	NA	--	10/10/18 SAS
CR940LCS-FS	Laboratory Control Sample	250.0	NA	--	10/10/18 SAS
J8613-FS	VC-PM553-GW0601A-1018	270.0	1	C	10/11/18 SAS
J8614-FS	VC-AQ-FB09-10012018	290.0	1	C	10/11/18 SAS
J8615-FS	VC-AQ-EB09-10012018	275.0	1	C	10/11/18 SAS
J8616-FS	VC-PM13-GW21-1018	275.0	1	C	10/11/18 SAS
J8617-FS	VC-KCHA16-GW01-1018	275.0	1	C	10/11/18 SAS
J8618-FS	VC-KCHA16-GW01P-1018	265.0	1	C	10/11/18 SAS
J8619-FS	VC-KCHA16-GW02-1018	270.0	1	C	10/11/18 SAS
J8620-FS	VC-KCHA16-GW03-1018	275.0	1	C	10/11/18 SAS
J8621-FS	VC-AQ-FB10-10022018	285.0	1	C	10/11/18 SAS
J8622-FS	VC-AQ-EB10-10022018	285.0	1	C	10/11/18 SAS
J8623-FS	VC-S09GW01-1018	270.0	1	C	10/11/18 SAS
J8624-FS	VC-S09GW04-1018	275.0	1	C	10/11/18 SAS
J8625-FS	VC-S09GW04P-1018	275.0	1	C	10/11/18 SAS
J8626-FS	VC-S09GW06-1018	275.0	1	C	10/11/18 SAS
J8627-FS	VC-S14GW01-1018	275.0	1	C	10/11/18 SAS
J8628-FS	VC-S14GW03-1018	280.0	1	C	10/11/18 SAS

Comments:

Samples Assigned By Jonathan Thorn Date : October 5, 2018

* - "C" = Sample is Consumed

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SURROGATE SPIKE FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Sample ID	Standard ID	Type	Vial No.	Vol Added (uL)	Date Spiked/ Spiked By	Witn'd By	Comment
CR939PB-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
CR940LCS-FS	JZ88	LCS/MS	1	150	10/10/18 SAS	SG	NA
CR940LCS-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8613-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8614-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8615-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8616-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8617-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8618-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8619-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8620-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8621-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8622-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8623-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8624-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8625-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8626-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8627-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA
J8628-FS	KB72	SIS	1	50	10/10/18 SAS	SG	NA

Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
JZ88	Pipette	B814657482
KB72	Pipette	B814659662

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SAMPLE EXTRACTION FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Sample ID	1st Extraction	2nd Extraction	3rd Extraction	Conc. ID	Turbo °C	Turbo PSI	KD °C	Comment
CR939PB-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
CR940LCS-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8613-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8614-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8615-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8616-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8617-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8618-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8619-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8620-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8621-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8622-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8623-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8624-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8625-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8626-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8627-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J8628-FS	10/10/18 SAS	NA	NA	NA	NA	NA	NA	NA

Solvents/Reagent Preparations:

Name	ID	Expires	Lot No	Procedure	Comments
0.4% NH3 in Methanol	RP-181010-3	10/10/18	183857	Per 100 mL, 3.5 mL ammonia solution brought to 100 mL with methanol	
0.4% NH3 in Methanol	RP-181010-3	10/10/18	SHBJ0412	Per 100 mL, 3.5 mL ammonia solution brought to 100 mL with methanol	
Pre-packed SPE Column	RP-181010-7	10/10/18	003537220A/ 0035	Pre-packed SPE Column	

Solvents/Reagents:

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
INTERNAL STANDARD SPIKING FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

(N/A Fraction)

Extract Id	Extr. Vol. (uL)	Added (uL)	Std. Id	Accm. (uL)	Vial No.	Pre Inj. Vol. (uL)^	Final Dilution *	Date Spiked/ Spiked By	Witn'd By
CR939PB-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
CR940LCS-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8613-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8613-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8613-FS-D(5)	951	49	KC03	50	1	1000	50.000	10/18/18 SAS	KB
J8613-FS-D(7)	965	35	KC03	50	1	1000	166.667	10/18/18 SAS	KB
J8614-FS(0)	950	50	KC03	50	1	1000	1.000	10/16/18 SAS	KB
J8615-FS(0)	950	50	KC03	50	1	1000	1.000	10/16/18 SAS	KB
J8616-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8616-FS-D(3)	953	47	KC03	50	1	1000	16.667	10/18/18 SAS	KB
J8617-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8617-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8618-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8618-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8619-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8619-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8620-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8620-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8621-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB

* - Final Dilution is any HPLC, dilutions, or other manipulation

^ - Pre Injection Volume (PIV) includes any RIS spikes.

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
INTERNAL STANDARD SPIKING FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

(N/A Fraction)

Extract Id	Extr. Vol. (uL)	Added (uL)	Std. Id	Accm. (uL)	Vial No.	Pre Inj. Vol. (uL)^	Final Dilution *	Date Spiked/ Spiked By	Witn'd By
J8622-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8623-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8623-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8624-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8624-FS-D(3)	953	47	KC03	50	1	1000	16.667	10/18/18 SAS	KB
J8625-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8625-FS-D(3)	953	47	KC03	50	1	1000	16.667	10/18/18 SAS	KB
J8626-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8626-FS-D(3)	955	45	KC03	50	1	1000	10.000	10/18/18 SAS	KB
J8627-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB
J8627-FS-D(3)	952	48	KC03	50	1	1000	25.000	10/18/18 SAS	KB
J8627-FS-D(5)	960	40	KC52	40	1	1000	125.000	10/24/18 LMG	KB
J8628-FS(0)	950	50	KC03	50	1	1000	1.000	10/18/18 SAS	KB

Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
KB72	Pipette	B814659662
KC03	Pipette	B814659662
KC19	Pipette	B814659662
KC52	Pipette	B814659662

* - Final Dilution is any HPLC, dilutions, or other manipulation

^ - Pre Injection Volume (PIV) includes any RIS spikes.

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
EXTRACT SPIKE FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Extract Id	DF	Std. ID	Type	Vial No.	Vol. Added (uL)	Conc (ug/mL)	Added (ng)	Date Spiked/ Spiked By	Witn'd By
J8613-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8613-FS-D(5)	50	KB72	SIS	1	49	0	0	10/18/18 SAS	KB
J8613-FS-D(7)	166.667	KB72	SIS	1	35	0	0	10/18/18 SAS	KB
J8616-FS-D(3)	16.667	KB72	SIS	1	47	0	0	10/18/18 SAS	KB
J8617-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8618-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8619-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8620-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8623-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8624-FS-D(3)	16.667	KB72	SIS	1	47	0	0	10/18/18 SAS	KB
J8625-FS-D(3)	16.667	KB72	SIS	1	47	0	0	10/18/18 SAS	KB
J8626-FS-D(3)	10	KB72	SIS	1	45	0	0	10/18/18 SAS	KB
J8627-FS-D(3)	25	KB72	SIS	1	48	0	0	10/18/18 SAS	KB
J8627-FS-D(5)	125	KC19	SIS	1	40	0	0	10/24/18 LMG	KB

Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
KB72	Pipette	B814659662
KC03	Pipette	B814659662
KC19	Pipette	B814659662
KC52	Pipette	B814659662

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
PREPARATION EXTRACT SPLIT FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					
CR939PB-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS
CR940LCS-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8613-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8613-FS	2	C	10/16/2018 3:32:00 PM	J8613-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8613-FS-D	3	--	10/16/2018 3:32:00 PM	J8613-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8613-FS	4	--	10/16/2018 3:33:00 PM	J8613-FS	2	900	880	1.023	1.136	10/16/18 SAS
J8613-FS-D	5	C	10/16/2018 3:33:00 PM	J8613-FS	2	900	20	45.000	50.000	10/16/18 SAS
J8613-FS-D	6	--	10/16/2018 3:34:00 PM	J8613-FS-D	5	1000	700	1.429	71.429	10/16/18 SAS
J8613-FS-D	7	--	10/16/2018 3:34:00 PM	J8613-FS-D	5	1000	300	3.333	166.667	10/16/18 SAS
J8614-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8615-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8616-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8616-FS	2	--	10/16/2018 3:36:00 PM	J8616-FS	0	1000	940	1.064	1.064	10/16/18 SAS
J8616-FS-D	3	--	10/16/2018 3:36:00 PM	J8616-FS	0	1000	60	16.667	16.667	10/16/18 SAS

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] * [Aliquot Weight (mg)]
Dilution Factor = [Sample Volume (uL) / Aliqot Volume (uL))] * Prior Dilution Factor
* - "C" = Extract is Consumed

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
PREPARATION EXTRACT SPLIT FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					
J8617-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8617-FS	2	--	10/16/2018 3:42:00 PM	J8617-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8617-FS-D	3	--	10/16/2018 3:42:00 PM	J8617-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8618-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8618-FS	2	--	10/16/2018 3:42:00 PM	J8618-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8618-FS-D	3	--	10/16/2018 3:42:00 PM	J8618-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8619-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8619-FS	2	--	10/16/2018 3:42:00 PM	J8619-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8619-FS-D	3	--	10/16/2018 3:42:00 PM	J8619-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8620-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8620-FS	2	--	10/16/2018 3:42:00 PM	J8620-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8620-FS-D	3	--	10/16/2018 3:42:00 PM	J8620-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8621-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8622-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] * [Aliquot Weight (mg)]
Dilution Factor = [Sample Volume (uL) / Aliqot Volume (uL))] * Prior Dilution Factor
* - "C" = Extract is Consumed

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
PREPARATION EXTRACT SPLIT FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					
J8623-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8623-FS	2	--	10/16/2018 3:42:00 PM	J8623-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8623-FS-D	3	--	10/16/2018 3:42:00 PM	J8623-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8624-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8624-FS	2	--	10/16/2018 3:45:00 PM	J8624-FS	0	1000	940	1.064	1.064	10/16/18 SAS
J8624-FS-D	3	--	10/16/2018 3:45:00 PM	J8624-FS	0	1000	60	16.667	16.667	10/16/18 SAS
J8625-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8625-FS	2	--	10/16/2018 3:45:00 PM	J8625-FS	0	1000	940	1.064	1.064	10/16/18 SAS
J8625-FS-D	3	--	10/16/2018 3:45:00 PM	J8625-FS	0	1000	60	16.667	16.667	10/16/18 SAS
J8626-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8626-FS	2	--	10/16/2018 3:42:00 PM	J8626-FS	0	1000	900	1.111	1.111	10/16/18 SAS
J8626-FS-D	3	--	10/16/2018 3:42:00 PM	J8626-FS	0	1000	100	10.000	10.000	10/16/18 SAS
J8627-FS	0	C	10/10/2018 10:44:00 AM	NA		NA	NA	1.000	1.000	10/10/18 SAS
J8627-FS	2	--	10/16/2018 3:47:00 PM	J8627-FS	0	1000	960	1.042	1.042	10/16/18 SAS

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] * [Aliquot Weight (mg)]
Dilution Factor = [Sample Volume (uL) / Aliqot Volume (uL))] * Prior Dilution Factor
* - "C" = Extract is Consumed

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
PREPARATION EXTRACT SPLIT FORM

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					
J8627-FS-D	3	C	10/16/2018 3:47:00 PM	J8627-FS	0	1000	40	25.000	25.000	10/16/18 SAS
J8627-FS-D	4	--	10/24/2018 10:32:00 AM	J8627-FS-D	3	1000	800	1.250	31.250	10/24/18 LMG
J8627-FS-D	5	--	10/24/2018 10:32:00 AM	J8627-FS-D	3	1000	200	5.000	125.000	10/24/18 LMG
J8628-FS	0	--	10/10/2018 2:03:00 PM	NA		NA	NA	1.000	1.000	10/10/18 SAS

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] * [Aliquot Weight (mg)]
Dilution Factor = [Sample Volume (uL) / Aliqot Volume (uL))] * Prior Dilution Factor
* - "C" = Extract is Consumed

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
EXTRACT - INSTRUMENT FACILITY CUSTODY PAGE

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Purpose:	LC-MS/MS TRANSFER	**Last Activity:**	Prep->Inst
Relinquished On/By:	Oct 22 2018 5:05PM KB	**Received On/By:**	Oct 22 2018 5:05PM LMG
Relinquished From:	Sample Preparation: NA	**Received Location:**	LC Laboratory: NA
Relinquish Comment:	NA	**Received Comment:**	NA

No.	BDO-ID:	PIV:	DF:	Condition:	Custody Comment:
1	CR939PB-FS(0)	1000	1	Intact	NA
2	CR940LCS-FS(0)	1000	1	Intact	NA
3	J8613-FS(0)	1000	1	Intact	NA
4	J8613-FS-D(3)	1000	10	Intact	NA
5	J8613-FS-D(5)	1000	50	Intact	NA
6	J8613-FS-D(7)	1000	166.667	Intact	NA
7	J8614-FS(0)	1000	1	Intact	NA
8	J8615-FS(0)	1000	1	Intact	NA
9	J8616-FS(0)	1000	1	Intact	NA
10	J8616-FS-D(3)	1000	16.667	Intact	NA
11	J8617-FS(0)	1000	1	Intact	NA
12	J8617-FS-D(3)	1000	10	Intact	NA
13	J8618-FS(0)	1000	1	Intact	NA
14	J8618-FS-D(3)	1000	10	Intact	NA
15	J8619-FS(0)	1000	1	Intact	NA
16	J8619-FS-D(3)	1000	10	Intact	NA
17	J8620-FS(0)	1000	1	Intact	NA
18	J8620-FS-D(3)	1000	10	Intact	NA
19	J8621-FS(0)	1000	1	Intact	NA
20	J8622-FS(0)	1000	1	Intact	NA
21	J8623-FS(0)	1000	1	Intact	NA
22	J8623-FS-D(3)	1000	10	Intact	NA
23	J8624-FS(0)	1000	1	Intact	NA
24	J8624-FS-D(3)	1000	16.667	Intact	NA
25	J8625-FS(0)	1000	1	Intact	NA
26	J8625-FS-D(3)	1000	16.667	Intact	NA
27	J8626-FS(0)	1000	1	Intact	NA
28	J8626-FS-D(3)	1000	10	Intact	NA

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
EXTRACT - INSTRUMENT FACILITY CUSTODY PAGE

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

29	J8627-FS(0)	1000	1	Intact	NA
30	J8627-FS-D(3)	1000	25	Intact	NA
31	J8628-FS(0)	1000	1	Intact	NA

Total Extracts: 31

Purpose: LC-MS/MS TRANSFER

Last Activity: Prep->Inst

Relinquished On/By: Oct 24 2018 11:00AM LMG

Relinquished From: Sample Preparation: NA

Relinquish Comment: NA

Received On/By: Oct 24 2018 11:00AM LMG

Received Location: LC Laboratory: NA

Received Comment: NA

No.	BDO-ID:	PIV:	DF:	Condition:	Custody Comment:
1	J8627-FS-D(5)	1000	125	Intact	NA

Total Extracts: 1

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
SAMPLE SPECIFIC COMMENTS

Project Title(s)
CTO-4164 Naval Base Ventura County, California

Project No.(s)
100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Sample ID:	Comment:	Date/Initials:
CR939PB-FS	Extraction started 2:03pm, extraction block 3, ended 3:03pm	10/10/18 SAS
CR940LCS-FS	Extraction started 2:03pm, extraction block 3, ended 3:03pm	10/10/18 SAS
J8613-FS	Extraction started 10:44am, extraction block 2, ended 11:50am	10/10/18 SAS
J8614-FS	Extraction started 2:03pm, extraction block 3, ended 3:10pm	10/10/18 SAS
J8615-FS	Extraction started 2:03pm, extraction block 3, ended 3:13pm	10/10/18 SAS
J8616-FS	Extraction started 10:44am, extraction block 2, ended 11:45am	10/10/18 SAS
J8617-FS	Extraction started 10:44am, extraction block 1, ended 12:02pm. Sample had floating particulates	10/10/18 SAS
J8618-FS	Extraction started 10:44am, extraction block 1, ended 12:08pm. Sample had floating particulates	10/10/18 SAS
J8619-FS	Extraction started 10:44am, extraction block 1, ended 12:11pm. Sample had floating particulates	10/10/18 SAS
J8620-FS	Extraction started 10:44am, extraction block 1, ended 12:09pm. Sample had floating particulates	10/10/18 SAS
J8621-FS	Extraction started 2:03pm, extraction block 3, ended 3:06pm	10/10/18 SAS
J8622-FS	Extraction started 2:03pm, extraction block 3, ended 3:14pm	10/10/18 SAS
J8623-FS	Extraction started 10:44am, extraction block 1, ended 11:41am. Sample had floating particulates	10/10/18 SAS
J8624-FS	Extraction started 10:44am, extraction block 1, ended 11:45am. Sample had floating particulates	10/10/18 SAS
J8625-FS	Extraction started 10:44am, extraction block 1, ended 11:55am. Sample had floating particulates	10/10/18 SAS
J8626-FS	Extraction started 10:44am, extraction block 1, ended 12:07pm. Sample had floating particulates	10/10/18 SAS
J8627-FS	Extraction started 10:44am, extraction block 2, ended 11:44am	10/10/18 SAS
J8628-FS	Extraction started 2:03pm, extraction block 3, ended 3:22pm	10/10/18 SAS

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
MISCELLANEOUS DOCUMENTATION FORM

Project Title(s)	**Project No.(s)**
CTO-4164 Naval Base Ventura County, California	100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Entered By:	**On:**

Task Leader Approval:	**On:**
SupervisorApproval:	**On:**
PM Approval:	**On:**

BATTELLE
It can be done

BATTELLE - NORWELL OPERATIONS
MISCELLANEOUS DOCUMENTATION FORM

Project Title(s)	**Project No.(s)**
CTO-4164 Naval Base Ventura County, California	100110125-01

18-0602

CTO-4164: Analysis of Non-Potable Waters

AQ, GW

Task Leader Approval:	**On:**
SupervisorApproval:	**On:**
PM Approval:	**On:**

Analytical Calibrations

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:59 PM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	MeOH		10/17/2018 7:36:00 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
2	KB73	L1	10/17/2018 7:46:52 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
3	KB74	L2	10/17/2018 7:57:45 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
4	KB75	L3	10/17/2018 8:08:39 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
5	KB76	L4	10/17/2018 8:19:32 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
6	KB77	L5	10/17/2018 8:30:23 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
7	KB78	L6	10/17/2018 8:41:14 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
8	KB79	L7	10/17/2018 8:52:06 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
9	KB80 IB	IB	10/17/2018 9:02:57 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
10	KB81 ICC	ICC	10/17/2018 9:13:49 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
11	KB89 Branch	BRANCH	10/17/2018 9:24:41 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	KB75 ISC	ISC	10/22/2018 7:32:46 PM	5-0369.dam	10222018_5500.wiff
2	KB80 IB	IB	10/22/2018 7:43:39 PM	5-0369.dam	10222018_5500.wiff
3	MeOH		10/22/2018 7:54:32 PM	5-0369.dam	10222018_5500.wiff
4	J8473-FS(3)		10/22/2018 8:05:24 PM	5-0369.dam	10222018_5500.wiff
5	J8665-FS(3)		10/22/2018 8:16:16 PM	5-0369.dam	10222018_5500.wiff
6	J8691-FS(3)		10/22/2018 8:27:08 PM	5-0369.dam	10222018_5500.wiff
7	J8692-FS(3)		10/22/2018 8:38:01 PM	5-0369.dam	10222018_5500.wiff
8	J8692-FS-D(5)		10/22/2018 8:48:53 PM	5-0369.dam	10222018_5500.wiff
9	J8693-FS(3)		10/22/2018 8:59:45 PM	5-0369.dam	10222018_5500.wiff
10	J8694-FS(3)		10/22/2018 9:10:36 PM	5-0369.dam	10222018_5500.wiff
11	J8695-FS(3)		10/22/2018 9:21:28 PM	5-0369.dam	10222018_5500.wiff
12	J8596-FS(3)		10/22/2018 9:32:20 PM	5-0369.dam	10222018_5500.wiff
13	KB77 CCV		10/22/2018 9:43:12 PM	5-0369.dam	10222018_5500.wiff
14	MeOH		10/22/2018 9:54:04 PM	5-0369.dam	10222018_5500.wiff
15	J8696-FS-D(5)		10/22/2018 10:04:57 PM	5-0369.dam	10222018_5500.wiff
16	J8697-FS(3)		10/22/2018 10:15:49 PM	5-0369.dam	10222018_5500.wiff
13	KB77 CCV		10/22/2018 10:26:39 PM	5-0369.dam	10222018_5500.wiff
14	MeOH		10/22/2018 10:37:32 PM	5-0369.dam	10222018_5500.wiff
17	CR992PB-FS(3)		10/22/2018 10:48:24 PM	5-0369.dam	10222018_5500.wiff
18	CR993LCS-FS(3)		10/22/2018 10:59:15 PM	5-0369.dam	10222018_5500.wiff
19	J6898-FS(3)		10/22/2018 11:10:06 PM	5-0369.dam	10222018_5500.wiff
20	J8699-FS(3)		10/22/2018 11:20:58 PM	5-0369.dam	10222018_5500.wiff
21	J8700-FS(3)		10/22/2018 11:31:51 PM	5-0369.dam	10222018_5500.wiff
22	J8701-FS(3)		10/22/2018 11:42:44 PM	5-0369.dam	10222018_5500.wiff
23	J8702-FS(3)		10/22/2018 11:53:34 PM	5-0369.dam	10222018_5500.wiff
24	J8703-FS(3)		10/23/2018 12:04:26 AM	5-0369.dam	10222018_5500.wiff
25	J8703-FS-D(5)		10/23/2018 12:15:17	5-0369.dam	10222018_5500.wiff

(1)

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File	
			AM			
26	J8703-FS-D(7)		10/23/2018 12:26:10 AM	5-0369.dam	10222018_5500.wiff	
27	KB76 CCV		10/23/2018 12:37:00 AM	5-0369.dam	10222018_5500.wiff	
28	MeOH		10/23/2018 12:47:52 AM	5-0369.dam	10222018_5500.wiff	
29	J8704-FS(3)		10/23/2018 12:58:45 AM	5-0369.dam	10222018_5500.wiff	
30	J8704-FS-D(5)		10/23/2018 1:09:37 AM	5-0369.dam	10222018_5500.wiff	
31	J8704-FS-D(7)		10/23/2018 1:20:29 AM	5-0369.dam	10222018_5500.wiff	
32	J8705-FS(3)		10/23/2018 1:31:20 AM	5-0369.dam	10222018_5500.wiff	
33	J8705-FS-D(5)		10/23/2018 1:42:11 AM	5-0369.dam	10222018_5500.wiff	
34	J8706-FS(3)		10/23/2018 1:53:02 AM	5-0369.dam	10222018_5500.wiff	
35	J8707-FS(3)		10/23/2018 2:03:54 AM	5-0369.dam	10222018_5500.wiff	(1)
36	J8708-FS(3)		10/23/2018 2:14:46 AM	5-0369.dam	10222018_5500.wiff	
37	J8709-FS(3)		10/23/2018 2:25:37 AM	5-0369.dam	10222018_5500.wiff	
38	J8710-FS(3)		10/23/2018 2:36:28 AM	5-0369.dam	10222018_5500.wiff	
39	KB77 CCV		10/23/2018 2:47:20 AM	5-0369.dam	10222018_5500.wiff	
40	MeOH		10/23/2018 2:58:12 AM	5-0369.dam	10222018_5500.wiff	
41	J8711-FS(3)		10/23/2018 3:09:04 AM	5-0369.dam	10222018_5500.wiff	
42	J8712MS-FS(3)		10/23/2018 3:19:56 AM	5-0369.dam	10222018_5500.wiff	
43	J8713MSD-FS(3)		10/23/2018 3:30:48 AM	5-0369.dam	10222018_5500.wiff	
44	KB76 CCV	CCV	10/23/2018 3:41:40 AM	5-0369.dam	10222018_5500.wiff	
45	MeOH		10/23/2018 3:52:32 AM	5-0369.dam	10222018_5500.wiff	
46	CR939PB-FS(0)	Procedural Blank	10/23/2018 4:03:23 AM	5-0369.dam	10222018_5500.wiff	
47	CR940LCS-FS(0)	Laboratory Control Sample	10/23/2018 4:14:16 AM	5-0369.dam	10222018_5500.wiff	(2)
48	J8613-FS(0)	VC-PM553-GW0601A-1018	10/23/2018 4:25:07 AM	5-0369.dam	10222018_5500.wiff	
49	J8613-FS-D(3)	VC-PM553-GW0601A-1018	10/23/2018 4:36:00 AM	5-0369.dam	10222018_5500.wiff	
50	J8613-FS-D(5)	VC-PM553-GW0601A-1018	10/23/2018 4:46:53 AM	5-0369.dam	10222018_5500.wiff	
51	J8613-FS-D(7)	VC-PM553-GW0601A-1018	10/23/2018 4:57:46 AM	5-0369.dam	10222018_5500.wiff	

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
52	J8614-FS(0)	VC-AQ-FB09-10012018	10/23/2018 5:08:37 AM	5-0369.dam	10222018_5500.wiff
53	J8615-FS(0)	VC-AQ-EB09-10012018	10/23/2018 5:19:29 AM	5-0369.dam	10222018_5500.wiff
54	KB77 CCV	CCV	10/23/2018 5:30:20 AM	5-0369.dam	10222018_5500.wiff
1	MeOH		10/23/2018 5:41:11 AM	5-0369.dam	10222018_5500.wiff
2	J8616-FS(0)	VC-PM13-GW21-1018	10/23/2018 5:52:05 AM	5-0369.dam	10222018_5500.wiff
3	~~J8616-FS-D(3)~~	~~VC-PM13-GW21-1018~~	10/23/2018 6:02:58 AM	~~5-0369.dam~~	~~10222018_5500.wiff~~ (3)
4	J8617-FS(0)	VC-KCHA16-GW01-1018	10/23/2018 6:13:52 AM	5-0369.dam	10222018_5500.wiff
5	J8617-FS-D(3)	VC-KCHA16-GW01-1018	10/23/2018 6:24:47 AM	5-0369.dam	10222018_5500.wiff
6	J8618-FS(0)	VC-KCHA16-GW01P-1018	10/23/2018 6:35:40 AM	5-0369.dam	10222018_5500.wiff
7	J8618-FS-D(3)	VC-KCHA16-GW01P-1018	10/23/2018 6:46:33 AM	5-0369.dam	10222018_5500.wiff
8	J8619-FS-(0)	VC-KCHA16-GW02-1018	10/23/2018 6:57:26 AM	5-0369.dam	10222018_5500.wiff
9	J8619-FS-D(3)	VC-KCHA16-GW02-1018	10/23/2018 7:08:19 AM	5-0369.dam	10222018_5500.wiff
10	J8620-FS(0)	VC-KCHA16-GW03-1018	10/23/2018 7:19:12 AM	5-0369.dam	10222018_5500.wiff
11	J8620-FS-D(3)	VC-KCHA16-GW03-1018	10/23/2018 7:30:04 AM	5-0369.dam	10222018_5500.wiff
12	KB76 CCV	CCV	10/23/2018 7:40:57 AM	5-0369.dam	10222018_5500.wiff
13	MeOH		10/23/2018 7:51:51 AM	5-0369.dam	10222018_5500.wiff
14	J8621-FS(0)	VC-AQ-FB10-10022018	10/23/2018 8:02:44 AM	5-0369.dam	10222018_5500.wiff
15	J8622-FS(0)	VC-AQ-EB10-10022018	10/23/2018 8:13:36 AM	5-0369.dam	10222018_5500.wiff
16	J8623-FS(0)	VC-S09GW01-1018	10/23/2018 8:24:29 AM	5-0369.dam	10222018_5500.wiff
17	J8623-FS-D(3)	VC-S09GW01-1018	10/23/2018 8:35:21 AM	5-0369.dam	10222018_5500.wiff
18	J8624-FS(0)	VC-S09GW04-1018	10/23/2018 8:46:14 AM	5-0369.dam	10222018_5500.wiff
19	J8624-FS-D(3)	VC-S09GW04-1018	10/23/2018 9:19:12 AM	5-0369.dam	10222018_5500.wiff
20	J8625-FS(0)	VC-S09GW04P-1018	10/23/2018 9:30:05 AM	5-0369.dam	10222018_5500.wiff
21	J8625-FS-D(3)	VC-S09GW04P-1018	10/23/2018 9:40:58 AM	5-0369.dam	10222018_5500.wiff
22	J8626-FS(0)	VC-S09GW06-1018	10/23/2018 9:51:52 AM	5-0369.dam	10222018_5500.wiff
23	J8626-FS-D(3)	VC-S09GW06-1018	10/23/2018 10:02:45 AM	5-0369.dam	10222018_5500.wiff
24	KB77 CCV	CCV	10/23/2018 10:13:38	5-0369.dam	10222018_5500.wiff

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
			AM		
25	MeOH		10/23/2018 10:24:31 AM	5-0369.dam	10222018_5500.wiff
26	J8627-FS(0)	VC-S14GW01-1018	10/23/2018 10:35:23 AM	5-0369.dam	10222018_5500.wiff
27	J8627-FS-D(3)	VC-S14GW01-1018	10/23/2018 10:46:16 AM	5-0369.dam	10222018_5500.wiff
28	J8628-FS(0)	VC-S14GW03-1018	10/23/2018 10:57:08 AM	5-0369.dam	10222018_5500.wiff
29	KB76 CCV	CCV	10/23/2018 11:08:03 AM	5-0369.dam	10222018_5500.wiff

(1) Samples do not apply to this batch. LMG 10/25/18

(2) The LCS was realiquoted and reanalyzed due to a high recovery. All data for the LCS was reported from the reanalysis. LMG 10/25/18

(3) This dilution was realiquoted and reanalyzed due to a high surrogate recovery. All data for this dilution was reported from the reanalysis. LMG 10/25/18

BATTELLE

Sequence Report

Created with Analyst Reporter
Printed: 25/10/2018 11:20:26 AM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	MeOH		10/17/2018 7:36:00 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
2	KB73	L1	10/17/2018 7:46:52 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
3	KB74	L2	10/17/2018 7:57:45 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
4	KB75	L3	10/17/2018 8:08:39 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
5	KB76	L4	10/17/2018 8:19:32 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
6	KB77	L5	10/17/2018 8:30:23 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
7	KB78	L6	10/17/2018 8:41:14 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
8	KB79	L7	10/17/2018 8:52:06 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
9	KB80 IB	IB	10/17/2018 9:02:57 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
10	KB81 ICC	ICC	10/17/2018 9:13:49 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
11	KB89 Branch	BRANCH	10/17/2018 9:24:41 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
1	MeOH		10/17/2018 9:35:33 PM	5-0369.dam	Data18-0590_18-01588_18-0589.wiff
25	KB75 ISC	ISC	10/24/2018 11:51:20 AM	5-0369.dam	AC_10242018_369.wiff
26	KB80 IB	IB	10/24/2018 12:02:11 PM	5-0369.dam	AC_10242018_369.wiff
27	CR940LCS-FS(0)	Laboratory Control Sample	10/24/2018 12:13:02 PM	5-0369.dam	AC_10242018_369.wiff
28	J8616-FS-D(3)	VC-PM13-GW21-1018	10/24/2018 12:23:52 PM	5-0369.dam	AC_10242018_369.wiff
29	J8627-FS-D(5)	VC-S14GW01-1018	10/24/2018 12:34:44 PM	5-0369.dam	AC_10242018_369.wiff
30	KB76	CCV	10/24/2018 12:56:27 PM	5-0369.dam	AC_10242018_369.wiff

BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFBS_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	298.9 / 80.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C3-PFBS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 2.90746 x + 0.06970 (r = 0.99940) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	104.635578	103.6
3	KB74	L2	True	252.50	249.370464	98.8
4	KB75	L3	True	505.00	495.098560	98.0
5	KB76	L4	True	1010.00	1091.288759	108.1
6	KB77	L5	True	2525.00	2332.691625	92.4
7	KB78	L6	True	10100.00	9811.581725	97.1
8	KB79	L7	True	20200.00	20608.833290	102.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFBS_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	298.9 / 99.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C3-PFBS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.85386 x + 0.01367 (r = 0.99964) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	109.976962	108.9
3	KB74	L2	True	252.50	242.180085	95.9
4	KB75	L3	True	505.00	475.121472	94.1
5	KB76	L4	True	1010.00	1086.381545	107.6
6	KB77	L5	True	2525.00	2339.199841	92.6
7	KB78	L6	True	10100.00	10143.472130	100.4
8	KB79	L7	True	20200.00	20297.167965	100.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHxA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	313.0 / 269.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C5-PFHxA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.98501 x + 0.13333 (r = 0.99994) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	98.968181	98.0
3	KB74	L2	True	252.50	269.910942	106.9
4	KB75	L3	True	505.00	490.752146	97.2
5	KB76	L4	True	1010.00	1004.472242	99.5
6	KB77	L5	True	2525.00	2495.865956	98.9
7	KB78	L6	True	10100.00	9993.499522	99.0
8	KB79	L7	True	20200.00	20340.031011	100.7



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHxA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	313.0 / 119.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C5-PFHxA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.07336 x + 0.01493 (r = 0.99989) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	113.183087	112.1
3	KB74	L2	True	252.50	230.645561	91.3
4	KB75	L3	True	505.00	514.901118	102.0
5	KB76	L4	True	1010.00	980.952621	97.1
6	KB77	L5	True	2525.00	2451.366061	97.1
7	KB78	L6	True	10100.00	10083.260482	99.8
8	KB79	L7	True	20200.00	20319.191070	100.6



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHpA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	363.0 / 319.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C4-PFHpA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.87722 x + 0.06760 (r = 0.99955) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	113.465532	113.5
3	KB74	L2	True	250.00	261.922885	104.8
4	KB75	L3	True	500.00	452.888167	90.6
5	KB76	L4	True	1000.00	963.772233	96.4
6	KB77	L5	True	2500.00	2376.996854	95.1
7	KB78	L6	True	10000.00	9765.162244	97.7
8	KB79	L7	True	20000.00	20415.792084	102.1



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHpA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	363.0 / 169.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C4-PFHpA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.01856 x + 0.00664 (r = 0.99907) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	103.311634	103.3
3	KB74	L2	True	250.00	175.823351	70.3
4	KB75	L3	True	500.00	618.850678	123.8
5	KB76	L4	True	1000.00	1051.701356	105.2
6	KB77	L5	True	2500.00	2469.602759	98.8
7	KB78	L6	True	10000.00	9796.218447	98.0
8	KB79	L7	True	20000.00	20134.491774	100.7



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHxS_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	399.0 / 80.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C3-PFHxS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 3.40443 x + 0.28942 (r = 0.99956) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	105.225987	104.2
3	KB74	L2	True	252.50	232.865372	92.2
4	KB75	L3	True	505.00	467.824000	92.6
5	KB76	L4	True	1010.00	1074.776448	106.4
6	KB77	L5	True	2525.00	2696.493262	106.8
7	KB78	L6	True	10100.00	9828.822898	97.3
8	KB79	L7	True	20200.00	20287.492034	100.4



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFHxS_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	399.0 / 99.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C3-PFHxS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.94916 x + 0.10161 (r = 0.99926) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	101.00	98.139274	97.2
3	KB74	L2	True	252.50	231.314187	91.6
4	KB75	L3	True	505.00	484.563583	96.0
5	KB76	L4	True	1010.00	1107.476178	109.7
6	KB77	L5	True	2525.00	2749.725828	108.9
7	KB78	L6	True	10100.00	9714.855919	96.2
8	KB79	L7	True	20200.00	20307.425032	100.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFOA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	413.0 / 369.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C8-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.00279 x + 0.06080 (r = 0.99969) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	113.608497	113.6
3	KB74	L2	True	250.00	242.259718	96.9
4	KB75	L3	True	500.00	467.456638	93.5
5	KB76	L4	True	1000.00	956.844142	95.7
6	KB77	L5	True	2500.00	2452.232813	98.1
7	KB78	L6	True	10000.00	10326.914009	103.3
8	KB79	L7	True	20000.00	19790.684183	99.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFOA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	413.0 / 169.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C8-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.06512 x + 0.00376 (r = 0.99971) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	104.849376	104.9
3	KB74	L2	True	250.00	273.238913	109.3
4	KB75	L3	True	500.00	421.303706	84.3
5	KB76	L4	True	1000.00	1028.034772	102.8
6	KB77	L5	True	2500.00	2444.883328	97.8
7	KB78	L6	True	10000.00	10121.411629	101.2
8	KB79	L7	True	20000.00	19956.278277	99.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFNA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	463.0 / 419.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C9-PFNA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.82839 x + 0.10849 (r = 0.99987) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	100.132667	100.1
3	KB74	L2	True	250.00	263.216231	105.3
4	KB75	L3	True	500.00	450.336433	90.1
5	KB76	L4	True	1000.00	1046.243999	104.6
6	KB77	L5	True	2500.00	2491.510736	99.7
7	KB78	L6	True	10000.00	10047.185055	100.5
8	KB79	L7	True	20000.00	19951.374879	99.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFNA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	463.0 / 219.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C9-PFNA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.25566 x + 0.02957 (r = 0.99963) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	111.702594	111.7
3	KB74	L2	True	250.00	220.916460	88.4
4	KB75	L3	True	500.00	448.287193	89.7
5	KB76	L4	True	1000.00	1050.052047	105.0
6	KB77	L5	True	2500.00	2632.290752	105.3
7	KB78	L6	True	10000.00	10108.558837	101.1
8	KB79	L7	True	20000.00	19778.192118	98.9



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFOS_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	499.0 / 80.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C8-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 4.79098 x + -0.03583 (r = 0.99927) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	112.998705	113.0
3	KB74	L2	True	250.00	245.807613	98.3
4	KB75	L3	True	500.00	508.954098	101.8
5	KB76	L4	True	1000.00	953.422687	95.3
6	KB77	L5	True	2500.00	2300.880982	92.0
7	KB78	L6	True	10000.00	9674.048541	96.7
8	KB79	L7	True	20000.00	20553.887374	102.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFOS_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	499.0 / 99.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C8-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.83172 x + -0.00135 (r = 0.99945) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	110.962710	111.0
3	KB74	L2	True	250.00	244.538081	97.8
4	KB75	L3	True	500.00	520.238814	104.1
5	KB76	L4	True	1000.00	958.547975	95.9
6	KB77	L5	True	2500.00	2271.299000	90.9
7	KB78	L6	True	10000.00	9849.093945	98.5
8	KB79	L7	True	20000.00	20395.319474	102.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFDA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	513.0 / 469.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C6-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.03992 x + 0.08967 (r = 0.99977) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	100.375074	100.4
3	KB74	L2	True	250.00	264.467518	105.8
4	KB75	L3	True	500.00	495.556524	99.1
5	KB76	L4	True	1000.00	987.592944	98.8
6	KB77	L5	True	2500.00	2418.646707	96.8
7	KB78	L6	True	10000.00	9760.929085	97.6
8	KB79	L7	True	20000.00	20322.432148	101.6



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFDA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	513.0 / 219.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C6-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.04284 x + 0.00408 (r = 0.99914) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	112.406023	112.4
3	KB74	L2	True	250.00	245.500787	98.2
4	KB75	L3	True	500.00	479.073328	95.8
5	KB76	L4	True	1000.00	1010.924950	101.1
6	KB77	L5	True	2500.00	2354.317167	94.2
7	KB78	L6	True	10000.00	9514.985150	95.2
8	KB79	L7	True	20000.00	20632.792594	103.2



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFUnA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	563.0 / 519.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C7-PFUnA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.04939 x + 0.04151 (r = 0.99953) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	111.527231	111.5
3	KB74	L2	True	250.00	244.824605	97.9
4	KB75	L3	True	500.00	423.690587	84.7
5	KB76	L4	True	1000.00	1026.343861	102.6
6	KB77	L5	True	2500.00	2538.187940	101.5
7	KB78	L6	True	10000.00	10323.155375	103.2
8	KB79	L7	True	20000.00	19682.270401	98.4



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFUnA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	563.0 / 269.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C7-PFUnA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.05225 x + -5.79528e-5 (r = 0.99994) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	102.220892	102.2
3	KB74	L2	True	250.00	242.442598	97.0
4	KB75	L3	True	500.00	501.608235	100.3
5	KB76	L4	True	1000.00	981.918474	98.2
6	KB77	L5	True	2500.00	2544.385755	101.8
7	KB78	L6	True	10000.00	10125.204796	101.3
8	KB79	L7	True	20000.00	19852.219250	99.3



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFDoA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	613.0 / 569.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFDoA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.90471 x + 0.10550 (r = 0.99952) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	90.939755	90.9
3	KB74	L2	True	250.00	247.072023	98.8
4	KB75	L3	True	500.00	506.360909	101.3
5	KB76	L4	True	1000.00	1050.478389	105.1
6	KB77	L5	True	2500.00	2565.539851	102.6
7	KB78	L6	True	10000.00	10368.355085	103.7
8	KB79	L7	True	20000.00	19521.253987	97.6



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFDoA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	613.0 / 319.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFDoA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.14074 x + 0.02292 (r = 0.99961) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	86.545687	86.6
3	KB74	L2	True	250.00	274.615048	109.9
4	KB75	L3	True	500.00	494.661881	98.9
5	KB76	L4	True	1000.00	971.395687	97.1
6	KB77	L5	True	2500.00	2712.576260	108.5
7	KB78	L6	True	10000.00	9996.454238	100.0
8	KB79	L7	True	20000.00	19813.751199	99.1



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFTrDA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	663.0 / 619.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFTeDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.95260 x + 0.12887 (r = 0.99908) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	85.967911	86.0
3	KB74	L2	True	250.00	241.563939	96.6
4	KB75	L3	True	500.00	488.700899	97.7
5	KB76	L4	True	1000.00	1126.764608	112.7
6	KB77	L5	True	2500.00	2664.921530	106.6
7	KB78	L6	True	10000.00	10336.521146	103.4
8	KB79	L7	True	20000.00	19405.559967	97.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFTrDA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	663.0 / 169.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFTeDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.06256 x + 0.00941 (r = 0.99887) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	78.198175	78.2
3	KB74	L2	True	250.00	247.986856	99.2
4	KB75	L3	True	500.00	495.036348	99.0
5	KB76	L4	True	1000.00	1164.067670	116.4
6	KB77	L5	True	2500.00	2684.712399	107.4
7	KB78	L6	True	10000.00	10280.911529	102.8
8	KB79	L7	True	20000.00	19399.087023	97.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFTeDA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	713.0 / 669.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFTeDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.07817 x + 0.18334 (r = 0.99951) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	81.728327	81.7
3	KB74	L2	True	250.00	253.645310	101.5
4	KB75	L3	True	500.00	515.154084	103.0
5	KB76	L4	True	1000.00	1095.645152	109.6
6	KB77	L5	True	2500.00	2603.218162	104.1
7	KB78	L6	True	10000.00	10217.289136	102.2
8	KB79	L7	True	20000.00	19583.319829	97.9



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	PFTeDA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	713.0 / 169.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	13C2-PFTeDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.05216 x + 0.01003 (r = 0.99933) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	89.886117	89.9
3	KB74	L2	True	250.00	238.937138	95.6
4	KB75	L3	True	500.00	490.005825	98.0
5	KB76	L4	True	1000.00	1092.239213	109.2
6	KB77	L5	True	2500.00	2680.771848	107.2
7	KB78	L6	True	10000.00	10258.453725	102.6
8	KB79	L7	True	20000.00	19499.706134	97.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	NMeFOSAA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	570.0 / 419.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	d3-MeFOSAA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.93472 x + 0.21431 (r = 0.99704) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	75.824991	75.8
3	KB74	L2	True	250.00	268.876365	107.6
4	KB75	L3	True	500.00	500.524214	100.1
5	KB76	L4	True	1000.00	1133.404189	113.3
6	KB77	L5	True	2500.00	2465.257046	98.6
7	KB78	L6	True	10000.00	11007.670601	110.1
8	KB79	L7	True	20000.00	18898.442593	94.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	NMeFOSAA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	570.0 / 512.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	d3-MeFOSAA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.50367 x + 0.21759 (r = 0.99687) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	False	100.00	3.453918	3.5
3	KB74	L2	True	250.00	218.419800	87.4
4	KB75	L3	True	500.00	439.990383	88.0
5	KB76	L4	True	1000.00	1163.925478	116.4
6	KB77	L5	True	2500.00	2616.439396	104.7
7	KB78	L6	True	10000.00	10905.550987	109.1
8	KB79	L7	True	20000.00	18905.673956	94.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	NEtFOSAA_1	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	584.0 / 419.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	d5-EtFOSAA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.90168 x + 0.04836 (r = 0.99948) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	100.00	104.771450	104.8
3	KB74	L2	True	250.00	251.018896	100.4
4	KB75	L3	True	500.00	508.056535	101.6
5	KB76	L4	True	1000.00	943.782521	94.4
6	KB77	L5	True	2500.00	2522.760733	100.9
7	KB78	L6	True	10000.00	9564.590700	95.7
8	KB79	L7	True	20000.00	20455.019164	102.3



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:10:09 PM

Analyte Name	NEtFOSAA_2	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	584.0 / 483.0	**Result Table**	18-0610_18-0611_BASE
Internal Standard	d5-EtFOSAA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.05241 x + 0.01326 (r = 0.99796) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	False	100.00	41.412646	41.4
3	KB74	L2	True	250.00	191.667209	76.7
4	KB75	L3	True	500.00	637.470069	127.5
5	KB76	L4	True	1000.00	945.207707	94.5
6	KB77	L5	True	2500.00	2637.389496	105.5
7	KB78	L6	True	10000.00	9326.284871	93.3
8	KB79	L7	True	20000.00	20511.980648	102.6



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C2-PFDoA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	615.0 / 570.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.97505 x (std. dev. = 0.03727) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	234.027671	93.6
3	KB74	L2	True	250.00	265.932266	106.4
4	KB75	L3	True	250.00	249.447374	99.8
5	KB76	L4	True	250.00	245.945631	98.4
6	KB77	L5	True	250.00	248.975274	99.6
7	KB78	L6	True	250.00	251.235272	100.5
8	KB79	L7	True	250.00	254.436513	101.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	d3-MeFOSAA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	573.0 / 419.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C4-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.51076 x (std. dev. = 0.04023) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	252.745204	101.1
3	KB74	L2	True	250.00	237.867686	95.2
4	KB75	L3	True	250.00	255.434971	102.2
5	KB76	L4	True	250.00	232.186572	92.9
6	KB77	L5	True	250.00	236.383719	94.6
7	KB78	L6	True	250.00	285.381847	114.2
8	KB79	L7	False	250.00	319.413436	127.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	d5-EtFOSAA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	589.0 / 419.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C4-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.57975 x (std. dev. = 0.04771) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	261.116785	104.5
3	KB74	L2	True	250.00	260.842878	104.3
4	KB75	L3	True	250.00	225.192349	90.1
5	KB76	L4	True	250.00	254.251713	101.7
6	KB77	L5	True	250.00	220.425706	88.2
7	KB78	L6	True	250.00	278.106701	111.2
8	KB79	L7	True	250.00	250.063868	100.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C5-PFHxA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	318.0 / 273.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.73994 x (std. dev. = 0.03937) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	262.567314	105.0
3	KB74	L2	True	250.00	254.187234	101.7
4	KB75	L3	True	250.00	230.325952	92.1
5	KB76	L4	True	250.00	235.058052	94.0
6	KB77	L5	True	250.00	257.145365	102.9
7	KB78	L6	True	250.00	264.503275	105.8
8	KB79	L7	True	250.00	246.212808	98.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C4-PFHpA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	367.0 / 322.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.84182 x (std. dev. = 0.03866) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	262.427043	105.0
3	KB74	L2	True	250.00	251.075240	100.4
4	KB75	L3	True	250.00	244.763156	97.9
5	KB76	L4	True	250.00	256.507439	102.6
6	KB77	L5	True	250.00	256.669677	102.7
7	KB78	L6	True	250.00	251.342992	100.5
8	KB79	L7	True	250.00	227.214453	90.9



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C8-PFOA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	421.0 / 376.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.02405 x (std. dev. = 0.06175) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	255.212785	102.1
3	KB74	L2	True	250.00	262.875155	105.2
4	KB75	L3	True	250.00	256.482263	102.6
5	KB76	L4	True	250.00	257.662297	103.1
6	KB77	L5	True	250.00	256.019871	102.4
7	KB78	L6	True	250.00	243.157749	97.3
8	KB79	L7	True	250.00	218.589880	87.4



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C9-PFNA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	472.0 / 427.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFOA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 1.16930 x (std. dev. = 0.04759) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	247.430463	99.0
3	KB74	L2	True	250.00	261.328420	104.5
4	KB75	L3	True	250.00	259.384215	103.8
5	KB76	L4	True	250.00	244.372857	97.8
6	KB77	L5	True	250.00	249.179395	99.7
7	KB78	L6	True	250.00	256.344825	102.5
8	KB79	L7	True	250.00	231.959824	92.8



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C6-PFDA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	519.0 / 474.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.98235 x (std. dev. = 0.06079) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	246.733475	98.7
3	KB74	L2	True	250.00	253.248818	101.3
4	KB75	L3	True	250.00	257.192135	102.9
5	KB76	L4	True	250.00	248.450649	99.4
6	KB77	L5	True	250.00	274.445566	109.8
7	KB78	L6	True	250.00	247.435316	99.0
8	KB79	L7	True	250.00	222.494042	89.0



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C7-PFUnA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	570.0 / 525.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.90743 x (std. dev. = 0.07148) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	247.760751	99.1
3	KB74	L2	True	250.00	262.960755	105.2
4	KB75	L3	True	250.00	286.201660	114.5
5	KB76	L4	True	250.00	234.954208	94.0
6	KB77	L5	True	250.00	249.135582	99.7
7	KB78	L6	True	250.00	242.726450	97.1
8	KB79	L7	True	250.00	226.260594	90.5



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C2-PFTeDA	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	715.0 / 670.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C2-PFDA	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.78723 x (std. dev. = 0.03476) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	250.00	242.249601	96.9
3	KB74	L2	True	250.00	262.922814	105.2
4	KB75	L3	True	250.00	245.975733	98.4
5	KB76	L4	True	250.00	231.048930	92.4
6	KB77	L5	True	250.00	252.097071	100.8
7	KB78	L6	True	250.00	258.482884	103.4
8	KB79	L7	True	250.00	257.222966	102.9



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C3-PFBS	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	302.0 / 99.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C4-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.93048 x (std. dev. = 0.11752) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	232.25	225.022462	96.9
3	KB74	L2	True	232.25	215.609121	92.8
4	KB75	L3	True	232.25	218.748838	94.2
5	KB76	L4	True	232.25	203.233433	87.5
6	KB77	L5	True	232.25	216.348971	93.2
7	KB78	L6	True	232.25	264.163371	113.7
8	KB79	L7	True	232.25	282.623804	121.7



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C3-PFHxS	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	402.0 / 99.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C4-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.84316 x (std. dev. = 0.07172) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	236.50	231.144826	97.7
3	KB74	L2	True	236.50	241.172849	102.0
4	KB75	L3	True	236.50	245.474666	103.8
5	KB76	L4	True	236.50	217.481590	92.0
6	KB77	L5	True	236.50	204.966129	86.7
7	KB78	L6	True	236.50	262.529838	111.0
8	KB79	L7	True	236.50	252.730102	106.9



BATTELLE


Calibration Summary Report

Created with Analyst Reporter
Printed: 24/10/2018 1:09:45 PM

Analyte Name	13C8-PFOS	**Data File**	Data18-0590_18-01588_18-0589.wiff
MRM Transition	507.0 / 99.0	**Result Table**	18-0610_18-0611_SIS
Internal Standard	13C4-PFOS	**Instrument Name**	QTRAP 5500
Acquisition Date	10/17/2018 7:36:00 PM	**Acquisition Method**	5-0369.dam

Regression Equation: y = 0.95250 x (std. dev. = 0.05159) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	KB73	L1	True	239.25	250.871078	104.9
3	KB74	L2	True	239.25	243.225761	101.7
4	KB75	L3	True	239.25	223.367998	93.4
5	KB76	L4	True	239.25	233.089271	97.4
6	KB77	L5	True	239.25	226.800752	94.8
7	KB78	L6	True	239.25	259.494294	108.5
8	KB79	L7	True	239.25	237.900846	99.4



BATTELLE

	Isotope Dilution Calibration Curve Concentrations (ng/L)						
	KB73	KB74	KB75	KB76	KB77	KB78	KB79
PFBA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFPeA	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
PFHxA	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
PFHpA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFOA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFNA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFDA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFUnA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFDoA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFTrDA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFTeDA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
NMeFOSAA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
NEtFOSAA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFOSA	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFBS	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
PFPeS	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFHxS (Branched)	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
PFHpS	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFOS (Branched)	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
PFDS	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
PFNS	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
4:2FTS	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
6:2FTS	100.00	250.00	500.00	1,000.00	2,500.00	10,000.00	20,000.00
8:2FTS	101.00	252.50	505.00	1,010.00	2,525.00	10,100.00	20,200.00
	Surrogates / Extracted Internal Standards						
13C4-PFBA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C5-PFPeA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C5-PFHxA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C4-PFHpA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C8-PFOA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C9-PFNA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C6-PFDA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C7-PFUnA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C2-PFDoA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C2-PFTeDA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
d3-MeFOSAA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
d5-EtFOSAA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C8-FOSA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C3-PFBS	232.25	232.25	232.25	232.25	232.25	232.25	232.25
13C3-PFHxS	236.50	236.50	236.50	236.50	236.50	236.50	236.50
13C8-PFOS	239.25	239.25	239.25	239.25	239.25	239.25	239.25
13C2-4:2FTS	233.75	233.75	233.75	233.75	233.75	233.75	233.75
13C2-6:2FTS	237.25	237.25	237.25	237.25	237.25	237.25	237.25
13C2-8:2FTS	239.50	239.50	239.50	239.50	239.50	239.50	239.50
	Internal Standards						
13C3-PFBA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C2-PFOA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C2-PFDA	250.00	250.00	250.00	250.00	250.00	250.00	250.00
13C4-PFOS	239.25	239.25	239.25	239.25	239.25	239.25	239.25

BATTELLE

Summary Quant Report

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.57	36035.22	104.635578	126.2	true
PFBS_2	298.9 / 99.0	1.57	10918.15	109.976962	78.2	true
PFHxA_1	313.0 / 269.0	1.90	32164.19	98.968181	3.3	false
PFHxA_2	313.0 / 119.0	1.91	2959.44	113.183087	3.5	false
PFHpA_1	363.0 / 319.0	2.32	32551.91	113.465532	39.8	false
PFHpA_2	363.0 / 169.0	2.32	1000.19	103.311634	29.1	false
PFHxS_1	399.0 / 80.0	2.34	44418.07	105.225987	134.5	false
PFHxS_2	399.0 / 99.0	2.34	12198.64	98.139274	86.9	false
PFOA_1	413.0 / 369.0	2.73	42708.15	113.608497	67.5	false
PFOA_2	413.0 / 169.0	2.73	2568.95	104.849376	46.6	false
PFNA_1	463.0 / 419.0	3.14	40302.36	100.132667	94.3	false
PFNA_2	463.0 / 219.0	3.14	13162.53	111.702594	118.3	false
PFOS_1	499.0 / 80.0	3.13	65728.52	112.998705	102.9	false
PFOS_2	499.0 / 99.0	3.13	11345.50	110.962710	100.1	false
PFDA_1	513.0 / 469.0	3.50	49242.60	100.375074	136.7	false
PFDA_2	513.0 / 219.0	3.49	2266.31	112.406023	88.7	false
PFUnA_1	563.0 / 519.0	3.83	45897.49	111.527231	163.3	false
PFUnA_2	563.0 / 269.0	3.82	1918.57	102.220892	42.6	true
PFDoA_1	613.0 / 569.0	4.11	39723.04	90.939755	179.1	false
PFDoA_2	613.0 / 319.0	4.11	6547.85	86.545687	145.5	false
PFTrDA_1	663.0 / 619.0	4.37	34866.78	85.967911	309.6	false
PFTrDA_2	663.0 / 169.0	4.36	2214.00	78.198175	110.2	false
PFTeDA_1	713.0 / 669.0	4.59	40930.04	81.728327	685.4	false
PFTeDA_2	713.0 / 169.0	4.58	2198.76	89.886117	212.0	false
NMeFOSAA_1	570.0 / 419.0	3.65	8081.03	75.824991	298.4	false
NMeFOSAA_2	570.0 / 512.0	3.65	3645.21	3.453918	224.2	false
NEtFOSAA_1	584.0 / 419.0	3.82	8180.80	104.771450	524.6	false
NEtFOSAA_2	584.0 / 483.0	3.82	421.19	41.412646	48.5	false

BATTELLE

Summary Quant Report

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.57	93283.23	249.370464	246.3	false
PFBS_2	298.9 / 99.0	1.57	26423.99	242.180085	146.1	false
PFHxA_1	313.0 / 269.0	1.90	81008.49	269.910942	7.1	false
PFHxA_2	313.0 / 119.0	1.89	5592.03	230.645561	6.5	false
PFHpA_1	363.0 / 319.0	2.31	75049.99	261.922885	57.6	false
PFHpA_2	363.0 / 169.0	2.31	1497.90	175.823351	28.9	true
PFHxS_1	399.0 / 80.0	2.33	108184.43	232.865372	188.9	false
PFHxS_2	399.0 / 99.0	2.33	30598.50	231.314187	310.4	false
PFOA_1	413.0 / 369.0	2.73	100033.19	242.259718	128.9	false
PFOA_2	413.0 / 169.0	2.73	7259.40	273.238913	110.3	false
PFNA_1	463.0 / 419.0	3.13	107846.09	263.216231	192.5	false
PFNA_2	463.0 / 219.0	3.12	28095.55	220.916460	213.9	false
PFOS_1	499.0 / 80.0	3.12	165271.40	245.807613	161.0	false
PFOS_2	499.0 / 99.0	3.12	28706.92	244.538081	211.7	true
PFDA_1	513.0 / 469.0	3.49	122994.62	264.467518	236.7	false
PFDA_2	513.0 / 219.0	3.48	4770.77	245.500787	791.7	false
PFUnA_1	563.0 / 519.0	3.82	106014.14	244.824605	237.3	false
PFUnA_2	563.0 / 269.0	3.81	5018.00	242.442598	81.5	false
PFDoA_1	613.0 / 569.0	4.10	107705.01	247.072023	318.6	false
PFDoA_2	613.0 / 319.0	4.10	19126.54	274.615048	276.9	false
PFTrDA_1	663.0 / 619.0	4.35	90249.73	241.563939	481.4	false
PFTrDA_2	663.0 / 169.0	4.35	6147.30	247.986856	208.6	false
PFTeDA_1	713.0 / 669.0	4.58	109852.37	253.645310	931.4	false
PFTeDA_2	713.0 / 169.0	4.58	5150.42	238.937138	369.4	false
NMeFOSAA_1	570.0 / 419.0	3.65	21539.90	268.876365	843.1	false
NMeFOSAA_2	570.0 / 512.0	3.65	11614.86	218.419800	528.2	false
NEtFOSAA_1	584.0 / 419.0	3.81	21525.62	251.018896	397.7	false
NEtFOSAA_2	584.0 / 483.0	3.80	1206.24	191.667209	290.9	false

BATTELLE

Summary Quant Report

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.57	163617.84	495.098560	296.3	false
PFBS_2	298.9 / 99.0	1.57	45956.46	475.121472	212.6	false
PFHxA_1	313.0 / 269.0	1.89	123711.77	490.752146	10.9	false
PFHxA_2	313.0 / 119.0	1.89	9937.29	514.901118	10.8	false
PFHpA_1	363.0 / 319.0	2.31	119885.50	452.888167	85.7	false
PFHpA_2	363.0 / 169.0	2.31	3804.54	618.850678	67.0	false
PFHxS_1	399.0 / 80.0	2.33	185842.71	467.824000	256.2	false
PFHxS_2	399.0 / 99.0	2.33	54144.25	484.563583	238.2	false
PFOA_1	413.0 / 369.0	2.72	178566.76	467.456638	160.8	false
PFOA_2	413.0 / 169.0	2.72	10469.56	421.303706	138.9	false
PFNA_1	463.0 / 419.0	3.12	170504.46	450.336433	203.9	false
PFNA_2	463.0 / 219.0	3.12	51980.47	448.287193	301.6	false
PFOS_1	499.0 / 80.0	3.12	275067.90	508.954098	204.8	false
PFOS_2	499.0 / 99.0	3.12	48946.70	520.238814	287.5	false
PFDA_1	513.0 / 469.0	3.48	213151.42	495.556524	301.2	false
PFDA_2	513.0 / 219.0	3.48	8539.13	479.073328	239.1	false
PFUnA_1	563.0 / 519.0	3.81	185383.49	423.690587	269.5	false
PFUnA_2	563.0 / 269.0	3.81	10671.72	501.608235	165.4	false
PFDoA_1	613.0 / 569.0	4.09	184868.00	506.360909	343.9	false
PFDoA_2	613.0 / 319.0	4.09	28751.08	494.661881	331.3	false
PFTrDA_1	663.0 / 619.0	4.35	151211.59	488.700899	597.4	false
PFTrDA_2	663.0 / 169.0	4.34	10123.70	495.036348	322.3	false
PFTeDA_1	713.0 / 669.0	4.57	182656.44	515.154084	1232.8	false
PFTeDA_2	713.0 / 169.0	4.56	8526.49	490.005825	501.7	false
NMeFOSAA_1	570.0 / 419.0	3.63	34656.73	500.524214	2338.0	false
NMeFOSAA_2	570.0 / 512.0	3.64	18345.00	439.990383	658.5	false
NEtFOSAA_1	584.0 / 419.0	3.80	31971.45	508.056535	633.3	false
NEtFOSAA_2	584.0 / 483.0	3.80	2497.16	637.470069	8056.1	true

BATTELLE

Summary Quant Report

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	336045.72	1091.288759	552.2	false
PFBS_2	298.9 / 99.0	1.56	98081.89	1086.381545	407.4	false
PFHxA_1	313.0 / 269.0	1.89	240693.36	1004.472242	18.1	false
PFHxA_2	313.0 / 119.0	1.89	17814.42	980.952621	16.5	false
PFHpA_1	363.0 / 319.0	2.30	251954.77	963.772233	144.8	false
PFHpA_2	363.0 / 169.0	2.30	6187.13	1051.701356	84.1	false
PFHxS_1	399.0 / 80.0	2.32	369801.30	1074.776448	333.3	false
PFHxS_2	399.0 / 99.0	2.32	106671.06	1107.476178	416.8	false
PFOA_1	413.0 / 369.0	2.71	347996.75	956.844142	237.9	false
PFOA_2	413.0 / 169.0	2.71	24237.17	1028.034772	244.5	false
PFNA_1	463.0 / 419.0	3.11	345587.51	1046.243999	410.7	false
PFNA_2	463.0 / 219.0	3.11	106652.75	1050.052047	427.7	false
PFOS_1	499.0 / 80.0	3.11	549475.41	953.422687	298.0	false
PFOS_2	499.0 / 99.0	3.11	96044.01	958.547975	432.6	false
PFDA_1	513.0 / 469.0	3.47	409723.46	987.592944	454.6	false
PFDA_2	513.0 / 219.0	3.47	17306.29	1010.924950	319.9	false
PFUnA_1	563.0 / 519.0	3.80	370869.35	1026.343861	315.7	false
PFUnA_2	563.0 / 269.0	3.80	17491.40	981.918474	206.5	false
PFDoA_1	613.0 / 569.0	4.09	374694.74	1050.478389	498.6	false
PFDoA_2	613.0 / 319.0	4.09	54643.36	971.395687	458.0	false
PFTrDA_1	663.0 / 619.0	4.34	321677.62	1126.764608	812.9	false
PFTrDA_2	663.0 / 169.0	4.34	21874.98	1164.067670	449.0	false
PFTeDA_1	713.0 / 669.0	4.56	357047.01	1095.645152	1545.8	false
PFTeDA_2	713.0 / 169.0	4.56	17306.66	1092.239213	701.8	false
NMeFOSAA_1	570.0 / 419.0	3.63	68030.30	1133.404189	751.5	false
NMeFOSAA_2	570.0 / 512.0	3.63	39158.03	1163.925478	709.7	false
NEtFOSAA_1	584.0 / 419.0	3.80	66289.88	943.782521	754.6	false
NEtFOSAA_2	584.0 / 483.0	3.80	4059.44	945.207707	43600.8	false

BATTELLE

Summary Quant Report

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.57	776289.58	2332.691625	927.0	false
PFBS_2	298.9 / 99.0	1.56	228434.75	2339.199841	734.2	false
PFHxA_1	313.0 / 269.0	1.89	609668.03	2495.865956	39.8	false
PFHxA_2	313.0 / 119.0	1.89	44913.51	2451.366061	32.3	false
PFHpA_1	363.0 / 319.0	2.30	584043.68	2376.996854	232.0	false
PFHpA_2	363.0 / 169.0	2.30	13193.99	2469.602759	211.4	false
PFHxS_1	399.0 / 80.0	2.32	886495.01	2696.493262	388.0	false
PFHxS_2	399.0 / 99.0	2.32	252473.06	2749.725828	571.1	false
PFOA_1	413.0 / 369.0	2.71	834165.44	2452.232813	357.8	false
PFOA_2	413.0 / 169.0	2.71	53993.69	2444.883328	444.0	false
PFNA_1	463.0 / 419.0	3.11	783461.97	2491.510736	552.9	false
PFNA_2	463.0 / 219.0	3.11	254909.55	2632.290752	603.7	false
PFOS_1	499.0 / 80.0	3.11	1306443.89	2300.880982	452.4	false
PFOS_2	499.0 / 99.0	3.11	224020.69	2271.299000	545.6	false
PFDA_1	513.0 / 469.0	3.47	986588.88	2418.646707	564.2	false
PFDA_2	513.0 / 219.0	3.47	39607.95	2354.317167	415.6	false
PFUnA_1	563.0 / 519.0	3.80	871741.30	2538.187940	460.9	false
PFUnA_2	563.0 / 269.0	3.80	43333.05	2544.385755	300.0	false
PFDoA_1	613.0 / 569.0	4.09	821796.76	2565.539851	662.4	false
PFDoA_2	613.0 / 319.0	4.09	135656.47	2712.576260	522.0	false
PFTrDA_1	663.0 / 619.0	4.33	735746.86	2664.921530	986.7	false
PFTrDA_2	663.0 / 169.0	4.33	48743.98	2684.712399	618.7	false
PFTeDA_1	713.0 / 669.0	4.56	816378.76	2603.218162	1929.9	false
PFTeDA_2	713.0 / 169.0	4.55	40737.36	2680.771848	1193.3	false
NMeFOSAA_1	570.0 / 419.0	3.63	151508.36	2465.257046	2245.1	false
NMeFOSAA_2	570.0 / 512.0	3.63	88173.15	2616.439396	615.9	false
NEtFOSAA_1	584.0 / 419.0	3.79	155433.42	2522.760733	648.0	false
NEtFOSAA_2	584.0 / 483.0	3.79	9620.27	2637.389496	1231310.6	false

BATTELLE

Summary Quant Report

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	3735985.71	9811.581725	1503.9	false
PFBS_2	298.9 / 99.0	1.56	1134069.29	10143.472130	1449.0	false
PFHxA_1	313.0 / 269.0	1.88	2658827.36	9993.499522	90.8	false
PFHxA_2	313.0 / 119.0	1.88	200130.60	10083.260482	76.5	false
PFHpA_1	363.0 / 319.0	2.30	2497856.31	9765.162244	436.5	false
PFHpA_2	363.0 / 169.0	2.30	53385.07	9796.218447	441.5	false
PFHxS_1	399.0 / 80.0	2.32	3854430.51	9828.822898	984.2	false
PFHxS_2	399.0 / 99.0	2.32	1062752.95	9714.855919	1104.6	false
PFOA_1	413.0 / 369.0	2.71	3551933.53	10326.914009	921.2	false
PFOA_2	413.0 / 169.0	2.71	226064.95	10121.411629	875.1	false
PFNA_1	463.0 / 419.0	3.11	3442576.47	10047.185055	870.1	false
PFNA_2	463.0 / 219.0	3.11	1068510.97	10108.558837	1018.0	false
PFOS_1	499.0 / 80.0	3.10	5888305.73	9674.048541	719.2	false
PFOS_2	499.0 / 99.0	3.10	1040866.54	9849.093945	1187.7	false
PFDA_1	513.0 / 469.0	3.47	4121337.28	9760.929085	867.9	false
PFDA_2	513.0 / 219.0	3.46	165545.81	9514.985150	1080.3	false
PFUnA_1	563.0 / 519.0	3.79	3980687.38	10323.155375	894.2	false
PFUnA_2	563.0 / 269.0	3.79	194191.57	10125.204796	674.2	false
PFDoA_1	613.0 / 569.0	4.08	3840657.97	10368.355085	1057.4	false
PFDoA_2	613.0 / 319.0	4.08	576764.61	9996.454238	1020.7	false
PFTrDA_1	663.0 / 619.0	4.33	3350407.66	10336.521146	1444.2	false
PFTrDA_2	663.0 / 169.0	4.33	218945.29	10280.911529	1007.9	false
PFTeDA_1	713.0 / 669.0	4.55	3751651.69	10217.289136	2725.0	false
PFTeDA_2	713.0 / 169.0	4.54	182332.73	10258.453725	1759.6	false
NMeFOSAA_1	570.0 / 419.0	3.62	753927.92	11007.670601	2278.1	false
NMeFOSAA_2	570.0 / 512.0	3.62	404364.11	10905.550987	1426.8	false
NEtFOSAA_1	584.0 / 419.0	3.79	707833.72	9564.590700	1144.3	false
NEtFOSAA_2	584.0 / 483.0	3.79	40332.26	9326.284871	2770.2	false

BATTELLE

Summary Quant Report

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	7654212.63	20608.833290	2863.5	false
PFBS_2	298.9 / 99.0	1.56	2213700.05	20297.167965	1861.8	false
PFHxA_1	313.0 / 269.0	1.89	5068064.70	20340.031011	123.0	false
PFHxA_2	313.0 / 119.0	1.88	377383.63	20319.191070	98.6	false
PFHpA_1	363.0 / 319.0	2.29	4752932.81	20415.792084	545.6	false
PFHpA_2	363.0 / 169.0	2.30	99502.18	20134.491774	637.6	false
PFHxS_1	399.0 / 80.0	2.32	6929917.90	20287.492034	1348.5	false
PFHxS_2	399.0 / 99.0	2.32	1934460.53	20307.425032	1153.9	false
PFOA_1	413.0 / 369.0	2.71	6162786.15	19790.684183	1174.3	false
PFOA_2	413.0 / 169.0	2.71	403546.52	19956.278277	1429.0	false
PFNA_1	463.0 / 419.0	3.11	6224206.25	19951.374879	1314.8	false
PFNA_2	463.0 / 219.0	3.11	1903895.81	19778.192118	1655.3	false
PFOS_1	499.0 / 80.0	3.10	10402856.43	20553.887374	904.0	true
PFOS_2	499.0 / 99.0	3.10	1792140.82	20395.319474	1641.5	false
PFDA_1	513.0 / 469.0	3.47	7455074.70	20322.432148	1181.3	false
PFDA_2	513.0 / 219.0	3.47	311824.53	20632.792594	1380.7	false
PFUnA_1	563.0 / 519.0	3.79	6840458.50	19682.270401	830.9	false
PFUnA_2	563.0 / 269.0	3.79	343323.53	19852.219250	642.0	false
PFDoA_1	613.0 / 569.0	4.07	7074590.20	19521.253987	1036.7	false
PFDoA_2	613.0 / 319.0	4.07	1117673.36	19813.751199	984.9	false
PFTrDA_1	663.0 / 619.0	4.32	6045542.29	19405.559967	1420.4	false
PFTrDA_2	663.0 / 169.0	4.32	396998.36	19399.087023	1127.9	false
PFTeDA_1	713.0 / 669.0	4.54	6908120.06	19583.319829	2951.8	false
PFTeDA_2	713.0 / 169.0	4.54	332887.97	19499.706134	2522.7	false
NMeFOSAA_1	570.0 / 419.0	3.62	1331189.45	18898.442593	2363.5	false
NMeFOSAA_2	570.0 / 512.0	3.62	719500.36	18905.673956	1612.0	false
NEtFOSAA_1	584.0 / 419.0	3.78	1231583.02	20455.019164	1168.4	false
NEtFOSAA_2	584.0 / 483.0	3.78	71957.40	20511.980648	877.0	false

Summary Quant Report

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.10	91402.83	234.027671	1231.7	false
d3-MeFOSAA	573.0 / 419.0	3.65	16104.22	252.745204	203.5	false
d5-EtFOSAA	589.0 / 419.0	3.81	18885.04	261.116785	197.5	false
13C5-PFHxA	318.0 / 273.0	1.89	61468.01	262.567314	846.3	false
13C4-PFHpA	367.0 / 322.0	2.30	69894.37	262.427043	1233.4	false
13C8-PFOA	421.0 / 376.0	2.72	82686.75	255.212785	7328.7	false
13C9-PFNA	472.0 / 427.0	3.12	91536.04	247.430463	984.5	false
13C6-PFDA	519.0 / 474.0	3.48	97086.56	246.733475	862.5	false
13C7-PFUnA	570.0 / 525.0	3.81	90055.69	247.760751	681.9	false
13C2-PFTeDA	715.0 / 670.0	4.58	76389.03	242.249601	2670.6	false
13C3-PFBS	302.0 / 99.0	1.55	26120.14	225.022462	521.2	false
13C3-PFHxS	402.0 / 99.0	2.33	24312.83	231.144826	504.5	false
13C8-PFOS	507.0 / 99.0	3.12	29809.56	250.871078	289.5	false

BATTELLE

Summary Quant Report

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.09	107747.25	265.932266	1391.1	false
d3-MeFOSAA	573.0 / 419.0	3.64	17700.53	237.867686	243.9	false
d5-EtFOSAA	589.0 / 419.0	3.80	22032.11	260.842878	257.9	false
13C5-PFHxA	318.0 / 273.0	1.89	67688.32	254.187234	641.8	false
13C4-PFHpA	367.0 / 322.0	2.30	76065.72	251.075240	798.7	false
13C8-PFOA	421.0 / 376.0	2.72	96880.10	262.875155	404.7	false
13C9-PFNA	472.0 / 427.0	3.11	109970.74	261.328420	1372.3	false
13C6-PFDA	519.0 / 474.0	3.48	103376.35	253.248818	1556.3	false
13C7-PFUnA	570.0 / 525.0	3.80	99154.47	262.960755	932.2	false
13C2-PFTeDA	715.0 / 670.0	4.57	86008.02	262.922814	1938.6	false
13C3-PFBS	302.0 / 99.0	1.55	29228.78	215.609121	465.4	false
13C3-PFHxS	402.0 / 99.0	2.32	29626.05	241.172849	441.9	false
13C8-PFOS	507.0 / 99.0	3.11	33752.70	243.225761	275.4	false

BATTELLE

Summary Quant Report

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.08	95394.66	249.447374	1919.0	false
d3-MeFOSAA	573.0 / 419.0	3.63	16732.69	255.434971	259.5	false
d5-EtFOSAA	589.0 / 419.0	3.79	16744.24	225.192349	304.7	false
13C5-PFHxA	318.0 / 273.0	1.88	59853.55	230.325952	495.7	false
13C4-PFHpA	367.0 / 322.0	2.29	72363.27	244.763156	5134.6	false
13C8-PFOA	421.0 / 376.0	2.71	92242.14	256.482263	3147.3	false
13C9-PFNA	472.0 / 427.0	3.10	106517.52	259.384215	1506.1	false
13C6-PFDA	519.0 / 474.0	3.47	99092.64	257.192135	1016.5	false
13C7-PFUnA	570.0 / 525.0	3.79	101859.96	286.201660	947.8	false
13C2-PFTeDA	715.0 / 670.0	4.56	75947.40	245.975733	2310.8	false
13C3-PFBS	302.0 / 99.0	1.55	26105.02	218.748838	604.7	false
13C3-PFHxS	402.0 / 99.0	2.32	26545.24	245.474666	396.8	false
13C8-PFOS	507.0 / 99.0	3.10	27286.93	223.367998	263.5	false

BATTELLE

Summary Quant Report

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.08	95903.63	245.945631	1244.5	false
d3-MeFOSAA	573.0 / 419.0	3.62	15347.55	232.186572	188.7	false
d5-EtFOSAA	589.0 / 419.0	3.79	19076.22	254.251713	210.8	false
13C5-PFHxA	318.0 / 273.0	1.88	58835.13	235.058052	1009.1	false
13C4-PFHpA	367.0 / 322.0	2.29	73044.35	256.507439	778.8	false
13C8-PFOA	421.0 / 376.0	2.70	89256.01	257.662297	1337.8	false
13C9-PFNA	472.0 / 427.0	3.10	96659.60	244.372857	1496.3	false
13C6-PFDA	519.0 / 474.0	3.46	97605.58	248.450649	1104.6	false
13C7-PFUnA	570.0 / 525.0	3.78	85263.94	234.954208	671.4	false
13C2-PFTeDA	715.0 / 670.0	4.55	72740.35	231.048930	2036.8	false
13C3-PFBS	302.0 / 99.0	1.55	24473.16	203.233433	669.1	false
13C3-PFHxS	402.0 / 99.0	2.32	23731.17	217.481590	485.7	false
13C8-PFOS	507.0 / 99.0	3.10	28732.45	233.089271	315.7	false

BATTELLE

Summary Quant Report

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.07	87520.82	248.975274	984.0	false
d3-MeFOSAA	573.0 / 419.0	3.62	15905.32	236.383719	163.7	false
d5-EtFOSAA	589.0 / 419.0	3.78	16835.01	220.425706	271.8	false
13C5-PFHxA	318.0 / 273.0	1.88	61167.94	257.145365	677.6	false
13C4-PFHpA	367.0 / 322.0	2.29	69461.61	256.669677	849.4	false
13C8-PFOA	421.0 / 376.0	2.70	84283.75	256.019871	1295.1	false
13C9-PFNA	472.0 / 427.0	3.10	93667.25	249.179395	1439.8	false
13C6-PFDA	519.0 / 474.0	3.46	97196.35	274.445566	2304.2	false
13C7-PFUnA	570.0 / 525.0	3.78	81503.66	249.135582	792.6	false
13C2-PFTeDA	715.0 / 670.0	4.55	71548.15	252.097071	2268.3	false
13C3-PFBS	302.0 / 99.0	1.55	26519.94	216.348971	787.4	false
13C3-PFHxS	402.0 / 99.0	2.31	22766.78	204.966129	549.1	false
13C8-PFOS	507.0 / 99.0	3.10	28458.87	226.800752	280.5	false

BATTELLE

Summary Quant Report

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.07	102072.61	251.235272	905.7	false
d3-MeFOSAA	573.0 / 419.0	3.61	18026.94	285.381847	140.7	false
d5-EtFOSAA	589.0 / 419.0	3.78	19940.39	278.106701	245.9	false
13C5-PFHxA	318.0 / 273.0	1.87	67298.42	264.503275	733.1	false
13C4-PFHpA	367.0 / 322.0	2.28	72755.48	251.342992	508.8	false
13C8-PFOA	421.0 / 376.0	2.70	85622.31	243.157749	1201.3	false
13C9-PFNA	472.0 / 427.0	3.09	103069.19	256.344825	1142.1	false
13C6-PFDA	519.0 / 474.0	3.45	101281.21	247.435316	1442.1	false
13C7-PFUnA	570.0 / 525.0	3.78	91776.59	242.726450	856.9	false
13C2-PFTeDA	715.0 / 670.0	4.54	84788.29	258.482884	1567.2	false
13C3-PFBS	302.0 / 99.0	1.54	30399.13	264.163371	680.5	false
13C3-PFHxS	402.0 / 99.0	2.31	27375.94	262.529838	495.9	false
13C8-PFOS	507.0 / 99.0	3.09	30568.34	259.494294	315.3	false

BATTELLE

Summary Quant Report

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.07	99995.01	254.436513	1141.6	false
d3-MeFOSAA	573.0 / 419.0	3.62	18400.21	319.413436	102.1	false
d5-EtFOSAA	589.0 / 419.0	3.78	16351.10	250.063868	251.7	false
13C5-PFHxA	318.0 / 273.0	1.87	63134.84	246.212808	622.1	false
13C4-PFHpA	367.0 / 322.0	2.28	66285.64	227.214453	928.8	false
13C8-PFOA	421.0 / 376.0	2.70	77573.52	218.589880	452635.8	false
13C9-PFNA	472.0 / 427.0	3.09	93994.33	231.959824	113653.9	false
13C6-PFDA	519.0 / 474.0	3.45	88095.93	222.494042	731.4	false
13C7-PFUnA	570.0 / 525.0	3.78	82754.96	226.260594	1283.3	false
13C2-PFTeDA	715.0 / 670.0	4.54	81617.66	257.222966	1991.1	false
13C3-PFBS	302.0 / 99.0	1.54	29660.00	282.623804	815.0	false
13C3-PFHxS	402.0 / 99.0	2.31	24033.73	252.730102	656.4	false
13C8-PFOS	507.0 / 99.0	3.09	25557.23	237.900846	289.4	false

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:24 PM

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.57	PFBS			
PFBS_2	298.9 / 99.0	1.57	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.90	PFHxA			
PFHxA_2	313.0 / 119.0	1.91	PFHxA	0.090	0.077	ü
PFHpA_1	363.0 / 319.0	2.32	PFHpA			
PFHpA_2	363.0 / 169.0	2.32	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.34	PFHxS			
PFHxS_2	399.0 / 99.0	2.34	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.73	PFOA			
PFOA_2	413.0 / 169.0	2.73	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.14	PFNA			
PFNA_2	463.0 / 219.0	3.14	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.13	PFOS			
PFOS_2	499.0 / 99.0	3.13	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.50	PFDA			
PFDA_2	513.0 / 219.0	3.49	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.83	PFUnA			
PFUnA_2	563.0 / 269.0	3.82	PFUnA	0.040	0.049	ü
PFDoA_1	613.0 / 569.0	4.11	PFDoA			
PFDoA_2	613.0 / 319.0	4.11	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.37	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.36	PFTrDA	0.060	0.066	ü
PFTeDA_1	713.0 / 669.0	4.59	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.58	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.65	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.65	NMeFOSAA	0.450	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.82	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.82	NEtFOSAA	0.050	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:29 PM

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.57	PFBS			
PFBS_2	298.9 / 99.0	1.57	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.90	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.31	PFHpA			
PFHpA_2	363.0 / 169.0	2.31	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.33	PFHxS			
PFHxS_2	399.0 / 99.0	2.33	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.73	PFOA			
PFOA_2	413.0 / 169.0	2.73	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.13	PFNA			
PFNA_2	463.0 / 219.0	3.12	PFNA	0.260	0.306	ü
PFOS_1	499.0 / 80.0	3.12	PFOS			
PFOS_2	499.0 / 99.0	3.12	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.49	PFDA			
PFDA_2	513.0 / 219.0	3.48	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.82	PFUnA			
PFUnA_2	563.0 / 269.0	3.81	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.10	PFDoA			
PFDoA_2	613.0 / 319.0	4.10	PFDoA	0.180	0.160	ü
PFTrDA_1	663.0 / 619.0	4.35	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.35	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.58	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.58	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.65	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.65	NMeFOSAA	0.540	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.81	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.80	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:33 PM

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.57	PFBS			
PFBS_2	298.9 / 99.0	1.57	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.31	PFHpA			
PFHpA_2	363.0 / 169.0	2.31	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.33	PFHxS			
PFHxS_2	399.0 / 99.0	2.33	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.72	PFOA			
PFOA_2	413.0 / 169.0	2.72	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.12	PFNA			
PFNA_2	463.0 / 219.0	3.12	PFNA	0.300	0.306	ü
PFOS_1	499.0 / 80.0	3.12	PFOS			
PFOS_2	499.0 / 99.0	3.12	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.48	PFDA			
PFDA_2	513.0 / 219.0	3.48	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.81	PFUnA			
PFUnA_2	563.0 / 269.0	3.81	PFUnA	0.060	0.049	ü
PFDoA_1	613.0 / 569.0	4.09	PFDoA			
PFDoA_2	613.0 / 319.0	4.09	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.35	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.34	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.57	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.56	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.64	NMeFOSAA	0.530	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.80	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.80	NEtFOSAA	0.080	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:38 PM

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.11	PFOS			
PFOS_2	499.0 / 99.0	3.11	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.47	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.80	PFUnA			
PFUnA_2	563.0 / 269.0	3.80	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.09	PFDoA			
PFDoA_2	613.0 / 319.0	4.09	PFDoA	0.150	0.160	ü
PFTrDA_1	663.0 / 619.0	4.34	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.34	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.56	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.56	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.63	NMeFOSAA	0.580	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.80	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.80	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:42 PM

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.57	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.11	PFOS			
PFOS_2	499.0 / 99.0	3.11	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.47	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.80	PFUnA			
PFUnA_2	563.0 / 269.0	3.80	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.09	PFDoA			
PFDoA_2	613.0 / 319.0	4.09	PFDoA	0.170	0.160	ü
PFTrDA_1	663.0 / 619.0	4.33	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.33	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.56	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.55	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.63	NMeFOSAA	0.580	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.79	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.79	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:46 PM

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.88	PFHxA			
PFHxA_2	313.0 / 119.0	1.88	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.10	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.46	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.79	PFUnA			
PFUnA_2	563.0 / 269.0	3.79	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.08	PFDoA			
PFDoA_2	613.0 / 319.0	4.08	PFDoA	0.150	0.160	ü
PFTrDA_1	663.0 / 619.0	4.33	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.33	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.55	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.54	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.62	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.62	NMeFOSAA	0.540	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.79	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.79	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 24/10/2018 1:12:50 PM

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.88	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.29	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.10	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.47	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.79	PFUnA			
PFUnA_2	563.0 / 269.0	3.79	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	4.07	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.32	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.32	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.54	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.54	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.62	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.62	NMeFOSAA	0.540	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.78	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.78	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Internal Standard Report

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.57	13C3-PFBS	302.0 / 99.0	26120.14	232.25
PFBS_2	298.9 / 99.0	1.57	13C3-PFBS	302.0 / 99.0	26120.14	232.25
PFHxA_1	313.0 / 269.0	1.90	13C5-PFHxA	318.0 / 273.0	61468.01	250.00
PFHxA_2	313.0 / 119.0	1.91	13C5-PFHxA	318.0 / 273.0	61468.01	250.00
PFHpA_1	363.0 / 319.0	2.32	13C4-PFHpA	367.0 / 322.0	69894.37	250.00
PFHpA_2	363.0 / 169.0	2.32	13C4-PFHpA	367.0 / 322.0	69894.37	250.00
PFHxS_1	399.0 / 80.0	2.34	13C3-PFHxS	402.0 / 99.0	24619.85	236.50
PFHxS_2	399.0 / 99.0	2.34	13C3-PFHxS	402.0 / 99.0	24619.85	236.50
PFOA_1	413.0 / 369.0	2.73	13C8-PFOA	421.0 / 376.0	82686.75	250.00
PFOA_2	413.0 / 169.0	2.73	13C8-PFOA	421.0 / 376.0	82686.75	250.00
PFNA_1	463.0 / 419.0	3.14	13C9-PFNA	472.0 / 427.0	91536.04	250.00
PFNA_2	463.0 / 219.0	3.14	13C9-PFNA	472.0 / 427.0	91536.04	250.00
PFOS_1	499.0 / 80.0	3.13	13C8-PFOS	507.0 / 99.0	29514.87	239.25
PFOS_2	499.0 / 99.0	3.13	13C8-PFOS	507.0 / 99.0	29514.87	239.25
PFDA_1	513.0 / 469.0	3.50	13C6-PFDA	519.0 / 474.0	97086.56	250.00
PFDA_2	513.0 / 219.0	3.49	13C6-PFDA	519.0 / 474.0	97086.56	250.00
PFUnA_1	563.0 / 519.0	3.83	13C7-PFUnA	570.0 / 525.0	90055.69	250.00
PFUnA_2	563.0 / 269.0	3.82	13C7-PFUnA	570.0 / 525.0	90055.69	250.00
PFDoA_1	613.0 / 569.0	4.11	13C2-PFDoA	615.0 / 570.0	91402.83	250.00
PFDoA_2	613.0 / 319.0	4.11	13C2-PFDoA	615.0 / 570.0	91402.83	250.00
PFTrDA_1	663.0 / 619.0	4.37	13C2-PFTeDA	715.0 / 670.0	76389.03	250.00
PFTrDA_2	663.0 / 169.0	4.36	13C2-PFTeDA	715.0 / 670.0	76389.03	250.00
PFTeDA_1	713.0 / 669.0	4.59	13C2-PFTeDA	715.0 / 670.0	76389.03	250.00
PFTeDA_2	713.0 / 169.0	4.58	13C2-PFTeDA	715.0 / 670.0	76389.03	250.00
NMeFOSAA_1	570.0 / 419.0	3.65	d3-MeFOSAA	573.0 / 419.0	16233.26	250.00
NMeFOSAA_2	570.0 / 512.0	3.65	d3-MeFOSAA	573.0 / 419.0	16233.26	250.00
NEtFOSAA_1	584.0 / 419.0	3.82	d5-EtFOSAA	589.0 / 419.0	19192.99	250.00
NEtFOSAA_2	584.0 / 483.0	3.82	d5-EtFOSAA	589.0 / 419.0	19192.99	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.57	13C3-PFBS	302.0 / 99.0	29228.78	232.25
PFBS_2	298.9 / 99.0	1.57	13C3-PFBS	302.0 / 99.0	29228.78	232.25
PFHxA_1	313.0 / 269.0	1.90	13C5-PFHxA	318.0 / 273.0	67688.32	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	67688.32	250.00
PFHpA_1	363.0 / 319.0	2.31	13C4-PFHpA	367.0 / 322.0	76065.72	250.00
PFHpA_2	363.0 / 169.0	2.31	13C4-PFHpA	367.0 / 322.0	76065.72	250.00
PFHxS_1	399.0 / 80.0	2.33	13C3-PFHxS	402.0 / 99.0	29708.48	236.50
PFHxS_2	399.0 / 99.0	2.33	13C3-PFHxS	402.0 / 99.0	29708.48	236.50
PFOA_1	413.0 / 369.0	2.73	13C8-PFOA	421.0 / 376.0	96880.10	250.00
PFOA_2	413.0 / 169.0	2.73	13C8-PFOA	421.0 / 376.0	96880.10	250.00
PFNA_1	463.0 / 419.0	3.13	13C9-PFNA	472.0 / 427.0	109970.74	250.00
PFNA_2	463.0 / 219.0	3.12	13C9-PFNA	472.0 / 427.0	109970.74	250.00
PFOS_1	499.0 / 80.0	3.12	13C8-PFOS	507.0 / 99.0	33822.33	239.25
PFOS_2	499.0 / 99.0	3.12	13C8-PFOS	507.0 / 99.0	33822.33	239.25
PFDA_1	513.0 / 469.0	3.49	13C6-PFDA	519.0 / 474.0	103376.35	250.00
PFDA_2	513.0 / 219.0	3.48	13C6-PFDA	519.0 / 474.0	103376.35	250.00
PFUnA_1	563.0 / 519.0	3.82	13C7-PFUnA	570.0 / 525.0	99154.47	250.00
PFUnA_2	563.0 / 269.0	3.81	13C7-PFUnA	570.0 / 525.0	99154.47	250.00
PFDoA_1	613.0 / 569.0	4.10	13C2-PFDoA	615.0 / 570.0	107747.25	250.00
PFDoA_2	613.0 / 319.0	4.10	13C2-PFDoA	615.0 / 570.0	107747.25	250.00
PFTrDA_1	663.0 / 619.0	4.35	13C2-PFTeDA	715.0 / 670.0	86008.02	250.00
PFTrDA_2	663.0 / 169.0	4.35	13C2-PFTeDA	715.0 / 670.0	86008.02	250.00
PFTeDA_1	713.0 / 669.0	4.58	13C2-PFTeDA	715.0 / 670.0	86008.02	250.00
PFTeDA_2	713.0 / 169.0	4.58	13C2-PFTeDA	715.0 / 670.0	86008.02	250.00
NMeFOSAA_1	570.0 / 419.0	3.65	d3-MeFOSAA	573.0 / 419.0	17661.42	250.00
NMeFOSAA_2	570.0 / 512.0	3.65	d3-MeFOSAA	573.0 / 419.0	17661.42	250.00
NEtFOSAA_1	584.0 / 419.0	3.81	d5-EtFOSAA	589.0 / 419.0	22570.27	250.00
NEtFOSAA_2	584.0 / 483.0	3.80	d5-EtFOSAA	589.0 / 419.0	22570.27	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.57	13C3-PFBS	302.0 / 99.0	26105.02	232.25
PFBS_2	298.9 / 99.0	1.57	13C3-PFBS	302.0 / 99.0	26105.02	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	59853.55	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	59853.55	250.00
PFHpA_1	363.0 / 319.0	2.31	13C4-PFHpA	367.0 / 322.0	72363.27	250.00
PFHpA_2	363.0 / 169.0	2.31	13C4-PFHpA	367.0 / 322.0	72363.27	250.00
PFHxS_1	399.0 / 80.0	2.33	13C3-PFHxS	402.0 / 99.0	26459.07	236.50
PFHxS_2	399.0 / 99.0	2.33	13C3-PFHxS	402.0 / 99.0	26459.07	236.50
PFOA_1	413.0 / 369.0	2.72	13C8-PFOA	421.0 / 376.0	92242.14	250.00
PFOA_2	413.0 / 169.0	2.72	13C8-PFOA	421.0 / 376.0	92242.14	250.00
PFNA_1	463.0 / 419.0	3.12	13C9-PFNA	472.0 / 427.0	106517.52	250.00
PFNA_2	463.0 / 219.0	3.12	13C9-PFNA	472.0 / 427.0	106517.52	250.00
PFOS_1	499.0 / 80.0	3.12	13C8-PFOS	507.0 / 99.0	27084.38	239.25
PFOS_2	499.0 / 99.0	3.12	13C8-PFOS	507.0 / 99.0	27084.38	239.25
PFDA_1	513.0 / 469.0	3.48	13C6-PFDA	519.0 / 474.0	99092.64	250.00
PFDA_2	513.0 / 219.0	3.48	13C6-PFDA	519.0 / 474.0	99092.64	250.00
PFUnA_1	563.0 / 519.0	3.81	13C7-PFUnA	570.0 / 525.0	101859.96	250.00
PFUnA_2	563.0 / 269.0	3.81	13C7-PFUnA	570.0 / 525.0	101859.96	250.00
PFDoA_1	613.0 / 569.0	4.09	13C2-PFDoA	615.0 / 570.0	95394.66	250.00
PFDoA_2	613.0 / 319.0	4.09	13C2-PFDoA	615.0 / 570.0	95394.66	250.00
PFTrDA_1	663.0 / 619.0	4.35	13C2-PFTeDA	715.0 / 670.0	75947.40	250.00
PFTrDA_2	663.0 / 169.0	4.34	13C2-PFTeDA	715.0 / 670.0	75947.40	250.00
PFTeDA_1	713.0 / 669.0	4.57	13C2-PFTeDA	715.0 / 670.0	75947.40	250.00
PFTeDA_2	713.0 / 169.0	4.56	13C2-PFTeDA	715.0 / 670.0	75947.40	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	16616.33	250.00
NMeFOSAA_2	570.0 / 512.0	3.64	d3-MeFOSAA	573.0 / 419.0	16616.33	250.00
NEtFOSAA_1	584.0 / 419.0	3.80	d5-EtFOSAA	589.0 / 419.0	16999.01	250.00
NEtFOSAA_2	584.0 / 483.0	3.80	d5-EtFOSAA	589.0 / 419.0	16999.01	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	24473.16	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	24473.16	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	58835.13	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	58835.13	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	73044.35	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	73044.35	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	23463.21	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	23463.21	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	89256.01	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	89256.01	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	96659.60	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	96659.60	250.00
PFOS_1	499.0 / 80.0	3.11	13C8-PFOS	507.0 / 99.0	28834.11	239.25
PFOS_2	499.0 / 99.0	3.11	13C8-PFOS	507.0 / 99.0	28834.11	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	97605.58	250.00
PFDA_2	513.0 / 219.0	3.47	13C6-PFDA	519.0 / 474.0	97605.58	250.00
PFUnA_1	563.0 / 519.0	3.80	13C7-PFUnA	570.0 / 525.0	85263.94	250.00
PFUnA_2	563.0 / 269.0	3.80	13C7-PFUnA	570.0 / 525.0	85263.94	250.00
PFDoA_1	613.0 / 569.0	4.09	13C2-PFDoA	615.0 / 570.0	95903.63	250.00
PFDoA_2	613.0 / 319.0	4.09	13C2-PFDoA	615.0 / 570.0	95903.63	250.00
PFTrDA_1	663.0 / 619.0	4.34	13C2-PFTeDA	715.0 / 670.0	72740.35	250.00
PFTrDA_2	663.0 / 169.0	4.34	13C2-PFTeDA	715.0 / 670.0	72740.35	250.00
PFTeDA_1	713.0 / 669.0	4.56	13C2-PFTeDA	715.0 / 670.0	72740.35	250.00
PFTeDA_2	713.0 / 169.0	4.56	13C2-PFTeDA	715.0 / 670.0	72740.35	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	15280.98	250.00
NMeFOSAA_2	570.0 / 512.0	3.63	d3-MeFOSAA	573.0 / 419.0	15280.98	250.00
NEtFOSAA_1	584.0 / 419.0	3.80	d5-EtFOSAA	589.0 / 419.0	19201.47	250.00
NEtFOSAA_2	584.0 / 483.0	3.80	d5-EtFOSAA	589.0 / 419.0	19201.47	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.57	13C3-PFBS	302.0 / 99.0	26519.94	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	26519.94	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	61167.94	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	61167.94	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	69461.61	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	69461.61	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	22669.25	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	22669.25	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	84283.75	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	84283.75	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	93667.25	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	93667.25	250.00
PFOS_1	499.0 / 80.0	3.11	13C8-PFOS	507.0 / 99.0	28376.74	239.25
PFOS_2	499.0 / 99.0	3.11	13C8-PFOS	507.0 / 99.0	28376.74	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	97196.35	250.00
PFDA_2	513.0 / 219.0	3.47	13C6-PFDA	519.0 / 474.0	97196.35	250.00
PFUnA_1	563.0 / 519.0	3.80	13C7-PFUnA	570.0 / 525.0	81503.66	250.00
PFUnA_2	563.0 / 269.0	3.80	13C7-PFUnA	570.0 / 525.0	81503.66	250.00
PFDoA_1	613.0 / 569.0	4.09	13C2-PFDoA	615.0 / 570.0	87520.82	250.00
PFDoA_2	613.0 / 319.0	4.09	13C2-PFDoA	615.0 / 570.0	87520.82	250.00
PFTrDA_1	663.0 / 619.0	4.33	13C2-PFTeDA	715.0 / 670.0	71548.15	250.00
PFTrDA_2	663.0 / 169.0	4.33	13C2-PFTeDA	715.0 / 670.0	71548.15	250.00
PFTeDA_1	713.0 / 669.0	4.56	13C2-PFTeDA	715.0 / 670.0	71548.15	250.00
PFTeDA_2	713.0 / 169.0	4.55	13C2-PFTeDA	715.0 / 670.0	71548.15	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	16063.93	250.00
NMeFOSAA_2	570.0 / 512.0	3.63	d3-MeFOSAA	573.0 / 419.0	16063.93	250.00
NEtFOSAA_1	584.0 / 419.0	3.79	d5-EtFOSAA	589.0 / 419.0	16992.28	250.00
NEtFOSAA_2	584.0 / 483.0	3.79	d5-EtFOSAA	589.0 / 419.0	16992.28	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	30399.13	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	30399.13	232.25
PFHxA_1	313.0 / 269.0	1.88	13C5-PFHxA	318.0 / 273.0	67298.42	250.00
PFHxA_2	313.0 / 119.0	1.88	13C5-PFHxA	318.0 / 273.0	67298.42	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	72755.48	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	72755.48	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	27186.76	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	27186.76	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	85622.31	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	85622.31	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	103069.19	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	103069.19	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	30401.17	239.25
PFOS_2	499.0 / 99.0	3.10	13C8-PFOS	507.0 / 99.0	30401.17	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	101281.21	250.00
PFDA_2	513.0 / 219.0	3.46	13C6-PFDA	519.0 / 474.0	101281.21	250.00
PFUnA_1	563.0 / 519.0	3.79	13C7-PFUnA	570.0 / 525.0	91776.59	250.00
PFUnA_2	563.0 / 269.0	3.79	13C7-PFUnA	570.0 / 525.0	91776.59	250.00
PFDoA_1	613.0 / 569.0	4.08	13C2-PFDoA	615.0 / 570.0	102072.61	250.00
PFDoA_2	613.0 / 319.0	4.08	13C2-PFDoA	615.0 / 570.0	102072.61	250.00
PFTrDA_1	663.0 / 619.0	4.33	13C2-PFTeDA	715.0 / 670.0	84788.29	250.00
PFTrDA_2	663.0 / 169.0	4.33	13C2-PFTeDA	715.0 / 670.0	84788.29	250.00
PFTeDA_1	713.0 / 669.0	4.55	13C2-PFTeDA	715.0 / 670.0	84788.29	250.00
PFTeDA_2	713.0 / 169.0	4.54	13C2-PFTeDA	715.0 / 670.0	84788.29	250.00
NMeFOSAA_1	570.0 / 419.0	3.62	d3-MeFOSAA	573.0 / 419.0	18223.77	250.00
NMeFOSAA_2	570.0 / 512.0	3.62	d3-MeFOSAA	573.0 / 419.0	18223.77	250.00
NEtFOSAA_1	584.0 / 419.0	3.79	d5-EtFOSAA	589.0 / 419.0	20489.99	250.00
NEtFOSAA_2	584.0 / 483.0	3.79	d5-EtFOSAA	589.0 / 419.0	20489.99	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	29660.00	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	29660.00	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	63134.84	250.00
PFHxA_2	313.0 / 119.0	1.88	13C5-PFHxA	318.0 / 273.0	63134.84	250.00
PFHpA_1	363.0 / 319.0	2.29	13C4-PFHpA	367.0 / 322.0	66285.64	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	66285.64	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	23705.87	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	23705.87	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	77573.52	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	77573.52	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	93994.33	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	93994.33	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	25276.93	239.25
PFOS_2	499.0 / 99.0	3.10	13C8-PFOS	507.0 / 99.0	25276.93	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	88095.93	250.00
PFDA_2	513.0 / 219.0	3.47	13C6-PFDA	519.0 / 474.0	88095.93	250.00
PFUnA_1	563.0 / 519.0	3.79	13C7-PFUnA	570.0 / 525.0	82754.96	250.00
PFUnA_2	563.0 / 269.0	3.79	13C7-PFUnA	570.0 / 525.0	82754.96	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	99995.01	250.00
PFDoA_2	613.0 / 319.0	4.07	13C2-PFDoA	615.0 / 570.0	99995.01	250.00
PFTrDA_1	663.0 / 619.0	4.32	13C2-PFTeDA	715.0 / 670.0	81617.66	250.00
PFTrDA_2	663.0 / 169.0	4.32	13C2-PFTeDA	715.0 / 670.0	81617.66	250.00
PFTeDA_1	713.0 / 669.0	4.54	13C2-PFTeDA	715.0 / 670.0	81617.66	250.00
PFTeDA_2	713.0 / 169.0	4.54	13C2-PFTeDA	715.0 / 670.0	81617.66	250.00
NMeFOSAA_1	570.0 / 419.0	3.62	d3-MeFOSAA	573.0 / 419.0	18782.71	250.00
NMeFOSAA_2	570.0 / 512.0	3.62	d3-MeFOSAA	573.0 / 419.0	18782.71	250.00
NEtFOSAA_1	584.0 / 419.0	3.78	d5-EtFOSAA	589.0 / 419.0	16682.63	250.00
NEtFOSAA_2	584.0 / 483.0	3.78	d5-EtFOSAA	589.0 / 419.0	16682.63	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.10	13C2-PFDA	515.0 / 470.0	100139.46	250.00
d3-MeFOSAA	573.0 / 419.0	3.65	13C4-PFOS	503.0 / 99.0	29846.55	239.25
d5-EtFOSAA	589.0 / 419.0	3.81	13C4-PFOS	503.0 / 99.0	29846.55	239.25
13C5-PFHxA	318.0 / 273.0	1.89	13C2-PFOA	415.0 / 370.0	79095.63	250.00
13C4-PFHpA	367.0 / 322.0	2.30	13C2-PFOA	415.0 / 370.0	79095.63	250.00
13C8-PFOA	421.0 / 376.0	2.72	13C2-PFOA	415.0 / 370.0	79095.63	250.00
13C9-PFNA	472.0 / 427.0	3.12	13C2-PFOA	415.0 / 370.0	79095.63	250.00
13C6-PFDA	519.0 / 474.0	3.48	13C2-PFDA	515.0 / 470.0	100139.46	250.00
13C7-PFUnA	570.0 / 525.0	3.81	13C2-PFDA	515.0 / 470.0	100139.46	250.00
13C2-PFTeDA	715.0 / 670.0	4.58	13C2-PFDA	515.0 / 470.0	100139.46	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	29846.55	239.25
13C3-PFHxS	402.0 / 99.0	2.33	13C4-PFOS	503.0 / 99.0	29846.55	239.25
13C8-PFOS	507.0 / 99.0	3.12	13C4-PFOS	503.0 / 99.0	29846.55	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.09	13C2-PFDA	515.0 / 470.0	103883.83	250.00
d3-MeFOSAA	573.0 / 419.0	3.64	13C4-PFOS	503.0 / 99.0	34856.85	239.25
d5-EtFOSAA	589.0 / 419.0	3.80	13C4-PFOS	503.0 / 99.0	34856.85	239.25
13C5-PFHxA	318.0 / 273.0	1.89	13C2-PFOA	415.0 / 370.0	89971.31	250.00
13C4-PFHpA	367.0 / 322.0	2.30	13C2-PFOA	415.0 / 370.0	89971.31	250.00
13C8-PFOA	421.0 / 376.0	2.72	13C2-PFOA	415.0 / 370.0	89971.31	250.00
13C9-PFNA	472.0 / 427.0	3.11	13C2-PFOA	415.0 / 370.0	89971.31	250.00
13C6-PFDA	519.0 / 474.0	3.48	13C2-PFDA	515.0 / 470.0	103883.83	250.00
13C7-PFUnA	570.0 / 525.0	3.80	13C2-PFDA	515.0 / 470.0	103883.83	250.00
13C2-PFTeDA	715.0 / 670.0	4.57	13C2-PFDA	515.0 / 470.0	103883.83	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	34856.85	239.25
13C3-PFHxS	402.0 / 99.0	2.32	13C4-PFOS	503.0 / 99.0	34856.85	239.25
13C8-PFOS	507.0 / 99.0	3.11	13C4-PFOS	503.0 / 99.0	34856.85	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.08	13C2-PFDA	515.0 / 470.0	98052.33	250.00
d3-MeFOSAA	573.0 / 419.0	3.63	13C4-PFOS	503.0 / 99.0	30684.77	239.25
d5-EtFOSAA	589.0 / 419.0	3.79	13C4-PFOS	503.0 / 99.0	30684.77	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	87799.30	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	87799.30	250.00
13C8-PFOA	421.0 / 376.0	2.71	13C2-PFOA	415.0 / 370.0	87799.30	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	87799.30	250.00
13C6-PFDA	519.0 / 474.0	3.47	13C2-PFDA	515.0 / 470.0	98052.33	250.00
13C7-PFUnA	570.0 / 525.0	3.79	13C2-PFDA	515.0 / 470.0	98052.33	250.00
13C2-PFTeDA	715.0 / 670.0	4.56	13C2-PFDA	515.0 / 470.0	98052.33	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	30684.77	239.25
13C3-PFHxS	402.0 / 99.0	2.32	13C4-PFOS	503.0 / 99.0	30684.77	239.25
13C8-PFOS	507.0 / 99.0	3.10	13C4-PFOS	503.0 / 99.0	30684.77	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 24/10/2018 1:13:33 PM

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.08	13C2-PFDA	515.0 / 470.0	99978.99	250.00
d3-MeFOSAA	573.0 / 419.0	3.62	13C4-PFOS	503.0 / 99.0	30962.75	239.25
d5-EtFOSAA	589.0 / 419.0	3.79	13C4-PFOS	503.0 / 99.0	30962.75	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	84567.91	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	84567.91	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	84567.91	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	84567.91	250.00
13C6-PFDA	519.0 / 474.0	3.46	13C2-PFDA	515.0 / 470.0	99978.99	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	99978.99	250.00
13C2-PFTeDA	715.0 / 670.0	4.55	13C2-PFDA	515.0 / 470.0	99978.99	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	30962.75	239.25
13C3-PFHxS	402.0 / 99.0	2.32	13C4-PFOS	503.0 / 99.0	30962.75	239.25
13C8-PFOS	507.0 / 99.0	3.10	13C4-PFOS	503.0 / 99.0	30962.75	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.07	13C2-PFDA	515.0 / 470.0	90129.70	250.00
d3-MeFOSAA	573.0 / 419.0	3.62	13C4-PFOS	503.0 / 99.0	31518.26	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	31518.26	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	80369.12	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	80369.12	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	80369.12	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	80369.12	250.00
13C6-PFDA	519.0 / 474.0	3.46	13C2-PFDA	515.0 / 470.0	90129.70	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	90129.70	250.00
13C2-PFTeDA	715.0 / 670.0	4.55	13C2-PFDA	515.0 / 470.0	90129.70	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	31518.26	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	31518.26	239.25
13C8-PFOS	507.0 / 99.0	3.10	13C4-PFOS	503.0 / 99.0	31518.26	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.07	13C2-PFDA	515.0 / 470.0	104169.70	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	29589.19	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	29589.19	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	85964.25	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	85964.25	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	85964.25	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	85964.25	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	104169.70	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	104169.70	250.00
13C2-PFTeDA	715.0 / 670.0	4.54	13C2-PFDA	515.0 / 470.0	104169.70	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	29589.19	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	29589.19	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	29589.19	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 24/10/2018 1:13:45 PM

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0610_18-0611_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.07	13C2-PFDA	515.0 / 470.0	100765.46	250.00
d3-MeFOSAA	573.0 / 419.0	3.62	13C4-PFOS	503.0 / 99.0	26984.04	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	26984.04	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	86636.81	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	86636.81	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	86636.81	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	86636.81	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	100765.46	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	100765.46	250.00
13C2-PFTeDA	715.0 / 670.0	4.54	13C2-PFDA	515.0 / 470.0	100765.46	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	26984.04	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	26984.04	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	26984.04	239.25

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.56	877.420602	1010.00	86.87
PFBS_2	298.9 / 99.0	1.56	838.659062	1010.00	83.04
PFHxA_1	313.0 / 269.0	1.89	1006.652936	1010.00	99.67
PFHxA_2	313.0 / 119.0	1.89	896.376115	1010.00	88.75
PFHpA_1	363.0 / 319.0	2.30	856.292979	1000.00	85.63
PFHpA_2	363.0 / 169.0	2.29	976.724688	1000.00	97.67
PFHxS_1	399.0 / 80.0	2.32	998.233289	1010.00	98.83
PFHxS_2	399.0 / 99.0	2.32	992.688608	1010.00	98.29
PFOA_1	413.0 / 369.0	2.71	930.879127	1000.00	93.09
PFOA_2	413.0 / 169.0	2.71	872.168076	1000.00	87.22
PFNA_1	463.0 / 419.0	3.11	1034.052055	1000.00	103.41
PFNA_2	463.0 / 219.0	3.10	1037.625087	1000.00	103.76
PFOS_1	499.0 / 80.0	3.10	906.626239	1000.00	90.66
PFOS_2	499.0 / 99.0	3.10	919.881718	1000.00	91.99
PFDA_1	513.0 / 469.0	3.46	983.979238	1000.00	98.40
PFDA_2	513.0 / 219.0	3.46	976.593938	1000.00	97.66
PFUnA_1	563.0 / 519.0	3.79	923.159866	1000.00	92.32
PFUnA_2	563.0 / 269.0	3.79	975.741234	1000.00	97.57
PFDoA_1	613.0 / 569.0	4.07	980.552902	1000.00	98.06
PFDoA_2	613.0 / 319.0	4.07	934.947666	1000.00	93.49
PFTrDA_1	663.0 / 619.0	4.32	1073.075185	1000.00	107.31
PFTrDA_2	663.0 / 169.0	4.32	1041.820014	1000.00	104.18
PFTeDA_1	713.0 / 669.0	4.54	1020.055697	1000.00	102.01
PFTeDA_2	713.0 / 169.0	4.54	1079.815842	1000.00	107.98
NMeFOSAA_1	570.0 / 419.0	3.62	945.129517	1000.00	94.51
NMeFOSAA_2	570.0 / 512.0	3.62	965.108458	1000.00	96.51
NEtFOSAA_1	584.0 / 419.0	3.78	1244.646852	1000.00	124.46
NEtFOSAA_2	584.0 / 483.0	3.78	1233.660110	1000.00	123.37

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.55	970.502253	1010.00	96.09
PFBS_2	298.9 / 99.0	1.55	975.811899	1010.00	96.62
PFHxA_1	313.0 / 269.0	1.87	1021.796315	1010.00	101.17
PFHxA_2	313.0 / 119.0	1.87	1040.344218	1010.00	103.00
PFHpA_1	363.0 / 319.0	2.28	1025.565325	1000.00	102.56
PFHpA_2	363.0 / 169.0	2.27	925.115481	1000.00	92.51
PFHxS_1	399.0 / 80.0	2.30	1028.011162	1010.00	101.78
PFHxS_2	399.0 / 99.0	2.30	1027.510384	1010.00	101.73
PFOA_1	413.0 / 369.0	2.69	1089.152356	1000.00	108.92
PFOA_2	413.0 / 169.0	2.69	972.981810	1000.00	97.30
PFNA_1	463.0 / 419.0	3.09	1096.274135	1000.00	109.63
PFNA_2	463.0 / 219.0	3.09	1176.854645	1000.00	117.69
PFOS_1	499.0 / 80.0	3.08	966.409606	1000.00	96.64
PFOS_2	499.0 / 99.0	3.08	1005.667125	1000.00	100.57
PFDA_1	513.0 / 469.0	3.44	930.784113	1000.00	93.08
PFDA_2	513.0 / 219.0	3.44	949.281973	1000.00	94.93
PFUnA_1	563.0 / 519.0	3.76	929.129524	1000.00	92.91
PFUnA_2	563.0 / 269.0	3.76	1022.431921	1000.00	102.24
PFDoA_1	613.0 / 569.0	4.04	1021.513275	1000.00	102.15
PFDoA_2	613.0 / 319.0	4.04	1009.419650	1000.00	100.94
PFTrDA_1	663.0 / 619.0	4.28	1040.929741	1000.00	104.09
PFTrDA_2	663.0 / 169.0	4.28	1132.866294	1000.00	113.29
PFTeDA_1	713.0 / 669.0	4.50	1052.743365	1000.00	105.27
PFTeDA_2	713.0 / 169.0	4.49	1055.013855	1000.00	105.50
NMeFOSAA_1	570.0 / 419.0	3.60	966.391560	1000.00	96.64
NMeFOSAA_2	570.0 / 512.0	3.59	1055.504283	1000.00	105.55
NEtFOSAA_1	584.0 / 419.0	3.76	891.166983	1000.00	89.12
NEtFOSAA_2	584.0 / 483.0	3.76	747.072688	1000.00	74.71

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.55	2369.216159	2525.00	93.83
PFBS_2	298.9 / 99.0	1.55	2325.560026	2525.00	92.10
PFHxA_1	313.0 / 269.0	1.87	2531.440864	2525.00	100.26
PFHxA_2	313.0 / 119.0	1.87	2480.277034	2525.00	98.23
PFHpA_1	363.0 / 319.0	2.28	2346.788435	2500.00	93.87
PFHpA_2	363.0 / 169.0	2.28	2132.608097	2500.00	85.30
PFHxS_1	399.0 / 80.0	2.30	2589.465547	2525.00	102.55
PFHxS_2	399.0 / 99.0	2.30	2612.549586	2525.00	103.47
PFOA_1	413.0 / 369.0	2.69	2443.383300	2500.00	97.74
PFOA_2	413.0 / 169.0	2.69	2302.848780	2500.00	92.11
PFNA_1	463.0 / 419.0	3.08	2693.385208	2500.00	107.74
PFNA_2	463.0 / 219.0	3.08	2781.799068	2500.00	111.27
PFOS_1	499.0 / 80.0	3.08	2538.244013	2500.00	101.53
PFOS_2	499.0 / 99.0	3.08	2528.818431	2500.00	101.15
PFDA_1	513.0 / 469.0	3.44	2445.623865	2500.00	97.82
PFDA_2	513.0 / 219.0	3.44	2561.411468	2500.00	102.46
PFUnA_1	563.0 / 519.0	3.76	2544.039408	2500.00	101.76
PFUnA_2	563.0 / 269.0	3.76	2658.216564	2500.00	106.33
PFDoA_1	613.0 / 569.0	4.04	2464.099799	2500.00	98.56
PFDoA_2	613.0 / 319.0	4.03	2509.000369	2500.00	100.36
PFTrDA_1	663.0 / 619.0	4.28	2597.444191	2500.00	103.90
PFTrDA_2	663.0 / 169.0	4.28	2688.889101	2500.00	107.56
PFTeDA_1	713.0 / 669.0	4.49	2618.964718	2500.00	104.76
PFTeDA_2	713.0 / 169.0	4.49	2653.683642	2500.00	106.15
NMeFOSAA_1	570.0 / 419.0	3.59	2743.692779	2500.00	109.75
NMeFOSAA_2	570.0 / 512.0	3.59	2724.156179	2500.00	108.97
NEtFOSAA_1	584.0 / 419.0	3.75	2101.345910	2500.00	84.05
NEtFOSAA_2	584.0 / 483.0	3.75	2090.305165	2500.00	83.61

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.55	937.856403	1010.00	92.86
PFBS_2	298.9 / 99.0	1.54	968.192887	1010.00	95.86
PFHxA_1	313.0 / 269.0	1.86	1086.523152	1010.00	107.58
PFHxA_2	313.0 / 119.0	1.86	1103.709203	1010.00	109.28
PFHpA_1	363.0 / 319.0	2.28	1039.170109	1000.00	103.92
PFHpA_2	363.0 / 169.0	2.27	1046.993654	1000.00	104.70
PFHxS_1	399.0 / 80.0	2.30	1007.725009	1010.00	99.77
PFHxS_2	399.0 / 99.0	2.30	1048.915738	1010.00	103.85
PFOA_1	413.0 / 369.0	2.68	1025.695194	1000.00	102.57
PFOA_2	413.0 / 169.0	2.68	1028.429117	1000.00	102.84
PFNA_1	463.0 / 419.0	3.08	1163.399637	1000.00	116.34
PFNA_2	463.0 / 219.0	3.08	1153.948123	1000.00	115.39
PFOS_1	499.0 / 80.0	3.08	943.683329	1000.00	94.37
PFOS_2	499.0 / 99.0	3.08	958.325620	1000.00	95.83
PFDA_1	513.0 / 469.0	3.43	933.785010	1000.00	93.38
PFDA_2	513.0 / 219.0	3.43	1063.890173	1000.00	106.39
PFUnA_1	563.0 / 519.0	3.75	993.537932	1000.00	99.35
PFUnA_2	563.0 / 269.0	3.75	1101.395376	1000.00	110.14
PFDoA_1	613.0 / 569.0	4.03	1007.158409	1000.00	100.72
PFDoA_2	613.0 / 319.0	4.03	996.597536	1000.00	99.66
PFTrDA_1	663.0 / 619.0	4.27	1018.173134	1000.00	101.82
PFTrDA_2	663.0 / 169.0	4.27	1059.125697	1000.00	105.91
PFTeDA_1	713.0 / 669.0	4.49	959.697415	1000.00	95.97
PFTeDA_2	713.0 / 169.0	4.48	1014.852413	1000.00	101.49
NMeFOSAA_1	570.0 / 419.0	3.59	1069.232180	1000.00	106.92
NMeFOSAA_2	570.0 / 512.0	3.59	1014.400975	1000.00	101.44
NEtFOSAA_1	584.0 / 419.0	3.74	1048.041519	1000.00	104.80
NEtFOSAA_2	584.0 / 483.0	3.74	1262.828513	1000.00	126.28

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.54	2554.341908	2525.00	101.16
PFBS_2	298.9 / 99.0	1.54	2694.424182	2525.00	106.71
PFHxA_1	313.0 / 269.0	1.87	2508.662068	2525.00	99.35
PFHxA_2	313.0 / 119.0	1.87	2367.348932	2525.00	93.76
PFHpA_1	363.0 / 319.0	2.28	2343.276633	2500.00	93.73
PFHpA_2	363.0 / 169.0	2.28	2190.024774	2500.00	87.60
PFHxS_1	399.0 / 80.0	2.30	2628.145020	2525.00	104.08
PFHxS_2	399.0 / 99.0	2.30	2766.086324	2525.00	109.55
PFOA_1	413.0 / 369.0	2.69	2465.981561	2500.00	98.64
PFOA_2	413.0 / 169.0	2.69	2279.640034	2500.00	91.19
PFNA_1	463.0 / 419.0	3.08	2588.359315	2500.00	103.53
PFNA_2	463.0 / 219.0	3.08	2582.453086	2500.00	103.30
PFOS_1	499.0 / 80.0	3.08	2252.922822	2500.00	90.12
PFOS_2	499.0 / 99.0	3.08	2267.070593	2500.00	90.68
PFDA_1	513.0 / 469.0	3.44	2397.385279	2500.00	95.90
PFDA_2	513.0 / 219.0	3.44	2370.547522	2500.00	94.82
PFUnA_1	563.0 / 519.0	3.76	2527.205775	2500.00	101.09
PFUnA_2	563.0 / 269.0	3.75	2606.459377	2500.00	104.26
PFDoA_1	613.0 / 569.0	4.04	2539.812648	2500.00	101.59
PFDoA_2	613.0 / 319.0	4.03	2623.183307	2500.00	104.93
PFTrDA_1	663.0 / 619.0	4.28	2587.646951	2500.00	103.51
PFTrDA_2	663.0 / 169.0	4.28	2553.838640	2500.00	102.15
PFTeDA_1	713.0 / 669.0	4.49	2547.298838	2500.00	101.89
PFTeDA_2	713.0 / 169.0	4.49	2632.693066	2500.00	105.31
NMeFOSAA_1	570.0 / 419.0	3.59	2581.273085	2500.00	103.25
NMeFOSAA_2	570.0 / 512.0	3.59	2825.539313	2500.00	113.02
NEtFOSAA_1	584.0 / 419.0	3.75	3021.344211	2500.00	120.85
NEtFOSAA_2	584.0 / 483.0	3.75	2764.388678	2500.00	110.58

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.54	1023.297764	1010.00	101.32
PFBS_2	298.9 / 99.0	1.54	1038.805130	1010.00	102.85
PFHxA_1	313.0 / 269.0	1.86	972.410253	1010.00	96.28
PFHxA_2	313.0 / 119.0	1.86	1008.383332	1010.00	99.84
PFHpA_1	363.0 / 319.0	2.27	1049.693219	1000.00	104.97
PFHpA_2	363.0 / 169.0	2.27	802.769767	1000.00	80.28
PFHxS_1	399.0 / 80.0	2.30	1065.296555	1010.00	105.47
PFHxS_2	399.0 / 99.0	2.30	1036.573130	1010.00	102.63
PFOA_1	413.0 / 369.0	2.68	1045.281242	1000.00	104.53
PFOA_2	413.0 / 169.0	2.68	1020.410103	1000.00	102.04
PFNA_1	463.0 / 419.0	3.08	1072.625459	1000.00	107.26
PFNA_2	463.0 / 219.0	3.08	1035.444285	1000.00	103.54
PFOS_1	499.0 / 80.0	3.08	1098.155304	1000.00	109.82
PFOS_2	499.0 / 99.0	3.08	1125.768088	1000.00	112.58
PFDA_1	513.0 / 469.0	3.43	1032.247251	1000.00	103.22
PFDA_2	513.0 / 219.0	3.43	1006.854350	1000.00	100.69
PFUnA_1	563.0 / 519.0	3.75	1105.835673	1000.00	110.58
PFUnA_2	563.0 / 269.0	3.75	1225.866868	1000.00	122.59
PFDoA_1	613.0 / 569.0	4.03	989.418445	1000.00	98.94
PFDoA_2	613.0 / 319.0	4.03	1031.563159	1000.00	103.16
PFTrDA_1	663.0 / 619.0	4.27	1075.257908	1000.00	107.53
PFTrDA_2	663.0 / 169.0	4.27	1124.290370	1000.00	112.43
PFTeDA_1	713.0 / 669.0	4.48	1018.432195	1000.00	101.84
PFTeDA_2	713.0 / 169.0	4.48	1039.803977	1000.00	103.98
NMeFOSAA_1	570.0 / 419.0	3.59	1086.267292	1000.00	108.63
NMeFOSAA_2	570.0 / 512.0	3.58	1135.498183	1000.00	113.55
NEtFOSAA_1	584.0 / 419.0	3.74	1051.066582	1000.00	105.11
NEtFOSAA_2	584.0 / 483.0	3.75	815.431703	1000.00	81.54

BATTELLE

QC Summary Report

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.56	530.901041	505.00	105.13
PFBS_2	298.9 / 99.0	1.56	534.776119	505.00	105.90
PFHxA_1	313.0 / 269.0	1.89	502.113179	505.00	99.43
PFHxA_2	313.0 / 119.0	1.89	482.200374	505.00	95.49
PFHpA_1	363.0 / 319.0	2.30	457.057610	500.00	91.41
PFHpA_2	363.0 / 169.0	2.30	378.164561	500.00	75.63
PFHxS_1	399.0 / 80.0	2.33	536.828284	505.00	106.30
PFHxS_2	399.0 / 99.0	2.33	554.677122	505.00	109.84
PFOA_1	413.0 / 369.0	2.72	450.142357	500.00	90.03
PFOA_2	413.0 / 169.0	2.72	512.989124	500.00	102.60
PFNA_1	463.0 / 419.0	3.12	496.359036	500.00	99.27
PFNA_2	463.0 / 219.0	3.12	517.431336	500.00	103.49
PFOS_1	499.0 / 80.0	3.12	447.719596	500.00	89.54
PFOS_2	499.0 / 99.0	3.12	466.506237	500.00	93.30
PFDA_1	513.0 / 469.0	3.48	468.705336	500.00	93.74
PFDA_2	513.0 / 219.0	3.48	371.360252	500.00	74.27
PFUnA_1	563.0 / 519.0	3.81	520.994498	500.00	104.20
PFUnA_2	563.0 / 269.0	3.81	561.839176	500.00	112.37
PFDoA_1	613.0 / 569.0	4.09	500.460376	500.00	100.09
PFDoA_2	613.0 / 319.0	4.09	510.989012	500.00	102.20
PFTrDA_1	663.0 / 619.0	4.34	471.768916	500.00	94.35
PFTrDA_2	663.0 / 169.0	4.34	489.753722	500.00	97.95
PFTeDA_1	713.0 / 669.0	4.56	491.209993	500.00	98.24
PFTeDA_2	713.0 / 169.0	4.55	490.024479	500.00	98.00
NMeFOSAA_1	570.0 / 419.0	3.63	567.846378	500.00	113.57
NMeFOSAA_2	570.0 / 512.0	3.64	454.660860	500.00	90.93
NEtFOSAA_1	584.0 / 419.0	3.80	527.932133	500.00	105.59
NEtFOSAA_2	584.0 / 483.0	3.80	443.589214	500.00	88.72

BATTELLE

QC Summary Report

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.55	1056.907488	1010.00	104.64
PFBS_2	298.9 / 99.0	1.55	1088.959526	1010.00	107.82
PFHxA_1	313.0 / 269.0	1.88	989.291134	1010.00	97.95
PFHxA_2	313.0 / 119.0	1.88	1052.356805	1010.00	104.19
PFHpA_1	363.0 / 319.0	2.29	958.465724	1000.00	95.85
PFHpA_2	363.0 / 169.0	2.29	913.087787	1000.00	91.31
PFHxS_1	399.0 / 80.0	2.31	1075.568492	1010.00	106.49
PFHxS_2	399.0 / 99.0	2.31	1052.112491	1010.00	104.17
PFOA_1	413.0 / 369.0	2.70	1001.747464	1000.00	100.17
PFOA_2	413.0 / 169.0	2.70	1016.828405	1000.00	101.68
PFNA_1	463.0 / 419.0	3.10	1026.735994	1000.00	102.67
PFNA_2	463.0 / 219.0	3.10	1100.617485	1000.00	110.06
PFOS_1	499.0 / 80.0	3.10	842.683724	1000.00	84.27
PFOS_2	499.0 / 99.0	3.10	891.919250	1000.00	89.19
PFDA_1	513.0 / 469.0	3.46	971.172969	1000.00	97.12
PFDA_2	513.0 / 219.0	3.46	930.773630	1000.00	93.08
PFUnA_1	563.0 / 519.0	3.78	951.925138	1000.00	95.19
PFUnA_2	563.0 / 269.0	3.78	978.637092	1000.00	97.86
PFDoA_1	613.0 / 569.0	4.06	971.698824	1000.00	97.17
PFDoA_2	613.0 / 319.0	4.06	1035.201696	1000.00	103.52
PFTrDA_1	663.0 / 619.0	4.30	1011.007598	1000.00	101.10
PFTrDA_2	663.0 / 169.0	4.30	1031.813406	1000.00	103.18
PFTeDA_1	713.0 / 669.0	4.52	1002.406141	1000.00	100.24
PFTeDA_2	713.0 / 169.0	4.52	1002.626566	1000.00	100.26
NMeFOSAA_1	570.0 / 419.0	3.61	1036.828058	1000.00	103.68
NMeFOSAA_2	570.0 / 512.0	3.61	976.810878	1000.00	97.68
NEtFOSAA_1	584.0 / 419.0	3.78	1114.947872	1000.00	111.49
NEtFOSAA_2	584.0 / 483.0	3.77	1055.074970	1000.00	105.51

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.06	231.611603	250.00	92.64
d3-MeFOSAA	573.0 / 419.0	3.61	244.423694	250.00	97.77
d5-EtFOSAA	589.0 / 419.0	3.78	176.181506	250.00	70.47
13C5-PFHxA	318.0 / 273.0	1.88	233.905404	250.00	93.56
13C4-PFHpA	367.0 / 322.0	2.29	244.843738	250.00	97.94
13C8-PFOA	421.0 / 376.0	2.70	255.222211	250.00	102.09
13C9-PFNA	472.0 / 427.0	3.09	253.115276	250.00	101.25
13C6-PFDA	519.0 / 474.0	3.45	223.796875	250.00	89.52
13C7-PFUnA	570.0 / 525.0	3.77	237.910809	250.00	95.16
13C2-PFTeDA	715.0 / 670.0	4.54	218.862217	250.00	87.54
13C3-PFBS	302.0 / 99.0	1.55	206.860264	232.25	89.07
13C3-PFHxS	402.0 / 99.0	2.31	210.154754	236.50	88.86
13C8-PFOS	507.0 / 99.0	3.09	222.736043	239.25	93.10

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.03	251.496245	250.00	100.60
d3-MeFOSAA	573.0 / 419.0	3.59	206.513415	250.00	82.61
d5-EtFOSAA	589.0 / 419.0	3.75	215.522596	250.00	86.21
13C5-PFHxA	318.0 / 273.0	1.86	243.853168	250.00	97.54
13C4-PFHpA	367.0 / 322.0	2.27	251.499403	250.00	100.60
13C8-PFOA	421.0 / 376.0	2.68	252.596966	250.00	101.04
13C9-PFNA	472.0 / 427.0	3.07	235.055860	250.00	94.02
13C6-PFDA	519.0 / 474.0	3.42	263.529523	250.00	105.41
13C7-PFUnA	570.0 / 525.0	3.74	264.759638	250.00	105.90
13C2-PFTeDA	715.0 / 670.0	4.49	274.563043	250.00	109.83
13C3-PFBS	302.0 / 99.0	1.53	208.968650	232.25	89.98
13C3-PFHxS	402.0 / 99.0	2.29	203.875863	236.50	86.21
13C8-PFOS	507.0 / 99.0	3.07	211.103660	239.25	88.24

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.03	256.730088	250.00	102.69
d3-MeFOSAA	573.0 / 419.0	3.59	222.354343	250.00	88.94
d5-EtFOSAA	589.0 / 419.0	3.74	250.424424	250.00	100.17
13C5-PFHxA	318.0 / 273.0	1.85	233.925126	250.00	93.57
13C4-PFHpA	367.0 / 322.0	2.27	236.268478	250.00	94.51
13C8-PFOA	421.0 / 376.0	2.68	258.188550	250.00	103.28
13C9-PFNA	472.0 / 427.0	3.07	224.830199	250.00	89.93
13C6-PFDA	519.0 / 474.0	3.42	237.785773	250.00	95.11
13C7-PFUnA	570.0 / 525.0	3.74	255.936822	250.00	102.37
13C2-PFTeDA	715.0 / 670.0	4.48	275.760706	250.00	110.30
13C3-PFBS	302.0 / 99.0	1.53	248.039058	232.25	106.80
13C3-PFHxS	402.0 / 99.0	2.29	227.456951	236.50	96.18
13C8-PFOS	507.0 / 99.0	3.07	223.947248	239.25	93.60

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.02	252.058922	250.00	100.82
d3-MeFOSAA	573.0 / 419.0	3.58	221.573454	250.00	88.63
d5-EtFOSAA	589.0 / 419.0	3.74	202.252273	250.00	80.90
13C5-PFHxA	318.0 / 273.0	1.85	252.147535	250.00	100.86
13C4-PFHpA	367.0 / 322.0	2.27	245.071359	250.00	98.03
13C8-PFOA	421.0 / 376.0	2.67	268.592304	250.00	107.44
13C9-PFNA	472.0 / 427.0	3.06	228.680180	250.00	91.47
13C6-PFDA	519.0 / 474.0	3.42	255.818078	250.00	102.33
13C7-PFUnA	570.0 / 525.0	3.74	258.228174	250.00	103.29
13C2-PFTeDA	715.0 / 670.0	4.48	279.224041	250.00	111.69
13C3-PFBS	302.0 / 99.0	1.53	243.217823	232.25	104.72
13C3-PFHxS	402.0 / 99.0	2.29	237.294654	236.50	100.34
13C8-PFOS	507.0 / 99.0	3.06	243.456749	239.25	101.76

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.03	258.457777	250.00	103.38
d3-MeFOSAA	573.0 / 419.0	3.58	227.146712	250.00	90.86
d5-EtFOSAA	589.0 / 419.0	3.74	195.584216	250.00	78.23
13C5-PFHxA	318.0 / 273.0	1.85	244.075531	250.00	97.63
13C4-PFHpA	367.0 / 322.0	2.27	244.112187	250.00	97.64
13C8-PFOA	421.0 / 376.0	2.68	253.044813	250.00	101.22
13C9-PFNA	472.0 / 427.0	3.07	228.128348	250.00	91.25
13C6-PFDA	519.0 / 474.0	3.42	255.081498	250.00	102.03
13C7-PFUnA	570.0 / 525.0	3.74	251.783994	250.00	100.71
13C2-PFTeDA	715.0 / 670.0	4.48	276.141791	250.00	110.46
13C3-PFBS	302.0 / 99.0	1.53	233.953222	232.25	100.73
13C3-PFHxS	402.0 / 99.0	2.29	224.876965	236.50	95.09
13C8-PFOS	507.0 / 99.0	3.07	247.797379	239.25	103.57

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.02	253.149534	250.00	101.26
d3-MeFOSAA	573.0 / 419.0	3.58	250.939491	250.00	100.38
d5-EtFOSAA	589.0 / 419.0	3.73	247.294781	250.00	98.92
13C5-PFHxA	318.0 / 273.0	1.85	279.995422	250.00	112.00
13C4-PFHpA	367.0 / 322.0	2.26	257.493253	250.00	103.00
13C8-PFOA	421.0 / 376.0	2.67	273.722234	250.00	109.49
13C9-PFNA	472.0 / 427.0	3.06	242.330211	250.00	96.93
13C6-PFDA	519.0 / 474.0	3.42	235.027234	250.00	94.01
13C7-PFUnA	570.0 / 525.0	3.73	230.811093	250.00	92.32
13C2-PFTeDA	715.0 / 670.0	4.48	261.993503	250.00	104.80
13C3-PFBS	302.0 / 99.0	1.53	300.659096	232.25	129.45
13C3-PFHxS	402.0 / 99.0	2.28	249.197160	236.50	105.37
13C8-PFOS	507.0 / 99.0	3.06	240.290397	239.25	100.43

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.08	265.065125	250.00	106.03
d3-MeFOSAA	573.0 / 419.0	3.63	235.936014	250.00	94.37
d5-EtFOSAA	589.0 / 419.0	3.79	236.159897	250.00	94.46
13C5-PFHxA	318.0 / 273.0	1.88	270.928650	250.00	108.37
13C4-PFHpA	367.0 / 322.0	2.29	274.081656	250.00	109.63
13C8-PFOA	421.0 / 376.0	2.71	262.586014	250.00	105.03
13C9-PFNA	472.0 / 427.0	3.11	235.605943	250.00	94.24
13C6-PFDA	519.0 / 474.0	3.47	258.018855	250.00	103.21
13C7-PFUnA	570.0 / 525.0	3.79	242.404039	250.00	96.96
13C2-PFTeDA	715.0 / 670.0	4.55	294.609831	250.00	117.84
13C3-PFBS	302.0 / 99.0	1.54	217.345232	232.25	93.58
13C3-PFHxS	402.0 / 99.0	2.32	213.712943	236.50	90.36
13C8-PFOS	507.0 / 99.0	3.10	241.527136	239.25	100.95

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFDoA	615.0 / 570.0	4.05	291.232104	250.00	116.49
d3-MeFOSAA	573.0 / 419.0	3.61	241.106040	250.00	96.44
d5-EtFOSAA	589.0 / 419.0	3.77	216.895753	250.00	86.76
13C5-PFHxA	318.0 / 273.0	1.86	259.302943	250.00	103.72
13C4-PFHpA	367.0 / 322.0	2.28	259.486408	250.00	103.79
13C8-PFOA	421.0 / 376.0	2.69	241.060106	250.00	96.42
13C9-PFNA	472.0 / 427.0	3.09	236.389690	250.00	94.56
13C6-PFDA	519.0 / 474.0	3.45	263.299625	250.00	105.32
13C7-PFUnA	570.0 / 525.0	3.76	279.076498	250.00	111.63
13C2-PFTeDA	715.0 / 670.0	4.51	312.016305	250.00	124.81
13C3-PFBS	302.0 / 99.0	1.53	220.539729	232.25	94.96
13C3-PFHxS	402.0 / 99.0	2.30	215.830304	236.50	91.26
13C8-PFOS	507.0 / 99.0	3.08	261.103354	239.25	109.13

BATTELLE

Summary Quant Report

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	3.00e5	877.420602	544.9	false
PFBS_2	298.9 / 99.0	1.56	8.39e4	838.659062	365.9	false
PFHxA_1	313.0 / 269.0	1.89	2.42e5	1006.652936	19.8	false
PFHxA_2	313.0 / 119.0	1.89	1.64e4	896.376115	16.0	false
PFHpA_1	363.0 / 319.0	2.30	2.16e5	856.292979	126.4	false
PFHpA_2	363.0 / 169.0	2.29	5.56e3	976.724688	99.7	false
PFHxS_1	399.0 / 80.0	2.32	3.62e5	998.233289	454.9	false
PFHxS_2	399.0 / 99.0	2.32	1.01e5	992.688608	416.4	false
PFOA_1	413.0 / 369.0	2.71	3.38e5	930.879127	304.4	false
PFOA_2	413.0 / 169.0	2.71	2.06e4	872.168076	223.7	false
PFNA_1	463.0 / 419.0	3.11	3.57e5	1034.052055	304.0	false
PFNA_2	463.0 / 219.0	3.10	1.10e5	1037.625087	466.5	false
PFOS_1	499.0 / 80.0	3.10	5.40e5	906.626239	343.1	false
PFOS_2	499.0 / 99.0	3.10	9.52e4	919.881718	450.1	false
PFDA_1	513.0 / 469.0	3.46	3.80e5	983.979238	327.0	false
PFDA_2	513.0 / 219.0	3.46	1.56e4	976.593938	361.3	false
PFUnA_1	563.0 / 519.0	3.79	3.50e5	923.159866	376.6	false
PFUnA_2	563.0 / 269.0	3.79	1.82e4	975.741234	188.3	false
PFDoA_1	613.0 / 569.0	4.07	3.41e5	980.552902	460.6	false
PFDoA_2	613.0 / 319.0	4.07	5.13e4	934.947666	412.2	false
PFTrDA_1	663.0 / 619.0	4.32	3.01e5	1073.075185	751.7	false
PFTrDA_2	663.0 / 169.0	4.32	1.93e4	1041.820014	533.2	false
PFTeDA_1	713.0 / 669.0	4.54	3.27e5	1020.055697	1259.1	false
PFTeDA_2	713.0 / 169.0	4.54	1.68e4	1079.815842	804.1	false
NMeFOSAA_1	570.0 / 419.0	3.62	6.69e4	945.129517	505.3	false
NMeFOSAA_2	570.0 / 512.0	3.62	3.86e4	965.108458	1105.4	false
NEtFOSAA_1	584.0 / 419.0	3.78	6.69e4	1244.646852	673.7	true
NEtFOSAA_2	584.0 / 483.0	3.78	4.01e3	1233.660110	1226.3	false

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	3.09e5	970.502253	411.5	false
PFBS_2	298.9 / 99.0	1.55	9.09e4	975.811899	376.8	false
PFHxA_1	313.0 / 269.0	1.87	2.21e5	1021.796315	53.4	false
PFHxA_2	313.0 / 119.0	1.87	1.70e4	1040.344218	47.1	false
PFHpA_1	363.0 / 319.0	2.28	2.28e5	1025.565325	162.7	false
PFHpA_2	363.0 / 169.0	2.27	4.69e3	925.115481	120.2	false
PFHxS_1	399.0 / 80.0	2.30	3.34e5	1028.011162	412.3	false
PFHxS_2	399.0 / 99.0	2.30	9.34e4	1027.510384	591.9	false
PFOA_1	413.0 / 369.0	2.69	3.37e5	1089.152356	347.8	false
PFOA_2	413.0 / 169.0	2.69	1.96e4	972.981810	214.5	false
PFNA_1	463.0 / 419.0	3.09	3.02e5	1096.274135	338.6	false
PFNA_2	463.0 / 219.0	3.09	9.96e4	1176.854645	486.1	false
PFOS_1	499.0 / 80.0	3.08	5.14e5	966.409606	205.8	false
PFOS_2	499.0 / 99.0	3.08	9.30e4	1005.667125	445.2	false
PFDA_1	513.0 / 469.0	3.44	3.77e5	930.784113	382.9	false
PFDA_2	513.0 / 219.0	3.44	1.59e4	949.281973	243.4	false
PFUnA_1	563.0 / 519.0	3.76	3.48e5	929.129524	312.6	false
PFUnA_2	563.0 / 269.0	3.76	1.89e4	1022.431921	223.3	false
PFDoA_1	613.0 / 569.0	4.04	3.43e5	1021.513275	488.7	false
PFDoA_2	613.0 / 319.0	4.04	5.33e4	1009.419650	394.7	false
PFTrDA_1	663.0 / 619.0	4.28	3.26e5	1040.929741	681.7	false
PFTrDA_2	663.0 / 169.0	4.28	2.33e4	1132.866294	418.1	false
PFTeDA_1	713.0 / 669.0	4.50	3.76e5	1052.743365	1384.7	false
PFTeDA_2	713.0 / 169.0	4.49	1.83e4	1055.013855	610.0	false
NMeFOSAA_1	570.0 / 419.0	3.60	5.30e4	966.391560	582.1	false
NMeFOSAA_2	570.0 / 512.0	3.59	3.24e4	1055.504283	440.3	false
NEtFOSAA_1	584.0 / 419.0	3.76	5.33e4	891.166983	869.0	false
NEtFOSAA_2	584.0 / 483.0	3.76	2.78e3	747.072688	249.5	false

BATTELLE

Summary Quant Report

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	7.40e5	2369.216159	815.3	false
PFBS_2	298.9 / 99.0	1.55	2.13e5	2325.560026	500.2	false
PFHxA_1	313.0 / 269.0	1.87	5.20e5	2531.440864	93.0	false
PFHxA_2	313.0 / 119.0	1.87	3.82e4	2480.277034	64.3	false
PFHpA_1	363.0 / 319.0	2.28	4.91e5	2346.788435	266.5	false
PFHpA_2	363.0 / 169.0	2.28	9.75e3	2132.608097	212.0	false
PFHxS_1	399.0 / 80.0	2.30	7.72e5	2589.465547	664.3	false
PFHxS_2	399.0 / 99.0	2.30	2.17e5	2612.549586	762.5	false
PFOA_1	413.0 / 369.0	2.69	7.75e5	2443.383300	471.5	false
PFOA_2	413.0 / 169.0	2.69	4.75e4	2302.848780	373.2	false
PFNA_1	463.0 / 419.0	3.08	7.06e5	2693.385208	484.8	false
PFNA_2	463.0 / 219.0	3.08	2.25e5	2781.799068	637.2	false
PFOS_1	499.0 / 80.0	3.08	1.19e6	2538.244013	394.7	false
PFOS_2	499.0 / 99.0	3.08	2.05e5	2528.818431	768.5	false
PFDA_1	513.0 / 469.0	3.44	8.20e5	2445.623865	624.6	false
PFDA_2	513.0 / 219.0	3.44	3.54e4	2561.411468	432.2	false
PFUnA_1	563.0 / 519.0	3.76	8.52e5	2544.039408	454.2	false
PFUnA_2	563.0 / 269.0	3.76	4.41e4	2658.216564	385.7	false
PFDoA_1	613.0 / 569.0	4.04	7.73e5	2464.099799	632.2	false
PFDoA_2	613.0 / 319.0	4.03	1.23e5	2509.000369	520.9	false
PFTrDA_1	663.0 / 619.0	4.28	7.45e5	2597.444191	930.7	false
PFTrDA_2	663.0 / 169.0	4.28	5.07e4	2688.889101	532.8	false
PFTeDA_1	713.0 / 669.0	4.49	8.52e5	2618.964718	1896.3	false
PFTeDA_2	713.0 / 169.0	4.49	4.19e4	2653.683642	857.6	false
NMeFOSAA_1	570.0 / 419.0	3.59	1.32e5	2743.692779	748.2	false
NMeFOSAA_2	570.0 / 512.0	3.59	7.17e4	2724.156179	1383.8	false
NEtFOSAA_1	584.0 / 419.0	3.75	1.18e5	2101.345910	1094.2	false
NEtFOSAA_2	584.0 / 483.0	3.75	6.99e3	2090.305165	435.4	false

BATTELLE

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	3.12e5	937.856403	445.8	false
PFBS_2	298.9 / 99.0	1.54	9.44e4	968.192887	379.6	false
PFHxA_1	313.0 / 269.0	1.86	2.45e5	1086.523152	54.4	false
PFHxA_2	313.0 / 119.0	1.86	1.88e4	1103.709203	45.6	false
PFHpA_1	363.0 / 319.0	2.28	2.28e5	1039.170109	179.8	false
PFHpA_2	363.0 / 169.0	2.27	5.17e3	1046.993654	139.4	false
PFHxS_1	399.0 / 80.0	2.30	3.40e5	1007.725009	519.1	false
PFHxS_2	399.0 / 99.0	2.30	9.92e4	1048.915738	590.0	false
PFOA_1	413.0 / 369.0	2.68	3.41e5	1025.695194	247.7	false
PFOA_2	413.0 / 169.0	2.68	2.22e4	1028.429117	218.9	false
PFNA_1	463.0 / 419.0	3.08	3.15e5	1163.399637	311.5	false
PFNA_2	463.0 / 219.0	3.08	9.61e4	1153.948123	381.7	false
PFOS_1	499.0 / 80.0	3.08	5.10e5	943.683329	314.5	false
PFOS_2	499.0 / 99.0	3.08	9.01e4	958.325620	659.2	false
PFDA_1	513.0 / 469.0	3.43	3.67e5	933.785010	395.1	false
PFDA_2	513.0 / 219.0	3.43	1.72e4	1063.890173	321.4	false
PFUnA_1	563.0 / 519.0	3.75	3.63e5	993.537932	343.3	false
PFUnA_2	563.0 / 269.0	3.75	1.98e4	1101.395376	241.0	false
PFDoA_1	613.0 / 569.0	4.03	3.39e5	1007.158409	354.5	false
PFDoA_2	613.0 / 319.0	4.03	5.28e4	996.597536	358.5	false
PFTrDA_1	663.0 / 619.0	4.27	3.24e5	1018.173134	654.0	false
PFTrDA_2	663.0 / 169.0	4.27	2.22e4	1059.125697	406.6	false
PFTeDA_1	713.0 / 669.0	4.49	3.49e5	959.697415	1330.5	false
PFTeDA_2	713.0 / 169.0	4.48	1.79e4	1014.852413	544.6	false
NMeFOSAA_1	570.0 / 419.0	3.59	5.65e4	1069.232180	828.5	false
NMeFOSAA_2	570.0 / 512.0	3.59	3.03e4	1014.400975	1094.0	false
NEtFOSAA_1	584.0 / 419.0	3.74	5.35e4	1048.041519	525.4	false
NEtFOSAA_2	584.0 / 483.0	3.74	3.88e3	1262.828513	998.5	false

BATTELLE

Summary Quant Report

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	8.25e5	2554.341908	706.9	false
PFBS_2	298.9 / 99.0	1.54	2.55e5	2694.424182	618.9	false
PFHxA_1	313.0 / 269.0	1.87	5.98e5	2508.662068	92.7	false
PFHxA_2	313.0 / 119.0	1.87	4.23e4	2367.348932	71.3	false
PFHpA_1	363.0 / 319.0	2.28	5.63e5	2343.276633	262.5	false
PFHpA_2	363.0 / 169.0	2.28	1.15e4	2190.024774	158.4	false
PFHxS_1	399.0 / 80.0	2.30	8.47e5	2628.145020	753.7	false
PFHxS_2	399.0 / 99.0	2.30	2.49e5	2766.086324	881.7	false
PFOA_1	413.0 / 369.0	2.69	8.52e5	2465.981561	436.8	false
PFOA_2	413.0 / 169.0	2.69	5.12e4	2279.640034	337.7	false
PFNA_1	463.0 / 419.0	3.08	7.65e5	2588.359315	515.0	false
PFNA_2	463.0 / 219.0	3.08	2.35e5	2582.453086	613.2	false
PFOS_1	499.0 / 80.0	3.08	1.28e6	2252.922822	381.6	false
PFOS_2	499.0 / 99.0	3.08	2.24e5	2267.070593	567.3	false
PFDA_1	513.0 / 469.0	3.44	9.46e5	2397.385279	580.3	false
PFDA_2	513.0 / 219.0	3.44	3.86e4	2370.547522	471.4	false
PFUnA_1	563.0 / 519.0	3.76	9.13e5	2527.205775	467.2	false
PFUnA_2	563.0 / 269.0	3.75	4.67e4	2606.459377	326.8	false
PFDoA_1	613.0 / 569.0	4.04	8.79e5	2539.812648	621.1	false
PFDoA_2	613.0 / 319.0	4.03	1.42e5	2623.183307	619.9	false
PFTrDA_1	663.0 / 619.0	4.28	8.15e5	2587.646951	951.9	false
PFTrDA_2	663.0 / 169.0	4.28	5.29e4	2553.838640	475.7	false
PFTeDA_1	713.0 / 669.0	4.49	9.11e5	2547.298838	1805.6	false
PFTeDA_2	713.0 / 169.0	4.49	4.56e4	2632.693066	1089.8	false
NMeFOSAA_1	570.0 / 419.0	3.59	1.38e5	2581.273085	899.4	false
NMeFOSAA_2	570.0 / 512.0	3.59	8.25e4	2825.539313	2378.0	false
NEtFOSAA_1	584.0 / 419.0	3.75	1.47e5	3021.344211	765.6	false
NEtFOSAA_2	584.0 / 483.0	3.75	7.98e3	2764.388678	404.7	false

BATTELLE

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	2.41e5	1023.297764	247.7	false
PFBS_2	298.9 / 99.0	1.54	7.17e4	1038.805130	258.3	false
PFHxA_1	313.0 / 269.0	1.86	1.56e5	972.410253	28.0	false
PFHxA_2	313.0 / 119.0	1.86	1.22e4	1008.383332	29.3	false
PFHpA_1	363.0 / 319.0	2.27	1.54e5	1049.693219	127.2	false
PFHpA_2	363.0 / 169.0	2.27	2.72e3	802.769767	65.9	false
PFHxS_1	399.0 / 80.0	2.30	2.15e5	1065.296555	365.1	false
PFHxS_2	399.0 / 99.0	2.30	5.85e4	1036.573130	375.2	false
PFOA_1	413.0 / 369.0	2.68	2.26e5	1045.281242	178.4	false
PFOA_2	413.0 / 169.0	2.68	1.43e4	1020.410103	162.7	false
PFNA_1	463.0 / 419.0	3.08	1.97e5	1072.625459	219.2	false
PFNA_2	463.0 / 219.0	3.08	5.85e4	1035.444285	268.6	false
PFOS_1	499.0 / 80.0	3.08	3.42e5	1098.155304	195.0	false
PFOS_2	499.0 / 99.0	3.08	6.09e4	1125.768088	336.5	false
PFDA_1	513.0 / 469.0	3.43	2.45e5	1032.247251	316.2	false
PFDA_2	513.0 / 219.0	3.43	9.87e3	1006.854350	210.6	false
PFUnA_1	563.0 / 519.0	3.75	2.37e5	1105.835673	316.4	false
PFUnA_2	563.0 / 269.0	3.75	1.30e4	1225.866868	199.6	true
PFDoA_1	613.0 / 569.0	4.03	2.20e5	989.418445	350.0	false
PFDoA_2	613.0 / 319.0	4.03	3.61e4	1031.563159	288.4	false
PFTrDA_1	663.0 / 619.0	4.27	2.11e5	1075.257908	539.9	false
PFTrDA_2	663.0 / 169.0	4.27	1.45e4	1124.290370	354.8	false
PFTeDA_1	713.0 / 669.0	4.48	2.28e5	1018.432195	1059.8	false
PFTeDA_2	713.0 / 169.0	4.48	1.13e4	1039.803977	442.4	false
NMeFOSAA_1	570.0 / 419.0	3.59	3.61e4	1086.267292	505.0	false
NMeFOSAA_2	570.0 / 512.0	3.58	2.12e4	1135.498183	856.9	false
NEtFOSAA_1	584.0 / 419.0	3.74	3.67e4	1051.066582	1098.2	false
NEtFOSAA_2	584.0 / 483.0	3.75	1.76e3	815.431703	79.0	false

BATTELLE

Summary Quant Report

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	172617.74	530.901041	259.6	false
PFBS_2	298.9 / 99.0	1.56	50885.76	534.776119	224.8	false
PFHxA_1	313.0 / 269.0	1.89	131873.89	502.113179	40.0	false
PFHxA_2	313.0 / 119.0	1.89	9769.12	482.200374	29.9	false
PFHpA_1	363.0 / 319.0	2.30	120129.47	457.057610	109.7	false
PFHpA_2	363.0 / 169.0	2.30	2494.87	378.164561	61.5	false
PFHxS_1	399.0 / 80.0	2.33	182635.60	536.828284	221.1	false
PFHxS_2	399.0 / 99.0	2.33	53027.45	554.677122	282.2	false
PFOA_1	413.0 / 369.0	2.72	156343.01	450.142357	177.5	false
PFOA_2	413.0 / 169.0	2.72	11508.26	512.989124	159.4	false
PFNA_1	463.0 / 419.0	3.12	150463.14	496.359036	175.7	false
PFNA_2	463.0 / 219.0	3.12	47949.01	517.431336	198.4	false
PFOS_1	499.0 / 80.0	3.12	259389.10	447.719596	173.9	false
PFOS_2	499.0 / 99.0	3.12	47068.98	466.506237	323.4	false
PFDA_1	513.0 / 469.0	3.48	209565.47	468.705336	279.0	false
PFDA_2	513.0 / 219.0	3.48	6958.61	371.360252	138.2	false
PFUnA_1	563.0 / 519.0	3.81	198730.28	520.994498	223.1	false
PFUnA_2	563.0 / 269.0	3.81	10465.75	561.839176	132.5	false
PFDoA_1	613.0 / 569.0	4.09	200825.40	500.460376	326.2	false
PFDoA_2	613.0 / 319.0	4.09	32543.91	510.989012	215.5	false
PFTrDA_1	663.0 / 619.0	4.34	181146.11	471.768916	424.3	false
PFTrDA_2	663.0 / 169.0	4.34	12409.68	489.753722	237.5	false
PFTeDA_1	713.0 / 669.0	4.56	216434.22	491.209993	994.5	false
PFTeDA_2	713.0 / 169.0	4.55	10556.87	490.024479	428.3	false
NMeFOSAA_1	570.0 / 419.0	3.63	36220.02	567.846378	440.0	false
NMeFOSAA_2	570.0 / 512.0	3.64	17565.88	454.660860	291.1	false
NEtFOSAA_1	584.0 / 419.0	3.80	35787.03	527.932133	513.4	false
NEtFOSAA_2	584.0 / 483.0	3.80	1947.57	443.589214	183.3	false

BATTELLE

Summary Quant Report

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	338042.04	1056.907488	321.1	false
PFBS_2	298.9 / 99.0	1.55	102098.19	1088.959526	324.4	false
PFHxA_1	313.0 / 269.0	1.88	254265.01	989.291134	52.9	false
PFHxA_2	313.0 / 119.0	1.88	20419.73	1052.356805	45.8	false
PFHpA_1	363.0 / 319.0	2.29	246361.81	958.465724	172.9	false
PFHpA_2	363.0 / 169.0	2.29	5343.79	913.087787	110.7	false
PFHxS_1	399.0 / 80.0	2.31	350967.44	1075.568492	363.9	false
PFHxS_2	399.0 / 99.0	2.31	96220.85	1052.112491	356.8	false
PFOA_1	413.0 / 369.0	2.70	331016.67	1001.747464	317.5	false
PFOA_2	413.0 / 169.0	2.70	21799.64	1016.828405	215.6	false
PFNA_1	463.0 / 419.0	3.10	319004.34	1026.735994	333.7	false
PFNA_2	463.0 / 219.0	3.10	104960.31	1100.617485	384.9	false
PFOS_1	499.0 / 80.0	3.10	522248.71	842.683724	234.7	false
PFOS_2	499.0 / 99.0	3.10	96122.75	891.919250	470.3	false
PFDA_1	513.0 / 469.0	3.46	396789.70	971.172969	467.2	false
PFDA_2	513.0 / 219.0	3.46	15717.49	930.773630	224.2	false
PFUnA_1	563.0 / 519.0	3.78	379820.86	951.925138	329.8	false
PFUnA_2	563.0 / 269.0	3.78	19235.11	978.637092	205.6	false
PFDoA_1	613.0 / 569.0	4.06	382081.51	971.698824	431.6	false
PFDoA_2	613.0 / 319.0	4.06	63895.81	1035.201696	314.0	false
PFTrDA_1	663.0 / 619.0	4.30	363283.78	1011.007598	656.2	false
PFTrDA_2	663.0 / 169.0	4.30	24421.20	1031.813406	312.5	false
PFTeDA_1	713.0 / 669.0	4.52	411209.17	1002.406141	1371.4	false
PFTeDA_2	713.0 / 169.0	4.52	20004.39	1002.626566	494.9	false
NMeFOSAA_1	570.0 / 419.0	3.61	64127.47	1036.828058	766.6	false
NMeFOSAA_2	570.0 / 512.0	3.61	34260.23	976.810878	461.4	false
NEtFOSAA_1	584.0 / 419.0	3.78	65588.86	1114.947872	645.7	false
NEtFOSAA_2	584.0 / 483.0	3.77	3778.43	1055.074970	601.8	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:34:09 PM

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.06	93406.68	231.611603	1689.5	false
d3-MeFOSAA	573.0 / 419.0	3.61	17579.80	244.423694	175.3	false
d5-EtFOSAA	589.0 / 419.0	3.78	14383.26	176.181506	200.6	false
13C5-PFHxA	318.0 / 273.0	1.88	59013.59	233.905404	685.2	false
13C4-PFHpA	367.0 / 322.0	2.29	70279.06	244.843738	1118.6	false
13C8-PFOA	421.0 / 376.0	2.70	89115.92	255.222211	9011.3	false
13C9-PFNA	472.0 / 427.0	3.09	100916.15	253.115276	1888.1	false
13C6-PFDA	519.0 / 474.0	3.45	90930.64	223.796875	1073.3	false
13C7-PFUnA	570.0 / 525.0	3.77	89293.12	237.910809	829.4	false
13C2-PFTeDA	715.0 / 670.0	4.54	71262.97	218.862217	1767.1	false
13C3-PFBS	302.0 / 99.0	1.55	27104.45	206.860264	624.1	false
13C3-PFHxS	402.0 / 99.0	2.31	24951.95	210.154754	577.2	false
13C8-PFOS	507.0 / 99.0	3.09	29875.11	222.736043	244.5	false

BATTELLE

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	90196.83	251.496245	868.5	false
d3-MeFOSAA	573.0 / 419.0	3.59	13699.36	206.513415	220.4	false
d5-EtFOSAA	589.0 / 419.0	3.75	16228.23	215.522596	229.9	false
13C5-PFHxA	318.0 / 273.0	1.86	53053.80	243.853168	883.4	false
13C4-PFHpA	367.0 / 322.0	2.27	62251.55	251.499403	822.5	false
13C8-PFOA	421.0 / 376.0	2.68	76057.36	252.596966	4120.8	false
13C9-PFNA	472.0 / 427.0	3.07	80814.57	235.055860	883.7	false
13C6-PFDA	519.0 / 474.0	3.42	95219.88	263.529523	914.8	false
13C7-PFUnA	570.0 / 525.0	3.74	88368.55	264.759638	626.1	false
13C2-PFTeDA	715.0 / 670.0	4.49	79501.87	274.563043	1411.1	false
13C3-PFBS	302.0 / 99.0	1.53	25253.76	208.968650	588.5	false
13C3-PFHxS	402.0 / 99.0	2.29	22326.08	203.875863	301.6	false
13C8-PFOS	507.0 / 99.0	3.07	26115.36	211.103660	229.0	false

BATTELLE

Summary Quant Report

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	85629.59	256.730088	1327.9	false
d3-MeFOSAA	573.0 / 419.0	3.59	12250.10	222.354343	156.9	false
d5-EtFOSAA	589.0 / 419.0	3.74	15660.19	250.424424	343.0	false
13C5-PFHxA	318.0 / 273.0	1.85	51469.81	233.925126	802.8	false
13C4-PFHpA	367.0 / 322.0	2.27	59143.44	236.268478	587.6	false
13C8-PFOA	421.0 / 376.0	2.68	78620.84	258.188550	1250.9	false
13C9-PFNA	472.0 / 427.0	3.07	78173.73	224.830199	1167.3	false
13C6-PFDA	519.0 / 474.0	3.42	79904.56	237.785773	22007.2	false
13C7-PFUnA	570.0 / 525.0	3.74	79444.91	255.936822	1148.1	false
13C2-PFTeDA	715.0 / 670.0	4.48	74260.00	275.760706	1540.3	false
13C3-PFBS	302.0 / 99.0	1.53	24894.70	248.039058	473.1	false
13C3-PFHxS	402.0 / 99.0	2.29	20686.54	227.456951	252.3	false
13C8-PFOS	507.0 / 99.0	3.07	23008.48	223.947248	201.2	false

BATTELLE

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	90392.05	252.058922	1749.1	false
d3-MeFOSAA	573.0 / 419.0	3.58	13210.94	221.573454	151.4	false
d5-EtFOSAA	589.0 / 419.0	3.74	13687.87	202.252273	158.1	false
13C5-PFHxA	318.0 / 273.0	1.85	55459.53	252.147535	587.9	false
13C4-PFHpA	367.0 / 322.0	2.27	61325.22	245.071359	1056.5	false
13C8-PFOA	421.0 / 376.0	2.67	81759.84	268.592304	1141.0	false
13C9-PFNA	472.0 / 427.0	3.06	79484.14	228.680180	820.6	false
13C6-PFDA	519.0 / 474.0	3.42	92426.82	255.818078	2162.1	false
13C7-PFUnA	570.0 / 525.0	3.74	86182.29	258.228174	694.8	false
13C2-PFTeDA	715.0 / 670.0	4.48	80845.62	279.224041	1594.9	false
13C3-PFBS	302.0 / 99.0	1.53	26418.26	243.217823	555.3	false
13C3-PFHxS	402.0 / 99.0	2.29	23356.00	237.294654	383.7	false
13C8-PFOS	507.0 / 99.0	3.06	27069.86	243.456749	262.7	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:34:27 PM

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	94601.53	258.457777	1212.7	false
d3-MeFOSAA	573.0 / 419.0	3.58	13723.35	227.146712	159.2	false
d5-EtFOSAA	589.0 / 419.0	3.74	13412.63	195.584216	193.5	false
13C5-PFHxA	318.0 / 273.0	1.85	59656.55	244.075531	706.3	false
13C4-PFHpA	367.0 / 322.0	2.27	67881.03	244.112187	892.5	false
13C8-PFOA	421.0 / 376.0	2.68	85596.56	253.044813	1262.7	false
13C9-PFNA	472.0 / 427.0	3.07	88113.74	228.128348	2648.5	false
13C6-PFDA	519.0 / 474.0	3.42	94064.58	255.081498	1040.0	false
13C7-PFUnA	570.0 / 525.0	3.74	85767.53	251.783994	766.6	false
13C2-PFTeDA	715.0 / 670.0	4.48	81604.89	276.141791	1401.6	false
13C3-PFBS	302.0 / 99.0	1.53	25749.91	233.953222	450.3	false
13C3-PFHxS	402.0 / 99.0	2.29	22428.15	224.876965	347.2	false
13C8-PFOS	507.0 / 99.0	3.07	27918.93	247.797379	283.4	false

BATTELLE

Summary Quant Report

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	59754.02	253.149534	1359.3	false
d3-MeFOSAA	573.0 / 419.0	3.58	8495.37	250.939491	165.1	false
d5-EtFOSAA	589.0 / 419.0	3.73	9502.87	247.294781	140.0	false
13C5-PFHxA	318.0 / 273.0	1.85	39264.94	279.995422	453.7	false
13C4-PFHpA	367.0 / 322.0	2.26	41081.38	257.493253	649.2	false
13C8-PFOA	421.0 / 376.0	2.67	53123.80	273.722234	3663.6	false
13C9-PFNA	472.0 / 427.0	3.06	53702.20	242.330211	2138.1	false
13C6-PFDA	519.0 / 474.0	3.42	55891.63	235.027234	1222.5	false
13C7-PFUnA	570.0 / 525.0	3.73	50702.91	230.811093	580.0	false
13C2-PFTeDA	715.0 / 670.0	4.48	49929.36	261.993503	1423.1	false
13C3-PFBS	302.0 / 99.0	1.53	18543.03	300.659096	385.6	true
13C3-PFHxS	402.0 / 99.0	2.28	13926.80	249.197160	276.0	false
13C8-PFOS	507.0 / 99.0	3.06	15170.44	240.290397	163.8	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:35:33 PM

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.08	104783.48	265.065125	1153.3	false
d3-MeFOSAA	573.0 / 419.0	3.63	15315.60	235.936014	235.7	false
d5-EtFOSAA	589.0 / 419.0	3.79	17400.92	236.159897	206.0	false
13C5-PFHxA	318.0 / 273.0	1.88	62450.03	270.928650	706.0	false
13C4-PFHpA	367.0 / 322.0	2.29	71875.82	274.081656	965.9	false
13C8-PFOA	421.0 / 376.0	2.71	83767.26	262.586014	2063.3	false
13C9-PFNA	472.0 / 427.0	3.11	85821.17	235.605943	1275.2	false
13C6-PFDA	519.0 / 474.0	3.47	102761.45	258.018855	3892.5	false
13C7-PFUnA	570.0 / 525.0	3.79	89179.75	242.404039	795.8	false
13C2-PFTeDA	715.0 / 670.0	4.55	94029.24	294.609831	1529.4	false
13C3-PFBS	302.0 / 99.0	1.54	25702.93	217.345232	425.2	false
13C3-PFHxS	402.0 / 99.0	2.32	22901.56	213.712943	313.3	false
13C8-PFOS	507.0 / 99.0	3.10	29238.42	241.527136	247.5	false

BATTELLE

Summary Quant Report

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.05	105491.92	291.232104	1421.8	false
d3-MeFOSAA	573.0 / 419.0	3.61	15251.83	241.106040	146.8	false
d5-EtFOSAA	589.0 / 419.0	3.77	15573.68	216.895753	175.7	false
13C5-PFHxA	318.0 / 273.0	1.86	63074.80	259.302943	691.6	false
13C4-PFHpA	367.0 / 322.0	2.28	71810.53	259.486408	915.7	false
13C8-PFOA	421.0 / 376.0	2.69	81151.91	241.060106	17478.8	false
13C9-PFNA	472.0 / 427.0	3.09	90867.26	236.389690	2200.2	false
13C6-PFDA	519.0 / 474.0	3.45	96087.92	263.299625	1564.3	false
13C7-PFUnA	570.0 / 525.0	3.76	94078.29	279.076498	1245.4	false
13C2-PFTeDA	715.0 / 670.0	4.51	91249.97	312.016305	1607.2	false
13C3-PFBS	302.0 / 99.0	1.53	25415.19	220.539729	416.6	false
13C3-PFHxS	402.0 / 99.0	2.30	22538.28	215.830304	294.6	false
13C8-PFOS	507.0 / 99.0	3.08	30801.68	261.103354	263.4	false

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:36:39 PM

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.29	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.10	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.10	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.46	PFDA			
PFDA_2	513.0 / 219.0	3.46	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.79	PFUnA			
PFUnA_2	563.0 / 269.0	3.79	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	4.07	PFDoA	0.150	0.160	ü
PFTrDA_1	663.0 / 619.0	4.32	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.32	PFTrDA	0.060	0.066	ü
PFTeDA_1	713.0 / 669.0	4.54	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.54	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.62	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.62	NMeFOSAA	0.580	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.78	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.78	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:36:43 PM

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.09	PFNA			
PFNA_2	463.0 / 219.0	3.09	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	3.76	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	4.04	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.50	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.49	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.60	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.59	NMeFOSAA	0.610	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.76	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.76	NEtFOSAA	0.050	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:36:48 PM

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.28	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.320	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	3.76	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	4.03	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.49	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.49	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.59	NMeFOSAA	0.540	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.75	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.75	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:36:53 PM

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.43	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	3.75	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	4.03	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.27	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.49	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.48	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.59	NMeFOSAA	0.540	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.74	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.74	NEtFOSAA	0.070	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:36:58 PM

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.310	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.28	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	3.75	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	4.03	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.060	0.066	ü
PFTeDA_1	713.0 / 669.0	4.49	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.49	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.59	NMeFOSAA	0.600	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.75	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.75	NEtFOSAA	0.050	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:37:03 PM

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.300	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.43	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	3.75	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	4.03	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.27	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.48	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.48	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.58	NMeFOSAA	0.590	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.74	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.75	NEtFOSAA	0.050	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:37:42 PM

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.89	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.33	PFHxS			
PFHxS_2	399.0 / 99.0	2.33	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.72	PFOA			
PFOA_2	413.0 / 169.0	2.72	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.12	PFNA			
PFNA_2	463.0 / 219.0	3.12	PFNA	0.320	0.306	ü
PFOS_1	499.0 / 80.0	3.12	PFOS			
PFOS_2	499.0 / 99.0	3.12	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.48	PFDA			
PFDA_2	513.0 / 219.0	3.48	PFDA	0.030	0.041	ü
PFUnA_1	563.0 / 519.0	3.81	PFUnA			
PFUnA_2	563.0 / 269.0	3.81	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.09	PFDoA			
PFDoA_2	613.0 / 319.0	4.09	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.34	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.34	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.56	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.55	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.64	NMeFOSAA	0.480	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.80	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.80	NEtFOSAA	0.050	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:37:47 PM

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.88	PFHxA			
PFHxA_2	313.0 / 119.0	1.88	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.29	PFHpA			
PFHpA_2	363.0 / 169.0	2.29	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.31	PFHxS			
PFHxS_2	399.0 / 99.0	2.31	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.70	PFOA			
PFOA_2	413.0 / 169.0	2.70	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.10	PFNA			
PFNA_2	463.0 / 219.0	3.10	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.10	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.46	PFDA			
PFDA_2	513.0 / 219.0	3.46	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.78	PFUnA			
PFUnA_2	563.0 / 269.0	3.78	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.06	PFDoA			
PFDoA_2	613.0 / 319.0	4.06	PFDoA	0.170	0.160	ü
PFTrDA_1	663.0 / 619.0	4.30	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.30	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.52	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.52	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.61	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.61	NMeFOSAA	0.530	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.78	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.77	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Internal Standard Report

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	27104.45	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	27104.45	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	59013.59	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	59013.59	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	70279.06	250.00
PFHpA_2	363.0 / 169.0	2.29	13C4-PFHpA	367.0 / 322.0	70279.06	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	24692.35	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	24692.35	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	89115.92	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	89115.92	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	100916.15	250.00
PFNA_2	463.0 / 219.0	3.10	13C9-PFNA	472.0 / 427.0	100916.15	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	29786.30	239.25
PFOS_2	499.0 / 99.0	3.10	13C8-PFOS	507.0 / 99.0	29786.30	239.25
PFDA_1	513.0 / 469.0	3.46	13C6-PFDA	519.0 / 474.0	90930.64	250.00
PFDA_2	513.0 / 219.0	3.46	13C6-PFDA	519.0 / 474.0	90930.64	250.00
PFUnA_1	563.0 / 519.0	3.79	13C7-PFUnA	570.0 / 525.0	89293.12	250.00
PFUnA_2	563.0 / 269.0	3.79	13C7-PFUnA	570.0 / 525.0	89293.12	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	93406.68	250.00
PFDoA_2	613.0 / 319.0	4.07	13C2-PFDoA	615.0 / 570.0	93406.68	250.00
PFTrDA_1	663.0 / 619.0	4.32	13C2-PFTeDA	715.0 / 670.0	71262.97	250.00
PFTrDA_2	663.0 / 169.0	4.32	13C2-PFTeDA	715.0 / 670.0	71262.97	250.00
PFTeDA_1	713.0 / 669.0	4.54	13C2-PFTeDA	715.0 / 670.0	71262.97	250.00
PFTeDA_2	713.0 / 169.0	4.54	13C2-PFTeDA	715.0 / 670.0	71262.97	250.00
NMeFOSAA_1	570.0 / 419.0	3.62	d3-MeFOSAA	573.0 / 419.0	17861.22	250.00
NMeFOSAA_2	570.0 / 512.0	3.62	d3-MeFOSAA	573.0 / 419.0	17861.22	250.00
NEtFOSAA_1	584.0 / 419.0	3.78	d5-EtFOSAA	589.0 / 419.0	14753.24	250.00
NEtFOSAA_2	584.0 / 483.0	3.78	d5-EtFOSAA	589.0 / 419.0	14753.24	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	25253.76	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	25253.76	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	53053.80	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	53053.80	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	62251.55	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	62251.55	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22115.08	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22115.08	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	76057.36	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	76057.36	250.00
PFNA_1	463.0 / 419.0	3.09	13C9-PFNA	472.0 / 427.0	80814.57	250.00
PFNA_2	463.0 / 219.0	3.09	13C9-PFNA	472.0 / 427.0	80814.57	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	26604.59	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	26604.59	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	95219.88	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	95219.88	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	88368.55	250.00
PFUnA_2	563.0 / 269.0	3.76	13C7-PFUnA	570.0 / 525.0	88368.55	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	90196.83	250.00
PFDoA_2	613.0 / 319.0	4.04	13C2-PFDoA	615.0 / 570.0	90196.83	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	79501.87	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	79501.87	250.00
PFTeDA_1	713.0 / 669.0	4.50	13C2-PFTeDA	715.0 / 670.0	79501.87	250.00
PFTeDA_2	713.0 / 169.0	4.49	13C2-PFTeDA	715.0 / 670.0	79501.87	250.00
NMeFOSAA_1	570.0 / 419.0	3.60	d3-MeFOSAA	573.0 / 419.0	13839.03	250.00
NMeFOSAA_2	570.0 / 512.0	3.59	d3-MeFOSAA	573.0 / 419.0	13839.03	250.00
NEtFOSAA_1	584.0 / 419.0	3.76	d5-EtFOSAA	589.0 / 419.0	16338.86	250.00
NEtFOSAA_2	584.0 / 483.0	3.76	d5-EtFOSAA	589.0 / 419.0	16338.86	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	24894.70	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	24894.70	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	51469.81	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	51469.81	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	59143.44	250.00
PFHpA_2	363.0 / 169.0	2.28	13C4-PFHpA	367.0 / 322.0	59143.44	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	20543.88	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	20543.88	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	78620.84	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	78620.84	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	78173.73	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	78173.73	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	23379.08	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	23379.08	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	79904.56	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	79904.56	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	79444.91	250.00
PFUnA_2	563.0 / 269.0	3.76	13C7-PFUnA	570.0 / 525.0	79444.91	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	85629.59	250.00
PFDoA_2	613.0 / 319.0	4.03	13C2-PFDoA	615.0 / 570.0	85629.59	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	74260.00	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	74260.00	250.00
PFTeDA_1	713.0 / 669.0	4.49	13C2-PFTeDA	715.0 / 670.0	74260.00	250.00
PFTeDA_2	713.0 / 169.0	4.49	13C2-PFTeDA	715.0 / 670.0	74260.00	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	12558.47	250.00
NMeFOSAA_2	570.0 / 512.0	3.59	d3-MeFOSAA	573.0 / 419.0	12558.47	250.00
NEtFOSAA_1	584.0 / 419.0	3.75	d5-EtFOSAA	589.0 / 419.0	15493.20	250.00
NEtFOSAA_2	584.0 / 483.0	3.75	d5-EtFOSAA	589.0 / 419.0	15493.20	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	26418.26	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	26418.26	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	55459.53	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	55459.53	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	61325.22	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	61325.22	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	23003.92	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	23003.92	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	81759.84	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	81759.84	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	79484.14	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	79484.14	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	27054.88	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	27054.88	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	92426.82	250.00
PFDA_2	513.0 / 219.0	3.43	13C6-PFDA	519.0 / 474.0	92426.82	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	86182.29	250.00
PFUnA_2	563.0 / 269.0	3.75	13C7-PFUnA	570.0 / 525.0	86182.29	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	90392.05	250.00
PFDoA_2	613.0 / 319.0	4.03	13C2-PFDoA	615.0 / 570.0	90392.05	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	80845.62	250.00
PFTrDA_2	663.0 / 169.0	4.27	13C2-PFTeDA	715.0 / 670.0	80845.62	250.00
PFTeDA_1	713.0 / 669.0	4.49	13C2-PFTeDA	715.0 / 670.0	80845.62	250.00
PFTeDA_2	713.0 / 169.0	4.48	13C2-PFTeDA	715.0 / 670.0	80845.62	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	13418.68	250.00
NMeFOSAA_2	570.0 / 512.0	3.59	d3-MeFOSAA	573.0 / 419.0	13418.68	250.00
NEtFOSAA_1	584.0 / 419.0	3.74	d5-EtFOSAA	589.0 / 419.0	13962.46	250.00
NEtFOSAA_2	584.0 / 483.0	3.74	d5-EtFOSAA	589.0 / 419.0	13962.46	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	25749.91	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	25749.91	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	59656.55	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	59656.55	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	67881.03	250.00
PFHpA_2	363.0 / 169.0	2.28	13C4-PFHpA	367.0 / 322.0	67881.03	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22230.55	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22230.55	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	85596.56	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	85596.56	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	88113.74	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	88113.74	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	28433.07	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	28433.07	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	94064.58	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	94064.58	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	85767.53	250.00
PFUnA_2	563.0 / 269.0	3.75	13C7-PFUnA	570.0 / 525.0	85767.53	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	94601.53	250.00
PFDoA_2	613.0 / 319.0	4.03	13C2-PFDoA	615.0 / 570.0	94601.53	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	81604.89	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	81604.89	250.00
PFTeDA_1	713.0 / 669.0	4.49	13C2-PFTeDA	715.0 / 670.0	81604.89	250.00
PFTeDA_2	713.0 / 169.0	4.49	13C2-PFTeDA	715.0 / 670.0	81604.89	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	13961.67	250.00
NMeFOSAA_2	570.0 / 512.0	3.59	d3-MeFOSAA	573.0 / 419.0	13961.67	250.00
NEtFOSAA_1	584.0 / 419.0	3.75	d5-EtFOSAA	589.0 / 419.0	13455.19	250.00
NEtFOSAA_2	584.0 / 483.0	3.75	d5-EtFOSAA	589.0 / 419.0	13455.19	250.00

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	18707.64	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	18707.64	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	39264.94	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	39264.94	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	41081.38	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	41081.38	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	13736.84	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	13736.84	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	53123.80	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	53123.80	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	53702.20	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	53702.20	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	15576.11	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	15576.11	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	55891.63	250.00
PFDA_2	513.0 / 219.0	3.43	13C6-PFDA	519.0 / 474.0	55891.63	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	50702.91	250.00
PFUnA_2	563.0 / 269.0	3.75	13C7-PFUnA	570.0 / 525.0	50702.91	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	59754.02	250.00
PFDoA_2	613.0 / 319.0	4.03	13C2-PFDoA	615.0 / 570.0	59754.02	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	49929.36	250.00
PFTrDA_2	663.0 / 169.0	4.27	13C2-PFTeDA	715.0 / 670.0	49929.36	250.00
PFTeDA_1	713.0 / 669.0	4.48	13C2-PFTeDA	715.0 / 670.0	49929.36	250.00
PFTeDA_2	713.0 / 169.0	4.48	13C2-PFTeDA	715.0 / 670.0	49929.36	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	8449.93	250.00
NMeFOSAA_2	570.0 / 512.0	3.58	d3-MeFOSAA	573.0 / 419.0	8449.93	250.00
NEtFOSAA_1	584.0 / 419.0	3.74	d5-EtFOSAA	589.0 / 419.0	9549.31	250.00
NEtFOSAA_2	584.0 / 483.0	3.75	d5-EtFOSAA	589.0 / 419.0	9549.31	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	25702.93	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	25702.93	232.25
PFHxA_1	313.0 / 269.0	1.89	13C5-PFHxA	318.0 / 273.0	62450.03	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	62450.03	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	71875.82	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	71875.82	250.00
PFHxS_1	399.0 / 80.0	2.33	13C3-PFHxS	402.0 / 99.0	22780.75	236.50
PFHxS_2	399.0 / 99.0	2.33	13C3-PFHxS	402.0 / 99.0	22780.75	236.50
PFOA_1	413.0 / 369.0	2.72	13C8-PFOA	421.0 / 376.0	83767.26	250.00
PFOA_2	413.0 / 169.0	2.72	13C8-PFOA	421.0 / 376.0	83767.26	250.00
PFNA_1	463.0 / 419.0	3.12	13C9-PFNA	472.0 / 427.0	85821.17	250.00
PFNA_2	463.0 / 219.0	3.12	13C9-PFNA	472.0 / 427.0	85821.17	250.00
PFOS_1	499.0 / 80.0	3.12	13C8-PFOS	507.0 / 99.0	29047.77	239.25
PFOS_2	499.0 / 99.0	3.12	13C8-PFOS	507.0 / 99.0	29047.77	239.25
PFDA_1	513.0 / 469.0	3.48	13C6-PFDA	519.0 / 474.0	102761.45	250.00
PFDA_2	513.0 / 219.0	3.48	13C6-PFDA	519.0 / 474.0	102761.45	250.00
PFUnA_1	563.0 / 519.0	3.81	13C7-PFUnA	570.0 / 525.0	89179.75	250.00
PFUnA_2	563.0 / 269.0	3.81	13C7-PFUnA	570.0 / 525.0	89179.75	250.00
PFDoA_1	613.0 / 569.0	4.09	13C2-PFDoA	615.0 / 570.0	104783.48	250.00
PFDoA_2	613.0 / 319.0	4.09	13C2-PFDoA	615.0 / 570.0	104783.48	250.00
PFTrDA_1	663.0 / 619.0	4.34	13C2-PFTeDA	715.0 / 670.0	94029.24	250.00
PFTrDA_2	663.0 / 169.0	4.34	13C2-PFTeDA	715.0 / 670.0	94029.24	250.00
PFTeDA_1	713.0 / 669.0	4.56	13C2-PFTeDA	715.0 / 670.0	94029.24	250.00
PFTeDA_2	713.0 / 169.0	4.55	13C2-PFTeDA	715.0 / 670.0	94029.24	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	15495.78	250.00
NMeFOSAA_2	570.0 / 512.0	3.64	d3-MeFOSAA	573.0 / 419.0	15495.78	250.00
NEtFOSAA_1	584.0 / 419.0	3.80	d5-EtFOSAA	589.0 / 419.0	18329.11	250.00
NEtFOSAA_2	584.0 / 483.0	3.80	d5-EtFOSAA	589.0 / 419.0	18329.11	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	25415.19	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	25415.19	232.25
PFHxA_1	313.0 / 269.0	1.88	13C5-PFHxA	318.0 / 273.0	63074.80	250.00
PFHxA_2	313.0 / 119.0	1.88	13C5-PFHxA	318.0 / 273.0	63074.80	250.00
PFHpA_1	363.0 / 319.0	2.29	13C4-PFHpA	367.0 / 322.0	71810.53	250.00
PFHpA_2	363.0 / 169.0	2.29	13C4-PFHpA	367.0 / 322.0	71810.53	250.00
PFHxS_1	399.0 / 80.0	2.31	13C3-PFHxS	402.0 / 99.0	22252.14	236.50
PFHxS_2	399.0 / 99.0	2.31	13C3-PFHxS	402.0 / 99.0	22252.14	236.50
PFOA_1	413.0 / 369.0	2.70	13C8-PFOA	421.0 / 376.0	81151.91	250.00
PFOA_2	413.0 / 169.0	2.70	13C8-PFOA	421.0 / 376.0	81151.91	250.00
PFNA_1	463.0 / 419.0	3.10	13C9-PFNA	472.0 / 427.0	90867.26	250.00
PFNA_2	463.0 / 219.0	3.10	13C9-PFNA	472.0 / 427.0	90867.26	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	31014.44	239.25
PFOS_2	499.0 / 99.0	3.10	13C8-PFOS	507.0 / 99.0	31014.44	239.25
PFDA_1	513.0 / 469.0	3.46	13C6-PFDA	519.0 / 474.0	96087.92	250.00
PFDA_2	513.0 / 219.0	3.46	13C6-PFDA	519.0 / 474.0	96087.92	250.00
PFUnA_1	563.0 / 519.0	3.78	13C7-PFUnA	570.0 / 525.0	94078.29	250.00
PFUnA_2	563.0 / 269.0	3.78	13C7-PFUnA	570.0 / 525.0	94078.29	250.00
PFDoA_1	613.0 / 569.0	4.06	13C2-PFDoA	615.0 / 570.0	105491.92	250.00
PFDoA_2	613.0 / 319.0	4.06	13C2-PFDoA	615.0 / 570.0	105491.92	250.00
PFTrDA_1	663.0 / 619.0	4.30	13C2-PFTeDA	715.0 / 670.0	91249.97	250.00
PFTrDA_2	663.0 / 169.0	4.30	13C2-PFTeDA	715.0 / 670.0	91249.97	250.00
PFTeDA_1	713.0 / 669.0	4.52	13C2-PFTeDA	715.0 / 670.0	91249.97	250.00
PFTeDA_2	713.0 / 169.0	4.52	13C2-PFTeDA	715.0 / 670.0	91249.97	250.00
NMeFOSAA_1	570.0 / 419.0	3.61	d3-MeFOSAA	573.0 / 419.0	15675.74	250.00
NMeFOSAA_2	570.0 / 512.0	3.61	d3-MeFOSAA	573.0 / 419.0	15675.74	250.00
NEtFOSAA_1	584.0 / 419.0	3.78	d5-EtFOSAA	589.0 / 419.0	16116.46	250.00
NEtFOSAA_2	584.0 / 483.0	3.77	d5-EtFOSAA	589.0 / 419.0	16116.46	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.06	13C2-PFDA	515.0 / 470.0	103402.36	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	33690.55	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	33690.55	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	85242.44	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	85242.44	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	85242.44	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	85242.44	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	103402.36	250.00
13C7-PFUnA	570.0 / 525.0	3.77	13C2-PFDA	515.0 / 470.0	103402.36	250.00
13C2-PFTeDA	715.0 / 670.0	4.54	13C2-PFDA	515.0 / 470.0	103402.36	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	33690.55	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	33690.55	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	33690.55	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	91954.41	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	31073.45	239.25
d5-EtFOSAA	589.0 / 419.0	3.75	13C4-PFOS	503.0 / 99.0	31073.45	239.25
13C5-PFHxA	318.0 / 273.0	1.86	13C2-PFOA	415.0 / 370.0	73507.58	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	73507.58	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	73507.58	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	73507.58	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	91954.41	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	91954.41	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	91954.41	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	31073.45	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	31073.45	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	31073.45	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	85518.46	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	25806.64	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	25806.64	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	74339.52	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	74339.52	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	74339.52	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	74339.52	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	85518.46	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	85518.46	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	85518.46	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	25806.64	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	25806.64	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	25806.64	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	91947.72	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	27928.87	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	27928.87	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	74313.12	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	74313.12	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	74313.12	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	74313.12	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	91947.72	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	91947.72	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	91947.72	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	27928.87	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	27928.87	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	27928.87	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	93847.21	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	28300.31	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	28300.31	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	82580.58	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	82580.58	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	82580.58	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	82580.58	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	93847.21	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	93847.21	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	93847.21	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	28300.31	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	28300.31	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	28300.31	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	60520.54	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	15858.09	239.25
d5-EtFOSAA	589.0 / 419.0	3.73	13C4-PFOS	503.0 / 99.0	15858.09	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	47380.33	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	47380.33	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	47380.33	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	47380.33	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	60520.54	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	60520.54	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	60520.54	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	15858.09	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	15858.09	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	15858.09	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:39:53 PM

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.08	13C2-PFDA	515.0 / 470.0	101356.83	250.00
d3-MeFOSAA	573.0 / 419.0	3.63	13C4-PFOS	503.0 / 99.0	30407.25	239.25
d5-EtFOSAA	589.0 / 419.0	3.79	13C4-PFOS	503.0 / 99.0	30407.25	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	77879.25	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	77879.25	250.00
13C8-PFOA	421.0 / 376.0	2.71	13C2-PFOA	415.0 / 370.0	77879.25	250.00
13C9-PFNA	472.0 / 427.0	3.11	13C2-PFOA	415.0 / 370.0	77879.25	250.00
13C6-PFDA	519.0 / 474.0	3.47	13C2-PFDA	515.0 / 470.0	101356.83	250.00
13C7-PFUnA	570.0 / 525.0	3.79	13C2-PFDA	515.0 / 470.0	101356.83	250.00
13C2-PFTeDA	715.0 / 670.0	4.55	13C2-PFDA	515.0 / 470.0	101356.83	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	30407.25	239.25
13C3-PFHxS	402.0 / 99.0	2.32	13C4-PFOS	503.0 / 99.0	30407.25	239.25
13C8-PFOS	507.0 / 99.0	3.10	13C4-PFOS	503.0 / 99.0	30407.25	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.05	13C2-PFDA	515.0 / 470.0	92873.70	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	29631.33	239.25
d5-EtFOSAA	589.0 / 419.0	3.77	13C4-PFOS	503.0 / 99.0	29631.33	239.25
13C5-PFHxA	318.0 / 273.0	1.86	13C2-PFOA	415.0 / 370.0	82184.98	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	82184.98	250.00
13C8-PFOA	421.0 / 376.0	2.69	13C2-PFOA	415.0 / 370.0	82184.98	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	82184.98	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	92873.70	250.00
13C7-PFUnA	570.0 / 525.0	3.76	13C2-PFDA	515.0 / 470.0	92873.70	250.00
13C2-PFTeDA	715.0 / 670.0	4.51	13C2-PFDA	515.0 / 470.0	92873.70	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	29631.33	239.25
13C3-PFHxS	402.0 / 99.0	2.30	13C4-PFOS	503.0 / 99.0	29631.33	239.25
13C8-PFOS	507.0 / 99.0	3.08	13C4-PFOS	503.0 / 99.0	29631.33	239.25

Raw Analytical Data

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:42 PM

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.57	4270.08	7.967207	23.4	true
PFBS_2	298.9 / 99.0	1.57	1751.65	15.187213	22.1	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	N/A	N/A	N/A	N/A	true
PFHxS_2	399.0 / 99.0	N/A	N/A	N/A	N/A	true
PFOA_1	413.0 / 369.0	N/A	N/A	N/A	N/A	true
PFOA_2	413.0 / 169.0	N/A	N/A	N/A	N/A	true
PFNA_1	463.0 / 419.0	3.11	5116.99	< 0	29.4	false
PFNA_2	463.0 / 219.0	3.11	1412.81	< 0	28.4	false
PFOS_1	499.0 / 80.0	3.10	10007.21	20.457445	28.3	false
PFOS_2	499.0 / 99.0	3.11	2107.71	23.036162	36.0	false
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.08	4132.45	< 0	50.7	false
PFDoA_2	613.0 / 319.0	4.08	648.21	< 0	25.1	false
PFTrDA_1	663.0 / 619.0	4.32	3827.82	< 0	101.8	false
PFTrDA_2	663.0 / 169.0	4.34	292.14	< 0	21.8	false
PFTeDA_1	713.0 / 669.0	4.54	4005.90	< 0	149.6	false
PFTeDA_2	713.0 / 169.0	4.54	211.66	< 0	28.3	false
NMeFOSAA_1	570.0 / 419.0	3.63	1091.61	< 0	277.3	false
NMeFOSAA_2	570.0 / 512.0	3.61	558.08	< 0	27.6	false
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:11:42 AM

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.58	2620.38	3.363745	18.0	true
PFBS_2	298.9 / 99.0	1.86	4524.10	48.786916	86.9	false
PFHxA_1	313.0 / 269.0	1.92	10499.59	11.339442	4.3	false
PFHxA_2	313.0 / 119.0	1.91	1053.00	9.946906	3.6	false
PFHpA_1	363.0 / 319.0	2.34	5963.93	6.614413	14.5	false
PFHpA_2	363.0 / 169.0	2.49	146.49	< 0	8.8	false
PFHxS_1	399.0 / 80.0	2.36	6822.95	2.081422	33.1	false
PFHxS_2	399.0 / 99.0	2.37	1666.19	< 0	22.8	false
PFOA_1	413.0 / 369.0	2.75	8135.53	11.076595	22.3	false
PFOA_2	413.0 / 169.0	2.75	715.87	21.127193	15.2	false
PFNA_1	463.0 / 419.0	3.16	6357.44	< 0	19.8	false
PFNA_2	463.0 / 219.0	3.15	1267.46	< 0	16.6	false
PFOS_1	499.0 / 80.0	N/A	N/A	N/A	N/A	true
PFOS_2	499.0 / 99.0	N/A	N/A	N/A	N/A	true
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:46 PM

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	N/A	N/A	N/A	N/A	true
PFHxS_2	399.0 / 99.0	N/A	N/A	N/A	N/A	true
PFOA_1	413.0 / 369.0	2.69	79551.79	239.182133	128.2	false
PFOA_2	413.0 / 169.0	2.69	4182.33	191.484186	74.2	false
PFNA_1	463.0 / 419.0	3.09	4533.64	< 0	24.6	false
PFNA_2	463.0 / 219.0	3.08	1416.90	< 0	33.0	false
PFOS_1	499.0 / 80.0	3.08	5998.62	14.549685	14.9	true
PFOS_2	499.0 / 99.0	3.04	965.83	12.222160	6.6	false
PFDA_1	513.0 / 469.0	3.43	1886.69	< 0	14.2	false
PFDA_2	513.0 / 219.0	3.47	219.62	< 0	10.3	true
PFUnA_1	563.0 / 519.0	3.77	2614.06	< 0	20.1	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.29	869.13	< 0	30.2	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:55 PM

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	21817513.87	80210.116161	510.2	false
PFBS_2	298.9 / 99.0	1.54	6502260.94	81400.038998	671.1	false
PFHxA_1	313.0 / 269.0	1.87	87506233.36	607742.293640	331.4	false
PFHxA_2	313.0 / 119.0	1.86	24831190.35	2315639.602391	861.0	false
PFHpA_1	363.0 / 319.0	2.28	77792422.27	456519.341180	706.7	false
PFHpA_2	363.0 / 169.0	2.27	3397630.10	942529.222078	924.9	false
PFHxS_1	399.0 / 80.0	2.30	107794078.13	456258.252485	771.5	false
PFHxS_2	399.0 / 99.0	2.30	42494455.44	645142.776421	1602.5	false
PFOA_1	413.0 / 369.0	2.65	120340307.31	866177.562273	426.8	false
PFOA_2	413.0 / 169.0	2.61	32224315.88	3571649.277222	748.2	false
PFNA_1	463.0 / 419.0	3.04	335674.57	1449.929641	139.4	false
PFNA_2	463.0 / 219.0	3.07	54070.43	744.950403	86.6	true
PFOS_1	499.0 / 80.0	2.95	12923224.41	32703.726029	341.2	false
PFOS_2	499.0 / 99.0	3.00	1379036.68	20101.726076	423.9	true
PFDA_1	513.0 / 469.0	3.44	12546.42	16.149935	22.2	true
PFDA_2	513.0 / 219.0	3.45	631.49	22.251317	17.5	true
PFUnA_1	563.0 / 519.0	3.76	11006.89	22.372577	37.0	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	10917.69	7.878150	74.7	false
PFDoA_2	613.0 / 319.0	4.04	1754.18	< 0	36.7	true
PFTrDA_1	663.0 / 619.0	4.28	11293.14	11.812107	137.1	false
PFTrDA_2	663.0 / 169.0	4.28	773.90	9.996468	41.4	false
PFTeDA_1	713.0 / 669.0	4.49	13626.25	6.134138	165.5	false
PFTeDA_2	713.0 / 169.0	4.48	564.35	< 0	53.4	false
NMeFOSAA_1	570.0 / 419.0	3.59	2389.64	< 0	54.1	false
NMeFOSAA_2	570.0 / 512.0	3.60	1316.80	< 0	33.5	false
NEtFOSAA_1	584.0 / 419.0	3.74	1519.16	20.099108	50.5	false
NEtFOSAA_2	584.0 / 483.0	3.74	252.16	32.416049	28.7	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:00 PM

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	3498110.91	8296.932585	544.8	false
PFBS_2	298.9 / 99.0	1.54	1072726.39	8665.707068	587.9	false
PFHxA_1	313.0 / 269.0	1.86	49081764.26	199206.975369	588.3	false
PFHxA_2	313.0 / 119.0	1.86	4058105.79	221136.228110	681.0	false
PFHpA_1	363.0 / 319.0	2.27	26489241.55	102357.333054	1069.3	false
PFHpA_2	363.0 / 169.0	2.27	562793.94	102735.897018	1019.8	false
PFHxS_1	399.0 / 80.0	2.30	20411289.00	62801.092020	1351.6	false
PFHxS_2	399.0 / 99.0	2.30	5770645.80	63678.555807	1835.5	false
PFOA_1	413.0 / 369.0	2.68	45220204.32	162167.102257	1041.6	false
PFOA_2	413.0 / 169.0	2.64	5027691.96	277651.724379	856.4	true
PFNA_1	463.0 / 419.0	3.07	59006.75	165.409544	103.2	false
PFNA_2	463.0 / 219.0	3.07	14236.80	126.000934	84.7	true
PFOS_1	499.0 / 80.0	2.94	1697961.23	2994.805805	212.2	false
PFOS_2	499.0 / 99.0	2.99	178055.05	1808.319952	231.9	true
PFDA_1	513.0 / 469.0	3.43	9670.56	1.852228	32.0	false
PFDA_2	513.0 / 219.0	3.42	676.49	15.932645	18.6	false
PFUnA_1	563.0 / 519.0	3.76	4506.36	1.627549	30.7	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	2198.95	< 0	39.5	false
PFDoA_2	613.0 / 319.0	4.02	302.68	< 0	15.7	true
PFTrDA_1	663.0 / 619.0	4.28	2212.05	< 0	45.8	false
PFTrDA_2	663.0 / 169.0	4.27	264.17	< 0	14.1	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:04 PM

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	693565.06	1651.926249	412.2	false
PFBS_2	298.9 / 99.0	1.54	219227.85	1780.247307	486.8	false
PFHxA_1	313.0 / 269.0	1.86	11031705.33	44230.764187	419.9	false
PFHxA_2	313.0 / 119.0	1.86	822912.62	44284.033738	372.5	false
PFHpA_1	363.0 / 319.0	2.27	5588441.48	19653.174335	762.1	false
PFHpA_2	363.0 / 169.0	2.27	117148.43	19405.538077	647.5	false
PFHxS_1	399.0 / 80.0	2.30	4507395.44	12693.384317	980.1	false
PFHxS_2	399.0 / 99.0	2.30	1228716.04	12405.401960	1329.2	false
PFOA_1	413.0 / 369.0	2.68	12459517.63	38538.540244	1011.3	false
PFOA_2	413.0 / 169.0	2.64	1172298.86	55843.736134	850.1	false
PFNA_1	463.0 / 419.0	3.06	15742.25	20.268342	52.4	false
PFNA_2	463.0 / 219.0	3.09	3607.40	10.449877	31.1	true
PFOS_1	499.0 / 80.0	2.94	393218.70	643.172105	133.9	false
PFOS_2	499.0 / 99.0	2.98	39411.37	370.685414	156.6	true
PFDA_1	513.0 / 469.0	3.43	3069.41	< 0	21.0	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.76	1554.76	< 0	16.1	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.04	852.63	< 0	13.0	false
PFDoA_2	613.0 / 319.0	4.04	127.67	< 0	5.7	true
PFTrDA_1	663.0 / 619.0	4.28	840.57	< 0	33.7	false
PFTrDA_2	663.0 / 169.0	4.27	82.22	< 0	8.3	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	178751.95	491.523422	303.6	false
PFBS_2	298.9 / 99.0	1.54	55467.89	521.515742	240.7	false
PFHxA_1	313.0 / 269.0	1.86	3031716.32	12316.246807	218.8	false
PFHxA_2	313.0 / 119.0	1.86	226303.98	12327.151206	183.6	false
PFHpA_1	363.0 / 319.0	2.28	1578720.72	6344.235352	442.9	false
PFHpA_2	363.0 / 169.0	2.27	31786.87	5967.526621	547.8	false
PFHxS_1	399.0 / 80.0	2.30	1281073.66	3769.141157	697.2	false
PFHxS_2	399.0 / 99.0	2.30	370760.68	3908.172139	820.6	false
PFOA_1	413.0 / 369.0	2.68	3775464.47	10899.060226	652.5	false
PFOA_2	413.0 / 169.0	2.64	340783.80	15155.546304	477.0	false
PFNA_1	463.0 / 419.0	3.08	12127.42	5.485572	44.6	false
PFNA_2	463.0 / 219.0	3.08	4695.27	19.045135	64.7	false
PFOS_1	499.0 / 80.0	2.94	123395.53	191.884778	82.9	false
PFOS_2	499.0 / 99.0	2.97	11173.31	99.539371	70.7	true
PFDA_1	513.0 / 469.0	3.43	12435.81	7.839294	57.5	false
PFDA_2	513.0 / 219.0	3.43	925.97	29.315206	33.6	false
PFUnA_1	563.0 / 519.0	3.76	2088.08	< 0	22.6	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.04	2775.79	< 0	35.4	false
PFDoA_2	613.0 / 319.0	4.02	857.40	< 0	24.9	false
PFTrDA_1	663.0 / 619.0	4.27	577.39	< 0	23.9	false
PFTrDA_2	663.0 / 169.0	4.27	299.41	< 0	12.6	false
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:11 PM

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	2.29	9552.51	28.498504	21.3	false
PFHpA_2	363.0 / 169.0	2.29	316.72	< 0	22.9	false
PFHxS_1	399.0 / 80.0	2.30	5675.96	< 0	35.9	false
PFHxS_2	399.0 / 99.0	2.31	1943.78	< 0	16.6	false
PFOA_1	413.0 / 369.0	2.69	81115.94	265.300213	118.6	false
PFOA_2	413.0 / 169.0	2.69	4979.80	250.708424	81.0	false
PFNA_1	463.0 / 419.0	3.08	2221.22	< 0	17.1	false
PFNA_2	463.0 / 219.0	3.10	758.54	< 0	16.5	false
PFOS_1	499.0 / 80.0	N/A	N/A	N/A	N/A	true
PFOS_2	499.0 / 99.0	N/A	N/A	N/A	N/A	true
PFDA_1	513.0 / 469.0	3.41	1516.48	< 0	12.0	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.75	1639.78	< 0	15.0	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	531.42	< 0	16.1	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	705.09	< 0	25.1	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:14 PM

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	2.28	5595.81	9.996192	15.3	false
PFHpA_2	363.0 / 169.0	2.30	143.81	< 0	9.1	false
PFHxS_1	399.0 / 80.0	2.30	3934.30	< 0	21.4	false
PFHxS_2	399.0 / 99.0	2.29	885.37	< 0	19.8	false
PFOA_1	413.0 / 369.0	2.68	80444.54	261.163445	121.3	false
PFOA_2	413.0 / 169.0	2.68	3759.31	184.421885	79.4	false
PFNA_1	463.0 / 419.0	3.08	2560.76	< 0	20.1	false
PFNA_2	463.0 / 219.0	3.08	1570.45	< 0	24.4	false
PFOS_1	499.0 / 80.0	3.08	12390.68	28.238329	31.9	false
PFOS_2	499.0 / 99.0	3.06	2542.29	31.647210	21.8	false
PFDA_1	513.0 / 469.0	3.43	1580.52	< 0	14.6	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.76	2279.81	< 0	19.1	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	3532.00	< 0	39.5	true
PFDoA_2	613.0 / 319.0	4.02	928.42	< 0	16.3	true
PFTrDA_1	663.0 / 619.0	4.27	3848.46	< 0	87.8	false
PFTrDA_2	663.0 / 169.0	4.28	227.91	< 0	14.0	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	3.74	523.74	< 0	27.8	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:31:56 AM

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	31265150.91	127217.892650	452.5	false
PFBS_2	298.9 / 99.0	1.53	9293040.06	128759.111077	804.8	false
PFHxA_1	313.0 / 269.0	1.85	18785350.89	77803.029958	181.7	false
PFHxA_2	313.0 / 119.0	1.86	1120899.89	62309.840597	363.2	false
PFHpA_1	363.0 / 319.0	2.27	3356519.25	13403.289247	153.0	false
PFHpA_2	363.0 / 169.0	2.24	82478.55	15502.694871	305.2	false
PFHxS_1	399.0 / 80.0	2.29	74399276.83	233751.067310	509.4	false
PFHxS_2	399.0 / 99.0	2.29	24788671.80	279345.568113	1268.8	false
PFOA_1	413.0 / 369.0	2.68	4868226.92	17468.873309	233.0	false
PFOA_2	413.0 / 169.0	2.68	225799.98	12473.153904	521.1	true
PFNA_1	463.0 / 419.0	3.08	75127.85	274.353657	58.0	false
PFNA_2	463.0 / 219.0	3.08	27667.15	337.538160	122.1	false
PFOS_1	499.0 / 80.0	2.95	14520228.67	37877.188970	343.6	false
PFOS_2	499.0 / 99.0	3.00	1658233.23	24916.251689	466.0	true
PFDA_1	513.0 / 469.0	3.43	4127.57	< 0	14.4	true
PFDA_2	513.0 / 219.0	3.41	374.45	1.574719	10.9	true
PFUnA_1	563.0 / 519.0	3.75	4893.98	3.893522	20.9	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	3081.18	< 0	22.9	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	1586.89	< 0	36.1	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	15312845.30	34710.971411	432.9	false
PFBS_2	298.9 / 99.0	1.54	4560643.50	35203.578507	726.2	false
PFHxA_1	313.0 / 269.0	1.86	29011565.45	82531.435863	284.6	false
PFHxA_2	313.0 / 119.0	1.87	1773646.02	67724.423953	472.5	false
PFHpA_1	363.0 / 319.0	2.27	1223639.66	3585.283150	131.2	false
PFHpA_2	363.0 / 169.0	2.24	32362.44	4417.212099	178.9	false
PFHxS_1	399.0 / 80.0	2.30	17173562.63	38996.438600	344.3	false
PFHxS_2	399.0 / 99.0	2.30	4717314.14	38415.165581	955.8	false
PFOA_1	413.0 / 369.0	2.67	638817.95	1826.581770	145.8	false
PFOA_2	413.0 / 169.0	2.68	21926.10	958.989929	173.6	true
PFNA_1	463.0 / 419.0	3.08	16032.34	20.467797	28.4	false
PFNA_2	463.0 / 219.0	3.08	4949.30	24.313790	43.1	false
PFOS_1	499.0 / 80.0	2.95	1390451.74	2457.415234	126.1	false
PFOS_2	499.0 / 99.0	3.05	180594.48	1837.591240	226.6	true
PFDA_1	513.0 / 469.0	3.43	2113.27	< 0	10.2	false
PFDA_2	513.0 / 219.0	3.41	266.65	< 0	10.8	false
PFUnA_1	563.0 / 519.0	3.75	2765.11	< 0	15.2	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	1535.15	< 0	16.8	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	812.02	< 0	25.2	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:10 AM

Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	1791381.55	4212.957365	467.6	false
PFBS_2	298.9 / 99.0	1.54	541706.76	4340.011642	563.3	false
PFHxA_1	313.0 / 269.0	1.86	2420596.92	9300.455758	132.7	false
PFHxA_2	313.0 / 119.0	1.87	152178.29	7828.453648	148.7	false
PFHpA_1	363.0 / 319.0	2.27	116989.12	439.371024	86.2	false
PFHpA_2	363.0 / 169.0	2.24	2809.84	431.261037	49.4	false
PFHxS_1	399.0 / 80.0	2.30	1546790.82	4888.272158	410.1	false
PFHxS_2	399.0 / 99.0	2.30	417499.50	4726.589324	739.8	false
PFOA_1	413.0 / 369.0	2.68	102991.63	286.395814	106.2	false
PFOA_2	413.0 / 169.0	2.67	5842.28	248.987170	119.0	true
PFNA_1	463.0 / 419.0	3.08	36170.52	86.616835	99.1	false
PFNA_2	463.0 / 219.0	3.08	9888.58	76.821920	96.7	false
PFOS_1	499.0 / 80.0	2.96	150156.59	276.003781	90.2	false
PFOS_2	499.0 / 99.0	3.07	22153.54	233.430242	88.1	true
PFDA_1	513.0 / 469.0	3.44	9840.65	4.982657	57.2	false
PFDA_2	513.0 / 219.0	3.42	927.66	36.913928	33.3	false
PFUnA_1	563.0 / 519.0	3.76	4907.73	3.114027	30.0	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.05	2516.17	< 0	44.0	false
PFDoA_2	613.0 / 319.0	4.00	449.69	< 0	10.1	true
PFTrDA_1	663.0 / 619.0	4.28	984.20	< 0	26.3	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:15 AM

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	15379638.59	36190.069213	403.2	false
PFBS_2	298.9 / 99.0	1.54	4513427.07	36165.785357	795.9	false
PFHxA_1	313.0 / 269.0	1.85	28311181.59	83000.652798	261.8	false
PFHxA_2	313.0 / 119.0	1.86	1713119.46	67412.243159	448.5	false
PFHpA_1	363.0 / 319.0	2.27	1221626.77	3553.761757	138.2	false
PFHpA_2	363.0 / 169.0	2.22	36279.25	4926.726937	185.4	false
PFHxS_1	399.0 / 80.0	2.30	16462487.30	38752.298790	421.1	false
PFHxS_2	399.0 / 99.0	2.29	4480245.89	37821.970703	1308.6	false
PFOA_1	413.0 / 369.0	2.67	598190.82	1573.275551	143.1	false
PFOA_2	413.0 / 169.0	2.67	21302.66	856.636466	169.6	true
PFNA_1	463.0 / 419.0	3.08	7287.19	< 0	21.8	false
PFNA_2	463.0 / 219.0	3.07	3049.55	5.055380	38.5	false
PFOS_1	499.0 / 80.0	2.96	1255961.77	1986.522570	121.0	false
PFOS_2	499.0 / 99.0	3.06	171223.21	1558.988433	205.5	true
PFDA_1	513.0 / 469.0	3.44	2479.22	< 0	11.0	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	1724.43	< 0	21.8	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.29	667.15	< 0	21.5	true
PFTrDA_2	663.0 / 169.0	4.28	129.68	< 0	7.9	false
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:19 AM

Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	1460458.47	3292.763275	435.6	false
PFBS_2	298.9 / 99.0	1.54	449774.37	3455.084079	523.6	false
PFHxA_1	313.0 / 269.0	1.86	2026559.84	8079.143504	126.0	false
PFHxA_2	313.0 / 119.0	1.87	122090.57	6511.795016	118.6	false
PFHpA_1	363.0 / 319.0	2.27	88890.46	323.864937	72.1	false
PFHpA_2	363.0 / 169.0	2.24	2934.56	446.023831	40.4	false
PFHxS_1	399.0 / 80.0	2.30	1227727.21	3903.239125	394.3	false
PFHxS_2	399.0 / 99.0	2.30	336628.94	3833.121017	700.7	false
PFOA_1	413.0 / 369.0	2.67	62727.11	161.626343	85.4	false
PFOA_2	413.0 / 169.0	2.67	2506.65	94.363647	54.1	true
PFNA_1	463.0 / 419.0	3.08	6767.75	< 0	33.3	false
PFNA_2	463.0 / 219.0	3.08	1396.25	< 0	29.5	false
PFOS_1	499.0 / 80.0	2.95	130253.23	218.425286	90.8	false
PFOS_2	499.0 / 99.0	3.05	17684.95	169.818273	112.2	false
PFDA_1	513.0 / 469.0	3.43	2377.02	< 0	19.8	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1305.48	< 0	18.7	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	939.04	< 0	24.3	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:25 AM

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	4653632.38	13276.691757	289.1	false
PFBS_2	298.9 / 99.0	1.54	1367947.03	13290.879200	483.9	false
PFHxA_1	313.0 / 269.0	1.86	15383585.90	54861.985922	213.5	false
PFHxA_2	313.0 / 119.0	1.86	968870.05	46371.296136	364.0	false
PFHpA_1	363.0 / 319.0	2.27	3000555.95	10565.209094	210.8	false
PFHpA_2	363.0 / 169.0	2.24	79783.46	13215.066378	370.4	false
PFHxS_1	399.0 / 80.0	2.30	26996520.56	77129.628118	565.3	false
PFHxS_2	399.0 / 99.0	2.30	7519497.61	77051.026293	1658.2	false
PFOA_1	413.0 / 369.0	2.67	3414041.81	10760.147355	339.4	false
PFOA_2	413.0 / 169.0	2.68	129331.33	6271.196670	360.5	true
PFNA_1	463.0 / 419.0	3.08	18327.64	40.696795	33.2	false
PFNA_2	463.0 / 219.0	3.08	6844.22	59.951630	61.2	true
PFOS_1	499.0 / 80.0	2.95	2270296.34	4842.570271	166.9	false
PFOS_2	499.0 / 99.0	3.05	312543.63	3839.139953	285.9	true
PFDA_1	513.0 / 469.0	3.43	2303.40	< 0	12.2	true
PFDA_2	513.0 / 219.0	3.41	133.58	< 0	6.6	true
PFUnA_1	563.0 / 519.0	3.76	2081.17	< 0	14.6	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1213.68	< 0	14.9	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	467.40	< 0	20.4	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:29 AM

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	455402.13	1139.051134	279.5	false
PFBS_2	298.9 / 99.0	1.54	129320.60	1103.057214	277.9	false
PFHxA_1	313.0 / 269.0	1.86	1112179.51	4421.492029	89.3	false
PFHxA_2	313.0 / 119.0	1.87	71574.84	3798.959957	81.3	false
PFHpA_1	363.0 / 319.0	2.27	233512.49	969.274297	134.8	false
PFHpA_2	363.0 / 169.0	2.25	6380.68	1187.456661	129.5	false
PFHxS_1	399.0 / 80.0	2.30	2141812.82	6776.643933	510.4	false
PFHxS_2	399.0 / 99.0	2.30	585739.58	6641.674150	971.8	false
PFOA_1	413.0 / 369.0	2.67	373986.83	1086.719401	253.6	false
PFOA_2	413.0 / 169.0	2.68	14370.56	637.558931	218.9	true
PFNA_1	463.0 / 419.0	3.08	39899.73	113.673324	89.4	false
PFNA_2	463.0 / 219.0	3.08	11559.97	108.540629	136.8	false
PFOS_1	499.0 / 80.0	2.94	231846.90	407.371660	129.7	false
PFOS_2	499.0 / 99.0	3.03	27978.04	282.317697	98.9	false
PFDA_1	513.0 / 469.0	3.43	9810.20	4.193101	51.5	false
PFDA_2	513.0 / 219.0	3.45	533.28	10.159506	15.7	false
PFUnA_1	563.0 / 519.0	3.76	4389.72	1.518633	30.7	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	870.22	< 0	18.2	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	598.33	< 0	22.0	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:34 AM

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	7919499.01	23301.490896	374.6	false
PFBS_2	298.9 / 99.0	1.54	2286681.08	22911.362024	629.0	false
PFHxA_1	313.0 / 269.0	1.86	19899172.03	75423.146804	247.1	false
PFHxA_2	313.0 / 119.0	1.86	1244646.90	63319.871865	407.6	false
PFHpA_1	363.0 / 319.0	2.27	2249470.10	8824.169907	182.1	false
PFHpA_2	363.0 / 169.0	2.24	62490.99	11524.411175	281.6	false
PFHxS_1	399.0 / 80.0	2.30	13626786.81	39559.085380	422.4	false
PFHxS_2	399.0 / 99.0	2.30	3723244.95	38762.978187	1082.1	false
PFOA_1	413.0 / 369.0	2.67	6410657.52	22104.756042	440.1	false
PFOA_2	413.0 / 169.0	2.68	231729.02	12298.036011	455.9	true
PFNA_1	463.0 / 419.0	3.08	13412.35	22.635084	29.9	false
PFNA_2	463.0 / 219.0	3.08	4158.88	26.727733	50.3	false
PFOS_1	499.0 / 80.0	2.95	1439732.66	3015.986796	176.7	false
PFOS_2	499.0 / 99.0	3.03	176703.68	2131.380394	218.8	true
PFDA_1	513.0 / 469.0	3.43	1273.93	< 0	10.0	true
PFDA_2	513.0 / 219.0	3.43	225.45	< 0	11.5	false
PFUnA_1	563.0 / 519.0	3.76	3249.49	0.405844	18.2	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.00	1083.26	< 0	14.9	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	1055.55	< 0	34.1	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:39 AM

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	832781.06	1918.919195	365.3	false
PFBS_2	298.9 / 99.0	1.54	252030.93	1979.470118	426.5	false
PFHxA_1	313.0 / 269.0	1.86	1527858.37	6309.245132	106.2	false
PFHxA_2	313.0 / 119.0	1.87	100629.75	5558.583149	124.0	false
PFHpA_1	363.0 / 319.0	2.27	186422.39	690.620950	114.9	false
PFHpA_2	363.0 / 169.0	2.25	3762.11	587.754610	51.1	false
PFHxS_1	399.0 / 80.0	2.30	1206031.61	3763.703893	368.2	false
PFHxS_2	399.0 / 99.0	2.30	323614.13	3616.377717	591.4	false
PFOA_1	413.0 / 369.0	2.67	660549.15	1816.153885	281.7	false
PFOA_2	413.0 / 169.0	2.67	28704.38	1211.010385	216.4	true
PFNA_1	463.0 / 419.0	3.08	17731.51	28.980945	55.1	false
PFNA_2	463.0 / 219.0	3.08	6169.93	40.680657	99.0	false
PFOS_1	499.0 / 80.0	2.95	146275.86	272.494510	129.5	false
PFOS_2	499.0 / 99.0	3.03	18633.54	199.027048	99.6	true
PFDA_1	513.0 / 469.0	3.43	4923.37	< 0	29.4	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.75	3850.34	0.825027	26.9	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.05	1187.82	< 0	17.5	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	1052.38	< 0	29.1	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:56 PM

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	2.27	6135.52	8.111438	12.1	false
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	N/A	N/A	N/A	N/A	true
PFHxS_2	399.0 / 99.0	N/A	N/A	N/A	N/A	true
PFOA_1	413.0 / 369.0	2.68	181942.92	536.801198	192.5	false
PFOA_2	413.0 / 169.0	2.68	10848.11	492.347847	156.8	false
PFNA_1	463.0 / 419.0	3.08	3156.06	< 0	21.1	false
PFNA_2	463.0 / 219.0	3.08	1073.84	< 0	24.8	false
PFOS_1	499.0 / 80.0	N/A	N/A	N/A	N/A	true
PFOS_2	499.0 / 99.0	3.10	796.33	9.680405	11.3	true
PFDA_1	513.0 / 469.0	3.42	1259.48	< 0	11.8	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.75	1099.19	< 0	15.6	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.04	1197.00	< 0	21.9	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.26	712.71	< 0	26.3	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:45:01 PM

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	2556.34	2.529745	15.7	true
PFBS_2	298.9 / 99.0	1.54	1079.48	7.924707	18.2	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	2.30	9468.09	9.276577	48.6	false
PFHxS_2	399.0 / 99.0	2.29	2681.40	4.527854	31.3	false
PFOA_1	413.0 / 369.0	2.68	101944.34	316.015295	142.1	false
PFOA_2	413.0 / 169.0	2.68	6723.36	321.906647	106.4	false
PFNA_1	463.0 / 419.0	3.07	3284.21	< 0	22.8	false
PFNA_2	463.0 / 219.0	3.09	700.67	< 0	13.5	false
PFOS_1	499.0 / 80.0	3.07	35276.19	69.195210	55.4	false
PFOS_2	499.0 / 99.0	3.08	6334.50	70.110333	64.0	false
PFDA_1	513.0 / 469.0	3.45	2209.43	< 0	18.1	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.76	2170.77	< 0	19.9	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	1931.30	< 0	27.1	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	1388.11	< 0	41.2	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:43 AM

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.53	1677936.12	7428.700562	114.2	true
PFBS_2	298.9 / 99.0	1.53	497758.20	7505.716603	193.2	false
PFHxA_1	313.0 / 269.0	1.85	3777160.10	15969.484469	93.4	false
PFHxA_2	313.0 / 119.0	1.85	232811.99	13193.393130	150.1	false
PFHpA_1	363.0 / 319.0	2.26	1369682.30	5068.794370	115.3	false
PFHpA_2	363.0 / 169.0	2.24	33444.47	5783.737810	271.2	false
PFHxS_1	399.0 / 80.0	2.29	7584263.88	23485.026301	286.5	false
PFHxS_2	399.0 / 99.0	2.29	2045924.45	22717.511157	664.0	false
PFOA_1	413.0 / 369.0	2.67	19986565.96	68049.992797	588.3	false
PFOA_2	413.0 / 169.0	2.67	1044918.92	54782.059573	629.0	true
PFNA_1	463.0 / 419.0	3.04	163107.86	680.447663	81.6	false
PFNA_2	463.0 / 219.0	3.06	28411.87	373.626912	131.3	true
PFOS_1	499.0 / 80.0	3.05	8352888.31	19096.316309	245.6	false
PFOS_2	499.0 / 99.0	3.07	1339749.80	17642.171538	690.7	false
PFDA_1	513.0 / 469.0	3.43	21545.29	41.927792	38.6	false
PFDA_2	513.0 / 219.0	3.44	1028.84	49.771518	23.0	false
PFUnA_1	563.0 / 519.0	3.75	6125.30	8.333942	29.9	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	2309.39	< 0	22.4	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	1109.24	< 0	29.9	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	3.74	1355.63	14.313201	39.8	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:32:48 AM

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	333499.07	723.510112	141.6	false
PFBS_2	298.9 / 99.0	1.54	91254.28	675.576862	185.7	false
PFHxA_1	313.0 / 269.0	1.86	429325.28	1565.323559	39.8	false
PFHxA_2	313.0 / 119.0	1.86	26664.46	1282.684354	43.7	false
PFHpA_1	363.0 / 319.0	2.27	139786.26	478.064187	60.0	false
PFHpA_2	363.0 / 169.0	2.27	3419.68	485.671115	66.1	false
PFHxS_1	399.0 / 80.0	2.30	694027.96	2130.603953	333.4	false
PFHxS_2	399.0 / 99.0	2.30	187332.44	2056.886323	547.1	false
PFOA_1	413.0 / 369.0	2.68	2148183.63	5951.557203	410.9	false
PFOA_2	413.0 / 169.0	2.67	111812.24	4767.877983	311.9	true
PFNA_1	463.0 / 419.0	3.08	56997.16	178.375324	101.7	false
PFNA_2	463.0 / 219.0	3.08	17798.43	184.702546	144.2	false
PFOS_1	499.0 / 80.0	3.06	962875.53	1735.791752	211.0	false
PFOS_2	499.0 / 99.0	3.08	155254.56	1610.923664	500.4	false
PFDA_1	513.0 / 469.0	3.44	12711.41	10.894269	50.3	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.75	5143.94	3.548213	35.2	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1635.57	< 0	24.9	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	1240.19	< 0	40.1	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:45:12 PM

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.53	3195276.79	14029.436070	104.3	true
PFBS_2	298.9 / 99.0	1.53	901118.67	13473.866252	231.4	false
PFHxA_1	313.0 / 269.0	1.85	9400207.79	46500.746976	116.8	false
PFHxA_2	313.0 / 119.0	1.85	625403.44	41518.819620	245.3	false
PFHpA_1	363.0 / 319.0	2.27	8271127.14	38174.166428	208.5	false
PFHpA_2	363.0 / 169.0	2.26	160703.48	34991.277191	407.1	false
PFHxS_1	399.0 / 80.0	2.29	58072412.87	198733.437808	513.4	false
PFHxS_2	399.0 / 99.0	2.29	18170717.92	223035.218093	1132.4	false
PFOA_1	413.0 / 369.0	2.68	30186724.92	126497.658789	470.3	false
PFOA_2	413.0 / 169.0	2.65	2830394.08	182647.860052	768.0	false
PFNA_1	463.0 / 419.0	3.08	610576.60	2687.199478	128.0	false
PFNA_2	463.0 / 219.0	3.07	204717.79	2926.050081	300.4	false
PFOS_1	499.0 / 80.0	3.06	45299972.81	139132.470112	397.8	false
PFOS_2	499.0 / 99.0	3.08	7285201.06	128888.565191	954.6	false
PFDA_1	513.0 / 469.0	3.43	104632.63	296.271828	70.5	false
PFDA_2	513.0 / 219.0	3.42	3299.19	219.454682	49.1	false
PFUnA_1	563.0 / 519.0	3.75	4771.62	4.581210	19.4	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	1211.67	< 0	17.5	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	1693.53	< 0	34.3	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:45:16 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	351341.26	708.329297	130.3	true
PFBS_2	298.9 / 99.0	1.56	120889.52	832.695193	180.8	false
PFHxA_1	313.0 / 269.0	1.88	792142.62	2524.937279	34.1	false
PFHxA_2	313.0 / 119.0	1.89	56988.52	2420.803712	45.9	false
PFHpA_1	363.0 / 319.0	2.30	589220.46	1888.618209	98.4	false
PFHpA_2	363.0 / 169.0	2.30	15218.23	2239.990336	151.7	false
PFHxS_1	399.0 / 80.0	2.32	4282543.17	11833.313369	438.2	false
PFHxS_2	399.0 / 99.0	2.32	1188751.51	11776.219771	698.4	false
PFOA_1	413.0 / 369.0	2.71	2186849.05	6300.443414	374.1	false
PFOA_2	413.0 / 169.0	2.68	211126.07	9374.620217	493.9	false
PFNA_1	463.0 / 419.0	3.11	58487.31	145.328332	65.6	false
PFNA_2	463.0 / 219.0	3.11	19902.42	167.430565	99.9	false
PFOS_1	499.0 / 80.0	3.10	3346089.71	5696.247655	294.2	false
PFOS_2	499.0 / 99.0	3.11	530839.60	5204.236954	568.4	false
PFDA_1	513.0 / 469.0	3.46	10126.91	2.880484	28.6	false
PFDA_2	513.0 / 219.0	3.49	1410.82	58.826762	19.2	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.07	2028.85	< 0	22.4	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.31	704.27	< 0	27.2	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 2:52:42 PM

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	3149899.41	14286.644720	127.6	true
PFBS_2	298.9 / 99.0	1.53	890300.58	13751.435841	223.9	false
PFHxA_1	313.0 / 269.0	1.85	9452992.63	49421.145096	108.7	false
PFHxA_2	313.0 / 119.0	1.86	612345.21	42963.661191	248.5	false
PFHpA_1	363.0 / 319.0	2.27	7821364.14	34194.030141	182.7	false
PFHpA_2	363.0 / 169.0	2.27	156327.17	32237.605467	441.6	false
PFHxS_1	399.0 / 80.0	2.30	54187884.94	184438.445481	472.8	false
PFHxS_2	399.0 / 99.0	2.29	16866497.88	205908.255126	1408.5	false
PFOA_1	413.0 / 369.0	2.69	28923735.76	126990.325662	465.6	false
PFOA_2	413.0 / 169.0	2.65	2712500.13	183394.854988	964.5	false
PFNA_1	463.0 / 419.0	3.09	576549.35	2710.095670	125.7	false
PFNA_2	463.0 / 219.0	3.08	193032.78	2946.667855	316.3	false
PFOS_1	499.0 / 80.0	3.07	41625953.94	117726.561042	397.8	false
PFOS_2	499.0 / 99.0	3.09	6915755.28	112665.601771	1041.6	false
PFDA_1	513.0 / 469.0	3.44	98127.64	272.282584	80.4	false
PFDA_2	513.0 / 219.0	3.43	7529.63	523.520679	105.1	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1508.53	< 0	14.4	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.29	1358.72	< 0	25.1	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 2:52:45 PM

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	392755.49	809.270970	142.2	true
PFBS_2	298.9 / 99.0	1.55	118295.67	831.969756	200.3	false
PFHxA_1	313.0 / 269.0	1.87	926289.55	2976.428902	52.8	false
PFHxA_2	313.0 / 119.0	1.87	65830.05	2821.615753	71.4	false
PFHpA_1	363.0 / 319.0	2.28	659569.11	2125.015086	142.4	false
PFHpA_2	363.0 / 169.0	2.28	13742.78	2022.620995	202.4	false
PFHxS_1	399.0 / 80.0	2.30	4445641.24	11439.334235	529.2	false
PFHxS_2	399.0 / 99.0	2.30	1256967.42	11596.086786	1001.9	false
PFOA_1	413.0 / 369.0	2.69	2350658.04	6693.722284	428.7	false
PFOA_2	413.0 / 169.0	2.66	213981.14	9389.729171	333.9	false
PFNA_1	463.0 / 419.0	3.09	56902.72	151.590008	90.1	false
PFNA_2	463.0 / 219.0	3.09	19661.34	177.464138	153.8	false
PFOS_1	499.0 / 80.0	3.08	3539145.58	6310.854718	314.2	false
PFOS_2	499.0 / 99.0	3.08	597586.55	6136.792104	660.3	false
PFDA_1	513.0 / 469.0	3.44	12947.87	9.258984	43.9	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.76	3059.15	< 0	25.4	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.04	2453.37	< 0	31.4	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.30	952.18	< 0	25.2	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.53	5557235.71	25086.986843	157.6	true
PFBS_2	298.9 / 99.0	1.53	1650031.43	25365.337367	314.7	false
PFHxA_1	313.0 / 269.0	1.85	7662709.09	37081.566168	127.7	false
PFHxA_2	313.0 / 119.0	1.86	477302.76	30990.712518	216.3	false
PFHpA_1	363.0 / 319.0	2.27	1348588.50	4939.557023	122.8	false
PFHpA_2	363.0 / 169.0	2.23	35489.35	6079.545229	184.2	false
PFHxS_1	399.0 / 80.0	2.28	5629483.43	18224.451463	197.9	false
PFHxS_2	399.0 / 99.0	2.30	1502486.08	17440.175955	456.6	false
PFOA_1	413.0 / 369.0	2.69	14547956.03	51044.320502	456.6	false
PFOA_2	413.0 / 169.0	2.64	1395898.49	75427.315761	661.7	true
PFNA_1	463.0 / 419.0	3.00	46947.05	167.255240	45.1	false
PFNA_2	463.0 / 219.0	2.97	15360.26	183.115725	75.2	true
PFOS_1	499.0 / 80.0	3.02	259213.51	561.396669	38.4	false
PFOS_2	499.0 / 99.0	3.08	35607.01	443.187660	110.8	true
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	925.42	< 0	12.6	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	774.32	< 0	17.4	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	1104981.08	2503.172825	222.7	true
PFBS_2	298.9 / 99.0	1.54	335770.72	2592.072793	267.7	false
PFHxA_1	313.0 / 269.0	1.86	1094117.72	3427.981507	58.8	false
PFHxA_2	313.0 / 119.0	1.86	72100.57	3012.184080	82.3	false
PFHpA_1	363.0 / 319.0	2.27	137851.06	408.374005	57.2	false
PFHpA_2	363.0 / 169.0	2.24	5480.98	714.310735	67.8	false
PFHxS_1	399.0 / 80.0	2.27	597749.77	1652.240906	193.2	false
PFHxS_2	399.0 / 99.0	2.30	159519.60	1575.442771	274.6	false
PFOA_1	413.0 / 369.0	2.68	1644037.33	4298.468902	326.3	false
PFOA_2	413.0 / 169.0	2.64	157595.18	6352.908526	261.1	false
PFNA_1	463.0 / 419.0	3.07	9846.18	< 0	33.0	false
PFNA_2	463.0 / 219.0	3.05	3546.21	8.398866	31.0	false
PFOS_1	499.0 / 80.0	3.00	32005.57	60.056801	25.8	false
PFOS_2	499.0 / 99.0	3.08	4832.92	51.069855	46.0	false
PFDA_1	513.0 / 469.0	3.44	3330.04	< 0	16.6	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.74	2326.63	< 0	18.2	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	870.77	< 0	12.6	false
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	651.98	< 0	23.0	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:45:35 PM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.53	6693612.63	29667.496838	158.2	true
PFBS_2	298.9 / 99.0	1.53	1909709.05	28822.990337	361.2	false
PFHxA_1	313.0 / 269.0	1.85	16707792.09	91471.915223	155.5	false
PFHxA_2	313.0 / 119.0	1.85	1041505.34	76538.597941	319.4	false
PFHpA_1	363.0 / 319.0	2.26	2669879.88	12542.343500	145.8	false
PFHpA_2	363.0 / 169.0	2.23	71395.29	15790.236308	297.1	false
PFHxS_1	399.0 / 80.0	2.29	97138522.22	318478.739105	531.8	false
PFHxS_2	399.0 / 99.0	2.29	43245608.53	508560.245462	1199.3	false
PFOA_1	413.0 / 369.0	2.68	13827332.76	54131.758372	387.7	false
PFOA_2	413.0 / 169.0	2.62	1405927.10	84763.265395	586.6	false
PFNA_1	463.0 / 419.0	3.02	109926.23	502.249030	62.3	false
PFNA_2	463.0 / 219.0	2.97	42665.22	643.904152	95.4	false
PFOS_1	499.0 / 80.0	2.99	151804795.13	444096.669566	476.7	false
PFOS_2	499.0 / 99.0	3.08	27322153.67	460417.841403	992.3	true
PFDA_1	513.0 / 469.0	3.39	16356.02	24.041409	21.5	false
PFDA_2	513.0 / 219.0	3.44	727.04	25.383265	33.4	true
PFUnA_1	563.0 / 519.0	3.76	5114.38	5.197873	20.3	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1295.67	< 0	18.6	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.27	1293.98	< 0	28.3	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	3.76	472.43	< 0	17.0	false
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:45:39 PM

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	546453.57	1152.086378	245.4	true
PFBS_2	298.9 / 99.0	1.54	161730.67	1162.940696	260.1	false
PFHxA_1	313.0 / 269.0	1.86	957747.20	3337.320652	66.4	false
PFHxA_2	313.0 / 119.0	1.87	64428.74	2994.103533	77.1	false
PFHpA_1	363.0 / 319.0	2.27	143456.89	471.966218	81.6	false
PFHpA_2	363.0 / 169.0	2.21	4964.42	714.139064	54.7	false
PFHxS_1	399.0 / 80.0	2.30	7680946.54	23936.080455	1040.9	false
PFHxS_2	399.0 / 99.0	2.30	2199736.61	24582.848549	1545.7	false
PFOA_1	413.0 / 369.0	2.68	702136.77	1882.518374	257.9	false
PFOA_2	413.0 / 169.0	2.62	74590.21	3089.894128	279.9	true
PFNA_1	463.0 / 419.0	3.07	12079.11	8.795006	43.3	false
PFNA_2	463.0 / 219.0	3.07	2881.65	3.197331	45.4	false
PFOS_1	499.0 / 80.0	3.00	10193946.85	19894.046802	385.3	false
PFOS_2	499.0 / 99.0	3.08	1510766.05	16982.256900	1134.5	true
PFDA_1	513.0 / 469.0	3.43	3197.30	< 0	14.4	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	3.75	1651.69	< 0	16.4	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.02	816.27	< 0	17.6	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.28	548.60	< 0	20.9	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	3.59	191.76	< 0	266.4	false
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID	VC-S14GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	2.29	83388.81	442.856715	22.4	true
PFHxS_2	399.0 / 99.0	2.29	21901.08	410.804741	45.4	true
PFOA_1	413.0 / 369.0	2.68	181912.52	1240.914050	49.5	true
PFOA_2	413.0 / 169.0	2.64	15817.86	1667.414261	79.9	true
PFNA_1	463.0 / 419.0	N/A	N/A	N/A	N/A	true
PFNA_2	463.0 / 219.0	N/A	N/A	N/A	N/A	true
PFOS_1	499.0 / 80.0	3.03	163762.87	613.662707	35.6	false
PFOS_2	499.0 / 99.0	3.08	25540.25	550.077504	90.7	false
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	3.44	415.70	27.578612	14.9	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:47:18 PM

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	3132.50	2.055020	21.8	true
PFBS_2	298.9 / 99.0	1.55	1266.82	6.778415	25.0	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	N/A	N/A	N/A	N/A	true
PFHxS_2	399.0 / 99.0	N/A	N/A	N/A	N/A	true
PFOA_1	413.0 / 369.0	2.72	2364.24	< 0	10.5	true
PFOA_2	413.0 / 169.0	2.72	695.71	12.515214	18.8	true
PFNA_1	463.0 / 419.0	N/A	N/A	N/A	N/A	true
PFNA_2	463.0 / 219.0	N/A	N/A	N/A	N/A	true
PFOS_1	499.0 / 80.0	3.09	6672.20	11.773449	14.1	false
PFOS_2	499.0 / 99.0	N/A	N/A	N/A	N/A	true
PFDA_1	513.0 / 469.0	3.48	1658.41	< 0	13.3	false
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.07	929.07	< 0	17.6	true
PFDoA_2	613.0 / 319.0	4.08	133.37	< 0	5.0	true
PFTrDA_1	663.0 / 619.0	4.34	1084.47	< 0	41.9	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	4.54	987.13	< 0	54.7	false
PFTeDA_2	713.0 / 169.0	4.55	116.03	< 0	10.8	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:47:22 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.56	2802211.96	7357.560978	971.7	false
PFBS_2	298.9 / 99.0	1.55	820778.50	7339.955964	845.2	false
PFHxA_1	313.0 / 269.0	1.88	2083886.90	8834.728933	191.3	false
PFHxA_2	313.0 / 119.0	1.88	146769.87	8335.876988	132.0	false
PFHpA_1	363.0 / 319.0	2.30	1837829.84	8134.488931	412.1	false
PFHpA_2	363.0 / 169.0	2.30	37859.70	7851.041692	295.1	false
PFHxS_1	399.0 / 80.0	2.32	2584599.60	7984.878228	752.1	false
PFHxS_2	399.0 / 99.0	2.32	781856.57	8660.283515	1238.8	false
PFOA_1	413.0 / 369.0	2.71	2528289.38	8534.901110	720.6	false
PFOA_2	413.0 / 169.0	2.71	159402.69	8286.407687	604.9	false
PFNA_1	463.0 / 419.0	3.11	2478682.75	8324.934510	1065.7	false
PFNA_2	463.0 / 219.0	3.11	795937.37	8667.104490	1330.0	false
PFOS_1	499.0 / 80.0	3.11	3957209.28	7476.135549	500.4	false
PFOS_2	499.0 / 99.0	3.11	688242.55	7488.498663	704.3	false
PFDA_1	513.0 / 469.0	3.47	2919015.35	8021.356130	897.7	false
PFDA_2	513.0 / 219.0	3.47	118583.00	7907.840759	506.0	false
PFUnA_1	563.0 / 519.0	3.80	2935159.52	8790.954297	881.0	false
PFUnA_2	563.0 / 269.0	3.79	139626.07	8409.394746	459.4	false
PFDoA_1	613.0 / 569.0	4.07	3030400.01	8147.586847	775.6	false
PFDoA_2	613.0 / 319.0	4.07	483858.81	8351.719646	644.7	false
PFTrDA_1	663.0 / 619.0	4.32	2671935.87	8398.273701	1130.9	false
PFTrDA_2	663.0 / 169.0	4.32	180446.97	8632.914895	752.5	false
PFTeDA_1	713.0 / 669.0	4.54	3114757.09	8642.183882	2192.9	false
PFTeDA_2	713.0 / 169.0	4.53	155161.51	8894.154607	1269.6	false
NMeFOSAA_1	570.0 / 419.0	3.63	524718.05	9678.153304	1838.0	false
NMeFOSAA_2	570.0 / 512.0	3.62	269424.58	9168.876742	971.8	false
NEtFOSAA_1	584.0 / 419.0	3.79	466875.44	8201.027537	973.5	false
NEtFOSAA_2	584.0 / 483.0	3.79	32526.12	9782.684787	814.8	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:47:26 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.55	3353246.29	6559.445088	426.5	false
PFBS_2	298.9 / 99.0	1.55	1003579.87	6686.620167	572.5	false
PFHxA_1	313.0 / 269.0	1.87	1321995.88	4052.416011	87.8	false
PFHxA_2	313.0 / 119.0	1.88	86366.80	3533.543143	92.2	false
PFHpA_1	363.0 / 319.0	2.29	222525.12	623.663728	114.3	false
PFHpA_2	363.0 / 169.0	2.27	5469.41	657.556510	68.3	false
PFHxS_1	399.0 / 80.0	2.32	5128871.55	12473.608323	628.2	false
PFHxS_2	399.0 / 99.0	2.32	1465332.88	12777.679531	1076.6	false
PFOA_1	413.0 / 369.0	2.70	344003.02	823.633026	194.9	false
PFOA_2	413.0 / 169.0	2.69	18716.22	688.316928	164.8	false
PFNA_1	463.0 / 419.0	3.11	29763.09	50.855327	51.3	false
PFNA_2	463.0 / 219.0	3.11	9682.18	59.206766	94.5	false
PFOS_1	499.0 / 80.0	2.98	1040896.70	1566.094392	235.0	false
PFOS_2	499.0 / 99.0	3.03	122751.83	1063.032921	220.6	false
PFDA_1	513.0 / 469.0	3.47	12886.82	5.885652	46.0	false
PFDA_2	513.0 / 219.0	3.50	674.50	11.047735	17.1	false
PFUnA_1	563.0 / 519.0	3.79	9977.09	10.899885	40.7	false
PFUnA_2	563.0 / 269.0	3.80	447.62	19.011891	10.4	false
PFDoA_1	613.0 / 569.0	4.07	7714.79	< 0	68.8	false
PFDoA_2	613.0 / 319.0	4.07	939.42	< 0	30.8	false
PFTrDA_1	663.0 / 619.0	4.32	6351.73	< 0	125.1	false
PFTrDA_2	663.0 / 169.0	4.32	619.63	< 0	31.4	true
PFTeDA_1	713.0 / 669.0	4.53	7511.63	< 0	106.5	false
PFTeDA_2	713.0 / 169.0	4.54	490.20	< 0	45.7	false
NMeFOSAA_1	570.0 / 419.0	3.62	1889.30	< 0	57.2	false
NMeFOSAA_2	570.0 / 512.0	3.63	600.06	< 0	15.6	false
NEtFOSAA_1	584.0 / 419.0	3.78	1467.81	6.856760	49.1	false
NEtFOSAA_2	584.0 / 483.0	3.78	88.51	< 0	24.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:47:31 PM

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	N/A	N/A	N/A	N/A	true
PFBS_2	298.9 / 99.0	N/A	N/A	N/A	N/A	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	2.31	1423912.96	3956.663262	607.4	false
PFHxS_2	399.0 / 99.0	2.31	398345.26	3965.042765	874.9	false
PFOA_1	413.0 / 369.0	N/A	N/A	N/A	N/A	true
PFOA_2	413.0 / 169.0	N/A	N/A	N/A	N/A	true
PFNA_1	463.0 / 419.0	N/A	N/A	N/A	N/A	true
PFNA_2	463.0 / 219.0	N/A	N/A	N/A	N/A	true
PFOS_1	499.0 / 80.0	N/A	N/A	N/A	N/A	true
PFOS_2	499.0 / 99.0	N/A	N/A	N/A	N/A	true
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:02 PM

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.07	91698.47	244.195426	1281.1	false
d3-MeFOSAA	573.0 / 419.0	3.61	16302.89	260.004961	195.5	false
d5-EtFOSAA	589.0 / 419.0	3.78	17051.45	239.580705	212.6	false
13C5-PFHxA	318.0 / 273.0	1.87	61987.66	244.293373	826.7	false
13C4-PFHpA	367.0 / 322.0	2.28	70438.43	244.000657	1152.0	false
13C8-PFOA	421.0 / 376.0	2.70	88535.23	252.114367	1604.4	false
13C9-PFNA	472.0 / 427.0	3.09	100138.19	249.732890	1149.5	false
13C6-PFDA	519.0 / 474.0	3.45	97258.51	257.077741	4005.3	false
13C7-PFUnA	570.0 / 525.0	3.78	88898.52	254.380465	800.7	false
13C2-PFTeDA	715.0 / 670.0	4.54	70451.22	232.374509	2131.0	false
13C3-PFBS	302.0 / 99.0	1.55	25201.92	220.626542	798.1	false
13C3-PFHxS	402.0 / 99.0	2.31	25094.63	242.439381	606.8	false
13C8-PFOS	507.0 / 99.0	3.09	26831.98	229.467599	276.5	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 01/11/2018 6:22:18 AM

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.13	97310.49	273.295459	980.3	false
d3-MeFOSAA	573.0 / 419.0	3.67	16307.83	281.326372	235.0	false
d5-EtFOSAA	589.0 / 419.0	3.83	16106.82	244.792248	231.9	false
13C5-PFHxA	318.0 / 273.0	1.90	58984.04	281.414077	800.8	false
13C4-PFHpA	367.0 / 322.0	2.32	65679.00	275.430947	858.6	false
13C8-PFOA	421.0 / 376.0	2.74	77309.52	266.513537	15903.8	false
13C9-PFNA	472.0 / 427.0	3.14	83779.99	252.942124	2153.8	false
13C6-PFDA	519.0 / 474.0	3.51	92389.14	257.546154	881.9	false
13C7-PFUnA	570.0 / 525.0	3.83	86079.46	259.768255	929.9	false
13C2-PFTeDA	715.0 / 670.0	4.61	88378.85	307.429555	1950.9	false
13C3-PFBS	302.0 / 99.0	1.56	23436.89	221.932706	376.1	false
13C3-PFHxS	402.0 / 99.0	2.35	21461.26	224.271849	408.3	false
13C8-PFOS	507.0 / 99.0	3.14	25206.28	233.171040	198.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:07 PM

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	87903.18	220.007997	1158.4	false
d3-MeFOSAA	573.0 / 419.0	3.58	14467.94	227.619110	225.9	false
d5-EtFOSAA	589.0 / 419.0	3.74	13557.74	187.915611	154.5	false
13C5-PFHxA	318.0 / 273.0	1.85	55762.03	229.155805	590.8	false
13C4-PFHpA	367.0 / 322.0	2.27	66522.02	240.288275	856.6	false
13C8-PFOA	421.0 / 376.0	2.68	77976.53	231.542732	928.5	false
13C9-PFNA	472.0 / 427.0	3.07	84602.02	220.010047	1634.6	false
13C6-PFDA	519.0 / 474.0	3.42	90666.51	225.238265	13937.4	false
13C7-PFUnA	570.0 / 525.0	3.74	80182.41	215.638732	815.2	false
13C2-PFTeDA	715.0 / 670.0	4.49	75890.10	235.257515	2121.3	false
13C3-PFBS	302.0 / 99.0	1.53	25188.47	217.525858	418.7	false
13C3-PFHxS	402.0 / 99.0	2.29	21758.45	207.364955	349.9	false
13C8-PFOS	507.0 / 99.0	3.07	23264.61	196.267958	234.7	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:15 PM

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	81470.47	216.658664	904.9	false
d3-MeFOSAA	573.0 / 419.0	3.58	11851.76	223.322630	137.2	false
d5-EtFOSAA	589.0 / 419.0	3.74	12327.91	204.650493	161.8	false
13C5-PFHxA	318.0 / 273.0	1.85	36542.29	46.409789	103.5	false
13C4-PFHpA	367.0 / 322.0	2.26	48561.65	54.210337	328.7	false
13C8-PFOA	421.0 / 376.0	2.68	34635.88	31.784477	352.6	false
13C9-PFNA	472.0 / 427.0	3.07	68324.46	54.911021	441.4	false
13C6-PFDA	519.0 / 474.0	3.42	79988.20	211.136411	708.7	false
13C7-PFUnA	570.0 / 525.0	3.74	81277.30	232.251700	869.0	false
13C2-PFTeDA	715.0 / 670.0	4.48	64949.82	213.933237	2015.1	false
13C3-PFBS	302.0 / 99.0	1.52	21726.38	224.721342	209.3	false
13C3-PFHxS	402.0 / 99.0	2.28	16501.04	188.350431	170.9	false
13C8-PFOS	507.0 / 99.0	3.07	19636.29	198.408868	156.4	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:19 PM

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	102365.92	257.090187	1352.0	false
d3-MeFOSAA	573.0 / 419.0	3.58	14001.31	233.809265	174.4	false
d5-EtFOSAA	589.0 / 419.0	3.74	14655.80	215.613658	230.7	false
13C5-PFHxA	318.0 / 273.0	1.85	62523.36	193.999638	284.9	false
13C4-PFHpA	367.0 / 322.0	2.26	73739.89	201.111069	881.2	false
13C8-PFOA	421.0 / 376.0	2.68	69511.85	155.844857	1394.9	false
13C9-PFNA	472.0 / 427.0	3.07	89868.53	176.455732	848.7	false
13C6-PFDA	519.0 / 474.0	3.42	99308.03	247.557402	1087.5	false
13C7-PFUnA	570.0 / 525.0	3.74	93211.96	251.544930	942.9	false
13C2-PFTeDA	715.0 / 670.0	4.48	87922.52	273.498386	1552.5	false
13C3-PFBS	302.0 / 99.0	1.52	33656.28	308.507861	410.2	false
13C3-PFHxS	402.0 / 99.0	2.29	22981.63	232.476574	272.1	false
13C8-PFOS	507.0 / 99.0	3.06	28154.99	252.115793	219.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:23 PM

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	108169.44	277.937733	1135.6	false
d3-MeFOSAA	573.0 / 419.0	3.58	16998.02	259.329253	213.4	false
d5-EtFOSAA	589.0 / 419.0	3.74	16081.73	216.152227	193.6	false
13C5-PFHxA	318.0 / 273.0	1.85	63253.79	244.564522	590.3	false
13C4-PFHpA	367.0 / 322.0	2.26	80959.19	275.136351	1077.5	false
13C8-PFOA	421.0 / 376.0	2.68	80568.43	225.085130	1244.0	false
13C9-PFNA	472.0 / 427.0	3.07	89621.40	219.274521	822.8	false
13C6-PFDA	519.0 / 474.0	3.42	108190.24	275.925890	825.7	false
13C7-PFUnA	570.0 / 525.0	3.74	101183.17	279.360498	1137.4	false
13C2-PFTeDA	715.0 / 670.0	4.48	93296.57	296.915683	2179.1	false
13C3-PFBS	302.0 / 99.0	1.52	33425.41	279.922060	478.7	false
13C3-PFHxS	402.0 / 99.0	2.29	24903.92	230.158185	331.6	false
13C8-PFOS	507.0 / 99.0	3.06	30511.70	249.615411	283.7	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:27 PM

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	98187.64	285.417050	1622.9	false
d3-MeFOSAA	573.0 / 419.0	3.58	15988.37	243.125225	243.5	false
d5-EtFOSAA	589.0 / 419.0	3.74	15969.71	213.942181	251.9	false
13C5-PFHxA	318.0 / 273.0	1.85	62304.44	272.465161	611.8	false
13C4-PFHpA	367.0 / 322.0	2.27	70703.83	271.775177	926.9	false
13C8-PFOA	421.0 / 376.0	2.68	86239.74	272.504948	1741.4	false
13C9-PFNA	472.0 / 427.0	3.06	95741.06	264.947443	894.1	false
13C6-PFDA	519.0 / 474.0	3.42	101696.33	293.420063	1512.0	false
13C7-PFUnA	570.0 / 525.0	3.74	100969.04	315.373463	913.7	false
13C2-PFTeDA	715.0 / 670.0	4.48	84991.30	306.000460	1681.2	false
13C3-PFBS	302.0 / 99.0	1.53	28724.76	239.767039	416.1	false
13C3-PFHxS	402.0 / 99.0	2.29	23990.45	220.988511	366.0	false
13C8-PFOS	507.0 / 99.0	3.06	31999.59	260.928789	276.4	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:31 PM

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	77356.10	214.992598	1529.8	false
d3-MeFOSAA	573.0 / 419.0	3.58	11604.10	188.360074	237.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	14434.41	206.419207	192.2	false
13C5-PFHxA	318.0 / 273.0	1.85	51377.14	212.712192	528.0	false
13C4-PFHpA	367.0 / 322.0	2.27	56998.41	207.424489	942.2	false
13C8-PFOA	421.0 / 376.0	2.67	72105.27	215.707077	9887.9	false
13C9-PFNA	472.0 / 427.0	3.07	70178.02	183.862440	507942.4	false
13C6-PFDA	519.0 / 474.0	3.42	84820.33	233.986233	1419.3	false
13C7-PFUnA	570.0 / 525.0	3.74	80003.86	238.920658	887.8	false
13C2-PFTeDA	715.0 / 670.0	4.48	67578.35	232.627508	1155.2	false
13C3-PFBS	302.0 / 99.0	1.52	23306.08	207.660443	394.7	false
13C3-PFHxS	402.0 / 99.0	2.29	21717.79	213.549348	400.0	false
13C8-PFOS	507.0 / 99.0	3.07	24228.85	210.892886	211.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:36 PM

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	82663.20	211.900624	1010.6	false
d3-MeFOSAA	573.0 / 419.0	3.58	12386.12	177.257779	197.3	false
d5-EtFOSAA	589.0 / 419.0	3.74	11551.88	145.645259	192.9	false
13C5-PFHxA	318.0 / 273.0	1.85	54899.59	238.339078	658.7	false
13C4-PFHpA	367.0 / 322.0	2.26	54502.65	207.978603	668.8	false
13C8-PFOA	421.0 / 376.0	2.67	72578.99	227.673271	2803.1	false
13C9-PFNA	472.0 / 427.0	3.06	74600.73	204.945682	847.4	false
13C6-PFDA	519.0 / 474.0	3.42	82984.85	211.144737	1125.1	false
13C7-PFUnA	570.0 / 525.0	3.74	78986.29	217.563248	763.4	false
13C2-PFTeDA	715.0 / 670.0	4.48	66887.03	212.366708	1977.9	false
13C3-PFBS	302.0 / 99.0	1.53	21606.32	169.729667	481.3	false
13C3-PFHxS	402.0 / 99.0	2.28	21199.00	183.776774	331.7	false
13C8-PFOS	507.0 / 99.0	3.06	23923.69	183.590340	199.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:40 PM

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	88560.66	233.064548	998.7	false
d3-MeFOSAA	573.0 / 419.0	3.58	14112.06	246.897265	157.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	16493.09	254.215088	159.7	false
13C5-PFHxA	318.0 / 273.0	1.85	61253.99	241.250318	137.0	false
13C4-PFHpA	367.0 / 322.0	2.26	71266.86	246.715266	233.8	false
13C8-PFOA	421.0 / 376.0	2.68	69415.90	197.545622	336.7	false
13C9-PFNA	472.0 / 427.0	3.07	73829.59	184.006615	306.4	false
13C6-PFDA	519.0 / 474.0	3.42	86045.59	224.763298	796.1	false
13C7-PFUnA	570.0 / 525.0	3.74	84587.82	239.197701	659.9	false
13C2-PFTeDA	715.0 / 670.0	4.48	76096.64	248.042101	1994.7	false
13C3-PFBS	302.0 / 99.0	1.52	19630.65	188.524395	233.7	false
13C3-PFHxS	402.0 / 99.0	2.28	22234.90	235.649383	173.4	false
13C8-PFOS	507.0 / 99.0	3.06	19275.38	180.834113	135.0	false

BATTELLE

Summary Quant Report

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	102547.94	227.921307	1238.4	false
d3-MeFOSAA	573.0 / 419.0	3.59	14453.01	227.322239	192.2	false
d5-EtFOSAA	589.0 / 419.0	3.74	12772.64	176.985580	160.7	false
13C5-PFHxA	318.0 / 273.0	1.85	89181.38	316.223847	219.2	false
13C4-PFHpA	367.0 / 322.0	2.27	96746.77	301.530871	422.9	false
13C8-PFOA	421.0 / 376.0	2.67	86472.58	221.551072	619.1	false
13C9-PFNA	472.0 / 427.0	3.07	90930.74	204.033116	684.9	false
13C6-PFDA	519.0 / 474.0	3.42	99312.33	219.089996	1117.7	false
13C7-PFUnA	570.0 / 525.0	3.74	100653.69	240.381726	942.9	false
13C2-PFTeDA	715.0 / 670.0	4.48	89623.80	246.720842	1342.4	false
13C3-PFBS	302.0 / 99.0	1.53	35233.91	304.194617	337.4	false
13C3-PFHxS	402.0 / 99.0	2.29	30765.68	293.126738	241.4	false
13C8-PFOS	507.0 / 99.0	3.07	28318.04	238.835246	200.5	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:52 PM

Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	99811.26	277.450863	960.9	false
d3-MeFOSAA	573.0 / 419.0	3.59	13563.23	234.653237	246.1	false
d5-EtFOSAA	589.0 / 419.0	3.74	14044.06	214.057119	163.6	false
13C5-PFHxA	318.0 / 273.0	1.85	65817.50	284.353006	420.7	false
13C4-PFHpA	367.0 / 322.0	2.27	72696.21	276.059759	673.7	false
13C8-PFOA	421.0 / 376.0	2.68	85146.36	265.801556	1614.5	false
13C9-PFNA	472.0 / 427.0	3.07	91454.28	250.028770	2463.6	false
13C6-PFDA	519.0 / 474.0	3.42	89135.49	245.934001	1407.0	false
13C7-PFUnA	570.0 / 525.0	3.74	89910.20	268.552547	912.9	false
13C2-PFTeDA	715.0 / 670.0	4.49	79893.80	275.070550	1238.4	false
13C3-PFBS	302.0 / 99.0	1.53	33921.02	322.136060	435.4	false
13C3-PFHxS	402.0 / 99.0	2.29	22116.23	231.782081	356.5	false
13C8-PFOS	507.0 / 99.0	3.07	27232.09	252.636489	242.2	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:56 PM

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	88531.24	188.582484	838.3	false
d3-MeFOSAA	573.0 / 419.0	3.58	14115.09	189.186507	177.3	false
d5-EtFOSAA	589.0 / 419.0	3.74	13154.31	155.327482	191.3	false
13C5-PFHxA	318.0 / 273.0	1.85	86536.62	295.003331	193.1	false
13C4-PFHpA	367.0 / 322.0	2.26	97439.72	291.969788	415.3	false
13C8-PFOA	421.0 / 376.0	2.67	93885.87	231.260924	600.6	false
13C9-PFNA	472.0 / 427.0	3.06	87794.90	189.393827	551.4	false
13C6-PFDA	519.0 / 474.0	3.42	99218.03	209.776490	1088.0	false
13C7-PFUnA	570.0 / 525.0	3.74	95079.09	217.622414	698.9	false
13C2-PFTeDA	715.0 / 670.0	4.48	79191.78	208.934148	1163.5	false
13C3-PFBS	302.0 / 99.0	1.52	33941.52	249.715262	283.0	false
13C3-PFHxS	402.0 / 99.0	2.29	29948.01	243.153167	215.5	false
13C8-PFOS	507.0 / 99.0	3.06	31270.57	224.747181	196.0	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:00 PM

Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	96209.85	255.286353	851.3	false
d3-MeFOSAA	573.0 / 419.0	3.58	11773.61	177.292390	201.8	false
d5-EtFOSAA	589.0 / 419.0	3.74	14065.10	186.593730	157.7	false
13C5-PFHxA	318.0 / 273.0	1.85	63398.55	256.506075	403.8	false
13C4-PFHpA	367.0 / 322.0	2.26	73829.73	262.557544	767.7	false
13C8-PFOA	421.0 / 376.0	2.67	88458.57	258.602826	1288.9	false
13C9-PFNA	472.0 / 427.0	3.07	85034.98	217.713565	1001.9	false
13C6-PFDA	519.0 / 474.0	3.42	95638.34	251.884541	2104.6	false
13C7-PFUnA	570.0 / 525.0	3.74	89341.16	254.726100	608.4	false
13C2-PFTeDA	715.0 / 670.0	4.48	78663.32	258.526368	1522.7	false
13C3-PFBS	302.0 / 99.0	1.52	35370.11	292.364065	417.6	false
13C3-PFHxS	402.0 / 99.0	2.29	21921.29	199.964085	286.8	false
13C8-PFOS	507.0 / 99.0	3.06	29594.99	238.973835	260.3	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:03 PM

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	73269.56	188.870959	957.3	false
d3-MeFOSAA	573.0 / 419.0	3.58	11713.50	179.346270	132.7	false
d5-EtFOSAA	589.0 / 419.0	3.74	13186.14	177.867648	182.7	false
13C5-PFHxA	318.0 / 273.0	1.85	71124.42	306.312777	252.6	false
13C4-PFHpA	367.0 / 322.0	2.26	80791.58	305.835158	379.5	false
13C8-PFOA	421.0 / 376.0	2.67	78989.42	245.804779	548.9	false
13C9-PFNA	472.0 / 427.0	3.07	75315.71	205.258672	615.5	false
13C6-PFDA	519.0 / 474.0	3.42	75466.65	193.089249	571.0	false
13C7-PFUnA	570.0 / 525.0	3.74	74274.11	205.727753	623.6	false
13C2-PFTeDA	715.0 / 670.0	4.48	62945.01	200.968254	1082.5	false
13C3-PFBS	302.0 / 99.0	1.53	27987.33	235.220049	303.6	false
13C3-PFHxS	402.0 / 99.0	2.29	24686.60	228.966700	230.7	false
13C8-PFOS	507.0 / 99.0	3.06	23059.19	189.322152	180.7	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:07 PM

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	86815.35	233.711322	1248.8	false
d3-MeFOSAA	573.0 / 419.0	3.58	13906.66	212.697048	177.5	false
d5-EtFOSAA	589.0 / 419.0	3.74	14481.02	195.124317	201.7	false
13C5-PFHxA	318.0 / 273.0	1.85	63357.10	278.620791	388.1	false
13C4-PFHpA	367.0 / 322.0	2.26	67320.94	260.221576	737.4	false
13C8-PFOA	421.0 / 376.0	2.67	84616.01	268.872105	1372.1	false
13C9-PFNA	472.0 / 427.0	3.06	82240.75	228.862585	1004.9	false
13C6-PFDA	519.0 / 474.0	3.42	91584.23	244.717686	1844.1	false
13C7-PFUnA	570.0 / 525.0	3.74	91666.24	265.159066	1044.8	false
13C2-PFTeDA	715.0 / 670.0	4.48	78581.48	262.016106	1807.0	false
13C3-PFBS	302.0 / 99.0	1.53	31781.59	266.821947	468.9	false
13C3-PFHxS	402.0 / 99.0	2.29	22195.74	205.642946	325.4	false
13C8-PFOS	507.0 / 99.0	3.06	28266.34	231.824843	232.6	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:11 PM

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	82646.80	210.203867	1218.0	false
d3-MeFOSAA	573.0 / 419.0	3.58	10264.02	181.027518	159.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	11801.20	183.369370	150.7	false
13C5-PFHxA	318.0 / 273.0	1.85	66932.34	272.850856	188.4	false
13C4-PFHpA	367.0 / 322.0	2.26	72492.47	259.750942	402.5	false
13C8-PFOA	421.0 / 376.0	2.67	72251.83	212.820345	683.3	false
13C9-PFNA	472.0 / 427.0	3.06	73093.75	188.555447	462.4	false
13C6-PFDA	519.0 / 474.0	3.42	77509.54	195.673145	814.1	false
13C7-PFUnA	570.0 / 525.0	3.74	75190.02	205.489024	523.6	false
13C2-PFTeDA	715.0 / 670.0	4.48	74520.78	234.755933	1815.2	false
13C3-PFBS	302.0 / 99.0	1.52	27142.61	262.776036	304.1	false
13C3-PFHxS	402.0 / 99.0	2.29	23963.67	256.026983	169.5	false
13C8-PFOS	507.0 / 99.0	3.06	23285.64	220.225150	171.5	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:14 PM

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	97301.17	264.349965	1060.8	false
d3-MeFOSAA	573.0 / 419.0	3.58	15567.57	240.786167	194.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	15287.18	208.310795	201.0	false
13C5-PFHxA	318.0 / 273.0	1.85	61134.03	257.805690	465.4	false
13C4-PFHpA	367.0 / 322.0	2.26	74841.70	277.413769	833.4	false
13C8-PFOA	421.0 / 376.0	2.67	89923.34	274.003976	1194.4	false
13C9-PFNA	472.0 / 427.0	3.06	86697.00	231.357231	1573.1	false
13C6-PFDA	519.0 / 474.0	3.42	94098.44	253.749394	1473.6	false
13C7-PFUnA	570.0 / 525.0	3.74	85607.61	249.912117	716.4	false
13C2-PFTeDA	715.0 / 670.0	4.48	83552.25	281.153715	1664.8	false
13C3-PFBS	302.0 / 99.0	1.52	34566.65	293.477779	545.3	false
13C3-PFHxS	402.0 / 99.0	2.29	22402.94	209.904216	303.1	false
13C8-PFOS	507.0 / 99.0	3.06	26459.95	219.458026	267.5	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:18 PM

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	93821.26	229.298466	919.0	false
d3-MeFOSAA	573.0 / 419.0	3.58	12532.29	169.938723	167.9	false
d5-EtFOSAA	589.0 / 419.0	3.74	15450.70	184.579587	202.6	false
13C5-PFHxA	318.0 / 273.0	1.85	57044.86	224.262937	687.5	false
13C4-PFHpA	367.0 / 322.0	2.27	63873.63	220.717848	866.1	false
13C8-PFOA	421.0 / 376.0	2.67	82178.36	233.439030	3085.6	false
13C9-PFNA	472.0 / 427.0	3.07	84493.92	210.201575	837.3	false
13C6-PFDA	519.0 / 474.0	3.42	86268.54	209.273250	863.8	false
13C7-PFUnA	570.0 / 525.0	3.74	80444.85	211.257384	943.7	false
13C2-PFTeDA	715.0 / 670.0	4.48	77728.09	235.289412	1681.6	false
13C3-PFBS	302.0 / 99.0	1.53	25131.27	187.061142	447.4	false
13C3-PFHxS	402.0 / 99.0	2.29	22099.18	181.527935	326.1	false
13C8-PFOS	507.0 / 99.0	3.06	24514.02	178.249426	225.8	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:21 PM

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	83541.13	211.501880	1027.9	false
d3-MeFOSAA	573.0 / 419.0	3.58	10982.59	158.949200	176.0	false
d5-EtFOSAA	589.0 / 419.0	3.73	12327.19	157.177849	173.8	false
13C5-PFHxA	318.0 / 273.0	1.85	55540.71	234.851981	508.2	false
13C4-PFHpA	367.0 / 322.0	2.26	58814.37	218.595449	769.5	false
13C8-PFOA	421.0 / 376.0	2.67	76742.66	234.473951	1297.1	false
13C9-PFNA	472.0 / 427.0	3.06	79056.09	211.537610	960.3	false
13C6-PFDA	519.0 / 474.0	3.42	82839.15	208.166542	2054.2	false
13C7-PFUnA	570.0 / 525.0	3.73	79244.29	215.573659	735.4	false
13C2-PFTeDA	715.0 / 670.0	4.48	65064.22	204.023704	1980.7	false
13C3-PFBS	302.0 / 99.0	1.53	25219.13	200.350719	527.9	false
13C3-PFHxS	402.0 / 99.0	2.29	22597.25	198.113780	264.3	false
13C8-PFOS	507.0 / 99.0	3.06	25711.20	199.539004	223.9	false

BATTELLE

Summary Quant Report

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	74290.96	176.937461	1049.8	false
d3-MeFOSAA	573.0 / 419.0	3.57	12242.79	206.471897	189.2	false
d5-EtFOSAA	589.0 / 419.0	3.74	13588.60	201.896573	169.3	false
13C5-PFHxA	318.0 / 273.0	1.84	59903.97	228.147859	162.4	false
13C4-PFHpA	367.0 / 322.0	2.26	76718.73	256.824843	351.0	false
13C8-PFOA	421.0 / 376.0	2.67	73205.39	201.455354	342.2	false
13C9-PFNA	472.0 / 427.0	3.06	69019.64	166.342409	327.6	false
13C6-PFDA	519.0 / 474.0	3.41	81585.93	192.868144	520.6	false
13C7-PFUnA	570.0 / 525.0	3.73	80073.69	204.921478	554.1	false
13C2-PFTeDA	715.0 / 670.0	4.48	65501.43	193.223122	1483.5	false
13C3-PFBS	302.0 / 99.0	1.51	18029.29	166.903860	204.1	false
13C3-PFHxS	402.0 / 99.0	2.28	22347.14	228.301088	200.2	false
13C8-PFOS	507.0 / 99.0	3.06	21381.76	193.364126	182.0	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:30 PM

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	94378.28	257.313349	1219.7	false
d3-MeFOSAA	573.0 / 419.0	3.58	13613.38	230.724293	182.9	false
d5-EtFOSAA	589.0 / 419.0	3.74	17667.86	263.806175	227.2	false
13C5-PFHxA	318.0 / 273.0	1.85	68138.74	300.884073	298.3	false
13C4-PFHpA	367.0 / 322.0	2.26	80104.14	310.910166	692.3	false
13C8-PFOA	421.0 / 376.0	2.67	89756.47	286.382081	880.8	false
13C9-PFNA	472.0 / 427.0	3.06	81476.57	227.670793	776.3	false
13C6-PFDA	519.0 / 474.0	3.42	94164.55	254.823284	2484.6	false
13C7-PFUnA	570.0 / 525.0	3.73	91192.69	267.155469	653.5	false
13C2-PFTeDA	715.0 / 670.0	4.48	79581.83	268.737750	1814.5	false
13C3-PFBS	302.0 / 99.0	1.52	36539.48	339.935851	363.6	false
13C3-PFHxS	402.0 / 99.0	2.28	22910.69	235.218242	288.8	false
13C8-PFOS	507.0 / 99.0	3.07	27567.47	250.539498	203.1	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:34 PM

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	84937.12	245.188685	806.8	false
d3-MeFOSAA	573.0 / 419.0	3.58	12693.72	272.490562	134.4	false
d5-EtFOSAA	589.0 / 419.0	3.73	14477.86	273.804592	133.2	false
13C5-PFHxA	318.0 / 273.0	1.84	51269.85	185.959091	98.0	false
13C4-PFHpA	367.0 / 322.0	2.25	61717.69	196.761410	197.1	false
13C8-PFOA	421.0 / 376.0	2.67	59485.51	155.898355	275.3	false
13C9-PFNA	472.0 / 427.0	3.06	67745.94	155.492049	356.6	false
13C6-PFDA	519.0 / 474.0	3.41	79142.97	226.765346	568.7	false
13C7-PFUnA	570.0 / 525.0	3.73	78553.70	243.659485	495.2	false
13C2-PFTeDA	715.0 / 670.0	4.48	69892.77	249.896223	1236.7	false
13C3-PFBS	302.0 / 99.0	1.52	18186.01	214.292549	192.5	false
13C3-PFHxS	402.0 / 99.0	2.28	20431.61	265.687209	121.6	false
13C8-PFOS	507.0 / 99.0	3.06	15839.65	182.330749	108.1	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:38 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.05	104011.35	279.262459	1307.9	false
d3-MeFOSAA	573.0 / 419.0	3.61	11579.11	165.351569	191.5	false
d5-EtFOSAA	589.0 / 419.0	3.78	13131.38	165.202717	176.9	false
13C5-PFHxA	318.0 / 273.0	1.88	78572.65	326.756221	271.0	false
13C4-PFHpA	367.0 / 322.0	2.29	88015.64	321.726788	645.3	false
13C8-PFOA	421.0 / 376.0	2.70	86324.44	259.394772	1036.0	false
13C9-PFNA	472.0 / 427.0	3.10	99122.80	260.852828	965.7	false
13C6-PFDA	519.0 / 474.0	3.45	99618.75	265.481550	749.6	false
13C7-PFUnA	570.0 / 525.0	3.77	93934.76	271.001651	929.7	false
13C2-PFTeDA	715.0 / 670.0	4.52	89235.46	296.752119	7644.8	false
13C3-PFBS	302.0 / 99.0	1.54	39312.91	308.159474	370.6	false
13C3-PFHxS	402.0 / 99.0	2.31	25020.64	216.439741	222.9	false
13C8-PFOS	507.0 / 99.0	3.09	29299.85	224.362536	228.4	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:42 PM

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	75285.88	215.895179	784.3	false
d3-MeFOSAA	573.0 / 419.0	3.59	11949.45	247.660616	129.6	false
d5-EtFOSAA	589.0 / 419.0	3.75	14952.12	273.014274	129.4	false
13C5-PFHxA	318.0 / 273.0	1.85	48513.17	192.083421	97.6	false
13C4-PFHpA	367.0 / 322.0	2.26	65151.03	226.739051	227.1	false
13C8-PFOA	421.0 / 376.0	2.68	56775.58	162.430203	299.9	false
13C9-PFNA	472.0 / 427.0	3.07	63436.48	158.942041	274.6	false
13C6-PFDA	519.0 / 474.0	3.42	80282.23	228.512649	595.9	false
13C7-PFUnA	570.0 / 525.0	3.74	76614.04	236.075846	587.2	false
13C2-PFTeDA	715.0 / 670.0	4.49	68631.64	243.768901	1365.0	false
13C3-PFBS	302.0 / 99.0	1.52	17605.10	200.287866	181.4	false
13C3-PFHxS	402.0 / 99.0	2.28	21672.42	272.095675	191.8	false
13C8-PFOS	507.0 / 99.0	3.07	17645.92	196.112393	119.4	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:47 PM

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	97309.37	254.357647	836.4	false
d3-MeFOSAA	573.0 / 419.0	3.59	13613.09	215.335011	184.5	false
d5-EtFOSAA	589.0 / 419.0	3.75	19470.81	271.341052	203.5	false
13C5-PFHxA	318.0 / 273.0	1.86	78098.37	315.557115	280.1	false
13C4-PFHpA	367.0 / 322.0	2.27	87662.25	311.331820	624.2	false
13C8-PFOA	421.0 / 376.0	2.68	87351.24	255.023432	1012.9	false
13C9-PFNA	472.0 / 427.0	3.07	93161.34	238.199713	760.8	false
13C6-PFDA	519.0 / 474.0	3.43	101005.94	262.058635	799.7	false
13C7-PFUnA	570.0 / 525.0	3.75	95017.72	266.875404	1386.3	false
13C2-PFTeDA	715.0 / 670.0	4.49	84876.94	274.792187	1588.9	false
13C3-PFBS	302.0 / 99.0	1.53	38502.79	334.316066	414.5	false
13C3-PFHxS	402.0 / 99.0	2.29	26946.80	258.208473	336.0	false
13C8-PFOS	507.0 / 99.0	3.07	27614.72	234.234211	290.0	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:51 PM

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	74184.66	178.622115	726.8	false
d3-MeFOSAA	573.0 / 419.0	3.59	12958.20	258.717133	155.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	12601.62	221.656445	158.9	false
13C5-PFHxA	318.0 / 273.0	1.85	52399.67	210.001911	121.7	false
13C4-PFHpA	367.0 / 322.0	2.26	77505.88	273.026125	276.9	false
13C8-PFOA	421.0 / 376.0	2.68	71032.21	205.695582	447.2	false
13C9-PFNA	472.0 / 427.0	3.07	70841.50	179.660217	446.7	false
13C6-PFDA	519.0 / 474.0	3.42	85172.75	203.555680	643.6	false
13C7-PFUnA	570.0 / 525.0	3.74	76205.68	197.161572	551.8	false
13C2-PFTeDA	715.0 / 670.0	4.49	61983.03	184.849575	1579.1	false
13C3-PFBS	302.0 / 99.0	1.52	17691.12	193.884392	171.1	false
13C3-PFHxS	402.0 / 99.0	2.28	21131.18	255.569845	157.8	false
13C8-PFOS	507.0 / 99.0	3.07	22996.67	246.205164	136.2	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:42:55 PM

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.03	98899.04	256.577683	1688.7	false
d3-MeFOSAA	573.0 / 419.0	3.58	15944.25	260.009756	305.2	false
d5-EtFOSAA	589.0 / 419.0	3.74	15741.98	226.161404	239.3	false
13C5-PFHxA	318.0 / 273.0	1.85	80215.79	319.329626	285.2	false
13C4-PFHpA	367.0 / 322.0	2.26	91868.65	321.456012	634.9	false
13C8-PFOA	421.0 / 376.0	2.67	95016.44	273.308463	1033.7	false
13C9-PFNA	472.0 / 427.0	3.07	92944.24	234.137691	1024.4	false
13C6-PFDA	519.0 / 474.0	3.42	99770.94	256.916681	904.1	false
13C7-PFUnA	570.0 / 525.0	3.74	96181.53	268.121906	758.5	false
13C2-PFTeDA	715.0 / 670.0	4.48	90902.35	292.096566	2028.0	false
13C3-PFBS	302.0 / 99.0	1.52	35183.62	314.943847	348.9	false
13C3-PFHxS	402.0 / 99.0	2.29	25147.82	248.422620	276.3	false
13C8-PFOS	507.0 / 99.0	3.06	27353.45	239.193492	237.8	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:00 PM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	80984.56	225.138294	1148.5	false
d3-MeFOSAA	573.0 / 419.0	3.58	13954.03	282.287474	142.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	14374.30	256.184190	149.3	false
13C5-PFHxA	318.0 / 273.0	1.84	46341.61	198.222300	100.4	false
13C4-PFHpA	367.0 / 322.0	2.26	60573.15	227.738487	231.9	false
13C8-PFOA	421.0 / 376.0	2.67	63664.07	196.766327	417.9	false
13C9-PFNA	472.0 / 427.0	3.06	62009.43	167.845108	375.3	false
13C6-PFDA	519.0 / 474.0	3.42	86229.83	237.939236	566.2	false
13C7-PFUnA	570.0 / 525.0	3.74	80755.15	241.229909	583.2	false
13C2-PFTeDA	715.0 / 670.0	4.48	66356.21	228.482653	1430.6	false
13C3-PFBS	302.0 / 99.0	1.52	18019.37	200.096182	183.0	false
13C3-PFHxS	402.0 / 99.0	2.28	21044.26	257.887966	183.3	false
13C8-PFOS	507.0 / 99.0	3.06	16456.37	178.516328	129.9	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:04 PM

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	98168.20	260.852142	1270.7	false
d3-MeFOSAA	573.0 / 419.0	3.58	14690.73	206.113006	197.0	false
d5-EtFOSAA	589.0 / 419.0	3.74	16583.75	204.983205	182.6	false
13C5-PFHxA	318.0 / 273.0	1.85	72106.09	294.358161	352.5	false
13C4-PFHpA	367.0 / 322.0	2.27	83228.08	298.640640	593.6	false
13C8-PFOA	421.0 / 376.0	2.68	92242.09	272.087313	1131.4	false
13C9-PFNA	472.0 / 427.0	3.07	87761.87	226.714617	1124.9	false
13C6-PFDA	519.0 / 474.0	3.42	103309.63	272.474469	1228.3	false
13C7-PFUnA	570.0 / 525.0	3.74	95482.74	272.622834	632.7	false
13C2-PFTeDA	715.0 / 670.0	4.48	85544.11	281.538761	1605.8	false
13C3-PFBS	302.0 / 99.0	1.53	37706.48	290.392282	489.0	false
13C3-PFHxS	402.0 / 99.0	2.29	22542.94	191.592245	320.5	false
13C8-PFOS	507.0 / 99.0	3.07	25339.57	190.639526	204.0	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:08 PM

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID		**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	49934.63	221.494569	694.6	false
d3-MeFOSAA	573.0 / 419.0	3.58	9599.04	280.286634	125.0	false
d5-EtFOSAA	589.0 / 419.0	3.73	9418.13	242.277345	108.5	false
13C5-PFHxA	318.0 / 273.0	1.84	21215.73	173.528039	68.5	false
13C4-PFHpA	367.0 / 322.0	2.25	33553.99	241.229653	191.3	false
13C8-PFOA	421.0 / 376.0	2.67	36105.85	213.385134	206.5	false
13C9-PFNA	472.0 / 427.0	3.06	40704.57	210.680476	210.5	false
13C6-PFDA	519.0 / 474.0	3.42	47197.29	207.797208	343.4	false
13C7-PFUnA	570.0 / 525.0	3.73	50168.76	239.115866	446.6	false
13C2-PFTeDA	715.0 / 670.0	4.48	34458.50	189.313734	969.8	false
13C3-PFBS	302.0 / 99.0	1.51	8767.75	140.530374	110.2	false
13C3-PFHxS	402.0 / 99.0	2.28	13040.58	230.662392	107.0	false
13C8-PFOS	507.0 / 99.0	3.06	12735.13	199.402152	84.2	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:46:37 PM

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.07	124281.55	283.925237	1368.7	false
d3-MeFOSAA	573.0 / 419.0	3.62	20250.11	267.818361	227.0	false
d5-EtFOSAA	589.0 / 419.0	3.78	18971.36	221.047251	230.4	false
13C5-PFHxA	318.0 / 273.0	1.87	77874.50	266.004193	882.3	false
13C4-PFHpA	367.0 / 322.0	2.29	87915.24	263.956465	846.9	false
13C8-PFOA	421.0 / 376.0	2.70	99152.64	244.721713	1307.2	false
13C9-PFNA	472.0 / 427.0	3.10	111508.83	241.030493	839.3	false
13C6-PFDA	519.0 / 474.0	3.46	108469.53	245.961139	1025.0	false
13C7-PFUnA	570.0 / 525.0	3.78	110417.79	271.050478	696.1	false
13C2-PFTeDA	715.0 / 670.0	4.54	106949.79	302.622889	1813.5	false
13C3-PFBS	302.0 / 99.0	1.54	32827.89	238.321404	465.2	false
13C3-PFHxS	402.0 / 99.0	2.31	29367.70	235.281781	396.4	false
13C8-PFOS	507.0 / 99.0	3.10	33587.26	238.198666	253.5	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:46:40 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.06	102412.42	264.845664	1187.8	false
d3-MeFOSAA	573.0 / 419.0	3.62	14401.34	235.330947	141.0	false
d5-EtFOSAA	589.0 / 419.0	3.78	15139.47	217.951734	129.7	false
13C5-PFHxA	318.0 / 273.0	1.87	59637.72	241.968216	512.6	false
13C4-PFHpA	367.0 / 322.0	2.29	64236.43	229.083352	762.5	false
13C8-PFOA	421.0 / 376.0	2.70	73720.34	216.122303	4438.7	false
13C9-PFNA	472.0 / 427.0	3.10	89503.09	229.797248	2474.0	false
13C6-PFDA	519.0 / 474.0	3.46	87249.50	223.956977	909.5	false
13C7-PFUnA	570.0 / 525.0	3.78	79452.86	220.781958	785.6	false
13C2-PFTeDA	715.0 / 670.0	4.53	83161.42	266.370808	1443.9	false
13C3-PFBS	302.0 / 99.0	1.54	30400.45	272.686255	425.1	false
13C3-PFHxS	402.0 / 99.0	2.31	22924.50	226.924508	337.7	false
13C8-PFOS	507.0 / 99.0	3.09	26653.92	233.554878	211.2	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:46:44 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.06	122058.68	269.417382	1006.1	false
d3-MeFOSAA	573.0 / 419.0	3.61	20262.41	269.879467	227.6	false
d5-EtFOSAA	589.0 / 419.0	3.77	19458.63	228.330834	214.7	false
13C5-PFHxA	318.0 / 273.0	1.87	82111.69	293.477244	327.3	false
13C4-PFHpA	367.0 / 322.0	2.28	98639.40	309.880852	558.2	false
13C8-PFOA	421.0 / 376.0	2.70	102244.02	264.047665	1011.5	false
13C9-PFNA	472.0 / 427.0	3.09	107445.94	243.012682	1622.9	false
13C6-PFDA	519.0 / 474.0	3.45	112886.63	247.320924	812.4	false
13C7-PFUnA	570.0 / 525.0	3.77	114328.90	271.160713	766.8	false
13C2-PFTeDA	715.0 / 670.0	4.53	107110.25	292.827940	1709.4	false
13C3-PFBS	302.0 / 99.0	1.53	40801.04	298.302645	340.2	false
13C3-PFHxS	402.0 / 99.0	2.31	28931.92	233.432513	306.2	false
13C8-PFOS	507.0 / 99.0	3.09	33242.30	237.422311	286.2	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:46:48 PM

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.05	98065.26	277.833614	799.3	false
d3-MeFOSAA	573.0 / 419.0	3.61	14742.30	254.155188	257.6	false
d5-EtFOSAA	589.0 / 419.0	3.77	15748.65	239.194258	214.5	false
13C5-PFHxA	318.0 / 273.0	1.86	64289.99	273.563867	451.1	false
13C4-PFHpA	367.0 / 322.0	2.28	70155.09	262.391330	685.7	false
13C8-PFOA	421.0 / 376.0	2.69	85385.52	262.527365	4764.4	false
13C9-PFNA	472.0 / 427.0	3.09	92550.97	249.210234	2271.7	false
13C6-PFDA	519.0 / 474.0	3.45	91658.72	257.753617	1007.4	false
13C7-PFUnA	570.0 / 525.0	3.77	92121.78	280.443713	651.0	false
13C2-PFTeDA	715.0 / 670.0	4.52	87321.85	306.419944	1502.9	false
13C3-PFBS	302.0 / 99.0	1.53	30785.58	291.332084	373.2	false
13C3-PFHxS	402.0 / 99.0	2.30	25301.18	264.228627	270.9	false
13C8-PFOS	507.0 / 99.0	3.08	27583.10	254.993203	228.0	false

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:11 PM

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.57	PFBS			
PFBS_2	298.9 / 99.0	1.57	PFBS	0.410	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	N/A	PFHxS			
PFHxS_2	399.0 / 99.0	N/A	PFHxS	N/A	0.282	ü
PFOA_1	413.0 / 369.0	N/A	PFOA			
PFOA_2	413.0 / 169.0	N/A	PFOA	N/A	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.280	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.11	PFOS	0.210	0.174	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.08	PFDoA			
PFDoA_2	613.0 / 319.0	4.08	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.32	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.34	PFTrDA	0.080	0.066	ü
PFTeDA_1	713.0 / 669.0	4.54	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.54	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.61	NMeFOSAA	0.510	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:12:24 AM

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.58	PFBS			
PFBS_2	298.9 / 99.0	1.86	PFBS	1.730	0.292	
PFHxA_1	313.0 / 269.0	1.92	PFHxA			
PFHxA_2	313.0 / 119.0	1.91	PFHxA	0.100	0.077	ü
PFHpA_1	363.0 / 319.0	2.34	PFHpA			
PFHpA_2	363.0 / 169.0	2.49	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.36	PFHxS			
PFHxS_2	399.0 / 99.0	2.37	PFHxS	0.240	0.282	ü
PFOA_1	413.0 / 369.0	2.75	PFOA			
PFOA_2	413.0 / 169.0	2.75	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.16	PFNA			
PFNA_2	463.0 / 219.0	3.15	PFNA	0.200	0.306	ü
PFOS_1	499.0 / 80.0	N/A	PFOS			
PFOS_2	499.0 / 99.0	N/A	PFOS	N/A	0.174	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	N/A	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:16 PM

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	N/A	PFHxS			
PFHxS_2	399.0 / 99.0	N/A	PFHxS	N/A	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.09	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.04	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.47	PFDA	0.120	0.041	
PFUnA_1	563.0 / 519.0	3.77	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	N/A	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	ü
PFTrDA_1	663.0 / 619.0	4.29	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:24 PM

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.280	0.077	
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.040	0.025	
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.390	0.282	ü
PFOA_1	413.0 / 369.0	2.65	PFOA			
PFOA_2	413.0 / 169.0	2.61	PFOA	0.270	0.065	
PFNA_1	463.0 / 419.0	3.04	PFNA			
PFNA_2	463.0 / 219.0	3.07	PFNA	0.160	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.00	PFOS	0.110	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.45	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	4.04	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.49	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.48	PFTeDA	0.040	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.60	NMeFOSAA	0.550	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.74	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.74	NEtFOSAA	0.170	0.062	

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:29 PM

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.310	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.080	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.110	0.065	
PFNA_1	463.0 / 419.0	3.07	PFNA			
PFNA_2	463.0 / 219.0	3.07	PFNA	0.240	0.306	ü
PFOS_1	499.0 / 80.0	2.94	PFOS			
PFOS_2	499.0 / 99.0	2.99	PFOS	0.100	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.42	PFDA	0.070	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	4.02	PFDoA	0.140	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.27	PFTrDA	0.120	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:34 PM

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.320	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.06	PFNA			
PFNA_2	463.0 / 219.0	3.09	PFNA	0.230	0.306	ü
PFOS_1	499.0 / 80.0	2.94	PFOS			
PFOS_2	499.0 / 99.0	2.98	PFOS	0.100	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	4.04	PFDoA	0.150	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.27	PFTrDA	0.100	0.066	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:38 PM

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.310	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.390	0.306	ü
PFOS_1	499.0 / 80.0	2.94	PFOS			
PFOS_2	499.0 / 99.0	2.97	PFOS	0.090	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.43	PFDA	0.070	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	4.02	PFDoA	0.310	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.27	PFTrDA	0.520	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:43 PM

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	2.29	PFHpA			
PFHpA_2	363.0 / 169.0	2.29	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.31	PFHxS	0.340	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.10	PFNA	0.340	0.306	ü
PFOS_1	499.0 / 80.0	N/A	PFOS			
PFOS_2	499.0 / 99.0	N/A	PFOS	N/A	0.174	ü
PFDA_1	513.0 / 469.0	3.41	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:48:47 PM

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.230	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.610	0.306	
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.06	PFOS	0.210	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	4.02	PFDoA	0.260	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.060	0.066	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.74	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:15 AM

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.29	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.330	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.370	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.00	PFOS	0.110	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.41	PFDA	0.090	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:24 AM

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.030	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.05	PFOS	0.130	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.41	PFDA	0.130	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:29 AM

Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.270	0.306	ü
PFOS_1	499.0 / 80.0	2.96	PFOS			
PFOS_2	499.0 / 99.0	3.07	PFOS	0.150	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.42	PFDA	0.090	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.05	PFDoA			
PFDoA_2	613.0 / 319.0	4.00	PFDoA	0.180	0.160	ü
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:34 AM

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.22	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.07	PFNA	0.420	0.306	ü
PFOS_1	499.0 / 80.0	2.96	PFOS			
PFOS_2	499.0 / 99.0	3.06	PFOS	0.140	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.29	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.28	PFTrDA	0.190	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:39 AM

Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.310	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.210	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.05	PFOS	0.140	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:44 AM

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.370	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.05	PFOS	0.140	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.41	PFDA	0.060	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:49 AM

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.25	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.290	0.306	ü
PFOS_1	499.0 / 80.0	2.94	PFOS			
PFOS_2	499.0 / 99.0	3.03	PFOS	0.120	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.45	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:54 AM

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.03	PFOS	0.120	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.43	PFDA	0.180	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.00	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:29:58 AM

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.25	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.040	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.350	0.306	ü
PFOS_1	499.0 / 80.0	2.95	PFOS			
PFOS_2	499.0 / 99.0	3.03	PFOS	0.130	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.05	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:49:34 PM

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	
PFHxS_1	399.0 / 80.0	N/A	PFHxS			
PFHxS_2	399.0 / 99.0	N/A	PFHxS	N/A	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.340	0.306	ü
PFOS_1	499.0 / 80.0	N/A	PFOS			
PFOS_2	499.0 / 99.0	3.10	PFOS	N/A	0.174	
PFDA_1	513.0 / 469.0	3.42	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.26	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:49:39 PM

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.420	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.070	0.065	ü
PFNA_1	463.0 / 419.0	3.07	PFNA			
PFNA_2	463.0 / 219.0	3.09	PFNA	0.210	0.306	ü
PFOS_1	499.0 / 80.0	3.07	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.45	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:30:02 AM

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.53	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.85	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.26	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.29	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.67	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.04	PFNA			
PFNA_2	463.0 / 219.0	3.06	PFNA	0.170	0.306	ü
PFOS_1	499.0 / 80.0	3.05	PFOS			
PFOS_2	499.0 / 99.0	3.07	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.74	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 6:30:06 AM

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.270	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.67	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.310	0.306	ü
PFOS_1	499.0 / 80.0	3.06	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:49:50 PM

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.53	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.85	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.26	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.29	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.310	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.65	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.08	PFNA			
PFNA_2	463.0 / 219.0	3.07	PFNA	0.340	0.306	ü
PFOS_1	499.0 / 80.0	3.06	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	3.42	PFDA	0.030	0.041	ü
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:49:55 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.56	PFBS	0.340	0.292	ü
PFHxA_1	313.0 / 269.0	1.88	PFHxA			
PFHxA_2	313.0 / 119.0	1.89	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.68	PFOA	0.100	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.340	0.306	ü
PFOS_1	499.0 / 80.0	3.10	PFOS			
PFOS_2	499.0 / 99.0	3.11	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	3.46	PFDA			
PFDA_2	513.0 / 219.0	3.49	PFDA	0.140	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.31	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 2:48:40 PM

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.280	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.310	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.65	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.09	PFNA			
PFNA_2	463.0 / 219.0	3.08	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.07	PFOS			
PFOS_2	499.0 / 99.0	3.09	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	3.43	PFDA	0.080	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.29	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 01/11/2018 2:48:45 PM

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.28	PFHpA			
PFHpA_2	363.0 / 169.0	2.28	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.66	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	3.09	PFNA			
PFNA_2	463.0 / 219.0	3.09	PFNA	0.350	0.306	ü
PFOS_1	499.0 / 80.0	3.08	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.04	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.30	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:50:08 PM

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.53	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.23	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.28	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.69	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.100	0.065	ü
PFNA_1	463.0 / 419.0	3.00	PFNA			
PFNA_2	463.0 / 219.0	2.97	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	3.02	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.140	0.174	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:50:11 PM

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.86	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.24	PFHpA	0.040	0.025	
PFHxS_1	399.0 / 80.0	2.27	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.270	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.100	0.065	ü
PFNA_1	463.0 / 419.0	3.07	PFNA			
PFNA_2	463.0 / 219.0	3.05	PFNA	0.360	0.306	ü
PFOS_1	499.0 / 80.0	3.00	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.150	0.174	ü
PFDA_1	513.0 / 469.0	3.44	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.74	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:50:15 PM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.53	PFBS			
PFBS_2	298.9 / 99.0	1.53	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.85	PFHxA			
PFHxA_2	313.0 / 119.0	1.85	PFHxA	0.060	0.077	ü
PFHpA_1	363.0 / 319.0	2.26	PFHpA			
PFHpA_2	363.0 / 169.0	2.23	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.29	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.450	0.282	
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.62	PFOA	0.100	0.065	
PFNA_1	463.0 / 419.0	3.02	PFNA			
PFNA_2	463.0 / 219.0	2.97	PFNA	0.390	0.306	ü
PFOS_1	499.0 / 80.0	2.99	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.180	0.174	ü
PFDA_1	513.0 / 469.0	3.39	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.76	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.03	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.27	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.76	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:50:20 PM

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.54	PFBS			
PFBS_2	298.9 / 99.0	1.54	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.86	PFHxA			
PFHxA_2	313.0 / 119.0	1.87	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.27	PFHpA			
PFHpA_2	363.0 / 169.0	2.21	PFHpA	0.030	0.025	ü
PFHxS_1	399.0 / 80.0	2.30	PFHxS			
PFHxS_2	399.0 / 99.0	2.30	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.62	PFOA	0.110	0.065	
PFNA_1	463.0 / 419.0	3.07	PFNA			
PFNA_2	463.0 / 219.0	3.07	PFNA	0.240	0.306	ü
PFOS_1	499.0 / 80.0	3.00	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.150	0.174	ü
PFDA_1	513.0 / 469.0	3.43	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	3.75	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	
PFDoA_1	613.0 / 569.0	4.02	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	
PFTrDA_1	663.0 / 619.0	4.28	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.59	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:50:24 PM

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID	VC-S14GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	2.29	PFHxS			
PFHxS_2	399.0 / 99.0	2.29	PFHxS	0.260	0.282	ü
PFOA_1	413.0 / 369.0	2.68	PFOA			
PFOA_2	413.0 / 169.0	2.64	PFOA	0.090	0.065	ü
PFNA_1	463.0 / 419.0	N/A	PFNA			
PFNA_2	463.0 / 219.0	N/A	PFNA	N/A	0.306	ü
PFOS_1	499.0 / 80.0	3.03	PFOS			
PFOS_2	499.0 / 99.0	3.08	PFOS	0.160	0.174	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	3.44	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	N/A	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:51:00 PM

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.400	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	N/A	PFHxS			
PFHxS_2	399.0 / 99.0	N/A	PFHxS	N/A	0.282	ü
PFOA_1	413.0 / 369.0	2.72	PFOA			
PFOA_2	413.0 / 169.0	2.72	PFOA	0.290	0.065	
PFNA_1	463.0 / 419.0	N/A	PFNA			
PFNA_2	463.0 / 219.0	N/A	PFNA	N/A	0.306	ü
PFOS_1	499.0 / 80.0	3.09	PFOS			
PFOS_2	499.0 / 99.0	N/A	PFOS	N/A	0.174	
PFDA_1	513.0 / 469.0	3.48	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	4.08	PFDoA	0.140	0.160	ü
PFTrDA_1	663.0 / 619.0	4.34	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	
PFTeDA_1	713.0 / 669.0	4.54	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.55	PFTeDA	0.120	0.049	
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:51:04 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.56	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.290	0.292	ü
PFHxA_1	313.0 / 269.0	1.88	PFHxA			
PFHxA_2	313.0 / 119.0	1.88	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.30	PFHpA			
PFHpA_2	363.0 / 169.0	2.30	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.300	0.282	ü
PFOA_1	413.0 / 369.0	2.71	PFOA			
PFOA_2	413.0 / 169.0	2.71	PFOA	0.060	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.320	0.306	ü
PFOS_1	499.0 / 80.0	3.11	PFOS			
PFOS_2	499.0 / 99.0	3.11	PFOS	0.170	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.47	PFDA	0.040	0.041	ü
PFUnA_1	563.0 / 519.0	3.80	PFUnA			
PFUnA_2	563.0 / 269.0	3.79	PFUnA	0.050	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	4.07	PFDoA	0.160	0.160	ü
PFTrDA_1	663.0 / 619.0	4.32	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.32	PFTrDA	0.070	0.066	ü
PFTeDA_1	713.0 / 669.0	4.54	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.53	PFTeDA	0.050	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.63	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.62	NMeFOSAA	0.510	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.79	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.79	NEtFOSAA	0.070	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:51:08 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.55	PFBS			
PFBS_2	298.9 / 99.0	1.55	PFBS	0.300	0.292	ü
PFHxA_1	313.0 / 269.0	1.87	PFHxA			
PFHxA_2	313.0 / 119.0	1.88	PFHxA	0.070	0.077	ü
PFHpA_1	363.0 / 319.0	2.29	PFHpA			
PFHpA_2	363.0 / 169.0	2.27	PFHpA	0.020	0.025	ü
PFHxS_1	399.0 / 80.0	2.32	PFHxS			
PFHxS_2	399.0 / 99.0	2.32	PFHxS	0.290	0.282	ü
PFOA_1	413.0 / 369.0	2.70	PFOA			
PFOA_2	413.0 / 169.0	2.69	PFOA	0.050	0.065	ü
PFNA_1	463.0 / 419.0	3.11	PFNA			
PFNA_2	463.0 / 219.0	3.11	PFNA	0.330	0.306	ü
PFOS_1	499.0 / 80.0	2.98	PFOS			
PFOS_2	499.0 / 99.0	3.03	PFOS	0.120	0.174	ü
PFDA_1	513.0 / 469.0	3.47	PFDA			
PFDA_2	513.0 / 219.0	3.50	PFDA	0.050	0.041	ü
PFUnA_1	563.0 / 519.0	3.79	PFUnA			
PFUnA_2	563.0 / 269.0	3.80	PFUnA	0.040	0.049	ü
PFDoA_1	613.0 / 569.0	4.07	PFDoA			
PFDoA_2	613.0 / 319.0	4.07	PFDoA	0.120	0.160	ü
PFTrDA_1	663.0 / 619.0	4.32	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.32	PFTrDA	0.100	0.066	ü
PFTeDA_1	713.0 / 669.0	4.53	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.54	PFTeDA	0.070	0.049	ü
NMeFOSAA_1	570.0 / 419.0	3.62	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.63	NMeFOSAA	0.320	0.551	ü
NEtFOSAA_1	584.0 / 419.0	3.78	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	3.78	NEtFOSAA	0.060	0.062	ü

BATTELLE

Summary Ion Ratio Report

Created with Analyst Reporter
Printed: 25/10/2018 1:51:12 PM

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	N/A	PFBS			
PFBS_2	298.9 / 99.0	N/A	PFBS	N/A	0.292	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.077	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.025	ü
PFHxS_1	399.0 / 80.0	2.31	PFHxS			
PFHxS_2	399.0 / 99.0	2.31	PFHxS	0.280	0.282	ü
PFOA_1	413.0 / 369.0	N/A	PFOA			
PFOA_2	413.0 / 169.0	N/A	PFOA	N/A	0.065	ü
PFNA_1	463.0 / 419.0	N/A	PFNA			
PFNA_2	463.0 / 219.0	N/A	PFNA	N/A	0.306	ü
PFOS_1	499.0 / 80.0	N/A	PFOS			
PFOS_2	499.0 / 99.0	N/A	PFOS	N/A	0.174	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.041	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.049	ü
PFDoA_1	613.0 / 569.0	N/A	PFDoA			
PFDoA_2	613.0 / 319.0	N/A	PFDoA	N/A	0.160	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.066	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.049	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.551	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSAA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSAA	N/A	0.062	ü

BATTELLE

Summary Internal Standard Report

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.57	13C3-PFBS	302.0 / 99.0	25201.92	232.25
PFBS_2	298.9 / 99.0	1.57	13C3-PFBS	302.0 / 99.0	25201.92	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	61987.66	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	61987.66	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	70438.43	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	70438.43	250.00
PFHxS_1	399.0 / 80.0	N/A	13C3-PFHxS	402.0 / 99.0	24683.15	236.50
PFHxS_2	399.0 / 99.0	N/A	13C3-PFHxS	402.0 / 99.0	24683.15	236.50
PFOA_1	413.0 / 369.0	N/A	13C8-PFOA	421.0 / 376.0	88535.23	250.00
PFOA_2	413.0 / 169.0	N/A	13C8-PFOA	421.0 / 376.0	88535.23	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	100138.19	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	100138.19	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	26769.68	239.25
PFOS_2	499.0 / 99.0	3.11	13C8-PFOS	507.0 / 99.0	26769.68	239.25
PFDA_1	513.0 / 469.0	N/A	13C6-PFDA	519.0 / 474.0	97258.51	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	97258.51	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	88898.52	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	88898.52	250.00
PFDoA_1	613.0 / 569.0	4.08	13C2-PFDoA	615.0 / 570.0	91698.47	250.00
PFDoA_2	613.0 / 319.0	4.08	13C2-PFDoA	615.0 / 570.0	91698.47	250.00
PFTrDA_1	663.0 / 619.0	4.32	13C2-PFTeDA	715.0 / 670.0	70451.22	250.00
PFTrDA_2	663.0 / 169.0	4.34	13C2-PFTeDA	715.0 / 670.0	70451.22	250.00
PFTeDA_1	713.0 / 669.0	4.54	13C2-PFTeDA	715.0 / 670.0	70451.22	250.00
PFTeDA_2	713.0 / 169.0	4.54	13C2-PFTeDA	715.0 / 670.0	70451.22	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	16454.46	250.00
NMeFOSAA_2	570.0 / 512.0	3.61	d3-MeFOSAA	573.0 / 419.0	16454.46	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	17612.55	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	17612.55	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.58	13C3-PFBS	302.0 / 99.0	23436.89	232.25
PFBS_2	298.9 / 99.0	1.86	13C3-PFBS	302.0 / 99.0	23436.89	232.25
PFHxA_1	313.0 / 269.0	1.92	13C5-PFHxA	318.0 / 273.0	58984.04	250.00
PFHxA_2	313.0 / 119.0	1.91	13C5-PFHxA	318.0 / 273.0	58984.04	250.00
PFHpA_1	363.0 / 319.0	2.34	13C4-PFHpA	367.0 / 322.0	65679.00	250.00
PFHpA_2	363.0 / 169.0	2.49	13C4-PFHpA	367.0 / 322.0	65679.00	250.00
PFHxS_1	399.0 / 80.0	2.36	13C3-PFHxS	402.0 / 99.0	21362.88	236.50
PFHxS_2	399.0 / 99.0	2.37	13C3-PFHxS	402.0 / 99.0	21362.88	236.50
PFOA_1	413.0 / 369.0	2.75	13C8-PFOA	421.0 / 376.0	77309.52	250.00
PFOA_2	413.0 / 169.0	2.75	13C8-PFOA	421.0 / 376.0	77309.52	250.00
PFNA_1	463.0 / 419.0	3.16	13C9-PFNA	472.0 / 427.0	83779.99	250.00
PFNA_2	463.0 / 219.0	3.15	13C9-PFNA	472.0 / 427.0	83779.99	250.00
PFOS_1	499.0 / 80.0	N/A	13C8-PFOS	507.0 / 99.0	24982.25	239.25
PFOS_2	499.0 / 99.0	N/A	13C8-PFOS	507.0 / 99.0	24982.25	239.25
PFDA_1	513.0 / 469.0	N/A	13C6-PFDA	519.0 / 474.0	92389.14	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	92389.14	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	86079.46	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	86079.46	250.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFDoA	615.0 / 570.0	97310.49	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	97310.49	250.00
PFTrDA_1	663.0 / 619.0	N/A	13C2-PFTeDA	715.0 / 670.0	88378.85	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	88378.85	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	88378.85	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	88378.85	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	16077.43	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	16077.43	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16446.12	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16446.12	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	25188.47	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	25188.47	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	55762.03	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	55762.03	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	66522.02	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	66522.02	250.00
PFHxS_1	399.0 / 80.0	N/A	13C3-PFHxS	402.0 / 99.0	21586.89	236.50
PFHxS_2	399.0 / 99.0	N/A	13C3-PFHxS	402.0 / 99.0	21586.89	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	77976.53	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	77976.53	250.00
PFNA_1	463.0 / 419.0	3.09	13C9-PFNA	472.0 / 427.0	84602.02	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	84602.02	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	23475.82	239.25
PFOS_2	499.0 / 99.0	3.04	13C8-PFOS	507.0 / 99.0	23475.82	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	90666.51	250.00
PFDA_2	513.0 / 219.0	3.47	13C6-PFDA	519.0 / 474.0	90666.51	250.00
PFUnA_1	563.0 / 519.0	3.77	13C7-PFUnA	570.0 / 525.0	80182.41	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	80182.41	250.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFDoA	615.0 / 570.0	87903.18	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	87903.18	250.00
PFTrDA_1	663.0 / 619.0	4.29	13C2-PFTeDA	715.0 / 670.0	75890.10	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	75890.10	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	75890.10	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	75890.10	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14529.19	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14529.19	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14011.50	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14011.50	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	21726.38	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	21726.38	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	36542.29	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	36542.29	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	48561.65	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	48561.65	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	16411.61	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	16411.61	236.50
PFOA_1	413.0 / 369.0	2.65	13C8-PFOA	421.0 / 376.0	34635.88	250.00
PFOA_2	413.0 / 169.0	2.61	13C8-PFOA	421.0 / 376.0	34635.88	250.00
PFNA_1	463.0 / 419.0	3.04	13C9-PFNA	472.0 / 427.0	68324.46	250.00
PFNA_2	463.0 / 219.0	3.07	13C9-PFNA	472.0 / 427.0	68324.46	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	19734.46	239.25
PFOS_2	499.0 / 99.0	3.00	13C8-PFOS	507.0 / 99.0	19734.46	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	79988.20	250.00
PFDA_2	513.0 / 219.0	3.45	13C6-PFDA	519.0 / 474.0	79988.20	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	81277.30	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	81277.30	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	81470.47	250.00
PFDoA_2	613.0 / 319.0	4.04	13C2-PFDoA	615.0 / 570.0	81470.47	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	64949.82	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	64949.82	250.00
PFTeDA_1	713.0 / 669.0	4.49	13C2-PFTeDA	715.0 / 670.0	64949.82	250.00
PFTeDA_2	713.0 / 169.0	4.48	13C2-PFTeDA	715.0 / 670.0	64949.82	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	12275.86	250.00
NMeFOSAA_2	570.0 / 512.0	3.60	d3-MeFOSAA	573.0 / 419.0	12275.86	250.00
NEtFOSAA_1	584.0 / 419.0	3.74	d5-EtFOSAA	589.0 / 419.0	12570.82	250.00
NEtFOSAA_2	584.0 / 483.0	3.74	d5-EtFOSAA	589.0 / 419.0	12570.82	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	33656.28	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	33656.28	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	62523.36	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	62523.36	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	73739.89	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	73739.89	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22570.98	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22570.98	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	69511.85	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	69511.85	250.00
PFNA_1	463.0 / 419.0	3.07	13C9-PFNA	472.0 / 427.0	89868.53	250.00
PFNA_2	463.0 / 219.0	3.07	13C9-PFNA	472.0 / 427.0	89868.53	250.00
PFOS_1	499.0 / 80.0	2.94	13C8-PFOS	507.0 / 99.0	28330.00	239.25
PFOS_2	499.0 / 99.0	2.99	13C8-PFOS	507.0 / 99.0	28330.00	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	99308.03	250.00
PFDA_2	513.0 / 219.0	3.42	13C6-PFDA	519.0 / 474.0	99308.03	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	93211.96	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	93211.96	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	102365.92	250.00
PFDoA_2	613.0 / 319.0	4.02	13C2-PFDoA	615.0 / 570.0	102365.92	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	87922.52	250.00
PFTrDA_2	663.0 / 169.0	4.27	13C2-PFTeDA	715.0 / 670.0	87922.52	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	87922.52	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	87922.52	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14158.08	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14158.08	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14738.16	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14738.16	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	33425.41	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	33425.41	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	63253.79	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	63253.79	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	80959.19	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	80959.19	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	24629.02	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	24629.02	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	80568.43	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	80568.43	250.00
PFNA_1	463.0 / 419.0	3.06	13C9-PFNA	472.0 / 427.0	89621.40	250.00
PFNA_2	463.0 / 219.0	3.09	13C9-PFNA	472.0 / 427.0	89621.40	250.00
PFOS_1	499.0 / 80.0	2.94	13C8-PFOS	507.0 / 99.0	30615.75	239.25
PFOS_2	499.0 / 99.0	2.98	13C8-PFOS	507.0 / 99.0	30615.75	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	108190.24	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	108190.24	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	101183.17	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	101183.17	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	108169.44	250.00
PFDoA_2	613.0 / 319.0	4.04	13C2-PFDoA	615.0 / 570.0	108169.44	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	93296.57	250.00
PFTrDA_2	663.0 / 169.0	4.27	13C2-PFTeDA	715.0 / 670.0	93296.57	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	93296.57	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	93296.57	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	17251.44	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	17251.44	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16222.61	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16222.61	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:54:39 PM

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	28724.76	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	28724.76	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	62304.44	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	62304.44	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	70703.83	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	70703.83	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	23485.92	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	23485.92	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	86239.74	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	86239.74	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	95741.06	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	95741.06	250.00
PFOS_1	499.0 / 80.0	2.94	13C8-PFOS	507.0 / 99.0	32415.74	239.25
PFOS_2	499.0 / 99.0	2.97	13C8-PFOS	507.0 / 99.0	32415.74	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	101696.33	250.00
PFDA_2	513.0 / 219.0	3.43	13C6-PFDA	519.0 / 474.0	101696.33	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	100969.04	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	100969.04	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	98187.64	250.00
PFDoA_2	613.0 / 319.0	4.02	13C2-PFDoA	615.0 / 570.0	98187.64	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	84991.30	250.00
PFTrDA_2	663.0 / 169.0	4.27	13C2-PFTeDA	715.0 / 670.0	84991.30	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	84991.30	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	84991.30	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	16345.04	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	16345.04	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16250.31	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16250.31	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:54:43 PM

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	23306.08	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	23306.08	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	51377.14	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	51377.14	250.00
PFHpA_1	363.0 / 319.0	2.29	13C4-PFHpA	367.0 / 322.0	56998.41	250.00
PFHpA_2	363.0 / 169.0	2.29	13C4-PFHpA	367.0 / 322.0	56998.41	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	21342.54	236.50
PFHxS_2	399.0 / 99.0	2.31	13C3-PFHxS	402.0 / 99.0	21342.54	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	72105.27	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	72105.27	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	70178.02	250.00
PFNA_2	463.0 / 219.0	3.10	13C9-PFNA	472.0 / 427.0	70178.02	250.00
PFOS_1	499.0 / 80.0	N/A	13C8-PFOS	507.0 / 99.0	24586.35	239.25
PFOS_2	499.0 / 99.0	N/A	13C8-PFOS	507.0 / 99.0	24586.35	239.25
PFDA_1	513.0 / 469.0	3.41	13C6-PFDA	519.0 / 474.0	84820.33	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	84820.33	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	80003.86	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	80003.86	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	77356.10	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	77356.10	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	67578.35	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	67578.35	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	67578.35	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	67578.35	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11815.95	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11815.95	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14677.13	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14677.13	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:54:47 PM

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	21606.32	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	21606.32	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	54899.59	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	54899.59	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	54502.65	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	54502.65	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	20949.43	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	20949.43	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	72578.99	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	72578.99	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	74600.73	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	74600.73	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	23394.61	239.25
PFOS_2	499.0 / 99.0	3.06	13C8-PFOS	507.0 / 99.0	23394.61	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	82984.85	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	82984.85	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	78986.29	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	78986.29	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	82663.20	250.00
PFDoA_2	613.0 / 319.0	4.02	13C2-PFDoA	615.0 / 570.0	82663.20	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	66887.03	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	66887.03	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	66887.03	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	66887.03	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	12612.24	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	12612.24	250.00
NEtFOSAA_1	584.0 / 419.0	3.74	d5-EtFOSAA	589.0 / 419.0	11535.06	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	11535.06	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	19630.65	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	19630.65	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	61253.99	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	61253.99	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	71266.86	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	71266.86	250.00
PFHxS_1	399.0 / 80.0	2.29	13C3-PFHxS	402.0 / 99.0	22108.74	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	22108.74	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	69415.90	250.00
PFOA_2	413.0 / 169.0	2.62	13C8-PFOA	421.0 / 376.0	69415.90	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	73829.59	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	73829.59	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	19144.51	239.25
PFOS_2	499.0 / 99.0	3.00	13C8-PFOS	507.0 / 99.0	19144.51	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	86045.59	250.00
PFDA_2	513.0 / 219.0	3.41	13C6-PFDA	519.0 / 474.0	86045.59	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	84587.82	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	84587.82	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	88560.66	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	88560.66	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	76096.64	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	76096.64	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	76096.64	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	76096.64	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14155.30	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14155.30	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16590.71	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16590.71	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:28 PM

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	35233.91	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	35233.91	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	89181.38	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	89181.38	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	96746.77	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	96746.77	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	30577.21	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	30577.21	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	86472.58	250.00
PFOA_2	413.0 / 169.0	2.58	13C8-PFOA	421.0 / 376.0	86472.58	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	90930.74	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	90930.74	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	28276.24	239.25
PFOS_2	499.0 / 99.0	3.05	13C8-PFOS	507.0 / 99.0	28276.24	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	99312.33	250.00
PFDA_2	513.0 / 219.0	3.41	13C6-PFDA	519.0 / 474.0	99312.33	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	100653.69	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	100653.69	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	102547.94	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	102547.94	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	89623.80	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	89623.80	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	89623.80	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	89623.80	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14636.12	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14636.12	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	12794.44	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	12794.44	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:32 PM

Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	33921.02	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	33921.02	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	65817.50	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	65817.50	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	72696.21	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	72696.21	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	21891.72	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	21891.72	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	85146.36	250.00
PFOA_2	413.0 / 169.0	2.63	13C8-PFOA	421.0 / 376.0	85146.36	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	91454.28	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	91454.28	250.00
PFOS_1	499.0 / 80.0	2.96	13C8-PFOS	507.0 / 99.0	27345.27	239.25
PFOS_2	499.0 / 99.0	3.07	13C8-PFOS	507.0 / 99.0	27345.27	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	89135.49	250.00
PFDA_2	513.0 / 219.0	3.42	13C6-PFDA	519.0 / 474.0	89135.49	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	89910.20	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	89910.20	250.00
PFDoA_1	613.0 / 569.0	4.05	13C2-PFDoA	615.0 / 570.0	99811.26	250.00
PFDoA_2	613.0 / 319.0	4.00	13C2-PFDoA	615.0 / 570.0	99811.26	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	79893.80	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	79893.80	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	79893.80	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	79893.80	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	13208.06	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	13208.06	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14080.80	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14080.80	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:37 PM

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	33941.52	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	33941.52	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	86536.62	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	86536.62	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	97439.72	250.00
PFHpA_2	363.0 / 169.0	2.22	13C4-PFHpA	367.0 / 322.0	97439.72	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	29495.72	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	29495.72	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	93885.87	250.00
PFOA_2	413.0 / 169.0	2.58	13C8-PFOA	421.0 / 376.0	93885.87	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	87794.90	250.00
PFNA_2	463.0 / 219.0	3.07	13C9-PFNA	472.0 / 427.0	87794.90	250.00
PFOS_1	499.0 / 80.0	2.96	13C8-PFOS	507.0 / 99.0	31601.10	239.25
PFOS_2	499.0 / 99.0	3.06	13C8-PFOS	507.0 / 99.0	31601.10	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	99218.03	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	99218.03	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	95079.09	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	95079.09	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	88531.24	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	88531.24	250.00
PFTrDA_1	663.0 / 619.0	4.29	13C2-PFTeDA	715.0 / 670.0	79191.78	250.00
PFTrDA_2	663.0 / 169.0	4.28	13C2-PFTeDA	715.0 / 670.0	79191.78	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	79191.78	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	79191.78	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14511.81	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14511.81	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	12945.04	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	12945.04	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:40 PM

Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	35370.11	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	35370.11	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	63398.55	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	63398.55	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	73829.73	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	73829.73	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	21738.61	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	21738.61	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	88458.57	250.00
PFOA_2	413.0 / 169.0	2.59	13C8-PFOA	421.0 / 376.0	88458.57	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	85034.98	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	85034.98	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	30025.21	239.25
PFOS_2	499.0 / 99.0	3.05	13C8-PFOS	507.0 / 99.0	30025.21	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	95638.34	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	95638.34	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	89341.16	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	89341.16	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	96209.85	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	96209.85	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	78663.32	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	78663.32	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	78663.32	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	78663.32	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11855.96	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11855.96	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14453.59	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14453.59	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:44 PM

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	27987.33	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	27987.33	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	71124.42	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	71124.42	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	80791.58	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	80791.58	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	24308.58	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	24308.58	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	78989.42	250.00
PFOA_2	413.0 / 169.0	2.59	13C8-PFOA	421.0 / 376.0	78989.42	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	75315.71	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	75315.71	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	23420.44	239.25
PFOS_2	499.0 / 99.0	3.05	13C8-PFOS	507.0 / 99.0	23420.44	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	75466.65	250.00
PFDA_2	513.0 / 219.0	3.41	13C6-PFDA	519.0 / 474.0	75466.65	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	74274.11	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	74274.11	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	73269.56	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	73269.56	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	62945.01	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	62945.01	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	62945.01	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	62945.01	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11837.96	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11837.96	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	13417.17	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	13417.17	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:48 PM

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	31781.59	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	31781.59	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	63357.10	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	63357.10	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	67320.94	250.00
PFHpA_2	363.0 / 169.0	2.25	13C4-PFHpA	367.0 / 322.0	67320.94	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	21891.05	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	21891.05	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	84616.01	250.00
PFOA_2	413.0 / 169.0	2.59	13C8-PFOA	421.0 / 376.0	84616.01	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	82240.75	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	82240.75	250.00
PFOS_1	499.0 / 80.0	2.94	13C8-PFOS	507.0 / 99.0	28546.33	239.25
PFOS_2	499.0 / 99.0	3.03	13C8-PFOS	507.0 / 99.0	28546.33	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	91584.23	250.00
PFDA_2	513.0 / 219.0	3.45	13C6-PFDA	519.0 / 474.0	91584.23	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	91666.24	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	91666.24	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	86815.35	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	86815.35	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	78581.48	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	78581.48	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	78581.48	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	78581.48	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14366.25	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14366.25	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14842.49	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14842.49	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:56:52 PM

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	27142.61	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	27142.61	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	66932.34	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	66932.34	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	72492.47	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	72492.47	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	23917.34	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	23917.34	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	72251.83	250.00
PFOA_2	413.0 / 169.0	2.58	13C8-PFOA	421.0 / 376.0	72251.83	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	73093.75	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	73093.75	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	23852.73	239.25
PFOS_2	499.0 / 99.0	3.03	13C8-PFOS	507.0 / 99.0	23852.73	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	77509.54	250.00
PFDA_2	513.0 / 219.0	3.43	13C6-PFDA	519.0 / 474.0	77509.54	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	75190.02	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	75190.02	250.00
PFDoA_1	613.0 / 569.0	4.00	13C2-PFDoA	615.0 / 570.0	82646.80	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	82646.80	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	74520.78	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	74520.78	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	74520.78	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	74520.78	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	10374.41	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	10374.41	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	12115.89	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	12115.89	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	34566.65	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	34566.65	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	61134.03	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	61134.03	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	74841.70	250.00
PFHpA_2	363.0 / 169.0	2.25	13C4-PFHpA	367.0 / 322.0	74841.70	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22141.94	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22141.94	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	89923.34	250.00
PFOA_2	413.0 / 169.0	2.58	13C8-PFOA	421.0 / 376.0	89923.34	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	86697.00	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	86697.00	250.00
PFOS_1	499.0 / 80.0	2.95	13C8-PFOS	507.0 / 99.0	26983.87	239.25
PFOS_2	499.0 / 99.0	3.03	13C8-PFOS	507.0 / 99.0	26983.87	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	94098.44	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	94098.44	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	85607.61	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	85607.61	250.00
PFDoA_1	613.0 / 569.0	4.05	13C2-PFDoA	615.0 / 570.0	97301.17	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	97301.17	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	83552.25	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	83552.25	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	83552.25	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	83552.25	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	15464.61	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	15464.61	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	15492.84	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	15492.84	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	25131.27	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	25131.27	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	57044.86	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	57044.86	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	63873.63	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	63873.63	250.00
PFHxS_1	399.0 / 80.0	N/A	13C3-PFHxS	402.0 / 99.0	21968.91	236.50
PFHxS_2	399.0 / 99.0	N/A	13C3-PFHxS	402.0 / 99.0	21968.91	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	82178.36	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	82178.36	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	84493.92	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	84493.92	250.00
PFOS_1	499.0 / 80.0	N/A	13C8-PFOS	507.0 / 99.0	24650.10	239.25
PFOS_2	499.0 / 99.0	3.10	13C8-PFOS	507.0 / 99.0	24650.10	239.25
PFDA_1	513.0 / 469.0	3.42	13C6-PFDA	519.0 / 474.0	86268.54	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	86268.54	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	80444.85	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	80444.85	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	93821.26	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	93821.26	250.00
PFTrDA_1	663.0 / 619.0	4.26	13C2-PFTeDA	715.0 / 670.0	77728.09	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	77728.09	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	77728.09	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	77728.09	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	12666.19	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	12666.19	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	15404.41	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	15404.41	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	25219.13	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	25219.13	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	55540.71	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	55540.71	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	58814.37	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	58814.37	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22385.41	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	22385.41	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	76742.66	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	76742.66	250.00
PFNA_1	463.0 / 419.0	3.07	13C9-PFNA	472.0 / 427.0	79056.09	250.00
PFNA_2	463.0 / 219.0	3.09	13C9-PFNA	472.0 / 427.0	79056.09	250.00
PFOS_1	499.0 / 80.0	3.07	13C8-PFOS	507.0 / 99.0	26134.40	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	26134.40	239.25
PFDA_1	513.0 / 469.0	3.45	13C6-PFDA	519.0 / 474.0	82839.15	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	82839.15	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	79244.29	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	79244.29	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	83541.13	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	83541.13	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	65064.22	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	65064.22	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	65064.22	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	65064.22	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11149.08	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11149.08	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	12416.82	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	12416.82	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.53	13C3-PFBS	302.0 / 99.0	18029.29	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	18029.29	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	59903.97	250.00
PFHxA_2	313.0 / 119.0	1.85	13C5-PFHxA	318.0 / 273.0	59903.97	250.00
PFHpA_1	363.0 / 319.0	2.26	13C4-PFHpA	367.0 / 322.0	76718.73	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	76718.73	250.00
PFHxS_1	399.0 / 80.0	2.29	13C3-PFHxS	402.0 / 99.0	22414.91	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	22414.91	236.50
PFOA_1	413.0 / 369.0	2.67	13C8-PFOA	421.0 / 376.0	73205.39	250.00
PFOA_2	413.0 / 169.0	2.62	13C8-PFOA	421.0 / 376.0	73205.39	250.00
PFNA_1	463.0 / 419.0	3.04	13C9-PFNA	472.0 / 427.0	69019.64	250.00
PFNA_2	463.0 / 219.0	3.06	13C9-PFNA	472.0 / 427.0	69019.64	250.00
PFOS_1	499.0 / 80.0	3.05	13C8-PFOS	507.0 / 99.0	21845.18	239.25
PFOS_2	499.0 / 99.0	3.07	13C8-PFOS	507.0 / 99.0	21845.18	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	81585.93	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	81585.93	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	80073.69	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	80073.69	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	74290.96	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	74290.96	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	65501.43	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	65501.43	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	65501.43	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	65501.43	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	12502.68	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	12502.68	250.00
NEtFOSAA_1	584.0 / 419.0	3.74	d5-EtFOSAA	589.0 / 419.0	13559.07	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	13559.07	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:57:12 PM

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	36539.48	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	36539.48	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	68138.74	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	68138.74	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	80104.14	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	80104.14	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22417.21	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22417.21	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	89756.47	250.00
PFOA_2	413.0 / 169.0	2.63	13C8-PFOA	421.0 / 376.0	89756.47	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	81476.57	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	81476.57	250.00
PFOS_1	499.0 / 80.0	3.06	13C8-PFOS	507.0 / 99.0	27729.91	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	27729.91	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	94164.55	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	94164.55	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	91192.69	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	91192.69	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	94378.28	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	94378.28	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	79581.83	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	79581.83	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	79581.83	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	79581.83	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	13446.34	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	13446.34	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	17487.39	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	17487.39	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:55:52 PM

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.53	13C3-PFBS	302.0 / 99.0	18186.01	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	18186.01	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	51269.85	250.00
PFHxA_2	313.0 / 119.0	1.85	13C5-PFHxA	318.0 / 273.0	51269.85	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	61717.69	250.00
PFHpA_2	363.0 / 169.0	2.26	13C4-PFHpA	367.0 / 322.0	61717.69	250.00
PFHxS_1	399.0 / 80.0	2.29	13C3-PFHxS	402.0 / 99.0	20297.44	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	20297.44	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	59485.51	250.00
PFOA_2	413.0 / 169.0	2.65	13C8-PFOA	421.0 / 376.0	59485.51	250.00
PFNA_1	463.0 / 419.0	3.08	13C9-PFNA	472.0 / 427.0	67745.94	250.00
PFNA_2	463.0 / 219.0	3.07	13C9-PFNA	472.0 / 427.0	67745.94	250.00
PFOS_1	499.0 / 80.0	3.06	13C8-PFOS	507.0 / 99.0	16259.34	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	16259.34	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	79142.97	250.00
PFDA_2	513.0 / 219.0	3.42	13C6-PFDA	519.0 / 474.0	79142.97	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	78553.70	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	78553.70	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	84937.12	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	84937.12	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	69892.77	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	69892.77	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	69892.77	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	69892.77	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	12061.78	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	12061.78	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	14686.12	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14686.12	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:55:57 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	39312.91	232.25
PFBS_2	298.9 / 99.0	1.56	13C3-PFBS	302.0 / 99.0	39312.91	232.25
PFHxA_1	313.0 / 269.0	1.88	13C5-PFHxA	318.0 / 273.0	78572.65	250.00
PFHxA_2	313.0 / 119.0	1.89	13C5-PFHxA	318.0 / 273.0	78572.65	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	88015.64	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	88015.64	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	25098.30	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	25098.30	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	86324.44	250.00
PFOA_2	413.0 / 169.0	2.68	13C8-PFOA	421.0 / 376.0	86324.44	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	99122.80	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	99122.80	250.00
PFOS_1	499.0 / 80.0	3.10	13C8-PFOS	507.0 / 99.0	29343.58	239.25
PFOS_2	499.0 / 99.0	3.11	13C8-PFOS	507.0 / 99.0	29343.58	239.25
PFDA_1	513.0 / 469.0	3.46	13C6-PFDA	519.0 / 474.0	99618.75	250.00
PFDA_2	513.0 / 219.0	3.49	13C6-PFDA	519.0 / 474.0	99618.75	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	93934.76	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	93934.76	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	104011.35	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	104011.35	250.00
PFTrDA_1	663.0 / 619.0	4.31	13C2-PFTeDA	715.0 / 670.0	89235.46	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	89235.46	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	89235.46	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	89235.46	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11213.24	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11213.24	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	13207.15	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	13207.15	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:56:01 PM

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	17605.10	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	17605.10	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	48513.17	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	48513.17	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	65151.03	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	65151.03	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	20407.49	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	20407.49	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	56775.58	250.00
PFOA_2	413.0 / 169.0	2.65	13C8-PFOA	421.0 / 376.0	56775.58	250.00
PFNA_1	463.0 / 419.0	3.09	13C9-PFNA	472.0 / 427.0	63436.48	250.00
PFNA_2	463.0 / 219.0	3.08	13C9-PFNA	472.0 / 427.0	63436.48	250.00
PFOS_1	499.0 / 80.0	3.07	13C8-PFOS	507.0 / 99.0	17657.30	239.25
PFOS_2	499.0 / 99.0	3.09	13C8-PFOS	507.0 / 99.0	17657.30	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	80282.23	250.00
PFDA_2	513.0 / 219.0	3.43	13C6-PFDA	519.0 / 474.0	80282.23	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	76614.04	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	76614.04	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	75285.88	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	75285.88	250.00
PFTrDA_1	663.0 / 619.0	4.29	13C2-PFTeDA	715.0 / 670.0	68631.64	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	68631.64	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	68631.64	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	68631.64	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	11925.58	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	11925.58	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	15331.17	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	15331.17	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	38502.79	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	38502.79	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	78098.37	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	78098.37	250.00
PFHpA_1	363.0 / 319.0	2.28	13C4-PFHpA	367.0 / 322.0	87662.25	250.00
PFHpA_2	363.0 / 169.0	2.28	13C4-PFHpA	367.0 / 322.0	87662.25	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	26949.91	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	26949.91	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	87351.24	250.00
PFOA_2	413.0 / 169.0	2.66	13C8-PFOA	421.0 / 376.0	87351.24	250.00
PFNA_1	463.0 / 419.0	3.09	13C9-PFNA	472.0 / 427.0	93161.34	250.00
PFNA_2	463.0 / 219.0	3.09	13C9-PFNA	472.0 / 427.0	93161.34	250.00
PFOS_1	499.0 / 80.0	3.08	13C8-PFOS	507.0 / 99.0	28013.11	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	28013.11	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	101005.94	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	101005.94	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	95017.72	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	95017.72	250.00
PFDoA_1	613.0 / 569.0	4.04	13C2-PFDoA	615.0 / 570.0	97309.37	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	97309.37	250.00
PFTrDA_1	663.0 / 619.0	4.30	13C2-PFTeDA	715.0 / 670.0	84876.94	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	84876.94	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	84876.94	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	84876.94	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14016.38	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14016.38	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	20120.98	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	20120.98	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:56:09 PM

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.53	13C3-PFBS	302.0 / 99.0	17691.12	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	17691.12	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	52399.67	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	52399.67	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	77505.88	250.00
PFHpA_2	363.0 / 169.0	2.23	13C4-PFHpA	367.0 / 322.0	77505.88	250.00
PFHxS_1	399.0 / 80.0	2.28	13C3-PFHxS	402.0 / 99.0	21434.90	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	21434.90	236.50
PFOA_1	413.0 / 369.0	2.69	13C8-PFOA	421.0 / 376.0	71032.21	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	71032.21	250.00
PFNA_1	463.0 / 419.0	3.00	13C9-PFNA	472.0 / 427.0	70841.50	250.00
PFNA_2	463.0 / 219.0	2.97	13C9-PFNA	472.0 / 427.0	70841.50	250.00
PFOS_1	499.0 / 80.0	3.02	13C8-PFOS	507.0 / 99.0	23131.42	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	23131.42	239.25
PFDA_1	513.0 / 469.0	N/A	13C6-PFDA	519.0 / 474.0	85172.75	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	85172.75	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	76205.68	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	76205.68	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	74184.66	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	74184.66	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	61983.03	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	61983.03	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	61983.03	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	61983.03	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	13381.71	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	13381.71	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	12899.82	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	12899.82	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:56:12 PM

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	35183.62	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	35183.62	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	80215.79	250.00
PFHxA_2	313.0 / 119.0	1.86	13C5-PFHxA	318.0 / 273.0	80215.79	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	91868.65	250.00
PFHpA_2	363.0 / 169.0	2.24	13C4-PFHpA	367.0 / 322.0	91868.65	250.00
PFHxS_1	399.0 / 80.0	2.27	13C3-PFHxS	402.0 / 99.0	24830.16	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	24830.16	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	95016.44	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	95016.44	250.00
PFNA_1	463.0 / 419.0	3.07	13C9-PFNA	472.0 / 427.0	92944.24	250.00
PFNA_2	463.0 / 219.0	3.05	13C9-PFNA	472.0 / 427.0	92944.24	250.00
PFOS_1	499.0 / 80.0	3.00	13C8-PFOS	507.0 / 99.0	27430.15	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	27430.15	239.25
PFDA_1	513.0 / 469.0	3.44	13C6-PFDA	519.0 / 474.0	99770.94	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	99770.94	250.00
PFUnA_1	563.0 / 519.0	3.74	13C7-PFUnA	570.0 / 525.0	96181.53	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	96181.53	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	98899.04	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	98899.04	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	90902.35	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	90902.35	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	90902.35	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	90902.35	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	16317.45	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	16317.45	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	15503.33	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	15503.33	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 2:57:39 PM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.53	13C3-PFBS	302.0 / 99.0	18019.37	232.25
PFBS_2	298.9 / 99.0	1.53	13C3-PFBS	302.0 / 99.0	18019.37	232.25
PFHxA_1	313.0 / 269.0	1.85	13C5-PFHxA	318.0 / 273.0	46341.61	250.00
PFHxA_2	313.0 / 119.0	1.85	13C5-PFHxA	318.0 / 273.0	46341.61	250.00
PFHpA_1	363.0 / 319.0	2.26	13C4-PFHpA	367.0 / 322.0	60573.15	250.00
PFHpA_2	363.0 / 169.0	2.23	13C4-PFHpA	367.0 / 322.0	60573.15	250.00
PFHxS_1	399.0 / 80.0	2.29	13C3-PFHxS	402.0 / 99.0	21187.02	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	21187.02	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	63664.07	250.00
PFOA_2	413.0 / 169.0	2.62	13C8-PFOA	421.0 / 376.0	63664.07	250.00
PFNA_1	463.0 / 419.0	3.02	13C9-PFNA	472.0 / 427.0	62009.43	250.00
PFNA_2	463.0 / 219.0	2.97	13C9-PFNA	472.0 / 427.0	62009.43	250.00
PFOS_1	499.0 / 80.0	2.99	13C8-PFOS	507.0 / 99.0	17070.16	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	17070.16	239.25
PFDA_1	513.0 / 469.0	3.39	13C6-PFDA	519.0 / 474.0	86229.83	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	86229.83	250.00
PFUnA_1	563.0 / 519.0	3.76	13C7-PFUnA	570.0 / 525.0	80755.15	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	80755.15	250.00
PFDoA_1	613.0 / 569.0	4.03	13C2-PFDoA	615.0 / 570.0	80984.56	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	80984.56	250.00
PFTrDA_1	663.0 / 619.0	4.27	13C2-PFTeDA	715.0 / 670.0	66356.21	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	66356.21	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	66356.21	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	66356.21	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14495.61	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14495.61	250.00
NEtFOSAA_1	584.0 / 419.0	3.76	d5-EtFOSAA	589.0 / 419.0	14454.05	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	14454.05	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.54	13C3-PFBS	302.0 / 99.0	37706.48	232.25
PFBS_2	298.9 / 99.0	1.54	13C3-PFBS	302.0 / 99.0	37706.48	232.25
PFHxA_1	313.0 / 269.0	1.86	13C5-PFHxA	318.0 / 273.0	72106.09	250.00
PFHxA_2	313.0 / 119.0	1.87	13C5-PFHxA	318.0 / 273.0	72106.09	250.00
PFHpA_1	363.0 / 319.0	2.27	13C4-PFHpA	367.0 / 322.0	83228.08	250.00
PFHpA_2	363.0 / 169.0	2.21	13C4-PFHpA	367.0 / 322.0	83228.08	250.00
PFHxS_1	399.0 / 80.0	2.30	13C3-PFHxS	402.0 / 99.0	22273.24	236.50
PFHxS_2	399.0 / 99.0	2.30	13C3-PFHxS	402.0 / 99.0	22273.24	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	92242.09	250.00
PFOA_2	413.0 / 169.0	2.62	13C8-PFOA	421.0 / 376.0	92242.09	250.00
PFNA_1	463.0 / 419.0	3.07	13C9-PFNA	472.0 / 427.0	87761.87	250.00
PFNA_2	463.0 / 219.0	3.07	13C9-PFNA	472.0 / 427.0	87761.87	250.00
PFOS_1	499.0 / 80.0	3.00	13C8-PFOS	507.0 / 99.0	25590.94	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	25590.94	239.25
PFDA_1	513.0 / 469.0	3.43	13C6-PFDA	519.0 / 474.0	103309.63	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	103309.63	250.00
PFUnA_1	563.0 / 519.0	3.75	13C7-PFUnA	570.0 / 525.0	95482.74	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	95482.74	250.00
PFDoA_1	613.0 / 569.0	4.02	13C2-PFDoA	615.0 / 570.0	98168.20	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	98168.20	250.00
PFTrDA_1	663.0 / 619.0	4.28	13C2-PFTeDA	715.0 / 670.0	85544.11	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	85544.11	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	85544.11	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	85544.11	250.00
NMeFOSAA_1	570.0 / 419.0	3.59	d3-MeFOSAA	573.0 / 419.0	15140.84	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	15140.84	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16379.03	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16379.03	250.00

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:56:23 PM

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID	VC-S14GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	8767.75	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	8767.75	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	21215.73	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	21215.73	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	33553.99	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	33553.99	250.00
PFHxS_1	399.0 / 80.0	2.29	13C3-PFHxS	402.0 / 99.0	12512.63	236.50
PFHxS_2	399.0 / 99.0	2.29	13C3-PFHxS	402.0 / 99.0	12512.63	236.50
PFOA_1	413.0 / 369.0	2.68	13C8-PFOA	421.0 / 376.0	36105.85	250.00
PFOA_2	413.0 / 169.0	2.64	13C8-PFOA	421.0 / 376.0	36105.85	250.00
PFNA_1	463.0 / 419.0	N/A	13C9-PFNA	472.0 / 427.0	40704.57	250.00
PFNA_2	463.0 / 219.0	N/A	13C9-PFNA	472.0 / 427.0	40704.57	250.00
PFOS_1	499.0 / 80.0	3.03	13C8-PFOS	507.0 / 99.0	13365.40	239.25
PFOS_2	499.0 / 99.0	3.08	13C8-PFOS	507.0 / 99.0	13365.40	239.25
PFDA_1	513.0 / 469.0	N/A	13C6-PFDA	519.0 / 474.0	47197.29	250.00
PFDA_2	513.0 / 219.0	3.44	13C6-PFDA	519.0 / 474.0	47197.29	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	50168.76	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	50168.76	250.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFDoA	615.0 / 570.0	49934.63	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	49934.63	250.00
PFTrDA_1	663.0 / 619.0	N/A	13C2-PFTeDA	715.0 / 670.0	34458.50	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	34458.50	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	34458.50	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	34458.50	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	9280.08	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	9280.08	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	9370.14	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	9370.14	250.00

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	32827.89	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	32827.89	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	77874.50	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	77874.50	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	87915.24	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	87915.24	250.00
PFHxS_1	399.0 / 80.0	N/A	13C3-PFHxS	402.0 / 99.0	29416.03	236.50
PFHxS_2	399.0 / 99.0	N/A	13C3-PFHxS	402.0 / 99.0	29416.03	236.50
PFOA_1	413.0 / 369.0	2.72	13C8-PFOA	421.0 / 376.0	99152.64	250.00
PFOA_2	413.0 / 169.0	2.72	13C8-PFOA	421.0 / 376.0	99152.64	250.00
PFNA_1	463.0 / 419.0	N/A	13C9-PFNA	472.0 / 427.0	111508.83	250.00
PFNA_2	463.0 / 219.0	N/A	13C9-PFNA	472.0 / 427.0	111508.83	250.00
PFOS_1	499.0 / 80.0	3.09	13C8-PFOS	507.0 / 99.0	33372.79	239.25
PFOS_2	499.0 / 99.0	N/A	13C8-PFOS	507.0 / 99.0	33372.79	239.25
PFDA_1	513.0 / 469.0	3.48	13C6-PFDA	519.0 / 474.0	108469.53	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	108469.53	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	110417.79	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	110417.79	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	124281.55	250.00
PFDoA_2	613.0 / 319.0	4.08	13C2-PFDoA	615.0 / 570.0	124281.55	250.00
PFTrDA_1	663.0 / 619.0	4.34	13C2-PFTeDA	715.0 / 670.0	106949.79	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	106949.79	250.00
PFTeDA_1	713.0 / 669.0	4.54	13C2-PFTeDA	715.0 / 670.0	106949.79	250.00
PFTeDA_2	713.0 / 169.0	4.55	13C2-PFTeDA	715.0 / 670.0	106949.79	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	19708.38	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	19708.38	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	19501.12	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	19501.12	250.00

Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.56	13C3-PFBS	302.0 / 99.0	30400.45	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	30400.45	232.25
PFHxA_1	313.0 / 269.0	1.88	13C5-PFHxA	318.0 / 273.0	59637.72	250.00
PFHxA_2	313.0 / 119.0	1.88	13C5-PFHxA	318.0 / 273.0	59637.72	250.00
PFHpA_1	363.0 / 319.0	2.30	13C4-PFHpA	367.0 / 322.0	64236.43	250.00
PFHpA_2	363.0 / 169.0	2.30	13C4-PFHpA	367.0 / 322.0	64236.43	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	22429.47	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	22429.47	236.50
PFOA_1	413.0 / 369.0	2.71	13C8-PFOA	421.0 / 376.0	73720.34	250.00
PFOA_2	413.0 / 169.0	2.71	13C8-PFOA	421.0 / 376.0	73720.34	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	89503.09	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	89503.09	250.00
PFOS_1	499.0 / 80.0	3.11	13C8-PFOS	507.0 / 99.0	26438.92	239.25
PFOS_2	499.0 / 99.0	3.11	13C8-PFOS	507.0 / 99.0	26438.92	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	87249.50	250.00
PFDA_2	513.0 / 219.0	3.47	13C6-PFDA	519.0 / 474.0	87249.50	250.00
PFUnA_1	563.0 / 519.0	3.80	13C7-PFUnA	570.0 / 525.0	79452.86	250.00
PFUnA_2	563.0 / 269.0	3.79	13C7-PFUnA	570.0 / 525.0	79452.86	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	102412.42	250.00
PFDoA_2	613.0 / 319.0	4.07	13C2-PFDoA	615.0 / 570.0	102412.42	250.00
PFTrDA_1	663.0 / 619.0	4.32	13C2-PFTeDA	715.0 / 670.0	83161.42	250.00
PFTrDA_2	663.0 / 169.0	4.32	13C2-PFTeDA	715.0 / 670.0	83161.42	250.00
PFTeDA_1	713.0 / 669.0	4.54	13C2-PFTeDA	715.0 / 670.0	83161.42	250.00
PFTeDA_2	713.0 / 169.0	4.53	13C2-PFTeDA	715.0 / 670.0	83161.42	250.00
NMeFOSAA_1	570.0 / 419.0	3.63	d3-MeFOSAA	573.0 / 419.0	14415.46	250.00
NMeFOSAA_2	570.0 / 512.0	3.62	d3-MeFOSAA	573.0 / 419.0	14415.46	250.00
NEtFOSAA_1	584.0 / 419.0	3.79	d5-EtFOSAA	589.0 / 419.0	15758.27	250.00
NEtFOSAA_2	584.0 / 483.0	3.79	d5-EtFOSAA	589.0 / 419.0	15758.27	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.55	13C3-PFBS	302.0 / 99.0	40801.04	232.25
PFBS_2	298.9 / 99.0	1.55	13C3-PFBS	302.0 / 99.0	40801.04	232.25
PFHxA_1	313.0 / 269.0	1.87	13C5-PFHxA	318.0 / 273.0	82111.69	250.00
PFHxA_2	313.0 / 119.0	1.88	13C5-PFHxA	318.0 / 273.0	82111.69	250.00
PFHpA_1	363.0 / 319.0	2.29	13C4-PFHpA	367.0 / 322.0	98639.40	250.00
PFHpA_2	363.0 / 169.0	2.27	13C4-PFHpA	367.0 / 322.0	98639.40	250.00
PFHxS_1	399.0 / 80.0	2.32	13C3-PFHxS	402.0 / 99.0	28517.83	236.50
PFHxS_2	399.0 / 99.0	2.32	13C3-PFHxS	402.0 / 99.0	28517.83	236.50
PFOA_1	413.0 / 369.0	2.70	13C8-PFOA	421.0 / 376.0	102244.02	250.00
PFOA_2	413.0 / 169.0	2.69	13C8-PFOA	421.0 / 376.0	102244.02	250.00
PFNA_1	463.0 / 419.0	3.11	13C9-PFNA	472.0 / 427.0	107445.94	250.00
PFNA_2	463.0 / 219.0	3.11	13C9-PFNA	472.0 / 427.0	107445.94	250.00
PFOS_1	499.0 / 80.0	2.98	13C8-PFOS	507.0 / 99.0	33228.76	239.25
PFOS_2	499.0 / 99.0	3.03	13C8-PFOS	507.0 / 99.0	33228.76	239.25
PFDA_1	513.0 / 469.0	3.47	13C6-PFDA	519.0 / 474.0	112886.63	250.00
PFDA_2	513.0 / 219.0	3.50	13C6-PFDA	519.0 / 474.0	112886.63	250.00
PFUnA_1	563.0 / 519.0	3.79	13C7-PFUnA	570.0 / 525.0	114328.90	250.00
PFUnA_2	563.0 / 269.0	3.80	13C7-PFUnA	570.0 / 525.0	114328.90	250.00
PFDoA_1	613.0 / 569.0	4.07	13C2-PFDoA	615.0 / 570.0	122058.68	250.00
PFDoA_2	613.0 / 319.0	4.07	13C2-PFDoA	615.0 / 570.0	122058.68	250.00
PFTrDA_1	663.0 / 619.0	4.32	13C2-PFTeDA	715.0 / 670.0	107110.25	250.00
PFTrDA_2	663.0 / 169.0	4.32	13C2-PFTeDA	715.0 / 670.0	107110.25	250.00
PFTeDA_1	713.0 / 669.0	4.53	13C2-PFTeDA	715.0 / 670.0	107110.25	250.00
PFTeDA_2	713.0 / 169.0	4.54	13C2-PFTeDA	715.0 / 670.0	107110.25	250.00
NMeFOSAA_1	570.0 / 419.0	3.62	d3-MeFOSAA	573.0 / 419.0	20341.58	250.00
NMeFOSAA_2	570.0 / 512.0	3.63	d3-MeFOSAA	573.0 / 419.0	20341.58	250.00
NEtFOSAA_1	584.0 / 419.0	3.78	d5-EtFOSAA	589.0 / 419.0	20083.34	250.00
NEtFOSAA_2	584.0 / 483.0	3.78	d5-EtFOSAA	589.0 / 419.0	20083.34	250.00

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	N/A	13C3-PFBS	302.0 / 99.0	30785.58	232.25
PFBS_2	298.9 / 99.0	N/A	13C3-PFBS	302.0 / 99.0	30785.58	232.25
PFHxA_1	313.0 / 269.0	N/A	13C5-PFHxA	318.0 / 273.0	64289.99	250.00
PFHxA_2	313.0 / 119.0	N/A	13C5-PFHxA	318.0 / 273.0	64289.99	250.00
PFHpA_1	363.0 / 319.0	N/A	13C4-PFHpA	367.0 / 322.0	70155.09	250.00
PFHpA_2	363.0 / 169.0	N/A	13C4-PFHpA	367.0 / 322.0	70155.09	250.00
PFHxS_1	399.0 / 80.0	2.31	13C3-PFHxS	402.0 / 99.0	24873.64	236.50
PFHxS_2	399.0 / 99.0	2.31	13C3-PFHxS	402.0 / 99.0	24873.64	236.50
PFOA_1	413.0 / 369.0	N/A	13C8-PFOA	421.0 / 376.0	85385.52	250.00
PFOA_2	413.0 / 169.0	N/A	13C8-PFOA	421.0 / 376.0	85385.52	250.00
PFNA_1	463.0 / 419.0	N/A	13C9-PFNA	472.0 / 427.0	92550.97	250.00
PFNA_2	463.0 / 219.0	N/A	13C9-PFNA	472.0 / 427.0	92550.97	250.00
PFOS_1	499.0 / 80.0	N/A	13C8-PFOS	507.0 / 99.0	27672.88	239.25
PFOS_2	499.0 / 99.0	N/A	13C8-PFOS	507.0 / 99.0	27672.88	239.25
PFDA_1	513.0 / 469.0	N/A	13C6-PFDA	519.0 / 474.0	91658.72	250.00
PFDA_2	513.0 / 219.0	N/A	13C6-PFDA	519.0 / 474.0	91658.72	250.00
PFUnA_1	563.0 / 519.0	N/A	13C7-PFUnA	570.0 / 525.0	92121.78	250.00
PFUnA_2	563.0 / 269.0	N/A	13C7-PFUnA	570.0 / 525.0	92121.78	250.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFDoA	615.0 / 570.0	98065.26	250.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFDoA	615.0 / 570.0	98065.26	250.00
PFTrDA_1	663.0 / 619.0	N/A	13C2-PFTeDA	715.0 / 670.0	87321.85	250.00
PFTrDA_2	663.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	87321.85	250.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFTeDA	715.0 / 670.0	87321.85	250.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFTeDA	715.0 / 670.0	87321.85	250.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	14735.20	250.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	14735.20	250.00
NEtFOSAA_1	584.0 / 419.0	N/A	d5-EtFOSAA	589.0 / 419.0	16222.53	250.00
NEtFOSAA_2	584.0 / 483.0	N/A	d5-EtFOSAA	589.0 / 419.0	16222.53	250.00

BATTELLE

Summary Internal Standard Report

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.07	13C2-PFDA	515.0 / 470.0	96280.28	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	29371.12	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	29371.12	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	85730.94	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	85730.94	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	85730.94	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	85730.94	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	96280.28	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	96280.28	250.00
13C2-PFTeDA	715.0 / 670.0	4.54	13C2-PFDA	515.0 / 470.0	96280.28	250.00
13C3-PFBS	302.0 / 99.0	1.55	13C4-PFOS	503.0 / 99.0	29371.12	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	29371.12	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	29371.12	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 01/11/2018 6:20:55 AM

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.13	13C2-PFDA	515.0 / 470.0	91293.54	250.00
d3-MeFOSAA	573.0 / 419.0	3.67	13C4-PFOS	503.0 / 99.0	27153.34	239.25
d5-EtFOSAA	589.0 / 419.0	3.83	13C4-PFOS	503.0 / 99.0	27153.34	239.25
13C5-PFHxA	318.0 / 273.0	1.90	13C2-PFOA	415.0 / 370.0	70816.22	250.00
13C4-PFHpA	367.0 / 322.0	2.32	13C2-PFOA	415.0 / 370.0	70816.22	250.00
13C8-PFOA	421.0 / 376.0	2.74	13C2-PFOA	415.0 / 370.0	70816.22	250.00
13C9-PFNA	472.0 / 427.0	3.14	13C2-PFOA	415.0 / 370.0	70816.22	250.00
13C6-PFDA	519.0 / 474.0	3.51	13C2-PFDA	515.0 / 470.0	91293.54	250.00
13C7-PFUnA	570.0 / 525.0	3.83	13C2-PFDA	515.0 / 470.0	91293.54	250.00
13C2-PFTeDA	715.0 / 670.0	4.61	13C2-PFDA	515.0 / 470.0	91293.54	250.00
13C3-PFBS	302.0 / 99.0	1.56	13C4-PFOS	503.0 / 99.0	27153.34	239.25
13C3-PFHxS	402.0 / 99.0	2.35	13C4-PFOS	503.0 / 99.0	27153.34	239.25
13C8-PFOS	507.0 / 99.0	3.14	13C4-PFOS	503.0 / 99.0	27153.34	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	102442.20	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	29773.88	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	29773.88	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	82215.13	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	82215.13	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	82215.13	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	82215.13	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	102442.20	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	102442.20	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	102442.20	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	29773.88	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	29773.88	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	29773.88	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:52:02 PM

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	96413.31	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	24859.22	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	24859.22	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	266029.78	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	266029.78	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	266029.78	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	266029.78	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	96413.31	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	96413.31	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	96413.31	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	24859.22	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	24859.22	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	24859.22	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	102089.87	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	28050.74	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	28050.74	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	108889.36	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	108889.36	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	108889.36	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	108889.36	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	102089.87	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	102089.87	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	102089.87	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	28050.74	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	28050.74	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	28050.74	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	99786.05	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	30703.24	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	30703.24	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	87385.04	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	87385.04	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	87385.04	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	87385.04	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	99786.05	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	99786.05	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	99786.05	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	30703.24	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	30703.24	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	30703.24	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	88204.29	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	30804.31	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	30804.31	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	77259.53	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	77259.53	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	77259.53	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	77259.53	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	88204.29	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	88204.29	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	88204.29	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	30804.31	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	30804.31	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	30804.31	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	92253.71	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	28857.59	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	28857.59	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	81605.91	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	81605.91	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	81605.91	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	81605.91	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	92253.71	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	92253.71	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	92253.71	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	28857.59	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	28857.59	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	28857.59	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	100021.36	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	32731.62	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	32731.62	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	77824.78	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	77824.78	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	77824.78	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	77824.78	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	100021.36	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	100021.36	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	100021.36	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	32731.62	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	32731.62	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	32731.62	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	97426.57	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	26773.90	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	26773.90	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	85784.84	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	85784.84	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	85784.84	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	85784.84	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	97426.57	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	97426.57	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	97426.57	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	26773.90	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	26773.90	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	26773.90	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	115359.89	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	29782.00	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	29782.00	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	95284.80	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	95284.80	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	95284.80	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	95284.80	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	115359.89	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	115359.89	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	115359.89	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	29782.00	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	29782.00	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	29782.00	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	92237.23	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	27075.35	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	27075.35	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	78203.75	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	78203.75	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	78203.75	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	78203.75	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	92237.23	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	92237.23	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	92237.23	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	27075.35	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	27075.35	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	27075.35	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	120367.15	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	34948.68	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	34948.68	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	99109.90	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	99109.90	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	99109.90	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	99109.90	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	120367.15	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	120367.15	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	120367.15	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	34948.68	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	34948.68	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	34948.68	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	96628.38	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	31106.92	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	31106.92	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	83507.54	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	83507.54	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	83507.54	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	83507.54	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	96628.38	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	96628.38	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	96628.38	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	31106.92	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	31106.92	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	31106.92	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	99465.21	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	30593.68	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	30593.68	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	78450.85	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	78450.85	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	78450.85	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	78450.85	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	99465.21	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	99465.21	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	99465.21	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	30593.68	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	30593.68	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	30593.68	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	95242.23	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	30626.59	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	30626.59	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	76829.11	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	76829.11	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	76829.11	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	76829.11	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	95242.23	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	95242.23	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	95242.23	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	30626.59	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	30626.59	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	30626.59	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	100808.73	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	26558.92	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	26558.92	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	82880.95	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	82880.95	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	82880.95	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	82880.95	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	100808.73	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	100808.73	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	100808.73	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	26558.92	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	26558.92	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	26558.92	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	94373.84	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	30284.94	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	30284.94	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	80118.83	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	80118.83	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	80118.83	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	80118.83	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	94373.84	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	94373.84	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	94373.84	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	30284.94	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	30284.94	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	30284.94	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	104909.04	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	34544.23	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	34544.23	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	85941.52	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	85941.52	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	85941.52	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	85941.52	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	104909.04	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	104909.04	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	104909.04	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	34544.23	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	34544.23	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	34544.23	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:53:13 PM

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	101274.22	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	32365.60	239.25
d5-EtFOSAA	589.0 / 419.0	3.73	13C4-PFOS	503.0 / 99.0	32365.60	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	79902.67	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	79902.67	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	79902.67	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	79902.67	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	101274.22	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	101274.22	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	101274.22	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	32365.60	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	32365.60	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	32365.60	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	107653.70	250.00
d3-MeFOSAA	573.0 / 419.0	3.57	13C4-PFOS	503.0 / 99.0	27775.17	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	27775.17	239.25
13C5-PFHxA	318.0 / 273.0	1.84	13C2-PFOA	415.0 / 370.0	88712.19	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	88712.19	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	88712.19	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	88712.19	250.00
13C6-PFDA	519.0 / 474.0	3.41	13C2-PFDA	515.0 / 470.0	107653.70	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	107653.70	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	107653.70	250.00
13C3-PFBS	302.0 / 99.0	1.51	13C4-PFOS	503.0 / 99.0	27775.17	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	27775.17	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	27775.17	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	94042.15	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	27638.21	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	27638.21	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	76513.66	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	76513.66	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	76513.66	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	76513.66	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	94042.15	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	94042.15	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	94042.15	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	27638.21	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	27638.21	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	27638.21	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	88819.82	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	21821.01	239.25
d5-EtFOSAA	589.0 / 419.0	3.73	13C4-PFOS	503.0 / 99.0	21821.01	239.25
13C5-PFHxA	318.0 / 273.0	1.84	13C2-PFOA	415.0 / 370.0	93151.25	250.00
13C4-PFHpA	367.0 / 322.0	2.25	13C2-PFOA	415.0 / 370.0	93151.25	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	93151.25	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	93151.25	250.00
13C6-PFDA	519.0 / 474.0	3.41	13C2-PFDA	515.0 / 470.0	88819.82	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	88819.82	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	88819.82	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	21821.01	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	21821.01	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	21821.01	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:53:27 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.05	13C2-PFDA	515.0 / 470.0	95495.07	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	32802.29	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	32802.29	239.25
13C5-PFHxA	318.0 / 273.0	1.88	13C2-PFOA	415.0 / 370.0	81244.05	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	81244.05	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	81244.05	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	81244.05	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	95495.07	250.00
13C7-PFUnA	570.0 / 525.0	3.77	13C2-PFDA	515.0 / 470.0	95495.07	250.00
13C2-PFTeDA	715.0 / 670.0	4.52	13C2-PFDA	515.0 / 470.0	95495.07	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	32802.29	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	32802.29	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	32802.29	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	89409.45	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	22601.04	239.25
d5-EtFOSAA	589.0 / 419.0	3.75	13C4-PFOS	503.0 / 99.0	22601.04	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	85332.38	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	85332.38	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	85332.38	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	85332.38	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	89409.45	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	89409.45	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	89409.45	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	22601.04	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	22601.04	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	22601.04	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	98089.53	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	29612.80	239.25
d5-EtFOSAA	589.0 / 419.0	3.75	13C4-PFOS	503.0 / 99.0	29612.80	239.25
13C5-PFHxA	318.0 / 273.0	1.86	13C2-PFOA	415.0 / 370.0	83619.59	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	83619.59	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	83619.59	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	83619.59	250.00
13C6-PFDA	519.0 / 474.0	3.43	13C2-PFDA	515.0 / 470.0	98089.53	250.00
13C7-PFUnA	570.0 / 525.0	3.75	13C2-PFDA	515.0 / 470.0	98089.53	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	98089.53	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	29612.80	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	29612.80	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	29612.80	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	106485.79	250.00
d3-MeFOSAA	573.0 / 419.0	3.59	13C4-PFOS	503.0 / 99.0	23461.57	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	23461.57	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	84304.23	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	84304.23	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	84304.23	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	84304.23	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	106485.79	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	106485.79	250.00
13C2-PFTeDA	715.0 / 670.0	4.49	13C2-PFDA	515.0 / 470.0	106485.79	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	23461.57	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	23461.57	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	23461.57	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.03	13C2-PFDA	515.0 / 470.0	98829.36	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	28724.47	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	28724.47	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	84872.05	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	84872.05	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	84872.05	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	84872.05	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	98829.36	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	98829.36	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	98829.36	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	28724.47	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	28724.47	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	28724.47	239.25

BATTELLE

Summary Internal Standard Report

Created with Analyst Reporter
Printed: 25/10/2018 1:53:46 PM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	92228.61	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	23155.03	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	23155.03	239.25
13C5-PFHxA	318.0 / 273.0	1.84	13C2-PFOA	415.0 / 370.0	78988.28	250.00
13C4-PFHpA	367.0 / 322.0	2.26	13C2-PFOA	415.0 / 370.0	78988.28	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	78988.28	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	78988.28	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	92228.61	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	92228.61	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	92228.61	250.00
13C3-PFBS	302.0 / 99.0	1.52	13C4-PFOS	503.0 / 99.0	23155.03	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	23155.03	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	23155.03	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	96491.53	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	33386.85	239.25
d5-EtFOSAA	589.0 / 419.0	3.74	13C4-PFOS	503.0 / 99.0	33386.85	239.25
13C5-PFHxA	318.0 / 273.0	1.85	13C2-PFOA	415.0 / 370.0	82763.70	250.00
13C4-PFHpA	367.0 / 322.0	2.27	13C2-PFOA	415.0 / 370.0	82763.70	250.00
13C8-PFOA	421.0 / 376.0	2.68	13C2-PFOA	415.0 / 370.0	82763.70	250.00
13C9-PFNA	472.0 / 427.0	3.07	13C2-PFOA	415.0 / 370.0	82763.70	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	96491.53	250.00
13C7-PFUnA	570.0 / 525.0	3.74	13C2-PFDA	515.0 / 470.0	96491.53	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	96491.53	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	33386.85	239.25
13C3-PFHxS	402.0 / 99.0	2.29	13C4-PFOS	503.0 / 99.0	33386.85	239.25
13C8-PFOS	507.0 / 99.0	3.07	13C4-PFOS	503.0 / 99.0	33386.85	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID		**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.02	13C2-PFDA	515.0 / 470.0	57803.16	250.00
d3-MeFOSAA	573.0 / 419.0	3.58	13C4-PFOS	503.0 / 99.0	16042.16	239.25
d5-EtFOSAA	589.0 / 419.0	3.73	13C4-PFOS	503.0 / 99.0	16042.16	239.25
13C5-PFHxA	318.0 / 273.0	1.84	13C2-PFOA	415.0 / 370.0	41307.84	250.00
13C4-PFHpA	367.0 / 322.0	2.25	13C2-PFOA	415.0 / 370.0	41307.84	250.00
13C8-PFOA	421.0 / 376.0	2.67	13C2-PFOA	415.0 / 370.0	41307.84	250.00
13C9-PFNA	472.0 / 427.0	3.06	13C2-PFOA	415.0 / 370.0	41307.84	250.00
13C6-PFDA	519.0 / 474.0	3.42	13C2-PFDA	515.0 / 470.0	57803.16	250.00
13C7-PFUnA	570.0 / 525.0	3.73	13C2-PFDA	515.0 / 470.0	57803.16	250.00
13C2-PFTeDA	715.0 / 670.0	4.48	13C2-PFDA	515.0 / 470.0	57803.16	250.00
13C3-PFBS	302.0 / 99.0	1.51	13C4-PFOS	503.0 / 99.0	16042.16	239.25
13C3-PFHxS	402.0 / 99.0	2.28	13C4-PFOS	503.0 / 99.0	16042.16	239.25
13C8-PFOS	507.0 / 99.0	3.06	13C4-PFOS	503.0 / 99.0	16042.16	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.07	13C2-PFDA	515.0 / 470.0	112231.68	250.00
d3-MeFOSAA	573.0 / 419.0	3.62	13C4-PFOS	503.0 / 99.0	35418.03	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	35418.03	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	98912.42	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	98912.42	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	98912.42	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	98912.42	250.00
13C6-PFDA	519.0 / 474.0	3.46	13C2-PFDA	515.0 / 470.0	112231.68	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	112231.68	250.00
13C2-PFTeDA	715.0 / 670.0	4.54	13C2-PFDA	515.0 / 470.0	112231.68	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	35418.03	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	35418.03	239.25
13C8-PFOS	507.0 / 99.0	3.10	13C4-PFOS	503.0 / 99.0	35418.03	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.06	13C2-PFDA	515.0 / 470.0	99145.40	250.00
d3-MeFOSAA	573.0 / 419.0	3.62	13C4-PFOS	503.0 / 99.0	28665.62	239.25
d5-EtFOSAA	589.0 / 419.0	3.78	13C4-PFOS	503.0 / 99.0	28665.62	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	83273.49	250.00
13C4-PFHpA	367.0 / 322.0	2.29	13C2-PFOA	415.0 / 370.0	83273.49	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	83273.49	250.00
13C9-PFNA	472.0 / 427.0	3.10	13C2-PFOA	415.0 / 370.0	83273.49	250.00
13C6-PFDA	519.0 / 474.0	3.46	13C2-PFDA	515.0 / 470.0	99145.40	250.00
13C7-PFUnA	570.0 / 525.0	3.78	13C2-PFDA	515.0 / 470.0	99145.40	250.00
13C2-PFTeDA	715.0 / 670.0	4.53	13C2-PFDA	515.0 / 470.0	99145.40	250.00
13C3-PFBS	302.0 / 99.0	1.54	13C4-PFOS	503.0 / 99.0	28665.62	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	28665.62	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	28665.62	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.06	13C2-PFDA	515.0 / 470.0	116159.80	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	35168.89	239.25
d5-EtFOSAA	589.0 / 419.0	3.77	13C4-PFOS	503.0 / 99.0	35168.89	239.25
13C5-PFHxA	318.0 / 273.0	1.87	13C2-PFOA	415.0 / 370.0	94531.08	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	94531.08	250.00
13C8-PFOA	421.0 / 376.0	2.70	13C2-PFOA	415.0 / 370.0	94531.08	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	94531.08	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	116159.80	250.00
13C7-PFUnA	570.0 / 525.0	3.77	13C2-PFDA	515.0 / 470.0	116159.80	250.00
13C2-PFTeDA	715.0 / 670.0	4.53	13C2-PFDA	515.0 / 470.0	116159.80	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	35168.89	239.25
13C3-PFHxS	402.0 / 99.0	2.31	13C4-PFOS	503.0 / 99.0	35168.89	239.25
13C8-PFOS	507.0 / 99.0	3.09	13C4-PFOS	503.0 / 99.0	35168.89	239.25

BATTELLE

Summary Internal Standard Report

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFDoA	615.0 / 570.0	4.05	13C2-PFDA	515.0 / 470.0	90498.88	250.00
d3-MeFOSAA	573.0 / 419.0	3.61	13C4-PFOS	503.0 / 99.0	27170.88	239.25
d5-EtFOSAA	589.0 / 419.0	3.77	13C4-PFOS	503.0 / 99.0	27170.88	239.25
13C5-PFHxA	318.0 / 273.0	1.86	13C2-PFOA	415.0 / 370.0	79401.49	250.00
13C4-PFHpA	367.0 / 322.0	2.28	13C2-PFOA	415.0 / 370.0	79401.49	250.00
13C8-PFOA	421.0 / 376.0	2.69	13C2-PFOA	415.0 / 370.0	79401.49	250.00
13C9-PFNA	472.0 / 427.0	3.09	13C2-PFOA	415.0 / 370.0	79401.49	250.00
13C6-PFDA	519.0 / 474.0	3.45	13C2-PFDA	515.0 / 470.0	90498.88	250.00
13C7-PFUnA	570.0 / 525.0	3.77	13C2-PFDA	515.0 / 470.0	90498.88	250.00
13C2-PFTeDA	715.0 / 670.0	4.52	13C2-PFDA	515.0 / 470.0	90498.88	250.00
13C3-PFBS	302.0 / 99.0	1.53	13C4-PFOS	503.0 / 99.0	27170.88	239.25
13C3-PFHxS	402.0 / 99.0	2.30	13C4-PFOS	503.0 / 99.0	27170.88	239.25
13C8-PFOS	507.0 / 99.0	3.08	13C4-PFOS	503.0 / 99.0	27170.88	239.25

Chromatograms

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:37 PM


Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1



PFOA_2



PFNA_1




PFNA_2

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:37 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:02:37 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:42 PM

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:



BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:42 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:02:42 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:47 PM


Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1



PFOA_2



PFNA_1




PFNA_2

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:47 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:02:47 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:53 PM

Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:



BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:53 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:02:53 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:58 PM


Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1



PFOA_2



PFNA_1




PFNA_2

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:58 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:02:58 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:03 PM

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:



BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:03 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:03 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:08 PM

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:



BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:08 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:08 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:13 PM

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:



BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:13 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:13 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:19 PM


Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1



PFOA_2



PFNA_1




PFNA_2

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:19 PM




Internal Standards:


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:19 PM




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:14:22 AM

Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:14:22 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:14:22 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:33 PM

Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:33 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:33 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:43 PM

Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:43 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:43 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:52 PM

Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:52 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:52 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:57 PM

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:57 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:57 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:04:02 PM

Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:02 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:02 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:04:07 PM

Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:07 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:07 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:04:12 PM

Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:12 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:12 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:04:17 PM

Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:17 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:17 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:04:23 PM


Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:23 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:04:23 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:33:47 AM

Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:33:47 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:33:47 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:33:59 AM

Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:33:59 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:33:59 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:05 AM


Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:05 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:05 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:11 AM

Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:11 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:11 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:17 AM


Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:17 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:17 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:23 AM

Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:23 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:23 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:29 AM

Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:29 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:29 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:35 AM

Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:35 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:35 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:40 AM

Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:40 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:40 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:05:23 PM

Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:23 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:23 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:05:33 PM

Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:33 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:33 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:05:39 PM

Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:39 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:39 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:46 AM

Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:46 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:46 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:53 AM

Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:53 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:53 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:05:56 PM

Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:56 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:05:56 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM

Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:01 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:06 PM

Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:06 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:06 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:11 PM

Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:11 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:11 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:16 PM

Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:16 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:16 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:21 PM

Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:21 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:21 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:26 PM


Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:26 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:26 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:34:59 AM

Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:59 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:34:59 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:35:05 AM

Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:35:05 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 01/11/2018 6:35:05 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:45 PM

Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID	VC-S14GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:45 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:45 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:06:50 PM

Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:50 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:06:50 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:28 PM

Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:28 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:28 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:33 PM

Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:33 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:33 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:38 PM


Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:38 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:38 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:44 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:44 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:44 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:50 PM

Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:50 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:50 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:12:55 PM


Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_BASE
Sample Comment			

Chromatograms


Target Analytes:




PFBS_1




PFBS_2




PFHxA_1




PFHxA_2




PFHpA_1




PFHpA_2




PFHxS_1




PFHxS_2




PFOA_1




PFOA_2




PFNA_1




PFNA_2




PFOS_1



PFOS_2



PFDA_1




PFDA_2


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:55 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:12:55 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:05:23 AM


Sample Name	KB80 IB	**Injection Vial**	2
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-22T19:43:39	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 01/11/2018 6:05:23 AM


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:19 PM

Sample Name	KB73	**Injection Vial**	2
Sample ID	L1	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:46:52	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:19 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:23 PM

Sample Name	KB74	**Injection Vial**	3
Sample ID	L2	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T19:57:45	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:23 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:27 PM

Sample Name	KB75	**Injection Vial**	4
Sample ID	L3	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:08:39	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:27 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:31 PM


Sample Name	KB76	**Injection Vial**	5
Sample ID	L4	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:19:32	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:31 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:35 PM

Sample Name	KB77	**Injection Vial**	6
Sample ID	L5	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:30:23	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:35 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:40 PM

Sample Name	KB78	**Injection Vial**	7
Sample ID	L6	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:41:14	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:40 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:45 PM

Sample Name	KB79	**Injection Vial**	8
Sample ID	L7	**Injection Volume**	10.00
Sample Type	Standard	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T20:52:06	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:45 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:49 PM

Sample Name	KB80 IB	**Injection Vial**	9
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:02:57	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:49 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:58:54 PM

Sample Name	KB81 ICC	**Injection Vial**	10
Sample ID	ICC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-17T21:13:49	**Data File**	Data18-0590_18-01588_18-0589.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 1:58:54 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS




13C3-PFHxS




13C8-PFOS


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:08 PM


Sample Name	KB76 CCV	**Injection Vial**	44
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T03:41:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:08 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:16 PM


Sample Name	CR939PB-FS(0)	**Injection Vial**	46
Sample ID	Procedural Blank	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:03:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:16 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:24 PM


Sample Name	J8613-FS(0)	**Injection Vial**	48
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:25:07	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:24 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:28 PM


Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:28 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:32 PM


Sample Name	J8613-FS-D(5)	**Injection Vial**	50
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:46:53	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:32 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:37 PM


Sample Name	J8613-FS-D(7)	**Injection Vial**	51
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:57:46	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:37 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:41 PM


Sample Name	J8614-FS(0)	**Injection Vial**	52
Sample ID	VC-AQ-FB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:08:37	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:41 PM


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:45 PM


Sample Name	J8615-FS(0)	**Injection Vial**	53
Sample ID	VC-AQ-EB09-10012018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:19:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:45 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:50 PM


Sample Name	KB77 CCV	**Injection Vial**	54
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:30:20	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:50 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:59 PM


Sample Name	J8616-FS(0)	**Injection Vial**	2
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T05:52:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:59 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:08 PM


Sample Name	J8617-FS(0)	**Injection Vial**	4
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:13:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:08 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:14 PM


Sample Name	J8617-FS-D(3)	**Injection Vial**	5
Sample ID	VC-KCHA16-GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:24:47	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:14 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:18 PM


Sample Name	J8618-FS(0)	**Injection Vial**	6
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:35:40	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:18 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:23 PM


Sample Name	J8618-FS-D(3)	**Injection Vial**	7
Sample ID	VC-KCHA16-GW01P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:46:33	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:23 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:28 PM


Sample Name	J8619-FS(0)	**Injection Vial**	8
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:57:26	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:28 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:32 PM


Sample Name	J8619-FS-D(3)	**Injection Vial**	9
Sample ID	VC-KCHA16-GW02-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:08:19	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:32 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:36 PM


Sample Name	J8620-FS(0)	**Injection Vial**	10
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:36 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:40 PM


Sample Name	J8620-FS-D(3)	**Injection Vial**	11
Sample ID	VC-KCHA16-GW03-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:30:04	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:40 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:44 PM


Sample Name	KB76 CCV	**Injection Vial**	12
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T07:40:57	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:44 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:53 PM


Sample Name	J8621-FS(0)	**Injection Vial**	14
Sample ID	VC-AQ-FB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:02:44	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:53 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:58 PM


Sample Name	J8622-FS(0)	**Injection Vial**	15
Sample ID	VC-AQ-EB10-10022018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:13:36	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:58 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:03 PM


Sample Name	J8623-FS(0)	**Injection Vial**	16
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:24:29	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:03 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:08 PM


Sample Name	J8623-FS-D(3)	**Injection Vial**	17
Sample ID	VC-S09GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:35:21	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:08 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:12 PM


Sample Name	J8624-FS(0)	**Injection Vial**	18
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T08:46:14	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:12 PM


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:17 PM


Sample Name	J8624-FS-D(3)	**Injection Vial**	19
Sample ID	VC-S09GW04-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:19:12	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:17 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:21 PM


Sample Name	J8625-FS(0)	**Injection Vial**	20
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:30:05	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:21 PM


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:25 PM


Sample Name	J8625-FS-D(3)	**Injection Vial**	21
Sample ID	VC-S09GW04P-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:40:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:25 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:31 PM


Sample Name	J8626-FS(0)	**Injection Vial**	22
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T09:51:52	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:31 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:36 PM


Sample Name	J8626-FS-D(3)	**Injection Vial**	23
Sample ID	VC-S09GW06-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:02:45	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:36 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:40 PM


Sample Name	KB77 CCV	**Injection Vial**	24
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:13:38	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:40 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:49 PM


Sample Name	J8627-FS(0)	**Injection Vial**	26
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:35:23	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:49 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:53 PM


Sample Name	J8627-FS-D(3)	**Injection Vial**	27
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:46:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:53 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:57 PM


Sample Name	J8628-FS(0)	**Injection Vial**	28
Sample ID		**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T10:57:08	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:01:57 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:01 PM


Sample Name	KB76 CCV	**Injection Vial**	29
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T11:08:03	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:02:01 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:37 PM


Sample Name	KB75 ISC	**Injection Vial**	25
Sample ID	ISC	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T11:51:20	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:37 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:41 PM


Sample Name	KB80 IB	**Injection Vial**	26
Sample ID	IB	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:02:11	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:41 PM


Internal Standards:



13C2-PFOA



13C4-PFOS



13C2-PFDA

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:45 PM


Sample Name	CR940LCS-FS(0)	**Injection Vial**	27
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:13:02	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:45 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:50 PM


Sample Name	J8616-FS-D(3)	**Injection Vial**	28
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:23:52	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:50 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:54 PM


Sample Name	J8627-FS-D(5)	**Injection Vial**	29
Sample ID	VC-S14GW01-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:34:44	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:54 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:58 PM


Sample Name	KB76	**Injection Vial**	30
Sample ID	CCV	**Injection Volume**	10.00
Sample Type	Quality Control	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-24T12:56:27	**Data File**	AC_10242018_369.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602A_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:09:58 PM

Internal Standards:



Unused Data

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:43:51 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	47
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:14:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	3009948.76	8045.772781	1352.1	false
PFBS_2	298.9 / 99.0	1.54	866807.21	7891.381571	1251.2	false
PFHxA_1	313.0 / 269.0	1.86	2090654.80	8811.531783	173.8	false
PFHxA_2	313.0 / 119.0	1.86	151704.86	8567.209778	121.6	false
PFHpA_1	363.0 / 319.0	2.27	1851204.27	7707.457081	481.5	false
PFHpA_2	363.0 / 169.0	2.27	38845.23	7575.272812	422.5	false
PFHxS_1	399.0 / 80.0	2.30	2801632.83	8655.446300	1157.8	false
PFHxS_2	399.0 / 99.0	2.30	791182.73	8762.242411	1423.7	false
PFOA_1	413.0 / 369.0	2.68	2843553.06	8514.504834	862.4	false
PFOA_2	413.0 / 169.0	2.68	176986.82	8160.713963	626.0	false
PFNA_1	463.0 / 419.0	3.08	2682113.29	9356.558585	938.6	true
PFNA_2	463.0 / 219.0	3.08	798504.86	9028.633620	923.2	false
PFOS_1	499.0 / 80.0	3.08	4209401.24	7417.741055	537.3	false
PFOS_2	499.0 / 99.0	3.08	752041.78	7632.330311	1113.2	false
PFDA_1	513.0 / 469.0	3.44	3195415.70	8324.332605	833.3	false
PFDA_2	513.0 / 219.0	3.43	134869.58	8527.318964	740.9	false
PFUnA_1	563.0 / 519.0	3.76	3233522.76	8603.247531	772.7	false
PFUnA_2	563.0 / 269.0	3.76	149106.06	7977.872851	515.1	false
PFDoA_1	613.0 / 569.0	4.03	3098454.32	8270.008062	832.7	false
PFDoA_2	613.0 / 319.0	4.03	500437.03	8575.719960	671.6	false
PFTrDA_1	663.0 / 619.0	4.28	2638459.01	7997.223346	1408.5	false
PFTrDA_2	663.0 / 169.0	4.27	181234.66	8361.809783	820.9	false
PFTeDA_1	713.0 / 669.0	4.49	3215529.97	8605.080504	2468.4	false
PFTeDA_2	713.0 / 169.0	4.49	153660.98	8493.494367	1721.5	false
NMeFOSAA_1	570.0 / 419.0	3.59	500817.50	9030.646144	1189.2	false
NMeFOSAA_2	570.0 / 512.0	3.59	259839.66	8642.364396	1300.0	false
NEtFOSAA_1	584.0 / 419.0	3.75	451332.49	7396.864492	953.2	false
NEtFOSAA_2	584.0 / 483.0	3.75	29154.91	8172.385561	1100.1	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:44:22 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	3
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:02:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.54	3568672.52	6196.044445	525.9	false
PFBS_2	298.9 / 99.0	1.54	1057036.12	6251.086446	683.7	false
PFHxA_1	313.0 / 269.0	1.86	1360996.92	4242.319495	77.9	false
PFHxA_2	313.0 / 119.0	1.86	87556.04	3642.805255	93.6	false
PFHpA_1	363.0 / 319.0	2.27	222797.27	628.862209	104.8	false
PFHpA_2	363.0 / 169.0	2.25	6518.83	806.994834	102.6	false
PFHxS_1	399.0 / 80.0	2.30	5472268.43	13979.459929	719.4	false
PFHxS_2	399.0 / 99.0	2.30	1544802.92	14149.773825	1121.1	false
PFOA_1	413.0 / 369.0	2.68	382132.32	917.565843	201.7	false
PFOA_2	413.0 / 169.0	2.63	34463.09	1280.899321	171.3	true
PFNA_1	463.0 / 419.0	3.08	23946.99	36.824578	54.2	false
PFNA_2	463.0 / 219.0	3.08	8623.78	52.264216	77.8	false
PFOS_1	499.0 / 80.0	2.95	1103533.72	1820.566681	255.5	false
PFOS_2	499.0 / 99.0	3.01	127889.71	1214.547978	229.8	true
PFDA_1	513.0 / 469.0	3.43	4491.96	< 0	24.2	false
PFDA_2	513.0 / 219.0	3.48	613.95	7.849537	16.1	false
PFUnA_1	563.0 / 519.0	3.76	4373.78	< 0	30.3	false
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	4.03	1275.48	< 0	25.6	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	4.29	886.01	< 0	34.8	false
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:11 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	47
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:14:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	103167.70	246.434154	1257.9	false
d3-MeFOSAA	573.0 / 419.0	3.58	14333.23	207.938320	129.4	false
d5-EtFOSAA	589.0 / 419.0	3.74	16606.27	212.244366	215.1	false
13C5-PFHxA	318.0 / 273.0	1.85	59988.32	229.841566	882.0	false
13C4-PFHpA	367.0 / 322.0	2.27	68279.88	229.947926	773.2	false
13C8-PFOA	421.0 / 376.0	2.68	83111.12	230.089080	1214.3	false
13C9-PFNA	472.0 / 427.0	3.07	86207.80	209.015225	1117.1	false
13C6-PFDA	519.0 / 474.0	3.42	92043.83	218.229387	828.5	false
13C7-PFUnA	570.0 / 525.0	3.74	89436.90	229.555474	898.1	false
13C2-PFTeDA	715.0 / 670.0	4.48	86220.32	255.088753	1350.4	false
13C3-PFBS	302.0 / 99.0	1.53	29862.92	237.809886	549.0	false
13C3-PFHxS	402.0 / 99.0	2.29	22850.15	200.809792	375.4	false
13C8-PFOS	507.0 / 99.0	3.07	28141.42	218.921417	250.3	false

BATTELLE

Summary Quant Report

Created with Analyst Reporter
Printed: 25/10/2018 1:41:44 PM

Sample Name	J8616-FS-D(3)	**Injection Vial**	3
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:02:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFDoA	615.0 / 570.0	4.02	114717.80	235.918771	1119.6	false
d3-MeFOSAA	573.0 / 419.0	3.58	15960.35	248.891766	223.4	false
d5-EtFOSAA	589.0 / 419.0	3.74	17510.77	240.573151	206.1	false
13C5-PFHxA	318.0 / 273.0	1.85	80779.98	288.921361	362.1	false
13C4-PFHpA	367.0 / 322.0	2.26	97967.91	307.988602	596.9	false
13C8-PFOA	421.0 / 376.0	2.67	102138.66	263.961799	1108.0	false
13C9-PFNA	472.0 / 427.0	3.07	103885.54	235.125939	891.6	false
13C6-PFDA	519.0 / 474.0	3.42	113128.74	230.922397	1200.8	false
13C7-PFUnA	570.0 / 525.0	3.74	110379.48	243.912370	871.8	false
13C2-PFTeDA	715.0 / 670.0	4.48	100537.80	256.085854	1621.3	false
13C3-PFBS	302.0 / 99.0	1.52	45966.72	393.476568	485.3	false
13C3-PFHxS	402.0 / 99.0	2.29	27374.15	258.591504	401.8	false
13C8-PFOS	507.0 / 99.0	3.06	29706.85	248.414262	243.7	false

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 02/11/2018 11:20:50 AM

Sample Name	J8613-FS-D(3)	**Injection Vial**	49
Sample ID	VC-PM553-GW0601A-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:36:00	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 02/11/2018 11:20:50 AM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 02/11/2018 11:20:50 AM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:03:47 PM

Sample Name	CR940LCS-FS(0)	**Injection Vial**	47
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:14:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_BASE
Sample Comment			

Chromatograms


Target Analytes:




BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:47 PM




PFUnA_1




PFUnA_2




PFDoA_1




PFDoA_2




PFTrDA_1




PFTrDA_2




PFTeDA_1




PFTeDA_2




NMeFOSAA_1



NMeFOSAA_2




NEtFOSAA_1




NEtFOSAA_2


Internal Standards:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA



13C8-PFOA



13C9-PFNA




13C6-PFDA


BATTELLE

Chromatogram Report


Created with Analyst Reporter
Printed: 25/10/2018 2:03:47 PM




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS



13C8-PFOS

BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:20 PM


Sample Name	CR940LCS-FS(0)	**Injection Vial**	47
Sample ID	Laboratory Control Sample	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T04:14:16	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 1:59:20 PM

Internal Standards:



BATTELLE


Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:04 PM


Sample Name	J8616-FS-D(3)	**Injection Vial**	3
Sample ID	VC-PM13-GW21-1018	**Injection Volume**	10.00
Sample Type	Unknown	**Instrument Name**	QTRAP 5500
Acquisition Date	2018-10-23T06:02:58	**Data File**	10222018_5500.wiff
Acquisition Method	5-0369.dam	**Result Table**	18-0602_SIS
Sample Comment			

Chromatograms


Target Analytes:




13C2-PFDoA




d3-MeFOSAA




d5-EtFOSAA




13C5-PFHxA




13C4-PFHpA




13C8-PFOA




13C9-PFNA




13C6-PFDA




13C7-PFUnA




13C2-PFTeDA



13C3-PFBS



13C3-PFHxS




13C8-PFOS

BATTELLE

Chromatogram Report

Created with Analyst Reporter
Printed: 25/10/2018 2:00:04 PM

Internal Standards:



Contract_I	DO_CTO_N	Phase	Installatior	Sample_Ni	CH2M_Coc	Analysis_G	Analytical_	PRC_Code	Lab_Code	Lab_Name	Leachate_I	SAMPLE_B	Extraction_	Result_Typ	Lab_QC_ty	SAMPLE_N	QC_Level	DateTime_	Date_Rece	Leachate_I	Leachate_	Extraction_	Extraction_	Analysis_D	Analysis_T	Lab_Sampl	Dilution	Run_Numt	PERCENT_I	PERCENT_I	Chem_Nan	Analyte_ID	Analyte_Vi	Original_A	Result_Uni	Lab_Qualif	Validator_	GC_Colum	Analysis_R	Result_Nat	QC_Contrc	QC_Accura	QC_Accura	Control_Lir	QC_Narrat	MDL		Detection_Q	SM_Vers	DL		LOD		LOQ		SDG	Analysis_B	Validator_	Val_Date
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoroh	307-24-4	0.5	NG_L	U		TRG					20171116		0.19	0.19	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoro	375-85-9	0.5	NG_L	U		TRG					20171116		0.16	0.16	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoroo	335-67-1	0.96	NG_L	J		TRG					20171116		0.18	0.18	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoro	375-95-1	1	NG_L	U		TRG					20171116		0.26	0.26	5.1		70	1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorod	335-76-2	0.5	NG_L	U		TRG					20171116		0.16	0.16	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorou	2058-94-8	1	NG_L	U		TRG					20171116		0.29	0.29	5.1		70	1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorod	307-55-1	0.5	NG_L	U		TRG					20171116		0.18	0.18	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorot	r72629-94-8	0.5	NG_L	U		TRG					20171116		0.15	0.15	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorot	376-06-7	1	NG_L	U		TRG					20171116		0.25	0.25	5.1		70	1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			N-Methyl	F2355-31-9	2	NG_L	U		TRG					20171116		0.56	0.56	5.1		70	2		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			N-Ethyl Pe	r2991-50-6	1	NG_L	U		TRG					20171116		0.49	0.49	5.1		70	1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluorob	375-73-5	0.5	NG_L	U		TRG					20171116		0.13	0.13	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoro	355-46-4	0.4	NG_L	U		TRG					20171116		0.11	0.11	5.1		70	0.4		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			Perfluoroo	1763-23-1	0.5	NG_L	U		TRG					20171116		0.19	0.19	5.1		70	0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C5-PFH	x BDO-2217	92	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C4-PFH	p BDO-2218	96	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C8-PFO	A BDO-2219	93	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C9-PFN	A BDO-2221	88	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C6-PFD	A BDO-2222	90	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C7-PFU	n BDO-2223	86	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C2-PFD	o BDO-2112	88	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C2-PFT	e BDO-2224	94	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			d3-MeFOS	BDO-1838	91	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			d5-EtFOSA	BDO-1839	75	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C3-PFB	S BDO-2226	94	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C3-PFH	x BDO-2227	88	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR939PB-F	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BLK	W	4	20181010	20181010			20181010	14:03:00	20181023	04:03:23	CR939PB-F	1	1			13C8-PFO	S BDO-2228	82	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoroh	307-24-4	117	PCT_REC			TRG	LSA	137	51	20171116		0.19	0.19	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoro	375-85-9	108	PCT_REC			TRG	LSA	136	48	20171116		0.16	0.16	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoroo	335-67-1	114	PCT_REC			TRG	LSA	141	49	20171116		0.18	0.18	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoro	375-95-1	111	PCT_REC			TRG	LSA	122	58	20171116		0.26	0.26	5.1			1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorod	335-76-2	107	PCT_REC			TRG	LSA	135	59	20171116		0.16	0.16	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorou	2058-94-8	117	PCT_REC			TRG	LSA	134	64	20171116		0.29	0.29	5.1			1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorod	307-55-1	109	PCT_REC			TRG	LSA	131	75	20171116		0.18	0.18	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorot	r72629-94-8	112	PCT_REC			TRG	LSA	148	42	20171116		0.15	0.15	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorot	376-06-7	115	PCT_REC			TRG	LSA	158	42	20171116		0.25	0.25	5.1			1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			N-Methyl	F2355-31-9	129	PCT_REC			TRG	LSA	146	50	20171116		0.56	0.56	5.1			2		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			N-Ethyl Pe	r2991-50-6	109	PCT_REC			TRG	LSA	131	51	20171116		0.49	0.49	5.1			1		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluorob	375-73-5	97	PCT_REC			TRG	LSA	134	56	20171116		0.13	0.13	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoro	355-46-4	105	PCT_REC			TRG	LSA	128	52	20171116		0.11	0.11	5.1			0.4		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			Perfluoroo	1763-23-1	100	PCT_REC			TRG	LSA	144	40	20171116		0.19	0.19	5.1			0.5		5	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C5-PFH	x BDO-2217	97	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C4-PFH	p BDO-2218	92	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C8-PFO	A BDO-2219	86	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C9-PFN	A BDO-2221	92	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C6-PFD	A BDO-2222	90	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C7-PFU	n BDO-2223	88	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C2-PFD	o BDO-2112	106	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C2-PFT	e BDO-2224	107	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			d3-MeFOS	BDO-1838	94	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			d5-EtFOSA	BDO-1839	87	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C3-PFB	S BDO-2226	117	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C3-PFH	x BDO-2227	96	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	CR940LCS-	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	BS	W	4	20181010	20181010			20181010	14:03:00	20181024	12:13:02	CR940LCS-	1	1			13C8-PFO	S BDO-2228	98	PCT_REC			SURR	SLSP	150	50	20171116				5.1						18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:57:46	JB613-FS	166.67	1			Perfluoroh	307-24-4	760.64	NG_L	D		TRG					20171116		29.32	29.32	5.1		70	77.16	771.61	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:57:46	JB613-FS	166.67	1			Perfluoro	375-85-9	3916.2	NG_L	D		TRG					20171116		24.69	24.69	5.1		70	77.16	771.61	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:57:46	JB613-FS	166.67	1			Perfluoroo	335-67-1	6727.83	NG_L	D		TRG					20171116		27.78	27.78	5.1		70	77.16	771.61	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluoro	375-95-1	5.37	NG_L			TRG					20171116		0.24	0.24	5.1		70	0.93	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluorod	335-76-2	0.46	NG_L	U		TRG					20171116		0.15	0.15	5.1		70	0.46	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluorou	2058-94-8	0.93	NG_L	U		TRG					20171116		0.27	0.27	5.1		70	0.93	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluorod	307-55-1	0.46	NG_L	U		TRG					20171116		0.17	0.17	5.1		70	0.46	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluorot	r72629-94-8	0.46	NG_L	U		TRG					20171116		0.14	0.14	5.1		70	0.46	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS	1	1			Perfluorot	376-06-7	0.93	NG_L	U		TRG					20171116		0.17	0.17	5.1		70	0.93	4.63	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-PM553	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	04:25:07	JB613-FS

Contract_I	DO_CTO_N	Phase	Installatior	Sample_Ni	CH2M_Cor	Analysis_G	Analytical_	PRC_Code	Lab_Code	Lab_Name	Leachate_I	SAMPLE_B	Extraction_	Result_Typ	Lab_QC_ty	SAMPLE_N	QC_Level	DateTime_	Date_Rece	Leachate_I	Leachate_	Extraction_	Extraction_	Analysis_D	Analysis_T	Lab_Sampl	Dilution	Run_Numt	PERCENT_I	PERCENT_I	Chem_Nan	Analyte_ID	Analyte_Vi	Original_A	Result_Uni	Lab_Qualif	Validator_	QC_Colum	Analysis_R	Result_Nat	QC_Contrc	QC_Accura	QC_Accura	Control_Lir	QC_Narrat	MDL		Detection_Q	SM_Vers	DL		LOD		LOQ		SDG	Analysis_B	Validator_	Val_Date
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			Perfluorot	376-06-7		0.91	NG_L	U		TRG					20171116			0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			N-Methyl	FOSA-N-ME		1.82	NG_L	U		TRG					20171116			0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			N-Ethyl Pe	r2991-50-6		0.91	NG_L	U		TRG					20171116			0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			Perfluorob	375-73-5		0.45	NG_L	U		TRG					20171116			0.12	0.12	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			Perfluoroh	355-46-4		0.36	NG_L	U		TRG					20171116			0.1	0.1	5.1	70	0.36	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			Perfluoroo	1763-23-1		0.45	NG_L	U		TRG					20171116			0.17	0.17	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C5-PFH	x BDO-2217		95	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C4-PFH	p BDO-2218		83	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C8-PFO	A BDO-2219		91	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C9-PFN	A BDO-2221		82	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C6-PFD	A BDO-2222		84	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C7-PFU	n BDO-2223		87	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C2-PFD	o BDO-2112		85	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C2-PFT	e BDO-2224		85	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			d3-MeFOS	BDO-1838		71	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			d5-EtFOS	A BDO-1839		58	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C3-PFB	S BDO-2226		73	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C3-PFH	x BDO-2227		78	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-AQ-EBC	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	14:03:00	20181023	05:19:29	J8615-FS	1	1			13C8-PFO	S BDO-2228		77	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			Perfluorob	307-24-4		245.61	NG_L	D		TRG					20171116			2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluoroh	375-85-9		48.74	NG_L			TRG					20171116			0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluoroo	335-67-1		63.52	NG_L			TRG					20171116			0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluoron	375-95-1		1	NG_L	J		TRG					20171116			0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluorod	335-76-2		0.45	NG_L	U		TRG					20171116			0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U		TRG					20171116			0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U		TRG					20171116			0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U		TRG					20171116			0.14	0.14	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			Perfluorot	376-06-7		0.91	NG_L	U		TRG					20171116			0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			N-Methyl	FOSA-N-ME		1.82	NG_L	U		TRG					20171116			0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			N-Ethyl Pe	r2991-50-6		0.91	NG_L	U		TRG					20171116			0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			Perfluorob	375-73-5		397.55	NG_L	D		TRG					20171116			1.97	1.97	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			Perfluoroh	355-46-4		755.99	NG_L	D		TRG					20171116			1.67	1.67	5.1	70	6.06	75.76	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			Perfluoroo	1763-23-1		94.92	NG_L	D		TRG					20171116			2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			13C5-PFH	x BDO-2217		117	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C4-PFH	p BDO-2218		99	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C8-PFO	A BDO-2219		79	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C9-PFN	A BDO-2221		74	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C6-PFD	A BDO-2222		90	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C7-PFU	n BDO-2223		96	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C2-PFD	o BDO-2112		93	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			13C2-PFT	e BDO-2224		99	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			d3-MeFOS	BDO-1838		99	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181001	20181005			20181010	10:44:00	20181023	05:52:05	J8616-FS	1	1			d5-EtFOS	A BDO-1839		102	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			13C3-PFB	S BDO-2226		128	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			13C3-PFH	x BDO-2227		99	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-PM13-(	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181001	20181005			20181010	10:44:00	20181024	12:23:52	J8616-FS	16.67	1			13C8-PFO	S BDO-2228		99	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:24:47	J8617-FS	10	1			Perfluorob	307-24-4		338.2	NG_L	D		TRG					20171116			1.73	1.73	5.1	70	4.55	45.45	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluoroh	375-85-9		13.04	NG_L			TRG					20171116			0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluoroo	335-67-1		6.64	NG_L	B		TRG					20171116			0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluoron	375-95-1		0.91	NG_L	U		TRG					20171116			0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluorod	335-76-2		0.45	NG_L	U		TRG					20171116			0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U		TRG					20171116			0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U		TRG					20171116			0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U		TRG					20171116			0.14	0.14	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluorot	376-06-7		0.91	NG_L	U		TRG					20171116			0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			N-Methyl	FOSA-N-ME		1.82	NG_L	U		TRG					20171116			0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			N-Ethyl Pe	r2991-50-6		0.91	NG_L	U		TRG					20171116			0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:24:47	J8617-FS	10	1			Perfluorob	375-73-5		153.2	NG_L	D		TRG					20171116			1.18	1.18	5.1	70	4.55	45.45	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			Perfluoroo	1763-23-1		8.94	NG_L			TRG					20171116			0.17	0.17	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:24:47	J8617-FS	10	1			Perfluoroh	355-46-4		177.76	NG_L	D		TRG					20171116			1	1	5.1	70	3.64	45.45	18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			13C5-PFH	x BDO-2217		114	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			13C4-PFH	p BDO-2218		121	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:13:52	J8617-FS	1	1			13C8-PFO	A BDO-2219		89	PCT_REC			SURR	SLSP	150	50	20171116					5.1				18-0602	DP-18-0303		
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000

Contract_I	DO_CTO_N	Phase	Installatior	Sample_Ni	CH2M_Coc	Analysis_G	Analytical_	PRC_Code	Lab_Code	Lab_Name	Leachate_I	SAMPLE_B	Extraction_	Result_Typ	Lab_QC_ty	SAMPLE_N	QC_Level	DateTime_	Date_Rece	Leachate_I	Leachate_	Extraction_	Extraction_	Analysis_D	Analysis_T	Lab_Sampl	Dilution	Run_Numt	PERCENT_I	PERCENT_I	Chem_Nan	Analyte_ID	Analyte_Vi	Original_A	Result_Uni	Lab_Qualif	Validator_	GC_Colum	Analysis_R	Result_Nat	QC_Contrc	QC_Accura	QC_Accura	Control_Lir	QC_Narrat	MDL	Detection_Q	SM_Vers	DL	LOD	LOQ	SDG	Analysis_B	Validator_	Val_Date
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C8-PFO	A BDO-2219		98	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C9-PFN	A BDO-2221		82	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C6-PFD	A BDO-2222		77	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C7-PFU	n BDO-2223		82	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C2-PFD	o BDO-2112		76	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C2-PFTe	BDO-2224		80	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			d3-MeFOS	BDO-1838		72	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			d5-EtFOS	A BDO-1839		71	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C3-PFB	S BDO-2226		101	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:08:19	J8619-FS	10	1			13C3-PFH	x BDO-2227		87	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	06:57:26	J8619-FS	1	1			13C8-PFO	S BDO-2228		79	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			Perfluoroh	307-24-4		229.43	NG_L	D		TRG				20171116		1.73	1.73	5.1	70	4.55	45.45	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoroh	375-85-9		32.09	NG_L			TRG				20171116		0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			Perfluoroo	335-67-1		66.04	NG_L	D		TRG				20171116		1.64	1.64	5.1	70	4.55	45.45	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoro	375-95-1		0.91	NG_L	U		TRG				20171116		0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoro	d 335-76-2		0.45	NG_L	U		TRG				20171116		0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U		TRG				20171116		0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoro	d 307-55-1		0.45	NG_L	U		TRG				20171116		0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U		TRG				20171116		0.14	0.14	5.1	70	0.45	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoro	n 376-06-7		0.91	NG_L	U		TRG				20171116		0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			N-Methyl	F 2355-31-9		1.82	NG_L	U		TRG				20171116		0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			N-Ethyl P	er 2991-50-6		0.91	NG_L	U		TRG				20171116		0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			Perfluorob	375-73-5		69.78	NG_L	D		TRG				20171116		1.18	1.18	5.1	70	4.55	45.45	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			Perfluoro	h 355-46-4		136.86	NG_L	D		TRG				20171116		1	1	5.1	70	3.64	45.45	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			Perfluoroo	1763-23-1		10.97	NG_L			TRG				20171116		0.17	0.17	5.1	70	0.45	4.55	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			13C5-PFH	x BDO-2217		103	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C4-PFH	p BDO-2218		104	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			13C8-PFO	A BDO-2219		110	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C9-PFN	A BDO-2221		75	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C6-PFD	A BDO-2222		78	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C7-PFU	n BDO-2223		82	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C2-PFD	o BDO-2112		84	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C2-PFTe	BDO-2224		94	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			d3-MeFOS	BDO-1838		72	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			d5-EtFOS	A BDO-1839		73	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			13C3-PFB	S BDO-2226		126	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:30:04	J8620-FS	10	1			13C3-PFH	x BDO-2227		89	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-KCHA1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	07:19:12	J8620-FS	1	1			13C8-PFO	S BDO-2228		92	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoroh	307-24-4		0.44	NG_L	U		TRG				20171116		0.17	0.17	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoroh	375-85-9		0.44	NG_L	U		TRG				20171116		0.14	0.14	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoroo	335-67-1		1.88	NG_L	J		TRG				20171116		0.16	0.16	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoro	375-95-1		0.88	NG_L	U		TRG				20171116		0.23	0.23	5.1	70	0.88	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoro	d 335-76-2		0.44	NG_L	U		TRG				20171116		0.14	0.14	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluorou	2058-94-8		0.88	NG_L	U		TRG				20171116		0.25	0.25	5.1	70	0.88	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoro	d 307-55-1		0.44	NG_L	U		TRG				20171116		0.16	0.16	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluorot	72629-94-8		0.44	NG_L	U		TRG				20171116		0.13	0.13	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoro	n 376-06-7		0.88	NG_L	U		TRG				20171116		0.22	0.22	5.1	70	0.88	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			N-Methyl	F 2355-31-9		1.75	NG_L	U		TRG				20171116		0.49	0.49	5.1	70	1.75	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			N-Ethyl P	er 2991-50-6		0.88	NG_L	U		TRG				20171116		0.43	0.43	5.1	70	0.88	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluorob	375-73-5		0.44	NG_L	U		TRG				20171116		0.11	0.11	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoro	h 355-46-4		0.35	NG_L	U		TRG				20171116		0.1	0.1	5.1	70	0.35	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			Perfluoroo	1763-23-1		0.44	NG_L	U		TRG				20171116		0.17	0.17	5.1	70	0.44	4.39	18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C5-PFH	x BDO-2217		90	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C4-PFH	p BDO-2218		88	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C8-PFO	A BDO-2219		93	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C9-PFN	A BDO-2221		84	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C6-PFD	A BDO-2222		84	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C7-PFU	n BDO-2223		85	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:02:44	J8621-FS	1	1			13C2-PFD	o BDO-2112		92	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:24:29	J8623-FS	1	1			13C2-PFTe	BDO-2224		94	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:24:29	J8623-FS	1	1			d3-MeFOS	BDO-1838		68	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:24:29	J8623-FS	1	1			d5-EtFOS	A BDO-1839		74	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	14:03:00	20181023	08:24:29	J8623-FS	1	1			13C3-PFB	S BDO-2226		80	PCT_REC			SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303			
N6247016	4164		POINT_MU	VC-AQ-FB1	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181

Contract_I	DO_CTO_N	Phase	Installatior	Sample_Nu	CH2M_Cor	Analysis_G	Analytical_	PRC_Code	Lab_Code	Lab_Name	Leachate_I	SAMPLE_B	Extraction_	Result_Typ	Lab_QC_ty	SAMPLE_N	QC_Level	DateTime_	Date_Rece	Leachate_I	Leachate_	Extraction_	Extraction_	Analysis_D	Analysis_T	Lab_Sampl	Dilution	Run_Numb	PERCENT_I	PERCENT_I	Chem_Nan	Analyte_ID	Analyte_Vi	Original_A	Result_Uni	Lab_Qualif	Validator_	QC_Colum	Analysis_R	Result_Nar	QC_Contro	QC_Accura	QC_Accura	Control_Lir	QC_Narrat	MDL	Detection_	QSM_Vers	DL	LOD	LOQ	SDG	Analysis_B	Validator_	Val_Date
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	08:24:29	J8623-FS	1	1			13C3-PFBS	BDO-2226		72	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	08:35:21	J8623-FS	10	1			13C3-PFHx	BDO-2227		99	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	08:24:29	J8623-FS	1	1			13C8-PFOS	BDO-2228		81	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			Perfluoroh	307-24-4		153.03	NG_L	D			TRG				20171116		2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			Perfluoroh	375-85-9		114.46	NG_L	D			TRG				20171116		2.42	2.42	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			Perfluoroo	335-67-1		381.85	NG_L	D			TRG				20171116		2.73	2.73	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluoro	375-95-1		9.77	NG_L				TRG				20171116		0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorod	335-76-2		1.08	NG_L	J			TRG				20171116		0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U			TRG				20171116		0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U			TRG				20171116		0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U			TRG				20171116		0.14	0.14	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorot	376-06-7		0.91	NG_L	U			TRG				20171116		0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			N-Methyl	FZ355-31-9		1.82	NG_L	U			TRG				20171116		0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			N-Ethyl	Per2991-50-6		0.91	NG_L	U			TRG				20171116		0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			Perfluorob	375-73-5		51.02	NG_L				TRG				20171116		0.12	0.12	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			Perfluoroh	355-46-4		717.18	NG_L	D			TRG				20171116		1.67	1.67	5.1	70	6.06	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			Perfluoroo	1763-23-1		345.23	NG_L	D			TRG				20171116		2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			13C5-PFHx	BDO-2217		131	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			13C4-PFHp	BDO-2218		129	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			13C8-PFOA	BDO-2219		104	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C9-PFNA	BDO-2221		62	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C6-PFDA	BDO-2222		91	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C7-PFUn	BDO-2223		97	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C2-PFDo	BDO-2112		98	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C2-PFTe	BDO-2224		100	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			d3-MeFOS	BDO-1838		109	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			d5-EtFOSA	BDO-1839		110	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	08:46:14	J8624-FS	1	1			13C3-PFBS	BDO-2226		92	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			13C3-PFHx	BDO-2227		92	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:19:12	J8624-FS	16.67	1			13C8-PFOS	BDO-2228		94	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			Perfluoroh	307-24-4		180.39	NG_L	D			TRG				20171116		2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			Perfluoroh	375-85-9		128.79	NG_L	D			TRG				20171116		2.42	2.42	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			Perfluoroo	335-67-1		405.69	NG_L	D			TRG				20171116		2.73	2.73	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluoro	375-95-1		9.85	NG_L				TRG				20171116		0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorod	335-76-2		0.99	NG_L	J			TRG				20171116		0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U			TRG				20171116		0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U			TRG				20171116		0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U			TRG				20171116		0.14	0.14	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorot	376-06-7		0.91	NG_L	U			TRG				20171116		0.23	0.23	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			N-Methyl	FZ355-31-9		1.82	NG_L	U			TRG				20171116		0.51	0.51	5.1	70	1.82	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			N-Ethyl	Per2991-50-6		0.91	NG_L	U			TRG				20171116		0.45	0.45	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			Perfluorob	375-73-5		51.95	NG_L				TRG				20171116		0.12	0.12	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			Perfluoroh	355-46-4		693.31	NG_L	D			TRG				20171116		1.67	1.67	5.1	70	6.06	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			Perfluoroo	1763-23-1		382.48	NG_L	D			TRG				20171116		2.88	2.88	5.1	70	7.58	75.76	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			13C5-PFHx	BDO-2217		126	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			13C4-PFHp	BDO-2218		125	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			13C8-PFOA	BDO-2219		102	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C9-PFNA	BDO-2221		64	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C6-PFDA	BDO-2222		91	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C7-PFUn	BDO-2223		94	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C2-PFDo	BDO-2112		86	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C2-PFTe	BDO-2224		98	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			d3-MeFOS	BDO-1838		99	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			d5-EtFOSA	BDO-1839		109	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:30:05	J8625-FS	1	1			13C3-PFBS	BDO-2226		86	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			13C3-PFHx	BDO-2227		109	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181002	20181005			20181010	10:44:00	20181023	09:40:58	J8625-FS	16.67	1			13C8-PFOS	BDO-2228		98	PCT_REC				SURR	SLSP	150	50	20171116				5.1					18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	10:02:45	J8626-FS	10	1			Perfluoroh	307-24-4		124.65	NG_L	D			TRG				20171116		1.73	1.73	5.1	70	4.55	45.45	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	DL1	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	10	1			Perfluoroo	335-67-1		156.31	NG_L	D			TRG				20171116		1.64	1.64	5.1	70	4.55	45.45	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	1	1			Perfluoro	375-95-1		0.61	NG_L	J			TRG				20171116		0.24	0.24	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	1	1			Perfluorod	335-76-2		0.45	NG_L	U			TRG				20171116		0.15	0.15	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	1	1			Perfluorou	2058-94-8		0.91	NG_L	U			TRG				20171116		0.26	0.26	5.1	70	0.91	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U			TRG				20171116		0.16	0.16	5.1	70	0.45	4.55	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S09GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	10:44:00	20181023	09:51:52	J8626-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L

Contract_I	DO_CTO_N	Phase	Installatior	Sample_Nu	CH2M_Coc	Analysis_G	Analytical_	PRC_Code	Lab_Code	Lab_Name	Leachate_I	SAMPLE_B	Extraction_	Result_Typ	Lab_QC_ty	SAMPLE_N	QC_Level	DateTime_	Date_Rece	Leachate_I	Leachate_	Extraction_	Extraction_	Analysis_D	Analysis_T	Lab_Sampl	Dilution	Run_Numb	PERCENT_I	PERCENT_I	Chem_Nan	Analyte_ID	Analyte_Vi	Original_A	Result_Uni	Lab_Qualif	Validator_	GC_Colum	Analysis_R	Result_Nat	QC_Contrc	QC_Accura	QC_Accura	Control_Lir	QC_Narrat	MDL	Detection_	QSM_Vers	DL	LOD	LOQ	SDG	Analysis_B	Validator_	Val_Date
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluorou	2058-94-8		0.89	NG_L	U		TRG					20171116		0.26	0.26	5.1	70	0.89	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluorod	307-55-1		0.45	NG_L	U		TRG					20171116		0.16	0.16	5.1	70	0.45	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluorot	72629-94-8		0.45	NG_L	U		TRG					20171116		0.13	0.13	5.1	70	0.45	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluorot	376-06-7		0.89	NG_L	U		TRG					20171116		0.22	0.22	5.1	70	0.89	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			N-Methyl F	2355-31-9		1.79	NG_L	U		TRG					20171116		0.5	0.5	5.1	70	1.79	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			N-Ethyl Pe	2991-50-6		0.89	NG_L	U		TRG					20171116		0.44	0.44	5.1	70	0.89	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluorob	375-73-5		0.45	NG_L	U		TRG					20171116		0.12	0.12	5.1	70	0.45	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluoroh	355-46-4		1.58	NG_L	J		TRG					20171116		0.1	0.1	5.1	70	0.36	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			Perfluoroo	1763-23-1		2.19	NG_L	J		TRG					20171116		0.17	0.17	5.1	70	0.45	4.46	18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C5-PFHx	BDO-2217		69	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C4-PFHp	BDO-2218		96	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C8-PFOA	BDO-2219		85	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C9-PFNA	BDO-2221		84	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C6-PFDA	BDO-2222		83	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C7-PFUn	BDO-2223		96	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C2-PFDo	BDO-2112		89	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C2-PFTe	BDO-2224		76	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			d3-MeFOS	BDO-1838		112	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			d5-EtFOSA	BDO-1839		97	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C3-PFBS	BDO-2226		60	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C3-PFHx	BDO-2227		98	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		
N6247016I	4164		POINT_MU	VC-S14GW	NONS	SVOA	537_MOD	ORG	BMSL_NO	FBATTELLE,	NULL	N	METHOD	000	REG	W	4	20181003	20181005			20181010	14:03:00	20181023	10:57:08	J8628-FS	1	1			13C8-PFOS	BDO-2228		83	PCT_REC			SURR	SLSP	150	50	20171116				5.1				18-0602	DP-18-0303		

ENVIRONMENTAL
Data Services, Inc.

DATA VALIDATION SUMMARY REPORT
NAVAL BASE VENTURA COUNTY, CALIFORNIA

Client: CH2M HILL, Inc., Corvallis, Oregon
SDG: 18-0602
Laboratory: Battelle Norwell Operations, Norwell, Massachusetts
Site: Naval Base Ventura County, CTO-4164, California
Date: January 11, 2019

PFCs			
EDS ID	Client Sample ID	Laboratory Sample ID	Matrix
1	VC-PM553-GW0601A-1018	J8613-FS	Water
2	VC-AQ-FB09-10012018	J8614-FS	Water
3	VC-AQ-EB09-10012018	J8615-FS	Water
4	VC-PM13-GW21-1018	J8616-FS	Water
5	VC-KCHA16-GW01-1018	J8617-FS	Water
6	VC-KCHA16-GW01P-1018	J8618-FS	Water
7	VC-KCHA16-GW02-1018	J8619-FS	Water
8	VC-KCHA16-GW03-1018	J8620-FS	Water
9	VC-AQ-FB10-10022018	J8621-FS	Water
10	VC-AQ-EB10-10022018	J8622-FS	Water
11	VC-S09GW01-1018	J8623-FS	Water
12	VC-S09GW04-1018	J8624-FS	Water
13	VC-S09GW04P-1018	J8625-FS	Water
14	VC-S09GW06-1018	J8626-FS	Water
15	VC-S14GW01-1018	J8627-FS	Water
16	VC-S14GW03-1018	J8628-FS	Water

A full data validation was performed on the analytical data for twelve water samples, two aqueous field blank samples, and two aqueous equipment blank samples collected on October 1-3, 2018 by CH2M HILL at the Naval Base Ventura County site in California. The samples were analyzed under the Battelle SOP Method for "Analysis of Perfluoroalkyl Substances in Environmental Samples by Liquid Chromatography and Tandem Mass Spectrometry (LC-MS/MS)".

Specific method references are as follows:

Analysis	*Method References*
PFCs	SOP 5-369

The data have been validated according to the protocols and quality control (QC) requirements of the analytical method, the DoD Quality Systems Manual for Environmental Laboratories, Version 5.1, February 2018, the Final Sampling and Analysis Plan Basewide Preliminary Assessments/Site Inspections of Per- and Polyfluoroalkyl Substances, August 2018, and the USEPA National Functional Guidelines for Organic Data Review as follows:

- The USEPA "Contract Laboratories Program National Functional Guidelines for Organic Superfund Methods Data Review," January 2017;
- and the reviewer's professional judgment.

The following data quality indicators were reviewed for this report:

Organics

- Date Completeness, Case Narrative & Custody Documentation
- Holding times
- Liquid Chromatography/Mass Spectrometry (LC/MS) Tuning
- Initial and continuing calibration summaries
- Method blank and field QC blank contamination
- Surrogate Spike recoveries
- Matrix Spike/Matrix Spike Duplicate (MS/MSD) recoveries
- Laboratory Control Sample/Laboratory Control Sample Duplicate (LCS/LCSD) recoveries
- Internal standard area and retention time summary forms
- Target Compound Identification
- Compound Quantitation
- Field Duplicate sample precision

A full (Level IV) data validation was performed with this review including a recalculation of 10% of the detected results in the samples.

Data Usability Assessment

There were no rejections of data.

Overall the data is acceptable for the intended purposes as qualified for the deficiencies detailed in this report.

Please note that any results qualified (U) due to blank contamination may be then qualified (J) due to another action. Therefore, the results may be qualified (UJ) due to the culmination of the blank contaminations and actions from other exceedances of QC criteria.

Perfluorinated Compounds (PFCs)

Data Completeness, Case Narrative & Custody Documentation

- The case narrative and chain-of-custody documentation were included in the data package as required. All criteria were met.

Holding Times

- All samples were extracted within 14 days for water samples and analyzed within 28 days.

LC/MS Tuning

- All criteria were met.

Initial Calibration

- All relative standard deviation (%RSD) and/or correlation coefficients criteria were met.

Continuing Calibration

- All percent recovery (%R) criteria were met.

Method Blank

- The method blanks exhibited the following contamination.

Blank ID	Compound	Conc. ng/L	Qualifier	Affected Samples
CR939PB-FS	PFOA	0.96	U	2-3, 5-6, 9-10, 16

Field QC Blank

- Field QC samples are summarized below.

Blank ID	Compound	Conc. ng/L	Qualifier	Affected Samples
VC-AQ-EB09-10012018	None - ND	-	-	-
VC-AQ-FB09-10012018	None - ND	-	-	-
VC-AQ-EB10-10022018	PFOS	0.24	None	All Associated ND or >10X
VC-AQ-FB10-10022018	None - ND	-	-	-
VC-SO-EB11-10062018	PFHxA	0.23	None	All Associated ND or >10X
VC-SO-FB11-10062018	PFHxS	0.33	U	16

Surrogate Spike Recoveries

- All samples exhibited acceptable surrogate %R values.

Matrix Spike/Matrix Spike Duplicate (MS/MSD) Recoveries

- MS/MSD samples were not analyzed.

Laboratory Control Samples

- The LCS samples exhibited acceptable percent recoveries (%R).

Internal Standard (IS) Area Performance

- All internal standards met response and retention time (RT) criteria.

Target Compound Identification

- All mass spectra and quantitation criteria were met.

Compound Quantitation

- Several samples were analyzed at various dilutions due to high concentrations of target compounds. The reporting limits were adjusted accordingly. No action was required.

Field Duplicate Sample Precision

- Field duplicate results are summarized below. The precision was acceptable.

Compound	VC-KCHA16-GW01-1018 ng/L	VC-KCHA16-GW01P-1018 ng/L	RPD	Qualifier
PFHxA	338.20	304.87	10%	None
PFHpA	13.04	13.41	3%	
PFBS	153.20	124.26	21%	
PFHxS	177.76	147.29	19%	
PFOS	8.94	7.50	18%	

Compound	VC-S09GW04-1018 ng/L	VC-S09GW04P-1018 ng/L	RPD	Qualifier
PFHxA	153.03	180.39	16%	None
PFHpA	114.46	128.79	12%	
PFOA	381.85	405.69	6%	
PFNA	9.77	9.85	1%	
PFDA	1.08	0.99	9%	
PFBS	51.02	51.95	2%	
PFHxS	717.18	693.31	3%	
PFOS	345.23	382.48	10%	

Please contact the undersigned at (757) 564-0090 if you have any questions or need further information.

Signed: Nancy Weaver Dated: 1/11/19

Nancy Weaver
Senior Chemist

Data Qualifier	Definition
U	The analyte was analyzed for, but was not detected above the level of the reported sample quantitation limit.
J	The analyte is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
J+	The result is an estimated quantity, but the result may be biased high.
J-	The result is an estimated quantity, but the result may be biased low.
NJ	The analysis has been "tentatively identified" or "presumptively" as present and the associated numerical value is the estimated concentration in the samples.
UJ	The analyte was analyzed for but was not detected. The reported quantitation limits is approximate and may be inaccurate or imprecise.
R	The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the samples.

BATTELLE
it can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM553-GW0601A-1018			
Battelle ID		J8613-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.270			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	7602.64 D	29.32	77.16	771.61
PFHpA	375-85-9	3916.20 D	24.69	77.16	771.61
PFOA	335-67-1	6727.83 D	27.78	77.16	771.61
PFNA	375-95-1	5.37	0.24	0.93	4.63
PFDA	335-76-2	0.46 U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93 U	0.27	0.93	4.63
PFDoA	307-55-1	0.46 U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46 U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93 U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85 U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93 U	0.45	0.93	4.63
PFBS	375-73-5	305.91 D	6.02	23.15	231.48
PFHxS	355-46-4	2350.63 D	5.09	18.52	231.48
PFOS	1763-23-1	110.92 D	1.76	4.63	46.30
Surrogate Recoveries (%)					
13C5-PFHxA		109 D			
13C4-PFHpA		109 D			
13C8-PFOA		109 D			
13C9-PFNA		106 D			
13C6-PFDA		84			
13C7-PFUnA		93			
13C2-PFDoA		87			
13C2-PFTeDA		86			
d3-MeFOSAA		89			
d5-EtFOSAA		82			
13C3-PFBS		120 D			
13C3-PFHxS		97 D			
13C8-PFOS		105 D			

BATTELLE
it can be done

2

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB09-10012018			
Battelle ID		J8614-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.290			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.43 U	0.16	0.43	4.31
PFHpA	375-85-9	0.43 U	0.14	0.43	4.31
PFOA	335-67-1	0.91 U	0.16	0.43	4.31 MBL
PFNA	375-95-1	0.86 U	0.22	0.86	4.31
PFDA	335-76-2	0.43 U	0.14	0.43	4.31
PFUnA	2058-94-8	0.86 U	0.25	0.86	4.31
PFDoA	307-55-1	0.43 U	0.16	0.43	4.31
PFTrDA	72629-94-8	0.43 U	0.13	0.43	4.31
PFTeDA	376-06-7	0.86 U	0.22	0.86	4.31
NMeFOSAA	2355-31-9	1.72 U	0.48	1.72	4.31
NEtFOSAA	2991-50-6	0.86 U	0.42	0.86	4.31
PFBS	375-73-5	0.43 U	0.11	0.43	4.31
PFHxS	355-46-4	0.34 U	0.09	0.34	4.31
PFOS	1763-23-1	0.43 U	0.16	0.43	4.31

Surrogate Recoveries (%)

13C5-PFHxA	85
13C4-PFHpA	83
13C8-PFOA	86
13C9-PFNA	74
13C6-PFDA	94
13C7-PFUnA	96
13C2-PFDoA	86
13C2-PFTeDA	93
d3-MeFOSAA	75
d5-EtFOSAA	83
13C3-PFBS	89
13C3-PFHxS	90
13C8-PFOS	88

BATTELLE
It can be done

3

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB09-10012018			
Battelle ID		J8615-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.45 U	0.17	0.45	4.55
PFHpA	375-85-9	0.45 U	0.15	0.45	4.55
PFOA	335-67-1	0.95 J U	0.16	0.45	4.55 MBL
PFNA	375-95-1	0.91 U	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	0.45 U	0.12	0.45	4.55
PFHxS	355-46-4	0.36 U	0.10	0.36	4.55
PFOS	1763-23-1	0.45 U	0.17	0.45	4.55

Surrogate Recoveries (%)

13C5-PFHxA	95
13C4-PFHpA	83
13C8-PFOA	91
13C9-PFNA	82
13C6-PFDA	84
13C7-PFUnA	87
13C2-PFDoA	85
13C2-PFTeDA	85
d3-MeFOSAA	71
d5-EtFOSAA	58
13C3-PFBS	73
13C3-PFHxS	78
13C8-PFOS	77

BATTELLE
It can be done

4

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-PM13-GW21-1018			
Battelle ID		J8616-FS			
Sample Type		SA			
Collection Date		10/01/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	245.61 D	2.88	7.58	75.76
PFHpA	375-85-9	48.74	0.15	0.45	4.55
PFOA	335-67-1	63.52	0.16	0.45	4.55
PFNA	375-95-1	1.00 J	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	397.55 D	1.97	7.58	75.76
PFHxS	355-46-4	755.99 D	1.67	6.06	75.76
PFOS	1763-23-1	94.92 D	2.88	7.58	75.76
Surrogate Recoveries (%)					
13C5-PFHxA		117 D			
13C4-PFHpA		99			
13C8-PFOA		79			
13C9-PFNA		74			
13C6-PFDA		90			
13C7-PFUnA		96			
13C2-PFDoA		93			
13C2-PFTeDA		99			
d3-MeFOSAA		99			
d5-EtFOSAA		102			
13C3-PFBS		128 D			
13C3-PFHxS		99 D			
13C8-PFOS		99 D			

BATTELLE
it can be done

5

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01-1018			
Battelle ID		J8617-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	338.20 D	1.73	4.55	45.45
PFHpA	375-85-9	13.04	0.15	0.45	4.55
PFOA	335-67-1	6.64 U	0.16	0.45	4.55
PFNA	375-95-1	0.91 U	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	153.20 D	1.18	4.55	45.45
PFHxS	355-46-4	177.76 D	1.00	3.64	45.45
PFOS	1763-23-1	8.94	0.17	0.45	4.55
Surrogate Recoveries (%)					
13C5-PFHxA		114			
13C4-PFHpA		121			
13C8-PFOA		89			
13C9-PFNA		82			
13C6-PFDA		88			
13C7-PFUnA		96			
13C2-PFDoA		91			
13C2-PFTeDA		99			
d3-MeFOSAA		91			
d5-EtFOSAA		71			
13C3-PFBS		139 D			
13C3-PFHxS		98 D			
13C8-PFOS		100			

MBL

BATTELLE
it can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW01P-1018			
Battelle ID		J8618-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.265			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	304.87 D	1.79	4.72	47.17
PFHpA	375-85-9	13.41	0.15	0.47	4.72
PFOA	335-67-1	5.94 U	0.17	0.47	4.72
PFNA	375-95-1	0.94 U	0.25	0.94	4.72
PFDA	335-76-2	0.47 U	0.15	0.47	4.72
PFUnA	2058-94-8	0.94 U	0.27	0.94	4.72
PFDoA	307-55-1	0.47 U	0.17	0.47	4.72
PFTrDA	72629-94-8	0.47 U	0.14	0.47	4.72
PFTeDA	376-06-7	0.94 U	0.24	0.94	4.72
NMeFOSAA	2355-31-9	1.89 U	0.53	1.89	4.72
NEtFOSAA	2991-50-6	0.94 U	0.46	0.94	4.72
PFBS	375-73-5	124.26 D	1.23	4.72	47.17
PFHxS	355-46-4	147.29 D	1.04	3.77	47.17
PFOS	1763-23-1	7.50	0.18	0.47	4.72
Surrogate Recoveries (%)					
13C5-PFHxA		103 D			
13C4-PFHpA		117			
13C8-PFOA		93			
13C9-PFNA		76			
13C6-PFDA		84			
13C7-PFUnA		87			
13C2-PFDoA		75			
13C2-PFTeDA		84			
d3-MeFOSAA		76			
d5-EtFOSAA		62			
13C3-PFBS		126 D			
13C3-PFHxS		85 D			
13C8-PFOS		94			

BATTELLE
it can be done

7

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW02-1018			
Battelle ID		J8619-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.270			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	163.76	1.76	4.63	46.30
PFHpA	375-85-9	39.13	0.15	0.46	4.63
PFOA	335-67-1	39.85	0.17	0.46	4.63
PFNA	375-95-1	0.93 U	0.24	0.93	4.63
PFDA	335-76-2	0.46 U	0.15	0.46	4.63
PFUnA	2058-94-8	0.93 U	0.27	0.93	4.63
PFDoA	307-55-1	0.46 U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46 U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93 U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85 U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93 U	0.45	0.93	4.63
PFBS	375-73-5	49.17	0.12	0.46	4.63
PFHxS	355-46-4	250.99	1.02	3.70	46.30
PFOS	1763-23-1	17.94	0.18	0.46	4.63
Surrogate Recoveries (%)					
13C5-PFHxA		111			
13C4-PFHpA		122			
13C8-PFOA		98			
13C9-PFNA		82			
13C6-PFDA		77			
13C7-PFUnA		82			
13C2-PFDoA		76			
13C2-PFTeDA		80			
d3-MeFOSAA		72			
d5-EtFOSAA		71			
13C3-PFBS		101			
13C3-PFHxS		87			
13C8-PFOS		79			

BATTELLE
it can be done

8

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-KCHA16-GW03-1018			
Battelle ID		J8620-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	229.43 Q	1.73	4.55	45.45
PFHpA	375-85-9	32.09	0.15	0.45	4.55
PFOA	335-67-1	66.04 Q	1.64	4.55	45.45
PFNA	375-95-1	0.91 U	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	69.78 Q	1.18	4.55	45.45
PFHxS	355-46-4	136.86 Q	1.00	3.64	45.45
PFOS	1763-23-1	10.97	0.17	0.45	4.55
Surrogate Recoveries (%)					
13C5-PFHxA		103 Q			
13C4-PFHpA		104			
13C8-PFOA		110 Q			
13C9-PFNA		75			
13C6-PFDA		78			
13C7-PFUnA		82			
13C2-PFDoA		84			
13C2-PFTeDA		94			
d3-MeFOSAA		72			
d5-EtFOSAA		73			
13C3-PFBS		126 Q			
13C3-PFHxS		89 Q			
13C8-PFOS		92			

BATTELLE
It can be done

9

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-FB10-10022018			
Battelle ID		J8621-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	375-85-9	0.44 U	0.14	0.44	4.39
PFOA	335-67-1	1.88 Yu	0.16	0.44	4.39
PFNA	375-95-1	0.88 U	0.23	0.88	4.39
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	2058-94-8	0.88 U	0.25	0.88	4.39
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	72629-94-8	0.44 U	0.13	0.44	4.39
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	2355-31-9	1.75 U	0.49	1.75	4.39
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	375-73-5	0.44 U	0.11	0.44	4.39
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	1763-23-1	0.44 U	0.17	0.44	4.39

MBL

Surrogate Recoveries (%)

13C5-PFHxA	90
13C4-PFHpA	88
13C8-PFOA	93
13C9-PFNA	84
13C6-PFDA	84
13C7-PFUnA	85
13C2-PFDoA	92
13C2-PFTeDA	94
d3-MeFOSAA	68
d5-EtFOSAA	74
13C3-PFBS	80
13C3-PFHxS	77
13C8-PFOS	75

BATTELLE
it can be done

16

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-AQ-EB10-10022018			
Battelle ID		J8622-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		AQ			
Sample Size		0.285			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.44 U	0.17	0.44	4.39
PFHpA	**375-85-9**	**0.44 U**	**0.14**	**0.44**	**4.39**
PFOA	335-67-1	1.11 Yu	0.16	0.44	4.39
PFNA	**375-95-1**	**0.88 U**	**0.23**	**0.88**	**4.39**
PFDA	335-76-2	0.44 U	0.14	0.44	4.39
PFUnA	**2058-94-8**	**0.88 U**	**0.25**	**0.88**	**4.39**
PFDoA	307-55-1	0.44 U	0.16	0.44	4.39
PFTrDA	**72629-94-8**	**0.44 U**	**0.13**	**0.44**	**4.39**
PFTeDA	376-06-7	0.88 U	0.22	0.88	4.39
NMeFOSAA	**2355-31-9**	**1.75 U**	**0.49**	**1.75**	**4.39**
NEtFOSAA	2991-50-6	0.88 U	0.43	0.88	4.39
PFBS	**375-73-5**	**0.44 U**	**0.11**	**0.44**	**4.39**
PFHxS	355-46-4	0.35 U	0.10	0.35	4.39
PFOS	**1763-23-1**	**0.24 J**	**0.17**	**0.44**	**4.39**

MBL

Surrogate Recoveries (%)

13C5-PFHxA	94
13C4-PFHpA	**87**
13C8-PFOA	94
13C9-PFNA	**85**
13C6-PFDA	83
13C7-PFUnA	**86**
13C2-PFDoA	85
13C2-PFTeDA	**82**
d3-MeFOSAA	64
d5-EtFOSAA	**63**
13C3-PFBS	86
13C3-PFHxS	**84**
13C8-PFOS	83

BATTELLE
it can be done

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW01-1018			
Battelle ID		J8623-FS			
Sample Type		SA			
Collection Date		10/03/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.270			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	59.15	0.18	0.46	4.63
PFHpA	375-85-9	18.77	0.15	0.46	4.63
PFOA	335-67-1	220.43	1.67	4.63	46.30
PFNA	375-95-1	2.52 J	0.24	0.93	4.63
PFDA	335-76-2	0.16 J	0.15	0.46	4.63
PFUnA	2058-94-8	0.93 U	0.27	0.93	4.63
PFDoA	307-55-1	0.46 U	0.17	0.46	4.63
PFTrDA	72629-94-8	0.46 U	0.14	0.46	4.63
PFTeDA	376-06-7	0.93 U	0.23	0.93	4.63
NMeFOSAA	2355-31-9	1.85 U	0.52	1.85	4.63
NEtFOSAA	2991-50-6	0.93 U	0.45	0.93	4.63
PFBS	375-73-5	27.51	0.12	0.46	4.63
PFHxS	355-46-4	78.91	1.02	3.70	46.30
PFOS	1763-23-1	70.73	0.18	0.46	4.63
Surrogate Recoveries (%)					
13C5-PFHxA		91			
13C4-PFHpA		103			
13C8-PFOA		115			
13C9-PFNA		67			
13C6-PFDA		77			
13C7-PFUnA		82			
13C2-PFDoA		71			
13C2-PFTeDA		77			
d3-MeFOSAA		83			
d5-EtFOSAA		81			
13C3-PFBS		72			
13C3-PFHxS		99			
13C8-PFOS		81			

BATTELLE
it can be done

12

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04-1018			
Battelle ID		J8624-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	153.03 D	2.88	7.58	75.76
PFHpA	375-85-9	114.46 D	2.42	7.58	75.76
PFOA	335-67-1	381.85 D	2.73	7.58	75.76
PFNA	375-95-1	9.77	0.24	0.91	4.55
PFDA	335-76-2	1.08 J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	51.02	0.12	0.45	4.55
PFHxS	355-46-4	717.18 D	1.67	6.06	75.76
PFOS	1763-23-1	345.23 D	2.88	7.58	75.76
Surrogate Recoveries (%)					
13C5-PFHxA		131 D			
13C4-PFHpA		129 D			
13C8-PFOA		104 D			
13C9-PFNA		62			
13C6-PFDA		91			
13C7-PFUnA		97			
13C2-PFDoA		98			
13C2-PFTeDA		100			
d3-MeFOSAA		109			
d5-EtFOSAA		110			
13C3-PFBS		92			
13C3-PFHxS		92 D			
13C8-PFOS		94 D			

BATTELLE
it can be done

13

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW04P-1018			
Battelle ID		J8625-FS			
Sample Type		SA			
Collection Date		10/02/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	180.39 D	2.88	7.58	75.76
PFHpA	375-85-9	128.79 D	2.42	7.58	75.76
PFOA	335-67-1	405.69 D	2.73	7.58	75.76
PFNA	375-95-1	9.85	0.24	0.91	4.55
PFDA	335-76-2	0.99 J	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	51.95	0.12	0.45	4.55
PFHxS	355-46-4	693.31 D	1.67	6.06	75.76
PFOS	1763-23-1	382.48 D	2.88	7.58	75.76
Surrogate Recoveries (%)					
13C5-PFHxA		126 D			
13C4-PFHpA		125 D			
13C8-PFOA		102 D			
13C9-PFNA		64			
13C6-PFDA		91			
13C7-PFUnA		94			
13C2-PFDoA		86			
13C2-PFTeDA		98			
d3-MeFOSAA		99			
d5-EtFOSAA		109			
13C3-PFBS		86			
13C3-PFHxS		109 D			
13C8-PFOS		98 D			

BATTELLE
it can be done

14

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S09GW06-1018			
Battelle ID		J8626-FS			
Sample Type		SA			
Collection Date		10/03/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	124.65	1.73	4.55	45.45
PFHpA	375-85-9	17.96	0.15	0.45	4.55
PFOA	335-67-1	156.31	1.64	4.55	45.45
PFNA	375-95-1	0.61 J	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	91.23	0.12	0.45	4.55
PFHxS	355-46-4	66.27	0.10	0.36	4.55
PFOS	1763-23-1	2.04 J	0.17	0.45	4.55
Surrogate Recoveries (%)					
13C5-PFHxA		128			
13C4-PFHpA		109			
13C8-PFOA		109			
13C9-PFNA		72			
13C6-PFDA		81			
13C7-PFUnA		79			
13C2-PFDoA		71			
13C2-PFTeDA		74			
d3-MeFOSAA		103			
d5-EtFOSAA		89			
13C3-PFBS		83			
13C3-PFHxS		108			
13C8-PFOS		103			

MW 12/28/18

BATTELLE
It can be done

15

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW01-1018			
Battelle ID		J8627-FS			
Sample Type		SA			
Collection Date		10/03/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.275			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	303.39 D	4.32	11.36	113.64
PFHpA	375-85-9	45.61	0.15	0.45	4.55
PFOA	335-67-1	171.14 D	4.09	11.36	113.64
PFNA	375-95-1	1.83 J	0.24	0.91	4.55
PFDA	335-76-2	0.45 U	0.15	0.45	4.55
PFUnA	2058-94-8	0.91 U	0.26	0.91	4.55
PFDoA	307-55-1	0.45 U	0.16	0.45	4.55
PFTrDA	72629-94-8	0.45 U	0.14	0.45	4.55
PFTeDA	376-06-7	0.91 U	0.23	0.91	4.55
NMeFOSAA	2355-31-9	1.82 U	0.51	1.82	4.55
NEtFOSAA	2991-50-6	0.91 U	0.45	0.91	4.55
PFBS	375-73-5	107.88	0.12	0.45	4.55
PFHxS	355-46-4	1798.48 D	12.50	45.45	568.18
PFOS	1763-23-1	1808.55 D	4.32	11.36	113.64
Surrogate Recoveries (%)					
13C5-PFHxA		118 D			
13C4-PFHpA		91			
13C8-PFOA		109 D			
13C9-PFNA		67			
13C6-PFDA		95			
13C7-PFUnA		96			
13C2-PFDoA		90			
13C2-PFTeDA		91			
d3-MeFOSAA		113			
d5-EtFOSAA		102			
13C3-PFBS		86			
13C3-PFHxS		112 D			
13C8-PFOS		80 D			

BATTELLE
It can be done

16

Project Client: CH2M
Project Name: CTO-4164 Naval Base Ventura County, California
Project No.: 100110125-01

Client ID		VC-S14GW03-1018			
Battelle ID		J8628-FS			
Sample Type		SA			
Collection Date		10/03/2018			
Extraction Date		10/10/2018			
Analysis Date		10/23/2018			
Analytical Instrument		Sciex 5500 LC/MS/MS			
% Moisture		NA			
Matrix		GW			
Sample Size		0.280			
Size Unit-Basis		L			
Units		ng/L	MDL	LOD	LOQ
PFHxA	307-24-4	0.45 U	0.17	0.45	4.46
PFHpA	375-85-9	0.45 U	0.14	0.45	4.46
PFOA	335-67-1	4.43 J U	0.16	0.45	4.46 MBL
PFNA	375-95-1	0.89 U	0.23	0.89	4.46
PFDA	335-76-2	0.45 U	0.14	0.45	4.46
PFUnA	2058-94-8	0.89 U	0.26	0.89	4.46
PFDoA	307-55-1	0.45 U	0.16	0.45	4.46
PFTrDA	72629-94-8	0.45 U	0.13	0.45	4.46
PFTeDA	376-06-7	0.89 U	0.22	0.89	4.46
NMeFOSAA	2355-31-9	1.79 U	0.50	1.79	4.46
NEtFOSAA	2991-50-6	0.89 U	0.44	0.89	4.46
PFBS	375-73-5	0.45 U	0.12	0.45	4.46
PFHxS	355-46-4	1.58 J U	0.10	0.36	4.46 FBL
PFOS	1763-23-1	2.19 J	0.17	0.45	4.46

Surrogate Recoveries (%)

13C5-PFHxA	69
13C4-PFHpA	96
13C8-PFOA	85
13C9-PFNA	84
13C6-PFDA	83
13C7-PFUnA	96
13C2-PFDoA	89
13C2-PFTeDA	76
d3-MeFOSAA	112
d5-EtFOSAA	97
13C3-PFBS	60
13C3-PFHxS	98
13C8-PFOS	83

LOCATION_NAME	SITE_NAME	INSTALLATION_ID	LOCATION_TYPE	LOCATION_TYPE_DESC	SDG	COORD_X	COORD_Y	ANALYTICAL_METHOD_GRP_DESC	SAMPLE_NAME	SAMPLE_MATRIX	SAMPLE_MATRIX_DESC	COLLECT_DATE
		POINT_MUGU_NAS			18-0602			Perfluoroalkyl Compounds	VC-AQ-EB09-10012018	WQ	Water for QC samples	01-Oct-18
		POINT_MUGU_NAS			18-0602			Perfluoroalkyl Compounds	VC-AQ-FB09-10012018	WQ	Water for QC samples	01-Oct-18
06W01A	BASEWIDE PFAS	POINT_MUGU_NAS	WLM	Monitoring well	18-0602	6222083.7	1865764	Perfluoroalkyl Compounds	VC-PM553-GW0601A-1018	WG	Ground water	01-Oct-18
PM13-MW21	BASEWIDE PFAS	POINT_MUGU_NAS	WLM	Monitoring well	18-0602	6226537.5	1865618.4	Perfluoroalkyl Compounds	VC-PM13-GW21-1018	WG	Ground water	01-Oct-18