



**Off-Base Drinking Water Sample Results,  
Combined Level 2 and Level 4 Laboratory Report,  
Electronic Data Deliverable, Data Validation Report,  
and the Sample Location Figure, SDG 18-0313**

*Naval Air Warfare Center Warminster  
Warminster, Pennsylvania*

August 2019

N62269\_001182  
WARMINSTER\_NAWC  
SSIC 5000-33c

**LABORATORY DATA PACKAGE, 18-0313, NAS WILLOW GROVE NAWC  
WARMINSTER PA**  
05/27/2018  
BATTELLE

Approved for public release: distribution unlimited.

**Naval Air Station Joint Reserve Base Willow Grove,  
PA**

**Project No 100117920-WE04  
PFAS in drinking water**

*DW*

*Batch 18-0313*

*Package DP-18-0112*

Submitted to:

Tetra Tech

661 Anderson Drive Foster Plaza 7

Pittsburgh, PA 15220 USA

Submitted by:

Battelle Norwell Operations  
141 Longwater Drive Suite 202  
Norwell, MA 02061

***BATTELLE***

**It can be done**

**Naval Air Station Joint Reserve Base Willow Grove,  
PA**

**Project No 100117920-WE04**

**PFAS in drinking water**

*DW*

*Batch 18-0313*

*Package DP-18-0112*

Submitted to:

Tetra Tech

661 Anderson Drive Foster Plaza 7

Pittsburgh, PA 15220 USA

NELAP Accreditation Number: E87856 (Florida Department of Health)

Submitted by:

Battelle Norwell Operations

141 Longwater Drive Suite 202

Norwell, MA 02061

Analyst Approval:



schumitzd@battelle.org  
2018.05.17 14:39:04 -04'00'

QC Chemist Approval:



Digitally signed by devinec@battelle.org  
DN: cn=devinec@battelle.org  
Date: 2018.05.25 17:02:46 -04'00'

Project Manager Approval:



Digitally signed by Jonathan Thorn  
Date: 2018.05.27 12:07:09 -04'00'

**BATTELLE**

**It can be done**



# Naval Air Station Joint Reserve Base Willow Grove, PA

## Project No 100117920-WE04 PFAS in drinking water DW

*Batch 18-0313*


















*Package DP-18-0112*

<b>1</b>	<b><i>Work Plan</i></b> Laboratory Work Plan, Addendums To Work Plan, Memos From Project Manager, Special Instructions, Chain-of-Custody Reports.	<b>1</b>
<b>2</b>	<b><i>Tables</i></b> Analytical Data Tables, Qualifier Definitions.	<b>18</b>
<b>3</b>	<b><i>Miscellaneous Documentation</i></b> Case Narrative, Miscellaneous Documentation Form, Quality Control Summary, Example Calculations, Internal Standard Recovery Report, Retention Time Window Report.	<b>34</b>
<b>4</b>	<b><i>Sample Preparation Records</i></b> Sample Preparation Records, Dilution Worksheets, Standard Preparation Records, Certificates Of Analysis, GPC Check Report.	<b>138</b>
<b>5</b>	<b><i>Analytical Calibrations</i></b> Analytical Sequence, Analytical Method, Tune Report, Initial Calibration, Pesticide Degradation Report, RF Summary, Calibration Verifications, Independent Calibration Verification Check.	<b>152</b>
<b>6</b>	<b><i>Analytical Data</i></b> Raw Data Quantification Reports.	<b>261</b>
<b>7</b>	<b><i>Chromatograms</i></b> Sample And Standard Chromatograms.	<b>318</b>
<b>8</b>	<b><i>Unused Data</i></b>	<b>631</b>

**BATTELLE**

It can be done

## Signature Page

Battelle 2018 (1 of 2) Signature Page			
Name (Printed)	Signature	Initials	Date
Jonathan Thorn		JRT	4/4/2018
Robert Lizotte, Jr.		BL	4-4-2018
FRANC PALA		FP	4-4-2018
Carla Devine		CRD	4/4/18
Denise Schumitz		DNS	4/4/18
Carol Ann McManis		CM	4/4/2018
Rich Rostucci		RR	4/4/2018
Michael Mena		MM	4/4/2018
Christie Usher		CU	4/4/18
Kevin Matroney		KM	4/4/18
Stephanie Schmitz		SAS	4/4/18
Jordan Tower		JT	4/4/18
KRISTEN NICHOLS		KN	4/4/18
Quimiao H Brown		CB	4/4/18
Matt Schumitz		MS	4-4-18
Sam Guimaraes		SG	4-4-18
Lauren Griffith		LMG	4.4.18



# Work Plan



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### 1.0 GENERAL PROJECT INFORMATION

**Project Title:** WE04 PFAS Analysis  
**Project Number:** 100117920-WE04  
**Client:** Tetra Tech  
 661 Anderson Drive Foster Plaza 7  
 Pittsburgh, PA 15220  
 USA  
  
**Client Contact Information:** Andrew Frebowitz  
 Project Manager  
 (610) 382-1170(V)  
 NA  
 andy.frebowitz@tetrattech.com  
  
**Effective Date of QAPP:** 5/4/2018  
**Version Number:** 100117920-WE04(L)-01  
**Project Manager:** Thorn, Jonathan  
**Laboratory Task Manager:** Thorn, Jonathan  
**Deliverable Due Date:** 5/22/2018

### 2.0 SCOPE OF WORK

**Overview:** Analysis of drinking water samples collected at Naval Air Station Joint Reserve Base Willow Grove.  
**Matrix:** Water

### 2.1 TECHNICAL APPROACH

#### 2.1.1 Sample Receipt, Storage, and Handling

The list of samples for this project plan are presented in Attachment 1.

**Storage Directions:** Store refrigerated.  
**Sub\_Sampling:** None  
**Procedures:** NA  
**Contact:** NA  
**Comment:** NA  
**Archiving:** Store for six months after delivery of final data. Notify client prior to disposal of samples.  
**Disposal:** Dispose of samples in the proper waste stream.



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### 2.1.2 Sample Preparation

Up to 150 field samples and 150 field reagent blanks (FRB) per quarter. FRB samples will only be analyzed if the corresponding sample has detected levels of any PFAS analyte at or above the LOQ.

Samples Expected:	Samples Per Batch:	Batches Expected:
300	20	15

Batch quality control samples are defined in Table 1.

Target samples are presented in Attachment 1.

**Table 1: Quality Control Samples**

Type:	Description:	Count:	Rgt:	Reference:	Comment:
PB	Laboratory control reagent blank.	1 per batch	--	NA	Millipore water with Trizma
LCS	Laboratory Control Sample	1 per batch	No	NA	Millipore water with Trizma
MS	Spiked field sample for determining method accuracy in the presence of matrix.	1 per batch	--	NA	MS/MSD indicated on COC
MSD	Spiked field sample for determining method accuracy and precision in the presence of matrix.	1 per batch	--	NA	MS/MSD indicated on COC

### 2.1.3 Extraction/Preparation

#### 2.1.3.1 Extraction

SOP No.-Rev:	<b>5-371-03</b>
SOP Title:	<i>ANALYSIS OF POLY AND PERFLUOROALKYL SUBSTANCES IN DRINKING WATER SAMPLES BY LIQUID CHROMATOGRAPHY AND TANDEM MASS SPECTROMETRY (LC-MS/MS) FOLLOWING EPA METHOD 537.1</i>
Sample Size:	250 ml
SIS and LCS/MS Compounds:	Defined in Table 2.
Deviations:	None
Comments:	<ul style="list-style-type: none"> <li>• MQO requirements per SOP 5-371 (EPA Method 537 Version 1.1).</li> <li>• FRB samples will only be analyzed if associated field sample has hits above the LOQ for any individual analyte.</li> </ul>

**Table 2: SIS and LCS/MS Spiking Level**



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

Standard Type	Standard Contents	Spike Amount (ng)	Volume (uL)	Comment
PFAS - 537.1 Surrogate Solution	JV60 SIS	~ 0.100 - 0.40 ng	50 uL	NA
PFAS - 537.1 Second Source LCS/MS Solution	JV41 LCS/MS	~ 2.00 - 2.50	50 uL	LCS samples - vary each batch (50, 75, 100, 150 µL spikes)
PFAS - 537.1 Second Source LCS/MS Solution	JV41 LCS/MS	~ 3.00 - 3.8 ng	75 uL	MS/MSD samples - vary each batch (75, 100, 150 µL spikes)

### 2.1.3.2 Cleanup

None.

RIS spiking levels are presented in Table 3.

Extract PIV (uL): 1000

**Table 3: RIS Spiking Level**

Standard Type	Standard Contents	Spike Amount (ng)	Volume (uL)	Comment
PFAS - 537.1 Internal Standard Solution	JV59 RIS	~ 0.100 - 0.40 ng	50 uL	NA

### 2.1.4 Instrumental Analysis

The list of analytes along with data quality criteria are presented in Attachment 2.

- 1) SOP\_No-Rev: **5-371-03**
- SOP\_Title: *ANALYSIS OF POLY AND PERFLUOROALKYL SUBSTANCES IN DRINKING WATER SAMPLES BY LIQUID CHROMATOGRAPHY AND TANDEM MASS SPECTROMETRY (LC-MS/MS) FOLLOWING EPA METHOD 537.1*
- Deviations: None
- Comments:
  - MQO requirements per SOP 5-371 (EPA Method 537 Version 1.1).
  - FRB samples will only be analyzed if associated field sample has hits above the LOQ for any individual analyte.



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### 2.2. DELIVERABLES

<b>Deliverables Due:</b>	5/22/2018
<b>LIMS Reports:</b>	Yes
<b>Histograms:</b>	No
<b>Excel Tables:</b>	Yes
<b>EICs:</b>	No
<b>Chromatograms:</b>	No
<b>EDDs:</b>	Yes
<b>Comments:</b>	Each data set will be due 21 days from receipt of samples Full QSM data package showing all aspects of Table B-15 Tetra Tech EDD format

### 3.0 QUALITY

The Method Quality Objectives are defined in Attachment 3.

### 4.0 ORGANIZATION AND COMMUNICATION

#### 4.1 ORGANIZATION

The project team is defined in Table 4. Supervisors may make substitutions with Project Manager concurrence.

**Table 4: Project Team and Roles**

Staff Member	Role	Comment
Jonathan R. Thorn	Project Manager	NA
Stephanie A. Schultz	Sample Preparation	NA
Denise M. Schumitz	LC-MS/MS Analysis	NA
Matt D. Schumitz	Sample Custody	NA
Carla R. Devine	Quality Control Officer	NA
Zachary J. Willenberg	Quality Assurance Officer	NA

#### 4.2 COMMUNICATION

A kick-off meeting will be held to discuss project scope and goals.

### 5.0 SCHEDULE

The project schedule is presented in Table 5.





It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

**Table 5. Schedule of Laboratory Activities**

<b>Activity:</b>	<b>Start Date:</b>	<b>End Date:</b>	<b>TAT (days):</b>	<b>Comment:</b>
Sample Receipt	05/01/2018	12/31/2018	244	NA
Sample Preparation	05/04/2018	01/31/2019	272	NA
Instrument Analysis	05/07/2018	01/31/2019	269	NA
Quality Control Review	05/14/2018	01/31/2019	262	NA
Final Data Reporting	05/18/2018	01/31/2019	258	NA
Quality Assurance Review	05/21/2018	01/31/2019	255	NA

### 6.0 BUDGET

The labor budget for the analytical task is presented in Table 6.

**Table 6. Labor Budget (Laboratory Analytical Task)**

<b>Labor Activity:</b>	<b>Hours/ Batch:</b>	<b>Batches:</b>	<b>Total Hours:</b>	<b>Comment:</b>
Sample Receipt	2	1	2	All labor hours are based on a batch of 20 field samples.
Sample Preparation	8	1	8	All labor hours are based on a batch of 20 field samples.
Instrument Analysis	8	1	8	All labor hours are based on a batch of 20 field samples.
Quality Control Review	3	1	3	All labor hours are based on a batch of 20 field samples.
Final Data Reporting	1	1	1	All labor hours are based on a batch of 20 field samples.
Quality Assurance Review	1	1	1	All labor hours are based on a batch of 20 field samples.

### 7.0 STAFF DEVELOPMENT

None anticipated



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### Attachment 1: Target Samples

**Shipment:** SHP-180501-01  
**Status:** Approved  
**Description:** WE04  
**Range:** J5964-J5969  
**Comment:** NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J5964	WGNA-043018-RW-3103	04/30/2018 10:10 am	DW	R0118 (NA)			MSMSD
2	J5965	WGNA-043018-FRB-3103	04/30/2018 10:05 am	DW	R0118 (NA)			
3	J5966	NAWC-043018-RW-207	04/30/2018 10:40 am	DW	R0118 (NA)			
4	J5967	NAWC-043018-FRB-207	04/30/2018 10:35 am	DW	R0118 (NA)			
5	J5968	WGNA-043018-RW-3409	04/30/2018 1:40 pm	DW	R0118 (NA)			
6	J5969	WGNA-043018-FRB-3409	04/30/2018 1:35 pm	DW	R0118 (NA)			

**Shipment:** SHP-180502-02  
**Status:** Approved  
**Description:** WE04  
**Range:** J5970-J5977  
**Comment:** NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J5970	WGNA-050118-RW-3385	05/01/2018 9:10 am	DW	R0118 (NA)			
2	J5971	WGNA-050118-FRB-3385	05/01/2018 9:05 am	DW	R0118 (NA)			
3	J5972	WGNA-050118-RW-3178	05/01/2018 9:40 am	DW	R0118 (NA)			
4	J5973	WGNA-050118-FRB-3178	05/01/2018 9:35 am	DW	R0118 (NA)			
5	J5974	NAWC-050118-RW-304	05/01/2018 10:10 am	DW	R0118 (NA)			
6	J5975	NAWC-050118-FRB-304	05/01/2018 10:05 am	DW	R0118 (NA)			
7	J5976	NAWC-050118-RW-098	05/01/2018 10:40 am	DW	R0118 (NA)			
8	J5977	NAWC-050118-FRB-098	05/01/2018 10:35 am	DW	R0118 (NA)			

**Shipment:** SHP-180508-02  
**Status:** Pending  
**Description:** WE04  
**Range:** J6148-J6170  
**Comment:** NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J6148	NAWC-050718-RW-316	05/07/2018 10:10 am	DW	R0118 (NA)			
2	J6149	NAWC-050718-FRB-316	05/07/2018 10:05 am	DW	R0118 (NA)			
3	J6150	NAWC-050718-RW-180	05/07/2018 10:40 am	DW	R0118 (NA)			
4	J6151	NAWC-050718-FRB-180	05/07/2018 10:35 am	DW	R0118 (NA)			
5	J6152	NAWC-050718-RW-275	05/07/2018 11:10 am	DW	R0118 (NA)			
6	J6153	NAWC-050718-FRB-275	05/07/2018 11:05 am	DW	R0118 (NA)			
7	J6154	NAWC-050718-RW-145	05/07/2018 12:40 pm	DW	R0118 (NA)			



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

**Shipment:** SHP-180508-02  
**Status:** Pending  
**Description:** WE04  
**Range:** J6148-J6170  
**Comment:** NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
8	J6155	NAWC-050718-FRB-145	05/07/2018 12:35 pm	DW	R0118 (NA)			
9	J6156	NAWC-050718-RW-357	05/07/2018 1:10 pm	DW	R0118 (NA)			
10	J6157	NAWC-050718-FRB-357	05/07/2018 1:05 pm	DW	R0118 (NA)			
11	J6158	NAWC-050718-RW-162	05/07/2018 1:40 pm	DW	R0118 (NA)			
12	J6159	NAWC-050718-FRB-162	05/07/2018 1:35 pm	DW	R0118 (NA)			
13	J6160	WGNA-050718-RW-0800	05/07/2018 2:10 pm	DW	R0118 (NA)			
14	J6161	WGNA-050718-FRB-0800	05/07/2018 2:05 pm	DW	R0118 (NA)			
15	J6162	WGNA-050718-RW-0335	05/07/2018 2:40 pm	DW	R0118 (NA)			
16	J6163	WGNA-050718-FRB-0335	05/07/2018 2:35 pm	DW	R0118 (NA)			
17	J6164	WGNA-050718-RW-3556	05/07/2018 3:10 pm	DW	R0118 (NA)			
18	J6165	WGNA-050718-FRB-3556	05/07/2018 3:05 pm	DW	R0118 (NA)			
19	J6166	NAWC-050718-RW-356	05/07/2018 3:40 pm	DW	R0118 (NA)			
20	J6167	NAWC-050718-FRB-356	05/07/2018 3:35 pm	DW	R0118 (NA)			
21	J6168	NAWC-050718-RW-289	05/07/2018 4:10 pm	DW	R0118 (NA)			
22	J6169	NAWC-050718-FRB-289	05/07/2018 4:05 pm	DW	R0118 (NA)			
23	J6170	WGNA-050718-DUP-35	05/07/2018 7:00 am	DW	R0118 (NA)			

**Shipment:** SHP-180511-02  
**Status:** Pending  
**Description:** WE04  
**Range:** J6204-J6212  
**Comment:** NA

No:	BDO Id:	Client Sample ID:	Collection Date:	Matrix:	Storage Facility:	Location:	No:	Comments:
1	J6204	NAWC-051018-RW-303	05/10/2018 9:10 am	DW	R0118 (NA)			
2	J6206	WGNA-051018-RW-3220	05/10/2018 9:40 am	DW	R0118 (NA)			
3	J6208	NAWC-051018-RW-177	05/10/2018 10:40 am	DW	R0118 (NA)			
4	J6210	WGNA-051018-RW-3295	05/10/2018 3:10 pm	DW	R0118 (NA)			
5	J6212	WGNA-051018-DUP-36	05/10/2018 7:00 am	DW	R0118 (NA)			



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

## Attachment 2: Test Codes

<b>Project Test Code Name:</b>	Master_371
<b>SOP Reference:</b>	5-371 - ANALYSIS OF POLY AND PERFLUOROALKYL SUBSTANCES IN DRINKING WATER SAMPLES BY LIQUID CHROMATOGRAPHY AND TANDEM MASS SPECTROMETRY (LC-MS/MS) FOLLOWING EPA METHOD 537.1
<b>Description:</b>	PFAS in drinking water
<b>Matrix:</b>	L - Liquid Samples, like water or sea water, prepared and analyzed under the same class of detection limits.
<b>Detection Limit Study:</b>	5-371
<b>Instrument:</b>	LC-MS/MS
<b>MQO Criteria</b>	Universal_LC
<b>Standard Report:</b>	Standard Result Report

Method Specific Reporting		Holding Times (days)		Data Flags
<b>Result Units:</b>	ng/L	<b>Unit Conversion:</b>	(none)	<b>Sample:</b> 14 <b>DL_Flag:</b> U
<b>Weight Basis:</b>	Liquid	<b>Result Format:</b>	Fixed Digits	<b>Frozen:</b> 40 <b>RL_Flag:</b> J
<b>Standard Basis:</b>	RIS	<b># of Figures/Digits:</b>	2	<b>Extract:</b> 28 <b>PB_Flag:</b> B
<b>Oil Weight Basis:</b>	No	<b>Oil Weight Source:</b>	Oil Weight	<b>DIL_Flag:</b> D
<b>U-Value Substitution:</b>	U-Flag=MD	<b>Histograms:</b>	No	<b>HT_Flag:</b> T
<b>ECD_Reporting:</b>	No			

No:	Analyte:	Report Name:	Type	RIS	SIS	Hidden:	Graph:
1	Perfluoro-n-hexanoic acid	PFHxA	T	13C2-PFOA		No	No
2	Perfluoro-n-heptanoic Acid	PFHpA	T	13C2-PFOA		No	No
3	Perfluoro-n-octanoic Acid	PFOA	T	13C2-PFOA		No	No
4	Perfluorononanoic Acid	PFNA	T	13C2-PFOA		No	No
5	Perfluoro-n-decanoic Acid	PFDA	T	13C2-PFOA		No	No
6	Perfluoro-n-undecanoic acid	PFUnA	T	13C2-PFOA		No	No
7	Perfluoro-n-dodecanoic acid	PFDoA	T	13C2-PFOA		No	No
8	Perfluoro-n-tridecanoic acid	PFTTrDA	T	13C2-PFOA		No	No
9	Perfluoro-n-tetradecanoic acid	PFTeDA	T	13C2-PFOA		No	No
10	N-methylperfluoro-1-octanesulfonamidoacetic acid	NMeFOSAA	T	d3-MeFOSAA		No	No
11	N-ethylperfluoro-octanesulfonamidoacetic acid	NEtFOSAA	T	d3-MeFOSAA		No	No
12	Perfluoro-1-butanefulfonate	PFBS	T	13C4-PFOS		No	No
13	Perfluoro-1-octanesulfonate	PFOS	T	13C4-PFOS		No	No
14	Perfluoro-1-hexanesulfonate	PFHxS	T	13C4-PFOS		No	No
1	13C2-PFHxA	13C2-PFHxA	SIS			No	No
2	13C2-PFDA	13C2-PFDA	SIS			No	No



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### Attachment 2: Test Codes

**Project Test Code Name:** Master\_371

No:	Analyte:	Report Name:	Type	RIS	SIS	Hidden:	Graph:
3	d5-EtFOSAA	d5-EtFOSAA	SIS			No	No
<b>Total Analytes:</b>		17					

**Subtract Peaks:**

None

**Sum Peaks:**

None



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### Attachment 2: Test Codes

**Project Test Code Name:** Master\_371

**ICAL Acceptance Criteria:**

Curve Fit:	Limit Mean(%):	Mean Qual:	Limit Ind.:	Ind. Qual:	Min Points:	Points Qual:	Comments:
Linear	NA	NA	0.995	N	5	N	NA
Average RF	15	N	25	N	5	N	NA
Linear (0,0)	NA	NA	0.995	N	5	N	NA
Quadratic	NA	NA	0.995	N	6	N	NA
Quadratic (0,0)	NA	NA	0.995	N	6	N	NA

**Continuing Calibration Verification Criteria:**

CCV Name: Standard							
Frequency Hrs:	Mean PD(%):	Individual PD(%):	RIS/SIS RT Window (min):	Area Limit Low(%):	Area Limit High(%):	Comment:	
12 (N)	20 (N)	25 (N)	0.07 (N)	-50	100 (N)	Lab Default Continuing Calibration Verification Criteria	

**Independent Calibration Verification:**

ICC Name: Standard							
Mean PD Limit(%):	Ind. PD Limit(%):	RIS/SIS Window Limit (Secs):	Area Limit High(%):	Area Limit Low(%):	Comment:		
15 (N)	20 (N)	0.07 (N)	-50	100 (N)	Standard laboratory criteria for ICCs		

**Mass Discrimination Criteria:**

*None*

**Degradation Check Criteria:**

*None*



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### Attachment 3: Method Quality Objectives

MQO Application	<i>Universal_LC</i>		
MQO:	Acceptance Criteria	Qual:	Corrective Action:
Procedural Blank	Samples must be greater than five times the blank concentration (>5xPB).	B	Review with Project Manager; re-analyze or justify results in project records.
PB Measurement Quality Objective	Organic results in the Procedural Blank are less than 1/2 times the LOQ (<1/2xLOQ)	N	Review with Project Manager; re-analyze or justify results in project records.
Laboratory Control Sample	Recovery values 70-130%.	N	Review with project manager; re-analyze or justify reporting the results in project records.
Matrix Spike / Matrix Spike Duplicate Recovery	Organics 70-130%. Analyte concentration in MS/MSD must be greater than five times reported background concentration. Organics Results in the Target is less than 5 times the Original	N n	Review with Project Manager; re-analyze or justify reporting results in the project records.
Matrix Spike/Spike Duplicate Precision	Organics results less than 30% Relative Percent Difference (RPD). Analyte concentration in MS/MSD must be greater than five times reported background concentration. Organics Results in the Target is less than 5 times the Original	N n	Review with Project Manager; re-analyze or justify reporting results in the project records.
Standard Reference Material Accuracy	Organics Percent Difference less than 30% from a range of certified values on average. Analyte concentration must be greater than five times the Method Detection Limit (>5xMDL). Organics Results in the Target is less than 5 times the MDL	N n	Review with Project Manager; re-analyze or justify reporting results in the project records.
Analytical Duplicate Precision	Organics results less than 30% Relative Percent Difference (RPD). Analyte concentration must be > 5x MDL. Organics Results in the Original is less than 5 times the MDL	N n	Review with Project Manager; re-analyze or justify reporting results in the project records.



It can be done

## WORK/QUALITY ASSURANCE PROJECT PLAN

### Attachment 3: Method Quality Objectives

<b>MQO Application</b>	<i>Universal_LC</i>		
<b>MQO:</b>	<b>Acceptance Criteria</b>	<b>Qual:</b>	<b>Corrective Action:</b>
Analytical Triplicate Precision	Organics results less than 30% Relative Standard Deviation (RSD). Analyte concentration must be > 5x MDL.  Organics Results in the Original is less than 5 times the MDL	N  n	Review with Project Manager; re-analyze or justify reporting results in the project records.
Surrogate Compound Recovery	Recovery results between 50% and 150%.	N	Review with Project Manager; re-analyze or justify reporting results in the project records.
Control Oil	RPD < 30% for at least 90% of analytes	N	Results examined by project manager, task leader, or subcontractor lab manager. Reextraction, reanalysis, or justification documented.
Instrument Calibration	5-371-3: R-squared greater than or equal to 0.995 Mean RSD less than or equal to 15%, Individual RSD less than or equal to 25%	N	Results examined by project manager, task leader, or subcontractor lab manager. Reextraction, reanalysis, or justification documented.
Independent Calibration Check Solution	5-371-3: Individual PD less than or equal to 20%. Mean Percent Difference less than or equal to 15%.	N	Review with Project Manager; re-analyze or justify in project records.
Continuing Calibration Verification	5-371-3: Individual PD less than or equal to 25%. Mean Percent Difference less than or equal to 20%.	N	Review with Project Manager; re-analyze or justify in project records.



It can be done

Battelle Project No:

**Sample Receipt Form**Approved:  Authorized 

Project Number: 112G08005-WE04

Client: Tetrattech

Received by: Schumitz, Matt

Date/Time Received: Tuesday, May 08, 2018 10:30 AM

No. of Shipping Containers: 2

**SHIPMENT**

Method of Delivery: Commercial Carrier

Tracking Number: Fed Ex

COC Forms:  Shipped with samples  No Forms**Cooler(s)/Box(es)**

Cntr	Type	Tracking No.	Seal	Seal	Container	Therm.	Temp C	Smps
1 of 2	Cooler	7721 6462 7174	Custody Seal:	Intact	Intact	Therm_1	0.9	12
2 of 2	Cooler	7721 6462 7152	Custody Seal:	Intact	Intact	Therm_1	1.2	11

**Samples**

Sample Labels:

- Sample labels agree with COC forms  
 Discrepancies (see Sample Custody Corrective Action Form)

Container Seals:

- Tape  Custody Seals  Other Seals (See sample Log)  
 Seals intact for each shipping container  
 Seals broken (See sample log for impacted samples)

Condition of Samples:

- Sample containers intact  
 Sample containers broken/leaking (See Custody Corrective Action Form)

Temperature upon receipt (°C): 1.2 Temperature Blank used  Yes  No*(Note: If temperature upon receipt differs from required conditions, see sample log comment field)*

Samples Acidified:

- Yes  No  Unknown

Initial pH 5-9?:

- Yes  No  NA

*If no, individual sample adjustments on the Auxiliary Sample Receipt Form*Total Residual Chlorine Present?:  Yes  No  NA*If yes, individual sample adjustments on the Auxiliary Sample Receipt Form*Head Space <1% in samples for water VOC analysis:  Yes  No  NA*Individual sample deviations noted on sample log*

Samples Containers:

Samples returned in PC-grade jars:  Yes  No  Unknown /Lot No.: Unknown

Storage Location: Custody: Refrigerator - R0118 (NA)

BDO IDs Assigned: J6148 - J6170

Samples logged in by: Schumitz, Matt

Date/Time: 05/08/2018 10:30 AM

Approved By:

Approved On:

Authorized By:

Authorized On:



It can be done

ShpNo SHP-180508-02

Battelle Project No:

Sample Receipt Form Details

Approved:  Authorized

Project Number: 112G08005-WE04 Client: Tetrattech

Received by: Schumitz, Matt Date/Time Received: Tuesday, May 08, 2018 10:30 AM

No. of Shipping Containers: 2

BDO Id:	Client Sample ID:	Collection Date:	Login Date:	Ctrs:	Matrix:	Temp:	pH:	TRC:	VOC:	Stored In:	Loc:	No:	Comments:
J6148	NAWC-050718-RW-316	05/07/18 10:10	05/08/18 12:33	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6149	NAWC-050718-FRB-316	05/07/18 10:05	05/08/18 12:34	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6150	NAWC-050718-RW-180	05/07/18 10:40	05/08/18 12:34	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6151	NAWC-050718-FRB-180	05/07/18 10:35	05/08/18 12:34	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6152	NAWC-050718-RW-275	05/07/18 11:10	05/08/18 12:34	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6153	NAWC-050718-FRB-275	05/07/18 11:05	05/08/18 12:35	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6154	NAWC-050718-RW-145	05/07/18 12:40	05/08/18 12:35	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6155	NAWC-050718-FRB-145	05/07/18 12:35	05/08/18 12:36	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6156	NAWC-050718-RW-357	05/07/18 13:10	05/08/18 12:36	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6157	NAWC-050718-FRB-357	05/07/18 13:05	05/08/18 12:36	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6158	NAWC-050718-RW-162	05/07/18 13:40	05/08/18 12:37	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6159	NAWC-050718-FRB-162	05/07/18 13:35	05/08/18 12:37	2	DW	0.9	NA	NA	NA	R0118 (NA)			
J6160	WGNA-050718-RW-0800	05/07/18 14:10	05/08/18 12:38	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6161	WGNA-050718-FRB-0800	05/07/18 14:05	05/08/18 12:38	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6162	WGNA-050718-RW-0335	05/07/18 14:40	05/08/18 12:38	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6163	WGNA-050718-FRB-0335	05/07/18 14:35	05/08/18 12:39	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6164	WGNA-050718-RW-3556	05/07/18 15:10	05/08/18 12:39	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6165	WGNA-050718-FRB-3556	05/07/18 15:05	05/08/18 12:40	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6166	NAWC-050718-RW-356	05/07/18 15:40	05/08/18 12:40	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6167	NAWC-050718-FRB-356	05/07/18 15:35	05/08/18 12:41	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6168	NAWC-050718-RW-289	05/07/18 16:10	05/08/18 12:41	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6169	NAWC-050718-FRB-289	05/07/18 16:05	05/08/18 12:41	2	DW	1.2	NA	NA	NA	R0118 (NA)			
J6170	WGNA-050718-DUP-35	05/07/18 7:00	05/08/18 12:42	2	DW	1.2	NA	NA	NA	R0118 (NA)			

Total Samples: 23

# Battelle

The Business of Innovation

## Chain-of-Custody

<b>Client Contact Information</b> Andy Frebowitz 234 Mall Boulevard, Suite 260 King of Prussia, PA 19406 610-382-1170		<b>Project Manager: Jonathan Thorn</b> Sampler Information (print name): Mary Kay Bond Phone: 610-382-1169 Email: mary.bond@tetrattech.com		Sampling Site: WE04		Site Information: NAS JRB Willow Grove/NAWC Warminster	
Project Name: WE04 Project No.: 112G08005-WE04		Turnaround Time (TAT) Requested: 21 days Normal <input checked="" type="checkbox"/> Priority <input type="checkbox"/> RUSH <input type="checkbox"/>		Preservation Trizma		COC #	
Time Zone: Eastern		Analysis PFAS EPA 537 14 analytes		Page# 1 of 1			
Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.		
NAWC-050718-RW-316 J6148	5/7/2018	10:10	G	DW	2	X	
NAWC-050718-FRB-316 J6149	5/7/2018	10:05	G	DW	2	X	
NAWC-050718-RW-180 J6150	5/7/2018	10:40	G	DW	2	X	Field Reagent Blank
NAWC-050718-FRB-180 J6151	5/7/2018	10:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-275 J6152	5/7/2018	11:10	G	DW	2	X	
NAWC-050718-FRB-275 J6153	5/7/2018	11:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-145 J6154	5/7/2018	12:40	G	DW	2	X	
NAWC-050718-FRB-145 J6155	5/7/2018	12:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-357 J6156	5/7/2018	13:10	G	DW	2	X	
NAWC-050718-FRB-357 J6157	5/7/2018	13:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-162 J6158	5/7/2018	13:40	G	DW	2	X	
NAWC-050718-FRB-162 J6159	5/7/2018	13:35	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-0800 J6160	5/7/2018	14:10	G	DW	2	X	
WGNA-050718-FRB-0800 J6161	5/7/2018	14:05	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-0335 J6162	5/7/2018	14:40	G	DW	2	X	
WGNA-050718-FRB-0335 J6163	5/7/2018	14:35	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-3556 J6164	5/7/2018	15:10	G	DW	2	X	
WGNA-050718-FRB-3556 J6165	5/7/2018	15:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-356 J6166	5/7/2018	15:40	G	DW	2	X	
NAWC-050718-FRB-356 J6167	5/7/2018	15:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-289 J6168	5/7/2018	16:10	G	DW	2	X	
NAWC-050718-FRB-289 J6169	5/7/2018	16:05	G	DW	2	X	Field Reagent Blank
WGNA-050718-DUP-35 J6170	5/7/2018	7:00	G	DW	2	X	Duplicate
Receipt Temperature: (°C) 0.9, 1.2		Samples Intact: Yes No		Samples on Ice: Yes No		Receipt Comments:	
Relinquished by (Print/Sign): Mary Kay Bond		Company: Tetra Tech Date/Time: 05/07/2018 16:00		Received by (Print/Sign): Matt Schwandt		Company: Battelle Date/Time: 5/8/18 1030	
Relinquished by (Print/Sign):		Company:		Received by (Print/Sign):		Company:	
Relinquished by (Print/Sign):		Company:		Received by (Print/Sign):		Company:	
Comments: FedEx Tracking #7721 6462 7152							

FedEx  
MPS# 7721 6462 7174  
0263

TUE - 08 MAY 10:30A  
PRIORITY OVERNIGHT

EM XPUA 0.9

02061  
MA-US BOS

EXP 01/19



#112573 05/07 552J2/782B/DCA5

Therm-1  
MS  
S/B  
8/18  
10:30

RT 246  
ST 2.2  
1 10:30  
B 7174  
05.08


FedEx  
TRK# 7721 6462 7152  
0201

TUE - 08 MAY 10:30A  
PRIORITY OVERNIGHT

EM XPUA 1.2

02061  
MA-US BOS

EXP 01/19



#112573 05/07 552J2/782B/DCA5

Therm-1  
MS  
S/B  
1030

RT 246  
ST 2.2  
1 10:30  
B 7152  
05.08

# Data Tables



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-316				
Battelle ID	J6148-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.70	0.19	0.44	2.19	
PFHpA	7.04	0.30	0.88	2.19	
PFOA	16.15	0.33	0.88	2.19	
PFNA	2.26	0.32	0.88	2.19	
PFDA	0.88 U	0.34	0.88	2.19	
PFUnA	0.88 U	0.33	0.88	2.19	
PFDaA	0.88 U	0.37	0.88	2.19	
PFTTrDA	0.88 U	0.37	0.88	2.19	
PFTeDA	1.32 U	0.64	1.32	2.19	
NMeFOSAA	0.88 U	0.37	0.88	2.19	
NEtFOSAA	0.88 U	0.39	0.88	2.19	
PFBS	6.83	0.18	0.44	2.19	
PFHxS	25.12	0.30	0.88	2.19	
PFOS	32.92	0.26	0.88	2.19	

**Surrogate Recoveries (%)**

13C2-PFHxA	104
13C2-PFDA	87
d5-EtFOSAA	90



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-180				
Battelle ID	J6150-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.93	0.20	0.45	2.23	
PFHpA	8.88	0.30	0.89	2.23	
PFOA	19.77	0.34	0.89	2.23	
PFNA	3.33	0.33	0.89	2.23	
PFDA	0.37 J	0.35	0.89	2.23	
PFUnA	0.89 U	0.34	0.89	2.23	
PFDaA	0.89 U	0.38	0.89	2.23	
PFTTrDA	0.89 U	0.38	0.89	2.23	
PFTeDA	1.34 U	0.65	1.34	2.23	
NMeFOSAA	0.89 U	0.38	0.89	2.23	
NEtFOSAA	0.89 U	0.39	0.89	2.23	
PFBS	14.92	0.19	0.45	2.23	
PFHxS	12.68	0.30	0.89	2.23	
PFOS	28.97	0.27	0.89	2.23	

**Surrogate Recoveries (%)**

13C2-PFHxA	128
13C2-PFDA	101
d5-EtFOSAA	98



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-275				
Battelle ID	J6152-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.14	0.20	0.45	2.23	
PFHpA	5.44	0.30	0.89	2.23	
PFOA	15.09	0.34	0.89	2.23	
PFNA	2.85	0.33	0.89	2.23	
PFDA	1.00 J	0.35	0.89	2.23	
PFUnA	0.89 U	0.34	0.89	2.23	
PFDaA	0.89 U	0.38	0.89	2.23	
PFTTrDA	0.89 U	0.38	0.89	2.23	
PFTeDA	1.34 U	0.65	1.34	2.23	
NMeFOSAA	0.89 U	0.38	0.89	2.23	
NEtFOSAA	0.89 U	0.39	0.89	2.23	
PFBS	12.85	0.19	0.45	2.23	
PFHxS	7.69	0.30	0.89	2.23	
PFOS	19.02	0.27	0.89	2.23	

**Surrogate Recoveries (%)**

13C2-PFHxA	138 N
13C2-PFDA	114
d5-EtFOSAA	113





Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-145			
Battelle ID	J6154-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	9.19	0.20	0.45	2.23
PFHpA	6.54	0.30	0.89	2.23
PFOA	18.64	0.34	0.89	2.23
PFNA	2.88	0.33	0.89	2.23
PFDA	0.57 J	0.35	0.89	2.23
PFUnA	0.89 U	0.34	0.89	2.23
PFDaA	0.89 U	0.38	0.89	2.23
PFTTrDA	0.89 U	0.38	0.89	2.23
PFTeDA	1.34 U	0.65	1.34	2.23
NMeFOSAA	0.89 U	0.38	0.89	2.23
NEtFOSAA	0.89 U	0.39	0.89	2.23
PFBS	10.27	0.19	0.45	2.23
PFHxS	8.56	0.30	0.89	2.23
PFOS	17.04	0.27	0.89	2.23

**Surrogate Recoveries (%)**

13C2-PFHxA	128
13C2-PFDA	96
d5-EtFOSAA	98



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-357				
Battelle ID	J6156-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.06	0.20	0.46	2.31	
PFHpA	4.28	0.31	0.93	2.31	
PFOA	11.61	0.35	0.93	2.31	
PFNA	1.31 J	0.34	0.93	2.31	
PFDA	0.42 J	0.36	0.93	2.31	
PFUnA	0.93 U	0.35	0.93	2.31	
PFDaA	0.93 U	0.39	0.93	2.31	
PFTTrDA	0.93 U	0.39	0.93	2.31	
PFTeDA	1.39 U	0.68	1.39	2.31	
NMeFOSAA	0.93 U	0.39	0.93	2.31	
NEtFOSAA	0.93 U	0.41	0.93	2.31	
PFBS	7.50	0.19	0.46	2.31	
PFHxS	12.06	0.31	0.93	2.31	
PFOS	9.78	0.28	0.93	2.31	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	137 N				
13C2-PFDA	108				
d5-EtFOSAA	80				



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-162				
Battelle ID	J6158-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	6.78	0.20	0.45	2.23	
PFHpA	3.94	0.30	0.89	2.23	
PFOA	12.87	0.34	0.89	2.23	
PFNA	1.83 J	0.33	0.89	2.23	
PFDA	0.89 U	0.35	0.89	2.23	
PFUnA	0.89 U	0.34	0.89	2.23	
PFDaA	0.89 U	0.38	0.89	2.23	
PFTTrDA	0.89 U	0.38	0.89	2.23	
PFTeDA	1.34 U	0.65	1.34	2.23	
NMeFOSAA	0.89 U	0.38	0.89	2.23	
NEtFOSAA	0.89 U	0.39	0.89	2.23	
PFBS	9.10	0.19	0.45	2.23	
PFHxS	13.69	0.30	0.89	2.23	
PFOS	30.68	0.27	0.89	2.23	

**Surrogate Recoveries (%)**

13C2-PFHxA	132 N
13C2-PFDA	95
d5-EtFOSAA	90



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID WGNA-050718-RW-0800

Battelle ID J6160-FS  
 Sample Type SA  
 Collection Date 05/07/2018  
 Extraction Date 05/10/2018  
 Analysis Date 05/14/2018  
 Analytical Instrument Sciex 5500 LC/MS/MS  
 % Moisture NA  
 Matrix DW  
 Sample Size 0.285  
 Size Unit-Basis L  
 Units ng/L MDL LOD LOQ

	ng/L	MDL	LOD	LOQ
PFHxA	8.53	0.19	0.44	2.19
PFHpA	5.38	0.30	0.88	2.19
PFOA	14.73	0.33	0.88	2.19
PFNA	1.89 J	0.32	0.88	2.19
PFDA	0.45 J	0.34	0.88	2.19
PFUnA	0.88 U	0.33	0.88	2.19
PFDaA	0.88 U	0.37	0.88	2.19
PFTTrDA	0.88 U	0.37	0.88	2.19
PFTeDA	1.32 U	0.64	1.32	2.19
NMeFOSAA	0.88 U	0.37	0.88	2.19
NEtFOSAA	0.88 U	0.39	0.88	2.19
PFBS	23.13	0.18	0.44	2.19
PFHxS	9.29	0.30	0.88	2.19
PFOS	19.76	0.26	0.88	2.19

**Surrogate Recoveries (%)**

13C2-PFHxA	121
13C2-PFDA	107
d5-EtFOSAA	97



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID WGNA-050718-RW-0335

Battelle ID	J6162-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.270			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	12.83	0.20	0.46	2.31
PFHpA	7.28	0.31	0.93	2.31
PFOA	22.49	0.35	0.93	2.31
PFNA	2.80	0.34	0.93	2.31
PFDA	0.93 U	0.36	0.93	2.31
PFUnA	0.93 U	0.35	0.93	2.31
PFDaA	0.93 U	0.39	0.93	2.31
PFTTrDA	0.93 U	0.39	0.93	2.31
PFTeDA	1.39 U	0.68	1.39	2.31
NMeFOSAA	0.93 U	0.39	0.93	2.31
NEtFOSAA	0.93 U	0.41	0.93	2.31
PFBS	11.74	0.19	0.46	2.31
PFHxS	7.86	0.31	0.93	2.31
PFOS	19.22	0.28	0.93	2.31

**Surrogate Recoveries (%)**

13C2-PFHxA	126
13C2-PFDA	92
d5-EtFOSAA	89



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID WGNA-050718-RW-3556

Battelle ID	J6164-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.260			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	5.69	0.21	0.48	2.40
PFHpA	3.27	0.33	0.96	2.40
PFOA	7.77	0.37	0.96	2.40
PFNA	1.26 J	0.36	0.96	2.40
PFDA	0.96 U	0.38	0.96	2.40
PFUnA	0.96 U	0.37	0.96	2.40
PFDaA	0.96 U	0.40	0.96	2.40
PFTTrDA	0.96 U	0.40	0.96	2.40
PFTeDA	1.44 U	0.70	1.44	2.40
NMeFOSAA	0.96 U	0.40	0.96	2.40
NEtFOSAA	0.96 U	0.42	0.96	2.40
PFBS	8.69	0.20	0.48	2.40
PFHxS	2.38 J	0.33	0.96	2.40
PFOS	5.94	0.29	0.96	2.40

**Surrogate Recoveries (%)**

13C2-PFHxA	117
13C2-PFDA	83
d5-EtFOSAA	91



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-356			
Battelle ID	J6166-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.275			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	4.26	0.20	0.45	2.27
PFHpA	2.74	0.31	0.91	2.27
PFOA	9.45	0.35	0.91	2.27
PFNA	1.06 J	0.34	0.91	2.27
PFDA	0.91 U	0.35	0.91	2.27
PFUnA	0.91 U	0.35	0.91	2.27
PFDaA	0.91 U	0.38	0.91	2.27
PFTTrDA	0.91 U	0.38	0.91	2.27
PFTeDA	1.36 U	0.66	1.36	2.27
NMeFOSAA	0.91 U	0.38	0.91	2.27
NEtFOSAA	0.91 U	0.40	0.91	2.27
PFBS	3.18	0.19	0.45	2.27
PFHxS	3.92	0.31	0.91	2.27
PFOS	7.94	0.27	0.91	2.27
<b>Surrogate Recoveries (%)</b>				
13C2-PFHxA	132	N		
13C2-PFDA	109			
d5-EtFOSAA	105			



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-RW-289				
Battelle ID	J6168-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	2.11 J	0.20	0.45	2.23	
PFHpA	1.50 J	0.30	0.89	2.23	
PFOA	5.47	0.34	0.89	2.23	
PFNA	0.94 J	0.33	0.89	2.23	
PFDA	0.89 U	0.35	0.89	2.23	
PFUnA	0.89 U	0.34	0.89	2.23	
PFDaA	0.89 U	0.38	0.89	2.23	
PFTTrDA	0.89 U	0.38	0.89	2.23	
PFTeDA	1.34 U	0.65	1.34	2.23	
NMeFOSAA	0.89 U	0.38	0.89	2.23	
NEtFOSAA	0.89 U	0.39	0.89	2.23	
PFBS	4.15	0.19	0.45	2.23	
PFHxS	3.75	0.30	0.89	2.23	
PFOS	10.99	0.27	0.89	2.23	

**Surrogate Recoveries (%)**

13C2-PFHxA	112
13C2-PFDA	94
d5-EtFOSAA	93





Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	WGNA-050718-DUP-35				
Battelle ID	J6170-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	9.11	0.20	0.45	2.23	
PFHpA	5.07	0.30	0.89	2.23	
PFOA	13.21	0.34	0.89	2.23	
PFNA	2.59	0.33	0.89	2.23	
PFDA	0.91 J	0.35	0.89	2.23	
PFUnA	0.89 U	0.34	0.89	2.23	
PFDaA	0.89 U	0.38	0.89	2.23	
PFTTrDA	0.89 U	0.38	0.89	2.23	
PFTeDA	1.34 U	0.65	1.34	2.23	
NMeFOSAA	0.89 U	0.38	0.89	2.23	
NEtFOSAA	0.89 U	0.39	0.89	2.23	
PFBS	12.58	0.19	0.45	2.23	
PFHxS	7.85	0.30	0.89	2.23	
PFOS	19.67	0.27	0.89	2.23	

**Surrogate Recoveries (%)**

13C2-PFHxA	134 N
13C2-PFDA	105
d5-EtFOSAA	110



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	Procedural Blank			
Battelle ID	CQ755PB-FS			
Sample Type	PB			
Collection Date	05/10/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	WATER			
Sample Size	0.250			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	0.50 U	0.22	0.50	2.50
PFHpA	1.00 U	0.34	1.00	2.50
PFOA	1.00 U	0.38	1.00	2.50
PFNA	1.00 U	0.37	1.00	2.50
PFDA	1.00 U	0.39	1.00	2.50
PFUnA	1.00 U	0.38	1.00	2.50
PFDaA	1.00 U	0.42	1.00	2.50
PFTTrDA	1.00 U	0.42	1.00	2.50
PFTeDA	1.50 U	0.73	1.50	2.50
NMeFOSAA	1.00 U	0.42	1.00	2.50
NEtFOSAA	1.00 U	0.44	1.00	2.50
PFBS	0.50 U	0.21	0.50	2.50
PFHxS	1.00 U	0.34	1.00	2.50
PFOS	1.00 U	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	107
13C2-PFDA	108
d5-EtFOSAA	91



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	Laboratory Control Sample					
Battelle ID	CQ756LCS-FS					
Sample Type	LCS					
Collection Date	05/10/2018					
Extraction Date	05/10/2018					
Analysis Date	05/14/2018					
Analytical Instrument	Sciex 5500 LC/MS/MS					
% Moisture	NA					
Matrix	WATER					
Sample Size	0.250					
Size Unit-Basis	L					
Units	ng/L	Target	Recovery	Qual	Control Limits	
					Lower	Upper
PFHxA	16.05	15.00	107		70	130
PFHpA	14.89	15.00	99		70	130
PFOA	15.17	15.00	101		70	130
PFNA	15.73	15.00	105		70	130
PFDA	15.51	15.00	103		70	130
PFUnA	14.75	15.00	98		70	130
PFDoA	14.90	15.00	99		70	130
PFTTrDA	14.59	15.00	97		70	130
PFTeDA	16.76	15.00	112		70	130
NMeFOSAA	17.70	15.00	118		70	130
NEtFOSAA	16.48	15.00	110		70	130
PFBS	14.28	13.28	108		70	130
PFHxS	15.14	14.18	107		70	130
PFOS	13.81	14.33	96		70	130

**Surrogate Recoveries (%)**

13C2-PFHxA	105
13C2-PFDA	104
d5-EtFOSAA	95



## Glossary of Data Qualifiers

Flag: Application:

---

B	Analyte found in the sample at a concentration <10x the level found in the procedural blank
D	Dilution Run. Initial run outside the initial calibration range of the instrument
E	Estimate, result is greater than the highest concentration level in the calibration
H	Surrogate diluted out. Used when surrogate recovery is affected by excessive dilution of the sample extract.
J	Analyte detected below the Limit of Quantitation (LOQ)
ME	Significant Matrix Interference - Estimated value.
MI	Significant Matrix Interference - value could not be determined.
n	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO), but meets secondary criteria
N	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO)
NA	Not Applicable
T	Holding Time (HT) exceeded
U	Analyte not detected or detected below the Method detection limit (MDL) value, Limit of Detection (LOD) reported

# Miscellaneous Documentation



Norwell Operations  
 141 Longwater Drive, Suite 202  
 Norwell, Massachusetts 02061  
 Telephone: 781-681-5400

July 13<sup>th</sup>, 2018

This data package has been revised to include the following updates to the reporting format:

- Use of LOD values for non-detected values (in place of the MDL value that was used in the original report).
- Use of sample specific MDL, LOD, and LOQ values (adjusted for dilution and sample size variations as compared to the MDL, LOD, and LOQ studies)

In addition to non-detect (“U” qualified) data changing to use the sample specific LOD value (not included in the table below), the information in the following table changed from the original report to the new report. The reason for these changes is the variation in sample size for individual samples when using sample specific values. This table includes information on all SDG updated and resubmitted on 7/13/2018.

SDG	Lab Sample ID	Client ID	Analyte	New Result	New Qual	Old Result	Old Qual
18-0299	J5972-FS	WGNA-050118-RW-3178	PFHpA	2.25		2.25	J
18-0313	J6148-FS	NAWC-050718-RW-316	PFNA	2.26		2.26	J
18-0313	J6150-FS	NAWC-050718-RW-180	PFDA	0.37	J	0.39	U
18-0323	J6209-FS	NAWC-051018-FRB-177	PFOA	0.38	J	0.38	U
18-0343	J6264-FS	WGNA-052918-RW-3978	PFNA	2.34		2.34	J
18-0343	J6273-FS	NAWC-053018-RW-231	PFHxS	37.20	JD	37.20	D
18-0343	J6275-FS	WGNA-053018-RW-3933	PFNA	2.35		2.35	J
18-0343	J6285-FS	NAWC-053018-RW-196	PFHxS	2.31		2.31	J
18-0360	J6583-FS	NAWC-060418-FRB-230	PFHxS	0.33	J	1.00	U
18-0360	J6643-FS	WGNA-060718-FRB-0626	PFOS	2.48	B	2.48	J

The original data tables have been moved to the unused data section of this complete data package. For SDG 18-0313, the original MQO report and case narrative were moved to the unused data section of the full data package.

## QA/QC Summary Batch 18-0313

Project:	CTO-WE04 Naval Air Station Joint Reserve Base Willow Grove
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	DW
Data Set:	DP-18-0112
Analytical SOP:	5-371
Method Reference:	USEPA 537 rev. 1.1, QSM 5.1

### Sample Custody

Collection Date	Receipt Date	Temp (°C)
5/7/2018	5/8/2018	0.9, 1.2

Corrective Actions	None
Sample Storage	The water samples were stored refrigerated until extraction.
Related samples	FRB samples associated with these samples are extracted in SDG 18-0316

### METHOD SUMMARIES

Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with methanol. Extracts were split and concentrated to dryness under nitrogen with a water bath set between 60 °C and 65 °C, reconstituted with 96:4 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.
Prep comments	None.
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.
Analysis Comments	Samples analyzed on the Sciex 5500. The confirmation ion ratio was above 50% RPD for the following samples and analytes: NAWC-050718-RW-316 for PFOA and PFDA NAWC-050718-RW-357 for PFOA NAWC-050718-RW-162 for PFDA WGNA-050718-RW-0335 for PFOA WGNA-050718-RW-3556 for PFHpA and PFOA NAWC -050718-RW-356 for PFOA and PFDA NAWC -050718-RW-289 for PFOA

Holding Times	Extraction Date(s)	Analysis Date(s)
	5/10/2018	5/14/2018

**QA/QC Summary**  
**Batch 18-0313**

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
$\leq 1/3$ the MRL	No exceedances noted. No comments.
Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
70-130% of true value	No exceedances noted. No comments.
Matrix Spike (MS) / Duplicate (MSD)	A MS/MSD were prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy. The relative percent difference was calculated to measure precision.
70-130% of true value, RPD $\leq 30\%$	No exceedances noted. MS/MSD samples were not collected for this SDG.
Surrogates Standard Analytes	Labelled surrogate compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
70-130% of true value	5 exceedances noted. 13C2-PFHxA recovered above 130% in five field samples (recoveries are less than 138% in all cases, DoD requirement is 50-150% of true value).
Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
ICAL high and low points RPD $\leq 20\%$ , 50-150% of average area of the ICAL and 70-140% of most recent CCV	No exceedances noted. No Comments.
Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadratic curve fitting.
R <sup>2</sup> >0.99	No exceedances noted.
Target and SIS compounds +/- 30% of true value, Low point 50-150% of true value	No comments.



**QA/QC Summary**  
**Batch 18-0313**

Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
Target and SIS compounds +/- 30% of true value	No exceedances noted.
	No comments.
Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.
Target and SIS compounds +/- 30% of true value Low point 50-150% of true value	No exceedances noted.
	No comments.



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Number: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0112  
 Test Code: Master\_371

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	0	None
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	N/A	None
Matrix Spike / Matrix Spike Duplicate Precision	N/A	None
Extracted Internal Standard Analytes (Surrogates)	5	Surrogate 13C2-PFHxA recovered above the recommended limit in five of the authentic samples. Integrations and prep records were verified.
Instrument Calibration	0	None
Instrument Blank	0	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None



# BATTELLE - NORWELL OPERATIONS MISCELLANEOUS DOCUMENTATION FORM

**Project Title:** Naval Air Station Joint Reserve Base Wi      **Data Set Number:** DP-18-0112  
**Project Number:** 100117920-WE04      **Prep Batch Number:** 18-0313  
**Entered By:** Denise Schumitz      **Entered On:** 05/17/2018  
**Test Code (Matrix Type):** Master\_371(L)

Samples that were manually integrated are noted on the quant reports with the comment (TRUE).  
DMS 6/21/2018

JV64 is not being used in the calibration curve for PFHxA, PFOA, PFDA, and PFTeDA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/21/2018

JV65 is not being used in the calibration curve for PFHxA and PFDA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/21/2018

JV64 has an ion ratio <50% for NMeFOSAA and PFUnA.  
DMS 6/21/2018

Sample J6164 has an ion ratio <50% for PFHpA.  
DMS 6/21/2018

Samples J6148, J6156, J6162, J6164, J6166 and J6168 have an ion ratio <50% for PFOA.  
DMS 6/21/2018

Sample J6148, J6158 and J6166 have an ion ratio <50% for PFDA.  
DMS 6/21/2018

**Task Leader Approval:**

**Supervisor Approval:**

Digitally signed by Jonathan  
Thorn  
Date: 2018.05.27 12:05:59 -04'00'

**PM Approval:**



## Example Calculation for PFAS

Calculation of final concentration from area:

$$\text{Concentration} = \left[ \frac{PA - b}{m} \right] * C_{IS} * PIV * DF / S$$

Where:

PA = Area of target / area of internal standard  
 b = y intercept from calibration curve  
 CIS = concentration of internal standard (ng/L)  
 m = slope of calibration  
 DF = dilution factor  
 S = Sample Size  
 PIV = Pre-injection volume (L)

Sample ID: J6154-FS(0)  
 Client Sample ID: NAWC-050718-RW-145  
 Sample Size: 0.28  
 Units: L  
 Dilution Factor: 1  
 PIV (L): 0.001  
 Target Analyte: PFHxA  
 MRM Transition: 313.0 / 269.0  
 Data file: 18-0313.wiff  
 Result table: 18-0313\_Base  
 Area: 773618  
 IS Name: 13C2-PFOA  
 IS Area: 33518.97  
 IS Amount (ng/L): 100  
 y-intercept 0.46617  
 slope 0.87836

$$\text{Concentration} = \frac{[(773617.98/33518.97) - 0.46617]}{0.87836} * 100 * 0.001 * 1 / 0.28$$

$$\text{ng/L} = 9.19$$



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0112

	CQ755PB-FS (Procedural Blank)	CQ756LCS-FS (Laboratory Control Sample)	J6148-FS (NAWC-050718-RW-316)	J6150-FS (NAWC-050718-RW-180)	J6152-FS (NAWC-050718-RW-275)	J6154-FS (NAWC-050718-RW-145)
PFHxA	-	L	L	L	L	L
PFHpA	-	L	L	L	L	L
PFOA	-	L	L	L	L	L
PFNA	-	L	L	L	L	L
PFDA	-	L	-	-	L	L
PFUnA	-	L	-	-	-	-
PFDoA	-	L	-	-	-	-
PFTTrDA	-	L	-	-	-	-
PFTeDA	-	L	-	-	-	-
NMeFOSAA	-	L	-	-	-	-
NEtFOSAA	-	L	-	-	-	-
PFBS	-	L	L	L	L	L
PFHxS	-	L	L	L	L	L
PFOS	-	L	L/Br	L/Br	L/Br	L/Br

"L": Linear

"Br": branched

"L/Br": Linear/Branched

"-": Not detected



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Preparation Batch: 18-0313

Data Set: DP-18-0:

	J6156-FS (NAWC-050718-RW-357)	J6158-FS (NAWC-050718-RW-162)	J6160-FS (WGNNA-050718-RW-0800)	J6162-FS (WGNNA-050718-RW-0335)	J6164-FS (WGNNA-050718-RW-3556)	J6166-FS (NAWC-050718-RW-356)	J6168-FS (NAWC-050718-RW-289)	J6170-FS (WGNNA-050718-DUP-35)
PFHxA	L	L	L	L	L	L	L	L
PFHpA	L	L	L	L	L	L	L	L
PFOA	L	L	L	L	L	L	L	L
PFNA	L	L	L	L	L	L	L	L
PFDA	L	L	L	L	L	L	L	L
PFUnA	-	-	-	-	-	-	-	-
PFDoA	-	-	-	-	-	-	-	-
PFTTrDA	-	-	-	-	-	-	-	-
PFTeDA	-	-	-	-	-	-	-	-
NMeFOSAA	-	-	-	-	-	-	-	-
NEtFOSAA	-	-	-	-	-	-	-	-
PFBS	L	L	L	L	L	L	L	L
PFHxS	L	L	L	L	L	L	L	L
PFOS	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br

"L": Linear

"Br": branched

"L/Br": Linear/Branched

"-": Not detected



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	13C2-PFOA	34,698.30	-
JV65	L2	5/14/18 11:38	13C2-PFOA	33,573.03	-
JV66	L3	5/14/18 11:47	13C2-PFOA	29,649.37	-
JV67	L4	5/14/18 11:56	13C2-PFOA	41,395.56	-
JV68	L5	5/14/18 12:05	13C2-PFOA	30,280.34	-
JV69	L6	5/14/18 12:14	13C2-PFOA	37,534.68	-
JV70	L7	5/14/18 12:23	13C2-PFOA	29,423.02	-
JV71	L8	5/14/18 12:32	13C2-PFOA	39,891.91	-
JV72	L9	5/14/18 12:41	13C2-PFOA	40,992.20	16.6

PASS

Average 35,270.93 Lower 17,635.47 Upper 52,906.40

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	13C2-PFOA	34,698.30	17,635.47	52,906.40		21,196.24	42,392.48	
JV65	L2	5/14/18 11:38	13C2-PFOA	33,573.03	17,635.47	52,906.40		21,196.24	42,392.48	
JV66	L3	5/14/18 11:47	13C2-PFOA	29,649.37	17,635.47	52,906.40		21,196.24	42,392.48	
JV67	L4	5/14/18 11:56	13C2-PFOA	41,395.56	17,635.47	52,906.40		21,196.24	42,392.48	
JV68	L5	5/14/18 12:05	13C2-PFOA	30,280.34	17,635.47	52,906.40		21,196.24	42,392.48	
JV69	L6	5/14/18 12:14	13C2-PFOA	37,534.68	17,635.47	52,906.40		21,196.24	42,392.48	
JV70	L7	5/14/18 12:23	13C2-PFOA	29,423.02	17,635.47	52,906.40		21,196.24	42,392.48	
JV71	L8	5/14/18 12:32	13C2-PFOA	39,891.91	17,635.47	52,906.40		21,196.24	42,392.48	
JV72	L9	5/14/18 12:41	13C2-PFOA	40,992.20	17,635.47	52,906.40		21,196.24	42,392.48	
JV63 ICC	ICC	5/14/18 12:50	13C2-PFOA	37,388.34	17,635.47	52,906.40		21,196.24	42,392.48	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	13C2-PFOA	35,848.65	17,635.47	52,906.40		21,196.24	42,392.48	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	13C2-PFOA	29,877.62	17,635.47	52,906.40		21,196.24	42,392.48	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	13C2-PFOA	36,858.30	17,635.47	52,906.40		21,196.24	42,392.48	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	13C2-PFOA	34,064.10	17,635.47	52,906.40		21,196.24	42,392.48	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	13C2-PFOA	24,285.52	17,635.47	52,906.40		21,196.24	42,392.48	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	13C2-PFOA	33,518.97	17,635.47	52,906.40		21,196.24	42,392.48	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	13C2-PFOA	24,248.05	17,635.47	52,906.40		21,196.24	42,392.48	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	13C2-PFOA	35,845.05	17,635.47	52,906.40		21,196.24	42,392.48	
JV69 CCV	CCV	5/14/18 14:19	13C2-PFOA	37,224.64	17,635.47	52,906.40		21,196.24	42,392.48	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	13C2-PFOA	31,443.44	17,635.47	52,906.40		26,057.25	52,114.50	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	13C2-PFOA	37,992.76	17,635.47	52,906.40		26,057.25	52,114.50	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	13C2-PFOA	30,729.65	17,635.47	52,906.40		26,057.25	52,114.50	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	13C2-PFOA	27,240.70	17,635.47	52,906.40		26,057.25	52,114.50	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	13C2-PFOA	34,960.30	17,635.47	52,906.40		26,057.25	52,114.50	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	13C2-PFOA	31,796.45	17,635.47	52,906.40		26,057.25	52,114.50	
JV68 CCV	CCV	5/14/18 15:30	13C2-PFOA	37,508.63	17,635.47	52,906.40		26,057.25	52,114.50	



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WF04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	13C4-PFOS	104,334.94	-
JV65	L2	5/14/18 11:38	13C4-PFOS	98,871.27	-
JV66	L3	5/14/18 11:47	13C4-PFOS	90,792.83	-
JV67	L4	5/14/18 11:56	13C4-PFOS	116,199.13	-
JV68	L5	5/14/18 12:05	13C4-PFOS	92,291.39	-
JV69	L6	5/14/18 12:14	13C4-PFOS	111,784.92	-
JV70	L7	5/14/18 12:23	13C4-PFOS	88,194.77	-
JV71	L8	5/14/18 12:32	13C4-PFOS	110,166.97	-
JV72	L9	5/14/18 12:41	13C4-PFOS	116,123.32	10.7

PASS

Average      Lower      Upper  
103,195.50    51,597.75    154,793.25

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	13C4-PFOS	104,334.94	51,597.75	154,793.25		64,603.97	129,207.95	
JV65	L2	5/14/18 11:38	13C4-PFOS	98,871.27	51,597.75	154,793.25		64,603.97	129,207.95	
JV66	L3	5/14/18 11:47	13C4-PFOS	90,792.83	51,597.75	154,793.25		64,603.97	129,207.95	
JV67	L4	5/14/18 11:56	13C4-PFOS	116,199.13	51,597.75	154,793.25		64,603.97	129,207.95	
JV68	L5	5/14/18 12:05	13C4-PFOS	92,291.39	51,597.75	154,793.25		64,603.97	129,207.95	
JV69	L6	5/14/18 12:14	13C4-PFOS	111,784.92	51,597.75	154,793.25		64,603.97	129,207.95	
JV70	L7	5/14/18 12:23	13C4-PFOS	88,194.77	51,597.75	154,793.25		64,603.97	129,207.95	
JV71	L8	5/14/18 12:32	13C4-PFOS	110,166.97	51,597.75	154,793.25		64,603.97	129,207.95	
JV72	L9	5/14/18 12:41	13C4-PFOS	116,123.32	51,597.75	154,793.25		64,603.97	129,207.95	
JV63 ICC	ICC	5/14/18 12:50	13C4-PFOS	112,208.04	51,597.75	154,793.25		64,603.97	129,207.95	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	13C4-PFOS	111,731.01	51,597.75	154,793.25		64,603.97	129,207.95	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	13C4-PFOS	80,170.42	51,597.75	154,793.25		64,603.97	129,207.95	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	13C4-PFOS	97,479.82	51,597.75	154,793.25		64,603.97	129,207.95	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	13C4-PFOS	97,536.34	51,597.75	154,793.25		64,603.97	129,207.95	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	13C4-PFOS	79,659.10	51,597.75	154,793.25		64,603.97	129,207.95	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	13C4-PFOS	93,168.05	51,597.75	154,793.25		64,603.97	129,207.95	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	13C4-PFOS	71,904.49	51,597.75	154,793.25		64,603.97	129,207.95	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	13C4-PFOS	100,729.16	51,597.75	154,793.25		64,603.97	129,207.95	
JV69 CCV	CCV	5/14/18 14:19	13C4-PFOS	113,429.56	51,597.75	154,793.25		64,603.97	129,207.95	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	13C4-PFOS	91,010.65	51,597.75	154,793.25		79,400.69	158,801.38	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	13C4-PFOS	106,801.49	51,597.75	154,793.25		79,400.69	158,801.38	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	13C4-PFOS	84,581.13	51,597.75	154,793.25		79,400.69	158,801.38	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	13C4-PFOS	86,631.83	51,597.75	154,793.25		79,400.69	158,801.38	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	13C4-PFOS	95,267.28	51,597.75	154,793.25		79,400.69	158,801.38	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	13C4-PFOS	92,164.56	51,597.75	154,793.25		79,400.69	158,801.38	
JV68 CCV	CCV	5/14/18 15:30	13C4-PFOS	110,468.54	51,597.75	154,793.25		79,400.69	158,801.38	





Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	d3-MeFOSAA	35,968.44	-
JV65	L2	5/14/18 11:38	d3-MeFOSAA	30,241.50	-
JV66	L3	5/14/18 11:47	d3-MeFOSAA	28,007.84	-
JV67	L4	5/14/18 11:56	d3-MeFOSAA	36,150.49	-
JV68	L5	5/14/18 12:05	d3-MeFOSAA	30,317.42	-
JV69	L6	5/14/18 12:14	d3-MeFOSAA	36,296.34	-
JV70	L7	5/14/18 12:23	d3-MeFOSAA	27,471.55	-
JV71	L8	5/14/18 12:32	d3-MeFOSAA	34,679.62	-
JV72	L9	5/14/18 12:41	d3-MeFOSAA	37,693.37	4.7

PASS

Average 32,980.73 Lower 16,490.37 Upper 49,471.10

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	d3-MeFOSAA	35,968.44	16,490.37	49,471.10		21,222.19	42,444.39	
JV65	L2	5/14/18 11:38	d3-MeFOSAA	30,241.50	16,490.37	49,471.10		21,222.19	42,444.39	
JV66	L3	5/14/18 11:47	d3-MeFOSAA	28,007.84	16,490.37	49,471.10		21,222.19	42,444.39	
JV67	L4	5/14/18 11:56	d3-MeFOSAA	36,150.49	16,490.37	49,471.10		21,222.19	42,444.39	
JV68	L5	5/14/18 12:05	d3-MeFOSAA	30,317.42	16,490.37	49,471.10		21,222.19	42,444.39	
JV69	L6	5/14/18 12:14	d3-MeFOSAA	36,296.34	16,490.37	49,471.10		21,222.19	42,444.39	
JV70	L7	5/14/18 12:23	d3-MeFOSAA	27,471.55	16,490.37	49,471.10		21,222.19	42,444.39	
JV71	L8	5/14/18 12:32	d3-MeFOSAA	34,679.62	16,490.37	49,471.10		21,222.19	42,444.39	
JV72	L9	5/14/18 12:41	d3-MeFOSAA	37,693.37	16,490.37	49,471.10		21,222.19	42,444.39	
JV63 ICC	ICC	5/14/18 12:50	d3-MeFOSAA	34,401.63	16,490.37	49,471.10		21,222.19	42,444.39	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	d3-MeFOSAA	35,000.18	16,490.37	49,471.10		21,222.19	42,444.39	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	d3-MeFOSAA	26,521.45	16,490.37	49,471.10		21,222.19	42,444.39	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	d3-MeFOSAA	28,397.36	16,490.37	49,471.10		21,222.19	42,444.39	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	d3-MeFOSAA	30,315.36	16,490.37	49,471.10		21,222.19	42,444.39	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	d3-MeFOSAA	25,531.81	16,490.37	49,471.10		21,222.19	42,444.39	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	d3-MeFOSAA	30,882.72	16,490.37	49,471.10		21,222.19	42,444.39	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	d3-MeFOSAA	25,387.66	16,490.37	49,471.10		21,222.19	42,444.39	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	d3-MeFOSAA	35,413.81	16,490.37	49,471.10		21,222.19	42,444.39	
JV69 CCV	CCV	5/14/18 14:19	d3-MeFOSAA	31,738.59	16,490.37	49,471.10		21,222.19	42,444.39	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	d3-MeFOSAA	26,555.02	16,490.37	49,471.10		22,217.01	44,434.03	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	d3-MeFOSAA	33,982.85	16,490.37	49,471.10		22,217.01	44,434.03	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	d3-MeFOSAA	24,586.57	16,490.37	49,471.10		22,217.01	44,434.03	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	d3-MeFOSAA	24,181.64	16,490.37	49,471.10		22,217.01	44,434.03	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	d3-MeFOSAA	30,500.70	16,490.37	49,471.10		22,217.01	44,434.03	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	d3-MeFOSAA	31,134.03	16,490.37	49,471.10		22,217.01	44,434.03	
JV68 CCV	CCV	5/14/18 15:30	d3-MeFOSAA	33,881.63	16,490.37	49,471.10		22,217.01	44,434.03	

<b>Sample Name</b>	JV70	<b>Injection Vial</b>	8
<b>Sample ID</b>	L7	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 12:23:30 PM	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

**Results Summary**

<b>Analyte</b>	<b>MRM Transition</b>	<b>RT</b>	<b>Asymmetry Factor</b>	<b>Passing Range</b>
PFBS_1	298.9 / 80.0	1.47	1.48	0.8 – 1.5
PFHxA_1	313.0 / 269.0	1.75	1.34	0.8 – 1.5

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/14/2018 12:23:30 PM	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
PFBS_1	298.9 / 80.0	1.47	36	>10
PFBS_2	298.9 / 99.0	1.47	35	>10
PFHxA_1	313.0 / 269.0	1.75	102	>10
PFHxA_2	313.0 / 119.0	1.75	41	>10
PFHpA_1	363.0 / 319.0	2.11	32	>10
PFHpA_2	363.0 / 169.0	2.11	32	>10
PFHxS_1	399.0 / 80.0	2.12	60	>10
PFHxS_2	399.0 / 99.0	2.12	56	>10
PFOA_1	413.0 / 369.0	2.49	61	>10
PFOA_2	413.0 / 169.0	2.49	33	>10
PFNA_1	463.0 / 419.0	2.86	56	>10
PFNA_2	463.0 / 219.0	2.86	55	>10
PFOS_1	499.0 / 80.0	2.86	65	>10
PFOS_2	499.0 / 99.0	2.86	50	>10
PFDA_1	513.0 / 469.0	3.22	55	>10
PFDA_2	513.0 / 219.0	3.21	38	>10
PFUnA_1	563.0 / 519.0	3.53	55	>10
PFUnA_2	563.0 / 269.0	3.53	36	>10
PFDaA_1	613.0 / 569.0	3.82	67	>10
PFDaA_2	613.0 / 319.0	3.82	48	>10
PFTrDA_1	663.0 / 619.0	4.08	76	>10
PFTrDA_2	663.0 / 169.0	4.08	33	>10
PFTeDA_1	713.0 / 669.0	4.30	61	>10
PFTeDA_2	713.0 / 169.0	4.30	51	>10
NMeFOSAA_1	570.0 / 419.0	3.36	57	>10
NMeFOSAA_2	570.0 / 512.0	3.36	58	>10
NEtFOSAA_1	584.0 / 419.0	3.53	54	>10
NEtFOSAA_2	584.0 / 483.0	3.52	35	>10



## Precision and Bias at the LOQ for PFAS in Drinking Water

Analyte	CAS No.	Average (ng/L)	ST DEV	3 Sigma	n
PFHxA	307-24-4	10.85	1.18	3.54	10
PFHpA	375-85-9	11.32	1.25	3.75	10
PFOA	335-67-1	11.36	1.21	3.63	10
PFNA	375-95-1	11.19	1.18	3.54	10
PFDA	335-76-2	11.20	1.25	3.75	10
PFUnA	2058-94-8	11.01	1.52	4.56	10
PFDoA	307-55-1	10.77	1.77	5.31	10
PFTTrDA	72629-94-8	10.72	1.69	5.07	10
PFTeDA	376-06-7	11.83	1.43	4.29	10
NMeFOSAA	2355-31-9	10.84	0.95	2.85	10
NEtFOSAA	2991-50-6	10.16	1.00	3.00	10
PFBS	375-73-5	9.01	1.35	4.05	10
PFHxS	355-46-4	10.58	1.28	3.84	10
PFOS	1763-23-1	9.81	1.25	3.75	10

# BATTELLE DETECTION LIMITS FOR PFAS IN DRINKING WATER

Battelle SOP 5-371 (EPA Method 537 Version 1.1)

Analyte	CAS No.	MDL (ng/L)	LOD (ng/L)	LOQ (ng/L)	MRL (ng/L)
<b>PFHxA</b>	307-24-4	0.22	0.5	2.5	2.5
<b>PFHpA</b>	375-85-9	0.34	1.0	2.5	2.5
<b>PFOA</b>	335-67-1	0.38	1.0	2.5	2.5
<b>PFNA</b>	375-95-1	0.37	1.0	2.5	2.5
<b>PFDA</b>	335-76-2	0.39	1.0	2.5	2.5
<b>PFUnA</b>	2058-94-8	0.38	1.0	2.5	2.5
<b>PFDoA</b>	307-55-1	0.42	1.0	2.5	2.5
<b>PFTTrDA</b>	72629-94-8	0.42	1.0	2.5	2.5
<b>PFTeDA</b>	376-06-7	0.73	1.5	2.5	2.5
<b>NMeFOSAA</b>	2355-31-9	0.42	1.0	2.5	2.5
<b>NEtFOSAA</b>	2991-50-6	0.44	1.0	2.5	2.5
<b>PFBS</b>	375-73-5	0.21	0.5	2.5	2.5
<b>PFHxS</b>	3871-99-6	0.34	1.0	2.5	2.5
<b>PFOS</b>	1763-23-1	0.30	1.0	2.5	2.5

*Analytes on NELAP and ELAP QSM 5.1 Scope of accreditation*

## Analytical Transitions for PFAS in drinking water

## SOP 5-371 (EPA 537 Version 1.1)

Analyte	CAS No.	Type	Primary Transition	Secondary Transition
<b>PFHxA</b>	307-24-4	Target	313.0 / 269.0	313.0 / 119.0
<b>PFHpA</b>	375-85-9	Target	363.0 / 319.0	363.0 / 169.0
<b>PFOA</b>	335-67-1	Target	413.0 / 369.0	413.0 / 169.0
<b>PFNA</b>	375-95-1	Target	463.0 / 419.0	463.0 / 219.0
<b>PFDA</b>	335-76-2	Target	513.0 / 469.0	513.0 / 219.0
<b>PFUnA</b>	2058-94-8	Target	563.0 / 519.0	563.0 / 269.0
<b>PFDoA</b>	307-55-1	Target	613.0 / 569.0	613.0 / 319.0
<b>PFTTrDA</b>	72629-94-8	Target	663.0 / 619.0	663.0 / 169.0
<b>PFTeDA</b>	376-06-7	Target	713.0 / 669.0	713.0 / 169.0
<b>NMeFOSAA</b>	2355-31-9	Target	570.0 / 419.0	570.0 / 512.0
<b>NEtFOSAA</b>	2991-50-6	Target	584.0 / 419.0	584.0 / 483.0
<b>PFBS</b>	375-73-5	Target	298.9.0 / 80.0	298.9.0 / 99.0
<b>PFHxS</b>	355-46-4	Target	399.0 / 80.0	399.0 / 99.0
<b>PFOS</b>	1763-23-1	Target	499.0 / 80.0	499.0 / 99.0
<b><sup>13</sup>C<sub>2</sub>-PFHxA</b>	NA	SIS	315.0 / 270.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFDA</b>	NA	SIS	515.0 / 470.0	NA
<b>d<sub>5</sub>-EtFOSAA</b>	NA	SIS	589.0 / 419.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFOA</b>	NA	IS	415.0 / 270.0	NA
<b><sup>13</sup>C<sub>4</sub>-PFOS</b>	NA	IS	503.0 / 80.0	NA
<b>d<sub>3</sub>-MeFOSAA</b>	NA	IS	573.0 / 419.0	NA



**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

# QTRAP 5500

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

## QTRAP 5500 Preventive Maintenance Checklist

<b>Preventive Maintenance Date:</b>	22-Feb-2017
<b>Request ID:</b>	3683
<b>Company Name:</b>	Battelle Memorial Institute
<b>Instrument ID:</b>	X60666
<b>Instrument Model:</b>	QTRAP 5500
<b>Instrument Serial Number:</b>	AU23051004

**PASS**       **FAIL**

**Any failure will lead to an automatic Service Call being open to investigate fault.**

Preventive Maintenance is performed twice every year unless specified in the Service Contract. It is designed to help maintain optimum system performance and to help diagnose any system deficiencies.

Engineer is required the assigned Request ID for this PM otherwise making this job invalid.

**Comments:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Performed By:**           Kaustubh Dhayagude                **Date:**           22-Feb-2017          

**Approved By :** \_\_\_\_\_      **Date:** \_\_\_\_\_

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

# QTRAP 5500

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

## PRE PM PPG PERFORMANCE EVALUATION:

- Consult Customer concerning the unit overall performance.
- Check Logbook for Services recently performed.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.5	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	1.9	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.4	Read Only
<input checked="" type="checkbox"/> CAD High	3.4	Read Only
<input checked="" type="checkbox"/> CAD 12	3.4	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Check for Front end contamination symptoms. Run Q1 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop
- Check for Q3 contamination symptoms. Run Q3 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop

**Pre PM PPG Test:** Perform each of the following tests. Optimize ion source position only. The specifications listed for these Pre PM tests are guidelines only, not required to be met.

- Perform Q1 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 175.133	1.64 e6	Read Only	0.8095	Read Only
Q1 500.380	2.40 e7	Read Only	0.8592	Read Only
Q1 906.673	2.86 e7	Read Only	0.9633	Read Only

- Perform Q3 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 175.133	1.26 e6	Read Only	0.6252	Read Only
Q3 500.380	2.19 e7	Read Only	0.7275	Read Only
Q3 906.673	3.02 e7	Read Only	0.7662	Read Only



**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform MSMS POS in Product Ion scan with 609.3 parent and record daughter 195.1 using Reserpine 0.167 pmol/ul at the scan rate of 10 Da/s for 10 MCA. Calculate transmission efficiency comparing Q1POS 609 intensity. Transmission Efficiency: : 19.51% (Read Only)

Mass	MSMS Intensity		MSMS Width Value	Width Specs
	Value	Spec		
Q1 609.3	7.43 e7	Read Only	0.9981	Read Only
MS/MS 195.1	1.45 e7	Read Only	0.6582	Read Only

Perform Q1 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 933.636	1.43 e7	Read Only	0.7330	Read Only

Perform Q3 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 933.636	2.22 e7	Read Only	0.8138	Read Only

Perform Product Ion scan using NEG PPG 3e-5M. Record 10 mca.

Mass	Scan Rate	MCA	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.35 e6	Read Only	0.6495	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

# QTRAP 5500

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

## PREVENTIVE MAINTENANCE CHECKLIST:

- Check Cooling Fans for Turbo Pumps while MS is ON.
- Check QJet and QPS tuning voltage for reference.
- Record AC input Voltage while MS is OFF: \_\_\_\_\_(200-240VAC).  
If Out-of-Range, notify customer.
  
- Clean Interface
  - Curtain Plate
  - Orifice Plate
  - QJet
  - Q0 Rods.
  
- Replace Roughing Pump Oil.
- Inspect Oil Exhaust Filter, if Applicable.  N/A
- Clean and inspect built-in divert valve if used.  N/A
- Check Multiplier Voltage, optimize if necessary.
- Replace four Air Filters at the bottom of the mass spectrometer.
  
- Pump down overnight if possible.  N/A
  
- Perform Maintenance on Turbo V source.
  
- Replace Electrode, if necessary.  N/A
- Check Turbo heaters resistances.
- Check if Temperature is reached at 500C with TIS Probe installed.
- Check if Temperature is reached at 500C with APCI Probe installed.  N/A

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**POST PM PPG PERFORMANCE TESTS:**

- Set-up Sample for Infusion.
- Check spray and adjust sprayer's position of the TIS source.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.8	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	2.1	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.6	Read Only
<input checked="" type="checkbox"/> CAD High	3.7	Read Only
<input checked="" type="checkbox"/> CAD 12	3.7	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Perform Q1 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q1 175.133	5.94 e6	≥1.2 <sup>e6</sup>	0.6933	0.6 to 0.8
Q1 500.380	2.25 e7	≥9.0 <sup>e6</sup>	0.7444	0.6 to 0.8
Q1 906.673	2.74 e7	≥1.4 <sup>e7</sup>	0.7347	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q1 906.673	1.33 e8	≥6.8 <sup>e7</sup>	0.7656	0.6 to 0.8

- Perform Q3 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q3 175.133	4.54 e6	≥1.2 <sup>e6</sup>	0.6390	0.6 to 0.8
Q3 500.380	2.13 e7	≥9.0 <sup>e6</sup>	0.7008	0.6 to 0.8
Q3 906.673	3.04 e7	≥1.4 <sup>e7</sup>	0.7683	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q3 906.673	1.51 e8	≥6.8 <sup>e7</sup>	0.7118	0.6 to 0.8

- Perform "Product of 609.3" POS and record product ion 195.1 using Reserpine 0.167pmol/uL. Record 10 mca. Calculate Transmission efficiency comparing Q1POS 609 intensity.

Transmission Efficiency: 16.93% (≥ 10.0%)

Mass	MSMS Intensity		Width Value	Width Specs
	Value	Spec		
Q1 609.3	5.74 e7	N/A	0.7667	Read Only
MS/MS 195.1	9.72 e6	N/A	0.6751	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform Q1 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q1 Intensity		Q1 Width Value	Width Specs
			Value	Spec		
Q1 933.636	10	10	1.31 e7	$\geq 1.0^{e7}$	0.6895	0.6 to 0.8
Q1 933.636	1000	50	6.32 e7	$\geq 4.0^{e7}$	0.6740	0.6 to 0.8

Perform Q3 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q3 Intensity		Q3 Width Value	Width Specs
			Value	Spec		
Q3 933.636	10	10	1.70 e7	$\geq 8.0^{e6}$	0.7665	0.6 to 0.8
Q3 933.636	1000	50	7.41 e7	$\geq 4.0^{e7}$	0.7292	0.6 to 0.8

Perform Product Ion scan using NEG PPG 3e-5M.

Mass	Scan Rate	Mca	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.33 e6	Read Only	0.6387	Read Only

Perform ER POS 118.087 and 922.01 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 118.087	0.05	8.08 e6	$\geq 7.2^{e6}$	0.1302	<0.35
ER 922.010	0.05	3.89 e7	$\geq 2.8^{e6}$	0.2603	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 118.087	0.05	2.55 e7	$\geq 2.4^{e7}$	0.3740	<0.65
ER 922.010	0.05	2.37 e8	$\geq 6.8^{e7}$	0.5407	<0.65

Perform ER NEG 431.982 and 601.978 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 431.982	0.05	1.05 e8	$\geq 4.4^{e7}$	0.1840	<0.35
ER 601.978	0.05	7.74 e7	$\geq 5.6^{e7}$	0.1849	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 431.982	0.05	3.43 e8	$\geq 1.2^{e8}$	0.4382	<0.65
ER 601.978	0.05	2.55 e8	$\geq 1.6^{e8}$	0.6205	<0.65

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

- Perform EPI POS 397.2 using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Q0 Trapping OFF		Q0 Trapping ON	
		Intensity	Spec	Intensity	Spec
EPI 397.2	10000	> 3.5 e6	≥2.0 e6	> 4.0 e7	≥6.4 e6

- Perform MS3 POS full scan Fragmentation ON & OFF using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Fragamentation OFF		Fragmentation ON	
		Intensity	Spec	Intensity	Spec
MS3 397.2	1000	3.2 e7	Contains only 397.2	N/A	N/A
<input type="checkbox"/> 236 OR <input checked="" type="checkbox"/> 365	1000	1.19 e8	Fragment Intensity	> 4.4 e6	≥1.6x 10 <sup>e6</sup>

**REVIEW:**

- Attach all spectrums printouts to this procedure.
- If any parameter setting access modes were changed during the PM, ensure they are returned to their normal access mode and that their offsets are adjusted to match optimized values from the post-PM acquisition files.
- Empty tuning cache folder, if necessary.  N/A
- Update Service Work Order status
- Fill and replace PM Label.

**END OF PREVENTIVE MAINTENANCE CHECKLIST****Document history:**

06 OCT 2016: Appendix ZEFPM003-2L: Removed requirements to fit Manufacturer's testing criteria.

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV35**

Description: PFAS - 537.1 Internal Standard Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180425-01	EPA-537IS	Neat	~2.66666 6	12/13/22	---	---	1000 uL	1	10	~0.3000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: \_\_\_\_\_ Date: \_\_\_\_\_



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: JV35

Description: PFAS - 537.1 Internal Standard Stock

Stock Id: 180425-01

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	1000	1.00	1	100.000	1	10	0.10000
13C4-PFOS	1000	2.87	1	100.000	1	10	0.28700
d3-MeFOSAA	1000	4.00	1	100.000	1	10	0.40000

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFOA	.10000
13C4-PFOS	.28700
d3-MeFOSAA	.40000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180425-01	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: \_\_\_\_\_ Date: \_\_\_\_\_



It can be done

## Standard Solution Prep Form II

Approved:

Standard Laboratory ID Number: **JV37**

Description: PFAS - 537.1 Surrogate Standard Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180425-02	EPA-537SS	Neat	~2.00000 0	11/08/22	---	---	1000 uL	1	10	~0.2000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: \_\_\_\_\_ Date: \_\_\_\_\_





It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV37**

Description: PFAS - 537.1 Surrogate Standard Stock

Stock Id: **180425-02**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	1000	1.00	1	100.000	1	10	0.10000
13C2-PFHxA	1000	1.00	1	100.000	1	10	0.10000
d5-EtFOSAA	1000	4.00	1	100.000	1	10	0.40000

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.10000
13C2-PFHxA	.10000
d5-EtFOSAA	.40000

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180425-02	Pipette	C0982448K

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: \_\_\_\_\_ Date: \_\_\_\_\_

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV41**

Description: PFAS - 537.1 Second Source LCS/MS Solution

Assigned Lab ID (from receipt log)	Chemical Name:.	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180425-04	EPA-537PDS-L (second source)	Neat	~2.00000 0	03/05/23	---	---	500 uL	1	20	~0.0500

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Thorn, Jonathan Date: 5/3/2018 8:26:00 AM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV41

Description: PFAS - 537.1 Second Source LCS/MS Solution

Stock Id: 180425-04

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	500	2.00	1	100.000	1	20	0.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-1-butanefluorobutane	500	1.77	1	100.000	1	20	0.04425
Perfluoro-1-hexanesulfonate	500	1.89	1	100.000	1	20	0.04725
Perfluoro-1-octanesulfonate	500	1.91	1	100.000	1	20	0.04775
Perfluoro-n-decanoic Acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-dodecanoic acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-heptanoic Acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-hexanoic acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-octanoic Acid	500	2.00	1	100.000	1	20	0.05000
Perfluorononanoic Acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-tetradecanoic acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-tridecanoic acid	500	2.00	1	100.000	1	20	0.05000
Perfluoro-n-undecanoic acid	500	2.00	1	100.000	1	20	0.05000

## Final Concentrations:

Analyte:	Conc (ug/mL):
N-ethylperfluoro-octanesulfonamidoacetic acid	.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.05000
Perfluoro-1-butanefluorobutane	.04425
Perfluoro-1-hexanesulfonate	.04725
Perfluoro-1-octanesulfonate	.04775
Perfluoro-n-decanoic Acid	.05000
Perfluoro-n-dodecanoic acid	.05000
Perfluoro-n-heptanoic Acid	.05000
Perfluoro-n-hexanoic acid	.05000
Perfluoro-n-octanoic Acid	.05000
Perfluorononanoic Acid	.05000
Perfluoro-n-tetradecanoic acid	.05000
Perfluoro-n-tridecanoic acid	.05000
Perfluoro-n-undecanoic acid	.05000

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180425-04	Pipette	B1100330B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Thorn, Jonathan Date: 5/3/2018 8:26:00 AM

It can be done

## Standard Solution Prep Form II

Approved: 

Standard Laboratory ID Number: JV42

Description: PFAS - 537.1 High ICAL Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180425-03	EPA-537PDS (calibration)	Neat	~2.00000 0	03/05/23	---	---	250 uL	1	10	~0.0500

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)  
Approved By: Schumitz, Denise Date: 5/3/2018 3:21:00 PM





It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV42

Description: PFAS - 537.1 High ICAL Stock

Stock Id: 180425-03

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	250	2.00	1	100.000	1	10	0.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-1-butanefluoride	250	1.77	1	100.000	1	10	0.04425
Perfluoro-1-hexanesulfonate	250	1.82	1	100.000	1	10	0.04560
Perfluoro-1-octanesulfonate	250	1.85	1	100.000	1	10	0.04628
Perfluoro-n-decanoic Acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-dodecanoic acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-heptanoic Acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-hexanoic acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-nonanoic Acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-octanoic Acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-tetradecanoic acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-tridecanoic acid	250	2.00	1	100.000	1	10	0.05000
Perfluoro-n-undecanoic acid	250	2.00	1	100.000	1	10	0.05000

## Final Concentrations:

Analyte:	Conc (ug/mL):
N-ethylperfluoro-octanesulfonamidoacetic acid	.05000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.05000
Perfluoro-1-butanefluoride	.04425
Perfluoro-1-hexanesulfonate	.04560
Perfluoro-1-octanesulfonate	.04628
Perfluoro-n-decanoic Acid	.05000
Perfluoro-n-dodecanoic acid	.05000
Perfluoro-n-heptanoic Acid	.05000
Perfluoro-n-hexanoic acid	.05000
Perfluoro-n-nonanoic Acid	.05000
Perfluoro-n-octanoic Acid	.05000
Perfluoro-n-tetradecanoic acid	.05000
Perfluoro-n-tridecanoic acid	.05000
Perfluoro-n-undecanoic acid	.05000

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180425-03	Pipette	B1100330B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:21:00 PM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV43**

Description: PFAS - 537.1 Low ICAL Stock

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
180425-03	EPA-537PDS (calibration)	Neat	~2.00000 0	03/05/23	---	---	250 uL	1	100	~0.0050

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 4 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:21:00 PM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV43

Description: PFAS - 537.1 Low ICAL Stock

Stock Id: 180425-03

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	250	2.00	1	100.000	1	100	0.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-1-butanefulfonate	250	1.77	1	100.000	1	100	0.00443
Perfluoro-1-hexanesulfonate	250	1.82	1	100.000	1	100	0.00456
Perfluoro-1-octanesulfonate	250	1.85	1	100.000	1	100	0.00463
Perfluoro-n-decanoic Acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-dodecanoic acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-heptanoic Acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-hexanoic acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-nonanoic Acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-octanoic Acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-tetradecanoic acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-tridecanoic acid	250	2.00	1	100.000	1	100	0.00500
Perfluoro-n-undecanoic acid	250	2.00	1	100.000	1	100	0.00500

## Final Concentrations:

Analyte:	Conc (ug/mL):
N-ethylperfluoro-octanesulfonamidoacetic acid	.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00500
Perfluoro-1-butanefulfonate	.00443
Perfluoro-1-hexanesulfonate	.00456
Perfluoro-1-octanesulfonate	.00463
Perfluoro-n-decanoic Acid	.00500
Perfluoro-n-dodecanoic acid	.00500
Perfluoro-n-heptanoic Acid	.00500
Perfluoro-n-hexanoic acid	.00500
Perfluoro-n-nonanoic Acid	.00500
Perfluoro-n-octanoic Acid	.00500
Perfluoro-n-tetradecanoic acid	.00500
Perfluoro-n-tridecanoic acid	.00500
Perfluoro-n-undecanoic acid	.00500

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
180425-03	Pipette	B1100330B

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 4 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:21:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV59**

Description: PFAS - 537.1 Internal Standard Solution

Assigned Lab ID (from receipt lcg)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV35	PFAS - 537.1 Internal Standard Stock	Solution	~0	05/02/19	---	---	500 uL	1	25	~0.0000

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)  
Approved By: Thorn, Jonathan Date: 5/3/2018 8:27:00 AM





It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV59**

Description: PFAS - 537.1 Internal Standard Solution

Stock Id: **JV35**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	500	0.10	---	---	1	25	0.00200
13C4-PFOS	500	0.29	---	---	1	25	0.00574
d3-MeFOSAA	500	0.40	---	---	1	25	0.00800

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFOA	.00200
13C4-PFOS	.00574
d3-MeFOSAA	.00800

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV35	Pipette	I0400533B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Thorn, Jonathan      Date: 5/3/2018 8:27:00 AM

It can be done

## Standard Solution Prep Form II

Approved:

Standard Laboratory ID Number: **JV60**

Description: PFAS - 537.1 Surrogate Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV37	PFAS - 537.1 Surrogate Standard Stock	Solution	~0	05/02/19	---	---	500 uL	1	25	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Thorn, Jonathan Date: 5/3/2018 8:27:00 AM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV60**

Description: PFAS - 537.1 Surrogate Solution

Stock Id: **JV37**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	500	0.10	---	---	1	25	0.00200
13C2-PFHxA	500	0.10	---	---	1	25	0.00200
d5-EtFOSAA	500	0.40	---	---	1	25	0.00800

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00200
13C2-PFHxA	.00200
d5-EtFOSAA	.00800

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV37	Pipette	I0400533B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Thorn, Jonathan Date: 5/3/2018 8:27:00 AM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV61**

Description: PFAS - 537.1 Internal Standard Calibration Stock Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV35	PFAS - 537.1 Internal Standard Stock	Solution	~0	05/02/19	---	---	1000 uL	1	5	~0.0000

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM



It can be done

**Standard Solution Concentrations** Approved:

Standard Laboratory ID Number: **JV61**  
 Description: PFAS - 537.1 Internal Standard Calibration Stock Solution

Stock Id: **JV35**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	1000	0.10	---	---	1	5	0.02000
13C4-PFOS	1000	0.29	---	---	1	5	0.05740
d3-MeFOSAA	1000	0.40	---	---	1	5	0.08000

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFOA	.02000
13C4-PFOS	.05740
d3-MeFOSAA	.08000

Syringes/Pipettes:

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM



It can be done

## Standard Solution Prep Form II

Approved:

Standard Laboratory ID Number: **JV62**

Description: PFAS - 537.1 Surrogate Calibration Stock Solution

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV37	PFAS - 537.1 Surrogate Standard Stock	Solution	~0	05/02/19	---	---	1000 uL	1	5	~0.0000

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV62**

Description: PFAS - 537.1 Surrogate Calibration Stock Solution

Stock Id: **JV37**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	1000	0.10	---	---	1	5	0.02000
13C2-PFHxA	1000	0.10	---	---	1	5	0.02000
d5-EtFOSAA	1000	0.40	---	---	1	5	0.08000

Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.02000
13C2-PFHxA	.02000
d5-EtFOSAA	.08000

Syringes/Pipettes:

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV63**

Description: PFAS - 537.1 ICC

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV41	PFAS - 537.1 Second Source LCS/MS Solution	Solution	~0	05/02/19	---	---	200 uL	1	10	~0.0000
JV59	PFAS - 537.1 Internal Standard Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV60	PFAS - 537.1 Surrogate Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM





It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV63

Description: PFAS - 537.1 ICC

## Stock Id: JV41

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-1-butanedisulfonate	200	0.04	---	---	1	10	0.00089
Perfluoro-1-hexanesulfonate	200	0.05	---	---	1	10	0.00095
Perfluoro-1-octanesulfonate	200	0.05	---	---	1	10	0.00095
Perfluoro-n-decanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-dodecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-heptanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-hexanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-octanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluorononanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tetradecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tridecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-undecanoic acid	200	0.05	---	---	1	10	0.00100

## Stock Id: JV59

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.00	---	---	1	10	0.00001
13C4-PFOS	50	0.01	---	---	1	10	0.00003
d3-MeFOSAA	50	0.01	---	---	1	10	0.00004

## Stock Id: JV60

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.00	---	---	1	10	0.00001
13C2-PFHxA	50	0.00	---	---	1	10	0.00001
d5-EtFOSAA	50	0.01	---	---	1	10	0.00004

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00001
13C2-PFHxA	.00001
13C2-PFOA	.00001
13C4-PFOS	.00003
d3-MeFOSAA	.00004
d5-EtFOSAA	.00004

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:23:00 PM

**BATTELLE**

It can be done

**Standard Solution Concentrations**Approved: Standard Laboratory ID Number: **JV63**

Description: PFAS - 537.1 ICC

N-ethylperfluoro-octanesulfonamidoacetic acid	.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00100
Perfluoro-1-butanefulfonate	.00089
Perfluoro-1-hexanesulfonate	.00095
Perfluoro-1-octanesulfonate	.00095
Perfluoro-n-decanoic Acid	.00100
Perfluoro-n-dodecanoic acid	.00100
Perfluoro-n-heptanoic Acid	.00100
Perfluoro-n-hexanoic acid	.00100
Perfluoro-n-octanoic Acid	.00100
Perfluorononanoic Acid	.00100
Perfluoro-n-tetradecanoic acid	.00100
Perfluoro-n-tridecanoic acid	.00100
Perfluoro-n-undecanoic acid	.00100

**Syringes/Pipettes:**

Stock ID:	Type:	Battelle ID:
JV41	Pipette	G0792979B
JV59	Pipette	I0793912B
JV60	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1      Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:23:00 PM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV64**

Description: PFAS - 537.1 ICAL L1

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV43	PFAS - 537.1 Low ICAL Stock	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM





It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV64**

Description: PFAS - 537.1 ICAL L1

**Stock Id: JV43**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	50	0.01	---	---	1	10	0.00003
N-methylperfluoro-1-octanesulfonamidoacetic acid	50	0.01	---	---	1	10	0.00003
Perfluoro-1-butanedisulfonate	50	0.00	---	---	1	10	0.00002
Perfluoro-1-hexanesulfonate	50	0.00	---	---	1	10	0.00002
Perfluoro-1-octanesulfonate	50	0.00	---	---	1	10	0.00002
Perfluoro-n-decanoic Acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-dodecanoic acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-heptanoic Acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-hexanoic acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-nonanoic Acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-octanoic Acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-tetradecanoic acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-tridecanoic acid	50	0.01	---	---	1	10	0.00003
Perfluoro-n-undecanoic acid	50	0.01	---	---	1	10	0.00003

**Stock Id: JV61**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

**Stock Id: JV62**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV64**

Description: PFAS - 537.1 ICAL L1

N-ethylperfluoro-octanesulfonamidoacetic acid	.00003
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00003
Perfluoro-1-butanefulfonate	.00002
Perfluoro-1-hexanesulfonate	.00002
Perfluoro-1-octanesulfonate	.00002
Perfluoro-n-decanoic Acid	.00003
Perfluoro-n-dodecanoic acid	.00003
Perfluoro-n-heptanoic Acid	.00003
Perfluoro-n-hexanoic acid	.00003
Perfluoro-n-nonanoic Acid	.00003
Perfluoro-n-octanoic Acid	.00003
Perfluoro-n-tetradecanoic acid	.00003
Perfluoro-n-tridecanoic acid	.00003
Perfluoro-n-undecanoic acid	.00003

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV43	Pipette	I0793912B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV65**

Description: PFAS - 537.1 ICAL L2

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV43	PFAS - 537.1 Low ICAL Stock	Solution	~0	05/02/19	---	---	100 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM





It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV65**

Description: PFAS - 537.1 ICAL L2

**Stock Id: JV43**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	100	0.01	---	---	1	10	0.00005
N-methylperfluoro-1-octanesulfonamidoacetic acid	100	0.01	---	---	1	10	0.00005
Perfluoro-1-butanefulfonate	100	0.00	---	---	1	10	0.00004
Perfluoro-1-hexanesulfonate	100	0.00	---	---	1	10	0.00005
Perfluoro-1-octanesulfonate	100	0.00	---	---	1	10	0.00005
Perfluoro-n-decanoic Acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-dodecanoic acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-heptanoic Acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-hexanoic acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-nonanoic Acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-octanoic Acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-tetradecanoic acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-tridecanoic acid	100	0.01	---	---	1	10	0.00005
Perfluoro-n-undecanoic acid	100	0.01	---	---	1	10	0.00005

**Stock Id: JV61**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

**Stock Id: JV62**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:22:00 PM



It can be done

**Standard Solution Concentrations** Approved:

Standard Laboratory ID Number: **JV65**

Description: PFAS - 537.1 ICAL L2

N-ethylperfluoro-octanesulfonamidoacetic acid	.00005
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00005
Perfluoro-1-butanefulfonate	.00004
Perfluoro-1-hexanesulfonate	.00005
Perfluoro-1-octanesulfonate	.00005
Perfluoro-n-decanoic Acid	.00005
Perfluoro-n-dodecanoic acid	.00005
Perfluoro-n-heptanoic Acid	.00005
Perfluoro-n-hexanoic acid	.00005
Perfluoro-n-nonanoic Acid	.00005
Perfluoro-n-octanoic Acid	.00005
Perfluoro-n-tetradecanoic acid	.00005
Perfluoro-n-tridecanoic acid	.00005
Perfluoro-n-undecanoic acid	.00005

**Syringes/Pipettes:**

Stock ID:	Type:	Battelle ID:
JV43	Pipette	I0793912B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV66**

Description: PFAS - 537.1 ICAL L3

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV43	PFAS - 537.1 Low ICAL Stock	Solution	~0	05/02/19	---	---	200 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV66

Description: PFAS - 537.1 ICAL L3

## Stock Id: JV43

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	200	0.01	---	---	1	10	0.00010
N-methylperfluoro-1-octanesulfonamidoacetic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-1-butanedisulfonate	200	0.00	---	---	1	10	0.00009
Perfluoro-1-hexanesulfonate	200	0.00	---	---	1	10	0.00009
Perfluoro-1-octanesulfonate	200	0.00	---	---	1	10	0.00009
Perfluoro-n-decanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-dodecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-heptanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-hexanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-nonanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-octanoic Acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-tetradecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-tridecanoic acid	200	0.01	---	---	1	10	0.00010
Perfluoro-n-undecanoic acid	200	0.01	---	---	1	10	0.00010

## Stock Id: JV61

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

## Stock Id: JV62

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV66**

Description: PFAS - 537.1 ICAL L3

N-ethylperfluoro-octanesulfonamidoacetic acid	.00010
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00010
Perfluoro-1-butanefulfonate	.00009
Perfluoro-1-hexanesulfonate	.00009
Perfluoro-1-octanesulfonate	.00009
Perfluoro-n-decanoic Acid	.00010
Perfluoro-n-dodecanoic acid	.00010
Perfluoro-n-heptanoic Acid	.00010
Perfluoro-n-hexanoic acid	.00010
Perfluoro-n-nonanoic Acid	.00010
Perfluoro-n-octanoic Acid	.00010
Perfluoro-n-tetradecanoic acid	.00010
Perfluoro-n-tridecanoic acid	.00010
Perfluoro-n-undecanoic acid	.00010

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV43	Pipette	G0792979B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV67**

Description: PFAS - 537.1 ICAL L4

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV43	PFAS - 537.1 Low ICAL Stock	Solution	~0	05/02/19	---	---	500 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV67**

Description: PFAS - 537.1 ICAL L4

**Stock Id: JV43**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	500	0.01	---	---	1	10	0.00025
N-methylperfluoro-1-octanesulfonamidoacetic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-1-butanedisulfonate	500	0.00	---	---	1	10	0.00022
Perfluoro-1-hexanesulfonate	500	0.00	---	---	1	10	0.00023
Perfluoro-1-octanesulfonate	500	0.00	---	---	1	10	0.00023
Perfluoro-n-decanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-dodecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-heptanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-hexanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-nonanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-octanoic Acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-tetradecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-tridecanoic acid	500	0.01	---	---	1	10	0.00025
Perfluoro-n-undecanoic acid	500	0.01	---	---	1	10	0.00025

**Stock Id: JV61**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

**Stock Id: JV62**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:22:00 PM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV67**

Description: PFAS - 537.1 ICAL L4

N-ethylperfluoro-octanesulfonamidoacetic acid	.00025
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00025
Perfluoro-1-butanefulfonate	.00022
Perfluoro-1-hexanesulfonate	.00023
Perfluoro-1-octanesulfonate	.00023
Perfluoro-n-decanoic Acid	.00025
Perfluoro-n-dodecanoic acid	.00025
Perfluoro-n-heptanoic Acid	.00025
Perfluoro-n-hexanoic acid	.00025
Perfluoro-n-nonanoic Acid	.00025
Perfluoro-n-octanoic Acid	.00025
Perfluoro-n-tetradecanoic acid	.00025
Perfluoro-n-tridecanoic acid	.00025
Perfluoro-n-undecanoic acid	.00025

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV43	Pipette	I0400533B
JV61	Pipette	D1075429B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV68**

Description: PFAS - 537.1 ICAL L5

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV43	PFAS - 537.1 Low ICAL Stock	Solution	~0	05/02/19	---	---	1000 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date:

5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID:

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise

Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV68

Description: PFAS - 537.1 ICAL L5

## Stock Id: JV43

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	0.01	---	---	1	10	0.00050
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-1-butanedisulfonate	1000	0.00	---	---	1	10	0.00044
Perfluoro-1-hexanesulfonate	1000	0.00	---	---	1	10	0.00046
Perfluoro-1-octanesulfonate	1000	0.00	---	---	1	10	0.00046
Perfluoro-n-decanoic Acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-dodecanoic acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-heptanoic Acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-hexanoic acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-nonanoic Acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-octanoic Acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-tetradecanoic acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-tridecanoic acid	1000	0.01	---	---	1	10	0.00050
Perfluoro-n-undecanoic acid	1000	0.01	---	---	1	10	0.00050

## Stock Id: JV61

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

## Stock Id: JV62

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM





It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV68**

Description: PFAS - 537.1 ICAL L5

N-ethylperfluoro-octanesulfonamidoacetic acid	.00050
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00050
Perfluoro-1-butanefulfonate	.00044
Perfluoro-1-hexanesulfonate	.00046
Perfluoro-1-octanesulfonate	.00046
Perfluoro-n-decanoic Acid	.00050
Perfluoro-n-dodecanoic acid	.00050
Perfluoro-n-heptanoic Acid	.00050
Perfluoro-n-hexanoic acid	.00050
Perfluoro-n-nonanoic Acid	.00050
Perfluoro-n-octanoic Acid	.00050
Perfluoro-n-tetradecanoic acid	.00050
Perfluoro-n-tridecanoic acid	.00050
Perfluoro-n-undecanoic acid	.00050

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV43	Pipette	I0400533B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM

It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV69**

Description: PFAS - 537.1 ICAL L6

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV42	PFAS - 537.1 High ICAL Stock	Solution	~0	05/02/19	---	---	200 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV69**

Description: PFAS - 537.1 ICAL L6

Stock Id: **JV42**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-1-butanedisulfonate	200	0.04	---	---	1	10	0.00089
Perfluoro-1-hexanesulfonate	200	0.05	---	---	1	10	0.00091
Perfluoro-1-octanesulfonate	200	0.05	---	---	1	10	0.00093
Perfluoro-n-decanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-dodecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-heptanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-hexanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-nonanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-octanoic Acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tetradecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-tridecanoic acid	200	0.05	---	---	1	10	0.00100
Perfluoro-n-undecanoic acid	200	0.05	---	---	1	10	0.00100

Stock Id: **JV61**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

Stock Id: **JV62**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV69**

Description: PFAS - 537.1 ICAL L6

N-ethylperfluoro-octanesulfonamidoacetic acid	.00100
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00100
Perfluoro-1-butanefulfonate	.00089
Perfluoro-1-hexanesulfonate	.00091
Perfluoro-1-octanesulfonate	.00093
Perfluoro-n-decanoic Acid	.00100
Perfluoro-n-dodecanoic acid	.00100
Perfluoro-n-heptanoic Acid	.00100
Perfluoro-n-hexanoic acid	.00100
Perfluoro-n-nonanoic Acid	.00100
Perfluoro-n-octanoic Acid	.00100
Perfluoro-n-tetradecanoic acid	.00100
Perfluoro-n-tridecanoic acid	.00100
Perfluoro-n-undecanoic acid	.00100

Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV42	Pipette	G0792979B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107		

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV70**

Description: PFAS - 537.1 ICAL L7

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV42	PFAS - 537.1 High ICAL Stock	Solution	~0	05/02/19	---	---	500 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV70

Description: PFAS - 537.1 ICAL L7

## Stock Id: JV42

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	500	0.05	---	---	1	10	0.00250
N-methylperfluoro-1-octanesulfonamidoacetic acid	500	0.05	---	---	1	10	0.00250
Perfluoro-1-butanefulfonate	500	0.04	---	---	1	10	0.00221
Perfluoro-1-hexanesulfonate	500	0.05	---	---	1	10	0.00228
Perfluoro-1-octanesulfonate	500	0.05	---	---	1	10	0.00231
Perfluoro-n-decanoic Acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-dodecanoic acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-heptanoic Acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-hexanoic acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-nonanoic Acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-octanoic Acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-tetradecanoic acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-tridecanoic acid	500	0.05	---	---	1	10	0.00250
Perfluoro-n-undecanoic acid	500	0.05	---	---	1	10	0.00250

## Stock Id: JV61

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

## Stock Id: JV62

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie Date Prepared: 5/2/2018 Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV70**

Description: PFAS - 537.1 ICAL L7

N-ethylperfluoro-octanesulfonamidoacetic acid	.00250
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00250
Perfluoro-1-butanefulfonate	.00221
Perfluoro-1-hexanesulfonate	.00228
Perfluoro-1-octanesulfonate	.00231
Perfluoro-n-decanoic Acid	.00250
Perfluoro-n-dodecanoic acid	.00250
Perfluoro-n-heptanoic Acid	.00250
Perfluoro-n-hexanoic acid	.00250
Perfluoro-n-nonanoic Acid	.00250
Perfluoro-n-octanoic Acid	.00250
Perfluoro-n-tetradecanoic acid	.00250
Perfluoro-n-tridecanoic acid	.00250
Perfluoro-n-undecanoic acid	.00250

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV42	Pipette	I0400533B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Prep Form II

Approved: Standard Laboratory ID Number: **JV71**

Description: PFAS - 537.1 ICAL L8

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV42	PFAS - 537.1 High ICAL Stock	Solution	~0	05/02/19	---	---	1000 uL	1	10	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	50 uL	1	10	~0.0000

Solution Prepared By: Schultz, Stephanie

Date Prepared: 5/2/2018

Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials

Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved:

Standard Laboratory ID Number: **JV71**

Description: PFAS - 537.1 ICAL L8

### Stock Id: JV42

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	0.05	---	---	1	10	0.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-1-butanefulfonate	1000	0.04	---	---	1	10	0.00443
Perfluoro-1-hexanesulfonate	1000	0.05	---	---	1	10	0.00456
Perfluoro-1-octanesulfonate	1000	0.05	---	---	1	10	0.00463
Perfluoro-n-decanoic Acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-dodecanoic acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-heptanoic Acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-hexanoic acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-nonanoic Acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-octanoic Acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-tetradecanoic acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-tridecanoic acid	1000	0.05	---	---	1	10	0.00500
Perfluoro-n-undecanoic acid	1000	0.05	---	---	1	10	0.00500

### Stock Id: JV61

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	50	0.02	---	---	1	10	0.00010
13C4-PFOS	50	0.06	---	---	1	10	0.00029
d3-MeFOSAA	50	0.08	---	---	1	10	0.00040

### Stock Id: JV62

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	50	0.02	---	---	1	10	0.00010
13C2-PFHxA	50	0.02	---	---	1	10	0.00010
d5-EtFOSAA	50	0.08	---	---	1	10	0.00040

### Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:22:00 PM



It can be done

## Standard Solution Concentrations

Approved: 

Standard Laboratory ID Number: JV71

Description: PFAS - 537.1 ICAL L8

N-ethylperfluoro-octanesulfonamidoacetic acid	.00500
N-methylperfluoro-1-octanesulfonamidoacetic acid	.00500
Perfluoro-1-butanedisulfonate	.00443
Perfluoro-1-hexanesulfonate	.00456
Perfluoro-1-octanesulfonate	.00463
Perfluoro-n-decanoic Acid	.00500
Perfluoro-n-dodecanoic acid	.00500
Perfluoro-n-heptanoic Acid	.00500
Perfluoro-n-hexanoic acid	.00500
Perfluoro-n-nonanoic Acid	.00500
Perfluoro-n-octanoic Acid	.00500
Perfluoro-n-tetradecanoic acid	.00500
Perfluoro-n-tridecanoic acid	.00500
Perfluoro-n-undecanoic acid	.00500

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV42	Pipette	I0400533B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:22:00 PM

It can be done

## Standard Solution Prep Form II

Approved:

Standard Laboratory ID Number: **JV72**

Description: PFAS - 537.1 ICAL L9

Assigned Lab ID (from receipt log)	Chemical Name:	Source	Stock (ug/mL)	Expir. Date	Purity (%)	Density (g/mL)	Amount Taken	Conv. Fact.	Final Vol. (mL)	Std. Conc. (ug/mL)
JV42	PFAS - 537.1 High ICAL Stock	Solution	~0	05/02/19	---	---	1000 uL	1	5	~0.0000
JV61	PFAS - 537.1 Internal Standard Calibration Stock Solution	Solution	~0	05/02/19	---	---	25 uL	1	5	~0.0000
JV62	PFAS - 537.1 Surrogate Calibration Stock Solution	Solution	~0	05/02/19	---	---	25 uL	1	5	~0.0000

Solution Prepared By: Schultz, Stephanie	Date Prepared: 5/2/2018	Expiration Date: 5/2/2019
Solution Volume 40 mL X 1 Vials	Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107	

Balance ID: \_\_\_\_\_

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise Date: 5/3/2018 3:21:00 PM





It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV72**

Description: PFAS - 537.1 ICAL L9

Stock Id: **JV42**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
N-ethylperfluoro-octanesulfonamidoacetic acid	1000	0.05	---	---	1	5	0.01000
N-methylperfluoro-1-octanesulfonamidoacetic acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-1-butanedisulfonate	1000	0.04	---	---	1	5	0.00885
Perfluoro-1-hexanesulfonate	1000	0.05	---	---	1	5	0.00912
Perfluoro-1-octanesulfonate	1000	0.05	---	---	1	5	0.00925
Perfluoro-n-decanoic Acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-dodecanoic acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-heptanoic Acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-hexanoic acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-nonanoic Acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-octanoic Acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-tetradecanoic acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-tridecanoic acid	1000	0.05	---	---	1	5	0.01000
Perfluoro-n-undecanoic acid	1000	0.05	---	---	1	5	0.01000

Stock Id: **JV61**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFOA	25	0.02	---	---	1	5	0.00010
13C4-PFOS	25	0.06	---	---	1	5	0.00029
d3-MeFOSAA	25	0.08	---	---	1	5	0.00040

Stock Id: **JV62**

Chemical Name	Stock Amount uL	Initial Conc. (ug/mL)	Density (g/mL)	Purity	Conv. Factor	Final Vol mL	Concentration (ug/mL)
13C2-PFDA	25	0.02	---	---	1	5	0.00010
13C2-PFHxA	25	0.02	---	---	1	5	0.00010
d5-EtFOSAA	25	0.08	---	---	1	5	0.00040

## Final Concentrations:

Analyte:	Conc (ug/mL):
13C2-PFDA	.00010
13C2-PFHxA	.00010
13C2-PFOA	.00010
13C4-PFOS	.00029
d3-MeFOSAA	.00040
d5-EtFOSAA	.00040

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:21:00 PM



It can be done

## Standard Solution Concentrations

Approved: Standard Laboratory ID Number: **JV72**

Description: PFAS - 537.1 ICAL L9

N-ethylperfluoro-octanesulfonamidoacetic acid	.01000
N-methylperfluoro-1-octanesulfonamidoacetic acid	.01000
Perfluoro-1-butanefulfonate	.00885
Perfluoro-1-hexanesulfonate	.00912
Perfluoro-1-octanesulfonate	.00925
Perfluoro-n-decanoic Acid	.01000
Perfluoro-n-dodecanoic acid	.01000
Perfluoro-n-heptanoic Acid	.01000
Perfluoro-n-hexanoic acid	.01000
Perfluoro-n-nonanoic Acid	.01000
Perfluoro-n-octanoic Acid	.01000
Perfluoro-n-tetradecanoic acid	.01000
Perfluoro-n-tridecanoic acid	.01000
Perfluoro-n-undecanoic acid	.01000

## Syringes/Pipettes:

Stock ID:	Type:	Battelle ID:
JV42	Pipette	I0400533B
JV61	Pipette	I0793912B
JV62	Pipette	I0793912B

Solution Prepared By: Schultz, Stephanie      Date Prepared: 5/2/2018      Expiration Date: 5/2/2019

Solution Volume 40 mL X 1 Vials Refrigerator/Freezer No: LC Laboratory: Refrigerator - R0107

Comment: 96/4 methanol/milli-q (RP-180502-2)

Approved By: Schumitz, Denise      Date: 5/3/2018 3:21:00 PM

It can be done

BDO Id: 180425-01

## Reagent Receipt Report

Approved:  Authorized

Name: EPA-537IS Received: 4/25/2018  
Vendor: Wellington Laboratories Custodian: Schumitz, Matt  
Catalogue No: EPA-537IS Expires: 12/13/2022  
Type: Solution Consumed: \_\_\_\_\_  
Lot No: 537IS1217 Stored In: AqChem Laboratory - R0124  
Quantity: 1 ea mL % Moisture: \_\_\_\_\_  
Description: EPA-537IS

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
d3-N-MeFOSAA	BDO-1838	4.0000	100.00	--	--	<input type="checkbox"/>			
M2PFOA	BDO-1842	1.0000	100.00	--	--	<input type="checkbox"/>			
MPFOS	BDO-1840	2.8700	100.00	--	--	<input type="checkbox"/>			

Total Analytes: 3

Notes:

Approved by: \_\_\_\_\_ Approved on: \_\_\_\_\_  
Authorized by: \_\_\_\_\_ Authorized on: \_\_\_\_\_

**WELLINGTON  
LABORATORIES****CERTIFICATE OF ANALYSIS  
DOCUMENTATION****EPA-537IS****Internal Standard  
Primary Dilution Standard**

**PRODUCT CODE:** EPA-537IS  
**LOT NUMBER:** 537IS1217  
**SOLVENT(S):** Methanol / Water (<1%)  
**DATE PREPARED:** (mm/dd/yyyy) 12/13/2017  
**LAST TESTED:** (mm/dd/yyyy) 12/13/2017  
**EXPIRY DATE:** (mm/dd/yyyy) 12/13/2022  
**RECOMMENDED STORAGE:** Refrigerate ampoule

**DESCRIPTION:**

EPA-537IS is a solution/mixture of a mass-labelled (<sup>13</sup>C) perfluoroalkylcarboxylic acid, a mass-labelled (<sup>13</sup>C) perfluoroalkylsulfonate, and a mass-labelled (<sup>2</sup>H) perfluorooctanesulfonamidoacetic acid. The components and their concentrations are given in Table A.

The mass-labelled perfluoroalkylcarboxylic acid and the mass-labelled perfluoroalkylsulfonate both have chemical purities of >98% and isotopic purities of ≥99%. The mass-labelled perfluorooctanesulfonamidoacetic acid has a chemical purity of >98% and an isotopic purity of ≥98%.

**DOCUMENTATION/ DATA ATTACHED:**

Table A: Components and Concentrations of the Solution/Mixture  
Figure 1: LC/MS Data (TIC)  
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

**ADDITIONAL INFORMATION:**

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.

**FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE**

**Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA  
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com**

**INTENDED USE:**

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

**HAZARDS:**

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

**SYNTHESIS / CHARACTERIZATION:**

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

**HOMOGENEITY:**

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

**UNCERTAINTY:**

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty,  $u_c(y)$ , of a value  $y$  and the uncertainty of the independent parameters  $x_1, x_2, \dots, x_n$  on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where  $x$  is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of  $\pm 5\%$  (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

**TRACEABILITY:**

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using calibrated NIST and/or NRC traceable external weights. All volumetric glassware used is calibrated, of Class A tolerance, and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

**EXPIRY DATE / PERIOD OF VALIDITY:**

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

**LIMITED WARRANTY:**

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

**QUALITY MANAGEMENT:**

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).




\*\*For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at [www.well-labs.com](http://www.well-labs.com) or contact us directly at [info@well-labs.com](mailto:info@well-labs.com)\*\*




**Table A: EPA-537IS; Components and Concentrations (ng/ml;  $\pm$  5% in Methanol / Water (<1%))**

Compound	Abbreviation	Concentration (ng/ml)		Peak Assignment in Figure 1
		as the salt	as the anion	
Perfluoro-n-[1,2- <sup>13</sup> C <sub>2</sub> ]octanoic acid	M2PFOA	1000		A
N-methyl-d <sub>3</sub> -perfluoro-1-octanesulfonamidoacetic acid	d3-N-MeFOSAA	4000		C
Compound	Abbreviation	Concentration (ng/ml)		Peak Assignment in Figure 1
		as the salt	as the anion	
Sodium perfluoro-1-[1,2,3,4- <sup>13</sup> C <sub>4</sub> ]octanesulfonate	MPFOS	3000	2870	B

Certified By: \_\_\_\_\_


  
B.G. Chittim, General Manager
Date: 12/22/2017  
(mm/dd/yyyy)

**Figure 1: EPA-537IS; LC/MS Data (Total Ion Current Chromatogram)****Conditions for Figure 1:**

**LC:** Waters Acquity Ultra Performance LC  
**MS:** Micromass Quattro *micro* API MS

**Chromatographic Conditions**

Column: Acquity UPLC BEH Shield RP<sub>18</sub>  
1.7  $\mu$ m, 2.1 x 100 mm


Mobile phase: Gradient  
Start: 40% (80:20 MeOH:ACN) / 60% H<sub>2</sub>O  
(both with 10 mM NH<sub>4</sub>OAc buffer)  
Ramp to 90% organic over 7 min  
and hold for 2 min before returning  
to initial conditions in 0.5 min.  
Time: 10 min

Flow: 300  $\mu$ l/min

**MS Parameters**

Experiment: Full Scan (150 - 850 amu)

Source: Electrospray (negative)  
Capillary Voltage (kV) = 3.00  
Cone Voltage (V) = 25.00  
Cone Gas Flow (l/hr) = 100  
Desolvation Gas Flow (l/hr) = 750

**Figure 2: EPA-537IS; LC/MS/MS Data (Selected MRM Transitions)****Conditions for Figure 2:**

Injection: On-column (EPA-537IS)

Mobile phase: Same as Figure 1

Flow: 300  $\mu$ l/min**MS Parameters**

Collision Gas (mbar) = 3.28e-3

Collision Energy (eV) = 11-40 (variable)

It can be done

BDO Id: 180425-02

## Reagent Receipt Report

Approved:  Authorized 

Name: EPA-537SS Received: 4/25/2018  
Vendor: Wellington Laboratories Custodian: Schumitz, Matt  
Catalogue No: EPA-537SS Expires: 11/8/2022  
Type: Solution Consumed: \_\_\_\_\_  
Lot No: 537SS1117 Stored In: AqChem Laboratory - R0124  
Quantity: 1 ea ml % Moisture: \_\_\_\_\_  
Description: EPA-537SS

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
13C2-PFDA	BDO-2110	1.0000	100.00	--	--	<input type="checkbox"/>			
13C2-PFHxA	BDO-2106	1.0000	100.00	--	--	<input type="checkbox"/>			
d5-EtFOSAA	BDO-1839	4.0000	100.00	--	--	<input type="checkbox"/>			

Total Analytes: 3

Notes:

Approved by: Thorn, Jonathan Approved on: 5/2/2018 10:00:00 AM  
Authorized by: \_\_\_\_\_ Authorized on: \_\_\_\_\_

**WELLINGTON**  
LABORATORIES**CERTIFICATE OF ANALYSIS**  
DOCUMENTATION**EPA-537SS****Surrogate Primary Dilution Standard**

**PRODUCT CODE:** EPA-537SS  
**LOT NUMBER:** 537SS1117  
**SOLVENT(S):** Methanol / Water (<1%)  
**DATE PREPARED:** (mm/dd/yyyy) 11/06/2017  
**LAST TESTED:** (mm/dd/yyyy) 11/08/2017  
**EXPIRY DATE:** (mm/dd/yyyy) 11/08/2022  
**RECOMMENDED STORAGE:** Refrigerate ampoule

**DESCRIPTION:**

EPA-537SS is a solution/mixture of two mass-labelled (<sup>13</sup>C) perfluoroalkylcarboxylic acids and a mass-labelled (<sup>2</sup>H) perfluorooctanesulfonamidoacetic acid. The components and their concentrations are given in Table A.

The mass-labelled perfluoroalkylcarboxylic acids both have chemical purities of >98% and isotopic purities of ≥99%. The mass-labelled perfluorooctanesulfonamidoacetic acid has a chemical purity of >98% and an isotopic purity of ≥98%.

**DOCUMENTATION/ DATA ATTACHED:**

Table A: Components and Concentrations of the Solution/Mixture  
Figure 1: LC/MS Data (TIC)  
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

**ADDITIONAL INFORMATION:**

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acids to their respective methyl esters.

**FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE**

**Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA**  
**519-822-2436 • Fax: 519-822-2849 • info@well-labs.com**

**INTENDED USE:**

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

**HAZARDS:**

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

**SYNTHESIS / CHARACTERIZATION:**

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

**HOMOGENEITY:**

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

**UNCERTAINTY:**

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty,  $u_c(y)$ , of a value  $y$  and the uncertainty of the independent parameters

$x_1, x_2, \dots, x_n$  on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where  $x$  is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of  $\pm 5\%$  (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

**TRACEABILITY:**

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using calibrated NIST and/or NRC traceable external weights. All volumetric glassware used is calibrated, of Class A tolerance, and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

**EXPIRY DATE / PERIOD OF VALIDITY:**

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

**LIMITED WARRANTY:**

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

**QUALITY MANAGEMENT:**

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



\*\*For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at [www.well-labs.com](http://www.well-labs.com) or contact us directly at [info@well-labs.com](mailto:info@well-labs.com)\*\*


**Table A: EPA-537SS; Components and Concentrations (ng/ml; ± 5% in Methanol / Water (<1%))**

Compound	Abbreviation	Concentration (ng/ml)	Peak Assignment in Figure 1
Perfluoro-n-[1,2- <sup>13</sup> C <sub>2</sub> ]hexanoic acid	MPFHxA	1000	A
Perfluoro-n-[1,2- <sup>13</sup> C <sub>2</sub> ]decanoic acid	MPFDA	1000	B
N-ethyl-d <sub>5</sub> -perfluoro-1-octanesulfonamidoacetic acid	d5-N-EtFOSAA	4000	C

Certified By: 

B.G. Chittim, General Manager

Date: 11/13/2017  
(mm/dd/yyyy)

**Figure 1: EPA-537SS; LC/MS Data (Total Ion Current Chromatogram)****Conditions for Figure 1:****LC:** Waters Acquity Ultra Performance LC**MS:** Micromass Quattro *micro* API MS**Chromatographic Conditions**Column: Acquity UPLC BEH Shield RP<sub>18</sub>  
1.7  $\mu$ m, 2.1 x 100 mm

Mobile phase: Gradient

Start: 45% (80:20 MeOH:ACN) / 55% H<sub>2</sub>O  
(both with 10 mM NH<sub>4</sub>OAc buffer)Ramp to 90% organic over 7 min  
and hold for 2 min before returning  
to initial conditions in 0.5 min.

Time: 10 min

Flow: 300  $\mu$ l/min**MS Parameters**

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)


Capillary Voltage (kV) = 3.00

Cone Voltage (V) = 25.00

Cone Gas Flow (l/hr) = 100

Desolvation Gas Flow (l/hr) = 750



**Figure 2: EPA-537SS; LC/MS/MS Data (Selected MRM Transitions)****Conditions for Figure 2:**

Injection: On-column (EPA-537SS)

Mobile phase: Same as Figure 1

Flow: 300  $\mu$ l/min**MS Parameters**

Collision Gas (mbar) = 3.50e-3

Collision Energy (eV) = 9-40 (variable)



It can be done

BDO Id: 180425-03

## Reagent Receipt Report

Approved:  Authorized

**Name:** EPA-537PDS (calibration) **Received:** 4/25/2018  
**Vendor:** Wellington Laboratories **Custodian:** Schumitz, Matt  
**Catalogue No:** EPA-537PDS **Expires:** 3/5/2023  
**Type:** Solution **Consumed:** \_\_\_\_\_  
**Lot No:** 537PDS0318 **Stored In:** AqChem Laboratory - R0124  
**Quantity:** 1 ea ml **% Moisture:** \_\_\_\_\_  
**Description:** EPA-537PDS

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
N-ethylperfluoro-octanesulfonamidoa	2991-50-6	2.0000	100.00	--	--	<input type="checkbox"/>			1
N-methylperfluoro-1-octanesulfonami	2355-31-9	2.0000	100.00	--	--	<input type="checkbox"/>			2
Perfluoro-1-butanefulfonate	375-73-5	1.7700	100.00	--	--	<input type="checkbox"/>			3
Perfluoro-1-hexanesulfonate	355-46-4	1.8240	100.00	--	--	<input type="checkbox"/>			4
Perfluoro-1-octanesulfonate	1763-23-1	1.8510	100.00	--	--	<input type="checkbox"/>			5
Perfluoro-n-decanoic Acid	335-76-2	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-dodecanoic acid	307-55-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-heptanoic Acid	375-85-9	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-hexanoic acid	307-24-4	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-nonanoic Acid	375-95-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-octanoic Acid	335-67-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-tetradecanoic acid	376-06-7	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-tridecanoic acid	72629-94-8	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-undecanoic acid	2058-94-8	2.0000	100.00	--	--	<input type="checkbox"/>			

**Total Analytes:** 14

**Notes:**

Analyte:	Comment:
1 N-ethylperfluoro-octanesulfonamidoacetic acid	sum of branched and linear isomers
2 N-methylperfluoro-1-octanesulfonamidoacetic acid	sum of branched and linear isomers
3 Perfluoro-1-butanefulfonate	2000 ng/ml as the salt, 1770 ng/ml as the anion
4 Perfluoro-1-hexanesulfonate	1998 ng/ml as the salt, 1824 ng/ml as the anion. sum of branched and linear isomers.
5 Perfluoro-1-octanesulfonate	2002 ng/ml as the salt, 1851 ng/ml as the anion. sum of branched and linear isomers.

**Approved by:** Thorn, Jonathan **Approved on:** 5/2/2018 10:05:00 AM  
**Authorized by:** \_\_\_\_\_ **Authorized on:** \_\_\_\_\_

**WELLINGTON**  
LABORATORIES**CERTIFICATE OF ANALYSIS**  
DOCUMENTATION**EPA-537PDS****Native PFAS Primary Dilution  
Standard Solution/Mixture**

**PRODUCT CODE:** EPA-537PDS  
**LOT NUMBER:** 537PDS0318  
**SOLVENT(S):** Methanol / Water (<1%)  
**DATE PREPARED:** (mm/dd/yyyy) 03/02/2018  
**LAST TESTED:** (mm/dd/yyyy) 03/05/2018  
**EXPIRY DATE:** (mm/dd/yyyy) 03/05/2023  
**RECOMMENDED STORAGE:** Refrigerate ampoule

for calibration  
JNT 5/2/2018

**DESCRIPTION:**

EPA-537PDS is a solution/mixture of nine native linear perfluoroalkylcarboxylic acids (C<sub>6</sub>-C<sub>14</sub>), three native perfluoroalkylsulfonates (C<sub>4</sub> linear; C<sub>6</sub> and C<sub>8</sub> linear and branched), and two native perfluorooctanesulfonamidoacetic acids (linear and branched). The components and their concentrations are given in Table A.

The native perfluoroalkylcarboxylic acids, native perfluoroalkylsulfonates, and native perfluorooctanesulfonamidoacetic acids have chemical purities of >98%.

**DOCUMENTATION/ DATA ATTACHED:**

Table A: Components and Concentrations of the Solution/Mixture  
Table B: Isomeric Components and Percent Composition of N-MeFOSAA  
Table C: Isomeric Components and Percent Composition of N-EtFOSAA  
Table D: Isomeric Components and Percent Composition of PFHxSK  
Table E: Isomeric Components and Percent Composition of PFOSK  
Figure 1: LC/MS Data (SIR)  
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

**ADDITIONAL INFORMATION:**

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acids to their respective methyl esters.

**FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE**

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA  
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

**INTENDED USE:**

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

**HANDLING:**

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

**SYNTHESIS / CHARACTERIZATION:**

Our products are synthesized using single-product unambiguous routes whenever possible. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

**HOMOGENEITY:**

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS, and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner. This further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

**UNCERTAINTY:**

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty,  $u_c(y)$ , of a value  $y$  and the uncertainty of the independent parameters  $x_1, x_2, \dots, x_n$  on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where  $x$  is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of  $\pm 5\%$  (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

**TRACEABILITY:**

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly calibrated by an external ISO/IEC 17025 accredited laboratory. In addition, their calibration is verified prior to each weighing using calibrated external weights traceable to an ISO/IEC 17025 accredited laboratory. All volumetric glassware used is calibrated, of Class A tolerance, and traceable to an ISO/IEC 17025 accredited laboratory. For certain products, traceability to international interlaboratory studies has also been established.

**EXPIRY DATE / PERIOD OF VALIDITY:**

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

**LIMITED WARRANTY:**

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

**QUALITY MANAGEMENT:**

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO 17034 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



\*\*For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at [www.well-labs.com](http://www.well-labs.com) or contact us directly at [info@well-labs.com](mailto:info@well-labs.com)\*\*

**Table A:** EPA-537PDS; Components and Concentrations (ng/ml;  $\pm$  5% in Methanol / Water (<1%))

Compound	Abbreviation	Concentration *		Peak Assignment in Figure 1
		(ng/ml)		
Perfluoro-n-hexanoic acid ✓	PFHxA	2000		B
Perfluoro-n-heptanoic acid ✓	PFHpA	2000		C
Perfluoro-n-octanoic acid ✓	PFOA	2000		F
Perfluoro-n-nonanoic acid ✓	PFNA	2000		G
Perfluoro-n-decanoic acid ✓	PFDA	2000		J
Perfluoro-n-undecanoic acid ✓	PFUdA	2000		O
Perfluoro-n-dodecanoic acid ✓	PFDoA	2000		P
Perfluoro-n-tridecanoic acid ✓	PFTrDA	2000		Q
Perfluoro-n-tetradecanoic acid ✓	PFTeDA	2000		R
N-methylperfluorooctanesulfonamidoacetic acid <sup>a</sup> ✓	N-MeFOSAA: linear isomer ✓	1520		L
	N-MeFOSAA: $\Sigma$ branched isomers	480		K
N-ethylperfluorooctanesulfonamidoacetic acid <sup>b</sup> ✓	N-EtFOSAA: linear isomer ✓	1550		N
	N-EtFOSAA: $\Sigma$ branched isomers	450		M
Compound	Abbreviation	Concentration *		Peak Assignment in Figure 1
		as the salt	as the anion	
Potassium perfluoro-1-butanesulfonate ✓	L-PFBS ✓	2000	1770	A
Potassium perfluorohexanesulfonate <sup>c</sup>	PFHxSK: linear isomer	1620	1480	E
	PFHxSK: $\Sigma$ branched isomers	378	344	D
Potassium perfluorooctanesulfonate <sup>d</sup>	PFOSK: linear isomer	1580	1460	I
	PFOSK: $\Sigma$ branched isomers	422	391	H

<sup>a</sup> See Table B for percent composition of linear and branched N-MeFOSAA isomers.

<sup>b</sup> See Table C for percent composition of linear and branched N-EtFOSAA isomers.

<sup>c</sup> See Table D for percent composition of linear and branched PFHxSK isomers.

<sup>d</sup> See Table E for percent composition of linear and branched PFOSK isomers.

\* Concentrations have been rounded to three significant figures.

**Table B: N-MeFOSAA; Isomeric Components and Percent Composition (by <sup>19</sup>F-NMR)\***

Isomer	Name	Structure	Percent Composition by <sup>19</sup> F-NMR	
1	N-methylperfluoro-1-octanesulfonamidoacetic acid	$\text{CF}_3(\text{CF}_2)_7\text{SO}_2\text{NCH}_2\text{CO}_2\text{H}$ $\quad \quad \quad  $ $\quad \quad \quad \text{CH}_3$	76.0	76.0
2	N-methylperfluoro-3-methylheptanesulfonamidoacetic acid	$\text{CF}_3(\text{CF}_2)_3\text{CF}(\text{CF}_2)_2\text{SO}_2\text{NCH}_2\text{CO}_2\text{H}$ $\quad \quad \quad   \quad \quad \quad  $ $\quad \quad \quad \text{CF}_3 \quad \quad \quad \text{CH}_3$	0.7	24.0
3	N-methylperfluoro-4-methylheptanesulfonamidoacetic acid	$\text{CF}_3(\text{CF}_2)_2\text{CF}(\text{CF}_2)_3\text{SO}_2\text{NCH}_2\text{CO}_2\text{H}$ $\quad \quad \quad   \quad \quad \quad  $ $\quad \quad \quad \text{CF}_3 \quad \quad \quad \text{CH}_3$	2.0	
4	N-methylperfluoro-5-methylheptanesulfonamidoacetic acid	$\text{CF}_3\text{CF}_2\text{CF}(\text{CF}_2)_4\text{SO}_2\text{NCH}_2\text{CO}_2\text{H}$ $\quad \quad \quad   \quad \quad \quad  $ $\quad \quad \quad \text{CF}_3 \quad \quad \quad \text{CH}_3$	6.0	
5	N-methylperfluoro-6-methylheptanesulfonamidoacetic acid	$\text{CF}_3\text{CF}(\text{CF}_2)_5\text{SO}_2\text{NCH}_2\text{CO}_2\text{H}$ $\quad \quad \quad   \quad \quad \quad  $ $\quad \quad \quad \text{CF}_3 \quad \quad \quad \text{CH}_3$	14.0	
6	N-methylperfluoro-5,5-dimethylhexanesulfonamidoacetic acid	$\begin{array}{c} \text{CF}_3 \\   \\ \text{CF}_3\text{C}(\text{CF}_2)_4\text{SO}_2\text{NCH}_2\text{CO}_2\text{H} \\   \\ \text{CF}_3 \end{array}$ $\quad \quad \quad  $ $\quad \quad \quad \text{CH}_3$	0.2	
7	Other Unidentified Isomers		1.1	

\* Percent of total N-methylperfluorooctanesulfonamidoacetic acid isomers only.





**Table D: PFHxSK; Isomeric Components and Percent Composition (by <sup>19</sup>F-NMR)\***

Isomer	Name	Structure	Percent Composition by <sup>19</sup> F-NMR	
1	Potassium perfluoro-1-hexanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>	81.1	81.1
2	Potassium 1-trifluoromethylperfluoropentanesulfonate**	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )K <sup>+</sup>   CF <sub>3</sub>	2.9	18.9
3	Potassium 2-trifluoromethylperfluoropentanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF(CF <sub>3</sub> )SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	1.4	
4	Potassium 3-trifluoromethylperfluoropentanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF(CF <sub>3</sub> )CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	5.0	
5	Potassium 4-trifluoromethylperfluoropentanesulfonate	CF <sub>3</sub> CF(CF <sub>3</sub> )CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	8.9	
6	Potassium 3,3-di(trifluoromethyl)perfluorobutanesulfonate	CF <sub>3</sub>   CF <sub>3</sub> CCF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	0.2	
7	Other Unidentified Isomers		0.5	

\* Percent of total perfluorohexanesulfonate isomers only.  
 \*\* Systematic Name: Potassium perfluorohexane-2-sulfonate.


**Table E: PFOSK; Isomeric Components and Percent Composition (by <sup>19</sup>F-NMR)\***

Isomer	Name	Structure	Percent Composition by <sup>19</sup> F-NMR	
1	Potassium perfluoro-1-octanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>	78.8	78.8
2	Potassium 1-trifluoromethylperfluoroheptanesulfonate**	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )K <sup>+</sup>   CF <sub>3</sub>	1.2	21.1
3	Potassium 2-trifluoromethylperfluoroheptanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> K <sup>+</sup>   CF <sub>3</sub>	0.6	
4	Potassium 3-trifluoromethylperfluoroheptanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF <sub>2</sub> K <sup>+</sup>   CF <sub>3</sub>	1.9	
5	Potassium 4-trifluoromethylperfluoroheptanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> K <sup>+</sup>   CF <sub>3</sub>	2.2	
6	Potassium 5-trifluoromethylperfluoroheptanesulfonate	CF <sub>3</sub> CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> K <sup>+</sup>   CF <sub>3</sub>	4.5	
7	Potassium 6-trifluoromethylperfluoroheptanesulfonate	CF <sub>3</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> K <sup>+</sup>   CF <sub>3</sub>	10.0	
8	Potassium 5,5-di(trifluoromethyl)perfluorohexanesulfonate	CF <sub>3</sub>   CF <sub>3</sub> CCF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	0.2	
9	Potassium 4,4-di(trifluoromethyl)perfluorohexanesulfonate	CF <sub>3</sub>   CF <sub>3</sub> CF <sub>2</sub> CCF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	0.03	
10	Potassium 4,5-di(trifluoromethyl)perfluorohexanesulfonate	CF <sub>3</sub>   CF <sub>3</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF <sub>2</sub> CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	0.4	
11	Potassium 3,5-di(trifluoromethyl)perfluorohexanesulfonate	CF <sub>3</sub>   CF <sub>3</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> CF(SO <sub>3</sub> <sup>-</sup> )CF <sub>2</sub> SO <sub>3</sub> <sup>-</sup> K <sup>+</sup>   CF <sub>3</sub>	0.07	

\* Percent of total perfluorooctanesulfonate isomers only.

\*\* Systematic Name: Potassium perfluorooctane-2-sulfonate.


Certified By:



B.G. Chittim, General Manager

Date: 04/02/2018

(mm/dd/yyyy)

**Figure 1: EPA-537PDS; LC/MS Data (SIR)****Conditions for Figure 1:****LC:** Waters Acquity Ultra Performance LC**MS:** Micromass Quattro *micro* API MS**Chromatographic Conditions**Column: Acquity UPLC BEH Shield RP<sub>18</sub>  
1.7  $\mu$ m, 2.1 x 100 mm

Mobile phase: Gradient

Start: 40% (80:20 MeOH:ACN) / 60% H<sub>2</sub>O  
(both with 10 mM NH<sub>4</sub>OAc buffer)

Ramp to 55% organic over 3.5 min.

Ramp to 70% organic over 6.5 min.

Ramp to 85% organic over 5 min and hold for

1 min before returning to initial conditions in 0.5 min.

Time: 17 min

Flow: 300  $\mu$ l/min**MS Parameters**

Experiment: SIR


Source: Electrospray (negative)


Capillary Voltage (kV) = 3.00

Cone Voltage (V) = variable (15-60)

Cone Gas Flow (l/hr) = 100

Desolvation Gas Flow (l/hr) = 750

**Figure 2: EPA-537PDS; LC/MS/MS Data (Selected MRM Transitions)**

**Figure 2:** EPA-537PDS; LC/MS/MS Data (Selected MRM Transitions)**Conditions for Figure 2:**

Injection: On-column (EPA-537PDS)

Mobile phase: Same as Figure 1

Flow: 300  $\mu$ l/min**MS Parameters**

Collision Gas (mbar) = 3.10e-3

Collision Energy (eV) = 10-40 (variable)

It can be done

BDO Id: 180425-04

## Reagent Receipt Report

Approved:  Authorized 

Name: EPA-537PDS-L (second source) Received: 4/25/2018  
Vendor: Wellington Laboratories Custodian: Schumitz, Matt  
Catalogue No: EPA-537PDS-L Expires: 3/5/2023  
Type: Solution Consumed: \_\_\_\_\_  
Lot No: 537PDSL0318 Stored In: AqChem Laboratory - R0124  
Quantity: 1 ea ml % Moisture: \_\_\_\_\_  
Description: EPA-537PDS-L

Analyte:	CAS No:	Concentration (ug/mL):	Purity:	Density:	Density Units:	Cert	Cert Val:	Lower Limit:	Upper Limit:
N-ethylperfluoro-octanesulfonamidoa	2991-50-6	2.0000	100.00	--	--	<input type="checkbox"/>			
N-methylperfluoro-1-octanesulfonami	2355-31-9	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-1-butanefulfonic Acid	375-73-5	2.0000	100.00	--	--	<input type="checkbox"/>			1
Perfluoro-n-decanoic Acid	335-76-2	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-dodecanoic acid	307-55-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-heptanoic Acid	375-85-9	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-hexanoic acid	307-24-4	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-octanoic Acid	335-67-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluorononanoic Acid	375-95-1	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-tetradecanoic acid	376-06-7	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-tridecanoic acid	72629-94-8	2.0000	100.00	--	--	<input type="checkbox"/>			
Perfluoro-n-undecanoic acid	2058-94-8	2.0000	100.00	--	--	<input type="checkbox"/>			
Sodium perfluoro-1-hexanesulfonate	82382-12-5	1.8900	100.00	--	--	<input type="checkbox"/>			2
Sodium perfluoro-1-octanesulfonate	4021-47-0	1.9100	100.00	--	--	<input type="checkbox"/>			3

Total Analytes: 14

## Notes:

Analyte:	Comment:
1 Perfluoro-1-butanefulfonic Acid	2000 ng/ml as the salt, 1770 ng/ml as the anion
2 Sodium perfluoro-1-hexanesulfonate	2000 ng/ml as the salt, 1890 ng/ml as the anion
3 Sodium perfluoro-1-octanesulfonate	2000 ng/ml as the salt, 1910 ng/ml as the anion

Approved by: \_\_\_\_\_ Approved on: \_\_\_\_\_  
Authorized by: \_\_\_\_\_ Authorized on: \_\_\_\_\_



**WELLINGTON**  
LABORATORIES**CERTIFICATE OF ANALYSIS**  
DOCUMENTATION**EPA-537PDS-L****Native PFAS Linear Primary Dilution  
Standard Solution/Mixture**

**PRODUCT CODE:** EPA-537PDS-L  
**LOT NUMBER:** 537PDSL0318  
**SOLVENT(S):** Methanol / Water (<1%)  
**DATE PREPARED:** (mm/dd/yyyy) 03/02/2018  
**LAST TESTED:** (mm/dd/yyyy) 03/05/2018  
**EXPIRY DATE:** (mm/dd/yyyy) 03/05/2023  
**RECOMMENDED STORAGE:** Refrigerate ampoule

**DESCRIPTION:**

EPA-537PDS-L is a solution/mixture of native linear perfluoroalkylcarboxylic acids (C<sub>6</sub>-C<sub>14</sub>), native linear perfluoroalkylsulfonates (C<sub>4</sub>, C<sub>6</sub>, and C<sub>8</sub>), and native linear perfluorooctanesulfonamidoacetic acids. The components and their concentrations are given in Table A.

The native perfluoroalkylcarboxylic acids, native perfluoroalkylsulfonates, and native perfluorooctanesulfonamidoacetic acids have chemical purities of >98%.

**DOCUMENTATION/ DATA ATTACHED:**

Table A: Components and Concentrations of the Solution/Mixture  
Figure 1: LC/MS Data (SIR)  
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

**ADDITIONAL INFORMATION:**

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acids to their respective methyl esters.

**FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE**

**Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA**  
519-822-2436 • Fax: 519-822-2849 • [info@well-labs.com](mailto:info@well-labs.com)

**INTENDED USE:**

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compounds it contains.

**HANDLING:**

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

**SYNTHESIS / CHARACTERIZATION:**

Our products are synthesized using single-product unambiguous routes whenever possible. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

**HOMOGENEITY:**

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS, and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products, as well as mixtures and calibration solutions, are compared to older lots in a similar manner. This further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers. In order to maintain the integrity of the assigned value(s), and associated uncertainty, the dilution or injection of a subsample of this product should be performed using calibrated measuring equipment.

**UNCERTAINTY:**

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty,  $u_c(y)$ , of a value  $y$  and the uncertainty of the independent parameters

$x_1, x_2, \dots, x_n$  on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where  $x$  is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of  $\pm 5\%$  (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

**TRACEABILITY:**

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly calibrated by an external ISO/IEC 17025 accredited laboratory. In addition, their calibration is verified prior to each weighing using calibrated external weights traceable to an ISO/IEC 17025 accredited laboratory. All volumetric glassware used is calibrated, of Class A tolerance, and traceable to an ISO/IEC 17025 accredited laboratory. For certain products, traceability to international interlaboratory studies has also been established.

**EXPIRY DATE / PERIOD OF VALIDITY:**

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

**LIMITED WARRANTY:**

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.

**QUALITY MANAGEMENT:**

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO 17034 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



\*\*For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at [www.well-labs.com](http://www.well-labs.com) or contact us directly at [info@well-labs.com](mailto:info@well-labs.com)\*\*

**Table A: EPA-537PDS-L; Components and Concentrations (ng/ml; ± 5% in Methanol / Water (<1%))**


Compound	Abbreviation	Concentration *		Peak Assignment in Figure 1
		(ng/ml)		
Perfluoro-n-hexanoic acid ✓	PFHxA	2000		B
Perfluoro-n-heptanoic acid ✓	PFHpA	2000		C
Perfluoro-n-octanoic acid ✓	PFOA	2000		E
Perfluoro-n-nonanoic acid ✓	PFNA	2000		F
Perfluoro-n-decanoic acid ✓	PFDA	2000		H
Perfluoro-n-undecanoic acid ✓	PFUDA	2000		K
Perfluoro-n-dodecanoic acid ✓	PFDoA	2000		L
Perfluoro-n-tridecanoic acid ✓	PFTDA	2000		M
Perfluoro-n-tetradecanoic acid ✓	PFTeDA	2000		N
N-methylperfluoro-1-octanesulfonamidoacetic acid ✓	N-MeFOSAA	2000		I
N-ethylperfluoro-1-octanesulfonamidoacetic acid ✓	N-EtFOSAA	2000		J
Compound	Abbreviation	Concentration (ng/ml)		Peak Assignment in Figure 1
		as the salt	as the anion	
Potassium perfluoro-1-butanefluorobutanesulfonate ✓	L-PFBS	2000	1770	A
Sodium perfluoro-1-hexanesulfonate ✓	L-PFHxS	2000	1890	D
Sodium perfluoro-1-octanesulfonate ✓	L-PFOS	2000	1910	G

\* Concentrations have been rounded to three significant figures.

Certified By: \_\_\_\_\_

B.G. Chittim, General Manager

Date: 04/02/2018  
(mm/dd/yyyy)

**Figure 1: EPA-537PDS-L; LC/MS Data (SIR)****Conditions for Figure 1:**

**LC:** Waters Acquity Ultra Performance LC  
**MS:** Micromass Quattro *micro* API MS

**Chromatographic Conditions**

Column: Acquity UPLC BEH Shield RP<sub>18</sub>  
1.7  $\mu$ m, 2.1 x 100 mm


Mobile phase: Gradient  
Start: 40% (80:20 MeOH:ACN) / 60% H<sub>2</sub>O  
(both with 10 mM NH<sub>4</sub>OAc buffer)  
Ramp to 55% organic over 3.5 min.  
Ramp to 70% organic over 6.5 min.  
Ramp to 85% organic over 5 min and hold for  
1 min before returning to initial conditions in 0.5 min.  
Time: 17 min


Flow: 300  $\mu$ l/min

**MS Parameters**

Experiment: SIR

Source: Electrospray (negative)  
Capillary Voltage (kV) = 3.00  
Cone Voltage (V) = variable (15-60)  
Cone Gas Flow (l/hr) = 100  
Desolvation Gas Flow (l/hr) = 750

**Figure 2: EPA-537PDS-L; LC/MS/MS Data (Selected MRM Transitions)**


**Figure 2: EPA-537PDS-L; LC/MS/MS Data (Selected MRM Transitions)****Conditions for Figure 2:**

Injection: On-column (EPA-537PDS-L)

Mobile phase: Same as Figure 1

Flow: 300  $\mu$ l/min**MS Parameters**

Collision Gas (mbar) = 3.17e-3

Collision Energy (eV) = 10-40 (variable)

# Sample Preparation





It can be done

**BATTELLE - NORWELL OPERATIONS  
SAMPLE PREPARATION RECORDS**

<b><u>Project Title(s)</u></b>	<b><u>Project No.(s)</u></b>
Naval Air Station Joint Reserve Base Willow Grove, PA	100117920-WE04
<b>18-0313</b>	
<b>WE04 PFAS Analysis</b>	
<b>DW</b>	
SOP Numbers (see workplan for modifications)	
VOASOP No.	5-371

<b>This Batch Contains The Following Samples:</b>		
CQ755PB-FS	J6156-FS	J6168-FS
CQ756LCS-FS	J6158-FS	J6170-FS
J6148-FS	J6160-FS	
J6150-FS	J6162-FS	
J6152-FS	J6164-FS	
J6154-FS	J6166-FS	

Laboratory Preparation Records  
COMPLETE AND VALIDATED

Prep Task Leader: Stephanie Schultz

Approved By:	Date	Initials
Denise Schumitz	05/16/2018	DMS



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE IDENTIFICATION PAGE

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	Description
CQ755PB-FS	Procedural Blank
CQ756LCS-FS	Laboratory Control Sample
J6148-FS	NAWC-050718-RW-316
J6150-FS	NAWC-050718-RW-180
J6152-FS	NAWC-050718-RW-275
J6154-FS	NAWC-050718-RW-145
J6156-FS	NAWC-050718-RW-357
J6158-FS	NAWC-050718-RW-162
J6160-FS	WGNA-050718-RW-0800
J6162-FS	WGNA-050718-RW-0335
J6164-FS	WGNA-050718-RW-3556
J6166-FS	NAWC-050718-RW-356
J6168-FS	NAWC-050718-RW-289
J6170-FS	WGNA-050718-DUP-35

Samples Assigned By:

Stephanie Schultz

Date :

May 9, 2018

Comments:



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE CUSTODY LOG

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

<b>Requested On/By:</b> 05/10/2018 SAS	<b>Purpose:</b> Sample Preparation
<b>Relinquished On/By:</b> 05/10/2018 MDS	<b>Last Activity:</b> Transfer
<b>Accepted On/By:</b> 05/10/2018 SAS <b>Stored In Facility:</b> Sample Preparation <b>Stored Until:</b> 05/10/2018 <b>Stored Comment:</b> NA	<b>Returned On/To:</b> <b>Returned To Facility:</b> <b>Returned Comment:</b> NA

No.	BDO-ID:	Ctrs	*	Condition:	Custody Comment:
1	J6148	1	C	Consumed	NA
2	J6150	1	C	Consumed	NA
3	J6152	1	C	Consumed	NA
4	J6154	1	C	Consumed	NA
5	J6156	1	C	Consumed	NA
6	J6158	1	C	Consumed	NA
7	J6160	1	C	Consumed	NA
8	J6162	1	C	Consumed	NA
9	J6164	1	C	Consumed	NA
10	J6166	1	C	Consumed	NA
11	J6168	1	C	Consumed	NA
12	J6170	1	C	Consumed	NA
<b>Total Samples</b>		12		* "C" = Consumed Container	



It can be done

## BATTELLE - NORWELL OPERATIONS LIQUID SAMPLE ID FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	Description	Volume (mL)	Bottles	*	Date Initials
CQ755PB-FS	Procedural Blank	250.0	NA	--	05/10/18 SAS
CQ756LCS-FS	Laboratory Control Sample	250.0	NA	--	05/10/18 SAS
J6148-FS	NAWC-050718-RW-316	285.0	1	C	05/11/18 SAS
J6150-FS	NAWC-050718-RW-180	280.0	1	C	05/11/18 SAS
J6152-FS	NAWC-050718-RW-275	280.0	1	C	05/11/18 SAS
J6154-FS	NAWC-050718-RW-145	280.0	1	C	05/11/18 SAS
J6156-FS	NAWC-050718-RW-357	270.0	1	C	05/11/18 SAS
J6158-FS	NAWC-050718-RW-162	280.0	1	C	05/11/18 SAS
J6160-FS	WGNA-050718-RW-0800	285.0	1	C	05/11/18 SAS
J6162-FS	WGNA-050718-RW-0335	270.0	1	C	05/11/18 SAS
J6164-FS	WGNA-050718-RW-3556	260.0	1	C	05/11/18 SAS
J6166-FS	NAWC-050718-RW-356	275.0	1	C	05/11/18 SAS
J6168-FS	NAWC-050718-RW-289	280.0	1	C	05/11/18 SAS
J6170-FS	WGNA-050718-DUP-35	280.0	1	C	05/11/18 SAS

**Comments:**

Sample ID:	Comments:
CQ755PB-FS	1.27g Trizma(170526-01) weighed on BAL-009
CQ756LCS-FS	1.23g Trizma(170526-01) weighed on BAL-009

Samples Assigned By

Stephanie Schultz

Date :

May 9, 2018

\* - "C" = Sample is Consumed



It can be done

## BATTELLE - NORWELL OPERATIONS SURROGATE SPIKE FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	Standard ID	Type	Vial No.	Vol Added (uL)	Date Spiked/ Spiked By	Witn'd By	Comment
CQ755PB-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
CQ756LCS-FS	JV41	LCS/MS	1	75	05/10/18 SAS	JCT	NA
CQ756LCS-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6148-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6150-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6152-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6154-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6156-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6158-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6160-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6162-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6164-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6166-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6168-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6170-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA

## Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
JV41	Pipette	I0793912B
JV60	Pipette	I0793912B



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE EXTRACTION FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	1st Extraction	2nd Extraction	3rd Extraction	Conc. ID	Turbo °C	Turbo PSI	KD °C	Comment
CQ755PB-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
CQ756LCS-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6148-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6150-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6152-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6154-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6156-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6158-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6160-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6162-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6164-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6166-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6168-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6170-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA

**Solvents/Reagent Preparations:**

Name	ID	Expires	Lot No	Procedure	Comments
Pre-packed SPE Column	RP-180510-1	05/10/18	S214-0071	Pre-packed SPE Column	

**Solvents/Reagents:**

Name	Lot No	Comments
Methanol (HPLC) (180227-02)	178212	



It can be done

## BATTELLE - NORWELL OPERATIONS INTERNAL STANDARD SPIKING FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW****(N/A Fraction)**

Extract Id	Extr. Vol. (uL)	Added (uL)	Std. Id	Accm . (uL)	Vial No.	Pre Inj. Vol. (uL)^	Final Dilution*	Date Spiked/ Spiked By	Witn'd By
CQ755PB-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
CQ756LCS-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6148-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6150-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6152-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6154-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6156-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6158-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6160-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6162-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6164-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6166-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6168-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6170-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT

\* - Final Dilution is any HPLC, dilutions, or other manipulation

^ - Pre Injection Volume (PIV) includes any RIS spikes.





It can be done

## BATTELLE - NORWELL OPERATIONS INTERNAL STANDARD SPIKING FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW****(N/A Fraction)**

Extract Id	Extr. Vol. (uL)	Added (uL)	Std. Id	Accm . (uL)	Vial No.	Pre Inj. Vol. (uL)^	Final Dilution *	Date Spiked/ Spiked By	Witn'd By
------------	-----------------	------------	---------	-------------	----------	---------------------	------------------	------------------------	-----------

Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
JV59	Pipette	I0793912B

<b>Extract Id:</b>	<b>Comments:</b>
CQ755PB-FS	Samples reconstituted in 96/4 Methanol/Milli-q Water (RP-180511-2)

\* - Final Dilution is any HPLC, dilutions, or other manipulation

^ - Pre Injection Volume (PIV) includes any RIS spikes.



It can be done

## BATTELLE - NORWELL OPERATIONS PREPARATION EXTRACT SPLIT FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					
CQ755PB-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
CQ756LCS-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6148-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6150-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6152-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6154-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6156-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6158-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6160-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6162-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6164-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6166-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6168-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS
J6170-FS	0	--	5/10/2018 10:53:00 AM	NA		NA	NA	1.000	1.000	05/10/18 SAS

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] \* [Aliquot Weight (mg)]

Dilution Factor = [Sample Volume (uL) / Aliquot Volume (uL)] \* Prior Dilution Factor

\* - "C" = Extract is Consumed



It can be done

**BATTELLE - NORWELL OPERATIONS  
PREPARATION EXTRACT SPLIT FORM**

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**

100117920-  
WE04

**18-0313**

**WE04 PFAS Analysis**

**DW**

Extract		*	Extract Date	Source		Initial Extract Vol (uL)	Extract Split	Extract Split	Total Dilution	Date/Initials
Name	#			Name	#					

Total Oil = [Sample Volume (uL) / Aliquot Volume (uL)] \* [Aliquot Weight (mg)]

Dilution Factor = [Sample Volume (uL) / Aliquot Volume (uL)] \* Prior Dilution Factor

\* - "C" = Extract is Consumed



It can be done

## BATTELLE - NORWELL OPERATIONS EXTRACT - INSTRUMENT FACILITY CUSTODY PAGE

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

<b>Purpose:</b> LC-MS/MS TRANSFER		<b>Last Activity:</b> Prep->Inst	
<b>Relinquished On/By:</b> May 14 2018 10:33AM SAS		<b>Received On/By:</b> May 14 2018 2:46PM DMS	
<b>Relinquished From:</b> Sample Preparation: NA		<b>Received Location:</b> LC Laboratory: NA	
<b>Relinquish Comment:</b> NA		<b>Received Comment:</b> NA	

No.	BDO-ID:	PIV:	DF:	Condition:	Custody Comment:
1	CQ755PB-FS(0)	1000	1	Intact	NA
2	CQ756LCS-FS(0)	1000	1	Intact	NA
3	J6148-FS(0)	1000	1	Intact	NA
4	J6150-FS(0)	1000	1	Intact	NA
5	J6152-FS(0)	1000	1	Intact	NA
6	J6154-FS(0)	1000	1	Intact	NA
7	J6156-FS(0)	1000	1	Intact	NA
8	J6158-FS(0)	1000	1	Intact	NA
9	J6160-FS(0)	1000	1	Intact	NA
10	J6162-FS(0)	1000	1	Intact	NA
11	J6164-FS(0)	1000	1	Intact	NA
12	J6166-FS(0)	1000	1	Intact	NA
13	J6168-FS(0)	1000	1	Intact	NA
14	J6170-FS(0)	1000	1	Intact	NA

**Total Extracts:** 14



It can be done

**BATTELLE - NORWELL OPERATIONS  
MISCELLANEOUS DOCUMENTATION FORM**

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**

100117920-  
WE04

**18-0313**

**WE04 PFAS Analysis**

**DW**

---

Entered By:

On:

---

---

Task Leader Approval:

On:

SupervisorApproval:

On:

PM Approval:

On:

---



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE SPECIFIC COMMENTS

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID:	Comment:	Date/Initials:
CQ755PB-FS	Sample extraction began at 10:53am for all samples.	05/10/18 SAS
CQ755PB-FS	Sample extraction ended at 11:28am	05/10/18 SAS
CQ756LCS-FS	Sample extraction ended at 11:20am	05/10/18 SAS
J6148-FS	Sample extraction ended at 11:25am	05/10/18 SAS
J6150-FS	Sample extraction ended at 11:23am	05/10/18 SAS
J6152-FS	Sample extraction ended at 11:25am	05/10/18 SAS
J6154-FS	Sample extraction ended at 11:25am	05/10/18 SAS
J6156-FS	Sample extraction ended at 11:25am	05/10/18 SAS
J6158-FS	Sample extraction ended at 11:29am	05/10/18 SAS
J6160-FS	Sample extraction ended at 11:34am	05/10/18 SAS
J6162-FS	Sample extraction ended at 11:27am	05/10/18 SAS
J6164-FS	Sample extraction ended at 11:27am	05/10/18 SAS
J6166-FS	Sample extraction ended at 11:27am	05/10/18 SAS
J6168-FS	Sample extraction ended at 11:27am	05/10/18 SAS
J6170-FS	Sample extraction ended at 11:29am	05/10/18 SAS

# Analytical Calibrations



Sequence Report

Created with Analyst Reporter  
 Printed: 17/05/2018 11:43:52 AM

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	MeOH		5/14/2018 11:20:57 AM	5-0371.dam	18-0313.wiff
2	JV64	L1	5/14/2018 11:29:54 AM	5-0371.dam	18-0313.wiff
3	JV65	L2	5/14/2018 11:38:49 AM	5-0371.dam	18-0313.wiff
4	JV66	L3	5/14/2018 11:47:45 AM	5-0371.dam	18-0313.wiff
5	JV67	L4	5/14/2018 11:56:42 AM	5-0371.dam	18-0313.wiff
6	JV68	L5	5/14/2018 12:05:38 PM	5-0371.dam	18-0313.wiff
7	JV69	L6	5/14/2018 12:14:33 PM	5-0371.dam	18-0313.wiff
8	JV70	L7	5/14/2018 12:23:30 PM	5-0371.dam	18-0313.wiff
9	JV71	L8	5/14/2018 12:32:26 PM	5-0371.dam	18-0313.wiff
10	JV72	L9	5/14/2018 12:41:22 PM	5-0371.dam	18-0313.wiff
11	JV63 ICC	ICC	5/14/2018 12:50:20 PM	5-0371.dam	18-0313.wiff
1	MeOH		5/14/2018 12:59:16 PM	5-0371.dam	18-0313.wiff
12	CQ755PB-FS(0)	Procedural Blank	5/14/2018 1:08:14 PM	5-0371.dam	18-0313.wiff
13	CQ756LCS-FS(0)	Laboratory Control Sample	5/14/2018 1:17:09 PM	5-0371.dam	18-0313.wiff
14	J6148-FS(0)	NAWC-050718-RW-316	5/14/2018 1:26:05 PM	5-0371.dam	18-0313.wiff
15	J6150-FS(0)	NAWC-050718-RW-180	5/14/2018 1:35:01 PM	5-0371.dam	18-0313.wiff
16	J6152-FS(0)	NAWC-050718-RW-275	5/14/2018 1:43:56 PM	5-0371.dam	18-0313.wiff
17	J6154-FS(0)	NAWC-050718-RW-145	5/14/2018 1:52:52 PM	5-0371.dam	18-0313.wiff
18	J6156-FS(0)	NAWC-050718-RW-357	5/14/2018 2:01:46 PM	5-0371.dam	18-0313.wiff
19	J6158-FS(0)	NAWC-050718-RW-162	5/14/2018 2:10:41 PM	5-0371.dam	18-0313.wiff
7	JV69 CCV	CCV	5/14/2018 2:19:36 PM	5-0371.dam	18-0313.wiff
1	MeOH		5/14/2018 2:28:30 PM	5-0371.dam	18-0313.wiff
20	J6160-FS(0)	WGNA-050718-RW-0800	5/14/2018 2:37:26 PM	5-0371.dam	18-0313.wiff
21	J6162-FS(0)	WGNA-050718-RW-0335	5/14/2018 2:46:22 PM	5-0371.dam	18-0313.wiff
22	J6164-FS(0)	WGNA-050718-RW-3556	5/14/2018 2:55:18 PM	5-0371.dam	18-0313.wiff
23	J6166-FS(0)	NAWC-050718-RW-356	5/14/2018 3:04:13 PM	5-0371.dam	18-0313.wiff
24	J6168-FS(0)	NAWC-050718-RW-	5/14/2018 3:13:09	5-0371.dam	18-0313.wiff

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
		289	PM		
25	J6170-FS(0)	WGNA-050718-DUP-35	5/14/2018 3:22:03 PM	5-0371.dam	18-0313.wiff
6	JV68 CCV	CCV	5/14/2018 3:30:58 PM	5-0371.dam	18-0313.wiff




## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFBS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	298.9 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.74054 x + 0.01745$  ( $r = 0.99950$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.15	22.804779	103.0
3	JV65	L2	True	44.30	38.779769	87.5
4	JV66	L3	True	88.60	83.589043	94.3
5	JV67	L4	True	221.50	248.105496	112.0
6	JV68	L5	True	443.00	480.862156	108.6
7	JV69	L6	True	885.00	867.003654	98.0
8	JV70	L7	True	2212.50	2093.047764	94.6
9	JV71	L8	True	4425.00	4522.222161	102.2
10	JV72	L9	True	8850.00	8835.635177	99.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFBS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	298.9 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.22816x + 0.01086$  (r = 0.99931) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.15	20.300377	91.7
3	JV65	L2	True	44.30	46.075180	104.0
4	JV66	L3	True	88.60	81.857187	92.4
5	JV67	L4	True	221.50	251.465679	113.5
6	JV68	L5	True	443.00	467.377936	105.5
7	JV69	L6	True	885.00	869.552481	98.3
8	JV70	L7	True	2212.50	2036.632463	92.1
9	JV71	L8	True	4425.00	4512.772938	102.0
10	JV72	L9	True	8850.00	8906.015758	100.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	313.0 / 269.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.87836x + 0.46617$  ( $r = 0.99905$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	False	50.00	2.575840	5.2
4	JV66	L3	True	100.00	78.986357	79.0
5	JV67	L4	True	250.00	251.390112	100.6
6	JV68	L5	True	500.00	541.056673	108.2
7	JV69	L6	True	1000.00	1089.329496	108.9
8	JV70	L7	True	2500.00	2659.394351	106.4
9	JV71	L8	True	5000.00	4963.910989	99.3
10	JV72	L9	True	10000.00	9765.932022	97.7






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	313.0 / 119.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06343 x + 0.00880$  ( $r = 0.99997$ ) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	11.923343	47.7
3	JV65	L2	False	50.00	45.290432	90.6
4	JV66	L3	True	100.00	100.733035	100.7
5	JV67	L4	True	250.00	251.163821	100.5
6	JV68	L5	True	500.00	492.679342	98.5
7	JV69	L6	True	1000.00	1000.547914	100.1
8	JV70	L7	True	2500.00	2487.113791	99.5
9	JV71	L8	True	5000.00	5054.860411	101.1
10	JV72	L9	True	10000.00	9962.901685	99.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHpA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	363.0 / 319.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.84607 x + 0.04250$  ( $r = 0.99981$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	26.704376	106.8
3	JV65	L2	True	50.00	42.475329	85.0
4	JV66	L3	True	100.00	108.532311	108.5
5	JV67	L4	True	250.00	243.123776	97.3
6	JV68	L5	True	500.00	493.515333	98.7
7	JV69	L6	True	1000.00	1007.476703	100.8
8	JV70	L7	True	2500.00	2589.189046	103.6
9	JV71	L8	True	5000.00	5029.188658	100.6
10	JV72	L9	True	10000.00	9884.794468	98.9








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHpA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	363.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.01637 x + 0.00763$  (r = 0.99929) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	N/A	N/A
3	JV65	L2	False	50.00	21.602557	43.2
4	JV66	L3	True	100.00	85.606667	85.6
5	JV67	L4	True	250.00	285.895686	114.4
6	JV68	L5	True	500.00	504.096919	100.8
7	JV69	L6	True	1000.00	938.783794	93.9
8	JV70	L7	True	2500.00	2665.862470	106.6
9	JV71	L8	True	5000.00	5000.525187	100.0
10	JV72	L9	True	10000.00	9869.229277	98.7






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	399.0 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.84495x + 0.00149$  ( $r = 0.99992$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	23.034700	101.0
3	JV65	L2	True	45.60	44.607323	97.8
4	JV66	L3	True	91.20	90.397278	99.1
5	JV67	L4	True	228.00	241.239806	105.8
6	JV68	L5	True	456.00	439.233831	96.3
7	JV69	L6	True	912.00	910.143123	99.8
8	JV70	L7	True	2280.00	2258.577450	99.1
9	JV71	L8	True	4560.00	4626.551146	101.5
10	JV72	L9	True	9120.00	9081.815342	99.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	399.0 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.24827 x + 0.00372$  (r = 0.99990) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	26.747945	117.3
3	JV65	L2	True	45.60	39.964480	87.6
4	JV66	L3	True	91.20	89.427525	98.1
5	JV67	L4	True	228.00	227.854457	99.9
6	JV68	L5	True	456.00	453.922143	99.5
7	JV69	L6	True	912.00	893.198688	97.9
8	JV70	L7	True	2280.00	2238.129172	98.2
9	JV71	L8	True	4560.00	4621.692989	101.4
10	JV72	L9	True	9120.00	9124.662600	100.1






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	413.0 / 369.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.96808x + 0.19816$  ( $r = 0.99938$ ) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	11.153654	44.6
3	JV65	L2	True	50.00	46.676599	93.4
4	JV66	L3	True	100.00	95.581325	95.6
5	JV67	L4	True	250.00	245.690478	98.3
6	JV68	L5	True	500.00	519.420773	103.9
7	JV69	L6	True	1000.00	1037.136408	103.7
8	JV70	L7	True	2500.00	2669.265130	106.8
9	JV71	L8	True	5000.00	5055.859138	101.1
10	JV72	L9	True	10000.00	9730.370148	97.3






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	413.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06473 x + 0.01078$  (r = 0.99767) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	22.005638	88.0
3	JV65	L2	True	50.00	46.540813	93.1
4	JV66	L3	True	100.00	82.992508	83.0
5	JV67	L4	True	250.00	261.053614	104.4
6	JV68	L5	True	500.00	517.072576	103.4
7	JV69	L6	True	1000.00	1070.067858	107.0
8	JV70	L7	True	2500.00	2733.741648	109.4
9	JV71	L8	True	5000.00	5284.814305	105.7
10	JV72	L9	True	10000.00	9403.716678	94.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFNA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	463.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.92776 x + 0.06645$  (r = 0.99904) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.411469	97.7
3	JV65	L2	True	50.00	41.743202	83.5
4	JV66	L3	True	100.00	98.979777	99.0
5	JV67	L4	True	250.00	251.354806	100.5
6	JV68	L5	True	500.00	542.040106	108.4
7	JV69	L6	True	1000.00	1053.161797	105.3
8	JV70	L7	True	2500.00	2705.262997	108.2
9	JV71	L8	True	5000.00	5033.084124	100.7
10	JV72	L9	True	10000.00	9674.961722	96.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFNA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	463.0 / 219.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.27039x + 0.02695$  ( $r = 0.99905$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	20.218296	80.9
3	JV65	L2	True	50.00	45.960397	91.9
4	JV66	L3	True	100.00	101.932105	101.9
5	JV67	L4	True	250.00	255.674001	102.3
6	JV68	L5	True	500.00	572.533150	114.5
7	JV69	L6	True	1000.00	1042.596703	104.3
8	JV70	L7	True	2500.00	2675.736272	107.0
9	JV71	L8	True	5000.00	5010.507489	100.2
10	JV72	L9	True	10000.00	9699.841587	97.0








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	499.0 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.16765x + 0.02406$  ( $r = 0.99981$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	23.361262	100.9
3	JV65	L2	True	46.30	42.153986	91.1
4	JV66	L3	True	92.60	88.414386	95.5
5	JV67	L4	True	231.50	248.496910	107.3
6	JV68	L5	True	463.00	472.188043	102.0
7	JV69	L6	True	925.60	954.991600	103.2
8	JV70	L7	True	2314.00	2301.735042	99.5
9	JV71	L8	True	4628.00	4717.827738	101.9
10	JV72	L9	True	9256.00	9130.981033	98.7






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	499.0 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.22592 x + 0.00424$  (r = 0.99977) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	25.835005	111.6
3	JV65	L2	True	46.30	37.607900	81.2
4	JV66	L3	True	92.60	92.009960	99.4
5	JV67	L4	True	231.50	248.837656	107.5
6	JV68	L5	True	463.00	455.023640	98.3
7	JV69	L6	True	925.60	923.897296	99.8
8	JV70	L7	True	2314.00	2352.785599	101.7
9	JV71	L8	True	4628.00	4719.092631	102.0
10	JV72	L9	True	9256.00	9125.060312	98.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	513.0 / 469.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.99919x + 0.21544$  ( $r = 0.99832$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	10.368703	41.5
3	JV65	L2	False	50.00	31.340650	62.7
4	JV66	L3	True	100.00	85.661929	85.7
5	JV67	L4	True	250.00	230.445171	92.2
6	JV68	L5	True	500.00	529.199399	105.8
7	JV69	L6	True	1000.00	1093.183122	109.3
8	JV70	L7	True	2500.00	2780.129248	111.2
9	JV71	L8	True	5000.00	4948.282902	99.0
10	JV72	L9	True	10000.00	9683.098229	96.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	513.0 / 219.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.03953 x + 0.01934$  (r = 0.99980) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	False	50.00	< 0	N/A
4	JV66	L3	True	100.00	109.576482	109.6
5	JV67	L4	True	250.00	225.898181	90.4
6	JV68	L5	True	500.00	485.438921	97.1
7	JV69	L6	True	1000.00	1002.397258	100.2
8	JV70	L7	True	2500.00	2575.357213	103.0
9	JV71	L8	True	5000.00	5020.912753	100.4
10	JV72	L9	True	10000.00	9930.419191	99.3






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFUnA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	563.0 / 519.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.09588x + 0.04114$  ( $r = 0.99970$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.333102	97.3
3	JV65	L2	True	50.00	47.670455	95.3
4	JV66	L3	True	100.00	97.714264	97.7
5	JV67	L4	True	250.00	255.004410	102.0
6	JV68	L5	True	500.00	512.965489	102.6
7	JV69	L6	True	1000.00	1017.920510	101.8
8	JV70	L7	True	2500.00	2640.922188	105.6
9	JV71	L8	True	5000.00	4930.392260	98.6
10	JV72	L9	True	10000.00	9898.077320	99.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFUnA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	563.0 / 269.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.05167 x + 0.00535$  (r = 0.99953) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	54.213071	216.9
3	JV65	L2	True	50.00	42.011553	84.0
4	JV66	L3	True	100.00	100.910898	100.9
5	JV67	L4	True	250.00	250.795438	100.3
6	JV68	L5	True	500.00	562.545175	112.5
7	JV69	L6	True	1000.00	1000.374015	100.0
8	JV70	L7	True	2500.00	2617.419524	104.7
9	JV71	L8	True	5000.00	4924.516855	98.5
10	JV72	L9	True	10000.00	9901.426543	99.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFD <sub>o</sub> A_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	613.0 / 569.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.09433x + 0.06074$  ( $r = 0.99915$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.809504	99.2
3	JV65	L2	True	50.00	39.835097	79.7
4	JV66	L3	True	100.00	101.286712	101.3
5	JV67	L4	True	250.00	254.074231	101.6
6	JV68	L5	True	500.00	526.830587	105.4
7	JV69	L6	True	1000.00	1087.157146	108.7
8	JV70	L7	True	2500.00	2674.514086	107.0
9	JV71	L8	True	5000.00	4994.806618	99.9
10	JV72	L9	True	10000.00	9721.686021	97.2








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFD <sub>o</sub> A_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	613.0 / 319.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.17249x + 0.01513$  ( $r = 0.99963$ ) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	26.532344	106.1
3	JV65	L2	True	50.00	42.462843	84.9
4	JV66	L3	True	100.00	96.476731	96.5
5	JV67	L4	True	250.00	263.226499	105.3
6	JV68	L5	True	500.00	506.075964	101.2
7	JV69	L6	True	1000.00	1020.038165	102.0
8	JV70	L7	True	2500.00	2621.333582	104.9
9	JV71	L8	True	5000.00	5061.671458	101.2
10	JV72	L9	True	10000.00	9787.182415	97.9






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTrDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	663.0 / 619.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.07282x + 0.05063$  ( $r = 0.99932$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.203429	100.8
3	JV65	L2	True	50.00	43.998106	88.0
4	JV66	L3	True	100.00	97.376116	97.4
5	JV67	L4	True	250.00	245.734116	98.3
6	JV68	L5	True	500.00	515.430124	103.1
7	JV69	L6	True	1000.00	1085.235268	108.5
8	JV70	L7	True	2500.00	2661.601187	106.5
9	JV71	L8	True	5000.00	4994.249530	99.9
10	JV72	L9	True	10000.00	9756.172124	97.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTrDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	663.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06800 x + 0.01002$  (r = 0.99958) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	27.141621	108.6
3	JV65	L2	True	50.00	52.520940	105.0
4	JV66	L3	True	100.00	80.738653	80.7
5	JV67	L4	True	250.00	254.680571	101.9
6	JV68	L5	True	500.00	475.884832	95.2
7	JV69	L6	True	1000.00	1065.280060	106.5
8	JV70	L7	True	2500.00	2604.777878	104.2
9	JV71	L8	True	5000.00	4924.491415	98.5
10	JV72	L9	True	10000.00	9939.484030	99.4






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTeDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	713.0 / 669.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.04679x + 0.07024$  (r = 0.99948) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	23.929635	95.7
3	JV65	L2	True	50.00	43.700214	87.4
4	JV66	L3	True	100.00	99.162152	99.2
5	JV67	L4	True	250.00	246.757950	98.7
6	JV68	L5	True	500.00	520.679296	104.1
7	JV69	L6	True	1000.00	1080.886811	108.1
8	JV70	L7	True	2500.00	2628.108012	105.1
9	JV71	L8	True	5000.00	4957.832384	99.2
10	JV72	L9	True	10000.00	9822.873181	98.2






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTeDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	713.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.04914 x + 0.00899$  (r = 0.99930) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	1.189372	4.8
3	JV65	L2	True	50.00	53.808904	107.6
4	JV66	L3	True	100.00	81.366098	81.4
5	JV67	L4	True	250.00	242.225032	96.9
6	JV68	L5	True	500.00	517.332905	103.5
7	JV69	L6	True	1000.00	1064.697063	106.5
8	JV70	L7	True	2500.00	2674.759141	107.0
9	JV71	L8	True	5000.00	4954.132013	99.1
10	JV72	L9	True	10000.00	9811.678844	98.1






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NMeFOSAA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	570.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.94846 x + 0.02409$  (r = 0.99947) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.648457	98.6
3	JV65	L2	True	50.00	52.409646	104.8
4	JV66	L3	True	100.00	100.762487	100.8
5	JV67	L4	True	250.00	253.928545	101.6
6	JV68	L5	True	500.00	452.857852	90.6
7	JV69	L6	True	1000.00	979.361667	97.9
8	JV70	L7	True	2500.00	2623.979649	105.0
9	JV71	L8	True	5000.00	5141.553388	102.8
10	JV72	L9	True	10000.00	9795.498310	98.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NMeFOSAA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	570.0 / 512.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.58675x + 0.04231$  (r = 0.99949) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	True	50.00	59.265228	118.5
4	JV66	L3	True	100.00	88.696165	88.7
5	JV67	L4	True	250.00	244.943682	98.0
6	JV68	L5	True	500.00	476.879086	95.4
7	JV69	L6	True	1000.00	970.886314	97.1
8	JV70	L7	True	2500.00	2480.966798	99.2
9	JV71	L8	True	5000.00	5230.915871	104.6
10	JV72	L9	True	10000.00	9847.446856	98.5








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NEtFOSAA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	584.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.90138x + 0.01437$  ( $r = 0.99896$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	21.319053	85.3
3	JV65	L2	True	50.00	64.299976	128.6
4	JV66	L3	True	100.00	76.835912	76.8
5	JV67	L4	True	250.00	265.899538	106.4
6	JV68	L5	True	500.00	514.051353	102.8
7	JV69	L6	True	1000.00	951.455825	95.2
8	JV70	L7	True	2500.00	2567.794899	102.7
9	JV71	L8	True	5000.00	5262.702481	105.3
10	JV72	L9	True	10000.00	9700.640962	97.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NEtFOSAA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	584.0 / 483.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06071 x + 0.00441$  (r = 0.99828) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	N/A	N/A
3	JV65	L2	False	50.00	65.966292	131.9
4	JV66	L3	False	100.00	94.303580	94.3
5	JV67	L4	True	250.00	251.980572	100.8
6	JV68	L5	True	500.00	430.232147	86.1
7	JV69	L6	True	1000.00	1043.578594	104.4
8	JV70	L7	True	2500.00	2744.280900	109.8
9	JV71	L8	True	5000.00	5123.296865	102.5
10	JV72	L9	True	10000.00	9656.630922	96.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 17/05/2018 1:54:09 PM

<b>Analyte Name</b>	13C2-PFHxA	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	315.0 / 270.0	<b>Result Table</b>	18-0313_SIS
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.87369 x$  (std. dev. = 0.06436) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	100.00	110.844174	110.8
3	JV65	L2	True	100.00	103.204010	103.2
4	JV66	L3	True	100.00	105.369867	105.4
5	JV67	L4	True	100.00	89.287891	89.3
6	JV68	L5	True	100.00	97.060668	97.1
7	JV69	L6	True	100.00	106.310194	106.3
8	JV70	L7	True	100.00	100.563720	100.6
9	JV71	L8	True	100.00	97.800376	97.8
10	JV72	L9	True	100.00	89.559100	89.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 17/05/2018 1:54:09 PM

<b>Analyte Name</b>	13C2-PFDA	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	515.0 / 470.0	<b>Result Table</b>	18-0313_SIS
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.04874 x$  (std. dev. = 0.06072) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	100.00	107.538910	107.5
3	JV65	L2	True	100.00	95.325906	95.3
4	JV66	L3	True	100.00	101.867049	101.9
5	JV67	L4	True	100.00	89.177646	89.2
6	JV68	L5	True	100.00	102.918905	102.9
7	JV69	L6	True	100.00	101.874206	101.9
8	JV70	L7	True	100.00	99.661495	99.7
9	JV71	L8	True	100.00	106.078813	106.1
10	JV72	L9	True	100.00	95.557070	95.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 17/05/2018 1:54:09 PM

<b>Analyte Name</b>	d5-EtFOSAA	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	589.0 / 419.0	<b>Result Table</b>	18-0313_SIS
<b>Internal Standard</b>	D3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.07146 x$  (std. dev. = 0.08115) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	400.00	395.628105	98.9
3	JV65	L2	True	400.00	421.352731	105.3
4	JV66	L3	True	400.00	441.863763	110.5
5	JV67	L4	True	400.00	415.907008	104.0
6	JV68	L5	True	400.00	365.942783	91.5
7	JV69	L6	True	400.00	401.971401	100.5
8	JV70	L7	True	400.00	410.733534	102.7
9	JV71	L8	True	400.00	406.137800	101.5
10	JV72	L9	True	400.00	340.462875	85.1





Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.49	7960.07	22.804779	240.9	true
PFBS_2	298.9 / 99.0	1.49	2817.25	20.300377	126.5	true
PFHxA_1	313.0 / 269.0	1.79	11169.61	< 0	129.9	true
PFHxA_2	313.0 / 119.0	1.77	567.61	11.923343	61.0	true
PFHpA_1	363.0 / 319.0	2.12	9314.42	26.704376	36.2	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	2.14	7230.87	23.034700	75.5	false
PFHxS_2	399.0 / 99.0	2.13	2802.83	26.747945	52.0	false
PFOA_1	413.0 / 369.0	2.50	10622.56	11.153654	43.6	true
PFOA_2	413.0 / 169.0	2.49	868.43	22.005638	28.5	true
PFNA_1	463.0 / 419.0	2.89	10164.04	24.411469	42.6	true
PFNA_2	463.0 / 219.0	2.88	2831.82	20.218296	46.1	true
PFOS_1	499.0 / 80.0	2.88	12427.11	23.361262	63.0	false
PFOS_2	499.0 / 99.0	2.89	2564.10	25.835005	70.1	false
PFDA_1	513.0 / 469.0	3.24	11070.16	10.368703	49.8	false
PFDA_2	513.0 / 219.0	3.26	648.37	< 0	25.8	false
PFUnA_1	563.0 / 519.0	3.56	10680.04	24.333102	46.6	false
PFUnA_2	563.0 / 269.0	3.60	1157.79	54.213071	19.1	true
PFDoA_1	613.0 / 569.0	3.85	11528.04	24.809504	58.3	false
PFDoA_2	613.0 / 319.0	3.85	2113.12	26.532344	47.0	false
PFTTrDA_1	663.0 / 619.0	4.11	11138.66	25.203429	99.3	false
PFTTrDA_2	663.0 / 169.0	4.10	987.98	27.141621	50.3	false
PFTeDA_1	713.0 / 669.0	4.33	11128.90	23.929635	126.7	false
PFTeDA_2	713.0 / 169.0	4.32	332.33	1.189372	24.5	false
NMeFOSAA_1	570.0 / 419.0	3.38	2968.73	24.648457	150.4	false
NMeFOSAA_2	570.0 / 512.0	3.37	749.48	< 0	28.4	false
NEtFOSAA_1	584.0 / 419.0	3.56	2244.94	21.319053	82.0	false
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true





## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:34 PM

<b>Sample Name</b>	JV64	<b>Injection Vial</b>	2
<b>Sample ID</b>	L1	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:29:54	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.75	33602.97	110.844174	917.1	false
13C2-PFDA	515.0 / 470.0	3.23	39133.03	107.538910	566.8	false
d5-EtFOSAA	589.0 / 419.0	3.54	38117.48	395.628105	240.2	false



Sample Name	JV65	Injection Vial	3
Sample ID	L2	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:38:49	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.49	11618.68	38.779769	307.5	true
PFBS_2	298.9 / 99.0	1.48	4695.68	46.075180	167.2	false
PFHxA_1	313.0 / 269.0	1.79	16410.25	2.575840	156.1	true
PFHxA_2	313.0 / 119.0	1.79	1259.79	45.290432	84.9	true
PFHpA_1	363.0 / 319.0	2.12	13492.10	42.475329	43.8	true
PFHpA_2	363.0 / 169.0	2.11	374.77	21.602557	31.1	false
PFHxS_1	399.0 / 80.0	2.14	13131.65	44.607323	104.5	false
PFHxS_2	399.0 / 99.0	2.13	3786.47	39.964480	67.5	false
PFOA_1	413.0 / 369.0	2.50	21823.50	46.676599	67.0	false
PFOA_2	413.0 / 169.0	2.51	1373.43	46.540813	68.7	false
PFNA_1	463.0 / 419.0	2.88	15232.88	41.743202	48.9	false
PFNA_2	463.0 / 219.0	2.88	5076.79	45.960397	76.0	false
PFOS_1	499.0 / 80.0	2.87	19335.81	42.153986	92.6	false
PFOS_2	499.0 / 99.0	2.87	3346.10	37.607900	86.3	false
PFDA_1	513.0 / 469.0	3.24	17746.35	31.340650	60.3	false
PFDA_2	513.0 / 219.0	3.22	641.07	< 0	37.4	false
PFUnA_1	563.0 / 519.0	3.56	18919.95	47.670455	68.1	false
PFUnA_2	563.0 / 269.0	3.50	908.57	42.011553	26.9	false
PFDoA_1	613.0 / 569.0	3.85	16674.57	39.835097	72.2	false
PFDoA_2	613.0 / 319.0	3.84	2967.12	42.462843	56.5	false
PFTTrDA_1	663.0 / 619.0	4.10	17546.86	43.998106	122.0	false
PFTTrDA_2	663.0 / 169.0	4.09	1535.33	52.520940	63.2	false
PFTeDA_1	713.0 / 669.0	4.33	17716.16	43.700214	211.8	false
PFTeDA_2	713.0 / 169.0	4.32	1189.70	53.808904	87.0	false
NMeFOSAA_1	570.0 / 419.0	3.39	4486.71	52.409646	166.0	false
NMeFOSAA_2	570.0 / 512.0	3.41	3908.56	59.265228	81.0	false
NEtFOSAA_1	584.0 / 419.0	3.55	4816.55	64.299976	187.0	true
NEtFOSAA_2	584.0 / 483.0	3.60	436.15	65.966292	24.0	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:36 PM

<b>Sample Name</b>	JV65	<b>Injection Vial</b>	3
<b>Sample ID</b>	L2	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:38:49	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.74	30272.18	103.204010	677.2	false
13C2-PFDA	515.0 / 470.0	3.22	33563.80	95.325906	501.9	false
d5-EtFOSAA	589.0 / 419.0	3.54	34132.22	421.352731	182.3	false



Sample Name	JV66	Injection Vial	4
Sample ID	L3	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:47:45	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.49	21166.84	83.589043	428.6	false
PFBS_2	298.9 / 99.0	1.48	6894.77	81.857187	203.3	false
PFHxA_1	313.0 / 269.0	1.80	34391.90	78.986357	249.3	true
PFHxA_2	313.0 / 119.0	1.78	2155.28	100.733035	164.1	true
PFHpA_1	363.0 / 319.0	2.12	28485.92	108.532311	69.0	false
PFHpA_2	363.0 / 169.0	2.11	641.57	85.606667	51.2	false
PFHxS_1	399.0 / 80.0	2.13	24298.38	90.397278	133.7	false
PFHxS_2	399.0 / 99.0	2.14	7362.03	89.427525	98.6	false
PFOA_1	413.0 / 369.0	2.50	33310.09	95.581325	87.1	false
PFOA_2	413.0 / 169.0	2.50	1912.46	82.992508	80.9	false
PFNA_1	463.0 / 419.0	2.88	29197.03	98.979777	77.6	false
PFNA_2	463.0 / 219.0	2.88	8970.62	101.932105	137.0	false
PFOS_1	499.0 / 80.0	2.87	34843.98	88.414386	98.7	false
PFOS_2	499.0 / 99.0	2.87	6960.83	92.009960	129.3	false
PFDA_1	513.0 / 469.0	3.23	31765.19	85.661929	94.2	false
PFDA_2	513.0 / 219.0	3.23	1857.53	109.576482	91.5	false
PFUnA_1	563.0 / 519.0	3.55	32969.06	97.714264	82.7	false
PFUnA_2	563.0 / 269.0	3.56	1704.75	100.910898	51.4	true
PFDaA_1	613.0 / 569.0	3.84	34664.50	101.286712	115.9	false
PFDaA_2	613.0 / 319.0	3.84	5382.73	96.476731	94.0	false
PFTTrDA_1	663.0 / 619.0	4.10	32474.88	97.376116	171.6	false
PFTTrDA_2	663.0 / 169.0	4.09	1924.80	80.738653	74.2	false
PFTeDA_1	713.0 / 669.0	4.32	32859.26	99.162152	253.0	false
PFTeDA_2	713.0 / 169.0	4.32	1452.17	81.366098	94.1	false
NMeFOSAA_1	570.0 / 419.0	3.38	7366.45	100.762487	228.3	false
NMeFOSAA_2	570.0 / 512.0	3.38	4829.01	88.696165	82.8	false
NEtFOSAA_1	584.0 / 419.0	3.54	5251.99	76.835912	157.5	true
NEtFOSAA_2	584.0 / 483.0	3.47	524.38	94.303580	42.7	true



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:38 PM

<b>Sample Name</b>	JV66	<b>Injection Vial</b>	4
<b>Sample ID</b>	L3	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:47:45	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.75	27295.34	105.369867	803.4	false
13C2-PFDA	515.0 / 470.0	3.22	31675.17	101.867049	484.9	false
d5-EtFOSAA	589.0 / 419.0	3.54	33150.00	441.863763	220.4	false





Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:56:42	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	76416.12	248.105496	798.2	false
PFBS_2	298.9 / 99.0	1.47	24492.22	251.465679	371.1	false
PFHxA_1	313.0 / 269.0	1.76	110703.46	251.390112	415.3	false
PFHxA_2	313.0 / 119.0	1.76	6959.18	251.163821	279.9	true
PFHpA_1	363.0 / 319.0	2.11	86909.74	243.123776	122.9	false
PFHpA_2	363.0 / 169.0	2.11	2252.81	285.895686	122.7	false
PFHxS_1	399.0 / 80.0	2.13	82700.70	241.239806	217.5	false
PFHxS_2	399.0 / 99.0	2.13	23336.84	227.854457	164.6	false
PFOA_1	413.0 / 369.0	2.49	106661.54	245.690478	189.5	false
PFOA_2	413.0 / 169.0	2.49	7441.05	261.053614	203.6	false
PFNA_1	463.0 / 419.0	2.87	99284.04	251.354806	159.7	false
PFNA_2	463.0 / 219.0	2.87	29732.59	255.674001	222.0	false
PFOS_1	499.0 / 80.0	2.86	120273.75	248.496910	205.2	false
PFOS_2	499.0 / 99.0	2.86	23253.63	248.837656	233.5	false
PFDA_1	513.0 / 469.0	3.22	104234.75	230.445171	161.3	false
PFDA_2	513.0 / 219.0	3.22	4496.66	225.898181	145.8	false
PFUnA_1	563.0 / 519.0	3.54	117384.21	255.004410	143.2	false
PFUnA_2	563.0 / 269.0	3.54	5586.16	250.795438	148.5	false
PFDoA_1	613.0 / 569.0	3.83	117610.75	254.074231	185.4	false
PFDoA_2	613.0 / 319.0	3.83	19421.62	263.226499	158.6	false
PFTTrDA_1	663.0 / 619.0	4.08	111226.21	245.734116	237.5	false
PFTTrDA_2	663.0 / 169.0	4.08	7583.57	254.680571	160.0	false
PFTeDA_1	713.0 / 669.0	4.31	109834.15	246.757950	411.4	false
PFTeDA_2	713.0 / 169.0	4.30	5299.77	242.225032	215.9	false
NMeFOSAA_1	570.0 / 419.0	3.37	22637.14	253.928545	429.0	false
NMeFOSAA_2	570.0 / 512.0	3.36	14518.41	244.943682	156.1	false
NEtFOSAA_1	584.0 / 419.0	3.53	22180.63	265.899538	282.9	false
NEtFOSAA_2	584.0 / 483.0	3.53	1541.91	251.980572	96.1	false



## Summary Quant Report

Created with Analyst Reporter  
 Printed: 17/05/2018 2:08:40 PM

<b>Sample Name</b>	JV67	<b>Injection Vial</b>	5
<b>Sample ID</b>	L4	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:56:42	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.74	32292.59	89.287891	1173.5	false
13C2-PFDA	515.0 / 470.0	3.21	38715.02	89.177646	510.4	false
d5-EtFOSAA	589.0 / 419.0	3.52	40274.10	415.907008	224.4	false



Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:05:38	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.48	116121.62	480.862156	978.9	false
PFBS_2	298.9 / 99.0	1.47	35294.80	467.377936	429.5	false
PFHxA_1	313.0 / 269.0	1.76	158021.14	541.056673	432.9	false
PFHxA_2	313.0 / 119.0	1.75	9729.47	492.679342	262.3	true
PFHpA_1	363.0 / 319.0	2.11	127721.74	493.515333	132.6	false
PFHpA_2	363.0 / 169.0	2.11	2729.35	504.096919	131.7	false
PFHxS_1	399.0 / 80.0	2.12	119482.60	439.233831	221.0	false
PFHxS_2	399.0 / 99.0	2.13	36584.23	453.922143	230.6	false
PFOA_1	413.0 / 369.0	2.49	158262.24	519.420773	215.2	false
PFOA_2	413.0 / 169.0	2.49	10460.83	517.072576	275.2	false
PFNA_1	463.0 / 419.0	2.87	154287.21	542.040106	211.2	false
PFNA_2	463.0 / 219.0	2.87	47691.62	572.533150	258.0	false
PFOS_1	499.0 / 80.0	2.86	179520.31	472.188043	221.3	false
PFOS_2	499.0 / 99.0	2.86	33448.63	455.023640	229.5	false
PFDA_1	513.0 / 469.0	3.22	166636.76	529.199399	205.3	false
PFDA_2	513.0 / 219.0	3.23	6395.56	485.438921	307.3	false
PFUnA_1	563.0 / 519.0	3.54	171465.70	512.965489	169.9	false
PFUnA_2	563.0 / 269.0	3.54	8964.01	562.545175	179.2	false
PFDoA_1	613.0 / 569.0	3.83	176413.04	526.830587	229.8	false
PFDoA_2	613.0 / 319.0	3.83	26890.78	506.075964	176.5	false
PFTrDA_1	663.0 / 619.0	4.08	168972.30	515.430124	276.3	false
PFTrDA_2	663.0 / 169.0	4.08	10101.97	475.884832	213.7	false
PFTeDA_1	713.0 / 669.0	4.31	167167.82	520.679296	432.4	false
PFTeDA_2	713.0 / 169.0	4.30	7970.43	517.332905	293.7	false
NMeFOSAA_1	570.0 / 419.0	3.37	33284.91	452.857852	361.9	false
NMeFOSAA_2	570.0 / 512.0	3.36	22490.32	476.879086	198.0	false
NEtFOSAA_1	584.0 / 419.0	3.53	35555.08	514.051353	283.6	false
NEtFOSAA_2	584.0 / 483.0	3.52	2113.27	430.232147	228.6	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:41 PM

<b>Sample Name</b>	JV68	<b>Injection Vial</b>	6
<b>Sample ID</b>	L5	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:05:38	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.74	25677.96	97.060668	993.5	false
13C2-PFDA	515.0 / 470.0	3.21	32683.28	102.918905	628.3	false
d5-EtFOSAA	589.0 / 419.0	3.53	29718.09	365.942783	226.8	false





Sample Name	JV69	Injection Vial	7
Sample ID	L6	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:14:33	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	252025.46	867.003654	1659.1	false
PFBS_2	298.9 / 99.0	1.47	78490.43	869.552481	587.8	false
PFHxA_1	313.0 / 269.0	1.75	376639.02	1089.329496	566.5	false
PFHxA_2	313.0 / 119.0	1.75	24152.30	1000.547914	395.4	true
PFHpA_1	363.0 / 319.0	2.11	321538.42	1007.476703	203.9	false
PFHpA_2	363.0 / 169.0	2.11	6053.77	938.783794	191.1	false
PFHxS_1	399.0 / 80.0	2.12	299696.82	910.143123	307.7	false
PFHxS_2	399.0 / 99.0	2.13	86790.25	893.198688	298.7	false
PFOA_1	413.0 / 369.0	2.49	384297.45	1037.136408	335.6	false
PFOA_2	413.0 / 169.0	2.49	26401.85	1070.067858	323.3	false
PFNA_1	463.0 / 419.0	2.87	369239.58	1053.161797	266.9	false
PFNA_2	463.0 / 219.0	2.87	106823.54	1042.596703	420.9	false
PFOS_1	499.0 / 80.0	2.86	437014.14	954.991600	305.3	false
PFOS_2	499.0 / 99.0	2.86	81772.01	923.897296	273.8	false
PFDA_1	513.0 / 469.0	3.22	418075.93	1093.183122	279.4	false
PFDA_2	513.0 / 219.0	3.22	15597.25	1002.397258	335.9	false
PFUnA_1	563.0 / 519.0	3.54	420249.35	1017.920510	223.6	false
PFUnA_2	563.0 / 269.0	3.54	19603.23	1000.374015	240.2	false
PFDoA_1	613.0 / 569.0	3.83	448832.28	1087.157146	263.6	false
PFDoA_2	613.0 / 319.0	3.83	66608.66	1020.038165	269.5	false
PFTTrDA_1	663.0 / 619.0	4.08	438902.32	1085.235268	351.7	false
PFTTrDA_2	663.0 / 169.0	4.08	27565.40	1065.280060	282.8	false
PFTeDA_1	713.0 / 669.0	4.30	427327.92	1080.886811	611.4	false
PFTeDA_2	713.0 / 169.0	4.30	19976.23	1064.697063	419.9	false
NMeFOSAA_1	570.0 / 419.0	3.37	85161.88	979.361667	414.3	false
NMeFOSAA_2	570.0 / 512.0	3.36	53227.50	970.886314	315.1	false
NEtFOSAA_1	584.0 / 419.0	3.53	78343.08	951.455825	324.5	false
NEtFOSAA_2	584.0 / 483.0	3.53	5908.65	1043.578594	267.7	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:43 PM

<b>Sample Name</b>	JV69	<b>Injection Vial</b>	7
<b>Sample ID</b>	L6	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:14:33	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.73	34862.96	106.310194	1035.7	false
13C2-PFDA	515.0 / 470.0	3.21	40102.06	101.874206	546.5	false
d5-EtFOSAA	589.0 / 419.0	3.52	39081.70	401.971401	238.9	false



Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	477846.79	2093.047764	2211.1	true
PFBS_2	298.9 / 99.0	1.47	143755.92	2036.632463	752.1	true
PFHxA_1	313.0 / 269.0	1.75	701011.98	2659.394351	643.9	false
PFHxA_2	313.0 / 119.0	1.75	46677.58	2487.113791	469.8	true
PFHpA_1	363.0 / 319.0	2.11	645799.76	2589.189046	283.3	false
PFHpA_2	363.0 / 169.0	2.11	13062.90	2665.862470	344.5	false
PFHxS_1	399.0 / 80.0	2.12	586574.34	2258.577450	361.2	false
PFHxS_2	399.0 / 99.0	2.12	171085.72	2238.129172	334.0	false
PFOA_1	413.0 / 369.0	2.49	766139.03	2669.265130	408.4	false
PFOA_2	413.0 / 169.0	2.49	52379.81	2733.741648	440.6	false
PFNA_1	463.0 / 419.0	2.86	740426.75	2705.262997	386.3	false
PFNA_2	463.0 / 219.0	2.86	213664.13	2675.736272	342.5	false
PFOS_1	499.0 / 80.0	2.86	828026.06	2301.735042	341.0	false
PFOS_2	499.0 / 99.0	2.86	163716.29	2352.785599	379.5	false
PFDA_1	513.0 / 469.0	3.22	823672.64	2780.129248	354.6	false
PFDA_2	513.0 / 219.0	3.21	30519.44	2575.357213	381.9	false
PFUnA_1	563.0 / 519.0	3.53	852749.88	2640.922188	272.3	false
PFUnA_2	563.0 / 269.0	3.53	39951.56	2617.419524	266.6	false
PFDoA_1	613.0 / 569.0	3.82	862938.82	2674.514086	359.8	false
PFDoA_2	613.0 / 319.0	3.82	133482.00	2621.333582	294.6	false
PFTTrDA_1	663.0 / 619.0	4.08	841640.24	2661.601187	349.2	false
PFTTrDA_2	663.0 / 169.0	4.08	52409.55	2604.777878	304.0	false
PFTeDA_1	713.0 / 669.0	4.30	811518.80	2628.108012	673.1	false
PFTeDA_2	713.0 / 169.0	4.30	38939.15	2674.759141	528.7	false
NMeFOSAA_1	570.0 / 419.0	3.36	171584.84	2623.979649	424.5	false
NMeFOSAA_2	570.0 / 512.0	3.36	101138.00	2480.966798	340.3	false
NEtFOSAA_1	584.0 / 419.0	3.53	159356.02	2567.794899	365.8	false
NEtFOSAA_2	584.0 / 483.0	3.52	11562.66	2744.280900	350.2	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:45 PM

<b>Sample Name</b>	JV70	<b>Injection Vial</b>	8
<b>Sample ID</b>	L7	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:23:30	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.73	25851.47	100.563720	968.2	false
13C2-PFDA	515.0 / 470.0	3.20	30752.78	99.661495	494.6	false
d5-EtFOSAA	589.0 / 419.0	3.52	30224.48	410.733534	226.2	false



Sample Name	JV71	Injection Vial	9
Sample ID	L8	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:32:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	1287413.24	4522.222161	3112.7	false
PFBS_2	298.9 / 99.0	1.46	396436.89	4512.772938	1144.5	false
PFHxA_1	313.0 / 269.0	1.74	1757928.43	4963.910989	673.7	false
PFHxA_2	313.0 / 119.0	1.74	128260.98	5054.860411	556.0	false
PFHpA_1	363.0 / 319.0	2.10	1699109.37	5029.188658	389.9	false
PFHpA_2	363.0 / 169.0	2.10	32954.72	5000.525187	345.7	false
PFHxS_1	399.0 / 80.0	2.12	1500735.87	4626.551146	401.9	false
PFHxS_2	399.0 / 99.0	2.12	440867.38	4621.692989	400.2	false
PFOA_1	413.0 / 369.0	2.48	1960403.01	5055.859138	410.4	false
PFOA_2	413.0 / 169.0	2.48	136886.88	5284.814305	449.0	false
PFNA_1	463.0 / 419.0	2.86	1865407.06	5033.084124	385.6	false
PFNA_2	463.0 / 219.0	2.86	541521.86	5010.507489	374.5	false
PFOS_1	499.0 / 80.0	2.85	2117234.27	4717.827738	415.8	false
PFOS_2	499.0 / 99.0	2.85	409712.35	4719.092631	479.9	false
PFDA_1	513.0 / 469.0	3.21	1980955.65	4948.282902	483.3	false
PFDA_2	513.0 / 219.0	3.21	79938.74	5020.912753	618.6	false
PFUnA_1	563.0 / 519.0	3.54	2157042.52	4930.392260	331.5	false
PFUnA_2	563.0 / 269.0	3.53	101722.85	4924.516855	285.7	false
PFDoA_1	613.0 / 569.0	3.82	2182897.19	4994.806618	321.3	false
PFDoA_2	613.0 / 319.0	3.82	348893.44	5061.671458	345.1	false
PFTTrDA_1	663.0 / 619.0	4.08	2139400.50	4994.249530	426.2	false
PFTTrDA_2	663.0 / 169.0	4.08	133982.05	4924.491415	430.0	false
PFTeDA_1	713.0 / 669.0	4.30	2073121.17	4957.832384	672.7	false
PFTeDA_2	713.0 / 169.0	4.30	97478.07	4954.132013	566.4	false
NMeFOSAA_1	570.0 / 419.0	3.36	423626.27	5141.553388	403.9	false
NMeFOSAA_2	570.0 / 512.0	3.36	267565.67	5230.915871	421.7	false
NEtFOSAA_1	584.0 / 419.0	3.53	411771.90	5262.702481	431.1	false
NEtFOSAA_2	584.0 / 483.0	3.52	27117.60	5123.296865	450.2	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:47 PM

<b>Sample Name</b>	JV71	<b>Injection Vial</b>	9
<b>Sample ID</b>	L8	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:32:26	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.73	34086.46	97.800376	977.8	false
13C2-PFDA	515.0 / 470.0	3.20	44379.58	106.078813	696.6	false
d5-EtFOSAA	589.0 / 419.0	3.52	37727.95	406.137800	253.2	false





Sample Name	JV72	Injection Vial	10
Sample ID	L9	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:41:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	2649446.73	8835.635177	4187.4	false
PFBS_2	298.9 / 99.0	1.45	823445.55	8906.015758	1713.3	false
PFHxA_1	313.0 / 269.0	1.74	3535431.08	9765.932022	696.8	false
PFHxA_2	313.0 / 119.0	1.74	259419.13	9962.901685	675.9	false
PFHpA_1	363.0 / 319.0	2.10	3430003.03	9884.794468	433.8	false
PFHpA_2	363.0 / 169.0	2.10	66530.24	9869.229277	511.6	false
PFHxS_1	399.0 / 80.0	2.11	3105019.55	9081.815342	440.8	false
PFHxS_2	399.0 / 99.0	2.11	917048.30	9124.662600	438.6	false
PFOA_1	413.0 / 369.0	2.48	3869492.59	9730.370148	634.9	false
PFOA_2	413.0 / 169.0	2.48	249948.05	9403.716678	604.4	false
PFNA_1	463.0 / 419.0	2.85	3682212.88	9674.961722	457.3	false
PFNA_2	463.0 / 219.0	2.85	1076213.30	9699.841587	497.3	false
PFOS_1	499.0 / 80.0	2.84	4316675.88	9130.981033	400.6	false
PFOS_2	499.0 / 99.0	2.85	834613.26	9125.060312	489.8	false
PFDA_1	513.0 / 469.0	3.20	3974922.58	9683.098229	523.8	false
PFDA_2	513.0 / 219.0	3.20	161689.38	9930.419191	460.1	false
PFUnA_1	563.0 / 519.0	3.52	4448141.68	9898.077320	410.6	false
PFUnA_2	563.0 / 269.0	3.52	209947.36	9901.426543	425.9	false
PFDoA_1	613.0 / 569.0	3.81	4363531.65	9721.686021	414.0	false
PFDoA_2	613.0 / 319.0	3.81	692643.75	9787.182415	397.1	false
PFTTrDA_1	663.0 / 619.0	4.07	4292571.57	9756.172124	513.9	false
PFTTrDA_2	663.0 / 169.0	4.06	277467.05	9939.484030	493.6	false
PFTeDA_1	713.0 / 669.0	4.29	4217906.59	9822.873181	748.6	false
PFTeDA_2	713.0 / 169.0	4.28	198019.08	9811.678844	649.1	false
NMeFOSAA_1	570.0 / 419.0	3.35	876392.36	9795.498310	437.7	false
NMeFOSAA_2	570.0 / 512.0	3.35	546070.83	9847.446856	414.2	false
NEtFOSAA_1	584.0 / 419.0	3.52	824514.70	9700.640962	430.2	false
NEtFOSAA_2	584.0 / 483.0	3.51	55407.23	9656.630922	322.1	false



## Summary Quant Report

Created with Analyst Reporter  
Printed: 17/05/2018 2:08:49 PM

<b>Sample Name</b>	JV72	<b>Injection Vial</b>	10
<b>Sample ID</b>	L9	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:41:22	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.72	33342.76	89.559100	986.7	false
13C2-PFDA	515.0 / 470.0	3.19	42703.92	95.557070	777.9	false
d5-EtFOSAA	589.0 / 419.0	3.51	34375.59	340.462875	245.2	false



<b>Sample Name</b>	JV64	<b>Injection Vial</b>	2
<b>Sample ID</b>	L1	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:29:54	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.49	PFBS	0.354	0.327	ü
PFHxA_1	313.0 / 269.0	1.79	PFHxA			
PFHxA_2	313.0 / 119.0	1.77	PFHxA	0.051	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.54	PFHpA	0.027	0.021	ü
PFHxS_1	399.0 / 80.0	2.14	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.388	0.304	ü
PFOA_1	413.0 / 369.0	2.90	PFOA			
PFOA_2	413.0 / 169.0	2.90	PFOA	0.087	0.066	ü
PFNA_1	463.0 / 419.0	3.29	PFNA			
PFNA_2	463.0 / 219.0	3.28	PFNA	0.338	0.305	ü
PFOS_1	499.0 / 80.0	2.88	PFOS			
PFOS_2	499.0 / 99.0	2.89	PFOS	0.206	0.192	ü
PFDA_1	513.0 / 469.0	3.24	PFDA			
PFDA_2	513.0 / 219.0	3.26	PFDA	0.059	0.042	ü
PFUnA_1	563.0 / 519.0	3.56	PFUnA			
PFUnA_2	563.0 / 269.0	3.60	PFUnA	0.108	0.048	
PFDaA_1	613.0 / 569.0	3.85	PFDaA			
PFDaA_2	613.0 / 319.0	3.85	PFDaA	0.183	0.162	ü
PFTrDA_1	663.0 / 619.0	4.11	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.10	PFTrDA	0.089	0.068	ü
PFTeDA_1	713.0 / 669.0	4.33	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.030	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.38	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.37	NMeFOSAA	0.253	0.664	
NEtFOSAA_1	584.0 / 419.0	3.56	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.36	NEtFOSSA	0.062	0.080	ü

<b>Sample Name</b>	JV65	<b>Injection Vial</b>	3
<b>Sample ID</b>	L2	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:38:49	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.48	PFBS	0.404	0.327	ü
PFHxA_1	313.0 / 269.0	1.79	PFHxA			
PFHxA_2	313.0 / 119.0	1.79	PFHxA	0.077	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.028	0.021	ü
PFHxS_1	399.0 / 80.0	2.14	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.288	0.304	ü
PFOA_1	413.0 / 369.0	2.50	PFOA			
PFOA_2	413.0 / 169.0	2.51	PFOA	0.063	0.066	ü
PFNA_1	463.0 / 419.0	2.88	PFNA			
PFNA_2	463.0 / 219.0	2.88	PFNA	0.333	0.305	ü
PFOS_1	499.0 / 80.0	2.87	PFOS			
PFOS_2	499.0 / 99.0	2.87	PFOS	0.173	0.192	ü
PFDA_1	513.0 / 469.0	3.24	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.036	0.042	ü
PFUnA_1	563.0 / 519.0	3.56	PFUnA			
PFUnA_2	563.0 / 269.0	3.50	PFUnA	0.048	0.048	ü
PFDaA_1	613.0 / 569.0	3.85	PFDaA			
PFDaA_2	613.0 / 319.0	3.84	PFDaA	0.178	0.162	ü
PFTrDA_1	663.0 / 619.0	4.10	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.09	PFTrDA	0.088	0.068	ü
PFTeDA_1	713.0 / 669.0	4.33	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.067	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.39	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.41	NMeFOSAA	0.871	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.55	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.60	NEtFOSSA	0.089	0.080	ü

<b>Sample Name</b>	JV66	<b>Injection Vial</b>	4
<b>Sample ID</b>	L3	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:47:45	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.48	PFBS	0.326	0.327	ü
PFHxA_1	313.0 / 269.0	1.80	PFHxA			
PFHxA_2	313.0 / 119.0	1.78	PFHxA	0.063	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.023	0.021	ü
PFHxS_1	399.0 / 80.0	2.13	PFHxS			
PFHxS_2	399.0 / 99.0	2.14	PFHxS	0.303	0.304	ü
PFOA_1	413.0 / 369.0	2.50	PFOA			
PFOA_2	413.0 / 169.0	2.50	PFOA	0.057	0.066	ü
PFNA_1	463.0 / 419.0	2.88	PFNA			
PFNA_2	463.0 / 219.0	2.88	PFNA	0.307	0.299	ü
PFOS_1	499.0 / 80.0	2.87	PFOS			
PFOS_2	499.0 / 99.0	2.87	PFOS	0.200	0.192	ü
PFDA_1	513.0 / 469.0	3.23	PFDA			
PFDA_2	513.0 / 219.0	3.23	PFDA	0.059	0.042	ü
PFUnA_1	563.0 / 519.0	3.55	PFUnA			
PFUnA_2	563.0 / 269.0	3.56	PFUnA	0.052	0.048	ü
PFDaA_1	613.0 / 569.0	3.84	PFDaA			
PFDaA_2	613.0 / 319.0	3.84	PFDaA	0.155	0.162	ü
PFTrDA_1	663.0 / 619.0	4.10	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.09	PFTrDA	0.059	0.068	ü
PFTeDA_1	713.0 / 669.0	4.32	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.044	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.38	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.38	NMeFOSAA	0.656	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.54	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.47	NEtFOSSA	0.100	0.068	ü

<b>Sample Name</b>	JV67	<b>Injection Vial</b>	5
<b>Sample ID</b>	L4	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:56:42	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.321	0.327	ü
PFHxA_1	313.0 / 269.0	1.76	PFHxA			
PFHxA_2	313.0 / 119.0	1.76	PFHxA	0.063	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.026	0.021	ü
PFHxS_1	399.0 / 80.0	2.13	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.282	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.070	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.300	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.193	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.043	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.048	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.165	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.068	0.068	ü
PFTeDA_1	713.0 / 669.0	4.31	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.641	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.53	NEtFOSSA	0.070	0.080	ü



<b>Sample Name</b>	JV68	<b>Injection Vial</b>	6
<b>Sample ID</b>	L5	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:05:38	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.48	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.304	0.327	ü
PFHxA_1	313.0 / 269.0	1.76	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.062	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.021	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.306	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.066	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.309	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.186	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.23	PFDA	0.038	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.052	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.152	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.060	0.068	ü
PFTeDA_1	713.0 / 669.0	4.31	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.676	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.059	0.080	ü

<b>Sample Name</b>	JV69	<b>Injection Vial</b>	7
<b>Sample ID</b>	L6	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:14:33	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.064	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.290	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.069	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.289	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.187	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.037	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.148	0.162	ü
PFTTrDA_1	663.0 / 619.0	4.08	PFTTrDA			
PFTTrDA_2	663.0 / 169.0	4.08	PFTTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.625	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.53	NEtFOSSA	0.075	0.080	ü

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.301	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.067	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.020	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.292	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.068	0.066	ü
PFNA_1	463.0 / 419.0	2.86	PFNA			
PFNA_2	463.0 / 219.0	2.86	PFNA	0.289	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.198	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.037	0.042	ü
PFUnA_1	563.0 / 519.0	3.53	PFUnA			
PFUnA_2	563.0 / 269.0	3.53	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.82	PFDaA			
PFDaA_2	613.0 / 319.0	3.82	PFDaA	0.155	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.062	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.589	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.073	0.080	ü

<b>Sample Name</b>	JV71	<b>Injection Vial</b>	9
<b>Sample ID</b>	L8	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:32:26	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.308	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.294	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.070	0.066	ü
PFNA_1	463.0 / 419.0	2.86	PFNA			
PFNA_2	463.0 / 219.0	2.86	PFNA	0.290	0.305	ü
PFOS_1	499.0 / 80.0	2.85	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.194	0.192	ü
PFDA_1	513.0 / 469.0	3.21	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.040	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.53	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.82	PFDaA			
PFDaA_2	613.0 / 319.0	3.82	PFDaA	0.160	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.632	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.066	0.080	ü

<b>Sample Name</b>	JV72	<b>Injection Vial</b>	10
<b>Sample ID</b>	L9	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:41:22	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.295	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.065	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.292	0.305	ü
PFOS_1	499.0 / 80.0	2.84	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.193	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.041	0.042	ü
PFUnA_1	563.0 / 519.0	3.52	PFUnA			
PFUnA_2	563.0 / 269.0	3.52	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.81	PFDaA			
PFDaA_2	613.0 / 319.0	3.81	PFDaA	0.159	0.162	ü
PFTrDA_1	663.0 / 619.0	4.07	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.06	PFTrDA	0.065	0.068	ü
PFTeDA_1	713.0 / 669.0	4.29	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.28	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.35	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.35	NMeFOSAA	0.623	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.52	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.51	NEtFOSSA	0.067	0.080	ü

Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.49	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFBS_2	298.9 / 99.0	1.49	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFHxA_1	313.0 / 269.0	1.79	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFHxA_2	313.0 / 119.0	1.77	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFHpA_1	363.0 / 319.0	2.12	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFHpA_2	363.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFHxS_1	399.0 / 80.0	2.14	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFHxS_2	399.0 / 99.0	2.13	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFOA_1	413.0 / 369.0	2.90	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFOA_2	413.0 / 169.0	2.90	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFNA_1	463.0 / 419.0	3.29	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFNA_2	463.0 / 219.0	3.28	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFOS_1	499.0 / 80.0	2.88	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFOS_2	499.0 / 99.0	2.89	13C4-PFOS	503.0 / 80.0	104334.94	287.00
PFDA_1	513.0 / 469.0	3.24	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFDA_2	513.0 / 219.0	3.26	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFUnA_1	563.0 / 519.0	3.56	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFUnA_2	563.0 / 269.0	3.60	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFDaA_1	613.0 / 569.0	3.85	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFDaA_2	613.0 / 319.0	3.85	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFTTrDA_1	663.0 / 619.0	4.11	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFTTrDA_2	663.0 / 169.0	4.10	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFTeDA_1	713.0 / 669.0	4.33	13C2-PFOA	415.0 / 370.0	34698.30	100.00
PFTeDA_2	713.0 / 169.0	4.32	13C2-PFOA	415.0 / 370.0	34698.30	100.00
NMeFOSAA_1	570.0 / 419.0	3.38	d3-MeFOSAA	573.0 / 419.0	35968.44	400.00
NMeFOSAA_2	570.0 / 512.0	3.37	d3-MeFOSAA	573.0 / 419.0	35968.44	400.00
NEtFOSAA_1	584.0 / 419.0	3.56	d3-MeFOSAA	573.0 / 419.0	35968.44	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	35968.44	400.00

Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.75	13C2-PFOA	415.0 / 370.0	34698.30	100.00
13C2-PFDA	515.0 / 470.0	3.23	13C2-PFOA	415.0 / 370.0	34698.30	100.00
d5-EtFOSAA	589.0 / 419.0	3.54	d3-MeFOSAA	573.0 / 419.0	35968.44	400.00

Sample Name	JV65	Injection Vial	3
Sample ID	L2	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:38:49	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.49	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFBS_2	298.9 / 99.0	1.48	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFHxA_1	313.0 / 269.0	1.79	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFHxA_2	313.0 / 119.0	1.79	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFHpA_1	363.0 / 319.0	2.12	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFHxS_1	399.0 / 80.0	2.14	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFHxS_2	399.0 / 99.0	2.13	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFOA_1	413.0 / 369.0	2.50	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFOA_2	413.0 / 169.0	2.51	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFNA_1	463.0 / 419.0	2.88	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFNA_2	463.0 / 219.0	2.88	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFOS_1	499.0 / 80.0	2.87	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFOS_2	499.0 / 99.0	2.87	13C4-PFOS	503.0 / 80.0	98871.27	287.00
PFDA_1	513.0 / 469.0	3.24	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFDA_2	513.0 / 219.0	3.22	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFUnA_1	563.0 / 519.0	3.56	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFUnA_2	563.0 / 269.0	3.50	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFDaA_1	613.0 / 569.0	3.85	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFDaA_2	613.0 / 319.0	3.84	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFTTrDA_1	663.0 / 619.0	4.10	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFTTrDA_2	663.0 / 169.0	4.09	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFTeDA_1	713.0 / 669.0	4.33	13C2-PFOA	415.0 / 370.0	33573.03	100.00
PFTeDA_2	713.0 / 169.0	4.32	13C2-PFOA	415.0 / 370.0	33573.03	100.00
NMeFOSAA_1	570.0 / 419.0	3.39	d3-MeFOSAA	573.0 / 419.0	30241.50	400.00
NMeFOSAA_2	570.0 / 512.0	3.41	d3-MeFOSAA	573.0 / 419.0	30241.50	400.00
NEtFOSAA_1	584.0 / 419.0	3.55	d3-MeFOSAA	573.0 / 419.0	30241.50	400.00
NEtFOSAA_2	584.0 / 483.0	3.60	d3-MeFOSAA	573.0 / 419.0	30241.50	400.00



Sample Name	JV65	Injection Vial	3
Sample ID	L2	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:38:49	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.74	13C2-PFOA	415.0 / 370.0	33573.03	100.00
13C2-PFDA	515.0 / 470.0	3.22	13C2-PFOA	415.0 / 370.0	33573.03	100.00
d5-EtFOSAA	589.0 / 419.0	3.54	d3-MeFOSAA	573.0 / 419.0	30241.50	400.00

Sample Name	JV66	Injection Vial	4
Sample ID	L3	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:47:45	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.49	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFBS_2	298.9 / 99.0	1.48	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFHxA_1	313.0 / 269.0	1.80	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFHxA_2	313.0 / 119.0	1.78	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFHpA_1	363.0 / 319.0	2.12	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFHxS_1	399.0 / 80.0	2.13	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFHxS_2	399.0 / 99.0	2.14	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFOA_1	413.0 / 369.0	2.50	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFOA_2	413.0 / 169.0	2.50	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFNA_1	463.0 / 419.0	2.88	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFNA_2	463.0 / 219.0	2.88	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFOS_1	499.0 / 80.0	2.87	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFOS_2	499.0 / 99.0	2.87	13C4-PFOS	503.0 / 80.0	90792.83	287.00
PFDA_1	513.0 / 469.0	3.23	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFDA_2	513.0 / 219.0	3.23	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFUnA_1	563.0 / 519.0	3.55	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFUnA_2	563.0 / 269.0	3.56	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFDaA_1	613.0 / 569.0	3.84	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFDaA_2	613.0 / 319.0	3.84	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFTTrDA_1	663.0 / 619.0	4.10	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFTTrDA_2	663.0 / 169.0	4.09	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFTeDA_1	713.0 / 669.0	4.32	13C2-PFOA	415.0 / 370.0	29649.37	100.00
PFTeDA_2	713.0 / 169.0	4.32	13C2-PFOA	415.0 / 370.0	29649.37	100.00
NMeFOSAA_1	570.0 / 419.0	3.38	d3-MeFOSAA	573.0 / 419.0	28007.84	400.00
NMeFOSAA_2	570.0 / 512.0	3.38	d3-MeFOSAA	573.0 / 419.0	28007.84	400.00
NEtFOSAA_1	584.0 / 419.0	3.54	d3-MeFOSAA	573.0 / 419.0	28007.84	400.00
NEtFOSAA_2	584.0 / 483.0	3.47	d3-MeFOSAA	573.0 / 419.0	28007.84	400.00

Sample Name	JV66	Injection Vial	4
Sample ID	L3	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:47:45	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.75	13C2-PFOA	415.0 / 370.0	29649.37	100.00
13C2-PFDA	515.0 / 470.0	3.22	13C2-PFOA	415.0 / 370.0	29649.37	100.00
d5-EtFOSAA	589.0 / 419.0	3.54	d3-MeFOSAA	573.0 / 419.0	28007.84	400.00

Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:56:42	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFHxA_1	313.0 / 269.0	1.76	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFHxA_2	313.0 / 119.0	1.76	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFHpA_1	363.0 / 319.0	2.11	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFHxS_1	399.0 / 80.0	2.13	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFHxS_2	399.0 / 99.0	2.13	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFOA_1	413.0 / 369.0	2.49	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFOA_2	413.0 / 169.0	2.49	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFNA_1	463.0 / 419.0	2.87	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFNA_2	463.0 / 219.0	2.87	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFOS_1	499.0 / 80.0	2.86	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFOS_2	499.0 / 99.0	2.86	13C4-PFOS	503.0 / 80.0	116199.13	287.00
PFDA_1	513.0 / 469.0	3.22	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFDA_2	513.0 / 219.0	3.22	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFUnA_1	563.0 / 519.0	3.54	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFUnA_2	563.0 / 269.0	3.54	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFDaA_1	613.0 / 569.0	3.83	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFDaA_2	613.0 / 319.0	3.83	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFTTrDA_1	663.0 / 619.0	4.08	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFTTrDA_2	663.0 / 169.0	4.08	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFTeDA_1	713.0 / 669.0	4.31	13C2-PFOA	415.0 / 370.0	41395.56	100.00
PFTeDA_2	713.0 / 169.0	4.30	13C2-PFOA	415.0 / 370.0	41395.56	100.00
NMeFOSAA_1	570.0 / 419.0	3.37	d3-MeFOSAA	573.0 / 419.0	36150.49	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	36150.49	400.00
NEtFOSAA_1	584.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	36150.49	400.00
NEtFOSAA_2	584.0 / 483.0	3.53	d3-MeFOSAA	573.0 / 419.0	36150.49	400.00

Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:56:42	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.74	13C2-PFOA	415.0 / 370.0	41395.56	100.00
13C2-PFDA	515.0 / 470.0	3.21	13C2-PFOA	415.0 / 370.0	41395.56	100.00
d5-EtFOSAA	589.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	36150.49	400.00

Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:05:38	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.48	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFHxA_1	313.0 / 269.0	1.76	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFHpA_1	363.0 / 319.0	2.11	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFHxS_2	399.0 / 99.0	2.13	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFOA_1	413.0 / 369.0	2.49	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFOA_2	413.0 / 169.0	2.49	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFNA_1	463.0 / 419.0	2.87	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFNA_2	463.0 / 219.0	2.87	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFOS_1	499.0 / 80.0	2.86	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFOS_2	499.0 / 99.0	2.86	13C4-PFOS	503.0 / 80.0	92291.39	287.00
PFDA_1	513.0 / 469.0	3.22	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFDA_2	513.0 / 219.0	3.23	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFUnA_1	563.0 / 519.0	3.54	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFUnA_2	563.0 / 269.0	3.54	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFDaA_1	613.0 / 569.0	3.83	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFDaA_2	613.0 / 319.0	3.83	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFTTrDA_1	663.0 / 619.0	4.08	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFTTrDA_2	663.0 / 169.0	4.08	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFTeDA_1	713.0 / 669.0	4.31	13C2-PFOA	415.0 / 370.0	30280.34	100.00
PFTeDA_2	713.0 / 169.0	4.30	13C2-PFOA	415.0 / 370.0	30280.34	100.00
NMeFOSAA_1	570.0 / 419.0	3.37	d3-MeFOSAA	573.0 / 419.0	30317.42	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	30317.42	400.00
NEtFOSAA_1	584.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	30317.42	400.00
NEtFOSAA_2	584.0 / 483.0	3.52	d3-MeFOSAA	573.0 / 419.0	30317.42	400.00

Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:05:38	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.74	13C2-PFOA	415.0 / 370.0	30280.34	100.00
13C2-PFDA	515.0 / 470.0	3.21	13C2-PFOA	415.0 / 370.0	30280.34	100.00
d5-EtFOSAA	589.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	30317.42	400.00

Sample Name	JV69	Injection Vial	7
Sample ID	L6	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:14:33	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFHpA_1	363.0 / 319.0	2.11	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFHxS_2	399.0 / 99.0	2.13	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFOA_1	413.0 / 369.0	2.49	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFOA_2	413.0 / 169.0	2.49	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFNA_1	463.0 / 419.0	2.87	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFNA_2	463.0 / 219.0	2.87	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFOS_1	499.0 / 80.0	2.86	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFOS_2	499.0 / 99.0	2.86	13C4-PFOS	503.0 / 80.0	111784.92	287.00
PFDA_1	513.0 / 469.0	3.22	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFDA_2	513.0 / 219.0	3.22	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFUnA_1	563.0 / 519.0	3.54	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFUnA_2	563.0 / 269.0	3.54	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFDaA_1	613.0 / 569.0	3.83	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFDaA_2	613.0 / 319.0	3.83	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFTTrDA_1	663.0 / 619.0	4.08	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFTTrDA_2	663.0 / 169.0	4.08	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFTeDA_1	713.0 / 669.0	4.30	13C2-PFOA	415.0 / 370.0	37534.68	100.00
PFTeDA_2	713.0 / 169.0	4.30	13C2-PFOA	415.0 / 370.0	37534.68	100.00
NMeFOSAA_1	570.0 / 419.0	3.37	d3-MeFOSAA	573.0 / 419.0	36296.34	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	36296.34	400.00
NEtFOSAA_1	584.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	36296.34	400.00
NEtFOSAA_2	584.0 / 483.0	3.53	d3-MeFOSAA	573.0 / 419.0	36296.34	400.00



<b>Sample Name</b>	JV69	<b>Injection Vial</b>	7
<b>Sample ID</b>	L6	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:14:33	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.73	13C2-PFOA	415.0 / 370.0	37534.68	100.00
13C2-PFDA	515.0 / 470.0	3.21	13C2-PFOA	415.0 / 370.0	37534.68	100.00
d5-EtFOSAA	589.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	36296.34	400.00

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFHpA_1	363.0 / 319.0	2.11	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFHpA_2	363.0 / 169.0	2.11	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFHxS_2	399.0 / 99.0	2.12	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFOA_1	413.0 / 369.0	2.49	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFOA_2	413.0 / 169.0	2.49	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFNA_1	463.0 / 419.0	2.86	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFNA_2	463.0 / 219.0	2.86	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFOS_1	499.0 / 80.0	2.86	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFOS_2	499.0 / 99.0	2.86	13C4-PFOS	503.0 / 80.0	88194.77	287.00
PFDA_1	513.0 / 469.0	3.22	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFDA_2	513.0 / 219.0	3.21	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFUnA_1	563.0 / 519.0	3.53	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFUnA_2	563.0 / 269.0	3.53	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFDaA_1	613.0 / 569.0	3.82	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFDaA_2	613.0 / 319.0	3.82	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFTTrDA_1	663.0 / 619.0	4.08	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFTTrDA_2	663.0 / 169.0	4.08	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFTeDA_1	713.0 / 669.0	4.30	13C2-PFOA	415.0 / 370.0	29423.02	100.00
PFTeDA_2	713.0 / 169.0	4.30	13C2-PFOA	415.0 / 370.0	29423.02	100.00
NMeFOSAA_1	570.0 / 419.0	3.36	d3-MeFOSAA	573.0 / 419.0	27471.55	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	27471.55	400.00
NEtFOSAA_1	584.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	27471.55	400.00
NEtFOSAA_2	584.0 / 483.0	3.52	d3-MeFOSAA	573.0 / 419.0	27471.55	400.00

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.73	13C2-PFOA	415.0 / 370.0	29423.02	100.00
13C2-PFDA	515.0 / 470.0	3.20	13C2-PFOA	415.0 / 370.0	29423.02	100.00
d5-EtFOSAA	589.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	27471.55	400.00

Sample Name	JV71	Injection Vial	9
Sample ID	L8	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:32:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFHpA_2	363.0 / 169.0	2.10	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFHxS_2	399.0 / 99.0	2.12	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFOA_1	413.0 / 369.0	2.48	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFOA_2	413.0 / 169.0	2.48	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFNA_1	463.0 / 419.0	2.86	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFNA_2	463.0 / 219.0	2.86	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFOS_1	499.0 / 80.0	2.85	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFOS_2	499.0 / 99.0	2.85	13C4-PFOS	503.0 / 80.0	110166.97	287.00
PFDA_1	513.0 / 469.0	3.21	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFDA_2	513.0 / 219.0	3.21	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFUnA_1	563.0 / 519.0	3.54	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFUnA_2	563.0 / 269.0	3.53	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFDaA_1	613.0 / 569.0	3.82	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFDaA_2	613.0 / 319.0	3.82	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFTTrDA_1	663.0 / 619.0	4.08	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFTTrDA_2	663.0 / 169.0	4.08	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFTeDA_1	713.0 / 669.0	4.30	13C2-PFOA	415.0 / 370.0	39891.91	100.00
PFTeDA_2	713.0 / 169.0	4.30	13C2-PFOA	415.0 / 370.0	39891.91	100.00
NMeFOSAA_1	570.0 / 419.0	3.36	d3-MeFOSAA	573.0 / 419.0	34679.62	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	34679.62	400.00
NEtFOSAA_1	584.0 / 419.0	3.53	d3-MeFOSAA	573.0 / 419.0	34679.62	400.00
NEtFOSAA_2	584.0 / 483.0	3.52	d3-MeFOSAA	573.0 / 419.0	34679.62	400.00

Sample Name	JV71	Injection Vial	9
Sample ID	L8	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:32:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.73	13C2-PFOA	415.0 / 370.0	39891.91	100.00
13C2-PFDA	515.0 / 470.0	3.20	13C2-PFOA	415.0 / 370.0	39891.91	100.00
d5-EtFOSAA	589.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	34679.62	400.00

Sample Name	JV72	Injection Vial	10
Sample ID	L9	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:41:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFHpA_2	363.0 / 169.0	2.10	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFOA_1	413.0 / 369.0	2.48	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFOA_2	413.0 / 169.0	2.48	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFOS_1	499.0 / 80.0	2.84	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFOS_2	499.0 / 99.0	2.85	13C4-PFOS	503.0 / 80.0	116123.32	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFUnA_1	563.0 / 519.0	3.52	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFUnA_2	563.0 / 269.0	3.52	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFDaA_1	613.0 / 569.0	3.81	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFDaA_2	613.0 / 319.0	3.81	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFTTrDA_1	663.0 / 619.0	4.07	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFTTrDA_2	663.0 / 169.0	4.06	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFTTeDA_1	713.0 / 669.0	4.29	13C2-PFOA	415.0 / 370.0	40992.20	100.00
PFTTeDA_2	713.0 / 169.0	4.28	13C2-PFOA	415.0 / 370.0	40992.20	100.00
NMeFOSAA_1	570.0 / 419.0	3.35	d3-MeFOSAA	573.0 / 419.0	37693.37	400.00
NMeFOSAA_2	570.0 / 512.0	3.35	d3-MeFOSAA	573.0 / 419.0	37693.37	400.00
NEtFOSAA_1	584.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	37693.37	400.00
NEtFOSAA_2	584.0 / 483.0	3.51	d3-MeFOSAA	573.0 / 419.0	37693.37	400.00

Sample Name	JV72	Injection Vial	10
Sample ID	L9	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:41:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

**Results Summary**

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
13C2-PFHxA	315.0 / 270.0	1.72	13C2-PFOA	415.0 / 370.0	42612.32	100.00
13C2-PFDA	515.0 / 470.0	3.19	13C2-PFOA	415.0 / 370.0	42612.32	100.00
d5-EtFOSAA	589.0 / 419.0	3.51	d3-MeFOSAA	573.0 / 419.0	37693.37	400.00

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.46	805.907678	885.00	91.06
PFBS_2	298.9 / 99.0	1.46	825.071704	885.00	93.23
PFHxA_1	313.0 / 269.0	1.75	962.249904	1000.00	96.22
PFHxA_2	313.0 / 119.0	1.74	1015.156998	1000.00	101.52
PFHpA_1	363.0 / 319.0	2.10	933.247533	1000.00	93.32
PFHpA_2	363.0 / 169.0	2.10	933.903106	1000.00	93.39
PFHxS_1	399.0 / 80.0	2.12	861.806904	912.00	94.50
PFHxS_2	399.0 / 99.0	2.12	826.048162	912.00	90.58
PFOA_1	413.0 / 369.0	2.48	941.626058	1000.00	94.16
PFOA_2	413.0 / 169.0	2.48	1065.520264	1000.00	106.55
PFNA_1	463.0 / 419.0	2.85	969.501428	1000.00	96.95
PFNA_2	463.0 / 219.0	2.85	1015.813698	1000.00	101.58
PFOS_1	499.0 / 80.0	2.85	830.716293	925.60	89.75
PFOS_2	499.0 / 99.0	2.85	936.586590	925.60	101.19
PFDA_1	513.0 / 469.0	3.21	952.414243	1000.00	95.24
PFDA_2	513.0 / 219.0	3.21	866.424928	1000.00	86.64
PFUnA_1	563.0 / 519.0	3.53	983.154309	1000.00	98.32
PFUnA_2	563.0 / 269.0	3.53	953.567429	1000.00	95.36
PFDoA_1	613.0 / 569.0	3.81	985.864788	1000.00	98.59
PFDoA_2	613.0 / 319.0	3.81	953.497275	1000.00	95.35
PFTTrDA_1	663.0 / 619.0	4.07	976.486543	1000.00	97.65
PFTTrDA_2	663.0 / 169.0	4.07	961.413908	1000.00	96.14
PFTeDA_1	713.0 / 669.0	4.29	938.482516	1000.00	93.85
PFTeDA_2	713.0 / 169.0	4.28	952.593795	1000.00	95.26
NMeFOSAA_1	570.0 / 419.0	3.36	1109.562943	1000.00	110.96
NMeFOSAA_2	570.0 / 512.0	3.36	980.360841	1000.00	98.04
NEtFOSAA_1	584.0 / 419.0	3.52	1088.808929	1000.00	108.88
NEtFOSAA_2	584.0 / 483.0	3.52	957.541377	1000.00	95.75



Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.73	92.204529	100.00	92.20
13C2-PFDA	515.0 / 470.0	3.19	92.949919	100.00	92.95
d5-EtFOSAA	589.0 / 419.0	3.51	381.212971	400.00	95.30

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.46	899.387330	885.00	101.63
PFBS_2	298.9 / 99.0	1.45	900.526623	885.00	101.75
PFHxA_1	313.0 / 269.0	1.74	1183.308886	1000.00	118.33
PFHxA_2	313.0 / 119.0	1.73	1192.266265	1000.00	119.23
PFHpA_1	363.0 / 319.0	2.09	1094.466247	1000.00	109.45
PFHpA_2	363.0 / 169.0	2.09	1068.460245	1000.00	106.85
PFHxS_1	399.0 / 80.0	2.10	913.078700	912.00	100.12
PFHxS_2	399.0 / 99.0	2.11	872.644722	912.00	95.68
PFOA_1	413.0 / 369.0	2.47	1115.848591	1000.00	111.58
PFOA_2	413.0 / 169.0	2.47	1136.973471	1000.00	113.70
PFNA_1	463.0 / 419.0	2.84	1171.510532	1000.00	117.15
PFNA_2	463.0 / 219.0	2.84	1212.351411	1000.00	121.24
PFOS_1	499.0 / 80.0	2.84	946.924957	925.60	102.30
PFOS_2	499.0 / 99.0	2.84	945.562142	925.60	102.16
PFDA_1	513.0 / 469.0	3.20	1186.150677	1000.00	118.62
PFDA_2	513.0 / 219.0	3.20	1102.602568	1000.00	110.26
PFUnA_1	563.0 / 519.0	3.51	1152.561588	1000.00	115.26
PFUnA_2	563.0 / 269.0	3.51	1141.269070	1000.00	114.13
PFDoA_1	613.0 / 569.0	3.80	1144.819746	1000.00	114.48
PFDoA_2	613.0 / 319.0	3.80	1111.307165	1000.00	111.13
PFTTrDA_1	663.0 / 619.0	4.05	1134.100929	1000.00	113.41
PFTTrDA_2	663.0 / 169.0	4.05	1115.502894	1000.00	111.55
PFTeDA_1	713.0 / 669.0	4.27	1145.940772	1000.00	114.59
PFTeDA_2	713.0 / 169.0	4.27	1129.570434	1000.00	112.96
NMeFOSAA_1	570.0 / 419.0	3.34	1244.481183	1000.00	124.45
NMeFOSAA_2	570.0 / 512.0	3.34	1216.883830	1000.00	121.69
NEtFOSAA_1	584.0 / 419.0	3.51	1169.751144	1000.00	116.98
NEtFOSAA_2	584.0 / 483.0	3.49	1069.898281	1000.00	106.99

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	107.339437	100.00	107.34
13C2-PFDA	515.0 / 470.0	3.18	100.831039	100.00	100.83
d5-EtFOSAA	589.0 / 419.0	3.50	427.992074	400.00	107.00

Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.45	440.416958	443.00	99.42
PFBS_2	298.9 / 99.0	1.45	434.150471	443.00	98.00
PFHxA_1	313.0 / 269.0	1.73	499.036672	500.00	99.81
PFHxA_2	313.0 / 119.0	1.73	528.285027	500.00	105.66
PFHpA_1	363.0 / 319.0	2.09	471.408594	500.00	94.28
PFHpA_2	363.0 / 169.0	2.08	558.410072	500.00	111.68
PFHxS_1	399.0 / 80.0	2.10	441.193340	456.00	96.75
PFHxS_2	399.0 / 99.0	2.10	428.872944	456.00	94.05
PFOA_1	413.0 / 369.0	2.46	469.777523	500.00	93.96
PFOA_2	413.0 / 169.0	2.46	515.299469	500.00	103.06
PFNA_1	463.0 / 419.0	2.83	523.788059	500.00	104.76
PFNA_2	463.0 / 219.0	2.83	505.535719	500.00	101.11
PFOS_1	499.0 / 80.0	2.83	460.723951	463.00	99.51
PFOS_2	499.0 / 99.0	2.83	477.544848	463.00	103.14
PFDA_1	513.0 / 469.0	3.18	489.972812	500.00	97.99
PFDA_2	513.0 / 219.0	3.18	465.489745	500.00	93.10
PFUnA_1	563.0 / 519.0	3.50	518.325361	500.00	103.67
PFUnA_2	563.0 / 269.0	3.50	471.585271	500.00	94.32
PFDoA_1	613.0 / 569.0	3.79	476.947097	500.00	95.39
PFDoA_2	613.0 / 319.0	3.79	505.027230	500.00	101.01
PFTTrDA_1	663.0 / 619.0	4.04	479.806082	500.00	95.96
PFTTrDA_2	663.0 / 169.0	4.03	501.939831	500.00	100.39
PFTeDA_1	713.0 / 669.0	4.25	485.629530	500.00	97.13
PFTeDA_2	713.0 / 169.0	4.25	486.803296	500.00	97.36
NMeFOSAA_1	570.0 / 419.0	3.34	487.813835	500.00	97.56
NMeFOSAA_2	570.0 / 512.0	3.34	448.600228	500.00	89.72
NEtFOSAA_1	584.0 / 419.0	3.50	482.711585	500.00	96.54
NEtFOSAA_2	584.0 / 483.0	3.49	625.585443	500.00	125.12

Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.71	95.522162	100.00	95.52
13C2-PFDA	515.0 / 470.0	3.17	92.313155	100.00	92.31
d5-EtFOSAA	589.0 / 419.0	3.49	357.700451	400.00	89.43

<b>Sample Name</b>	JV63 ICC	<b>Injection Vial</b>	11
<b>Sample ID</b>	ICC	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:50:20	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	235290.46	805.907678	1408.2	false
PFBS_2	298.9 / 99.0	1.46	74819.62	825.071704	582.0	false
PFHxA_1	313.0 / 269.0	1.75	333436.98	962.249904	534.0	false
PFHxA_2	313.0 / 119.0	1.74	24404.61	1015.156998	359.9	false
PFHpA_1	363.0 / 319.0	2.10	296803.84	933.247533	222.8	false
PFHpA_2	363.0 / 169.0	2.10	6000.30	933.903106	246.2	false
PFHxS_1	399.0 / 80.0	2.12	284863.44	861.806904	425.4	false
PFHxS_2	399.0 / 99.0	2.12	80600.62	826.048162	342.2	false
PFOA_1	413.0 / 369.0	2.48	348229.30	941.626058	281.2	false
PFOA_2	413.0 / 169.0	2.48	26188.87	1065.520264	339.0	false
PFNA_1	463.0 / 419.0	2.85	338780.27	969.501428	265.0	false
PFNA_2	463.0 / 219.0	2.85	103699.47	1015.813698	399.5	false
PFOS_1	499.0 / 80.0	2.85	381934.75	830.716293	307.1	false
PFOS_2	499.0 / 99.0	2.85	83202.35	936.586590	350.3	false
PFDA_1	513.0 / 469.0	3.21	363857.52	952.414243	274.5	false
PFDA_2	513.0 / 219.0	3.21	13527.04	866.424928	375.3	false
PFUnA_1	563.0 / 519.0	3.53	404366.10	983.154309	270.3	false
PFUnA_2	563.0 / 269.0	3.53	18622.53	953.567429	269.8	false
PFDoA_1	613.0 / 569.0	3.81	405638.50	985.864788	283.6	false
PFDoA_2	613.0 / 319.0	3.81	62057.69	953.497275	267.5	false
PFTTrDA_1	663.0 / 619.0	4.07	393570.96	976.486543	366.6	false
PFTTrDA_2	663.0 / 169.0	4.07	24817.26	961.413908	316.7	false
PFTeDA_1	713.0 / 669.0	4.29	369927.89	938.482516	548.1	false
PFTeDA_2	713.0 / 169.0	4.28	17838.62	952.593795	428.9	false
NMeFOSAA_1	570.0 / 419.0	3.36	91336.98	1109.562943	541.3	false
NMeFOSAA_2	570.0 / 512.0	3.36	50927.07	980.360841	366.8	false
NEtFOSAA_1	584.0 / 419.0	3.52	84901.41	1088.808929	427.7	false
NEtFOSAA_2	584.0 / 483.0	3.52	5151.02	957.541377	279.7	false

<b>Sample Name</b>	JV63 ICC	<b>Injection Vial</b>	11
<b>Sample ID</b>	ICC	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:50:20	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

**Results Summary**

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.73	30119.31	92.204529	727.6	false
13C2-PFDA	515.0 / 470.0	3.19	36446.42	92.949919	504.0	false
d5-EtFOSAA	589.0 / 419.0	3.51	35128.71	381.212971	272.6	false



Summary Quant Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:47:29 PM

<b>Sample Name</b>	JV69 CCV	<b>Injection Vial</b>	7
<b>Sample ID</b>	CCV	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:19:36	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	290061.89	984.294108	1295.9	false
PFBS_2	298.9 / 99.0	1.45	90162.89	986.186808	648.0	false
PFHxA_1	313.0 / 269.0	1.74	404256.15	1183.308886	478.1	false
PFHxA_2	313.0 / 119.0	1.73	28479.75	1192.266265	359.5	false
PFHpA_1	363.0 / 319.0	2.09	346279.47	1094.466247	231.7	false
PFHpA_2	363.0 / 169.0	2.09	6793.86	1068.460245	177.9	false
PFHxS_1	399.0 / 80.0	2.10	333673.22	998.682032	291.3	false
PFHxS_2	399.0 / 99.0	2.11	94113.27	954.815460	266.0	false
PFOA_1	413.0 / 369.0	2.47	409488.17	1115.848591	313.1	false
PFOA_2	413.0 / 169.0	2.47	27795.81	1136.973471	310.0	false
PFNA_1	463.0 / 419.0	2.84	407062.15	1171.510532	247.4	false
PFNA_2	463.0 / 219.0	2.84	123027.12	1212.351411	337.6	false
PFOS_1	499.0 / 80.0	2.84	480923.23	1036.206519	271.7	false
PFOS_2	499.0 / 99.0	2.84	92865.60	1034.666374	366.9	false
PFDA_1	513.0 / 469.0	3.20	449201.35	1186.150677	276.6	false
PFDA_2	513.0 / 219.0	3.20	16942.76	1102.602568	328.8	false
PFUnA_1	563.0 / 519.0	3.51	471703.07	1152.561588	258.4	false
PFUnA_2	563.0 / 269.0	3.51	22151.39	1141.269070	285.5	false
PFDoA_1	613.0 / 569.0	3.80	468614.33	1144.819746	299.1	false
PFDoA_2	613.0 / 319.0	3.80	71918.72	1111.307165	268.8	false
PFTTrDA_1	663.0 / 619.0	4.05	454791.64	1134.100929	372.3	false
PFTTrDA_2	663.0 / 169.0	4.05	28608.97	1115.502894	310.2	false
PFTeDA_1	713.0 / 669.0	4.27	449147.42	1145.940772	535.3	false
PFTeDA_2	713.0 / 169.0	4.27	20997.95	1129.570434	391.3	false
NMeFOSAA_1	570.0 / 419.0	3.34	94420.04	1244.481183	364.0	false
NMeFOSAA_2	570.0 / 512.0	3.34	57996.41	1216.883830	262.4	false
NEtFOSAA_1	584.0 / 419.0	3.51	84118.25	1169.751144	361.5	false
NEtFOSAA_2	584.0 / 483.0	3.49	5293.48	1069.898281	213.1	false





<b>Sample Name</b>	JV69 CCV	<b>Injection Vial</b>	7
<b>Sample ID</b>	CCV	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:19:36	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

**Results Summary**

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.72	37547.30	107.339437	1132.9	false
13C2-PFDA	515.0 / 470.0	3.18	42337.65	100.831039	423.3	false
d5-EtFOSAA	589.0 / 419.0	3.50	39770.08	427.992074	225.6	false

Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.45	127463.77	440.416958	953.8	false
PFBS_2	298.9 / 99.0	1.45	39328.13	434.150471	467.5	false
PFHxA_1	313.0 / 269.0	1.73	181898.79	499.036672	428.3	false
PFHxA_2	313.0 / 119.0	1.73	12899.18	528.285027	317.8	false
PFHpA_1	363.0 / 319.0	2.09	151194.99	471.408594	160.8	false
PFHpA_2	363.0 / 169.0	2.08	3714.33	558.410072	195.0	false
PFHxS_1	399.0 / 80.0	2.10	143652.45	441.193340	261.1	false
PFHxS_2	399.0 / 99.0	2.10	41395.86	428.872944	242.1	false
PFOA_1	413.0 / 369.0	2.46	178015.30	469.777523	235.9	false
PFOA_2	413.0 / 169.0	2.46	12914.91	515.299469	230.9	false
PFNA_1	463.0 / 419.0	2.83	184765.95	523.788059	206.0	false
PFNA_2	463.0 / 219.0	2.83	52281.42	505.535719	280.8	false
PFOS_1	499.0 / 80.0	2.83	209725.14	460.723951	276.1	false
PFOS_2	499.0 / 99.0	2.83	41994.88	477.544848	367.4	false
PFDA_1	513.0 / 469.0	3.18	191713.64	489.972812	229.0	false
PFDA_2	513.0 / 219.0	3.18	7626.50	465.489745	212.3	false
PFUnA_1	563.0 / 519.0	3.50	214599.88	518.325361	224.8	false
PFUnA_2	563.0 / 269.0	3.50	9340.88	471.585271	215.5	false
PFDoA_1	613.0 / 569.0	3.79	198049.51	476.947097	235.4	false
PFDoA_2	613.0 / 319.0	3.79	33242.09	505.027230	214.4	false
PFTrDA_1	663.0 / 619.0	4.04	194973.00	479.806082	297.6	false
PFTrDA_2	663.0 / 169.0	4.03	13177.98	501.939831	222.2	false
PFTeDA_1	713.0 / 669.0	4.25	193311.02	485.629530	443.5	false
PFTeDA_2	713.0 / 169.0	4.25	9310.33	486.803296	312.4	false
NMeFOSAA_1	570.0 / 419.0	3.34	40006.28	487.813835	461.5	false
NMeFOSAA_2	570.0 / 512.0	3.34	23728.90	448.600228	218.3	false
NEtFOSAA_1	584.0 / 419.0	3.50	37342.24	482.711585	371.5	false
NEtFOSAA_2	584.0 / 483.0	3.49	3366.23	625.585443	180.3	false

<b>Sample Name</b>	JV68 CCV	<b>Injection Vial</b>	6
<b>Sample ID</b>	CCV	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:30:58	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

**Results Summary**

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
13C2-PFHxA	315.0 / 270.0	1.71	31303.44	95.522162	661.6	false
13C2-PFDA	515.0 / 470.0	3.17	36313.20	92.313155	491.6	false
d5-EtFOSAA	589.0 / 419.0	3.49	32463.79	357.700451	321.9	false

<b>Sample Name</b>	JV63 ICC	<b>Injection Vial</b>	11
<b>Sample ID</b>	ICC	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:50:20	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.318	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.020	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.283	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.075	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.306	0.305	ü
PFOS_1	499.0 / 80.0	2.85	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.218	0.192	ü
PFDA_1	513.0 / 469.0	3.21	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.037	0.042	ü
PFUnA_1	563.0 / 519.0	3.53	PFUnA			
PFUnA_2	563.0 / 269.0	3.53	PFUnA	0.046	0.048	ü
PFDaA_1	613.0 / 569.0	3.81	PFDaA			
PFDaA_2	613.0 / 319.0	3.81	PFDaA	0.153	0.162	ü
PFTrDA_1	663.0 / 619.0	4.07	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.07	PFTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.29	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.28	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.558	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.52	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.061	0.080	ü

<b>Sample Name</b>	JV69 CCV	<b>Injection Vial</b>	7
<b>Sample ID</b>	CCV	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:19:36	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.070	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.09	PFHpA	0.020	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.282	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.47	PFOA	0.068	0.066	ü
PFNA_1	463.0 / 419.0	2.84	PFNA			
PFNA_2	463.0 / 219.0	2.84	PFNA	0.302	0.305	ü
PFOS_1	499.0 / 80.0	2.84	PFOS			
PFOS_2	499.0 / 99.0	2.84	PFOS	0.193	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.038	0.042	ü
PFUnA_1	563.0 / 519.0	3.51	PFUnA			
PFUnA_2	563.0 / 269.0	3.51	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.80	PFDaA			
PFDaA_2	613.0 / 319.0	3.80	PFDaA	0.154	0.162	ü
PFTrDA_1	663.0 / 619.0	4.05	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.05	PFTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.27	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.27	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.34	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.34	NMeFOSAA	0.614	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.51	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.49	NEtFOSSA	0.063	0.080	ü

<b>Sample Name</b>	JV68 CCV	<b>Injection Vial</b>	6
<b>Sample ID</b>	CCV	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:30:58	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.45	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.309	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.071	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.08	PFHpA	0.025	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.10	PFHxS	0.288	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.46	PFOA	0.073	0.066	ü
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.283	0.305	ü
PFOS_1	499.0 / 80.0	2.83	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.200	0.192	ü
PFDA_1	513.0 / 469.0	3.18	PFDA			
PFDA_2	513.0 / 219.0	3.18	PFDA	0.040	0.042	ü
PFUnA_1	563.0 / 519.0	3.50	PFUnA			
PFUnA_2	563.0 / 269.0	3.50	PFUnA	0.044	0.048	ü
PFDaA_1	613.0 / 569.0	3.79	PFDaA			
PFDaA_2	613.0 / 319.0	3.79	PFDaA	0.168	0.162	ü
PFTrDA_1	663.0 / 619.0	4.04	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.03	PFTrDA	0.068	0.068	ü
PFTeDA_1	713.0 / 669.0	4.25	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.25	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.34	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.34	NMeFOSAA	0.593	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.50	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.49	NEtFOSSA	0.090	0.080	ü

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFHpA_2	363.0 / 169.0	2.10	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFHxS_2	399.0 / 99.0	2.12	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFOA_1	413.0 / 369.0	2.48	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFOA_2	413.0 / 169.0	2.48	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFOS_1	499.0 / 80.0	2.85	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFOS_2	499.0 / 99.0	2.85	13C4-PFOS	503.0 / 80.0	112208.04	287.00
PFDA_1	513.0 / 469.0	3.21	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFDA_2	513.0 / 219.0	3.21	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFUnA_1	563.0 / 519.0	3.53	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFUnA_2	563.0 / 269.0	3.53	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFDaA_1	613.0 / 569.0	3.81	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFDaA_2	613.0 / 319.0	3.81	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFTTrDA_1	663.0 / 619.0	4.07	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFTTrDA_2	663.0 / 169.0	4.07	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFTTeDA_1	713.0 / 669.0	4.29	13C2-PFOA	415.0 / 370.0	37388.34	100.00
PFTTeDA_2	713.0 / 169.0	4.28	13C2-PFOA	415.0 / 370.0	37388.34	100.00
NMeFOSAA_1	570.0 / 419.0	3.36	d3-MeFOSAA	573.0 / 419.0	34401.63	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	34401.63	400.00
NEtFOSAA_1	584.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	34401.63	400.00
NEtFOSAA_2	584.0 / 483.0	3.52	d3-MeFOSAA	573.0 / 419.0	34401.63	400.00





Created with Analyst Reporter  
 Summary Internal Standard Report Printed: 23/05/2018 11:57:49 AM

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFHpA_2	363.0 / 169.0	2.09	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFOA_2	413.0 / 169.0	2.47	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFNA_1	463.0 / 419.0	2.84	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFNA_2	463.0 / 219.0	2.84	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFOS_1	499.0 / 80.0	2.84	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFOS_2	499.0 / 99.0	2.84	13C4-PFOS	503.0 / 80.0	113429.56	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFUnA_1	563.0 / 519.0	3.51	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFUnA_2	563.0 / 269.0	3.51	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFDoA_1	613.0 / 569.0	3.80	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFDoA_2	613.0 / 319.0	3.80	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFTTrDA_1	663.0 / 619.0	4.05	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFTTrDA_2	663.0 / 169.0	4.05	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFTeDA_1	713.0 / 669.0	4.27	13C2-PFOA	415.0 / 370.0	37224.64	100.00
PFTeDA_2	713.0 / 169.0	4.27	13C2-PFOA	415.0 / 370.0	37224.64	100.00
NMeFOSAA_1	570.0 / 419.0	3.34	d3-MeFOSAA	573.0 / 419.0	31738.59	400.00
NMeFOSAA_2	570.0 / 512.0	3.34	d3-MeFOSAA	573.0 / 419.0	31738.59	400.00
NEtFOSAA_1	584.0 / 419.0	3.51	d3-MeFOSAA	573.0 / 419.0	31738.59	400.00
NEtFOSAA_2	584.0 / 483.0	3.49	d3-MeFOSAA	573.0 / 419.0	31738.59	400.00



Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.45	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFHxA_1	313.0 / 269.0	1.73	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFHpA_2	363.0 / 169.0	2.08	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFHxS_2	399.0 / 99.0	2.10	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFOA_2	413.0 / 169.0	2.46	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFOS_1	499.0 / 80.0	2.83	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	110468.54	287.00
PFDA_1	513.0 / 469.0	3.18	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFDA_2	513.0 / 219.0	3.18	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFUnA_1	563.0 / 519.0	3.50	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFUnA_2	563.0 / 269.0	3.50	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFDaA_1	613.0 / 569.0	3.79	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFDaA_2	613.0 / 319.0	3.79	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFTTrDA_1	663.0 / 619.0	4.04	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFTTrDA_2	663.0 / 169.0	4.03	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFTeDA_1	713.0 / 669.0	4.25	13C2-PFOA	415.0 / 370.0	37508.63	100.00
PFTeDA_2	713.0 / 169.0	4.25	13C2-PFOA	415.0 / 370.0	37508.63	100.00
NMeFOSAA_1	570.0 / 419.0	3.34	d3-MeFOSAA	573.0 / 419.0	33881.63	400.00
NMeFOSAA_2	570.0 / 512.0	3.34	d3-MeFOSAA	573.0 / 419.0	33881.63	400.00
NEtFOSAA_1	584.0 / 419.0	3.50	d3-MeFOSAA	573.0 / 419.0	33881.63	400.00
NEtFOSAA_2	584.0 / 483.0	3.49	d3-MeFOSAA	573.0 / 419.0	33881.63	400.00

# Raw Analytical Data

Sample Name	CQ755PB-FS(0)	Injection Vial	12
Sample ID	Procedural Blank	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:08:14	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	3618.96	5.789735	104.0	true
PFBS_2	298.9 / 99.0	1.46	1659.51	5.018315	69.3	true
PFHxA_1	313.0 / 269.0	N/A	N/A	N/A	N/A	true
PFHxA_2	313.0 / 119.0	N/A	N/A	N/A	N/A	true
PFHpA_1	363.0 / 319.0	N/A	N/A	N/A	N/A	true
PFHpA_2	363.0 / 169.0	N/A	N/A	N/A	N/A	true
PFHxS_1	399.0 / 80.0	N/A	N/A	N/A	N/A	true
PFHxS_2	399.0 / 99.0	N/A	N/A	N/A	N/A	true
PFOA_1	413.0 / 369.0	N/A	N/A	N/A	N/A	true
PFOA_2	413.0 / 169.0	N/A	N/A	N/A	N/A	true
PFNA_1	463.0 / 419.0	N/A	N/A	N/A	N/A	true
PFNA_2	463.0 / 219.0	N/A	N/A	N/A	N/A	true
PFOS_1	499.0 / 80.0	N/A	N/A	N/A	N/A	true
PFOS_2	499.0 / 99.0	N/A	N/A	N/A	N/A	true
PFDA_1	513.0 / 469.0	N/A	N/A	N/A	N/A	true
PFDA_2	513.0 / 219.0	N/A	N/A	N/A	N/A	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	CQ755PB-FS(0)	<b>Injection Vial</b>	12
<b>Sample ID</b>	Procedural Blank	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:08:14	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	107.373755	100.00	107.37
13C2-PFDA	515.0 / 470.0	3.20	107.599408	100.00	107.60
d5-EtFOSAA	589.0 / 419.0	3.51	364.611822	400.00	91.15

Sample Name	CQ756LCS-FS(0)	Injection Vial	13
Sample ID	Laboratory Control Sample	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:17:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	739940.00	3570.218503	1998.9	false
PFBS_2	298.9 / 99.0	1.46	230442.24	3601.947556	926.6	false
PFHxA_1	313.0 / 269.0	1.75	1066705.51	4011.593241	664.2	false
PFHxA_2	313.0 / 119.0	1.75	75479.78	3968.798756	489.5	false
PFHpA_1	363.0 / 319.0	2.10	942104.96	3721.881102	336.4	false
PFHpA_2	363.0 / 169.0	2.10	18626.34	3762.241261	380.4	false
PFHxS_1	399.0 / 80.0	2.12	893205.14	3783.824766	446.7	false
PFHxS_2	399.0 / 99.0	2.12	257388.16	3706.971183	419.3	false
PFOA_1	413.0 / 369.0	2.48	1102933.95	3792.757672	415.2	false
PFOA_2	413.0 / 169.0	2.48	76771.59	3953.190615	419.7	false
PFNA_1	463.0 / 419.0	2.85	1092069.39	3932.574515	371.2	false
PFNA_2	463.0 / 219.0	2.85	322479.50	3981.841831	423.3	false
PFOS_1	499.0 / 80.0	2.85	1127662.32	3451.357989	400.2	false
PFOS_2	499.0 / 99.0	2.85	250384.64	3962.140224	511.6	false
PFDA_1	513.0 / 469.0	3.21	1164254.84	3878.351394	447.5	false
PFDA_2	513.0 / 219.0	3.21	45034.24	3764.527111	433.3	false
PFUnA_1	563.0 / 519.0	3.52	1208903.31	3688.433102	324.5	false
PFUnA_2	563.0 / 269.0	3.52	56760.63	3666.217814	299.1	false
PFDoA_1	613.0 / 569.0	3.81	1219557.94	3724.450905	332.1	false
PFDoA_2	613.0 / 319.0	3.81	183591.34	3553.639405	302.3	false
PFTTrDA_1	663.0 / 619.0	4.06	1170703.15	3647.643772	414.8	false
PFTTrDA_2	663.0 / 169.0	4.06	75894.02	3720.851486	366.8	false
PFTeDA_1	713.0 / 669.0	4.28	1312778.30	4190.732313	775.1	false
PFTeDA_2	713.0 / 169.0	4.28	62529.43	4240.479647	642.7	false
NMeFOSAA_1	570.0 / 419.0	3.36	278836.52	4423.837299	466.8	false
NMeFOSAA_2	570.0 / 512.0	3.36	151766.72	3872.273482	380.6	false
NEtFOSAA_1	584.0 / 419.0	3.52	246636.87	4120.418777	293.3	false
NEtFOSAA_2	584.0 / 483.0	3.52	14476.71	3567.615885	233.2	false

<b>Sample Name</b>	CQ756LCS-FS(0)	<b>Injection Vial</b>	13
<b>Sample ID</b>	Laboratory Control Sample	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:17:09	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.73	105.351947	100.00	105.35
13C2-PFDA	515.0 / 470.0	3.19	103.605682	100.00	103.61
d5-EtFOSAA	589.0 / 419.0	3.51	379.680338	400.00	94.92



Sample Name	J6148-FS(0)	Injection Vial	14
Sample ID	NAWC-050718-RW-316	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:26:05	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	491659.18	1947.953430	311.9	false
PFBS_2	298.9 / 99.0	1.46	151011.73	1934.967632	478.4	false
PFHxA_1	313.0 / 269.0	1.74	1373573.21	4189.631616	404.5	false
PFHxA_2	313.0 / 119.0	1.74	93862.93	4000.785589	447.4	false
PFHpA_1	363.0 / 319.0	2.10	626938.85	2005.387237	226.2	false
PFHpA_2	363.0 / 169.0	2.09	16244.42	2646.052841	273.7	false
PFHxS_1	399.0 / 80.0	2.11	2054505.64	7158.366521	385.8	false
PFHxS_2	399.0 / 99.0	2.11	580157.25	6875.560120	418.2	false
PFOA_1	413.0 / 369.0	2.47	1649246.42	4601.630824	390.2	false
PFOA_2	413.0 / 169.0	2.45	164270.03	6868.953710	522.3	false
PFNA_1	463.0 / 419.0	2.85	222849.34	644.524888	185.0	false
PFNA_2	463.0 / 219.0	2.85	66331.69	655.612956	335.4	false
PFOS_1	499.0 / 80.0	2.80	3723209.66	9382.050761	370.9	false
PFOS_2	499.0 / 99.0	2.84	612334.86	7974.569699	394.8	false
PFDA_1	513.0 / 469.0	3.21	13865.55	16.087824	51.4	false
PFDA_2	513.0 / 219.0	3.22	1163.45	30.932675	44.5	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6148-FS(0)	<b>Injection Vial</b>	14
<b>Sample ID</b>	NAWC-050718-RW-316	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:26:05	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	103.889075	100.00	103.89
13C2-PFDA	515.0 / 470.0	3.19	86.512111	100.00	86.51
d5-EtFOSAA	589.0 / 419.0	3.51	359.644244	400.00	89.91

Sample Name	J6150-FS(0)	Injection Vial	15
Sample ID	NAWC-050718-RW-180	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:35:01	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	1052997.47	4177.270954	528.8	false
PFBS_2	298.9 / 99.0	1.46	299212.89	3845.095023	545.4	false
PFHxA_1	313.0 / 269.0	1.74	1266530.46	4179.895272	278.5	false
PFHxA_2	313.0 / 119.0	1.74	89660.84	4135.626213	335.3	false
PFHpA_1	363.0 / 319.0	2.09	717760.11	2485.423630	179.0	false
PFHpA_2	363.0 / 169.0	2.08	16847.16	2975.030226	227.9	false
PFHxS_1	399.0 / 80.0	2.11	1019368.18	3549.398177	252.9	false
PFHxS_2	399.0 / 99.0	2.11	291632.31	3452.047528	368.3	false
PFOA_1	413.0 / 369.0	2.47	1832522.86	5536.546085	419.9	false
PFOA_2	413.0 / 169.0	2.46	179301.23	8115.501545	455.7	false
PFNA_1	463.0 / 419.0	2.85	297033.80	932.716835	201.8	false
PFNA_2	463.0 / 219.0	2.85	90370.27	971.199647	299.0	false
PFOS_1	499.0 / 80.0	2.81	3221068.80	8111.210343	364.4	false
PFOS_2	499.0 / 99.0	2.84	523970.34	6819.045160	397.5	false
PFDA_1	513.0 / 469.0	3.20	42519.17	103.361219	101.9	false
PFDA_2	513.0 / 219.0	3.20	2480.19	135.280580	78.7	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

Sample Name	J6150-FS(0)	Injection Vial	15
Sample ID	NAWC-050718-RW-180	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:35:01	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	127.942285	100.00	127.94
13C2-PFDA	515.0 / 470.0	3.19	100.563028	100.00	100.56
d5-EtFOSAA	589.0 / 419.0	3.50	392.609589	400.00	98.15

Sample Name	J6152-FS(0)	Injection Vial	16
Sample ID	NAWC-050718-RW-275	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:43:56	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.47	740858.34	3597.646490	510.5	false
PFBS_2	298.9 / 99.0	1.47	211021.34	3318.487969	501.8	false
PFHxA_1	313.0 / 269.0	1.75	616878.28	2838.794437	210.0	false
PFHxA_2	313.0 / 119.0	1.75	45038.13	2909.761657	268.1	false
PFHpA_1	363.0 / 319.0	2.10	314268.63	1524.473118	116.6	false
PFHpA_2	363.0 / 169.0	2.09	8267.82	2033.360363	176.4	false
PFHxS_1	399.0 / 80.0	2.12	504810.17	2152.001942	199.7	false
PFHxS_2	399.0 / 99.0	2.11	143415.53	2076.874656	276.3	false
PFOA_1	413.0 / 369.0	2.47	997876.15	4223.950350	294.9	false
PFOA_2	413.0 / 169.0	2.46	88629.33	5621.658851	350.0	false
PFNA_1	463.0 / 419.0	2.85	181435.27	798.100022	134.7	false
PFNA_2	463.0 / 219.0	2.85	50015.61	751.712850	230.9	false
PFOS_1	499.0 / 80.0	2.82	1727685.49	5324.954281	271.5	false
PFOS_2	499.0 / 99.0	2.84	270926.78	4315.201535	373.6	false
PFDA_1	513.0 / 469.0	3.20	72882.30	278.788592	123.8	false
PFDA_2	513.0 / 219.0	3.20	3069.98	270.895390	85.9	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

Sample Name	J6152-FS(0)	Injection Vial	16
Sample ID	NAWC-050718-RW-275	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:43:56	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.73	138.099910	100.00	138.10
13C2-PFDA	515.0 / 470.0	3.19	113.558121	100.00	113.56
d5-EtFOSAA	589.0 / 419.0	3.50	451.454614	400.00	112.86

Sample Name	J6154-FS(0)	Injection Vial	17
Sample ID	NAWC-050718-RW-145	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:52:52	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	692865.80	2875.386216	568.7	false
PFBS_2	298.9 / 99.0	1.45	204013.62	2740.729404	538.3	false
PFHxA_1	313.0 / 269.0	1.73	773617.98	2574.545420	251.4	false
PFHxA_2	313.0 / 119.0	1.73	54361.14	2542.877145	312.7	false
PFHpA_1	363.0 / 319.0	2.09	520689.16	1831.019544	140.7	false
PFHpA_2	363.0 / 169.0	2.08	14390.65	2576.418927	267.7	false
PFHxS_1	399.0 / 80.0	2.11	657824.94	2397.749137	254.9	false
PFHxS_2	399.0 / 99.0	2.11	183902.44	2277.451508	338.0	false
PFOA_1	413.0 / 369.0	2.47	1700177.77	5219.066960	377.8	false
PFOA_2	413.0 / 169.0	2.45	164816.77	7580.135692	436.4	false
PFNA_1	463.0 / 419.0	2.84	253363.34	807.572480	178.7	false
PFNA_2	463.0 / 219.0	2.84	76908.78	838.626177	272.0	false
PFOS_1	499.0 / 80.0	2.78	1810729.34	4771.085409	264.5	true
PFOS_2	499.0 / 99.0	2.83	279256.34	3802.310425	416.6	false
PFDA_1	513.0 / 469.0	3.20	60877.89	160.208558	105.4	false
PFDA_2	513.0 / 219.0	3.20	3317.75	201.495536	93.7	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

Sample Name	J6154-FS(0)	Injection Vial	17
Sample ID	NAWC-050718-RW-145	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:52:52	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	128.245538	100.00	128.25
13C2-PFDA	515.0 / 470.0	3.19	96.267031	100.00	96.27
d5-EtFOSAA	589.0 / 419.0	3.50	390.764748	400.00	97.69



Sample Name	J6156-FS(0)	Injection Vial	18
Sample ID	NAWC-050718-RW-357	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:01:46	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	376958.24	2024.992866	488.7	false
PFBS_2	298.9 / 99.0	1.47	93256.98	1617.730744	411.7	false
PFHxA_1	313.0 / 269.0	1.75	589580.04	2715.094361	232.0	false
PFHxA_2	313.0 / 119.0	1.75	41885.64	2709.321114	278.7	false
PFHpA_1	363.0 / 319.0	2.10	238089.87	1155.513698	114.7	false
PFHpA_2	363.0 / 169.0	2.07	7361.56	1808.232979	146.1	false
PFHxS_1	399.0 / 80.0	2.11	689197.83	3255.160349	189.4	false
PFHxS_2	399.0 / 99.0	2.11	200400.88	3217.449450	327.3	false
PFOA_1	413.0 / 369.0	2.47	740744.24	3135.120947	286.7	false
PFOA_2	413.0 / 169.0	2.44	78278.79	4970.886430	355.6	false
PFNA_1	463.0 / 419.0	2.84	81424.66	354.782672	109.9	false
PFNA_2	463.0 / 219.0	2.84	27512.20	409.660235	170.0	false
PFOS_1	499.0 / 80.0	2.79	774088.06	2640.165754	284.4	false
PFOS_2	499.0 / 99.0	2.84	118636.20	2090.596109	348.1	false
PFDA_1	513.0 / 469.0	3.19	32453.08	112.385411	85.4	false
PFDA_2	513.0 / 219.0	3.20	1414.54	98.662746	76.5	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6156-FS(0)	<b>Injection Vial</b>	18
<b>Sample ID</b>	NAWC-050718-RW-357	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:01:46	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.73	137.169796	100.00	137.17
13C2-PFDA	515.0 / 470.0	3.18	108.201724	100.00	108.20
d5-EtFOSAA	589.0 / 419.0	3.50	320.444490	400.00	80.11

Sample Name	J6158-FS(0)	Injection Vial	19
Sample ID	NAWC-050718-RW-162	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:10:41	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	664193.93	2548.725884	417.4	false
PFBS_2	298.9 / 99.0	1.45	176542.83	2190.930097	521.0	false
PFHxA_1	313.0 / 269.0	1.74	614150.47	1897.544136	203.4	false
PFHxA_2	313.0 / 119.0	1.73	40643.41	1773.651028	242.7	false
PFHpA_1	363.0 / 319.0	2.09	336063.61	1103.097686	108.5	false
PFHpA_2	363.0 / 169.0	2.07	9175.60	1517.329565	202.7	false
PFHxS_1	399.0 / 80.0	2.10	1136496.70	3831.842207	245.3	false
PFHxS_2	399.0 / 99.0	2.11	324423.28	3718.803038	379.8	false
PFOA_1	413.0 / 369.0	2.46	1257625.99	3603.724955	297.1	false
PFOA_2	413.0 / 169.0	2.45	123648.04	5312.732676	425.3	false
PFNA_1	463.0 / 419.0	2.84	173191.97	513.626720	146.5	false
PFNA_2	463.0 / 219.0	2.83	51177.56	518.071074	197.8	false
PFOS_1	499.0 / 80.0	2.80	3522729.75	8590.013727	327.7	false
PFOS_2	499.0 / 99.0	2.83	562524.86	7088.965433	408.7	false
PFDA_1	513.0 / 469.0	3.19	21265.83	37.814056	63.6	false
PFDA_2	513.0 / 219.0	3.18	1405.11	50.246883	44.4	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6158-FS(0)	<b>Injection Vial</b>	19
<b>Sample ID</b>	NAWC-050718-RW-162	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:10:41	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	132.343573	100.00	132.34
13C2-PFDA	515.0 / 470.0	3.18	95.178040	100.00	95.18
d5-EtFOSAA	589.0 / 419.0	3.50	361.400056	400.00	90.35

Sample Name	J6160-FS(0)	Injection Vial	20
Sample ID	WGNA-050718-RW-0800	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:37:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	1549791.92	6592.804920	592.0	false
PFBS_2	298.9 / 99.0	1.46	473836.70	6535.265229	720.0	false
PFHxA_1	313.0 / 269.0	1.74	685906.75	2430.411100	261.3	false
PFHxA_2	313.0 / 119.0	1.74	45318.75	2258.283854	321.9	false
PFHpA_1	363.0 / 319.0	2.09	409310.98	1533.550069	149.5	false
PFHpA_2	363.0 / 169.0	2.07	11429.59	2174.213487	190.2	false
PFHxS_1	399.0 / 80.0	2.11	709744.59	2648.371629	240.1	false
PFHxS_2	399.0 / 99.0	2.11	202700.95	2570.310347	336.6	false
PFOA_1	413.0 / 369.0	2.47	1283978.61	4197.630527	361.5	false
PFOA_2	413.0 / 169.0	2.45	125393.44	6144.527286	404.6	false
PFNA_1	463.0 / 419.0	2.84	158917.85	537.598466	161.0	false
PFNA_2	463.0 / 219.0	2.84	49036.59	566.805964	261.0	false
PFOS_1	499.0 / 80.0	2.80	2087046.96	5630.574654	331.7	false
PFOS_2	499.0 / 99.0	2.83	336830.70	4696.214020	380.9	false
PFDA_1	513.0 / 469.0	3.19	47463.83	129.511268	108.6	false
PFDA_2	513.0 / 219.0	3.19	2104.59	120.412277	90.0	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6160-FS(0)	<b>Injection Vial</b>	20
<b>Sample ID</b>	WGNA-050718-RW-0800	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:37:26	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	121.380570	100.00	121.38
13C2-PFDA	515.0 / 470.0	3.18	106.902854	100.00	106.90
d5-EtFOSAA	589.0 / 419.0	3.50	386.001260	400.00	96.50

Sample Name	J6162-FS(0)	Injection Vial	21
Sample ID	WGNA-050718-RW-0335	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:46:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.45	875168.13	3169.002981	452.4	false
PFBS_2	298.9 / 99.0	1.45	239737.63	2809.867771	529.1	false
PFHxA_1	313.0 / 269.0	1.73	1173995.84	3464.896768	250.9	false
PFHxA_2	313.0 / 119.0	1.73	75847.32	3133.364717	318.7	false
PFHpA_1	363.0 / 319.0	2.08	633809.39	1966.730947	149.2	false
PFHpA_2	363.0 / 169.0	2.07	18590.11	2942.859335	288.4	false
PFHxS_1	399.0 / 80.0	2.10	667287.59	2121.701116	185.3	false
PFHxS_2	399.0 / 99.0	2.10	203190.16	2194.942516	343.8	false
PFOA_1	413.0 / 369.0	2.46	2240843.32	6072.091194	372.3	false
PFOA_2	413.0 / 169.0	2.45	225427.51	9150.310032	431.1	false
PFNA_1	463.0 / 419.0	2.83	269165.50	756.465486	174.7	false
PFNA_2	463.0 / 219.0	2.83	79761.19	766.468157	242.3	false
PFOS_1	499.0 / 80.0	2.78	2257255.11	5188.923366	344.5	true
PFOS_2	499.0 / 99.0	2.82	339225.91	4029.562082	397.4	false
PFDA_1	513.0 / 469.0	3.18	36331.39	74.143597	84.8	false
PFDA_2	513.0 / 219.0	3.19	1986.19	83.336061	55.4	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6162-FS(0)	<b>Injection Vial</b>	21
<b>Sample ID</b>	WGNA-050718-RW-0335	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:46:22	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.71	125.989562	100.00	125.99
13C2-PFDA	515.0 / 470.0	3.18	92.243518	100.00	92.24
d5-EtFOSAA	589.0 / 419.0	3.49	355.515117	400.00	88.88



Sample Name	J6164-FS(0)	Injection Vial	22
Sample ID	WGNA-050718-RW-3556	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:55:18	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	494358.31	2258.416754	527.6	false
PFBS_2	298.9 / 99.0	1.46	142447.20	2104.766402	480.8	false
PFHxA_1	313.0 / 269.0	1.74	413823.06	1480.072997	209.6	false
PFHxA_2	313.0 / 119.0	1.74	28162.91	1430.938479	270.9	false
PFHpA_1	363.0 / 319.0	2.09	222169.07	849.493859	114.1	false
PFHpA_2	363.0 / 169.0	2.04	8692.03	1681.526460	164.3	false
PFHxS_1	399.0 / 80.0	2.10	153978.92	617.852042	145.3	false
PFHxS_2	399.0 / 99.0	2.10	44108.24	598.524484	193.0	false
PFOA_1	413.0 / 369.0	2.46	607230.36	2020.727704	262.8	false
PFOA_2	413.0 / 169.0	2.44	64520.14	3227.163242	388.7	false
PFNA_1	463.0 / 419.0	2.83	95168.81	326.648508	119.0	false
PFNA_2	463.0 / 219.0	2.83	29242.71	341.978577	157.9	false
PFOS_1	499.0 / 80.0	2.77	533776.31	1545.237950	289.0	true
PFOS_2	499.0 / 99.0	2.83	81070.66	1212.248039	324.5	false
PFDA_1	513.0 / 469.0	3.18	26083.73	63.389017	78.9	false
PFDA_2	513.0 / 219.0	3.17	1283.89	56.776296	34.9	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6164-FS(0)	<b>Injection Vial</b>	22
<b>Sample ID</b>	WGNA-050718-RW-3556	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:55:18	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	116.857208	100.00	116.86
13C2-PFDA	515.0 / 470.0	3.18	82.503571	100.00	82.50
d5-EtFOSAA	589.0 / 419.0	3.49	362.383536	400.00	90.60

Sample Name	J6166-FS(0)	Injection Vial	23
Sample ID	NAWC-050718-RW-356	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:04:13	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	197206.89	875.460976	382.7	false
PFBS_2	298.9 / 99.0	1.46	54348.55	775.457930	421.6	false
PFHxA_1	313.0 / 269.0	1.74	293302.80	1172.740488	232.8	false
PFHxA_2	313.0 / 119.0	1.74	19615.22	1121.312617	252.2	false
PFHpA_1	363.0 / 319.0	2.09	174598.62	752.537168	110.0	false
PFHpA_2	363.0 / 169.0	2.06	4865.44	1044.629337	104.3	false
PFHxS_1	399.0 / 80.0	2.10	275010.76	1077.756044	177.7	false
PFHxS_2	399.0 / 99.0	2.11	78120.51	1038.099203	243.7	false
PFOA_1	413.0 / 369.0	2.46	690721.52	2598.761547	317.0	false
PFOA_2	413.0 / 169.0	2.44	73360.29	4144.000680	321.5	false
PFNA_1	463.0 / 419.0	2.83	75567.17	291.842620	111.3	false
PFNA_2	463.0 / 219.0	2.83	22257.35	292.216622	127.4	false
PFOS_1	499.0 / 80.0	2.78	771263.09	2182.319387	310.1	true
PFOS_2	499.0 / 99.0	2.83	122267.80	1787.534482	317.8	false
PFDA_1	513.0 / 469.0	3.20	4238.14	< 0	30.3	false
PFDA_2	513.0 / 219.0	3.15	310.75	< 0	18.4	true
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

Sample Name	J6166-FS(0)	Injection Vial	23
Sample ID	NAWC-050718-RW-356	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:04:13	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	131.880494	100.00	131.88
13C2-PFDA	515.0 / 470.0	3.18	108.867027	100.00	108.87
d5-EtFOSAA	589.0 / 419.0	3.49	421.132292	400.00	105.28

Sample Name	J6168-FS(0)	Injection Vial	24
Sample ID	NAWC-050718-RW-289	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:13:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.45	287344.43	1162.179888	478.1	false
PFBS_2	298.9 / 99.0	1.45	82724.53	1078.591443	440.9	false
PFHxA_1	313.0 / 269.0	1.73	197761.56	590.938194	200.0	false
PFHxA_2	313.0 / 119.0	1.73	13247.94	583.530667	232.8	false
PFHpA_1	363.0 / 319.0	2.08	125398.09	418.922555	98.6	false
PFHpA_2	363.0 / 169.0	2.06	3829.87	622.701231	104.9	false
PFHxS_1	399.0 / 80.0	2.10	294454.00	1049.340496	196.9	false
PFHxS_2	399.0 / 99.0	2.10	80597.88	973.671404	254.2	false
PFOA_1	413.0 / 369.0	2.46	525397.14	1531.923267	283.3	false
PFOA_2	413.0 / 169.0	2.44	53687.66	2355.908265	368.0	false
PFNA_1	463.0 / 419.0	2.83	87701.35	263.230182	128.1	false
PFNA_2	463.0 / 219.0	2.83	25716.87	262.089402	162.6	false
PFOS_1	499.0 / 80.0	2.79	1195178.06	3077.681798	299.4	false
PFOS_2	499.0 / 99.0	2.82	186196.99	2477.491062	393.7	false
PFDA_1	513.0 / 469.0	3.18	16579.71	25.901664	60.6	false
PFDA_2	513.0 / 219.0	3.16	810.00	9.689869	28.0	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

Sample Name	J6168-FS(0)	Injection Vial	24
Sample ID	NAWC-050718-RW-289	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:13:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	112.048828	100.00	112.05
13C2-PFDA	515.0 / 470.0	3.17	94.333578	100.00	94.33
d5-EtFOSAA	589.0 / 419.0	3.49	373.593815	400.00	93.40

Sample Name	J6170-FS(0)	Injection Vial	25
Sample ID	WGNA-050718-DUP-35	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:22:03	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.45	839374.24	3522.841322	537.8	false
PFBS_2	298.9 / 99.0	1.45	246017.58	3343.990435	553.8	false
PFHxA_1	313.0 / 269.0	1.73	727583.75	2552.064762	228.4	false
PFHxA_2	313.0 / 119.0	1.73	50465.90	2488.255940	308.1	false
PFHpA_1	363.0 / 319.0	2.08	382959.66	1418.515287	125.5	false
PFHpA_2	363.0 / 169.0	2.07	9582.53	1794.652635	168.6	false
PFHxS_1	399.0 / 80.0	2.10	596420.08	2197.558142	196.4	false
PFHxS_2	399.0 / 99.0	2.10	175386.41	2195.482684	358.5	false
PFOA_1	413.0 / 369.0	2.46	1145118.85	3699.685594	304.6	false
PFOA_2	413.0 / 169.0	2.44	104608.75	5066.209703	426.7	false
PFNA_1	463.0 / 419.0	2.83	216127.79	725.485204	144.6	false
PFNA_2	463.0 / 219.0	2.83	62576.72	717.894187	263.3	false
PFOS_1	499.0 / 80.0	2.80	2067764.56	5508.581505	296.4	false
PFOS_2	499.0 / 99.0	2.82	324766.48	4471.061567	383.1	false
PFDA_1	513.0 / 469.0	3.18	87828.75	254.885084	141.3	false
PFDA_2	513.0 / 219.0	3.18	3578.57	235.814380	124.4	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true

<b>Sample Name</b>	J6170-FS(0)	<b>Injection Vial</b>	25
<b>Sample ID</b>	WGNA-050718-DUP-35	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:22:03	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.71	133.638484	100.00	133.64
13C2-PFDA	515.0 / 470.0	3.17	104.605324	100.00	104.61
d5-EtFOSAA	589.0 / 419.0	3.48	439.768769	400.00	109.94



<b>Sample Name</b>	CQ755PB-FS(0)	<b>Injection Vial</b>	12
<b>Sample ID</b>	Procedural Blank	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:08:14	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.459	0.327	ü
PFHxA_1	313.0 / 269.0	N/A	PFHxA			
PFHxA_2	313.0 / 119.0	N/A	PFHxA	N/A	0.066	ü
PFHpA_1	363.0 / 319.0	N/A	PFHpA			
PFHpA_2	363.0 / 169.0	N/A	PFHpA	N/A	0.021	ü
PFHxS_1	399.0 / 80.0	N/A	PFHxS			
PFHxS_2	399.0 / 99.0	N/A	PFHxS	N/A	0.304	ü
PFOA_1	413.0 / 369.0	N/A	PFOA			
PFOA_2	413.0 / 169.0	N/A	PFOA	N/A	0.066	ü
PFNA_1	463.0 / 419.0	N/A	PFNA			
PFNA_2	463.0 / 219.0	N/A	PFNA	N/A	0.305	ü
PFOS_1	499.0 / 80.0	N/A	PFOS			
PFOS_2	499.0 / 99.0	N/A	PFOS	N/A	0.192	ü
PFDA_1	513.0 / 469.0	N/A	PFDA			
PFDA_2	513.0 / 219.0	N/A	PFDA	N/A	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	CQ756LCS-FS(0)	<b>Injection Vial</b>	13
<b>Sample ID</b>	Laboratory Control Sample	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:17:09	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.071	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.020	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.288	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.070	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.295	0.305	ü
PFOS_1	499.0 / 80.0	2.85	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.222	0.192	ü
PFDA_1	513.0 / 469.0	3.21	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.039	0.042	ü
PFUnA_1	563.0 / 519.0	3.52	PFUnA			
PFUnA_2	563.0 / 269.0	3.52	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.81	PFDaA			
PFDaA_2	613.0 / 319.0	3.81	PFDaA	0.151	0.162	ü
PFTrDA_1	663.0 / 619.0	4.06	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.06	PFTrDA	0.065	0.068	ü
PFTeDA_1	713.0 / 669.0	4.28	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.28	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.544	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.52	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.059	0.080	ü

<b>Sample Name</b>	J6148-FS(0)	<b>Injection Vial</b>	14
<b>Sample ID</b>	NAWC-050718-RW-316	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:26:05	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.307	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.068	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.09	PFHpA	0.026	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.282	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.100	0.066	
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.298	0.305	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.84	PFOS	0.165	0.192	ü
PFDA_1	513.0 / 469.0	3.21	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.084	0.042	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6150-FS(0)	<b>Injection Vial</b>	15
<b>Sample ID</b>	NAWC-050718-RW-180	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:35:01	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.284	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.071	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.08	PFHpA	0.024	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.286	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.46	PFOA	0.098	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.304	0.305	ü
PFOS_1	499.0 / 80.0	2.81	PFOS			
PFOS_2	499.0 / 99.0	2.84	PFOS	0.163	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.058	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6152-FS(0)	<b>Injection Vial</b>	16
<b>Sample ID</b>	NAWC-050718-RW-275	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:43:56	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.285	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.09	PFHpA	0.026	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.284	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.46	PFOA	0.089	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.276	0.305	ü
PFOS_1	499.0 / 80.0	2.82	PFOS			
PFOS_2	499.0 / 99.0	2.84	PFOS	0.157	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.042	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6154-FS(0)	<b>Injection Vial</b>	17
<b>Sample ID</b>	NAWC-050718-RW-145	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:52:52	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.294	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.070	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.08	PFHpA	0.028	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.280	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.097	0.066	ü
PFNA_1	463.0 / 419.0	2.84	PFNA			
PFNA_2	463.0 / 219.0	2.84	PFNA	0.304	0.305	ü
PFOS_1	499.0 / 80.0	2.78	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.154	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.055	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6156-FS(0)	<b>Injection Vial</b>	18
<b>Sample ID</b>	NAWC-050718-RW-357	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:01:46	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.247	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.071	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.031	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.291	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.106	0.066	
PFNA_1	463.0 / 419.0	2.84	PFNA			
PFNA_2	463.0 / 219.0	2.84	PFNA	0.338	0.305	ü
PFOS_1	499.0 / 80.0	2.79	PFOS			
PFOS_2	499.0 / 99.0	2.84	PFOS	0.153	0.192	ü
PFDA_1	513.0 / 469.0	3.19	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.044	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTTrDA_1	663.0 / 619.0	N/A	PFTTrDA			
PFTTrDA_2	663.0 / 169.0	N/A	PFTTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6158-FS(0)	<b>Injection Vial</b>	19
<b>Sample ID</b>	NAWC-050718-RW-162	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:10:41	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.266	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.066	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.027	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.286	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.098	0.066	ü
PFNA_1	463.0 / 419.0	2.84	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.296	0.305	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.160	0.192	ü
PFDA_1	513.0 / 469.0	3.19	PFDA			
PFDA_2	513.0 / 219.0	3.18	PFDA	0.066	0.042	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü



<b>Sample Name</b>	J6160-FS(0)	<b>Injection Vial</b>	20
<b>Sample ID</b>	WGNA-050718-RW-0800	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:37:26	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.306	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.066	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.028	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.286	0.304	ü
PFOA_1	413.0 / 369.0	2.47	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.098	0.066	ü
PFNA_1	463.0 / 419.0	2.84	PFNA			
PFNA_2	463.0 / 219.0	2.84	PFNA	0.309	0.305	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.161	0.192	ü
PFDA_1	513.0 / 469.0	3.19	PFDA			
PFDA_2	513.0 / 219.0	3.19	PFDA	0.044	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6162-FS(0)	<b>Injection Vial</b>	21
<b>Sample ID</b>	WGNA-050718-RW-0335	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:46:22	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.45	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.274	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.065	0.066	ü
PFHpA_1	363.0 / 319.0	2.08	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.029	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.10	PFHxS	0.305	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.101	0.066	
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.296	0.305	ü
PFOS_1	499.0 / 80.0	2.78	PFOS			
PFOS_2	499.0 / 99.0	2.82	PFOS	0.150	0.192	ü
PFDA_1	513.0 / 469.0	3.18	PFDA			
PFDA_2	513.0 / 219.0	3.19	PFDA	0.055	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6164-FS(0)	<b>Injection Vial</b>	22
<b>Sample ID</b>	WGNA-050718-RW-3556	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T14:55:18	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.288	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.068	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.04	PFHpA	0.039	0.021	
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.10	PFHxS	0.287	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.106	0.066	
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.307	0.305	ü
PFOS_1	499.0 / 80.0	2.77	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.152	0.192	ü
PFDA_1	513.0 / 469.0	3.18	PFDA			
PFDA_2	513.0 / 219.0	3.17	PFDA	0.049	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6166-FS(0)	<b>Injection Vial</b>	23
<b>Sample ID</b>	NAWC-050718-RW-356	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:04:13	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.276	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.067	0.066	ü
PFHpA_1	363.0 / 319.0	2.09	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.028	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.284	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.106	0.066	
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.295	0.305	ü
PFOS_1	499.0 / 80.0	2.78	PFOS			
PFOS_2	499.0 / 99.0	2.83	PFOS	0.159	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.15	PFDA	0.073	0.042	
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6168-FS(0)	<b>Injection Vial</b>	24
<b>Sample ID</b>	NAWC-050718-RW-289	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:13:09	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.45	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.288	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.067	0.066	ü
PFHpA_1	363.0 / 319.0	2.08	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.031	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.10	PFHxS	0.274	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.102	0.066	
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.293	0.305	ü
PFOS_1	499.0 / 80.0	2.79	PFOS			
PFOS_2	499.0 / 99.0	2.82	PFOS	0.156	0.192	ü
PFDA_1	513.0 / 469.0	3.18	PFDA			
PFDA_2	513.0 / 219.0	3.16	PFDA	0.049	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

<b>Sample Name</b>	J6170-FS(0)	<b>Injection Vial</b>	25
<b>Sample ID</b>	WGNA-050718-DUP-35	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Unknown	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T15:22:03	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.45	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.293	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.069	0.066	ü
PFHpA_1	363.0 / 319.0	2.08	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.025	0.021	ü
PFHxS_1	399.0 / 80.0	2.10	PFHxS			
PFHxS_2	399.0 / 99.0	2.10	PFHxS	0.294	0.304	ü
PFOA_1	413.0 / 369.0	2.46	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.091	0.066	ü
PFNA_1	463.0 / 419.0	2.83	PFNA			
PFNA_2	463.0 / 219.0	2.83	PFNA	0.290	0.305	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.82	PFOS	0.157	0.192	ü
PFDA_1	513.0 / 469.0	3.18	PFDA			
PFDA_2	513.0 / 219.0	3.18	PFDA	0.041	0.042	ü
PFUnA_1	563.0 / 519.0	N/A	PFUnA			
PFUnA_2	563.0 / 269.0	N/A	PFUnA	N/A	0.048	ü
PFDaA_1	613.0 / 569.0	N/A	PFDaA			
PFDaA_2	613.0 / 319.0	N/A	PFDaA	N/A	0.162	ü
PFTrDA_1	663.0 / 619.0	N/A	PFTrDA			
PFTrDA_2	663.0 / 169.0	N/A	PFTrDA	N/A	0.068	ü
PFTeDA_1	713.0 / 669.0	N/A	PFTeDA			
PFTeDA_2	713.0 / 169.0	N/A	PFTeDA	N/A	0.050	ü
NMeFOSAA_1	570.0 / 419.0	N/A	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	N/A	NMeFOSAA	N/A	0.664	ü
NEtFOSAA_1	584.0 / 419.0	N/A	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	N/A	NEtFOSSA	N/A	0.080	ü

Sample Name	CQ755PB-FS(0)	Injection Vial	12
Sample ID	Procedural Blank	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:08:14	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFHxA_1	313.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFHxA_2	313.0 / 119.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFHpA_1	363.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFHpA_2	363.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFHxS_1	399.0 / 80.0	N/A	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFHxS_2	399.0 / 99.0	N/A	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFOA_1	413.0 / 369.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFOA_2	413.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFNA_1	463.0 / 419.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFNA_2	463.0 / 219.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFOS_1	499.0 / 80.0	N/A	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFOS_2	499.0 / 99.0	N/A	13C4-PFOS	503.0 / 80.0	111731.01	287.00
PFDA_1	513.0 / 469.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFDA_2	513.0 / 219.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35848.65	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	35000.18	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	35000.18	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	35000.18	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	35000.18	400.00



Sample Name	CQ756LCS-FS(0)	Injection Vial	13
Sample ID	Laboratory Control Sample	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:17:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFHpA_2	363.0 / 169.0	2.10	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFHxS_2	399.0 / 99.0	2.12	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFOA_1	413.0 / 369.0	2.48	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFOA_2	413.0 / 169.0	2.48	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFOS_1	499.0 / 80.0	2.85	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFOS_2	499.0 / 99.0	2.85	13C4-PFOS	503.0 / 80.0	80170.42	287.00
PFDA_1	513.0 / 469.0	3.21	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFDA_2	513.0 / 219.0	3.21	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFUnA_1	563.0 / 519.0	3.52	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFUnA_2	563.0 / 269.0	3.52	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFDoA_1	613.0 / 569.0	3.81	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFDoA_2	613.0 / 319.0	3.81	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFTTrDA_1	663.0 / 619.0	4.06	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFTTrDA_2	663.0 / 169.0	4.06	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFTeDA_1	713.0 / 669.0	4.28	13C2-PFOA	415.0 / 370.0	29877.62	100.00
PFTeDA_2	713.0 / 169.0	4.28	13C2-PFOA	415.0 / 370.0	29877.62	100.00
NMeFOSAA_1	570.0 / 419.0	3.36	d3-MeFOSAA	573.0 / 419.0	26521.45	400.00
NMeFOSAA_2	570.0 / 512.0	3.36	d3-MeFOSAA	573.0 / 419.0	26521.45	400.00
NEtFOSAA_1	584.0 / 419.0	3.52	d3-MeFOSAA	573.0 / 419.0	26521.45	400.00
NEtFOSAA_2	584.0 / 483.0	3.52	d3-MeFOSAA	573.0 / 419.0	26521.45	400.00



Sample Name	J6148-FS(0)	Injection Vial	14
Sample ID	NAWC-050718-RW-316	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:26:05	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFHpA_2	363.0 / 169.0	2.09	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFOA_2	413.0 / 169.0	2.45	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFOS_1	499.0 / 80.0	2.80	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFOS_2	499.0 / 99.0	2.84	13C4-PFOS	503.0 / 80.0	97479.82	287.00
PFDA_1	513.0 / 469.0	3.21	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFDA_2	513.0 / 219.0	3.22	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	36858.30	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	28397.36	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	28397.36	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	28397.36	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	28397.36	400.00

Sample Name	J6150-FS(0)	Injection Vial	15
Sample ID	NAWC-050718-RW-180	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:35:01	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFHpA_2	363.0 / 169.0	2.08	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFOA_2	413.0 / 169.0	2.46	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFOS_1	499.0 / 80.0	2.81	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFOS_2	499.0 / 99.0	2.84	13C4-PFOS	503.0 / 80.0	97536.34	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	34064.10	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30315.36	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	30315.36	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30315.36	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	30315.36	400.00

Sample Name	J6152-FS(0)	Injection Vial	16
Sample ID	NAWC-050718-RW-275	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:43:56	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.47	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFHpA_2	363.0 / 169.0	2.09	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFHxS_1	399.0 / 80.0	2.12	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFOA_2	413.0 / 169.0	2.46	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFNA_1	463.0 / 419.0	2.85	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFNA_2	463.0 / 219.0	2.85	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFOS_1	499.0 / 80.0	2.82	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFOS_2	499.0 / 99.0	2.84	13C4-PFOS	503.0 / 80.0	79659.10	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	24285.52	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	25531.81	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	25531.81	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	25531.81	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	25531.81	400.00

Sample Name	J6154-FS(0)	Injection Vial	17
Sample ID	NAWC-050718-RW-145	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:52:52	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFHxA_1	313.0 / 269.0	1.73	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFHpA_2	363.0 / 169.0	2.08	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFOA_2	413.0 / 169.0	2.45	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFNA_1	463.0 / 419.0	2.84	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFNA_2	463.0 / 219.0	2.84	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFOS_1	499.0 / 80.0	2.78	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	93168.05	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	33518.97	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30882.72	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	30882.72	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30882.72	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	30882.72	400.00

Sample Name	J6156-FS(0)	Injection Vial	18
Sample ID	NAWC-050718-RW-357	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:01:46	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFBS_2	298.9 / 99.0	1.47	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFHxA_1	313.0 / 269.0	1.75	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFHxA_2	313.0 / 119.0	1.75	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFHpA_1	363.0 / 319.0	2.10	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFHpA_2	363.0 / 169.0	2.07	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFOA_2	413.0 / 169.0	2.44	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFNA_1	463.0 / 419.0	2.84	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFNA_2	463.0 / 219.0	2.84	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFOS_1	499.0 / 80.0	2.79	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFOS_2	499.0 / 99.0	2.84	13C4-PFOS	503.0 / 80.0	71904.49	287.00
PFDA_1	513.0 / 469.0	3.19	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFDA_2	513.0 / 219.0	3.20	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	24248.05	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	25387.66	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	25387.66	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	25387.66	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	25387.66	400.00

Sample Name	J6158-FS(0)	Injection Vial	19
Sample ID	NAWC-050718-RW-162	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:10:41	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFHpA_2	363.0 / 169.0	2.07	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFOA_2	413.0 / 169.0	2.45	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFNA_1	463.0 / 419.0	2.84	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFOS_1	499.0 / 80.0	2.80	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	100729.16	287.00
PFDA_1	513.0 / 469.0	3.19	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFDA_2	513.0 / 219.0	3.18	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	35845.05	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	35413.81	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	35413.81	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	35413.81	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	35413.81	400.00



Sample Name	J6160-FS(0)	Injection Vial	20
Sample ID	WGNA-050718-RW-0800	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:37:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFHpA_2	363.0 / 169.0	2.07	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFHxS_1	399.0 / 80.0	2.11	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFOA_1	413.0 / 369.0	2.47	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFOA_2	413.0 / 169.0	2.45	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFNA_1	463.0 / 419.0	2.84	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFNA_2	463.0 / 219.0	2.84	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFOS_1	499.0 / 80.0	2.80	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	91010.65	287.00
PFDA_1	513.0 / 469.0	3.19	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFDA_2	513.0 / 219.0	3.19	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	31443.44	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	26555.02	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	26555.02	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	26555.02	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	26555.02	400.00

Sample Name	J6162-FS(0)	Injection Vial	21
Sample ID	WGNA-050718-RW-0335	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:46:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.45	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFHxA_1	313.0 / 269.0	1.73	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFHpA_1	363.0 / 319.0	2.08	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFHpA_2	363.0 / 169.0	2.07	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFHxS_2	399.0 / 99.0	2.10	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFOA_2	413.0 / 169.0	2.45	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFOS_1	499.0 / 80.0	2.78	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFOS_2	499.0 / 99.0	2.82	13C4-PFOS	503.0 / 80.0	106801.49	287.00
PFDA_1	513.0 / 469.0	3.18	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFDA_2	513.0 / 219.0	3.19	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	37992.76	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	33982.85	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	33982.85	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	33982.85	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	33982.85	400.00



Sample Name	J6164-FS(0)	Injection Vial	22
Sample ID	WGNA-050718-RW-3556	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:55:18	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFHpA_2	363.0 / 169.0	2.04	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFHxS_2	399.0 / 99.0	2.10	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFOA_2	413.0 / 169.0	2.44	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFOS_1	499.0 / 80.0	2.77	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	84581.13	287.00
PFDA_1	513.0 / 469.0	3.18	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFDA_2	513.0 / 219.0	3.17	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	30729.65	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	24586.57	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	24586.57	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	24586.57	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	24586.57	400.00

Sample Name	J6166-FS(0)	Injection Vial	23
Sample ID	NAWC-050718-RW-356	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:04:13	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.46	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFBS_2	298.9 / 99.0	1.46	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFHxA_1	313.0 / 269.0	1.74	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFHxA_2	313.0 / 119.0	1.74	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFHpA_1	363.0 / 319.0	2.09	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFHpA_2	363.0 / 169.0	2.06	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFHxS_2	399.0 / 99.0	2.11	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFOA_2	413.0 / 169.0	2.44	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFOS_1	499.0 / 80.0	2.78	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFOS_2	499.0 / 99.0	2.83	13C4-PFOS	503.0 / 80.0	86631.83	287.00
PFDA_1	513.0 / 469.0	3.20	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFDA_2	513.0 / 219.0	3.15	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	27240.70	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	24181.64	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	24181.64	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	24181.64	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	24181.64	400.00

Sample Name	J6168-FS(0)	Injection Vial	24
Sample ID	NAWC-050718-RW-289	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:13:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.45	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFHxA_1	313.0 / 269.0	1.73	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFHpA_1	363.0 / 319.0	2.08	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFHpA_2	363.0 / 169.0	2.06	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFHxS_2	399.0 / 99.0	2.10	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFOA_2	413.0 / 169.0	2.44	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFOS_1	499.0 / 80.0	2.79	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFOS_2	499.0 / 99.0	2.82	13C4-PFOS	503.0 / 80.0	95267.28	287.00
PFDA_1	513.0 / 469.0	3.18	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFDA_2	513.0 / 219.0	3.16	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	34960.30	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30500.70	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	30500.70	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	30500.70	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	30500.70	400.00

Sample Name	J6170-FS(0)	Injection Vial	25
Sample ID	WGNA-050718-DUP-35	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:22:03	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			



## Results Summary

Analyte	MRM Transition	RT	IS	IS MRM Transition	IS Area	IS Conc. (ng/L)
PFBS_1	298.9 / 80.0	1.45	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFBS_2	298.9 / 99.0	1.45	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFHxA_1	313.0 / 269.0	1.73	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFHxA_2	313.0 / 119.0	1.73	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFHpA_1	363.0 / 319.0	2.08	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFHpA_2	363.0 / 169.0	2.07	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFHxS_1	399.0 / 80.0	2.10	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFHxS_2	399.0 / 99.0	2.10	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFOA_1	413.0 / 369.0	2.46	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFOA_2	413.0 / 169.0	2.44	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFNA_1	463.0 / 419.0	2.83	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFNA_2	463.0 / 219.0	2.83	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFOS_1	499.0 / 80.0	2.80	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFOS_2	499.0 / 99.0	2.82	13C4-PFOS	503.0 / 80.0	92164.56	287.00
PFDA_1	513.0 / 469.0	3.18	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFDA_2	513.0 / 219.0	3.18	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFUnA_1	563.0 / 519.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFUnA_2	563.0 / 269.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFDoA_1	613.0 / 569.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFDoA_2	613.0 / 319.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFTTrDA_1	663.0 / 619.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFTTrDA_2	663.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFTeDA_1	713.0 / 669.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
PFTeDA_2	713.0 / 169.0	N/A	13C2-PFOA	415.0 / 370.0	31796.45	100.00
NMeFOSAA_1	570.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	31134.03	400.00
NMeFOSAA_2	570.0 / 512.0	N/A	d3-MeFOSAA	573.0 / 419.0	31134.03	400.00
NEtFOSAA_1	584.0 / 419.0	N/A	d3-MeFOSAA	573.0 / 419.0	31134.03	400.00
NEtFOSAA_2	584.0 / 483.0	N/A	d3-MeFOSAA	573.0 / 419.0	31134.03	400.00

# Chromatograms


Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0




PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0




PFHpA\_2 363.0 / 169.0



PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0







PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0






Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

### Chromatograms







d5-EtFOSAA 589.0 / 419.0




Sample Name	JV65	Injection Vial	3
Sample ID	L2	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:38:49	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms




PFHxA\_1 313.0 / 269.0




PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0




PFHpA\_2 363.0 / 169.0






PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0









PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFU<sub>n</sub>A\_2 563.0 / 269.0



PFD<sub>o</sub>A\_1 613.0 / 569.0




PFD<sub>o</sub>A\_2 613.0 / 319.0







PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV65	Injection Vial	3
Sample ID	L2	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:38:49	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms







d5-EtFOSAA 589.0 / 419.0




Sample Name	JV66	Injection Vial	4
Sample ID	L3	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:47:45	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms




PFHxA\_1 313.0 / 269.0




PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0




PFHpA\_2 363.0 / 169.0



PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0







PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0






PFDoA\_1 613.0 / 569.0




PFDoA\_2 613.0 / 319.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0






Sample Name	JV66	Injection Vial	4
Sample ID	L3	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:47:45	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms








d5-EtFOSAA 589.0 / 419.0



Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:56:42	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms









PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0



NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0






NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:56:42	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0






Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:05:38	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms




PFHxA\_1 313.0 / 269.0




PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0




PFHpA\_2 363.0 / 169.0






PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0









PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0




PFU<sub>n</sub>A\_2 563.0 / 269.0



PFD<sub>o</sub>A\_1 613.0 / 569.0




PFD<sub>o</sub>A\_2 613.0 / 319.0







PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:05:38	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0






Sample Name	JV69	Injection Vial	7
Sample ID	L6	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:14:33	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms




PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0




PFDoA\_1 613.0 / 569.0




PFDoA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0



NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0






NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV69	Injection Vial	7
Sample ID	L6	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:14:33	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms








d5-EtFOSAA 589.0 / 419.0



Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			


## Chromatograms









PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0



PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0







PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0






PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:23:30	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms










d5-EtFOSAA 589.0 / 419.0




Sample Name	JV71	Injection Vial	9
Sample ID	L8	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:32:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHpA\_2 363.0 / 169.0



PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0







PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0






PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0









PFTTrDA\_1 663.0 / 619.0




PFTTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0




PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0






Sample Name	JV71	Injection Vial	9
Sample ID	L8	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:32:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	JV72	Injection Vial	10
Sample ID	L9	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:41:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0




PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0




PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0



PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0







PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0




PFU<sub>n</sub>A\_2 563.0 / 269.0



PFD<sub>o</sub>A\_1 613.0 / 569.0




PFD<sub>o</sub>A\_2 613.0 / 319.0







PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV72	Injection Vial	10
Sample ID	L9	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:41:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	CQ755PB-FS(0)	Injection Vial	12
Sample ID	Procedural Blank	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:08:14	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0



PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0







PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	CQ755PB-FS(0)	Injection Vial	12
Sample ID	Procedural Blank	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:08:14	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	CQ756LCS-FS(0)	Injection Vial	13
Sample ID	Laboratory Control Sample	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:17:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFU<sub>n</sub>A\_2 563.0 / 269.0




PFDoA\_1 613.0 / 569.0




PFDoA\_2 613.0 / 319.0




PFTrDA\_1 663.0 / 619.0



PFTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	CQ756LCS-FS(0)	Injection Vial	13
Sample ID	Laboratory Control Sample	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:17:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0






Sample Name	J6148-FS(0)	Injection Vial	14
Sample ID	NAWC-050718-RW-316	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:26:05	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms












PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFU<sub>n</sub>A\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6148-FS(0)	Injection Vial	14
Sample ID	NAWC-050718-RW-316	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:26:05	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6150-FS(0)	Injection Vial	15
Sample ID	NAWC-050718-RW-180	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:35:01	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6150-FS(0)	Injection Vial	15
Sample ID	NAWC-050718-RW-180	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:35:01	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6152-FS(0)	Injection Vial	16
Sample ID	NAWC-050718-RW-275	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:43:56	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




<b>Sample Name</b>	J6152-FS(0)	<b>Injection Vial</b>	16
<b>Sample ID</b>	NAWC-050718-RW-275	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Quality Control	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T13:43:56	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_SIS
<b>Sample Comment</b>			

## Chromatograms




d5-EtFOSAA 589.0 / 419.0




Sample Name	J6154-FS(0)	Injection Vial	17
Sample ID	NAWC-050718-RW-145	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:52:52	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms

PFBS\_1 298.9 / 80.0




PFBS\_2 298.9 / 99.0







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6154-FS(0)	Injection Vial	17
Sample ID	NAWC-050718-RW-145	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:52:52	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6156-FS(0)	Injection Vial	18
Sample ID	NAWC-050718-RW-357	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:01:46	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6156-FS(0)	Injection Vial	18
Sample ID	NAWC-050718-RW-357	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:01:46	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0






Sample Name	J6158-FS(0)	Injection Vial	19
Sample ID	NAWC-050718-RW-162	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:10:41	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			




## Chromatograms














PFOA\_1 413.0 / 369.0






PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0









PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6158-FS(0)	Injection Vial	19
Sample ID	NAWC-050718-RW-162	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:10:41	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0







PFU<sub>n</sub>A\_2 563.0 / 269.0




PFD<sub>o</sub>A\_1 613.0 / 569.0




PFD<sub>o</sub>A\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6160-FS(0)	Injection Vial	20
Sample ID	WGNA-050718-RW-0800	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:37:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0






PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0









PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6160-FS(0)	Injection Vial	20
Sample ID	WGNA-050718-RW-0800	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:37:26	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6162-FS(0)	Injection Vial	21
Sample ID	WGNA-050718-RW-0335	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:46:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0






PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0









PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6162-FS(0)	Injection Vial	21
Sample ID	WGNA-050718-RW-0335	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:46:22	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6164-FS(0)	Injection Vial	22
Sample ID	WGNA-050718-RW-3556	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:55:18	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0






PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0









PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFU<sub>n</sub>A\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0






PFDoA\_1 613.0 / 569.0




PFDoA\_2 613.0 / 319.0









PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6164-FS(0)	Injection Vial	22
Sample ID	WGNA-050718-RW-3556	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:55:18	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6166-FS(0)	Injection Vial	23
Sample ID	NAWC-050718-RW-356	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:04:13	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0






PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0









PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6166-FS(0)	Injection Vial	23
Sample ID	NAWC-050718-RW-356	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:04:13	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6168-FS(0)	Injection Vial	24
Sample ID	NAWC-050718-RW-289	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:13:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0



PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0







PFDA\_1 513.0 / 469.0




PFDA\_2 513.0 / 219.0




PFU<sub>n</sub>A\_1 563.0 / 519.0




PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTrDA\_1 663.0 / 619.0



PFTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6168-FS(0)	Injection Vial	24
Sample ID	NAWC-050718-RW-289	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:13:09	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0




Sample Name	J6170-FS(0)	Injection Vial	25
Sample ID	WGNA-050718-DUP-35	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:22:03	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms







PFHxA\_1 313.0 / 269.0



PFHxA\_2 313.0 / 119.0




PFHpA\_1 363.0 / 319.0







PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0




PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0




PFNA\_2 463.0 / 219.0




PFOS\_1 499.0 / 80.0




PFOS\_2 499.0 / 99.0




PFDA\_1 513.0 / 469.0



PFDA\_2 513.0 / 219.0




PFUnA\_1 563.0 / 519.0







PFUnA\_2 563.0 / 269.0




PFDaA\_1 613.0 / 569.0




PFDaA\_2 613.0 / 319.0




PFTrDA\_1 663.0 / 619.0



PFTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	J6170-FS(0)	Injection Vial	25
Sample ID	WGNA-050718-DUP-35	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:22:03	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms





d5-EtFOSAA 589.0 / 419.0






Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Chromatograms












PFHpA\_2 363.0 / 169.0




PFHxS\_1 399.0 / 80.0




PFHxS\_2 399.0 / 99.0




PFOA\_1 413.0 / 369.0






PFOA\_2 413.0 / 169.0




PFNA\_1 463.0 / 419.0







PFDA\_1 513.0 / 469.0






PFDA\_2 513.0 / 219.0




PFAUnA\_1 563.0 / 519.0









PFTTrDA\_1 663.0 / 619.0



PFTTrDA\_2 663.0 / 169.0




PFTeDA\_1 713.0 / 669.0







PFTeDA\_2 713.0 / 169.0




NMeFOSAA\_1 570.0 / 419.0




NMeFOSAA\_2 570.0 / 512.0



NEtFOSAA\_1 584.0 / 419.0





NEtFOSAA\_2 584.0 / 483.0




Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Chromatograms



d5-EtFOSAA 589.0 / 419.0



# Unused Data



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-316				
Battelle ID	J6148-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.70	0.22	0.50	2.50	
PFHpA	7.04	0.34	1.00	2.50	
PFOA	16.15	0.38	1.00	2.50	
PFNA	2.26 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	6.83	0.21	0.50	2.50	
PFHxS	25.12	0.34	1.00	2.50	
PFOS	32.92	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		104
13C5-PFPeA	40	87
13C5-PFHxA		90





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-180				
Battelle ID	J6150-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.93	0.22	0.50	2.50	
PFHpA	8.88	0.34	1.00	2.50	
PFOA	19.77	0.38	1.00	2.50	
PFNA	3.33	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	14.92	0.21	0.50	2.50	
PFHxS	12.68	0.34	1.00	2.50	
PFOS	28.97	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		128
13C5-PFPeA	40	101
13C5-PFHxA		98



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-275				
Battelle ID	J6152-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.14	0.22	0.50	2.50	
PFHpA	5.44	0.34	1.00	2.50	
PFOA	15.09	0.38	1.00	2.50	
PFNA	2.85	0.37	1.00	2.50	
PFDA	1.00 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	12.85	0.21	0.50	2.50	
PFHxS	7.69	0.34	1.00	2.50	
PFOS	19.02	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		138
13C5-PFPeA	40	114
13C5-PFHxA		113



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-145			
Battelle ID	J6154-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	9.19	0.22	0.50	2.50
PFHpA	6.54	0.34	1.00	2.50
PFOA	18.64	0.38	1.00	2.50
PFNA	2.88	0.37	1.00	2.50
PFDA	0.57 J	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	10.27	0.21	0.50	2.50
PFHxS	8.56	0.34	1.00	2.50
PFOS	17.04	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C4-PFBA		128
13C5-PFPeA	40	96
13C5-PFHxA		98



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-357				
Battelle ID	J6156-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.06	0.22	0.50	2.50	
PFHpA	4.28	0.34	1.00	2.50	
PFOA	11.61	0.38	1.00	2.50	
PFNA	1.31 J	0.37	1.00	2.50	
PFDA	0.42 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	7.50	0.21	0.50	2.50	
PFHxS	12.06	0.34	1.00	2.50	
PFOS	9.78	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C4-PFBA		137			
13C5-PFPeA	40	108			
13C5-PFHxA		80			



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-162				
Battelle ID	J6158-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	6.78	0.22	0.50	2.50	
PFHpA	3.94	0.34	1.00	2.50	
PFOA	12.87	0.38	1.00	2.50	
PFNA	1.83 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	9.10	0.21	0.50	2.50	
PFHxS	13.69	0.34	1.00	2.50	
PFOS	30.68	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		132
13C5-PFPeA	40	95
13C5-PFHxA		90



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0800				
Battelle ID	J6160-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	8.53	0.22	0.50	2.50	
PFHpA	5.38	0.34	1.00	2.50	
PFOA	14.73	0.38	1.00	2.50	
PFNA	1.89 J	0.37	1.00	2.50	
PFDA	0.45 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	23.13	0.21	0.50	2.50	
PFHxS	9.29	0.34	1.00	2.50	
PFOS	19.76	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		121
13C5-PFPeA	40	107
13C5-PFHxA		97





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0335				
Battelle ID	J6162-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	12.83	0.22	0.50	2.50	
PFHpA	7.28	0.34	1.00	2.50	
PFOA	22.49	0.38	1.00	2.50	
PFNA	2.80	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	11.74	0.21	0.50	2.50	
PFHxS	7.86	0.34	1.00	2.50	
PFOS	19.22	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		126
13C5-PFPeA	40	92
13C5-PFHxA		89



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-3556			
Battelle ID	J6164-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.260			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	5.69	0.22	0.50	2.50
PFHpA	3.27	0.34	1.00	2.50
PFOA	7.77	0.38	1.00	2.50
PFNA	1.26 J	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	8.69	0.21	0.50	2.50
PFHxS	2.38 J	0.34	1.00	2.50
PFOS	5.94	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C4-PFBA		117
13C5-PFPeA	40	83
13C5-PFHxA		91



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-356				
Battelle ID	J6166-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.275				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	4.26	0.22	0.50	2.50	
PFHpA	2.74	0.34	1.00	2.50	
PFOA	9.45	0.38	1.00	2.50	
PFNA	1.06 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	3.18	0.21	0.50	2.50	
PFHxS	3.92	0.34	1.00	2.50	
PFOS	7.94	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		132
13C5-PFPeA	40	109
13C5-PFHxA		105



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-289				
Battelle ID	J6168-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	2.11 J	0.22	0.50	2.50	
PFHpA	1.50 J	0.34	1.00	2.50	
PFOA	5.47	0.38	1.00	2.50	
PFNA	0.94 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	4.15	0.21	0.50	2.50	
PFHxS	3.75	0.34	1.00	2.50	
PFOS	10.99	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		112
13C5-PFPeA	40	94
13C5-PFHxA		93



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-DUP-35				
Battelle ID	J6170-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	9.11	0.22	0.50	2.50	
PFHpA	5.07	0.34	1.00	2.50	
PFOA	13.21	0.38	1.00	2.50	
PFNA	2.59	0.37	1.00	2.50	
PFDA	0.91 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	12.58	0.21	0.50	2.50	
PFHxS	7.85	0.34	1.00	2.50	
PFOS	19.67	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C4-PFBA		134
13C5-PFPeA	40	105
13C5-PFHxA		110



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Client ID	Procedural Blank			
Battelle ID	CQ755PB-FS			
Sample Type	PB			
Collection Date	05/10/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	WATER			
Sample Size	0.250			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	0.22 U	0.22	0.50	2.50
PFHpA	0.34 U	0.34	1.00	2.50
PFOA	0.38 U	0.38	1.00	2.50
PFNA	0.37 U	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	0.21 U	0.21	0.50	2.50
PFHxS	0.34 U	0.34	1.00	2.50
PFOS	0.30 U	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA		107
13C2-PFDA	40	108
d5-EtFOSAA		91





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	Laboratory Control Sample					
Battelle ID	CQ756LCS-FS					
Sample Type	LCS					
Collection Date	05/10/2018					
Extraction Date	05/10/2018					
Analysis Date	05/14/2018					
Analytical Instrument	Sciex 5500 LC/MS/MS					
% Moisture	NA					
Matrix	WATER					
Sample Size	0.250					
Size Unit-Basis	L					
Units	ng/L	Target	Recovery	Qual	Control Limits Lower	Upper
PFHxA	16.05	15.00	107		70	130
PFHpA	14.89	15.00	99		70	130
PFOA	15.17	15.00	101		70	130
PFNA	15.73	15.00	105		70	130
PFDA	15.51	15.00	103		70	130
PFUnA	14.75	15.00	98		70	130
PFDoA	14.90	15.00	99		70	130
PFTTrDA	14.59	15.00	97		70	130
PFTTeDA	16.76	15.00	112		70	130
NMeFOSAA	17.70	15.00	118		70	130
NEtFOSAA	16.48	15.00	110		70	130
PFBS	14.28	13.28	108		70	130
PFHxS	15.14	14.18	107		70	130
PFOS	13.81	14.33	96		70	130
<b>Surrogate Recoveries (%)</b>						
13C4-PFBA	105					
13C5-PFPeA	104					
13C5-PFHxA	95					



## Glossary of Data Qualifiers

Flag:      Application:

---

B	Analyte found in the sample at a concentration <10x the level found in the procedural blank
D	Dilution Run. Initial run outside the initial calibration range of the instrument
E	Estimate, result is greater than the highest concentration level in the calibration
H	Surrogate diluted out. Used when surrogate recovery is affected by excessive dilution of the sample extract.
J	Analyte detected below the Limit of Quantitation (LOQ)
ME	Significant Matrix Interference - Estimated value.
MI	Significant Matrix Interference - value could not be determined.
n	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO), but meets secondary criteria
N	Quality Control (QC) value is outside the accuracy or precision Data Quality Objective (DQO)
NA	Not Applicable
T	Holding Time (HT) exceeded
U	Analyte not detected or detected below the Method detection limit (MDL) value, MDL reported

## QA/QC Summary Batch 18-0313

Project:	CTO-WE04 Naval Air Station Joint Reserve Base Willow Grove
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	DW
Data Set:	DP-18-0112
Analytical SOP:	5-371
Method Reference:	USEPA 537 rev. 1.1, QSM 5.1

### Sample Custody

Collection Date	Receipt Date	Temp (°C)
5/7/2018	5/8/2018	0.9, 1.2

Corrective Actions	None
Sample Storage	The water samples were stored refrigerated until extraction.
Related samples	FRB samples associated with these samples are extracted in SDG 18-0316

### METHOD SUMMARIES

Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with methanol. Extracts were split and concentrated to dryness under nitrogen with a water bath set between 60 °C and 65 °C, reconstituted with 96:4 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.
Prep comments	None.
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.
Analysis Comments	Samples analyzed on the Sciex 5500. The confirmation ion ratio was above 50% RPD for the following samples and analytes: NAWC-050718-RW-316 for PFOA and PFDA NAWC-050718-RW-357 for PFOA NAWC-050718-RW-162 for PFDA WGNA-050718-RW-0335 for PFOA WGNA-050718-RW-3556 for PFHpA and PFOA NAWC -050718-RW-356 for PFOA and PFDA NAWC -050718-RW-289 for PFOA

Holding Times	Extraction Date(s)	Analysis Date(s)
	5/10/2018	5/14/2018

**QA/QC Summary**  
**Batch 18-0313**

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
$\leq 1/3$ the MRL	No exceedances noted. No comments.
Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
70-130% of true value	No exceedances noted. No comments.
Matrix Spike (MS) / Duplicate (MSD)	A MS/MSD were prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy. The relative percent difference was calculated to measure precision.
70-130% of true value, RPD $\leq 30\%$	No exceedances noted. MS/MSD samples were not collected for this SDG.
Surrogates Standard Analytes	Labelled surrogate compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
70-130% of true value	No exceedances noted. No Comments.
Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
ICal high and low points RPD $\leq 20\%$ , 50-150% of average area of the ICAL and 70-140% of most recent CCV	No exceedances noted. No Comments.
Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadratic curve fitting.
R <sup>2</sup> >0.99 Target and SIS compounds +/- 30% of true value, Low point 50-150% of true value	No exceedances noted. No comments.

**QA/QC Summary**  
**Batch 18-0313**

Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
Target and SIS compounds +/- 30% of true value	No exceedances noted.
	No comments.
Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.
Target and SIS compounds +/- 30% of true value Low point 50-150% of true value	No exceedances noted.
	No comments.



**It can be done**

Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Number: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0112  
 Test Code: Master\_371

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	0	None
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	NA	None
Matrix Spike / Matrix Spike Duplicate Precision	NA	None
Extracted Internal Standard Analytes (Surrogates)	0	None
Instrument Calibration	0	None
Instrument Blank	NA	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None



"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"307-24-4"	"PFHxA"	".500000"		
"ng/L"	"U"	".220000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	".500000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"375-85-9"	"PFHpA"	"1.000000"		
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"335-67-1"	"PFOA"	"1.000000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"375-95-1"	"PFNA"	"1.000000"		
"ng/L"	"U"	".370000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"335-76-2"	"PFDA"	"1.000000"		
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"2058-94-8"	"PFUnA"	"1.000000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"307-55-1"	"PFDoA"	"1.000000"		
"ng/L"	"U"	".420000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	"1.000000"		
"ng/L"	"U"	".420000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"376-06-7"	"PFTeDA"	"1.500000"		
"ng/L"	"U"	".730000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.500000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"			
"1.000000"	"ng/L"	"U"	".420000"	"MDL"	""	"T"	""	""	"2.500000"
"LOQ"	"YES"	"-99.000000"	""						
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	"1.000000"		
"ng/L"	"U"	".440000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"375-73-5"	"PFBS"	".500000"	"ng/L"	
"U"	".210000"	"MDL"	""	"T"	""	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".000500"	".500000"	""							".250000"
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"1763-23-1"	"PFOS"	"1.000000"		
"ng/L"	"U"	".300000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"355-46-4"	"PFHxS"	"1.000000"		
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	"1.000000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"	".430000"		
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"107.00"	""	"-99.000000"	"NA"
"YES"	"-99.000000"	""							
".250000"	".000500"	".500000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".430000"		
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"108.00"	""	"-99.000000"	"NA"
"YES"	"-99.000000"	""							
".250000"	".000500"	".500000"	""						
"CQ755PB-FS"	"SOP 5-369"	"Initial"	"CQ755PB-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"			
"1.460000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"91.00"	""	"-99.000000"
"YES"	"-99.000000"	""							
".250000"	".000500"	".500000"	""						
"CQ756LCS-FS"	"SOP 5-369"	"Initial"	"CQ756LCS-FS"	"BNO"	"307-24-4"	"PFHxA"	"16.050000"		
"ng/L"	""	".220000"	"MDL"	""	"T"	"107.00"	""	"2.500000"	"LOQ"
"YES"	"-99.000000"	""							
".250000"	".000500"	".500000"	""						









".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"2058-94-8"	"PFUnA"	".890000"		
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"307-55-1"	"PFDoA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"376-06-7"	"PFTeDA"			
"1.340000"	"ng/L"	"U"	".650000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	"1.340000"	""				
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"			
".890000"	"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".890000"		
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"375-73-5"	"PFBS"	"10.270000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".450000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"1763-23-1"	"PFOS"	"17.040000"		
"ng/L"	""	".270000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"355-46-4"	"PFHxS"	"8.560000"		
"ng/L"	""	".300000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"			
".460000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"128.00"	""	"-99.000000"
".360000"	""	".280000"	".000500"	".500000"	""				
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".340000"		
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"96.00"	""	"-99.000000"	"NA"
""	".280000"	".000500"	".500000"	""					
"NAWC-050718-RW-145"	"SOP 5-369"	"Initial"	"J6154-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"			
"1.400000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"98.00"	""	"-99.000000"
"1.430000"	""	".280000"	".000500"	".500000"	""				
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"307-24-4"	"PFHxA"	"10.060000"		
"ng/L"	""	".200000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
".270000"	".000500"	".460000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"375-85-9"	"PFHpA"	"4.280000"		
"ng/L"	""	".310000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
".270000"	".000500"	".930000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"335-67-1"	"PFOA"	"11.610000"		
"ng/L"	""	".350000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
".270000"	".000500"	".930000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"375-95-1"	"PFNA"	"1.310000"		
"ng/L"	"J"	".340000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
".270000"	".000500"	".930000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"335-76-2"	"PFDA"	".420000"		
"ng/L"	"J"	".360000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
".270000"	".000500"	".930000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"2058-94-8"	"PFUnA"	".930000"		
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"



".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"307-55-1"	"PFDoA"	".930000"			
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"72629-94-8"	"PFTrDA"	".930000"			
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"376-06-7"	"PFTeDA"				
"1.390000"	"ng/L"	"U"	".680000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
"-99.000000"	""	".270000"	".000500"	"1.390000"	""					
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"				
".930000"	"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"
"-99.000000"	""	".270000"	".000500"	".930000"	""					
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".930000"			
"ng/L"	"U"	".410000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"375-73-5"	"PFBS"	"7.500000"			
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".460000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"1763-23-1"	"PFOS"	"9.780000"			
"ng/L"	""	".280000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"355-46-4"	"PFHxS"	"12.060000"			
"ng/L"	""	".310000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"
".270000"	".000500"	".930000"	""							
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"				
".510000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"137.00"	""	"-99.000000"	"NA"
".370000"	""	".270000"	".000500"	".500000"	""					
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".400000"			
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"108.00"	""	"-99.000000"	"NA"	"YES"
""	".270000"	".000500"	".500000"	""						
"NAWC-050718-RW-357"	"SOP 5-369"	"Initial"	"J6156-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"				
"1.190000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"80.00"	""	"-99.000000"	"NA"
"1.480000"	""	".270000"	".000500"	".500000"	""					
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"307-24-4"	"PFHxA"	"6.780000"			
"ng/L"	""	".200000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".450000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"375-85-9"	"PFHpA"	"3.940000"			
"ng/L"	""	".300000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".890000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"335-67-1"	"PFOA"	"12.870000"			
"ng/L"	""	".340000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".890000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"375-95-1"	"PFNA"	"1.830000"			
"ng/L"	"J"	".330000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".890000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"335-76-2"	"PFDA"	".890000"			
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".890000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"2058-94-8"	"PFUnA"	".890000"			
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"
".280000"	".000500"	".890000"	""							
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"307-55-1"	"PFDoA"	".890000"			
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"	"-99.000000"

".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"376-06-7"	"PFTeDA"			
"1.340000"	"ng/L"	"U"	".650000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	".1.340000"	""				
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"			
".890000"	"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".890000"		
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"375-73-5"	"PFBS"	"9.100000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".450000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"1763-23-1"	"PFOS"	"30.680000"		
"ng/L"	""	".270000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"355-46-4"	"PFHxS"	"13.690000"		
"ng/L"	""	".300000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"			
".470000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"132.00"	""	"-99.000000"
".360000"	""	".280000"	".000500"	".500000"	""				
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".340000"		
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"95.00"	"-99.000000"	"NA"	"YES"
""	".280000"	".000500"	".500000"	""					
"NAWC-050718-RW-162"	"SOP 5-369"	"Initial"	"J6158-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"			
"1.290000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"90.00"	""	"-99.000000"
"1.430000"	""	".280000"	".000500"	".500000"	""				
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"307-24-4"	"PFHxA"	"8.530000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".440000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"375-85-9"	"PFHpA"	"5.380000"		
"ng/L"	""	".300000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"335-67-1"	"PFOA"	"14.730000"		
"ng/L"	""	".330000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"375-95-1"	"PFNA"	"1.890000"		
"ng/L"	"J"	".320000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"335-76-2"	"PFDA"	".450000"		
"ng/L"	"J"	".340000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"2058-94-8"	"PFUnA"	".880000"		
"ng/L"	"U"	".330000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"307-55-1"	"PFDoA"	".880000"		
"ng/L"	"U"	".370000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"
".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".880000"		
"ng/L"	"U"	".370000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"

".285000"	".000500"	".880000"	""								
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"376-06-7"	"PFTeDA"					
"1.320000"	"ng/L"	"U"	".640000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	
"-99.000000"	""	".285000"	".000500"	".1320000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"					
".880000"	"ng/L"	"U"	".370000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	
"-99.000000"	""	".285000"	".000500"	".880000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".880000"				
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	"-99.000000"	
".285000"	".000500"	".880000"	""								
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"375-73-5"	"PFBS"	"23.130000"				
"ng/L"	""	".180000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	"-99.000000"	
".285000"	".000500"	".440000"	""								
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"1763-23-1"	"PFOS"	"19.760000"				
"ng/L"	""	".260000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	"-99.000000"	
".285000"	".000500"	".880000"	""								
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"355-46-4"	"PFHxS"	"9.290000"				
"ng/L"	""	".300000"	"MDL"	""	"T"	""	"2.190000"	"LOQ"	"YES"	"-99.000000"	
".285000"	".000500"	".880000"	""								
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"					
".430000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"121.00"	""	"-99.000000"	"NA"	"YES"
".350000"	""	".285000"	".000500"	".500000"	""						
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".380000"				
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"107.00"	""	"-99.000000"	"NA"	"YES"	".350000"
""	".285000"	".000500"	".500000"	""							
"WGNA-050718-RW-0800"	"SOP 5-369"	"Initial"	"J6160-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"					
"1.350000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"97.00"	""	"-99.000000"	"NA"	"YES"
"1.400000"	""	".285000"	".000500"	".500000"	""						
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"307-24-4"	"PFHxA"	"12.830000"				
"ng/L"	""	".200000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".460000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"375-85-9"	"PFHpA"	"7.280000"				
"ng/L"	""	".310000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"335-67-1"	"PFOA"	"22.490000"				
"ng/L"	""	".350000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"375-95-1"	"PFNA"	"2.800000"				
"ng/L"	""	".340000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"335-76-2"	"PFDA"	".930000"				
"ng/L"	"U"	".360000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"2058-94-8"	"PFUnA"	".930000"				
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"307-55-1"	"PFDoA"	".930000"				
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".930000"				
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	"-99.000000"	
".270000"	".000500"	".930000"	""								
"WGNA-050718-RW-0335"	"SOP 5-369"	"Initial"	"J6162-FS"	"BNO"	"376-06-7"	"PFTeDA"					
"1.390000"	"ng/L"	"U"	".680000"	"MDL"	""	"T"	""	"2.310000"	"LOQ"	"YES"	

"-99.000000" "" ".270000" ".000500" "1.390000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "2355-31-9" "NMeFOSAA"  
 ".930000" "ng/L" "U" ".390000" "MDL" "" "T" "" "" "2.310000" "LOQ" "YES"  
 "-99.000000" "" ".270000" ".000500" ".930000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "2991-50-6" "NEtFOSAA" ".930000"  
 "ng/L" "U" ".410000" "MDL" "" "T" "" "" "2.310000" "LOQ" "YES" "-99.000000" ""  
 ".270000" ".000500" ".930000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "375-73-5" "PFBS" "11.740000"  
 "ng/L" "" ".190000" "MDL" "" "T" "" "" "2.310000" "LOQ" "YES" "-99.000000" ""  
 ".270000" ".000500" ".460000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "1763-23-1" "PFOS" "19.220000"  
 "ng/L" "" ".280000" "MDL" "" "T" "" "" "2.310000" "LOQ" "YES" "-99.000000" ""  
 ".270000" ".000500" ".930000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "355-46-4" "PFHxS" "7.860000"  
 "ng/L" "" ".310000" "MDL" "" "T" "" "" "2.310000" "LOQ" "YES" "-99.000000" ""  
 ".270000" ".000500" ".930000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "BDO-2106" "13C2-PFHxA"  
 ".470000" "ng/L" "" "-99.000000" "NA" "" "SIS" "126.00" "" "-99.000000" "NA" "YES"  
 ".370000" "" ".270000" ".000500" ".500000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "BDO-2110" "13C2-PFDA" ".340000"  
 "ng/L" "" "-99.000000" "NA" "" "SIS" "92.00" "" "-99.000000" "NA" "YES" ".370000"  
 "" ".270000" ".000500" ".500000" ""  
 "WGNA-050718-RW-0335" "SOP 5-369" "Initial" "J6162-FS" "BNO" "BDO-1839" "d5-EtFOSAA"  
 "1.320000" "ng/L" "" "-99.000000" "NA" "" "SIS" "89.00" "" "-99.000000" "NA" "YES"  
 "1.480000" "" ".270000" ".000500" ".500000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "307-24-4" "PFHxA" "5.690000"  
 "ng/L" "" ".210000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".480000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "375-85-9" "PFHpA" "3.270000"  
 "ng/L" "" ".330000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "335-67-1" "PFOA" "7.770000"  
 "ng/L" "" ".370000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "375-95-1" "PFNA" "1.260000"  
 "ng/L" "J" ".360000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "335-76-2" "PFDA" ".960000"  
 "ng/L" "U" ".380000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "2058-94-8" "PFUnA" ".960000"  
 "ng/L" "U" ".370000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "307-55-1" "PFDoA" ".960000"  
 "ng/L" "U" ".400000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "72629-94-8" "PFTTrDA" ".960000"  
 "ng/L" "U" ".400000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES" "-99.000000" ""  
 ".260000" ".000500" ".960000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "376-06-7" "PFTeDA"  
 "1.440000" "ng/L" "U" ".700000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES"  
 "-99.000000" "" ".260000" ".000500" "1.440000" ""  
 "WGNA-050718-RW-3556" "SOP 5-369" "Initial" "J6164-FS" "BNO" "2355-31-9" "NMeFOSAA"  
 ".960000" "ng/L" "U" ".400000" "MDL" "" "T" "" "" "2.400000" "LOQ" "YES"

"-99.000000"	""	".260000"	".000500"	".960000"	""						
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".960000"				
"ng/L"	"U"	".420000"	"MDL"	""	"T"	""	"2.400000"	"LOQ"	"YES"	"-99.000000"	""
".260000"	".000500"	".960000"	""								
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"375-73-5"	"PFBS"	"8.690000"				
"ng/L"	""	".200000"	"MDL"	""	"T"	""	"2.400000"	"LOQ"	"YES"	"-99.000000"	""
".260000"	".000500"	".480000"	""								
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"1763-23-1"	"PFOS"	"5.940000"				
"ng/L"	""	".290000"	"MDL"	""	"T"	""	"2.400000"	"LOQ"	"YES"	"-99.000000"	""
".260000"	".000500"	".960000"	""								
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"355-46-4"	"PFHxS"	"2.380000"				
"ng/L"	"J"	".330000"	"MDL"	""	"T"	""	"2.400000"	"LOQ"	"YES"	"-99.000000"	""
".260000"	".000500"	".960000"	""								
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"					
".450000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"117.00"	""	"-99.000000"	"NA"	"YES"
".380000"	""	".260000"	".000500"	".500000"	""						
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".320000"				
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"83.00"	""	"-99.000000"	"NA"	"YES"	".380000"
""	".260000"	".000500"	".500000"	""							
"WGNA-050718-RW-3556"	"SOP 5-369"	"Initial"	"J6164-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"					
"1.390000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"91.00"	""	"-99.000000"	"NA"	"YES"
"1.540000"	""	".260000"	".000500"	".500000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"307-24-4"	"PFHxA"	"4.260000"				
"ng/L"	""	".200000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".450000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"375-85-9"	"PFHpA"	"2.740000"				
"ng/L"	""	".310000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"335-67-1"	"PFOA"	"9.450000"				
"ng/L"	""	".350000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"375-95-1"	"PFNA"	"1.060000"				
"ng/L"	"J"	".340000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"335-76-2"	"PFDA"	".910000"				
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"2058-94-8"	"PFUnA"	".910000"				
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"307-55-1"	"PFDoA"	".910000"				
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"72629-94-8"	"PFTrDA"	".910000"				
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""
".275000"	".000500"	".910000"	""								
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"376-06-7"	"PFTeDA"					
"1.360000"	"ng/L"	"U"	".660000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	
"-99.000000"	""	".275000"	".000500"	"1.360000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"					
".910000"	"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	
"-99.000000"	""	".275000"	".000500"	".910000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".910000"				
"ng/L"	"U"	".400000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"	"-99.000000"	""

".275000"	".000500"	".910000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"375-73-5"	"PFBS"	"3.180000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"
".275000"	".000500"	".450000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"1763-23-1"	"PFOS"	"7.940000"		
"ng/L"	""	".270000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"
".275000"	".000500"	".910000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"355-46-4"	"PFHxS"	"3.920000"		
"ng/L"	""	".310000"	"MDL"	""	"T"	""	"2.270000"	"LOQ"	"YES"
".275000"	".000500"	".910000"	""						
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"			
".480000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"132.00"	""	"-99.000000"
".360000"	""	".275000"	".000500"	".500000"	""				
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".400000"		
"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"109.00"	""	"-99.000000"	"NA"
""	".275000"	".000500"	".500000"	""					
"NAWC-050718-RW-356"	"SOP 5-369"	"Initial"	"J6166-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"			
"1.530000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"105.00"	""	"-99.000000"
"1.450000"	""	".275000"	".000500"	".500000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"307-24-4"	"PFHxA"	"2.110000"		
"ng/L"	"J"	".200000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".450000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"375-85-9"	"PFHpA"	"1.500000"		
"ng/L"	"J"	".300000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"335-67-1"	"PFOA"	"5.470000"		
"ng/L"	""	".340000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"375-95-1"	"PFNA"	".940000"		
"ng/L"	"J"	".330000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"335-76-2"	"PFDA"	".890000"		
"ng/L"	"U"	".350000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"2058-94-8"	"PFUnA"	".890000"		
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"307-55-1"	"PFDoA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"376-06-7"	"PFTeDA"			
"1.340000"	"ng/L"	"U"	".650000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	"1.340000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"			
".890000"	"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".890000"		
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"
".280000"	".000500"	".890000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"375-73-5"	"PFBS"	"4.150000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	"2.230000"	"LOQ"	"YES"



".280000"	".000500"	".450000"	""						
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"1763-23-1"	"PFOS"	"10.990000"		
"ng/L"	""	".270000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"355-46-4"	"PFHxS"	"3.750000"		
"ng/L"	""	".300000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"BDO-2106"	"13C2-PFHxA"			
".400000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"112.00"	""	"-99.000000"
"NA"	"YES"								
".360000"	""	".280000"	".000500"	".500000"	""				
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"BDO-2110"	"13C2-PFDA"	".340000"		
"ng/L"	""	"-99.000000"	"NA"	"SIS"	"94.00"	"-99.000000"	"NA"	"YES"	".360000"
""	".280000"	".000500"	".500000"	""					
"NAWC-050718-RW-289"	"SOP 5-369"	"Initial"	"J6168-FS"	"BNO"	"BDO-1839"	"d5-EtFOSAA"			
"1.330000"	"ng/L"	""	"-99.000000"	"NA"	""	"SIS"	"93.00"	""	"-99.000000"
"NA"	"YES"								
"1.430000"	""	".280000"	".000500"	".500000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"307-24-4"	"PFHxA"	"9.110000"		
"ng/L"	""	".200000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".450000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"375-85-9"	"PFHpA"	"5.070000"		
"ng/L"	""	".300000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"335-67-1"	"PFOA"	"13.210000"		
"ng/L"	""	".340000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"375-95-1"	"PFNA"	"2.590000"		
"ng/L"	""	".330000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"335-76-2"	"PFDA"	".910000"		
"ng/L"	"J"	".350000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"2058-94-8"	"PFUnA"	".890000"		
"ng/L"	"U"	".340000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"307-55-1"	"PFDoA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"72629-94-8"	"PFTTrDA"	".890000"		
"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"376-06-7"	"PFTeDA"			
"1.340000"	"ng/L"	"U"	".650000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"
"YES"									
"-99.000000"	""	".280000"	".000500"	".1.340000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"2355-31-9"	"NMeFOSAA"			
".890000"	"ng/L"	"U"	".380000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"
"YES"									
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"2991-50-6"	"NEtFOSAA"	".890000"		
"ng/L"	"U"	".390000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".890000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"375-73-5"	"PFBS"	"12.580000"		
"ng/L"	""	".190000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""	".280000"	".000500"	".450000"	""				
"WGNA-050718-DUP-35"	"SOP 5-369"	"Initial"	"J6170-FS"	"BNO"	"1763-23-1"	"PFOS"	"19.670000"		
"ng/L"	""	".270000"	"MDL"	""	"T"	""	".2.230000"	"LOQ"	"YES"
"-99.000000"	""								

".280000" ".000500" ".890000" ""  
"WGNA-050718-DUP-35" "SOP 5-369" "Initial" "J6170-FS" "BNO" "355-46-4" "PFHxS" "7.850000"  
"ng/L" "" ".300000" "MDL" "" "T" "" "" "2.230000" "LOQ" "YES" "-99.000000" ""  
".280000" ".000500" ".890000" ""  
"WGNA-050718-DUP-35" "SOP 5-369" "Initial" "J6170-FS" "BNO" "BDO-2106" "13C2-PFHxA"  
".480000" "ng/L" "" "-99.000000" "NA" "" "SIS" "134.00" "" "-99.000000" "NA" "YES"  
".360000" "" ".280000" ".000500" ".500000" ""  
"WGNA-050718-DUP-35" "SOP 5-369" "Initial" "J6170-FS" "BNO" "BDO-2110" "13C2-PFDA" ".370000"  
"ng/L" "" "-99.000000" "NA" "" "SIS" "105.00" "" "-99.000000" "NA" "YES" ".360000"  
"" ".280000" ".000500" ".500000" ""  
"WGNA-050718-DUP-35" "SOP 5-369" "Initial" "J6170-FS" "BNO" "BDO-1839" "d5-EtFOSAA"  
"1.570000" "ng/L" "" "-99.000000" "NA" "" "SIS" "110.00" "" "-99.000000" "NA" "YES"  
"1.430000" "" ".280000" ".000500" ".500000" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "CQ755PB-FS" "" "WATER" "CQ755PB-FS"  
"Method Bla" "" "-99.000000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53" "05/14/2018 13:08"  
"BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313" "18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/10/2018 10:53" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "CQ756LCS-FS" "" "WATER" "CQ756LCS-FS"  
"LCS" "" "-99.000000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53" "05/14/2018 13:17" "BNO"  
"COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313" "18-0313" "DP-18-0112"  
"DP-18-0112" "18-0313" "05/10/2018 10:53" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-316" "05/07/2018 10:10" "DW"  
"J6148-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 13:26" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-180" "05/07/2018 10:40" "DW"  
"J6150-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 13:35" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-275" "05/07/2018 11:10" "DW"  
"J6152-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 13:43" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-145" "05/07/2018 12:40" "DW"  
"J6154-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 13:52" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-357" "05/07/2018 13:10" "DW"  
"J6156-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 14:01" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-162" "05/07/2018 13:40" "DW"  
"J6158-FS" "NM" "SHP-180508-02" ".900000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 14:10" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "WGNA-050718-RW-0800" "05/07/2018 14:10" "DW"  
"J6160-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 14:37" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "WGNA-050718-RW-0335" "05/07/2018 14:40" "DW"  
"J6162-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 14:46" "BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30" "07/13/2018 14:27" ""  
"112G08005-WE04" "WE04 NAS Willow Grove" "WGNA-050718-RW-3556" "05/07/2018 15:10" "DW"

"J6164-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 14:55""BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30""07/13/2018 14:27""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-356" "05/07/2018 15:40""DW"  
"J6166-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 15:04""BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30""07/13/2018 14:27""  
"112G08005-WE04" "WE04 NAS Willow Grove" "NAWC-050718-RW-289" "05/07/2018 16:10""DW"  
"J6168-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 15:13""BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30""07/13/2018 14:27""  
"112G08005-WE04" "WE04 NAS Willow Grove" "WGNA-050718-DUP-35" "05/07/2018 07:00""DW"  
"J6170-FS" "NM" "SHP-180508-02" "1.200000" "SOP 5-369" "Gen Prep" "Initial" "05/10/2018 10:53"  
"05/14/2018 15:22""BNO" "COA" "NA" "T" "1.000" "NA" "NA" "" "100.000000" "18-0313"  
"18-0313" "DP-18-0112" "DP-18-0112" "18-0313" "05/08/2018 10:30""07/13/2018 14:27""



**Minor**

The surrogate spike recoveries for 13C2-PFHxA were greater than the 130% quality control limit for samples NAWC-050718-RW-275, NAWC-050718-RW-357, NAWC-050718-RW-162, NAWC-050718-RW-356 and WGNA-050718-DUP-35. The detected results reported in the affected samples were qualified as estimated (J).

Detected results reported below the limit of quantitation (LOQ) but above the detection limit (DL) were qualified as estimated (J).

**Notes**

The laboratory uses a primary transition for the quantitation of each analyte and a secondary transition for confirmation.

Samples with detections and their associated FRBs are summarized below. No detected results were present in the FRBs.

<b><u>Sample</u></b>	<b><u>Associated FRB</u></b>
NAWC-050718-RW-145	NAWC-050718-FRB-145
NAWC-050718-RW-162	NAWC-050718-FRB-162
NAWC-050718-RW-180	NAWC-050718-FRB-180
NAWC-050718-RW-275	NAWC-050718-FRB-275
NAWC-050718-RW-289	NAWC-050718-FRB-289
NAWC-050718-RW-316	NAWC-050718-FRB-316
NAWC-050718-RW-356	NAWC-050718-FRB-356
NAWC-050718-RW-357	NAWC-050718-FRB-357
WGNA-050718-DUP-35	NAWC-050718-FRB-275
WGNA-050718-RW-0335	WGNA-050718-FRB-0335
WGNA-050718-RW-0800	WGNA-050718-FRB-0800
WGNA-050718-RW-3556	WGNA-050718-FRB-3556

Non-detected results were reported to the Limit of Detection (LOD).

The buffering agent Trizma was added to all drinking water samples.

**Executive Summary**

**Laboratory Performance:** The surrogate recovery for 13C2-PFHxA was above the 130% quality control limit for several samples.

**Other Factors Affecting Data Quality:** Results below the RL were estimated.

TO: A. FREBOWITZ  
SDGS: 18-0313; 18-0316

PAGE 3

The data for these analyses were reviewed with reference to the Environmental Protection Agency document EPA/600/R-08/092, Method 537, "Determination of Selected Perfluorinated Alkyl Acids in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS)", (September 2009), US EPA National Functional Guidelines for Organic Data Review (January 2017), and the Department of Defense (DoD) document entitled "Quality Systems Manual (QSM) for Environmental Laboratories version 5.1" (2017) as applicable. The text of this report has been formulated to address only those areas affecting data quality.



---

Tetra Tech, Inc.  
Terri L. Solomon  
Chemist/Data Validator



---

Tetra Tech, Inc.  
Joseph A. Samchuck  
Data Validation Manager

Attachments:  
Appendix A – Qualified Analytical Results  
Appendix B – Results as Reported by the Laboratory  
Appendix C – Support Documentation



### Data Qualifier Definitions

The following definitions provide brief explanations of the validation qualifiers assigned to results in the data review process.

<b>U</b>	The analyte was analyzed for, but was not detected at a level greater than or equal to the level of the adjusted method detection limit for sample and method.
<b>J</b>	The analyte was positively identified and the associated numerical value is the approximate concentration of the analyte in the sample (due either to the quality of the data generated because certain quality control criteria were not met, or the concentration of the analyte was below the reporting limit).
<b>J+</b>	The result is an estimated quantity, but the result may be biased high.
<b>J-</b>	The result is an estimated quantity, but the result may be biased low.
<b>UJ</b>	The analyte was analyzed for, but was not detected. The reported detection limit is approximate and may be inaccurate or imprecise.
<b>R</b>	The sample result (detected) is unusable due to the quality of the data generated because certain criteria were not met. The analyte may or may not be present in the sample.
<b>UR</b>	The sample result (nondetected) is unusable due to the quality of the data generated because certain criteria were not met. The analyte may or may not be present in the sample.

**Appendix A**

Qualified Analytical Results

**Qualifier Codes:**

- A = Lab Blank Contamination
- B = Field Blank Contamination
- C = Calibration Noncompliance (i.e., % RSDs, %Ds, ICVs, CCVs, RRFs, etc.)
- C01 = GC/MS Tuning Noncompliance
- D = MS/MSD Recovery Noncompliance
- E = LCS/LCSD Recovery Noncompliance
- F = Lab Duplicate Imprecision
- G = Field Duplicate Imprecision
- H = Holding Time Exceedance
- I = ICP Serial Dilution Noncompliance
- J = ICP PDS Recovery Noncompliance; MSA's  $r < 0.995$
- K = ICP Interference - includes ICS % R Noncompliance
- L = Instrument Calibration Range Exceedance
- M = Sample Preservation Noncompliance
- N = Internal Standard Noncompliance
- N01 = Internal Standard Recovery Noncompliance Dioxins
- N02 = Recovery Standard Noncompliance Dioxins
- N03 = Clean-up Standard Noncompliance Dioxins
- O = Poor Instrument Performance (i.e., base-time drifting)
- P = Uncertainty near detection limit ( $< 2 \times$  IDL for inorganics and  $<$ CRQL for organics)
- Q = Other problems (can encompass a number of issues; i.e.chromatography,interferences, etc.)
- R = Surrogates Recovery Noncompliance
- S = Pesticide/PCB Resolution
- T = % Breakdown Noncompliance for DDT and Endrin
- U = RPD between columns/detectors  $>40\%$  for positive results determined via GC/HPLC
- V = Non-linear calibrations; correlation coefficient  $r < 0.995$
- W = EMPC result
- X = Signal to noise response drop
- Y = Percent solids  $<30\%$
- Z = Uncertainty at 2 standard deviations is greater than sample activity
- Z1 = Tentatively Identified Compound considered presumptively present
- Z2 = Tentatively Identified Compound column bleed
- Z3 = Tentatively Identified Compound aldol condensate
- Z4 = Sample activity is less than the at uncertainty at 3 standard deviations and greater than the MDC
- Z5 = Sample activity is less than the at uncertainty at 3 standard deviations and less than the MDC

<b>PROJ_NO: 08005-WE04</b> <b>SDG: 18-0313</b> <b>FRACTION: PFAS</b> <b>MEDIA: WATER</b>	NSAMPLE	NAWC-050718-RW-145			NAWC-050718-RW-162			NAWC-050718-RW-180			NAWC-050718-RW-275		
	LAB_ID	J6154-FS			J6158-FS			J6150-FS			J6152-FS		
	SAMP_DATE	5/7/2018			5/7/2018			5/7/2018			5/7/2018		
	QC_TYPE	NM			NM			NM			NM		
	UNITS	NG/L			NG/L			NG/L			NG/L		
	PCT_SOLIDS	0.0			0.0			0.0			0.0		
	DUP_OF												
PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	
N-ETHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		1	U		
N-METHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		1	U		
PENTADEC AFLUOROOCCTANOIC ACID (PFOA)	18.64			12.87	J	R	19.77			15.09	J	R	
PERFLUOROBUTANESULFONIC ACID (PFBS)	10.27			9.1	J	R	14.92			12.85	J	R	
PERFLUORODECANOIC ACID (PFDA)	0.57	J	P	1	U		1	U		1	J	RP	
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		1	U		
PERFLUOROHEPTANOIC ACID (PFHPA)	6.54			3.94	J	R	8.88			5.44	J	R	
PERFLUOROHEXANESULFONIC ACID (PFHXS)	8.56			13.69	J	R	12.68			7.69	J	R	
PERFLUOROHEXANOIC ACID (PFHXA)	9.19			6.78	J	R	14.93			10.14	J	R	
PERFLUORONONANOIC ACID (PFNA)	2.88			1.83	J	PR	3.33			2.85	J	R	
PERFLUOROOCCTANESULFONIC ACID (PFOS)	17.04			30.68	J	R	28.97			19.02	J	R	
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		1.5	U		
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		1	U		
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		1	U		

PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD			
PROJ_NO: 08005-WE04	NSAMPLE			NAWC-050718-RW-289			NAWC-050718-RW-316			NAWC-050718-RW-356			NAWC-050718-RW-357		
SDG: 18-0313	LAB_ID			J6168-FS			J6148-FS			J6166-FS			J6156-FS		
FRACTION: PFAS	SAMP_DATE			5/7/2018			5/7/2018			5/7/2018			5/7/2018		
MEDIA: WATER	QC_TYPE			NM			NM			NM			NM		
	UNITS			NG/L			NG/L			NG/L			NG/L		
	PCT_SOLIDS			0.0			0.0			0.0			0.0		
	DUP_OF														
N-ETHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		1	U				
N-METHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		1	U				
PENTADEC AFLUOROOCCTANOIC ACID (PFOA)	5.47			16.15			9.45	J	R	11.61	J	R			
PERFLUOROBUTANESULFONIC ACID (PFBS)	4.15			6.83			3.18	J	R	7.5	J	R			
PERFLUORODECANOIC ACID (PFDA)	1	U		1	U		1	U		0.42	J	PR			
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		1	U	R			
PERFLUOROHEPTANOIC ACID (PFHPA)	1.5	J	P	7.04			2.74	J	R	4.28	J	R			
PERFLUOROHEXANESULFONIC ACID (PFHXS)	3.75			25.12			3.92	J	R	12.06	J	R			
PERFLUOROHEXANOIC ACID (PFHXA)	2.11	J	P	14.7			4.26	J	R	10.06	J	R			
PERFLUORONONANOIC ACID (PFNA)	0.94	J	P	2.26	J	P	1.06	J	PR	1.31	J	PR			
PERFLUOROOCCTANESULFONIC ACID (PFOS)	10.99			32.92			7.94	J	R	9.78	J	R			
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		1.5	U				
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		1	U				
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		1	U				

<b>PROJ_NO: 08005-WE04</b> <b>SDG: 18-0313</b> <b>FRACTION: PFAS</b> <b>MEDIA: WATER</b>	NSAMPLE	WGNA-050718-DUP-35			WGNA-050718-RW-0335			WGNA-050718-RW-0800			WGNA-050718-RW-3556		
	LAB_ID	J6170-FS			J6162-FS			J6160-FS			J6164-FS		
	SAMP_DATE	5/7/2018			5/7/2018			5/7/2018			5/7/2018		
	QC_TYPE	NM			NM			NM			NM		
	UNITS	NG/L			NG/L			NG/L			NG/L		
	PCT_SOLIDS	0.0			0.0			0.0			0.0		
	DUP_OF	NAWC-050718-RW-275											
PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	
N-ETHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		1	U		
N-METHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		1	U		
PENTADEC AFLUOROOCCTANOIC ACID (PFOA)	13.21	J	R	22.49			14.73			7.77			
PERFLUOROBUTANESULFONIC ACID (PFBS)	12.58	J	R	11.74			23.13			8.69			
PERFLUORODECANOIC ACID (PFDA)	0.91	J	P	1	U		0.45	J	P	1	U		
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		1	U		
PERFLUOROHEPTANOIC ACID (PFHPA)	5.07	J	R	7.28			5.38			3.27			
PERFLUOROHEXANESULFONIC ACID (PFHXS)	7.85	J	R	7.86			9.29			2.38	J	P	
PERFLUOROHEXANOIC ACID (PFHXA)	9.11	J	R	12.83			8.53			5.69			
PERFLUORONONANOIC ACID (PFNA)	2.59	J	R	2.8			1.89	J	P	1.26	J	P	
PERFLUOROOCCTANESULFONIC ACID (PFOS)	19.67	J	R	19.22			19.76			5.94			
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		1.5	U		
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		1	U		
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		1	U		



<b>PROJ_NO: 08005-WE04</b> <b>SDG: 18-0316</b> <b>FRACTION: PFAS</b> <b>MEDIA: WATER</b>	NSAMPLE	NAWC-050718-FRB-145			NAWC-050718-FRB-162			NAWC-050718-FRB-180			NAWC-050718-FRB-275		
	LAB_ID	J6155-FS			J6159-FS			J6151-FS			J6153-FS		
	SAMP_DATE	5/7/2018			5/7/2018			5/7/2018			5/7/2018		
	QC_TYPE	NM			NM			NM			NM		
	UNITS	NG/L			NG/L			NG/L			NG/L		
	PCT_SOLIDS	0.0			0.0			0.0			0.0		
	DUP_OF												
PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	
N-ETHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		1	U		
N-METHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		1	U		
PENTADEC AFLUOROOCCTANOIC ACID (PFOA)	1	U		1	U		1	U		1	U		
PERFLUOROBUTANESULFONIC ACID (PFBS)	0.5	U		0.5	U		0.5	U		0.5	U		
PERFLUORODECANOIC ACID (PFDA)	1	U		1	U		1	U		1	U		
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		1	U		
PERFLUOROHEPTANOIC ACID (PFHPA)	1	U		1	U		1	U		1	U		
PERFLUOROHEXANESULFONIC ACID (PFHXS)	1	U		1	U		1	U		1	U		
PERFLUOROHEXANOIC ACID (PFHXA)	0.5	U		0.5	U		0.5	U		0.5	U		
PERFLUORONONANOIC ACID (PFNA)	1	U		1	U		1	U		1	U		
PERFLUOROOCCTANESULFONIC ACID (PFOS)	1	U		1	U		1	U		1	U		
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		1.5	U		
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		1	U		
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		1	U		

<b>PROJ_NO: 08005-WE04</b> <b>SDG: 18-0316</b> <b>FRACTION: PFAS</b> <b>MEDIA: WATER</b>	NSAMPLE	NAWC-050718-FRB-289			NAWC-050718-FRB-316			NAWC-050718-FRB-356			NAWC-050718-FRB-357		
	LAB_ID	J6169-FS			J6149-FS			J6167-FS			J6157-FS		
	SAMP_DATE	5/7/2018			5/7/2018			5/7/2018			5/7/2018		
	QC_TYPE	NM			NM			NM			NM		
	UNITS	NG/L			NG/L			NG/L			NG/L		
	PCT_SOLIDS	0.0			0.0			0.0			0.0		
	DUP_OF												
PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	
N-ETHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		1	U		
N-METHYLPERFLUOROOCCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		1	U		
PENTADEC AFLUOROOCCTANOIC ACID (PFOA)	1	U		1	U		1	U		1	U		
PERFLUOROBUTANESULFONIC ACID (PFBS)	0.5	U		0.5	U		0.5	U		0.5	U		
PERFLUORODECANOIC ACID (PFDA)	1	U		1	U		1	U		1	U		
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		1	U		
PERFLUOROHEPTANOIC ACID (PFHPA)	1	U		1	U		1	U		1	U		
PERFLUOROHEXANESULFONIC ACID (PFHXS)	1	U		1	U		1	U		1	U		
PERFLUOROHEXANOIC ACID (PFHXA)	0.5	U		0.5	U		0.5	U		0.5	U		
PERFLUORONONANOIC ACID (PFNA)	1	U		1	U		1	U		1	U		
PERFLUOROOCCTANESULFONIC ACID (PFOS)	1	U		1	U		1	U		1	U		
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		1.5	U		
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		1	U		
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		1	U		

<b>PROJ_NO: 08005-WE04</b> <b>SDG: 18-0316</b> <b>FRACTION: PFAS</b> <b>MEDIA: WATER</b>	NSAMPLE	WGNA-050718-FRB-0335			WGNA-050718-FRB-0800			WGNA-050718-FRB-3556		
	LAB_ID	J6163-FS			J6161-FS			J6165-FS		
	SAMP_DATE	5/7/2018			5/7/2018			5/7/2018		
	QC_TYPE	NM			NM			NM		
	UNITS	NG/L			NG/L			NG/L		
	PCT_SOLIDS	0.0			0.0			0.0		
	DUP_OF									
PARAMETER	RESULT	VQL	QLCD	RESULT	VQL	QLCD	RESULT	VQL	QLCD	
N-ETHYLPERFLUOROOCTANE SULFONAMIDOACETATE(NEFOSA)	1	U		1	U		1	U		
N-METHYLPERFLUOROOCTANE SULFONAMIDOACETATE(NMFOSA)	1	U		1	U		1	U		
PENTADECAFLUOROOCTANOIC ACID (PFOA)	1	U		1	U		1	U		
PERFLUOROBUTANESULFONIC ACID (PFBS)	0.5	U		0.5	U		0.5	U		
PERFLUORODECANOIC ACID (PFDA)	1	U		1	U		1	U		
PERFLUORODODECANOIC ACID (PFDOA)	1	U		1	U		1	U		
PERFLUOROHEPTANOIC ACID (PFHPA)	1	U		1	U		1	U		
PERFLUOROHEXANESULFONIC ACID (PFHXS)	1	U		1	U		1	U		
PERFLUOROHEXANOIC ACID (PFHXA)	0.5	U		0.5	U		0.5	U		
PERFLUORONONANOIC ACID (PFNA)	1	U		1	U		1	U		
PERFLUOROOCTANESULFONIC ACID (PFOS)	1	U		1	U		1	U		
PERFLUOROTETRADECANOIC ACID (PFTEA)	1.5	U		1.5	U		1.5	U		
PERFLUOROTRIDECANOIC ACID (PFTRIA)	1	U		1	U		1	U		
PERFLUOROUNDECANOIC ACID (PFUNA)	1	U		1	U		1	U		



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-316				
Battelle ID	J6148-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.70	0.22	0.50	2.50	
PFHpA	7.04	0.34	1.00	2.50	
PFOA	16.15	0.38	1.00	2.50	
PFNA	2.26 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	6.83	0.21	0.50	2.50	
PFHxS	25.12	0.34	1.00	2.50	
PFOS	32.92	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	104
13C2-PFPNA	87
d5-EtFOSAA	90

*Wesley L. Selman*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-180			
Battelle ID	J6150-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	14.93	0.22	0.50	2.50
PFHpA	8.88	0.34	1.00	2.50
PFOA	19.77	0.38	1.00	2.50
PFNA	3.33	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	14.92	0.21	0.50	2.50
PFHxS	12.68	0.34	1.00	2.50
PFOS	28.97	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	128
13C2-PFNA	101
d5-EtFOSAA	98

*Denise L. Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-275				
Battelle ID	J6152-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.14	J	0.22	0.50	2.50
PFHpA	5.44	J	0.34	1.00	2.50
PFOA	15.09	J	0.38	1.00	2.50
PFNA	2.85	J	0.37	1.00	2.50
PFDA	1.00	J	0.39	1.00	2.50
PFUnA	0.38	U	0.38	1.00	2.50
PFDoA	0.42	U	0.42	1.00	2.50
PFTTrDA	0.42	U	0.42	1.00	2.50
PFTeDA	0.73	U	0.73	1.50	2.50
NMeFOSAA	0.42	U	0.42	1.00	2.50
NEtFOSAA	0.44	U	0.44	1.00	2.50
PFBS	12.85	J	0.21	0.50	2.50
PFHxS	7.69	J	0.34	1.00	2.50
PFOS	19.02	J	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	138
13C2-PFNA	114
d5-EtFOSAA	113

*Marie L. Selmer*  
 06/08/2018





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-145				
Battelle ID	J6154-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	9.19	0.22	0.50	2.50	
PFHpA	6.54	0.34	1.00	2.50	
PFOA	18.64	0.38	1.00	2.50	
PFNA	2.88	0.37	1.00	2.50	
PFDA	0.57 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	10.27	0.21	0.50	2.50	
PFHxS	8.56	0.34	1.00	2.50	
PFOS	17.04	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	128
13C2-PFNA	96
d5-EtFOSAA	98

*Denise L. Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-357				
Battelle ID	J6156-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.06 J	0.22	0.50	2.50	
PFHpA	4.28 J	0.34	1.00	2.50	
PFOA	11.61 J	0.38	1.00	2.50	
PFNA	1.31 J	0.37	1.00	2.50	
PFDA	0.42 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	7.50 J	0.21	0.50	2.50	
PFHxS	12.06 J	0.34	1.00	2.50	
PFOS	9.78 J	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	137				
13C2-PFNA	108				
d5-EtFOSAA	80				

*Steve L. Salomon*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-162				
Battelle ID	J6158-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	6.78 J	0.22	0.50	2.50	
PFHpA	3.94 J	0.34	1.00	2.50	
PFOA	12.87 J	0.38	1.00	2.50	
PFNA	1.83 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	9.10 J	0.21	0.50	2.50	
PFHxS	13.69 J	0.34	1.00	2.50	
PFOS	30.68 J	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	132
13C2-PFNA	95
d5-EtFOSAA	90

*Denise Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0800				
Battelle ID	J6160-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	8.53	0.22	0.50	2.50	
PFHpA	5.38	0.34	1.00	2.50	
PFOA	14.73	0.38	1.00	2.50	
PFNA	1.89 J	0.37	1.00	2.50	
PFDA	0.45 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	23.13	0.21	0.50	2.50	
PFHxS	9.29	0.34	1.00	2.50	
PFOS	19.76	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	121
13C2-PFNA	107
d5-EtFOSAA	97

*Wesley L. Selman*  
 06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0335				
Battelle ID	J6162-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	12.83	0.22	0.50	2.50	
PFHpA	7.28	0.34	1.00	2.50	
PFOA	22.49	0.38	1.00	2.50	
PFNA	2.80	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	11.74	0.21	0.50	2.50	
PFHxS	7.86	0.34	1.00	2.50	
PFOS	19.22	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	126
13C2-PFNA	92
d5-EtFOSAA	89

*Denise Schumitz*

06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-3556				
Battelle ID	J6164-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.260				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	5.69 J	0.22	0.50	2.50	
PFHpA	3.27 J	0.34	1.00	2.50	
PFOA	7.77 J	0.38	1.00	2.50	
PFNA	1.26 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	8.69 J	0.21	0.50	2.50	
PFHxS	2.38 J	0.34	1.00	2.50	
PFOS	5.94 J	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	117
13C2-PFNA	83
d5-EtFOSAA	91

*Denise L. Schumitz*

06/08/2018





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-356				
Battelle ID	J6166-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.275				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	4.26	0.22	0.50	2.50	
PFHpA	2.74	0.34	1.00	2.50	
PFOA	9.45	0.38	1.00	2.50	
PFNA	1.06 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	3.18	0.21	0.50	2.50	
PFHxS	3.92	0.34	1.00	2.50	
PFOS	7.94	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	132
13C2-PFNA	109
d5-EtFOSAA	105

*Denise L. Schumitz*

06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-289				
Battelle ID	J6168-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	2.11 J	0.22	0.50	2.50	
PFHpA	1.50 J	0.34	1.00	2.50	
PFOA	5.47	0.38	1.00	2.50	
PFNA	0.94 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	4.15	0.21	0.50	2.50	
PFHxS	3.75	0.34	1.00	2.50	
PFOS	10.99	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	112
13C2-PFNA	94
d5-EtFOSAA	93

*Handwritten signature: Denise Schumitz*

06/08/2018



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-DUP-35				
Battelle ID	J6170-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	9.11 J	0.22	0.50	2.50	
PFHpA	5.07 J	0.34	1.00	2.50	
PFOA	13.21 J	0.38	1.00	2.50	
PFNA	2.59 J	0.37	1.00	2.50	
PFDA	0.91 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	12.58 J	0.21	0.50	2.50	
PFHxS	7.85 J	0.34	1.00	2.50	
PFOS	19.67 J	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	134
13C2-PFNA	105
d5-EtFOSAA	110

*Denise Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-316				
Battelle ID	J6149-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.255				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	110
13C2-PFDA	94
d5-EtFOSAA	103

*Denise L. Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-180				
Battelle ID	J6151-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	117
13C2-PFDA	103
d5-EtFOSAA	90

*Steve L. Selman*  
 06/08/2018




Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-275				
Battelle ID	J6153-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	113
13C2-PFDA	103
d5-EtFOSAA	102

  
 06/08/2018





Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-145				
Battelle ID	J6155-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	79
13C2-PFDA	90
d5-EtFOSAA	93

*Denise L. Schumitz*

06/08/2018




Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-357				
Battelle ID	J6157-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	108
13C2-PFDA	103
d5-EtFOSAA	101

  
 06/08/2018




Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-162				
Battelle ID	J6159-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	113
13C2-PFDA	98
d5-EtFOSAA	87

  
 06/08/2018



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	WGNA-050718-FRB-0800				
Battelle ID	J6161-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.255				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	109
13C2-PFDA	100
d5-EtFOSAA	103

*Denise L. Schumitz*  
 06/08/2018



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID WGNA-050718-FRB-0335

Battelle ID J6163-FS  
 Sample Type SA  
 Collection Date 05/07/2018  
 Extraction Date 05/15/2018  
 Analysis Date 05/17/2018  
 Analytical Instrument Sciex 5500 LC/MS/MS  
 % Moisture NA  
 Matrix DW  
 Sample Size 0.250  
 Size Unit-Basis L  
 Units ng/L MDL LOD LOQ

	ng/L	MDL	LOD	LOQ
PFHxA	0.22 U	0.22	0.50	2.50
PFHpA	0.34 U	0.34	1.00	2.50
PFOA	0.38 U	0.38	1.00	2.50
PFNA	0.37 U	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	0.21 U	0.21	0.50	2.50
PFHxS	0.34 U	0.34	1.00	2.50
PFOS	0.30 U	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	115
13C2-PFDA	107
d5-EtFOSAA	85

*Wesley L. Salomon*  
 06/08/2018



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID WGNA-050718-FRB-3556

Battelle ID J6165-FS  
 Sample Type SA  
 Collection Date 05/07/2018  
 Extraction Date 05/15/2018  
 Analysis Date 05/17/2018  
 Analytical Instrument Sciex 5500 LC/MS/MS  
 % Moisture NA  
 Matrix DW  
 Sample Size 0.250  
 Size Unit-Basis L  
 Units ng/L MDL LOD LOQ

	ng/L	MDL	LOD	LOQ
PFHxA	0.22 U	0.22	0.50	2.50
PFHpA	0.34 U	0.34	1.00	2.50
PFOA	0.38 U	0.38	1.00	2.50
PFNA	0.37 U	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDaA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	0.21 U	0.21	0.50	2.50
PFHxS	0.34 U	0.34	1.00	2.50
PFOS	0.30 U	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	90
13C2-PFDA	91
d5-EtFOSAA	100

06/08/2018






Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-356				
Battelle ID	J6167-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	116
13C2-PFDA	98
d5-EtFOSAA	85

  
 06/08/2018




Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-289				
Battelle ID	J6169-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	105
13C2-PFDA	94
d5-EtFOSAA	95

  
 06/08/2018

**Appendix B**

Results as Reported by the Laboratory



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-316				
Battelle ID	J6148-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	14.70	0.22	0.50	2.50	
PFHpA	7.04	0.34	1.00	2.50	
PFOA	16.15	0.38	1.00	2.50	
PFNA	2.26 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	6.83	0.21	0.50	2.50	
PFHxS	25.12	0.34	1.00	2.50	
PFOS	32.92	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	104				
13C2-PFPNA	87				
d5-EtFOSAA	90				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-180			
Battelle ID	J6150-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	14.93	0.22	0.50	2.50
PFHpA	8.88	0.34	1.00	2.50
PFOA	19.77	0.38	1.00	2.50
PFNA	3.33	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	14.92	0.21	0.50	2.50
PFHxS	12.68	0.34	1.00	2.50
PFOS	28.97	0.30	1.00	2.50
<b>Surrogate Recoveries (%)</b>				
13C2-PFHxA	128			
13C2-PFNA	101			
d5-EtFOSAA	98			



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-275				
Battelle ID	J6152-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.14	0.22	0.50	2.50	
PFHpA	5.44	0.34	1.00	2.50	
PFOA	15.09	0.38	1.00	2.50	
PFNA	2.85	0.37	1.00	2.50	
PFDA	1.00 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	12.85	0.21	0.50	2.50	
PFHxS	7.69	0.34	1.00	2.50	
PFOS	19.02	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	138				
13C2-PFNA	114				
d5-EtFOSAA	113				





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-145			
Battelle ID	J6154-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	9.19	0.22	0.50	2.50
PFHpA	6.54	0.34	1.00	2.50
PFOA	18.64	0.38	1.00	2.50
PFNA	2.88	0.37	1.00	2.50
PFDA	0.57 J	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	10.27	0.21	0.50	2.50
PFHxS	8.56	0.34	1.00	2.50
PFOS	17.04	0.30	1.00	2.50
<b>Surrogate Recoveries (%)</b>				
13C2-PFHxA	128			
13C2-PFNA	96			
d5-EtFOSAA	98			



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-357				
Battelle ID	J6156-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	10.06	0.22	0.50	2.50	
PFHpA	4.28	0.34	1.00	2.50	
PFOA	11.61	0.38	1.00	2.50	
PFNA	1.31 J	0.37	1.00	2.50	
PFDA	0.42 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	7.50	0.21	0.50	2.50	
PFHxS	12.06	0.34	1.00	2.50	
PFOS	9.78	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	137				
13C2-PFNA	108				
d5-EtFOSAA	80				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-162			
Battelle ID	J6158-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.280			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	6.78	0.22	0.50	2.50
PFHpA	3.94	0.34	1.00	2.50
PFOA	12.87	0.38	1.00	2.50
PFNA	1.83 J	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	9.10	0.21	0.50	2.50
PFHxS	13.69	0.34	1.00	2.50
PFOS	30.68	0.30	1.00	2.50
<b>Surrogate Recoveries (%)</b>				
13C2-PFHxA	132			
13C2-PFNA	95			
d5-EtFOSAA	90			



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0800				
Battelle ID	J6160-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.285				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	8.53	0.22	0.50	2.50	
PFHpA	5.38	0.34	1.00	2.50	
PFOA	14.73	0.38	1.00	2.50	
PFNA	1.89 J	0.37	1.00	2.50	
PFDA	0.45 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	23.13	0.21	0.50	2.50	
PFHxS	9.29	0.34	1.00	2.50	
PFOS	19.76	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	121				
13C2-PFNA	107				
d5-EtFOSAA	97				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-0335				
Battelle ID	J6162-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.270				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	12.83	0.22	0.50	2.50	
PFHpA	7.28	0.34	1.00	2.50	
PFOA	22.49	0.38	1.00	2.50	
PFNA	2.80	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	11.74	0.21	0.50	2.50	
PFHxS	7.86	0.34	1.00	2.50	
PFOS	19.22	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	126				
13C2-PFNA	92				
d5-EtFOSAA	89				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-RW-3556			
Battelle ID	J6164-FS			
Sample Type	SA			
Collection Date	05/07/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	DW			
Sample Size	0.260			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	5.69	0.22	0.50	2.50
PFHpA	3.27	0.34	1.00	2.50
PFOA	7.77	0.38	1.00	2.50
PFNA	1.26 J	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	8.69	0.21	0.50	2.50
PFHxS	2.38 J	0.34	1.00	2.50
PFOS	5.94	0.30	1.00	2.50
<b>Surrogate Recoveries (%)</b>				
13C2-PFHxA	117			
13C2-PFNA	83			
d5-EtFOSAA	91			





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-356				
Battelle ID	J6166-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.275				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	4.26	0.22	0.50	2.50	
PFHpA	2.74	0.34	1.00	2.50	
PFOA	9.45	0.38	1.00	2.50	
PFNA	1.06 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	3.18	0.21	0.50	2.50	
PFHxS	3.92	0.34	1.00	2.50	
PFOS	7.94	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	132				
13C2-PFNA	109				
d5-EtFOSAA	105				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	NAWC-050718-RW-289				
Battelle ID	J6168-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	2.11 J	0.22	0.50	2.50	
PFHpA	1.50 J	0.34	1.00	2.50	
PFOA	5.47	0.38	1.00	2.50	
PFNA	0.94 J	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	4.15	0.21	0.50	2.50	
PFHxS	3.75	0.34	1.00	2.50	
PFOS	10.99	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	112				
13C2-PFNA	94				
d5-EtFOSAA	93				



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	WGNA-050718-DUP-35				
Battelle ID	J6170-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/10/2018				
Analysis Date	05/14/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.280				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	9.11	0.22	0.50	2.50	
PFHpA	5.07	0.34	1.00	2.50	
PFOA	13.21	0.38	1.00	2.50	
PFNA	2.59	0.37	1.00	2.50	
PFDA	0.91 J	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDoA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	12.58	0.21	0.50	2.50	
PFHxS	7.85	0.34	1.00	2.50	
PFOS	19.67	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	134				
13C2-PFNA	105				
d5-EtFOSAA	110				



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-316				
Battelle ID	J6149-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.255				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	110				
13C2-PFDA	94				
d5-EtFOSAA	103				



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-180				
Battelle ID	J6151-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	117
13C2-PFDA	103
d5-EtFOSAA	90



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-275				
Battelle ID	J6153-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	113				
13C2-PFDA	103				
d5-EtFOSAA	102				





Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-145				
Battelle ID	J6155-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	79
13C2-PFDA	90
d5-EtFOSAA	93



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-357				
Battelle ID	J6157-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	108
13C2-PFDA	103
d5-EtFOSAA	101



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-162				
Battelle ID	J6159-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	113
13C2-PFDA	98
d5-EtFOSAA	87



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	WGNA-050718-FRB-0800				
Battelle ID	J6161-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.255				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	
<b>Surrogate Recoveries (%)</b>					
13C2-PFHxA	109				
13C2-PFDA	100				
d5-EtFOSAA	103				



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	WGNA-050718-FRB-0335				
Battelle ID	J6163-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	115
13C2-PFDA	107
d5-EtFOSAA	85



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	WGNA-050718-FRB-3556				
Battelle ID	J6165-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	90
13C2-PFDA	91
d5-EtFOSAA	100





Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-356				
Battelle ID	J6167-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFUnA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	116
13C2-PFDA	98
d5-EtFOSAA	85



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04

Client ID	NAWC-050718-FRB-289				
Battelle ID	J6169-FS				
Sample Type	SA				
Collection Date	05/07/2018				
Extraction Date	05/15/2018				
Analysis Date	05/17/2018				
Analytical Instrument	Sciex 5500 LC/MS/MS				
% Moisture	NA				
Matrix	DW				
Sample Size	0.250				
Size Unit-Basis	L				
Units	ng/L	MDL	LOD	LOQ	
PFHxA	0.22 U	0.22	0.50	2.50	
PFHpA	0.34 U	0.34	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFNA	0.37 U	0.37	1.00	2.50	
PFDA	0.39 U	0.39	1.00	2.50	
PFOA	0.38 U	0.38	1.00	2.50	
PFDaA	0.42 U	0.42	1.00	2.50	
PFTTrDA	0.42 U	0.42	1.00	2.50	
PFTeDA	0.73 U	0.73	1.50	2.50	
NMeFOSAA	0.42 U	0.42	1.00	2.50	
NEtFOSAA	0.44 U	0.44	1.00	2.50	
PFBS	0.21 U	0.21	0.50	2.50	
PFHxS	0.34 U	0.34	1.00	2.50	
PFOS	0.30 U	0.30	1.00	2.50	

**Surrogate Recoveries (%)**

13C2-PFHxA	105
13C2-PFDA	94
d5-EtFOSAA	95

**Appendix C**

Support Documentation

ANALYTE	ORIGINAL 050718-	DUPLICATE 050718-	RL	RPD	RPD > 50%	ORIGINAL	DUPLICATE SAMPLE	DIFFERENCE >2XRL
	RW-275	DUP-35				SAMPLE CONC	CONC >2xRL	
PENTADEC AFLUOROOCTANOIC ACID (PFOA)	15.09	13.21	2.5	13.286	FALSE	TRUE	TRUE	FALSE
PERFLUOROBUTANESULFONIC ACID (PFBS)	12.85	12.58	2.5	2.123	FALSE	TRUE	TRUE	FALSE
PERFLUORODECANOIC ACID (PFDA)	1	0.91	2.5	9.424	FALSE	FALSE	FALSE	FALSE
PERFLUOROHEPTANOIC ACID (PFHPA)	5.44	5.07	2.5	7.041	FALSE	TRUE	TRUE	FALSE
PERFLUOROHEXANESULFONIC ACID (PFHXS)	7.69	7.85	2.5	2.059	FALSE	TRUE	TRUE	FALSE
PERFLUOROHEXANOIC ACID (PFHXA)	10.14	9.11	2.5	10.701	FALSE	TRUE	TRUE	FALSE
PERFLUORONONANOIC ACID (PFNA)	2.85	2.59	2.5	9.559	FALSE	FALSE	FALSE	FALSE
PERFLUOROOCTANESULFONIC ACID (PFOS)	19.02	19.67	2.5	3.360	FALSE	TRUE	TRUE	FALSE

# Battelle

The Business of Innovation

## Chain-of-Custody

<b>Client Contact Information</b> Andy Frebowitz 234 Mall Boulevard, Suite 260 King of Prussia, PA 19406 610-382-1170		<b>Project Manager: Jonathan Thorn</b> Sampler Information (print name): Mary Kay Bond Phone: 610-382-1169 Email: mary.bond@tetrattech.com Turnaround Time (TAT) Requested: 21 days		Sampling Site: WE04 Site Information: NAS JRB Willow Grove/NAWC Warminster		COC #	
Project Name: WE04 Project No.: 112G08005-WE04		Normal <input checked="" type="checkbox"/> Priority <input type="checkbox"/> RUSH <input type="checkbox"/> Time Zone: Eastern		Analysis PFAS EPA 537 14 analytes		Page# 1 of 1	
Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Total # of Cont.		
NAWC-050718-RW-316 J6148	5/7/2018	10:10	G	DW	2	X	
NAWC-050718-FRB-316 J6149	5/7/2018	10:05	G	DW	2	X	
NAWC-050718-RW-180 J6150	5/7/2018	10:40	G	DW	2	X	Field Reagent Blank
NAWC-050718-FRB-180 J6151	5/7/2018	10:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-275 J6152	5/7/2018	11:10	G	DW	2	X	
NAWC-050718-FRB-275 J6153	5/7/2018	11:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-145 J6154	5/7/2018	12:40	G	DW	2	X	
NAWC-050718-FRB-145 J6155	5/7/2018	12:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-357 J6156	5/7/2018	13:10	G	DW	2	X	
NAWC-050718-FRB-357 J6157	5/7/2018	13:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-162 J6158	5/7/2018	13:40	G	DW	2	X	
NAWC-050718-FRB-162 J6159	5/7/2018	13:35	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-0800 J6160	5/7/2018	14:10	G	DW	2	X	
WGNA-050718-FRB-0800 J6161	5/7/2018	14:05	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-0335 J6162	5/7/2018	14:40	G	DW	2	X	
WGNA-050718-FRB-0335 J6163	5/7/2018	14:35	G	DW	2	X	Field Reagent Blank
WGNA-050718-RW-3556 J6164	5/7/2018	15:10	G	DW	2	X	
WGNA-050718-FRB-3556 J6165	5/7/2018	15:05	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-356 J6166	5/7/2018	15:40	G	DW	2	X	
NAWC-050718-FRB-356 J6167	5/7/2018	15:35	G	DW	2	X	Field Reagent Blank
NAWC-050718-RW-289 J6168	5/7/2018	16:10	G	DW	2	X	
NAWC-050718-FRB-289 J6169	5/7/2018	16:05	G	DW	2	X	Field Reagent Blank
WGNA-050718-DUP-35 J6170	5/7/2018	7:00	G	DW	2	X	Duplicate
Receipt Temperature: (°C) 0.9, 1.2		Samples Intact: Yes No		Samples on Ice: Yes No		Receipt Comments:	
Relinquished by (Print/Sign): Mary Kay Bond		Company: Tetra Tech Date/Time: 05/07/2018 16:00		Received by (Print/Sign): Matt Schwandt		Company: Battelle Date/Time: 5/8/18 1030	
Relinquished by (Print/Sign):		Company:		Received by (Print/Sign):		Company:	
Relinquished by (Print/Sign):		Company:		Received by (Print/Sign):		Company:	
Comments: FedEx Tracking #7721 6462 7152							

## QA/QC Summary Batch 18-0313

Project:	CTO-WE04 Naval Air Station Joint Reserve Base Willow Grove
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	DW
Data Set:	DP-18-0112
Analytical SOP:	5-371
Method Reference:	USEPA 537 rev. 1.1, QSM 5.1

### Sample Custody

Collection Date	Receipt Date	Temp (°C)
5/7/2018	5/8/2018	0.9, 1.2

Corrective Actions	None
Sample Storage	The water samples were stored refrigerated until extraction.
Related samples	FRB samples associated with these samples are extracted in SDG 18-0316

### METHOD SUMMARIES

Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with methanol. Extracts were split and concentrated to dryness under nitrogen with a water bath set between 60 °C and 65 °C, reconstituted with 96:4 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.
Prep comments	None.
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.
Analysis Comments	Samples analyzed on the Sciex 5500. The confirmation ion ratio was above 50% RPD for the following samples and analytes: NAWC-050718-RW-316 for PFOA and PFDA NAWC-050718-RW-357 for PFOA NAWC-050718-RW-162 for PFDA WGNA-050718-RW-0335 for PFOA WGNA-050718-RW-3556 for PFHpA and PFOA NAWC -050718-RW-356 for PFOA and PFDA NAWC -050718-RW-289 for PFOA

Holding Times	Extraction Date(s)	Analysis Date(s)
	5/10/2018	5/14/2018



**QA/QC Summary**  
**Batch 18-0313**

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
$\leq 1/3$ the MRL	No exceedances noted. No comments.
Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
70-130% of true value	No exceedances noted. No comments.
Matrix Spike (MS) / Duplicate (MSD)	A MS/MSD were prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy. The relative percent difference was calculated to measure precision.
70-130% of true value, RPD $\leq 30\%$	No exceedances noted. MS/MSD samples were not collected for this SDG.
Surrogates Standard Analytes	Labelled surrogate compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
70-130% of true value	No exceedances noted. No Comments.
Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
ICal high and low points RPD $\leq 20\%$ , 50-150% of average area of the ICAL and 70-140% of most recent CCV	No exceedances noted. No Comments.
Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadratic curve fitting.
$R^2 > 0.99$ Target and SIS compounds +/- 30% of true value, Low point 50-150% of true value	No exceedances noted. No comments.

**QA/QC Summary**  
**Batch 18-0313**

Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
Target and SIS compounds +/- 30% of true value	No exceedances noted.
	No comments.
Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.
Target and SIS compounds +/- 30% of true value Low point 50-150% of true value	No exceedances noted.
	No comments.



It can be done

BATTELLE - NORWELL OPERATIONS  
MISCELLANEOUS DOCUMENTATION FORM

<b>Project Title:</b>	Naval Air Station Joint Reserve Base Wi	<b>Data Set Number:</b>	DP-18-0112
<b>Project Number:</b>	100117920-WE04	<b>Prep Batch Number:</b>	18-0313
<b>Entered By:</b>	Denise Schumitz	<b>Entered On:</b>	05/17/2018
<b>Test Code (Matrix Type):</b>	Master_371(L)		

Samples that were manually integrated are noted on the quant reports with the comment (TRUE).  
DMS 6/21/2018

JV64 is not being used in the calibration curve for PFHxA, PFOA, PFDA, and PFTeDA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/21/2018

JV65 is not being used in the calibration curve for PFHxA and PFDA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/21/2018

JV64 has an ion ratio <50% for NMeFOSAA and PFUnA.  
DMS 6/21/2018

Sample J6164 has an ion ratio <50% for PFHpA.  
DMS 6/21/2018

Samples J6148, J6156, J6162, J6164, J6166 and J6168 have an ion ratio <50% for PFOA.  
DMS 6/21/2018

Sample J6148, J6158 and J6166 have an ion ratio <50% for PFDA.  
DMS 6/21/2018

**Task Leader Approval:**

**Supervisor Approval:**

Digitally signed by Jonathan  
Thorn  
Date: 2018.05.27 12:05:59 -04'00'

**PM Approval:**



**It can be done**

Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Number: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0112  
 Test Code: Master\_371

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	0	None
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	NA	None
Matrix Spike / Matrix Spike Duplicate Precision	NA	None
Extracted Internal Standard Analytes (Surrogates)	0	None
Instrument Calibration	0	None
Instrument Blank	NA	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None



It can be done

**BATTELLE - NORWELL OPERATIONS  
SAMPLE IDENTIFICATION PAGE**

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**

100117920-  
WE04

**18-0313**

**WE04 PFAS Analysis**

**DW**

<b>Sample ID</b>	<b>Description</b>
CQ755PB-FS	Procedural Blank
CQ756LCS-FS	Laboratory Control Sample
J6148-FS	NAWC-050718-RW-316
J6150-FS	NAWC-050718-RW-180
J6152-FS	NAWC-050718-RW-275
J6154-FS	NAWC-050718-RW-145
J6156-FS	NAWC-050718-RW-357
J6158-FS	NAWC-050718-RW-162
J6160-FS	WGNA-050718-RW-0800
J6162-FS	WGNA-050718-RW-0335
J6164-FS	WGNA-050718-RW-3556
J6166-FS	NAWC-050718-RW-356
J6168-FS	NAWC-050718-RW-289
J6170-FS	WGNA-050718-DUP-35

Samples Assigned By:

Stephanie Schultz

Date :

May 9, 2018

Comments:



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0112

	CQ755PB-FS (Procedural Blank)	CQ756LCS-FS (Laboratory Control Sample)	J6148-FS (NAWC-050718-RW-316)	J6150-FS (NAWC-050718-RW-180)	J6152-FS (NAWC-050718-RW-275)	J6154-FS (NAWC-050718-RW-145)
PFHxA	-	L	L	L	L	L
PFHpA	-	L	L	L	L	L
PFOA	-	L	L	L	L	L
PFNA	-	L	L	L	L	L
PFDA	-	L	-	-	L	L
PFUnA	-	L	-	-	-	-
PFDoA	-	L	-	-	-	-
PFTTrDA	-	L	-	-	-	-
PFTeDA	-	L	-	-	-	-
NMeFOSAA	-	L	-	-	-	-
NEtFOSAA	-	L	-	-	-	-
PFBS	-	L	L	L	L	L
PFHxS	-	L	L	L	L	L
PFOS	-	L	L/Br	L/Br	L/Br	L/Br

"L": Linear  
 "Br": branched  
 "L/Br": Linear/Branched  
 "-": Not detected





Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04  
 Preparation Batch: 18-0313  
 Data Set: DP-18-0:

	J6156-FS (NAWC-050718-RW-357)	J6158-FS (NAWC-050718-RW-162)	J6160-FS (WGNNA-050718-RW-0800)	J6162-FS (WGNNA-050718-RW-0335)	J6164-FS (WGNNA-050718-RW-3556)	J6166-FS (NAWC-050718-RW-356)	J6168-FS (NAWC-050718-RW-289)	J6170-FS (WGNNA-050718-DUP-35)
PFHxA	L	L	L	L	L	L	L	L
PFHpA	L	L	L	L	L	L	L	L
PFOA	L	L	L	L	L	L	L	L
PFNA	L	L	L	L	L	L	L	L
PFDA	L	L	L	L	L	L	L	L
PFUnA	-	-	-	-	-	-	-	-
PFDoA	-	-	-	-	-	-	-	-
PFTTrDA	-	-	-	-	-	-	-	-
PFTeDA	-	-	-	-	-	-	-	-
NMeFOSAA	-	-	-	-	-	-	-	-
NEtFOSAA	-	-	-	-	-	-	-	-
PFBS	L	L	L	L	L	L	L	L
PFHxS	L	L	L	L	L	L	L	L
PFOS	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br	L/Br

"L": Linear  
 "Br": branched  
 "L/Br": Linear/Branched  
 "-": Not detected



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Client ID	Procedural Blank			
Battelle ID	CQ755PB-FS			
Sample Type	PB			
Collection Date	05/10/2018			
Extraction Date	05/10/2018			
Analysis Date	05/14/2018			
Analytical Instrument	Sciex 5500 LC/MS/MS			
% Moisture	NA			
Matrix	WATER			
Sample Size	0.250			
Size Unit-Basis	L			
Units	ng/L	MDL	LOD	LOQ
PFHxA	0.22 U	0.22	0.50	2.50
PFHpA	0.34 U	0.34	1.00	2.50
PFOA	0.38 U	0.38	1.00	2.50
PFNA	0.37 U	0.37	1.00	2.50
PFDA	0.39 U	0.39	1.00	2.50
PFUnA	0.38 U	0.38	1.00	2.50
PFDoA	0.42 U	0.42	1.00	2.50
PFTTrDA	0.42 U	0.42	1.00	2.50
PFTeDA	0.73 U	0.73	1.50	2.50
NMeFOSAA	0.42 U	0.42	1.00	2.50
NEtFOSAA	0.44 U	0.44	1.00	2.50
PFBS	0.21 U	0.21	0.50	2.50
PFHxS	0.34 U	0.34	1.00	2.50
PFOS	0.30 U	0.30	1.00	2.50

**Surrogate Recoveries (%)**

13C2-PFHxA	107
13C2-PFDA	108
d5-EtFOSAA	91



Project Client: Tetra Tech  
 Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Client: 100117920-WE04

Client ID	Laboratory Control Sample					
Battelle ID	CQ756LCS-FS					
Sample Type	LCS					
Collection Date	05/10/2018					
Extraction Date	05/10/2018					
Analysis Date	05/14/2018					
Analytical Instrument	Sciex 5500 LC/MS/MS					
% Moisture	NA					
Matrix	WATER					
Sample Size	0.250					
Size Unit-Basis	L					
Units	ng/L	Target	Recovery	Qual	Control Limits Lower	Upper
PFHxA	16.05	15.00	107		70	130
PFHpA	14.89	15.00	99		70	130
PFOA	15.17	15.00	101		70	130
PFNA	15.73	15.00	105		70	130
PFDA	15.51	15.00	103		70	130
PFUnA	14.75	15.00	98		70	130
PFDoA	14.90	15.00	99		70	130
PFTTrDA	14.59	15.00	97		70	130
PFTeDA	16.76	15.00	112		70	130
NMeFOSAA	17.70	15.00	118		70	130
NEtFOSAA	16.48	15.00	110		70	130
PFBS	14.28	13.28	108		70	130
PFHxS	15.14	14.18	107		70	130
PFOS	13.81	14.33	96		70	130
<b>Surrogate Recoveries (%)</b>						
13C2-PFHxA	105					
13C2-PFDA	104					
d5-EtFOSAA	95					



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	13C2-PFOA	34,698.30	-
JV65	L2	5/14/18 11:38	13C2-PFOA	33,573.03	-
JV66	L3	5/14/18 11:47	13C2-PFOA	29,649.37	-
JV67	L4	5/14/18 11:56	13C2-PFOA	41,395.56	-
JV68	L5	5/14/18 12:05	13C2-PFOA	30,280.34	-
JV69	L6	5/14/18 12:14	13C2-PFOA	37,534.68	-
JV70	L7	5/14/18 12:23	13C2-PFOA	29,423.02	-
JV71	L8	5/14/18 12:32	13C2-PFOA	39,891.91	-
JV72	L9	5/14/18 12:41	13C2-PFOA	40,992.20	16.6

PASS

Average 35,270.93 Lower 17,635.47 Upper 52,906.40

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	13C2-PFOA	34,698.30	17,635.47	52,906.40		21,196.24	42,392.48	
JV65	L2	5/14/18 11:38	13C2-PFOA	33,573.03	17,635.47	52,906.40		21,196.24	42,392.48	
JV66	L3	5/14/18 11:47	13C2-PFOA	29,649.37	17,635.47	52,906.40		21,196.24	42,392.48	
JV67	L4	5/14/18 11:56	13C2-PFOA	41,395.56	17,635.47	52,906.40		21,196.24	42,392.48	
JV68	L5	5/14/18 12:05	13C2-PFOA	30,280.34	17,635.47	52,906.40		21,196.24	42,392.48	
JV69	L6	5/14/18 12:14	13C2-PFOA	37,534.68	17,635.47	52,906.40		21,196.24	42,392.48	
JV70	L7	5/14/18 12:23	13C2-PFOA	29,423.02	17,635.47	52,906.40		21,196.24	42,392.48	
JV71	L8	5/14/18 12:32	13C2-PFOA	39,891.91	17,635.47	52,906.40		21,196.24	42,392.48	
JV72	L9	5/14/18 12:41	13C2-PFOA	40,992.20	17,635.47	52,906.40		21,196.24	42,392.48	
JV63 ICC	ICC	5/14/18 12:50	13C2-PFOA	37,388.34	17,635.47	52,906.40		21,196.24	42,392.48	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	13C2-PFOA	35,848.65	17,635.47	52,906.40		21,196.24	42,392.48	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	13C2-PFOA	29,877.62	17,635.47	52,906.40		21,196.24	42,392.48	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	13C2-PFOA	36,858.30	17,635.47	52,906.40		21,196.24	42,392.48	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	13C2-PFOA	34,064.10	17,635.47	52,906.40		21,196.24	42,392.48	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	13C2-PFOA	24,285.52	17,635.47	52,906.40		21,196.24	42,392.48	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	13C2-PFOA	33,518.97	17,635.47	52,906.40		21,196.24	42,392.48	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	13C2-PFOA	24,248.05	17,635.47	52,906.40		21,196.24	42,392.48	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	13C2-PFOA	35,845.05	17,635.47	52,906.40		21,196.24	42,392.48	
JV69 CCV	CCV	5/14/18 14:19	13C2-PFOA	37,224.64	17,635.47	52,906.40		21,196.24	42,392.48	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	13C2-PFOA	31,443.44	17,635.47	52,906.40		26,057.25	52,114.50	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	13C2-PFOA	37,992.76	17,635.47	52,906.40		26,057.25	52,114.50	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	13C2-PFOA	30,729.65	17,635.47	52,906.40		26,057.25	52,114.50	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	13C2-PFOA	27,240.70	17,635.47	52,906.40		26,057.25	52,114.50	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	13C2-PFOA	34,960.30	17,635.47	52,906.40		26,057.25	52,114.50	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	13C2-PFOA	31,796.45	17,635.47	52,906.40		26,057.25	52,114.50	
JV68 CCV	CCV	5/14/18 15:30	13C2-PFOA	37,508.63	17,635.47	52,906.40		26,057.25	52,114.50	



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	13C4-PFOS	104,334.94	-
JV65	L2	5/14/18 11:38	13C4-PFOS	98,871.27	-
JV66	L3	5/14/18 11:47	13C4-PFOS	90,792.83	-
JV67	L4	5/14/18 11:56	13C4-PFOS	116,199.13	-
JV68	L5	5/14/18 12:05	13C4-PFOS	92,291.39	-
JV69	L6	5/14/18 12:14	13C4-PFOS	111,784.92	-
JV70	L7	5/14/18 12:23	13C4-PFOS	88,194.77	-
JV71	L8	5/14/18 12:32	13C4-PFOS	110,166.97	-
JV72	L9	5/14/18 12:41	13C4-PFOS	116,123.32	10.7

PASS

Average 103,195.50 Lower 51,597.75 Upper 154,793.25

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	13C4-PFOS	104,334.94	51,597.75	154,793.25		64,603.97	129,207.95	
JV65	L2	5/14/18 11:38	13C4-PFOS	98,871.27	51,597.75	154,793.25		64,603.97	129,207.95	
JV66	L3	5/14/18 11:47	13C4-PFOS	90,792.83	51,597.75	154,793.25		64,603.97	129,207.95	
JV67	L4	5/14/18 11:56	13C4-PFOS	116,199.13	51,597.75	154,793.25		64,603.97	129,207.95	
JV68	L5	5/14/18 12:05	13C4-PFOS	92,291.39	51,597.75	154,793.25		64,603.97	129,207.95	
JV69	L6	5/14/18 12:14	13C4-PFOS	111,784.92	51,597.75	154,793.25		64,603.97	129,207.95	
JV70	L7	5/14/18 12:23	13C4-PFOS	88,194.77	51,597.75	154,793.25		64,603.97	129,207.95	
JV71	L8	5/14/18 12:32	13C4-PFOS	110,166.97	51,597.75	154,793.25		64,603.97	129,207.95	
JV72	L9	5/14/18 12:41	13C4-PFOS	116,123.32	51,597.75	154,793.25		64,603.97	129,207.95	
JV63 ICC	ICC	5/14/18 12:50	13C4-PFOS	112,208.04	51,597.75	154,793.25		64,603.97	129,207.95	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	13C4-PFOS	111,731.01	51,597.75	154,793.25		64,603.97	129,207.95	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	13C4-PFOS	80,170.42	51,597.75	154,793.25		64,603.97	129,207.95	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	13C4-PFOS	97,479.82	51,597.75	154,793.25		64,603.97	129,207.95	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	13C4-PFOS	97,536.34	51,597.75	154,793.25		64,603.97	129,207.95	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	13C4-PFOS	79,659.10	51,597.75	154,793.25		64,603.97	129,207.95	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	13C4-PFOS	93,168.05	51,597.75	154,793.25		64,603.97	129,207.95	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	13C4-PFOS	71,904.49	51,597.75	154,793.25		64,603.97	129,207.95	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	13C4-PFOS	100,729.16	51,597.75	154,793.25		64,603.97	129,207.95	
JV69 CCV	CCV	5/14/18 14:19	13C4-PFOS	113,429.56	51,597.75	154,793.25		64,603.97	129,207.95	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	13C4-PFOS	91,010.65	51,597.75	154,793.25		79,400.69	158,801.38	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	13C4-PFOS	106,801.49	51,597.75	154,793.25		79,400.69	158,801.38	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	13C4-PFOS	84,581.13	51,597.75	154,793.25		79,400.69	158,801.38	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	13C4-PFOS	86,631.83	51,597.75	154,793.25		79,400.69	158,801.38	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	13C4-PFOS	95,267.28	51,597.75	154,793.25		79,400.69	158,801.38	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	13C4-PFOS	92,164.56	51,597.75	154,793.25		79,400.69	158,801.38	
JV68 CCV	CCV	5/14/18 15:30	13C4-PFOS	110,468.54	51,597.75	154,793.25		79,400.69	158,801.38	



Project Client: Tetra Tech

Project Client: Naval Air Station Joint Reserve Base Willow Grove, PA

Project Client: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/14/18 11:29	d3-MeFOSAA	35,968.44	-
JV65	L2	5/14/18 11:38	d3-MeFOSAA	30,241.50	-
JV66	L3	5/14/18 11:47	d3-MeFOSAA	28,007.84	-
JV67	L4	5/14/18 11:56	d3-MeFOSAA	36,150.49	-
JV68	L5	5/14/18 12:05	d3-MeFOSAA	30,317.42	-
JV69	L6	5/14/18 12:14	d3-MeFOSAA	36,296.34	-
JV70	L7	5/14/18 12:23	d3-MeFOSAA	27,471.55	-
JV71	L8	5/14/18 12:32	d3-MeFOSAA	34,679.62	-
JV72	L9	5/14/18 12:41	d3-MeFOSAA	37,693.37	4.7

PASS

Average 32,980.73 Lower 16,490.37 Upper 49,471.10

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/14/18 11:29	d3-MeFOSAA	35,968.44	16,490.37	49,471.10		21,222.19	42,444.39	
JV65	L2	5/14/18 11:38	d3-MeFOSAA	30,241.50	16,490.37	49,471.10		21,222.19	42,444.39	
JV66	L3	5/14/18 11:47	d3-MeFOSAA	28,007.84	16,490.37	49,471.10		21,222.19	42,444.39	
JV67	L4	5/14/18 11:56	d3-MeFOSAA	36,150.49	16,490.37	49,471.10		21,222.19	42,444.39	
JV68	L5	5/14/18 12:05	d3-MeFOSAA	30,317.42	16,490.37	49,471.10		21,222.19	42,444.39	
JV69	L6	5/14/18 12:14	d3-MeFOSAA	36,296.34	16,490.37	49,471.10		21,222.19	42,444.39	
JV70	L7	5/14/18 12:23	d3-MeFOSAA	27,471.55	16,490.37	49,471.10		21,222.19	42,444.39	
JV71	L8	5/14/18 12:32	d3-MeFOSAA	34,679.62	16,490.37	49,471.10		21,222.19	42,444.39	
JV72	L9	5/14/18 12:41	d3-MeFOSAA	37,693.37	16,490.37	49,471.10		21,222.19	42,444.39	
JV63 ICC	ICC	5/14/18 12:50	d3-MeFOSAA	34,401.63	16,490.37	49,471.10		21,222.19	42,444.39	
CQ755PB-FS(0)	Procedural Blank	5/14/18 13:08	d3-MeFOSAA	35,000.18	16,490.37	49,471.10		21,222.19	42,444.39	
CQ756LCS-FS(0)	Laboratory Control Sample	5/14/18 13:17	d3-MeFOSAA	26,521.45	16,490.37	49,471.10		21,222.19	42,444.39	
J6148-FS(0)	NAWC-050718-RW-316	5/14/18 13:26	d3-MeFOSAA	28,397.36	16,490.37	49,471.10		21,222.19	42,444.39	
J6150-FS(0)	NAWC-050718-RW-180	5/14/18 13:35	d3-MeFOSAA	30,315.36	16,490.37	49,471.10		21,222.19	42,444.39	
J6152-FS(0)	NAWC-050718-RW-275	5/14/18 13:43	d3-MeFOSAA	25,531.81	16,490.37	49,471.10		21,222.19	42,444.39	
J6154-FS(0)	NAWC-050718-RW-145	5/14/18 13:52	d3-MeFOSAA	30,882.72	16,490.37	49,471.10		21,222.19	42,444.39	
J6156-FS(0)	NAWC-050718-RW-357	5/14/18 14:01	d3-MeFOSAA	25,387.66	16,490.37	49,471.10		21,222.19	42,444.39	
J6158-FS(0)	NAWC-050718-RW-162	5/14/18 14:10	d3-MeFOSAA	35,413.81	16,490.37	49,471.10		21,222.19	42,444.39	
JV69 CCV	CCV	5/14/18 14:19	d3-MeFOSAA	31,738.59	16,490.37	49,471.10		21,222.19	42,444.39	
J6160-FS(0)	WGNA-050718-RW-0800	5/14/18 14:37	d3-MeFOSAA	26,555.02	16,490.37	49,471.10		22,217.01	44,434.03	
J6162-FS(0)	WGNA-050718-RW-0335	5/14/18 14:46	d3-MeFOSAA	33,982.85	16,490.37	49,471.10		22,217.01	44,434.03	
J6164-FS(0)	WGNA-050718-RW-3556	5/14/18 14:55	d3-MeFOSAA	24,586.57	16,490.37	49,471.10		22,217.01	44,434.03	
J6166-FS(0)	NAWC-050718-RW-356	5/14/18 15:04	d3-MeFOSAA	24,181.64	16,490.37	49,471.10		22,217.01	44,434.03	
J6168-FS(0)	NAWC-050718-RW-289	5/14/18 15:13	d3-MeFOSAA	30,500.70	16,490.37	49,471.10		22,217.01	44,434.03	
J6170-FS(0)	WGNA-050718-DUP-35	5/14/18 15:22	d3-MeFOSAA	31,134.03	16,490.37	49,471.10		22,217.01	44,434.03	
JV68 CCV	CCV	5/14/18 15:30	d3-MeFOSAA	33,881.63	16,490.37	49,471.10		22,217.01	44,434.03	



## Summary Asymmetry Report

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/14/2018 12:23:30 PM	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Asymmetry Factor	Passing Range
PFBS_1	298.9 / 80.0	1.47	1.48	0.8 – 1.5
PFHxA_1	313.0 / 269.0	1.75	1.34	0.8 – 1.5

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/14/2018 12:23:30 PM	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
PFBS_1	298.9 / 80.0	1.47	36	>10
PFBS_2	298.9 / 99.0	1.47	35	>10
PFHxA_1	313.0 / 269.0	1.75	102	>10
PFHxA_2	313.0 / 119.0	1.75	41	>10
PFHpA_1	363.0 / 319.0	2.11	32	>10
PFHpA_2	363.0 / 169.0	2.11	32	>10
PFHxS_1	399.0 / 80.0	2.12	60	>10
PFHxS_2	399.0 / 99.0	2.12	56	>10
PFOA_1	413.0 / 369.0	2.49	61	>10
PFOA_2	413.0 / 169.0	2.49	33	>10
PFNA_1	463.0 / 419.0	2.86	56	>10
PFNA_2	463.0 / 219.0	2.86	55	>10
PFOS_1	499.0 / 80.0	2.86	65	>10
PFOS_2	499.0 / 99.0	2.86	50	>10
PFDA_1	513.0 / 469.0	3.22	55	>10
PFDA_2	513.0 / 219.0	3.21	38	>10
PFUnA_1	563.0 / 519.0	3.53	55	>10
PFUnA_2	563.0 / 269.0	3.53	36	>10
PFDaA_1	613.0 / 569.0	3.82	67	>10
PFDaA_2	613.0 / 319.0	3.82	48	>10
PFTrDA_1	663.0 / 619.0	4.08	76	>10
PFTrDA_2	663.0 / 169.0	4.08	33	>10
PFTeDA_1	713.0 / 669.0	4.30	61	>10
PFTeDA_2	713.0 / 169.0	4.30	51	>10
NMeFOSAA_1	570.0 / 419.0	3.36	57	>10
NMeFOSAA_2	570.0 / 512.0	3.36	58	>10
NEtFOSAA_1	584.0 / 419.0	3.53	54	>10
NEtFOSAA_2	584.0 / 483.0	3.52	35	>10

Analytical Transitions for PFAS in drinking water

SOP 5-371 (EPA 537 Version 1.1)

Analyte	CAS No.	Type	Primary Transition	Secondary Transition
<b>PFHxA</b>	307-24-4	Target	313.0 / 269.0	313.0 / 119.0
<b>PFHpA</b>	375-85-9	Target	363.0 / 319.0	363.0 / 169.0
<b>PFOA</b>	335-67-1	Target	413.0 / 369.0	413.0 / 169.0
<b>PFNA</b>	375-95-1	Target	463.0 / 419.0	463.0 / 219.0
<b>PFDA</b>	335-76-2	Target	513.0 / 469.0	513.0 / 219.0
<b>PFUnA</b>	2058-94-8	Target	563.0 / 519.0	563.0 / 269.0
<b>PFDoA</b>	307-55-1	Target	613.0 / 569.0	613.0 / 319.0
<b>PFTTrDA</b>	72629-94-8	Target	663.0 / 619.0	663.0 / 169.0
<b>PFTeDA</b>	376-06-7	Target	713.0 / 669.0	713.0 / 169.0
<b>NMeFOSAA</b>	2355-31-9	Target	570.0 / 419.0	570.0 / 512.0
<b>NEtFOSAA</b>	2991-50-6	Target	584.0 / 419.0	584.0 / 483.0
<b>PFBS</b>	375-73-5	Target	298.9.0 / 80.0	298.9.0 / 99.0
<b>PFHxS</b>	355-46-4	Target	399.0 / 80.0	399.0 / 99.0
<b>PFOS</b>	1763-23-1	Target	499.0 / 80.0	499.0 / 99.0
<b><sup>13</sup>C<sub>2</sub>-PFHxA</b>	NA	SIS	315.0 / 270.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFDA</b>	NA	SIS	515.0 / 470.0	NA
<b>d<sub>5</sub>-EtFOSAA</b>	NA	SIS	589.0 / 419.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFOA</b>	NA	IS	415.0 / 270.0	NA
<b><sup>13</sup>C<sub>4</sub>-PFOS</b>	NA	IS	503.0 / 80.0	NA
<b>d<sub>3</sub>-MeFOSAA</b>	NA	IS	573.0 / 419.0	NA

# BATTELLE DETECTION LIMITS FOR PFAS IN DRINKING WATER

Battelle SOP 5-371 (EPA Method 537 Version 1.1)

Analyte	CAS No.	MDL (ng/L)	LOD (ng/L)	LOQ (ng/L)	MRL (ng/L)
<b>PFHxA</b>	307-24-4	0.22	0.5	2.5	2.5
<b>PFHpA</b>	375-85-9	0.34	1.0	2.5	2.5
<b>PFOA</b>	335-67-1	0.38	1.0	2.5	2.5
<b>PFNA</b>	375-95-1	0.37	1.0	2.5	2.5
<b>PFDA</b>	335-76-2	0.39	1.0	2.5	2.5
<b>PFUnA</b>	2058-94-8	0.38	1.0	2.5	2.5
<b>PFDoA</b>	307-55-1	0.42	1.0	2.5	2.5
<b>PFTrDA</b>	72629-94-8	0.42	1.0	2.5	2.5
<b>PFTeDA</b>	376-06-7	0.73	1.5	2.5	2.5
<b>NMeFOSAA</b>	2355-31-9	0.42	1.0	2.5	2.5
<b>NEtFOSAA</b>	2991-50-6	0.44	1.0	2.5	2.5
<b>PFBS</b>	375-73-5	0.21	0.5	2.5	2.5
<b>PFHxS</b>	3871-99-6	0.34	1.0	2.5	2.5
<b>PFOS</b>	1763-23-1	0.30	1.0	2.5	2.5

*Analytes on NELAP and ELAP QSM 5.1 Scope of accreditation*



**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

# QTRAP 5500

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Mass Calibration and Tune Check

## QTRAP 5500 Preventive Maintenance Checklist

<b>Preventive Maintenance Date:</b>	22-Feb-2017
<b>Request ID:</b>	3683
<b>Company Name:</b>	Battelle Memorial Institute
<b>Instrument ID:</b>	X60666
<b>Instrument Model:</b>	QTRAP 5500
<b>Instrument Serial Number:</b>	AU23051004

**PASS**       **FAIL**

**Any failure will lead to an automatic Service Call being open to investigate fault.**

Preventive Maintenance is performed twice every year unless specified in the Service Contract. It is designed to help maintain optimum system performance and to help diagnose any system deficiencies.

Engineer is required the assigned Request ID for this PM otherwise making this job invalid.

**Comments:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Performed By:**           Kaustubh Dhayagude                **Date:**           22-Feb-2017          

**Approved By :** \_\_\_\_\_      **Date:** \_\_\_\_\_

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500**

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**PRE PM PPG PERFORMANCE EVALUATION:**

- Consult Customer concerning the unit overall performance.
- Check Logbook for Services recently performed.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.5	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	1.9	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.4	Read Only
<input checked="" type="checkbox"/> CAD High	3.4	Read Only
<input checked="" type="checkbox"/> CAD 12	3.4	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Check for Front end contamination symptoms. Run Q1 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop
- Check for Q3 contamination symptoms. Run Q3 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop

**Pre PM PPG Test:** Perform each of the following tests. Optimize ion source position only. The specifications listed for these Pre PM tests are guidelines only, not required to be met.

- Perform Q1 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 175.133	1.64 e6	Read Only	0.8095	Read Only
Q1 500.380	2.40 e7	Read Only	0.8592	Read Only
Q1 906.673	2.86 e7	Read Only	0.9633	Read Only

- Perform Q3 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 175.133	1.26 e6	Read Only	0.6252	Read Only
Q3 500.380	2.19 e7	Read Only	0.7275	Read Only
Q3 906.673	3.02 e7	Read Only	0.7662	Read Only



**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform MSMS POS in Product Ion scan with 609.3 parent and record daughter 195.1 using Reserpine 0.167 pmol/ul at the scan rate of 10 Da/s for 10 MCA. Calculate transmission efficiency comparing Q1POS 609 intensity. Transmission Efficiency: : 19.51% (Read Only)

Mass	MSMS Intensity		MSMS Width Value	Width Specs
	Value	Spec		
Q1 609.3	7.43 e7	Read Only	0.9981	Read Only
MS/MS 195.1	1.45 e7	Read Only	0.6582	Read Only

Perform Q1 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 933.636	1.43 e7	Read Only	0.7330	Read Only

Perform Q3 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 933.636	2.22 e7	Read Only	0.8138	Read Only

Perform Product Ion scan using NEG PPG 3e-5M. Record 10 mca.

Mass	Scan Rate	MCA	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.35 e6	Read Only	0.6495	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**PREVENTIVE MAINTENANCE CHECKLIST:**

- Check Cooling Fans for Turbo Pumps while MS is ON.
- Check QJet and QPS tuning voltage for reference.
- Record AC input Voltage while MS is OFF: \_\_\_\_\_(200-240VAC).  
If Out-of-Range, notify customer.
  
- Clean Interface
  - Curtain Plate
  - Orifice Plate
  - QJet
  - Q0 Rods.
  
- Replace Roughing Pump Oil.
- Inspect Oil Exhaust Filter, if Applicable.  N/A
- Clean and inspect built-in divert valve if used.  N/A
- Check Multiplier Voltage, optimize if necessary.
- Replace four Air Filters at the bottom of the mass spectrometer.
  
- Pump down overnight if possible.  N/A
  
- Perform Maintenance on Turbo V source.
  
- Replace Electrode, if necessary.  N/A
- Check Turbo heaters resistances.
- Check if Temperature is reached at 500C with TIS Probe installed.
- Check if Temperature is reached at 500C with APCI Probe installed.  N/A

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**POST PM PPG PERFORMANCE TESTS:**

- Set-up Sample for Infusion.
- Check spray and adjust sprayer's position of the TIS source.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.8	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	2.1	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.6	Read Only
<input checked="" type="checkbox"/> CAD High	3.7	Read Only
<input checked="" type="checkbox"/> CAD 12	3.7	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Perform Q1 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q1 175.133	5.94 e6	≥1.2 <sup>e6</sup>	0.6933	0.6 to 0.8
Q1 500.380	2.25 e7	≥9.0 <sup>e6</sup>	0.7444	0.6 to 0.8
Q1 906.673	2.74 e7	≥1.4 <sup>e7</sup>	0.7347	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q1 906.673	1.33 e8	≥6.8 <sup>e7</sup>	0.7656	0.6 to 0.8

- Perform Q3 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q3 175.133	4.54 e6	≥1.2 <sup>e6</sup>	0.6390	0.6 to 0.8
Q3 500.380	2.13 e7	≥9.0 <sup>e6</sup>	0.7008	0.6 to 0.8
Q3 906.673	3.04 e7	≥1.4 <sup>e7</sup>	0.7683	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q3 906.673	1.51 e8	≥6.8 <sup>e7</sup>	0.7118	0.6 to 0.8

- Perform "Product of 609.3" POS and record product ion 195.1 using Reserpine 0.167pmol/uL. Record 10 mca. Calculate Transmission efficiency comparing Q1POS 609 intensity.

Transmission Efficiency: 16.93% (≥ 10.0%)

Mass	MSMS Intensity		Width Value	Width Specs
	Value	Spec		
Q1 609.3	5.74 e7	N/A	0.7667	Read Only
MS/MS 195.1	9.72 e6	N/A	0.6751	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform Q1 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q1 Intensity		Q1 Width Value	Width Specs
			Value	Spec		
Q1 933.636	10	10	1.31 e7	$\geq 1.0^{e7}$	0.6895	0.6 to 0.8
Q1 933.636	1000	50	6.32 e7	$\geq 4.0^{e7}$	0.6740	0.6 to 0.8

Perform Q3 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q3 Intensity		Q3 Width Value	Width Specs
			Value	Spec		
Q3 933.636	10	10	1.70 e7	$\geq 8.0^{e6}$	0.7665	0.6 to 0.8
Q3 933.636	1000	50	7.41 e7	$\geq 4.0^{e7}$	0.7292	0.6 to 0.8

Perform Product Ion scan using NEG PPG 3e-5M.

Mass	Scan Rate	Mca	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.33 e6	Read Only	0.6387	Read Only

Perform ER POS 118.087 and 922.01 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 118.087	0.05	8.08 e6	$\geq 7.2^{e6}$	0.1302	<0.35
ER 922.010	0.05	3.89 e7	$\geq 2.8^{e6}$	0.2603	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 118.087	0.05	2.55 e7	$\geq 2.4^{e7}$	0.3740	<0.65
ER 922.010	0.05	2.37 e8	$\geq 6.8^{e7}$	0.5407	<0.65

Perform ER NEG 431.982 and 601.978 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 431.982	0.05	1.05 e8	$\geq 4.4^{e7}$	0.1840	<0.35
ER 601.978	0.05	7.74 e7	$\geq 5.6^{e7}$	0.1849	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 431.982	0.05	3.43 e8	$\geq 1.2^{e8}$	0.4382	<0.65
ER 601.978	0.05	2.55 e8	$\geq 1.6^{e8}$	0.6205	<0.65

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

- Perform EPI POS 397.2 using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Q0 Trapping OFF		Q0 Trapping ON	
		Intensity	Spec	Intensity	Spec
EPI 397.2	10000	> 3.5 e6	≥2.0 e6	> 4.0 e7	≥6.4 e6

- Perform MS3 POS full scan Fragmentation ON & OFF using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Fragamentation OFF		Fragmentation ON	
		Intensity	Spec	Intensity	Spec
MS3 397.2	1000	3.2 e7	Contains only 397.2	N/A	N/A
<input type="checkbox"/> 236 OR <input checked="" type="checkbox"/> 365	1000	1.19 e8	Fragment Intensity	> 4.4 e6	≥1.6x 10 <sup>e6</sup>

**REVIEW:**

- Attach all spectrums printouts to this procedure.
- If any parameter setting access modes were changed during the PM, ensure they are returned to their normal access mode and that their offsets are adjusted to match optimized values from the post-PM acquisition files.
- Empty tuning cache folder, if necessary.  N/A
- Update Service Work Order status
- Fill and replace PM Label.

**END OF PREVENTIVE MAINTENANCE CHECKLIST****Document history:**

06 OCT 2016: Appendix ZEFPM003-2L: Removed requirements to fit Manufacturer's testing criteria.



It can be done

**BATTELLE - NORWELL OPERATIONS  
SAMPLE PREPARATION RECORDS**

<b><u>Project Title(s)</u></b>	<b><u>Project No.(s)</u></b>
Naval Air Station Joint Reserve Base Willow Grove, PA	100117920-WE04
<b>18-0313</b>	
<b>WE04 PFAS Analysis</b>	
<b>DW</b>	
SOP Numbers (see workplan for modifications)	
VOASOP No.	5-371

<b>This Batch Contains The Following Samples:</b>		
CQ755PB-FS	J6156-FS	J6168-FS
CQ756LCS-FS	J6158-FS	J6170-FS
J6148-FS	J6160-FS	
J6150-FS	J6162-FS	
J6152-FS	J6164-FS	
J6154-FS	J6166-FS	

Laboratory Preparation Records  
COMPLETE AND VALIDATED

Prep Task Leader: Stephanie Schultz

Approved By:	Date	Initials
Denise Schumitz	05/16/2018	DMS





It can be done

BATTELLE - NORWELL OPERATIONS  
LIQUID SAMPLE ID FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**

100117920-WE04

18-0313

WE04 PFAS Analysis

DW

Sample ID	Description	Volume (mL)	Bottles	*	Date Initials
CQ755PB-FS	Procedural Blank	250.0	NA	--	05/10/18 SAS
CQ756LCS-FS	Laboratory Control Sample	250.0	NA	--	05/10/18 SAS
J6148-FS	NAWC-050718-RW-316	285.0	1	C	05/11/18 SAS
J6150-FS	NAWC-050718-RW-180	280.0	1	C	05/11/18 SAS
J6152-FS	NAWC-050718-RW-275	280.0	1	C	05/11/18 SAS
J6154-FS	NAWC-050718-RW-145	280.0	1	C	05/11/18 SAS
J6156-FS	NAWC-050718-RW-357	270.0	1	C	05/11/18 SAS
J6158-FS	NAWC-050718-RW-162	280.0	1	C	05/11/18 SAS
J6160-FS	WGNA-050718-RW-0800	285.0	1	C	05/11/18 SAS
J6162-FS	WGNA-050718-RW-0335	270.0	1	C	05/11/18 SAS
J6164-FS	WGNA-050718-RW-3556	260.0	1	C	05/11/18 SAS
J6166-FS	NAWC-050718-RW-356	275.0	1	C	05/11/18 SAS
J6168-FS	NAWC-050718-RW-289	280.0	1	C	05/11/18 SAS
J6170-FS	WGNA-050718-DUP-35	280.0	1	C	05/11/18 SAS

**Comments:**

Sample ID:	Comments:
CQ755PB-FS	1.27g Trizma(170526-01) weighed on BAL-009
CQ756LCS-FS	1.23g Trizma(170526-01) weighed on BAL-009

Samples Assigned By

Stephanie Schultz

Date :

May 9, 2018

\* - "C" = Sample is Consumed



It can be done

## BATTELLE - NORWELL OPERATIONS SURROGATE SPIKE FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	Standard ID	Type	Vial No.	Vol Added (uL)	Date Spiked/ Spiked By	Witn'd By	Comment
CQ755PB-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
CQ756LCS-FS	JV41	LCS/MS	1	75	05/10/18 SAS	JCT	NA
CQ756LCS-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6148-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6150-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6152-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6154-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6156-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6158-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6160-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6162-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6164-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6166-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6168-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA
J6170-FS	JV60	SIS	1	50	05/10/18 SAS	JCT	NA

## Syringes/Pipettes Used:

Std ID	Type	Syr/Pip
JV41	Pipette	I0793912B
JV60	Pipette	I0793912B



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE EXTRACTION FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW**

Sample ID	1st Extraction	2nd Extraction	3rd Extraction	Conc. ID	Turbo °C	Turbo PSI	KD °C	Comment
CQ755PB-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
CQ756LCS-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6148-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6150-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6152-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6154-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6156-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6158-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6160-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6162-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6164-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6166-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6168-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA
J6170-FS	05/10/18 SAS	NA	NA	NA	NA	NA	NA	NA

**Solvents/Reagent Preparations:**

Name	ID	Expires	Lot No	Procedure	Comments
Pre-packed SPE Column	RP-180510-1	05/10/18	S214-0071	Pre-packed SPE Column	

**Solvents/Reagents:**

Name	Lot No	Comments
Methanol (HPLC) (180227-02)	178212	



It can be done

## BATTELLE - NORWELL OPERATIONS INTERNAL STANDARD SPIKING FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0313****WE04 PFAS Analysis****DW****(N/A Fraction)**

Extract Id	Extr. Vol. (uL)	Added (uL)	Std. Id	Accm . (uL)	Vial No.	Pre Inj. Vol. (uL)^	Final Dilution*	Date Spiked/ Spiked By	Witn'd By
CQ755PB-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
CQ756LCS-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6148-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6150-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6152-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6154-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6156-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6158-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6160-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6162-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6164-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6166-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6168-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT
J6170-FS(0)	950	50	JV59	50	1	1000	1.000	05/14/18 SAS	JCT

\* - Final Dilution is any HPLC, dilutions, or other manipulation

^ - Pre Injection Volume (PIV) includes any RIS spikes.

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	MeOH		5/14/2018 11:20:57 AM	5-0371.dam	18-0313.wiff
2	JV64	L1	5/14/2018 11:29:54 AM	5-0371.dam	18-0313.wiff
3	JV65	L2	5/14/2018 11:38:49 AM	5-0371.dam	18-0313.wiff
4	JV66	L3	5/14/2018 11:47:45 AM	5-0371.dam	18-0313.wiff
5	JV67	L4	5/14/2018 11:56:42 AM	5-0371.dam	18-0313.wiff
6	JV68	L5	5/14/2018 12:05:38 PM	5-0371.dam	18-0313.wiff
7	JV69	L6	5/14/2018 12:14:33 PM	5-0371.dam	18-0313.wiff
8	JV70	L7	5/14/2018 12:23:30 PM	5-0371.dam	18-0313.wiff
9	JV71	L8	5/14/2018 12:32:26 PM	5-0371.dam	18-0313.wiff
10	JV72	L9	5/14/2018 12:41:22 PM	5-0371.dam	18-0313.wiff
11	JV63 ICC	ICC	5/14/2018 12:50:20 PM	5-0371.dam	18-0313.wiff
1	MeOH		5/14/2018 12:59:16 PM	5-0371.dam	18-0313.wiff
12	CQ755PB-FS(0)	Procedural Blank	5/14/2018 1:08:14 PM	5-0371.dam	18-0313.wiff
13	CQ756LCS-FS(0)	Laboratory Control Sample	5/14/2018 1:17:09 PM	5-0371.dam	18-0313.wiff
14	J6148-FS(0)	NAWC-050718-RW-316	5/14/2018 1:26:05 PM	5-0371.dam	18-0313.wiff
15	J6150-FS(0)	NAWC-050718-RW-180	5/14/2018 1:35:01 PM	5-0371.dam	18-0313.wiff
16	J6152-FS(0)	NAWC-050718-RW-275	5/14/2018 1:43:56 PM	5-0371.dam	18-0313.wiff
17	J6154-FS(0)	NAWC-050718-RW-145	5/14/2018 1:52:52 PM	5-0371.dam	18-0313.wiff
18	J6156-FS(0)	NAWC-050718-RW-357	5/14/2018 2:01:46 PM	5-0371.dam	18-0313.wiff
19	J6158-FS(0)	NAWC-050718-RW-162	5/14/2018 2:10:41 PM	5-0371.dam	18-0313.wiff
7	JV69 CCV	CCV	5/14/2018 2:19:36 PM	5-0371.dam	18-0313.wiff
1	MeOH		5/14/2018 2:28:30 PM	5-0371.dam	18-0313.wiff
20	J6160-FS(0)	WGNA-050718-RW-0800	5/14/2018 2:37:26 PM	5-0371.dam	18-0313.wiff
21	J6162-FS(0)	WGNA-050718-RW-0335	5/14/2018 2:46:22 PM	5-0371.dam	18-0313.wiff
22	J6164-FS(0)	WGNA-050718-RW-3556	5/14/2018 2:55:18 PM	5-0371.dam	18-0313.wiff
23	J6166-FS(0)	NAWC-050718-RW-356	5/14/2018 3:04:13 PM	5-0371.dam	18-0313.wiff
24	J6168-FS(0)	NAWC-050718-RW-	5/14/2018 3:13:09	5-0371.dam	18-0313.wiff

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
		289	PM		
25	J6170-FS(0)	WGNA-050718-DUP-35	5/14/2018 3:22:03 PM	5-0371.dam	18-0313.wiff
6	JV68 CCV	CCV	5/14/2018 3:30:58 PM	5-0371.dam	18-0313.wiff






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFBS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	298.9 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.74054 x + 0.01745$  ( $r = 0.99950$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.15	22.804779	103.0
3	JV65	L2	True	44.30	38.779769	87.5
4	JV66	L3	True	88.60	83.589043	94.3
5	JV67	L4	True	221.50	248.105496	112.0
6	JV68	L5	True	443.00	480.862156	108.6
7	JV69	L6	True	885.00	867.003654	98.0
8	JV70	L7	True	2212.50	2093.047764	94.6
9	JV71	L8	True	4425.00	4522.222161	102.2
10	JV72	L9	True	8850.00	8835.635177	99.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFBS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	298.9 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.22816x + 0.01086$  ( $r = 0.99931$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.15	20.300377	91.7
3	JV65	L2	True	44.30	46.075180	104.0
4	JV66	L3	True	88.60	81.857187	92.4
5	JV67	L4	True	221.50	251.465679	113.5
6	JV68	L5	True	443.00	467.377936	105.5
7	JV69	L6	True	885.00	869.552481	98.3
8	JV70	L7	True	2212.50	2036.632463	92.1
9	JV71	L8	True	4425.00	4512.772938	102.0
10	JV72	L9	True	8850.00	8906.015758	100.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	313.0 / 269.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.87836 x + 0.46617$  (r = 0.99905) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	False	50.00	2.575840	5.2
4	JV66	L3	True	100.00	78.986357	79.0
5	JV67	L4	True	250.00	251.390112	100.6
6	JV68	L5	True	500.00	541.056673	108.2
7	JV69	L6	True	1000.00	1089.329496	108.9
8	JV70	L7	True	2500.00	2659.394351	106.4
9	JV71	L8	True	5000.00	4963.910989	99.3
10	JV72	L9	True	10000.00	9765.932022	97.7






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	313.0 / 119.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06343 x + 0.00880$  (r = 0.99997) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	11.923343	47.7
3	JV65	L2	False	50.00	45.290432	90.6
4	JV66	L3	True	100.00	100.733035	100.7
5	JV67	L4	True	250.00	251.163821	100.5
6	JV68	L5	True	500.00	492.679342	98.5
7	JV69	L6	True	1000.00	1000.547914	100.1
8	JV70	L7	True	2500.00	2487.113791	99.5
9	JV71	L8	True	5000.00	5054.860411	101.1
10	JV72	L9	True	10000.00	9962.901685	99.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHpA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	363.0 / 319.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.84607x + 0.04250$  ( $r = 0.99981$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	26.704376	106.8
3	JV65	L2	True	50.00	42.475329	85.0
4	JV66	L3	True	100.00	108.532311	108.5
5	JV67	L4	True	250.00	243.123776	97.3
6	JV68	L5	True	500.00	493.515333	98.7
7	JV69	L6	True	1000.00	1007.476703	100.8
8	JV70	L7	True	2500.00	2589.189046	103.6
9	JV71	L8	True	5000.00	5029.188658	100.6
10	JV72	L9	True	10000.00	9884.794468	98.9






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHpA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	363.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.01637 x + 0.00763$  ( $r = 0.99929$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	N/A	N/A
3	JV65	L2	False	50.00	21.602557	43.2
4	JV66	L3	True	100.00	85.606667	85.6
5	JV67	L4	True	250.00	285.895686	114.4
6	JV68	L5	True	500.00	504.096919	100.8
7	JV69	L6	True	1000.00	938.783794	93.9
8	JV70	L7	True	2500.00	2665.862470	106.6
9	JV71	L8	True	5000.00	5000.525187	100.0
10	JV72	L9	True	10000.00	9869.229277	98.7








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	399.0 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.84495x + 0.00149$  ( $r = 0.99992$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	23.034700	101.0
3	JV65	L2	True	45.60	44.607323	97.8
4	JV66	L3	True	91.20	90.397278	99.1
5	JV67	L4	True	228.00	241.239806	105.8
6	JV68	L5	True	456.00	439.233831	96.3
7	JV69	L6	True	912.00	910.143123	99.8
8	JV70	L7	True	2280.00	2258.577450	99.1
9	JV71	L8	True	4560.00	4626.551146	101.5
10	JV72	L9	True	9120.00	9081.815342	99.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFHxS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	399.0 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.24827 x + 0.00372$  (r = 0.99990) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	26.747945	117.3
3	JV65	L2	True	45.60	39.964480	87.6
4	JV66	L3	True	91.20	89.427525	98.1
5	JV67	L4	True	228.00	227.854457	99.9
6	JV68	L5	True	456.00	453.922143	99.5
7	JV69	L6	True	912.00	893.198688	97.9
8	JV70	L7	True	2280.00	2238.129172	98.2
9	JV71	L8	True	4560.00	4621.692989	101.4
10	JV72	L9	True	9120.00	9124.662600	100.1






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	413.0 / 369.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.96808x + 0.19816$  ( $r = 0.99938$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	11.153654	44.6
3	JV65	L2	True	50.00	46.676599	93.4
4	JV66	L3	True	100.00	95.581325	95.6
5	JV67	L4	True	250.00	245.690478	98.3
6	JV68	L5	True	500.00	519.420773	103.9
7	JV69	L6	True	1000.00	1037.136408	103.7
8	JV70	L7	True	2500.00	2669.265130	106.8
9	JV71	L8	True	5000.00	5055.859138	101.1
10	JV72	L9	True	10000.00	9730.370148	97.3






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	413.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06473 x + 0.01078$  (r = 0.99767) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	22.005638	88.0
3	JV65	L2	True	50.00	46.540813	93.1
4	JV66	L3	True	100.00	82.992508	83.0
5	JV67	L4	True	250.00	261.053614	104.4
6	JV68	L5	True	500.00	517.072576	103.4
7	JV69	L6	True	1000.00	1070.067858	107.0
8	JV70	L7	True	2500.00	2733.741648	109.4
9	JV71	L8	True	5000.00	5284.814305	105.7
10	JV72	L9	True	10000.00	9403.716678	94.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFNA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	463.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.92776 x + 0.06645$  (r = 0.99904) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.411469	97.7
3	JV65	L2	True	50.00	41.743202	83.5
4	JV66	L3	True	100.00	98.979777	99.0
5	JV67	L4	True	250.00	251.354806	100.5
6	JV68	L5	True	500.00	542.040106	108.4
7	JV69	L6	True	1000.00	1053.161797	105.3
8	JV70	L7	True	2500.00	2705.262997	108.2
9	JV71	L8	True	5000.00	5033.084124	100.7
10	JV72	L9	True	10000.00	9674.961722	96.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFNA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	463.0 / 219.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.27039x + 0.02695$  ( $r = 0.99905$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	20.218296	80.9
3	JV65	L2	True	50.00	45.960397	91.9
4	JV66	L3	True	100.00	101.932105	101.9
5	JV67	L4	True	250.00	255.674001	102.3
6	JV68	L5	True	500.00	572.533150	114.5
7	JV69	L6	True	1000.00	1042.596703	104.3
8	JV70	L7	True	2500.00	2675.736272	107.0
9	JV71	L8	True	5000.00	5010.507489	100.2
10	JV72	L9	True	10000.00	9699.841587	97.0








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOS_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	499.0 / 80.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.16765x + 0.02406$  (r = 0.99981) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	23.361262	100.9
3	JV65	L2	True	46.30	42.153986	91.1
4	JV66	L3	True	92.60	88.414386	95.5
5	JV67	L4	True	231.50	248.496910	107.3
6	JV68	L5	True	463.00	472.188043	102.0
7	JV69	L6	True	925.60	954.991600	103.2
8	JV70	L7	True	2314.00	2301.735042	99.5
9	JV71	L8	True	4628.00	4717.827738	101.9
10	JV72	L9	True	9256.00	9130.981033	98.7






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFOS_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	499.0 / 99.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.22592 x + 0.00424$  (r = 0.99977) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	25.835005	111.6
3	JV65	L2	True	46.30	37.607900	81.2
4	JV66	L3	True	92.60	92.009960	99.4
5	JV67	L4	True	231.50	248.837656	107.5
6	JV68	L5	True	463.00	455.023640	98.3
7	JV69	L6	True	925.60	923.897296	99.8
8	JV70	L7	True	2314.00	2352.785599	101.7
9	JV71	L8	True	4628.00	4719.092631	102.0
10	JV72	L9	True	9256.00	9125.060312	98.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	513.0 / 469.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.99919x + 0.21544$  ( $r = 0.99832$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	10.368703	41.5
3	JV65	L2	False	50.00	31.340650	62.7
4	JV66	L3	True	100.00	85.661929	85.7
5	JV67	L4	True	250.00	230.445171	92.2
6	JV68	L5	True	500.00	529.199399	105.8
7	JV69	L6	True	1000.00	1093.183122	109.3
8	JV70	L7	True	2500.00	2780.129248	111.2
9	JV71	L8	True	5000.00	4948.282902	99.0
10	JV72	L9	True	10000.00	9683.098229	96.8






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	513.0 / 219.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.03953 x + 0.01934$  (r = 0.99980) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	False	50.00	< 0	N/A
4	JV66	L3	True	100.00	109.576482	109.6
5	JV67	L4	True	250.00	225.898181	90.4
6	JV68	L5	True	500.00	485.438921	97.1
7	JV69	L6	True	1000.00	1002.397258	100.2
8	JV70	L7	True	2500.00	2575.357213	103.0
9	JV71	L8	True	5000.00	5020.912753	100.4
10	JV72	L9	True	10000.00	9930.419191	99.3






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFUnA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	563.0 / 519.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.09588x + 0.04114$  ( $r = 0.99970$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.333102	97.3
3	JV65	L2	True	50.00	47.670455	95.3
4	JV66	L3	True	100.00	97.714264	97.7
5	JV67	L4	True	250.00	255.004410	102.0
6	JV68	L5	True	500.00	512.965489	102.6
7	JV69	L6	True	1000.00	1017.920510	101.8
8	JV70	L7	True	2500.00	2640.922188	105.6
9	JV71	L8	True	5000.00	4930.392260	98.6
10	JV72	L9	True	10000.00	9898.077320	99.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFUnA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	563.0 / 269.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.05167 x + 0.00535$  (r = 0.99953) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	54.213071	216.9
3	JV65	L2	True	50.00	42.011553	84.0
4	JV66	L3	True	100.00	100.910898	100.9
5	JV67	L4	True	250.00	250.795438	100.3
6	JV68	L5	True	500.00	562.545175	112.5
7	JV69	L6	True	1000.00	1000.374015	100.0
8	JV70	L7	True	2500.00	2617.419524	104.7
9	JV71	L8	True	5000.00	4924.516855	98.5
10	JV72	L9	True	10000.00	9901.426543	99.0








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFD <sub>o</sub> A_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	613.0 / 569.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.09433x + 0.06074$  ( $r = 0.99915$ ) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.809504	99.2
3	JV65	L2	True	50.00	39.835097	79.7
4	JV66	L3	True	100.00	101.286712	101.3
5	JV67	L4	True	250.00	254.074231	101.6
6	JV68	L5	True	500.00	526.830587	105.4
7	JV69	L6	True	1000.00	1087.157146	108.7
8	JV70	L7	True	2500.00	2674.514086	107.0
9	JV71	L8	True	5000.00	4994.806618	99.9
10	JV72	L9	True	10000.00	9721.686021	97.2






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFDaA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	613.0 / 319.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.17249x + 0.01513$  ( $r = 0.99963$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	26.532344	106.1
3	JV65	L2	True	50.00	42.462843	84.9
4	JV66	L3	True	100.00	96.476731	96.5
5	JV67	L4	True	250.00	263.226499	105.3
6	JV68	L5	True	500.00	506.075964	101.2
7	JV69	L6	True	1000.00	1020.038165	102.0
8	JV70	L7	True	2500.00	2621.333582	104.9
9	JV71	L8	True	5000.00	5061.671458	101.2
10	JV72	L9	True	10000.00	9787.182415	97.9






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTrDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	663.0 / 619.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.07282x + 0.05063$  ( $r = 0.99932$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.203429	100.8
3	JV65	L2	True	50.00	43.998106	88.0
4	JV66	L3	True	100.00	97.376116	97.4
5	JV67	L4	True	250.00	245.734116	98.3
6	JV68	L5	True	500.00	515.430124	103.1
7	JV69	L6	True	1000.00	1085.235268	108.5
8	JV70	L7	True	2500.00	2661.601187	106.5
9	JV71	L8	True	5000.00	4994.249530	99.9
10	JV72	L9	True	10000.00	9756.172124	97.6






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTrDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	663.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06800 x + 0.01002$  (r = 0.99958) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	27.141621	108.6
3	JV65	L2	True	50.00	52.520940	105.0
4	JV66	L3	True	100.00	80.738653	80.7
5	JV67	L4	True	250.00	254.680571	101.9
6	JV68	L5	True	500.00	475.884832	95.2
7	JV69	L6	True	1000.00	1065.280060	106.5
8	JV70	L7	True	2500.00	2604.777878	104.2
9	JV71	L8	True	5000.00	4924.491415	98.5
10	JV72	L9	True	10000.00	9939.484030	99.4






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTeDA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	713.0 / 669.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.04679x + 0.07024$  ( $r = 0.99948$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	23.929635	95.7
3	JV65	L2	True	50.00	43.700214	87.4
4	JV66	L3	True	100.00	99.162152	99.2
5	JV67	L4	True	250.00	246.757950	98.7
6	JV68	L5	True	500.00	520.679296	104.1
7	JV69	L6	True	1000.00	1080.886811	108.1
8	JV70	L7	True	2500.00	2628.108012	105.1
9	JV71	L8	True	5000.00	4957.832384	99.2
10	JV72	L9	True	10000.00	9822.873181	98.2






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	PFTeDA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	713.0 / 169.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.04914 x + 0.00899$  (r = 0.99930) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	1.189372	4.8
3	JV65	L2	True	50.00	53.808904	107.6
4	JV66	L3	True	100.00	81.366098	81.4
5	JV67	L4	True	250.00	242.225032	96.9
6	JV68	L5	True	500.00	517.332905	103.5
7	JV69	L6	True	1000.00	1064.697063	106.5
8	JV70	L7	True	2500.00	2674.759141	107.0
9	JV71	L8	True	5000.00	4954.132013	99.1
10	JV72	L9	True	10000.00	9811.678844	98.1








## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NMeFOSAA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	570.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.94846 x + 0.02409$  ( $r = 0.99947$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.648457	98.6
3	JV65	L2	True	50.00	52.409646	104.8
4	JV66	L3	True	100.00	100.762487	100.8
5	JV67	L4	True	250.00	253.928545	101.6
6	JV68	L5	True	500.00	452.857852	90.6
7	JV69	L6	True	1000.00	979.361667	97.9
8	JV70	L7	True	2500.00	2623.979649	105.0
9	JV71	L8	True	5000.00	5141.553388	102.8
10	JV72	L9	True	10000.00	9795.498310	98.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NMeFOSAA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	570.0 / 512.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.58675x + 0.04231$  ( $r = 0.99949$ ) (weighting:  $1/x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	True	50.00	59.265228	118.5
4	JV66	L3	True	100.00	88.696165	88.7
5	JV67	L4	True	250.00	244.943682	98.0
6	JV68	L5	True	500.00	476.879086	95.4
7	JV69	L6	True	1000.00	970.886314	97.1
8	JV70	L7	True	2500.00	2480.966798	99.2
9	JV71	L8	True	5000.00	5230.915871	104.6
10	JV72	L9	True	10000.00	9847.446856	98.5






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NEtFOSAA_1	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	584.0 / 419.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.90138x + 0.01437$  ( $r = 0.99896$ ) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	21.319053	85.3
3	JV65	L2	True	50.00	64.299976	128.6
4	JV66	L3	True	100.00	76.835912	76.8
5	JV67	L4	True	250.00	265.899538	106.4
6	JV68	L5	True	500.00	514.051353	102.8
7	JV69	L6	True	1000.00	951.455825	95.2
8	JV70	L7	True	2500.00	2567.794899	102.7
9	JV71	L8	True	5000.00	5262.702481	105.3
10	JV72	L9	True	10000.00	9700.640962	97.0






## Calibration Summary Report

Created with Analyst Reporter  
Printed: 23/05/2018 2:44:19 PM

<b>Analyte Name</b>	NEtFOSAA_2	<b>Data File</b>	18-0313.wiff
<b>MRM Transition</b>	584.0 / 483.0	<b>Result Table</b>	18-0313_Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/14/2018 11:20:57 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06071 x + 0.00441$  ( $r = 0.99828$ ) (weighting:  $1 / x$ )

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	N/A	N/A
3	JV65	L2	False	50.00	65.966292	131.9
4	JV66	L3	False	100.00	94.303580	94.3
5	JV67	L4	True	250.00	251.980572	100.8
6	JV68	L5	True	500.00	430.232147	86.1
7	JV69	L6	True	1000.00	1043.578594	104.4
8	JV70	L7	True	2500.00	2744.280900	109.8
9	JV71	L8	True	5000.00	5123.296865	102.5
10	JV72	L9	True	10000.00	9656.630922	96.6





Sample Name	JV64	Injection Vial	2
Sample ID	L1	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T11:29:54	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.49	PFBS	0.354	0.327	ü
PFHxA_1	313.0 / 269.0	1.79	PFHxA			
PFHxA_2	313.0 / 119.0	1.77	PFHxA	0.051	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.54	PFHpA	0.027	0.021	ü
PFHxS_1	399.0 / 80.0	2.14	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.388	0.304	ü
PFOA_1	413.0 / 369.0	2.90	PFOA			
PFOA_2	413.0 / 169.0	2.90	PFOA	0.087	0.066	ü
PFNA_1	463.0 / 419.0	3.29	PFNA			
PFNA_2	463.0 / 219.0	3.28	PFNA	0.338	0.305	ü
PFOS_1	499.0 / 80.0	2.88	PFOS			
PFOS_2	499.0 / 99.0	2.89	PFOS	0.206	0.192	ü
PFDA_1	513.0 / 469.0	3.24	PFDA			
PFDA_2	513.0 / 219.0	3.26	PFDA	0.059	0.042	ü
PFUnA_1	563.0 / 519.0	3.56	PFUnA			
PFUnA_2	563.0 / 269.0	3.60	PFUnA	0.108	0.048	
PFDaA_1	613.0 / 569.0	3.85	PFDaA			
PFDaA_2	613.0 / 319.0	3.85	PFDaA	0.183	0.162	ü
PFTrDA_1	663.0 / 619.0	4.11	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.10	PFTrDA	0.089	0.068	ü
PFTeDA_1	713.0 / 669.0	4.33	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.030	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.38	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.37	NMeFOSAA	0.253	0.664	
NEtFOSAA_1	584.0 / 419.0	3.56	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.36	NEtFOSSA	0.062	0.080	ü



<b>Sample Name</b>	JV65	<b>Injection Vial</b>	3
<b>Sample ID</b>	L2	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:38:49	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.48	PFBS	0.404	0.327	ü
PFHxA_1	313.0 / 269.0	1.79	PFHxA			
PFHxA_2	313.0 / 119.0	1.79	PFHxA	0.077	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.028	0.021	ü
PFHxS_1	399.0 / 80.0	2.14	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.288	0.304	ü
PFOA_1	413.0 / 369.0	2.50	PFOA			
PFOA_2	413.0 / 169.0	2.51	PFOA	0.063	0.066	ü
PFNA_1	463.0 / 419.0	2.88	PFNA			
PFNA_2	463.0 / 219.0	2.88	PFNA	0.333	0.305	ü
PFOS_1	499.0 / 80.0	2.87	PFOS			
PFOS_2	499.0 / 99.0	2.87	PFOS	0.173	0.192	ü
PFDA_1	513.0 / 469.0	3.24	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.036	0.042	ü
PFUnA_1	563.0 / 519.0	3.56	PFUnA			
PFUnA_2	563.0 / 269.0	3.50	PFUnA	0.048	0.048	ü
PFDaA_1	613.0 / 569.0	3.85	PFDaA			
PFDaA_2	613.0 / 319.0	3.84	PFDaA	0.178	0.162	ü
PFTrDA_1	663.0 / 619.0	4.10	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.09	PFTrDA	0.088	0.068	ü
PFTeDA_1	713.0 / 669.0	4.33	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.067	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.39	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.41	NMeFOSAA	0.871	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.55	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.60	NEtFOSSA	0.089	0.080	ü

<b>Sample Name</b>	JV66	<b>Injection Vial</b>	4
<b>Sample ID</b>	L3	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:47:45	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.49	PFBS			
PFBS_2	298.9 / 99.0	1.48	PFBS	0.326	0.327	ü
PFHxA_1	313.0 / 269.0	1.80	PFHxA			
PFHxA_2	313.0 / 119.0	1.78	PFHxA	0.063	0.066	ü
PFHpA_1	363.0 / 319.0	2.12	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.023	0.021	ü
PFHxS_1	399.0 / 80.0	2.13	PFHxS			
PFHxS_2	399.0 / 99.0	2.14	PFHxS	0.303	0.304	ü
PFOA_1	413.0 / 369.0	2.50	PFOA			
PFOA_2	413.0 / 169.0	2.50	PFOA	0.057	0.066	ü
PFNA_1	463.0 / 419.0	2.88	PFNA			
PFNA_2	463.0 / 219.0	2.88	PFNA	0.307	0.299	ü
PFOS_1	499.0 / 80.0	2.87	PFOS			
PFOS_2	499.0 / 99.0	2.87	PFOS	0.200	0.192	ü
PFDA_1	513.0 / 469.0	3.23	PFDA			
PFDA_2	513.0 / 219.0	3.23	PFDA	0.059	0.042	ü
PFUnA_1	563.0 / 519.0	3.55	PFUnA			
PFUnA_2	563.0 / 269.0	3.56	PFUnA	0.052	0.048	ü
PFDaA_1	613.0 / 569.0	3.84	PFDaA			
PFDaA_2	613.0 / 319.0	3.84	PFDaA	0.155	0.162	ü
PFTrDA_1	663.0 / 619.0	4.10	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.09	PFTrDA	0.059	0.068	ü
PFTeDA_1	713.0 / 669.0	4.32	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.32	PFTeDA	0.044	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.38	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.38	NMeFOSAA	0.656	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.54	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.47	NEtFOSSA	0.100	0.068	ü

<b>Sample Name</b>	JV67	<b>Injection Vial</b>	5
<b>Sample ID</b>	L4	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T11:56:42	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.321	0.327	ü
PFHxA_1	313.0 / 269.0	1.76	PFHxA			
PFHxA_2	313.0 / 119.0	1.76	PFHxA	0.063	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.026	0.021	ü
PFHxS_1	399.0 / 80.0	2.13	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.282	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.070	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.300	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.193	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.043	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.048	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.165	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.068	0.068	ü
PFTeDA_1	713.0 / 669.0	4.31	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.641	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.53	NEtFOSSA	0.070	0.080	ü

<b>Sample Name</b>	JV68	<b>Injection Vial</b>	6
<b>Sample ID</b>	L5	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:05:38	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.48	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.304	0.327	ü
PFHxA_1	313.0 / 269.0	1.76	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.062	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.021	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.306	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.066	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.309	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.186	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.23	PFDA	0.038	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.052	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.152	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.060	0.068	ü
PFTeDA_1	713.0 / 669.0	4.31	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.676	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.059	0.080	ü

<b>Sample Name</b>	JV69	<b>Injection Vial</b>	7
<b>Sample ID</b>	L6	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:14:33	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.064	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.13	PFHxS	0.290	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.069	0.066	ü
PFNA_1	463.0 / 419.0	2.87	PFNA			
PFNA_2	463.0 / 219.0	2.87	PFNA	0.289	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.187	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.22	PFDA	0.037	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.54	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.83	PFDaA			
PFDaA_2	613.0 / 319.0	3.83	PFDaA	0.148	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.37	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.625	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.53	NEtFOSSA	0.075	0.080	ü

<b>Sample Name</b>	JV70	<b>Injection Vial</b>	8
<b>Sample ID</b>	L7	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:23:30	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.301	0.327	ü
PFHxA_1	313.0 / 269.0	1.75	PFHxA			
PFHxA_2	313.0 / 119.0	1.75	PFHxA	0.067	0.066	ü
PFHpA_1	363.0 / 319.0	2.11	PFHpA			
PFHpA_2	363.0 / 169.0	2.11	PFHpA	0.020	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.292	0.304	ü
PFOA_1	413.0 / 369.0	2.49	PFOA			
PFOA_2	413.0 / 169.0	2.49	PFOA	0.068	0.066	ü
PFNA_1	463.0 / 419.0	2.86	PFNA			
PFNA_2	463.0 / 219.0	2.86	PFNA	0.289	0.305	ü
PFOS_1	499.0 / 80.0	2.86	PFOS			
PFOS_2	499.0 / 99.0	2.86	PFOS	0.198	0.192	ü
PFDA_1	513.0 / 469.0	3.22	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.037	0.042	ü
PFUnA_1	563.0 / 519.0	3.53	PFUnA			
PFUnA_2	563.0 / 269.0	3.53	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.82	PFDaA			
PFDaA_2	613.0 / 319.0	3.82	PFDaA	0.155	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.062	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.048	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.589	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.073	0.080	ü



<b>Sample Name</b>	JV71	<b>Injection Vial</b>	9
<b>Sample ID</b>	L8	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:32:26	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.308	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.12	PFHxS			
PFHxS_2	399.0 / 99.0	2.12	PFHxS	0.294	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.070	0.066	ü
PFNA_1	463.0 / 419.0	2.86	PFNA			
PFNA_2	463.0 / 219.0	2.86	PFNA	0.290	0.305	ü
PFOS_1	499.0 / 80.0	2.85	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.194	0.192	ü
PFDA_1	513.0 / 469.0	3.21	PFDA			
PFDA_2	513.0 / 219.0	3.21	PFDA	0.040	0.042	ü
PFUnA_1	563.0 / 519.0	3.54	PFUnA			
PFUnA_2	563.0 / 269.0	3.53	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.82	PFDaA			
PFDaA_2	613.0 / 319.0	3.82	PFDaA	0.160	0.162	ü
PFTrDA_1	663.0 / 619.0	4.08	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.08	PFTrDA	0.063	0.068	ü
PFTeDA_1	713.0 / 669.0	4.30	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.30	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.36	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.36	NMeFOSAA	0.632	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.53	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.52	NEtFOSSA	0.066	0.080	ü

<b>Sample Name</b>	JV72	<b>Injection Vial</b>	10
<b>Sample ID</b>	L9	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-14T12:41:22	<b>Data File</b>	18-0313.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0313_Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.311	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.74	PFHxA	0.073	0.066	ü
PFHpA_1	363.0 / 319.0	2.10	PFHpA			
PFHpA_2	363.0 / 169.0	2.10	PFHpA	0.019	0.021	ü
PFHxS_1	399.0 / 80.0	2.11	PFHxS			
PFHxS_2	399.0 / 99.0	2.11	PFHxS	0.295	0.304	ü
PFOA_1	413.0 / 369.0	2.48	PFOA			
PFOA_2	413.0 / 169.0	2.48	PFOA	0.065	0.066	ü
PFNA_1	463.0 / 419.0	2.85	PFNA			
PFNA_2	463.0 / 219.0	2.85	PFNA	0.292	0.305	ü
PFOS_1	499.0 / 80.0	2.84	PFOS			
PFOS_2	499.0 / 99.0	2.85	PFOS	0.193	0.192	ü
PFDA_1	513.0 / 469.0	3.20	PFDA			
PFDA_2	513.0 / 219.0	3.20	PFDA	0.041	0.042	ü
PFUnA_1	563.0 / 519.0	3.52	PFUnA			
PFUnA_2	563.0 / 269.0	3.52	PFUnA	0.047	0.048	ü
PFDaA_1	613.0 / 569.0	3.81	PFDaA			
PFDaA_2	613.0 / 319.0	3.81	PFDaA	0.159	0.162	ü
PFTrDA_1	663.0 / 619.0	4.07	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.06	PFTrDA	0.065	0.068	ü
PFTeDA_1	713.0 / 669.0	4.29	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.28	PFTeDA	0.047	0.050	ü
NMeFOSAA_1	570.0 / 419.0	3.35	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.35	NMeFOSAA	0.623	0.664	ü
NEtFOSAA_1	584.0 / 419.0	3.52	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.51	NEtFOSSA	0.067	0.080	ü

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.46	805.907678	885.00	91.06
PFBS_2	298.9 / 99.0	1.46	825.071704	885.00	93.23
PFHxA_1	313.0 / 269.0	1.75	962.249904	1000.00	96.22
PFHxA_2	313.0 / 119.0	1.74	1015.156998	1000.00	101.52
PFHpA_1	363.0 / 319.0	2.10	933.247533	1000.00	93.32
PFHpA_2	363.0 / 169.0	2.10	933.903106	1000.00	93.39
PFHxS_1	399.0 / 80.0	2.12	861.806904	912.00	94.50
PFHxS_2	399.0 / 99.0	2.12	826.048162	912.00	90.58
PFOA_1	413.0 / 369.0	2.48	941.626058	1000.00	94.16
PFOA_2	413.0 / 169.0	2.48	1065.520264	1000.00	106.55
PFNA_1	463.0 / 419.0	2.85	969.501428	1000.00	96.95
PFNA_2	463.0 / 219.0	2.85	1015.813698	1000.00	101.58
PFOS_1	499.0 / 80.0	2.85	830.716293	925.60	89.75
PFOS_2	499.0 / 99.0	2.85	936.586590	925.60	101.19
PFDA_1	513.0 / 469.0	3.21	952.414243	1000.00	95.24
PFDA_2	513.0 / 219.0	3.21	866.424928	1000.00	86.64
PFUnA_1	563.0 / 519.0	3.53	983.154309	1000.00	98.32
PFUnA_2	563.0 / 269.0	3.53	953.567429	1000.00	95.36
PFDoA_1	613.0 / 569.0	3.81	985.864788	1000.00	98.59
PFDoA_2	613.0 / 319.0	3.81	953.497275	1000.00	95.35
PFTTrDA_1	663.0 / 619.0	4.07	976.486543	1000.00	97.65
PFTTrDA_2	663.0 / 169.0	4.07	961.413908	1000.00	96.14
PFTeDA_1	713.0 / 669.0	4.29	938.482516	1000.00	93.85
PFTeDA_2	713.0 / 169.0	4.28	952.593795	1000.00	95.26
NMeFOSAA_1	570.0 / 419.0	3.36	1109.562943	1000.00	110.96
NMeFOSAA_2	570.0 / 512.0	3.36	980.360841	1000.00	98.04
NEtFOSAA_1	584.0 / 419.0	3.52	1088.808929	1000.00	108.88
NEtFOSAA_2	584.0 / 483.0	3.52	957.541377	1000.00	95.75

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T12:50:20	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.73	92.204529	100.00	92.20
13C2-PFDA	515.0 / 470.0	3.19	92.949919	100.00	92.95
d5-EtFOSAA	589.0 / 419.0	3.51	381.212971	400.00	95.30

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.46	899.387330	885.00	101.63
PFBS_2	298.9 / 99.0	1.45	900.526623	885.00	101.75
PFHxA_1	313.0 / 269.0	1.74	1183.308886	1000.00	118.33
PFHxA_2	313.0 / 119.0	1.73	1192.266265	1000.00	119.23
PFHpA_1	363.0 / 319.0	2.09	1094.466247	1000.00	109.45
PFHpA_2	363.0 / 169.0	2.09	1068.460245	1000.00	106.85
PFHxS_1	399.0 / 80.0	2.10	913.078700	912.00	100.12
PFHxS_2	399.0 / 99.0	2.11	872.644722	912.00	95.68
PFOA_1	413.0 / 369.0	2.47	1115.848591	1000.00	111.58
PFOA_2	413.0 / 169.0	2.47	1136.973471	1000.00	113.70
PFNA_1	463.0 / 419.0	2.84	1171.510532	1000.00	117.15
PFNA_2	463.0 / 219.0	2.84	1212.351411	1000.00	121.24
PFOS_1	499.0 / 80.0	2.84	946.924957	925.60	102.30
PFOS_2	499.0 / 99.0	2.84	945.562142	925.60	102.16
PFDA_1	513.0 / 469.0	3.20	1186.150677	1000.00	118.62
PFDA_2	513.0 / 219.0	3.20	1102.602568	1000.00	110.26
PFUnA_1	563.0 / 519.0	3.51	1152.561588	1000.00	115.26
PFUnA_2	563.0 / 269.0	3.51	1141.269070	1000.00	114.13
PFDoA_1	613.0 / 569.0	3.80	1144.819746	1000.00	114.48
PFDoA_2	613.0 / 319.0	3.80	1111.307165	1000.00	111.13
PFTTrDA_1	663.0 / 619.0	4.05	1134.100929	1000.00	113.41
PFTTrDA_2	663.0 / 169.0	4.05	1115.502894	1000.00	111.55
PFTeDA_1	713.0 / 669.0	4.27	1145.940772	1000.00	114.59
PFTeDA_2	713.0 / 169.0	4.27	1129.570434	1000.00	112.96
NMeFOSAA_1	570.0 / 419.0	3.34	1244.481183	1000.00	124.45
NMeFOSAA_2	570.0 / 512.0	3.34	1216.883830	1000.00	121.69
NEtFOSAA_1	584.0 / 419.0	3.51	1169.751144	1000.00	116.98
NEtFOSAA_2	584.0 / 483.0	3.49	1069.898281	1000.00	106.99

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T14:19:36	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.72	107.339437	100.00	107.34
13C2-PFDA	515.0 / 470.0	3.18	100.831039	100.00	100.83
d5-EtFOSAA	589.0 / 419.0	3.50	427.992074	400.00	107.00



Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.45	440.416958	443.00	99.42
PFBS_2	298.9 / 99.0	1.45	434.150471	443.00	98.00
PFHxA_1	313.0 / 269.0	1.73	499.036672	500.00	99.81
PFHxA_2	313.0 / 119.0	1.73	528.285027	500.00	105.66
PFHpA_1	363.0 / 319.0	2.09	471.408594	500.00	94.28
PFHpA_2	363.0 / 169.0	2.08	558.410072	500.00	111.68
PFHxS_1	399.0 / 80.0	2.10	441.193340	456.00	96.75
PFHxS_2	399.0 / 99.0	2.10	428.872944	456.00	94.05
PFOA_1	413.0 / 369.0	2.46	469.777523	500.00	93.96
PFOA_2	413.0 / 169.0	2.46	515.299469	500.00	103.06
PFNA_1	463.0 / 419.0	2.83	523.788059	500.00	104.76
PFNA_2	463.0 / 219.0	2.83	505.535719	500.00	101.11
PFOS_1	499.0 / 80.0	2.83	460.723951	463.00	99.51
PFOS_2	499.0 / 99.0	2.83	477.544848	463.00	103.14
PFDA_1	513.0 / 469.0	3.18	489.972812	500.00	97.99
PFDA_2	513.0 / 219.0	3.18	465.489745	500.00	93.10
PFUnA_1	563.0 / 519.0	3.50	518.325361	500.00	103.67
PFUnA_2	563.0 / 269.0	3.50	471.585271	500.00	94.32
PFDoA_1	613.0 / 569.0	3.79	476.947097	500.00	95.39
PFDoA_2	613.0 / 319.0	3.79	505.027230	500.00	101.01
PFTTrDA_1	663.0 / 619.0	4.04	479.806082	500.00	95.96
PFTTrDA_2	663.0 / 169.0	4.03	501.939831	500.00	100.39
PFTeDA_1	713.0 / 669.0	4.25	485.629530	500.00	97.13
PFTeDA_2	713.0 / 169.0	4.25	486.803296	500.00	97.36
NMeFOSAA_1	570.0 / 419.0	3.34	487.813835	500.00	97.56
NMeFOSAA_2	570.0 / 512.0	3.34	448.600228	500.00	89.72
NEtFOSAA_1	584.0 / 419.0	3.50	482.711585	500.00	96.54
NEtFOSAA_2	584.0 / 483.0	3.49	625.585443	500.00	125.12

Sample Name	JV68 CCV	Injection Vial	6
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T15:30:58	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.71	95.522162	100.00	95.52
13C2-PFDA	515.0 / 470.0	3.17	92.313155	100.00	92.31
d5-EtFOSAA	589.0 / 419.0	3.49	357.700451	400.00	89.43

## QA/QC Summary Batch 18-0316

Project:	CTO-WE04 Naval Air Station Joint Reserve Base Willow Grove
Parameters:	PFAS
Laboratory:	Battelle, Norwell, MA
Matrix:	DW
Data Set:	DP-18-0117
Analytical SOP:	5-371
Method Reference:	USEPA 537 rev. 1.1, QSM 5.1

### Sample Custody

Collection Date	Receipt Date	Temp (°C)
5/7/2018	5/8/2018	0.9, 1.2

Corrective Actions	None
Sample Storage	The water samples were stored refrigerated until extraction.
Related samples	Field samples related to these FRB samples reported in SDG 18-0313

### METHOD SUMMARIES

Sample Preparation	Water samples were spiked with surrogates in the original sample container from the field. The water was extracted using a weak ion exchange solid phase extraction (SPE) cartridge and eluted from the SPE with methanol. Extracts were split and concentrated to dryness under nitrogen with a water bath set between 60 °C and 65 °C, reconstituted with 96:4 methanol/water (V/V) and fortified with internal standard. Extracts were transferred for LC-MS/MS analysis.	
Prep comments	None.	
Analysis	PFAS were measured by liquid chromatography tandem mass spectrometry (LC-MS/MS) in the multiple reaction monitoring (MRM). An initial calibration consisting of representative target analytes, labelled analogs, and internal standards was analyzed prior to analysis to demonstrate the linear range of analysis. Calibration verification was performed at the beginning and end of 10 injections and at the end of each sequence. Target PFAS were quantified using the isotope dilution method. Samples are reported in ng/L concentrations.	
Analysis Comments	<p>Samples analyzed on the Sciex 5500.</p> <p>Individual ion ratios were outside of the 50% RPD criteria for the following samples and analytes, however, all hits were below the detection limits:</p> <p>Procedural Blank (CQ774PB) - PFOA  NAWC-050718-FRB-180 (J6151) - PFOA  NAWC-050718-FRB-145 (J6155) - PFOA  NAWC-050718-FRB-357 (J6157) – PFOA, PFNA  NAWC-050718-FRB-162 (J6159) - PFNA  WGNA-050718-FRB-0800 (J6161) - PFOA  WGNA-050718-FRB-3556 (J6165) - PFOA  NAWC-050718-FRB-356 (J6167) - PFOA</p>	
Holding Times	Extraction Date(s)	Analysis Date(s)
	5/15/2018	5/17/2018

**QA/QC Summary**  
**Batch 18-0316**

Procedural Blank (PB)	A PB was prepared with this analytical batch to ensure the sample extraction and analysis methods are free of contamination.
≤ 1/3 the MRL	No exceedances noted. No comments.
Laboratory Control Spike (LCS)	A LCS was prepared with this analytical batch. The percent recoveries of target analytes were calculated to measure accuracy.
70-130% of true value	No exceedances noted. No comments.
Surrogates Standard Analytes	Labelled surrogate compounds were added prior to extraction. The recoveries are calculated to measure extraction efficiency.
70-130% of true value	No exceedances noted. No Comments.
Internal Standard Analytes	Labelled analog compounds were added prior to analysis.
ICal high and low points RPD ≤20%, 50-150% of average area of the ICAL and 70-140% of most recent CCV	No exceedances noted. No Comments.
Initial Calibration (ICAL)	The LC-MS/MS was calibrated with multi-level calibration curve for all compounds using linear or quadratic curve fitting.
R <sup>2</sup> >0.99	No exceedances noted.
Target and SIS compounds +/- 30% of true value, Low point 50-150% of true value	No comments.
Independent Calibration Check (ICC)	The independent check was run after each initial calibration to verify the calibration. This standard is from a different source than the ICAL.
Target and SIS compounds +/- 30% of true value	No exceedances noted. No comments.
Continuing Calibration Verification (CCV)	Continuing calibration standards were run at the beginning and end of 10 injections and at the end of the sequence to ensure that initial calibration is still valid.

**QA/QC Summary  
Batch 18-0316**

Target and SIS compounds +/- 30% of true value Low point 50- 150% of true value	No exceedances noted.
	No comments.

**BATTELLE**

It can be done

**BATTELLE - NORWELL OPERATIONS  
MISCELLANEOUS DOCUMENTATION FORM**

**Project Title:** Naval Air Station Joint Reserve Base Wi      **Data Set Number:** DP-18-0117  
**Project Number:** 100117920-WE04      **Prep Batch Number:** 18-0316  
**Entered By:** Denise Schumitz      **Entered On:** 05/24/2018  
**Test Code (Matrix Type):** Master\_371(L)

Samples that were manually integrated are noted on the quant reports with the comment (TRUE).  
DMS 6/24/2018

JV64 is not being used in the calibration curve for PFBS and NMeFOSAA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/23/2018

JV65 is not being used in the calibration curve for NMeFOSAA. There is no impact on the data once this point is removed from the calibration.  
DMS 6/23/2018

Ion ratios greater than 50% for PFOA are noted in samples CQ774PB, J6151, J6155, J6157, J6161, J6165, J6167 and J6169.  
DMS 6/23/2018

Ion ratios greater than 50% for PFNA are noted in samples J6157 and J6159.  
DMS 6/23/2018

Ion ratio greater than 50% for PFDA is noted in JV66.  
DMS 6/23/2018

**Task Leader Approval:****Supervisor Approval:****PM Approval:**


Digitally signed by Jonathan Thorn

Date: 2018.05.29 11:07:31 -04'00'





**It can be done**

Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project Number: 100117920-WE04  
 Preparation Batch: 18-0316  
 Data Set: DP-18-0117  
 Test Code: Master\_371

QC Parameter:	Exceed:	Justification:
Procedural Blank	0	None
PB Measurement Quality Objective	0	None
Laboratory Control Sample	0	None
Matrix Spike / Matrix Spike Duplicate Recovery	NA	None
Matrix Spike / Matrix Spike Duplicate Precision	NA	None
Extracted Internal Standard Analytes (Surrogates)	0	None
Instrument Calibration	0	None
Instrument Blank	NA	None
Independent Calibration Check	0	None
Continuing Calibration Verification	0	None



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04  
 Preparation Batch: 18-0316  
 Data Set: DP-18-0117

	CQ774PB-FS (Procedural Blank)	CQ775LCS-FS (Laboratory Control Sample)	J6149-FS (NAWC-050718-FRB-316)	J6151-FS (NAWC-050718-FRB-180)	J6153-FS (NAWC-050718-FRB-275)	J6155-FS (NAWC-050718-FRB-145)
PFHxA	-	L	-	-	-	-
PFHpA	-	L	-	-	-	-
PFOA	-	L	-	-	-	-
PFNA	-	L	-	-	-	-
PFDA	-	L	-	-	-	-
PFUnA	-	L	-	-	-	-
PFDoA	-	L	-	-	-	-
PFTrDA	-	L	-	-	-	-
PFTeDA	-	L	-	-	-	-
NMeFOSAA	-	L	-	-	-	-
NEtFOSAA	-	L	-	-	-	-
PFBS	-	L	-	-	-	-
PFHxS	-	L	-	-	-	-
PFOS	-	L/Br	-	-	-	-

"L": Linear  
 "Br": branched  
 "L/Br": Linear/Branched  
 "-": Not detected



Project Client: Tetra Tech  
 Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA  
 Project No.: 100117920-WE04  
 Preparation Batch: 18-0316  
 Data Set: DP-18-0:

	J6157-FS (NAWC-050718-FRB-357)	J6159-FS (NAWC-050718-FRB-162)	J6161-FS (WGNA-050718-FRB-0800)	J6163-FS (WGNA-050718-FRB-0335)	J6165-FS (WGNA-050718-FRB-3556)	J6167-FS (NAWC-050718-FRB-356)	J6169-FS (NAWC-050718-FRB-289)
PFHxA	-	-	-	-	-	-	-
PFHpA	-	-	-	-	-	-	-
PFOA	-	-	-	-	-	-	-
PFNA	-	-	-	-	-	-	-
PFDA	-	-	-	-	-	-	-
PFUnA	-	-	-	-	-	-	-
PFDoA	-	-	-	-	-	-	-
PFTTrDA	-	-	-	-	-	-	-
PFTeDA	-	-	-	-	-	-	-
NMeFOSAA	-	-	-	-	-	-	-
NEtFOSAA	-	-	-	-	-	-	-
PFBS	-	-	-	-	-	-	-
PFHxS	-	-	-	-	-	-	-
PFOS	-	-	-	-	-	-	-

"L": Linear  
 "Br": branched  
 "L/Br": Linear/Branched  
 "-": Not detected



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE IDENTIFICATION PAGE

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0316****WE04 PFAS Analysis****DW**

Sample ID	Description
CQ774PB-FS	Procedural Blank
CQ775LCS-FS	Laboratory Control Sample
J6149-FS	NAWC-050718-FRB-316
J6151-FS	NAWC-050718-FRB-180
J6153-FS	NAWC-050718-FRB-275
J6155-FS	NAWC-050718-FRB-145
J6157-FS	NAWC-050718-FRB-357
J6159-FS	NAWC-050718-FRB-162
J6161-FS	WGNA-050718-FRB-0800
J6163-FS	WGNA-050718-FRB-0335
J6165-FS	WGNA-050718-FRB-3556
J6167-FS	NAWC-050718-FRB-356
J6169-FS	NAWC-050718-FRB-289

Samples Assigned By:

Stephanie Schultz

Date :

May 15, 2018

Comments:



Project Client: Tetra Tech

Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA

Project No.: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/17/18 9:20	13C4-PFOS	106,482.72	-
JV65	L2	5/17/18 9:29	13C4-PFOS	119,997.89	-
JV66	L3	5/17/18 9:38	13C4-PFOS	122,253.98	-
JV67	L4	5/17/18 9:47	13C4-PFOS	102,094.68	-
JV68	L5	5/17/18 9:56	13C4-PFOS	111,836.08	-
JV69	L6	5/17/18 10:05	13C4-PFOS	114,992.54	-
JV70	L7	5/17/18 10:14	13C4-PFOS	110,302.85	-
JV71	L8	5/17/18 10:23	13C4-PFOS	100,136.69	-
JV72	L9	5/17/18 10:31	13C4-PFOS	115,397.28	8.0

PASS

Average 111,499.41 Lower 55,749.71 Upper 167,249.12

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/17/18 9:20	13C4-PFOS	106,482.72	55,749.71	167,249.12		78,285.26	156,570.51	
JV65	L2	5/17/18 9:29	13C4-PFOS	119,997.89	55,749.71	167,249.12		78,285.26	156,570.51	
JV66	L3	5/17/18 9:38	13C4-PFOS	122,253.98	55,749.71	167,249.12		78,285.26	156,570.51	
JV67	L4	5/17/18 9:47	13C4-PFOS	102,094.68	55,749.71	167,249.12		78,285.26	156,570.51	
JV68	L5	5/17/18 9:56	13C4-PFOS	111,836.08	55,749.71	167,249.12		78,285.26	156,570.51	
JV69	L6	5/17/18 10:05	13C4-PFOS	114,992.54	55,749.71	167,249.12		78,285.26	156,570.51	
JV70	L7	5/17/18 10:14	13C4-PFOS	110,302.85	55,749.71	167,249.12		78,285.26	156,570.51	
JV71	L8	5/17/18 10:23	13C4-PFOS	100,136.69	55,749.71	167,249.12		78,285.26	156,570.51	
JV72	L9	5/17/18 10:31	13C4-PFOS	115,397.28	55,749.71	167,249.12		78,285.26	156,570.51	
JV63 ICC	ICC	5/17/18 10:40	13C4-PFOS	113,811.02	55,749.71	167,249.12		78,285.26	156,570.51	
CQ774PB-FS(0)	Procedural Blank	5/17/18 14:35	13C4-PFOS	105,807.87	55,749.71	167,249.12		78,285.26	156,570.51	
CQ775LCS-FS(0)	Laboratory Control Sample	5/17/18 14:44	13C4-PFOS	84,608.36	55,749.71	167,249.12		78,285.26	156,570.51	
J6149-FS(0)	NAWC-050718-FRB-316	5/17/18 14:53	13C4-PFOS	95,906.86	55,749.71	167,249.12		78,285.26	156,570.51	
J6151-FS(0)	NAWC-050718-FRB-180	5/17/18 15:02	13C4-PFOS	82,833.06	55,749.71	167,249.12		78,285.26	156,570.51	
J6153-FS(0)	NAWC-050718-FRB-275	5/17/18 15:11	13C4-PFOS	89,739.14	55,749.71	167,249.12		78,285.26	156,570.51	
J6155-FS(0)	NAWC-050718-FRB-145	5/17/18 15:20	13C4-PFOS	83,891.29	55,749.71	167,249.12		78,285.26	156,570.51	
J6157-FS(0)	NAWC-050718-FRB-357	5/17/18 15:29	13C4-PFOS	87,652.64	55,749.71	167,249.12		78,285.26	156,570.51	
J6159-FS(0)	NAWC-050718-FRB-162	5/17/18 15:38	13C4-PFOS	99,454.46	55,749.71	167,249.12		78,285.26	156,570.51	
JV70 CCV	CCV	5/17/18 15:47	13C4-PFOS	114,163.67	55,749.71	167,249.12		78,285.26	156,570.51	
J6161-FS(0)	WGNA-050718-FRB-0800	5/17/18 16:04	13C4-PFOS	109,563.37	55,749.71	167,249.12		79,914.57	159,829.14	
J6163-FS(0)	WGNA-050718-FRB-0335	5/17/18 16:13	13C4-PFOS	94,266.70	55,749.71	167,249.12		79,914.57	159,829.14	
J6165-FS(0)	WGNA-050718-FRB-3556	5/17/18 16:22	13C4-PFOS	84,277.54	55,749.71	167,249.12		79,914.57	159,829.14	
J6167-FS(0)	NAWC-050718-FRB-356	5/17/18 16:31	13C4-PFOS	98,363.28	55,749.71	167,249.12		79,914.57	159,829.14	
J6169-FS(0)	NAWC-050718-FRB-289	5/17/18 16:40	13C4-PFOS	85,338.73	55,749.71	167,249.12		79,914.57	159,829.14	
JV69 CCV	CCV	5/17/18 16:49	13C4-PFOS	121,671.75	55,749.71	167,249.12		79,914.57	159,829.14	



Project Client: Tetra Tech

Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA

Project No.: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/17/18 9:20	13C2-PFOA	28,448.80	-
JV65	L2	5/17/18 9:29	13C2-PFOA	30,346.55	-
JV66	L3	5/17/18 9:38	13C2-PFOA	31,920.71	-
JV67	L4	5/17/18 9:47	13C2-PFOA	28,765.39	-
JV68	L5	5/17/18 9:56	13C2-PFOA	33,859.37	-
JV69	L6	5/17/18 10:05	13C2-PFOA	32,336.53	-
JV70	L7	5/17/18 10:14	13C2-PFOA	32,706.57	-
JV71	L8	5/17/18 10:23	13C2-PFOA	29,031.43	-
JV72	L9	5/17/18 10:31	13C2-PFOA	33,993.78	17.8

PASS

Average 31,267.68 Lower 15,633.84 Upper 46,901.52

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/17/18 9:20	13C2-PFOA	28,448.80	15,633.84	46,901.52		23,701.56	47,403.12	
JV65	L2	5/17/18 9:29	13C2-PFOA	30,346.55	15,633.84	46,901.52		23,701.56	47,403.12	
JV66	L3	5/17/18 9:38	13C2-PFOA	31,920.71	15,633.84	46,901.52		23,701.56	47,403.12	
JV67	L4	5/17/18 9:47	13C2-PFOA	28,765.39	15,633.84	46,901.52		23,701.56	47,403.12	
JV68	L5	5/17/18 9:56	13C2-PFOA	33,859.37	15,633.84	46,901.52		23,701.56	47,403.12	
JV69	L6	5/17/18 10:05	13C2-PFOA	32,336.53	15,633.84	46,901.52		23,701.56	47,403.12	
JV70	L7	5/17/18 10:14	13C2-PFOA	32,706.57	15,633.84	46,901.52		23,701.56	47,403.12	
JV71	L8	5/17/18 10:23	13C2-PFOA	29,031.43	15,633.84	46,901.52		23,701.56	47,403.12	
JV72	L9	5/17/18 10:31	13C2-PFOA	33,993.78	15,633.84	46,901.52		23,701.56	47,403.12	
JV63 ICC	ICC	5/17/18 10:40	13C2-PFOA	35,067.86	15,633.84	46,901.52		23,701.56	47,403.12	
CQ774PB-FS(0)	Procedural Blank	5/17/18 14:35	13C2-PFOA	29,243.16	15,633.84	46,901.52		23,701.56	47,403.12	
CQ775LCS-FS(0)	Laboratory Control Sample	5/17/18 14:44	13C2-PFOA	25,000.07	15,633.84	46,901.52		23,701.56	47,403.12	
J6149-FS(0)	NAWC-050718-FRB-316	5/17/18 14:53	13C2-PFOA	29,366.38	15,633.84	46,901.52		23,701.56	47,403.12	
J6151-FS(0)	NAWC-050718-FRB-180	5/17/18 15:02	13C2-PFOA	23,903.88	15,633.84	46,901.52		23,701.56	47,403.12	
J6153-FS(0)	NAWC-050718-FRB-275	5/17/18 15:11	13C2-PFOA	25,373.77	15,633.84	46,901.52		23,701.56	47,403.12	
J6155-FS(0)	NAWC-050718-FRB-145	5/17/18 15:20	13C2-PFOA	24,418.17	15,633.84	46,901.52		23,701.56	47,403.12	
J6157-FS(0)	NAWC-050718-FRB-357	5/17/18 15:29	13C2-PFOA	24,844.34	15,633.84	46,901.52		23,701.56	47,403.12	
J6159-FS(0)	NAWC-050718-FRB-162	5/17/18 15:38	13C2-PFOA	27,428.30	15,633.84	46,901.52		23,701.56	47,403.12	
JV70 CCV	CCV	5/17/18 15:47	13C2-PFOA	31,530.88	15,633.84	46,901.52		23,701.56	47,403.12	
J6161-FS(0)	WGNA-050718-FRB-0800	5/17/18 16:04	13C2-PFOA	30,018.42	15,633.84	46,901.52		22,071.62	44,143.23	
J6163-FS(0)	WGNA-050718-FRB-0335	5/17/18 16:13	13C2-PFOA	25,903.21	15,633.84	46,901.52		22,071.62	44,143.23	
J6165-FS(0)	WGNA-050718-FRB-3556	5/17/18 16:22	13C2-PFOA	25,672.65	15,633.84	46,901.52		22,071.62	44,143.23	
J6167-FS(0)	NAWC-050718-FRB-356	5/17/18 16:31	13C2-PFOA	27,021.98	15,633.84	46,901.52		22,071.62	44,143.23	
J6169-FS(0)	NAWC-050718-FRB-289	5/17/18 16:40	13C2-PFOA	24,296.54	15,633.84	46,901.52		22,071.62	44,143.23	
JV69 CCV	CCV	5/17/18 16:49	13C2-PFOA	33,798.22	15,633.84	46,901.52		22,071.62	44,143.23	





Project Client: Tetra Tech

Project Name: Naval Air Station Joint Reserve Base Willow Grove, PA

Project No.: 100117920-WE04

Sample Name	Sample ID	Analysis Date	Analyte	Area	RPD (L1/L9)
JV64	L1	5/17/18 9:20	d3-MeFOSAA	16,123.52	-
JV65	L2	5/17/18 9:29	d3-MeFOSAA	17,769.57	-
JV66	L3	5/17/18 9:38	d3-MeFOSAA	19,084.74	-
JV67	L4	5/17/18 9:47	d3-MeFOSAA	17,771.51	-
JV68	L5	5/17/18 9:56	d3-MeFOSAA	18,199.88	-
JV69	L6	5/17/18 10:05	d3-MeFOSAA	17,045.59	-
JV70	L7	5/17/18 10:14	d3-MeFOSAA	18,272.74	-
JV71	L8	5/17/18 10:23	d3-MeFOSAA	15,851.69	-
JV72	L9	5/17/18 10:31	d3-MeFOSAA	18,290.15	12.6

PASS

Average      Lower      Upper  
17,601.04    8,800.52    26,401.56

Sample Name	Sample ID	Analysis Date	Analyte	Area	Lower	Upper	Qualifier	CCV Lower	CCV Upper	Qualifier
JV64	L1	5/17/18 9:20	d3-MeFOSAA	16,123.52	8,800.52	26,401.56		12,739.92	25,479.83	
JV65	L2	5/17/18 9:29	d3-MeFOSAA	17,769.57	8,800.52	26,401.56		12,739.92	25,479.83	
JV66	L3	5/17/18 9:38	d3-MeFOSAA	19,084.74	8,800.52	26,401.56		12,739.92	25,479.83	
JV67	L4	5/17/18 9:47	d3-MeFOSAA	17,771.51	8,800.52	26,401.56		12,739.92	25,479.83	
JV68	L5	5/17/18 9:56	d3-MeFOSAA	18,199.88	8,800.52	26,401.56		12,739.92	25,479.83	
JV69	L6	5/17/18 10:05	d3-MeFOSAA	17,045.59	8,800.52	26,401.56		12,739.92	25,479.83	
JV70	L7	5/17/18 10:14	d3-MeFOSAA	18,272.74	8,800.52	26,401.56		12,739.92	25,479.83	
JV71	L8	5/17/18 10:23	d3-MeFOSAA	15,851.69	8,800.52	26,401.56		12,739.92	25,479.83	
JV72	L9	5/17/18 10:31	d3-MeFOSAA	18,290.15	8,800.52	26,401.56		12,739.92	25,479.83	
JV63 ICC	ICC	5/17/18 10:40	d3-MeFOSAA	18,230.15	8,800.52	26,401.56		12,739.92	25,479.83	
CQ774PB-FS(0)	Procedural Blank	5/17/18 14:35	d3-MeFOSAA	16,567.01	8,800.52	26,401.56		12,739.92	25,479.83	
CQ775LCS-FS(0)	Laboratory Control Sample	5/17/18 14:44	d3-MeFOSAA	13,002.05	8,800.52	26,401.56		12,739.92	25,479.83	
J6149-FS(0)	NAWC-050718-FRB-316	5/17/18 14:53	d3-MeFOSAA	15,391.10	8,800.52	26,401.56		12,739.92	25,479.83	
J6151-FS(0)	NAWC-050718-FRB-180	5/17/18 15:02	d3-MeFOSAA	13,679.60	8,800.52	26,401.56		12,739.92	25,479.83	
J6153-FS(0)	NAWC-050718-FRB-275	5/17/18 15:11	d3-MeFOSAA	14,186.77	8,800.52	26,401.56		12,739.92	25,479.83	
J6155-FS(0)	NAWC-050718-FRB-145	5/17/18 15:20	d3-MeFOSAA	13,091.39	8,800.52	26,401.56		12,739.92	25,479.83	
J6157-FS(0)	NAWC-050718-FRB-357	5/17/18 15:29	d3-MeFOSAA	14,843.43	8,800.52	26,401.56		12,739.92	25,479.83	
J6159-FS(0)	NAWC-050718-FRB-162	5/17/18 15:38	d3-MeFOSAA	17,200.62	8,800.52	26,401.56		12,739.92	25,479.83	
JV70 CCV	CCV	5/17/18 15:47	d3-MeFOSAA	17,173.89	8,800.52	26,401.56		12,739.92	25,479.83	
J6161-FS(0)	WGNA-050718-FRB-0800	5/17/18 16:04	d3-MeFOSAA	17,311.56	8,800.52	26,401.56		12,021.72	24,043.45	
J6163-FS(0)	WGNA-050718-FRB-0335	5/17/18 16:13	d3-MeFOSAA	16,167.71	8,800.52	26,401.56		12,021.72	24,043.45	
J6165-FS(0)	WGNA-050718-FRB-3556	5/17/18 16:22	d3-MeFOSAA	12,300.76	8,800.52	26,401.56		12,021.72	24,043.45	
J6167-FS(0)	NAWC-050718-FRB-356	5/17/18 16:31	d3-MeFOSAA	17,743.88	8,800.52	26,401.56		12,021.72	24,043.45	
J6169-FS(0)	NAWC-050718-FRB-289	5/17/18 16:40	d3-MeFOSAA	13,101.61	8,800.52	26,401.56		12,021.72	24,043.45	
JV69 CCV	CCV	5/17/18 16:49	d3-MeFOSAA	17,006.92	8,800.52	26,401.56		12,021.72	24,043.45	

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/17/2018 10:14:08 AM	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Asymmetry Factor	Passing Range
PFBS_1	298.9 / 80.0	1.44	1.49	0.8 – 1.5
PFHxA_1	313.0 / 269.0	1.71	1.34	0.8 – 1.5

Mass Spectral Acquisition Rate Report

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/17/2018 10:14:08 AM	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
PFBS_1	298.9 / 80.0	1.44	40	>10
PFBS_2	298.9 / 99.0	1.43	74	>10
PFHxA_1	313.0 / 269.0	1.71	34	>10
PFHxA_2	313.0 / 119.0	1.71	52	>10
PFHpA_1	363.0 / 319.0	2.06	25	>10
PFHpA_2	363.0 / 169.0	2.06	28	>10
PFHxS_1	399.0 / 80.0	2.07	59	>10
PFHxS_2	399.0 / 99.0	2.07	41	>10
PFOA_1	413.0 / 369.0	2.43	33	>10
PFOA_2	413.0 / 169.0	2.43	34	>10
PFNA_1	463.0 / 419.0	2.80	30	>10
PFNA_2	463.0 / 219.0	2.80	30	>10
PFOS_1	499.0 / 80.0	2.80	59	>10
PFOS_2	499.0 / 99.0	2.80	51	>10
PFDA_1	513.0 / 469.0	3.15	56	>10
PFDA_2	513.0 / 219.0	3.15	38	>10
PFUnA_1	563.0 / 519.0	3.47	58	>10
PFUnA_2	563.0 / 269.0	3.47	28	>10
PFDaA_1	613.0 / 569.0	3.76	52	>10
PFDaA_2	613.0 / 319.0	3.76	36	>10
PFTrDA_1	663.0 / 619.0	4.01	59	>10
PFTrDA_2	663.0 / 169.0	4.01	44	>10
PFTeDA_1	713.0 / 669.0	4.23	62	>10
PFTeDA_2	713.0 / 169.0	4.23	57	>10
NMeFOSAA_1	570.0 / 419.0	3.30	48	>10
NMeFOSAA_2	570.0 / 512.0	3.30	52	>10
NEtFOSAA_1	584.0 / 419.0	3.46	37	>10
NEtFOSAA_2	584.0 / 483.0	3.46	40	>10

Sample Name	JV70	Injection Vial	8
Sample ID	L7	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	5/17/2018 10:14:08 AM	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Spectra Acquisition Rate	Passing Range
13C2-PFHxA	315.0 / 270.0	1.69	47	>10
13C2-PFDA	515.0 / 470.0	3.14	38	>10
d5-EtFOSAA	589.0 / 419.0	3.46	34	>10

# BATTELLE DETECTION LIMITS FOR PFAS IN DRINKING WATER

Battelle SOP 5-371 (EPA Method 537 Version 1.1)

Analyte	CAS No.	MDL (ng/L)	LOD (ng/L)	LOQ (ng/L)	MRL (ng/L)
<b>PFHxA</b>	307-24-4	0.22	0.5	2.5	2.5
<b>PFHpA</b>	375-85-9	0.34	1.0	2.5	2.5
<b>PFOA</b>	335-67-1	0.38	1.0	2.5	2.5
<b>PFNA</b>	375-95-1	0.37	1.0	2.5	2.5
<b>PFDA</b>	335-76-2	0.39	1.0	2.5	2.5
<b>PFUnA</b>	2058-94-8	0.38	1.0	2.5	2.5
<b>PFDoA</b>	307-55-1	0.42	1.0	2.5	2.5
<b>PFTrDA</b>	72629-94-8	0.42	1.0	2.5	2.5
<b>PFTeDA</b>	376-06-7	0.73	1.5	2.5	2.5
<b>NMeFOSAA</b>	2355-31-9	0.42	1.0	2.5	2.5
<b>NEtFOSAA</b>	2991-50-6	0.44	1.0	2.5	2.5
<b>PFBS</b>	375-73-5	0.21	0.5	2.5	2.5
<b>PFHxS</b>	3871-99-6	0.34	1.0	2.5	2.5
<b>PFOS</b>	1763-23-1	0.30	1.0	2.5	2.5

*Analytes on NELAP and ELAP QSM 5.1 Scope of accreditation*

Analytical Transitions for PFAS in drinking water

SOP 5-371 (EPA 537 Version 1.1)

Analyte	CAS No.	Type	Primary Transition	Secondary Transition
<b>PFHxA</b>	307-24-4	Target	313.0 / 269.0	313.0 / 119.0
<b>PFHpA</b>	375-85-9	Target	363.0 / 319.0	363.0 / 169.0
<b>PFOA</b>	335-67-1	Target	413.0 / 369.0	413.0 / 169.0
<b>PFNA</b>	375-95-1	Target	463.0 / 419.0	463.0 / 219.0
<b>PFDA</b>	335-76-2	Target	513.0 / 469.0	513.0 / 219.0
<b>PFUnA</b>	2058-94-8	Target	563.0 / 519.0	563.0 / 269.0
<b>PFDoA</b>	307-55-1	Target	613.0 / 569.0	613.0 / 319.0
<b>PFTTrDA</b>	72629-94-8	Target	663.0 / 619.0	663.0 / 169.0
<b>PFTeDA</b>	376-06-7	Target	713.0 / 669.0	713.0 / 169.0
<b>NMeFOSAA</b>	2355-31-9	Target	570.0 / 419.0	570.0 / 512.0
<b>NEtFOSAA</b>	2991-50-6	Target	584.0 / 419.0	584.0 / 483.0
<b>PFBS</b>	375-73-5	Target	298.9.0 / 80.0	298.9.0 / 99.0
<b>PFHxS</b>	355-46-4	Target	399.0 / 80.0	399.0 / 99.0
<b>PFOS</b>	1763-23-1	Target	499.0 / 80.0	499.0 / 99.0
<b><sup>13</sup>C<sub>2</sub>-PFHxA</b>	NA	SIS	315.0 / 270.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFDA</b>	NA	SIS	515.0 / 470.0	NA
<b>d<sub>5</sub>-EtFOSAA</b>	NA	SIS	589.0 / 419.0	NA
<b><sup>13</sup>C<sub>2</sub>-PFOA</b>	NA	IS	415.0 / 270.0	NA
<b><sup>13</sup>C<sub>4</sub>-PFOS</b>	NA	IS	503.0 / 80.0	NA
<b>d<sub>3</sub>-MeFOSAA</b>	NA	IS	573.0 / 419.0	NA





**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

# QTRAP 5500

**LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Mass Calibration and Tune Check

## QTRAP 5500 Preventive Maintenance Checklist

<b>Preventive Maintenance Date:</b>	22-Feb-2017
<b>Request ID:</b>	3683
<b>Company Name:</b>	Battelle Memorial Institute
<b>Instrument ID:</b>	X60666
<b>Instrument Model:</b>	QTRAP 5500
<b>Instrument Serial Number:</b>	AU23051004

**PASS**       **FAIL**

**Any failure will lead to an automatic Service Call being open to investigate fault.**

Preventive Maintenance is performed twice every year unless specified in the Service Contract. It is designed to help maintain optimum system performance and to help diagnose any system deficiencies.

Engineer is required the assigned Request ID for this PM otherwise making this job invalid.

**Comments:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Performed By:**           Kaustubh Dhayagude                **Date:**           22-Feb-2017          

**Approved By :** \_\_\_\_\_      **Date:** \_\_\_\_\_

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**PRE PM PPG PERFORMANCE EVALUATION:**

- Consult Customer concerning the unit overall performance.
- Check Logbook for Services recently performed.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.5	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	1.9	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.4	Read Only
<input checked="" type="checkbox"/> CAD High	3.4	Read Only
<input checked="" type="checkbox"/> CAD 12	3.4	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Check for Front end contamination symptoms. Run Q1 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop
- Check for Q3 contamination symptoms. Run Q3 POS PPG using PPG 2e-7for a few minutes and check for any TIC signal degradation or huge sensitivity drop where the sensitivity result can't pass specification
  - No degradation or Sensitivity drop

**Pre PM PPG Test:** Perform each of the following tests. Optimize ion source position only. The specifications listed for these Pre PM tests are guidelines only, not required to be met.

- Perform Q1 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 175.133	1.64 e6	Read Only	0.8095	Read Only
Q1 500.380	2.40 e7	Read Only	0.8592	Read Only
Q1 906.673	2.86 e7	Read Only	0.9633	Read Only

- Perform Q3 POS using POS PPG 2e-7M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 175.133	1.26 e6	Read Only	0.6252	Read Only
Q3 500.380	2.19 e7	Read Only	0.7275	Read Only
Q3 906.673	3.02 e7	Read Only	0.7662	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform MSMS POS in Product Ion scan with 609.3 parent and record daughter 195.1 using Reserpine 0.167 pmol/ul at the scan rate of 10 Da/s for 10 MCA. Calculate transmission efficiency comparing Q1POS 609 intensity. Transmission Efficiency: : 19.51% (Read Only)

Mass	MSMS Intensity		MSMS Width Value	Width Specs
	Value	Spec		
Q1 609.3	7.43 e7	Read Only	0.9981	Read Only
MS/MS 195.1	1.45 e7	Read Only	0.6582	Read Only

Perform Q1 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Q1 933.636	1.43 e7	Read Only	0.7330	Read Only

Perform Q3 NEG using NEG PPG 3e-5M. Scan Rate 10 Da/s. Record 10 mca.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Q3 933.636	2.22 e7	Read Only	0.8138	Read Only

Perform Product Ion scan using NEG PPG 3e-5M. Record 10 mca.

Mass	Scan Rate	MCA	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.35 e6	Read Only	0.6495	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**PREVENTIVE MAINTENANCE CHECKLIST:**

- Check Cooling Fans for Turbo Pumps while MS is ON.
- Check QJet and QPS tuning voltage for reference.
- Record AC input Voltage while MS is OFF: \_\_\_\_\_ (200-240VAC).  
If Out-of-Range, notify customer.
- Clean Interface
- Curtain Plate
  - Orifice Plate
  - QJet
  - Q0 Rods.
- Replace Roughing Pump Oil.
- Inspect Oil Exhaust Filter, if Applicable.  N/A
- Clean and inspect built-in divert valve if used.  N/A
- Check Multiplier Voltage, optimize if necessary.
- Replace four Air Filters at the bottom of the mass spectrometer.
- Pump down overnight if possible.  N/A
- Perform Maintenance on Turbo V source.
- Replace Electrode, if necessary.  N/A
- Check Turbo heaters resistances.
- Check if Temperature is reached at 500C with TIS Probe installed.
- Check if Temperature is reached at 500C with APCI Probe installed.  N/A

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

**POST PM PPG PERFORMANCE TESTS:**

- Set-up Sample for Infusion.
- Check spray and adjust sprayer's position of the TIS source.
- Check Vacuum Pressure:

CAD Settings	Vacuum Reading ( x 10 <sup>-5</sup> Torr)	Acceptance Criteria
<input checked="" type="checkbox"/> CAD 0	0.8	0.4 to 1.1 x10 <sup>-5</sup> Torr
<input checked="" type="checkbox"/> CAD Low	2.1	Read Only
<input checked="" type="checkbox"/> CAD Medium	2.6	Read Only
<input checked="" type="checkbox"/> CAD High	3.7	Read Only
<input checked="" type="checkbox"/> CAD 12	3.7	2.4 to 4.5 x10 <sup>-5</sup> Torr

- Perform Q1 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q1 Intensity		Q1 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q1 175.133	5.94 e6	≥1.2 <sup>e6</sup>	0.6933	0.6 to 0.8
Q1 500.380	2.25 e7	≥9.0 <sup>e6</sup>	0.7444	0.6 to 0.8
Q1 906.673	2.74 e7	≥1.4 <sup>e7</sup>	0.7347	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q1 906.673	1.33 e8	≥6.8 <sup>e7</sup>	0.7656	0.6 to 0.8

- Perform Q3 POS using POS PPG 2e-7M. Mass calibrate to less than 0.1 amu.

Mass	Q3 Intensity		Q3 Width Value	Width Specs
	Value	Spec		
Scan Rate 10 Da/s Record 10 mca				
Q3 175.133	4.54 e6	≥1.2 <sup>e6</sup>	0.6390	0.6 to 0.8
Q3 500.380	2.13 e7	≥9.0 <sup>e6</sup>	0.7008	0.6 to 0.8
Q3 906.673	3.04 e7	≥1.4 <sup>e7</sup>	0.7683	0.6 to 0.8
Scan Rate 1000 Da/s Record 50 mca				
Q3 906.673	1.51 e8	≥6.8 <sup>e7</sup>	0.7118	0.6 to 0.8

- Perform "Product of 609.3" POS and record product ion 195.1 using Reserpine 0.167pmol/uL. Record 10 mca. Calculate Transmission efficiency comparing Q1POS 609 intensity.

Transmission Efficiency: 16.93% (≥ 10.0%)

Mass	MSMS Intensity		Width Value	Width Specs
	Value	Spec		
Q1 609.3	5.74 e7	N/A	0.7667	Read Only
MS/MS 195.1	9.72 e6	N/A	0.6751	Read Only

**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

Perform Q1 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q1 Intensity		Q1 Width Value	Width Specs
			Value	Spec		
Q1 933.636	10	10	1.31 e7	$\geq 1.0^{e7}$	0.6895	0.6 to 0.8
Q1 933.636	1000	50	6.32 e7	$\geq 4.0^{e7}$	0.6740	0.6 to 0.8

Perform Q3 NEG using NEG PPG 3e-5M. Mass calibrate to less than 0.1 amu.

Mass	Scan Rate	Mca	Q3 Intensity		Q3 Width Value	Width Specs
			Value	Spec		
Q3 933.636	10	10	1.70 e7	$\geq 8.0^{e6}$	0.7665	0.6 to 0.8
Q3 933.636	1000	50	7.41 e7	$\geq 4.0^{e7}$	0.7292	0.6 to 0.8

Perform Product Ion scan using NEG PPG 3e-5M.

Mass	Scan Rate	Mca	MSMS Intensity		MSMS Width Value	Width Specs
			Value	Spec		
MSMS 45	10	10	3.33 e6	Read Only	0.6387	Read Only

Perform ER POS 118.087 and 922.01 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 118.087	0.05	8.08 e6	$\geq 7.2^{e6}$	0.1302	<0.35
ER 922.010	0.05	3.89 e7	$\geq 2.8^{e6}$	0.2603	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 118.087	0.05	2.55 e7	$\geq 2.4^{e7}$	0.3740	<0.65
ER 922.010	0.05	2.37 e8	$\geq 6.8^{e7}$	0.5407	<0.65

Perform ER NEG 431.982 and 601.978 using ESI Tuning Mix 1:100 in ES Tuning Dilution Solvent. Apply suggested Scan Rate and Record number of MCA. Mass calibrate to less than 0.1 amu.

Mass	Fill Time (ms)	ER Intensity		ER Width Value	Width Specs
		Value	Spec		
ScanRate : 1000 Da/s ; 50 Mca					
ER 431.982	0.05	1.05 e8	$\geq 4.4^{e7}$	0.1840	<0.35
ER 601.978	0.05	7.74 e7	$\geq 5.6^{e7}$	0.1849	<0.35
ScanRate : 10000 Da/s ; 50 Mca					
ER 431.982	0.05	3.43 e8	$\geq 1.2^{e8}$	0.4382	<0.65
ER 601.978	0.05	2.55 e8	$\geq 1.6^{e8}$	0.6205	<0.65



**Zef Scientific Inc.**

12707 High Bluff Dr.  
Suite 200  
San Diego, CA  
USA 92130

1975 Hymus Blvd.  
Suite 230  
Dorval, QC  
Canada H9P 1J8

Phone: 1.866.854.7988

**QTRAP 5500****LC/MS/MS Detector System**

Appendix ZEFPM003-2L

- Perform EPI POS 397.2 using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Q0 Trapping OFF		Q0 Trapping ON	
		Intensity	Spec	Intensity	Spec
EPI 397.2	10000	> 3.5 e6	≥2.0 e6	> 4.0 e7	≥6.4 e6

- Perform MS3 POS full scan Fragmentation ON & OFF using Reserpine 0.167pmol/uL. Record 20 mca.

Mass	Scan Rate (Da/s)	Fragamentation OFF		Fragmentation ON	
		Intensity	Spec	Intensity	Spec
MS3 397.2	1000	3.2 e7	Contains only 397.2	N/A	N/A
<input type="checkbox"/> 236 OR <input checked="" type="checkbox"/> 365	1000	1.19 e8	Fragment Intensity	> 4.4 e6	≥1.6x 10 <sup>e6</sup>

**REVIEW:**

- Attach all spectrums printouts to this procedure.
- If any parameter setting access modes were changed during the PM, ensure they are returned to their normal access mode and that their offsets are adjusted to match optimized values from the post-PM acquisition files.
- Empty tuning cache folder, if necessary.  N/A
- Update Service Work Order status
- Fill and replace PM Label.

**END OF PREVENTIVE MAINTENANCE CHECKLIST****Document history:**

06 OCT 2016: Appendix ZEFPM003-2L: Removed requirements to fit Manufacturer's testing criteria.



It can be done

## BATTELLE - NORWELL OPERATIONS SAMPLE PREPARATION RECORDS

<b><u>Project Title(s)</u></b>	<b><u>Project No.(s)</u></b>
Naval Air Station Joint Reserve Base Willow Grove, PA	100117920-WE04
<b>18-0316</b>	
<b>WE04 PFAS Analysis</b>	
<b>DW</b>	
SOP Numbers (see workplan for modifications)	
VOASOP No.	5-371

This Batch Contains The Following Samples:		
CQ774PB-FS	J6157-FS	J6169-FS
CQ775LCS-FS	J6159-FS	
J6149-FS	J6161-FS	
J6151-FS	J6163-FS	
J6153-FS	J6165-FS	
J6155-FS	J6167-FS	

Laboratory Preparation Records  
COMPLETE AND VALIDATED

Prep Task Leader: Stephanie Schultz

Approved By:	Date	Initials
Denise Schumitz	05/21/2018	DMS



It can be done

## BATTELLE - NORWELL OPERATIONS LIQUID SAMPLE ID FORM

**Project Title(s)**

Naval Air Station Joint Reserve Base Willow Grove, PA

**Project No.(s)**100117920-  
WE04**18-0316****WE04 PFAS Analysis****DW**

Sample ID	Description	Volume (mL)	Bottles	*	Date Initials
CQ774PB-FS	Procedural Blank	250.0	NA	--	05/15/18 SAS
CQ775LCS-FS	Laboratory Control Sample	250.0	NA	--	05/15/18 SAS
J6149-FS	NAWC-050718-FRB-316	255.0	1	C	05/16/18 SAS
J6151-FS	NAWC-050718-FRB-180	250.0	1	C	05/16/18 SAS
J6153-FS	NAWC-050718-FRB-275	250.0	1	C	05/16/18 SAS
J6155-FS	NAWC-050718-FRB-145	250.0	1	C	05/16/18 SAS
J6157-FS	NAWC-050718-FRB-357	250.0	1	C	05/16/18 SAS
J6159-FS	NAWC-050718-FRB-162	250.0	1	C	05/16/18 SAS
J6161-FS	WGNA-050718-FRB-0800	255.0	1	C	05/16/18 SAS
J6163-FS	WGNA-050718-FRB-0335	250.0	1	C	05/16/18 SAS
J6165-FS	WGNA-050718-FRB-3556	250.0	1	C	05/16/18 SAS
J6167-FS	NAWC-050718-FRB-356	250.0	1	C	05/16/18 SAS
J6169-FS	NAWC-050718-FRB-289	250.0	1	C	05/16/18 SAS

**Comments:**

Sample ID:	Comments:
CQ774PB-FS	1.23g Trizma(170526-01) weighed on BAL-009
CQ775LCS-FS	1.27g Trizma(170526-01) weighed on BAL-009

Samples Assigned By

Stephanie Schultz

Date :

May 15, 2018

\* - "C" = Sample is Consumed


Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
1	MeOH		5/17/2018 9:11:38 AM	5-0371.dam	18-0287.wiff
2	JV64	L1	5/17/2018 9:20:34 AM	5-0371.dam	18-0287.wiff
3	JV65	L2	5/17/2018 9:29:29 AM	5-0371.dam	18-0287.wiff
4	JV66	L3	5/17/2018 9:38:24 AM	5-0371.dam	18-0287.wiff
5	JV67	L4	5/17/2018 9:47:21 AM	5-0371.dam	18-0287.wiff
6	JV68	L5	5/17/2018 9:56:17 AM	5-0371.dam	18-0287.wiff
7	JV69	L6	5/17/2018 10:05:12 AM	5-0371.dam	18-0287.wiff
8	JV70	L7	5/17/2018 10:14:08 AM	5-0371.dam	18-0287.wiff
9	JV71	L8	5/17/2018 10:23:03 AM	5-0371.dam	18-0287.wiff
10	JV72	L9	5/17/2018 10:31:58 AM	5-0371.dam	18-0287.wiff
11	JV63 ICC	ICC	5/17/2018 10:40:54 AM	5-0371.dam	18-0287.wiff
1	MeOH		5/17/2018 10:49:49 AM	5-0371.dam	18-0287.wiff
15	CQ774PB-FS(0)	Procedural Blank	5/17/2018 2:35:34 PM	5-0371.dam	18-0315.wiff
16	CQ775LCS-FS(0)	Laboratory Control Sample	5/17/2018 2:44:30 PM	5-0371.dam	18-0315.wiff
17	J6149-FS(0)	NAWC-050718-FRB-316	5/17/2018 2:53:26 PM	5-0371.dam	18-0315.wiff
18	J6151-FS(0)	NAWC-050718-FRB-180	5/17/2018 3:02:22 PM	5-0371.dam	18-0315.wiff
19	J6153-FS(0)	NAWC-050718-FRB-275	5/17/2018 3:11:18 PM	5-0371.dam	18-0315.wiff
20	J6155-FS(0)	NAWC-050718-FRB-145	5/17/2018 3:20:15 PM	5-0371.dam	18-0315.wiff
21	J6157-FS(0)	NAWC-050718-FRB-357	5/17/2018 3:29:12 PM	5-0371.dam	18-0315.wiff
22	J6159-FS(0)	NAWC-050718-FRB-162	5/17/2018 3:38:10 PM	5-0371.dam	18-0315.wiff
8	JV70 CCV	CCV	5/17/2018 3:47:06 PM	5-0371.dam	18-0315.wiff
1	MeOH		5/17/2018 3:56:03 PM	5-0371.dam	18-0315.wiff
23	J6161-FS(0)	WGNA-050718-FRB-0800	5/17/2018 4:04:58 PM	5-0371.dam	18-0315.wiff
24	J6163-FS(0)	WGNA-050718-FRB-0335	5/17/2018 4:13:55 PM	5-0371.dam	18-0315.wiff
25	J6165-FS(0)	WGNA-050718-FRB-3556	5/17/2018 4:22:52 PM	5-0371.dam	18-0315.wiff
26	J6167-FS(0)	NAWC-050718-FRB-356	5/17/2018 4:31:47 PM	5-0371.dam	18-0315.wiff
27	J6169-FS(0)	NAWC-050718-FRB-	5/17/2018 4:40:42	5-0371.dam	18-0315.wiff

Vial	Laboratory Sample ID	Client Sample ID	Acquisition Date	Acquisition Method	Data File
		289	PM		
7	JV69 CCV	CCV	5/17/2018 4:49:36 PM	5-0371.dam	18-0315.wiff

<b>Analyte Name</b>	PFBS_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	298.9 / 80.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.76645x + -8.51923e-4$  (r = 0.99924) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	22.15	23.483107	106.0
3	JV65	L2	True	44.30	38.690092	87.3
4	JV66	L3	True	88.60	77.119919	87.0
5	JV67	L4	True	221.50	270.479947	122.1
6	JV68	L5	True	443.00	494.830183	111.7
7	JV69	L6	True	885.00	836.047360	94.5
8	JV70	L7	True	2212.50	2161.273577	97.7
9	JV71	L8	True	4425.00	4377.980502	98.9
10	JV72	L9	True	8850.00	8913.478420	100.7






<b>Analyte Name</b>	PFBS_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	298.9 / 99.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.24667 x + 6.62243e-4$  (r = 0.99862) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.15	18.520807	83.6
3	JV65	L2	True	44.30	49.623047	112.0
4	JV66	L3	True	88.60	77.085462	87.0
5	JV67	L4	True	221.50	198.376185	89.6
6	JV68	L5	True	443.00	551.358297	124.5
7	JV69	L6	True	885.00	888.633944	100.4
8	JV70	L7	True	2212.50	2357.741880	106.6
9	JV71	L8	True	4425.00	4327.956354	97.8
10	JV72	L9	True	8850.00	8722.754023	98.6



<b>Analyte Name</b>	PFHxA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	313.0 / 269.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.94368x + 0.09434$  (r = 0.99980) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.353288	101.4
3	JV65	L2	True	50.00	48.230678	96.5
4	JV66	L3	True	100.00	92.729568	92.7
5	JV67	L4	True	250.00	242.616322	97.1
6	JV68	L5	True	500.00	547.608857	109.5
7	JV69	L6	True	1000.00	1021.164999	102.1
8	JV70	L7	True	2500.00	2539.791620	101.6
9	JV71	L8	True	5000.00	5004.441317	100.1
10	JV72	L9	True	10000.00	9903.063352	99.0



<b>Analyte Name</b>	PFHxA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	313.0 / 119.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06557 x + -0.01141$  (r = 0.99883) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	32.175804	128.7
3	JV65	L2	True	50.00	62.207726	124.4
4	JV66	L3	True	100.00	71.281276	71.3
5	JV67	L4	True	250.00	187.631333	75.1
6	JV68	L5	True	500.00	501.920797	100.4
7	JV69	L6	True	1000.00	939.982193	94.0
8	JV70	L7	True	2500.00	2662.599478	106.5
9	JV71	L8	True	5000.00	4998.914969	100.0
10	JV72	L9	True	10000.00	9968.286425	99.7



<b>Analyte Name</b>	PFHpA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	363.0 / 319.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.86846 x + 0.01868$  (r = 0.99968) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	30.745121	123.0
3	JV65	L2	True	50.00	45.261637	90.5
4	JV66	L3	True	100.00	94.999992	95.0
5	JV67	L4	True	250.00	230.401025	92.2
6	JV68	L5	True	500.00	507.718530	101.5
7	JV69	L6	True	1000.00	958.469829	95.9
8	JV70	L7	True	2500.00	2585.103762	103.4
9	JV71	L8	True	5000.00	4881.799777	97.6
10	JV72	L9	True	10000.00	10090.500327	100.9



<b>Analyte Name</b>	PFHpA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	363.0 / 169.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.01781 x + 0.00179$  (r = 0.99935) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	28.647339	114.6
3	JV65	L2	True	50.00	44.765389	89.5
4	JV66	L3	True	100.00	84.456300	84.5
5	JV67	L4	True	250.00	230.303823	92.1
6	JV68	L5	True	500.00	549.021210	109.8
7	JV69	L6	True	1000.00	1101.666841	110.2
8	JV70	L7	True	2500.00	2552.091209	102.1
9	JV71	L8	True	5000.00	4890.698476	97.8
10	JV72	L9	True	10000.00	9943.349413	99.4



<b>Analyte Name</b>	PFHxS_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	399.0 / 80.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.81452 x + -0.00421$  (r = 0.99944) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	24.720628	108.4
3	JV65	L2	True	45.60	40.923316	89.7
4	JV66	L3	True	91.20	84.560191	92.7
5	JV67	L4	True	228.00	234.300477	102.8
6	JV68	L5	True	456.00	490.468001	107.6
7	JV69	L6	True	912.00	867.632050	95.1
8	JV70	L7	True	2280.00	2426.969740	106.5
9	JV71	L8	True	4560.00	4439.462493	97.4
10	JV72	L9	True	9120.00	9106.563104	99.9






<b>Analyte Name</b>	PFHxS_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	399.0 / 99.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.24236 x + 0.00960$  (r = 0.99962) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	22.80	25.925035	113.7
3	JV65	L2	True	45.60	42.715296	93.7
4	JV66	L3	True	91.20	79.822448	87.5
5	JV67	L4	True	228.00	232.286071	101.9
6	JV68	L5	True	456.00	484.419318	106.2
7	JV69	L6	True	912.00	887.969901	97.4
8	JV70	L7	True	2280.00	2321.581504	101.8
9	JV71	L8	True	4560.00	4397.947134	96.5
10	JV72	L9	True	9120.00	9242.933292	101.4



<b>Analyte Name</b>	PFOA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	413.0 / 369.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.03975x + 0.07616$  (r = 0.99957) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	26.708862	106.8
3	JV65	L2	True	50.00	42.899544	85.8
4	JV66	L3	True	100.00	94.094405	94.1
5	JV67	L4	True	250.00	251.494624	100.6
6	JV68	L5	True	500.00	525.460651	105.1
7	JV69	L6	True	1000.00	1039.091021	103.9
8	JV70	L7	True	2500.00	2636.449272	105.5
9	JV71	L8	True	5000.00	5012.599247	100.3
10	JV72	L9	True	10000.00	9796.202374	98.0



<b>Analyte Name</b>	PFOA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	413.0 / 169.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.06919x + -6.10299e-4$  (r = 0.99963) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	30.244231	121.0
3	JV65	L2	True	50.00	39.807292	79.6
4	JV66	L3	True	100.00	91.316603	91.3
5	JV67	L4	True	250.00	253.258834	101.3
6	JV68	L5	True	500.00	501.014564	100.2
7	JV69	L6	True	1000.00	1073.277778	107.3
8	JV70	L7	True	2500.00	2461.920179	98.5
9	JV71	L8	True	5000.00	5103.925253	102.1
10	JV72	L9	True	10000.00	9870.235267	98.7



<b>Analyte Name</b>	PFNA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	463.0 / 419.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.05388x + 0.02887$  (r = 0.99958) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	22.448588	89.8
3	JV65	L2	True	50.00	45.828516	91.7
4	JV66	L3	True	100.00	101.744372	101.7
5	JV67	L4	True	250.00	252.109959	100.8
6	JV68	L5	True	500.00	548.360364	109.7
7	JV69	L6	True	1000.00	1039.938157	104.0
8	JV70	L7	True	2500.00	2614.669594	104.6
9	JV71	L8	True	5000.00	4970.858264	99.4
10	JV72	L9	True	10000.00	9829.042185	98.3



<b>Analyte Name</b>	PFNA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	463.0 / 219.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.31494 x + 0.00205$  (r = 0.99965) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	17.606402	70.4
3	JV65	L2	True	50.00	55.585100	111.2
4	JV66	L3	True	100.00	100.725248	100.7
5	JV67	L4	True	250.00	263.453692	105.4
6	JV68	L5	True	500.00	532.054934	106.4
7	JV69	L6	True	1000.00	1044.980309	104.5
8	JV70	L7	True	2500.00	2573.326453	102.9
9	JV71	L8	True	5000.00	5008.271225	100.2
10	JV72	L9	True	10000.00	9828.996638	98.3



<b>Analyte Name</b>	PFOS_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	499.0 / 80.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.16999x + 0.01017$  (r = 0.99966) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	23.088724	99.7
3	JV65	L2	True	46.30	42.127787	91.0
4	JV66	L3	True	92.60	87.472091	94.5
5	JV67	L4	True	231.50	234.830537	101.4
6	JV68	L5	True	463.00	526.238813	113.7
7	JV69	L6	True	925.60	911.789563	98.5
8	JV70	L7	True	2314.00	2367.443455	102.3
9	JV71	L8	True	4628.00	4622.929105	99.9
10	JV72	L9	True	9256.00	9164.229925	99.0






<b>Analyte Name</b>	PFOS_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	499.0 / 99.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C4-PFOS	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.22845x + 0.00142$  (r = 0.99952) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	23.15	19.073086	82.4
3	JV65	L2	True	46.30	42.754548	92.3
4	JV66	L3	True	92.60	95.652663	103.3
5	JV67	L4	True	231.50	246.896117	106.7
6	JV68	L5	True	463.00	534.354893	115.4
7	JV69	L6	True	925.60	920.517935	99.5
8	JV70	L7	True	2314.00	2338.375483	101.1
9	JV71	L8	True	4628.00	4674.443945	101.0
10	JV72	L9	True	9256.00	9108.081330	98.4



<b>Analyte Name</b>	PFDA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	513.0 / 469.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.15138x + 0.01830$  (r = 0.99962) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.780783	103.1
3	JV65	L2	True	50.00	47.741972	95.5
4	JV66	L3	True	100.00	97.400854	97.4
5	JV67	L4	True	250.00	240.150008	96.1
6	JV68	L5	True	500.00	512.308119	102.5
7	JV69	L6	True	1000.00	1009.776029	101.0
8	JV70	L7	True	2500.00	2663.600645	106.5
9	JV71	L8	True	5000.00	4966.639934	99.3
10	JV72	L9	True	10000.00	9861.601655	98.6



<b>Analyte Name</b>	PFDA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	513.0 / 219.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.04195x + 0.02344$  (r = 0.99882) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	< 0	N/A
3	JV65	L2	False	50.00	< 0	N/A
4	JV66	L3	True	100.00	120.282787	120.3
5	JV67	L4	True	250.00	181.707950	72.7
6	JV68	L5	True	500.00	500.313071	100.1
7	JV69	L6	True	1000.00	1031.439826	103.1
8	JV70	L7	True	2500.00	2669.043842	106.8
9	JV71	L8	True	5000.00	4859.355694	97.2
10	JV72	L9	True	10000.00	9987.856830	99.9



<b>Analyte Name</b>	PFUnA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	563.0 / 519.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.20316 x + 0.01632$  (r = 0.99933) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	21.476403	85.9
3	JV65	L2	True	50.00	46.942062	93.9
4	JV66	L3	True	100.00	105.088967	105.1
5	JV67	L4	True	250.00	259.283618	103.7
6	JV68	L5	True	500.00	528.597710	105.7
7	JV69	L6	True	1000.00	1005.384599	100.5
8	JV70	L7	True	2500.00	2708.602972	108.3
9	JV71	L8	True	5000.00	4930.949311	98.6
10	JV72	L9	True	10000.00	9818.674359	98.2



<b>Analyte Name</b>	PFUnA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	563.0 / 269.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.05105 x + 0.00264$  (r = 0.99930) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	18.744641	75.0
3	JV65	L2	True	50.00	49.810694	99.6
4	JV66	L3	True	100.00	112.965030	113.0
5	JV67	L4	True	250.00	264.209742	105.7
6	JV68	L5	True	500.00	544.666363	108.9
7	JV69	L6	True	1000.00	929.274388	92.9
8	JV70	L7	True	2500.00	2660.578162	106.4
9	JV71	L8	True	5000.00	5001.977143	100.0
10	JV72	L9	True	10000.00	9842.773836	98.4



<b>Analyte Name</b>	PFD <sub>o</sub> A_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	613.0 / 569.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.21825x + 0.01097$  (r = 0.99971) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	22.899081	91.6
3	JV65	L2	True	50.00	49.230961	98.5
4	JV66	L3	True	100.00	100.630285	100.6
5	JV67	L4	True	250.00	251.821712	100.7
6	JV68	L5	True	500.00	543.205579	108.6
7	JV69	L6	True	1000.00	993.526410	99.4
8	JV70	L7	True	2500.00	2578.802928	103.2
9	JV71	L8	True	5000.00	4858.808287	97.2
10	JV72	L9	True	10000.00	10026.074758	100.3






<b>Analyte Name</b>	PFD <sub>o</sub> A_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	613.0 / 319.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.18668x + -5.27718e-4$  (r = 0.99978) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	21.198571	84.8
3	JV65	L2	True	50.00	48.544516	97.1
4	JV66	L3	True	100.00	104.886111	104.9
5	JV67	L4	True	250.00	268.108635	107.2
6	JV68	L5	True	500.00	540.076896	108.0
7	JV69	L6	True	1000.00	977.473712	97.8
8	JV70	L7	True	2500.00	2546.512918	101.9
9	JV71	L8	True	5000.00	4918.186485	98.4
10	JV72	L9	True	10000.00	10000.012156	100.0



<b>Analyte Name</b>	PFTrDA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	663.0 / 619.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.14025x + -0.00563$  (r = 0.99975) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	23.303615	93.2
3	JV65	L2	True	50.00	49.441089	98.9
4	JV66	L3	True	100.00	101.743185	101.7
5	JV67	L4	True	250.00	250.416518	100.2
6	JV68	L5	True	500.00	536.739235	107.4
7	JV69	L6	True	1000.00	961.587378	96.2
8	JV70	L7	True	2500.00	2599.762848	104.0
9	JV71	L8	True	5000.00	4947.640990	99.0
10	JV72	L9	True	10000.00	9954.365142	99.5



<b>Analyte Name</b>	PFTTrDA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	663.0 / 169.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.07119x + 0.00179$  (r = 0.99985) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	28.219861	112.9
3	JV65	L2	True	50.00	44.534197	89.1
4	JV66	L3	True	100.00	104.781408	104.8
5	JV67	L4	True	250.00	242.644756	97.1
6	JV68	L5	True	500.00	505.553916	101.1
7	JV69	L6	True	1000.00	939.542406	94.0
8	JV70	L7	True	2500.00	2511.903429	100.5
9	JV71	L8	True	5000.00	5019.349063	100.4
10	JV72	L9	True	10000.00	10028.470965	100.3



<b>Analyte Name</b>	PFTeDA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	713.0 / 669.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.17996 x + 0.00995$  (r = 0.99970) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	24.735841	98.9
3	JV65	L2	True	50.00	47.529454	95.1
4	JV66	L3	True	100.00	103.219818	103.2
5	JV67	L4	True	250.00	248.573936	99.4
6	JV68	L5	True	500.00	512.080787	102.4
7	JV69	L6	True	1000.00	972.640447	97.3
8	JV70	L7	True	2500.00	2644.160657	105.8
9	JV71	L8	True	5000.00	4918.111573	98.4
10	JV72	L9	True	10000.00	9953.947488	99.5



<b>Analyte Name</b>	PFTeDA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	713.0 / 169.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	13C2-PFOA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.05522 x + -2.53563e-4$  (r = 0.99873) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.372335	101.5
3	JV65	L2	True	50.00	41.421057	82.8
4	JV66	L3	True	100.00	105.755853	105.8
5	JV67	L4	True	250.00	257.279859	102.9
6	JV68	L5	True	500.00	534.349917	106.9
7	JV69	L6	True	1000.00	939.252895	93.9
8	JV70	L7	True	2500.00	2771.350611	110.9
9	JV71	L8	True	5000.00	4784.927639	95.7
10	JV72	L9	True	10000.00	9965.289833	99.7



<b>Analyte Name</b>	NMeFOSAA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	570.0 / 419.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.06117x + -0.04402$  (r = 0.99943) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	False	25.00	33.804476	135.2
3	JV65	L2	False	50.00	75.638920	151.3
4	JV66	L3	True	100.00	96.476906	96.5
5	JV67	L4	True	250.00	223.748795	89.5
6	JV68	L5	True	500.00	543.743514	108.8
7	JV69	L6	True	1000.00	1052.710998	105.3
8	JV70	L7	True	2500.00	2582.869250	103.3
9	JV71	L8	True	5000.00	4818.449837	96.4
10	JV72	L9	True	10000.00	10032.000701	100.3






<b>Analyte Name</b>	NMeFOSAA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	570.0 / 512.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.63652 x + 2.66585e-4$  (r = 0.99944) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	20.981286	83.9
3	JV65	L2	True	50.00	51.560465	103.1
4	JV66	L3	True	100.00	111.874753	111.9
5	JV67	L4	True	250.00	230.145813	92.1
6	JV68	L5	True	500.00	563.501107	112.7
7	JV69	L6	True	1000.00	984.601410	98.5
8	JV70	L7	True	2500.00	2502.038421	100.1
9	JV71	L8	True	5000.00	4817.598080	96.4
10	JV72	L9	True	10000.00	10142.698664	101.4



<b>Analyte Name</b>	NEtFOSAA_1	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	584.0 / 419.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 1.02015x + -0.00244$  (r = 0.99960) (weighting: 1 / x)


Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	19.766835	79.1
3	JV65	L2	True	50.00	59.290846	118.6
4	JV66	L3	True	100.00	92.324906	92.3
5	JV67	L4	True	250.00	238.820153	95.5
6	JV68	L5	True	500.00	533.733349	106.8
7	JV69	L6	True	1000.00	1071.478343	107.2
8	JV70	L7	True	2500.00	2565.600521	102.6
9	JV71	L8	True	5000.00	4953.962734	99.1
10	JV72	L9	True	10000.00	9890.022314	98.9



<b>Analyte Name</b>	NEtFOSAA_2	<b>Data File</b>	18-0287.wiff
<b>MRM Transition</b>	584.0 / 483.0	<b>Result Table</b>	18-0316Base
<b>Internal Standard</b>	d3-MeFOSAA	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	5/17/2018 9:11:38 AM	<b>Acquisition Method</b>	5-0371.dam

Regression Equation:  $y = 0.07082 x + 4.14020e-5$  (r = 0.99757) (weighting: 1 / x)

Vial	Sample Name	Sample ID	Used for ICAL	Target Conc. (ng/L)	Calculated Conc. (ng/L)	Recovery (%)
2	JV64	L1	True	25.00	25.182661	100.7
3	JV65	L2	True	50.00	32.517054	65.0
4	JV66	L3	True	100.00	113.389105	113.4
5	JV67	L4	True	250.00	226.023527	90.4
6	JV68	L5	True	500.00	535.608828	107.1
7	JV69	L6	True	1000.00	1182.728888	118.3
8	JV70	L7	True	2500.00	2687.401954	107.5
9	JV71	L8	True	5000.00	5132.452168	102.7
10	JV72	L9	True	10000.00	9489.695816	94.9





<b>Sample Name</b>	JV64	<b>Injection Vial</b>	2
<b>Sample ID</b>	L1	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T09:20:34	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.46	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.268	0.327	ü
PFHxA_1	313.0 / 269.0	1.74	PFHxA			
PFHxA_2	313.0 / 119.0	1.73	PFHxA	0.029	0.056	ü
PFHpA_1	363.0 / 319.0	2.08	PFHpA			
PFHpA_2	363.0 / 169.0	2.08	PFHpA	0.024	0.022	ü
PFHxS_1	399.0 / 80.0	2.09	PFHxS			
PFHxS_2	399.0 / 99.0	2.09	PFHxS	0.478	0.335	ü
PFOA_1	413.0 / 369.0	2.45	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.057	0.062	ü
PFNA_1	463.0 / 419.0	2.81	PFNA			
PFNA_2	463.0 / 219.0	2.82	PFNA	0.217	0.294	ü
PFOS_1	499.0 / 80.0	2.81	PFOS			
PFOS_2	499.0 / 99.0	2.81	PFOS	0.159	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.19	PFDA	0.023	0.041	ü
PFUnA_1	563.0 / 519.0	3.48	PFUnA			
PFUnA_2	563.0 / 269.0	3.46	PFUnA	0.044	0.044	ü
PFDaA_1	613.0 / 569.0	3.77	PFDaA			
PFDaA_2	613.0 / 319.0	3.78	PFDaA	0.135	0.152	ü
PFTrDA_1	663.0 / 619.0	4.03	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.02	PFTrDA	0.084	0.064	ü
PFTeDA_1	713.0 / 669.0	4.25	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.25	PFTeDA	0.046	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.34	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.34	NMeFOSAA	0.737	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.48	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.48	NEtFOSSA	0.094	0.071	ü

<b>Sample Name</b>	JV65	<b>Injection Vial</b>	3
<b>Sample ID</b>	L2	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T09:29:29	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.45	PFBS	0.423	0.327	ü
PFHxA_1	313.0 / 269.0	1.72	PFHxA			
PFHxA_2	313.0 / 119.0	1.70	PFHxA	0.054	0.056	ü
PFHpA_1	363.0 / 319.0	2.07	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.024	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.08	PFHxS	0.408	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.43	PFOA	0.052	0.062	ü
PFNA_1	463.0 / 419.0	2.81	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.346	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.82	PFOS	0.195	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.19	PFDA	0.031	0.041	ü
PFUnA_1	563.0 / 519.0	3.48	PFUnA			
PFUnA_2	563.0 / 269.0	3.48	PFUnA	0.048	0.044	ü
PFDaA_1	613.0 / 569.0	3.77	PFDaA			
PFDaA_2	613.0 / 319.0	3.77	PFDaA	0.148	0.152	ü
PFTrDA_1	663.0 / 619.0	4.02	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.01	PFTrDA	0.060	0.064	ü
PFTeDA_1	713.0 / 669.0	4.24	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.24	PFTeDA	0.040	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.32	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.32	NMeFOSAA	0.526	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.48	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.51	NEtFOSSA	0.039	0.071	ü



<b>Sample Name</b>	JV66	<b>Injection Vial</b>	4
<b>Sample ID</b>	L3	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T09:38:24	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

## Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.48	PFBS			
PFBS_2	298.9 / 99.0	1.47	PFBS	0.326	0.327	ü
PFHxA_1	313.0 / 269.0	1.72	PFHxA			
PFHxA_2	313.0 / 119.0	1.72	PFHxA	0.036	0.056	ü
PFHpA_1	363.0 / 319.0	2.07	PFHpA			
PFHpA_2	363.0 / 169.0	2.08	PFHpA	0.020	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.09	PFHxS	0.327	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.059	0.062	ü
PFNA_1	463.0 / 419.0	2.81	PFNA			
PFNA_2	463.0 / 219.0	2.81	PFNA	0.290	0.294	ü
PFOS_1	499.0 / 80.0	2.81	PFOS			
PFOS_2	499.0 / 99.0	2.81	PFOS	0.212	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.17	PFDA	0.065	0.041	
PFUnA_1	563.0 / 519.0	3.48	PFUnA			
PFUnA_2	563.0 / 269.0	3.48	PFUnA	0.047	0.044	ü
PFDaA_1	613.0 / 569.0	3.77	PFDaA			
PFDaA_2	613.0 / 319.0	3.77	PFDaA	0.158	0.152	ü
PFTrDA_1	663.0 / 619.0	4.02	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.02	PFTrDA	0.066	0.064	ü
PFTeDA_1	713.0 / 669.0	4.24	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.24	PFTeDA	0.047	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.32	NMeFOSAA	0.841	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.48	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.53	NEtFOSSA	0.086	0.071	ü

Sample Name	JV67	Injection Vial	5
Sample ID	L4	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T09:47:21	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.47	PFBS			
PFBS_2	298.9 / 99.0	1.46	PFBS	0.237	0.327	ü
PFHxA_1	313.0 / 269.0	1.73	PFHxA			
PFHxA_2	313.0 / 119.0	1.72	PFHxA	0.047	0.056	ü
PFHpA_1	363.0 / 319.0	2.07	PFHpA			
PFHpA_2	363.0 / 169.0	2.07	PFHpA	0.021	0.022	ü
PFHxS_1	399.0 / 80.0	2.09	PFHxS			
PFHxS_2	399.0 / 99.0	2.09	PFHxS	0.311	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.45	PFOA	0.065	0.062	ü
PFNA_1	463.0 / 419.0	2.81	PFNA			
PFNA_2	463.0 / 219.0	2.81	PFNA	0.310	0.294	ü
PFOS_1	499.0 / 80.0	2.81	PFOS			
PFOS_2	499.0 / 99.0	2.81	PFOS	0.205	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.16	PFDA	0.036	0.041	ü
PFUnA_1	563.0 / 519.0	3.48	PFUnA			
PFUnA_2	563.0 / 269.0	3.48	PFUnA	0.044	0.044	ü
PFDaA_1	613.0 / 569.0	3.77	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.162	0.152	ü
PFTrDA_1	663.0 / 619.0	4.02	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.02	PFTrDA	0.061	0.064	ü
PFTeDA_1	713.0 / 669.0	4.24	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.048	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.31	NMeFOSAA	0.667	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.48	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.45	NEtFOSSA	0.066	0.071	ü

Sample Name	JV68	Injection Vial	6
Sample ID	L5	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T09:56:17	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.44	PFBS			
PFBS_2	298.9 / 99.0	1.44	PFBS	0.359	0.327	ü
PFHxA_1	313.0 / 269.0	1.71	PFHxA			
PFHxA_2	313.0 / 119.0	1.71	PFHxA	0.060	0.056	ü
PFHpA_1	363.0 / 319.0	2.06	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.023	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.08	PFHxS	0.302	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.43	PFOA	0.063	0.062	ü
PFNA_1	463.0 / 419.0	2.80	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.289	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.80	PFOS	0.198	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.16	PFDA	0.039	0.041	ü
PFUnA_1	563.0 / 519.0	3.47	PFUnA			
PFUnA_2	563.0 / 269.0	3.47	PFUnA	0.044	0.044	ü
PFDaA_1	613.0 / 569.0	3.76	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.152	0.152	ü
PFTTrDA_1	663.0 / 619.0	4.02	PFTTrDA			
PFTTrDA_2	663.0 / 169.0	4.02	PFTTrDA	0.059	0.064	ü
PFTeDA_1	713.0 / 669.0	4.23	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.049	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.30	NMeFOSAA	0.641	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.47	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.48	NEtFOSSA	0.070	0.071	ü

Sample Name	JV69	Injection Vial	7
Sample ID	L6	Injection Volume	10.00
Sample Type	Standard	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T10:05:12	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.44	PFBS			
PFBS_2	298.9 / 99.0	1.44	PFBS	0.343	0.327	ü
PFHxA_1	313.0 / 269.0	1.71	PFHxA			
PFHxA_2	313.0 / 119.0	1.71	PFHxA	0.062	0.056	ü
PFHpA_1	363.0 / 319.0	2.06	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.024	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.08	PFHxS	0.309	0.335	ü
PFOA_1	413.0 / 369.0	2.43	PFOA			
PFOA_2	413.0 / 169.0	2.43	PFOA	0.068	0.062	ü
PFNA_1	463.0 / 419.0	2.80	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.300	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.80	PFOS	0.197	0.194	ü
PFDA_1	513.0 / 469.0	3.15	PFDA			
PFDA_2	513.0 / 219.0	3.15	PFDA	0.039	0.041	ü
PFUnA_1	563.0 / 519.0	3.47	PFUnA			
PFUnA_2	563.0 / 269.0	3.47	PFUnA	0.039	0.044	ü
PFDaA_1	613.0 / 569.0	3.76	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.151	0.152	ü
PFTrDA_1	663.0 / 619.0	4.01	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.01	PFTrDA	0.061	0.064	ü
PFTeDA_1	713.0 / 669.0	4.23	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.045	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.31	NMeFOSAA	0.570	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.47	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.45	NEtFOSSA	0.077	0.071	ü

<b>Sample Name</b>	JV70	<b>Injection Vial</b>	8
<b>Sample ID</b>	L7	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T10:14:08	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.44	PFBS			
PFBS_2	298.9 / 99.0	1.43	PFBS	0.351	0.327	ü
PFHxA_1	313.0 / 269.0	1.71	PFHxA			
PFHxA_2	313.0 / 119.0	1.71	PFHxA	0.072	0.056	ü
PFHpA_1	363.0 / 319.0	2.06	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.020	0.022	ü
PFHxS_1	399.0 / 80.0	2.07	PFHxS			
PFHxS_2	399.0 / 99.0	2.07	PFHxS	0.286	0.335	ü
PFOA_1	413.0 / 369.0	2.43	PFOA			
PFOA_2	413.0 / 169.0	2.43	PFOA	0.062	0.062	ü
PFNA_1	463.0 / 419.0	2.80	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.294	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.80	PFOS	0.193	0.194	ü
PFDA_1	513.0 / 469.0	3.15	PFDA			
PFDA_2	513.0 / 219.0	3.15	PFDA	0.037	0.041	ü
PFUnA_1	563.0 / 519.0	3.47	PFUnA			
PFUnA_2	563.0 / 269.0	3.47	PFUnA	0.042	0.044	ü
PFDaA_1	613.0 / 569.0	3.76	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.151	0.152	ü
PFTrDA_1	663.0 / 619.0	4.01	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.01	PFTrDA	0.060	0.064	ü
PFTeDA_1	713.0 / 669.0	4.23	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.049	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.30	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.30	NMeFOSAA	0.585	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.46	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.46	NEtFOSSA	0.073	0.071	ü

<b>Sample Name</b>	JV71	<b>Injection Vial</b>	9
<b>Sample ID</b>	L8	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T10:23:03	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.45	PFBS			
PFBS_2	298.9 / 99.0	1.44	PFBS	0.318	0.327	ü
PFHxA_1	313.0 / 269.0	1.71	PFHxA			
PFHxA_2	313.0 / 119.0	1.71	PFHxA	0.069	0.056	ü
PFHpA_1	363.0 / 319.0	2.07	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.021	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.08	PFHxS	0.296	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.068	0.062	ü
PFNA_1	463.0 / 419.0	2.81	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.301	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.80	PFOS	0.197	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.15	PFDA	0.036	0.041	ü
PFUnA_1	563.0 / 519.0	3.47	PFUnA			
PFUnA_2	563.0 / 269.0	3.47	PFUnA	0.043	0.044	ü
PFDaA_1	613.0 / 569.0	3.76	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.155	0.152	ü
PFTrDA_1	663.0 / 619.0	4.01	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.01	PFTrDA	0.063	0.064	ü
PFTeDA_1	713.0 / 669.0	4.23	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.046	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.31	NMeFOSAA	0.602	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.47	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.46	NEtFOSSA	0.072	0.071	ü



<b>Sample Name</b>	JV72	<b>Injection Vial</b>	10
<b>Sample ID</b>	L9	<b>Injection Volume</b>	10.00
<b>Sample Type</b>	Standard	<b>Instrument Name</b>	QTRAP 5500
<b>Acquisition Date</b>	2018-05-17T10:31:58	<b>Data File</b>	18-0287.wiff
<b>Acquisition Method</b>	5-0371.dam	<b>Result Table</b>	18-0316Base
<b>Sample Comment</b>			

### Results Summary

Analyte	MRM Transition	RT	Ratio Group	Calculated Ion ratio	Expected Ion Ratio	Ratio OK
PFBS_1	298.9 / 80.0	1.44	PFBS			
PFBS_2	298.9 / 99.0	1.44	PFBS	0.315	0.327	ü
PFHxA_1	313.0 / 269.0	1.71	PFHxA			
PFHxA_2	313.0 / 119.0	1.71	PFHxA	0.070	0.056	ü
PFHpA_1	363.0 / 319.0	2.06	PFHpA			
PFHpA_2	363.0 / 169.0	2.06	PFHpA	0.020	0.022	ü
PFHxS_1	399.0 / 80.0	2.08	PFHxS			
PFHxS_2	399.0 / 99.0	2.08	PFHxS	0.302	0.335	ü
PFOA_1	413.0 / 369.0	2.44	PFOA			
PFOA_2	413.0 / 169.0	2.44	PFOA	0.067	0.062	ü
PFNA_1	463.0 / 419.0	2.80	PFNA			
PFNA_2	463.0 / 219.0	2.80	PFNA	0.299	0.294	ü
PFOS_1	499.0 / 80.0	2.80	PFOS			
PFOS_2	499.0 / 99.0	2.80	PFOS	0.194	0.194	ü
PFDA_1	513.0 / 469.0	3.16	PFDA			
PFDA_2	513.0 / 219.0	3.16	PFDA	0.037	0.041	ü
PFUnA_1	563.0 / 519.0	3.48	PFUnA			
PFUnA_2	563.0 / 269.0	3.47	PFUnA	0.043	0.044	ü
PFDaA_1	613.0 / 569.0	3.76	PFDaA			
PFDaA_2	613.0 / 319.0	3.76	PFDaA	0.153	0.152	ü
PFTrDA_1	663.0 / 619.0	4.01	PFTrDA			
PFTrDA_2	663.0 / 169.0	4.01	PFTrDA	0.063	0.064	ü
PFTeDA_1	713.0 / 669.0	4.23	PFTeDA			
PFTeDA_2	713.0 / 169.0	4.23	PFTeDA	0.047	0.046	ü
NMeFOSAA_1	570.0 / 419.0	3.31	NMeFOSAA			
NMeFOSAA_2	570.0 / 512.0	3.31	NMeFOSAA	0.608	0.642	ü
NEtFOSAA_1	584.0 / 419.0	3.47	NEtFOSSA			
NEtFOSAA_2	584.0 / 483.0	3.46	NEtFOSSA	0.067	0.071	ü

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T10:40:54	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.44	901.851676	885.00	101.90
PFBS_2	298.9 / 99.0	1.44	918.307737	885.00	103.76
PFHxA_1	313.0 / 269.0	1.71	951.549680	1000.00	95.15
PFHxA_2	313.0 / 119.0	1.71	942.165072	1000.00	94.22
PFHpA_1	363.0 / 319.0	2.06	872.686188	1000.00	87.27
PFHpA_2	363.0 / 169.0	2.06	946.636265	1000.00	94.66
PFHxS_1	399.0 / 80.0	2.08	896.220615	912.00	98.27
PFHxS_2	399.0 / 99.0	2.08	883.956996	912.00	96.93
PFOA_1	413.0 / 369.0	2.43	898.740895	1000.00	89.87
PFOA_2	413.0 / 169.0	2.43	903.113647	1000.00	90.31
PFNA_1	463.0 / 419.0	2.80	933.155583	1000.00	93.32
PFNA_2	463.0 / 219.0	2.80	912.518493	1000.00	91.25
PFOS_1	499.0 / 80.0	2.80	877.977513	925.60	94.85
PFOS_2	499.0 / 99.0	2.80	949.188735	925.60	102.55
PFDA_1	513.0 / 469.0	3.15	935.035030	1000.00	93.50
PFDA_2	513.0 / 219.0	3.15	853.902376	1000.00	85.39
PFUnA_1	563.0 / 519.0	3.47	899.041959	1000.00	89.90
PFUnA_2	563.0 / 269.0	3.47	921.161237	1000.00	92.12
PFDoA_1	613.0 / 569.0	3.76	939.203264	1000.00	93.92
PFDoA_2	613.0 / 319.0	3.76	933.405461	1000.00	93.34
PFTTrDA_1	663.0 / 619.0	4.01	907.522118	1000.00	90.75
PFTTrDA_2	663.0 / 169.0	4.01	881.226616	1000.00	88.12
PFTeDA_1	713.0 / 669.0	4.23	878.562702	1000.00	87.86
PFTeDA_2	713.0 / 169.0	4.23	896.542887	1000.00	89.65
NMeFOSAA_1	570.0 / 419.0	3.31	1037.863357	1000.00	103.79
NMeFOSAA_2	570.0 / 512.0	3.31	978.255006	1000.00	97.83
NEtFOSAA_1	584.0 / 419.0	3.47	1101.385563	1000.00	110.14
NEtFOSAA_2	584.0 / 483.0	3.46	760.930454	1000.00	76.09

Sample Name	JV63 ICC	Injection Vial	11
Sample ID	ICC	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T10:40:54	Data File	18-0287.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.70	95.781622	100.00	95.78
13C2-PFDA	515.0 / 470.0	3.14	96.598799	100.00	96.60
d5-EtFOSAA	589.0 / 419.0	3.46	454.100896	400.00	113.53

Sample Name	JV70 CCV	Injection Vial	8
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T15:47:06	Data File	18-0315.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.45	2389.280653	2212.50	107.99
PFBS_2	298.9 / 99.0	1.44	2336.102989	2212.50	105.59
PFHxA_1	313.0 / 269.0	1.71	2895.179038	2500.00	115.81
PFHxA_2	313.0 / 119.0	1.71	2757.167581	2500.00	110.29
PFHpA_1	363.0 / 319.0	2.07	2689.724123	2500.00	107.59
PFHpA_2	363.0 / 169.0	2.07	2957.239736	2500.00	118.29
PFHxS_1	399.0 / 80.0	2.08	2362.465731	2280.00	103.62
PFHxS_2	399.0 / 99.0	2.08	2385.168075	2280.00	104.61
PFOA_1	413.0 / 369.0	2.44	2782.126442	2500.00	111.29
PFOA_2	413.0 / 169.0	2.44	2852.993105	2500.00	114.12
PFNA_1	463.0 / 419.0	2.81	2907.125755	2500.00	116.29
PFNA_2	463.0 / 219.0	2.81	2920.415932	2500.00	116.82
PFOS_1	499.0 / 80.0	2.80	2507.720205	2314.00	108.37
PFOS_2	499.0 / 99.0	2.80	2473.922064	2314.00	106.91
PFDA_1	513.0 / 469.0	3.15	2773.485338	2500.00	110.94
PFDA_2	513.0 / 219.0	3.15	3034.005413	2500.00	121.36
PFUnA_1	563.0 / 519.0	3.47	2679.223951	2500.00	107.17
PFUnA_2	563.0 / 269.0	3.47	2779.161095	2500.00	111.17
PFDoA_1	613.0 / 569.0	3.76	2772.264301	2500.00	110.89
PFDoA_2	613.0 / 319.0	3.76	2887.210254	2500.00	115.49
PFTTrDA_1	663.0 / 619.0	4.01	2674.074282	2500.00	106.96
PFTTrDA_2	663.0 / 169.0	4.01	2725.782470	2500.00	109.03
PFTeDA_1	713.0 / 669.0	4.23	2609.373592	2500.00	104.37
PFTeDA_2	713.0 / 169.0	4.23	2706.104274	2500.00	108.24
NMeFOSAA_1	570.0 / 419.0	3.31	2597.213173	2500.00	103.89
NMeFOSAA_2	570.0 / 512.0	3.30	2695.202915	2500.00	107.81
NEtFOSAA_1	584.0 / 419.0	3.47	2612.985642	2500.00	104.52
NEtFOSAA_2	584.0 / 483.0	3.46	2545.488986	2500.00	101.82

Sample Name	JV70 CCV	Injection Vial	8
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T15:47:06	Data File	18-0315.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.70	110.534591	100.00	110.53
13C2-PFDA	515.0 / 470.0	3.14	110.333291	100.00	110.33
d5-EtFOSAA	589.0 / 419.0	3.46	439.232673	400.00	109.81

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T16:49:36	Data File	18-0315.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
PFBS_1	298.9 / 80.0	1.45	847.605926	885.00	95.77
PFBS_2	298.9 / 99.0	1.45	870.715187	885.00	98.39
PFHxA_1	313.0 / 269.0	1.72	1022.006730	1000.00	102.20
PFHxA_2	313.0 / 119.0	1.72	930.277456	1000.00	93.03
PFHpA_1	363.0 / 319.0	2.07	933.467426	1000.00	93.35
PFHpA_2	363.0 / 169.0	2.07	1082.067179	1000.00	108.21
PFHxS_1	399.0 / 80.0	2.08	827.092317	912.00	90.69
PFHxS_2	399.0 / 99.0	2.08	833.132346	912.00	91.35
PFOA_1	413.0 / 369.0	2.44	958.563866	1000.00	95.86
PFOA_2	413.0 / 169.0	2.44	1030.464497	1000.00	103.05
PFNA_1	463.0 / 419.0	2.80	980.054277	1000.00	98.01
PFNA_2	463.0 / 219.0	2.80	1003.390952	1000.00	100.34
PFOS_1	499.0 / 80.0	2.80	903.095244	925.60	97.57
PFOS_2	499.0 / 99.0	2.80	917.814249	925.60	99.16
PFDA_1	513.0 / 469.0	3.15	999.131055	1000.00	99.91
PFDA_2	513.0 / 219.0	3.15	948.230292	1000.00	94.82
PFUnA_1	563.0 / 519.0	3.47	970.375319	1000.00	97.04
PFUnA_2	563.0 / 269.0	3.47	907.593320	1000.00	90.76
PFDoA_1	613.0 / 569.0	3.76	959.999246	1000.00	96.00
PFDoA_2	613.0 / 319.0	3.76	919.087660	1000.00	91.91
PFTTrDA_1	663.0 / 619.0	4.01	919.035872	1000.00	91.90
PFTTrDA_2	663.0 / 169.0	4.01	939.185778	1000.00	93.92
PFTeDA_1	713.0 / 669.0	4.22	893.932506	1000.00	89.39
PFTeDA_2	713.0 / 169.0	4.23	916.714584	1000.00	91.67
NMeFOSAA_1	570.0 / 419.0	3.31	985.067572	1000.00	98.51
NMeFOSAA_2	570.0 / 512.0	3.30	1057.975878	1000.00	105.80
NEtFOSAA_1	584.0 / 419.0	3.46	978.433968	1000.00	97.84
NEtFOSAA_2	584.0 / 483.0	3.46	695.333625	1000.00	69.53

Not used for samples.

Sample Name	JV69 CCV	Injection Vial	7
Sample ID	CCV	Injection Volume	10.00
Sample Type	Quality Control	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-17T16:49:36	Data File	18-0315.wiff
Acquisition Method	5-0371.dam	Result Table	18-0316_SIS
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Conc. (ng/L)	Target Conc. (ng/L)	Recovery (%)
13C2-PFHxA	315.0 / 270.0	1.71	104.338054	100.00	104.34
13C2-PFDA	515.0 / 470.0	3.14	101.559171	100.00	101.56
d5-EtFOSAA	589.0 / 419.0	3.45	417.852873	400.00	104.46



Sample Calculation

Sample Name	J6148-FS(0)	Injection Vial	14
Sample ID	NAWC-050718-RW-316	Injection Volume	10.00
Sample Type	Unknown	Instrument Name	QTRAP 5500
Acquisition Date	2018-05-14T13:26:05	Data File	18-0313.wiff
Acquisition Method	5-0371.dam	Result Table	18-0313_Base
Sample Comment			

## Results Summary

Analyte	MRM Transition	RT	Area	Conc. (ng/L)	Signal/Noise Ratio	Modified
PFBS_1	298.9 / 80.0	1.46	491659.18	1947.953430	311.9	false
PFBS_2	298.9 / 99.0	1.46	151011.73	1934.967632	478.4	false
PFHxA_1	313.0 / 269.0	1.74	1373573.21	4189.631616	404.5	false
PFHxA_2	313.0 / 119.0	1.74	93862.93	4000.785589	447.4	false
PFHpA_1	363.0 / 319.0	2.10	626938.85	2005.387237	226.2	false
PFHpA_2	363.0 / 169.0	2.09	16244.42	2646.052841	273.7	false
PFHxS_1	399.0 / 80.0	2.11	2054505.64	7158.366521	385.8	false
PFHxS_2	399.0 / 99.0	2.11	580157.25	6875.560120	418.2	false
PFOA_1	413.0 / 369.0	2.47	1649246.42	4601.630824	390.2	false
PFOA_2	413.0 / 169.0	2.45	164270.03	6868.953710	522.3	false
PFNA_1	463.0 / 419.0	2.85	222849.34	644.524888	185.0	false
PFNA_2	463.0 / 219.0	2.85	66331.69	655.612956	335.4	false
PFOS_1	499.0 / 80.0	2.80	3723209.66	9382.050761	370.9	false
PFOS_2	499.0 / 99.0	2.84	612334.86	7974.569699	394.8	false
PFDA_1	513.0 / 469.0	3.21	13865.55	16.087824	51.4	false
PFDA_2	513.0 / 219.0	3.22	1163.45	30.932675	44.5	false
PFUnA_1	563.0 / 519.0	N/A	N/A	N/A	N/A	true
PFUnA_2	563.0 / 269.0	N/A	N/A	N/A	N/A	true
PFDoA_1	613.0 / 569.0	N/A	N/A	N/A	N/A	true
PFDoA_2	613.0 / 319.0	N/A	N/A	N/A	N/A	true
PFTTrDA_1	663.0 / 619.0	N/A	N/A	N/A	N/A	true
PFTTrDA_2	663.0 / 169.0	N/A	N/A	N/A	N/A	true
PFTeDA_1	713.0 / 669.0	N/A	N/A	N/A	N/A	true
PFTeDA_2	713.0 / 169.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_1	570.0 / 419.0	N/A	N/A	N/A	N/A	true
NMeFOSAA_2	570.0 / 512.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_1	584.0 / 419.0	N/A	N/A	N/A	N/A	true
NEtFOSAA_2	584.0 / 483.0	N/A	N/A	N/A	N/A	true


Sample calc PFOA 16.15 ng/L  
y=0.96808x+0.0.19816

$$((1649246.42/36858.30)-0.19816)/0.96808 * 100 * 0.001 / 0.285 = 16.146 \text{ ng/L}$$

LCS PFOA 101%      15.17/15\*100=101%



C:\AI\Projects\112008005\WE04\F.S.DR.03\NAWC\_201901.mxd MKB 1/24/2019



**LEGEND**

- PFOA & PFOS BELOW DETECTION LIMITS
- SUM OF PFOA & PFOS CONCENTRATIONS AT OR BELOW 40 PPT
- SUM OF PFOA & PFOS CONCENTRATIONS BETWEEN 40 PPT AND 70 PPT
- SUM OF PFOA & PFOS CONCENTRATIONS ABOVE 70 PPT
- 👉 PUBLIC WATER CONNECTION (SOME NOT BY NAVY)
- SAMPLING AREA
- FORMER NAWC WARRINSTER
- TOWNSHIP BOUNDARY



**PFC SAMPLE RESULTS MAP**  
 FORMER NAWC WARRINSTER  
 MONTGOMERY COUNTY, PENNSYLVANIA

SCALE AS NOTED	
FILE	
NAWC	
REV 0	DATE 01/24/19
FIGURE NUMBER	
FIGURE 1	