



**Off-base Drinking Water Sample Results,
Level 2 Laboratory Report, Level 4 Laboratory Report,
Electronic Data Deliverable, Data Validation Report,
and the Sample Location Figure, SDG 320-23971**

Ault Field

Naval Air Station Whidbey Island

Oak Harbor, Washington

February 2019

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Sacramento

880 Riverside Parkway
West Sacramento, CA 95605

Tel: (916)373-5600

TestAmerica Job ID: 320-23971-1

Client Project/Site: Whidbey Island

For:

CH2M Hill Constructors, Inc.
1100 NE Circle Blvd
Corvallis, Oregon 97330

Attn: Tiffany Hill



Authorized for release by:

12/13/2016 6:04:47 PM

Laura Turpen, Project Manager I

(916)374-4414

laura.turpen@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?

Ask
The
Expert

Visit us at:

www.testamericainc.com

The test results in this report meet all 2003 NELAC and 2009 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Detection Summary	6
Client Sample Results	7
Surrogate Summary	10
QC Sample Results	11
QC Association Summary	12
Lab Chronicle	13
Certification Summary	15
Method Summary	16
Sample Summary	17
Chain of Custody	18
Receipt Checklists	19

Definitions/Glossary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Undetected at the Limit of Detection.
M	Manual integrated compound.

Glossary

Abbreviation

These commonly used abbreviations may or may not be present in this report.

□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Case Narrative

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Job ID: 320-23971-1

Laboratory: TestAmerica Sacramento

Narrative

CASE NARRATIVE

Client: CH2M Hill Constructors, Inc.

Project: Whidbey Island

Report Number: 320-23971-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

TestAmerica Sacramento attests to the validity of the laboratory data generated by TestAmerica facilities reported herein. All analyses performed by TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the applicable methods. TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

TestAmerica utilizes USEPA approved methods and DOD QSM, where applicable, in all analytical work. The samples presented in this report were analyzed for the parameter(s) listed on the analytical methods summary page in accordance with the method(s) indicated. A summary of QC data for these analyses is included at the back of the report.

All parameters for which TestAmerica Sacramento has certification were evaluated to the QSM specified reporting convention or to the client specified format if different from QSM. Parameters not certified under QSM, if any, were evaluated to the detection limit (DL) and include qualified results where applicable.

The sample(s) that contain constituents flagged with U are undetected. The result associated with this flag is the limit of detection (LOD).

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of TestAmerica and its client.

RECEIPT

The samples were received on 12/02/2016; the samples arrived in good condition, properly preserved and on ice. The temperature of the coolers at receipt was 0.4 C.

Receipt Exceptions

The container label for the following sample did not match the information listed on the Chain-of-Custody (COC): WI-AF-3FB12-1116 (320-23971-10). The container labels list WI-AF-3FB12-1116, while the COC lists WI-AF-3FB11-1116. Based on the label ID, the sample time, and the location of the sample on the COC, it appears that the sample is supposed to be the field blank for sample WI-AF-3RW12-1116. This indicates that the sample label is correct and the COC contained a typo; therefore the sample was logged in according to the label ID. The client confirmed that the sample label was correct for this sample, not the COC.

Case Narrative

Client: CH2M Hill Constructors, Inc.

Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Job ID: 320-23971-1 (Continued)

Laboratory: TestAmerica Sacramento (Continued)

PFOA/PFOS

Samples WI-AF-2RW07-1116 (320-23971-1), WI-AF-2FB07-1116 (320-23971-2), WI-AF-2RW08-1116 (320-23971-3), WI-AF-2FB08-1116 (320-23971-4), WI-AF-3RW10-1116 (320-23971-5), WI-AF-3FB10-1116 (320-23971-6), WI-AF-3RW11-1116 (320-23971-7), WI-AF-3FB11-1116 (320-23971-8), WI-AF-3RW12-1116 (320-23971-9) and WI-AF-3FB12-1116 (320-23971-10) were analyzed for PFOA/PFOS in accordance with 537. The samples were prepared on 12/05/2016 and analyzed on 12/11/2016.

These samples have a pH 9:

WI-AF-2FB07-1116 (320-23971-2), WI-AF-2RW08-1116 (320-23971-3), WI-AF-2FB08-1116 (320-23971-4), WI-AF-3RW10-1116 (320-23971-5), WI-AF-3FB10-1116 (320-23971-6), WI-AF-3RW11-1116 (320-23971-7), WI-AF-3FB11-1116 (320-23971-8) and WI-AF-3FB12-1116 (320-23971-10)

Due to the excessive amount of sediment in the sample bottles, the aqueous portion of these samples was decanted to new bottles prior to spiking and the extraction.

WI-AF-3RW12-1116 (320-23971-9)

Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with preparation batch 320-140632.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

Detection Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116	Lab Sample ID: 320-23971-1
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-2FB07-1116	Lab Sample ID: 320-23971-2
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-2RW08-1116	Lab Sample ID: 320-23971-3
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-2FB08-1116	Lab Sample ID: 320-23971-4
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3RW10-1116	Lab Sample ID: 320-23971-5
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3FB10-1116	Lab Sample ID: 320-23971-6
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3RW11-1116	Lab Sample ID: 320-23971-7
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3FB11-1116	Lab Sample ID: 320-23971-8
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3RW12-1116	Lab Sample ID: 320-23971-9
<input type="checkbox"/> No Detections.	
Client Sample ID: WI-AF-3FB12-1116	Lab Sample ID: 320-23971-10
<input type="checkbox"/> No Detections.	

This Detection Summary does not include radiochemical test results.

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Date Collected: 11/30/16 09:10
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-1

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.047	U M	0.058	0.015	ug/L		12/05/16 11:42	12/11/16 15:00	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.029	0.0091	ug/L		12/05/16 11:42	12/11/16 15:00	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L		12/05/16 11:42	12/11/16 15:00	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	103		70 - 130				12/05/16 11:42	12/11/16 15:00	1
13C2 PFDA	118		70 - 130				12/05/16 11:42	12/11/16 15:00	1

Client Sample ID: WI-AF-2FB07-1116

Date Collected: 11/30/16 09:12
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-2

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.015	ug/L		12/05/16 11:42	12/11/16 15:30	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 15:30	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 15:30	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	112		70 - 130				12/05/16 11:42	12/11/16 15:30	1
13C2 PFDA	112		70 - 130				12/05/16 11:42	12/11/16 15:30	1

Client Sample ID: WI-AF-2RW08-1116

Date Collected: 11/30/16 09:30
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-3

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.014	ug/L		12/05/16 11:42	12/11/16 15:59	1
Perfluoroctanoic acid (PFOA)	0.022	U M	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 15:59	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 15:59	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	96		70 - 130				12/05/16 11:42	12/11/16 15:59	1
13C2 PFDA	108		70 - 130				12/05/16 11:42	12/11/16 15:59	1

Client Sample ID: WI-AF-2FB08-1116

Date Collected: 11/30/16 09:32
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-4

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 16:29	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 16:29	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 16:29	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	108		70 - 130				12/05/16 11:42	12/11/16 16:29	1
13C2 PFDA	105		70 - 130				12/05/16 11:42	12/11/16 16:29	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-5

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 16:58	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 16:58	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 16:58	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/05/16 11:42	12/11/16 16:58	1
13C2 PFDA	111		70 - 130				12/05/16 11:42	12/11/16 16:58	1

Client Sample ID: WI-AF-3FB10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-6

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.044	U M	0.056	0.014	ug/L		12/05/16 11:42	12/11/16 17:28	1
Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.0087	ug/L		12/05/16 11:42	12/11/16 17:28	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 17:28	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	117		70 - 130				12/05/16 11:42	12/11/16 17:28	1
13C2 PFDA	106		70 - 130				12/05/16 11:42	12/11/16 17:28	1

Client Sample ID: WI-AF-3RW11-1116

Date Collected: 11/30/16 10:30

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-7

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L		12/05/16 11:42	12/11/16 17:58	1
Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 17:58	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 17:58	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/05/16 11:42	12/11/16 17:58	1
13C2 PFDA	109		70 - 130				12/05/16 11:42	12/11/16 17:58	1

Client Sample ID: WI-AF-3FB11-1116

Date Collected: 11/30/16 10:31

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-8

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U M	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 19:56	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 19:56	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 19:56	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	108		70 - 130				12/05/16 11:42	12/11/16 19:56	1
13C2 PFDA	108		70 - 130				12/05/16 11:42	12/11/16 19:56	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW12-1116

Date Collected: 11/30/16 17:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-9

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.059	0.015	ug/L		12/05/16 11:42	12/11/16 20:26	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.0092	ug/L		12/05/16 11:42	12/11/16 20:26	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L		12/05/16 11:42	12/11/16 20:26	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	119		70 - 130				12/05/16 11:42	12/11/16 20:26	1
13C2 PFDA	130		70 - 130				12/05/16 11:42	12/11/16 20:26	1

Client Sample ID: WI-AF-3FB12-1116

Date Collected: 11/30/16 17:01

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-10

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 20:55	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 20:55	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 20:55	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	99		70 - 130				12/05/16 11:42	12/11/16 20:55	1
13C2 PFDA	105		70 - 130				12/05/16 11:42	12/11/16 20:55	1

Surrogate Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	3C2 PFHxA (70-130)	3C2 PFDA (70-130)											
320-23971-1	WI-AF-2RW07-1116	103	118											
320-23971-2	WI-AF-2FB07-1116	112	112											
320-23971-3	WI-AF-2RW08-1116	96	108											
320-23971-4	WI-AF-2FB08-1116	108	105											
320-23971-5	WI-AF-3RW10-1116	110	111											
320-23971-6	WI-AF-3FB10-1116	117	106											
320-23971-7	WI-AF-3RW11-1116	110	109											
320-23971-8	WI-AF-3FB11-1116	108	108											
320-23971-9	WI-AF-3RW12-1116	119	130											
320-23971-10	WI-AF-3FB12-1116	99	105											
LCS 320-140632/2-A	Lab Control Sample	113	110											
LCSD 320-140632/3-A	Lab Control Sample Dup	117	111											
MB 320-140632/1-A	Method Blank	106	104											

Surrogate Legend

13C2 PFHxA = 13C2 PFHxA

13C2 PFDA = 13C2 PFDA

QC Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 320-140632/1-A

Matrix: Water

Analysis Batch: 141573

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 140632

Analyte	MB		LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.016	ug/L		12/05/16 11:42	12/11/16 13:31	1
Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.0094	ug/L		12/05/16 11:42	12/11/16 13:31	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L		12/05/16 11:42	12/11/16 13:31	1

Surrogate	MB		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C2 PFHxA	106		70 - 130	12/05/16 11:42	12/11/16 13:31	1
13C2 PFDA	104		70 - 130	12/05/16 11:42	12/11/16 13:31	1

Lab Sample ID: LCS 320-140632/2-A

Matrix: Water

Analysis Batch: 141573

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 140632

Analyte	Spike		LCS Result	LCS Qualifier	Unit	D	%Rec.	Limits
	Added	Result						
Perfluorooctanesulfonic acid (PFOS)	0.160	0.124	ug/L			77	70 - 130	
Perfluorooctanoic acid (PFOA)		0.0811	ug/L	0.0619		76	70 - 130	
Perfluorobutanesulfonic acid (PFBS)		0.359	ug/L	0.275		77	70 - 130	

Surrogate	LCS		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C2 PFHxA	113		70 - 130	12/05/16 11:42	12/11/16 13:31	1
13C2 PFDA	110		70 - 130	12/05/16 11:42	12/11/16 13:31	1

Lab Sample ID: LCSD 320-140632/3-A

Matrix: Water

Analysis Batch: 141573

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 140632

Analyte	Spike		LCSD Result	LCSD Qualifier	Unit	D	%Rec.	RPD	Limit
	Added	Result							
Perfluorooctanesulfonic acid (PFOS)	0.160	0.129	ug/L			80	70 - 130	4	30
Perfluorooctanoic acid (PFOA)		0.0811	ug/L	0.0627		77	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)		0.359	ug/L	0.294		82	70 - 130	7	30

Surrogate	LCSD		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C2 PFHxA	117		70 - 130	12/05/16 11:42	12/11/16 13:31	1
13C2 PFDA	111		70 - 130	12/05/16 11:42	12/11/16 13:31	1

TestAmerica Sacramento

QC Association Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

LCMS

Prep Batch: 140632

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-1	WI-AF-2RW07-1116	Total/NA	Water	537	
320-23971-2	WI-AF-2FB07-1116	Total/NA	Water	537	
320-23971-3	WI-AF-2RW08-1116	Total/NA	Water	537	
320-23971-4	WI-AF-2FB08-1116	Total/NA	Water	537	
320-23971-5	WI-AF-3RW10-1116	Total/NA	Water	537	
320-23971-6	WI-AF-3FB10-1116	Total/NA	Water	537	
320-23971-7	WI-AF-3RW11-1116	Total/NA	Water	537	
320-23971-8	WI-AF-3FB11-1116	Total/NA	Water	537	
320-23971-9	WI-AF-3RW12-1116	Total/NA	Water	537	
320-23971-10	WI-AF-3FB12-1116	Total/NA	Water	537	
MB 320-140632/1-A	Method Blank	Total/NA	Water	537	
LCS 320-140632/2-A	Lab Control Sample	Total/NA	Water	537	
LCSD 320-140632/3-A	Lab Control Sample Dup	Total/NA	Water	537	

Analysis Batch: 141573

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-1	WI-AF-2RW07-1116	Total/NA	Water	537	140632
320-23971-2	WI-AF-2FB07-1116	Total/NA	Water	537	140632
320-23971-3	WI-AF-2RW08-1116	Total/NA	Water	537	140632
320-23971-4	WI-AF-2FB08-1116	Total/NA	Water	537	140632
320-23971-5	WI-AF-3RW10-1116	Total/NA	Water	537	140632
320-23971-6	WI-AF-3FB10-1116	Total/NA	Water	537	140632
320-23971-7	WI-AF-3RW11-1116	Total/NA	Water	537	140632
MB 320-140632/1-A	Method Blank	Total/NA	Water	537	140632
LCS 320-140632/2-A	Lab Control Sample	Total/NA	Water	537	140632
LCSD 320-140632/3-A	Lab Control Sample Dup	Total/NA	Water	537	140632

Analysis Batch: 141574

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-8	WI-AF-3FB11-1116	Total/NA	Water	537	140632
320-23971-9	WI-AF-3RW12-1116	Total/NA	Water	537	140632
320-23971-10	WI-AF-3FB12-1116	Total/NA	Water	537	140632

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Date Collected: 11/30/16 09:10

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			257.9 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 15:00	JRB	TAL SAC

Client Sample ID: WI-AF-2FB07-1116

Date Collected: 11/30/16 09:12

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			266.2 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 15:30	JRB	TAL SAC

Client Sample ID: WI-AF-2RW08-1116

Date Collected: 11/30/16 09:30

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			269 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 15:59	JRB	TAL SAC

Client Sample ID: WI-AF-2FB08-1116

Date Collected: 11/30/16 09:32

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			263.7 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 16:29	JRB	TAL SAC

Client Sample ID: WI-AF-3RW10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			264 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 16:58	JRB	TAL SAC

Client Sample ID: WI-AF-3FB10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			270.1 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 17:28	JRB	TAL SAC

TestAmerica Sacramento

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW11-1116

Date Collected: 11/30/16 10:30
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			268.9 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141573	12/11/16 17:58	JRB	TAL SAC

Client Sample ID: WI-AF-3FB11-1116

Date Collected: 11/30/16 10:31
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			265.1 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141574	12/11/16 19:56	JRB	TAL SAC

Client Sample ID: WI-AF-3RW12-1116

Date Collected: 11/30/16 17:00
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-9

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			256.4 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141574	12/11/16 20:26	JRB	TAL SAC

Client Sample ID: WI-AF-3FB12-1116

Date Collected: 11/30/16 17:01
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-10

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			265.2 mL	1.0 mL	140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1			141574	12/11/16 20:55	JRB	TAL SAC

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

TestAmerica Sacramento

Certification Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Laboratory: TestAmerica Sacramento

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
A2LA	DoD ELAP		2928-01	01-31-17
Analysis Method	Prep Method	Matrix	Analyte	

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

TestAmerica Sacramento

Method Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method	Method Description	Protocol	Laboratory
537	Perfluorinated Alkyl Acids (LC/MS)	EPA	TAL SAC

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Sample Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
320-23971-1	WI-AF-2RW07-1116	Water	11/30/16 09:10	12/02/16 09:40
320-23971-2	WI-AF-2FB07-1116	Water	11/30/16 09:12	12/02/16 09:40
320-23971-3	WI-AF-2RW08-1116	Water	11/30/16 09:30	12/02/16 09:40
320-23971-4	WI-AF-2FB08-1116	Water	11/30/16 09:32	12/02/16 09:40
320-23971-5	WI-AF-3RW10-1116	Water	11/30/16 10:00	12/02/16 09:40
320-23971-6	WI-AF-3FB10-1116	Water	11/30/16 10:00	12/02/16 09:40
320-23971-7	WI-AF-3RW11-1116	Water	11/30/16 10:30	12/02/16 09:40
320-23971-8	WI-AF-3FB11-1116	Water	11/30/16 10:31	12/02/16 09:40
320-23971-9	WI-AF-3RW12-1116	Water	11/30/16 17:00	12/02/16 09:40
320-23971-10	WI-AF-3FB12-1116	Water	11/30/16 17:01	12/02/16 09:40

TestAmerica Sacramento

Login Sample Receipt Checklist

Client: CH2M Hill Constructors, Inc.

Job Number: 320-23971-1

Login Number: 23971

List Source: TestAmerica Sacramento

List Number: 1

Creator: Turpen, Troy

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	Seals
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	
There are no discrepancies between the containers received and the COC.	False	Refer to Job Narrative for details.
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	N/A	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

ANALYTICAL REPORT


Job Number: 320-23971-1

Job Description: Whidbey Island

For:

CH2M Hill Constructors, Inc.
1100 NE Circle Blvd
Corvallis, OR 97330

Attention: Tiffany Hill



Approved for release.
Laura Turpen
Project Manager I
12/13/2016 6:05 PM

Laura Turpen, Project Manager I
880 Riverside Parkway, West Sacramento, CA, 95605
(916)374-4414
laura.turpen@testamericainc.com
12/13/2016

Table of Contents

Cover Title Page	1
Data Summaries	4
Definitions	4
Case Narrative	5
Detection Summary	7
Client Sample Results	8
Default Detection Limits	11
Surrogate Summary	12
QC Sample Results	13
QC Association	14
Chronicle	15
Certification Summary	17
Method Summary	18
Sample Summary	19
Manual Integration Summary	20
Reagent Traceability	24
COAs	32
Organic Sample Data	82
LCMS	82
Method 537 DOD	82
Method 537 DOD QC Summary	83
Method 537 DOD Sample Data	92
Standards Data	138
Method 537 DOD ICAL Data	138
Method 537 DOD CCAL Data	161
Raw QC Data	185

Table of Contents

Method 537 DOD Blank Data	185
Method 537 DOD LCS/LCSD Data	191
Method 537 DOD Run Logs	199
Method 537 DOD Prep Data	202
Shipping and Receiving Documents	214
Client Chain of Custody	215
Sample Receipt Checklist	216

Definitions/Glossary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Undetected at the Limit of Detection.
M	Manual integrated compound.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
¤	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

CASE NARRATIVE

Client: CH2M Hill Constructors, Inc.

Project: Whidbey Island

Report Number: 320-23971-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

TestAmerica Sacramento attests to the validity of the laboratory data generated by TestAmerica facilities reported herein. All analyses performed by TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the applicable methods. TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

TestAmerica utilizes USEPA approved methods and DOD QSM, where applicable, in all analytical work. The samples presented in this report were analyzed for the parameter(s) listed on the analytical methods summary page in accordance with the method(s) indicated. A summary of QC data for these analyses is included at the back of the report.

All parameters for which TestAmerica Sacramento has certification were evaluated to the QSM specified reporting convention or to the client specified format if different from QSM. Parameters not certified under QSM, if any, were evaluated to the detection limit (DL) and include qualified results where applicable.

The sample(s) that contain constituents flagged with U are undetected. The result associated with this flag is the limit of detection (LOD).

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of TestAmerica and its client.

RECEIPT

The samples were received on 12/02/2016; the samples arrived in good condition, properly preserved and on ice. The temperature of the coolers at receipt was 0.4 C.

Receipt Exceptions

The container label for the following sample did not match the information listed on the Chain-of-Custody (COC): WI-AF-3FB12-1116 (320-23971-10). The container labels list WI-AF-3FB12-1116, while the COC lists WI-AF-3FB11-1116. Based on the label ID, the sample time, and the location of the sample on the COC, it appears that the sample is supposed to be the field blank for sample WI-AF-3RW12-1116. This indicates that the sample label is correct and the COC contained a typo; therefore the sample was logged in according to the label ID. The client confirmed that the sample label was correct for this sample, not the COC.

PFOA/PFOS

Samples WI-AF-2RW07-1116 (320-23971-1), WI-AF-2FB07-1116 (320-23971-2), WI-AF-2RW08-1116 (320-23971-3), WI-AF-2FB08-1116 (320-23971-4), WI-AF-3RW10-1116 (320-23971-5), WI-AF-3FB10-1116 (320-23971-6), WI-AF-3RW11-1116 (320-23971-7), WI-AF-3FB11-1116 (320-23971-8), WI-AF-3RW12-1116 (320-23971-9) and WI-AF-3FB12-1116 (320-23971-10) were analyzed for PFOA/PFOS in accordance with 537. The samples were prepared on 12/05/2016 and analyzed on 12/11/2016.

These samples have a pH 9:

WI-AF-2FB07-1116 (320-23971-2), WI-AF-2RW08-1116 (320-23971-3), WI-AF-2FB08-1116 (320-23971-4), WI-AF-3RW10-1116 (320-23971-5), WI-AF-3FB10-1116 (320-23971-6), WI-AF-3RW11-1116 (320-23971-7), WI-AF-3FB11-1116 (320-23971-8) and WI-AF-3FB12-1116 (320-23971-10)

Due to the excessive amount of sediment in the sample bottles, the aqueous portion of these samples was decanted to new bottles prior to spiking and the extraction.

WI-AF-3RW12-1116 (320-23971-9)

Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with preparation batch 320-140632.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

Detection Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Lab Sample ID: 320-23971-1

No Detections.

Client Sample ID: WI-AF-2FB07-1116

Lab Sample ID: 320-23971-2

No Detections.

Client Sample ID: WI-AF-2RW08-1116

Lab Sample ID: 320-23971-3

No Detections.

Client Sample ID: WI-AF-2FB08-1116

Lab Sample ID: 320-23971-4

No Detections.

Client Sample ID: WI-AF-3RW10-1116

Lab Sample ID: 320-23971-5

No Detections.

Client Sample ID: WI-AF-3FB10-1116

Lab Sample ID: 320-23971-6

No Detections.

Client Sample ID: WI-AF-3RW11-1116

Lab Sample ID: 320-23971-7

No Detections.

Client Sample ID: WI-AF-3FB11-1116

Lab Sample ID: 320-23971-8

No Detections.

Client Sample ID: WI-AF-3RW12-1116

Lab Sample ID: 320-23971-9

No Detections.

Client Sample ID: WI-AF-3FB12-1116

Lab Sample ID: 320-23971-10

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Date Collected: 11/30/16 09:10
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-1

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.047	U M	0.058	0.015	ug/L	D	12/05/16 11:42	12/11/16 15:00	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.029	0.0091	ug/L		12/05/16 11:42	12/11/16 15:00	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L		12/05/16 11:42	12/11/16 15:00	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	103		70 - 130				12/05/16 11:42	12/11/16 15:00	1
13C2 PFDA	118		70 - 130				12/05/16 11:42	12/11/16 15:00	1

Client Sample ID: WI-AF-2FB07-1116

Date Collected: 11/30/16 09:12
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-2

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U M	0.056	0.015	ug/L	D	12/05/16 11:42	12/11/16 15:30	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 15:30	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 15:30	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	112		70 - 130				12/05/16 11:42	12/11/16 15:30	1
13C2 PFDA	112		70 - 130				12/05/16 11:42	12/11/16 15:30	1

Client Sample ID: WI-AF-2RW08-1116

Date Collected: 11/30/16 09:30
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-3

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U M	0.056	0.014	ug/L	D	12/05/16 11:42	12/11/16 15:59	1
Perfluoroctanoic acid (PFOA)	0.022	U M	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 15:59	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 15:59	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	96		70 - 130				12/05/16 11:42	12/11/16 15:59	1
13C2 PFDA	108		70 - 130				12/05/16 11:42	12/11/16 15:59	1

Client Sample ID: WI-AF-2FB08-1116

Date Collected: 11/30/16 09:32
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-4

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.046	U	0.057	0.015	ug/L	D	12/05/16 11:42	12/11/16 16:29	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 16:29	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 16:29	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	108		70 - 130				12/05/16 11:42	12/11/16 16:29	1
13C2 PFDA	105		70 - 130				12/05/16 11:42	12/11/16 16:29	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW10-1116

Date Collected: 11/30/16 10:00
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-5
Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 16:58	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 16:58	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 16:58	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/05/16 11:42	12/11/16 16:58	1
13C2 PFDA	111		70 - 130				12/05/16 11:42	12/11/16 16:58	1

Client Sample ID: WI-AF-3FB10-1116

Date Collected: 11/30/16 10:00
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-6
Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.044	U M	0.056	0.014	ug/L		12/05/16 11:42	12/11/16 17:28	1
Perfluoroctanoic acid (PFOA)	0.022	U	0.028	0.0087	ug/L		12/05/16 11:42	12/11/16 17:28	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 17:28	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	117		70 - 130				12/05/16 11:42	12/11/16 17:28	1
13C2 PFDA	106		70 - 130				12/05/16 11:42	12/11/16 17:28	1

Client Sample ID: WI-AF-3RW11-1116

Date Collected: 11/30/16 10:30
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-7
Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L		12/05/16 11:42	12/11/16 17:58	1
Perfluoroctanoic acid (PFOA)	0.022	U	0.028	0.0088	ug/L		12/05/16 11:42	12/11/16 17:58	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/05/16 11:42	12/11/16 17:58	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/05/16 11:42	12/11/16 17:58	1
13C2 PFDA	109		70 - 130				12/05/16 11:42	12/11/16 17:58	1

Client Sample ID: WI-AF-3FB11-1116

Date Collected: 11/30/16 10:31
Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-8
Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U M	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 19:56	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 19:56	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 19:56	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	108		70 - 130				12/05/16 11:42	12/11/16 19:56	1
13C2 PFDA	108		70 - 130				12/05/16 11:42	12/11/16 19:56	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW12-1116

Date Collected: 11/30/16 17:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-9

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.047	U	0.059	0.015	ug/L		12/05/16 11:42	12/11/16 20:26	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.029	0.0092	ug/L		12/05/16 11:42	12/11/16 20:26	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L		12/05/16 11:42	12/11/16 20:26	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	119		70 - 130				12/05/16 11:42	12/11/16 20:26	1
13C2 PFDA	130		70 - 130				12/05/16 11:42	12/11/16 20:26	1

Client Sample ID: WI-AF-3FB12-1116

Date Collected: 11/30/16 17:01

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-10

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroctanesulfonic acid (PFOS)	0.045	U	0.057	0.015	ug/L		12/05/16 11:42	12/11/16 20:55	1
Perfluoroctanoic acid (PFOA)	0.023	U	0.028	0.0089	ug/L		12/05/16 11:42	12/11/16 20:55	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/05/16 11:42	12/11/16 20:55	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	99		70 - 130				12/05/16 11:42	12/11/16 20:55	1
13C2 PFDA	105		70 - 130				12/05/16 11:42	12/11/16 20:55	1

Default Detection Limits

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Prep: 537

Analyte	LOQ	DL	Units	Method
Perfluorobutanesulfonic acid (PFBS)	0.14	0.048	ug/L	537
Perfluorooctanesulfonic acid (PFOS)	0.060	0.016	ug/L	537
Perfluorooctanoic acid (PFOA)	0.030	0.0094	ug/L	537

Surrogate Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

3C2 PFHx 3C2 PFDA

(70-130) (70-130)

Lab Sample ID	Client Sample ID	3C2 PFHx (70-130)	3C2 PFDA (70-130)
320-23971-1	WI-AF-2RW07-1116	103	118
320-23971-2	WI-AF-2FB07-1116	112	112
320-23971-3	WI-AF-2RW08-1116	96	108
320-23971-4	WI-AF-2FB08-1116	108	105
320-23971-5	WI-AF-3RW10-1116	110	111
320-23971-6	WI-AF-3FB10-1116	117	106
320-23971-7	WI-AF-3RW11-1116	110	109
320-23971-8	WI-AF-3FB11-1116	108	108
320-23971-9	WI-AF-3RW12-1116	119	130
320-23971-10	WI-AF-3FB12-1116	99	105
LCS 320-140632/2-A	Lab Control Sample	113	110
LCSD 320-140632/3-A	Lab Control Sample Dup	117	111
MB 320-140632/1-A	Method Blank	106	104

Surrogate Legend

13C2 PFHxA = 13C2 PFHxA

13C2 PFDA = 13C2 PFDA

QC Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 320-140632/1-A

Matrix: Water

Analysis Batch: 141573

Analyte	MB		LOQ	DL	Unit	D	Prepared		Dil Fac
	Result	Qualifier					Prepared	Analyzed	
Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.016	ug/L	12/05/16 11:42	12/11/16 13:31	1	
Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.0094	ug/L	12/05/16 11:42	12/11/16 13:31	1	
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L	12/05/16 11:42	12/11/16 13:31	1	
Surrogate	MB		%Recovery	Qualifier	Limits	D	Prepared		Dil Fac
	LCS	LCSD					Prepared	Analyzed	
13C2 PFHxA	106		70 - 130			12/05/16 11:42	12/11/16 13:31	1	
13C2 PFDA	104		70 - 130			12/05/16 11:42	12/11/16 13:31	1	

Lab Sample ID: LCS 320-140632/2-A

Matrix: Water

Analysis Batch: 141573

Analyte	Spike		LCS Result	LCS Qualifier	Unit	D	%Rec.		Limits
	Added	Result					%Rec	Limit	
Perfluorooctanesulfonic acid (PFOS)	0.160	0.124	ug/L	77	70 - 130				
Perfluorooctanoic acid (PFOA)	0.0811	0.0619	ug/L	76	70 - 130				
Perfluorobutanesulfonic acid (PFBS)	0.359	0.275	ug/L	77	70 - 130				
Surrogate	LCS		%Recovery	Qualifier	Limits	D	LCS		Limits
	LCS	LCSD					%Rec	Limit	
13C2 PFHxA	113		70 - 130						
13C2 PFDA	110		70 - 130						

Lab Sample ID: LCSD 320-140632/3-A

Matrix: Water

Analysis Batch: 141573

Analyte	Spike		LCSD Result	LCSD Qualifier	Unit	D	%Rec.		RPD	Limit
	Added	Result					%Rec	Limit		
Perfluorooctanesulfonic acid (PFOS)	0.160	0.129	ug/L	80	70 - 130	4	30			
Perfluorooctanoic acid (PFOA)	0.0811	0.0627	ug/L	77	70 - 130	1	30			
Perfluorobutanesulfonic acid (PFBS)	0.359	0.294	ug/L	82	70 - 130	7	30			
Surrogate	LCSD		%Recovery	Qualifier	Limits	D	LCSD		RPD	Limit
	LCSD	LCSD					%Rec	Limit		
13C2 PFHxA	117		70 - 130							
13C2 PFDA	111		70 - 130							

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 140632

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 140632

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 140632

QC Association Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

LCMS

Prep Batch: 140632

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-1	WI-AF-2RW07-1116	Total/NA	Water	537	
320-23971-2	WI-AF-2FB07-1116	Total/NA	Water	537	
320-23971-3	WI-AF-2RW08-1116	Total/NA	Water	537	
320-23971-4	WI-AF-2FB08-1116	Total/NA	Water	537	
320-23971-5	WI-AF-3RW10-1116	Total/NA	Water	537	
320-23971-6	WI-AF-3FB10-1116	Total/NA	Water	537	
320-23971-7	WI-AF-3RW11-1116	Total/NA	Water	537	
320-23971-8	WI-AF-3FB11-1116	Total/NA	Water	537	
320-23971-9	WI-AF-3RW12-1116	Total/NA	Water	537	
320-23971-10	WI-AF-3FB12-1116	Total/NA	Water	537	
MB 320-140632/1-A	Method Blank	Total/NA	Water	537	
LCS 320-140632/2-A	Lab Control Sample	Total/NA	Water	537	
LCSD 320-140632/3-A	Lab Control Sample Dup	Total/NA	Water	537	

Analysis Batch: 141573

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-1	WI-AF-2RW07-1116	Total/NA	Water	537	140632
320-23971-2	WI-AF-2FB07-1116	Total/NA	Water	537	140632
320-23971-3	WI-AF-2RW08-1116	Total/NA	Water	537	140632
320-23971-4	WI-AF-2FB08-1116	Total/NA	Water	537	140632
320-23971-5	WI-AF-3RW10-1116	Total/NA	Water	537	140632
320-23971-6	WI-AF-3FB10-1116	Total/NA	Water	537	140632
320-23971-7	WI-AF-3RW11-1116	Total/NA	Water	537	140632
MB 320-140632/1-A	Method Blank	Total/NA	Water	537	140632
LCS 320-140632/2-A	Lab Control Sample	Total/NA	Water	537	140632
LCSD 320-140632/3-A	Lab Control Sample Dup	Total/NA	Water	537	140632

Analysis Batch: 141574

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-23971-8	WI-AF-3FB11-1116	Total/NA	Water	537	140632
320-23971-9	WI-AF-3RW12-1116	Total/NA	Water	537	140632
320-23971-10	WI-AF-3FB12-1116	Total/NA	Water	537	140632

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Date Collected: 11/30/16 09:10

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 15:00	JRB	TAL SAC

Client Sample ID: WI-AF-2FB07-1116

Date Collected: 11/30/16 09:12

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 15:30	JRB	TAL SAC

Client Sample ID: WI-AF-2RW08-1116

Date Collected: 11/30/16 09:30

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 15:59	JRB	TAL SAC

Client Sample ID: WI-AF-2FB08-1116

Date Collected: 11/30/16 09:32

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 16:29	JRB	TAL SAC

Client Sample ID: WI-AF-3RW10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 16:58	JRB	TAL SAC

Client Sample ID: WI-AF-3FB10-1116

Date Collected: 11/30/16 10:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 17:28	JRB	TAL SAC

TestAmerica Sacramento

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Client Sample ID: WI-AF-3RW11-1116

Date Collected: 11/30/16 10:30

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141573	12/11/16 17:58	JRB	TAL SAC

Client Sample ID: WI-AF-3FB11-1116

Date Collected: 11/30/16 10:31

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141574	12/11/16 19:56	JRB	TAL SAC

Client Sample ID: WI-AF-3RW12-1116

Date Collected: 11/30/16 17:00

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-9

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141574	12/11/16 20:26	JRB	TAL SAC

Client Sample ID: WI-AF-3FB12-1116

Date Collected: 11/30/16 17:01

Date Received: 12/02/16 09:40

Lab Sample ID: 320-23971-10

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			140632	12/05/16 11:42	NS1	TAL SAC
Total/NA	Analysis	537		1	141574	12/11/16 20:55	JRB	TAL SAC

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Certification Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Laboratory: TestAmerica Sacramento

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
A2LA	DoD ELAP		2928-01	01-31-17
Analysis Method	Prep Method	Matrix	Analyte	

Method Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Method	Method Description	Protocol	Laboratory
537	Perfluorinated Alkyl Acids (LC/MS)	EPA	TAL SAC

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Sample Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-23971-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
320-23971-1	WI-AF-2RW07-1116	Water	11/30/16 09:10	12/02/16 09:40
320-23971-2	WI-AF-2FB07-1116	Water	11/30/16 09:12	12/02/16 09:40
320-23971-3	WI-AF-2RW08-1116	Water	11/30/16 09:30	12/02/16 09:40
320-23971-4	WI-AF-2FB08-1116	Water	11/30/16 09:32	12/02/16 09:40
320-23971-5	WI-AF-3RW10-1116	Water	11/30/16 10:00	12/02/16 09:40
320-23971-6	WI-AF-3FB10-1116	Water	11/30/16 10:00	12/02/16 09:40
320-23971-7	WI-AF-3RW11-1116	Water	11/30/16 10:30	12/02/16 09:40
320-23971-8	WI-AF-3FB11-1116	Water	11/30/16 10:31	12/02/16 09:40
320-23971-9	WI-AF-3RW12-1116	Water	11/30/16 17:00	12/02/16 09:40
320-23971-10	WI-AF-3FB12-1116	Water	11/30/16 17:01	12/02/16 09:40

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.: _____

Instrument ID: A6

Analysis Batch Number: 140688

Lab Sample ID: STD 320-140688/2 IC

Client Sample ID: _____

Date Analyzed: 12/05/16 17:26

Lab File ID: 05DEC2016A6A_004.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.37	Split Peak	barnettj	12/06/16 10:00
Perfluorooctanoic acid (PFOA)	20.05	Split Peak	barnettj	12/06/16 10:00

Lab Sample ID: STD 320-140688/3 IC

Client Sample ID: _____

Date Analyzed: 12/05/16 17:55

Lab File ID: 05DEC2016A6A_005.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.38	Split Peak	barnettj	12/06/16 10:03
Perfluorooctanoic acid (PFOA)	20.05	Split Peak	barnettj	12/06/16 10:03

Lab Sample ID: CCV 320-140688/9 CCVL

Client Sample ID: _____

Date Analyzed: 12/05/16 20:53

Lab File ID: 05DEC2016A6A_011.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.38	Split Peak	barnettj	12/06/16 10:08
Perfluorooctanoic acid (PFOA)	20.05	Split Peak	barnettj	12/06/16 10:08

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Instrument ID: A6

Analysis Batch Number: 141573

Lab Sample ID: MB 320-140632/1-A

Client Sample ID:

Date Analyzed: 12/11/16 13:31

Lab File ID: 11DEC2016A6A_006.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.05	Split Peak	barnettj	12/12/16 15:00
Perfluorooctanesulfonic acid (PFOS)	20.66	Missed Peak	barnettj	12/12/16 15:00

Lab Sample ID: 320-23971-1

Client Sample ID: WI-AF-2RW07-1116

Date Analyzed: 12/11/16 15:00

Lab File ID: 11DEC2016A6A_009.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanesulfonic acid (PFOS)	20.68	Missed Peak	barnettj	12/12/16 15:22

Lab Sample ID: 320-23971-2

Client Sample ID: WI-AF-2FB07-1116

Date Analyzed: 12/11/16 15:30

Lab File ID: 11DEC2016A6A_010.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanesulfonic acid (PFOS)	20.70	Missed Peak	barnettj	12/12/16 15:23

Lab Sample ID: 320-23971-3

Client Sample ID: WI-AF-2RW08-1116

Date Analyzed: 12/11/16 15:59

Lab File ID: 11DEC2016A6A_011.d

GC Column: Acquity ID: 2.1(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.07	Split Peak	barnettj	12/12/16 15:24
Perfluorooctanesulfonic acid (PFOS)	20.68	Missed Peak	barnettj	12/12/16 15:24

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica SacramentoJob No.: 320-23971-1

SDG No.: _____

Instrument ID: A6Analysis Batch Number: 141573Lab Sample ID: 320-23971-6Client Sample ID: WI-AF-3FB10-1116Date Analyzed: 12/11/16 17:28Lab File ID: 11DEC2016A6A_014.dGC Column: AcquityID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanesulfonic acid (PFOS)	20.67	Missed Peak	barnettj	12/12/16 15:38

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica SacramentoJob No.: 320-23971-1

SDG No.: _____

Instrument ID: A6Analysis Batch Number: 141574Lab Sample ID: 320-23971-8Client Sample ID: WI-AF-3FB11-1116Date Analyzed: 12/11/16 19:56Lab File ID: 11DEC2016A6A_019.dGC Column: AcquityID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanesulfonic acid (PFOS)	20.68	Missed Peak	barnettj	12/12/16 15:41

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
LC537-ICV_00017	01/13/17	08/09/16	MeOH/H2O, Lot 067374	10 mL	LC537-IS_00018	200 uL	13C2-PFOA	10 ng/mL
.LC537-IS_00018	01/13/17	07/13/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00004	100 uL	13C4 PFOS	28.68 ng/mL
.LCM2PFOA_00004	03/19/17	Wellington Laboratories, Lot M2PFOA0312			LCMPFOS_00013	300 uL	13C2-PFOA	0.5 ug/mL
.LCMPFOS_00013	01/22/21	Wellington Laboratories, Lot MPFOS0116			(Purchased Reagent)		13C4 PFOS	1.434 ug/mL
LC537-ICV_00017	01/13/17	08/09/16	MeOH/H2O, Lot 067374	10 mL	LC537-SU_00017	500 uL	13C2 PFDA	50 ug/mL
					LC537ICIM_00013	25 uL	13C2 PFHxA	47.8 ug/mL
.LC537-SU_00017	01/19/17	07/19/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	Perfluorobutanesulfonic acid (PFBS)	10 ng/mL
.LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			LCMPFHxA_00009	100 uL	Perfluoroctanoic acid (PFOA)	114.77 ng/mL
.LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		Perfluoroctanesulfonic acid (PFOS)	25.0965 ng/mL
.LC537ICIM_00013	02/05/17	08/09/16	Methanol, Lot 090285	25 mL	LC537-PFBS2_00005	0.5 mL	Perfluorobutanesulfonic acid (PFBS)	27.2389 ng/mL
					LC537-PFOA2_00007	0.13 mL	Perfluoroctanoic acid (PFOA)	10.2 ug/mL
					LC537-PFOS2_00005	0.22 mL	Perfluoroctanesulfonic acid (PFOS)	10.0386 ug/mL
..LC537-PFBS2_00005	03/01/17	02/29/16	Methanol, Lot 090285	10 mL	LC537_PFBS2_00001	0.023 g	Perfluorobutanesulfonic acid (PFBS)	10.8956 ug/mL
...LC537_PFBS2_00001	08/09/17	Santa Cruz Biotechnology, Lot H0112			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	2295.4 ug/mL
..LC537-PFOA2_00007	07/25/17	08/05/16	Methanol, Lot 090285	10 mL	LC537_PFOA2_00001	0.0195 g	Perfluoroctanoic acid (PFOA)	0.998 g/g
...LC537_PFOA2_00001	07/25/17	Afla Aesar, Lot D24Y026			(Purchased Reagent)		Perfluoroctanoic acid (PFOA)	1930.5 ug/mL
..LC537-PFOS2_00005	03/01/17	02/29/16	Methanol, Lot 090285	10 mL	LC537_PFOS2_00001	0.0159 g	Perfluoroctanesulfonic acid (PFOS)	0.99 g/g
...LC537_PFOS2_00001	07/26/17	Sigma, Lot BCBF5116V			(Purchased Reagent)		Perfluoroctanesulfonic acid (PFOS)	1238.13 ug/mL
LC537-IS_00025	03/19/17	11/21/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00003	100 uL	Perfluoroctanoic acid (PFOA)	0.7787 g/g
.LCM2PFOA_00003	03/19/17	Wellington Laboratories, Lot M2PFOA0312			LCMPFOS_00018	300 uL	Perfluoroctanesulfonic acid (PFOS)	47.8 ug/mL
.LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		Perfluoroctanoic acid (PFOA)	10 ng/mL
LC537-L1_00015	01/13/17	07/28/16	MeOH/H2O, Lot 090285	5 mL	LC537-IS_00018	100 uL	13C2-PFOA	28.68 ng/mL
					LC537-MSP_00012	24.4 uL	13C4 PFOS	10 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	8.76058 ng/mL
							Perfluoroheptanoic acid	0.993847 ng/mL
							Perfluorohexanesulfonic acid	2.9532 ng/mL
							Perfluorononanoic acid	1.91737 ng/mL
							Perfluoroctanoic acid (PFOA)	3.91048 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
.LC537-IS_00018	01/13/17	07/13/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00004	100 uL	13C2_PFHxA	10 ng/mL
					LCMPFOS_00013	300 uL	13C2-PFOA	0.5 ug/mL
..LCM2PFOA_00004	03/19/17		Wellington Laboratories, Lot M2PFOA0312		(Purchased Reagent)		13C4_PFOS	1.434 ug/mL
..LCMPFOS_00013	01/22/21		Wellington Laboratories, Lot MPFOS0116		(Purchased Reagent)		13C2-PFOA	50 ug/mL
..LC537-MSP_00012	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00012	200 uL	Perfluorobutanesulfonic acid (PFBS)	47.8 ug/mL
							Perfluoroheptanoic acid	1795.2 ng/mL
							Perfluorohexanesulfonic acid	203.657 ng/mL
							Perfluorononanoic acid	605.164 ng/mL
							Perfluoroctanoic acid (PFOA)	392.904 ng/mL
							Perfluoroctanesulfonic acid (PFOS)	405.594 ng/mL
							Perfluoroctanesulfonic acid (PFOS)	801.328 ng/mL
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
					LC537-PFHxA_00010	0.1 mL	Perfluoroheptanoic acid	10.1829 ug/mL
					LC537-PFHxS_00008	0.3 mL	Perfluorohexanesulfonic acid	30.2582 ug/mL
					LC537-PFNA_00008	0.2 mL	Perfluorononanoic acid	19.6452 ug/mL
					LC537-PFOA_00009	0.098 mL	Perfluoroctanoic acid (PFOA)	20.2797 ug/mL
					LC537-PFOS_00006	0.4 mL	Perfluoroctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537_PFBS_00002	04/01/18		Sigma, Lot MKBP8842V		(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g
...LC537-PFHxA_00010	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFHxA_00002	0.0072 g	Perfluoroheptanoic acid	1018.29 ug/mL
....LC537_PFHxA_00002	04/01/18		Aldrich, Lot BCBM2579V		(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537_PFHxS_00002	04/01/18		Sigma, Lot BCBL3545V		(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
...LC537_PFNA_00008	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFNA_00002	0.0051 g	Perfluorononanoic acid	982.26 ug/mL
....LC537_PFNA_00002	04/01/18		TCI America, Lot QN44F		(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
...LC537-PFOA_00009	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Perfluoroctanoic acid (PFOA)	2069.36 ug/mL
....LC537_PFOA_00002	11/04/18		Fluka, Lot SZBD308XV		(Purchased Reagent)		Perfluoroctanoic acid (PFOA)	0.999 g/g
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluoroctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17		Fluka, Lot SZBC222XV		(Purchased Reagent)		Perfluoroctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-SU_00017	01/19/17	07/19/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	13C2_PFDA	0.2 ug/mL
					LCMPFHxA_00009	100 uL	13C2_PFHxA	0.2 ug/mL
..LCMPFDA_00008	08/19/20		Wellington Laboratories, Lot MPFDA0815		(Purchased Reagent)		13C2_PFDA	50 ug/mL
..LCMPFHxA_00009	04/09/20		Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)		13C2_PFHxA	50 ug/mL
LC537-L2_00014	01/13/17	07/28/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00010	34 uL	Perfluorobutanesulfonic acid (PFBS)	22.8888 ng/mL
							Perfluoroheptanoic acid	2.59663 ng/mL
							Perfluorohexanesulfonic acid	7.71585 ng/mL
							Perfluorononanoic acid	5.00953 ng/mL
							Perfluoroctanoic acid (PFOA)	5.17132 ng/mL
							Perfluoroctanesulfonic acid (PFOS)	10.2169 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
.LC537-HSP_00010	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537-IS_00018	100 uL	13C2-PFOA	10 ng/mL
					LC537-SU_00017	250 uL	13C4 PFOS	28.68 ng/mL
							13C2 PFDA	10 ng/mL
							13C2 PFHxA	10 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	381.857 ng/mL
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	Perfluorohexanesulfonic acid	1134.68 ng/mL
					LC537-PFHxA_00010	0.1 mL	Perfluorononanoic acid	736.695 ng/mL
					LC537-PFHxS_00008	0.3 mL	Perfluoroctanoic acid (PFOA)	760.489 ng/mL
					LC537-PFNA_00008	0.2 mL	Perfluoroctanesulfonic acid (PFOS)	1502.49 ng/mL
					LC537-PFOA_00009	0.098 mL	Perfluorobutanesulfonic acid	89.76 ug/mL
					LC537-PFOS_00006	0.4 mL	Perfluoroheptanoic acid	10.1829 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorohexanesulfonic acid	30.2582 ug/mL
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorononanoic acid	19.6452 ug/mL
...LC537-PFHxA_00010	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFHxA_00002	0.0072 g	Perfluoroctanoic acid (PFOA)	20.2797 ug/mL
....LC537_PFHxA_00002	04/01/18	Aldrich, Lot BCBM2579V			(Purchased Reagent)		Perfluoroctanesulfonic acid	40.0664 ug/mL
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	Perfluorobutanesulfonic acid (PFBS)	1 g/g
....LC537_PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorobutanesulfonic acid	1018.29 ug/mL
...LC537-PFNA_00008	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFNA_00002	0.0051 g	Perfluorohexanesulfonic acid	0.99 g/g
....LC537_PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorononanoic acid	982.26 ug/mL
...LC537-PFOA_00009	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Perfluorooctanoic acid (PFOA)	0.963 g/g
....LC537_PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	2069.36 ug/mL
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluoroctanesulfonic acid	0.999 g/g
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFOS)	1001.66 ug/mL
.LC537-IS_00018	01/13/17	07/13/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00004	100 uL	13C2-PFOA	0.5 ug/mL
					LCMPFOS_00013	300 uL	13C4 PFOS	1.434 ug/mL
..LCM2PFOA_00004	03/19/17	Wellington Laboratories, Lot M2PFOA0312			(Purchased Reagent)		13C2-PFOA	50 ug/mL
..LCMPFOS_00013	01/22/21	Wellington Laboratories, Lot MPFOS0116			(Purchased Reagent)		13C4 PFOS	47.8 ug/mL
.LC537-SU_00017	01/19/17	07/19/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	13C2 PFDA	0.2 ug/mL
					LCMPFHxA_00009	100 uL	13C2 PFHxA	0.2 ug/mL
..LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFDA	50 ug/mL
..LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	50 ug/mL
LC537-L3_00016	01/28/17	11/07/16	MeOH/H ₂ O, Lot 090285	5 mL	LC537-HSP_00010	67 uL	Perfluorobutanesulfonic acid (PFBS)	45.1044 ng/mL
							Perfluoroheptanoic acid	5.11689 ng/mL
							Perfluorohexanesulfonic acid	15.2048 ng/mL
							Perfluorononanoic acid	9.87171 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration	
					Reagent ID	Volume Added			
.LC537-HSP_00010	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00012	375 uL	Perfluorooctanoic acid (PFOA)	10.1905 ng/mL	
							Perfluorooctanesulfonic acid (PFOS)	20.1334 ng/mL	
							13C2-PFOA	10 ng/mL	
							13C4 PFOS	28.68 ng/mL	
							13C2 PFDA	10 ng/mL	
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	13C2 PFHxA	10 ng/mL	
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL	
							Perfluoroheptanoic acid	381.857 ng/mL	
							Perfluorohexanesulfonic acid	1134.68 ng/mL	
							Perfluorononanoic acid	736.695 ng/mL	
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluoroctanoic acid (PFOA)	760.489 ng/mL	
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL	
							Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL	
							Perfluoroheptanoic acid	10.1829 ug/mL	
							Perfluorohexanesulfonic acid	30.2582 ug/mL	
....LC537_PFBS_00002	04/01/18		Sigma, Lot MKBP8842V		(Purchased Reagent)		Perfluorononanoic acid	19.6452 ug/mL	
							Perfluoroctanoic acid (PFOA)	20.2797 ug/mL	
							Perfluorooctanesulfonic acid	40.0664 ug/mL	
							Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL	
							Perfluorobutanesulfonic acid (PFBS)	1 g/g	
....LC537-PFHpA_00010	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFHpA_00002	0.0072 g	Perfluoroheptanoic acid	1018.29 ug/mL	
							Perfluoroheptanoic acid	0.99 g/g	
....LC537_PFHxS_00008	04/01/18		Aldrich, Lot BCBM2579V		(Purchased Reagent)		Perfluorohexanesulfonic acid	1008.61 ug/mL	
							Perfluorohexanesulfonic acid	0.9094 g/g	
....LC537_PFHxS_00002	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	(Purchased Reagent)	982.26 ug/mL	
							Perfluorononanoic acid	0.963 g/g	
....LC537_PFNNA_00008	04/01/18		Sigma, Lot BCBL3545V		LC537_PFNNA_00002	0.0051 g	(Purchased Reagent)	2069.36 ug/mL	
							Perfluorobutanesulfonic acid (PFBS)	0.999 g/g	
....LC537_PFNNA_00002	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFNNA_00002	0.0051 g	(Purchased Reagent)	1001.66 ug/mL	
							Perfluorobutanesulfonic acid (PFBS)	0.9106 g/g	
....LC537_PFOA_00009	04/01/18		TCI America, Lot QN44F		(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	50 ug/mL	
							Perfluorooctanoic acid (PFOA)	47.8 ug/mL	
....LC537_PFOA_00002	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	(Purchased Reagent)	0.5 ug/mL	
							Perfluorooctanoic acid (PFOA)	1.434 ug/mL	
....LC537_PFOA_00006	11/04/18		Fluka, Lot SZBD308XV		(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.2 ug/mL	
							Perfluorooctanoic acid (PFOA)	0.2 ug/mL	
....LC537_PFOA_00002	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOA_00002	0.0066 g	(Purchased Reagent)	0.9106 g/g	
							Perfluorooctanesulfonic acid (PFOS)	50 ug/mL	
....LC537_IS_00024	03/19/17	11/03/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00003	100 uL	13C2-PFOA	50 ug/mL	
							13C4 PFOS	47.8 ug/mL	
....LCMPFOS_00018	08/03/21		Wellington Laboratories, Lot M2PFOA0312		(Purchased Reagent)		13C2-PFOA	0.5 ug/mL	
							13C4 PFOS	47.8 ug/mL	
....LC537-SU_00020	04/07/17	10/07/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	13C2 PFDA	0.2 ug/mL	
							13C2 PFHxA	0.2 ug/mL	
....LCMPFDA_00008	08/19/20		Wellington Laboratories, Lot MPFDA0815		(Purchased Reagent)		13C2 PFDA	50 ug/mL	
							13C2 PFHxA	50 ug/mL	
....LCMPFHxA_00009	04/09/20		Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)		13C2 PFHxA	50 ug/mL	
							13C2 PFHxA	50 ug/mL	
LC537-L4_00015	01/13/17	07/28/16	MeOH/H ₂ O, Lot 090285	5 mL	LC537-HSP_00010	135 uL	Perfluorobutanesulfonic acid (PFBS)	90.882 ng/mL	

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
.LC537-HSP_00010	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00012	375 uL	Perfluoroheptanoic acid	10.3101 ng/mL
							Perfluorohexanesulfonic acid	30.6364 ng/mL
							Perfluorononanoic acid	19.8908 ng/mL
							Perfluoroctanoic acid (PFOA)	20.5332 ng/mL
							Perfluoroctanesulfonic acid (PFOS)	40.5672 ng/mL
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	381.857 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	736.695 ng/mL
							Perfluoroctanoic acid (PFOA)	760.489 ng/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBs_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
							Perfluoroheptanoic acid	2040 ug/mL
							Perfluorohexanesulfonic acid	1 g/g (PFBS)
							Perfluorononanoic acid	0.99 g/g
							Perfluoroctanoic acid (PFOA)	0.9094 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFHxS_00002	0.0072 g	Perfluoroheptanoic acid	1018.29 ug/mL
....LC537_PFHxS_00002	04/01/18		Aldrich, Lot BCBM2579V			(Purchased Reagent)	Perfluoroheptanoic acid	0.99 g/g
....LC537_PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537_PFHxS_00002	04/01/18		Sigma, Lot BCBL3545V			(Purchased Reagent)	Perfluorohexanesulfonic acid	0.9094 g/g
....LC537_PFNNA_00008	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFNNA_00002	0.0051 g	Perfluorononanoic acid	982.26 ug/mL
....LC537_PFNNA_00002	04/01/18		TCI America, Lot QN44F			(Purchased Reagent)	Perfluorononanoic acid	0.963 g/g
....LC537_PFOA_00009	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Perfluoroctanoic acid (PFOA)	2069.36 ug/mL
....LC537_PFOA_00002	11/04/18		Fluka, Lot SZBD308XV			(Purchased Reagent)	Perfluoroctanoic acid (PFOA)	0.999 g/g
....LC537_PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluoroctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17		Fluka, Lot SZBC222XV			(Purchased Reagent)	Perfluoroctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-IS_00018	01/13/17	07/13/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00004	100 uL	13C2-PFOA	0.5 ug/mL
					LCMPFOS_00013	300 uL	13C4 PFOS	1.434 ug/mL
..LCM2PFOA_00004	03/19/17		Wellington Laboratories, Lot M2PFOA0312			(Purchased Reagent)	13C2-PFOA	50 ug/mL
..LCMPFOS_00013	01/22/21		Wellington Laboratories, Lot MPFOS0116			(Purchased Reagent)	13C4 PFOS	47.8 ug/mL
.LC537-SU_00017	01/19/17	07/19/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	13C2 PFDA	0.2 ug/mL
					LCMPFHxA_00009	100 uL	13C2 PFHxA	0.2 ug/mL
..LCMPFDA_00008	08/19/20		Wellington Laboratories, Lot MFDA0815			(Purchased Reagent)	13C2 PFDA	50 ug/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
..LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)		13C2 PFHxA		50 ug/mL
LC537-L5_00017	01/28/17	11/07/16	MeOH/H ₂ O, Lot 090285	5 mL	LC537-HSP_00010	200 uL	Perfluorobutanesulfonic acid (PFBS)	134.64 ng/mL
					LC537-IS_00024	100 uL	Perfluoroheptanoic acid	15.2743 ng/mL
					LC537-SU_00020	250 uL	Perfluorohexanesulfonic acid	45.3873 ng/mL
							Perfluorononanoic acid	29.4678 ng/mL
LC537-HSP_00010	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00012	375 uL	Perfluoroheptanoic acid (PFOA)	30.4196 ng/mL
							Perfluorooctanoic acid (PFOS)	60.0996 ng/mL
							13C2-PFOA	10 ng/mL
							13C4 PFOS	28.68 ng/mL
							13C2 PFDA	10 ng/mL
							13C2 PFHxA	10 ng/mL
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	381.857 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	736.695 ng/mL
							Perfluoroctanoic acid (PFOA)	760.489 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V		(Purchased Reagent)			Perfluorobutanesulfonic acid (PFBS)	1 g/g
....LC537-PFHxA_00010	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFHxA_00002	0.0072 g	Perfluoroheptanoic acid	1018.29 ug/mL
....LC537_PFHxA_00002	04/01/18	Aldrich, Lot BCBM2579V		(Purchased Reagent)			Perfluoroheptanoic acid	0.99 g/g
....LC537-PFHxA_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxA_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537_PFHxA_00002	04/01/18	Sigma, Lot BCBL3545V		(Purchased Reagent)			Perfluorohexanesulfonic acid	0.9094 g/g
....LC537-PFNA_00008	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFNA_00002	0.0051 g	Perfluorononanoic acid	982.26 ug/mL
....LC537_PFNA_00002	04/01/18	TCI America, Lot QN44F		(Purchased Reagent)			Perfluorononanoic acid	0.963 g/g
....LC537-PFOA_00009	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Perfluoroctanoic acid (PFOA)	2069.36 ug/mL
....LC537_PFOA_00002	11/04/18	Fluka, Lot SZBD308XV		(Purchased Reagent)			Perfluoroctanoic acid (PFOA)	0.999 g/g
....LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV		(Purchased Reagent)			Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
..LC537-IS_00024	03/19/17	11/03/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00003	100 uL	13C2-PFOA	0.5 ug/mL
..LCM2PFOA_00003	03/19/17	Wellington Laboratories, Lot M2PFOA0312		(Purchased Reagent)			13C4 PFOS	1.434 ug/mL
..LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816		(Purchased Reagent)			13C2-PFOA	50 ug/mL
..LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816		(Purchased Reagent)			13C4 PFOS	47.8 ug/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration		
					Reagent ID	Volume Added				
.LC537-SU_00020	04/07/17	10/07/16	Methanol, Lot 104453	25000 uL	LCMPFDA_00008	100 uL	13C2 PFDA	0.2 ug/mL		
					LCMPFHxA_00009	100 uL	13C2 PFHxA	0.2 ug/mL		
..LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFDA	50 ug/mL		
..LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	50 ug/mL		
LC537-L6_00014	01/13/17	07/28/16	MeOH/H ₂ O, Lot 090285	5 mL	LC537-HSP_00010	265 uL	Perfluorobutanesulfonic acid (PFBS)	178.398 ng/mL		
							Perfluoroheptanoic acid	20.2384 ng/mL		
							Perfluorohexanesulfonic acid	60.1382 ng/mL		
					LC537-IS_00018	100 uL	Perfluorononanoic acid	39.0448 ng/mL		
							Perfluoroctanoic acid (PFOA)	40.3059 ng/mL		
					LC537-SU_00017	250 uL	Perfluoroctanesulfonic acid (PFOS)	79.632 ng/mL		
LC537-HSP_00010	01/28/17	07/28/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00012	375 uL	13C2-PFOA	10 ng/mL		
							13C4 PFOS	28.68 ng/mL		
							13C2 PFDA	10 ng/mL		
					LC537-SU_00017	250 uL	13C2 PFHxA	10 ng/mL		
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL		
							Perfluoroheptanoic acid	381.857 ng/mL		
..LC537SPIM_00012	01/28/17	07/28/16	Methanol, Lot 104453	10 mL	LC537-PFBS_00006	0.44 mL	Perfluorohexanesulfonic acid	1134.68 ng/mL		
							Perfluorononanoic acid	736.695 ng/mL		
							Perfluoroctanoic acid (PFOA)	760.489 ng/mL		
					LC537-PFOS_00006	0.4 mL	Perfluoroctanesulfonic acid (PFOS)	1502.49 ng/mL		
							Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL		
							Perfluoroheptanoic acid	10.1829 ug/mL		
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537-PFHxA_00010	0.1 mL	Perfluorohexanesulfonic acid	30.2582 ug/mL		
							Perfluorononanoic acid	19.6452 ug/mL		
							Perfluoroctanoic acid (PFOA)	20.2797 ug/mL		
					LC537-PFNA_00008	0.098 mL	Perfluoroctanesulfonic acid (PFOS)	40.0664 ug/mL		
							Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL		
							Perfluoroheptanoic acid	0.99 g/g		
....LC537_PFBs_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorohexanesulfonic acid	1018.29 ug/mL		
					LC537-PFHxA_00002	0.0072 g	Perfluoroheptanoic acid	0.99 g/g		
							Perfluorohexanesulfonic acid	1008.61 ug/mL		
					LC537-PFNA_00008	0.0061 g	Perfluorobutanesulfonic acid	0.9094 g/g		
							Perfluorononanoic acid	982.26 ug/mL		
							Perfluoroctanoic acid	0.963 g/g		
....LC537_PFHxA_00002	04/01/18	Aldrich, Lot BCBM2579V			(Purchased Reagent)		Perfluoroctanoic acid (PFOA)	2069.36 ug/mL		
					LC537-PFNA_00002	0.0051 g	Perfluorobutanesulfonic acid	0.999 g/g		
							Perfluorononanoic acid	1001.66 ug/mL		
					LC537-PFOS_00006	0.0066 g	Perfluoroheptanoic acid (PFOS)	0.9106 g/g		
							Perfluoroctanesulfonic acid (PFOS)	0.5 ug/mL		
							Fluka, Lot SZBD308XV	0.5 ug/mL		
....LC537_PFOA_00009	07/28/17	07/28/16	Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Fluka, Lot SZBC222XV	0.999 g/g		
							(Purchased Reagent)	0.999 g/g		
....LC537_PFOA_00002	11/04/18	TCI America, Lot QN44F			(Purchased Reagent)		Fluka, Lot SZBC222XV	0.999 g/g		
					LC537_PFOA_00002	0.0066 g	(Purchased Reagent)	0.999 g/g		
....LC537_PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Fluka, Lot SZBC222XV	0.999 g/g		
							(Purchased Reagent)	0.999 g/g		
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Fluka, Lot SZBC222XV	0.999 g/g		
					LC537_PFOA_00004	100 uL	(Purchased Reagent)	0.999 g/g		
....LC537_IS_00018	01/13/17	07/13/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00004	100 uL	13C2-PFOA	0.5 ug/mL		

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration		
					Reagent ID	Volume Added				
				LCMPFOS_00013	300 uL	13C4 PFOS	1.434 ug/mL			
..LCM2PFOA_00004	03/19/17	Wellington Laboratories, Lot M2PFOA0312		(Purchased Reagent)	13C2-PFOA	50 ug/mL				
..LCMPFOS_00013	01/22/21	Wellington Laboratories, Lot MPFOS0116		(Purchased Reagent)	13C4 PFOS	47.8 ug/mL				
.LC537-SU_00017	01/19/17	07/19/16 Methanol, Lot 104453	25000 uL	LCMPFDA_00008 LCMPFHxA_00009	100 uL 100 uL	13C2 PFDA 13C2 PFHxA	0.2 ug/mL 0.2 ug/mL			
..LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815		(Purchased Reagent)	13C2 PFDA	50 ug/mL				
..LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)	13C2 PFHxA	50 ug/mL				
LC537-MSP_00014	03/14/17	09/14/16 Methanol, Lot 090285	10000 uL	LC537SPIM_00013	200 uL	Perfluorobutane Sulfonate	1795.2 ng/mL			
						Perfluorobutanesulfonic acid (PFBS)	1795.2 ng/mL			
						Perfluoroheptanoic acid	203.657 ng/mL			
						Perfluorohexanesulfonic acid	605.164 ng/mL			
						Perfluorononanoic acid	392.904 ng/mL			
						Perfluoroctanoic acid (PFOA)	405.594 ng/mL			
						Perfluoroctanesulfonic acid (PFOS)	801.328 ng/mL			
.LC537SPIM_00013	03/14/17	09/14/16 Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutane Sulfonate	89760 ng/mL			
						Perfluorobutanesulfonic acid (PFBS)	89760 ng/mL			
						Perfluoroheptanoic acid	10182.9 ng/mL			
						Perfluorohexanesulfonic acid	30258.2 ng/mL			
						Perfluorononanoic acid	19645.2 ng/mL			
						Perfluoroctanoic acid (PFOA)	20279.7 ng/mL			
..LC537-PFBS_00006	07/28/17	07/28/16 Methanol, Lot 090285	5 mL	LC537_PFBs_00002	0.0102 g	Perfluorobutane Sulfonate	2040 ug/mL			
						Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL			
...LC537_PFBs_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorobutane Sulfonate	1 g/g		
							Perfluorobutanesulfonic acid (PFBS)	1 g/g		
..LC537-PFHxA_00010	07/28/17	07/28/16 Methanol, Lot 090285	7 mL	LC537_PFHxA_00002	0.0072 g	Perfluoroheptanoic acid	1018.29 ug/mL			
...LC537_PFHxA_00002	04/01/18	Aldrich, Lot BCBM2579V		(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g			
..LC537-PFHxS_00008	07/28/17	07/28/16 Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL			
...LC537_PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V		(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g			
..LC537-PFNA_00008	07/28/17	07/28/16 Methanol, Lot 090285	5 mL	LC537_PFNA_00002	0.0051 g	Perfluorononanoic acid	982.26 ug/mL			
...LC537_PFNA_00002	04/01/18	TCI America, Lot QN44F		(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g			
..LC537-PFOA_00009	07/28/17	07/28/16 Methanol, Lot 090285	7 mL	LC537_PFOA_00002	0.0145 g	Perfluoroctanoic acid (PFOA)	2069.36 ug/mL			
...LC537_PFOA_00002	11/04/18	Fluka, Lot SZBD308XV		(Purchased Reagent)		Perfluoroctanoic acid (PFOA)	0.999 g/g			
..LC537-PFOS_00006	07/28/17	07/28/16 Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluoroctanesulfonic acid (PFOS)	1001.66 ug/mL			
...LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV		(Purchased Reagent)		Perfluoroctanesulfonic acid (PFOS)	0.9106 g/g			
LC537-SU_00022	05/21/17	11/21/16 Methanol, Lot 104453	20000 uL	LCMPFDA_00008 LCMPFHxA_00009	80 uL 80 uL	13C2 PFDA	0.2 ug/mL			
						13C2 PFHxA	0.2 ug/mL			
..LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815		(Purchased Reagent)	13C2 PFDA	50 ug/mL				
..LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)	13C2 PFHxA	50 ug/mL				

Reagent

LC537_PFBS_00002

C: 4/1/15 SPV

SIGMA-ALDRICH®

sigma-aldrich.com

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com


Outside USA: eurtechserv@sial.com

Product Name:

Certificate of Analysis

Nonafluorobutane-1-sulfonic acid - 97%

Product Number: 562629
Batch Number: MKBP8842V
Brand: ALDRICH
CAS Number: 375-73-5
MDL Number: MFCD01320794
Formula: C₄H₉O₃S
Formula Weight: 300.10 g/mol
Storage Temperature: Store at 2 - 8 °C
Quality Release Date: 11 OCT 2013



PFBS

Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Fluorine NMR Spectrum	Conforms to Structure	Conforms
Purity (Titration by NaOH)	96.5 - 103.5 %	101.6 %


Jamie Gleason

Jamie Gleason, Manager
 Quality Control
 Milwaukee, Wisconsin US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFB2_00001



The Power to Question

CERTIFICATE OF ANALYSIS

Catalog Number: sc-236187
Product Name: Nonanfluorobutane-1-sulfonic acid
CAS Number: 375-73-5
Molecular Formula: C₄H₉F₉O₃S
Molecular Weight: 300.10
Lot Number: H0112

Test	Result
Refractive Index	1.3200 to 1.3290
Purity (Titration)	min. 98.0%

Test Conditions: Refractive Index: n_{20/D}

Reagent

LC537_PFHpA_00002

R: 4/1/15 SV

SIGMA-ALDRICH®

3050 Spruce Street, Saint Louis, MO 63103 USA
Email USA: techserv@sial.com Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name: PERFLUOROHEPTANOIC ACID
99 %
Product Number: 342041
Batch Number: BCBM2579V
Brand: Aldrich
CAS Number: 375-85-9
Formula: $\text{CF}_3(\text{CF}_2)_5\text{CO}_2\text{H}$
Formula Weight: 364.06
Quality Release Date: 06 DEC 2013
Recommended Retest Date: OCT 2018

PFH₇A

TEST	SPECIFICATION	RESULT
APPEARANCE (COLOR)	COLORLESS OR WHITE	WHITE
APPEARANCE (FORM)	LIQUID OR SOLID	SOLID
TITRATION	98.5 - 101.5 %	99.8 %
TITRATION (METHOD)	-	BACK TITRATION
PURITY (GC AREA %)	≥ 98.5 %	99.5 %
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS



Dr. Claudia Geitner
Manager Quality Control
Buchs, Switzerland

Sigma-Aldrich warrants that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFHxS_00002

R: 4/11/15 SW

SIGMA-ALDRICH®3050 Spruce Street, Saint Louis, MO 63103 USA
Email USA: techserv@sial.com Outside USA: eurtechserv@sial.com**Certificate of Analysis**

Product Name: TRIDECAFLUOROHEXANE-1-SULFONIC ACID POTASSIUM SALT
Spec: >= 98.0 % T
Product Number: 50929
Batch Number: BCBL3545V
Brand: Aldrich
CAS Number: 3871-99-6
Formula: C₆F₁₃KO₃S
Formula Weight: 438.20
Quality Release Date: 20 JUN 2013

PFH_xS-K

TEST	SPECIFICATION	RESULT
APPEARANCE (COLOR)	WHITE TO FAINT BEIGE	WHITE
APPEARANCE (FORM)	POWDER OR CRYSTALS	POWDER
TITRATION (ION EXCHANGE)	≥ 98.0 %	99.5 %
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS

Dr. Claudia Geitner
 Manager Quality Control
 Buchs, Switzerland

$$\text{MW corr} = \frac{(k_{\text{form}}) - (k) + (n)}{(438.20 - 391.0 + 1.0)} = 0.91307 \quad (\text{anion form})$$

Purity = 90.94 % w/m.w correction

SW 4/11/15

Sigma-Aldrich warrants that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFNA_00002

R: 4/1/15 SKV



Certificate of Analysis

Apr 2, 2015 (JST)

TOKYO CHEMICAL INDUSTRY CO.,LTD.
4-10-1 Nihonbashi-Honcho, Chuo-ku, Tokyo 103-0023 Japan

Chemical Name: Heptadecafluororonanoic Acid		
Product Number: H0843 CAS: 375-95-1	Lot: QN44F	

Tests	Results	Specifications
Purity(GC)	96.3 %	min. 95.0 %
Purity(Neutralization titration)	98.1 %	min. 95.0 %
Melting point	63.3 deg-C	62.0 to 67.0 deg-C

TCI Lot numbers are 4-5 characters in length.
Characters listed after the first 4-5 characters are control numbers for internal purpose only.

Customer service:

TCI AMERICA
Tel: +1-800-423-8616 / +1-503-283-1681
Fax: +1-866-520-1075 / +1-503-283-1987
E-mail: Sales-US@TCIchemicals.com

PFNA

Reagent

LC537_PFOA_00002

13/21/15 PV

SIGMA-ALDRICH®

CERTIFICATE OF ANALYSIS

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze
Telefon: +49 5137 8238-150

Seelze, 13.11.2013/505378/13/24029

Order-No.:

Customer-No.:

Order-Code:

Quantity:

Production Date: 04.Nov.2013

Expiry Date: 04.Nov.2018

Article/Product: 33824

Batch : SZBD308XV

Pentadecafluorooctanoic acid OEKANAL®

PFCA

Reference Material (RM)

1. General Information

Formula: C₈HF₁₅O₂

Molar mass: 414.07 g/Mole

CAS-No.: [335-67-1]

Recomm. storage temp.: roomtemp.

Usage : PFOA

The estimated uncertainty of a single measurement of the assay can be expected to be 0.5 % relative (confidence level = 95%, n= 6) whereby the assay measurements are calculated by 100% minus found impurities.

2. Batch Analysis

Identify (GC-MS)

complying

Assay (GCMS)

99.4 %

Date of Analysis

13.Nov.2013

3. Advice and Remarks

- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- Whenever the container is opened for removal of aliquot portions of the substance, the person handling the substance must assure, that the integrity of the substance is maintained and proper records of all its handlings are kept. Special care has to be taken to avoid any contamination or adulteration of the substance.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Sigma-Aldrich Laborchemikalien GmbH
Quality Management SA-LC

GC/MS-Method**Analytical Department****Article:** Pentadecafluoroctanoic acid OEKANAL**Article-No.:** 33824**Batch:** SZBD308XV

Column: XTI-5 (Restek); 30 m; fs cap.; I.D.:0.25 mm; 1 µm df

Injector: Split mode

Injection: approx. 1 µl of reaction mixture with MSTFA (approx. 10 mg + 200 µl MSTFA)

Inj.-temp.: 280°C

Oven-temp.: 40°C (for 2 min) to 320°C (6°C/min) hold for 2 min

Split: 1:100

Flow: 1 ml He/min (Constant flow mode)


Detector: MSD

Mass range: 10-600 amu (Scan mode)


Evaluation: Purity: Total Ion Chromatogram
(MSTFA and reaction products blinded out in report)

Identity: Mass spectrum complies

Operator: Ahrens / 2013-11-13

Total Ion Chromatogram:

Ret.time Area Area-% Com

11.54 565.1670 99.4 Pentadecafluoroctanoic acid (as TMS-ester)
11.95 3.6792 0.64**Mass spectrum (rt = 11.54 min):**

Reagent

LC537_PFOA2_00001

Certificate of Analysis

Alfa Aesar®
A Johnson Matthey Company

Product No.: L08862

Product: Perfluorooctanoic acid, 95%

PFOA

Lot No.: D24Y026

Appearance White solid

Melting point 58 - 60°C

Assay 99 %

Identity Matches reference

This document has been electronically generated and does not require a signature.

www.alfa.com

NORTH AMERICA
Tel: +1-800-343-0660 or
+1-978-521-6300
Fax: +1-800-322-4757
Email: info@alfa.com

GERMANY
Tel: 00800 4566 4566 or
+49 721 84007 280
Fax: 00800 4577 4577 or
+49 721 84007 300
.Email: Eurosales@alfa.com

UNITED KINGDOM
Tel: 0800-801812 or
+44 (0)1524-850506
Fax: +44 (0)1524-850608
Email: UKsales@alfa.com

FRANCE
Tel: 0800 03 51 47 or
+33 (0)3 8862 2690
Fax: 0800 10 20 67 or
+33 (0)3 8862 6864
Email: frventes@alfa.com

INDIA
Tel: +91 8008 812424 or
+91 8008 812525 or
+91 8008 812626
Fax: +91 8418 260060
Email: India@alfa.com

CHINA
Tel: +86 (010) 8567-8600
Fax: +86 (010) 8567-8601
Email: saleschina@alfa-asia.com

KOREA
Tel: +82-2-3140-6000
Fax: +82-2-3140-6002
Email: saleskorea@alfa-asia.com

Reagent

LC537_PFOS_00002

F: 4/115 SV

SIGMA-ALDRICH®

CERTIFICATE OF ANALYSIS

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze
Telefon: +49 5137 8238-150

Seelze, 13.08.2012/419060/12/17583

Order-No.:

Customer-No.:

Order-Code:

Quantity:

Production Date: 09.Aug.2012

Expiry Date: 09.Aug.2017 - ~~exp date~~

Article/Product: 33829

Batch : SZBC222XV

Heptadecafluorooctanesulfonic acid potassium salt OEKANAL®

PFOS-K+

Reference Material (RM)

1. General Information

Formula: C₈F₁₇KO₃S

Molar mass: 538.22 g/Mole

CAS-No.: [2795-39-3]

Recomm. storage temp.: roomtemp.

Usage : PFOS

The estimated uncertainty of a single measurement of the assay can be expected to be 0.5 % relative (confidence level = 95%, n= 6) whereby the assay measurements are calculated by 100% minus found impurities.

2. Batch Analysis

Identity

complying

Assay (LC-MS)

98 %

Date of Analysis

10.Aug.2012

W-Correction: $\frac{538.22 - 39.10 + 1.01}{538.22} = \frac{500.13}{538.22} = 0.92923$

Purity: 91.06 %

3. Advice and Remarks

- The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- Whenever the container is opened for removal of aliquot portions of the substance, the person handling the substance must assure, that the integrity of the substance is maintained and proper records of all its handlings are kept. Special care has to be taken to avoid any contamination or adulteration of the substance.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Sigma-Aldrich Laborchemikalien GmbH
Quality Management SA-LC

Reagent

LC537_PFOS2_00001

Certificate of Analysis

Inv 820
12LCMS 0579

Product Name: HEPTADECAFLUOROOCTANESULFONIC ACID TETRAETHYLMAMMONIUM SALT
98 %

Product Number: 365289

Product Brand: Aldrich

Molecular Formula: C₁₆H₂₀F₁₇NO₃S

Molecular Mass: 629.37

CAS Number: 56773-42-3

TEST	SPECIFICATION	LOT BCBF5116V RESULTS
APPEARANCE (COLOR)	OFF-WHITE TO WHITE	WHITE
APPEARANCE (FORM)	POWDER, LUMPS OR CHUNKS	POWDER WITH LUMPS
CARBON CONTENT	29.77 % - 31.29 %	30.52
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS


QC RELEASE DATE 13/APR/11

$$\text{MW correction} = \frac{500.125}{629.37} = 0.7946$$

= 79.46% 04/11/12

Purity + MW Correction = 77.87%

E. Schwärzler
Edeltraud Schwärzler, Manager
Quality Control
Buchs, Switzerland



<u>C₈F₁₇SO₃H</u>		<u>C₈H₂₀N</u>
C = 12.011	96.088	96.088
F = 18.998	322.966	-
S = 32.066	32.066	-
O = 15.999	47.997	-
H = 1.008	1.008	20.160
N = 14.007	-	14.007
	<u>500.125</u>	<u>130.255</u> →

Sigma-Aldrich warrants, that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice for additional terms and conditions of sale. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

Certificate of Origin

Product Name: Heptadecafluoroctanesulfonic acid tetraethylammonium salt
98 %

Product Number: 365289

Product Brand: Aldrich

Lot: BCBF5116V

Molecular Formula: C₁₆H₂₀F₁₇NO₃S

Molecular Mass: 629.37

CAS Number: 56773-42-3

Date of Issue: 30-MAR-11

Country of Origin China

product is of synthetic origin yes

only synthetic materials used in the manufacturing process yes

compounds of animal origin used no

genetically modified organisms used no

allergenic materials used no

procedures in place to avoid cross contamination with residue of animal, human, GMO or allergenes in manufacturing process yes

Sigma-Aldrich has quality systems and procedures in place for monitoring the production process, traceability and batch consistency.

Document issued by Sigma-Aldrich Corporation "Sigma-Aldrich". This document is valid without signature and has been produced digitally.

This information is to be used for the purpose of determining animal or other biological origin only and not to be confused with "Country of Origin" for import/export purposes. Data provided on this document are property of Sigma-Aldrich.

This information is considered accurate and reliable as of the date appearing on the document and is presented in good faith.

Sigma-Aldrich shall not be held liable for any damage resulting from handling or from processing the above product(s). This document does not make any warranty, express or implied, of fitness for any particular use of the product(s). Purchaser must determine the suitability of the product(s) for its use under the applicable law and regulations.

For further questions please contact your local Sigma-Aldrich representative.

We are committed to the success of our Customers, Employees and Shareholders through leadership in Life Science, High Technology and Service.

Reagent

LCM2PFOA_00003



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:

M2PFOA


LOT NUMBER: M2PFOA0312

COMPOUND:

Perfluoro-n-[1,2-¹³C₂]octanoic acid

STRUCTURE:

CAS #: Not available



MOLECULAR FORMULA:

¹³C₂¹²C₆HF₁₅O₂

MOLECULAR WEIGHT: 416.05

CONCENTRATION:

50 ± 2.5 µg/ml

SOLVENT(S): Methanol

CHEMICAL PURITY:

>98%

ISOTOPIC PURITY: >99% ¹³C

LAST TESTED: (mm/dd/yyyy)

03/19/2012

(1,2-¹³C₂)

EXPIRY DATE: (mm/dd/yyyy)

03/19/2017

RECOMMENDED STORAGE: Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 01/09/2013

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. They are designed to be used as reference standards for the identification and/or quantification of specific chemical compound(s).

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Material Safety Data Sheets (MSDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product, unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, x-ray crystallography and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS and/or LC/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external, ISO/IEC 17025:2005 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration for the period of time specified by the expiry date in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.


QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to ISO 9001:2008 by SAI Global, ISO/IEC 17025:2005 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34:2009 by ACCLASS (certificate number AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: M2PFOA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro micro API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm


Mobile phase: Gradient
Start: 60% (80:20 MeOH:ACN) / 40% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 6.5 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min

Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (250 - 850 amu)
Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: M2PFOA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml M2PFOA)

MS Parameters

Collision Gas (mbar) = 3.35e-3
Collision Energy (eV) = 11

Mobile phase: Isocratic 70% (80:20 MeOH:ACN) / 30% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

Reagent

LCM2PFOA_00004



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:

M2PFOA


LOT NUMBER: M2PFOA0312

COMPOUND:

Perfluoro-n-[1,2-¹³C₂]octanoic acid

STRUCTURE:

CAS #: Not available



MOLECULAR FORMULA:

¹³C₂¹²C₆HF₁₅O₂

MOLECULAR WEIGHT: 416.05

CONCENTRATION:

50 ± 2.5 µg/ml

SOLVENT(S): Methanol

CHEMICAL PURITY:

>98%

ISOTOPIC PURITY: >99% ¹³C

LAST TESTED: (mm/dd/yyyy)

03/19/2012

(^{1,2-13}C₂)

EXPIRY DATE: (mm/dd/yyyy)

03/19/2017

RECOMMENDED STORAGE: Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 01/09/2013

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. They are designed to be used as reference standards for the identification and/or quantification of specific chemical compound(s).

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Material Safety Data Sheets (MSDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product, unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, x-ray crystallography and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS and/or LC/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external, ISO/IEC 17025:2005 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration for the period of time specified by the expiry date in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.


QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to ISO 9001:2008 by SAI Global, ISO/IEC 17025:2005 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34:2009 by ACCLASS (certificate number AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: M2PFOA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro micro API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 60% (80:20 MeOH:ACN) / 40% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 6.5 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (250 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: M2PFOA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml M2PFOA)

MS Parameters

Collision Gas (mbar) = 3.35e-3
Collision Energy (eV) = 11

Mobile phase: Isocratic 70% (80:20 MeOH:ACN) / 30% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

Reagent

LCMPFDA_00008



605243

ID: LCMPFDA_00008

Exp: 08/19/20 Prod: CBW

13C2-Perfluorodecanoic a

Rec. 3/29/16 JEB ✓



WELLINGTON
LABORATORIES


**CERTIFICATE OF ANALYSIS
DOCUMENTATION**

PRODUCT CODE:

MPFDA

LOT NUMBER: MPFDA0815**COMPOUND:**Perfluoro-n-[1,2-¹³C₂]decanoic acid**STRUCTURE:****CAS #:**

Not available

**MOLECULAR FORMULA:**¹³C₂¹²C₈HF₁₉O₂**CONCENTRATION:**

50 ± 2.5 µg/ml

MOLECULAR WEIGHT: 516.07**CHEMICAL PURITY:**

>98%

SOLVENT(S): Methanol**LAST TESTED:** (mm/dd/yyyy)

08/19/2015

Water (<1%)

EXPIRY DATE: (mm/dd/yyyy)

08/19/2020

ISOTOPIC PURITY: >99% ¹³C**RECOMMENDED STORAGE:**

Store ampoule in a cool, dark place

(1,2-¹³C₂)**DOCUMENTATION/ DATA ATTACHED:**

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.
- Contains < 0.1% of ¹³C₁-PFNA.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 08/21/2015

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFDA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acuity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acuity UPLC BEH Shield RP₁₈
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (150 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFDA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml MPFDA)

MS Parameters


Collision Gas (mbar) = 3.35e-3
Collision Energy (eV) = 13

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

Reagent

LCMPFHxA_00009



WELLINGTON
LABORATORIES



605244

ID: LCMPFHxA_00009

Exp: 04/09/20 Prp: CBW

13C2-Perfluorohexanoic ac

Rec. 3/29/16 JRB ✓

CERTIFICATE OF ANALYSIS
DOCUMENTATION

PRODUCT CODE:

MPFHxA


COMPOUND:

Perfluoro-n-[1,2-¹³C₂]hexanoic acid

LOT NUMBER: MPFHxA0415

STRUCTURE:

CAS #: Not available



MOLECULAR FORMULA:

¹³C₂¹²C₄HF₁₁O₂

CONCENTRATION:

50 ± 2.5 µg/ml

MOLECULAR WEIGHT: 316.04

CHEMICAL PURITY:

>98%

SOLVENT(S): Methanol

LAST TESTED: (mm/dd/yyyy)

04/09/2015

Water (<1%)

EXPIRY DATE: (mm/dd/yyyy)

04/09/2020

ISOTOPIC PURITY: >99%¹³C

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

(1,2-¹³C₂)

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.
- Contains < 0.1% of perfluoro-n-hexanoic acid and ~ 0.3% of perfluoro-n-octanoic acid.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 04/14/2015

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:


Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.


QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFHxA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acuity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acuity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 2 min
before returning to initial conditions over 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFHxA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml MPFHxA)

MS Parameters

Collision Gas (mbar) = 3.20e-3
Collision Energy (eV) = 10

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

Reagent

LCMPFOS_00013



WELLINGTON LABORATORIES



606227

ID: LCMPFOS_00012
Exp: 01/22/21 Prd: CBW
13C4-Perfluoroctanesulfonate

Rec. 3/29/16 JRB ✓



606228

ID: LCMPFOS_00013
Exp: 01/22/21 Prd: CBW
13C4-Perfluoroctanesulfonate


CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:

MPFOS

COMPOUND:Sodium perfluoro-1-[1,2,3,4-¹³C₄]octanesulfonateLOT NUMBER: MPFOS0116STRUCTURE:CAS #:

Not available

MOLECULAR FORMULA:¹³C₄¹²C₄F₁₇SO₃NaCONCENTRATION:

50.0 ± 2.5 µg/ml (Na salt)

47.8 ± 2.4 µg/ml (MPFOS anion)

CHEMICAL PURITY:

>98%

LAST TESTED: (mm/dd/yyyy)

01/22/2016

EXPIRY DATE: (mm/dd/yyyy)

01/22/2021

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

MOLECULAR WEIGHT: 526.08SOLVENT(S): MethanolISOTOPIC PURITY: ≥99% ¹³C
(1,2,3,4-¹³C₄)DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains ~ 0.8% Sodium perfluoro-1-[1,2,3-¹³C₃]heptanesulfonate.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 02/01/2016

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(x_i, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.


QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFOS; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acuity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acuity UPLC BEH Shield RP₁₈
1.7 µm, 2.1 x 100 mm

Mobile phase: Gradient
Start: 55% (80:20 MeOH:ACN) / 45% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 µl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 60.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFOS; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 µl (500 ng/ml MPFOS)

MS Parameters

Collision Gas (mbar) = 3.70e-3
Collision Energy (eV) = 40

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H₂O
(both with 10 mM NH₄OAc buffer)

Flow: 300 µl/min

Reagent

LCMPFOS_00018

R: SBe 9/22/16



738686
ID: LCMPFOS_00018
Exp: 08/03/21 Ppd: SBC
13C4-Perfluorooctanesulfonate




WELLINGTON
LABORATORIES

CERTIFICATE OF ANALYSIS
DOCUMENTATION

PRODUCT CODE: MPFOS LOT NUMBER: MPFOS0816
COMPOUND: Sodium perfluoro-1-[1,2,3,4-¹³C₄]octanesulfonate

STRUCTURE: CAS #: Not available



MOLECULAR FORMULA: ¹³C₄¹²C₄F₁₇SO₃Na MOLECULAR WEIGHT: 526.08
CONCENTRATION: 50.0 ± 2.5 µg/ml (Na salt) SOLVENT(S): Methanol
47.8 ± 2.4 µg/ml (MPFOS anion)
CHEMICAL PURITY: >98% ISOTOPIC PURITY: ≥99% ¹³C
LAST TESTED: (mm/dd/yyyy) 08/03/2016 (1,2,3,4-¹³C₄)
EXPIRY DATE: (mm/dd/yyyy) 08/03/2021
RECOMMENDED STORAGE: Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains ~ 0.8% Sodium perfluoro-1-[1,2,3-¹³C₃]heptanesulfonate.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 08/05/2016

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

x_1, x_2, \dots, x_n on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:


Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.


QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFOS; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μm, 2.1 x 100 mm

Mobile phase: Gradient

Start: 45% (80:20 MeOH:ACN) / 55% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 1.5 min
before returning to initial conditions in 0.5 min.

Time: 10 min


Flow: 300 μl/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 3.00
Cone Voltage (V) = 60.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFOS; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
 10 μl (500 ng/ml MPFOS)

MS Parameters

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
 (both with 10 mM NH_4OAc buffer)

Collision Gas (mbar) = 3.46e-3
 Collision Energy (eV) = 40

Flow: 300 $\mu\text{l}/\text{min}$

Method 537 DOD

Perfluorinated Alkyl Acids (LC/MS)
by Method 537 DOD

FORM II
LCMS SURROGATE RECOVERY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Matrix: Water Level: Low
GC Column (1): Acquity ID: 2.1 (mm)

Client Sample ID	Lab Sample ID	PFHxA #	PFDA #
WI-AF-2RW07-1116	320-23971-1	103	118
WI-AF-2FB07-1116	320-23971-2	112	112
WI-AF-2RW08-1116	320-23971-3	96	108
WI-AF-2FB08-1116	320-23971-4	108	105
WI-AF-3RW10-1116	320-23971-5	110	111
WI-AF-3FB10-1116	320-23971-6	117	106
WI-AF-3RW11-1116	320-23971-7	110	109
WI-AF-3FB11-1116	320-23971-8	108	108
WI-AF-3RW12-1116	320-23971-9	119	130
WI-AF-3FB12-1116	320-23971-10	99	105
	MB 320-140632/1-A	106	104
	LCS 320-140632/2-A	113	110
	LCSD 320-140632/3-A	117	111

PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA

QC LIMITS
70-130
70-130

Column to be used to flag recovery values

FORM III
LCMS LAB CONTROL SAMPLE RECOVERY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Matrix: Water Level: Low Lab File ID: 11DEC2016A6A_007.d
Lab ID: LCS 320-140632/2-A Client ID: _____

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	LCS % REC	QC LIMITS REC	#
Perfluorooctanesulfonic acid (PFOS)	0.160	0.124	77	70-130	
Perfluorooctanoic acid (PFOA)	0.0811	0.0619	76	70-130	
Perfluorobutanesulfonic acid (PFBS)	0.359	0.275	77	70-130	

Column to be used to flag recovery and RPD values

FORM III 537

FORM III
LCMS LAB CONTROL SAMPLE DUPLICATE RECOVERY

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.: _____

Matrix: Water Level: Low Lab File ID: 11DEC2016A6A_008.d

Lab ID: LCSD 320-140632/3-A Client ID: _____

COMPOUND	SPIKE ADDED (ug/L)	LCSD CONCENTRATION (ug/L)	LCSD % REC	% RPD	QC LIMITS		#
					RPD	REC	
Perfluorooctanesulfonic acid (PFOS)	0.160	0.129	80	4	30	70-130	
Perfluorooctanoic acid (PFOA)	0.0811	0.0627	77	1	30	70-130	
Perfluorobutanesulfonic acid (PFBS)	0.359	0.294	82	7	30	70-130	

Column to be used to flag recovery and RPD values

FORM III 537

FORM IV
LCMS METHOD BLANK SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab File ID: 11DEC2016A6A_006.d Lab Sample ID: MB 320-140632/1-A
Matrix: Water Date Extracted: 12/05/2016 11:42
Instrument ID: A6 Date Analyzed: 12/11/2016 13:31
Level: (Low/Med) Low

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES:

CLIENT SAMPLE ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
	LCS 320-140632/2-A	11DEC2016A6 A 007.d	12/11/2016 14:01
	LCSD 320-140632/3-A	11DEC2016A6 A 008.d	12/11/2016 14:30
WI-AF-2RW07-1116	320-23971-1	11DEC2016A6 A 009.d	12/11/2016 15:00
WI-AF-2FB07-1116	320-23971-2	11DEC2016A6 A 010.d	12/11/2016 15:30
WI-AF-2RW08-1116	320-23971-3	11DEC2016A6 A 011.d	12/11/2016 15:59
WI-AF-2FB08-1116	320-23971-4	11DEC2016A6 A 012.d	12/11/2016 16:29
WI-AF-3RW10-1116	320-23971-5	11DEC2016A6 A 013.d	12/11/2016 16:58
WI-AF-3FB10-1116	320-23971-6	11DEC2016A6 A 014.d	12/11/2016 17:28
WI-AF-3RW11-1116	320-23971-7	11DEC2016A6 A 015.d	12/11/2016 17:58
WI-AF-3FB11-1116	320-23971-8	11DEC2016A6 A 019.d	12/11/2016 19:56
WI-AF-3RW12-1116	320-23971-9	11DEC2016A6 A 020.d	12/11/2016 20:26
WI-AF-3FB12-1116	320-23971-10	11DEC2016A6 A 021.d	12/11/2016 20:55

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Instrument ID: A6 Calibration Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.1 (mm) Calibration End Date: 12/05/2016 19:54
Calibration ID: 26888

		13PFOA		PFOS			
		AREA #	RT #	AREA #	RT #	AREA #	RT #
INITIAL CALIBRATION MEAN AREA AND MEAN RT		965911	20.05	2046916	20.67		
UPPER LIMIT		1448867	20.55	3070374	21.17		
LOWER LIMIT		482956	19.55	1023458	20.17		
LAB SAMPLE ID	CLIENT SAMPLE ID						
CCV 320-140688/9 CCVL		1025187	20.05	2358079	20.67		
ICV 320-140688/11		877210	20.05	2015178	20.67		
CCV 320-141573/3 CCVL		802153	20.06	1836390	20.68		
CCV 320-141573/4 CCVIS		900761	20.06	1778917	20.68		
MB 320-140632/1-A		764515	20.06	1993596	20.68		
LCS 320-140632/2-A		813368	20.06	2105344	20.68		
LCSD 320-140632/3-A		733684	20.06	1868957	20.68		
320-23971-1	WI-AF-2RW07-1116	670638	20.05	2176264	20.68		
320-23971-2	WI-AF-2FB07-1116	698931	20.05	2028913	20.68		
320-23971-3	WI-AF-2RW08-1116	779708	20.05	2254015	20.68		
320-23971-4	WI-AF-2FB08-1116	781277	20.05	2077587	20.68		
320-23971-5	WI-AF-3RW10-1116	675004	20.05	1937434	20.67		
320-23971-6	WI-AF-3FB10-1116	864715	20.06	2424671	20.68		
320-23971-7	WI-AF-3RW11-1116	731987	20.06	2119325	20.68		
CCV 320-141573/17 CCVIS		805687	20.05	1720352	20.68		
CCV 320-141574/17 CCVIS		805687	20.05	1720352	20.68		
320-23971-8	WI-AF-3FB11-1116	768857	20.06	2063347	20.68		
320-23971-9	WI-AF-3RW12-1116	597162	20.06	2146584	20.68		
320-23971-10	WI-AF-3FB12-1116	772993	20.06	2088535	20.68		
CCV 320-141574/29 CCVIS		886505	20.05	1829934	20.68		

13PFOA = 13C2-PFOA
PFOS = 13C4 PFOS

Area Limit = 50%-150% of internal standard area
RT Limit = ± 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Sample No.: CCV 320-141573/4 Date Analyzed: 12/11/2016 12:32
Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
Lab File ID (Standard): 11DEC2016A6A_004.d Heated Purge: (Y/N) N
Calibration ID: 26888

	13PFOA		PFOS		AREA #	RT #
	AREA #	RT #	AREA #	RT #		
12/24 HOUR STD	900761	20.06	1778917	20.68		
UPPER LIMIT	1261065	20.56	2490484	21.18		
LOWER LIMIT	630533	19.56	1245242	20.18		
LAB SAMPLE ID	CLIENT SAMPLE ID					
MB 320-140632/1-A		764515	20.06	1993596	20.68	
LCS 320-140632/2-A		813368	20.06	2105344	20.68	
LCSD 320-140632/3-A		733684	20.06	1868957	20.68	
320-23971-1	WI-AF-2RW07-1116	670638	20.05	2176264	20.68	
320-23971-2	WI-AF-2FB07-1116	698931	20.05	2028913	20.68	
320-23971-3	WI-AF-2RW08-1116	779708	20.05	2254015	20.68	
320-23971-4	WI-AF-2FB08-1116	781277	20.05	2077587	20.68	
320-23971-5	WI-AF-3RW10-1116	675004	20.05	1937434	20.67	
320-23971-6	WI-AF-3FB10-1116	864715	20.06	2424671	20.68	
320-23971-7	WI-AF-3RW11-1116	731987	20.06	2119325	20.68	

13PFOA = 13C2-PFOA
PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
RT Limit = ± 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Sample No.: CCV 320-141573/17 Date Analyzed: 12/11/2016 18:57
Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
Lab File ID (Standard): 11DEC2016A6A_017.d Heated Purge: (Y/N) N
Calibration ID: 26888

	13PFOA		PFOS		AREA #	RT #
	AREA #	RT #	AREA #	RT #		
12/24 HOUR STD	805687	20.05	1720352	20.68		
UPPER LIMIT	1127962	20.55	2408493	21.18		
LOWER LIMIT	563981	19.55	1204246	20.18		
LAB SAMPLE ID	CLIENT SAMPLE ID					
MB 320-140632/1-A		764515	20.06	1993596	20.68	
LCS 320-140632/2-A		813368	20.06	2105344	20.68	
LCSD 320-140632/3-A		733684	20.06	1868957	20.68	
320-23971-1	WI-AF-2RW07-1116	670638	20.05	2176264	20.68	
320-23971-2	WI-AF-2FB07-1116	698931	20.05	2028913	20.68	
320-23971-3	WI-AF-2RW08-1116	779708	20.05	2254015	20.68	
320-23971-4	WI-AF-2FB08-1116	781277	20.05	2077587	20.68	
320-23971-5	WI-AF-3RW10-1116	675004	20.05	1937434	20.67	
320-23971-6	WI-AF-3FB10-1116	864715	20.06	2424671	20.68	
320-23971-7	WI-AF-3RW11-1116	731987	20.06	2119325	20.68	

13PFOA = 13C2-PFOA
PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
RT Limit = ± 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Sample No.: CCV 320-141574/17 Date Analyzed: 12/11/2016 18:57
Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
Lab File ID (Standard): 11DEC2016A6A_017.d Heated Purge: (Y/N) N
Calibration ID: 26888

	13PFOA		PFOS		AREA #	RT #
	AREA #	RT #	AREA #	RT #		
12/24 HOUR STD	805687	20.05	1720352	20.68		
UPPER LIMIT	1127962	20.55	2408493	21.18		
LOWER LIMIT	563981	19.55	1204246	20.18		
LAB SAMPLE ID	CLIENT SAMPLE ID					
320-23971-8	WI-AF-3FB11-1116	768857	20.06	2063347	20.68	
320-23971-9	WI-AF-3RW12-1116	597162	20.06	2146584	20.68	
320-23971-10	WI-AF-3FB12-1116	772993	20.06	2088535	20.68	

13PFOA = 13C2-PFOA
PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
RT Limit = ± 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Sample No.: CCV 320-141574/29 Date Analyzed: 12/12/2016 00:52
Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
Lab File ID (Standard): 11DEC2016A6A_029.d Heated Purge: (Y/N) N
Calibration ID: 26888

	13PFOA		PFOS		AREA #	RT #
	AREA #	RT #	AREA #	RT #		
12/24 HOUR STD	886505	20.05	1829934	20.68		
UPPER LIMIT	1241107	20.55	2561908	21.18		
LOWER LIMIT	620554	19.55	1280954	20.18		
LAB SAMPLE ID	CLIENT SAMPLE ID					
320-23971-8	WI-AF-3FB11-1116	768857	20.06	2063347	20.68	
320-23971-9	WI-AF-3RW12-1116	597162	20.06	2146584	20.68	
320-23971-10	WI-AF-3FB12-1116	772993	20.06	2088535	20.68	

13PFOA = 13C2-PFOA
PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
RT Limit = ± 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2RW07-1116 Lab Sample ID: 320-23971-1
 Matrix: Water Lab File ID: 11DEC2016A6A_009.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:10
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 257.9 (mL) Date Analyzed: 12/11/2016 15:00
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U M	0.058	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0091
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	103		70-130
STL00996	13C2 PFDA	118		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_009.d
 Lims ID: 320-23971-A-1-A
 Client ID: WI-AF-2RW07-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:00:28 ALS Bottle#: 20 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-1-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:15:14

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.615 17.611 0.004 1.000 9267 0.1741 42.0
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 807916 10.3 26056
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.356 19.356 0.0 1.000 16813 0.2467 331
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.380 19.391 -0.011 1.000 2789 0.0342 1.1
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.058 -0.011 1.000 670638 10.0 17663
 6 Perfluoroctanoic acid
 413.0 > 369.0 20.059 20.058 0.001 1.000 8778 0.1258 1.7
 7 Perfluoroctane sulfonic acid
 499.0 > 80.0 20.679 20.619 0.060 1.000 12959 0.1636 244 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1.000 2176264 28.7 15062
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 1643 0.0216 6.6 M
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.480 -0.009 1.000 696041 11.8 22055

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_009.d
 Injection Date: 11-Dec-2016 15:00:28
 Lims ID: 320-23971-A-1-A
 Client ID: WI-AF-2RW07-1116
 Operator ID: CBW
 Injection Vol: 10.0 ul
 Method: 537_A6
 Instrument ID: A6
 Lab Sample ID: 320-23971-1
 ALS Bottle#: 20
 Dil. Factor: 1.0000
 Limit Group: LC 537 ICAL
 Worklist Smp#: 9



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_009.d
 Lims ID: 320-23971-A-1-A
 Client ID: WI-AF-2RW07-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:00:28 ALS Bottle#: 20 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-1-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:15:14

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.3	103.27
\$ 10 13C2 PFDA	10.0	11.8	118.44

TestAmerica Sacramento

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_009.d
 Injection Date: 11-Dec-2016 15:00:28 Instrument ID: A6
 Lims ID: 320-23971-A-1-A Lab Sample ID: 320-23971-1
 Client ID: WI-AF-2RW07-1116
 Operator ID: CBW ALS Bottle#: 20 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM


7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1

Not Detected


Expected RT: 20.62

Processing Integration Results



Manual Integration Results

RT: 20.68
 Area: 12959
 Amount: 0.163585
 Amount Units: ng/ml



Reviewer: barnettj, 12-Dec-2016 15:22:49

Audit Action: Manually Integrated

Audit Reason: Missed Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2FB07-1116 Lab Sample ID: 320-23971-2
 Matrix: Water Lab File ID: 11DEC2016A6A_010.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:12
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 266.2 (mL) Date Analyzed: 12/11/2016 15:30
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	112		70-130
STL00996	13C2 PFDA	112		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_010.d
 Lims ID: 320-23971-A-2-A
 Client ID: WI-AF-2FB07-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:30:05 ALS Bottle#: 21 Worklist Smp#: 10
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:23:24

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 916221 11.2 29466
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.058 -0.011 698931 10.0 18293
 7 Perfluorooctane sulfonic acid M
 499.0 > 80.0 20.702 20.619 0.083 1.000 663 0.008977 19.5 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2028913 28.7 52998
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 6160 0.0777 45.7
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.480 -0.009 1.000 687782 11.2 21530

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_010.d
 Injection Date: 11-Dec-2016 15:30:05 Instrument ID: A6
 Lims ID: 320-23971-A-2-A Lab Sample ID: 320-23971-2
 Client ID: WI-AF-2FB07-1116
 Operator ID: CBW ALS Bottle#: 21 Worklist Smp#: 10
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_010.d
 Lims ID: 320-23971-A-2-A
 Client ID: WI-AF-2FB07-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:30:05 ALS Bottle#: 21 Worklist Smp#: 10
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:23:24

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.2	112.38
\$ 10 13C2 PFDA	10.0	11.2	112.30

TestAmerica Sacramento

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_010.d
 Injection Date: 11-Dec-2016 15:30:05 Instrument ID: A6
 Lims ID: 320-23971-A-2-A Lab Sample ID: 320-23971-2
 Client ID: WI-AF-2FB07-1116
 Operator ID: CBW ALS Bottle#: 21 Worklist Smp#: 10
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM


7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1

Not Detected


Expected RT: 20.62

Processing Integration Results



RT: 20.70
 Area: 663
 Amount: 0.008977
 Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 12-Dec-2016 15:23:24

Audit Action: Manually Integrated

Audit Reason: Missed Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2RW08-1116 Lab Sample ID: 320-23971-3
 Matrix: Water Lab File ID: 11DEC2016A6A_011.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:30
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 269 (mL) Date Analyzed: 12/11/2016 15:59
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U M	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	96		70-130
STL00996	13C2 PFDA	108		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_011.d
 Lims ID: 320-23971-A-3-A
 Client ID: WI-AF-2RW08-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:59:42 ALS Bottle#: 22 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:24:23

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 869665 9.56 28087
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.356 19.356 0.0 1.000 22527 0.3192 106
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.320 19.391 -0.071 1.000 433 0.004570 0.2
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.058 -0.011 779708 10.0 11657
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.070 20.058 0.012 1.000 3138 0.0387 1.1 M
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.679 20.619 0.060 1.000 3225 0.0393 26.0 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2254015 28.7 21365
 9 Perfluorononanoic acid
 463.0 > 419.0 20.738 20.750 -0.012 1.000 8396 0.0949 23.4
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.480 -0.009 1.000 737074 10.8 23365

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_011.d
 Injection Date: 11-Dec-2016 15:59:42 Instrument ID: A6
 Lims ID: 320-23971-A-3-A Lab Sample ID: 320-23971-3
 Client ID: WI-AF-2RW08-1116
 Operator ID: CBW ALS Bottle#: 22 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_011.d
 Lims ID: 320-23971-A-3-A
 Client ID: WI-AF-2RW08-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 15:59:42 ALS Bottle#: 22 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:24:23

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	9.56	95.61
\$ 10 13C2 PFDA	10.0	10.8	107.88

TestAmerica Sacramento

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_011.d
 Injection Date: 11-Dec-2016 15:59:42 Instrument ID: A6
 Lims ID: 320-23971-A-3-A Lab Sample ID: 320-23971-3
 Client ID: WI-AF-2RW08-1116
 Operator ID: CBW ALS Bottle#: 22 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM


7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1

Not Detected


Expected RT: 20.62

Processing Integration Results



Manual Integration Results

RT: 20.68
 Area: 3225
 Amount: 0.039306
 Amount Units: ng/ml



Reviewer: barnettj, 12-Dec-2016 15:24:23

Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_011.d
 Injection Date: 11-Dec-2016 15:59:42 Instrument ID: A6
 Lims ID: 320-23971-A-3-A Lab Sample ID: 320-23971-3
 Client ID: WI-AF-2RW08-1116
 Operator ID: CBW ALS Bottle#: 22 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Processing Integration Results

RT: 20.30
 Area: 25696
 Amount: 0.316754
 Amount Units: ng/ml



Manual Integration Results

RT: 20.07
 Area: 3138
 Amount: 0.038682
 Amount Units: ng/ml



Reviewer: barnettj, 12-Dec-2016 15:24:23

Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2FB08-1116 Lab Sample ID: 320-23971-4
 Matrix: Water Lab File ID: 11DEC2016A6A_012.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:32
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 263.7 (mL) Date Analyzed: 12/11/2016 16:29
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	108		70-130
STL00996	13C2 PFDA	105		70-130

TestAmerica Sacramento
Target Compound Quantitation Report


Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_012.d
 Lims ID: 320-23971-A-4-A
 Client ID: WI-AF-2FB08-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 16:29:17 ALS Bottle#: 23 Worklist Smp#: 12
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-4-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:13

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 986599 10.8 31626
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.058 -0.011 781277 10.0 20511
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2077587 28.7 54101
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 5565 0.0628 164
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.480 -0.009 1.000 721905 10.5 22949

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_012.d
 Injection Date: 11-Dec-2016 16:29:17 Instrument ID: A6
 Lims ID: 320-23971-A-4-A Lab Sample ID: 320-23971-4
 Client ID: WI-AF-2FB08-1116
 Operator ID: CBW ALS Bottle#: 23 Worklist Smp#: 12
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_012.d
 Lims ID: 320-23971-A-4-A
 Client ID: WI-AF-2FB08-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 16:29:17 ALS Bottle#: 23 Worklist Smp#: 12
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-4-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:13

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.8	108.25
\$ 10 13C2 PFDA	10.0	10.5	105.45

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3RW10-1116 Lab Sample ID: 320-23971-5
 Matrix: Water Lab File ID: 11DEC2016A6A_013.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:00
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 264 (mL) Date Analyzed: 12/11/2016 16:58
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	111		70-130

TestAmerica Sacramento
Target Compound Quantitation Report


Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_013.d
 Lims ID: 320-23971-A-5-A
 Client ID: WI-AF-3RW10-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 16:58:53 ALS Bottle#: 24 Worklist Smp#: 13
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-5-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:33

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

\$ 2 13C2 PFHxA
 315.0 > 270.0 18.576 18.585 -0.009 1.000 867480 11.0 8055
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.225 19.391 -0.166 1.000 483 0.005889 0.2
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.058 -0.011 675004 10.0 17697
 * 8 13C4 PFOS
 503.0 > 80.0 20.667 20.679 -0.012 1937434 28.7 50819
 9 Perfluorononanoic acid
 463.0 > 419.0 20.738 20.750 -0.012 1.000 11292 0.1475 132
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.462 21.480 -0.018 1.000 654536 11.1 20476

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_013.d
 Injection Date: 11-Dec-2016 16:58:53 Instrument ID: A6
 Lims ID: 320-23971-A-5-A Lab Sample ID: 320-23971-5
 Client ID: WI-AF-3RW10-1116
 Operator ID: CBW ALS Bottle#: 24 Worklist Smp#: 13
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_013.d
 Lims ID: 320-23971-A-5-A
 Client ID: WI-AF-3RW10-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 16:58:53 ALS Bottle#: 24 Worklist Smp#: 13
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-5-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:33

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.0	110.17
\$ 10 13C2 PFDA	10.0	11.1	110.66

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3FB10-1116 Lab Sample ID: 320-23971-6
 Matrix: Water Lab File ID: 11DEC2016A6A_014.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:00
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 270.1 (mL) Date Analyzed: 12/11/2016 17:28
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.044	U M	0.056	0.044	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0087
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	117		70-130
STL00996	13C2 PFDA	106		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_014.d
 Lims ID: 320-23971-B-6-A
 Client ID: WI-AF-3FB10-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 17:28:30 ALS Bottle#: 25 Worklist Smp#: 14
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-b-6-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:58

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA								
315.0 > 270.0	18.585	18.585	0.0	1.000	1181965	11.7	37646	
* 5 13C2-PFOA								
415.0 > 370.0	20.058	20.058	0.0		864715	10.0	22754	
7 Perfluorooctane sulfonic acid								M
499.0 > 80.0	20.667	20.619	0.048	1.000	721	0.008169	15.8	M
* 8 13C4 PFOS								
503.0 > 80.0	20.679	20.679	0.0		2424671	28.7	63593	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.750	0.0	1.000	6326	0.0645	229	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.480	-0.009	1.000	806342	10.6	25182	

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_014.d
 Injection Date: 11-Dec-2016 17:28:30 Instrument ID: A6
 Lims ID: 320-23971-B-6-A Lab Sample ID: 320-23971-6
 Client ID: WI-AF-3FB10-1116
 Operator ID: CBW ALS Bottle#: 25 Worklist Smp#: 14
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_014.d
 Lims ID: 320-23971-B-6-A
 Client ID: WI-AF-3FB10-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 17:28:30 ALS Bottle#: 25 Worklist Smp#: 14
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-b-6-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:38:58

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.7	117.18
\$ 10 13C2 PFDA	10.0	10.6	106.42

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_014.d
 Injection Date: 11-Dec-2016 17:28:30 Instrument ID: A6
 Lims ID: 320-23971-B-6-A Lab Sample ID: 320-23971-6
 Client ID: WI-AF-3FB10-1116
 Operator ID: CBW ALS Bottle#: 25 Worklist Smp#: 14
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1


Not Detected
 Expected RT: 20.62

Processing Integration Results



RT: 20.67
 Area: 721
 Amount: 0.008169
 Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 12-Dec-2016 15:38:58

Audit Action: Manually Integrated

Audit Reason: Missed Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3RW11-1116 Lab Sample ID: 320-23971-7
 Matrix: Water Lab File ID: 11DEC2016A6A_015.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:30
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 268.9 (mL) Date Analyzed: 12/11/2016 17:58
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141573 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	109		70-130

TestAmerica Sacramento
Target Compound Quantitation Report


Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_015.d
 Lims ID: 320-23971-A-7-A
 Client ID: WI-AF-3RW11-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 17:58:05 ALS Bottle#: 26 Worklist Smp#: 15
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-7-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 13:51:39

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 940504 11.0 30201
 * 5 13C2-PFOA
 415.0 > 370.0 20.058 20.058 0.0 731987 10.0 19157
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2119325 28.7 55558
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 13436 0.1618 210
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 696464 10.9 21994

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_015.d
 Injection Date: 11-Dec-2016 17:58:05 Instrument ID: A6
 Lims ID: 320-23971-A-7-A Lab Sample ID: 320-23971-7
 Client ID: WI-AF-3RW11-1116
 Operator ID: CBW ALS Bottle#: 26 Worklist Smp#: 15
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_015.d
 Lims ID: 320-23971-A-7-A
 Client ID: WI-AF-3RW11-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 17:58:05 ALS Bottle#: 26 Worklist Smp#: 15
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-7-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 13:51:39

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.0	110.14
\$ 10 13C2 PFDA	10.0	10.9	108.58

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB11-1116 Lab Sample ID: 320-23971-8
 Matrix: Water Lab File ID: 11DEC2016A6A_019.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:31
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 265.1 (mL) Date Analyzed: 12/11/2016 19:56
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141574 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	108		70-130
STL00996	13C2 PFDA	108		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_019.d
 Lims ID: 320-23971-A-8-A
 Client ID: WI-AF-3FB11-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 19:56:26 ALS Bottle#: 27 Worklist Smp#: 19
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-8-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:39:27

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA								
315.0 > 270.0	18.585	18.585	0.0	1.000	964741	10.8	30960	
* 5 13C2-PFOA								
415.0 > 370.0	20.059	20.047	0.012		768857	10.0	20020	
7 Perfluorooctane sulfonic acid							M	
499.0 > 80.0	20.679	20.619	0.060	1.000	651	0.008667	20.5	M
* 8 13C4 PFOS								
503.0 > 80.0	20.679	20.679	0.0		2063347	28.7	27132	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.750	0.0	1.000	8357	0.0958	235	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.480	21.480	0.0	1.000	726581	10.8	22720	

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_019.d
 Injection Date: 11-Dec-2016 19:56:26 Instrument ID: A6
 Lims ID: 320-23971-A-8-A Lab Sample ID: 320-23971-8
 Client ID: WI-AF-3FB11-1116
 Operator ID: CBW ALS Bottle#: 27 Worklist Smp#: 19
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_019.d
 Lims ID: 320-23971-A-8-A
 Client ID: WI-AF-3FB11-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 19:56:26 ALS Bottle#: 27 Worklist Smp#: 19
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-8-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:39:27

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.8	107.56
\$ 10 13C2 PFDA	10.0	10.8	107.84

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_019.d
 Injection Date: 11-Dec-2016 19:56:26 Instrument ID: A6
 Lims ID: 320-23971-A-8-A Lab Sample ID: 320-23971-8
 Client ID: WI-AF-3FB11-1116
 Operator ID: CBW ALS Bottle#: 27 Worklist Smp#: 19
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1


Processing Integration Results

Not Detected
 Expected RT: 20.62



Manual Integration Results

RT: 20.68
 Area: 651
 Amount: 0.008667
 Amount Units: ng/ml



Reviewer: barnettj, 12-Dec-2016 15:41:05

Audit Action: Manually Integrated

Audit Reason: Missed Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3RW12-1116 Lab Sample ID: 320-23971-9
 Matrix: Water Lab File ID: 11DEC2016A6A_020.d
 Analysis Method: 537 Date Collected: 11/30/2016 17:00
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 256.4 (mL) Date Analyzed: 12/11/2016 20:26
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141574 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.059	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0092
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	119		70-130
STL00996	13C2 PFDA	130		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_020.d
 Lims ID: 320-23971-A-9-A
 Client ID: WI-AF-3RW12-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 20:26:02 ALS Bottle#: 28 Worklist Smp#: 20
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-9-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:41:52

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 828115 11.9 26495
 * 5 13C2-PFOA
 415.0 > 370.0 20.059 20.047 0.012 597162 10.0 15430
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2146584 28.7 37593
 9 Perfluorononanoic acid M
 463.0 > 419.0 20.762 20.750 0.012 1.000 1499 0.0221 14.9 M
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 680430 13.0 21575

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_020.d
 Injection Date: 11-Dec-2016 20:26:02 Instrument ID: A6
 Lims ID: 320-23971-A-9-A Lab Sample ID: 320-23971-9
 Client ID: WI-AF-3RW12-1116
 Operator ID: CBW ALS Bottle#: 28 Worklist Smp#: 20
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_020.d
 Lims ID: 320-23971-A-9-A
 Client ID: WI-AF-3RW12-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 20:26:02 ALS Bottle#: 28 Worklist Smp#: 20
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-9-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:41:52

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.9	118.88
\$ 10 13C2 PFDA	10.0	13.0	130.03

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3FB12-1116 Lab Sample ID: 320-23971-10
 Matrix: Water Lab File ID: 11DEC2016A6A_021.d
 Analysis Method: 537 Date Collected: 11/30/2016 17:01
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 265.2 (mL) Date Analyzed: 12/11/2016 20:55
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture:
 Analysis Batch No.: 141574 GPC Cleanup: (Y/N) N
 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	99		70-130
STL00996	13C2 PFDA	105		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_021.d
 Lims ID: 320-23971-A-10-A
 Client ID: WI-AF-3FB12-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 20:55:38 ALS Bottle#: 29 Worklist Smp#: 21
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-10-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:42:27

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 894651 9.92 28817
 * 5 13C2-PFOA
 415.0 > 370.0 20.058 20.047 0.011 772993 10.0 20214
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 2088535 28.7 27254
 9 Perfluorononanoic acid M
 463.0 > 419.0 20.750 20.750 0.0 1.000 4266 0.0487 119 M
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 713508 10.5 22429


QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_021.d
 Injection Date: 11-Dec-2016 20:55:38 Instrument ID: A6
 Lims ID: 320-23971-A-10-A Lab Sample ID: 320-23971-10
 Client ID: WI-AF-3FB12-1116
 Operator ID: CBW ALS Bottle#: 29 Worklist Smp#: 21
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_021.d
 Lims ID: 320-23971-A-10-A
 Client ID: WI-AF-3FB12-1116
 Sample Type: Client
 Inject. Date: 11-Dec-2016 20:55:38 ALS Bottle#: 29 Worklist Smp#: 21
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-23971-a-10-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:02:26 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:42:27

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	9.92	99.22
\$ 10 13C2 PFDA	10.0	10.5	105.34

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
CURVE EVALUATION

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

Analy Batch No.: 140688

SDG No.: _____

Instrument ID: A6 GC Column: Acquity ID: 2.1(mm) Heated Purge: (Y/N) N

Calibration Start Date: 12/05/2016 17:26 Calibration End Date: 12/05/2016 19:54 Calibration ID: 26888

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-140688/2	05DEC2016A6A_004.d
Level 2	STD 320-140688/3	05DEC2016A6A_005.d
Level 3	STD 320-140688/4	05DEC2016A6A_006.d
Level 4	STD 320-140688/5	05DEC2016A6A_007.d
Level 5	STD 320-140688/6	05DEC2016A6A_008.d
Level 6	STD 320-140688/7	05DEC2016A6A_009.d

ANALYTE	RRF					CURVE TYPE	COEFFICIENT			#	MIN RRF	%RSD	#	MAX %RSD	R^2 OR COD	#	MIN R^2 OR COD
	LVL 1	LVL 2	LVL 3	LVL 4	LVL 5		B	M1	M2								
Perfluorobutanesulfonic acid (PFBS)	0.7247 0.6563	0.6525	0.7178	0.7256	0.7321	Ave		0.7015				5.2	30.0				
Perfluorohexanesulfonic acid	0.8344 0.8930	0.7757	0.9290	0.9478	1.0082	Ave		0.8980				9.3	30.0				
Perfluoroheptanoic acid	1.4137 1.1078	1.1891	1.2161	1.1975	1.1665	Ave		1.2151				8.6	30.0				
Perfluorooctanoic acid (PFOA)	0.9720 1.0610	0.9049	1.0674	1.1235	1.1136	Ave		1.0404				8.2	30.0				
Perfluorooctanesulfonic acid (PFOS)	0.8855 1.0951	0.9020	1.0711	1.0966	1.2136	Ave		1.0440				12.1	30.0				
Perfluorononanoic acid	0.9735 1.1655	0.9961	1.1929	1.2321	1.2453	Ave		1.1342				10.5	30.0				
13C2 PFHxA	1.0366 1.2091	1.0515	1.1929	1.2298	1.2791	Ave		1.1665				8.5	30.0				
13C2 PFDA	0.8084 0.9456	0.7439	0.8674	0.9054	0.9868	Ave		0.8763				10.2	30.0				

Note: The m1 coefficient is the same as Ave RRF for an Ave curve type.

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
RESPONSE AND CONCENTRATION

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

Analy Batch No.: 140688

SDG No.: _____

Instrument ID: A6 GC Column: Acquity ID: 2.1(mm) Heated Purge: (Y/N) N

Calibration Start Date: 12/05/2016 17:26 Calibration End Date: 12/05/2016 19:54 Calibration ID: 26888

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-140688/2	05DEC2016A6A_004.d
Level 2	STD 320-140688/3	05DEC2016A6A_005.d
Level 3	STD 320-140688/4	05DEC2016A6A_006.d
Level 4	STD 320-140688/5	05DEC2016A6A_007.d
Level 5	STD 320-140688/6	05DEC2016A6A_008.d
Level 6	STD 320-140688/7	05DEC2016A6A_009.d

ANALYTE	IS REF	CURVE TYPE	RESPONSE					CONCENTRATION (NG/ML)				
			LVL 1 LVL 6	LVL 2	LVL 3	LVL 4	LVL 5	LVL 1 LVL 6	LVL 2	LVL 3	LVL 4	LVL 5
Perfluorobutanesulfonic acid (PFBS)	PFOS	Ave	437563 7753569	1227165	2489398	4401661	6630132	8.76 178	22.9	45.1	90.9	135
Perfluorohexanesulfonic acid	PFOS	Ave	169827 3556638	491809	1086082	1938237	3077974	2.95 60.1	7.72	15.2	30.6	45.4
Perfluoroheptanoic acid	13PF OA	Ave	126557 2032288	324913	658044	1121930	1727957	0.994 20.2	2.60	5.12	10.3	15.3
Perfluorooctanoic acid (PFOA)	13PF OA	Ave	173304 3876381	492431	1150281	2096404	3285195	1.98 40.3	5.17	10.2	20.5	30.4
Perfluorooctanesulfonic acid (PFOS)	PFOS	Ave	238662 5775285	757269	1658139	2969550	4906017	3.91 79.6	10.2	20.1	40.6	60.1
Perfluorononanoic acid	13PF OA	Ave	168128 4124664	525061	1245341	2227031	3558831	1.92 39.0	5.01	9.87	19.9	29.5
13C2 PFHxA	13PF OA	Ave	933751 1095977	1106485	1261522	1117585	1240474	10.0 10.0	10.0	10.0	10.0	10.0
13C2 PFDA	13PF OA	Ave	728204 857144	782778	917302	822787	957025	10.0 10.0	10.0	10.0	10.0	10.0

Curve Type Legend:

Ave = Average ISTD

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
READBACK PERCENT ERROR

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1 Analy Batch No.: 140688

SDG No.: _____


Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm) Heated Purge: (Y/N) N


Calibration Start Date: 12/05/2016 17:26 Calibration End Date: 12/05/2016 19:54 Calibration ID: 26888

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-140688/2	05DEC2016A6A_004.d
Level 2	STD 320-140688/3	05DEC2016A6A_005.d
Level 3	STD 320-140688/4	05DEC2016A6A_006.d
Level 4	STD 320-140688/5	05DEC2016A6A_007.d
Level 5	STD 320-140688/6	05DEC2016A6A_008.d
Level 6	STD 320-140688/7	05DEC2016A6A_009.d

ANALYTE	PERCENT ERROR						PERCENT ERROR LIMIT					
	LVL 1 #	LVL 2 #	LVL 3 #	LVL 4 #	LVL 5 #	LVL 6 #	LVL 1	LVL 2	LVL 3	LVL 4	LVL 5	LVL 6
Perfluorobutanesulfonic acid (PFBS)	3.3	-7.0	2.3	3.4	4.4	-6.4	50	50	50	50	50	50
Perfluorohexanesulfonic acid	-7.1	-13.6	3.4	5.5	12.3	-0.6	50	50	50	50	50	50
Perfluoroheptanoic acid	16.3	-2.1	0.1	-1.5	-4.0	-8.8	50	50	50	50	50	50
Perfluorooctanoic acid (PFOA)	-6.6	-13.0	2.6	8.0	7.0	2.0	50	50	50	50	50	50
Perfluorooctanesulfonic acid (PFOS)	-15.2	-13.6	2.6	5.0	16.2	4.9	50	50	50	50	50	50
Perfluorononanoic acid	-14.2	-12.2	5.2	8.6	9.8	2.8	50	50	50	50	50	50
13C2 PFHxA	-11.1	-9.9	2.3	5.4	9.7	3.7	30	30	30	30	30	30
13C2 PFDA	-7.7	-15.1	-1.0	3.3	12.6	7.9	30	30	30	30	30	30





TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_004.d
 Lims ID: STD L1
 Client ID:
 Sample Type: IC Calib Level: 1
 Inject. Date: 05-Dec-2016 17:26:03 ALS Bottle#: 1 Worklist Smp#: 2
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L1 L1
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:34 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024

First Level Reviewer: barnettj Date: 06-Dec-2016 10:00:02

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.576	17.581	-0.005	1.000	437563	9.05	466	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.567	0.0	1.000	933751	8.89	30467	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.332	19.342	-0.010	1.000	169827	2.74	4140	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.368	19.378	-0.010	1.000	126557	1.16	45.1	M
* 5 13C2-PFOA								
415.0 > 370.0	20.047	20.047	0.0		900764	10.0	23392	
6 Perfluorooctanoic acid								M
413.0 > 369.0	20.047	20.047	0.0	1.000	173304	1.85	35.0	M
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	238662	3.32	2941	
* 8 13C4 PFOS								
503.0 > 80.0	20.667	20.669	-0.002		1976615	28.7	40886	
9 Perfluorononanoic acid								
463.0 > 419.0	20.738	20.748	-0.010	1.000	168128	1.65	6043	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.474	-0.003	1.000	728204	9.23	22953	

QC Flag Legend

Review Flags

M - Manually Integrated


Reagents:

LC537-L1_00015

Amount Added: 1.00

Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_004.d
 Injection Date: 05-Dec-2016 17:26:03 Instrument ID: A6
 Lims ID: STD L1
 Client ID:
 Operator ID: CBW ALS Bottle#: 1 Worklist Smp#: 2
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_004.d
 Injection Date: 05-Dec-2016 17:26:03 Instrument ID: A6
 Lims ID: STD L1
 Client ID:
 Operator ID: CBW ALS Bottle#: 1 Worklist Smp#: 2
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector: F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


Processing Integration Results

RT: 19.37
 Area: 155591
 Amount: 1.476072
 Amount Units: ng/ml



Manual Integration Results

RT: 19.37
 Area: 126557
 Amount: 1.156251
 Amount Units: ng/ml



Reviewer: barnettj, 06-Dec-2016 10:00:02

Audit Action: Manually Integrated

Audit Reason: Split Peak

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_004.d
 Injection Date: 05-Dec-2016 17:26:03 Instrument ID: A6
 Lims ID: STD L1
 Client ID:
 Operator ID: CBW ALS Bottle#: 1 Worklist Smp#: 2
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector: F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Processing Integration Results

RT: 20.05
 Area: 186490
 Amount: 1.959453
 Amount Units: ng/ml



Manual Integration Results

RT: 20.05
 Area: 173304
 Amount: 1.849212
 Amount Units: ng/ml



Reviewer: barnettj, 06-Dec-2016 10:00:02

Audit Action: Manually Integrated

Audit Reason: Split Peak

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_005.d
 Lims ID: STD L2
 Client ID:
 Sample Type: IC Calib Level: 2
 Inject. Date: 05-Dec-2016 17:55:38 ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L2 L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:35 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024

First Level Reviewer: barnettj Date: 06-Dec-2016 09:58:24

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.582	17.581	0.001	1.000	1227165	21.3	5055	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.567	0.0	1.000	1106485	9.01	35678	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.344	19.342	0.002	1.000	491809	6.67	11495	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.380	19.378	0.002	1.000	324913	2.54	155	M
* 5 13C2-PFOA								
415.0 > 370.0	20.047	20.047	0.0		1052273	10.0	27645	
6 Perfluorooctanoic acid								M
413.0 > 369.0	20.047	20.047	0.0	1.000	492431	4.50	100	M
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	757269	8.83	8449	
* 8 13C4 PFOS								
503.0 > 80.0	20.667	20.669	-0.002		2356620	28.7	30757	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.748	0.002	1.000	525061	4.40	13911	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.480	21.474	0.006	1.000	782778	8.49	24678	

QC Flag Legend

Review Flags

M - Manually Integrated


Reagents:

LC537-L2_00014

Amount Added: 1.00

Units: mL

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_005.d
 Injection Date: 05-Dec-2016 17:55:38 Instrument ID: A6
 Lims ID: STD L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_005.d
 Injection Date: 05-Dec-2016 17:55:38 Instrument ID: A6
 Lims ID: STD L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector: F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


Processing Integration Results

RT: 19.38
 Area: 344811
 Amount: 2.670013
 Amount Units: ng/ml



Manual Integration Results

RT: 19.38
 Area: 324913
 Amount: 2.541065
 Amount Units: ng/ml



Reviewer: barnettj, 06-Dec-2016 10:03:30

Audit Action: Manually Integrated

Audit Reason: Split Peak

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_005.d
 Injection Date: 05-Dec-2016 17:55:38 Instrument ID: A6
 Lims ID: STD L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector: F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Processing Integration Results

RT: 20.05
 Area: 504990
 Amount: 4.595586
 Amount Units: ng/ml



Manual Integration Results

RT: 20.05
 Area: 492431
 Amount: 4.497863
 Amount Units: ng/ml



Reviewer: barnettj, 06-Dec-2016 10:03:30

Audit Action: Manually Integrated

Audit Reason: Split Peak

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_006.d
 Lims ID: STD L3
 Client ID:
 Sample Type: IC Calib Level: 3
 Inject. Date: 05-Dec-2016 18:25:13 ALS Bottle#: 3 Worklist Smp#: 4
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L3 L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:36 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024


First Level Reviewer: barnettj Date: 06-Dec-2016 09:58:05

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.582	17.581	0.001	1.000	2489398	46.2	1804	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.567	0.0	1.000	1261522	10.2	40506	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.344	19.342	0.002	1.000	1086082	15.7	25400	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.380	19.378	0.002	1.000	658044	5.12	4774	
* 5 13C2-PFOA								
415.0 > 370.0	20.047	20.047	0.0		1057506	10.0	27287	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.047	20.047	0.0	1.000	1150281	10.5	429	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	1658139	20.7	19019	
* 8 13C4 PFOS								
503.0 > 80.0	20.667	20.669	-0.002		2205243	28.7	57142	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.748	0.002	1.000	1245341	10.4	13210	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.474	-0.003	1.000	917302	9.90	28753	

Reagents:

LC537-L3_00016 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_006.d
 Injection Date: 05-Dec-2016 18:25:13 Instrument ID: A6
 Lims ID: STD L3
 Client ID:
 Operator ID: CBW ALS Bottle#: 3 Worklist Smp#: 4
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_007.d
 Lims ID: STD L4
 Client ID:
 Sample Type: ICISAV Calib Level: 4
 Inject. Date: 05-Dec-2016 18:54:48 ALS Bottle#: 4 Worklist Smp#: 5
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L4 L4
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:37 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024


First Level Reviewer: barnettj Date: 06-Dec-2016 13:43:03

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.579	17.581	-0.002	1.000	4401661	94.0	2768	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.567	0.0	1.000	1117585	10.5	28676	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.344	19.342	0.002	1.000	1938237	32.3	25196	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.380	19.378	0.002	1.000	1121930	10.2	12796	
* 5 13C2-PFOA								
415.0 > 370.0	20.047	20.047	0.0		908727	10.0	23744	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.047	20.047	0.0	1.000	2096404	22.2	516	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	2969550	42.6	9704	
* 8 13C4 PFOS								
503.0 > 80.0	20.667	20.669	-0.002		1914415	28.7	28032	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.748	0.002	1.000	2227031	21.6	23494	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.474	-0.003	1.000	822787	10.3	25796	

Reagents:

LC537-L4_00015 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_007.d
 Injection Date: 05-Dec-2016 18:54:48 Instrument ID: A6
 Lims ID: STD L4
 Client ID:
 Operator ID: CBW ALS Bottle#: 4 Worklist Smp#: 5
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_008.d
 Lims ID: STD L5
 Client ID:
 Sample Type: IC Calib Level: 5
 Inject. Date: 05-Dec-2016 19:24:23 ALS Bottle#: 5 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L5 L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:38 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024


Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.581 0.001 1.000 6630132 140.5 3208
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.567 18.567 0.0 1.000 1240474 11.0 39454
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.344 19.342 0.002 1.000 3077974 51.0 14553
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.380 19.378 0.002 1.000 1727957 14.7 6886
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.047 0.0 969779 10.0 24964
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.047 20.047 0.0 1.000 3285195 32.6 1114
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.620 20.619 0.001 1.000 4906017 69.9 10146
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.669 0.010 1929192 28.7 32805
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.748 0.002 1.000 3558831 32.4 16307
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.474 0.006 1.000 957025 11.3 30231

Reagents:

LC537-L5_00017 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_008.d
 Injection Date: 05-Dec-2016 19:24:23 Instrument ID: A6
 Lims ID: STD L5
 Client ID:
 Operator ID: CBW ALS Bottle#: 5 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Lims ID: STD L6
 Client ID:
 Sample Type: IC Calib Level: 6
 Inject. Date: 05-Dec-2016 19:54:00 ALS Bottle#: 6 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L6 L6
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:39 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.581 0.001 1.000 7753569 166.9 8570
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.567 18.567 0.0 1.000 1095977 10.4 34796
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.344 19.342 0.002 1.000 3556638 59.8 31299
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.380 19.378 0.002 1.000 2032288 18.5 6367
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.047 0.0 1.000 906416 10.0 23083
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.047 20.047 0.0 1.000 3876381 41.1 917
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 5775285 83.5 12991
 * 8 13C4 PFOS
 503.0 > 80.0 20.667 20.669 -0.002 1.000 1899408 28.7 17628
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.748 0.002 1.000 4124664 40.1 17939
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.474 -0.003 1.000 857144 10.8 26862

Reagents:


LC537-L6_00014 Amount Added: 1.00 Units: mL

Report Date: 06-Dec-2016 16:35:40

Chrom Revision: 2.2 05-Dec-2016 12:37:22

TestAmerica Sacramento

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_009.d
 Injection Date: 05-Dec-2016 19:54:00 Instrument ID: A6
 Lims ID: STD L6
 Client ID:
 Operator ID: CBW ALS Bottle#: 6 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-140688/9 Calibration Date: 12/05/2016 20:53
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 05DEC2016A6A_011.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.6306		20.6	22.9	-10.1	50.0
Perfluorohexanesulfonic acid	Ave	0.8980	0.7822		6.72	7.72	-12.9	50.0
Perfluoroheptanoic acid	Ave	1.215	1.239		2.65	2.60	1.9	50.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	0.9133		4.54	5.17	-12.2	50.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	0.8902		8.71	10.2	-14.7	50.0
Perfluorononanoic acid	Ave	1.134	1.093		4.83	5.01	-3.6	50.0
13C2 PFHxA	Ave	1.167	1.081		9.27	10.0	-7.3	30.0
13C2 PFDA	Ave	0.8763	0.8211		9.37	10.0	-6.3	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_011.d
 Lims ID: CCV L2
 Client ID:
 Sample Type: CCVL
 Inject. Date: 05-Dec-2016 20:53:12 ALS Bottle#: 2 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L2 CCV L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:35:40 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024

First Level Reviewer: barnettj Date: 06-Dec-2016 10:08:33

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.586	17.581	0.005	1.000	1186753	20.6	693	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.567	0.0	1.000	1108698	9.27	35970	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.344	19.342	0.002	1.000	496197	6.72	11535	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.380	19.378	0.002	1.000	329772	2.65	166	M
* 5 13C2-PFOA								
415.0 > 370.0	20.047	20.047	0.0		1025187	10.0	21492	
6 Perfluorooctanoic acid								M
413.0 > 369.0	20.047	20.047	0.0	1.000	484196	4.54	93.2	M
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	747766	8.71	8549	
* 8 13C4 PFOS								
503.0 > 80.0	20.667	20.669	-0.002		2358079	28.7	20478	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.748	0.002	1.000	561371	4.83	15032	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.474	-0.003	1.000	841818	9.37	26813	

QC Flag Legend

Review Flags

M - Manually Integrated


Reagents:

LC537-L2_00014

Amount Added: 1.00

Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_011.d
 Injection Date: 05-Dec-2016 20:53:12 Instrument ID: A6
 Lims ID: CCV L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_011.d
 Injection Date: 05-Dec-2016 20:53:12 Instrument ID: A6
 Lims ID: CCV L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


Processing Integration Results

RT: 19.38
 Area: 349162
 Amount: 2.802857
 Amount Units: ng/ml



Manual Integration Results

RT: 19.38
 Area: 329772
 Amount: 2.647206
 Amount Units: ng/ml



Reviewer: barnettj, 06-Dec-2016 10:08:33

Audit Action: Manually Integrated

Audit Reason: Split Peak

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_011.d
 Injection Date: 05-Dec-2016 20:53:12 Instrument ID: A6
 Lims ID: CCV L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector: F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.05
 Area: 520603
 Amount: 4.880820
 Amount Units: ng/ml

Processing Integration Results



RT: 20.05
 Area: 484196
 Amount: 4.539493
 Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 06-Dec-2016 10:08:33

Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: ICV 320-140688/11 Calibration Date: 12/05/2016 21:52
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 05DEC2016A6A_013.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.5756		94.2	115	-18.0	30.0
Perfluorohexanesulfonic acid	Ave	0.8980	0.6976		20.6	26.5	-22.3	30.0
Perfluoroheptanoic acid	Ave	1.215	1.155		11.9	12.5	-4.9	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	0.9604		23.2	25.1	-7.7	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	0.8424		22.0	27.2	-19.3	30.0
Perfluorononanoic acid	Ave	1.134	0.9316		20.6	25.1	-17.9	30.0
13C2 PFHxA	Ave	1.167	1.079		9.25	10.0	-7.5	30.0
13C2 PFDA	Ave	0.8763	0.8628		9.85	10.0	-1.5	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_013.d
 Lims ID: ICV
 Client ID:
 Sample Type: ICV
 Inject. Date: 05-Dec-2016 21:52:24 ALS Bottle#: 7 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: ICV ICV
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist:
 Method: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 06-Dec-2016 16:53:23 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK024


First Level Reviewer: barnettj Date: 06-Dec-2016 16:34:53

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0 17.582 17.581 0.001 1.000 4641388 94.2 8629								
\$ 2 13C2 PFHxA								
315.0 > 270.0 18.567 18.567 0.0 1.000 946677 9.25 29673								
3 Perfluorohexanesulfonic acid								
399.0 > 80.0 19.344 19.342 0.002 1.000 1298107 20.6 29738								
4 Perfluoroheptanoic acid								
363.0 > 319.0 19.380 19.378 0.002 1.000 1267011 11.9 9991								
* 5 13C2-PFOA								
415.0 > 370.0 20.047 20.047 0.0 1.000 877210 10.0 22431								
6 Perfluorooctanoic acid								
413.0 > 369.0 20.047 20.047 0.0 1.000 2114272 23.2 647								
7 Perfluorooctane sulfonic acid								
499.0 > 80.0 20.619 20.619 0.0 1.000 1612191 22.0 13496								
* 8 13C4 PFOS								
503.0 > 80.0 20.667 20.669 -0.002 1.000 2015178 28.7 51574								
9 Perfluorononanoic acid								
463.0 > 419.0 20.750 20.748 0.002 1.000 2051048 20.6 7161								
\$ 10 13C2 PFDA								
515.0 > 470.0 21.480 21.474 0.006 1.000 756809 9.85 23714								

Reagents:

LC537-ICV_00017 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_013.d
 Injection Date: 05-Dec-2016 21:52:24 Instrument ID: A6
 Lims ID: ICV
 Client ID:
 Operator ID: CBW ALS Bottle#: 7 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-141573/3 Calibration Date: 12/11/2016 12:02
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 11DEC2016A6A_003.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.6477		21.1	22.9	-7.7	50.0
Perfluorohexanesulfonic acid	Ave	0.8980	0.7642		6.57	7.72	-14.9	50.0
Perfluoroheptanoic acid	Ave	1.215	1.413		3.02	2.60	16.3	50.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	1.044		5.19	5.17	0.3	50.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	0.8884		8.69	10.2	-14.9	50.0
Perfluorononanoic acid	Ave	1.134	1.098		4.85	5.01	-3.2	50.0
13C2 PFHxA	Ave	1.167	1.108		9.50	10.0	-5.0	30.0
13C2 PFDA	Ave	0.8763	0.7903		9.02	10.0	-9.8	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_003.d
 Lims ID: CCV L2
 Client ID:
 Sample Type: CCVL
 Inject. Date: 11-Dec-2016 12:02:56 ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L2 CCV L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:28 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161205-37524.b\\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002


First Level Reviewer: westendorfc Date: 11-Dec-2016 12:36:03

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0 17.612 17.612 0.0	1.000	949190	21.1	569				
\$ 2 13C2 PFHxA								
315.0 > 270.0 18.585 18.585 0.0	1.000	889175	9.50	27996				
3 Perfluorohexanesulfonic acid								
399.0 > 80.0 19.356 19.356 0.0	1.000	377563	6.57	8685				
4 Perfluoroheptanoic acid								
363.0 > 319.0 19.391 19.391 0.0	1.000	294373	3.02	4026				
* 5 13C2-PFOA								
415.0 > 370.0 20.058 20.058 0.0		802153	10.0	20748				
6 Perfluorooctanoic acid								
413.0 > 369.0 20.058 20.058 0.0	1.000	432862	5.19	181				
7 Perfluorooctane sulfonic acid								
499.0 > 80.0 20.619 20.619 0.0	1.000	581183	8.69	5853				
* 8 13C4 PFOS								
503.0 > 80.0 20.679 20.679 0.0		1836390	28.7	27361				
9 Perfluorononanoic acid								
463.0 > 419.0 20.750 20.750 0.0	1.000	441361	4.85	3888				
\$ 10 13C2 PFDA								
515.0 > 470.0 21.480 21.480 0.0	1.000	633909	9.02	19884				

Reagents:

LC537-L2_00014 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_003.d
 Injection Date: 11-Dec-2016 12:02:56 Instrument ID: A6
 Lims ID: CCV L2
 Client ID:
 Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-141573/4 Calibration Date: 12/11/2016 12:32
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 11DEC2016A6A_004.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.7094		136	135	1.1	30.0
Perfluorohexanesulfonic acid	Ave	0.8980	1.006		50.9	45.4	12.1	30.0
Perfluoroheptanoic acid	Ave	1.215	1.198		15.1	15.3	-1.4	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	1.111		32.5	30.4	6.8	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	1.179		67.8	60.1	12.9	30.0
Perfluorononanoic acid	Ave	1.134	1.179		30.6	29.5	3.9	30.0
13C2 PFHxA	Ave	1.167	1.324		11.4	10.0	13.5	30.0
13C2 PFDA	Ave	0.8763	0.9513		10.9	10.0	8.6	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_004.d
 Lims ID: CCV L5
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 11-Dec-2016 12:32:31 ALS Bottle#: 5 Worklist Smp#: 4
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L5 CCV L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002


Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.611 17.611 0.0 1.000 5924235 136.2 7103
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 1192925 11.4 38299
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.356 19.356 0.0 1.000 2833340 50.9 25601
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.391 19.391 0.0 1.000 1648862 15.1 5901
 * 5 13C2-PFOA
 415.0 > 370.0 20.058 20.058 0.0 1.000 900761 10.0 23241
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.058 20.058 0.0 1.000 3044806 32.5 748
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 4393612 67.8 5235
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1.000 1778917 28.7 16519
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 3128230 30.6 15591
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 856896 10.9 27089

Reagents:

LC537-L5_00017 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_004.d
 Injection Date: 11-Dec-2016 12:32:31 Instrument ID: A6
 Lims ID: CCV L5
 Client ID:
 Operator ID: CBW ALS Bottle#: 5 Worklist Smp#: 4
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-141573/17 Calibration Date: 12/11/2016 18:57
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 11DEC2016A6A_017.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.7477		48.1	45.1	6.6	30.0
Perfluorohexanesulfonic acid	Ave	0.8980	0.9640		16.3	15.2	7.4	30.0
Perfluoroheptanoic acid	Ave	1.215	1.333		5.61	5.12	9.7	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	1.051		10.3	10.2	1.1	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	1.062		20.5	20.1	1.7	30.0
Perfluorononanoic acid	Ave	1.134	1.172		10.2	9.87	3.3	30.0
13C2 PFHxA	Ave	1.167	1.173		10.1	10.0	0.5	30.0
13C2 PFDA	Ave	0.8763	0.8911		10.2	10.0	1.7	30.0

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-141574/17 Calibration Date: 12/11/2016 18:57
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 11DEC2016A6A_017.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.7477		48.1	45.1	6.6	30.0
Perfluorohexanesulfonic acid	Ave	0.8980	0.9640		16.3	15.2	7.4	30.0
Perfluoroheptanoic acid	Ave	1.215	1.333		5.61	5.12	9.7	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	1.051		10.3	10.2	1.1	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	1.062		20.5	20.1	1.7	30.0
Perfluorononanoic acid	Ave	1.134	1.172		10.2	9.87	3.3	30.0
13C2 PFHxA	Ave	1.167	1.173		10.1	10.0	0.5	30.0
13C2 PFDA	Ave	0.8763	0.8911		10.2	10.0	1.7	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_017.d
 Lims ID: CCV L3
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 11-Dec-2016 18:57:16 ALS Bottle#: 3 Worklist Smp#: 17
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L3 CCV L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:44 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.608 17.608 0.0 1.000 2022947 48.1 841
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 944910 10.1 29815
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.356 19.356 0.0 1.000 879241 16.3 20597
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.391 19.391 0.0 1.000 549674 5.61 6137
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.047 0.0 1.000 805687 10.0 20791
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.058 20.058 0.0 1.000 863299 10.3 241
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 1282046 20.5 12589
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1.000 1720352 28.7 22446
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 931931 10.2 10954
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 717912 10.2 22552

Reagents:

LC537-L3_00016 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_017.d
 Lims ID: CCV L3
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 11-Dec-2016 18:57:16 ALS Bottle#: 3 Worklist Smp#: 17
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L3 CCV L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:44 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002


Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.608 17.608 0.0 1.000 2022947 48.1 841
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 944910 10.1 29815
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.356 19.356 0.0 1.000 879241 16.3 20597
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.391 19.391 0.0 1.000 549674 5.61 6137
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.047 0.0 1.000 805687 10.0 20791
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.058 20.058 0.0 1.000 863299 10.3 241
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 1282046 20.5 12589
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1.000 1720352 28.7 22446
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 931931 10.2 10954
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.480 21.480 0.0 1.000 717912 10.2 22552


Reagents:

LC537-L3_00016 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_017.d
 Injection Date: 11-Dec-2016 18:57:16 Instrument ID: A6
 Lims ID: CCV L3
 Client ID:
 Operator ID: CBW ALS Bottle#: 3 Worklist Smp#: 17
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_017.d
 Injection Date: 11-Dec-2016 18:57:16 Instrument ID: A6
 Lims ID: CCV L3
 Client ID:
 Operator ID: CBW ALS Bottle#: 3 Worklist Smp#: 17
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.: _____
Lab Sample ID: CCV 320-141574/29 Calibration Date: 12/12/2016 00:52
Instrument ID: A6 Calib Start Date: 12/05/2016 17:26
GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/05/2016 19:54
Lab File ID: 11DEC2016A6A_029.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7015	0.7417		142	135	5.7	30.0
Perfluorohexanesulfonic acid	Ave	0.8980	1.056		53.4	45.4	17.6	30.0
Perfluoroheptanoic acid	Ave	1.215	1.187		14.9	15.3	-2.3	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.040	1.100		32.2	30.4	5.7	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.044	1.251		72.0	60.1	19.8	30.0
Perfluorononanoic acid	Ave	1.134	1.223		31.8	29.5	7.8	30.0
13C2 PFHxA	Ave	1.167	1.297		11.1	10.0	11.2	30.0
13C2 PFDA	Ave	0.8763	0.9257		10.6	10.0	5.6	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_029.d
 Lims ID: CCV L5
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 12-Dec-2016 00:52:23 ALS Bottle#: 5 Worklist Smp#: 29
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L5 CCV L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 16:03:06 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002


Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.612 17.612 0.0 1.000 6371773 142.4 26361
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 1149765 11.1 37115
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.344 19.344 0.0 1.000 3058257 53.4 68628
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.380 19.380 0.0 1.000 1607457 14.9 23861
 * 5 13C2-PFOA
 415.0 > 370.0 20.047 20.047 0.0 1.000 886505 10.0 22969
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.047 20.047 0.0 1.000 2965495 32.2 1550
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 4797861 72.0 24477
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1.000 1829934 28.7 37378
 9 Perfluorononanoic acid
 463.0 > 419.0 20.750 20.750 0.0 1.000 3194560 31.8 47943
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.471 0.0 1.000 820622 10.6 25626

Reagents:

LC537-L5_00017 Amount Added: 1.00 Units: mL

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_029.d
 Injection Date: 12-Dec-2016 00:52:23 Instrument ID: A6
 Lims ID: CCV L5
 Client ID:
 Operator ID: CBW ALS Bottle#: 5 Worklist Smp#: 29
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: MB 320-140632/1-A
 Matrix: Water Lab File ID: 11DEC2016A6A_006.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 250 (mL) Date Analyzed: 12/11/2016 13:31
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	106		70-130
STL00996	13C2 PFDA	104		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_006.d
 Lims ID: MB 320-140632/1-A
 Client ID:
 Sample Type: MB
 Inject. Date: 11-Dec-2016 13:31:42 ALS Bottle#: 17 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: mb 320-140632/1-a BOX 17
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:00:54

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------


\$ 2 13C2 PFHxA
 315.0 > 270.0 18.585 18.585 0.0 1.000 941882 10.6 30302
 * 5 13C2-PFOA
 415.0 > 370.0 20.059 20.058 0.001 764515 10.0 20056
 6 Perfluorooctanoic acid M
 413.0 > 369.0 20.047 20.058 -0.011 1.000 1174 0.0148 0.3 M
 7 Perfluorooctane sulfonic acid M
 499.0 > 80.0 20.655 20.619 0.036 1.000 732 0.0101 14.8 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.679 20.679 0.0 1993596 28.7 51558
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.471 21.480 -0.009 1.000 696850 10.4 21920

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_006.d
 Injection Date: 11-Dec-2016 13:31:42 Instrument ID: A6
 Lims ID: MB 320-140632/1-A
 Client ID:
 Operator ID: CBW ALS Bottle#: 17 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_006.d
 Lims ID: MB 320-140632/1-A
 Client ID:
 Sample Type: MB
 Inject. Date: 11-Dec-2016 13:31:42 ALS Bottle#: 17 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: mb 320-140632/1-a BOX 17
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 15:00:54

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.6	105.61
\$ 10 13C2 PFDA	10.0	10.4	104.02

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_006.d
 Injection Date: 11-Dec-2016 13:31:42 Instrument ID: A6
 Lims ID: MB 320-140632/1-A
 Client ID:
 Operator ID: CBW ALS Bottle#: 17 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1


Not Detected
 Expected RT: 20.62

Processing Integration Results



RT: 20.66
 Area: 732
 Amount: 0.010087
 Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 12-Dec-2016 15:00:54

Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_006.d
 Injection Date: 11-Dec-2016 13:31:42 Instrument ID: A6
 Lims ID: MB 320-140632/1-A
 Client ID:
 Operator ID: CBW ALS Bottle#: 17 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL
 Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Processing Integration Results

RT: 20.30
 Area: 33028
 Amount: 0.415227
 Amount Units: ng/ml



Manual Integration Results

RT: 20.05
 Area: 1174
 Amount: 0.014759
 Amount Units: ng/ml



Reviewer: barnettj, 12-Dec-2016 15:00:54

Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: LCS 320-140632/2-A
 Matrix: Water Lab File ID: 11DEC2016A6A_007.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 250 (mL) Date Analyzed: 12/11/2016 14:01
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.124		0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.0619		0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.275		0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	113		70-130
STL00996	13C2 PFDA	110		70-130

TestAmerica Sacramento
Target Compound Quantitation Report


Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_007.d
 Lims ID: LCS 320-140632/2-A
 Client ID:
 Sample Type: LCS
 Inject. Date: 11-Dec-2016 14:01:18 ALS Bottle#: 18 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcs 320-140632/2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICAL File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 13:50:34

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	--------	--------	--------	----------	--------------	-----	-------

1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.608	17.611	-0.003	1.000	3545743	68.9	1681	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.585	18.585	0.0	1.000	1070896	11.3	22997	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.356	19.356	0.0	1.000	1595835	24.2	36853	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.392	19.391	0.001	1.000	827715	8.37	5358	
* 5 13C2-PFOA								
415.0 > 370.0	20.059	20.058	0.001		813368	10.0	21246	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.059	20.058	0.001	1.000	1308804	15.5	572	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	2375806	31.0	24256	
* 8 13C4 PFOS								
503.0 > 80.0	20.679	20.679	0.0		2105344	28.7	16826	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.750	0.0	1.000	1500116	16.3	15932	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.480	21.480	0.0	1.000	786329	11.0	24953	

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_007.d
 Injection Date: 11-Dec-2016 14:01:18 Instrument ID: A6
 Lims ID: LCS 320-140632/2-A
 Client ID:
 Operator ID: CBW ALS Bottle#: 18 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_007.d
 Lims ID: LCS 320-140632/2-A
 Client ID:
 Sample Type: LCS
 Inject. Date: 11-Dec-2016 14:01:18 ALS Bottle#: 18 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcs 320-140632/2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

First Level Reviewer: barnettj Date: 12-Dec-2016 13:50:34

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.3	112.87
\$ 10 13C2 PFDA	10.0	11.0	110.33

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: LCSD 320-140632/3-A
 Matrix: Water Lab File ID: 11DEC2016A6A_008.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 250 (mL) Date Analyzed: 12/11/2016 14:30
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.129		0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.0627		0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.294		0.14	0.11	0.048


CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	117		70-130
STL00996	13C2 PFDA	111		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_008.d
 Lims ID: LCSD 320-140632/3-A
 Client ID:
 Sample Type: LCSD
 Inject. Date: 11-Dec-2016 14:30:53 ALS Bottle#: 19 Worklist Smp#: 8
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcsm 320-140632/3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.612	17.611	0.001	1.000	3361963	73.5	1101	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.585	18.585	0.0	1.000	1000894	11.7	32019	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.356	19.356	0.0	1.000	1494371	25.5	22827	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.392	19.391	0.001	1.000	779470	8.74	5329	
* 5 13C2-PFOA								
415.0 > 370.0	20.058	20.058	0.0		733684	10.0	19232	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.058	20.058	0.0	1.000	1196673	15.7	496	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	2187840	32.2	12847	
* 8 13C4 PFOS								
503.0 > 80.0	20.679	20.679	0.0		1868957	28.7	27574	
9 Perfluorononanoic acid								
463.0 > 419.0	20.750	20.750	0.0	1.000	1424554	17.1	16768	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.471	21.480	-0.009	1.000	711593	11.1	22420	

TestAmerica Sacramento
 Data File: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b\\11DEC2016A6A_008.d
 Injection Date: 11-Dec-2016 14:30:53 Instrument ID: A6
 Lims ID: LCSD 320-140632/3-A
 Client ID:
 Operator ID: CBW ALS Bottle#: 19 Worklist Smp#: 8
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Method: 537_A6 Limit Group: LC 537 ICAL



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\11DEC2016A6A_008.d
 Lims ID: LCSD 320-140632/3-A
 Client ID:
 Sample Type: LCSD
 Inject. Date: 11-Dec-2016 14:30:53 ALS Bottle#: 19 Worklist Smp#: 8
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcsm 320-140632/3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20161211-37708.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 12-Dec-2016 15:39:30 Calib Date: 05-Dec-2016 19:54:00
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161205-37524.b\05DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK002

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.7	116.95
\$ 10 13C2 PFDA	10.0	11.1	110.68

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Instrument ID: A6

Start Date: 12/05/2016 17:26

Analysis Batch Number: 140688

End Date: 12/06/2016 02:48

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
STD 320-140688/2 IC		12/05/2016 17:26	1	05DEC2016A6A_00 4.d	Acquity 2.1 (mm)
STD 320-140688/3 IC		12/05/2016 17:55	1	05DEC2016A6A_00 5.d	Acquity 2.1 (mm)
STD 320-140688/4 IC		12/05/2016 18:25	1	05DEC2016A6A_00 6.d	Acquity 2.1 (mm)
STD 320-140688/5 ICISAV		12/05/2016 18:54	1	05DEC2016A6A_00 7.d	Acquity 2.1 (mm)
STD 320-140688/6 IC		12/05/2016 19:24	1	05DEC2016A6A_00 8.d	Acquity 2.1 (mm)
STD 320-140688/7 IC		12/05/2016 19:54	1	05DEC2016A6A_00 9.d	Acquity 2.1 (mm)
ZZZZZ		12/05/2016 20:23	1		Acquity 2.1 (mm)
CCV 320-140688/9 CCVL		12/05/2016 20:53	1	05DEC2016A6A_01 1.d	Acquity 2.1 (mm)
ZZZZZ		12/05/2016 21:22	1		Acquity 2.1 (mm)
ICV 320-140688/11		12/05/2016 21:52	1	05DEC2016A6A_01 3.d	Acquity 2.1 (mm)
ZZZZZ		12/05/2016 22:22	1		Acquity 2.1 (mm)
ZZZZZ		12/05/2016 22:51	1		Acquity 2.1 (mm)
ZZZZZ		12/05/2016 23:21	1		Acquity 2.1 (mm)
ZZZZZ		12/05/2016 23:50	1		Acquity 2.1 (mm)
ZZZZZ		12/06/2016 00:20	1		Acquity 2.1 (mm)
ZZZZZ		12/06/2016 00:49	1		Acquity 2.1 (mm)
ZZZZZ		12/06/2016 01:19	1		Acquity 2.1 (mm)
ZZZZZ		12/06/2016 01:49	1		Acquity 2.1 (mm)
ZZZZZ		12/06/2016 02:18	1		Acquity 2.1 (mm)
CCV 320-140688/21 CCVIS		12/06/2016 02:48	1		Acquity 2.1 (mm)

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica SacramentoJob No.: 320-23971-1

SDG No.:

Instrument ID: A6Start Date: 12/11/2016 12:02Analysis Batch Number: 141573End Date: 12/11/2016 18:57

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
CCV 320-141573/3 CCVL		12/11/2016 12:02	1	11DEC2016A6A_00 3.d	Acquity 2.1(mm)
CCV 320-141573/4 CCVIS		12/11/2016 12:32	1	11DEC2016A6A_00 4.d	Acquity 2.1(mm)
ZZZZZ		12/11/2016 13:02	1		Acquity 2.1(mm)
MB 320-140632/1-A		12/11/2016 13:31	1	11DEC2016A6A_00 6.d	Acquity 2.1(mm)
LCS 320-140632/2-A		12/11/2016 14:01	1	11DEC2016A6A_00 7.d	Acquity 2.1(mm)
LCSD 320-140632/3-A		12/11/2016 14:30	1	11DEC2016A6A_00 8.d	Acquity 2.1(mm)
320-23971-1		12/11/2016 15:00	1	11DEC2016A6A_00 9.d	Acquity 2.1(mm)
320-23971-2		12/11/2016 15:30	1	11DEC2016A6A_01 0.d	Acquity 2.1(mm)
320-23971-3		12/11/2016 15:59	1	11DEC2016A6A_01 1.d	Acquity 2.1(mm)
320-23971-4		12/11/2016 16:29	1	11DEC2016A6A_01 2.d	Acquity 2.1(mm)
320-23971-5		12/11/2016 16:58	1	11DEC2016A6A_01 3.d	Acquity 2.1(mm)
320-23971-6		12/11/2016 17:28	1	11DEC2016A6A_01 4.d	Acquity 2.1(mm)
320-23971-7		12/11/2016 17:58	1	11DEC2016A6A_01 5.d	Acquity 2.1(mm)
CCV 320-141573/17 CCVIS		12/11/2016 18:57	1	11DEC2016A6A_01 7.d	Acquity 2.1(mm)

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Instrument ID: A6

Start Date: 12/11/2016 18:57

Analysis Batch Number: 141574

End Date: 12/12/2016 00:52

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
CCV 320-141574/17 CCVIS		12/11/2016 18:57	1	11DEC2016A6A_01 7.d	Acquity 2.1 (mm)
ZZZZZ		12/11/2016 19:26	1		Acquity 2.1 (mm)
320-23971-8		12/11/2016 19:56	1	11DEC2016A6A_01 9.d	Acquity 2.1 (mm)
320-23971-9		12/11/2016 20:26	1	11DEC2016A6A_02 0.d	Acquity 2.1 (mm)
320-23971-10		12/11/2016 20:55	1	11DEC2016A6A_02 1.d	Acquity 2.1 (mm)
ZZZZZ		12/11/2016 21:25	1		Acquity 2.1 (mm)
ZZZZZ		12/11/2016 21:54	1		Acquity 2.1 (mm)
ZZZZZ		12/11/2016 22:24	1		Acquity 2.1 (mm)
ZZZZZ		12/11/2016 22:54	1		Acquity 2.1 (mm)
ZZZZZ		12/11/2016 23:23	1		Acquity 2.1 (mm)
ZZZZZ		12/11/2016 23:53	1		Acquity 2.1 (mm)
CCV 320-141574/29 CCVIS		12/12/2016 00:52	1	11DEC2016A6A_02 9.d	Acquity 2.1 (mm)

LCMS BATCH WORKSHEET

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Batch Number: 140632

Batch Start Date: 12/05/16 11:42

Batch Analyst: Sharifi, Nooshin

Batch Method: 537

Batch End Date: 12/06/16 18:30

Lab Sample ID	Client Sample ID	Method Chain	Basis	GrossWeight	TareWeight	InitialAmount	FinalAmount	ReceivedpH	LC537-IS 00025
MB 320-140632/1		537, 537				250 mL	1.0 mL	7 SU	20 uL
LCS 320-140632/2		537, 537				250 mL	1.0 mL	7 SU	20 uL
LCSD 320-140632/3		537, 537				250 mL	1.0 mL	7 SU	20 uL
320-23971-A-1	WI-AF-2RW07-1116	537, 537	T	295.87 g	38.00 g	257.9 mL	1.0 mL	7 SU	20 uL
320-23971-A-2	WI-AF-2FB07-1116	537, 537	T	292.77 g	26.56 g	266.2 mL	1.0 mL	9 SU	20 uL
320-23971-A-3	WI-AF-2RW08-1116	537, 537	T	296.15 g	27.13 g	269 mL	1.0 mL	9 SU	20 uL
320-23971-A-4	WI-AF-2FB08-1116	537, 537	T	290.78 g	27.11 g	263.7 mL	1.0 mL	9 SU	20 uL
320-23971-A-5	WI-AF-3RW10-1116	537, 537	T	291.63 g	27.59 g	264 mL	1.0 mL	9 SU	20 uL
320-23971-B-6	WI-AF-3FB10-1116	537, 537	T	297.66 g	27.53 g	270.1 mL	1.0 mL	9 SU	20 uL
320-23971-A-7	WI-AF-3RW11-1116	537, 537	T	296.13 g	27.21 g	268.9 mL	1.0 mL	9 SU	20 uL
320-23971-A-8	WI-AF-3FB11-1116	537, 537	T	292.51 g	27.39 g	265.1 mL	1.0 mL	9 SU	20 uL
320-23971-A-9	WI-AF-3RW12-1116	537, 537	T	282.97 g	26.61 g	256.4 mL	1.0 mL	7 SU	20 uL
320-23971-A-10	WI-AF-3FB12-1116	537, 537	T	292.38 g	27.21 g	265.2 mL	1.0 mL	9 SU	20 uL

Lab Sample ID	Client Sample ID	Method Chain	Basis	LC537-MSP 00014	LC537-SU 00022	AnalysisComment			
MB 320-140632/1		537, 537		50 uL	Chlorine ND				
LCS 320-140632/2		537, 537		50 uL	Chlorine ND				
LCSD 320-140632/3		537, 537		50 uL	Chlorine ND				
320-23971-A-1	WI-AF-2RW07-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-2	WI-AF-2FB07-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-3	WI-AF-2RW08-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-4	WI-AF-2FB08-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-5	WI-AF-3RW10-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-B-6	WI-AF-3FB10-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-7	WI-AF-3RW11-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-8	WI-AF-3FB11-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-9	WI-AF-3RW12-1116	537, 537	T	50 uL	Chlorine ND				
320-23971-A-10	WI-AF-3FB12-1116	537, 537	T	50 uL	Chlorine ND				

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

LCMS BATCH WORKSHEET

Lab Name: TestAmerica Sacramento

Job No.: 320-23971-1

SDG No.:

Batch Number: 140632

Batch Start Date: 12/05/16 11:42

Batch Analyst: Sharifi, Nooshin

Batch Method: 537

Batch End Date: 12/06/16 18:30

Batch Notes	
Manifold ID	2,4
Methanol ID	789820
Pipette ID	MD05306
Analyst ID - IS Reagent Drop	ERW
Analyst ID - IS Reagent Drop Witness	NSH
Analyst ID - SU Reagent Drop	NSH
Analyst ID - SU Reagent Drop Witness	VPM
Analyst ID - TA Reagent Drop	NSH
Analyst ID - TA Reagent Drop Witness	VPM
SPE Cartridge ID	6332578-03
Trizma ID	SLBN2122V
Reagent Water ID	11/29/16

Basis	Basis Description
T	Total/NA

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

Method 537 CCV/Data Review Checklist

A6

Job No: 23971, 23970 Instrument ID & Date: 12-11-16 ICAL Batch: 140688
 Extraction Batch: 140632 Worklist #: 37708 TALS Batch: 141573, 141574, 141758

Review Items	— Level 1 —			Level 2
	Yes	No	N/A	
Initial Calibration				
1. Is ICAL verified and locked in Chrom & TALS?	✓			✓
2. Is ICV properly linked in TALS?	✓			✓
Continuing Calibration				
1. Low-range CCV injected at start of analytical run? CCV injected after every 10 samples and at the end of the analytical run and alternated between Low-range, Mid-range and High-range?	✓			✓
2. If sequence was not after an ICAL was a low and mid range CCV injected at the start of the analytical run?	✓			✓
3. Native compounds and surrogates in control? Low-range within ±50% of true value Mid and High-range within ±30% of true value	✓			✓
4. Internal Standard areas in control? Areas ≥ 50% of average area of the ICAL and 70-140% of the most recent CCV.	✓			✓
Client Samples & QC Sample Results				
1. Were preparation and analysis done within holding times?	✓			✓
2. Are Chromatograms reviewed and spectra verified?	✓			✓
3. Are positive results within calibration range?	✓			✓
4. Dilutions due to target cpds? Dilutions due to non-targets?			✓	✓
5. All target compounds in MB < 1/3 RL? (Requires NCM if "no.")	✓			✓
6. Are target constituents in LCS/LCSD within method control limits?	✓			✓
7. Internal Standard areas in control for all samples and QC reported? ±50% from the average area of the ICAL and 70-140% of the most recent CCV	✓			✓
8. Do results (e.g., dilutions/trip blanks) make sense?	✓			✓
9. Are MS/MSD recoveries and RPDs within method control limits?			✓	✓
10. Are all QC samples properly linked in TALS?	✓		✓	✓
11. All manual integrations appropriate and completely documented?	✓			✓
12. Are nonconformances documented as NCMs?	✓			✓
13. Are all Chrom graphics uploaded?	✓			✓

 1st Level Reviewer / Date: JRB 12-13-16

 2nd Level Reviewer / Date: Murphy 12-13-2016

 NCM # and Comments: 72739

A6
 Instrument ID & Date: 12-5-16 Worklist#: 37524

ICAL Batch: 140688, 140689 Calibration ID number: 26888, 26889

Review Items	-- Level 1 --			Level 2
	Yes	No	N/A	
Initial Calibration				
1. Mass calibration, as needed, verified by full scan of PFC stock standard. All PFC ions used for quantitation are within 0.3 m/z of true mass?	✓			✓
2. Responses increase with increasing concentration?	✓			✓
3. Fit used (circle): <u>Average</u> Linear (1/x ²)Linear Quadratic (6 points minimum)				
4. Meets fit criteria? Intercept ≤ 1/2 RL RSD ≤ 30% for Average $R^2 \geq 0.990$ for Linear $R^2 \geq 0.990$ for Quadratic				
NOTE: "Force through Zero" must be used and weighted if needed	✓			✓
5. If quadratic fit used the curve does not "bend over".			✓	
6. Feed calibration points into the calculated curve. Are points ≤MRL within ±50% of true value? Are points >MRL within ±30% of true value?	✓			✓
7. Any carryover from the high calibration point must be < 1/3 RL	✓			✓
8. Asymmetry check meets criteria for the first two eluting peaks? (0.8 - 1.5).	✓			✓
9. Is the asymmetry check scanned and linked in TALS to the calibration point?	✓			✓
10. Is ICV (2 nd source) ± 30% of true value?	✓			✓
11. Is ICV (2 nd source) internal standards ±50% of average area of the ICAL?	✓			✓
12. ICAL locked in Chrom and uploaded to TALS?	✓			✓
13. ICAL locked in TALS and scanned?				✓

1st Level Reviewer / Date: JRB 12-6-16

2nd Level Reviewer / Date: R. Hurd, 12/7/16

NCM # and Comments: _____

TestAmerica Laboratories
Worklist QC Batch Report

Worklist Name: 11DEC2016A_A6 537 Worklist Number: 37708
 Instrument Name: A6 Chrom Method: 537_A6
 Data Directory: \\ChromNA\\Sacramento\\ChromData\\A6\\20161211-37708.b
 QC Batching: Enabled Limit Group Batching: Enabled

QC Batch: 1	LC 537 ICAL Raw Batch: 141573
# 1 RB	# 1 RB
# 2 RB	# 2 RB
# 3 CCV L2	# 3 CCV L2
# 4 CCV L5	# 4 CCV L5
# 5 RB	# 5 RB
# 6 MB 320-140632/1-A	# 6 MB 320-140632/1-A
# 7 LCS 320-140632/2-A	# 7 LCS 320-140632/2-A
# 8 LCSD 320-140632/3-A	# 8 LCSD 320-140632/3-A
# 9 320-23971-A-1-A	# 9 320-23971-A-1-A
#10 320-23971-A-2-A	#10 320-23971-A-2-A
#11 320-23971-A-3-A	#11 320-23971-A-3-A
#12 320-23971-A-4-A	#12 320-23971-A-4-A
#13 320-23971-A-5-A	#13 320-23971-A-5-A
#14 320-23971-B-6-A	#14 320-23971-B-6-A
#15 320-23971-A-7-A	#15 320-23971-A-7-A
#16 RB	#16 RB
#17 CCV L3	#17 CCV L3

QC Batch: 2	LC 537 ICAL Raw Batch: 141574
#17 CCV L3	#17 CCV L3
#18 RB	#18 RB
#19 320-23971-A-8-A	#19 320-23971-A-8-A
#20 320-23971-A-9-A	#20 320-23971-A-9-A
#21 320-23971-A-10-A	#21 320-23971-A-10-A
#22 320-23970-A-1-A	#22 320-23970-A-1-A
#23 320-23970-A-2-A	#23 320-23970-A-2-A
#24 320-23970-A-3-A	#24 320-23970-A-3-A
#25 320-23970-A-4-A	#25 320-23970-A-4-A
#26 320-23970-A-5-A	#26 320-23970-A-5-A
#27 320-23970-A-6-A	#27 320-23970-A-6-A
#28 RB	#28 RB
#29 CCV L5	#29 CCV L5

Sur out.

QC Batch: 3	LC 537 ICAL Raw Batch: 141575
#29 CCV L5	#29 CCV L5
#30 RB	#30 RB
#31 MB 320-140697/1-A	#31 MB 320-140697/1-A
#32 LCS 320-140697/2-A	#32 LCS 320-140697/2-A
#33 LCSD 320-140697/3-A	#33 LCSD 320-140697/3-A
#34 320-24005-A-1-A	#34 320-24005-A-1-A
#35 320-24005-A-2-A	#35 320-24005-A-2-A
#36 320-24005-A-3-A	#36 320-24005-A-3-A
#37 320-24005-A-4-A	#37 320-24005-A-4-A
#38 320-24005-A-5-A	#38 320-24005-A-5-A
#39 320-24005-A-6-A	#39 320-24005-A-6-A
#40 320-24005-A-7-A	#40 320-24005-A-7-A
#41 RB	#41 RB
#42 CCV L3	#42 CCV L3

QC Batch: 4	LC 537 ICAL Raw Batch: 141576
#42 CCV L3	#42 CCV L3
#43 RB	#43 RB
#44 320-24005-A-8-A	#44 320-24005-A-8-A
#45 320-24007-A-1-A	#45 320-24007-A-1-A
#46 320-24007-A-2-A	#46 320-24007-A-2-A
#47 RB	#47 RB
#48 CCV L5	#48 CCV L5

QC Batch: 5	LC 537 ICAL Raw Batch: 141577
#48 CCV L5	#48 CCV L5
#49 RB	#49 RB
#50 320-23719-A-4-A	#50 320-23719-A-4-A
#51 320-23719-A-5-A	#51 320-23719-A-5-A
#52 320-23719-A-6-A	#52 320-23719-A-6-A
#53 320-23719-A-7-A	#53 320-23719-A-7-A
#54 320-23719-A-8-A	#54 320-23719-A-8-A
#55 320-23719-A-9-A	#55 320-23719-A-9-A
#56 320-23720-A-1-A	#56 320-23720-A-1-A
#57 320-23720-A-2-A	#57 320-23720-A-2-A
#58 320-23721-A-1-A	#58 320-23721-A-1-A
#59 320-23722-A-1-A	#59 320-23722-A-1-A
#60 RB	#60 RB
#61 CCV L3	#61 CCV L3

QC Batch: 6	LC 537 ICAL Raw Batch: 141758
#61 CCV L3	#61 CCV L3
#79 RB	#79 RB
#62 320-23928-A-15-A	#62 320-23928-A-15-A
#63 320-23928-A-28-A	#63 320-23928-A-28-A
#64 320-23928-A-25-D MS	#64 320-23928-A-25-D MS
#65 320-23928-A-25-D MS	#65 320-23928-A-25-D MS
#66 320-23928-A-27-D MS	#66 320-23928-A-27-D MS
#67 320-23928-A-27-A	#67 320-23928-A-27-A
#68 320-23928-A-27-D MS	#68 320-23928-A-27-D MS
#69 320-23970-A-1-A	#69 320-23970-A-1-A
#70 RB	#70 RB
#71 CCV L5	#71 CCV L5

QC Batch: 7	LC 537 ICAL Raw Batch: 141759
#71 CCV L5	#71 CCV L5
#72 RB	#72 RB
#73 320-24005-A-1-A	#73 320-24005-A-1-A
#74 320-24005-A-3-A	#74 320-24005-A-3-A
#75 320-24005-A-8-A	#75 320-24005-A-8-A
#76 320-24007-A-1-A	#76 320-24007-A-1-A
#77 RB	#77 RB
#78 CCV L3	#78 CCV L3
#80 RB	#80 RB

Bushes

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Analyst: Sharifi, Nooshin

Batch Number: 320-140632

Method Code: 320-537_Prep-320

Batch Open: 12/5/2016 11:42:00AM
Batch End: 12/10/16 / 8:30

Extraction of Perfluorinated Alkyl Acids

Screened At 12/7/16
No dilutions needed

Input Sample Lab ID (Analytical Method)	SDG (Job #)	Gross Wt Tare Wt	Init Amnt Fin Amnt	Rcvd	PHs Adj1	Due Date Adj2	Analytical TAT	DIV Rank	Comments	Output Sample Lab ID
MB-320-140632/1 N/A	N/A	250 mL	7		N/A	N/A	N/A	N/A	Chlorine ND	
LCS-320-140632/2 N/A	N/A	250 mL	7		N/A	N/A	N/A	N/A	Chlorine ND	
LCS-D-320-140632/3 N/A	N/A	250 mL	7		N/A	N/A	N/A	N/A	Chlorine ND	
320-23971-A-1 (537_DOD5)	N/A (320-23971-1)	295.87 g	257.9 mL	7		12/6/16	5_Days	4	Chlorine ND	
320-23971-A-2 (537_DOD5)	N/A (320-23971-1)	292.77 g	266.2 mL	9		12/6/16	5_Days	4	Chlorine ND	
320-23971-A-3 (537_DOD5)	N/A (320-23971-1)	26.56 g	1.0 mL			12/6/16	5_Days	4	Chlorine ND	
320-23971-A-4 (537_DOD5)	N/A (320-23971-1)	296.15 g	269 mL	9		12/6/16	5_Days	4	Chlorine ND	
320-23971-A-5 (537_DOD5)	N/A (320-23971-1)	290.78 g	263.7 mL	9		12/6/16	5_Days	4	Chlorine ND	
320-23971-B-6 (537_DOD5)	N/A (320-23971-1)	291.63 g	264 mL	9		12/6/16	5_Days	4	Chlorine ND	
320-23971-A-7 (537_DOD5)	N/A (320-23971-1)	27.59 g	1.0 mL			12/6/16	5_Days	4	Chlorine ND	

12/13/2016

Printed : 12/5/2016

Page 1 of 5

TestAmerica Sacramento

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Analyst: Sharifi, Nooshin

Batch Open: 12/5/2016 11:42:00AM

Batch Number: 320-140632

Method Code: 320-537_Prep-320

Sample ID	Sample Description	Sample ID	Sample Description	Volume	Conc.	Start Date	End Date	Chlorine ND			
11 (537_DOD5)	320-23971-A-8 (320-23971-1) N/A	12 (537_DOD5)	320-23971-A-9 (320-23971-1) N/A	292.51 g 27.39 g	265.1 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND
13 (537_DOD5)	320-23971-A-10 (320-23971-1) N/A	14 (537_DOD5)	320-23970-A-1 (320-23970-1) N/A	282.97 g 26.61 g	256.4 mL 1.0 mL	7		12/6/16	5_Days	4	Chlorine ND
15 (537_DOD5)	320-23970-A-2 (320-23970-1) N/A	16 (537_DOD5)	320-23970-A-3 (320-23970-1) N/A	292.38 g 27.21 g	265.2 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND
17 (537_DOD5)	320-23970-A-4 (320-23970-1) N/A	18 (537_DOD5)	320-23970-A-5 (320-23970-1) N/A	293.55 g 27.83 g	255.7 mL 1.0 mL	7		12/6/16	5_Days	4	Chlorine ND
19 (537_DOD5)	320-23970-A-6 (320-23970-1) N/A			296.96 g 27.32 g	269.6 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND
				289.54 g 38.13 g	251.4 mL 1.0 mL	7		12/6/16	5_Days	4	Chlorine ND
				289.59 g 27.60 g	262 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND
				289.52 g 27.99 g	261.5 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND
				288.84 g 27.15 g	261.7 mL 1.0 mL	9		12/6/16	5_Days	4	Chlorine ND

Page 16 of 209 of 216

12/13/2016

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Analyst: Sharifi, Nooshin

Batch Open: 12/5/2016 11:42:00AM

Batch Number: 320-140632

Method Code: 320-537_Prep-320

Batch End:

Batch Notes	
Manifold ID	2,4
Trizma ID	SLBN2122V
SPE Cartridge ID	6332578-03
Methanol ID	789820
Reagent Water ID	11/29/16
Pipette ID	MD05306
Analyst ID - TA Reagent Drop	NSH
Analyst ID - TA Reagent Drop	✓PM
Witness	
Analyst ID - SU Reagent Drop	NSH
Analyst ID - SU Reagent Drop	✓PM
Witness	
Analyst ID - IS Reagent Drop	NSH
Analyst ID - IS Reagent Drop	✓PM
Witness	
Batch Comment	

Comments

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Analyst: Sharifi, Nooshin

Batch Open: 12/5/2016 11:42:00AM

Batch End:

Method Code: 320-537_Prep-320

Batch Number: 320-140632

Reagent Additions Worksheet

Lab ID	Reagent Code	Amount Added	Final Amount	By	Witness
MB 320-140632/1	LC537-SU_00022	50 uL	1.0 mL	N SH	12-5-16 VPM 12/6/16
LCS 320-140632/2	LC537-MSP_00014	50 uL	1.0 mL		
LCS 320-140632/2	LC537-SU_00022	50 uL	1.0 mL		
LCSD 320-140632/3	LC537-MSP_00014	50 uL	1.0 mL		
LCSD 320-140632/3	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-1	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-2	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-3	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-4	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-5	LC537-SU_00022	50 uL	1.0 mL		
320-23971-B-6	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-7	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-8	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-9	LC537-SU_00022	50 uL	1.0 mL		
320-23971-A-10	LC537-SU_00022	50 uL	1.0 mL		
320-23970-A-1	LC537-SU_00022	50 uL	1.0 mL		
320-23970-A-2	LC537-SU_00022	50 uL	1.0 mL		
320-23970-A-3	LC537-SU_00022	50 uL	1.0 mL		

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Batch Number: 320-140632

Method Code: 320-537 Prent-320

Analyst: Sharifi, Nooshin

Batch Open: 12/5/2016 11:42:00AM

Batch End

320-23970-A-4	LC537-SU_00022	50 uL	1.0 mL	N5A	12-5-16 VPM 12/05716
320-23970-A-5	LC537-SU_00022	50 uL	1.0 mL		
320-23970-A-6	LC537-SU_00022	50 uL	1.0 mL		

Reagent	Amount/Units	Lot#:
LC537-DS-00025	20 uL	
exp - 3/19/17		
0.5 - 1.434 ug/mL		

Sacramento
Preparation Data Review Checklist

Preparation Batch Number(s): 140632

Test: 537-1005 Rush

Earliest Holding Time: 12-14-16

Sample List Tab		1 st Level Reviewer	2 nd Level Reviewer
Samples identified to the correct method		/	/
All necessary NCMs filed (including holding time)		/	/
Method/sample/login/QAS checked and correct		/	/
Worksheet Tab		1 st Level Reviewer	2 nd Level Reviewer
All samples properly preserved		/	/
Weights in anticipated range and not targeted		/	/
All additional test requirements performed, documented, and uploaded to TALS correctly (e.g. final amount, initial amount, turbidity, and CI Check)		/	/
The pH is transcribed correctly in TALS		/	/
All additional information transcribed into TALS is correct and raw data is attached		/	/
Comments are transcribed correctly in TALS		/	/
Reagents Tab		1 st Level Reviewer	2 nd Level Reviewer
All necessary reagents not expired and entered into TALS		/	/
All spike amounts correct and added to necessary samples and QC		/	/
Batch Information		1 st Level Reviewer	2 nd Level Reviewer
Date and time accurate and entered into TALS correctly		/	/
All necessary 'batch information' complete and entered into TALS correctly		/	/

1st Level Reviewer: VPM

Date: 12/06/16

2nd Level Reviewer: ERW

Date: 12/6/16

Comments: _____

Shipping and Receiving Documents

TestAmerica Laboratories, Inc.

Form No. CA-C-WI-002, Rev. 4.3, dated 12/05/2013

Login Sample Receipt Checklist

Client: CH2M Hill Constructors, Inc.

Job Number: 320-23971-1

Login Number: 23971

List Source: TestAmerica Sacramento

List Number: 1

Creator: Turpen, Troy

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	Seals
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	
There are no discrepancies between the containers received and the COC.	False	Refer to Job Narrative for details.
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	N/A	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

Lab_Sample_ID	Contract_ID	DO_CTO_Number	Phase	Installation_ID	Sample_Name	Chem_Name	Analyte_ID	Analyte_Value	Original_Analyte_Units	Result_Lab_Qualifier	Validator_Qualifier	GC_Column_Type	Analysis_Result_Type	Result_Narrative	QC_Control_Limit_Type	QC_Limit_Code	QC_Accuracy_Upper	QC_Accuracy_Lower	Control_Limit	QC_Narrative	MDL	Detection_Limit_DL	QSM_Version	LOD	SDG	Analysis_Batch
320-23971-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW07-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.047	UG_L	U M	PR	TRG			00000000				5.0	0.015	0.047	0.058	320-23971-1	320-141573		
320-23971-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW07-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.023	UG_L	U	PR	TRG			00000000				5.0	0.0091	0.023	0.029	320-23971-1	320-141573		
320-23971-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW07-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.11	UG_L	U	PR	TRG			00000000				5.0	0.046	0.11	0.14	320-23971-1	320-141573		
320-23971-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW07-1116	13C2 PFHXA	103	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW07-1116	13C2 PFDA	118	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB07-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U M	PR	TRG			00000000				5.0	0.015	0.045	0.056	320-23971-1	320-141573		
320-23971-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB07-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.023	UG_L	U	PR	TRG			00000000				5.0	0.0088	0.023	0.028	320-23971-1	320-141573		
320-23971-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB07-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U	PR	TRG			00000000				5.0	0.045	0.10	0.13	320-23971-1	320-141573		
320-23971-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB07-1116	13C2 PFHXA	112	0.046	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB07-1116	13C2 PFDA	112	0.046	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW08-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U M	PR	TRG			00000000				5.0	0.014	0.045	0.056	320-23971-1	320-141573		
320-23971-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW08-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.022	UG_L	U M	PR	TRG			00000000				5.0	0.0088	0.022	0.028	320-23971-1	320-141573		
320-23971-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW08-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U	PR	TRG			00000000				5.0	0.044	0.10	0.13	320-23971-1	320-141573		
320-23971-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW08-1116	13C2 PFHXA	96	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2RW08-1116	13C2 PFDA	108	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0	0.015	0.046	0.057	320-23971-1	320-141573	
320-23971-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB08-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.046	UG_L	U	PR	TRG			00000000				5.0	0.0089	0.023	0.028	320-23971-1	320-141573		
320-23971-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB08-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.023	UG_L	U	PR	TRG			00000000				5.0	0.0089	0.023	0.028	320-23971-1	320-141573		
320-23971-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB08-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U	PR	TRG			00000000				5.0	0.045	0.10	0.13	320-23971-1	320-141573		
320-23971-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB08-1116	13C2 PFHXA	108	0.046	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-2FB08-1116	13C2 PFDA	105	0.046	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW10-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U	PR	TRG			00000000				5.0	0.015	0.045	0.057	320-23971-1	320-141573		
320-23971-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW10-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.023	UG_L	U	PR	TRG			00000000				5.0	0.0089	0.023	0.028	320-23971-1	320-141573		
320-23971-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW10-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U	PR	TRG			00000000				5.0	0.045	0.10	0.13	320-23971-1	320-141573		
320-23971-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW10-1116	13C2 PFHXA	110	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW10-1116	13C2 PFDA	111	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
320-23971-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB10-1116	Perfluoroctane Sulfonate (PFOS)	1763-23-1	0.044	UG_L	U M	PR	TRG			00000000				5.0	0.014	0.044	0.056	320-23971-1	320-141573		
320-23971-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB10-1116	Perfluoroctanoic acid (PFOA)	335-67-1	0.022	UG_L	U	PR	TRG			00000000				5.0	0.0087	0.022	0.028	320-23971-1	320-141573		
320-23971-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB10-1116	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U	PR	TRG			00000000				5.0	0.044	0.10	0.13	320-23971-1	320-141573		
320-23971-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB10-1116	13C2 PFHXA	117	0.045	PCT_REC		PR	SURR	SLSA	130	70	00000000				5.0				320-23971-1	320-141573	
32																										

**DATA VALIDATION SUMMARY REPORT
WHIDBEY ISLAND, WASHINGTON**

Client: CH2M HILL, Inc., Corvallis, Oregon
SDG: 320-23971
Laboratory: Test America, Sacramento, California
Site: Whidbey Island, CTO-0008, Washington
Date: December 21, 2016

PFCs			
EDS ID	Client Sample ID	Laboratory Sample ID	Matrix
1	WI-AF-2RW07-1116	320-23971-1	Water
2	WI-AF-2FB07-1116	320-23971-2	Water
3	WI-AF-2RW08-1116	320-23971-3	Water
4	WI-AF-2FB08-1116	320-23971-4	Water
5	WI-AF-3RW10-1116	320-23971-5	Water
6	WI-AF-3FB10-1116	320-23971-6	Water
7	WI-AF-3RW11-1116	320-23971-7	Water
8	WI-AF-3FB11-1116	320-23971-8	Water
9	WI-AF-3RW12-1116	320-23971-9	Water
10	WI-AF-3FB12-1116	320-23971-10	Water

A full data validation was performed on the analytical data for five water samples and five aqueous field blank samples collected on November 30, 2016 by CH2M HILL at the Whidbey Island site in Washington. The samples were analyzed under the EPA Method “Determination of Selected Perfluorinated Alkyl Acids in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS)”.

Specific method references are as follows:

Analysis

PFCs

Method References

USEPA Method 537 Rev 1.1 Modified

The data have been validated according to the protocols and quality control (QC) requirements of the analytical method, and the U.S. Department of Defense (DoD) Quality Systems Manual (QSM), Version 5.0 (DoD 2013) and the USEPA National Functional Guidelines for Organic Data Review as follows:

- The USEPA “Contract Laboratories Program National Functional Guidelines for Superfund Organic Methods Data Review,” August 2014;
- and the reviewer's professional judgment.

The following data quality indicators were reviewed for this report:

Organics

- Date Completeness, Case Narrative & Custody Documentation
- Holding times
- Gas Chromatography/Mass Spectrometry (GC/MS) Tuning
- Initial and continuing calibration summaries
- Method blank and field QC blank contamination
- Surrogate Spike recoveries
- Matrix Spike/Matrix Spike Duplicate (MS/MSD) recoveries
- Laboratory Control Sample/Laboratory Control Sample Duplicate (LCS/LCSD) recoveries
- Internal standard area and retention time summary forms
- Target Compound Identification
- Compound Quantitation
- Field Duplicate sample precision

A full (Level IV) data validation was performed with this review including a recalculation of 10% of the detected results in the samples.

Data Usability Assessment

There were no rejections of data.

Overall the data is acceptable for the intended purposes. There were no qualifications.

Perfluorinated Compounds (PFCs)

Data Completeness, Case Narrative & Custody Documentation

- The case narrative and chain-of-custody documentation were included in the data package as required.
- There was a discrepancy between the container label and the Chain-of-Custody for EDS Sample ID #10. The client and the laboratory resolved the issue and no action was required by the reviewer.

Holding Times

- All samples were extracted within 14 days for water samples and analyzed within 28 days.

GC/MS Tuning

- All criteria were met.

Initial Calibration

- All percent difference (%D) or correlation coefficients criteria were met.

Continuing Calibration

- All percent difference (%D) and RRF criteria were met.

Method Blank

- The method blanks were free of contamination.

Field QC Blank

- All field blank samples were free of contamination.

Surrogate Spike Recoveries

- All samples exhibited acceptable surrogate %R values.

Matrix Spike/Matrix Spike Duplicate (MS/MSD) Recoveries

- A MS/MSD sample was not collected.

Laboratory Control Samples/ Laboratory Control Sample Duplicates

- The LCS/LCSD samples exhibited acceptable percent recoveries (%R) and RPD values.

Internal Standard (IS) Area Performance

- All internal standards met response and retention time (RT) criteria.

Target Compound Identification

- All mass spectra and quantitation criteria were met.

Compound Quantitation

- All criteria were met. No action was required.

Field Duplicate Sample Precision

- Field duplicate samples were not collected.

Please contact the undersigned at (757) 564-0090 if you have any questions or need further information.

Signed:

Nancy Weaver

Nancy Weaver
Senior Chemist

Dated: 12/21/16

Data Qualifier	Definition
U	The analyte was analyzed for, but was not detected above the level of the reported sample quantitation limit.
J	The analyte is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
NJ	The analysis has been "tentatively identified" or "presumptively" as present and the associated numerical value is the estimated concentration in the samples.
UJ	The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
R	The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the samples.

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento	Job No.: 320-23971-1
SDG No.:	
Client Sample ID: WI-AF-2RW07-1116	Lab Sample ID: 320-23971-1
Matrix: Water	Lab File ID: 11DEC2016A6A_009.d
Analysis Method: 537	Date Collected: 11/30/2016 09:10
Extraction Method: 537	Date Extracted: 12/05/2016 11:42
Sample wt/vol: 257.9 (mL)	Date Analyzed: 12/11/2016 15:00
Con. Extract Vol.: 1.0 (mL)	Dilution Factor: 1
Injection Volume: 10 (uL)	GC Column: Acquity ID: 2.1 (mm)
% Moisture:	GPC Cleanup: (Y/N) N
Analysis Batch No.: 141573	Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U M	0.058	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0091
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	103		70-130
STL00996	13C2 PFDA	118		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2FB07-1116 Lab Sample ID: 320-23971-2
 Matrix: Water Lab File ID: 11DEC2016A6A_010.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:12
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 266.2 (mL) Date Analyzed: 12/11/2016 15:30
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	112		70-130
STL00996	13C2 PFDA	112		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-2RW08-1116 Lab Sample ID: 320-23971-3
 Matrix: Water Lab File ID: 11DEC2016A6A_011.d
 Analysis Method: 537 Date Collected: 11/30/2016 09:30
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 269 (mL) Date Analyzed: 12/11/2016 15:59
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U N	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U N	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	96		70-130
STL00996	13C2 PFDA	108		70-130

4

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.:
Client Sample ID: WI-AF-2FB08-1116 Lab Sample ID: 320-23971-4
Matrix: Water Lab File ID: 11DEC2016A6A_012.d
Analysis Method: 537 Date Collected: 11/30/2016 09:32
Extraction Method: 537 Date Extracted: 12/05/2016 11:42
Sample wt/vol: 263.7 (mL) Date Analyzed: 12/11/2016 16:29
Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
% Moisture: GPC Cleanup: (Y/N) N
Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	108		70-130
STL00996	13C2 PFDA	105		70-130

5

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
SDG No.:
Client Sample ID: WI-AF-3RW10-1116 Lab Sample ID: 320-23971-5
Matrix: Water Lab File ID: 11DEC2016A6A_013.d
Analysis Method: 537 Date Collected: 11/30/2016 10:00
Extraction Method: 537 Date Extracted: 12/05/2016 11:42
Sample wt/vol: 264 (mL) Date Analyzed: 12/11/2016 16:58
Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
% Moisture: GPC Cleanup: (Y/N) N
Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	111		70-130

6

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1

SDG No.:

Client Sample ID: WI-AF-3FB10-1116 Lab Sample ID: 320-23971-6

Matrix: Water Lab File ID: 11DEC2016A6A_014.d

Analysis Method: 537 Date Collected: 11/30/2016 10:00

Extraction Method: 537 Date Extracted: 12/05/2016 11:42

Sample wt/vol: 270.1 (mL) Date Analyzed: 12/11/2016 17:28

Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1

Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)

% Moisture: GPC Cleanup: (Y/N) N

Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.044	U M	0.056	0.044	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0087
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	117		70-130
STL00996	13C2 PFDA	106		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

7

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3RW11-1116 Lab Sample ID: 320-23971-7
 Matrix: Water Lab File ID: 11DEC2016A6A_015.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:30
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 268.9 (mL) Date Analyzed: 12/11/2016 17:58
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141573 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	109		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

8

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1
 SDG No.:
 Client Sample ID: WI-AF-3FB11-1116 Lab Sample ID: 320-23971-8
 Matrix: Water Lab File ID: 11DEC2016A6A_019.d
 Analysis Method: 537 Date Collected: 11/30/2016 10:31
 Extraction Method: 537 Date Extracted: 12/05/2016 11:42
 Sample wt/vol: 265.1 (mL) Date Analyzed: 12/11/2016 19:56
 Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 141574 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	108		70-130
STL00996	13C2 PFDA	108		70-130

9

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1

SDG No.:

Client Sample ID: WI-AF-3RW12-1116 Lab Sample ID: 320-23971-9

Matrix: Water Lab File ID: 11DEC2016A6A_020.d

Analysis Method: 537 Date Collected: 11/30/2016 17:00

Extraction Method: 537 Date Extracted: 12/05/2016 11:42

Sample wt/vol: 256.4 (mL) Date Analyzed: 12/11/2016 20:26

Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1

Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)

% Moisture: GPC Cleanup: (Y/N) N

Analysis Batch No.: 141574 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.059	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0092
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	% REC	Q	LIMITS
STL00993	13C2 PFHxA	119		70-130
STL00996	13C2 PFDA	130		70-130

10

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-23971-1

SDG No.:

Client Sample ID: WI-AF-3FB12-1116 Lab Sample ID: 320-23971-10

Matrix: Water Lab File ID: 11DEC2016A6A_021.d

Analysis Method: 537 Date Collected: 11/30/2016 17:01

Extraction Method: 537 Date Extracted: 12/05/2016 11:42

Sample wt/vol: 265.2 (mL) Date Analyzed: 12/11/2016 20:55

Con. Extract Vol.: 1.0 (mL) Dilution Factor: 1


Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)

% Moisture: GPC Cleanup: (Y/N) N

Analysis Batch No.: 141574 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.057	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	99		70-130
STL00996	13C2 PFDA	105		70-130



Legend

- 1 Mile Zone
- Surface Water
- Drainage Ditch
- Suspected Source Area
- Phase 1 Sampling Area
- Phase 2 Sampling Area
- Phase 3 Sampling Area
- Base Boundary
- Inferred Groundwater Flow Direction

Notes:

1. Results shown on this figure are for PFOA and PFOS. See text and Table 2 for PFBS results.
2. * Both results above the LHA are from the same well; the second sample collected was a confirmation sample.

N

0 0.25 0.5 Miles

1 inch = 0.5 mile

Imagery Source: Esri

Figure 2
Results for Drinking Water Well Sampling
Near Ault Field
Naval Air Station Whidbey Island
Oak Harbor, Washington

For Official Use Only

ch2m