



**Off-base Drinking Water Sample Results,
Level 2 Laboratory Report, Level 4 Laboratory Report,
Electronic Data Deliverable, Data Validation Report,
and the Sample Location Figure, SDG 320-24648**

Ault Field

Naval Air Station Whidbey Island

Oak Harbor, Washington

February 2019

TestAmerica

THE LEADER IN ENVIRONMENTAL TESTING

ANALYTICAL REPORT

TestAmerica Laboratories, Inc.

TestAmerica Sacramento

880 Riverside Parkway

West Sacramento, CA 95605

Tel: (916)373-5600

TestAmerica Job ID: 320-24648-1

Client Project/Site: Whidbey Island


For:

CH2M Hill Constructors, Inc.

1100 NE Circle Blvd

Corvallis, Oregon 97330

Attn: Tiffany Hill



Authorized for release by:

1/3/2017 4:12:28 PM

Laura Turpen, Project Manager I

(916)374-4414

laura.turpen@testamericainc.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.testamericainc.com

The test results in this report meet all 2003 NELAC and 2009 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15

Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Detection Summary	6
Client Sample Results	7
Surrogate Summary	10
QC Sample Results	11
QC Association Summary	12
Lab Chronicle	13
Certification Summary	15
Method Summary	16
Sample Summary	17
Chain of Custody	18
Receipt Checklists	19

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Definitions/Glossary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Undetected at the Limit of Detection.
M	Manual integrated compound.
J	Estimated: The analyte was positively identified; the quantitation is an estimation
E	Result exceeded calibration range.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Job ID: 320-24648-1

Laboratory: TestAmerica Sacramento

Narrative

CASE NARRATIVE

Client: CH2M Hill Constructors, Inc.

Project: Whidbey Island

Report Number: 320-24648-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

TestAmerica Sacramento attests to the validity of the laboratory data generated by TestAmerica facilities reported herein. All analyses performed by TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the applicable methods. TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

TestAmerica utilizes USEPA approved methods and DOD QSM, where applicable, in all analytical work. The samples presented in this report were analyzed for the parameter(s) listed on the analytical methods summary page in accordance with the method(s) indicated. A summary of QC data for these analyses is included at the back of the report.

All parameters for which TestAmerica Sacramento has certification were evaluated to the QSM specified reporting convention or to the client specified format if different from QSM. Parameters not certified under QSM, if any, were evaluated to the detection limit (DL) and include qualified results where applicable.

The sample(s) that contain constituents flagged with U are undetected. The result associated with this flag is the limit of detection (LOD).

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of TestAmerica and its client.

RECEIPT

The samples were received on 12/23/2016; the samples arrived in good condition, properly preserved and on ice. The temperature of the coolers at receipt was 2.0 C.

An extended TAT was requested by the client via an email on December 21. Sample results were requested to not be sent on December 29 or December 30.

PFOA/PFOS

Samples WI-AF-3RW38-1216 (320-24648-1), WI-AF-3FB38-1216 (320-24648-2), WI-AF-3RW39-1216 (320-24648-3), WI-AF-3FB39-1216 (320-24648-4), WI-AF-3RW40-1216 (320-24648-5), WI-AF-3RW40P-1216 (320-24648-6), WI-AF-3FB40-1216 (320-24648-7), WI-AF-3RW41-1216 (320-24648-8) and WI-AF-3FB41-1216 (320-24648-9) were analyzed for PFOA/PFOS in accordance with 537. The

Case Narrative

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Job ID: 320-24648-1 (Continued)

Laboratory: TestAmerica Sacramento (Continued)

samples were prepared on 12/23/2016 and analyzed on 01/02/2017.

Sample extract for WI-AF-3RW41-1216 (320-24648-8) was brown in color at final volume.

Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with 320-143783.

The laboratory control sample (LCS) has an E flag because it is spiked at the upper level of the calibration curve as specified in the method.

(LCS 320-143783/2-A)

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

Detection Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Lab Sample ID: 320-24648-1

☐ No Detections.

Client Sample ID: WI-AF-3FB38-1216

Lab Sample ID: 320-24648-2

☐ No Detections.

Client Sample ID: WI-AF-3RW39-1216

Lab Sample ID: 320-24648-3

☐ No Detections.

Client Sample ID: WI-AF-3FB39-1216

Lab Sample ID: 320-24648-4

☐ No Detections.

Client Sample ID: WI-AF-3RW40-1216

Lab Sample ID: 320-24648-5

☐ No Detections.

Client Sample ID: WI-AF-3RW40P-1216

Lab Sample ID: 320-24648-6

☐ No Detections.

Client Sample ID: WI-AF-3FB40-1216

Lab Sample ID: 320-24648-7

☐ No Detections.

Client Sample ID: WI-AF-3RW41-1216

Lab Sample ID: 320-24648-8

Analyte	Result	Qualifier	LOQ	DL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.045	ug/L	1		537	Total/NA

Client Sample ID: WI-AF-3FB41-1216

Lab Sample ID: 320-24648-9

☐ No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Date Collected: 12/21/16 09:18

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-1

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.015	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.0089	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	101		70 - 130				12/23/16 18:41	01/02/17 13:53	1
13C2 PFDA	119		70 - 130				12/23/16 18:41	01/02/17 13:53	1

Client Sample ID: WI-AF-3FB38-1216

Date Collected: 12/21/16 09:19

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-2

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.058	0.015	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.0091	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	100		70 - 130				12/23/16 18:41	01/02/17 14:23	1
13C2 PFDA	106		70 - 130				12/23/16 18:41	01/02/17 14:23	1

Client Sample ID: WI-AF-3RW39-1216

Date Collected: 12/21/16 09:25

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-3

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.015	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.0090	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	103		70 - 130				12/23/16 18:41	01/02/17 14:52	1
13C2 PFDA	101		70 - 130				12/23/16 18:41	01/02/17 14:52	1

Client Sample ID: WI-AF-3FB39-1216

Date Collected: 12/21/16 09:26

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-4

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.049	U	0.061	0.016	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Perfluorooctanoic acid (PFOA)	0.024	U	0.031	0.0096	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	101		70 - 130				12/23/16 18:41	01/02/17 15:22	1
13C2 PFDA	107		70 - 130				12/23/16 18:41	01/02/17 15:22	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW40-1216

Date Collected: 12/21/16 16:47

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-5

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L	—	12/23/16 18:41	01/02/17 15:52	1
Perfluorooctanoic acid (PFOA)	0.022	U M	0.028	0.0088	ug/L	—	12/23/16 18:41	01/02/17 15:52	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L	—	12/23/16 18:41	01/02/17 15:52	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/23/16 18:41	01/02/17 15:52	1
13C2 PFDA	123		70 - 130				12/23/16 18:41	01/02/17 15:52	1

Client Sample ID: WI-AF-3RW40P-1216

Date Collected: 12/21/16 16:52

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-6

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L	—	12/23/16 18:41	01/02/17 16:21	1
Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.0088	ug/L	—	12/23/16 18:41	01/02/17 16:21	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L	—	12/23/16 18:41	01/02/17 16:21	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	96		70 - 130				12/23/16 18:41	01/02/17 16:21	1
13C2 PFDA	107		70 - 130				12/23/16 18:41	01/02/17 16:21	1

Client Sample ID: WI-AF-3FB40-1216

Date Collected: 12/21/16 16:53

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-7

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.058	0.015	ug/L	—	12/23/16 18:41	01/02/17 16:51	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.0092	ug/L	—	12/23/16 18:41	01/02/17 16:51	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L	—	12/23/16 18:41	01/02/17 16:51	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	99		70 - 130				12/23/16 18:41	01/02/17 16:51	1
13C2 PFDA	108		70 - 130				12/23/16 18:41	01/02/17 16:51	1

Client Sample ID: WI-AF-3RW41-1216

Date Collected: 12/21/16 18:40

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-8

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.015	ug/L	—	12/23/16 18:41	01/02/17 18:20	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.0089	ug/L	—	12/23/16 18:41	01/02/17 18:20	1
Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.045	ug/L	—	12/23/16 18:41	01/02/17 18:20	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	91		70 - 130				12/23/16 18:41	01/02/17 18:20	1
13C2 PFDA	105		70 - 130				12/23/16 18:41	01/02/17 18:20	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3FB41-1216

Lab Sample ID: 320-24648-9

Date Collected: 12/21/16 18:41

Matrix: Water

Date Received: 12/23/16 10:50

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.015	ug/L		12/23/16 18:41	01/02/17 18:49	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.0090	ug/L		12/23/16 18:41	01/02/17 18:49	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/23/16 18:41	01/02/17 18:49	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFHxA	102		70 - 130	12/23/16 18:41	01/02/17 18:49	1
13C2 PFDA	108		70 - 130	12/23/16 18:41	01/02/17 18:49	1

Surrogate Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)	
Lab Sample ID	Client Sample ID	3C2 PFHx (70-130)	3C2 PFDA (70-130)
320-24648-1	WI-AF-3RW38-1216	101	119
320-24648-2	WI-AF-3FB38-1216	100	106
320-24648-3	WI-AF-3RW39-1216	103	101
320-24648-4	WI-AF-3FB39-1216	101	107
320-24648-5	WI-AF-3RW40-1216	110	123
320-24648-6	WI-AF-3RW40P-1216	96	107
320-24648-7	WI-AF-3FB40-1216	99	108
320-24648-8	WI-AF-3RW41-1216	91	105
320-24648-9	WI-AF-3FB41-1216	102	108
LCS 320-143783/2-A	Lab Control Sample	121	121
LCSD 320-143783/3-A	Lab Control Sample Dup	111	116
MB 320-143783/1-A	Method Blank	111	116

Surrogate Legend

13C2 PFHxA = 13C2 PFHxA

13C2 PFDA = 13C2 PFDA

QC Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 320-143783/1-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 143783

Analyte	MB Result	MB Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.016	ug/L		12/23/16 18:41	01/02/17 12:24	1
Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.0094	ug/L		12/23/16 18:41	01/02/17 12:24	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L		12/23/16 18:41	01/02/17 12:24	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFHxA	111		70 - 130	12/23/16 18:41	01/02/17 12:24	1
13C2 PFDA	116		70 - 130	12/23/16 18:41	01/02/17 12:24	1

Lab Sample ID: LCS 320-143783/2-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 143783

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Perfluorooctanesulfonic acid (PFOS)	0.300	0.318		ug/L		106	70 - 130
Perfluorooctanoic acid (PFOA)	0.146	0.159	E	ug/L		109	70 - 130
Perfluorobutanesulfonic acid (PFBS)	0.673	0.659		ug/L		98	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
13C2 PFHxA	121		70 - 130
13C2 PFDA	121		70 - 130

Lab Sample ID: LCSD 320-143783/3-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 143783

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	Limits	RPD	RPD Limit
Perfluorooctanesulfonic acid (PFOS)	0.300	0.307		ug/L		102	70 - 130	4	30
Perfluorooctanoic acid (PFOA)	0.146	0.155		ug/L		106	70 - 130	3	30
Perfluorobutanesulfonic acid (PFBS)	0.673	0.622		ug/L		92	70 - 130	6	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
13C2 PFHxA	111		70 - 130
13C2 PFDA	116		70 - 130

TestAmerica Sacramento

QC Association Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

LCMS

Prep Batch: 143783

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-1	WI-AF-3RW38-1216	Total/NA	Water	537	
320-24648-2	WI-AF-3FB38-1216	Total/NA	Water	537	
320-24648-3	WI-AF-3RW39-1216	Total/NA	Water	537	
320-24648-4	WI-AF-3FB39-1216	Total/NA	Water	537	
320-24648-5	WI-AF-3RW40-1216	Total/NA	Water	537	
320-24648-6	WI-AF-3RW40P-1216	Total/NA	Water	537	
320-24648-7	WI-AF-3FB40-1216	Total/NA	Water	537	
320-24648-8	WI-AF-3RW41-1216	Total/NA	Water	537	
320-24648-9	WI-AF-3FB41-1216	Total/NA	Water	537	
MB 320-143783/1-A	Method Blank	Total/NA	Water	537	
LCS 320-143783/2-A	Lab Control Sample	Total/NA	Water	537	
LCSD 320-143783/3-A	Lab Control Sample Dup	Total/NA	Water	537	

Analysis Batch: 144608

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-1	WI-AF-3RW38-1216	Total/NA	Water	537	143783
320-24648-2	WI-AF-3FB38-1216	Total/NA	Water	537	143783
320-24648-3	WI-AF-3RW39-1216	Total/NA	Water	537	143783
320-24648-4	WI-AF-3FB39-1216	Total/NA	Water	537	143783
320-24648-5	WI-AF-3RW40-1216	Total/NA	Water	537	143783
320-24648-6	WI-AF-3RW40P-1216	Total/NA	Water	537	143783
320-24648-7	WI-AF-3FB40-1216	Total/NA	Water	537	143783
MB 320-143783/1-A	Method Blank	Total/NA	Water	537	143783
LCS 320-143783/2-A	Lab Control Sample	Total/NA	Water	537	143783
LCSD 320-143783/3-A	Lab Control Sample Dup	Total/NA	Water	537	143783

Analysis Batch: 144610

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-8	WI-AF-3RW41-1216	Total/NA	Water	537	143783
320-24648-9	WI-AF-3FB41-1216	Total/NA	Water	537	143783

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Date Collected: 12/21/16 09:18

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			263.2 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 13:53	JRB	TAL SAC

Client Sample ID: WI-AF-3FB38-1216

Date Collected: 12/21/16 09:19

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			258.6 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 14:23	JRB	TAL SAC

Client Sample ID: WI-AF-3RW39-1216

Date Collected: 12/21/16 09:25

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			263 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 14:52	JRB	TAL SAC

Client Sample ID: WI-AF-3FB39-1216

Date Collected: 12/21/16 09:26

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			245.6 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 15:22	JRB	TAL SAC

Client Sample ID: WI-AF-3RW40-1216

Date Collected: 12/21/16 16:47

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			269 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 15:52	JRB	TAL SAC

Client Sample ID: WI-AF-3RW40P-1216

Date Collected: 12/21/16 16:52

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			267.5 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 16:21	JRB	TAL SAC

TestAmerica Sacramento

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3FB40-1216

Date Collected: 12/21/16 16:53

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			256.9 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144608	01/02/17 16:51	JRB	TAL SAC

Client Sample ID: WI-AF-3RW41-1216

Date Collected: 12/21/16 18:40

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			265.6 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144610	01/02/17 18:20	JRB	TAL SAC

Client Sample ID: WI-AF-3FB41-1216

Date Collected: 12/21/16 18:41

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-9

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			262.9 mL	1.00 mL	143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1			144610	01/02/17 18:49	JRB	TAL SAC

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Certification Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Laboratory: TestAmerica Sacramento

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
A2LA	DoD ELAP		2928-01	01-31-17
Analysis Method	Prep Method	Matrix	Analyte	

Method Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method	Method Description	Protocol	Laboratory
537	Perfluorinated Alkyl Acids (LC/MS)	EPA	TAL SAC

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Sample Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
320-24648-1	WI-AF-3RW38-1216	Water	12/21/16 09:18	12/23/16 10:50
320-24648-2	WI-AF-3FB38-1216	Water	12/21/16 09:19	12/23/16 10:50
320-24648-3	WI-AF-3RW39-1216	Water	12/21/16 09:25	12/23/16 10:50
320-24648-4	WI-AF-3FB39-1216	Water	12/21/16 09:26	12/23/16 10:50
320-24648-5	WI-AF-3RW40-1216	Water	12/21/16 16:47	12/23/16 10:50
320-24648-6	WI-AF-3RW40P-1216	Water	12/21/16 16:52	12/23/16 10:50
320-24648-7	WI-AF-3FB40-1216	Water	12/21/16 16:53	12/23/16 10:50
320-24648-8	WI-AF-3RW41-1216	Water	12/21/16 18:40	12/23/16 10:50
320-24648-9	WI-AF-3FB41-1216	Water	12/21/16 18:41	12/23/16 10:50

West Sacramento, CA 95605
phone 916.373.5600 fax

TestAmerica Laboratories, Inc.

Regulatory Program: ☐ DW ☐ NPDES ☐ RCRA ☐ Other:

COC No: 11 of 1 COCs

Date: 12/22/2016

Carrier: FedEx

Project Manager: Katie Tipping

Tell/Fax: (757) 671-6258

Client Contact

Project Chemist
1100 NE Circle Blvd Ste 300 Corvallis, OR 97330

(541) 768-3109
(541) 908-3794

Project Name: CTO-08

Site: NAS Whidbey Island

P O #: 100067106050 - 679580.06.FIFS

Analysis Turnaround Time
☒ CALENDAR DAYS ☐ WORKING DAYS
TAT if different from Below _____ 7-Day _____
☐ 2 weeks
☐ 1 week
☐ 2 days
☐ 1 day

Sample Identification

Sample Date

Sample Time

Sample Type (C=Comp, G=Grab)

Matrix

of Cont.

Filtered Sample (Y/N)

Perform MS / MSD (Y/N)

USEPA Method 537 (PFOA, PFOS, and PFBS)

Sample Specific Notes:

WI-AF-3RW38-1216

WI-AF-3FB38-1216

WI-AF-3RW39-1216

WI-AF-3FB39-1216

WI-AF-3RW40-1216

WI-AF-3RW40P-1216

WI-AF-3FB40-1216

WI-AF-3RW41-1216

WI-AF-3FB41-1216

320-24648 Chain of Custody

Preservation Used: 1= Ice, 2= HCl, 3= H2SO4, 4=HNO3; 5=NaOH; 6= Other Trizma

Possible Hazard Identification:

Are any samples from a listed EPA Hazardous Waste? Please List any EPA Waste Codes for the sample in the

Comments Section if the lab is to dispose of the sample.

☒ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☐ Unknown

Special Instructions/QC Requirements & Comments:

Custody Seals Intact: ☐ Yes ☐ No

Relinquished by: Eric Eppe

Relinquished by:

Relinquished by:

Company: CH2M

Company:

Company:

Company:

Company:

Company:

Company:

Company:

Company:

2.0°C

Form No. CA-C-WI-002, Rev. 4.3, dated 12/05/2013

Login Sample Receipt Checklist

Client: CH2M Hill Constructors, Inc.

Job Number: 320-24648-1

Login Number: 24648

List Source: TestAmerica Sacramento

List Number: 1

Creator: Edman, Connor M

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	N/A	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

ANALYTICAL REPORT

Job Number: 320-24648-1

Job Description: Whidbey Island

For:

CH2M Hill Constructors, Inc.

1100 NE Circle Blvd

Corvallis, OR 97330

Attention: Tiffany Hill



Approved for release.
Laura Turpen
Project Manager I
1/3/2017 4:14 PM

Laura Turpen, Project Manager I
880 Riverside Parkway, West Sacramento, CA, 95605
(916)374-4414
laura.turpen@testamericainc.com
01/03/2017

Table of Contents

Cover Title Page	1
Data Summaries	4
Definitions	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Default Detection Limits	10
Surrogate Summary	11
QC Sample Results	12
QC Association	13
Chronicle	14
Certification Summary	16
Method Summary	17
Sample Summary	18
Manual Integration Summary	19
Reagent Traceability	24
COAs	33
Organic Sample Data	78
LCMS	78
Method 537 DOD	78
Method 537 DOD QC Summary	79
Method 537 DOD Sample Data	88
Standards Data	133
Method 537 DOD ICAL Data	133
Method 537 DOD CCAL Data	164
Raw QC Data	192

Table of Contents

Method 537 DOD Blank Data	192
Method 537 DOD LCS/LCSD Data	198
Method 537 DOD Run Logs	206
Method 537 DOD Prep Data	209
Shipping and Receiving Documents	220
Client Chain of Custody	221
Sample Receipt Checklist	222

Definitions/Glossary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Qualifiers

LCMS

Qualifier	Qualifier Description
U	Undetected at the Limit of Detection.
M	Manual integrated compound.
J	Estimated: The analyte was positively identified; the quantitation is an estimation
E	Result exceeded calibration range.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains no Free Liquid
DER	Duplicate error ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision level concentration
MDA	Minimum detectable activity
EDL	Estimated Detection Limit
MDC	Minimum detectable concentration
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative error ratio
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

CASE NARRATIVE

Client: CH2M Hill Constructors, Inc.

Project: Whidbey Island

Report Number: 320-24648-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

TestAmerica Sacramento attests to the validity of the laboratory data generated by TestAmerica facilities reported herein. All analyses performed by TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the applicable methods. TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

TestAmerica utilizes USEPA approved methods and DOD QSM, where applicable, in all analytical work. The samples presented in this report were analyzed for the parameter(s) listed on the analytical methods summary page in accordance with the method(s) indicated. A summary of QC data for these analyses is included at the back of the report.

All parameters for which TestAmerica Sacramento has certification were evaluated to the QSM specified reporting convention or to the client specified format if different from QSM. Parameters not certified under QSM, if any, were evaluated to the detection limit (DL) and include qualified results where applicable.

The sample(s) that contain constituents flagged with U are undetected. The result associated with this flag is the limit of detection (LOD).

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of TestAmerica and its client.

RECEIPT

The samples were received on 12/23/2016; the samples arrived in good condition, properly preserved and on ice. The temperature of the coolers at receipt was 2.0 C.

An extended TAT was requested by the client via an email on December 21. Sample results were requested to not be sent on December 29 or December 30.

PFOA/PFOS

Samples WI-AF-3RW38-1216 (320-24648-1), WI-AF-3FB38-1216 (320-24648-2), WI-AF-3RW39-1216 (320-24648-3), WI-AF-3FB39-1216 (320-24648-4), WI-AF-3RW40-1216 (320-24648-5), WI-AF-3RW40P-1216 (320-24648-6), WI-AF-3FB40-1216 (320-24648-7), WI-AF-3RW41-1216 (320-24648-8) and WI-AF-3FB41-1216 (320-24648-9) were analyzed for PFOA/PFOS in accordance with 537. The samples were prepared on 12/23/2016 and analyzed on 01/02/2017.

Sample extract for WI-AF-3RW41-1216 (320-24648-8) was brown in color at final volume.

Insufficient sample volume was available to perform a matrix spike/matrix spike duplicate (MS/MSD) associated with 320-143783.

The laboratory control sample (LCS) has an E flag because it is spiked at the upper level of the calibration curve as specified in the method.

(LCS 320-143783/2-A)

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

Detection Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Lab Sample ID: 320-24648-1

☐ No Detections.

Client Sample ID: WI-AF-3FB38-1216

Lab Sample ID: 320-24648-2

☐ No Detections.

Client Sample ID: WI-AF-3RW39-1216

Lab Sample ID: 320-24648-3

☐ No Detections.

Client Sample ID: WI-AF-3FB39-1216

Lab Sample ID: 320-24648-4

☐ No Detections.

Client Sample ID: WI-AF-3RW40-1216

Lab Sample ID: 320-24648-5

☐ No Detections.

Client Sample ID: WI-AF-3RW40P-1216

Lab Sample ID: 320-24648-6

☐ No Detections.

Client Sample ID: WI-AF-3FB40-1216

Lab Sample ID: 320-24648-7

☐ No Detections.

Client Sample ID: WI-AF-3RW41-1216

Lab Sample ID: 320-24648-8

Analyte	Result	Qualifier	LOQ	DL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.045	ug/L	1		537	Total/NA

Client Sample ID: WI-AF-3FB41-1216

Lab Sample ID: 320-24648-9

☐ No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Date Collected: 12/21/16 09:18

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-1

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.015	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.0089	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L	—	12/23/16 18:41	01/02/17 13:53	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	101		70 - 130				12/23/16 18:41	01/02/17 13:53	1
13C2 PFDA	119		70 - 130				12/23/16 18:41	01/02/17 13:53	1

Client Sample ID: WI-AF-3FB38-1216

Date Collected: 12/21/16 09:19

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-2

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.058	0.015	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.0091	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L	—	12/23/16 18:41	01/02/17 14:23	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	100		70 - 130				12/23/16 18:41	01/02/17 14:23	1
13C2 PFDA	106		70 - 130				12/23/16 18:41	01/02/17 14:23	1

Client Sample ID: WI-AF-3RW39-1216

Date Collected: 12/21/16 09:25

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-3

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.015	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.0090	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L	—	12/23/16 18:41	01/02/17 14:52	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	103		70 - 130				12/23/16 18:41	01/02/17 14:52	1
13C2 PFDA	101		70 - 130				12/23/16 18:41	01/02/17 14:52	1

Client Sample ID: WI-AF-3FB39-1216

Date Collected: 12/21/16 09:26

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-4

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.049	U	0.061	0.016	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Perfluorooctanoic acid (PFOA)	0.024	U	0.031	0.0096	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L	—	12/23/16 18:41	01/02/17 15:22	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	101		70 - 130				12/23/16 18:41	01/02/17 15:22	1
13C2 PFDA	107		70 - 130				12/23/16 18:41	01/02/17 15:22	1

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW40-1216

Date Collected: 12/21/16 16:47

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-5

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L		12/23/16 18:41	01/02/17 15:52	1
Perfluorooctanoic acid (PFOA)	0.022	U M	0.028	0.0088	ug/L		12/23/16 18:41	01/02/17 15:52	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/23/16 18:41	01/02/17 15:52	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	110		70 - 130				12/23/16 18:41	01/02/17 15:52	1
13C2 PFDA	123		70 - 130				12/23/16 18:41	01/02/17 15:52	1

Client Sample ID: WI-AF-3RW40P-1216

Date Collected: 12/21/16 16:52

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-6

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.014	ug/L		12/23/16 18:41	01/02/17 16:21	1
Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.0088	ug/L		12/23/16 18:41	01/02/17 16:21	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.044	ug/L		12/23/16 18:41	01/02/17 16:21	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	96		70 - 130				12/23/16 18:41	01/02/17 16:21	1
13C2 PFDA	107		70 - 130				12/23/16 18:41	01/02/17 16:21	1

Client Sample ID: WI-AF-3FB40-1216

Date Collected: 12/21/16 16:53

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-7

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.058	0.015	ug/L		12/23/16 18:41	01/02/17 16:51	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.0092	ug/L		12/23/16 18:41	01/02/17 16:51	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.046	ug/L		12/23/16 18:41	01/02/17 16:51	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	99		70 - 130				12/23/16 18:41	01/02/17 16:51	1
13C2 PFDA	108		70 - 130				12/23/16 18:41	01/02/17 16:51	1

Client Sample ID: WI-AF-3RW41-1216

Date Collected: 12/21/16 18:40

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-8

Matrix: Water

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.015	ug/L		12/23/16 18:41	01/02/17 18:20	1
Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.0089	ug/L		12/23/16 18:41	01/02/17 18:20	1
Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.045	ug/L		12/23/16 18:41	01/02/17 18:20	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	91		70 - 130				12/23/16 18:41	01/02/17 18:20	1
13C2 PFDA	105		70 - 130				12/23/16 18:41	01/02/17 18:20	1

TestAmerica Sacramento

Client Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3FB41-1216

Lab Sample ID: 320-24648-9

Date Collected: 12/21/16 18:41

Matrix: Water

Date Received: 12/23/16 10:50

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.015	ug/L		12/23/16 18:41	01/02/17 18:49	1
Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.0090	ug/L		12/23/16 18:41	01/02/17 18:49	1
Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.045	ug/L		12/23/16 18:41	01/02/17 18:49	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	102		70 - 130				12/23/16 18:41	01/02/17 18:49	1
13C2 PFDA	108		70 - 130				12/23/16 18:41	01/02/17 18:49	1

Default Detection Limits

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Prep: 537

Analyte	LOQ	DL	Units	Method
Perfluorobutanesulfonic acid (PFBS)	0.14	0.048	ug/L	537
Perfluorooctanesulfonic acid (PFOS)	0.060	0.016	ug/L	537
Perfluorooctanoic acid (PFOA)	0.030	0.0094	ug/L	537

Surrogate Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)	
Lab Sample ID	Client Sample ID	3C2 PFHx	3C2 PFD/
		(70-130)	(70-130)
320-24648-1	WI-AF-3RW38-1216	101	119
320-24648-2	WI-AF-3FB38-1216	100	106
320-24648-3	WI-AF-3RW39-1216	103	101
320-24648-4	WI-AF-3FB39-1216	101	107
320-24648-5	WI-AF-3RW40-1216	110	123
320-24648-6	WI-AF-3RW40P-1216	96	107
320-24648-7	WI-AF-3FB40-1216	99	108
320-24648-8	WI-AF-3RW41-1216	91	105
320-24648-9	WI-AF-3FB41-1216	102	108
LCS 320-143783/2-A	Lab Control Sample	121	121
LCSD 320-143783/3-A	Lab Control Sample Dup	111	116
MB 320-143783/1-A	Method Blank	111	116

Surrogate Legend

13C2 PFHxA = 13C2 PFHxA

13C2 PFDA = 13C2 PFDA

QC Sample Results

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method: 537 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 320-143783/1-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 143783

Analyte	MB Result	MB Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.016	ug/L		12/23/16 18:41	01/02/17 12:24	1
Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.0094	ug/L		12/23/16 18:41	01/02/17 12:24	1
Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.048	ug/L		12/23/16 18:41	01/02/17 12:24	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFHxA	111		70 - 130	12/23/16 18:41	01/02/17 12:24	1
13C2 PFDA	116		70 - 130	12/23/16 18:41	01/02/17 12:24	1

Lab Sample ID: LCS 320-143783/2-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 143783

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Perfluorooctanesulfonic acid (PFOS)	0.300	0.318		ug/L		106	70 - 130
Perfluorooctanoic acid (PFOA)	0.146	0.159	E	ug/L		109	70 - 130
Perfluorobutanesulfonic acid (PFBS)	0.673	0.659		ug/L		98	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
13C2 PFHxA	121		70 - 130
13C2 PFDA	121		70 - 130

Lab Sample ID: LCSD 320-143783/3-A

Matrix: Water

Analysis Batch: 144608

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 143783

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Perfluorooctanesulfonic acid (PFOS)	0.300	0.307		ug/L		102	70 - 130	4	30
Perfluorooctanoic acid (PFOA)	0.146	0.155		ug/L		106	70 - 130	3	30
Perfluorobutanesulfonic acid (PFBS)	0.673	0.622		ug/L		92	70 - 130	6	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
13C2 PFHxA	111		70 - 130
13C2 PFDA	116		70 - 130

QC Association Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

LCMS

Prep Batch: 143783

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-1	WI-AF-3RW38-1216	Total/NA	Water	537	
320-24648-2	WI-AF-3FB38-1216	Total/NA	Water	537	
320-24648-3	WI-AF-3RW39-1216	Total/NA	Water	537	
320-24648-4	WI-AF-3FB39-1216	Total/NA	Water	537	
320-24648-5	WI-AF-3RW40-1216	Total/NA	Water	537	
320-24648-6	WI-AF-3RW40P-1216	Total/NA	Water	537	
320-24648-7	WI-AF-3FB40-1216	Total/NA	Water	537	
320-24648-8	WI-AF-3RW41-1216	Total/NA	Water	537	
320-24648-9	WI-AF-3FB41-1216	Total/NA	Water	537	
MB 320-143783/1-A	Method Blank	Total/NA	Water	537	
LCS 320-143783/2-A	Lab Control Sample	Total/NA	Water	537	
LCSD 320-143783/3-A	Lab Control Sample Dup	Total/NA	Water	537	

Analysis Batch: 144608

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-1	WI-AF-3RW38-1216	Total/NA	Water	537	143783
320-24648-2	WI-AF-3FB38-1216	Total/NA	Water	537	143783
320-24648-3	WI-AF-3RW39-1216	Total/NA	Water	537	143783
320-24648-4	WI-AF-3FB39-1216	Total/NA	Water	537	143783
320-24648-5	WI-AF-3RW40-1216	Total/NA	Water	537	143783
320-24648-6	WI-AF-3RW40P-1216	Total/NA	Water	537	143783
320-24648-7	WI-AF-3FB40-1216	Total/NA	Water	537	143783
MB 320-143783/1-A	Method Blank	Total/NA	Water	537	143783
LCS 320-143783/2-A	Lab Control Sample	Total/NA	Water	537	143783
LCSD 320-143783/3-A	Lab Control Sample Dup	Total/NA	Water	537	143783

Analysis Batch: 144610

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
320-24648-8	WI-AF-3RW41-1216	Total/NA	Water	537	143783
320-24648-9	WI-AF-3FB41-1216	Total/NA	Water	537	143783

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3RW38-1216

Date Collected: 12/21/16 09:18

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 13:53	JRB	TAL SAC

Client Sample ID: WI-AF-3FB38-1216

Date Collected: 12/21/16 09:19

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 14:23	JRB	TAL SAC

Client Sample ID: WI-AF-3RW39-1216

Date Collected: 12/21/16 09:25

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 14:52	JRB	TAL SAC

Client Sample ID: WI-AF-3FB39-1216

Date Collected: 12/21/16 09:26

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 15:22	JRB	TAL SAC

Client Sample ID: WI-AF-3RW40-1216

Date Collected: 12/21/16 16:47

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 15:52	JRB	TAL SAC

Client Sample ID: WI-AF-3RW40P-1216

Date Collected: 12/21/16 16:52

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 16:21	JRB	TAL SAC

TestAmerica Sacramento

Lab Chronicle

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Client Sample ID: WI-AF-3FB40-1216

Date Collected: 12/21/16 16:53

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144608	01/02/17 16:51	JRB	TAL SAC

Client Sample ID: WI-AF-3RW41-1216

Date Collected: 12/21/16 18:40

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144610	01/02/17 18:20	JRB	TAL SAC

Client Sample ID: WI-AF-3FB41-1216

Date Collected: 12/21/16 18:41

Date Received: 12/23/16 10:50

Lab Sample ID: 320-24648-9

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	537			143783	12/23/16 18:41	JER	TAL SAC
Total/NA	Analysis	537		1	144610	01/02/17 18:49	JRB	TAL SAC

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Certification Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Laboratory: TestAmerica Sacramento

Unless otherwise noted, all analytes for this laboratory were covered under each certification below.

Authority	Program	EPA Region	Certification ID	Expiration Date
A2LA	DoD ELAP		2928-01	01-31-17
Analysis Method	Prep Method	Matrix	Analyte	

Method Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Method	Method Description	Protocol	Laboratory
537	Perfluorinated Alkyl Acids (LC/MS)	EPA	TAL SAC

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

TAL SAC = TestAmerica Sacramento, 880 Riverside Parkway, West Sacramento, CA 95605, TEL (916)373-5600

Sample Summary

Client: CH2M Hill Constructors, Inc.
Project/Site: Whidbey Island

TestAmerica Job ID: 320-24648-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
320-24648-1	WI-AF-3RW38-1216	Water	12/21/16 09:18	12/23/16 10:50
320-24648-2	WI-AF-3FB38-1216	Water	12/21/16 09:19	12/23/16 10:50
320-24648-3	WI-AF-3RW39-1216	Water	12/21/16 09:25	12/23/16 10:50
320-24648-4	WI-AF-3FB39-1216	Water	12/21/16 09:26	12/23/16 10:50
320-24648-5	WI-AF-3RW40-1216	Water	12/21/16 16:47	12/23/16 10:50
320-24648-6	WI-AF-3RW40P-1216	Water	12/21/16 16:52	12/23/16 10:50
320-24648-7	WI-AF-3FB40-1216	Water	12/21/16 16:53	12/23/16 10:50
320-24648-8	WI-AF-3RW41-1216	Water	12/21/16 18:40	12/23/16 10:50
320-24648-9	WI-AF-3FB41-1216	Water	12/21/16 18:41	12/23/16 10:50

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Analysis Batch Number: 143828Lab Sample ID: STD 320-143828/4 IC Client Sample ID: _____Date Analyzed: 12/24/16 04:26 Lab File ID: 24DEC2016A6A_004.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.34	Baseline	phomsopha t	12/24/16 11:57
Perfluorooctanoic acid (PFOA)	20.00	Baseline	phomsopha t	12/24/16 11:57

Lab Sample ID: STD 320-143828/5 IC Client Sample ID: _____Date Analyzed: 12/24/16 04:55 Lab File ID: 24DEC2016A6A_005.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.00	Baseline	phomsopha t	12/24/16 11:59

Lab Sample ID: STD 320-143828/6 IC Client Sample ID: _____Date Analyzed: 12/24/16 05:25 Lab File ID: 24DEC2016A6A_006.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.34	Baseline	phomsopha t	12/24/16 12:04
Perfluorooctanoic acid (PFOA)	20.00	Baseline	phomsopha t	12/24/16 12:09

Lab Sample ID: STD 320-143828/7 ICISAV Client Sample ID: _____Date Analyzed: 12/24/16 05:54 Lab File ID: 24DEC2016A6A_007.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.36	Baseline	phomsopha t	12/24/16 12:07

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Analysis Batch Number: 143828Lab Sample ID: STD 320-143828/8 IC Client Sample ID: _____Date Analyzed: 12/24/16 06:24 Lab File ID: 24DEC2016A6A_008.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.36	Baseline	phomsopha t	12/24/16 12:10

Lab Sample ID: STD 320-143828/9 IC Client Sample ID: _____Date Analyzed: 12/24/16 06:54 Lab File ID: 24DEC2016A6A_009.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.36	Baseline	phomsopha t	12/24/16 12:11

Lab Sample ID: CCV 320-143828/11 CCVL Client Sample ID: _____Date Analyzed: 12/24/16 07:53 Lab File ID: 24DEC2016A6A_011.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.36	Baseline	phomsopha t	12/24/16 12:19
Perfluorooctanoic acid (PFOA)	20.01	Baseline	phomsopha t	12/24/16 12:19

Lab Sample ID: ICV 320-143828/13 Client Sample ID: _____Date Analyzed: 12/24/16 08:52 Lab File ID: 24DEC2016A6A_013.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.36	Baseline	phomsopha t	12/24/16 12:21

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Analysis Batch Number: 144608Lab Sample ID: CCV 320-144608/2 CCVL Client Sample ID: _____Date Analyzed: 01/02/17 10:56 Lab File ID: 02JAN2017A6A_002.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluoroheptanoic acid	19.34	Split Peak	barnettj	01/03/17 13:55

Lab Sample ID: MB 320-143783/1-A Client Sample ID: _____Date Analyzed: 01/02/17 12:24 Lab File ID: 02JAN2017A6A_005.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	19.97	Missed Peak	barnettj	01/03/17 14:38
Perfluorooctanesulfonic acid (PFOS)	20.63	Missed Peak	barnettj	01/03/17 14:38

Lab Sample ID: 320-24648-1 Client Sample ID: WI-AF-3RW38-1216Date Analyzed: 01/02/17 13:53 Lab File ID: 02JAN2017A6A_008.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.00	Baseline	barnettj	01/03/17 14:41
Perfluorooctanesulfonic acid (PFOS)	20.62	Missed Peak	barnettj	01/03/17 14:41

Lab Sample ID: 320-24648-2 Client Sample ID: WI-AF-3FB38-1216Date Analyzed: 01/02/17 14:23 Lab File ID: 02JAN2017A6A_009.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.00	Split Peak	barnettj	01/03/17 14:42
Perfluorooctanesulfonic acid (PFOS)	20.62	Missed Peak	barnettj	01/03/17 14:42

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Analysis Batch Number: 144608Lab Sample ID: 320-24648-3 Client Sample ID: WI-AF-3RW39-1216Date Analyzed: 01/02/17 14:52 Lab File ID: 02JAN2017A6A_010.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	20.00	Split Peak	barnettj	01/03/17 14:44

Lab Sample ID: 320-24648-5 Client Sample ID: WI-AF-3RW40-1216Date Analyzed: 01/02/17 15:52 Lab File ID: 02JAN2017A6A_012.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	19.97	Missed Peak	barnettj	01/03/17 14:45

LCMS MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Analysis Batch Number: 144610Lab Sample ID: 320-24648-8 Client Sample ID: WI-AF-3RW41-1216Date Analyzed: 01/02/17 18:20 Lab File ID: 02JAN2017A6A_017.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanoic acid (PFOA)	19.96	Split Peak	barnettj	01/03/17 14:07
Perfluorooctanesulfonic acid (PFOS)	20.34	Missed Peak	barnettj	01/03/17 14:07

Lab Sample ID: 320-24648-9 Client Sample ID: WI-AF-3FB41-1216Date Analyzed: 01/02/17 18:49 Lab File ID: 02JAN2017A6A_018.d GC Column: Acquity ID: 2.1 (mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Perfluorooctanesulfonic acid (PFOS)	20.60	Missed Peak	barnettj	01/03/17 14:09

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	375 uL	Perfluorobutane Sulfonate	3366 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	371.25 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	777.808 ng/mL
							Perfluorooctanoic acid (PFOA)	731.96 ng/mL
.LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutane Sulfonate	89.76 ug/mL
							Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
					LC537-PFHpA_00013	100 uL	Perfluoroheptanoic acid	9.9 ug/mL
					LC537-PFHxS_00008	300 uL	Perfluorohexanesulfonic acid	30.2582 ug/mL
					LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
					LC537-PFOA_00011	100 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
..LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537-PFOS_00006	400 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
							Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00002	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537-PFBS_00002	0.0102 g	Perfluorobutane Sulfonate	2040 ug/mL
							Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
...LC537-PFBS_00002	04/01/18		Sigma, Lot MKBP8842V		(Purchased Reagent)		Perfluorobutane Sulfonate	1 g/g
							Perfluorobutanesulfonic acid (PFBS)	1 g/g
..LC537-PFHpA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537-PFHpA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL
...LC537-PFHpA_00002	04/01/18		Aldrich, Lot BCBM2579V		(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
..LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537-PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
...LC537-PFHxS_00002	04/01/18		Sigma, Lot BCBL3545V		(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
..LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL
...LC537-PFNA_00002	04/01/18		TCI America, Lot QN44F		(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
..LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
...LC537-PFOA_00002	11/04/18		Fluka, Lot SZBD308XV		(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g
..LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537-PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
...LC537-PFOS_00002	08/09/17		Fluka, Lot SZBC222XV		(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
LC537-ICV_00019	03/01/17	12/20/16	MeOH/H2O, Lot 067374	10 mL	LC537-IS_00028	200 uL	13C2-PFOA	10 ng/mL
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C4 PFOS	28.68 ng/mL
					LCMPFOS_00018	300 uL	13C2-PFOA	0.5 ug/mL
..LCM2PFOA_00005	06/19/18		Wellington Laboratories, Lot M2PFOA0613		(Purchased Reagent)		13C4 PFOS	1.434 ug/mL
..LCMPFOS_00018	08/03/21		Wellington Laboratories, Lot MPFOS0816		(Purchased Reagent)		13C2-PFOA	50 ug/mL
LC537-ICV_00019	03/01/17	12/20/16	MeOH/H2O, Lot 067374	10 mL	LC537-SU_00027	500 uL	13C4 PFOS	47.8 ug/mL
							13C2 PFDA	10 ng/mL
					LC537ICIM_00014	25 uL	13C2 PFHxA	10 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	114.77 ng/mL
							Perfluorooctanoic acid (PFOA)	25.0232 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
							Perfluorooctanesulfonic acid (PFOS)	27.2389 ng/mL
.LC537-SU_00027	06/19/17	12/19/16	Methanol, Lot 104453	20000 uL	LCMPFDA 00008	80 uL	13C2 PFDA	0.2 ug/mL
..LCMPFDA 00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			LCMPFHxA 00009	80 uL	13C2 PFHxA	0.2 ug/mL
..LCMPFHxA 00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFDA	50 ug/mL
.LC537ICIM_00014	03/01/17	12/20/16	Methanol, Lot 090285	25 mL	LC537-PFBS2_00005	0.5 mL	13C2 PFHxA	50 ug/mL
					LC537-PFOA2_00008	0.142 mL	Perfluorobutanesulfonic acid (PFBS)	45.908 ug/mL
					LC537-PFOS2_00005	0.22 mL	Perfluorooctanoic acid (PFOA)	10.0093 ug/mL
..LC537-PFBS2_00005	03/01/17	02/29/16	Methanol, Lot 090285	10 mL	LC537_PFBS2_00001	0.023 g	Perfluorooctanesulfonic acid (PFOS)	10.8956 ug/mL
...LC537_PFBS2_00001	08/09/17	Santa Cruz Biotechnology, Lot H0112			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	2295.4 ug/mL
..LC537-PFOA2_00008	07/25/17	12/20/16	Methanol, Lot 090285	10 mL	LC537 PFOA2_00001	0.0178 g	Perfluorobutanesulfonic acid (PFBS)	0.998 g/g
...LC537 PFOA2_00001	07/25/17	Afla Aesar, Lot D24Y026			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	1762.2 ug/mL
..LC537-PFOS2_00005	03/01/17	02/29/16	Methanol, Lot 090285	10 mL	LC537_PFOS2_00001	0.0159 g	Perfluorooctanoic acid (PFOA)	0.99 g/g
...LC537_PFOS2_00001	07/26/17	Sigma, Lot BCBF5116V			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	1238.13 ug/mL
LC537-IS_00027	03/19/17	12/14/16	Methanol, Lot 090285	10000 uL	LCM2PFOA 00003	100 uL	Perfluorooctanesulfonic acid (PFOS)	0.7787 g/g
.LCM2PFOA 00003	03/19/17	Wellington Laboratories, Lot M2PFOA0312			LCMPFOS 00018	300 uL	13C2-PFOA	0.5 ug/mL
.LCMPFOS 00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C4 PFOS	1.434 ug/mL
					(Purchased Reagent)		13C2-PFOA	50 ug/mL
					(Purchased Reagent)		13C4 PFOS	47.8 ug/mL
LC537-L1_00017	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL
					LC537-MSP_00017	25 uL	13C4 PFOS	28.68 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	8.976 ng/mL
							Perfluoroheptanoic acid	0.99 ng/mL
							Perfluorohexanesulfonic acid	3.02582 ng/mL
							Perfluorononanoic acid	2.07415 ng/mL
							Perfluorooctanoic acid (PFOA)	1.95189 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	4.00664 ng/mL
					LC537-SU_00026	250 uL	13C2 PFDA	10 ng/mL
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA 00005	100 uL	13C2 PFHxA	10 ng/mL
..LCM2PFOA 00005	06/19/18	Wellington Laboratories, Lot M2PFOA0613			LCMPFOS 00018	300 uL	13C2-PFOA	0.5 ug/mL
..LCMPFOS 00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C4 PFOS	1.434 ug/mL
.LC537-MSP_00017	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	200 uL	13C2-PFOA	50 ug/mL
							(Purchased Reagent)	47.8 ug/mL
							Perfluorobutanesulfonic acid (PFBS)	1795.2 ng/mL
							Perfluoroheptanoic acid	198 ng/mL
							Perfluorohexanesulfonic acid	605.164 ng/mL
							Perfluorononanoic acid	414.831 ng/mL
							Perfluorooctanoic acid (PFOA)	390.378 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
							Perfluorooctanesulfonic acid (PFOS)	801.328 ng/mL
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
					LC537-PFHpA_00013	100 uL	Perfluoroheptanoic acid	9.9 ug/mL
					LC537-PFHxS_00008	300 uL	Perfluorohexanesulfonic acid	30.2582 ug/mL
					LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
					LC537-PFOA_00011	100 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
					LC537-PFOS_00006	400 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g
...LC537-PFHpA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537 PFHpA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL
....LC537 PFHpA_00002	04/01/18	Aldrich, Lot BCM2579V			(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537 PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537 PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL
....LC537 PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
....LC537 PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	0.2 ug/mL
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL	LCMPFDA_00008	80 uL	13C2 PFHxA	0.2 ug/mL
...LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			LCMPFHxA_00009	80 uL	13C2 PFDA	0.4 ug/mL
...LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	0.4 ug/mL
LC537-L2_00016	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00014	34 uL	Perfluorobutanesulfonic acid (PFBS)	22.8888 ng/mL
							Perfluoroheptanoic acid	2.5245 ng/mL
							Perfluorohexanesulfonic acid	7.71585 ng/mL
							Perfluorononanoic acid	5.28909 ng/mL
							Perfluorooctanoic acid (PFOA)	4.97733 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	10.2169 ng/mL
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL
							13C4 PFOS	28.68 ng/mL
							13C2 PFDA	10 ng/mL
							13C2 PFHxA	10 ng/mL
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537-SU_00026	250 uL	Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	371.25 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	777.808 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
							Perfluorooctanoic acid (PFOA)	731.96 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
					LC537-PFHpA_00013	100 uL	Perfluoroheptanoic acid	9.9 ug/mL
					LC537-PFHxS_00008	300 uL	Perfluorohexanesulfonic acid	30.2582 ug/mL
					LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
					LC537-PFOA_00011	100 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
					LC537-PFOS_00006	400 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g
...LC537-PFHpA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537 PFHpA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL
....LC537 PFHpA_00002	04/01/18	Aldrich, Lot BCM2579V			(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537 PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537 PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL
....LC537 PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
....LC537 PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C2-PFOA	0.5 ug/mL
					LCMPFOS_00018	300 uL	13C4 PFOS	1.434 ug/mL
..LCM2PFOA_00005	06/19/18	Wellington Laboratories, Lot M2PFOA0613			(Purchased Reagent)		13C2-PFOA	50 ug/mL
..LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C4 PFOS	47.8 ug/mL
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	0.2 ug/mL
							13C2 PFHxA	0.2 ug/mL
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL	LCMPFDA_00008	80 uL	13C2 PFDA	0.4 ug/mL
					LCMPFHxA_00009	80 uL	13C2 PFHxA	0.4 ug/mL
...LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFDA	50 ug/mL
...LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	50 ug/mL
LC537-L3_00018	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00014	67 uL	Perfluorobutanesulfonic acid (PFBS)	45.1044 ng/mL
							Perfluoroheptanoic acid	4.97475 ng/mL
							Perfluorohexanesulfonic acid	15.2048 ng/mL
							Perfluorononanoic acid	10.4226 ng/mL
							Perfluorooctanoic acid (PFOA)	9.80826 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	20.1334 ng/mL
					LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL
							13C4 PFOS	28.68 ng/mL
					LC537-SU_00026	250 uL	13C2 PFDA	10 ng/mL
							13C2 PFHxA	10 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	375 uL	Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	371.25 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	777.808 ng/mL
							Perfluorooctanoic acid (PFOA)	731.96 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
					LC537-PFHpA_00013	100 uL	Perfluoroheptanoic acid	9.9 ug/mL
					LC537-PFHxS_00008	300 uL	Perfluorohexanesulfonic acid	30.2582 ug/mL
					LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
					LC537-PFOA_00011	100 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
					LC537-PFOS_00006	400 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g
...LC537-PFHpA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537_PFHpA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL
....LC537_PFHpA_00002	04/01/18	Aldrich, Lot BCM2579V			(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537_PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537_PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537_PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL
....LC537_PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537_PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
....LC537_PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C2-PFOA	0.5 ug/mL
..LCM2PFOA_00005	06/19/18	Wellington Laboratories, Lot M2PFOA0613			LCMPFOS_00018	300 uL	13C4 PFOS	1.434 ug/mL
...LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C2-PFOA	50 ug/mL
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	0.2 ug/mL
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL	LCMPFDA_00008	80 uL	13C2 PFHxA	0.2 ug/mL
...LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			LCMPFHxA_00009	80 uL	13C2 PFHxA	0.4 ug/mL
....LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFDA	50 ug/mL
LC537-L4_00017	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00014	135 uL	Perfluorobutanesulfonic acid (PFBS)	90.882 ng/mL
							Perfluoroheptanoic acid	10.0238 ng/mL
							Perfluorohexanesulfonic acid	30.6364 ng/mL
							Perfluorononanoic acid	21.0008 ng/mL
							Perfluorooctanoic acid (PFOA)	19.7629 ng/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-24648-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
							Perfluorooctanesulfonic acid (PFOS)	40.5672 ng/mL
					LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL
							13C4 PFOS	28.68 ng/mL
					LC537-SU_00026	250 uL	13C2 PFDA	10 ng/mL
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	375 uL	13C2 PFHxA	10 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	371.25 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
							Perfluorononanoic acid	777.808 ng/mL
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorooctanoic acid (PFOA)	731.96 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
							Perfluoroheptanoic acid	9.9 ug/mL
							Perfluorohexanesulfonic acid	30.2582 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
							Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
							Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
							Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
							Perfluoroheptanoic acid	990 ug/mL
....LC537-PFHxA_00002	04/01/18		Sigma, Lot MKBP8842V		(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g
...LC537-PFHxA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537-PFHxA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL
....LC537-PFHxA_00002	04/01/18		Aldrich, Lot BCM2579V		(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537-PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL
....LC537-PFHxS_00002	04/01/18		Sigma, Lot BCBL3545V		(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL
....LC537-PFNA_00002	04/01/18		TCI America, Lot QN44F		(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
....LC537-PFOA_00002	11/04/18		Fluka, Lot SZBD308XV		(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537-PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
....LC537-PFOS_00002	08/09/17		Fluka, Lot SZBC222XV		(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C2-PFOA	0.5 ug/mL
..LCM2PFOA_00005	06/19/18		Wellington Laboratories, Lot M2PFOA0613		LCMPFOS_00018	300 uL	13C4 PFOS	1.434 ug/mL
...LCMPFOS_00018	08/03/21		Wellington Laboratories, Lot MPFOS0816		(Purchased Reagent)		13C2-PFOA	50 ug/mL
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	47.8 ug/mL
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL			13C2 PFHxA	0.2 ug/mL
					LCMPFDA_00008	80 uL	13C2 PFDA	0.2 ug/mL
					LCMPFHxA_00009	80 uL	13C2 PFHxA	0.4 ug/mL
...LCMPFDA_00008	08/19/20		Wellington Laboratories, Lot MPFDA0815		(Purchased Reagent)		13C2 PFDA	0.4 ug/mL
...LCMPFHxA_00009	04/09/20		Wellington Laboratories, Lot MPFHxA0415		(Purchased Reagent)		13C2 PFHxA	50 ug/mL

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-24648-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration		
					Reagent ID	Volume Added				
LC537-L5_00019	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00014	200 uL	Perfluorobutanesulfonic acid (PFBS)	134.64 ng/mL		
							Perfluoroheptanoic acid	14.85 ng/mL		
							Perfluorohexanesulfonic acid	45.3873 ng/mL		
							Perfluorononanoic acid	31.1123 ng/mL		
							Perfluorooctanoic acid (PFOA)	29.2784 ng/mL		
					Perfluorooctanesulfonic acid (PFOS)	60.0996 ng/mL				
					LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL		
							13C4 PFOS	28.68 ng/mL		
					LC537-SU_00026	250 uL	13C2 PFDA	10 ng/mL		
13C2 PFHxA	10 ng/mL									
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	375 uL	Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL		
							Perfluoroheptanoic acid	371.25 ng/mL		
							Perfluorohexanesulfonic acid	1134.68 ng/mL		
							Perfluorononanoic acid	777.808 ng/mL		
							Perfluorooctanoic acid (PFOA)	731.96 ng/mL		
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL		
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL		
							LC537-PFHpA_00013	100 uL	Perfluoroheptanoic acid	9.9 ug/mL
							LC537-PFHxS_00008	300 uL	Perfluorohexanesulfonic acid	30.2582 ug/mL
							LC537-PFNA_00011	200 uL	Perfluorononanoic acid	20.7415 ug/mL
							LC537-PFOA_00011	100 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
							LC537-PFOS_00006	400 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
...LC537-PFBS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537_PFBS_00002	0.0102 g	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL		
....LC537_PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluorobutanesulfonic acid (PFBS)	1 g/g		
...LC537-PFHpA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537 PFHpA_00002	0.0568 g	Perfluoroheptanoic acid	990 ug/mL		
....LC537 PFHpA_00002	04/01/18	Aldrich, Lot BCM2579V			(Purchased Reagent)		Perfluoroheptanoic acid	0.99 g/g		
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537 PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	1008.61 ug/mL		
....LC537 PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorohexanesulfonic acid	0.9094 g/g		
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFNA_00002	0.007 g	Perfluorononanoic acid	1037.08 ug/mL		
....LC537 PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorononanoic acid	0.963 g/g		
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537 PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	1951.89 ug/mL		
....LC537 PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	0.999 g/g		
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537_PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL		
....LC537_PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g		
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C2-PFOA	0.5 ug/mL		
					LCMPFOS_00018	300 uL	13C4 PFOS	1.434 ug/mL		
..LCM2PFOA_00005	06/19/18	Wellington Laboratories, Lot M2PFOA0613			(Purchased Reagent)		13C2-PFOA	50 ug/mL		
..LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C4 PFOS	47.8 ug/mL		
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	0.2 ug/mL		

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento

Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL	LCMPFDA_00008	80 uL	13C2 PFHxA	0.2 ug/mL
					LCMPFHxA_00009	80 uL	13C2 PFDA	0.4 ug/mL
...LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFHxA	0.4 ug/mL
...LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFDA	50 ug/mL
LC537-L6_00016	06/14/17	12/23/16	MeOH/H2O, Lot 090285	5 mL	LC537-HSP_00014	265 uL	Perfluorobutanesulfonic acid (PFBS)	178.398 ng/mL
							Perfluoroheptanoic acid	19.6763 ng/mL
							Perfluorohexanesulfonic acid	60.1382 ng/mL
							Perfluorononanoic acid	41.2238 ng/mL
							Perfluorooctanoic acid (PFOA)	38.7939 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	79.632 ng/mL
					LC537-IS_00028	100 uL	13C2-PFOA	10 ng/mL
					LC537-SU_00026	250 uL	13C4 PFOS	28.68 ng/mL
.LC537-HSP_00014	06/22/17	12/22/16	Methanol, Lot 090285	10000 uL	LC537SPIM_00018	375 uL	13C2 PFDA	10 ng/mL
							13C2 PFHxA	10 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	3366 ng/mL
							Perfluoroheptanoic acid	371.25 ng/mL
							Perfluorohexanesulfonic acid	1134.68 ng/mL
..LC537SPIM_00018	06/22/17	12/22/16	Methanol, Lot 104453	10000 uL	LC537-PFBS_00006	440 uL	Perfluorononanoic acid	777.808 ng/mL
							Perfluorooctanoic acid (PFOA)	731.96 ng/mL
							Perfluorooctanesulfonic acid (PFOS)	1502.49 ng/mL
							Perfluorobutanesulfonic acid (PFBS)	89.76 ug/mL
							Perfluoroheptanoic acid	9.9 ug/mL
							Perfluorohexanesulfonic acid	30.2582 ug/mL
...LC537PFBS_00006	07/28/17	07/28/16	Methanol, Lot 104453	5 mL	LC537-PFHxA_00013	100 uL	Perfluorononanoic acid	20.7415 ug/mL
					LC537-PFHxS_00008	300 uL	Perfluorooctanoic acid (PFOA)	19.5189 ug/mL
					LC537-PFNA_00011	200 uL	Perfluorooctanesulfonic acid (PFOS)	40.0664 ug/mL
					LC537-PFOA_00011	100 uL	Perfluorobutanesulfonic acid (PFBS)	2040 ug/mL
....LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	5 mL	LC537-PFOS_00006	400 uL	Perfluorobutanesulfonic acid (PFBS)	1 g/g
....LC537-PFBS_00002	04/01/18	Sigma, Lot MKBP8842V			(Purchased Reagent)		Perfluoroheptanoic acid	990 ug/mL
...LC537-PFHxA_00013	06/22/17	12/22/16	Methanol, Lot 090285	56.8 mL	LC537-PFHxA_00002	0.0568 g	Perfluoroheptanoic acid	0.99 g/g
...LC537-PFHxA_00002	04/01/18	Aldrich, Lot BCM2579V			(Purchased Reagent)		Perfluorohexanesulfonic acid	1008.61 ug/mL
...LC537-PFHxS_00008	07/28/17	07/28/16	Methanol, Lot 090285	5.5 mL	LC537-PFHxS_00002	0.0061 g	Perfluorohexanesulfonic acid	0.9094 g/g
...LC537-PFHxS_00002	04/01/18	Sigma, Lot BCBL3545V			(Purchased Reagent)		Perfluorononanoic acid	1037.08 ug/mL
...LC537-PFNA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFNA_00002	0.007 g	Perfluorononanoic acid	0.963 g/g
...LC537-PFNA_00002	04/01/18	TCI America, Lot QN44F			(Purchased Reagent)		Perfluorooctanoic acid (PFOA)	1951.89 ug/mL
...LC537-PFOA_00011	11/21/17	11/21/16	Methanol, Lot 090285	6.5 mL	LC537-PFOA_00002	0.0127 g	Perfluorooctanoic acid (PFOA)	0.999 g/g
...LC537-PFOA_00002	11/04/18	Fluka, Lot SZBD308XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	1001.66 ug/mL
...LC537-PFOS_00006	07/28/17	07/28/16	Methanol, Lot 090285	6 mL	LC537-PFOS_00002	0.0066 g	Perfluorooctanesulfonic acid (PFOS)	0.9106 g/g
....LC537-PFOS_00002	08/09/17	Fluka, Lot SZBC222XV			(Purchased Reagent)		Perfluorooctanesulfonic acid (PFOS)	

REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
.LC537-IS_00028	06/19/17	12/19/16	Methanol, Lot 090285	10000 uL	LCM2PFOA_00005	100 uL	13C2-PFOA	0.5 ug/mL
					LCMPFOS_00018	300 uL	13C4 PFOS	1.434 ug/mL
..LCM2PFOA_00005	06/19/18	Wellington Laboratories, Lot M2PFOA0613			(Purchased Reagent)		13C2-PFOA	50 ug/mL
..LCMPFOS_00018	08/03/21	Wellington Laboratories, Lot MPFOS0816			(Purchased Reagent)		13C4 PFOS	47.8 ug/mL
.LC537-SU_00026	06/14/17	12/16/16	Methanol, Lot 104453	20000 uL	LC537-SU_00025	10000 uL	13C2 PFDA	0.2 ug/mL
							13C2 PFHxA	0.2 ug/mL
..LC537-SU_00025	06/14/17	12/14/16	Methanol, Lot 104453	10000 uL	LCMPFDA_00008	80 uL	13C2 PFDA	0.4 ug/mL
					LCMPFHxA_00009	80 uL	13C2 PFHxA	0.4 ug/mL
...LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFDA	50 ug/mL
...LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	50 ug/mL
LC537-SU_00027	06/19/17	12/19/16	Methanol, Lot 104453	20000 uL	LCMPFDA_00008	80 uL	13C2 PFDA	0.2 ug/mL
					LCMPFHxA_00009	80 uL	13C2 PFHxA	0.2 ug/mL
.LCMPFDA_00008	08/19/20	Wellington Laboratories, Lot MPFDA0815			(Purchased Reagent)		13C2 PFDA	50 ug/mL
.LCMPFHxA_00009	04/09/20	Wellington Laboratories, Lot MPFHxA0415			(Purchased Reagent)		13C2 PFHxA	50 ug/mL

Reagent

LC537_PFB_00002

4/1/15 SPV

SIGMA-ALDRICH®

sigma-aldrich.com

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:

Nonafluorobutane-1-sulfonic acid - 97%

Product Number:

562629

Batch Number:

MKBP8842V

Brand:

ALDRICH

CAS Number:

375-73-5

MDL Number:

MFCD01320794

Formula:

C₄HF₉O₃S

Formula Weight:


300.10 g/mol

Storage Temperature:

Store at 2 - 8 °C

Quality Release Date:

11 OCT 2013



PFBS

Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Fluorine NMR Spectrum	Conforms to Structure	Conforms
Purity (Titration by NaOH)	96.5 - 103.5 %	101.6 %

Jamie Gleason

Jamie Gleason, Manager

Quality Control

Milwaukee, Wisconsin US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFB2_00001



The Power to Question

CERTIFICATE OF ANALYSIS

Catalog Number: sc-236187
Product Name: Nonafluorobutane-1-sulfonic acid
CAS Number: 375-73-5
Molecular Formula: $C_4HF_9O_3S$
Molecular Weight: 300.10
Lot Number: H0112

Test		Result
Refractive Index	1.3200 to 1.3290	1.3219
Purity (Titration)	min. 98.0%	99.8%

Test Conditions: Refractive Index: $n_{20/D}$

Reagent


LC537_PFH_pA_00002

Certificate of Analysis

Product Name: PERFLUOROHEPTANOIC ACID
Product Number: 342041
Batch Number: BCBM2579V
Brand: Aldrich
CAS Number: 375-85-9
Formula: $\text{CF}_3(\text{CF}_2)_5\text{CO}_2\text{H}$
Formula Weight: 364.06
Quality Release Date: 06 DEC 2013
Recommended Retest Date: OCT 2018

PFH₇A

TEST	SPECIFICATION	RESULT
APPEARANCE (COLOR)	COLORLESS OR WHITE	WHITE
APPEARANCE (FORM)	LIQUID OR SOLID	SOLID
TITRATION	98.5 - 101.5 %	99.8 %
TITRATION (METHOD)	-	BACK TITRATION
PURITY (GC AREA %)	≥ 98.5 %	99.5 %
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS



Dr. Claudia Geitner
Manager Quality Control
Buchs, Switzerland

Sigma-Aldrich warrants that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFHxS_00002

Y: 4/1/15 SW

SIGMA-ALDRICH

3050 Spruce Street, Saint Louis, MO 63103 USA
Email USA: techserv@sial.com Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name: TRIDECAFLUOROHXANE-1-SULFONIC ACID POTASSIUM SALT
Product Number: 50929
Batch Number: BCBL3545V
Brand: Aldrich
CAS Number: 3871-99-6
Formula: $C_{13}F_{27}KO_2S$
Formula Weight: 438.20
Quality Release Date: 20 JUN 2013

PFH₁₃S-K

TEST	SPECIFICATION	RESULT
APPEARANCE (COLOR)	WHITE TO FAINT BEIGE	WHITE
APPEARANCE (FORM)	POWDER OR CRYSTALS	POWDER
TITRATION (ION EXCHANGE)	≥ 98.0 %	99.5 %
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS

Claudia Geitner

Dr. Claudia Geitner
Manager Quality Control
Buchs, Switzerland

$$MW_{corr} = \frac{(k_{form}) - (K) + (H)}{(438.20 - 39.10 + 1.01)} = \frac{0.91307}{438.20 (k_{form})} = 0.91307 \text{ (anion form)}$$

$$Purity = 90.94 \% \text{ w/m.w correction}$$

SW 4/1/15

Sigma-Aldrich warrants that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Reagent

LC537_PFNA_00002

R: 4/1/15 SKV



Certificate of Analysis

Apr 2, 2015 (JST)

TOKYO CHEMICAL INDUSTRY CO., LTD.
4-10-1 Nihonbashi-Honcho, Chuo-ku, Tokyo 103-0023 Japan

Chemical Name: Heptadecafluorononanoic Acid		
Product Number: H0843	Lot: QN44F	
CAS: 375-95-1		

Tests	Results	Specifications
Purity(GC)	96.3 %	min. 95.0 %
Purity(Neutralization titration)	98.1 %	min. 95.0 %
Melting point	63.3 deg-C	62.0 to 67.0 deg-C

TCI Lot numbers are 4-5 characters in length.
Characters listed after the first 4-5 characters are control numbers for internal purpose only.

Customer service:

TCI AMERICA
Tel: +1-800-423-8616 / +1-503-283-1681
Fax: +1-888-520-1075 / +1-503-283-1987
E-mail: Sales-US@TCIchemicals.com

PFNA

Reagent

LC537_PFOA_00002

11/3/2015 2

SIGMA-ALDRICH®

CERTIFICATE OF ANALYSIS

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze
Telefon: +49 5137 8238-150

Seelze, 13.11.2013/505378/13/24029

Order-No.:

Customer-No.:

Order-Code:

Quantity:

Production Date: 04.Nov.2013

Expiry Date: 04.Nov.2018

Article/Product: 33824

Batch : SZBD308XV

Pentadecafluorooctanoic acid OEKANAL®

PFOA

Reference Material (RM)

1. General Information

Formula: C₈HF₁₅O₂

CAS-No.: [335-67-1]

Usage : PFOA

Molar mass: 414.07 g/Mole

Recomm. storage temp.: roomtemp.

The estimated uncertainty of a single measurement of the assay can be expected to be 0.5 % relative (confidence level = 95%, n= 6) whereby the assay measurements are calculated by 100% minus found impurities.

2. Batch Analysis

identity (GC-MS)

Assay (GCMS)

Date of Analysis

complying

99.4 %

13.Nov.2013

3. Advice and Remarks


- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- Whenever the container is opened for removal of aliquot portions of the substance, the person handling the substance must assure, that the integrity of the substance is maintained and proper records of all its handlings are kept. Special care has to be taken to avoid any contamination or adulteration of the substance.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Sigma-Aldrich Laborchemikalien GmbH
Quality Management SA-LC

This document was produced electronically and is valid without a signature

GC/MS-Method**Analytical Department****Article:** Pentadecafluorooctanoic acid OEKANAL**Article-No.:** 33824**Batch:** SZBD308XV**Column:** XTI-5 (Restek); 30 m; fs cap.; I.D.:0.25 mm; 1 µm df**Injector:** Split mode**Injection:** approx. 1 µl of reaction mixture with MSTFA (approx. 10 mg + 200 µl MSTFA)**Inj.-temp.:** 280°C**Oven-temp.:** 40°C (for 2 min) to 320°C (6°C/min) hold for 2 min**Split:** 1:100**Flow:** 1 ml He/min (Constant flow mode)**Detector:** MSD**Mass range:** 10-600 amu (Scan mode)**Evaluation:** Purity: Total Ion Chromatogram
(MSTFA and reaction products blinded out in report)


Identity: Mass spectrum complies

Operator: Ahrens / 2013-11-13**Total Ion Chromatogram:**

Ret.time	Area	Area-%	Com
----------	------	--------	-----

11.54	565.1670	99.4	Pentadecafluorooctanoic acid (as TMS-ester)
-------	----------	------	---

11.95	3.6792	0.64	
-------	--------	------	--

Mass spectrum (rt = 11.54 min):

Reagent

LC537_PFOA2_00001

Certificate of Analysis

Alfa Aesar
A Johnson Matthey Company

Product No.: L08862
Product: Perfluorooctanoic acid, 95%
Lot No.: D24Y026

PFOA

Appearance White solid
Melting point 58 - 60°C
Assay 99 %
Identity Matches reference

This document has been electronically generated and does not require a signature.

www.alfa.com

NORTH AMERICA
Tel: +1-800-343-0660 or
+1-978-521-6300
Fax: +1-800-322-4757
Email: info@alfa.com

GERMANY
Tel: 00800 4566 4566 or
+49 721 84007 280
Fax: 00800 4577 4577 or
+49 721 84007 300
Email: Eurosales@alfa.com

UNITED KINGDOM
Tel: 0800-801812 or
+44 (0)1524-850506
Fax: +44 (0)1524-850608
Email: UKsales@alfa.com

FRANCE
Tel: 0800 03 51 47 or
+33 (0)3 6862 2690
Fax: 0800 10 20 67 or
+33 (0)3 8862 6864
Email: frventes@alfa.com

INDIA
Tel: +91 8008 812424 or
+91 8008 812525 or
+91 8008 812626
Fax: +91 8418 260060
Email: India@alfa.com

CHINA
Tel: +86 (010) 8567-8600
Fax: +86 (010) 8567-8601
Email: saleschina@alfa-asia.com

KOREA
Tel: +82-2-3140-6000
Fax: +82-2-3140-6002
Email: saleskorea@alfa-asia.com

Reagent

LC537_PFOS_00002

F: 4/1/15 SV

SIGMA-ALDRICH®

CERTIFICATE OF ANALYSIS

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze
Telefon: +49 5137 8238-150

Seelze, 13.08.2012/419060/12/17583

Order-No.:

Customer-No.:

Order-Code:

Quantity:

Production Date: 09.Aug.2012

Expiry Date: 09.Aug.2017 - *not date*

Article/Product: 33829

Batch : SZBC222XV

Heptadecafluorooctanesulfonic acid potassium salt OEKANAL®

PFOS-K⁺

Reference Material (RM)

1. General Information

Formula: C8F17KO3S
CAS-No.: [2795-39-3]
Usage : PFOS

Molar mass: 538.22 g/Mole
Recomm. storage temp.: roomtemp.

The estimated uncertainty of a single measurement of the assay can be expected to be 0.5 % relative (confidence level = 95%, n= 6) whereby the assay measurements are calculated by 100% minus found impurities.

2. Batch Analysis

Identity
Assay (LC-MS)
Date of Analysis

complying
98 %
10.Aug.2012

$$\text{P.W.-Correction: } \frac{538.22 - 39.10 + 1.01}{538.22} = \frac{500.13}{538.22} = 0.92923$$

Purity = 91.06%

3. Advice and Remarks

- The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- Whenever the container is opened for removal of aliquot portions of the substance, the person handling the substance must assure, that the integrity of the substance is maintained and proper records of all its handlings are kept. Special care has to be taken to avoid any contamination or adulteration of the substance.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Sigma-Aldrich Laborchemikalien GmbH
Quality Management SA-LC

Reagent

LC537_PFOS2_00001

Certificate of Analysis

Inv 820
12LCMS 0579

Product Name: HEPTADEC AFLUORO OCTANESULFONIC ACID TETRAETHYLAMMONIUM SALT
98 %
Product Number: 365289
Product Brand: Aldrich
Molecular Formula: $C_{16}H_{20}F_{17}NO_3S$
Molecular Mass: 629.37
CAS Number: 56773-42-3

TEST	SPECIFICATION	LOT BCBF5116V RESULTS
APPEARANCE (COLOR)	OFF-WHITE TO WHITE	WHITE
APPEARANCE (FORM)	POWDER, LUMPS OR CHUNKS	POWDER WITH LUMPS
CARBON CONTENT	29.77 % - 31.29 %	30.52
INFRARED SPECTRUM	CONFORMS TO STRUCTURE	CONFORMS

QC RELEASE DATE 13/APR/11



E. Schwarz

Edeltraud Schwarzler, Manager
Quality Control
Buchs, Switzerland

$$MW \text{ correction} = \frac{500.125}{629.37} = 0.7946$$

~~79.46%~~ $\frac{500.125}{629.37} \times 100 = 79.46\%$ $\frac{500.125}{629.37} \times 100 = 79.46\%$ $\frac{500.125}{629.37} \times 100 = 79.46\%$

$$Purity + MW \text{ Correction} = 77.87\%$$



	$C_8F_{17}SO_3 + H$
C = 12.011	96.088
F = 18.998	322.966
S = 32.066	32.066
O = 15.999	47.997
H = 1.008	1.008
N = 14.007	-
	<u>500.125</u>

$C_8H_{20}N$
96.088
-
-
-
20.160
14.007
<u>130.255</u>
→

Sigma-Aldrich warrants, that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice for additional terms and conditions of sale. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

Certificate of Origin

Product Name: Heptadecafluorooctanesulfonic acid tetraethylammonium salt
98 %
Product Number: 365289
Product Brand: Aldrich
Lot: BCBF5116V
Molecular Formula: $C_{16}H_{20}F_{17}NO_3S$
Molecular Mass: 629.37
CAS Number: 56773-42-3
Date of Issue: 30-MAR-11

Country of Origin China

product is of synthetic origin	yes
only synthetic materials used in the manufacturing process	yes
compounds of animal origin used	no
genetically modified organisms used	no
allergenic materials used	no
procedures in place to avoid cross contamination with residue of animal, human, GMO or allergenes in manufacturing process	yes

Sigma-Aldrich has quality systems and procedures in place for monitoring the production process, traceability and batch consistency.

Document issued by Sigma-Aldrich Corporation "Sigma-Aldrich". This document is valid without signature and has been produced digitally.

This information is to be used for the purpose of determining animal or other biological origin only and not to be confused with "Country of Origin" for import/export purposes. Data provided on this document are property of Sigma-Aldrich.

This information is considered accurate and reliable as of the date appearing on the document and is presented in good faith.

Sigma-Aldrich shall not be held liable for any damage resulting from handling or from processing the above product(s). This document does not make any warranty, express or implied, of fitness for any particular use of the product(s). Purchaser must determine the suitability of the product(s) for its use under the applicable law and regulations.

For further questions please contact your local Sigma-Aldrich representative.

We are committed to the success of our Customers, Employees and Shareholders through leadership in Life Science, High Technology and Service.

Reagent

LCM2PFOA_00003



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:


M2PFOA

LOT NUMBER:

M2PFOA0312

COMPOUND:Perfluoro-n-[1,2-¹³C₂]octanoic acid**STRUCTURE:****CAS #:**

Not available

**MOLECULAR FORMULA:**¹³C₂¹²C₆HF₁₅O₂**CONCENTRATION:**

50 ± 2.5 µg/ml

MOLECULAR WEIGHT:

416.05

SOLVENT(S):

Methanol

Water (<1%)

CHEMICAL PURITY:

>98%

ISOTOPIC PURITY:≥99%¹³C**LAST TESTED:** (mm/dd/yyyy)

03/19/2012

(1,2-¹³C₂)**EXPIRY DATE:** (mm/dd/yyyy)

03/19/2017

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date:

01/09/2013
(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. They are designed to be used as reference standards for the identification and/or quantification of specific chemical compound(s).

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Material Safety Data Sheets (MSDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product, unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, x-ray crystallography and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS and/or LC/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \dots, x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external, ISO/IEC 17025:2005 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration for the period of time specified by the expiry date in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to ISO 9001:2008 by SAI Global, ISO/IEC 17025:2005 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34:2009 by ACLASS (certificate number AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: M2PFOA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 60% (80:20 MeOH:ACN) / 40% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 6.5 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (250 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: M2PFOA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml M2PFOA)

Mobile phase: Isocratic 70% (80:20 MeOH:ACN) / 30% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

MS Parameters

Collision Gas (mbar) = $3.35\text{e-}3$
Collision Energy (eV) = 11

Reagent

LCM2PFOA_00005



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:


M2PFOA

LOT NUMBER:

M2PFOA0613

COMPOUND:Perfluoro-n-[1,2-¹³C₂]octanoic acid**STRUCTURE:****CAS #:**

Not available

**MOLECULAR FORMULA:**¹³C₂¹²C₆H₁₅O₂**MOLECULAR WEIGHT:**

416.05

CONCENTRATION:

50 ± 2.5 µg/ml

SOLVENT(S):

Methanol

Water (<1%)

CHEMICAL PURITY:

>98%

ISOTOPIC PURITY:≥99% ¹³C**LAST TESTED:** (mm/dd/yyyy)

06/19/2013

(1,2-¹³C₂)**EXPIRY DATE:** (mm/dd/yyyy)

06/19/2018

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 07/16/2013

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. They are designed to be used as reference standards for the identification and/or quantification of specific chemical compound(s).

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Material Safety Data Sheets (MSDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product, unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, x-ray crystallography and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS and/or LC/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters

$x_1, x_2, \dots, x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external, ISO/IEC 17025:2005 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration for the period of time specified by the expiry date in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to ISO 9001:2008 by SAI Global, ISO/IEC 17025:2005 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34:2009 by ACLASS (certificate number AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: M2PFOA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 1.5 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (150 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: M2PFOA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml M2PFOA)

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

MS Parameters

Collision Gas (mbar) = 3.66e-3
Collision Energy (eV) = 11

Reagent

LCMPFDA_00008



605243

ID: LCMFPA_00008

Exp: 08/19/20 Ppd: CBW

13C2-Perfluorodecanoic acid

Rec. 3/29/16 JRB ✓



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:


MPFDA

LOT NUMBER:

MPFDA0815

COMPOUND:Perfluoro-n-[1,2-¹³C₂]decanoic acid**STRUCTURE:****CAS #:**

Not available

**MOLECULAR FORMULA:**¹³C₂¹²C₈H₁₉O₂**MOLECULAR WEIGHT:**

516.07

CONCENTRATION:

50 ± 2.5 µg/ml

SOLVENT(S):

Methanol

Water (<1%)

CHEMICAL PURITY:

>98%

ISOTOPIC PURITY:≥99% ¹³C(1,2-¹³C₂)**LAST TESTED:** (mm/dd/yyyy)

08/19/2015

EXPIRY DATE: (mm/dd/yyyy)

08/19/2020

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.
- Contains < 0.1% of ¹³C₁-PFNA.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 08/21/2015

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \dots, x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFDA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro *micro* API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 2 min
before returning to initial conditions in 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (150 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 50
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFDA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μ l (500 ng/ml MPFDA)

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 μ l/min

MS Parameters

Collision Gas (mbar) = 3.35e-3
Collision Energy (eV) = 13

Reagent

LCMPFHxA_00009



605244

ID: LCMPFHxA_00009

Exp: 04/09/20 Prod: CBW

13C2-Perfluorohexanoic ac

Rec. 3/29/16 JRB ✓



WELLINGTON LABORATORIES


CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE: MPFHxA
COMPOUND: Perfluoro-n-[1,2-¹³C₂]hexanoic acid

LOT NUMBER: MPFHxA0415

STRUCTURE:

CAS #: Not available



MOLECULAR FORMULA: ¹³C₂¹²C₄HF₁₁O₂
CONCENTRATION: 50 ± 2.5 µg/ml

MOLECULAR WEIGHT: 316.04
SOLVENT(S): Methanol
Water (<1%)

CHEMICAL PURITY: >98%
LAST TESTED: (mm/dd/yyyy) 04/09/2015
EXPIRY DATE: (mm/dd/yyyy) 04/09/2020
RECOMMENDED STORAGE: Store ampoule in a cool, dark place

ISOTOPIC PURITY: ≥99%¹³C
(1,2-¹³C₂)

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)
Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains 4 mole eq. of NaOH to prevent conversion of the carboxylic acid to the methyl ester.
- Contains < 0.1% of perfluoro-n-hexanoic acid and ~ 0.3% of perfluoro-n-octanoic acid.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 04/14/2015
(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \dots, x_n$ on which it depends is:

$$u_c(v(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:

Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFHxA; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro micro API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient
Start: 50% (80:20 MeOH:ACN) / 50% H₂O
(both with 10 mM NH₄OAc buffer)
Ramp to 90% organic over 7 min and hold for 2 min
before returning to initial conditions over 0.5 min.
Time: 10 min


Flow: 300 μ l/min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)
Capillary Voltage (kV) = 2.00
Cone Voltage (V) = 15.00
Cone Gas Flow (l/hr) = 100
Desolvation Gas Flow (l/hr) = 750

Figure 2: MPFHxA; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μ l (500 ng/ml MPFHxA)

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H₂O
(both with 10 mM NH₄OAc buffer)

Flow: 300 μ l/min

MS Parameters

Collision Gas (mbar) = 3.20e-3
Collision Energy (eV) = 10

Reagent

LCMPFOS_00018

R: SBC 9/22/16

738686
ID: LCMFOS_00018
Exp: 08/03/21 Prod: SBC
13C4-Perfluorooctanesulfo



WELLINGTON LABORATORIES

CERTIFICATE OF ANALYSIS DOCUMENTATION

PRODUCT CODE:

MPFOS

LOT NUMBER:

MPFOS0816


COMPOUND:

Sodium perfluoro-1-[1,2,3,4-¹³C₄]octanesulfonate

STRUCTURE:

CAS #:

Not available



MOLECULAR FORMULA:

¹³C₄¹²C₄F₁₇SO₃Na

CONCENTRATION:

50.0 ± 2.5 µg/ml (Na salt)

47.8 ± 2.4 µg/ml (MPFOS anion)

CHEMICAL PURITY:

>98%

LAST TESTED: (mm/dd/yyyy)

08/03/2016

EXPIRY DATE: (mm/dd/yyyy)

08/03/2021

RECOMMENDED STORAGE:

Store ampoule in a cool, dark place

MOLECULAR WEIGHT:

526.08

SOLVENT(S):

Methanol

ISOTOPIC PURITY:

≥99% ¹³C
(1,2,3,4-¹³C₄)

DOCUMENTATION/ DATA ATTACHED:

Figure 1: LC/MS Data (TIC and Mass Spectrum)

Figure 2: LC/MS/MS Data (Selected MRM Transitions)

ADDITIONAL INFORMATION:

- See page 2 for further details.
- Contains ~ 0.8% Sodium perfluoro-1-[1,2,3-¹³C₃]heptanesulfonate.

FOR LABORATORY USE ONLY: NOT FOR HUMAN OR DRUG USE

Certified By:

B.G. Chittim

Date: 08/05/2016

(mm/dd/yyyy)

Wellington Laboratories Inc., 345 Southgate Dr. Guelph ON N1G 3M5 CANADA
519-822-2436 • Fax: 519-822-2849 • info@well-labs.com

INTENDED USE:

The products prepared by Wellington Laboratories Inc. are for laboratory use only. This certified reference material (CRM) was designed to be used as a standard for the identification and/or quantification of the specific chemical compound it contains.

HAZARDS:

This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion. All procedures should be carried out in a well-functioning fume hood and suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed of according to national and regional regulations. Safety Data Sheets (SDSs) are available upon request.

SYNTHESIS / CHARACTERIZATION:

Where possible, all of our products are synthesized using single-product unambiguous routes. They are then characterized, and their structures and purities confirmed, using a combination of the most relevant techniques, such as NMR, GC/MS, LC/MS/MS, SFC/UV/MS/MS, x-ray crystallography, and melting point. Isotopic purities of mass-labelled compounds are also confirmed using HRGC/HRMS and/or LC/MS/MS.

HOMOGENEITY:

Prior to solution preparation, crystalline material is tested for homogeneity using a variety of techniques (as stated above) and its solubility in a given diluent is taken into consideration. Duplicate solutions of a new product are prepared from the same crystalline lot and, after the addition of an appropriate internal standard, they are compared by GC/MS, LC/MS/MS and/or SFC/UV/MS/MS. The relative response factors of the analyte of interest in each solution are required to be <5% RSD. New solution lots of existing products are compared to older lots in the same manner, which further confirms the homogeneity of the crystalline material as well as the stability and homogeneity of the solutions in the storage containers.

UNCERTAINTY:

The maximum combined relative standard uncertainty of our reference standard solutions is calculated using the following equation:

The combined relative standard uncertainty, $u_c(y)$, of a value y and the uncertainty of the independent parameters $x_1, x_2, \dots, x_n$ on which it depends is:

$$u_c(y(x_1, x_2, \dots, x_n)) = \sqrt{\sum_{i=1}^n u(y, x_i)^2}$$

where x is expressed as a relative standard uncertainty of the individual parameter.

The individual uncertainties taken into account include those associated with weights (calibration of the balance) and volumes (calibration of the volumetric glassware). An expanded maximum combined percent relative uncertainty of $\pm 5\%$ (calculated with a coverage factor of 2 and a level of confidence of 95%) is stated on the Certificate of Analysis for all of our products.

TRACEABILITY:

All reference standard solutions are traceable to specific crystalline lots. The microbalances used for solution preparation are regularly tested by an external ISO/IEC 17025 accredited calibration company. In addition, their calibration is verified prior to each weighing using NIST and/or NRC traceable external weights. All volumetric glassware used is of Class A tolerance and has been tested according to the appropriate ASTM procedures, which are ultimately traceable to NIST. For certain products, traceability to international interlaboratory studies has also been established.

EXPIRY DATE / PERIOD OF VALIDITY:


Ongoing stability studies of this product have demonstrated stability in its composition and concentration, until the specified expiry date, in the unopened ampoule. Monitoring for any degradation or change in concentration of the listed analyte(s) is performed on a routine basis.

LIMITED WARRANTY:

At the time of shipment, all products are warranted to be free of defects in material and workmanship and to conform to the stated technical and purity specifications.



QUALITY MANAGEMENT:

This product was produced using a Quality Management System registered to the latest versions of ISO 9001 by SAI Global, ISO/IEC 17025 by the Canadian Association for Laboratory Accreditation Inc. (CALA; A 1226), and ISO GUIDE 34 by ANSI-ASQ National Accreditation Board (ANAB; AR-1523).



For additional information or assistance concerning this or any other products from Wellington Laboratories Inc., please visit our website at www.well-labs.com or contact us directly at info@well-labs.com

Figure 1: MPFOS; LC/MS Data (TIC and Mass Spectrum)



Conditions for Figure 1:

LC: Waters Acquity Ultra Performance LC
MS: Micromass Quattro micro API MS

Chromatographic Conditions

Column: Acquity UPLC BEH Shield RP₁₈
1.7 μ m, 2.1 x 100 mm

Mobile phase: Gradient

Start: 45% (80:20 MeOH:ACN) / 55% H₂O
(both with 10 mM NH₄OAc buffer)

Ramp to 90% organic over 7 min and hold for 1.5 min
before returning to initial conditions in 0.5 min.

Time: 10 min

MS Parameters

Experiment: Full Scan (225 - 850 amu)

Source: Electrospray (negative)

Capillary Voltage (kV) = 3.00


Cone Voltage (V) = 60.00

Cone Gas Flow (l/hr) = 50

Desolvation Gas Flow (l/hr) = 750

Flow: 300 μ l/min

Figure 2: MPFOS; LC/MS/MS Data (Selected MRM Transitions)



Conditions for Figure 2:

Injection: Direct loop injection
10 μl (500 ng/ml MPFOS)

Mobile phase: Isocratic 80% (80:20 MeOH:ACN) / 20% H_2O
(both with 10 mM NH_4OAc buffer)

Flow: 300 $\mu\text{l}/\text{min}$

MS Parameters

Collision Gas (mbar) = 3.46e-3
Collision Energy (eV) = 40

Method 537 DOD

Perfluorinated Alkyl Acids (LC/MS)
by Method 537 DOD

FORM II
LCMS SURROGATE RECOVERY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Matrix: Water Level: Low

GC Column (1): Acquity ID: 2.1 (mm)

Client Sample ID	Lab Sample ID	PFHxA #	PFDA #
WI-AF-3RW38-1216	320-24648-1	101	119
WI-AF-3FB38-1216	320-24648-2	100	106
WI-AF-3RW39-1216	320-24648-3	103	101
WI-AF-3FB39-1216	320-24648-4	101	107
WI-AF-3RW40-1216	320-24648-5	110	123
WI-AF-3RW40P-1216	320-24648-6	96	107
WI-AF-3FB40-1216	320-24648-7	99	108
WI-AF-3RW41-1216	320-24648-8	91	105
WI-AF-3FB41-1216	320-24648-9	102	108
	MB 320-143783/1-A	111	116
	LCS 320-143783/2-A	121	121
	LCSD 320-143783/3-A	111	116

PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA

QC LIMITS
70-130
70-130

Column to be used to flag recovery values

FORM III
LCMS LAB CONTROL SAMPLE RECOVERY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
SDG No.: _____
Matrix: Water Level: Low Lab File ID: 02JAN2017A6A_006.d
Lab ID: LCS 320-143783/2-A Client ID: _____

COMPOUND	SPIKE ADDED (ug/L)	LCS CONCENTRATION (ug/L)	LCS % REC	QC LIMITS REC	#
Perfluorooctanesulfonic acid (PFOS)	0.300	0.318	106	70-130	
Perfluorooctanoic acid (PFOA)	0.146	0.159	109	70-130	E
Perfluorobutanesulfonic acid (PFBS)	0.673	0.659	98	70-130	

Column to be used to flag recovery and RPD values

FORM III
LCMS LAB CONTROL SAMPLE DUPLICATE RECOVERY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Matrix: Water Level: Low Lab File ID: 02JAN2017A6A_007.d
 Lab ID: LCSD 320-143783/3-A Client ID: _____

COMPOUND	SPIKE ADDED (ug/L)	LCSD CONCENTRATION (ug/L)	LCSD % REC	% RPD	QC LIMITS		#
					RPD	REC	
Perfluorooctanesulfonic acid (PFOS)	0.300	0.307	102	4	30	70-130	
Perfluorooctanoic acid (PFOA)	0.146	0.155	106	3	30	70-130	
Perfluorobutanesulfonic acid (PFBS)	0.673	0.622	92	6	30	70-130	

Column to be used to flag recovery and RPD values

FORM IV
LCMS METHOD BLANK SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab File ID: 02JAN2017A6A_005.d Lab Sample ID: MB 320-143783/1-A
 Matrix: Water Date Extracted: 12/23/2016 18:41
 Instrument ID: A6 Date Analyzed: 01/02/2017 12:24
 Level: (Low/Med) Low

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES:

CLIENT SAMPLE ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
	LCS 320-143783/2-A	02JAN2017A6 A 006.d	01/02/2017 12:54
	LCSD 320-143783/3-A	02JAN2017A6 A 007.d	01/02/2017 13:24
WI-AF-3RW38-1216	320-24648-1	02JAN2017A6 A 008.d	01/02/2017 13:53
WI-AF-3FB38-1216	320-24648-2	02JAN2017A6 A 009.d	01/02/2017 14:23
WI-AF-3RW39-1216	320-24648-3	02JAN2017A6 A 010.d	01/02/2017 14:52
WI-AF-3FB39-1216	320-24648-4	02JAN2017A6 A 011.d	01/02/2017 15:22
WI-AF-3RW40-1216	320-24648-5	02JAN2017A6 A 012.d	01/02/2017 15:52
WI-AF-3RW40P-1216	320-24648-6	02JAN2017A6 A 013.d	01/02/2017 16:21
WI-AF-3FB40-1216	320-24648-7	02JAN2017A6 A 014.d	01/02/2017 16:51
WI-AF-3RW41-1216	320-24648-8	02JAN2017A6 A 017.d	01/02/2017 18:20
WI-AF-3FB41-1216	320-24648-9	02JAN2017A6 A 018.d	01/02/2017 18:49

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Instrument ID: A6 Calibration Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.1 (mm) Calibration End Date: 12/24/2016 06:54
 Calibration ID: 27291

		13PFOA		PFOS			
		AREA #	RT #	AREA #	RT #	AREA #	RT #
INITIAL CALIBRATION MEAN AREA AND MEAN RT		587721	20.01	1475089	20.62		
UPPER LIMIT		881582	20.51	2212634	21.12		
LOWER LIMIT		293861	19.51	737545	20.12		
LAB SAMPLE ID	CLIENT SAMPLE ID						
CCV 320-143828/11 CCVL		652887	20.01	1599599	20.62		
ICV 320-143828/13		590506	20.01	1587807	20.62		
CCV 320-144608/2 CCVL		608470	19.99	1539571	20.61		
CCV 320-144608/3 CCVIS		577407	19.99	1447047	20.61		
MB 320-143783/1-A		520699	20.00	1544642	20.62		
LCS 320-143783/2-A		505010	20.00	1319151	20.62		
LCSD 320-143783/3-A		535790	20.00	1403740	20.62		
320-24648-1	WI-AF-3RW38-1216	483791	20.00	1598542	20.62		
320-24648-2	WI-AF-3FB38-1216	531363	20.00	1636955	20.62		
320-24648-3	WI-AF-3RW39-1216	468167	19.99	1566193	20.61		
320-24648-4	WI-AF-3FB39-1216	562524	20.00	1595430	20.62		
320-24648-5	WI-AF-3RW40-1216	467801	19.99	1555833	20.60		
320-24648-6	WI-AF-3RW40P-1216	502920	19.99	1518835	20.60		
320-24648-7	WI-AF-3FB40-1216	542670	20.00	1616522	20.61		
CCV 320-144608/15 CCVIS		558059	19.99	1349410	20.61		
CCV 320-144610/15 CCVIS		558059	19.99	1349410	20.61		
320-24648-8	WI-AF-3RW41-1216	506888	19.97	1586162	20.60		
320-24648-9	WI-AF-3FB41-1216	525654	19.99	1629374	20.61		
CCV 320-144610/27 CCVIS		533787	19.99	1460847	20.60		

13PFOA = 13C2-PFOA

PFOS = 13C4 PFOS

Area Limit = 50%-150% of internal standard area

RT Limit = $\pm$ 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Sample No.: CCV 320-144608/3 Date Analyzed: 01/02/2017 11:25
 Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
 Lab File ID (Standard): 02JAN2017A6A_003.d Heated Purge: (Y/N) N
 Calibration ID: 27291

		13PFOA		PFOS			
		AREA #	RT #	AREA #	RT #	AREA #	RT #
12/24 HOUR STD		577407	19.99	1447047	20.61		
UPPER LIMIT		808370	20.49	2025866	21.11		
LOWER LIMIT		404185	19.49	1012933	20.11		
LAB SAMPLE ID	CLIENT SAMPLE ID						
MB 320-143783/1-A		520699	20.00	1544642	20.62		
LCS 320-143783/2-A		505010	20.00	1319151	20.62		
LCSD 320-143783/3-A		535790	20.00	1403740	20.62		
320-24648-1	WI-AF-3RW38-1216	483791	20.00	1598542	20.62		
320-24648-2	WI-AF-3FB38-1216	531363	20.00	1636955	20.62		
320-24648-3	WI-AF-3RW39-1216	468167	19.99	1566193	20.61		
320-24648-4	WI-AF-3FB39-1216	562524	20.00	1595430	20.62		
320-24648-5	WI-AF-3RW40-1216	467801	19.99	1555833	20.60		
320-24648-6	WI-AF-3RW40P-1216	502920	19.99	1518835	20.60		
320-24648-7	WI-AF-3FB40-1216	542670	20.00	1616522	20.61		

13PFOA = 13C2-PFOA

PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area

RT Limit = $\pm$ 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Sample No.: CCV 320-144608/15 Date Analyzed: 01/02/2017 17:20
 Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
 Lab File ID (Standard): 02JAN2017A6A_015.d Heated Purge: (Y/N) N
 Calibration ID: 27291

	13PFOA		PFOS			
	AREA #	RT #	AREA #	RT #	AREA #	RT #
12/24 HOUR STD	558059	19.99	1349410	20.61		
UPPER LIMIT	781283	20.49	1889174	21.11		
LOWER LIMIT	390641	19.49	944587	20.11		
LAB SAMPLE ID	CLIENT SAMPLE ID					
MB 320-143783/1-A		520699	20.00	1544642	20.62	
LCS 320-143783/2-A		505010	20.00	1319151	20.62	
LCSD 320-143783/3-A		535790	20.00	1403740	20.62	
320-24648-1	WI-AF-3RW38-1216	483791	20.00	1598542	20.62	
320-24648-2	WI-AF-3FB38-1216	531363	20.00	1636955	20.62	
320-24648-3	WI-AF-3RW39-1216	468167	19.99	1566193	20.61	
320-24648-4	WI-AF-3FB39-1216	562524	20.00	1595430	20.62	
320-24648-5	WI-AF-3RW40-1216	467801	19.99	1555833	20.60	
320-24648-6	WI-AF-3RW40P-1216	502920	19.99	1518835	20.60	
320-24648-7	WI-AF-3FB40-1216	542670	20.00	1616522	20.61	

13PFOA = 13C2-PFOA
 PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
 RT Limit = $\pm$ 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Sample No.: CCV 320-144610/15 Date Analyzed: 01/02/2017 17:20
 Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
 Lab File ID (Standard): 02JAN2017A6A_015.d Heated Purge: (Y/N) N
 Calibration ID: 27291

	13PFOA		PFOS			
	AREA #	RT #	AREA #	RT #	AREA #	RT #
12/24 HOUR STD	558059	19.99	1349410	20.61		
UPPER LIMIT	781283	20.49	1889174	21.11		
LOWER LIMIT	390641	19.49	944587	20.11		
LAB SAMPLE ID	CLIENT SAMPLE ID					
320-24648-8	WI-AF-3RW41-1216		506888	19.97	1586162	20.60
320-24648-9	WI-AF-3FB41-1216		525654	19.99	1629374	20.61

13PFOA = 13C2-PFOA
 PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
 RT Limit = $\pm$ 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM VIII
LCMS INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Sample No.: CCV 320-144610/27 Date Analyzed: 01/02/2017 23:15
 Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm)
 Lab File ID (Standard): 02JAN2017A6A_027.d Heated Purge: (Y/N) N
 Calibration ID: 27291

	13PFOA		PFOS			
	AREA #	RT #	AREA #	RT #	AREA #	RT #
12/24 HOUR STD	533787	19.99	1460847	20.60		
UPPER LIMIT	747302	20.49	2045186	21.10		
LOWER LIMIT	373651	19.49	1022593	20.10		
LAB SAMPLE ID	CLIENT SAMPLE ID					
320-24648-8	WI-AF-3RW41-1216		506888	19.97	1586162	20.60
320-24648-9	WI-AF-3FB41-1216		525654	19.99	1629374	20.61

13PFOA = 13C2-PFOA
 PFOS = 13C4 PFOS

Area Limit = 70%-140% of internal standard area
 RT Limit = $\pm$ 0.5 minutes of internal standard RT

Column used to flag values outside QC limits

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW38-1216 Lab Sample ID: 320-24648-1
 Matrix: Water Lab File ID: 02JAN2017A6A_008.d
 Analysis Method: 537 Date Collected: 12/21/2016 09:18
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 263.2 (mL) Date Analyzed: 01/02/2017 13:53
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	101		70-130
STL00996	13C2 PFDA	119		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_008.d
 Lims ID: 320-24648-A-1-A
 Client ID: WI-AF-3RW38-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 13:53:43 ALS Bottle#: 15 Worklist Smp#: 8
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-1-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:41:03

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.579 17.582 -0.003 1.000 22298 0.5275 18.9
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.548 18.558 -0.010 1.000 580236 10.1 19213
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.308 19.308 0.0 1.000 15777 0.3010 402
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.332 19.332 0.0 1.000 19713 0.3335 8.8
 * 5 13C2-PFOA
 415.0 > 370.0 19.999 19.986 0.013 483791 10.0 12402
 6 Perfluorooctanoic acid M
 413.0 > 369.0 19.999 19.986 0.013 1.000 14685 0.3018 5.0 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.619 20.608 0.011 1598542 28.7 41855
 7 Perfluorooctane sulfonic acid M
 499.0 > 80.0 20.619 20.608 0.011 1.000 14938 0.2437 325 M
 9 Perfluorononanoic acid M
 463.0 > 419.0 20.691 20.679 0.012 1.000 4387 0.0789 94.2 M
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.409 21.400 0.009 1.000 581482 11.9 18540

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_008.d

Injection Date: 02-Jan-2017 13:53:43

Instrument ID: A6

Lims ID: 320-24648-A-1-A

Lab Sample ID: 320-24648-1

Client ID: WI-AF-3RW38-1216

Operator ID: CBW

ALS Bottle#: 15

Worklist Smp#: 8

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (M)

9 Perfluorononanoic acid (M)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_008.d
Lims ID: 320-24648-A-1-A
Client ID: WI-AF-3RW38-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 13:53:43 ALS Bottle#: 15 Worklist Smp#: 8
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-1-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:41:03

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.1	100.52
\$ 10 13C2 PFDA	10.0	11.9	119.21

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_008.d
Injection Date: 02-Jan-2017 13:53:43 Instrument ID: A6
Lims ID: 320-24648-A-1-A Lab Sample ID: 320-24648-1
Client ID: WI-AF-3RW38-1216
Operator ID: CBW ALS Bottle#: 15 Worklist Smp#: 8
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1


Signal: 1

Not Detected
Expected RT: 20.61

Processing Integration Results



Manual Integration Results



RT: 20.62
Area: 14938
Amount: 0.243732
Amount Units: ng/ml

Reviewer: barnettj, 03-Jan-2017 14:41:03
Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_008.d
Injection Date: 02-Jan-2017 13:53:43 Instrument ID: A6
Lims ID: 320-24648-A-1-A Lab Sample ID: 320-24648-1
Client ID: WI-AF-3RW38-1216
Operator ID: CBW ALS Bottle#: 15 Worklist Smp#: 8
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 11599
Amount: 0.238365
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 14685
Amount: 0.301784
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:41:03
Audit Action: Manually Integrated

Audit Reason: Baseline
Page 93 of 222

01/03/2017

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: <u>TestAmerica Sacramento</u>	Job No.: <u>320-24648-1</u>
SDG No.: _____	
Client Sample ID: <u>WI-AF-3FB38-1216</u>	Lab Sample ID: <u>320-24648-2</u>
Matrix: <u>Water</u>	Lab File ID: <u>02JAN2017A6A_009.d</u>
Analysis Method: <u>537</u>	Date Collected: <u>12/21/2016 09:19</u>
Extraction Method: <u>537</u>	Date Extracted: <u>12/23/2016 18:41</u>
Sample wt/vol: <u>258.6(mL)</u>	Date Analyzed: <u>01/02/2017 14:23</u>
Con. Extract Vol.: <u>1.00(mL)</u>	Dilution Factor: <u>1</u>
Injection Volume: <u>10(uL)</u>	GC Column: <u>Acquity</u> ID: <u>2.1(mm)</u>
% Moisture: _____	GPC Cleanup: (Y/N) <u>N</u>
Analysis Batch No.: <u>144608</u>	Units: <u>ug/L</u>

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.058	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.023	0.0091
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	100		70-130
STL00996	13C2 PFDA	106		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_009.d
 Lims ID: 320-24648-A-2-A
 Client ID: WI-AF-3FB38-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 14:23:20 ALS Bottle#: 16 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj Date: 03-Jan-2017 14:42:54

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.548 18.558 -0.010 1.000 636289 10.0 21012

3 Perfluorohexanesulfonic acid

M

399.0 > 80.0 19.308 19.308 0.0 1.000 1142 0.0213 10.8 M

4 Perfluoroheptanoic acid

363.0 > 319.0 19.344 19.332 0.012 1.000 11929 0.1837 6.3

* 5 13C2-PFOA

415.0 > 370.0 19.999 19.986 0.013 531363 10.0 13502

6 Perfluorooctanoic acid

M

413.0 > 369.0 19.999 19.986 0.013 1.000 19617 0.3670 7.1 M

* 8 13C4 PFOS

503.0 > 80.0 20.619 20.608 0.011 1636955 28.7 28470

7 Perfluorooctane sulfonic acid

M

499.0 > 80.0 20.619 20.608 0.011 1.000 9102 0.1450 182 M

9 Perfluorononanoic acid

M

463.0 > 419.0 20.690 20.679 0.011 1.000 575074 9.42 408 M

\$ 10 13C2 PFDA

515.0 > 470.0 21.409 21.400 0.009 1.000 567700 10.6 17967

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_009.d

Injection Date: 02-Jan-2017 14:23:20

Instrument ID: A6

Lims ID: 320-24648-A-2-A

Lab Sample ID: 320-24648-2

Client ID: WI-AF-3FB38-1216

Operator ID: CBW

ALS Bottle#: 16

Worklist Smp#: 9

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (M)



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (M)

9 Perfluorononanoic acid (M)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_009.d
Lims ID: 320-24648-A-2-A
Client ID: WI-AF-3FB38-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 14:23:20 ALS Bottle#: 16 Worklist Smp#: 9
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-2-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:42:54

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.0	100.36
\$ 10 13C2 PFDA	10.0	10.6	105.96

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_009.d
Injection Date: 02-Jan-2017 14:23:20 Instrument ID: A6
Lims ID: 320-24648-A-2-A Lab Sample ID: 320-24648-2
Client ID: WI-AF-3FB38-1216
Operator ID: CBW ALS Bottle#: 16 Worklist Smp#: 9
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1


Signal: 1

Not Detected
Expected RT: 20.61

Processing Integration Results



Manual Integration Results



RT: 20.62
Area: 9102
Amount: 0.145025
Amount Units: ng/ml

Reviewer: barnettj, 03-Jan-2017 14:42:54
Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_009.d
Injection Date: 02-Jan-2017 14:23:20 Instrument ID: A6
Lims ID: 320-24648-A-2-A Lab Sample ID: 320-24648-2
Client ID: WI-AF-3FB38-1216
Operator ID: CBW ALS Bottle#: 16 Worklist Smp#: 9
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 32798
Amount: 0.613671
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 19617
Amount: 0.367046
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:42:54
Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW39-1216 Lab Sample ID: 320-24648-3
 Matrix: Water Lab File ID: 02JAN2017A6A_010.d
 Analysis Method: 537 Date Collected: 12/21/2016 09:25
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 263(mL) Date Analyzed: 01/02/2017 14:52
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.029	0.023	0.0090
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	103		70-130
STL00996	13C2 PFDA	101		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_010.d
 Lims ID: 320-24648-A-3-A
 Client ID: WI-AF-3RW39-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 14:52:55 ALS Bottle#: 17 Worklist Smp#: 10
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:44:20

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.582 0.0 1.000 3304 0.0798 5.6
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.548 18.558 -0.010 1.000 572688 10.3 18943
 3 Perfluorohexanesulfonic acid M
 399.0 > 80.0 19.308 19.308 0.0 1.000 8174 0.1592 62.6 M
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.344 19.332 0.012 1.000 24445 0.4273 9.3
 * 5 13C2-PFOA
 415.0 > 370.0 19.986 19.986 0.0 468167 10.0 11923
 6 Perfluorooctanoic acid M
 413.0 > 369.0 19.999 19.986 0.013 1.000 35319 0.7500 11.8 M
 * 8 13C4 PFOS
 503.0 > 80.0 20.608 20.608 0.0 1566193 28.7 23536
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.608 20.608 0.0 1.000 48901 0.8144 668
 9 Perfluorononanoic acid M
 463.0 > 419.0 20.690 20.679 0.011 1.000 4915 0.0914 139 M
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.409 21.400 0.009 1.000 476996 10.1 14942

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_010.d

Injection Date: 02-Jan-2017 14:52:55

Instrument ID: A6

Lims ID: 320-24648-A-3-A

Lab Sample ID: 320-24648-3

Client ID: WI-AF-3RW39-1216

Operator ID: CBW

ALS Bottle#: 17

Worklist Smp#: 10

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (M)



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid (M)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_010.d
Lims ID: 320-24648-A-3-A
Client ID: WI-AF-3RW39-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 14:52:55 ALS Bottle#: 17 Worklist Smp#: 10
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-3-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:44:20

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.3	102.53
\$ 10 13C2 PFDA	10.0	10.1	101.05

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_010.d
Injection Date: 02-Jan-2017 14:52:55 Instrument ID: A6
Lims ID: 320-24648-A-3-A Lab Sample ID: 320-24648-3
Client ID: WI-AF-3RW39-1216
Operator ID: CBW ALS Bottle#: 17 Worklist Smp#: 10
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 47349
Amount: 1.005517
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 35319
Amount: 0.750044
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:44:20
Audit Action: Manually Integrated

Audit Reason: Split Peak
Page 104 of 222

01/03/2017

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB39-1216 Lab Sample ID: 320-24648-4
 Matrix: Water Lab File ID: 02JAN2017A6A_011.d
 Analysis Method: 537 Date Collected: 12/21/2016 09:26
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 245.6(mL) Date Analyzed: 01/02/2017 15:22
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.049	U	0.061	0.049	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.024	U	0.031	0.024	0.0096
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	101		70-130
STL00996	13C2 PFDA	107		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_011.d
 Lims ID: 320-24648-A-4-A
 Client ID: WI-AF-3FB39-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 15:22:29 ALS Bottle#: 18 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-4-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:44:48

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.549 18.558 -0.009 1.000 676109 10.1 22366

* 5 13C2-PFOA

415.0 > 370.0 19.999 19.986 0.013 562524 10.0 14305

* 8 13C4 PFOS

503.0 > 80.0 20.620 20.608 0.012 1595430 28.7 41552

9 Perfluorononanoic acid

463.0 > 419.0 20.679 20.679 0.0 1.000 857 0.0133 26.8 M

\$ 10 13C2 PFDA

515.0 > 470.0 21.409 21.400 0.009 1.000 604756 10.7 19151

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_011.d

Injection Date: 02-Jan-2017 15:22:29

Instrument ID: A6

Lims ID: 320-24648-A-4-A

Lab Sample ID: 320-24648-4

Client ID: WI-AF-3FB39-1216

Operator ID: CBW

ALS Bottle#: 18

Worklist Smp#: 11

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (ND)



4 Perfluoroheptanoic acid (ND)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (ND)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (ND)

9 Perfluorononanoic acid (M)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_011.d
Lims ID: 320-24648-A-4-A
Client ID: WI-AF-3FB39-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 15:22:29 ALS Bottle#: 18 Worklist Smp#: 11
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-4-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:44:48

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.1	100.74
\$ 10 13C2 PFDA	10.0	10.7	106.62

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW40-1216 Lab Sample ID: 320-24648-5
 Matrix: Water Lab File ID: 02JAN2017A6A_012.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:47
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 269(mL) Date Analyzed: 01/02/2017 15:52
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U M	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	123		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_012.d
 Lims ID: 320-24648-A-5-A
 Client ID: WI-AF-3RW40-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 15:52:03 ALS Bottle#: 19 Worklist Smp#: 12
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-5-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:45:29

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.539 18.558 -0.019 1.000 611627 11.0 20444

* 5 13C2-PFOA

415.0 > 370.0 19.986 19.986 0.0 467801 10.0 11756

6 Perfluorooctanoic acid

413.0 > 369.0 19.972 19.986 -0.014 1.000 781 0.0166 0.5 M

* 8 13C4 PFOS

503.0 > 80.0 20.596 20.608 -0.012 1555833 28.7 40550

\$ 10 13C2 PFDA

515.0 > 470.0 21.400 21.400 0.0 1.000 580157 12.3 18336

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_012.d

Injection Date: 02-Jan-2017 15:52:03

Instrument ID: A6

Lims ID: 320-24648-A-5-A

Lab Sample ID: 320-24648-5

Client ID: WI-AF-3RW40-1216

Operator ID: CBW

ALS Bottle#: 19

Worklist Smp#: 12

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (ND)



4 Perfluoroheptanoic acid (ND)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (ND)

9 Perfluorononanoic acid (ND)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_012.d
Lims ID: 320-24648-A-5-A
Client ID: WI-AF-3RW40-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 15:52:03 ALS Bottle#: 19 Worklist Smp#: 12
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-5-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:45:29

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.0	109.58
\$ 10 13C2 PFDA	10.0	12.3	123.00

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_012.d
Injection Date: 02-Jan-2017 15:52:03 Instrument ID: A6
Lims ID: 320-24648-A-5-A Lab Sample ID: 320-24648-5
Client ID: WI-AF-3RW40-1216
Operator ID: CBW ALS Bottle#: 19 Worklist Smp#: 12
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Not Detected
Expected RT: 19.99

Processing Integration Results



RT: 19.97
Area: 781
Amount: 0.016599
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:45:29
Audit Action: Manually Integrated

Audit Reason: Missed Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW40P-1216 Lab Sample ID: 320-24648-6
 Matrix: Water Lab File ID: 02JAN2017A6A_013.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:52
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 267.5 (mL) Date Analyzed: 01/02/2017 16:21
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	96		70-130
STL00996	13C2 PFDA	107		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_013.d
 Lims ID: 320-24648-A-6-A
 Client ID: WI-AF-3RW40P-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 16:21:37 ALS Bottle#: 20 Worklist Smp#: 13
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-6-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:45:51

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.539 18.558 -0.019 1.000 575326 9.59 19362

* 5 13C2-PFOA

415.0 > 370.0 19.986 19.986 0.0 502920 10.0 12752

* 8 13C4 PFOS

503.0 > 80.0 20.596 20.608 -0.012 1518835 28.7 39841

\$ 10 13C2 PFDA

515.0 > 470.0 21.391 21.400 -0.009 1.000 542632 10.7 17268

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_013.d

Injection Date: 02-Jan-2017 16:21:37

Instrument ID: A6

Lims ID: 320-24648-A-6-A

Lab Sample ID: 320-24648-6

Client ID: WI-AF-3RW40P-1216

Operator ID: CBW

ALS Bottle#: 20

Worklist Smp#: 13

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (ND)



4 Perfluoroheptanoic acid (ND)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (ND)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (ND)

9 Perfluorononanoic acid (ND)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_013.d
Lims ID: 320-24648-A-6-A
Client ID: WI-AF-3RW40P-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 16:21:37 ALS Bottle#: 20 Worklist Smp#: 13
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-6-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:45:51

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	9.59	95.88
\$ 10 13C2 PFDA	10.0	10.7	107.01

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB40-1216 Lab Sample ID: 320-24648-7
 Matrix: Water Lab File ID: 02JAN2017A6A_014.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:53
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 256.9(mL) Date Analyzed: 01/02/2017 16:51
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.058	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0092
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	99		70-130
STL00996	13C2 PFDA	108		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_014.d
 Lims ID: 320-24648-A-7-A
 Client ID: WI-AF-3FB40-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 16:51:14 ALS Bottle#: 21 Worklist Smp#: 14
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-7-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:46:11

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.539 18.558 -0.019 1.000 641807 9.91 21145

* 5 13C2-PFOA

415.0 > 370.0 19.999 19.986 0.013 542670 10.0 13819

* 8 13C4 PFOS

503.0 > 80.0 20.608 20.608 0.0 1616522 28.7 24302

\$ 10 13C2 PFDA

515.0 > 470.0 21.400 21.400 0.0 1.000 591519 10.8 18551

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_014.d

Injection Date: 02-Jan-2017 16:51:14

Instrument ID: A6

Lims ID: 320-24648-A-7-A

Lab Sample ID: 320-24648-7

Client ID: WI-AF-3FB40-1216

Operator ID: CBW

ALS Bottle#: 21

Worklist Smp#: 14

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (ND)



4 Perfluoroheptanoic acid (ND)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (ND)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (ND)

9 Perfluorononanoic acid (ND)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_014.d
Lims ID: 320-24648-A-7-A
Client ID: WI-AF-3FB40-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 16:51:14 ALS Bottle#: 21 Worklist Smp#: 14
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-7-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:46:11

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	9.91	99.13
\$ 10 13C2 PFDA	10.0	10.8	108.11

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW41-1216 Lab Sample ID: 320-24648-8
 Matrix: Water Lab File ID: 02JAN2017A6A_017.d
 Analysis Method: 537 Date Collected: 12/21/2016 18:40
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 265.6(mL) Date Analyzed: 01/02/2017 18:20
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144610 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	91		70-130
STL00996	13C2 PFDA	105		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_017.d
 Lims ID: 320-24648-A-8-A
 Client ID: WI-AF-3RW41-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 18:20:02 ALS Bottle#: 22 Worklist Smp#: 17
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-8-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:57:06 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:07:53

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.560	17.576	-0.016	1.000	659301	15.7	242	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.530	18.549	-0.019	1.000	552265	9.13	18587	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.285	19.308	-0.023	1.000	567518	10.9	11231	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.320	19.344	-0.024	1.000	57046	0.9210	23.4	M
* 5 13C2-PFOA								
415.0 > 370.0	19.972	19.986	-0.014		506888	10.0	12779	
6 Perfluorooctanoic acid								M
413.0 > 369.0	19.959	19.999	-0.040	1.000	57306	1.12	9.9	M
* 8 13C4 PFOS								
503.0 > 80.0	20.596	20.608	-0.012		1586162	28.7	41646	
7 Perfluorooctane sulfonic acid								M
499.0 > 80.0	20.335	20.608	-0.273	1.000	67432	1.11	481	M
\$ 10 13C2 PFDA								
515.0 > 470.0	21.391	21.409	-0.018	1.000	538811	10.5	16888	

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_017.d

Injection Date: 02-Jan-2017 18:20:02

Instrument ID: A6

Lims ID: 320-24648-A-8-A

Lab Sample ID: 320-24648-8

Client ID: WI-AF-3RW41-1216

Operator ID: CBW

ALS Bottle#: 22

Worklist Smp#: 17

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (M)

9 Perfluorononanoic acid (ND)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_017.d
Lims ID: 320-24648-A-8-A
Client ID: WI-AF-3RW41-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 18:20:02 ALS Bottle#: 22 Worklist Smp#: 17
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-8-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:57:06 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:07:53

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	9.13	91.32
\$ 10 13C2 PFDA	10.0	10.5	105.42

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_017.d
Injection Date: 02-Jan-2017 18:20:02 Instrument ID: A6
Lims ID: 320-24648-A-8-A Lab Sample ID: 320-24648-8
Client ID: WI-AF-3RW41-1216
Operator ID: CBW ALS Bottle#: 22 Worklist Smp#: 17
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1


Not Detected
Expected RT: 20.61

Processing Integration Results



RT: 20.34
Area: 67432
Amount: 1.108822
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:07:53

Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_017.d
Injection Date: 02-Jan-2017 18:20:02 Instrument ID: A6
Lims ID: 320-24648-A-8-A Lab Sample ID: 320-24648-8
Client ID: WI-AF-3RW41-1216
Operator ID: CBW ALS Bottle#: 22 Worklist Smp#: 17
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 19.96
Area: 59779
Amount: 1.172509
Amount Units: ng/ml

Processing Integration Results



RT: 19.96
Area: 57306
Amount: 1.124003
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:07:53

Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB41-1216 Lab Sample ID: 320-24648-9
 Matrix: Water Lab File ID: 02JAN2017A6A_018.d
 Analysis Method: 537 Date Collected: 12/21/2016 18:41
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 262.9(mL) Date Analyzed: 01/02/2017 18:49
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144610 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0090
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	102		70-130
STL00996	13C2 PFDA	108		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_018.d
 Lims ID: 320-24648-A-9-A
 Client ID: WI-AF-3FB41-1216
 Sample Type: Client
 Inject. Date: 02-Jan-2017 18:49:38 ALS Bottle#: 23 Worklist Smp#: 18
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: 320-24648-a-9-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:57:06 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:09:11

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.549 18.549 0.0 1.000 640511 10.2 21093

3 Perfluorohexanesulfonic acid

M

399.0 > 80.0 19.308 19.308 0.0 1.000 517 0.009676 7.7 M

* 5 13C2-PFOA

415.0 > 370.0 19.986 19.986 0.0 525654 10.0 13354

* 8 13C4 PFOS

503.0 > 80.0 20.608 20.608 0.0 1629374 28.7 42729

7 Perfluorooctane sulfonic acid

M

499.0 > 80.0 20.596 20.608 -0.012 1.000 15056 0.2410 470 M

\$ 10 13C2 PFDA

515.0 > 470.0 21.400 21.409 -0.009 1.000 570656 10.8 18092

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_018.d

Injection Date: 02-Jan-2017 18:49:38

Instrument ID: A6

Lims ID: 320-24648-A-9-A

Lab Sample ID: 320-24648-9

Client ID: WI-AF-3FB41-1216

Operator ID: CBW

ALS Bottle#: 23

Worklist Smp#: 18

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (M)



4 Perfluoroheptanoic acid (ND)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (ND)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (M)

9 Perfluorononanoic acid (ND)



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_018.d
Lims ID: 320-24648-A-9-A
Client ID: WI-AF-3FB41-1216
Sample Type: Client
Inject. Date: 02-Jan-2017 18:49:38 ALS Bottle#: 23 Worklist Smp#: 18
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: 320-24648-a-9-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:57:06 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:09:11

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	10.2	102.13
\$ 10 13C2 PFDA	10.0	10.8	107.67

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_018.d
Injection Date: 02-Jan-2017 18:49:38 Instrument ID: A6
Lims ID: 320-24648-A-9-A Lab Sample ID: 320-24648-9
Client ID: WI-AF-3FB41-1216
Operator ID: CBW ALS Bottle#: 23 Worklist Smp#: 18
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1


Signal: 1

Not Detected
Expected RT: 20.61

Processing Integration Results



Manual Integration Results



RT: 20.60
Area: 15056
Amount: 0.241008
Amount Units: ng/ml

Reviewer: barnettj, 03-Jan-2017 14:09:11
Audit Action: Manually Integrated

Audit Reason: Missed Peak
Page 132 of 222

01/03/2017

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
CURVE EVALUATION

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1 Analy Batch No.: 143828

SDG No.: _____

Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm) Heated Purge: (Y/N) N

Calibration Start Date: 12/24/2016 04:26 Calibration End Date: 12/24/2016 06:54 Calibration ID: 27291

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-143828/4	24DEC2016A6A_004.d
Level 2	STD 320-143828/5	24DEC2016A6A_005.d
Level 3	STD 320-143828/6	24DEC2016A6A_006.d
Level 4	STD 320-143828/7	24DEC2016A6A_007.d
Level 5	STD 320-143828/8	24DEC2016A6A_008.d
Level 6	STD 320-143828/9	24DEC2016A6A_009.d

ANALYTE	RRF					CURVE TYPE	COEFFICIENT			#	MIN RRF	%RSD	#	MAX %RSD	R^2 OR COD	#	MIN R^2 OR COD
	LVL 1 LVL 6	LVL 2	LVL 3	LVL 4	LVL 5		B	M1	M2								
Perfluorobutanesulfonic acid (PFBS)	0.7489 0.7091	0.7895	0.8086	0.7461	0.7483	Ave		0.7584				4.7		30.0			
Perfluorohexanesulfonic acid	0.7980 0.9951	0.9269	0.9870	0.9674	0.9683	Ave		0.9405				7.8		30.0			
Perfluoroheptanoic acid	1.0776 1.1996	1.3095	1.3597	1.2735	1.1115	Ave		1.2219				9.2		30.0			
Perfluorooctanoic acid (PFOA)	0.8779 1.1195	0.9865	1.0039	1.0531	0.9939	Ave		1.0058				8.0		30.0			
Perfluorooctanesulfonic acid (PFOS)	0.9077 1.1731	1.0347	1.1414	1.1293	1.2114	Ave		1.0996				10.1		30.0			
Perfluorononanoic acid	1.1003 1.2343	1.1975	1.1855	1.1319	1.0424	Ave		1.1487				6.2		30.0			
13C2 PFHxA	1.1206 1.3938	1.0547	1.1626	1.2662	1.1609	Ave		1.1931				10.1		30.0			
13C2 PFDA	0.9167 1.1683	0.9437	0.9678	1.0543	0.9989	Ave		1.0083				9.1		30.0			

Note: The m1 coefficient is the same as Ave RRF for an Ave curve type.

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
RESPONSE AND CONCENTRATION

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1 Analy Batch No.: 143828

SDG No.: _____

Instrument ID: A6 GC Column: Acquity ID: 2.1(mm) Heated Purge: (Y/N) N

Calibration Start Date: 12/24/2016 04:26 Calibration End Date: 12/24/2016 06:54 Calibration ID: 27291

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-143828/4	24DEC2016A6A_004.d
Level 2	STD 320-143828/5	24DEC2016A6A_005.d
Level 3	STD 320-143828/6	24DEC2016A6A_006.d
Level 4	STD 320-143828/7	24DEC2016A6A_007.d
Level 5	STD 320-143828/8	24DEC2016A6A_008.d
Level 6	STD 320-143828/9	24DEC2016A6A_009.d

ANALYTE	IS REF	CURVE TYPE	RESPONSE					CONCENTRATION (NG/ML)				
			LVL 1 LVL 6	LVL 2	LVL 3	LVL 4	LVL 5	LVL 1 LVL 6	LVL 2	LVL 3	LVL 4	LVL 5
Perfluorobutanesulfonic acid (PFBS)	PFOS	Ave	384512 5939448	999960	1880513	3444375	4710266	8.98 178	22.9	45.1	90.9	135
Perfluorohexanesulfonic acid	PFOS	Ave	138107 2809913	395737	773860	1505450	2054618	3.03 60.1	7.72	15.2	30.6	45.4
Perfluoroheptanoic acid	13PF OA	Ave	65406 1230238	205998	407699	733295	976632	0.990 19.7	2.52	4.97	10.0	14.9
Perfluorooctanoic acid (PFOA)	13PF OA	Ave	105059 2263615	305974	593502	1195609	1721874	1.95 38.8	4.98	9.81	19.8	29.3
Perfluorooctanesulfonic acid (PFOS)	PFOS	Ave	208018 4386351	584936	1184968	2327023	3403779	4.01 79.6	10.2	20.1	40.6	60.1
Perfluorononanoic acid	13PF OA	Ave	139923 2652055	394672	744764	1365537	1918933	2.07 41.2	5.29	10.4	21.0	31.1
13C2 PFHxA	13PF OA	Ave	687019 726485	657231	700737	727339	686886	10.0 10.0	10.0	10.0	10.0	10.0
13C2 PFDA	13PF OA	Ave	561986 608935	588071	583365	605628	591036	10.0 10.0	10.0	10.0	10.0	10.0

Curve Type Legend:

Ave = Average ISTD

FORM VI
LCMS BY INTERNAL STANDARD - INITIAL CALIBRATION DATA
READBACK PERCENT ERROR

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1 Analy Batch No.: 143828

SDG No.: _____


Instrument ID: A6 GC Column: Acquity ID: 2.1 (mm) Heated Purge: (Y/N) N


Calibration Start Date: 12/24/2016 04:26 Calibration End Date: 12/24/2016 06:54 Calibration ID: 27291

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	STD 320-143828/4	24DEC2016A6A_004.d
Level 2	STD 320-143828/5	24DEC2016A6A_005.d
Level 3	STD 320-143828/6	24DEC2016A6A_006.d
Level 4	STD 320-143828/7	24DEC2016A6A_007.d
Level 5	STD 320-143828/8	24DEC2016A6A_008.d
Level 6	STD 320-143828/9	24DEC2016A6A_009.d

ANALYTE	PERCENT ERROR						PERCENT ERROR LIMIT					
	LVL 1 #	LVL 2 #	LVL 3 #	LVL 4 #	LVL 5 #	LVL 6 #	LVL 1	LVL 2	LVL 3	LVL 4	LVL 5	LVL 6
Perfluorobutanesulfonic acid (PFBS)	-1.3	4.1	6.6	-1.6	-1.3	-6.5	50	50	50	50	50	50
Perfluorohexanesulfonic acid	-15.2	-1.4	5.0	2.9	3.0	5.8	50	50	50	50	50	50
Perfluoroheptanoic acid	-11.8	7.2	11.3	4.2	-9.0	-1.8	50	50	50	50	50	50
Perfluorooctanoic acid (PFOA)	-12.7	-1.9	-0.2	4.7	-1.2	11.3	50	50	50	50	50	50
Perfluorooctanesulfonic acid (PFOS)	-17.5	-5.9	3.8	2.7	10.2	6.7	50	50	50	50	50	50
Perfluorononanoic acid	-4.2	4.3	3.2	-1.5	-9.3	7.5	50	50	50	50	50	50
13C2 PFHxA	-6.1	-11.6	-2.6	6.1	-2.7	16.8	30	30	30	30	30	30
13C2 PFDA	-9.1	-6.4	-4.0	4.6	-0.9	15.9	30	30	30	30	30	30





TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_004.d
 Lims ID: STD L1
 Client ID:
 Sample Type: IC Calib Level: 1
 Inject. Date: 24-Dec-2016 04:26:10 ALS Bottle#: 1 Worklist Smp#: 4
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L1 L1
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:54:51 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 11:54:09

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.589	17.586	0.003	1.000	384512	8.86	135	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.558	18.558	0.0	1.000	687019	9.39	22481	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.320	19.318	0.002	1.000	138107	2.57	2684	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.344	19.350	-0.006	1.000	65406	0.8731	6.0	M
* 5 13C2-PFOA								
415.0 > 370.0	19.999	20.005	-0.006		613085	10.0	15656	
6 Perfluorooctanoic acid								M
413.0 > 369.0	19.999	20.004	-0.005	1.000	105059	1.70	29.0	M
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.620	20.619	0.001	1.000	208018	3.31	3500	
* 8 13C4 PFOS								
503.0 > 80.0	20.620	20.619	0.001		1640498	28.7	42687	
9 Perfluorononanoic acid								
463.0 > 419.0	20.691	20.697	-0.006	1.000	139923	1.99	3701	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.418	21.416	0.002	1.000	561986	9.09	17453	

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L1_00017

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_004.d

Injection Date: 24-Dec-2016 04:26:10

Instrument ID: A6

Lims ID: STD L1

Client ID:

Operator ID: CBW

ALS Bottle#: 1

Worklist Smp#: 4

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_004.d
Injection Date: 24-Dec-2016 04:26:10 Instrument ID: A6
Lims ID: STD L1
Client ID:
Operator ID: CBW ALS Bottle#: 1 Worklist Smp#: 4
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.34
Area: 63912
Amount: 0.784751
Amount Units: ng/ml

Processing Integration Results



RT: 19.34
Area: 65406
Amount: 0.873101
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 11:57:48

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_004.d
Injection Date: 24-Dec-2016 04:26:10 Instrument ID: A6
Lims ID: STD L1
Client ID:
Operator ID: CBW ALS Bottle#: 1 Worklist Smp#: 4
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 124416
Amount: 1.917532
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 105059
Amount: 1.703696
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 11:57:48

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_005.d
 Lims ID: STD L2
 Client ID:
 Sample Type: IC Calib Level: 2
 Inject. Date: 24-Dec-2016 04:55:44 ALS Bottle#: 2 Worklist Smp#: 5
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L2 L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:54:56 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 11:59:47

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.586 17.586 0.0 1.000 999960 23.8 332
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.558 18.558 0.0 1.000 657231 8.84 21447
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.309 19.318 -0.009 1.000 395737 7.60 9507
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.344 19.350 -0.006 1.000 205998 2.71 17.1
 * 5 13C2-PFOA
 415.0 > 370.0 19.999 20.005 -0.006 623136 10.0 15732
 6 Perfluorooctanoic acid M
 413.0 > 369.0 19.999 20.005 -0.006 1.000 305974 4.88 76.0 M
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.620 20.619 0.001 1.000 584936 9.61 9904
 * 8 13C4 PFOS
 503.0 > 80.0 20.620 20.619 0.001 1586953 28.7 41115
 9 Perfluorononanoic acid
 463.0 > 419.0 20.691 20.697 -0.006 1.000 394672 5.51 8417
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.409 21.416 -0.007 1.000 588071 9.36 18601

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L2_00016

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_005.d

Injection Date: 24-Dec-2016 04:55:44

Instrument ID: A6

Lims ID: STD L2

Client ID:

Operator ID: CBW

ALS Bottle#: 2

Worklist Smp#: 5

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_005.d
Injection Date: 24-Dec-2016 04:55:44 Instrument ID: A6
Lims ID: STD L2
Client ID:
Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 5
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 340214
Amount: 5.293744
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 305974
Amount: 4.881813
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 11:59:47

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_006.d
 Lims ID: STD L3
 Client ID:
 Sample Type: IC Calib Level: 3
 Inject. Date: 24-Dec-2016 05:25:20 ALS Bottle#: 3 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L3 L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:55:03 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 12:04:18

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.586 17.586 0.0 1.000 1880513 48.1 564
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.558 18.558 0.0 1.000 700737 9.74 22761
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.320 19.318 0.002 1.000 773860 16.0 18225
 4 Perfluoroheptanoic acid M
 363.0 > 319.0 19.344 19.350 -0.006 1.000 407699 5.54 39.5 M
 * 5 13C2-PFOA
 415.0 > 370.0 19.999 20.005 -0.006 602755 10.0 15228
 6 Perfluorooctanoic acid M
 413.0 > 369.0 19.999 20.005 -0.006 1.000 593502 9.79 208 M
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 1184968 20.9 19371
 * 8 13C4 PFOS
 503.0 > 80.0 20.619 20.619 0.0 1478857 28.7 38359
 9 Perfluorononanoic acid
 463.0 > 419.0 20.691 20.697 -0.006 1.000 744764 10.8 19653
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.418 21.416 0.002 1.000 583365 9.60 18358

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L3_00018

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_006.d

Injection Date: 24-Dec-2016 05:25:20

Instrument ID: A6

Lims ID: STD L3

Client ID:

Operator ID: CBW

ALS Bottle#: 3

Worklist Smp#: 6

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_006.d
Injection Date: 24-Dec-2016 05:25:20 Instrument ID: A6
Lims ID: STD L3
Client ID:
Operator ID: CBW ALS Bottle#: 3 Worklist Smp#: 6
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.34
Area: 491634
Amount: 6.121135
Amount Units: ng/ml

Processing Integration Results



RT: 19.34
Area: 407699
Amount: 5.535622
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:04:18

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_006.d
Injection Date: 24-Dec-2016 05:25:20 Instrument ID: A6
Lims ID: STD L3
Client ID:
Operator ID: CBW ALS Bottle#: 3 Worklist Smp#: 6
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.00
Area: 618796
Amount: 10.134867
Amount Units: ng/ml

Processing Integration Results



RT: 20.00
Area: 593502
Amount: 9.789506
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:09:12

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_007.d
 Lims ID: STD L4
 Client ID:
 Sample Type: ICISAV Calib Level: 4
 Inject. Date: 24-Dec-2016 05:54:58 ALS Bottle#: 4 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L4 L4
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:55:08 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 11:52:02

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.586	17.586	0.0	1.000	3444375	89.4	760	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.558	18.558	0.0	1.000	727339	10.6	23818	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.320	19.318	0.002	1.000	1505450	31.5	34932	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.356	19.350	0.006	1.000	733295	10.4	79.4	M
* 5 13C2-PFOA								
415.0 > 370.0	20.011	20.005	0.006		574449	10.0	14462	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.011	20.005	0.006	1.000	1195609	20.7	546	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.619	0.0	1.000	2327023	41.7	37125	
* 8 13C4 PFOS								
503.0 > 80.0	20.619	20.619	0.0		1456779	28.7	37378	
9 Perfluorononanoic acid								
463.0 > 419.0	20.702	20.697	0.005	1.000	1365537	20.7	20501	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.418	21.416	0.002	1.000	605628	10.5	19163	

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L4_00017

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_007.d

Injection Date: 24-Dec-2016 05:54:58

Instrument ID: A6

Lims ID: STD L4

Client ID:

Operator ID: CBW

ALS Bottle#: 4

Worklist Smp#: 7

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_007.d

Injection Date: 24-Dec-2016 05:54:58

Instrument ID: A6

Lims ID: STD L4

Client ID:

Operator ID: CBW

ALS Bottle#:

4

Worklist Smp#: 7

Injection Vol: 10.0 ul

Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

Column: Acquity BEH C18 (2.10 mm)


Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.36
Area: 824224
Amount: 11.158414
Amount Units: ng/ml

Processing Integration Results



RT: 19.36
Area: 733295
Amount: 10.447079
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:07:04

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_008.d
 Lims ID: STD L5
 Client ID:
 Sample Type: IC Calib Level: 5
 Inject. Date: 24-Dec-2016 06:24:33 ALS Bottle#: 5 Worklist Smp#: 8
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L5 L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:55:14 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 12:10:47

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.586 -0.004 1.000 4710266 132.8 8284
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.558 18.558 0.0 1.000 686886 9.73 22056
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.320 19.318 0.002 1.000 2054618 46.7 46922
 4 Perfluoroheptanoic acid M
 363.0 > 319.0 19.356 19.350 0.006 1.000 976632 13.5 82.2 M
 * 5 13C2-PFOA
 415.0 > 370.0 20.011 20.005 0.006 591689 10.0 15022
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.011 20.005 0.006 1.000 1721874 28.9 1197
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 3403779 66.2 19341
 * 8 13C4 PFOS
 503.0 > 80.0 20.619 20.619 0.0 1340813 28.7 22935
 9 Perfluorononanoic acid
 463.0 > 419.0 20.702 20.697 0.005 1.000 1918933 28.2 14288
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.418 21.416 0.002 1.000 591036 9.91 18532

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L5_00019

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_008.d

Injection Date: 24-Dec-2016 06:24:33

Instrument ID: A6

Lims ID: STD L5

Client ID:

Operator ID: CBW

ALS Bottle#: 5

Worklist Smp#: 8

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537_A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_008.d
Injection Date: 24-Dec-2016 06:24:33 Instrument ID: A6
Lims ID: STD L5
Client ID:
Operator ID: CBW ALS Bottle#: 5 Worklist Smp#: 8
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.36
Area: 1090608
Amount: 14.634076
Amount Units: ng/ml

Processing Integration Results



RT: 19.36
Area: 976632
Amount: 13.508437
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:10:47

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Lims ID: STD L6
 Client ID:
 Sample Type: IC Calib Level: 6
 Inject. Date: 24-Dec-2016 06:54:10 ALS Bottle#: 6 Worklist Smp#: 9
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: STD L6 L6
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:55:20 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 12:11:59

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.586 17.586 0.0 1.000 5939448 166.8 2454
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.567 18.558 0.009 1.000 726485 11.7 23627
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.320 19.318 0.002 1.000 2809913 63.6 63277
 4 Perfluoroheptanoic acid M
 363.0 > 319.0 19.356 19.350 0.006 1.000 1230238 19.3 129 M
 * 5 13C2-PFOA
 415.0 > 370.0 20.011 20.005 0.006 521213 10.0 13076
 6 Perfluorooctanoic acid
 413.0 > 369.0 20.011 20.005 0.006 1.000 2263615 43.2 940
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.619 20.619 0.0 1.000 4386351 85.0 67299
 * 8 13C4 PFOS
 503.0 > 80.0 20.619 20.619 0.0 1346636 28.7 34338
 9 Perfluorononanoic acid
 463.0 > 419.0 20.702 20.697 0.005 1.000 2652055 44.3 14613
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.418 21.416 0.002 1.000 608935 11.6 19393

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L6_00016

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d

Injection Date: 24-Dec-2016 06:54:10

Instrument ID: A6

Lims ID: STD L6

Client ID:

Operator ID: CBW

ALS Bottle#: 6

Worklist Smp#: 9

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Injection Date: 24-Dec-2016 06:54:10 Instrument ID: A6
Lims ID: STD L6
Client ID:
Operator ID: CBW ALS Bottle#: 6 Worklist Smp#: 9
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.36
Area: 1328838
Amount: 20.595205
Amount Units: ng/ml

Processing Integration Results



RT: 19.36
Area: 1230238
Amount: 19.317083
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:11:59

Audit Action: Manually Integrated

Audit Reason: Baseline

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-143828/11 Calibration Date: 12/24/2016 07:53
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 24DEC2016A6A_011.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.8007		24.2	22.9	5.6	50.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.9098		7.46	7.72	-3.3	50.0
Perfluoroheptanoic acid	Ave	1.222	1.308		2.70	2.52	7.0	50.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.037		5.13	4.98	3.1	50.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.106		10.3	10.2	0.5	50.0
Perfluorononanoic acid	Ave	1.149	1.267		5.83	5.29	10.3	50.0
13C2 PFHxA	Ave	1.193	1.160		9.72	10.0	-2.8	30.0
13C2 PFDA	Ave	1.008	0.9498		9.42	10.0	-5.8	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_011.d
 Lims ID: CCV L2
 Client ID:
 Sample Type: CCVL
 Inject. Date: 24-Dec-2016 07:53:20 ALS Bottle#: 2 Worklist Smp#: 11
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L2 CCV L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:56:28 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 12:19:01

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.586	17.586	0.0	1.000	1022169	24.2	348	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.558	18.559	-0.001	1.000	757155	9.72	24717	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.320	19.318	0.002	1.000	391515	7.46	9073	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.356	19.350	0.006	1.000	215552	2.70	18.7	M
* 5 13C2-PFOA								
415.0 > 370.0	20.011	20.005	0.006		652887	10.0	16566	
6 Perfluorooctanoic acid								M
413.0 > 369.0	20.011	20.005	0.006	1.000	336982	5.13	120	M
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.620	20.619	0.001	1.000	629983	10.3	10166	
* 8 13C4 PFOS								
503.0 > 80.0	20.620	20.619	0.001		1599599	28.7	41464	
9 Perfluorononanoic acid								
463.0 > 419.0	20.691	20.697	-0.006	1.000	437409	5.83	9269	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.418	21.416	0.002	1.000	620121	9.42	19537	

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L2_00016

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_011.d

Injection Date: 24-Dec-2016 07:53:20

Instrument ID: A6

Lims ID: CCV L2

Client ID:

Operator ID: CBW

ALS Bottle#: 2

Worklist Smp#: 11

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_011.d
Injection Date: 24-Dec-2016 07:53:20 Instrument ID: A6
Lims ID: CCV L2
Client ID:
Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 11
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.36
Area: 218316
Amount: 2.736624
Amount Units: ng/ml

Processing Integration Results



RT: 19.36
Area: 215552
Amount: 2.701977
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:19:01

Audit Action: Manually Integrated

Audit Reason: Baseline

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_011.d
Injection Date: 24-Dec-2016 07:53:20 Instrument ID: A6
Lims ID: CCV L2
Client ID:
Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 11
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


RT: 20.01
Area: 341171
Amount: 5.195334
Amount Units: ng/ml

Processing Integration Results



RT: 20.01
Area: 336982
Amount: 5.131545
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:19:01

Audit Action: Manually Integrated

Audit Reason: Baseline

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: ICV 320-143828/13 Calibration Date: 12/24/2016 08:52
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 24DEC2016A6A_013.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.5696		86.2	115	-24.9	30.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.6887		19.4	26.5	-26.8	30.0
Perfluoroheptanoic acid	Ave	1.222	1.215		12.5	12.6	-0.5	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.023		25.5	25.0	1.7	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	0.8190		20.3	27.2	-25.5	30.0
Perfluorononanoic acid	Ave	1.149	1.056		23.0	25.0	-8.1	30.0
13C2 PFHxA	Ave	1.193	1.179		9.88	10.0	-1.2	30.0
13C2 PFDA	Ave	1.008	0.9390		9.31	10.0	-6.9	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_013.d
 Lims ID: ICV
 Client ID:
 Sample Type: ICV
 Inject. Date: 24-Dec-2016 08:52:30 ALS Bottle#: 7 Worklist Smp#: 13
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: ICV ICV
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist:
 Method: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 27-Dec-2016 10:56:38 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK031

First Level Reviewer: phomsophat

Date: 24-Dec-2016 12:21:33

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.589	17.586	0.003	1.000	3619397	86.2	870	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.567	18.559	0.008	1.000	695964	9.88	22590	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.320	19.318	0.002	1.000	1009629	19.4	23583	
4 Perfluoroheptanoic acid								M
363.0 > 319.0	19.356	19.350	0.006	1.000	904351	12.5	87.5	M
* 5 13C2-PFOA								
415.0 > 370.0	20.011	20.005	0.006		590506	10.0	14954	
6 Perfluorooctanoic acid								
413.0 > 369.0	20.011	20.005	0.006	1.000	1511917	25.5	1505	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.631	20.619	0.012	1.000	1235004	20.3	19443	
* 8 13C4 PFOS								
503.0 > 80.0	20.620	20.619	0.001		1587807	28.7	41047	
9 Perfluorononanoic acid								
463.0 > 419.0	20.703	20.697	0.006	1.000	1559739	23.0	32800	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.418	21.416	0.002	1.000	554498	9.31	17342	

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-ICV_00019

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_013.d

Injection Date: 24-Dec-2016 08:52:30

Instrument ID: A6

Lims ID: ICV

Client ID:

Operator ID: CBW

ALS Bottle#: 7

Worklist Smp#: 13

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid




7 Perfluorooctane sulfonic acid

* 8 13C4 PFOS

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_013.d
Injection Date: 24-Dec-2016 08:52:30 Instrument ID: A6
Lims ID: ICV
Client ID:
Operator ID: CBW ALS Bottle#: 7 Worklist Smp#: 13
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.36
Area: 1014586
Amount: 14.061517
Amount Units: ng/ml

Processing Integration Results



RT: 19.36
Area: 904351
Amount: 12.533730
Amount Units: ng/ml

Manual Integration Results



Reviewer: phomsophat, 24-Dec-2016 12:21:33

Audit Action: Manually Integrated

Audit Reason: Baseline

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-144608/2 Calibration Date: 01/02/2017 10:56
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 02JAN2017A6A_002.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.7947		24.0	22.9	4.8	50.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.8573		7.03	7.72	-8.8	50.0
Perfluoroheptanoic acid	Ave	1.222	1.373		2.84	2.52	12.4	50.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.195		5.92	4.98	18.9	50.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.089		10.1	10.2	-1.0	50.0
Perfluorononanoic acid	Ave	1.149	1.261		5.81	5.29	9.8	50.0
13C2 PFHxA	Ave	1.193	1.143		9.58	10.0	-4.2	30.0
13C2 PFDA	Ave	1.008	1.003		9.94	10.0	-0.6	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_002.d
 Lims ID: CCV L2
 Client ID:
 Sample Type: CCVL
 Inject. Date: 02-Jan-2017 10:56:06 ALS Bottle#: 2 Worklist Smp#: 2
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L2 CCV L2
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:49 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: phomsophat

Date: 02-Jan-2017 12:26:04

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.582 0.0 1.000 976470 24.0 516
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.558 18.558 0.0 1.000 695267 9.58 22841
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.309 19.309 0.0 1.000 355087 7.03 8199
 4 Perfluoroheptanoic acid M
 363.0 > 319.0 19.344 19.344 0.0 1.000 210901 2.84 59.4 M
 * 5 13C2-PFOA
 415.0 > 370.0 19.986 19.986 0.0 608470 10.0 15353
 6 Perfluorooctanoic acid
 413.0 > 369.0 19.986 19.986 0.0 1.000 362039 5.92 219
 * 8 13C4 PFOS
 503.0 > 80.0 20.608 20.608 0.0 1539571 28.7 40162
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.608 20.608 0.0 1.000 597238 10.1 10223
 9 Perfluorononanoic acid
 463.0 > 419.0 20.679 20.679 0.0 1.000 405842 5.81 3619
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.400 21.400 0.0 1.000 610113 9.94 19095

QC Flag Legend

Review Flags

M - Manually Integrated

Reagents:

LC537-L2_00016

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_002.d

Injection Date: 02-Jan-2017 10:56:06

Instrument ID: A6

Lims ID: CCV L2

Client ID:

Operator ID: CBW

ALS Bottle#: 2

Worklist Smp#: 2

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_002.d
Injection Date: 02-Jan-2017 10:56:06 Instrument ID: A6
Lims ID: CCV L2
Client ID:
Operator ID: CBW ALS Bottle#: 2 Worklist Smp#: 2
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F3:MRM

4 Perfluoroheptanoic acid, CAS: 375-85-9

Signal: 1


RT: 19.34
Area: 247027
Amount: 3.322560
Amount Units: ng/ml

Processing Integration Results



RT: 19.34
Area: 210901
Amount: 2.836659
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 13:55:22

Audit Action: Manually Integrated

Audit Reason: Split Peak

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-144608/3 Calibration Date: 01/02/2017 11:25
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 02JAN2017A6A_003.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.8387		49.9	45.1	10.6	30.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.9512		15.4	15.2	1.1	30.0
Perfluoroheptanoic acid	Ave	1.222	1.392		5.67	4.97	13.9	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.143		11.1	9.81	13.6	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.095		20.1	20.1	-0.4	30.0
Perfluorononanoic acid	Ave	1.149	1.285		11.7	10.4	11.9	30.0
13C2 PFHxA	Ave	1.193	1.175		9.85	10.0	-1.5	30.0
13C2 PFDA	Ave	1.008	1.032		10.2	10.0	2.3	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_003.d
 Lims ID: CCV L3
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 02-Jan-2017 11:25:42 ALS Bottle#: 3 Worklist Smp#: 3
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: phomsophat

Date: 02-Jan-2017 12:30:11

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid
 299.0 > 80.0 17.582 17.582 0.0 1.000 1908634 49.9 793
 \$ 2 13C2 PFHxA
 315.0 > 270.0 18.558 18.558 0.0 1.000 678392 9.85 22291
 3 Perfluorohexanesulfonic acid
 399.0 > 80.0 19.308 19.308 0.0 1.000 729736 15.4 17055
 4 Perfluoroheptanoic acid
 363.0 > 319.0 19.332 19.332 0.0 1.000 399870 5.67 3483
 * 5 13C2-PFOA
 415.0 > 370.0 19.986 19.986 0.0 577407 10.0 14449
 6 Perfluorooctanoic acid
 413.0 > 369.0 19.986 19.986 0.0 1.000 647223 11.1 284
 * 8 13C4 PFOS
 503.0 > 80.0 20.608 20.608 0.0 1447047 28.7 29897
 7 Perfluorooctane sulfonic acid
 499.0 > 80.0 20.608 20.608 0.0 1.000 1112547 20.1 11973
 9 Perfluorononanoic acid
 463.0 > 419.0 20.679 20.679 0.0 1.000 773312 11.7 20531
 \$ 10 13C2 PFDA
 515.0 > 470.0 21.400 21.400 0.0 1.000 595615 10.2 18896

Reagents:

LC537-L3_00018

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_003.d

Injection Date: 02-Jan-2017 11:25:42

Instrument ID: A6

Lims ID: CCV L3

Client ID:

Operator ID: CBW

ALS Bottle#: 3

Worklist Smp#: 3

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-144608/15 Calibration Date: 01/02/2017 17:20
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 02JAN2017A6A_015.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.7470		133	135	-1.5	30.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.9878		47.7	45.4	5.0	30.0
Perfluoroheptanoic acid	Ave	1.222	1.179		14.3	14.9	-3.5	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.069		31.1	29.3	6.3	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.196		65.4	60.1	8.8	30.0
Perfluorononanoic acid	Ave	1.149	1.207		32.7	31.1	5.0	30.0
13C2 PFHxA	Ave	1.193	1.111		9.31	10.0	-6.9	30.0
13C2 PFDA	Ave	1.008	1.012		10.0	10.0	0.3	30.0

FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-144610/15 Calibration Date: 01/02/2017 17:20
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 02JAN2017A6A_015.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.7470		133	135	-1.5	30.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.9878		47.7	45.4	5.0	30.0
Perfluoroheptanoic acid	Ave	1.222	1.179		14.3	14.9	-3.5	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.069		31.1	29.3	6.3	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.196		65.4	60.1	8.8	30.0
Perfluorononanoic acid	Ave	1.149	1.207		32.7	31.1	5.0	30.0
13C2 PFHxA	Ave	1.193	1.111		9.31	10.0	-6.9	30.0
13C2 PFDA	Ave	1.008	1.012		10.0	10.0	0.3	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_015.d
 Lims ID: CCV L5
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 02-Jan-2017 17:20:51 ALS Bottle#: 5 Worklist Smp#: 15
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:47:20 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.576	17.576	0.0	1.000	4731838	132.6	987	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.549	18.549	0.0	1.000	619886	9.31	20497	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.308	19.308	0.0	1.000	2109420	47.7	48601	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.344	19.344	0.0	1.000	977037	14.3	3588	
* 5 13C2-PFOA								
415.0 > 370.0	19.986	19.986	0.0		558059	10.0	14034	
6 Perfluorooctanoic acid								
413.0 > 369.0	19.999	19.999	0.0	1.000	1746516	31.1	915	
* 8 13C4 PFOS								
503.0 > 80.0	20.608	20.608	0.0		1349410	28.7	34991	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.608	20.608	0.0	1.000	3383090	65.4	11050	
9 Perfluorononanoic acid								
463.0 > 419.0	20.679	20.679	0.0	1.000	2094866	32.7	15710	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.409	21.409	0.0	1.000	564527	10.0	17711	

Reagents:

LC537-L5_00019

Amount Added: 1.00

Units: mL

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_015.d
 Lims ID: CCV L5
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 02-Jan-2017 17:20:51 ALS Bottle#: 5 Worklist Smp#: 15
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L5
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:47:20 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.576	17.576	0.0	1.000	4731838	132.6	987	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.549	18.549	0.0	1.000	619886	9.31	20497	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.308	19.308	0.0	1.000	2109420	47.7	48601	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.344	19.344	0.0	1.000	977037	14.3	3588	
* 5 13C2-PFOA								
415.0 > 370.0	19.986	19.986	0.0		558059	10.0	14034	
6 Perfluorooctanoic acid								
413.0 > 369.0	19.999	19.999	0.0	1.000	1746516	31.1	915	
* 8 13C4 PFOS								
503.0 > 80.0	20.608	20.608	0.0		1349410	28.7	34991	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.608	20.608	0.0	1.000	3383090	65.4	11050	
9 Perfluorononanoic acid								
463.0 > 419.0	20.679	20.679	0.0	1.000	2094866	32.7	15710	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.409	21.409	0.0	1.000	564527	10.0	17711	

Reagents:

LC537-L5_00019

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_015.d

Injection Date: 02-Jan-2017 17:20:51

Instrument ID: A6

Lims ID: CCV L5

Client ID:

Operator ID: CBW

ALS Bottle#: 5

Worklist Smp#: 15

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_015.d

Injection Date: 02-Jan-2017 17:20:51

Instrument ID: A6

Lims ID: CCV L5

Client ID:

Operator ID: CBW

ALS Bottle#: 5

Worklist Smp#: 15

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



FORM VII
LCMS CONTINUING CALIBRATION DATA

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Lab Sample ID: CCV 320-144610/27 Calibration Date: 01/02/2017 23:15
 Instrument ID: A6 Calib Start Date: 12/24/2016 04:26
 GC Column: Acquity ID: 2.10 (mm) Calib End Date: 12/24/2016 06:54
 Lab File ID: 02JAN2017A6A_027.d Conc. Units: ng/mL

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Perfluorobutanesulfonic acid (PFBS)	Ave	0.7584	0.8100		48.2	45.1	6.8	30.0
Perfluorohexanesulfonic acid	Ave	0.9405	0.9291		15.0	15.2	-1.2	30.0
Perfluoroheptanoic acid	Ave	1.222	1.395		5.68	4.97	14.2	30.0
Perfluorooctanoic acid (PFOA)	Ave	1.006	1.109		10.8	9.81	10.3	30.0
Perfluorooctanesulfonic acid (PFOS)	Ave	1.100	1.103		20.2	20.1	0.3	30.0
Perfluorononanoic acid	Ave	1.149	1.312		11.9	10.4	14.2	30.0
13C2 PFHxA	Ave	1.193	1.195		10.0	10.0	0.2	30.0
13C2 PFDA	Ave	1.008	1.036		10.3	10.0	2.7	30.0

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_027.d
 Lims ID: CCV L3
 Client ID:
 Sample Type: CCVIS
 Inject. Date: 02-Jan-2017 23:15:58 ALS Bottle#: 3 Worklist Smp#: 27
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: CCV L3
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Sublist: chrom-537__A6*sub3
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:57:22 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.579	17.579	0.0	1.000	1861018	48.2	663	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.558	18.558	0.0	1.000	637870	10.0	21098	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.308	19.308	0.0	1.000	719583	15.0	16805	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.344	19.344	0.0	1.000	370496	5.68	1217	
* 5 13C2-PFOA								
415.0 > 370.0	19.986	19.986	0.0		533787	10.0	13371	
6 Perfluorooctanoic acid								
413.0 > 369.0	19.986	19.986	0.0	1.000	580860	10.8	355	
* 8 13C4 PFOS								
503.0 > 80.0	20.596	20.596	0.0		1460847	28.7	38080	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.596	20.596	0.0	1.000	1131009	20.2	18651	
9 Perfluorononanoic acid								
463.0 > 419.0	20.679	20.679	0.0	1.000	729833	11.9	8515	
\$ 10 13C2 PFDA								
515.0 > 470.0	21.391	21.391	0.0	1.000	552909	10.3	17606	

Reagents:

LC537-L3_00018

Amount Added: 1.00

Units: mL

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_027.d

Injection Date: 02-Jan-2017 23:15:58

Instrument ID: A6

Lims ID: CCV L3

Client ID:

Operator ID: CBW

ALS Bottle#: 3

Worklist Smp#: 27

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: MB 320-143783/1-A
 Matrix: Water Lab File ID: 02JAN2017A6A_005.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 250.00 (mL) Date Analyzed: 01/02/2017 12:24
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.048	U M	0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.024	U M	0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	111		70-130
STL00996	13C2 PFDA	116		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_005.d
 Lims ID: MB 320-143783/1-A
 Client ID:
 Sample Type: MB
 Inject. Date: 02-Jan-2017 12:24:53 ALS Bottle#: 12 Worklist Smp#: 5
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: mb 320-143783/1-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:38:32

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

\$ 2 13C2 PFHxA

315.0 > 270.0 18.548 18.558 -0.010 1.000 690592 11.1 22759

3 Perfluorohexanesulfonic acid M

399.0 > 80.0 19.344 19.308 0.036 1.000 172 0.003396 1.8 M

4 Perfluoroheptanoic acid M

363.0 > 319.0 19.380 19.332 0.048 1.000 816 0.0128 0.3 M

* 5 13C2-PFOA

415.0 > 370.0 19.999 19.986 0.013 520699 10.0 13323

6 Perfluorooctanoic acid M

413.0 > 369.0 19.972 19.986 -0.014 1.000 1014 0.0194 0.3 M

* 8 13C4 PFOS

503.0 > 80.0 20.619 20.608 0.011 1544642 28.7 27020

7 Perfluorooctane sulfonic acid M

499.0 > 80.0 20.631 20.608 0.023 1.000 1420 0.0240 39.5 M

9 Perfluorononanoic acid

463.0 > 419.0 20.691 20.679 0.012 1.000 189706 3.17 3343

\$ 10 13C2 PFDA

515.0 > 470.0 21.418 21.400 0.018 1.000 610144 11.6 19569

QC Flag Legend

Review Flags

M - Manually Integrated

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_005.d

Injection Date: 02-Jan-2017 12:24:53

Instrument ID: A6

Lims ID: MB 320-143783/1-A

Client ID:

Operator ID: CBW

ALS Bottle#: 12

Worklist Smp#: 5

Injection Vol: 10.0 ul




Dil. Factor: 1.0000

Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid (ND) \$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid (M)



4 Perfluoroheptanoic acid (M)

* 5 13C2-PFOA




6 Perfluorooctanoic acid (M)




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid (M)

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_005.d
Lims ID: MB 320-143783/1-A
Client ID:
Sample Type: MB
Inject. Date: 02-Jan-2017 12:24:53 ALS Bottle#: 12 Worklist Smp#: 5
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: mb 320-143783/1-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:38:32

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.1	111.16
\$ 10 13C2 PFDA	10.0	11.6	116.22

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_005.d
Injection Date: 02-Jan-2017 12:24:53 Instrument ID: A6
Lims ID: MB 320-143783/1-A
Client ID:
Operator ID: CBW ALS Bottle#: 12 Worklist Smp#: 5
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F5:MRM

7 Perfluorooctane sulfonic acid, CAS: 1763-23-1

Signal: 1


Not Detected
Expected RT: 20.61

Processing Integration Results



RT: 20.63
Area: 1420
Amount: 0.023978
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:38:32
Audit Action: Manually Integrated

Audit Reason: Missed Peak

TestAmerica Sacramento


Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_005.d
Injection Date: 02-Jan-2017 12:24:53 Instrument ID: A6
Lims ID: MB 320-143783/1-A
Client ID:
Operator ID: CBW ALS Bottle#: 12 Worklist Smp#: 5
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Method: 537__A6 Limit Group: LC 537 ICAL
Column: Acquity BEH C18 (2.10 mm) Detector F4:MRM

6 Perfluorooctanoic acid, CAS: 335-67-1

Signal: 1


Not Detected
Expected RT: 19.99

Processing Integration Results



RT: 19.97
Area: 1014
Amount: 0.019361
Amount Units: ng/ml

Manual Integration Results



Reviewer: barnettj, 03-Jan-2017 14:38:32
Audit Action: Manually Integrated

Audit Reason: Missed Peak
Page 197 of 222

01/03/2017

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: LCS 320-143783/2-A
 Matrix: Water Lab File ID: 02JAN2017A6A_006.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 250.00 (mL) Date Analyzed: 01/02/2017 12:54
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.318		0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.159	E	0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.659		0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	121		70-130
STL00996	13C2 PFDA	121		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_006.d
 Lims ID: LCS 320-143783/2-A
 Client ID:
 Sample Type: LCS
 Inject. Date: 02-Jan-2017 12:54:30 ALS Bottle#: 13 Worklist Smp#: 6
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcs 320-143783/2-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.589	17.582	0.007	1.000	5746726	164.7	8647	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.558	18.558	0.0	1.000	728692	12.1	23714	
3 Perfluorohexanesulfonic acid								E
399.0 > 80.0	19.308	19.308	0.0	1.000	2602692	60.2	29481	E
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.344	19.332	0.012	1.000	1148717	18.6	29407	
* 5 13C2-PFOA								
415.0 > 370.0	19.999	19.986	0.013		505010	10.0	12794	
6 Perfluorooctanoic acid								E
413.0 > 369.0	19.999	19.986	0.013	1.000	2021855	39.8	1417	E
* 8 13C4 PFOS								
503.0 > 80.0	20.619	20.608	0.011		1319151	28.7	34020	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.608	0.011	1.000	4021607	79.5	14721	
9 Perfluorononanoic acid								E
463.0 > 419.0	20.691	20.679	0.012	1.000	2553302	44.0	38313	E
\$ 10 13C2 PFDA								
515.0 > 470.0	21.409	21.400	0.009	1.000	617770	12.1	19658	

QC Flag Legend

Processing Flags

E - Exceeded Maximum Amount

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_006.d

Injection Date: 02-Jan-2017 12:54:30

Instrument ID: A6

Lims ID: LCS 320-143783/2-A

Client ID:

Operator ID: CBW

ALS Bottle#: 13

Worklist Smp#: 6

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_006.d
Lims ID: LCS 320-143783/2-A
Client ID:
Sample Type: LCS
Inject. Date: 02-Jan-2017 12:54:30 ALS Bottle#: 13 Worklist Smp#: 6
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: lcs 320-143783/2-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	12.1	120.94
\$ 10 13C2 PFDA	10.0	12.1	121.32

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: _____ Lab Sample ID: LCSD 320-143783/3-A
 Matrix: Water Lab File ID: 02JAN2017A6A_007.d
 Analysis Method: 537 Date Collected: _____
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 250.00 (mL) Date Analyzed: 01/02/2017 13:24
 Con. Extract Vol.: 1.00 (mL) Dilution Factor: 1
 Injection Volume: 10 (uL) GC Column: Acquity ID: 2.1 (mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.307		0.060	0.048	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.155		0.030	0.024	0.0094
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.622		0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	111		70-130
STL00996	13C2 PFDA	116		70-130

TestAmerica Sacramento
Target Compound Quantitation Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_007.d
 Lims ID: LCSD 320-143783/3-A
 Client ID:
 Sample Type: LCSD
 Inject. Date: 02-Jan-2017 13:24:07 ALS Bottle#: 14 Worklist Smp#: 7
 Injection Vol: 10.0 ul Dil. Factor: 1.0000
 Sample Info: lcsd 320-143783/3-a
 Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
 Operator ID: CBW Instrument ID: A6
 Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
 Limit Group: LC 537 ICAL
 Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
 Integrator: Picker
 Quant Method: Internal Standard Quant By: Initial Calibration
 Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
 Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
 Process Host: XAWRK018

First Level Reviewer: barnettj Date: 03-Jan-2017 14:39:50

Signal	RT	EXP RT	DLT RT	REL RT	Response	Amount ng/ml	S/N	Flags
--------	----	-----------	-----------	-----------	----------	-----------------	-----	-------

1 Perfluorobutanesulfonic acid								
299.0 > 80.0	17.582	17.582	0.0	1.000	5772962	155.5	1410	
\$ 2 13C2 PFHxA								
315.0 > 270.0	18.557	18.558	-0.001	1.000	707451	11.1	23428	
3 Perfluorohexanesulfonic acid								
399.0 > 80.0	19.308	19.308	0.0	1.000	2642246	57.4	48267	
4 Perfluoroheptanoic acid								
363.0 > 319.0	19.344	19.332	0.012	1.000	1136371	17.4	28968	
* 5 13C2-PFOA								
415.0 > 370.0	19.999	19.986	0.013		535790	10.0	13510	
6 Perfluorooctanoic acid								
413.0 > 369.0	19.999	19.986	0.013	1.000	2089950	38.8	1282	
* 8 13C4 PFOS								
503.0 > 80.0	20.619	20.608	0.011		1403740	28.7	29139	
7 Perfluorooctane sulfonic acid								
499.0 > 80.0	20.619	20.608	0.011	1.000	4124317	76.6	43633	
9 Perfluorononanoic acid								E
463.0 > 419.0	20.690	20.679	0.011	1.000	2701379	43.9	31534	E
\$ 10 13C2 PFDA								
515.0 > 470.0	21.409	21.400	0.009	1.000	628161	11.6	19768	

QC Flag Legend

Processing Flags

E - Exceeded Maximum Amount

TestAmerica Sacramento

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_007.d

Injection Date: 02-Jan-2017 13:24:07

Instrument ID: A6

Lims ID: LCSD 320-143783/3-A

Client ID:

Operator ID: CBW

ALS Bottle#: 14

Worklist Smp#: 7

Injection Vol: 10.0 ul

Dil. Factor: 1.0000




Method: 537__A6

Limit Group: LC 537 ICAL

1 Perfluorobutanesulfonic acid

\$ 2 13C2 PFHxA




3 Perfluorohexanesulfonic acid



4 Perfluoroheptanoic acid

* 5 13C2-PFOA




6 Perfluorooctanoic acid




* 8 13C4 PFOS

7 Perfluorooctane sulfonic acid

9 Perfluorononanoic acid



\$ 10 13C2 PFDA



TestAmerica Sacramento
Recovery Report

Data File: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\02JAN2017A6A_007.d
Lims ID: LCSD 320-143783/3-A
Client ID:
Sample Type: LCSD
Inject. Date: 02-Jan-2017 13:24:07 ALS Bottle#: 14 Worklist Smp#: 7
Injection Vol: 10.0 ul Dil. Factor: 1.0000
Sample Info: lcsd 320-143783/3-a
Misc. Info.: Acquity BEH 1.7um, 3X150mm T=35*C
Operator ID: CBW Instrument ID: A6
Method: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b\537__A6.m
Limit Group: LC 537 ICAL
Last Update: 03-Jan-2017 14:46:53 Calib Date: 24-Dec-2016 06:54:10
Integrator: Picker
Quant Method: Internal Standard Quant By: Initial Calibration
Last ICal File: \\ChromNA\Sacramento\ChromData\A6\20161224-38202.b\24DEC2016A6A_009.d
Column 1 : Acquity BEH C18 (2.10 mm) Det: F1:MRM
Process Host: XAWRK018

First Level Reviewer: barnettj

Date: 03-Jan-2017 14:39:50

Compound	Amount Added	Amount Recovered	% Rec.
\$ 2 13C2 PFHxA	10.0	11.1	110.67
\$ 10 13C2 PFDA	10.0	11.6	116.28

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica SacramentoJob No.: 320-24648-1

SDG No.: _____

Instrument ID: A6Start Date: 12/24/2016 03:26Analysis Batch Number: 143828End Date: 12/24/2016 15:56

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
ZZZZZ		12/24/2016 03:26	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 03:56	1		Acquity 2.1 (mm)
STD 320-143828/4 IC		12/24/2016 04:26	1	24DEC2016A6A_00 4.d	Acquity 2.1 (mm)
STD 320-143828/5 IC		12/24/2016 04:55	1	24DEC2016A6A_00 5.d	Acquity 2.1 (mm)
STD 320-143828/6 IC		12/24/2016 05:25	1	24DEC2016A6A_00 6.d	Acquity 2.1 (mm)
STD 320-143828/7 ICISAV		12/24/2016 05:54	1	24DEC2016A6A_00 7.d	Acquity 2.1 (mm)
STD 320-143828/8 IC		12/24/2016 06:24	1	24DEC2016A6A_00 8.d	Acquity 2.1 (mm)
STD 320-143828/9 IC		12/24/2016 06:54	1	24DEC2016A6A_00 9.d	Acquity 2.1 (mm)
ZZZZZ		12/24/2016 07:23	1		Acquity 2.1 (mm)
CCV 320-143828/11 CCVL		12/24/2016 07:53	1	24DEC2016A6A_01 1.d	Acquity 2.1 (mm)
ZZZZZ		12/24/2016 08:22	1		Acquity 2.1 (mm)
ICV 320-143828/13		12/24/2016 08:52	1	24DEC2016A6A_01 3.d	Acquity 2.1 (mm)
ZZZZZ		12/24/2016 09:22	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 11:00	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 11:29	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 11:59	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 12:29	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 12:58	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 13:28	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 13:57	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 14:27	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 14:57	1		Acquity 2.1 (mm)
ZZZZZ		12/24/2016 15:26	1		Acquity 2.1 (mm)
CCV 320-143828/26 CCVIS		12/24/2016 15:56	1		Acquity 2.1 (mm)

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica SacramentoJob No.: 320-24648-1

SDG No.: _____

Instrument ID: A6Start Date: 01/02/2017 10:56Analysis Batch Number: 144608End Date: 01/02/2017 17:20

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
CCV 320-144608/2 CCVL		01/02/2017 10:56	1	02JAN2017A6A_00 2.d	Acquity 2.1 (mm)
CCV 320-144608/3 CCVIS		01/02/2017 11:25	1	02JAN2017A6A_00 3.d	Acquity 2.1 (mm)
ZZZZZ		01/02/2017 11:55	1		Acquity 2.1 (mm)
MB 320-143783/1-A		01/02/2017 12:24	1	02JAN2017A6A_00 5.d	Acquity 2.1 (mm)
LCS 320-143783/2-A		01/02/2017 12:54	1	02JAN2017A6A_00 6.d	Acquity 2.1 (mm)
LCSD 320-143783/3-A		01/02/2017 13:24	1	02JAN2017A6A_00 7.d	Acquity 2.1 (mm)
320-24648-1		01/02/2017 13:53	1	02JAN2017A6A_00 8.d	Acquity 2.1 (mm)
320-24648-2		01/02/2017 14:23	1	02JAN2017A6A_00 9.d	Acquity 2.1 (mm)
320-24648-3		01/02/2017 14:52	1	02JAN2017A6A_01 0.d	Acquity 2.1 (mm)
320-24648-4		01/02/2017 15:22	1	02JAN2017A6A_01 1.d	Acquity 2.1 (mm)
320-24648-5		01/02/2017 15:52	1	02JAN2017A6A_01 2.d	Acquity 2.1 (mm)
320-24648-6		01/02/2017 16:21	1	02JAN2017A6A_01 3.d	Acquity 2.1 (mm)
320-24648-7		01/02/2017 16:51	1	02JAN2017A6A_01 4.d	Acquity 2.1 (mm)
CCV 320-144608/15 CCVIS		01/02/2017 17:20	1	02JAN2017A6A_01 5.d	Acquity 2.1 (mm)

LCMS ANALYSIS RUN LOG

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Instrument ID: A6 Start Date: 01/02/2017 17:20Analysis Batch Number: 144610 End Date: 01/02/2017 23:15

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
CCV 320-144610/15 CCVIS		01/02/2017 17:20	1	02JAN2017A6A_01 5.d	Acquity 2.1 (mm)
ZZZZZ		01/02/2017 17:50	1		Acquity 2.1 (mm)
320-24648-8		01/02/2017 18:20	1	02JAN2017A6A_01 7.d	Acquity 2.1 (mm)
320-24648-9		01/02/2017 18:49	1	02JAN2017A6A_01 8.d	Acquity 2.1 (mm)
ZZZZZ		01/02/2017 21:47	1		Acquity 2.1 (mm)
ZZZZZ		01/02/2017 22:16	1		Acquity 2.1 (mm)
ZZZZZ		01/02/2017 22:46	1		Acquity 2.1 (mm)
CCV 320-144610/27 CCVIS		01/02/2017 23:15	1	02JAN2017A6A_02 7.d	Acquity 2.1 (mm)

LCMS BATCH WORKSHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Batch Number: 143783 Batch Start Date: 12/23/16 18:41 Batch Analyst: Reed, Jonathan EBatch Method: 537 Batch End Date: 12/27/16 14:10

Lab Sample ID	Client Sample ID	Method Chain	Basis	GrossWeight	TareWeight	InitialAmount	FinalAmount	ReceivedpH	LC537-HSP 00014
MB 320-143783/1		537, 537				250.00 mL	1.00 mL	7 SU	
LCS 320-143783/2		537, 537				250.00 mL	1.00 mL	7 SU	50 uL
LCSD 320-143783/3		537, 537				250.00 mL	1.00 mL	7 SU	50 uL
320-24648-A-1	WI-AF-3RW38-1216	537, 537	T	290.60 g	27.45 g	263.2 mL	1.00 mL	7 SU	
320-24648-A-2	WI-AF-3FB38-1216	537, 537	T	285.25 g	26.68 g	258.6 mL	1.00 mL	7 SU	
320-24648-A-3	WI-AF-3RW39-1216	537, 537	T	290.36 g	27.39 g	263 mL	1.00 mL	7 SU	
320-24648-A-4	WI-AF-3FB39-1216	537, 537	T	272.31 g	26.72 g	245.6 mL	1.00 mL	7 SU	
320-24648-A-5	WI-AF-3RW40-1216	537, 537	T	296.11 g	27.08 g	269 mL	1.00 mL	7 SU	
320-24648-A-6	WI-AF-3RW40P-1216	537, 537	T	294.77 g	27.31 g	267.5 mL	1.00 mL	7 SU	
320-24648-A-7	WI-AF-3FB40-1216	537, 537	T	283.65 g	26.73 g	256.9 mL	1.00 mL	7 SU	
320-24648-A-8	WI-AF-3RW41-1216	537, 537	T	292.45 g	26.85 g	265.6 mL	1.00 mL	7 SU	
320-24648-A-9	WI-AF-3FB41-1216	537, 537	T	289.00 g	26.07 g	262.9 mL	1.00 mL	7 SU	

Lab Sample ID	Client Sample ID	Method Chain	Basis	LC537-IS 00027	LC537-SU 00027	AnalysisComment			
MB 320-143783/1		537, 537		20 uL	50 uL	Free Chlorine: ND			
LCS 320-143783/2		537, 537		20 uL	50 uL	Free Chlorine: ND			
LCSD 320-143783/3		537, 537		20 uL	50 uL	Free Chlorine: ND			
320-24648-A-1	WI-AF-3RW38-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-2	WI-AF-3FB38-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-3	WI-AF-3RW39-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-4	WI-AF-3FB39-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-5	WI-AF-3RW40-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-6	WI-AF-3RW40P-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-7	WI-AF-3FB40-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

LCMS BATCH WORKSHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Batch Number: 143783 Batch Start Date: 12/23/16 18:41 Batch Analyst: Reed, Jonathan EBatch Method: 537 Batch End Date: 12/27/16 14:10

Lab Sample ID	Client Sample ID	Method Chain	Basis	LC537-IS 00027	LC537-SU 00027	AnalysisComment			
320-24648-A-8	WI-AF-3RW41-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			
320-24648-A-9	WI-AF-3FB41-1216	537, 537	T	20 uL	50 uL	Free Chlorine: ND			

Batch Notes	
Manifold ID	3
Methanol ID	807182
Pipette ID	MD05306
Analyst ID - IS Reagent Drop	VPM
Analyst ID - IS Reagent Drop Witness	CCB
Analyst ID - SU Reagent Drop	JER
Analyst ID - SU Reagent Drop Witness	KMK
Analyst ID - TA Reagent Drop	JER
Analyst ID - TA Reagent Drop Witness	KMK
SPE Cartridge ID	6341059-01
Trizma ID	SLBR4303V
Reagent Water ID	12/23/16

Basis	Basis Description
T	Total/NA

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

Method 537 CCV/Data Review Checklist

Job No: 24648 Instrument ID & Date: 1-2-17 ICAL Batch: 143828 JRB
1-3-16
Extraction Batch: 143783 Worklist #: 38388 TALS Batch: 144608, 144609, 144610

Review Items	Level 1			Level 2
	Yes	No	N/A	
Initial Calibration				
1. Is ICAL verified and locked in Chrom & TALS?	✓			✓
2. Is ICV properly linked in TALS?	✓			✓
Continuing Calibration				
1. Low-range CCV injected at start of analytical run? CCV injected after every 10 samples and at the end of the analytical run and alternated between Low-range, Mid-range and High-range?	✓			✓
2. If sequence was not after an ICAL was a low and mid range CCV injected at the start of the analytical run? <u>144608</u>	✓			✓
3. Native compounds and surrogates in control? Low-range within ±50% of true value Mid and High-range within ±30% of true value	✓			✓
4. Internal Standard areas in control? Areas ≥ 50% of average area of the ICAL and 70-140% of the most recent CCV.	✓			✓
Client Samples & QC Sample Results				
1. Were preparation and analysis done within holding times?	✓			✓
2. Are Chromatograms reviewed and spectra verified?	✓			✓
3. Are positive results within calibration range?	✓			✓
4. Dilutions due to target cpds? _____ Dilutions due to non-targets? _____			✓	
5. All target compounds in MB < 1/3 RL ? (Requires NCM if "no.")	✓			✓
6. Are target constituents in LCS/LCSD within method control limits?	✓			✓
7. Internal Standard areas in control for all samples and QC reported? ±50% from the average area of the ICAL and 70-140% of the most recent CCV	✓			✓
8. Do results (e.g., dilutions/trip blanks) make sense?	✓			✓
9. Are MS/MSD recoveries and RPDs within method control limits?			✓	
10. Are all QC samples properly linked in TALS?	✓			✓
11. All manual integrations appropriate and completely documented?	✓			✓
12. Are nonconformances documented as NCMs?	✓			✓
13. Are all Chrom graphics uploaded?	✓			✓

1st Level Reviewer / Date: JRB 1-3-17

2nd Level Reviewer / Date: MLW 1/3/2017

NCM # and Comments: 74332

Instrument ID & Date: AL6 12/24/16 Worklist#: 39202

ICAL Batch: 143828, 143829 Calibration ID number: 27291, 27292

Review Items	-- Level 1 --			Level 2
	Yes	No	N/A	
Initial Calibration				
1. Mass calibration, as needed, verified by full scan of PFC stock standard. All PFC ions used for quantitation are within 0.3 m/z of true mass?	✓			✓
2. Responses increase with increasing concentration?	✓			✓
3. Fit used (circle): <u>Average</u> Linear (1/x ²) Linear Quadratic (6 points minimum)				
4. Meets fit criteria? Intercept ≤ ½ RL RSD ≤ 30% for Average R ² ≥ 0.990 for Linear R ² ≥ 0.990 for Quadratic NOTE: "Force through Zero" must be used and weighted if needed	✓			✓
5. If quadratic fit used the curve does not "bend over".			✓	
6. Feed calibration points into the calculated curve. Are points ≤ MRL within ±50% of true value? Are points > MRL within ±30% of true value?	✓			✓
7. Any carryover from the high calibration point must be < 1/3 RL	✓			✓
8. Asymmetry check meets criteria for the first two eluting peaks? (0.8 - 1.5).	✓			✓
9. Is the asymmetry check scanned and linked in TALS to the calibration point?	✓			✓
10. Is ICV (2 nd source) ± 30% of true value?	✓			✓
11. Is ICV (2 nd source) internal standards ±50% of average area of the ICAL?	✓			✓
12. ICAL locked in Chrom and uploaded to TALS?	✓			✓
13. ICAL locked in TALS and scanned?	✓			✓

1st Level Reviewer / Date: CBW 12/27/16

2nd Level Reviewer / Date: Murphy 12/27/2016

NCM # and Comments:

TestAmerica Laboratories
Worklist QC Batch Report

Worklist Name: 02JAN2017A_A6 537

Worklist Number: 38388

Instrument Name: A6

Chrom Method: 537_A6

Data Directory: \\ChromNA\Sacramento\ChromData\A6\20170102-38388.b

QC Batching: Enabled

Limit Group Batching: Enabled

QC Batch: 1	LC 537 ICAL Raw Batch: 144608	LC 537 CS ICAL Raw Batch: 144607
# 1 RB	# 1 RB	# 2 CCV L2
# 2 CCV L2	# 2 CCV L2	
# 3 CCV L3	# 3 CCV L3	
# 4 RB	# 4 RB	
# 5 MB 320-143783/1-A	# 5 MB 320-143783/1-A	
# 6 LCS 320-143783/2-A	# 6 LCS 320-143783/2-A	
# 7 LCSD 320-143783/3-A	# 7 LCSD 320-143783/3-A	
# 8 320-24648-A-1-A	# 8 320-24648-A-1-A	
# 9 320-24648-A-2-A	# 9 320-24648-A-2-A	
#10 320-24648-A-3-A	#10 320-24648-A-3-A	
#11 320-24648-A-4-A	#11 320-24648-A-4-A	
#12 320-24648-A-5-A	#12 320-24648-A-5-A	
#13 320-24648-A-6-A	#13 320-24648-A-6-A	
#14 320-24648-A-7-A	#14 320-24648-A-7-A	
#15 CCV L5	#15 CCV L5	

QC Batch: 2	LC 537 CS ICAL Raw Batch: 144609	LC 537 ICAL Raw Batch: 144610
#15 CCV L5	#15 CCV L5	#15 CCV L5
#16 RB	#16 RB	#16 RB
#17 320-24648-A-8-A		#17 320-24648-A-8-A
#18 320-24648-A-9-A		#18 320-24648-A-9-A
#19 MB 320-144356/1-A	#19 MB 320-144356/1-A	
#20 LCS 320-144356/2-A	#20 LCS 320-144356/2-A	
#21 280-91747-E-1-A	#21 280-91747-E-1-A	
#22 280-91747-B-1-A DU	#22 280-91747-B-1-A DU	
#23 280-91747-B-1-B MS	#23 280-91747-B-1-B MS	
#24 MB 320-143781/1-A		#24 MB 320-143781/1-A
#25 LCS 320-143781/2-A		#25 LCS 320-143781/2-A
#26 LCSD 320-143781/3-A		#26 LCSD 320-143781/3-A
#27 CCV L3	#27 CCV L3	#27 CCV L3

QC Batch: 3	LC 537 CS ICAL Raw Batch: 144611	LC 537 ICAL Raw Batch: 144612
#27 CCV L3	#27 CCV L3	#27 CCV L3
#28 RB		#28 RB
#29 320-24636-A-1-A		#29 320-24636-A-1-A
#30 320-24636-A-2-A		#30 320-24636-A-2-A
#31 320-24636-A-3-A		#31 320-24636-A-3-A
#32 320-24636-A-4-A		#32 320-24636-A-4-A
#33 320-24636-A-5-A		#33 320-24636-A-5-A
#34 320-24636-A-6-A		#34 320-24636-A-6-A
#35 320-24636-A-7-A		#35 320-24636-A-7-A
#36 320-24636-A-8-A		#36 320-24636-A-8-A
#37 320-24637-A-1-A		#37 320-24637-A-1-A
#38 320-24637-A-2-A		#38 320-24637-A-2-A
#39 CCV L5		#39 CCV L5

QC Batch: 4	LC 537 ICAL Raw Batch: 144614
#39 CCV L5	#39 CCV L5
#40 RB	#40 RB
#41 320-24637-B-3-A	#41 320-24637-B-3-A
#42 320-24637-A-4-A	#42 320-24637-A-4-A
#43 MB 320-144066/1-A	#43 MB 320-144066/1-A
#44 LCS 320-144066/2-A	#44 LCS 320-144066/2-A
#45 320-24673-A-1-A	#45 320-24673-A-1-A
#46 320-24673-A-2-A	#46 320-24673-A-2-A
#47 320-24673-A-3-A	#47 320-24673-A-3-A
#48 320-24673-A-4-A	#48 320-24673-A-4-A
#49 320-24673-A-5-A	#49 320-24673-A-5-A
#50 320-24673-A-6-A	#50 320-24673-A-6-A
#51 CCV L3	#51 CCV L3

QC Batch: 5	LC 537 ICAL Raw Batch: 144616	LC 537 CS ICAL Raw Batch: 144615
#51 CCV L3	#51 CCV L3	
#52 RB	#52 RB	
#61 QC LC537-SU_00028	#61 QC LC537-SU_00028	
#53 320-24673-A-7-A	#53 320-24673-A-7-A	
#54 320-24673-A-7-B MS	#54 320-24673-A-7-B MS	
#55 320-24673-A-7-C MSD	#55 320-24673-A-7-C MSD	
#56 320-24673-A-8-A	#56 320-24673-A-8-A	
#57 320-24673-A-9-A	#57 320-24673-A-9-A	
#58 320-24673-A-10-A	#58 320-24673-A-10-A	
#59 CCV L5	#59 CCV L5	#59 CCV L5

QC Batch: 6	LC 537 ICAL Raw Batch: 144764	LC 537 CS ICAL Raw Batch: 144714
#59 CCV L5	#59 CCV L5	#59 CCV L5
#60 RB		#60 RB
#62 320-24228-A-2-A		#62 320-24228-A-2-A
#63 320-24322-A-1-A		#63 320-24322-A-1-A
#64 320-24322-A-1-C DU		#64 320-24322-A-1-C DU
#65 320-24322-A-1-B MS		#65 320-24322-A-1-B MS
#66 320-24322-A-3-A		#66 320-24322-A-3-A
#67 320-24322-A-5-A		#67 320-24322-A-5-A
#68 CCV L3		#68 CCV L3
#69 RB		#69 RB

78

Rush

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Batch Number: 320-143783

Method Code: 320-537_Prep-320

Analyst: Reed, Jonathan E

Screen by 12/23/16
Sam AG 12/21/16

Batch Open: 12/23/2016 6:41:00PM

Batch End: 12/22/16 14:10

AG 1/22/17

Due 1/3

Extraction of Perfluorinated Alkyl Acids

Input Sample Lab ID (Analytical Method)	SDG (Job #)	GrossWt TareWt	InitAmt FinAmt	PHs		Due Date	Analytical TAT	Div Rank	Comments	Output Sample Lab ID
				Rcvd	Adj1 Adj2					
1 MB-320-143783/1 N/A	N/A		250.00 mL 1.00 mL	7		N/A	N/A	N/A		MB-320-143783/1-A
2 LCS-320-143783/2 N/A	N/A		250.00 mL 1.00 mL	7		N/A	N/A	N/A	Free Chlorine ND	LCS-320-143783/2-A
3 LCSD-320-143783/3 N/A	N/A		250.00 mL 1.00 mL	7		N/A	N/A	N/A		LCS-320-143783/3-A
4 320-24648-A-1 (537_DOD5)	N/A (320-24648-1)	290.60 g 27.45 g	263.2 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-1-A
5 320-24648-A-2 (537_DOD5)	N/A (320-24648-1)	285.25 g 26.68 g	258.6 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-2-A
6 320-24648-A-3 (537_DOD5)	N/A (320-24648-1)	290.36 g 27.39 g	263 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-3-A
7 320-24648-A-4 (537_DOD5)	N/A (320-24648-1)	272.31 g 26.72 g	245.6 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-4-A
8 320-24648-A-5 (537_DOD5)	N/A (320-24648-1)	296.11 g 27.08 g	269 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-5-A
9 320-24648-A-6 (537_DOD5)	N/A (320-24648-1)	294.77 g 27.31 g	267.5 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-6-A
10 320-24648-A-7 (537_DOD5)	N/A (320-24648-1)	283.65 g 26.73 g	256.9 mL 1.00 mL	7		12/28/16	8_Day_Rush	4		320-24648-A-7-A

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Batch Number: 320-143783

Method Code: 320-537_Prep-320

Analyst: Reed, Jonathan E

Batch Open: 12/23/2016 6:41:00PM

Batch End:

320-24648-A-8 (537_DOD5)	N/A (320-24648-1)	292.45 g	265.6 mL		12/28/16	8_Day_Rush	4	Free Chlorine: ND	320-24648-A-8-A
		26.85 g	1.00 mL	7					
320-24648-A-9 (537_DOD5)	N/A (320-24648-1)	289.00 g	262.9 mL	7	12/28/16	8_Day_Rush	4		320-24648-A-9-A
		26.07 g	1.00 mL						

Batch Notes

Manifold ID 3
 Trizma ID SLBR4303V
 SPE Cartridge ID 6341059-01
 Methanol ID 807182
 Reagent Water ID 12/23/16
 Pipette ID MD05306
 Analyst ID - TA Reagent Drop JER
 Analyst ID - TA Reagent Drop Witness KMK
 Analyst ID - SU Reagent Drop JER
 Analyst ID - SU Reagent Drop Witness KMK
 Analyst ID - IS Reagent Drop JPM (808145)
 Analyst ID - IS Reagent Drop Witness CCB
 Batch Comment

Comments

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Batch Number: 320-143783

Method Code: 320-537_Prep-320

Analyst: Reed, Jonathan E

Batch Open: 12/23/2016 6:41:00PM

Batch End:

Reagent Additions Worksheet

Lab ID	Reagent Code	Amount Added	Final Amount	By	Witness
MB 320-143783/1	LC537-SU_00027	50 uL	1.00 mL		KMK 12-23-16
LCS 320-143783/2	LC537-HSP_00014	50 uL	1.00 mL		
LCS 320-143783/2	LC537-SU_00027	50 uL	1.00 mL		
LCSD 320-143783/3	LC537-HSP_00014	50 uL	1.00 mL		
LCSD 320-143783/3	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-1	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-2	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-3	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-4	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-5	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-6	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-7	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-8	LC537-SU_00027	50 uL	1.00 mL		
320-24648-A-9	LC537-SU_00027	50 uL	1.00 mL		

Aqueous Extraction Analysis Sheet

(To Accompany Samples to Instruments)

Analyst: Reed, Jonathan E

Batch Open: 12/23/2016 6:41:00PM

Batch End:

Batch Number: 320-143783

Method Code: 320-537_Prep-320

Reagent	Other Reagents: Amount/Units	Lot#:

Preparation Batch Number(s): 143783 Test: 537-0005 (rush)

Earliest Holding Time: 1/04

Sample List Tab	1 st Level Reviewer	2 nd Level Reviewer
Samples identified to the correct method	✓	✓
All necessary NCMs filed (including holding time)	✓	✓
Method/sample/login/QAS checked and correct	✓	✓
Worksheet Tab	1 st Level Reviewer	2 nd Level Reviewer
All samples properly preserved	✓	✓
Weights in anticipated range and not targeted	✓	✓
All additional test requirements performed, documented, and uploaded to TALS correctly (e.g. final amount, initial amount, turbidity, and CI Check)	✓	✓
The pH is transcribed correctly in TALS	✓	✓
All additional information transcribed into TALS is correct and raw data is attached	✓	✓
Comments are transcribed correctly in TALS	✓	✓
Reagents Tab	1 st Level Reviewer	2 nd Level Reviewer
All necessary reagents not expired and entered into TALS	✓	✓
All spike amounts correct and added to necessary samples and QC	✓	✓
Batch Information	1 st Level Reviewer	2 nd Level Reviewer
Date and time accurate and entered into TALS correctly	✓	✓
All necessary 'batch information' complete and entered into TALS correctly	✓	✓

1st Level Reviewer: [Signature]

Date: 12/27/16

2nd Level Reviewer: [Signature]

Date: 12/27/16

Comments: _____

Shipping and Receiving Documents

Login Sample Receipt Checklist

Client: CH2M Hill Constructors, Inc.

Job Number: 320-24648-1

Login Number: 24648

List Source: TestAmerica Sacramento

List Number: 1

Creator: Edman, Connor M

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	True	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	N/A	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

Lab_Sample_ID	Contract_ID	DO_CTO_Number	Phase	Installation_ID	Sample_Name	CH2M_Code	Analysis_Group	Analytical_Method	PRC_Code	Lab_Code	Lab_Name	Leachate_Method	Sample_Basis	Extraction_Method	Result_Type	Lab_QC_Type	Sample_Medium	QC_Level	DateTime_Collected	Date_Received	Leachate_Deate	Leachate_Time	Extraction_Date	Extraction_Time	Analysis_Date	Analysis_Time	Dilution	Run_Number	Percent_Moisture	Percent_Lipid
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:18	12/23/2016			20161223	18:41:00	20170102	13:53:00	1	1		
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:18	12/23/2016			20161223	18:41:00	20170102	13:53:00	1	1		
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:18	12/23/2016			20161223	18:41:00	20170102	13:53:00	1	1		
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:18	12/23/2016			20161223	18:41:00	20170102	13:53:00	1	1		
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:18	12/23/2016			20161223	18:41:00	20170102	13:53:00	1	1		
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:19	12/23/2016			20161223	18:41:00	20170102	14:23:00	1	1		
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:19	12/23/2016			20161223	18:41:00	20170102	14:23:00	1	1		
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:19	12/23/2016			20161223	18:41:00	20170102	14:23:00	1	1		
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:19	12/23/2016			20161223	18:41:00	20170102	14:23:00	1	1		
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:19	12/23/2016			20161223	18:41:00	20170102	14:23:00	1	1		
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:25	12/23/2016			20161223	18:41:00	20170102	14:52:00	1	1		
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:25	12/23/2016			20161223	18:41:00	20170102	14:52:00	1	1		
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:25	12/23/2016			20161223	18:41:00	20170102	14:52:00	1	1		
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:25	12/23/2016			20161223	18:41:00	20170102	14:52:00	1	1		
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:25	12/23/2016			20161223	18:41:00	20170102	14:52:00	1	1		
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:26	12/23/2016			20161223	18:41:00	20170102	15:22:00	1	1		
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:26	12/23/2016			20161223	18:41:00	20170102	15:22:00	1	1		
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:26	12/23/2016			20161223	18:41:00	20170102	15:22:00	1	1		
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:26	12/23/2016			20161223	18:41:00	20170102	15:22:00	1	1		
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 09:26	12/23/2016			20161223	18:41:00	20170102	15:22:00	1	1		
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:47	12/23/2016			20161223	18:41:00	20170102	15:52:00	1	1		
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:47	12/23/2016			20161223	18:41:00	20170102	15:52:00	1	1		
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:47	12/23/2016			20161223	18:41:00	20170102	15:52:00	1	1		
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:47	12/23/2016			20161223	18:41:00	20170102	15:52:00	1	1		
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:52	12/23/2016			20161223	18:41:00	20170102	16:21:00	1	1		
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:52	12/23/2016			20161223	18:41:00	20170102	16:21:00	1	1		
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:52	12/23/2016			20161223	18:41:00	20170102	16:21:00	1	1		
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:52	12/23/2016			20161223	18:41:00	20170102	16:21:00	1	1		
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:53	12/23/2016			20161223	18:41:00	20170102	16:51:00	1	1		
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:53	12/23/2016			20161223	18:41:00	20170102	16:51:00	1	1		
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:53	12/23/2016			20161223	18:41:00	20170102	16:51:00	1	1		
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 16:53	12/23/2016			20161223	18:41:00	20170102	16:51:00	1	1		
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:40	12/23/2016			20161223	18:41:00	20170102	18:20:00	1	1		
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:40	12/23/2016			20161223	18:41:00	20170102	18:20:00	1	1		
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:40	12/23/2016			20161223	18:41:00	20170102	18:20:00	1	1		
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:40	12/23/2016			20161223	18:41:00	20170102	18:20:00	1	1		
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:41	12/23/2016			20161223	18:41:00	20170102	18:49:00	1	1		
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:41	12/23/2016			20161223	18:41:00	20170102	18:49:00	1	1		
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	REG	W	4	12/21/2016 18:41	12/23/2016			20161223	18:41:00	20170102	18:49:00	1	1		
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	BS	W	4	12/23/2016 18:41	12/23/2016			20161223	18:41:00	20170102	12:54:00	1	1		
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	BS	W	4	12/23/2016 18:41	12/23/2016			20161223	18:41:00	20170102	12:54:00	1	1		
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	BS	W	4	12/23/2016 18:41	12/23/2016			20161223	18:41:00	20170102	12:54:00	1	1		
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	NONE		537	ORG	TAMER	Test America	NONE	NA	METHOD	000	BS	W	4	12/23/2016 18:41	12/23/2016			20161223	18:41:00	20170102	12:54:00	1	1		
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	NONE		537	ORG																					

Lab_Sample_ID	Contract_ID	DO_CTO_Number	Phase	Installation_ID	Sample_Name	Chem_Name	Analyte_ID	Analyte_Value	Original_Analyte_Value	Result_Units	Lab_Qualifier	Validator_Qualifier	GC_Column_Type	Analysis_Result_Type	Result_Narrative	QC_Control_Limit_Code	QC_Accuracy_Upper	QC_Accuracy_Lower	Control_Limit_Date	QC_Narrative	MDL	Detection_Limit	QSM_Version	DL	LOD	LOQ	SDG	Analysis_Batch
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.046	UG_L	U M			PR	TRG					00000000				5.0	0.015	0.046	0.057	320-24648-1	320-144608
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U M			PR	TRG					00000000				5.0	0.0089	0.023	0.028	320-24648-1	320-144608
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U			PR	TRG					00000000				5.0	0.045	0.10	0.13	320-24648-1	320-144608
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	13C2 PFHXA	13C2 PFHXA	101	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-1	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW38-1216	13C2 PFDA	13C2 PFDA	119	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.046	UG_L	U M			PR	TRG					00000000				5.0	0.015	0.046	0.058	320-24648-1	320-144608
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U M			PR	TRG					00000000				5.0	0.0091	0.023	0.029	320-24648-1	320-144608
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.11	UG_L	U			PR	TRG					00000000				5.0	0.046	0.11	0.14	320-24648-1	320-144608
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	13C2 PFHXA	13C2 PFHXA	100	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-2	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB38-1216	13C2 PFDA	13C2 PFDA	106	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.046	UG_L	U			PR	TRG					00000000				5.0	0.015	0.046	0.057	320-24648-1	320-144608
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U M			PR	TRG					00000000				5.0	0.0090	0.023	0.029	320-24648-1	320-144608
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U			PR	TRG					00000000				5.0	0.045	0.10	0.13	320-24648-1	320-144608
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	13C2 PFHXA	13C2 PFHXA	103	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-3	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW39-1216	13C2 PFDA	13C2 PFDA	101	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.049	UG_L	U			PR	TRG					00000000				5.0	0.016	0.049	0.061	320-24648-1	320-144608
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.024	UG_L	U			PR	TRG					00000000				5.0	0.0096	0.024	0.031	320-24648-1	320-144608
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.11	UG_L	U			PR	TRG					00000000				5.0	0.048	0.11	0.14	320-24648-1	320-144608
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	13C2 PFHXA	13C2 PFHXA	101	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-4	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB39-1216	13C2 PFDA	13C2 PFDA	107	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U			PR	TRG					00000000				5.0	0.014	0.045	0.056	320-24648-1	320-144608
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.022	UG_L	U M			PR	TRG					00000000				5.0	0.0088	0.022	0.028	320-24648-1	320-144608
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U			PR	TRG					00000000				5.0	0.044	0.10	0.13	320-24648-1	320-144608
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	13C2 PFHXA	13C2 PFHXA	110	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-5	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40-1216	13C2 PFDA	13C2 PFDA	123	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U			PR	TRG					00000000				5.0	0.014	0.045	0.056	320-24648-1	320-144608
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.022	UG_L	U			PR	TRG					00000000				5.0	0.0088	0.022	0.028	320-24648-1	320-144608
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U			PR	TRG					00000000				5.0	0.044	0.10	0.13	320-24648-1	320-144608
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	13C2 PFHXA	13C2 PFHXA	96	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-6	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW40P-1216	13C2 PFDA	13C2 PFDA	107	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.047	UG_L	U			PR	TRG					00000000				5.0	0.015	0.047	0.058	320-24648-1	320-144608
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U			PR	TRG					00000000				5.0	0.0092	0.023	0.029	320-24648-1	320-144608
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.11	UG_L	U			PR	TRG					00000000				5.0	0.046	0.11	0.14	320-24648-1	320-144608
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	13C2 PFHXA	13C2 PFHXA	99	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-7	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB40-1216	13C2 PFDA	13C2 PFDA	108	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144608
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.045	UG_L	U M			PR	TRG					00000000				5.0	0.015	0.045	0.056	320-24648-1	320-144610
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U M			PR	TRG					00000000				5.0	0.0089	0.023	0.028	320-24648-1	320-144610
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.059	UG_L	J			PR	TRG					00000000				5.0	0.045	0.10	0.13	320-24648-1	320-144610
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	13C2 PFHXA	13C2 PFHXA	91	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144610
320-24648-8	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3RW41-1216	13C2 PFDA	13C2 PFDA	105	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144610
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.046	UG_L	U M			PR	TRG					00000000				5.0	0.015	0.046	0.057	320-24648-1	320-144610
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	Perfluorooctanoic acid (PFOA)	335-67-1	0.023	UG_L	U			PR	TRG					00000000				5.0	0.0090	0.023	0.029	320-24648-1	320-144610
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	Perfluorobutanesulfonic acid (PFBS)	375-73-5	0.10	UG_L	U			PR	TRG					00000000				5.0	0.045	0.10	0.13	320-24648-1	320-144610
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	13C2 PFHXA	13C2 PFHXA	102	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144610
320-24648-9	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	WI-AF-3FB41-1216	13C2 PFDA	13C2 PFDA	108	PCT_REC				PR	SURR		SLSA	130	70	00000000				5.0				320-24648-1	320-144610
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	Perfluorooctane Sulfonate (PFOS)	1763-23-1	0.06	PCT_REC				PR	TRG		LSA	130	70	00000000				5.0	0.016	0.048	0.060	320-24648-1	320-144608
LCS 320-143783/2-A	N6247016D9000	0008		WHIDBEY_ISLAND_NAS	LCS 320-143783/2-A	Perfluorooctanoic acid (PFOA)	335-67-1	0.09	PCT_REC	E			PR	TRG		LSA	130	70	00000000				5.0	0.0094	0.024	0.03		

**DATA VALIDATION SUMMARY REPORT
WHIDBEY ISLAND, WASHINGTON**

Client: CH2M HILL, Inc., Corvallis, Oregon
SDG: 320-24648-1
Laboratory: Test America, Sacramento, California
Site: Whidbey Island, CTO-0008, Washington
Date: January 13, 2017

PFCs			
EDS ID	Client Sample ID	Laboratory Sample ID	Matrix
1	WI-AF-3RW38-1216	320-24648-1	Water
2	WI-AF-3FB38-1216	320-24648-2	Water
3	WI-AF-3RW39-1216	320-24648-3	Water
4	WI-AF-3FB39-1216	320-24648-4	Water
5	WI-AF-3RW40-1216	320-24648-5	Water
6	WI-AF-3RW40P-1216	320-24648-6	Water
7	WI-AF-3FB40-1216	320-24648-7	Water
8	WI-AF-3RW41-1216	320-24648-8	Water
9	WI-AF-3FB41-1216	320-24648-9	Water

A full data validation was performed on the analytical data for five water samples and four aqueous field blank samples collected on December 21, 2016 by CH2M HILL at the Whidbey Island site in Washington. The samples were analyzed under the EPA Method "Determination of Selected Perfluorinated Alkyl Acids in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS)".

Specific method references are as follows:

Analysis
PFCs

Method References
USEPA Method 537 Rev 1.1 Modified

The data have been validated according to the protocols and quality control (QC) requirements of the analytical method, and the U.S. Department of Defense (DoD) Quality Systems Manual (QSM), Version 5.0 (DoD 2013) and the USEPA National Functional Guidelines for Organic Data Review as follows:

- The USEPA "Contract Laboratories Program National Functional Guidelines for Superfund Organic Methods Data Review," August 2014;
- and the reviewer's professional judgment.

The following data quality indicators were reviewed for this report:

Organics

- Date Completeness, Case Narrative & Custody Documentation
- Holding times
- Gas Chromatography/Mass Spectrometry (GC/MS) Tuning
- Initial and continuing calibration summaries
- Method blank and field QC blank contamination
- Surrogate Spike recoveries
- Matrix Spike/Matrix Spike Duplicate (MS/MSD) recoveries
- Laboratory Control Sample/Laboratory Control Sample Duplicate (LCS/LCSD) recoveries
- Internal standard area and retention time summary forms
- Target Compound Identification
- Compound Quantitation
- Field Duplicate sample precision

A full (Level IV) data validation was performed with this review including a recalculation of 10% of the detected results in the samples.

Data Usability Assessment

There were no rejections of data.

Overall the data is acceptable for the intended purposes. There were no qualifications.

Perfluorinated Compounds (PFCs)

Data Completeness, Case Narrative & Custody Documentation

- The case narrative and chain-of-custody documentation were included in the data package as required. All criteria were met.

Holding Times

- All samples were extracted within 14 days for water samples and analyzed within 28 days.

GC/MS Tuning

- All criteria were met.

Initial Calibration

- All relative standard deviation (%RSD) and/or correlation coefficients criteria were met.

Continuing Calibration

- All percent difference (%D) and RRF criteria were met.

Method Blank

- The method blanks were free of contamination.

Field QC Blank

- The field blank samples were free of contamination.

Surrogate Spike Recoveries

- All samples exhibited acceptable surrogate %R values.

Matrix Spike/Matrix Spike Duplicate (MS/MSD) Recoveries

- A MS/MSD sample was not collected.

Laboratory Control Samples/Laboratory Control Sample Duplicates

- The LCS/LCSD samples exhibited acceptable percent recoveries (%R) and RPD values.

Internal Standard (IS) Area Performance

- All internal standards met response and retention time (RT) criteria.

Target Compound Identification

- All mass spectra and quantitation criteria were met.

Compound Quantitation

- All criteria were met.


Field Duplicate Sample Precision

- Field duplicate results are summarized below.

Compound	WI-AF-3RW40-1216 ug/L	WI-AF-3RW40P-1216 ug/L	RPD	Qualifier
None	ND	ND	-	-

Please contact the undersigned at (757) 564-0090 if you have any questions or need further information.

Signed:


Nancy Weaver
Senior Chemist

Dated: 1/17/17

Data Qualifier	Definition
U	The analyte was analyzed for, but was not detected above the level of the reported sample quantitation limit.
J	The analyte is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
NJ	The analysis has been "tentatively identified" or "presumptively" as present and the associated numerical value is the estimated concentration in the samples.
UJ	The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
R	The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the samples.

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW38-1216 Lab Sample ID: 320-24648-1
 Matrix: Water Lab File ID: 02JAN2017A6A_008.d
 Analysis Method: 537 Date Collected: 12/21/2016 09:18
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 263.2(mL) Date Analyzed: 01/02/2017 13:53
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	101		70-130
STL00996	13C2 PFDA	119		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

2

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB38-1216 Lab Sample ID: 320-24648-2
 Matrix: Water Lab File ID: 02JAN2017A6A_009.d
 Analysis Method: 537 Date Collected: 12/21/2016 09:19
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 258.6(mL) Date Analyzed: 01/02/2017 14:23
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.058	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U N	0.029	0.023	0.0091
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	100		70-130
STL00996	13C2 PFDA	106		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

3

Lab Name: <u>TestAmerica Sacramento</u>	Job No.: <u>320-24648-1</u>
SDG No.: _____	
Client Sample ID: <u>WI-AF-3RW39-1216</u>	Lab Sample ID: <u>320-24648-3</u>
Matrix: <u>Water</u>	Lab File ID: <u>02JAN2017A6A_010.d</u>
Analysis Method: <u>537</u>	Date Collected: <u>12/21/2016 09:25</u>
Extraction Method: <u>537</u>	Date Extracted: <u>12/23/2016 18:41</u>
Sample wt/vol: <u>263(mL)</u>	Date Analyzed: <u>01/02/2017 14:52</u>
Con. Extract Vol.: <u>1.00(mL)</u>	Dilution Factor: <u>1</u>
Injection Volume: <u>10(uL)</u>	GC Column: <u>Acquity</u> ID: <u>2.1(mm)</u>
% Moisture: _____	GPC Cleanup: (Y/N) <u>N</u>
Analysis Batch No.: <u>144608</u>	Units: <u>ug/L</u>

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U ✗	0.029	0.023	0.0090
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	103		70-130
STL00996	13C2 PFDA	101		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

4

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1

SDG No.: _____

Client Sample ID: WI-AF-3FB39-1216 Lab Sample ID: 320-24648-4

Matrix: Water Lab File ID: 02JAN2017A6A_011.d

Analysis Method: 537 Date Collected: 12/21/2016 09:26

Extraction Method: 537 Date Extracted: 12/23/2016 18:41

Sample wt/vol: 245.6(mL) Date Analyzed: 01/02/2017 15:22

Con. Extract Vol.: 1.00(mL) Dilution Factor: 1

Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)

% Moisture: _____ GPC Cleanup: (Y/N) N

Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.049	U	0.061	0.049	0.016
335-67-1	Perfluorooctanoic acid (PFOA)	0.024	U	0.031	0.024	0.0096
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.048

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	101		70-130
STL00996	13C2 PFDA	107		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

5

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3RW40-1216 Lab Sample ID: 320-24648-5
 Matrix: Water Lab File ID: 02JAN2017A6A_012.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:47
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 269(mL) Date Analyzed: 01/02/2017 15:52
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U 1	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	110		70-130
STL00996	13C2 PFDA	123		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

6

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.:
 Client Sample ID: WI-AF-3RW40P-1216 Lab Sample ID: 320-24648-6
 Matrix: Water Lab File ID: 02JAN2017A6A_013.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:52
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 267.5(mL) Date Analyzed: 01/02/2017 16:21
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U	0.056	0.045	0.014
335-67-1	Perfluorooctanoic acid (PFOA)	0.022	U	0.028	0.022	0.0088
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.044

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	96		70-130
STL00996	13C2 PFDA	107		70-130

12/13/17

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

7

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.:
 Client Sample ID: WI-AF-3FB40-1216 Lab Sample ID: 320-24648-7
 Matrix: Water Lab File ID: 02JAN2017A6A_014.d
 Analysis Method: 537 Date Collected: 12/21/2016 16:53
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 256.9(mL) Date Analyzed: 01/02/2017 16:51
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144608 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.047	U	0.058	0.047	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0092
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.11	U	0.14	0.11	0.046

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	99		70-130
STL00996	13C2 PFDA	108		70-130

11/13/17

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET

8

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
SDG No.:
Client Sample ID: WI-AF-3RW41-1216 Lab Sample ID: 320-24648-8
Matrix: Water Lab File ID: 02JAN2017A6A_017.d
Analysis Method: 537 Date Collected: 12/21/2016 18:40
Extraction Method: 537 Date Extracted: 12/23/2016 18:41
Sample wt/vol: 265.6(mL) Date Analyzed: 01/02/2017 18:20
Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
% Moisture: GPC Cleanup: (Y/N) N
Analysis Batch No.: 144610 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.045	U M	0.056	0.045	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U M	0.028	0.023	0.0089
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.059	J	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	91		70-130
STL00996	13C2 PFDA	105		70-130

FORM I
LCMS ORGANICS ANALYSIS DATA SHEET


9

Lab Name: TestAmerica Sacramento Job No.: 320-24648-1
 SDG No.: _____
 Client Sample ID: WI-AF-3FB41-1216 Lab Sample ID: 320-24648-9
 Matrix: Water Lab File ID: 02JAN2017A6A_018.d
 Analysis Method: 537 Date Collected: 12/21/2016 18:41
 Extraction Method: 537 Date Extracted: 12/23/2016 18:41
 Sample wt/vol: 262.9(mL) Date Analyzed: 01/02/2017 18:49
 Con. Extract Vol.: 1.00(mL) Dilution Factor: 1
 Injection Volume: 10(uL) GC Column: Acquity ID: 2.1(mm)
 % Moisture: _____ GPC Cleanup: (Y/N) N
 Analysis Batch No.: 144610 Units: ug/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	0.046	U M	0.057	0.046	0.015
335-67-1	Perfluorooctanoic acid (PFOA)	0.023	U	0.029	0.023	0.0090
375-73-5	Perfluorobutanesulfonic acid (PFBS)	0.10	U	0.13	0.10	0.045

CAS NO.	SURROGATE	%REC	Q	LIMITS
STL00993	13C2 PFHxA	102		70-130
STL00996	13C2 PFDA	108		70-130

12/13/17



- Legend**
- 1 Mile Zone
 - Surface Water
 - Drainage Ditch
 - Suspected Source Area
 - Phase 1 Sampling Area
 - Phase 2 Sampling Area
 - Phase 3 Sampling Area
 - Base Boundary
 - Inferred Groundwater Flow Direction

Notes:

- Results shown on this figure are for PFOA and PFOS. See text and Table 2 for PFBS results
- * Both results above the LHA are from the same well; the second sample collected was a confirmation sample.




Figure 2
Results for Drinking Water Well Sampling
Near Ault Field
Naval Air Station Whidbey Island
Oak Harbor, Washington

For Official Use Only