N00174.AR.001959
NSWC INDIAN HEAD
5090.3a

VALIDATED DATA PACKAGE, 76380, NSWC INDIAN HEAD MD
6/26/2015
APPL LABS

908 North Temperance Ave. ∇ Clovis, CA 93611 ∇ Phone 559-275-2175 ∇ Fax 559-275-4422

Certification Number: CA1312
NELAP Certification number: CA00046
DoD-ELAP Certificate number: 74807

Summary Report

June 18, 2015

Tetra Tech
5700 Lake Wright Drive, Suite 309
Norfolk, Virginia 23502

Attn: Ed Corack

Title: Report of Data: Case 76380

Project: CTO JU11 112G02622 NSF Indian Head, MD

Contract #: Prime contract # for DoD: Navy CLEAN. N62470-08-D-1001
Subcontract: MSA # 1045497, Work Release # 08-JU11

Dear Mr. Corack:

Five soil samples were received May 8, 2015, at 3.0°C. Written results for the requested analyses are provided on this June 18, 2015.

Results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.

If you have any questions or require further information, please contact your APPL Project Manager, Cynthia Clark, cclark@applinc.com, at your convenience. Thank you for choosing APPL, Inc.

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. These test results meet all requirements of NELAC and DoD QSM. Release of the hard copy has been authorized by the Laboratory Manager or her designee, as verified by the following signature.

Sharon Dehmlow, QAO

for Sharon Dehmlow, Laboratory Director
APPL, Inc.

SD/rp
Enclosure
cc: File

Number of pages in this report: 75

Sample receipt information

ARF: 76380

Project: CTO JU11 112G02622 NSF Indian Head, MD

Sample Receipt Information:

The soil samples were received on May 8, 2015, at 3.0C. The samples were assigned Analytical Request Form (ARF) number 76380. The sample numbers and requested analyses were compared to the chain of custody. No exception was encountered.

Sample Table

CLIENT ID	APPL ID	Matrix	Date Sampled	Date Received
S67-SS030-0001	AZ16137	SOIL	5/7/2015	5/8/2015
S67-SS030P-0001	AZ16138	SOIL	5/7/2015	5/8/2015
S67-SB030-0607	AZ16139	SOIL	5/7/2015	5/8/2015
S67-SS032-0001	AZ16140	SOIL	5/7/2015	5/8/2015
S67-SB032-0607	AZ16141	SOIL	5/7/2015	5/8/2015

Percent moisture was determined using CLP 4.0.

The samples and blanks were screened for J-value responses between the detection limit (DL) and limit of quantitation (LOQ).

Laboratory control limits generated in house do not meet the control limits listed in DoD QSM 4.2 for all analytes. Laboratory control limits generated for this project meet all control limits listed in the DoD QSM 4.2 except where noted. A copy of our in house generated control limits is available upon request. In addition, a copy of our LOQ control limits, established using 7 data points, are also available upon request.

Only the portion of the injection log relative to these samples is included. A full sequence log is available upon request.

Measurement uncertainty can be reported upon request.

EPA Method 8270D
Semi-Volatile Organic Compounds

Sample Preparation and Analysis Information:

The soil samples were extracted according to EPA method 3550B. The samples were analyzed according to EPA Method 8270D using an Agilent GC/MS. All holding times were met.

Manual integration were performed in accordance with APPL's SOP. Phenol-D6 was manually integrated in the 100ug/mL standard. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance

Spike Recovery

A Laboratory Control Spike (LCS) was used for quality assurance. All LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blank.

Surrogates

Surrogate recoveries are summarized on Form 2 & 8. Sample S67-SB032-0607 recovered the surrogate 2-Fluorobiphenyl below the 45% lower control limit at 44.4%. The sample was not re extracted because the PAH recovery for this surrogate is 45.1% showing the low recovery is due to matrix. All other surrogate recoveries were within the control limits.

Calibration

The initial and continuing calibrations were performed according to the method. All method acceptance criteria were met.

Tuning:

The instrument was tuned using DFTPP. All method criteria were met.

Internal Standards

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8260. All method criteria were met.

Summary:

No other problem was encountered. All data generated are acceptable.

EPA Method 8270D-SIM
Polynuclear Aromatic Hydrocarbons

Sample Preparation and Analysis Information:

The soil samples were extracted according to EPA method 3550B. The samples were analyzed according to EPA Method 8270D using an Agilent GC/MS with selective ion monitoring. Samples S67-SS030-0001, S67-SS030P-0001 and S67-SS032-0001 were injected at a 10:1 dilution because of color and viscosity. Due to low surrogate recovery, sample S67-SB030-0607 was re-extracted, more than 14 days after collection. All other holding times were met.

Benzo(b)fluoranthene and Benzo(k)fluoranthene could not be chromatographically separated in all samples except S67-SB030-0607. As a result, each of these is reported from the sum of the two area counts and may be biased high.

Quality Control/Assurance

Spike Recovery

Laboratory Control Spikes (LCS) were used for quality assurance. The 150601A LCS for the re extract of sample S67-SB030-0607 recovered five compounds above their upper control limit. All other LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blanks.

Surrogates

Surrogate recoveries are summarized on Form 2 & 8. Samples S67-SS030-0001, S67-SS030P-0001, and S67-SS032-0001 recovered Nitrobenzene-D5 above the 100% upper control. The area counts for this standard were at the low end of the calibration curve. The samples were not re-extracted. Sample S67-SB030-0607 recovered 2-Fluorobiphenyl below the lower limit of 45% at 28.7%. The sample was re -extracted, outside of holding time but within two times the holding time. The re-extracted sample had a passing 2-FBP recovery. The two extracts had similar results. All other surrogate recoveries were within the control limits.

Calibration

The initial and continuing calibrations were performed according to the method. All method acceptance criteria were met.

Tuning:

The instrument was tuned using DFTPP. All method criteria were met.

Internal Standards

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8260. All method criteria were met.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

EPA Method 6850
Perchlorate by LC-Mass Spec

Sample Preparation and Analysis Information:

The soil samples were prepared according to the method. The samples were analyzed according to EPA Method 6850 using an Agilent 6460 Triple Quad LC/MS. The samples were prepared and analyzed within acceptable hold time.

Manual integrations were performed in accordance to APPL's SOP. Perchlorate was manually integrated in the calibration curve and in sample S67-SB032-0607. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance

Calibrations:

Calibrations were performed according to the method. All calibration acceptance criteria were met. The second source met acceptance criteria.

Blanks:

Perchlorate was not detected at or above one-half the limit of quantitation (LOQ) in the method blank.

Spikes:

Laboratory Control Spike (LCS) and Interference Check Sample (I.C.S.) were used for quality assurance. All LCS acceptance criteria were met..

The Interference Check Sample (I.C.S.) was prepared using a mixed anions solution. The ICS recovery met acceptance criteria.

Sample was S67-SS030-0001 selected by the laboratory for MS/MSD analysis. All spike recoveries were acceptable.

Internal Standards:

The area counts of the sample Internal Standards were compared to the average IS area counts of the initial calibration. The internal standard in the ICS recovered slightly below the lower control limit. All other internal standards were within the 50%D acceptance criteria.

Summary:

All data were acceptable. No other analytical exception is noted.

EPA Method 8330B
Energetics

Sample Preparation:

The samples were dried and extracted according to EPA method 8330B, without using incremental sampling procedures. All holding times were met.

Analysis:

The samples were analyzed according to EPA Method 8330B using an Agilent 1290 HPLC with DA detector. Manual integrations were performed in accordance with APPL's SOP. The following analytes were manually integrated in the calibration standards: HMX, Nitroglycerin, PETN, 1,2-Dinitrobenzene, 1,3- Dinitrobenzene, Nitrobenzene, 2,4,6-Trinitrotoluene, Tetryl, 2,6-Dinitrotoluene, 2,4-Dinitrotoluene, 2-Nitrotoluene, 4-Nitrotoluene, 3-Nitrotoluene, 4-Amino-2,6-Dinitrotoluene, 2-Amino-4,6-Dinitrotoluene and 3,5-Dinitroaniline. HMX was manually integrated in the LCS. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance:

Spike Recovery:

A Laboratory Control Spike (LCS) was used for quality assurance. All spike acceptance criteria was met.

Sample was S67-SS030-0001 selected by the laboratory for MS/MSD analysis. All spike recoveries were acceptable.

Method blanks:

No target analyte was detected at or above one-half the limit of quantitation (LOQ) in the method blank.

Surrogates:

Surrogate recoveries are summarized on the Form 2 & 8. All surrogates had acceptable recoveries.

Calibration:

The initial and continuing calibrations and second source were analyzed according to the method. All calibration criteria were met.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

EPA Method 6010C
Metals

Digestion and Analysis Information:

The soil samples were digested according to EPA method 3050B. The samples were analyzed for metals according to EPA method 6010C using a Perkin Elmer Optima 5300DV. All holding times were met.

Quality Control/Assurance:

Calibrations:

Calibrations were performed according to the methods for the initial calibration and the initial calibration verification. The initial calibration verification is prepared from a second source standard.

Blanks:

No target metal was detected above one half the LOQ in the method blank.

Spikes:

Laboratory Control Spike (LCS), Matrix Spikes (MS/MSD), Post-Digestion Spike (PDS), and a Dilution Test (DT) were used for quality assurance. All LCS acceptance criteria were met.

Sample S67-SS030-0001 was selected by the laboratory as the QC sample for the analytical batch. All target metals recovered outside the 80-120% control limits in the MS and/or MSD. Recoveries in the DT and PDS were acceptable.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

Inorganic Analyses

EPA Method 9045D and Walkley Black

Sample Preparation and Analysis

The samples were prepared and analyzed according to the methods. All holding times were met.

Quality Control/Assurance

Spike Recovery

A Laboratory Control Spike (LCS) was used for quality assurance. All LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected one-half the limit of quantitation (LOQ) in the method blank.

Calibration

Calibrations were performed according to the method for the initial calibration and the continuing calibration verification. The initial calibration verification is prepared from a second source standard. All calibration criteria were met.

Summary:

No analytical exception was noted.

CHEMTECH
CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. 76375 3.0
QUOTE NO.
COC Number 088451

CLIENT INFORMATION
REPORT TO BE SENT TO:
COMPANY: Tetra Tech
ADDRESS: 5700 Lake Wright Dr. St 309
CITY: Norfolk STATE: VA ZIP: 23502
ATTENTION: R Ed Corack
PHONE: 757-466-4908 FAX:

CLIENT PROJECT INFORMATION
PROJECT NAME: NSF-Indian Head Site 67
PROJECT NO.: 112G02622 (GO-JUL) LOCATION: Site 67
PROJECT MANAGER: Ed. Corack
e-mail: Ed.Corack@tetratech.com
PHONE: 757-466-4908 FAX:

CLIENT BILLING INFORMATION
BILL TO: See Contract PO#:
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION
FAX: _____ DAYS *
HARD COPY: _____ DAYS *
EDD: _____ DAYS *
PREAPPROVED TAT: ☐ YES ☐ NO
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION
☐ RESULTS ONLY ☐ USEPA CLP
☒ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☒ EDD FORMAT _____

ANALYSIS
1 Perchlorate (24 Hr TAT) 2 PAHs, Phthalates 3 Metals 4 Energetics 5 Perchlorate 6 pH 7 TOC 8 9

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
1.	S67-GW-E1-050715	GW		X	5-7-15	0920	1	E 1									
2.	S67-GW-F1-050715	GW		X	5-7-15	0950	1	E 1									Rush 24 HR TAT
3.	S67-SS030-0001	SO		X	5-7-15	1035	3		E		2		1				Rush 24 HR TAT
4.	S67-SS030P-0001	SO		X	5-7-15	1035	3		E		2		1				
5.	S67-SB030-0607	SO		X	5-7-15	1100	2		E		2		1		3		5-7-15
6.	S67-SS032-0001	SO		X	5-7-15	1420	3		E		2		1				
7.	S67-SB032-0607	SO		X	5-7-15	1440	2		E		2						
8.																	
9.																	
10.																	

5-7-15

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Jacob Birkett	5-7-15 1800	1. FedEx
RELINQUISHED BY:	DATE/TIME:	RECEIVED BY:
2.		2.
RELINQUISHED BY:	DATE/TIME:	RECEIVED FOR LAB BY:
3.	5/8/15 11:00	3. Young Jin

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant Cooler Temp. _____
MeOH extraction requires an additional 4 oz jar for percent solid.
Comments: Ice in Cooler?: _____

Page 1 of 1

SHIPPED VIA: CLIENT: ☐ HAND DELIVERED ☐ OVERNIGHT
CHEMTECH: ☐ PICKED UP ☐ OVERNIGHT
Shipment Complete: ☐ YES ☐ NO

Revision 8/2007

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	30	7.60	1.900	0.720	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	1.900 U	3.80	1.900	0.650	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	1.900 U	3.80	1.900	0.760	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	1.900 U	3.80	1.900	0.670	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	1.900 U	3.80	1.900	0.720	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	1.900 U	3.80	1.900	0.730	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	61.4	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	59.6	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	59.8	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	59.2	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	59.5	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	76.4	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y057
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 10
Initials: DA

Printed: 6/2/2015 12:11:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030P-0001

APPL ID: AZ16138

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	9.6	0.75	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.190 U	0.38	0.190	0.064	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.190 U	0.38	0.190	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.190 U	0.38	0.190	0.066	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.190 U	0.38	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.190 U	0.38	0.190	0.072	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	57.8	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	56.4	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	51.9	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	54.0	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	51.1	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	70.9	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y058
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 12:11:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB030-0607

APPL ID: AZ16139

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.200 U	0.80	0.200	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.200 U	0.40	0.200	0.068	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.200 U	0.40	0.200	0.080	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.200 U	0.40	0.200	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.200 U	0.40	0.200	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.200 U	0.40	0.200	0.076	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	65.6	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	47.5	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	54.0	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	55.9	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	57.9	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	87.8	30-125			%	5/21/2015	5/28/2015

Quant Method:	Y0520TT.M
Run #:	0520Y059
Instrument:	Yoda
Sequence:	Y150520
Dilution Factor:	1
Initials:	DA

Printed: 6/2/2015 12:11:46 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS032-0001

APPL ID: AZ16140

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.190 U	0.74	0.190	0.069	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.190 U	0.37	0.190	0.063	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.190 U	0.37	0.190	0.074	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.190 U	0.37	0.190	0.065	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.190 U	0.37	0.190	0.069	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.190 U	0.37	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	66.2	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	63.3	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	58.5	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	60.2	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	59.6	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	76.5	30-125			%	5/21/2015	5/28/2015

Quant Method:	Y0520TT.M
Run #:	0520Y060
Instrument:	Yoda
Sequence:	Y150520
Dilution Factor:	1
Initials:	DA

Printed: 6/2/2015 12:11:46 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB032-0607

APPL ID: AZ16141

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.220 U	0.86	0.220	0.081	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.220 U	0.43	0.220	0.073	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.220 U	0.43	0.220	0.086	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.220 U	0.43	0.220	0.076	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.220 U	0.43	0.220	0.081	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.220 U	0.43	0.220	0.082	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	62.7	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	44.4 #	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	60.6	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	59.8	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	62.4	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	72.0	30-125			%	5/21/2015	5/28/2015

= Recovery (or RPD) is outside QC limits.

Quant Method:	Y0520TT.M
Run #:	0520Y061
Instrument:	Yoda
Sequence:	Y150520
Dilution Factor:	1
Initials:	DA

Printed: 6/2/2015 12:11:46 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Method Blank
EPA 8270D SOILS

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150521S-16137 - 197602**
Batch ID: #87DJU-150521A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	BIS (2-ETHYLHEXYL) PHTHALAT	0.167 U	0.66	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	BUTYL BENZYL PHTHALATE	0.167 U	0.33	0.167	0.056	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-BUTYL PHTHALATE	0.167 U	0.33	0.167	0.066	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-OCTYL PHTHALATE	0.167 U	0.33	0.167	0.058	mg/kg	5/21/2015	5/28/2015
BLANK	DIETHYL PHTHALATE	0.167 U	0.33	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	DIMETHYL PHTHALATE	0.167 U	0.33	0.167	0.063	mg/kg	5/21/2015	5/28/2015
BLANK	SURROGATE: 2,4,6-TRIBROMOP	66.1	35-125			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	66.9	45-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUOROPHENO	70.2	35-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: NITROBENZENE-	67.7	35-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: PHENOL (S)	69.3	40-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: TERPHENYL-D14 (	86.8	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y055
Instrument: Yoda
Sequence: Y150520
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 12:12:10 PM

Laboratory Control Spike Recovery
EPA 8270D SOILS

APPL ID: **150521S-16137 LCS - 197602**
Batch ID: #87DJU-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
BIS (2-ETHYLHEXYL) PHTHALATE	1.67	1.35	80.8	45-125
BUTYL BENZYL PHTHALATE	1.67	1.31	78.4	50-125
DI-N-BUTYL PHTHALATE	1.67	1.30	77.8	55-110
DI-N-OCTYL PHTHALATE	1.67	1.39	83.2	40-130
DIETHYL PHTHALATE	1.67	1.19	71.3	50-115
DIMETHYL PHTHALATE	1.67	1.19	71.3	50-110
SURROGATE: 2,4,6-TRIBROMOPHENOL	6.67	5.07	76.0	35-125
SURROGATE: 2-FLUORBIPHENYL (S)	3.33	2.27	68.2	45-105
SURROGATE: 2-FLUOROPHENOL (S)	6.67	4.46	66.9	35-105
SURROGATE: NITROBENZENE-D5 (S)	3.33	2.22	66.7	35-100
SURROGATE: PHENOL (S)	6.67	4.66	69.9	40-100
SURROGATE: TERPHENYL-D14 (S)	3.33	2.98	89.5	30-125

Comments: _____

Primary	SPK
Quant Method :	Y0520TT.M
Extraction Date :	5/21/2015
Analysis Date :	5/28/2015
Instrument :	Yoda
Run :	0520Y056
Initials :	DA

Printed: 6/2/2015 12:11:55 PM
APPL Standard LCS

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: 2,4,6-TRIBROMOPHENOL (S)			SURROGATE: 2-FLUORBIPHENYL (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	35-125	66.1		45-105	66.9	
150521A-LCS	Lab Control Spike	35-125	76.0		45-105	68.2	
AZ16137	S67-SS030-0001	35-125	61.4		45-105	59.6	
AZ16138	S67-SS030P-0001	35-125	57.8		45-105	56.4	
AZ16139	S67-SB030-0607	35-125	65.6		45-105	47.5	
AZ16140	S67-SS032-0001	35-125	66.2		45-105	63.3	
AZ16141	S67-SB032-0607	35-125	62.7		45-105	44.4	#

Comments: Batch: #87DJU-150521A

= Recovery outside of Control Limits on Sample.

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: 2-FLUOROPHENOL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	35-105	70.2		35-100	67.7	
150521A-LCS	Lab Control Spike	35-105	66.9		35-100	66.7	
AZ16137	S67-SS030-0001	35-105	59.8		35-100	59.2	
AZ16138	S67-SS030P-0001	35-105	51.9		35-100	54.0	
AZ16139	S67-SB030-0607	35-105	54.0		35-100	55.9	
AZ16140	S67-SS032-0001	35-105	58.5		35-100	60.2	
AZ16141	S67-SB032-0607	35-105	60.6		35-100	59.8	

Comments: Batch: #87DJU-150521A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: PHENOL (S)			SURROGATE: TERPHENYL-D14 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	40-100	69.3		30-125	86.8	
150521A-LCS	Lab Control Spike	40-100	69.9		30-125	89.5	
AZ16137	S67-SS030-0001	40-100	59.5		30-125	76.4	
AZ16138	S67-SS030P-0001	40-100	51.1		30-125	70.9	
AZ16139	S67-SB030-0607	40-100	57.9		30-125	87.8	
AZ16140	S67-SS032-0001	40-100	59.6		30-125	76.5	
AZ16141	S67-SB032-0607	40-100	62.4		30-125	72.0	

Comments: Batch: #87DJU-150521A

EPA 8270D

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

Blank ID: 150521A-BLK Time Analyzed: 1449

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150521A-BLK	Blank	0520Y055	5/28/2015 1449
150521A-LCS	Lab Control Spike	0520Y056	5/28/2015 1518
AZ16137	S67-SS030-0001	0520Y057	5/28/2015 1547
AZ16138	S67-SS030P-0001	0520Y058	5/28/2015 1617
AZ16139	S67-SB030-0607	0520Y059	5/28/2015 1646
AZ16140	S67-SS032-0001	0520Y060	5/28/2015 1716
AZ16141	S67-SB032-0607	0520Y061	5/28/2015 1746

Comments: Batch: #87DJU-150521A

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.033 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0200 U	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.11	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.038 J	0.058	0.0200	0.0092	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.13	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.051 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.14	0.058	0.0200	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.076	0.058	0.0200	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.33	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.12	0.058	0.0200	0.0092	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.061	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.18	0.058	0.0200	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.034 J	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.16	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.012 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.11	0.058	0.0200	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.20	0.058	0.0200	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	68.8	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	117 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	87.0	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method:	P0518-42.M
Run #:	0515L237
Instrument:	Linus
Sequence:	L150515
Dilution Factor:	10
Initials:	DA

Printed: 6/2/2015 2:14:16 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030P-0001

APPL ID: AZ16138

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.032 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0190 U	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.11	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.033 J	0.057	0.0190	0.0091	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.12	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.042 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.11	0.057	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.073	0.057	0.0190	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.30	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.090	0.057	0.0190	0.0091	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.057	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.14	0.057	0.0190	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.033 J	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.14	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.011 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.087	0.057	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.16	0.057	0.0190	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	68.1	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	116 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	83.7	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method:	P0518-42.M
Run #:	0515L238
Instrument:	Linus
Sequence:	L150515
Dilution Factor:	10
Initials:	DA

Printed: 6/2/2015 2:14:16 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB030-0607

APPL ID: AZ16139

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0021 U	0.006	0.0021	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.0015 J	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	28.7 #	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	48.4	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	83.6	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method:	P0518-42.M
Run #:	0515L239
Instrument:	Linus
Sequence:	L150515
Dilution Factor:	1
Initials:	DA

Printed: 6/2/2015 2:14:16 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D LL SOILS RE-EXTRACT

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16139
QCG: #SIMDD-150601A-197777

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	ACENAPHTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	ACENAPHTHYLENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	ANTHRACENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (A) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0021 U	0.006	0.0021	0.0016	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	CHRYSENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	6/1/2015	6/1/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	FLUORANTHENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	6/1/2015	6/1/2015
8270D-LL	FLUORENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	NAPHTHALENE	0.0024 J	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	PHENANTHRENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	6/1/2015	6/1/2015
8270D-LL	PYRENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	6/1/2015	6/1/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	57.7	45-105			%	6/1/2015	6/1/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	57.6	35-100			%	6/1/2015	6/1/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	90.0	30-125			%	6/1/2015	6/1/2015

J = Estimated value.

Quant Method: P0518-42.M
Run #: 0515L264
Instrument: Linus
Sequence: L150515
Dilution Factor: 1
Initials: DA

Printed: 6/5/2015 2:20:03 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS032-0001

APPL ID: AZ16140

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.080	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.067	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.21	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.33	0.056	0.0190	0.0089	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.83	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.44	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	1.3	0.056	0.0190	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.40	0.056	0.0190	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	1.5	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.86	0.056	0.0190	0.0089	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.16	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	1.4	0.056	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.12	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.47	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.091	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	1.1	0.056	0.0190	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	1.3	0.056	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	53.8	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	112 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	72.0	30-125			%	5/21/2015	5/27/2015

= Recovery (or RPD) is outside QC limits.

Quant Method: P0518-42.M
Run #: 0515L240
Instrument: Linus
Sequence: L150515
Dilution Factor: 10
Initials: DA

Printed: 6/2/2015 2:14:16 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB032-0607

APPL ID: AZ16141

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0034 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0022 U	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.012	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.0025 J	0.006	0.0022	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0054 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.0022 U	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0022 U	0.006	0.0022	0.0014	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0022 U	0.006	0.0022	0.0017	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.019	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.0032 J	0.006	0.0022	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0022 U	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.0022 U	0.006	0.0022	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.0022 U	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.012	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.0021 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.0028 J	0.006	0.0022	0.0014	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.0053 J	0.006	0.0022	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	45.1	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	52.1	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	86.8	30-125			%	5/21/2015	5/27/2015

J = Estimated value.

Quant Method: P0518-42.M
Run #: 0515L241
Instrument: Linus
Sequence: L150515
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 2:14:16 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Method Blank
EPA 8270D LL SOILS

Blank Name/QCG: **150521S-16137 - 197621**
Batch ID: #SIMDD-150521A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	63.5	45-105			%	5/21/2015	5/27/2015
BLANK	SURROGATE: NITROBENZENE-	44.2	35-100			%	5/21/2015	5/27/2015
BLANK	SURROGATE: TERPHENYL-D14 (	106	30-125			%	5/21/2015	5/27/2015

Quant Method: P0518-42.M
Run #: 0515L235
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 2:14:38 PM

Method Blank

EPA 8270D LL SOILS RE-EXTRACT

Blank Name/QCG: **150601S-16139 - 197777**
Batch ID: #SIMDD-150601A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	77.1	45-105			%	6/1/2015	6/1/2015
BLANK	SURROGATE: NITROBENZENE-	68.4	35-100			%	6/1/2015	6/1/2015
BLANK	SURROGATE: TERPHENYL-D14	112	30-125			%	6/1/2015	6/1/2015

Quant Method: P0518-42.M
Run #: 0515L262
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/5/2015 2:20:33 PM

Laboratory Control Spike Recovery

EPA 8270D LL SOILS

APPL ID: **150521S-16137 LCS - 197621**
Batch ID: #SIMDD-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.133	0.0883	66.4	45-105
ACENAPHTHENE	0.133	0.127	95.5	45-110
ACENAPHTHYLENE	0.133	0.112	84.2	45-105
ANTHRACENE	0.133	0.126	94.7	55-105
BENZ (A) ANTHRACENE	0.133	0.139	105	50-110
BENZO (A) PYRENE	0.133	0.132	99.2	50-110
BENZO (B) FLUORANTHENE	0.133	0.140	105	45-115
BENZO (G,H,I) PERYLENE	0.133	0.135	102	40-125
BENZO (K) FLUORANTHENE	0.133	0.132	99.2	45-125
CHRYSENE	0.133	0.144	108	55-110
DIBENZ (A,H) ANTHRACENE	0.133	0.136	102	40-125
FLUORANTHENE	0.133	0.127	95.5	55-115
FLUORENE	0.133	0.120	90.2	50-110
INDENO (1,2,3-CD) PYRENE	0.133	0.145	109	40-120
NAPHTHALENE	0.133	0.0797	59.9	40-105
PHENANTHRENE	0.133	0.120	90.2	50-110
PYRENE	0.133	0.146	110	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0667	0.043	64.5	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0667	0.0313	46.9	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0667	0.0711	107	30-125

Comments: _____

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	5/21/2015
Analysis Date :	5/27/2015
Instrument :	Linus
Run :	0515L236
Initials :	DA

Printed: 6/2/2015 2:14:23 PM
APPL Standard LCS

Laboratory Control Spike Recovery
EPA 8270D LL SOILS RE-EXTRACT

APPL ID: **150601S-16139 LCS - 197777**
Batch ID: #SIMDD-150601A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.167	0.156	93.4	45-105
ACENAPHTHENE	0.167	0.178	107	45-110
ACENAPHTHYLENE	0.167	0.164	98.2	45-105
ANTHRACENE	0.167	0.170	102	55-105
BENZ (A) ANTHRACENE	0.167	0.192	115 #	50-110
BENZO (A) PYRENE	0.167	0.199	119 #	50-110
BENZO (B) FLUORANTHENE	0.167	0.205	123 #	45-115
BENZO (G,H,I) PERYLENE	0.167	0.194	116	40-125
BENZO (K) FLUORANTHENE	0.167	0.198	119	45-125
CHRYSENE	0.167	0.199	119 #	55-110
DIBENZ (A,H) ANTHRACENE	0.167	0.195	117	40-125
FLUORANTHENE	0.167	0.174	104	55-115
FLUORENE	0.167	0.165	98.8	50-110
INDENO (1,2,3-CD) PYRENE	0.167	0.189	113	40-120
NAPHTHALENE	0.167	0.179	107 #	40-105
PHENANTHRENE	0.167	0.165	98.8	50-110
PYRENE	0.167	0.207	124	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0828	0.0665	80.3	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0828	0.0577	69.7	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0828	0.0985	119	30-125

= Recovery is outside QC limits.

Comments: _____

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	6/1/2015
Analysis Date :	6/1/2015
Instrument :	Linus
Run :	0515L263
Initials :	DA

Printed: 6/5/2015 2:20:19 PM
APPL Standard LCS

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: 2-FLUORBIPHENYL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	45-105	63.5		35-100	44.2	
150521A-LCS	Lab Control Spike	45-105	64.5		35-100	46.9	
AZ16137	S67-SS030-0001	45-105	68.8		35-100	117	#
AZ16138	S67-SS030P-0001	45-105	68.1		35-100	116	#
AZ16139	S67-SB030-0607	45-105	28.7	#	35-100	48.4	
AZ16140	S67-SS032-0001	45-105	53.8		35-100	112	#
AZ16141	S67-SB032-0607	45-105	45.1		35-100	52.1	

Comments: Batch: #SIMDD-150521A

= Recovery outside of Control Limits on Sample.

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: TERPHENYL-D14 (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	30-125	106				
150521A-LCS	Lab Control Spike	30-125	107				
AZ16137	S67-SS030-0001	30-125	87.0				
AZ16138	S67-SS030P-0001	30-125	83.7				
AZ16139	S67-SB030-0607	30-125	83.6				
AZ16140	S67-SS032-0001	30-125	72.0				
AZ16141	S67-SB032-0607	30-125	86.8				

Comments: Batch: #SIMDD-150521A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: 2-FLUORBIPHENYL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150601A-BLK	Blank	45-105	77.1		35-100	68.4	
150601A-LCS	Lab Control Spike	45-105	80.3		35-100	69.7	
AZ16139	S67-SB030-0607	45-105	57.7		35-100	57.6	

Comments: Batch: #SIMDD-150601A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: TERPHENYL-D14 (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150601A-BLK	Blank	30-125	112				
150601A-LCS	Lab Control Spike	30-125	119				
AZ16139	S67-SB030-0607	30-125	90.0				

Comments: Batch: #SIMDD-150601A

8270D-LL

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

Blank ID: 150521A-BLK Time Analyzed: 1917

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150521A-BLK	Blank	0515L235	5/27/2015 1917
150521A-LCS	Lab Control Spike	0515L236	5/27/2015 1942
AZ16137	S67-SS030-0001	0515L237	5/27/2015 2008
AZ16138	S67-SS030P-0001	0515L238	5/27/2015 2033
AZ16139	S67-SB030-0607	0515L239	5/27/2015 2059
AZ16140	S67-SS032-0001	0515L240	5/27/2015 2124
AZ16141	S67-SB032-0607	0515L241	5/27/2015 2150

Comments: Batch: #SIMDD-150521A

8270D-LL

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

Blank ID: 150601A-BLK Time Analyzed: 1600

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150601A-BLK	Blank	0515L262	6/1/2015 1600
150601A-LCS	Lab Control Spike	0515L263	6/1/2015 1625
AZ16139	S67-SB030-0607	0515L264	6/1/2015 1651

Comments: Batch: #SIMDD-150601A

Printed: 6/5/2015 2:20:07 PM

Form 4, Blank Summary

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16137
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 13.4 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_050.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/18/15 12:51:21 PM
APPL-F1-SC-MCResOnly-REG MDLs

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16138
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 12.0 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_053.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/18/15 12:51:21 PM
APPL-F1-SC-MCResOnly-REG MDLs

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16139
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 17.5 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_054.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/18/15 12:51:21 PM
APPL-F1-SC-MCResOnly-REG MDLs

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16140
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 10.4 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.022	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_055.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/18/15 12:51:21 PM
APPL-F1-SC-MCResOnly-REG MDLs

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16141
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 23.3 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.013	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_060.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/18/15 12:51:21 PM
APPL-F1-SC-MCResOnly-REG MDLs

Method Blank

PERCHLORATE EPA 6850 - SOIL

Blank Name/QCG: **150518S-16137 - 198067**
Batch ID: #6850SM-150518A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method: QTLMFL2
Run #: T523_049.d
Instrument: AGIL_6460
Sequence: TQ052315
Initials: RP

GC SC-Blank-REG MDLs
Printed: 06/18/15 12:51:24 PM

Laboratory Control Spike Recovery
PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 LCS - 198067**
Batch ID: #6850SM-150518A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits
PERCHLORATE	0.00594	0.00556	93.6	80-120

Comments: _____

Primary	SPK
Quant Method :	QTLMFL2
Extraction Date :	05/18/15
Analysis Date :	05/24/15
Instrument :	AGIL_6460
Run :	T523_047.d
Initials :	RP

Printed: 06/18/15 12:51:26 PM
APPL Standard LCS

Matrix Spike Recoveries

PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 MS - 198067**
Batch ID: #6850SM-150518A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Lvl mg/Kg	Matrix Result mg/Kg	SPK Result mg/Kg	DUP Result mg/Kg	SPK % Recovery	DUP % Recovery	Recovery Limits	RPD %	RPD Limits
PERCHLORATE	0.00508	ND	0.00496	0.00529	97.6	104	80-120	6.4	15

Comments: _____

Primary	SPK	DUP
Quant Method :	QTLMFL2	QTLMFL2
Extraction Date :	05/18/15	05/18/15
Analysis Date :	05/24/15	05/24/15
Instrument :	AGIL_6460	AGIL_6460
Run :	T523_051.d	T523_052.d
Initials :	RP	

Printed: 06/18/15 12:51:29 PM
APPL MSD SCII

EPA 6850

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/24/15

Matrix: SOIL Instrument: AGIL_6460

Blank ID: 150518A-BLK Time Analyzed: 0628

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150518A-LCS	Lab Control Spike	T523_047.d	05/24/15 0550
150518A-BLK	Blank	T523_049.d	05/24/15 0628
AZ16137	S67-SS030-0001	T523_050.d	05/24/15 0646
150518A-MS	Matrix Spike	T523_051.d	05/24/15 0705
150518A-MSD	Matrix SpikeD	T523_052.d	05/24/15 0724
AZ16138	S67-SS030P-0001	T523_053.d	05/24/15 0742
AZ16139	S67-SB030-0607	T523_054.d	05/24/15 0801
AZ16140	S67-SS032-0001	T523_055.d	05/24/15 0819
AZ16141	S67-SB032-0607	T523_060.d	05/24/15 0953

Comments: Batch: #6850SM-150518A

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16137
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	96.4	70-130			%	05/20/15	05/23/15

Quant Method:	W150522.M
Run #:	0522_000052
Instrument:	Waldorf
Sequence:	150522
Dilution Factor:	1
Initials:	MP

Printed: 06/02/15 2:48:09 PM
APPL-F1-SC-NoMC-REG MDLs

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16138
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	97.6	70-130			%	05/20/15	05/23/15

Quant Method:	W150522.M
Run #:	0522_000053
Instrument:	Waldorf
Sequence:	150522
Dilution Factor:	1
Initials:	MP

Printed: 06/02/15 2:48:09 PM
APPL-F1-SC-NoMC-REG MDLs

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16139
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	96.2	70-130			%	05/20/15	05/23/15

Quant Method:	W150522.M
Run #:	0522_000054
Instrument:	Waldorf
Sequence:	150522
Dilution Factor:	1
Initials:	MP

Printed: 06/02/15 2:48:09 PM
APPL-F1-SC-NoMC-REG MDLs

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16140
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	95.5	70-130			%	05/20/15	05/23/15

Quant Method:	W150522.M
Run #:	0522_000055
Instrument:	Waldorf
Sequence:	150522
Dilution Factor:	1
Initials:	MP

Printed: 06/02/15 2:48:09 PM
APPL-F1-SC-NoMC-REG MDLs

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16141
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	93.9	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000056
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 06/02/15 2:48:09 PM
APPL-F1-SC-NoMC-REG MDLs

Method Blank
EPA 8330B SOIL

Blank Name/QCG: **150520S-16137 - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
BLANK	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
BLANK	SURROGATE: 1,2-DINITROBENZ	95.0	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_00005
Instrument: Waldorf
Sequence: 150522
Initials: MP

GC SC-Blank-REG MDLs
Printed: 06/02/15 2:47:41 PM

Laboratory Control Spike Recovery
EPA 8330B SOIL

APPL ID: **150520S-16137 LCS - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2,4-DINITROTOLUENE	1.99	1.83	92.0	80-125
2,6-DINITROTOLUENE	1.99	1.84	92.5	80-120
HMX	1.99	1.90	95.5	75-125
NITROGLYCERIN	1.99	1.83	92.0	68-131
RDX	1.99	1.83	92.0	70-135
TETRYL	1.99	1.97	99.0	10-150
SURROGATE: 1,2-DINITROBENZENE (S	1.99	1.82	91.5	70-130

Comments: _____

Primary	SPK
Quant Method :	W150522.M
Extraction Date :	05/20/15
Analysis Date :	05/23/15
Instrument :	Waldorf
Run :	0522_000048
Initials :	MP

Printed: 06/02/15 2:47:43 PM
APPL Standard LCS

Matrix Spike Recoveries

EPA 8330B SOIL

APPL ID: **150520S-16137 MS - 197310**
Batch ID: #83BJU-150520A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	SPK Lvl	DUP Lvl	Matrix Result	SPK Result	DUP Result	SPK %	DUP %	Recovery	RPD	RPD
	mg/kg	mg/kg	mg/kg	mg/kg	mg/kg	Recovery	Recovery	Limits	%	Limits
2,4-DINITROTOLUENE	1.99	2.00	ND	1.86	1.85	93.5	92.5	80-125	1.1	20
2,6-DINITROTOLUENE	1.99	2.00	ND	1.87	1.86	94.0	93.0	80-120	1.1	20
HMX	1.99	2.00	ND	1.85	1.83	93.0	91.5	75-125	1.6	20
NITROGLYCERIN	1.99	2.00	ND	1.90	1.93	95.5	96.5	68-131	1	20
RDX	1.99	2.00	ND	1.69	1.69	84.9	84.5	70-135	0.47	20
TETRYL	1.99	2.00	ND	1.77	1.81	88.9	90.5	10-150	1.8	20
SURROGATE: 1,2-DINITROBENZEN	1.99	2.00	NA	1.86	1.86	93.5	93.0	70-130		

Comments: _____

Primary	SPK	DUP
Quant Method :	W150522.M	W150522.M
Extraction Date :	05/20/15	05/20/15
Analysis Date :	05/23/15	05/23/15
Instrument :	Waldorf	Waldorf
Run :	0522_000049	0522_000050
Initials :	MP	

Printed: 06/02/15 2:48:00 PM
APPL MSD SCII

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/23/15

Matrix: SOIL Instrument: Waldorf

APPL ID.	Client Sample No.	SURROGATE: 1,2-DINITROBENZENE (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150520A-LCS	Lab Control Spike	70-130	91.5				
AZ16137-MS	Matrix Spike	70-130	93.5				
AZ16137-MSD	Matrix SpikeD	70-130	93.5				
150520A-BLK	Blank	70-130	95.0				
AZ16137	S67-SS030-0001	70-130	96.4				
AZ16138	S67-SS030P-0001	70-130	97.6				
AZ16139	S67-SB030-0607	70-130	96.2				
AZ16140	S67-SS032-0001	70-130	95.5				
AZ16141	S67-SB032-0607	70-130	93.9				

Comments: Batch: #83BJU-150520A

EPA 8330B

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/23/15

Matrix: SOIL Instrument: Waldorf

Blank ID: 150520A-BLK Time Analyzed: 1225

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150520A-LCS	Lab Control Spike	0522_000048	05/23/15 1033
150520A-MS	Matrix Spike	0522_000049	05/23/15 1110
150520A-MSD	Matrix SpikeD	0522_000050	05/23/15 1148
150520A-BLK	Blank	0522_000051	05/23/15 1225
AZ16137	S67-SS030-0001	0522_000052	05/23/15 1303
AZ16138	S67-SS030P-0001	0522_000053	05/23/15 1340
AZ16139	S67-SB030-0607	0522_000054	05/23/15 1418
AZ16140	S67-SS032-0001	0522_000055	05/23/15 1455
AZ16141	S67-SB032-0607	0522_000056	05/23/15 1533

Comments: Batch: #83BJU-150520A

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 05/07/15

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	5640	1150.0	92.00	46.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	BORON (B)	92.00U	115.0	92.00	36.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	12.000U	15.00	12.000	5.800	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	539	185.0	92.00	27.00	mg/Kg	20	06/04/15	06/12/15

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16138

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	3190	57.0	4.50	2.30	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.50U	5.7	4.50	1.80	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	91.00U	114.0	91.00	35.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	3.6	0.76	0.570	0.280	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	593	182.0	91.00	26.00	mg/Kg	20	06/04/15	06/12/15

Printed: 06/16/15 1:57:37 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
ARF: 76380
Sample ID: S67-SB030-0607
APPL ID: AZ16139
Sample Collection Date: 05/07/15

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	9150	1210.0	97.00	48.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.8J	6.1	4.80	1.90	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	9.6	0.81	0.610	0.300	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	21.5	9.7	4.80	1.40	mg/Kg	1	06/04/15	06/12/15

J = Estimated value.

Printed: 06/16/15 1:57:37 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS032-0001

APPL ID: AZ16140

Sample Collection Date: 05/07/15

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	2330	56.0	4.50	2.20	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.50U	5.6	4.50	1.70	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	2.7	0.75	0.560	0.280	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	150	8.9	4.50	1.30	mg/Kg	1	06/04/15	06/12/15

Printed: 06/16/15 1:57:37 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16141

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	1710	65.0	5.20	2.60	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	6.6	6.5	5.20	2.00	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	1.7	0.87	0.650	0.330	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	179	10.0	5.20	1.50	mg/Kg	1	06/04/15	06/12/15

Printed: 06/16/15 1:57:37 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

METALS BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
6010C	ALUMINUM (AL)	4.00 U	50.0	4.00	1.98	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	BORON (B)	4.00 U	5.0	4.00	1.55	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	LITHIUM (LI)	0.500 U	0.67	0.500	0.250	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	ZINC (ZN)	4.00 U	8.0	4.00	1.15	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Metals SC-Blank-REG MDLs
Printed: 06/16/15 1:57:49 PM

Laboratory Control Spike Recovery

METALS

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
EPA 6010C	ALUMINUM (AL)	200	186	93.0	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	BORON (B)	25.0	23.8	95.2	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	LITHIUM (LI)	10.00	10.2	102	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	ZINC (ZN)	50.0	46.7	93.4	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Comments:

Matrix Spike Recoveries

METALS

APPL ID: **150604S-16137 MS - 198024**

Sample ID: AZ16137

Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Lvl mg/Kg	Matrix Res mg/Kg	SPK Res mg/Kg	DUP Res mg/Kg	SPK % Recovery	DUP % Recovery	RPD	RPD Max	Recovery Limits	Extract Date-Spk	Analysis Date-Spk	Extract Date-Dup	Analysis Date-Dup	QC Group	QC Sample
EPA 6010C	ALUMINUM (AL)	396	4880	5220	5820	85.9	237 #	10.9	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	BORON (B)	49.5	12.7	48.0	49.5	71.3 #	74.3 #	3.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	LITHIUM (LI)	19.8	3.7	18.7	19.1	75.8 #	77.8 #	2.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	ZINC (ZN)	99.0	467	649	585	184 #	119	10.4	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137

= Recovery is outside QC limits.

Comments: _____

6010C/3050B^

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 06/12/15

Matrix: SOIL Instrument: Jules

Blank ID: 150604A-BLK Time Analyzed: 1426

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150604A-MSD	Matrix SpikeD	150612A	06/12/15 1655
150604A-MS	Matrix Spike	150612A	06/12/15 1648
150604A-LCS	Lab Control Spike	150612A	06/12/15 1432
AZ16139	S67-SB030-0607	150612A	06/12/15 1723
AZ16138	S67-SS030P-0001	150612A	06/12/15 1716
150604A-BLK	Blank	150612A	06/12/15 1426
AZ16141	S67-SB032-0607	150612A	06/12/15 1954
AZ16140	S67-SS032-0001	150612A	06/12/15 1947
AZ16139	S67-SB030-0607	150612A	06/12/15 1939
AZ16138	S67-SS030P-0001	150612A	06/12/15 1933
AZ16137	S67-SS030-0001	150612A	06/12/15 1641

Comments: Batch: #61CJU-150604A

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16137

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	11000	230.0	170.00	120.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/18/15 1:03:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030-0001
Sample Collection Date: 05/07/15

APPL ID: AZ16137
ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.66@21.4C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/18/15 12:54:46 PM

APPL-F1-SC-NoMC-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

APPL ID: AZ16138
ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	8200	230.0	170.00	110.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/18/15 1:03:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

APPL ID: AZ16138
ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.7@21.3C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/18/15 12:54:46 PM

APPL-F1-SC-NoMC-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS032-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16140

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	36000	220.0	170.00	110.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/18/15 1:03:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

APPL ID: AZ16140
ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.34@21.6C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/18/15 12:54:46 PM

APPL-F1-SC-NoMC-REG MDLs

WETLAB BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC C	150.00 U	200.0	150.00	100.00	mg/Kg	06/03/15	06/03/15	TOC-WB-150603A-AZ16137

Wetlab SC-Blank-REG MDLs
Printed: 06/18/15 1:03:49 PM

Laboratory Control Spike Recovery

WETLAB

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC CARBON	2000	1810	90.5	80-120	06/03/15	06/03/15	#TOC-WB-150603A-AZ16137

Comments: _____

WETLAB

Sample/Sample Duplicate Results

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: AZ16138
Client ID: S67-SS030P-0001

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

ARF: 76380

Method	Analyte	Sample ID	Sample Result	Sample Dup Result	RPD	RPD Max	MDL	PQL	Units	Sample Extract Date	Sample Analysis Date	Sample Dup Extract Date	Sample Dup Analysis Date
EPA 9045D	PH	AZ16138	7.7@21.3C	7.76@21.3C	0.8	20		NA	pH Units	05/08/15	05/08/15	05/08/15	05/08/15

Printed: 06/18/15 12:54:39 PM

Dup-SCII (NoMC)

WalkBlk

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 06/03/15

Matrix: SOIL Instrument: manual

Blank ID: WB-150603A-BLK Time Analyzed: 1100

APPL ID.	Client Sample No.	File ID.	Date Analyzed
AZ16140	S67-SS032-0001		06/03/15 1118
AZ16138	S67-SS030P-0001		06/03/15 1115
AZ16137	S67-SS030-0001		06/03/15 1106
WB-150603A-LCS	Lab Control Spike		06/03/15 1103
WB-150603A-BLK	Blank		06/03/15 1100

Comments: Batch: #TOC-WB-150603A

Printed: 06/18/15 1:03:34 PM

Form 4, Blank Summary

EPA 9045D

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed:

Matrix: SOIL Instrument: Tiamo

Blank ID: Time Analyzed:

APPL ID.	Client Sample No.	File ID.	Date Analyzed
AZ16137	S67-SS030-0001	3	05/08/15 1333
150508c-DUP	Duplicate	4	05/08/15 1342
AZ16138	S67-SS030P-0001	4	05/08/15 1339
AZ16140	S67-SS032-0001	6	05/08/15 1346

Comments: Batch: #9045D-150508c

908 North Temperance Ave. ∇ Clovis, CA 93611 ∇ Phone 559-275-2175 ∇ Fax 559-275-4422

Certification Number: CA1312
NELAP Certification number: CA00046
DoD-ELAP Certificate number: 74807

Data Validatable Report

June 18, 2015

Tetra Tech
5700 Lake Wright Drive, Suite 309
Norfolk, Virginia 23502

Attn: Ed Corack

Title: Report of Data: Case 76380

Project: CTO JU11 112G02622 NSF Indian Head, MD

Contract #: Prime contract # for DoD: Navy CLEAN. N62470-08-D-1001
Subcontract: MSA # 1045497, Work Release # 08-JU11

Dear Mr. Corack:

Five soil samples were received May 8, 2015, at 3.0°C. Written results for the requested analyses are provided on this June 18, 2015.

Results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.

If you have any questions or require further information, please contact your APPL Project Manager, Cynthia Clark, cclark@applinc.com, at your convenience. Thank you for choosing APPL, Inc.

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. These test results meet all requirements of NELAC and DoD QSM. Release of the hard copy has been authorized by the Laboratory Manager or her designee, as verified by the following signature.

Frances Ledain, QAO

for Sharon Dehmlow, Laboratory Director
APPL, Inc.

SD/rp
Enclosure
cc: File

Number of pages in this report: 753

Data Validation Package

for

CTO JU11 112G02622 NSF Indian Head

ARF 76380

TABLE OF CONTENTS

LABORATORY NAME: *APPL, Inc.*

Sample receipt information	4
Case Narrative	6
Chain of Custody and ARF	14
EPA Method 8270D	20
QC Summary	21
Sample Data	32
Calibration Data	52
Raw Data	78
EPA Method 8270D-SIM	97
QC Summary	98
Sample Data	114
Calibration Data	168
Raw Data	195
EPA Method 6850	221
QC Summary	222
Sample Data	243
Calibration Data	261
Raw Data	321

EPA Method 8330B 360
QC Summary 361
Sample Data 367
Calibration Data 385
Raw Data 600
EPA Method 6010C 648
QC Summary 649
Sample Data 654
Calibration Data 667
Raw Data 715
Inorganic Analyses 726
QC Summary 727
Sample Data 734
Calibration Data 742
Raw Data 745

SAMPLE RECEIPT INFORMATION

Sample receipt information

ARF: 76380

Project: CTO JU11 112G02622 NSF Indian Head, MD

Sample Receipt Information:

The soil samples were received on May 8, 2015, at 3.0C. The samples were assigned Analytical Request Form (ARF) number 76380. The sample numbers and requested analyses were compared to the chain of custody. No exception was encountered.

Sample Table

CLIENT ID	APPL ID	Matrix	Date Sampled	Date Received
S67-SS030-0001	AZ16137	SOIL	5/7/2015	5/8/2015
S67-SS030P-0001	AZ16138	SOIL	5/7/2015	5/8/2015
S67-SB030-0607	AZ16139	SOIL	5/7/2015	5/8/2015
S67-SS032-0001	AZ16140	SOIL	5/7/2015	5/8/2015
S67-SB032-0607	AZ16141	SOIL	5/7/2015	5/8/2015

Percent moisture was determined using CLP 4.0.

The samples and blanks were screened for J-value responses between the detection limit (DL) and limit of quantitation (LOQ).

Laboratory control limits generated in house do not meet the control limits listed in DoD QSM 4.2 for all analytes. Laboratory control limits generated for this project meet all control limits listed in the DoD QSM 4.2 except where noted. A copy of our in house generated control limits is available upon request. In addition, a copy of our LOQ control limits, established using 7 data points, are also available upon request.

Only the portion of the injection log relative to these samples is included. A full sequence log is available upon request.

Measurement uncertainty can be reported upon request.

CASE NARRATIVE

EPA Method 8270D
Semi-Volatile Organic Compounds

Sample Preparation and Analysis Information:

The soil samples were extracted according to EPA method 3550B. The samples were analyzed according to EPA Method 8270D using an Agilent GC/MS. All holding times were met.

Manual integration were performed in accordance with APPL's SOP. Phenol-D6 was manually integrated in the 100ug/mL standard. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance

Spike Recovery

A Laboratory Control Spike (LCS) was used for quality assurance. All LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blank.

Surrogates

Surrogate recoveries are summarized on Form 2 & 8. Sample S67-SB032-0607 recovered the surrogate 2-Fluorobiphenyl below the 45% lower control limit at 44.4%. The sample was not re extracted because the PAH recovery for this surrogate is 45.1% showing the low recovery is due to matrix. All other surrogate recoveries were within the control limits.

Calibration

The initial and continuing calibrations were performed according to the method. All method acceptance criteria were met.

Tuning:

The instrument was tuned using DFTPP. All method criteria were met.

Internal Standards

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8260. All method criteria were met.

Summary:

No other problem was encountered. All data generated are acceptable.

EPA Method 8270D-SIM
Polynuclear Aromatic Hydrocarbons

Sample Preparation and Analysis Information:

The soil samples were extracted according to EPA method 3550B. The samples were analyzed according to EPA Method 8270D using an Agilent GC/MS with selective ion monitoring. Samples S67-SS030-0001, S67-SS030P-0001 and S67-SS032-0001 were injected at a 10:1 dilution because of color and viscosity. Due to low surrogate recovery, sample S67-SB030-0607 was re-extracted, more than 14 days after collection. All other holding times were met.

Benzo(b)fluoranthene and Benzo(k)fluoranthene could not be chromatographically separated in all samples except S67-SB030-0607. As a result, each of these is reported from the sum of the two area counts and may be biased high.

Quality Control/Assurance

Spike Recovery

Laboratory Control Spikes (LCS) were used for quality assurance. The 150601A LCS for the re extract of sample S67-SB030-0607 recovered five compounds above their upper control limit. All other LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blanks.

Surrogates

Surrogate recoveries are summarized on Form 2 & 8. Samples S67-SS030-0001, S67-SS030P-0001, and S67-SS032-0001 recovered Nitrobenzene-D5 above the 100% upper control. The area counts for this standard were at the low end of the calibration curve. The samples were not re-extracted. Sample S67-SB030-0607 recovered 2-Fluorobiphenyl below the lower limit of 45% at 28.7%. The sample was re -extracted, outside of holding time but within two times the holding time. The re-extracted sample had a passing 2-FBP recovery. The two extracts had similar results. All other surrogate recoveries were within the control limits.

Calibration

The initial and continuing calibrations were performed according to the method. All method acceptance criteria were met.

Tuning:

The instrument was tuned using DFTPP. All method criteria were met.

Internal Standards

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8260. All method criteria were met.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

EPA Method 6850
Perchlorate by LC-Mass Spec

Sample Preparation and Analysis Information:

The soil samples were prepared according to the method. The samples were analyzed according to EPA Method 6850 using an Agilent 6460 Triple Quad LC/MS. The samples were prepared and analyzed within acceptable hold time.

Manual integrations were performed in accordance to APPL's SOP. Perchlorate was manually integrated in the calibration curve and in sample S67-SB032-0607. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance

Calibrations:

Calibrations were performed according to the method. All calibration acceptance criteria were met. The second source met acceptance criteria.

Blanks:

Perchlorate was not detected at or above one-half the limit of quantitation (LOQ) in the method blank.

Spikes:

Laboratory Control Spike (LCS) and Interference Check Sample (I.C.S.) were used for quality assurance. All LCS acceptance criteria were met..

The Interference Check Sample (I.C.S.) was prepared using a mixed anions solution. The ICS recovery met acceptance criteria.

Sample was S67-SS030-0001 selected by the laboratory for MS/MSD analysis. All spike recoveries were acceptable.

Internal Standards:

The area counts of the sample Internal Standards were compared to the average IS area counts of the initial calibration. The internal standard in the ICS recovered slightly below the lower control limit. All other internal standards were within the 50%D acceptance criteria.

Summary:

All data were acceptable. No other analytical exception is noted.

EPA Method 8330B
Energetics

Sample Preparation:

The samples were dried and extracted according to EPA method 8330B, without using incremental sampling procedures. All holding times were met.

Analysis:

The samples were analyzed according to EPA Method 8330B using an Agilent 1290 HPLC with DA detector. Manual integrations were performed in accordance with APPL's SOP. The following analytes were manually integrated in the calibration standards: HMX, Nitroglycerin, PETN, 1,2-Dinitrobenzene, 1,3- Dinitrobenzene, Nitrobenzene, 2,4,6-Trinitrotoluene, Tetryl, 2,6-Dinitrotoluene, 2,4-Dinitrotoluene, 2-Nitrotoluene, 4-Nitrotoluene, 3-Nitrotoluene, 4-Amino-2,6-Dinitrotoluene, 2-Amino-4,6-Dinitrotoluene and 3,5-Dinitroaniline. HMX was manually integrated in the LCS. Chromatograms of before and after manual integration are enclosed.

Quality Control/Assurance:

Spike Recovery:

A Laboratory Control Spike (LCS) was used for quality assurance. All spike acceptance criteria was met.

Sample was S67-SS030-0001 selected by the laboratory for MS/MSD analysis. All spike recoveries were acceptable.

Method blanks:

No target analyte was detected at or above one-half the limit of quantitation (LOQ) in the method blank.

Surrogates:

Surrogate recoveries are summarized on the Form 2 & 8. All surrogates had acceptable recoveries.

Calibration:

The initial and continuing calibrations and second source were analyzed according to the method. All calibration criteria were met.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

EPA Method 6010C
Metals

Digestion and Analysis Information:

The soil samples were digested according to EPA method 3050B. The samples were analyzed for metals according to EPA method 6010C using a Perkin Elmer Optima 5300DV. All holding times were met.

Quality Control/Assurance:

Calibrations:

Calibrations were performed according to the methods for the initial calibration and the initial calibration verification. The initial calibration verification is prepared from a second source standard.

Blanks:

No target metal was detected above one half the LOQ in the method blank.

Spikes:

Laboratory Control Spike (LCS), Matrix Spikes (MS/MSD), Post-Digestion Spike (PDS), and a Dilution Test (DT) were used for quality assurance. All LCS acceptance criteria were met.

Sample S67-SS030-0001 was selected by the laboratory as the QC sample for the analytical batch. All target metals recovered outside the 80-120% control limits in the MS and/or MSD. Recoveries in the DT and PDS were acceptable.

Summary:

No other analytical problem was encountered. The data generated are acceptable.

Inorganic Analyses

EPA Method 9045D and Walkley Black

Sample Preparation and Analysis

The samples were prepared and analyzed according to the methods. All holding times were met.

Quality Control/Assurance

Spike Recovery

A Laboratory Control Spike (LCS) was used for quality assurance. All LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Method blanks

No target analyte was detected one-half the limit of quantitation (LOQ) in the method blank.

Calibration

Calibrations were performed according to the method for the initial calibration and the continuing calibration verification. The initial calibration verification is prepared from a second source standard. All calibration criteria were met.

Summary:

No analytical exception was noted.

APPL Inc.
Abbreviations and Flags

FLAG	DESCRIPTION
#	Recovery or RPD outside control limits
*	Recovery or RPD outside control limits
B	Analyte detected in associated method blank
C1	Reason for correction: wrote incorrect response
C2	Reason for correction: calculated incorrectly
C3	Reason for correction: needs to be rechecked
C4	Reason for correction: data not usable
DO	Diluted out
E	Exceeds linear range
F	Estimated value
G1	Includes a wide range of hydrocarbons which does not match our gasoline standard
G10	Includes a match to hydrocarbon profiles within the range of mineral spirits
G11	Includes a match to hydrocarbon profiles within the range of JP-4
G12	Pattern does not match the gasoline standard; the carbon range for this sample is consistent with JP8
G13	Closely resembles the hydrocarbon profile of aviation gasoline
G14	Analyte concentration may be biased due to carry over
G2	Closely resembles the boiling point hydrocarbon profile consistent with weathered gasoline
G3	Includes higher boiling hydrocarbons
G4	Includes dominant peak(s) not indicative of petroleum hydrocarbons
G5	Is mainly dominant peak(s) not indicative of petroleum hydrocarbons
G6	Contains recognizable contaminant peak(s) which has been removed from quantitation
G7	Is mainly a match to hydrocarbons within the range of gasoline
G8	Closely resembles the boiling point hydrocarbon profile consistent with weathered gasoline
G9	Includes hydrocarbons within the range of kerosene
J	Estimated value
M	Matrix effect
MI1	Manual integration: integration does not follow baseline
MI2	Manual integration: non-target peak interference
MI3	Manual integration: to split a peak that was integrated as one peak by the computer.
MI4	Manual integration: to integrate a split peak
MI5	Manual integration: the whole peak or part of the peak was not integrated
MI6	Manual integration: computer integrated wrong peak
MI7	Manual integration: other – (See case narrative)
MDL	Method detection limit
ND	Not detected
NT	Non-target
Q	Acceptance criteria not met
T1 I	Includes wide range of hydrocarbons not indicative of diesel
T1 M	Is mainly wide range of hydrocarbons not necessarily indicative of diesel
T2 I	Includes lower boiling hydrocarbons, i.e. mineral spirits, kerosene, stoddard solvent, white gas
T2 M	Is mainly lower boiling hydrocarbons, i.e. mineral spirits, kerosene, stoddard solvent, white gas
T3 I	Includes higher boiling hydrocarbons, i.e. asphaltene, waster oil, motor oil, or weathered diesel fuel
T3 M	Is mainly higher boiling hydrocarbons, i.e. asphaltene, waster oil, motor oil, or weathered diesel fuel
T4 I	Includes dominant peak(s) not indicative of hydrocarbons
T4 M	Is mainly dominant peak(s) not indicative of hydrocarbons
T5	Contains recognizable contaminant peak(s) which has been removed from quantitation
T6	Is mainly a match to hydrocarbons within range of diesel fuel
T7	Closely resembles the boiling point hydrocarbon profile consistent with diesel fuel
T8	Includes a match to hydrocarbon profiles within range of diesel and kerosene fuel
T9 I	Includes non-diesel hydrocarbons within boiling point range of diesel fuel
T9 M	Is mainly non-diesel hydrocarbons within boiling point range of diesel fuel
U	Not detected
Y	Percent difference between primary and confirmation column > 40%

CHAIN OF CUSTODY,

ARF, CRF, AND

CLIENT COMMUNICATION

APPL - Analysis Request Form 76380

Client: Tetra Tech
Address: 5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502
Attn: Ed Corack
Phone: 757-466-4908 Fax: 757-461-4148
Job: CTO JU11 112G02622 NSF Indian Head
PO #: MSA #1045497 Release #08-CTO JU11
Chain of Custody (Y/N): Y # 088451
RAD Screen (Y/N): Y pH (Y/N): N
Turn Around Type: STD

Received by: YL
Date Received: 05/08/15 Time: 11:00
Delivered by: FED EX
Shuttle Custody Seals (Y/N): Y Time Zone: -4
Chest Temp(s): 3.0°C
Color: E-Brn
Samples Chilled until Placed in Refrig/Freezer: Y
Project Manager: Cynthia Clark
QC Report Type: DVP4/LEDD/MD
Due Date: 05/29/15

Comments:
login to ed.corack@ & tobrena.sedlmyer@tetratech.com.
21 calendar day TAT for final report
1HC DVP4 and summary report & 2 BOOKMARKED CDs of DVP4 and
Summary report (sample & QC results w/surrogate & blank summaries) to Amy Thomson (Pittsburgh)
Include original COC with report. NOTHING to Ed in VA office
EDD: TTEC LEDD to http://isg.applications.tetratech.com/LEDDChecker2010/
Guidance: DOD QSM v4.2: DOD Forms, LOD Database

Sample Distribution:
GC: 5-$87DJU11S, 5-$SIMDDODSM
Extractions: 5-_HPLC6850GROSS, 5-_MSE018, 5-_SON009, 5-_SON009S
LCMS: 5-$6850SM, 5-$83BJU11S
Metals: 5-$61CJU11S2(Al,B,Li,Zn)
Wetlab: 3-$9045D, 3-$TOC-WB, 5-MOIST
Other: 5-_M3050

Charges:

Invoice To:
ACCOUNTS PAYABLE
661 Andersen Dr, Foster Plaza 7
Pittsburgh, PA 15220-2745
invoice in triplicate per SOW

	Client ID	APPL ID	Sampled	Analyses Requested
1.	S67-SS030-0001	AZ16137S	05/07/15 10:35	$61CJU11S2(Al,B,Li,Zn), $6850SM, $83BJU11S, $87DJU11S, $9045D, $SIMDDODSM, $TOC-WB, MOIST
2.	S67-SS030P-0001	AZ16138S	05/07/15 10:35	$61CJU11S2(Al,B,Li,Zn), $6850SM, $83BJU11S, $87DJU11S, $9045D, $SIMDDODSM, $TOC-WB, MOIST
3.	S67-SB030-0607	AZ16139S	05/07/15 11:00	$61CJU11S2(Al,B,Li,Zn), $6850SM, $83BJU11S, $87DJU11S, $SIMDDODSM, MOIST

4.	S67-SS032-0001	AZ16140S	05/07/15	14:20	$61CJU11S2(Al,B,Li,Zn), $6850SM, $83BJU11S, $87DJU11S, $9045D, $SIMDDODSM, $TOC-WB, MOIST
5.	S67-SB032-0607	AZ16141S	05/07/15	14:40	$61CJU11S2(Al,B,Li,Zn), $6850SM, $83BJU11S, $87DJU11S, $SIMDDODSM, MOIST

APPL Sample Receipt Form

ARF# **76380**

Sample	Container Type	Count	pH
AZ16137	[20] 4oz Jar	1	NA
	[21] 8oz Jar	2	NA
AZ16138	[20] 4oz Jar	1	NA
	[21] 8oz Jar	2	NA
AZ16139	[21] 8oz Jar	2	NA
AZ16140	[20] 4oz Jar	1	NA
	[21] 8oz Jar	2	NA
AZ16141	[21] 8oz Jar	2	NA

Sample	Container Type	Count	pH

CHEMTECH
CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. 76375 3.0
QUOTE NO.
COC Number 088451

CLIENT INFORMATION
REPORT TO BE SENT TO:
COMPANY: Tetra Tech
ADDRESS: 5700 Lake Wright Dr. St 301
CITY: Norfolk STATE: VA ZIP: 23502
ATTENTION: R Ed Corack
PHONE: 757-466-4908 FAX:

CLIENT PROJECT INFORMATION
PROJECT NAME: NSF-Indian Head Site 67 GO-JUL
PROJECT NO.: 112G02622 LOCATION: Site 67
PROJECT MANAGER: Ed. Corack
e-mail: Ed.Corack@tetratech.com
PHONE: 757-466-4908 FAX:

CLIENT BILLING INFORMATION
BILL TO: See Contract PO#:
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION
FAX: _____ DAYS *
HARD COPY: _____ DAYS *
EDD: _____ DAYS *
PREAPPROVED TAT: ☐ YES ☐ NO
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION
☐ RESULTS ONLY ☐ USEPA CLP
☒ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☒ EDD FORMAT _____

ANALYSIS
1 Perchlorate (24 Hr TAT)
2 PAHs, Phthalates
3 Metals
4 Energetics
5 Perchlorate
6 pH
7 TOC
8
9

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
1.	S67-GW-E1-050715	GW		X	5-7-15	0920	1	E 1									
2.	S67-GW-F1-050715	GW		X	5-7-15	0950	1	E 1									Rush 24 HR TAT
3.	S67-SS030-0001	SO		X	5-7-15	1035	3		E 2				E 1				Rush 24 HR TAT
4.	S67-SS030P-0001	SO		X	5-7-15	1035	3		E 2				E 1				
5.	S67-SB030-0607	SO		X	5-7-15	1100	2		E 2				E 1		E 3		5-7-15
6.	S67-SS032-0001	SO		X	5-7-15	1420	3		E 2				E 1				
7.	S67-SB032-0607	SO		X	5-7-15	1440	2		E 2								
8.																	
9.																	
10.																	

5-7-15

← Specify Preservatives
A-HCl B-HNO_3
C-H_2SO_4 D-NaOH
E-ICE F-Other

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Jason Birkett	5-7-15 1800	1. FedEx
RELINQUISHED BY:	DATE/TIME:	RECEIVED BY:
2.		2.
RELINQUISHED BY:	DATE/TIME:	RECEIVED FOR LAB BY:
3.	5/8/15 11:00	3. Young Zo

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant Cooler Temp. _____
MeOH extraction requires an additional 4 oz jar for percent solid.
Comments: Ice in Cooler?: _____

Page 1 of 1

SHIPPED VIA: CLIENT: ☐ HAND DELIVERED ☐ OVERNIGHT
CHEMTECH: ☐ PICKED UP ☐ OVERNIGHT
Shipment Complete: ☐ YES ☐ NO

Revision 8/2007

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

COOLER RECEIPT FORM **ARF: 76380**

1) Project: CTO JU11 112G02622 NSF Indian Head Date Received: 05/08/15

2) Coolers: Number of Coolers: 1

3) YES Were custody seals present and intact?
How many? 2 Name/Date on seal? SEE BELOW

4) YES Was there a shipping slip? Carrier name: FED EX

5) Type of packing in cooler: X bubble wrap __popcorn __foam X plastic bags
X wet ice __dry ice __no ice __other

6) YES Were cooler temperatures acceptable?

7) Serial number of certified NIST thermometer use J5297

8) Cooler temp(s): In °C
1: 3.0 2: ____ 3: ____ 4: ____ 5: ____ 6: ____
7: ____ 8: ____ 9: ____ 10: ____ 11: ____ 12: ____

Chain of custody:

9) YES Was a chain of custody received?

10) YES Were the custody papers complete/signed in the appropriate places?

Sample Labels:

11) YES Were all sample labels complete (sample ID, date/time of sampling, etc.)?

12) YES Did all container labels agree with custody papers?

Sample Containers:

13) YES Were all containers sealed in separate bags?

14) YES Did all containers arrive in good condition:(unbroken, no leakage, no cracked/broken lids)?

15) YES Were correct containers and preservatives used for the tests indicated?

16) YES Was a sufficient amount of sample sent for tests indicated?

17) NA Were bubbles present in volatile samples?
If yes, the following were received with air bubbles:
Larger than a pea: ____
Smaller than a pea: ____



Preservation Hold time:

18) Yes Was a sufficient amount of holding time remaining to analyze the samples?

19) NA Was the pH taken of all non-VOA preserved samples and written on the sample container?

20) NA Was the pH of acid preserved non-VOA samples < 2?

21) NA Was the pH of sodium hydroxide preserved samples for Cyanide > 12 and Sulfide >9?

22) NO Were unpreserved VOA Vials received?

23) NA Are unpreserved VOA vials noted in the ADD TEST FIELD on the ARF?
pH strip lot number: ____
Lab notified if pH was not adequate: ____

Notes/Deficiencies:

Personnel receiving samples: BB Second reviewer: YL
Personnel labeling samples: ____
Project manager notified: ____ Date/Time of notification ____
Name of client notified: ____ Date/Time of notification ____

EPA METHOD 8270D
Semivolatile Organic Compounds

APPL, INC.

EPA METHOD 8270D
Semivolatile Organic Compounds
QC Summary

APPL, INC.

Method Blank
EPA 8270D SOILS

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150521S-16137 - 197602**
Batch ID: #87DJU-150521A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	BIS (2-ETHYLHEXYL) PHTHALAT	0.167 U	0.66	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	BUTYL BENZYL PHTHALATE	0.167 U	0.33	0.167	0.056	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-BUTYL PHTHALATE	0.167 U	0.33	0.167	0.066	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-OCTYL PHTHALATE	0.167 U	0.33	0.167	0.058	mg/kg	5/21/2015	5/28/2015
BLANK	DIETHYL PHTHALATE	0.167 U	0.33	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	DIMETHYL PHTHALATE	0.167 U	0.33	0.167	0.063	mg/kg	5/21/2015	5/28/2015
BLANK	SURROGATE: 2,4,6-TRIBROMOP	66.1	35-125			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	66.9	45-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUOROPHENO	70.2	35-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: NITROBENZENE-	67.7	35-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: PHENOL (S)	69.3	40-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: TERPHENYL-D14 (	86.8	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y055
Instrument: Yoda
Sequence: Y150520
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 12:12:12 PM

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: 2,4,6-TRIBROMOPHENOL (S)			SURROGATE: 2-FLUORBIPHENYL (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	35-125	66.1		45-105	66.9	
150521A-LCS	Lab Control Spike	35-125	76.0		45-105	68.2	
AZ16137	S67-SS030-0001	35-125	61.4		45-105	59.6	
AZ16138	S67-SS030P-0001	35-125	57.8		45-105	56.4	
AZ16139	S67-SB030-0607	35-125	65.6		45-105	47.5	
AZ16140	S67-SS032-0001	35-125	66.2		45-105	63.3	
AZ16141	S67-SB032-0607	35-125	62.7		45-105	44.4	#

Comments: Batch: #87DJU-150521A

= Recovery outside of Control Limits on Sample.

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: 2-FLUOROPHENOL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	35-105	70.2		35-100	67.7	
150521A-LCS	Lab Control Spike	35-105	66.9		35-100	66.7	
AZ16137	S67-SS030-0001	35-105	59.8		35-100	59.2	
AZ16138	S67-SS030P-0001	35-105	51.9		35-100	54.0	
AZ16139	S67-SB030-0607	35-105	54.0		35-100	55.9	
AZ16140	S67-SS032-0001	35-105	58.5		35-100	60.2	
AZ16141	S67-SB032-0607	35-105	60.6		35-100	59.8	

Comments: Batch: #87DJU-150521A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

APPL ID.	Client Sample No.	SURROGATE: PHENOL (S)			SURROGATE: TERPHENYL-D14 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	40-100	69.3		30-125	86.8	
150521A-LCS	Lab Control Spike	40-100	69.9		30-125	89.5	
AZ16137	S67-SS030-0001	40-100	59.5		30-125	76.4	
AZ16138	S67-SS030P-0001	40-100	51.1		30-125	70.9	
AZ16139	S67-SB030-0607	40-100	57.9		30-125	87.8	
AZ16140	S67-SS032-0001	40-100	59.6		30-125	76.5	
AZ16141	S67-SB032-0607	40-100	62.4		30-125	72.0	

Comments: Batch: #87DJU-150521A

Laboratory Control Spike Recovery
EPA 8270D SOILS

APPL ID: **150521S-16137 LCS - 197602**
Batch ID: #87DJU-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
BIS (2-ETHYLHEXYL) PHTHALATE	1.67	1.35	80.8	45-125
BUTYL BENZYL PHTHALATE	1.67	1.31	78.4	50-125
DI-N-BUTYL PHTHALATE	1.67	1.30	77.8	55-110
DI-N-OCTYL PHTHALATE	1.67	1.39	83.2	40-130
DIETHYL PHTHALATE	1.67	1.19	71.3	50-115
DIMETHYL PHTHALATE	1.67	1.19	71.3	50-110
SURROGATE: 2,4,6-TRIBROMOPHENOL	6.67	5.07	76.0	35-125
SURROGATE: 2-FLUORBIPHENYL (S)	3.33	2.27	68.2	45-105
SURROGATE: 2-FLUOROPHENOL (S)	6.67	4.46	66.9	35-105
SURROGATE: NITROBENZENE-D5 (S)	3.33	2.22	66.7	35-100
SURROGATE: PHENOL (S)	6.67	4.66	69.9	40-100
SURROGATE: TERPHENYL-D14 (S)	3.33	2.98	89.5	30-125

Comments: _____

Primary	SPK
Quant Method :	Y0520TT.M
Extraction Date :	5/21/2015
Analysis Date :	5/28/2015
Instrument :	Yoda
Run :	0520Y056
Initials :	DA

Printed: 6/2/2015 12:11:56 PM
APPL Standard LCS

EPA 8270D

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/28/2015

Matrix: SOIL Instrument: Yoda

Blank ID: 150521A-BLK Time Analyzed: 1449

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150521A-BLK	Blank	0520Y055	5/28/2015 1449
150521A-LCS	Lab Control Spike	0520Y056	5/28/2015 1518
AZ16137	S67-SS030-0001	0520Y057	5/28/2015 1547
AZ16138	S67-SS030P-0001	0520Y058	5/28/2015 1617
AZ16139	S67-SB030-0607	0520Y059	5/28/2015 1646
AZ16140	S67-SS032-0001	0520Y060	5/28/2015 1716
AZ16141	S67-SB032-0607	0520Y061	5/28/2015 1746

Comments: Batch: #87DJU-150521A

Form 5
Tune Summary

Lab Name: APPL Inc.
Case No: ~~0520Y005.D~~ 76380
Matrix: Soil mp6-1715
ID: SV Tune 5/14/15

SDG No: mp6-1715 ~~Yoda~~ 76380
Date Analyzed: 5/22/2015
Instrument: Yoda
Time Analyzed: 10:28

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		5ug/ml SVOC 05/22/15	0520Y006.D	5/22/2015 10:46
2		10ug/ml SVOC 05/22/1	0520Y007.D	5/22/2015 11:15
3		20ug/ml SVOC 05/22/1	0520Y008.D	5/22/2015 11:45
4		40ug/ml SVOC 05/22/1	0520Y009.D	5/22/2015 12:14
5		50ug/ml SVOC 05/22/1	0520Y010.D	5/22/2015 12:43
6		60ug/ml SVOC 05/22/1	0520Y011.D	5/22/2015 13:12
7		80ug/ml SVOC 05/22/1	0520Y012.D	5/22/2015 13:42
8		100ug/ml SVOC 05/22/	0520Y013.D	5/22/2015 14:11
9		50ug/ml SVOC SS 05/2	0520Y014.D	5/22/2015 14:41
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e		
51	29.95 - 60.04% of mass 198	31.1
68	0 - 2% of mass 69	0.0
70	0 - 2% of mass 69	0.1
127	40 - 60% of mass 198	51.6
197	0 - 1.4% of mass 198	0.1
198	100 - 100% of mass 198	100.0
199	5 - 9% of mass 198	6.8
275	10 - 30% of mass 198	27.8
365	1 - 100% of mass 198	3.1
441	0.01 - 100% of mass 443	76.5
442	40 - 150% of mass 198	105.2
443	17 - 23% of mass 442	18.9

Form 5
Tune Summary

Lab Name: APPL Inc.
Case No: 76380
Matrix: Soil
ID: SV TUNE 5/14/15

SDG No: 76380
Date Analyzed: 5/28/2015
Instrument: Yoda
Time Analyzed: 10:36

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		50ug/ml SVOC 05/28/1	0520Y048.D	5/28/2015 10:54
2	Blank	150521A BLK 1/30.00G	0520Y055.D	5/28/2015 14:49
3	Lab Control Spike	150521A LCS-1 1/30.0	0520Y056.D	5/28/2015 15:18
4	S67-SS030-0001	AZ16137S01 1/30.15G	0520Y057.D	5/28/2015 15:47
5	S67-SS030P-0001	AZ16138S01 1/30.26G	0520Y058.D	5/28/2015 16:17
6	S67-SB030-0607	AZ16139S01 1/30.17G	0520Y059.D	5/28/2015 16:46
7	S67-SS032-0001	AZ16140S01 1/30.04G	0520Y060.D	5/28/2015 17:16
8	S67-SB032-0607	AZ16141S01 1/30.33G	0520Y061.D	5/28/2015 17:46
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e		
51	29.95 - 60.04% of mass 198	30.0
68	0 - 2% of mass 69	0.1
70	0 - 2% of mass 69	0.4
127	40 - 60% of mass 198	48.9
197	0 - 1.4% of mass 198	0.0
198	100 - 100% of mass 198	100.0
199	5 - 9% of mass 198	6.6
275	10 - 30% of mass 198	28.7
365	1 - 100% of mass 198	3.0
441	0.01 - 100% of mass 443	75.6
442	40 - 150% of mass 198	88.4
443	17 - 23% of mass 442	19.4

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review

Lab Code: _____ SDG No.: 76380

Lab File ID (Standard): 0520Y010.D Date Analyzed: 05/22/15

Instrument ID: Yoda Time Analyzed: 12:43

GC Column: _____ ID: Heated Purge: (Y/N) _____

		1,4-dichlorobenzene-D4(IS)		Napthalene-D8(IS)		Acenaphthene-D10(IS)	
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	789097	4.80	3524030	6.23	2025700	8.22
	UPPER LIMIT	1578194	5.30	7048060	6.73	4051400	8.72
	LOWER LIMIT	394549	4.30	1762015	5.73	1012850	7.72
	SAMPLE NO.						
01	50ug/ml SVOC SS 05/2	477080	4.80	2254040	6.23	1379160	8.22
02	50ug/ml SVOC 05/28/15	937800	4.80	4119960	6.24	2216050	8.22
03	150521A BLK 1/30.00G	934649	4.80	4145960	6.24	2316230	8.22
04	150521A LCS-1 1/30.07	983964	4.80	4419820	6.23	2490320	8.22
05	AZ16137S01 1/30.15G	1023120	4.80	4664940	6.23	2549410	8.22
06	AZ16138S01 1/30.26G	892677	4.80	4006140	6.23	2237170	8.22
07	AZ16139S01 1/30.17G	904027	4.80	4054600	6.23	2272840	8.22
08	AZ16140S01 1/30.04G	853087	4.80	3891190	6.23	2134360	8.22
09	AZ16141S01 1/30.33G	875159	4.80	3871990	6.23	2150550	8.22
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
AREA LOWER LIMIT = -50% of internal standard area.
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review

Lab Code: _____ SDG No.: 76380

Lab File ID (Standard): 0520Y010.D Date Analyzed: 05/22/15

Instrument ID: Yoda Time Analyzed: 12:43

GC Column: _____ ID: Heated Purge: (Y/N) _____

		Phenanthrene-D10(IS)		Chrysene-D12(IS)		Perylene-D12(IS)	
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	3568970	9.95	3316560	13.01	2280870	14.64
	UPPER LIMIT	7137940	10.45	6633120	13.51	4561740	15.14
	LOWER LIMIT	1784485	9.45	1658280	12.51	1140435	14.14
	SAMPLE NO.						
01	50ug/ml SVOC SS 05/2	2658510	9.95	2791250	13.01	2129350	14.64
02	50ug/ml SVOC 05/28/15	3645220	9.94	2955800	13.00	1573180	14.64
03	150521A BLK 1/30.00G	4141420	9.94	3325730	13.01	2051460	14.64
04	150521A LCS-1 1/30.07	4281750	9.95	3572460	13.01	2131740	14.64
05	AZ16137S01 1/30.15G	4374110	9.95	3576400	13.00	2143870	14.64
06	AZ16138S01 1/30.26G	3914600	9.95	3278260	13.01	1937930	14.64
07	AZ16139S01 1/30.17G	4051140	9.94	3241450	13.01	1952890	14.65
08	AZ16140S01 1/30.04G	3681210	9.94	2974590	13.00	1837130	14.64
09	AZ16141S01 1/30.33G	3712440	9.94	3014510	13.01	1869380	14.64
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
AREA LOWER LIMIT = -50% of internal standard area.
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

EPA METHOD 8270D
Semivolatile Organic Compounds
Sample Data

APPL, INC.

EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	30	7.60	1.900	0.720	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	1.900 U	3.80	1.900	0.650	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	1.900 U	3.80	1.900	0.760	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	1.900 U	3.80	1.900	0.670	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	1.900 U	3.80	1.900	0.720	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	1.900 U	3.80	1.900	0.730	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	61.4	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	59.6	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	59.8	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	59.2	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	59.5	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	76.4	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y057
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 10
Initials: DA

Printed: 6/2/2015 12:11:46 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y057.D Vial: 57
Acq On : 28 May 15 15:47 Operator: csc
Sample : AZ16137S01 1/30.15G x10 Inst : Yoda
Misc : soil Multiplr: 331.68

Quant Time: Jun 2 10:16 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	1023120	40.00	ppb	0.00
7) Napthalene-D8(IS)	6.23	136	4664935	40.00	ppb	0.00
13) Acenaphthene-D10(IS)	8.22	164	2549406	40.00	ppb	0.00
23) Phenanthrene-D10(IS)	9.95	188	4374108	40.00	ppb	0.00
27) Chrysene-D12(IS)	13.00	240	3576400	40.00	ppb	0.00
32) Perylene-D12(IS)	14.64	264	2143865	40.00	ppb	0.00
System Monitoring Compounds						
2) 2-Fluorophenol (S)	4.06	112	452401	3964.82	ppb	0.02
Spiked Amount 6633.499			Recovery =	59.770%		
3) Phenol-D6 (S)	5.16	99	659889	3948.27	ppb	0.05
Spiked Amount 6633.499			Recovery =	59.520%		
8) Nitrobenzene-D5(S)	5.53	82	283762	1964.37	ppb	0.03
Spiked Amount 3316.750			Recovery =	59.226%		
16) 2-Fluorobiphenyl(S)	7.46	172	581808	1977.37	ppb	0.00
Spiked Amount 3316.750			Recovery =	59.618%		
22) 2,4,6-Tribromophenol(S)	9.23	330	85755	4073.19	ppb	0.00
Spiked Amount 6633.499			Recovery =	61.403%		
28) Terphenyl-D14(S)	11.74	244	676975	2532.45	ppb	0.00
Spiked Amount 3316.750			Recovery =	76.353%		
Target Compounds						Qvalue
26) Di-n-butylphthalate	10.54	149	6419	13.68	ppb #	94
30) Bis (2-ethylhexyl) phthala	12.82	149	6901534	25958.82	ppb #	93

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y057.D Vial: 57
Acq On : 28 May 15 15:47 Operator: csc
Sample : AZ16137S01 1/30.15G x10 Inst : Yoda
Misc : soil Multiplr: 331.68

Quant Time: Jun 2 10:16 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration





#26
Di-n-butylphthalate
Concen: 13.68 ppb
RT: 10.54 min Scan# 992
Delta R.T. 0.00 min
Lab File: 0520Y057.D
Acq: 28 May 15 15:47

Tgt Ion:149 Resp: 6419

Ion	Ratio	Lower	Upper
149	100		
150	11.3	6.4	11.8
104	8.0	4.3	7.9#





#30
Bis (2-ethylhexyl) phthalate
Concen: 25958.82 ppb
RT: 12.82 min Scan# 1237
Delta R.T. 0.00 min
Lab File: 0520Y057.D
Acq: 28 May 15 15:47

Tgt Ion:149 Resp: 6901534

Ion	Ratio	Lower	Upper
149	100		
167	29.9	23.1	42.9
279	5.7	7.1	13.3#



EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16138
QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	9.6	0.75	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.190 U	0.38	0.190	0.064	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.190 U	0.38	0.190	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.190 U	0.38	0.190	0.066	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.190 U	0.38	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.190 U	0.38	0.190	0.072	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	57.8	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	56.4	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	51.9	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	54.0	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	51.1	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	70.9	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y058
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 12:11:46 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y058.D Vial: 58
Acq On : 28 May 15 16:17 Operator: csc
Sample : AZ16138S01 1/30.26G x10 Inst : Yoda
Misc : soil Multiplr: 330.47

Quant Time: Jun 2 9:36 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	892677	40.00	ppb	0.00
7) Napthalene-D8(IS)	6.23	136	4006141	40.00	ppb	0.00
13) Acenaphthene-D10(IS)	8.22	164	2237173	40.00	ppb	0.00
23) Phenanthrene-D10(IS)	9.95	188	3914604	40.00	ppb	0.00
27) Chrysene-D12(IS)	13.01	240	3278256	40.00	ppb	0.00
32) Perylene-D12(IS)	14.64	264	1937928	40.00	ppb	0.00
System Monitoring Compounds						
2) 2-Fluorophenol (S)	4.08	112	343380	3433.27	ppb	0.04
Spiked Amount 6609.385			Recovery =	51.945%		
3) Phenol-D6 (S)	5.17	99	501658	3376.49	ppb	0.06
Spiked Amount 6609.385			Recovery =	51.086%		
8) Nitrobenzene-D5(S)	5.53	82	222327	1785.66	ppb	0.03
Spiked Amount 3304.693			Recovery =	54.034%		
16) 2-Fluorobiphenyl(S)	7.46	172	483241	1864.79	ppb	0.00
Spiked Amount 3304.693			Recovery =	56.429%		
22) 2,4,6-Tribromophenol(S)	9.24	330	69669	3817.23	ppb	0.02
Spiked Amount 6609.385			Recovery =	57.755%		
28) Terphenyl-D14(S)	11.74	244	575819	2341.40	ppb	0.00
Spiked Amount 3304.693			Recovery =	70.851%		
Target Compounds						Qvalue
29) Butyl benzylphthalate	12.27	149	3998	22.75	ppb	86
30) Bis (2-ethylhexyl) phthala	12.81	149	2076402	8489.30	ppb	97

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y058.D Vial: 58
Acq On : 28 May 15 16:17 Operator: csc
Sample : AZ16138S01 1/30.26G x10 Inst : Yoda
Misc : soil Multiplr: 330.47

Quant Time: Jun 2 9:36 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration





#29
Butyl benzylphthalate
Concen: 2.28 ppb
RT: 12.27 min Scan# 1178
Delta R.T. -0.00 min
Lab File: 0520Y058.D
Acq: 28 May 15 16:17

Tgt Ion:149 Resp: 3998

Ion	Ratio	Lower	Upper
149	100		
91	84.4	49.4	91.8
206	23.6	15.3	28.3





#30
Bis (2-ethylhexyl) phthalate
Concen: 848.93 ppb
RT: 12.81 min Scan# 1237
Delta R.T. -0.00 min
Lab File: 0520Y058.D
Acq: 28 May 15 16:17

Tgt Ion:149 Resp: 2076402

Ion	Ratio	Lower	Upper
149	100		
167	31.9	23.1	42.9
279	7.5	7.1	13.3



EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16139
QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.200 U	0.80	0.200	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.200 U	0.40	0.200	0.068	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.200 U	0.40	0.200	0.080	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.200 U	0.40	0.200	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.200 U	0.40	0.200	0.075	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.200 U	0.40	0.200	0.076	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	65.6	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	47.5	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	54.0	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	55.9	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	57.9	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	87.8	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y059
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 12:11:47 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y059.D Vial: 59
Acq On : 28 May 15 16:46 Operator: csc
Sample : AZ16139S01 1/30.17G Inst : Yoda
Misc : soil Multiplr: 33.15

Quant Time: Jun 2 10:19 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	904027	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	4054597	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2272843	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	4051136	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3241445	40.00	ppb		0.00
32) Perylene-D12(IS)	14.65	264	1952891	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	3.99	112	3592610	3581.30	ppb		-0.06
Spiked Amount 6629.102			Recovery =	54.024%			
3) Phenol-D6 (S)	5.09	99	5208705	3840.84	ppb		-0.02
Spiked Amount 6629.102			Recovery =	57.939%			
8) Nitrobenzene-D5(S)	5.50	82	2327216	1852.31	ppb		0.00
Spiked Amount 3314.551			Recovery =	55.884%			
16) 2-Fluorobiphenyl(S)	7.45	172	4132647	1574.41	ppb		0.00
Spiked Amount 3314.551			Recovery =	47.500%			
22) 2,4,6-Tribromophenol(S)	9.22	330	1001212	4350.27	ppb		0.00
Spiked Amount 6629.102			Recovery =	65.624%			
28) Terphenyl-D14(S)	11.75	244	7053551	2909.35	ppb		0.00
Spiked Amount 3314.551			Recovery =	87.775%			
Target Compounds							Qvalue
21) Diethyl phthalate	8.74	149	4346	19.46	ppb		93
30) Bis (2-ethylhexyl) phthala	12.81	149	11078	4.59	ppb	#	95

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y059.D Vial: 59
Acq On : 28 May 15 16:46 Operator: csc
Sample : AZ16139S01 1/30.17G Inst : Yoda
Misc : soil Multiplr: 33.15

Quant Time: Jun 2 10:19 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration





#21
Diethyl phthalate
Concen: 19.46 ppb
RT: 8.74 min Scan# 798
Delta R.T. 0.01 min
Lab File: 0520Y059.D
Acq: 28 May 15 16:46

Tgt Ion:149 Resp: 4346

Ion	Ratio	Lower	Upper
149	100		
177	17.3	15.1	28.0
150	10.3	8.3	15.5





#30
Bis (2-ethylhexyl) phthalate
Concen: 4.59 ppb
RT: 12.81 min Scan# 1237
Delta R.T. -0.00 min
Lab File: 0520Y059.D
Acq: 28 May 15 16:46

Tgt Ion:149 Resp: 11078

Ion	Ratio	Lower	Upper
149	100		
167	31.5	23.1	42.9
279	5.9	7.1	13.3#



EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS032-0001

APPL ID: AZ16140

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.190 U	0.74	0.190	0.069	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.190 U	0.37	0.190	0.063	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.190 U	0.37	0.190	0.074	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.190 U	0.37	0.190	0.065	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.190 U	0.37	0.190	0.069	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.190 U	0.37	0.190	0.070	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	66.2	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	63.3	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	58.5	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	60.2	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	59.6	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	76.5	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y060
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 12:11:47 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y060.D Vial: 60
Acq On : 28 May 15 17:16 Operator: csc
Sample : AZ16140S01 1/30.04G x10 Inst : Yoda
Misc : soil Multiplr: 332.89

Quant Time: Jun 2 9:39 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	853087	40.00	ppb	0.00
7) Napthalene-D8(IS)	6.23	136	3891193	40.00	ppb	0.00
13) Acenaphthene-D10(IS)	8.22	164	2134360	40.00	ppb	0.00
23) Phenanthrene-D10(IS)	9.94	188	3681208	40.00	ppb	0.00
27) Chrysene-D12(IS)	13.00	240	2974587	40.00	ppb	0.00
32) Perylene-D12(IS)	14.64	264	1837132	40.00	ppb	0.00
System Monitoring Compounds						
2) 2-Fluorophenol (S)	4.07	112	369489	3897.29	ppb	0.03
Spiked Amount 6657.790			Recovery =	58.537%		
3) Phenol-D6 (S)	5.16	99	550449	3964.52	ppb	0.05
Spiked Amount 6657.790			Recovery =	59.547%		
8) Nitrobenzene-D5(S)	5.53	82	240619	2004.23	ppb	0.03
Spiked Amount 3328.895			Recovery =	60.207%		
16) 2-Fluorobiphenyl(S)	7.46	172	516925	2106.17	ppb	0.00
Spiked Amount 3328.895			Recovery =	63.269%		
22) 2,4,6-Tribromophenol(S)	9.23	330	78753	4405.46	ppb	0.00
Spiked Amount 6657.790			Recovery =	66.170%		
28) Terphenyl-D14(S)	11.74	244	564343	2547.52	ppb	0.00
Spiked Amount 3328.895			Recovery =	76.528%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y060.D Vial: 60
Acq On : 28 May 15 17:16 Operator: csc
Sample : AZ16140S01 1/30.04G x10 Inst : Yoda
Misc : soil Multiplr: 332.89

Quant Time: Jun 2 9:39 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration



EPA 8270D SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB032-0607

APPL ID: AZ16141

Sample Collection Date: 5/7/2015

QCG: #87DJU-150521A-197602

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)								
EPA 8270D	BIS (2-ETHYLHEXYL) PHTHALATE	0.220 U	0.86	0.220	0.081	mg/kg	5/21/2015	5/28/2015
EPA 8270D	BUTYL BENZYL PHTHALATE	0.220 U	0.43	0.220	0.073	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-BUTYL PHTHALATE	0.220 U	0.43	0.220	0.086	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DI-N-OCTYL PHTHALATE	0.220 U	0.43	0.220	0.076	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIETHYL PHTHALATE	0.220 U	0.43	0.220	0.081	mg/kg	5/21/2015	5/28/2015
EPA 8270D	DIMETHYL PHTHALATE	0.220 U	0.43	0.220	0.082	mg/kg	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2,4,6-TRIBROMOPHEN	62.7	35-125			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUORBIPHENYL (S)	44.4 #	45-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: 2-FLUOROPHENOL (S)	60.6	35-105			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: NITROBENZENE-D5 (S)	59.8	35-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: PHENOL (S)	62.4	40-100			%	5/21/2015	5/28/2015
EPA 8270D	SURROGATE: TERPHENYL-D14 (S)	72.0	30-125			%	5/21/2015	5/28/2015

= Recovery (or RPD) is outside QC limits.

Quant Method: Y0520TT.M
Run #: 0520Y061
Instrument: Yoda
Sequence: Y150520
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 12:11:48 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y061.D Vial: 61
Acq On : 28 May 15 17:46 Operator: csc
Sample : AZ16141S01 1/30.33G Inst : Yoda
Misc : soil Multiplr: 32.97

Quant Time: Jun 2 10:20 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	875159	40.00	ppb	0.00
7) Napthalene-D8(IS)	6.23	136	3871991	40.00	ppb	0.00
13) Acenaphthene-D10(IS)	8.22	164	2150551	40.00	ppb	0.00
23) Phenanthrene-D10(IS)	9.94	188	3712436	40.00	ppb	0.00
27) Chrysene-D12(IS)	13.01	240	3014511	40.00	ppb	0.00
32) Perylene-D12(IS)	14.64	264	1869383	40.00	ppb	0.00
System Monitoring Compounds						
2) 2-Fluorophenol (S)	3.98	112	3898873	3993.93	ppb	-0.06
Spiked Amount 6594.131			Recovery =	60.568%		
3) Phenol-D6 (S)	5.08	99	5422620	4111.68	ppb	-0.03
Spiked Amount 6594.131			Recovery =	62.354%		
8) Nitrobenzene-D5(S)	5.50	82	2379121	1972.47	ppb	0.00
Spiked Amount 3297.066			Recovery =	59.825%		
16) 2-Fluorobiphenyl(S)	7.45	172	3654960	1463.85	ppb	0.00
Spiked Amount 3297.066			Recovery =	44.398%		
22) 2,4,6-Tribromophenol(S)	9.22	330	903895	4132.56	ppb	0.00
Spiked Amount 6594.131			Recovery =	62.670%		
28) Terphenyl-D14(S)	11.74	244	5382652	2374.70	ppb	0.00
Spiked Amount 3297.066			Recovery =	72.025%		
Target Compounds						Qvalue
21) Diethyl phthalate	8.74	149	3234	19.02	ppb	98
30) Bis (2-ethylhexyl) phthala	12.82	149	11749	5.21	ppb	97

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y061.D Vial: 61
Acq On : 28 May 15 17:46 Operator: csc
Sample : AZ16141S01 1/30.33G Inst : Yoda
Misc : soil Multiplr: 32.97

Quant Time: Jun 2 10:20 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration





#21
Diethyl phthalate
Concen: 19.02 ppb
RT: 8.74 min Scan# 798
Delta R.T. 0.01 min
Lab File: 0520Y061.D
Acq: 28 May 15 17:46

Tgt Ion:149 Resp: 3234

Ion	Ratio	Lower	Upper
149	100		
177	22.0	15.1	28.0
150	13.3	8.3	15.5





#30
Bis (2-ethylhexyl) phthalate
Concen: 5.21 ppb
RT: 12.82 min Scan# 1237
Delta R.T. 0.00 min
Lab File: 0520Y061.D
Acq: 28 May 15 17:46

Tgt Ion:149 Resp: 11749

Ion	Ratio	Lower	Upper
149	100		
167	35.1	23.1	42.9
279	9.9	7.1	13.3



EPA METHOD 8270D
Semivolatile Organic Compounds
Calibration Data

APPL, INC.

Form 6
Initial Calibration

Lab Name: APPL, Inc. SDG No:
Case No: Initial Cal. Date: 5/20/2015
Matrix: Instrument: Yoda Initials:

			0520Y006.D	0520Y007.D	0520Y008.D	0520Y009 D	0520Y010.D	0520Y011.D	0520Y012.D	0520Y013.D						
		Compound	5	10	20	40	50	60	80	100			Avg	%RSD		
1	I	1,4-dichlorobenzene-D4(IS)														
2	SL	2-Fluorophenol (S)	1.271	1.563	1.753	1.394	1.441	1.365	1.526	1.482			1.5	9.9	SL	0.994
3	SL	Phenol-D6 (S)	3.382	2.112	2.393	1.845	1.908	1.828	2.061	2.018			2.2	23	SL	0.992
4	*TM	Phenol	2.017	2.265	2.622	2.006	2.131	2.062	2.317	2.269			2.2	9.3	*TM	
5	*TM	1,4-DCB	1.832	2.003	2.305	1.777	1.864	1.769	1.907	1.878			1.9	9.1	*TM	
6	**TM	n-Nitrosodi-n-propylamine	0.7766	0.9711	1.105	0.8531	1.104	1.078	1.211	1.199			1.0	15	**TM	
7	I	Napthalene-D8(IS)														
8	S	Nitrobenzene-D5(S)	0.3446	0.4381	0.4895	0.3927	0.4074	0.3841	0.4259	0.4042			0.41	10	S	
9	*TM	2-Nitrophenol	0.1717	0.2118	0.2471	0.2039	0.2166	0.2034	0.2252	0.2187			0.21	10	*TM	
10	*TM	2,4-Dichlorophenol	0.5642	0.6336	0.7405	0.5908	0.6300	0.5492	0.6545	0.6392			0.63	9.6	*TM	
11	*TM	Hexachlorobutadiene	0.1647	0.1793	0.1992	0.1584	0.1685	0.1522	0.1680	0.1581			0.17	8.8	*TM	
12	*TM	4-Chloro-3-methylphenol	0.0236	0.0254	0.0286	0.0221	0.0227	0.0223	0.0243	0.0232			0.02	8.9	*TM	
13	I	Acenaphthene-D10(IS)														
14	**TMQ	Hexachlorocyclopentadiene	0.0044	0.0290	0.0704	0.1202	0.1586	0.1488	0.1594	0.1694			0.11	60	**TMQ	0.995
15	*TM	2,4,6-Trichlorophenol	0.3268	0.3622	0.4200	0.3384	0.3616	0.3330	0.3735	0.3510			0.36	8.3	*TM	
16	S	2-Fluorobiphenyl(S)	1.437	1.655	1.802	1.453	1.509	1.386	1.554	1.454			1.5	9.0	S	
17	TM	Dimethyl phthalate	1.459	1.611	1.803	1.414	1.555	1.463	1.640	1.600			1.6	8.0	TM	
18	*TM	Acenaphthene	1.245	1.374	1.532	1.223	1.316	1.220	1.367	1.279			1.3	7.9	*TM	
19	**TMQ	2,4-Dinitrophenol	0.0026	0.0242	0.1103	0.1202	0.1438	0.1470	0.1779	0.1805			0.11	59	**TMQ	0.996
20	**TM	4-Nitrophenol	0.3115	0.3716	0.4652	0.3603	0.3844	0.3658	0.4580	0.4232			0.39	13	**TM	
21	TML	Diethyl phthalate	1.375	1.610	1.829	1.464	1.532	1.460	1.656	1.592			1.6	9.0	TML	0.993
22	SL	2,4,6-Tribromophenol(S)	0.1897	0.1306	0.1456	0.1207	0.1282	0.1183	0.1430	0.1379			0.14	16	SL	0.988
23	I	Phenanthrene-D10(IS)														
24	*TM	n-Nitrosodiphenylamine	1.099	1.241	1.437	1.165	1.246	1.167	1.334	1.257			1.2	8.5	*TM	
25	*TMQ	Pentachlorophenol	0.0286	0.0379	0.0481	0.0469	0.0558	0.0560	0.0594	0.0589			0.05	22	*TMQ	0.997
26	TM	Di-n-butylphthalate	1.205	1.443	1.711	1.441	1.518	1.353	1.396	1.318			1.4	11	TM	
27	I	Chrysene-D12(IS)														
28	S	Terphenyl-D14(S)	0.9384	1.075	1.153	0.9403	0.9946	0.9238	0.9484	0.9601			0.99	8.2	S	
29	TM	Butyl benzylphthalate	0.6027	0.7028	0.8216	0.6812	0.7459	0.6870	0.7207	0.7064			0.71	8.7	TM	
30	TM	Bis (2-ethylhexyl) phthalate	0.7826	0.9641	1.153	0.9922	1.056	0.9378	1.017	0.9875			0.99	11	TM	
31	*TM	Di-n-octylphthalate	1.290	1.662	1.776	1.707	1.893	1.665	1.735	1.363			1.6	13	*TM	
32	I	Perylene-D12(IS)														
33																
34																
35																

Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y006.D Vial: 6
Acq On : 22 May 15 10:46 Operator: csc
Sample : 5ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	776980	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	3450438	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	1885022	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	3305748	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	2785847	40.00	ppb		0.00
32) Perylene-D12(IS)	14.65	264	1640376	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.12	112	246795	8.57	ppb		0.07
Spiked Amount 200.000			Recovery =	4.283%			
3) Phenol-D6 (S)	5.19	99	328498	7.39	ppb		0.08
Spiked Amount 200.000			Recovery =	3.695%			
8) Nitrobenzene-D5(S)	5.52	82	148646	4.19	ppb		0.02
Spiked Amount 100.000			Recovery =	4.194%			
16) 2-Fluorobiphenyl(S)	7.46	172	338544	4.69	ppb		0.00
Spiked Amount 100.000			Recovery =	4.692%			
22) 2,4,6-Tribromophenol(S)	9.24	330	44690	9.38	ppb		0.02
Spiked Amount 200.000			Recovery =	4.689%			
28) Terphenyl-D14(S)	11.75	244	326790	4.73	ppb		0.00
Spiked Amount 100.000			Recovery =	4.732%			
Target Compounds							Qvalue
4) Phenol	5.20	94	195890	4.56	ppb		96
5) 1,4-DCB	4.81	146	177898	4.78	ppb		96
6) n-Nitrosodi-n-propylamine	5.29	43	75421	3.74	ppb		95
9) 2-Nitrophenol	5.89	139	74035	4.04	ppb		91
10) 2,4-Dichlorophenol	6.52	162	243342	4.51	ppb		99
11) Hexachlorobutadiene	6.36	225	71039	4.89	ppb		98
12) 4-Chloro-3-methylphenol	7.21	107	10165	4.91	ppb	#	17
14) Hexachlorocyclopentadiene	7.17	237	1040	7.78	ppb	#	89
15) 2,4,6-Trichlorophenol	7.58	196	77012	4.56	ppb		94
17) Dimethyl phthalate	7.97	163	343740	4.65	ppb		94
18) Acenaphthene	8.26	154	293434	4.72	ppb		97
19) 2,4-Dinitrophenol	8.65	184	608	6.48	ppb		100
20) 4-Nitrophenol	9.20	65	73389	3.97	ppb		97
21) Diethyl phthalate	8.73	149	324049	4.86	ppb		97
24) n-Nitrosodiphenylamine	10.37	167	454279	4.42	ppb		99
25) Pentachlorophenol	9.98	266	11807	5.71	ppb		88
26) Di-n-butylphthalate	10.54	149	497919	4.23	ppb		98
29) Butyl benzylphthalate	12.27	149	209883	4.25	ppb		89
30) Bis (2-ethylhexyl) phthala	12.81	149	272539	3.97	ppb	#	94
31) Di-n-octylphthalate	13.50	149	449254	3.94	ppb		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y006.D Vial: 6
Acq On : 22 May 15 10:46 Operator: csc
Sample : 5ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y007.D Vial: 7
Acq On : 22 May 15 11:15 Operator: csc
Sample : 10ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	711729	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.24	136	3204797	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	1853344	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	3328897	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.00	240	3013322	40.00	ppb		0.00
32) Perylene-D12(IS)	14.65	264	1886518	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.10	112	556175	21.18	ppb		0.06
Spiked Amount 200.000			Recovery =	10.593%			
3) Phenol-D6 (S)	5.17	99	751709	20.26	ppb		0.06
Spiked Amount 200.000			Recovery =	10.130%			
8) Nitrobenzene-D5(S)	5.51	82	351037	10.66	ppb		0.01
Spiked Amount 100.000			Recovery =	10.665%			
16) 2-Fluorobiphenyl(S)	7.45	172	766903	10.81	ppb		0.00
Spiked Amount 100.000			Recovery =	10.810%			
22) 2,4,6-Tribromophenol(S)	9.23	330	121047	21.54	ppb		0.01
Spiked Amount 200.000			Recovery =	10.771%			
28) Terphenyl-D14(S)	11.74	244	809651	10.84	ppb		0.00
Spiked Amount 100.000			Recovery =	10.838%			
Target Compounds							Qvalue
4) Phenol	5.19	94	403000	10.24	ppb		94
5) 1,4-DCB	4.82	146	356311	10.45	ppb		99
6) n-Nitrosodi-n-propylamine	5.28	43	172784	9.36	ppb		96
9) 2-Nitrophenol	5.87	139	169704	9.98	ppb		97
10) 2,4-Dichlorophenol	6.51	162	507616	10.13	ppb		97
11) Hexachlorobutadiene	6.35	225	143633	10.64	ppb		99
12) 4-Chloro-3-methylphenol	7.21	107	20362	10.58	ppb	#	16
14) Hexachlorocyclopentadiene	7.17	237	13459	9.52	ppb		96
15) 2,4,6-Trichlorophenol	7.57	196	167811	10.10	ppb		94
17) Dimethyl phthalate	7.97	163	746233	10.27	ppb	#	84
18) Acenaphthene	8.25	154	636435	10.41	ppb		98
19) 2,4-Dinitrophenol	8.56	184	11232	8.19	ppb		100
20) 4-Nitrophenol	9.19	65	172161	9.47	ppb		98
21) Diethyl phthalate	8.73	149	745745	10.65	ppb		96
24) n-Nitrosodiphenylamine	10.36	167	1032422	9.98	ppb		99
25) Pentachlorophenol	9.97	266	31530	10.02	ppb		94
26) Di-n-butylphthalate	10.54	149	1200925	10.14	ppb	#	98
29) Butyl benzylphthalate	12.27	149	529447	9.92	ppb		99
30) Bis (2-ethylhexyl) phthala	12.82	149	726265	9.78	ppb		98
31) Di-n-octylphthalate	13.49	149	1251846	10.15	ppb		99

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y007.D Vial: 7
Acq On : 22 May 15 11:15 Operator: csc
Sample : 10ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y008.D Vial: 8
Acq On : 22 May 15 11:45 Operator: csc
Sample : 20ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	536321	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.24	136	2460518	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.21	164	1459456	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	2689962	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.00	240	2633883	40.00	ppb		0.00
32) Perylene-D12(IS)	14.65	264	1379249	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.08	112	940205	47.62	ppb		0.04
Spiked Amount 200.000			Recovery =	23.810%			
3) Phenol-D6 (S)	5.15	99	1283669	47.43	ppb		0.04
Spiked Amount 200.000			Recovery =	23.717%			
8) Nitrobenzene-D5(S)	5.50	82	602243	23.83	ppb		0.00
Spiked Amount 100.000			Recovery =	23.831%			
16) 2-Fluorobiphenyl(S)	7.45	172	1315033	23.54	ppb		0.00
Spiked Amount 100.000			Recovery =	23.539%			
22) 2,4,6-Tribromophenol(S)	9.23	330	212482	45.01	ppb		0.00
Spiked Amount 200.000			Recovery =	22.508%			
28) Terphenyl-D14(S)	11.74	244	1518105	23.25	ppb		0.00
Spiked Amount 100.000			Recovery =	23.249%			
Target Compounds							Qvalue
4) Phenol	5.17	94	703132	23.72	ppb		96
5) 1,4-DCB	4.82	146	618042	24.05	ppb		99
6) n-Nitrosodi-n-propylamine	5.28	43	296337	21.31	ppb		92
9) 2-Nitrophenol	5.87	139	303943	23.28	ppb		89
10) 2,4-Dichlorophenol	6.51	162	911038	23.69	ppb		96
11) Hexachlorobutadiene	6.35	225	245022	23.63	ppb		99
12) 4-Chloro-3-methylphenol	7.20	107	35158	23.80	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	51348	16.63	ppb		99
15) 2,4,6-Trichlorophenol	7.57	196	306460	23.43	ppb		98
17) Dimethyl phthalate	7.97	163	1315746	23.00	ppb	#	80
18) Acenaphthene	8.25	154	1117896	23.22	ppb		97
19) 2,4-Dinitrophenol	8.53	184	80477	21.77	ppb	#	100
20) 4-Nitrophenol	9.18	65	339484	23.71	ppb		99
21) Diethyl phthalate	8.73	149	1334829	23.52	ppb		95
24) n-Nitrosodiphenylamine	10.36	167	1932983	23.12	ppb		99
25) Pentachlorophenol	9.96	266	64717	20.45	ppb		95
26) Di-n-butylphthalate	10.54	149	2300904	24.04	ppb	#	97
29) Butyl benzylphthalate	12.27	149	1081948	23.19	ppb		99
30) Bis (2-ethylhexyl) phthala	12.82	149	1517895	23.37	ppb		98
31) Di-n-octylphthalate	13.49	149	2339398	21.71	ppb		99

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y008.D Vial: 8
Acq On : 22 May 15 11:45 Operator: csc
Sample : 20ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y009.D Vial: 9
Acq On : 22 May 15 12:14 Operator: csc
Sample : 40ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	781016	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	3505306	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2036207	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	3623295	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3382326	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2155209	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.05	112	2177862	75.79	ppb		0.00
Spiked Amount 200.000			Recovery =	37.897%			
3) Phenol-D6 (S)	5.12	99	2882043	73.78	ppb		0.01
Spiked Amount 200.000			Recovery =	36.892%			
8) Nitrobenzene-D5(S)	5.50	82	1376418	38.23	ppb		0.00
Spiked Amount 100.000			Recovery =	38.232%			
16) 2-Fluorobiphenyl(S)	7.45	172	2958571	37.96	ppb		0.00
Spiked Amount 100.000			Recovery =	37.957%			
22) 2,4,6-Tribromophenol(S)	9.22	330	491603	73.04	ppb		0.00
Spiked Amount 200.000			Recovery =	36.519%			
28) Terphenyl-D14(S)	11.74	244	3180432	37.93	ppb		0.00
Spiked Amount 100.000			Recovery =	37.929%			
Target Compounds							Qvalue
4) Phenol	5.14	94	1566459	36.29	ppb		99
5) 1,4-DCB	4.81	146	1387561	37.08	ppb		100
6) n-Nitrosodi-n-propylamine	5.29	43	666311	32.90	ppb	#	84
9) 2-Nitrophenol	5.87	139	714614	38.41	ppb		87
10) 2,4-Dichlorophenol	6.49	162	2071011	37.80	ppb		96
11) Hexachlorobutadiene	6.35	225	555140	37.59	ppb		99
12) 4-Chloro-3-methylphenol	7.20	107	77506	36.83	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	244842	37.11	ppb		99
15) 2,4,6-Trichlorophenol	7.56	196	689043	37.76	ppb		95
17) Dimethyl phthalate	7.97	163	2879340	36.07	ppb		92
18) Acenaphthene	8.26	154	2490439	37.08	ppb		97
19) 2,4-Dinitrophenol	8.52	184	244845	37.59	ppb	#	100
20) 4-Nitrophenol	9.17	65	733688	36.72	ppb		98
21) Diethyl phthalate	8.73	149	2980636	37.33	ppb		99
24) n-Nitrosodiphenylamine	10.35	167	4220469	37.48	ppb		99
25) Pentachlorophenol	9.95	266	169881	36.19	ppb		98
26) Di-n-butylphthalate	10.54	149	5221513	40.51	ppb		97
29) Butyl benzylphthalate	12.27	149	2303882	38.45	ppb		92
30) Bis (2-ethylhexyl) phthala	12.82	149	3355899	40.24	ppb	#	94
31) Di-n-octylphthalate	13.49	149	5773216	41.72	ppb		99

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y009.D Vial: 9
Acq On : 22 May 15 12:14 Operator: csc
Sample : 40ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y010.D Vial: 10
Acq On : 22 May 15 12:43 Operator: csc
Sample : 50ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	789097	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	3524029	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2025704	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	3568969	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3316564	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2280872	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.04	112	2842780	97.94	ppb		0.00
Spiked Amount 200.000			Recovery =	48.971%			
3) Phenol-D6 (S)	5.11	99	3764614	95.74	ppb		0.00
Spiked Amount 200.000			Recovery =	47.872%			
8) Nitrobenzene-D5(S)	5.50	82	1794638	49.58	ppb		0.00
Spiked Amount 100.000			Recovery =	49.584%			
16) 2-Fluorobiphenyl(S)	7.45	172	3820599	49.27	ppb		0.00
Spiked Amount 100.000			Recovery =	49.271%			
22) 2,4,6-Tribromophenol(S)	9.22	330	649441	96.19	ppb		0.00
Spiked Amount 200.000			Recovery =	48.094%			
28) Terphenyl-D14(S)	11.74	244	4123162	50.15	ppb		0.00
Spiked Amount 100.000			Recovery =	50.147%			
Target Compounds							Qvalue
4) Phenol	5.13	94	2101573	48.18	ppb		99
5) 1,4-DCB	4.81	146	1838737	48.63	ppb		99
6) n-Nitrosodi-n-propylamine	5.29	43	1088971	53.22	ppb		93
9) 2-Nitrophenol	5.87	139	954247	51.02	ppb		88
10) 2,4-Dichlorophenol	6.49	162	2775088	50.38	ppb		98
11) Hexachlorobutadiene	6.35	225	742040	49.98	ppb		97
12) 4-Chloro-3-methylphenol	7.20	107	99869	47.20	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	401575	54.58	ppb		99
15) 2,4,6-Trichlorophenol	7.56	196	915533	50.44	ppb		96
17) Dimethyl phthalate	7.97	163	3937522	49.58	ppb		96
18) Acenaphthene	8.26	154	3333346	49.88	ppb		97
19) 2,4-Dinitrophenol	8.51	184	364103	50.52	ppb	#	100
20) 4-Nitrophenol	9.17	65	973429	48.97	ppb		98
21) Diethyl phthalate	8.73	149	3880357	48.68	ppb		98
24) n-Nitrosodiphenylamine	10.35	167	5556568	50.09	ppb		98
25) Pentachlorophenol	9.94	266	248731	51.30	ppb		96
26) Di-n-butylphthalate	10.54	149	6772183	53.34	ppb		97
29) Butyl benzylphthalate	12.27	149	3092083	52.63	ppb		88
30) Bis (2-ethylhexyl) phthala	12.81	149	4376747	53.52	ppb	#	92
31) Di-n-octylphthalate	13.49	149	7847240	57.84	ppb		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y010.D Vial: 10
Acq On : 22 May 15 12:43 Operator: csc
Sample : 50ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y011.D Vial: 11
Acq On : 22 May 15 13:12 Operator: csc
Sample : 60ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	745684	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	3425046	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2045435	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	3680591	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3637880	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2525912	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.03	112	3054194	111.36	ppb		0.00
Spiked Amount 200.000			Recovery =	55.681%			
3) Phenol-D6 (S)	5.10	99	4088841	110.22	ppb		0.00
Spiked Amount 200.000			Recovery =	55.111%			
8) Nitrobenzene-D5(S)	5.51	82	1973572	56.10	ppb		0.00
Spiked Amount 100.000			Recovery =	56.103%			
16) 2-Fluorobiphenyl(S)	7.45	172	4252492	54.31	ppb		0.00
Spiked Amount 100.000			Recovery =	54.312%			
22) 2,4,6-Tribromophenol(S)	9.22	330	725939	106.22	ppb		0.00
Spiked Amount 200.000			Recovery =	53.109%			
28) Terphenyl-D14(S)	11.75	244	5041242	55.90	ppb		0.00
Spiked Amount 100.000			Recovery =	55.897%			
Target Compounds							Qvalue
4) Phenol	5.13	94	2306105	55.95	ppb		99
5) 1,4-DCB	4.81	146	1978564	55.37	ppb		99
6) n-Nitrosodi-n-propylamine	5.29	43	1205905	62.36	ppb		98
9) 2-Nitrophenol	5.87	139	1044787	57.48	ppb		89
10) 2,4-Dichlorophenol	6.49	162	2821647	52.70	ppb		98
11) Hexachlorobutadiene	6.35	225	781810	54.18	ppb		100
12) 4-Chloro-3-methylphenol	7.21	107	114328	55.60	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	456490	59.93	ppb		98
15) 2,4,6-Trichlorophenol	7.56	196	1021681	55.74	ppb		97
17) Dimethyl phthalate	7.98	163	4489137	55.98	ppb	#	80
18) Acenaphthene	8.26	154	3743229	55.47	ppb		97
19) 2,4-Dinitrophenol	8.51	184	451131	58.75	ppb	#	100
20) 4-Nitrophenol	9.18	65	1122316	55.92	ppb		99
21) Diethyl phthalate	8.74	149	4480539	55.59	ppb		97
24) n-Nitrosodiphenylamine	10.35	167	6444197	56.33	ppb		99
25) Pentachlorophenol	9.94	266	308980	60.45	ppb		96
26) Di-n-butylphthalate	10.54	149	7468640	57.04	ppb		99
29) Butyl benzylphthalate	12.28	149	3748651	58.17	ppb		92
30) Bis (2-ethylhexyl) phthala	12.81	149	5117380	57.05	ppb	#	91
31) Di-n-octylphthalate	13.50	149	9087213	61.06	ppb		96

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y011.D Vial: 11
Acq On : 22 May 15 13:12 Operator: csc
Sample : 60ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y012.D Vial: 12
Acq On : 22 May 15 13:42 Operator: csc
Sample : 80ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.79	152	471386	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.24	136	2201686	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	1325479	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	2575106	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	2797916	40.00	ppb		0.00
32) Perylene-D12(IS)	14.65	264	2057568	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.04	112	2878109	166.04	ppb		0.00
Spiked Amount 200.000			Recovery =	83.022%			
3) Phenol-D6 (S)	5.10	99	3885528	166.30	ppb		0.00
Spiked Amount 200.000			Recovery =	83.148%			
8) Nitrobenzene-D5(S)	5.50	82	1875307	82.93	ppb		0.00
Spiked Amount 100.000			Recovery =	82.931%			
16) 2-Fluorobiphenyl(S)	7.45	172	4118879	81.18	ppb		0.00
Spiked Amount 100.000			Recovery =	81.178%			
22) 2,4,6-Tribromophenol(S)	9.23	330	757950	169.64	ppb		0.00
Spiked Amount 200.000			Recovery =	84.823%			
28) Terphenyl-D14(S)	11.74	244	5307158	76.51	ppb		0.00
Spiked Amount 100.000			Recovery =	76.512%			
Target Compounds							Qvalue
4) Phenol	5.13	94	2184278	83.83	ppb		98
5) 1,4-DCB	4.81	146	1797847	79.60	ppb		98
6) n-Nitrosodi-n-propylamine	5.29	43	1141626	93.39	ppb	#	86
9) 2-Nitrophenol	5.87	139	991785	84.88	ppb		99
10) 2,4-Dichlorophenol	6.49	162	2881986	83.74	ppb		95
11) Hexachlorobutadiene	6.35	225	739732	79.74	ppb		100
12) 4-Chloro-3-methylphenol	7.20	107	107136	81.05	ppb	#	13
14) Hexachlorocyclopentadiene	7.17	237	422535	79.60	ppb		100
15) 2,4,6-Trichlorophenol	7.56	196	990190	83.36	ppb		99
17) Dimethyl phthalate	7.97	163	4348241	83.68	ppb		94
18) Acenaphthene	8.26	154	3625113	82.90	ppb		96
19) 2,4-Dinitrophenol	8.51	184	471642	83.27	ppb	#	100
20) 4-Nitrophenol	9.17	65	1214197	93.35	ppb		98
21) Diethyl phthalate	8.73	149	4389637	83.77	ppb		98
24) n-Nitrosodiphenylamine	10.35	167	6872196	85.86	ppb		98
25) Pentachlorophenol	9.93	266	306060	82.03	ppb		98
26) Di-n-butylphthalate	10.54	149	7191780	78.50	ppb		97
29) Butyl benzylphthalate	12.27	149	4033100	81.38	ppb		90
30) Bis (2-ethylhexyl) phthala	12.82	149	5693453	82.53	ppb	#	92
31) Di-n-octylphthalate	13.50	149	9711278	84.84	ppb		95

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y012.D Vial: 12
Acq On : 22 May 15 13:42 Operator: csc
Sample : 80ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:47 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y013.D Vial: 13
Acq On : 22 May 15 14:11 Operator: csc
Sample : 100ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:55 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	476024	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	2309482	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	1435509	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	2809978	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3038455	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2240653	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.03	112	3526527	201.49	ppb		-0.02
Spiked Amount 200.000			Recovery =	100.744%			
3) Phenol-D6 (S)	5.09	99	4804030m	203.87	ppb		-0.02
Spiked Amount 200.000			Recovery =	101.937%			
8) Nitrobenzene-D5(S)	5.50	82	2333800	98.39	ppb		0.00
Spiked Amount 100.000			Recovery =	98.390%			
16) 2-Fluorobiphenyl(S)	7.45	172	5217364	94.95	ppb		0.00
Spiked Amount 100.000			Recovery =	94.947%			
22) 2,4,6-Tribromophenol(S)	9.22	330	989432	203.98	ppb		0.00
Spiked Amount 200.000			Recovery =	101.990%			
28) Terphenyl-D14(S)	11.75	244	7293230	96.82	ppb		0.00
Spiked Amount 100.000			Recovery =	96.821%			
Target Compounds							Qvalue
4) Phenol	5.11	94	2700240	102.62	ppb		96
5) 1,4-DCB	4.81	146	2234575	97.97	ppb		96
6) n-Nitrosodi-n-propylamine	5.29	43	1427232	115.62	ppb		98
9) 2-Nitrophenol	5.87	139	1262527	103.01	ppb		87
10) 2,4-Dichlorophenol	6.49	162	3690751	102.24	ppb		95
11) Hexachlorobutadiene	6.35	225	913036	93.83	ppb		99
12) 4-Chloro-3-methylphenol	7.20	107	133875	96.56	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	607791	99.76	ppb		100
15) 2,4,6-Trichlorophenol	7.56	196	1259538	97.91	ppb		98
17) Dimethyl phthalate	7.98	163	5743504	102.06	ppb	#	84
18) Acenaphthene	8.26	154	4588873	96.90	ppb		97
19) 2,4-Dinitrophenol	8.51	184	647728	98.67	ppb	#	100
20) 4-Nitrophenol	9.18	65	1518697	107.82	ppb		98
21) Diethyl phthalate	8.74	149	5715083	100.60	ppb		99
24) n-Nitrosodiphenylamine	10.35	167	8829418	101.10	ppb		98
25) Pentachlorophenol	9.93	266	413809	98.82	ppb		98
26) Di-n-butylphthalate	10.54	149	9256493	92.59	ppb		97
29) Butyl benzylphthalate	12.28	149	5365726	99.70	ppb		93
30) Bis (2-ethylhexyl) phthala	12.82	149	7501525	100.13	ppb		99
31) Di-n-octylphthalate	13.49	149	10350012	83.26	ppb		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y013.D Vial: 13
Acq On : 22 May 15 14:11 Operator: csc
Sample : 100ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 10:55 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Initial Calibration



Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y013.D Vial: 13
Acq On : 22 May 15 14:11 Operator: csc
Sample : 100ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00
Quant Time: Jun 2 10:47 2015 Quant Results File: temp.res

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Multiple Level Calibration





(3) Phenol-D6 (S) (S)

5.09min 122.1724ppb

response 2890141

Ion	Exp%	Act%
99.00	100	100
71.00	32.80	36.88
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y013.D Vial: 13
Acq On : 22 May 15 14:11 Operator: csc
Sample : 100ug/ml SVOC 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00
Quant Time: Jun 2 10:55 2015 Quant Results File: temp.res

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 10:48:52 2015
Response via : Multiple Level Calibration





(3) Phenol-D6 (S) (S)

5.09min 203.8728ppb m

response 4804030

Ion	Exp%	Act%
99.00	100	100
71.00	32.80	22.19#
0.00	0.00	0.00
0.00	0.00	0.00

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No:
Date Analyzed: 5/22/2015
Instrument: Yoda
Initial Cal. Date: 5/20/2015
Data File: 0520Y014.D

		Compound	MEAN	CCRF	%D		%Drift	
1	*TM	Phenol	2.211	2.067	6.5	*TM		
2	*TM	1,4-DCB	1.917	2.011	4.9	*TM		
3	**TM	n-Nitrosodi-n-propylamine	1.037	0.8725	16	**TM		
4	*TM	2-Nitrophenol	0.2123	0.2020	4.8	*TM		
5	*TM	2,4-Dichlorophenol	0.6253	0.5989	4.2	*TM		
6	*TM	Hexachlorobutadiene	0.1685	0.1790	6.2	*TM		
7	*TM	4-Chloro-3-methylphenol	0.0240	0.0242	0.85	*TM		
8	**TMQ	Hexachlorocyclopentadiene	0.1075	0.1099	2.2	**TMQ	18	
9	*TM	2,4,6-Trichlorophenol	0.3583	0.3248	9.3	*TM		
10	TM	Dimethyl phthalate	1.568	1.490	5.0	TM		
11	*TM	Acenaphthene	1.320	1.261	4.5	*TM		
12	**TMQ	2,4-Dinitrophenol	0.1133	0.1263	11	**TMQ	8.4	
13	**TM	4-Nitrophenol	0.3925	0.3505	11	**TM		
14	TML	Diethyl phthalate	1.565	1.502	4.0	TML	4.5	
15	*TM	n-Nitrosodiphenylamine	1.243	1.216	2.2	*TM		
16	*TMQ	Pentachlorophenol	0.0489	0.0367	25	*TMQ	29	*
17	TM	Di-n-butylphthalate	1.423	1.386	2.6	TM		
18	TM	Butyl benzylphthalate	0.7085	0.6583	7.1	TM		
19	TM	Bis (2-ethylhexyl) phthalate	0.9863	0.9236	6.4	TM		
20	*TM	Di-n-octylphthalate	1.636	1.704	4.2	*TM		
21								
22								
23								
24								
25								
26								
27								
28								
29								
30								
31								
32								
33								
34								
35								
36								
37								
38								
39								
40								
		Average			6.9			

Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y014.D Vial: 14
Acq On : 22 May 15 14:41 Operator: csc
Sample : 50ug/ml SVOC SS 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 11:17 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 11:00:10 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	477080	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	2254044	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	1379161	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	2658514	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	2791249	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2129348	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	0.00	112	0	0.00	ppb		
Spiked Amount 200.000			Recovery =	0.000%			
3) Phenol-D6 (S)	0.00	99	0	0.00	ppb		
Spiked Amount 200.000			Recovery =	0.000%			
8) Nitrobenzene-D5(S)	0.00	82	0	0.00	ppb		
Spiked Amount 100.000			Recovery =	0.000%			
16) 2-Fluorobiphenyl(S)	0.00	172	0	0.00	ppb		
Spiked Amount 100.000			Recovery =	0.000%			
22) 2,4,6-Tribromophenol(S)	0.00	330	0	0.00	ppb		
Spiked Amount 200.000			Recovery =	0.000%			
28) Terphenyl-D14(S)	11.75	244	1177	0.02	ppb		0.00
Spiked Amount 100.000			Recovery =	0.017%			
Target Compounds							Qvalue
4) Phenol	5.17	94	1232753	46.75	ppb		89
5) 1,4-DCB	4.81	146	1199318	52.46	ppb		98
6) n-Nitrosodi-n-propylamine	5.28	43	520296	42.06	ppb		92
9) 2-Nitrophenol	5.87	139	569254	47.59	ppb		89
10) 2,4-Dichlorophenol	6.49	162	1687426	47.89	ppb		96
11) Hexachlorobutadiene	6.35	225	504457	53.12	ppb		100
12) 4-Chloro-3-methylphenol	7.21	107	68238	50.43	ppb	#	15
14) Hexachlorocyclopentadiene	7.17	237	189419	41.04	ppb		100
15) 2,4,6-Trichlorophenol	7.56	196	559947	45.33	ppb		94
17) Dimethyl phthalate	7.97	163	2568426	47.50	ppb		92
18) Acenaphthene	8.26	154	2173065	47.76	ppb		98
19) 2,4-Dinitrophenol	8.52	184	217659	45.81	ppb	#	100
20) 4-Nitrophenol	9.17	65	604234	44.65	ppb		97
21) Diethyl phthalate	8.73	149	2590069	47.74	ppb		100
24) n-Nitrosodiphenylamine	10.35	167	4039348	48.89	ppb		99
25) Pentachlorophenol	9.95	266	121801	35.47	ppb		98
26) Di-n-butylphthalate	10.54	149	4606544	48.70	ppb	#	97
29) Butyl benzylphthalate	12.27	149	2297025	46.46	ppb		88
30) Bis (2-ethylhexyl) phthala	12.81	149	3222474	46.82	ppb	#	93
31) Di-n-octylphthalate	13.49	149	5947012	52.08	ppb		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y014.D Vial: 14
Acq On : 22 May 15 14:41 Operator: csc
Sample : 50ug/ml SVOC SS 05/22/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 11:17 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 11:00:10 2015
Response via : Initial Calibration



Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No:
Date Analyzed: 5/28/2015
Instrument: Yoda
Initial Cal. Date: 5/20/2015
Data File: 0520Y048.D

		Compound	MEAN	CCRF	%D		%Drift
1	I	1,4-dichlorobenzene-D4(IS)	ISTD			I	
2	SL	2-Fluorophenol (S)	1.474	1.378	6.6	SL	6.4
3	SL	Phenol-D6 (S)	2.194	1.819	17	SL	8.8
4	*TM	Phenol	2.211	2.007	9.2	*TM	
5	*TM	1,4-DCB	1.917	1.794	6.4	*TM	
6	**TM	n-Nitrosodi-n-propylamine	1.037	1.042	0.48	**TM	
7	I	Napthalene-D8(IS)	ISTD			I	
8	S	Nitrobenzene-D5(S)	0.4108	0.3948	3.9	S	
9	*TM	2-Nitrophenol	0.2123	0.2025	4.6	*TM	
10	*TM	2,4-Dichlorophenol	0.6253	0.5768	7.7	*TM	
11	*TM	Hexachlorobutadiene	0.1685	0.1598	5.2	*TM	
12	*TM	4-Chloro-3-methylphenol	0.0240	0.0214	11	*TM	
13	I	Acenaphthene-D10(IS)	ISTD			I	
14	**TMQ	Hexachlorocyclopentadiene	0.1075	0.1627	51	**TMQ	11
15	*TM	2,4,6-Trichlorophenol	0.3583	0.3409	4.9	*TM	
16	S	2-Fluorobiphenyl(S)	1.531	1.500	2.0	S	
17	TM	Dimethyl phthalate	1.568	1.398	11	TM	
18	*TM	Acenaphthene	1.320	1.240	6.0	*TM	
19	**TMQ	2,4-Dinitrophenol	0.1133	0.1283	13	**TMQ	7.3
20	**TM	4-Nitrophenol	0.3925	0.3435	12	**TM	
21	TML	Diethyl phthalate	1.565	1.416	9.5	TML	9.9
22	SL	2,4,6-Tribromophenol(S)	0.1392	0.1158	17	SL	13
23	I	Phenanthrene-D10(IS)	ISTD			I	
24	*TM	n-Nitrosodiphenylamine	1.243	1.157	6.9	*TM	
25	*TMQ	Pentachlorophenol	0.0489	0.0586	20	*TMQ	7.3
26	TM	Di-n-butylphthalate	1.423	1.465	2.9	TM	
27	I	Chrysene-D12(IS)	ISTD			I	
28	S	Terphenyl-D14(S)	0.9916	1.030	3.9	S	
29	TM	Butyl benzylphthalate	0.7085	0.7305	3.1	TM	
30	TM	Bis (2-ethylhexyl) phthalate	0.9863	1.041	5.5	TM	
31	*TM	Di-n-octylphthalate	1.636	1.725	5.4	*TM	
32	I	Perylene-D12(IS)	ISTD			I	
33							
34							
35							
36							
37							
38							
39							
40							
		Average			9.5		

Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y048.D Vial: 48
Acq On : 28 May 15 10:54 Operator: csc
Sample : 50ug/ml SVOC 05/28/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 11:26 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 11:00:10 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	937800	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.24	136	4119963	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2216054	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	3645219	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.00	240	2955804	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	1573179	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	4.01	112	3229987	93.63	ppb		-0.04
Spiked Amount 200.000			Recovery =	46.817%			
3) Phenol-D6 (S)	5.07	99	4263552	91.18	ppb		-0.04
Spiked Amount 200.000			Recovery =	45.591%			
8) Nitrobenzene-D5(S)	5.50	82	2033215	48.05	ppb		0.00
Spiked Amount 100.000			Recovery =	48.050%			
16) 2-Fluorobiphenyl(S)	7.45	172	4156205	48.99	ppb		0.00
Spiked Amount 100.000			Recovery =	48.995%			
22) 2,4,6-Tribromophenol(S)	9.22	330	641274	87.06	ppb		0.00
Spiked Amount 200.000			Recovery =	43.529%			
28) Terphenyl-D14(S)	11.74	244	3807358	51.96	ppb		0.00
Spiked Amount 100.000			Recovery =	51.958%			
Target Compounds							Qvalue
4) Phenol	5.10	94	2352246	45.38	ppb		98
5) 1,4-DCB	4.82	146	2103462	46.81	ppb		99
6) n-Nitrosodi-n-propylamine	5.29	43	1221727	50.24	ppb		95
9) 2-Nitrophenol	5.86	139	1042994	47.70	ppb		97
10) 2,4-Dichlorophenol	6.49	162	2970632	46.13	ppb		98
11) Hexachlorobutadiene	6.35	225	822742	47.40	ppb		100
12) 4-Chloro-3-methylphenol	7.20	107	110074	44.50	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	450781	55.70	ppb		98
15) 2,4,6-Trichlorophenol	7.55	196	944240	47.57	ppb		98
17) Dimethyl phthalate	7.97	163	3873903	44.59	ppb		96
18) Acenaphthene	8.26	154	3434318	46.98	ppb		97
19) 2,4-Dinitrophenol	8.51	184	355370	46.35	ppb	#	100
20) 4-Nitrophenol	9.17	65	951637	43.76	ppb		99
21) Diethyl phthalate	8.73	149	3923146	45.03	ppb		99
24) n-Nitrosodiphenylamine	10.35	167	5271897	46.53	ppb		98
25) Pentachlorophenol	9.93	266	267156	53.63	ppb		96
26) Di-n-butylphthalate	10.54	149	6673345	51.46	ppb		97
29) Butyl benzylphthalate	12.27	149	2699065	51.55	ppb		90
30) Bis (2-ethylhexyl) phthala	12.82	149	3844926	52.76	ppb	#	93
31) Di-n-octylphthalate	13.49	149	6371887	52.69	ppb		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y048.D Vial: 48
Acq On : 28 May 15 10:54 Operator: csc
Sample : 50ug/ml SVOC 05/28/15 Inst : Yoda
Misc : Multiplr: 1.00

Quant Time: Jun 2 11:26 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Tue Jun 02 11:00:10 2015
Response via : Initial Calibration



EPA METHOD 8270D
Semivolatile Organic Compounds
Raw Data

APPL, INC.

Method Blank
EPA 8270D SOILS

Blank Name/QCG: **150521S-16137 - 197602**
Batch ID: #87DJU-150521A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	BIS (2-ETHYLHEXYL) PHTHALAT	0.167 U	0.66	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	BUTYL BENZYL PHTHALATE	0.167 U	0.33	0.167	0.056	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-BUTYL PHTHALATE	0.167 U	0.33	0.167	0.066	mg/kg	5/21/2015	5/28/2015
BLANK	DI-N-OCTYL PHTHALATE	0.167 U	0.33	0.167	0.058	mg/kg	5/21/2015	5/28/2015
BLANK	DIETHYL PHTHALATE	0.167 U	0.33	0.167	0.062	mg/kg	5/21/2015	5/28/2015
BLANK	DIMETHYL PHTHALATE	0.167 U	0.33	0.167	0.063	mg/kg	5/21/2015	5/28/2015
BLANK	SURROGATE: 2,4,6-TRIBROMOP	66.1	35-125			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	66.9	45-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: 2-FLUOROPHENO	70.2	35-105			%	5/21/2015	5/28/2015
BLANK	SURROGATE: NITROBENZENE-	67.7	35-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: PHENOL (S)	69.3	40-100			%	5/21/2015	5/28/2015
BLANK	SURROGATE: TERPHENYL-D14 (	86.8	30-125			%	5/21/2015	5/28/2015

Quant Method: Y0520TT.M
Run #: 0520Y055
Instrument: Yoda
Sequence: Y150520
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 12:12:08 PM

Quantitation Report (QT Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y055.D Vial: 55
Acq On : 28 May 15 14:49 Operator: csc
Sample : 150521A BLK 1/30.00G Inst : Yoda
Misc : soil Multiplr: 33.33333
MS Integration Params: rteint.p
Quant Time: Jun 2 10:36 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	934649	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.24	136	4145958	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2316228	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.94	188	4141417	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3325730	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2051463	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	3.97	112	4826646	4680.95	ppb		-0.07
Spiked Amount 6666.667			Recovery =	70.214%			
3) Phenol-D6 (S)	5.08	99	6426671	4617.42	ppb		-0.03
Spiked Amount 6666.667			Recovery =	69.261%			
8) Nitrobenzene-D5(S)	5.50	82	2882944	2256.79	ppb		0.00
Spiked Amount 3333.333			Recovery =	67.704%			
16) 2-Fluorobiphenyl(S)	7.45	172	5934326	2231.02	ppb		0.00
Spiked Amount 3333.333			Recovery =	66.931%			
22) 2,4,6-Tribromophenol(S)	9.22	330	1027477	4405.02	ppb		0.00
Spiked Amount 6666.667			Recovery =	66.075%			
28) Terphenyl-D14(S)	11.74	244	7155193	2892.78	ppb		0.00
Spiked Amount 3333.333			Recovery =	86.783%			
Target Compounds							Qvalue
21) Diethyl phthalate	8.74	149	3647	19.28	ppb		95
26) Di-n-butylphthalate	10.54	149	10722	2.43	ppb		97
30) Bis (2-ethylhexyl) phthala	12.82	149	13551	5.51	ppb	#	96
31) Di-n-octylphthalate	13.50	149	3285	0.80	ppb		93

(#) = qualifier out of range (m) = manual integration
0520Y055.D Y0520TT.M Tue Jun 02 10:37:01 2015

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y055.D
Acq On : 28 May 15 14:49
Sample : 150521A BLK 1/30.00G
Misc : soil

Vial: 55
Operator: csc
Inst : Yoda
Multiplr: 33.33

Quant Time: Jun 2 10:17 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration





#21
Diethyl phthalate
Concen: 19.28 ppb
RT: 8.74 min Scan# 798
Delta R.T. 0.01 min
Lab File: 0520Y055.D
Acq: 28 May 15 14:49

Tgt Ion:149 Resp: 3647

Ion	Ratio	Lower	Upper
149	100		
177	23.2	15.1	28.0
150	14.9	8.3	15.5





#26
Di-n-butylphthalate
Concen: 2.43 ppb
RT: 10.54 min Scan# 992
Delta R.T. 0.00 min
Lab File: 0520Y055.D
Acq: 28 May 15 14:49

Tgt Ion:149 Resp: 10722

Ion	Ratio	Lower	Upper
149	100		
150	8.4	6.4	11.8
104	7.2	4.3	7.9





#30
Bis (2-ethylhexyl) phthalate
Concen: 5.51 ppb
RT: 12.82 min Scan# 1237
Delta R.T. 0.00 min
Lab File: 0520Y055.D
Acq: 28 May 15 14:49

Tgt Ion:149 Resp: 13551

Ion	Ratio	Lower	Upper
149	100		
167	34.1	23.1	42.9
279	6.6	7.1	13.3#





#31
Di-n-octylphthalate
Concen: 0.80 ppb
RT: 13.50 min Scan# 1311
Delta R.T. 0.01 min
Lab File: 0520Y055.D
Acq: 28 May 15 14:49

Tgt Ion:149 Resp: 3285

Ion	Ratio	Lower	Upper
149	100		
150	6.7	6.4	12.0



Laboratory Control Spike Recovery
EPA 8270D SOILS

APPL ID: **150521S-16137 LCS - 197602**
Batch ID: #87DJU-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
BIS (2-ETHYLHEXYL) PHTHALATE	1.67	1.35	80.8	45-125
BUTYL BENZYL PHTHALATE	1.67	1.31	78.4	50-125
DI-N-BUTYL PHTHALATE	1.67	1.30	77.8	55-110
DI-N-OCTYL PHTHALATE	1.67	1.39	83.2	40-130
DIETHYL PHTHALATE	1.67	1.19	71.3	50-115
DIMETHYL PHTHALATE	1.67	1.19	71.3	50-110
SURROGATE: 2,4,6-TRIBROMOPHENOL	6.67	5.07	76.0	35-125
SURROGATE: 2-FLUORBIPHENYL (S)	3.33	2.27	68.2	45-105
SURROGATE: 2-FLUOROPHENOL (S)	6.67	4.46	66.9	35-105
SURROGATE: NITROBENZENE-D5 (S)	3.33	2.22	66.7	35-100
SURROGATE: PHENOL (S)	6.67	4.66	69.9	40-100
SURROGATE: TERPHENYL-D14 (S)	3.33	2.98	89.5	30-125

Comments: _____

Primary	SPK
Quant Method :	Y0520TT.M
Extraction Date :	5/21/2015
Analysis Date :	5/28/2015
Instrument :	Yoda
Run :	0520Y056
Initials :	DA

Printed: 6/2/2015 12:11:57 PM
APPL Standard LCS

Quantitation Report (Not Reviewed)

Data File : M:\YODA\DATA\Y150520\0520Y056.D
Acq On : 28 May 15 15:18
Sample : 150521A LCS-1 1/30.07G
Misc : soil

Vial: 56
Operator: csc
Inst : Yoda
Multiplr: 33.26

Quant Time: Jun 1 11:22 2015 Quant Results File: Y0520TT.RES

Quant Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration
DataAcq Meth : SVOC416

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) 1,4-dichlorobenzene-D4(IS)	4.80	152	983964	40.00	ppb		0.00
7) Napthalene-D8(IS)	6.23	136	4419817	40.00	ppb		0.00
13) Acenaphthene-D10(IS)	8.22	164	2490321	40.00	ppb		0.00
23) Phenanthrene-D10(IS)	9.95	188	4281753	40.00	ppb		0.00
27) Chrysene-D12(IS)	13.01	240	3572459	40.00	ppb		0.00
32) Perylene-D12(IS)	14.64	264	2131735	40.00	ppb		0.00
System Monitoring Compounds							
2) 2-Fluorophenol (S)	3.97	112	4854290	4461.29	ppb		-0.07
Spiked Amount 6651.147			Recovery =	67.076%			
3) Phenol-D6 (S)	5.06	99	6850450	4664.83	ppb		-0.05
Spiked Amount 6651.147			Recovery =	70.136%			
8) Nitrobenzene-D5(S)	5.50	82	3025830	2216.70	ppb		0.00
Spiked Amount 3325.574			Recovery =	66.656%			
16) 2-Fluorobiphenyl(S)	7.45	172	6501011	2267.92	ppb		0.00
Spiked Amount 3325.574			Recovery =	68.196%			
22) 2,4,6-Tribromophenol(S)	9.22	330	1278414	5073.03	ppb		0.00
Spiked Amount 6651.147			Recovery =	76.273%			
28) Terphenyl-D14(S)	11.75	244	7946573	2983.88	ppb		0.00
Spiked Amount 3325.574			Recovery =	89.725%			
Target Compounds							Qvalue
4) Phenol	5.07	94	1862536	1138.86	ppb		88
5) 1,4-DCB	4.81	146	1584760	1117.80	ppb		99
6) n-Nitrosodi-n-propylamine	5.29	43	973854	1269.25	ppb		97
9) 2-Nitrophenol	5.87	139	847441	1201.47	ppb		87
10) 2,4-Dichlorophenol	6.48	162	2434236	1171.73	ppb		99
11) Hexachlorobutadiene	6.35	225	658624	1176.18	ppb		98
12) 4-Chloro-3-methylphenol	7.20	107	90997	1140.46	ppb	#	14
14) Hexachlorocyclopentadiene	7.17	237	248717	1076.07	ppb		99
15) 2,4,6-Trichlorophenol	7.55	196	818478	1219.71	ppb		98
17) Dimethyl phthalate	7.97	163	3500272	1192.29	ppb		90
18) Acenaphthene	8.26	154	2931202	1186.55	ppb		97
19) 2,4-Dinitrophenol	8.51	184	258619	1123.90	ppb	#	100
20) 4-Nitrophenol	9.17	65	853905	1162.09	ppb		99
21) Diethyl phthalate	8.73	149	3487656	1188.43	ppb		100
24) n-Nitrosodiphenylamine	10.35	167	4799487	1199.35	ppb		98
25) Pentachlorophenol	9.93	266	184061	1115.79	ppb		90
26) Di-n-butylphthalate	10.54	149	5972882	1303.94	ppb	#	97
29) Butyl benzylphthalate	12.27	149	2488530	1307.83	ppb		93
30) Bis (2-ethylhexyl) phthala	12.81	149	3586234	1353.98	ppb	#	93
31) Di-n-octylphthalate	13.49	149	6108460	1389.95	ppb		100

$\frac{1662536 \times 40}{983964 \times 2.211} \times 33.26 = 1138.98$ TA 6/5/15

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\YODA\DATA\Y150520\0520Y056.D Vial: 56
Acq On : 28 May 15 15:18 Operator: csc
Sample : 150521A LCS-1 1/30.07G Inst : Yoda
Misc : soil Multiplr: 33.26

Quant Time: Jun 1 11:22 2015 Quant Results File: Y0520TT.RES

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C
Last Update : Fri May 29 12:06:14 2015
Response via : Initial Calibration



DFTPP

Data File : M:\YODA\DATA\Y150520\0520Y005.D Vial: 5
Acq On : 22 May 15 10:28 Operator: csc
Sample : SV Tune 5/14/15 Inst : Yoda
Misc : Multiplr: 1.00

Method : M:\YODA\DATA\Y150520\Y0520TT.M (RTE Integrator)
Title : EPA 8270C



AutoFind: Scans 282, 283, 284; Background Corrected with Scan 279

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	31.1	115997	PASS
68	69	0.00	2	0.0	0	PASS
70	69	0.00	2	0.1	127	PASS
127	198	40	60	51.6	192476	PASS
197	198	0.00	1	0.1	253	PASS
198	198	100	100	100.0	373053	PASS
199	198	5	9	6.8	25312	PASS
275	198	10	30	27.8	103725	PASS
365	198	1	100	3.1	11748	PASS
441	443	0.01	100	76.5	56619	PASS
442	198	40	150	105.2	392399	PASS
443	442	17	23	18.9	74045	PASS

0520Y005.D Y0520TT.M Tue Jun 02 10:05:34 2015

DFTPP

Data File : M:\YODA\DATA\Y150520\0520Y047.D Vial: 47
Acq On : 28 May 15 10:36 Operator: csc
Sample : SV TUNE 5/14/15 Inst : Yoda
Misc : Multiplr: 1.00

Method : M:\YODA\DATA\Y150520\Y0520.M (RTE Integrator)
Title : EPA 8270C



Spectrum Information: Average of 6.201 to 6.220 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	30.0	133405	PASS
68	69	0.00	2	0.1	168	PASS
70	69	0.00	2	0.4	723	PASS
127	198	40	60	48.9	217427	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	444517	PASS
199	198	5	9	6.6	29426	PASS
275	198	10	30	28.7	127578	PASS
365	198	1	100	3.0	13174	PASS
441	443	0.01	100	75.6	57498	PASS
442	198	40	150	88.4	393040	PASS
443	442	17	23	19.4	76072	PASS

INDOXACARB CURVE											
Prep: 5/14/15 Ex: 7/31/15-L.H.											
Supplier	**ID#**	**ug/mL**	**LOT#**	**DATE**	**EXP.DATE**	**µL**	**µL**	**µL**	**µL**	**µL**	**µL**
INDOXACARB STD.		1		07/31/14	07/31/15	5	50	100	150	200	250
						995	950	900	850	800	750
	Hexane		051914A		**FINAL VOL**	1000	1000	1000	1000	1000	1000

8270 Full Scan Second Source Stock						
	5/14/15-L.H.					
8270 Second Source Stock						
Exp:	7/21/15					
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	116070-02	2000	34696-3398	5/14/15	5/14/16	1000
o2si	110391-01	2000	34693-3401	5/14/15	5/14/16	1000
o2si	010337-01	1000	17950-3514	5/14/15	11/14/15	1000
o2si	110396-01	2000	34698-3515	5/14/15	5/14/16	1000
o2si	110397-01	2000	50011-3505	5/14/15	7/21/15	1000
o2si	110393-01	2000	34694-3515	5/14/15	5/14/16	1000
o2si	110394-01	2000	35110-3515	5/14/15	5/14/16	1000
o2si	110395-01	2000	34697-3402	5/14/15	5/14/16	1000
o2si	110392-01	2000	18075-3397	5/14/15	5/14/16	1000
EM Science	Methylene Chloride		55006	11/06/14		1000
				Final Vol.		10000

Semivolatile (SV) Tuning Solution -L.H.						
PREP DATE:	5/14/15					
Semivolatile (SV) Tune Solution 50ug/ml						
Exp:	5/14/16					
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Ultra Scient	Semivolatile	1000	CL-5063-3518	05/14/15	5/14/16	1000
EM Science	Methylene Chloride		47186	11/06/14		19000
				Final Vol.		20000

PCB CONGENERS SPIKE
Source: O2Si **Con.** 100mg/L
CAT: 130116-01-SS
LOT: 250062-35148
Open: 05/22/15-L.H.
Ex: 5/22/16

PCB CONGENERS SURR.					
Prep: 5/22/15-L.H. Ex: 11/22/15					
Compounds		**Int. Conc.**	**ALIQUOT**	**Fin. Vol.**	**Fin. Conc.**
TCMX		1000µg/mL	500µL	50mL	10µg/mL
Part: 70273		OP: 4/21/15			
Lot: 041111-30678		EX: 4/11/16			
Source: Absolute					
Solvent:					
Acetone:		051413C			

8270 Full Scan Standard Curve

PREP DATE:	05/22/15-L.H.												
8270 Standard Curve													
Exp:	05/29/15					5	10	20	40	50	60	80	100
		Conc.		Date									
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL	µL	µL	µL	µL	µL	µL	µL
Absolute	8270 Stock	200	VAR	04/27/15	10/27/15	5	5	10	20	25	30	40	50
o2si	8270 BN:A (200	VAR	242352-34963	05/21/15	5/21/16	5	5	10	20	25	30	40	50
EM Science	Methylene Chloride		54200	11/06/14		190	90	80	60	50	40	20	0
Supelco	SV Internal St	2000	XA10079V-34496	05/21/15	5/21/16	4	2	2	2	2	2	2	2
				Final Vol.		204	102	102	102	102	102	102	102

8270 PAH SIM CCV

PREP DATE:	05/22/15-LH					
8270 PAH SIM Continuing Calibration Verification 5ug/ml						
Exp.DATE:	05/29/15					5
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	8270 PAH SIM S	200	230271-34569	02/27/15	06/08/16	5
o2si	8270 BN:A (200	VAR	242352-34963	02/27/15	#######	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SIM SV Intern	125	XA10079V-34496	05/15/15	5/15/16	4
				Final Vol.		204

Tech. Chlordane Spike
Source: Absolute **Con.** 1000mg/L
CAT: 130116-01-SS
LOT: 101311-32223
Open: 05/22/15-L.H.
Ex: 5/22/16

GC/MS STANDARD PREPARATION BOOK # L PAGE #

104

04/27/15
ac

PREP DATE:	04/27/15					ac
8270 Stock/Spike Standard						
Exp:	10/27/15					
		Conc.				
Supplier	ID #	µg/mL	Lot #	Open Date	Exp.Date	µL
Absolute	10001	2000	053014-34968	04/01/15	05/03/17	1000
Absolute	10001	2000	053014-34967	04/01/15	05/03/17	1000
Absolute	10001	2000	053014-34966	04/01/15	05/03/17	1000
Absolute	10002	2000	111114-34973	04/01/15	11/11/17	1000
Absolute	10002	2000	111114-34972	04/01/15	11/11/17	1000
Absolute	10002	2000	111114-34971	04/01/15	11/11/17	1000
Absolute	10004	2000	010815-34978	04/01/15	01/08/20	1000
Absolute	10004	2000	010815-34977	04/01/15	01/08/20	1000
Absolute	10004	2000	010815-34976	04/01/15	01/08/20	1000
Absolute	10005	2000	110314-34983	04/01/15	11/03/19	1000
Absolute	10005	2000	110314-34982	04/01/15	11/03/19	1000
Absolute	10005	2000	110314-34981	04/01/15	11/03/19	1000
Absolute	10006	2000	031715-34988	04/01/15	03/17/18	1000
Absolute	10006	2000	031715-34987	04/01/15	03/17/18	1000
Absolute	10006	2000	031715-34986	04/01/15	03/17/18	1000
Absolute	10007	2000	020515-34993	04/01/15	02/05/20	1000
Absolute	10007	2000	020515-34992	04/01/15	02/05/20	1000
Absolute	10007	2000	020515-34991	04/01/15	02/05/20	1000
Absolute	10018	2000	020315-34998	04/01/15	02/03/20	1000
Absolute	10018	2000	020315-34997	04/01/15	02/03/20	1000
Absolute	10018	2000	020315-34996	04/01/15	02/03/20	1000
Absolute	70023	1000	112114-35003	04/01/15	11/21/19	1000
Absolute	70023	1000	112114-35002	04/01/15	11/21/19	1000
Absolute	70023	1000	112114-35001	04/01/15	11/21/19	1000
Absolute	82705	2000	031215-35008	04/01/15	03/12/18	1000
Absolute	82705	2000	031215-35007	04/01/15	03/12/18	1000
Absolute	82705	2000	031215-35006	04/01/15	03/12/18	1000
Absolute	94552	2000	082914-35013	04/01/15	08/29/19	1000
Absolute	94552	2000	082914-35012	04/01/15	08/29/19	1000
Absolute	94552	2000	082914-35011	04/01/15	08/29/19	1000
				Final Vol.		30000

04/21/15
csc

PREP DATE:	04/21/15					ac
8270 Continuing Calibration Verification 50ug/ml						50
Exp:	04/28/15					
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Absolute	8270 Stock	200	VAR	11/18/14	05/21/15	25
o2si	8270 BN:A Surrogate	VAR	237956-34559	02/27/15	10/08/16	25
EM Science	Methylene Chloride		54200	11/06/14		50
Supelco	SV Internal Standard	2000	XA10079V-34496	02/27/15	09/30/17	2
				Final Vol.		102

04/21/15
csc

PREP DATE:	04/21/15					ac
8270 Continuing Calibration Verification 50ug/ml						50
Exp:	04/28/15					
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Absolute	8270 Stock	200	VAR	11/18/14	05/21/15	25
o2si	8270 BN:A Surrogate	VAR	237956-34559	02/27/15	10/08/16	25
EM Science	Methylene Chloride		54200	11/06/14		50
Supelco	SV Internal Standard	2000	XA10079V-34496	02/27/15	09/30/17	2
				Final Vol.		102

SV Internal Stanard Stock Prep: 05/21/15-L.H. Ex: 5/21/16						
		Conc.				
Supplier	ID #	µg/mL	Lot #	Open Date	Exp.Date	µL
Supleco	CRM48902	2000	XA10079V-34489	5/21/15	5/21/16	1000

8270 BNA Surrogate Stock 5/21/15-L.H. Ex: 5/21/16						
		Conc.				
Supplier	ID #	µg/mL	Lot #	Open Date	Exp.Date	µL
o2SI	110004-17	200/400	242352-34956	5/21/15	5/21/16	1000

PREP DATE	05/21/15-L.H.					
8270 Lv. 1						
Exp:	05/28/15					5
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Absolute	8270 Stock	200	VAR	04/27/15	10/27/15	5
o2si	8270 BN:A (200:400)	VAR	242352-34956	05/21/15	05/21/16	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SV Internal S	2000	10079V-344	05/21/15	5/21/16	4
				Final Vol.		204

		AR 1254 CALIBRATION CURVE					
AR1254 STANDARD		Prep:4/16/15-L.H.					
		Exp: 10/1/15					
		final				Final Vol.	
	Initial	conc.		Aliquot		Solvent	Solvent
LEVELS ID	conc.	(ug/ml)		(uL)			(ml)
LEVEL 10	1ug/ml	0.010		10 µL			1.0
LEVEL 50		0.050		50 µL		HEXANE	1.0
LEVEL 100		0.100		100 µL			1.0
LEVEL 250		0.250		250 µL		LOT #05191	1.0
LEVEL 1000		1.000		100 µL			1.0

MITC CURVE Prep: 5/21/15-L.H. Ex: 6/11/15												
						1	2	3	4	5	6	7
SUPPLIER	ID#	ug/mL	LOT #	DATE	EXP.	mL	mL	mL	mL	mL	mL	mL
					Final Conc.	0.02	0.13	0.25	1	2.5	3.5	5
	MITC STD	5		05/05/15	06/11/15	4	25	50	20	50	70	100
VWR	ETHYL ACETATE		043014E			996	975	950	80	50	30	N/A
					Final VOL.	1000	1000	1000	100	100	100	100

Prep: 5/21/15 Exp: 8/21/15 GA Diesel 400ppm CCV							
STANDARD	INITIAL CONC.	SOURCE D	ALIQUOT	FINAL VOL	FINAL CON	SOLVENT	
DIESEL STD.	1000 µg/mL	Diesel Std.	2000µL	5mL	400 µg/ml	MC	
STANDARD	Prep:	12/05/14				55006	
DIESEL STD.	Exp:	11/24/15					

Prep: 5/22/15	Exp: 6/26/15	OCL Level 2 Standard			GA
Compound	Initial conc	ource Stoc	Aliquot	Final Conc	Final Vol.
Various	10 ug/ml	OCL Stock	125ul	0.050 ug/ml	25 mL
		Prep: 11/17/14			
		Exp: 6/26/15			
	Hexane:	051914A			

OP/F CURVE		Prep: 5/20/15 Exp: 10/24/15			
SUPPLIER	ID#	[µg/mL]	LOT #	DATE	EXP. DATE
	OP/F STD	5		04/24/15	######
VWR	Hexane		051914A		
					Final VOL

PCB Water Spike													
Prep: 5/22/15-L.H. Ex: 8/22/15						Acetone:051413C							
Compounds		Int. Conc.	ALIQUOT	Fin. Vol.	Fin. Conc.								
AR 1016		1000µg/mL	50µL	10mL	5µg/mL								
AR1260						IA	IB	1	2	3	4	5	6
Cat: 130011-03		OP: 5/22/15				mL	mL	mL	mL	mL	mL	mL	mL
Lot:235487-34300		EX: 5/22/16				2	5	10	50	20	100	70	100
Source: O2Si						998	995	990	950	80	100	30	NA
Solvent:						1000	1000	1000	1000	100	200	100	100

8270 Full Scan Second Source (SS)						
PREP DATE:	/022/15-L.	EX. 5/29/15				
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
o2si	8270 Stock S	200	VAR	05/14/15	7/21/15	25
EM Science	Methylene Chloride		54200	11/06/14		75
Supelco	SV Internal	2000	XA10079V-34496	02/27/15	09/30/17	2
				Final Vol.		102

072

02/24/15
csc

GC/MS STANDARD PREPARATION BOOK # L PAGE

PREP DATE:	02/24/15					
8270 PAH SIM Continuing Calibration Verification 5ug/ml						
Exp.DATE:	02/31/15					5
Supplier	ID #	Conc. µg/mL	Lot #	Date Code	Exp. Date	µL
o2si	8270 PAH SIM Stock	10	213931-33759	12/23/14	08/26/15	5
o2si	8270 BN:A Surrogate	VAR	226244-34103	12/23/14	04/10/16	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SIM Internal Standard	125	XA10079V-34507	12/23/14	09/30/17	4
				Final Vol.		204

02/27/15
csc

8270D PAH SIM Solution, 200 mg/L, 1 ml
110730-01
Lot # 230271 Storage ≤-10 °C Expiry 8/Jun/2016
Solv: Methylene Chloride

Recieved: 01/13/15
Opened: 02/27/15
Expired: 06/8/16

8270D PAH SIM
Lot #: 230271 - 34569
Rec: 1/13/15 MFR exp. 6/8/16

02/27/15
csc

8270 BN:A (200:400) Surrogate Solution, 1 mL
110804-17
Lot # 237956 Storage ≤-10 °C Expiry 8/Oct/2016
Solv: Methylene Chloride

Recieved: 01/13/15
Opened: 02/27/15
Expired: 10/8/16

8270 BN:A (200:400) Surrogate Solution
Lot #: 237956 - 34559
Rec: 1/13/15 MFR exp. 10/8/16

NOTEBOOK INSERT LABEL

Semivolatile Internal Std Mix **CRM48902**
Lot: XA10079V EXP: SEP/2017 STORAGE: REFRIGERATE 2 x 1ml

DATE RECEIVED:_____

SUPELCO
Solutions within.
595 North Harrison Road • Bellefonte, PA
16823-0048 USA • Phone 814-359-3441

STANDARD TRANSFER LABEL

Date of Preparation: **Exp. Date:**
Reference Number: **Storage:** EXP: SEP/2017
Description: REFRIGERATE
CHEMST

02/28/15
csc

Semivolatile Internal Standard Mix
Lot #: XA10079V - 34496
Rec: 12/11/14 MFR exp. 9/30/17

CRM48902 XA10079V
Semivolatile Internal Std Mix
2000ug/ml in dichloromethane

Recieved: 12/11/14
Opened: 02/27/15
Expired: 09/30/17

_SON009

Organic Extraction Worksheet

Method	8270 Sonication Ext. Methylene c 3550	Extraction Set	150521A	Extraction Method	SON009	Units	mL

Spiked ID 1	8270T Spike 4-27-15 EXP 10-27-15		Surrogate ID 1	8270 Surrogate 248398-35028	
Spiked ID 2			Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	NO	
Spiked ID 7			Ext. Start Time:	05/21/15 8:15	
Spiked ID 8			Ext. End Time:	05/21/15 14:45	
			GC Requires Extract By:	**05/29/15 0:00**	

pH1			Water Bath Temp Criteria	79,80 °C
pH2				
pH3				

Spiked By: IC **Date** 05/21/15 **Witnessed By:** DL **Date** 05/21/15

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	150521A Blk				1	1 equip	30.00g E-S5 E-WB7	1	NA	05/21/15 8:15	
2	150521A LCS-1		0.250	1	1	1 equip	30.07g E-S4 E-WB7	1	NA	05/21/15 8:15	
3	AZ16137	AZ16137S01			1	1 equip	30.15g E-S3.1 E-WB7	1	NA	05/21/15 8:15	76380
4	AZ16138	AZ16138S01			1	1 equip	30.26g E-S2 E-WB7	1	NA	05/21/15 8:15	76380
5	AZ16139	AZ16139S01			1	1 equip	30.17g E-S1.2 E-WB7	1	NA	05/21/15 8:15	76380
6	AZ16140	AZ16140S01			1	1 equip	30.04g E-S1.1 E-WB7	1	NA	05/21/15 8:15	76380
7	AZ16141	AZ16141S01			1	1 equip	30.33g E-S6 E-WB7	1	NA	05/21/15 8:15	76380
8	AZ16198	AZ16198S02			1	1 equip	30.11g E-S7 E-WB7	1	NA	05/21/15 8:15	76390
9	AZ16199	AZ16199S01			1	1 equip	30.01g E-S5 E-WB6	1	NA	05/21/15 8:15	76390
10	AZ16201	AZ16201S02			1	1 equip	30.10g E-S4 E-WB6	1	NA	05/21/15 8:15	76390
11	AZ16202	AZ16202S01			1	1 equip	30.18g E-S3.1 E-WB6	1	NA	05/21/15 8:15	76390
12	AZ16286	AZ16286S02			1	1 equip	30.10g E-S2 E-WB6	1	NA	05/21/15 8:15	76414
13	AZ16400	AZ16400S02			1	1 equip	30.10g E-S1.2 E-WB6	1	NA	05/21/15 8:15	76447
14	AZ16401	AZ16401S01			1	1 equip	30.34g E-S1.1 E-WB6	1	NA	05/21/15 8:15	76447
15	AZ16402	AZ16402S01			1	1 equip	30.17g E-S6 E-WB6	1	NA	05/21/15 8:15	76447

Solvent and Lot#	
B.Na2SO4	XE21E
MC	55006
Filter Paper	9624720
Acidified Na2SO4	4-13-15

KY 5/21/15

Extraction COC Transfer	
Extraction lab employee Initials	KY
GC analyst's initials	GA
Date	5/21/15
Time	1600
Refrigerator	Hobart

	Technician's Initials
Scanned By	KY
Sample Preparation	KY
Extraction	IC
Concentration	IC

Modified	05/21/15 3:46:20 PM

Reviewed By: KY **Date** 5/21/15

Injection Log

Directory: M:\YODA\DATA\Y150520

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	5	0520Y005.D	1	SV Tune 5/14/15		05/22/2015 10:28
2	6	0520Y006.D	1	5ug/ml SVOC 05/22/15		05/22/2015 10:46
3	7	0520Y007.D	1	10ug/ml SVOC 05/22/15		05/22/2015 11:15
4	8	0520Y008.D	1	20ug/ml SVOC 05/22/15		05/22/2015 11:45
5	9	0520Y009.D	1	40ug/ml SVOC 05/22/15		05/22/2015 12:14
6	10	0520Y010.D	1	50ug/ml SVOC 05/22/15		05/22/2015 12:43
7	11	0520Y011.D	1	60ug/ml SVOC 05/22/15		05/22/2015 13:12
8	12	0520Y012.D	1	80ug/ml SVOC 05/22/15		05/22/2015 13:42
9	13	0520Y013.D	1	100ug/ml SVOC 05/22/15		05/22/2015 14:11
10	14	0520Y014.D	1	50ug/ml SVOC SS 05/22/15		05/22/2015 14:41
11	47	0520Y047.D	1	SV TUNE 5/14/15		05/28/2015 10:36
12	48	0520Y048.D	1	50ug/ml SVOC 05/28/15		05/28/2015 10:54
13	55	0520Y055.D	33.33333	150521A BLK 1/30.00G	soil	05/28/2015 14:49
14	56	0520Y056.D	33.25574	150521A LCS-1 1/30.07G	soil	05/28/2015 15:18
15	57	0520Y057.D	331.675	AZ16137S01 1/30.15G x10	soil	05/28/2015 15:47
16	58	0520Y058.D	33.04693	AZ16138S01 1/30.26G x10	soil	05/28/2015 16:17
17	59	0520Y059.D	33.14551	AZ16139S01 1/30.17G	soil	05/28/2015 16:46
18	60	0520Y060.D	33.28895	AZ16140S01 1/30.04G x10	soil	05/28/2015 17:16
19	61	0520Y061.D	32.97066	AZ16141S01 1/30.33G	soil	05/28/2015 17:46

EPA METHOD 8270D
Polynuclear Aromatic Hydrocarbons

APPL, INC.

EPA METHOD 8270D
Polynuclear Aromatic Hydrocarbons
QC Summary

APPL, INC.

Method Blank
EPA 8270D LL SOILS

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150521S-16137 - 197621**
Batch ID: #SIMDD-150521A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	63.5	45-105			%	5/21/2015	5/27/2015
BLANK	SURROGATE: NITROBENZENE-	44.2	35-100			%	5/21/2015	5/27/2015
BLANK	SURROGATE: TERPHENYL-D14 (	106	30-125			%	5/21/2015	5/27/2015

Quant Method: P0518-42.M
Run #: 0515L235
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 2:14:35 PM

Method Blank

EPA 8270D LL SOILS RE-EXTRACT

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150601S-16139 - 197777**
Batch ID: #SIMDD-150601A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	77.1	45-105			%	6/1/2015	6/1/2015
BLANK	SURROGATE: NITROBENZENE-	68.4	35-100			%	6/1/2015	6/1/2015
BLANK	SURROGATE: TERPHENYL-D14	112	30-125			%	6/1/2015	6/1/2015

Quant Method: P0518-42.M
Run #: 0515L262
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/5/2015 2:20:31 PM

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: 2-FLUORBIPHENYL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	45-105	63.5		35-100	44.2	
150521A-LCS	Lab Control Spike	45-105	64.5		35-100	46.9	
AZ16137	S67-SS030-0001	45-105	68.8		35-100	117	#
AZ16138	S67-SS030P-0001	45-105	68.1		35-100	116	#
AZ16139	S67-SB030-0607	45-105	28.7	#	35-100	48.4	
AZ16140	S67-SS032-0001	45-105	53.8		35-100	112	#
AZ16141	S67-SB032-0607	45-105	45.1		35-100	52.1	

Comments: Batch: #SIMDD-150521A

= Recovery outside of Control Limits on Sample.

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: TERPHENYL-D14 (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150521A-BLK	Blank	30-125	106				
150521A-LCS	Lab Control Spike	30-125	107				
AZ16137	S67-SS030-0001	30-125	87.0				
AZ16138	S67-SS030P-0001	30-125	83.7				
AZ16139	S67-SB030-0607	30-125	83.6				
AZ16140	S67-SS032-0001	30-125	72.0				
AZ16141	S67-SB032-0607	30-125	86.8				

Comments: Batch: #SIMDD-150521A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: 2-FLUORBIPHENYL (S)			SURROGATE: NITROBENZENE-D5 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
150601A-BLK	Blank	45-105	77.1		35-100	68.4	
150601A-LCS	Lab Control Spike	45-105	80.3		35-100	69.7	
AZ16139	S67-SB030-0607	45-105	57.7		35-100	57.6	

Comments: Batch: #SIMDD-150601A

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: TERPHENYL-D14 (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150601A-BLK	Blank	30-125	112				
150601A-LCS	Lab Control Spike	30-125	119				
AZ16139	S67-SB030-0607	30-125	90.0				

Comments: Batch: #SIMDD-150601A

Laboratory Control Spike Recovery
EPA 8270D LL SOILS

APPL ID: **150521S-16137 LCS - 197621**
Batch ID: #SIMDD-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.133	0.0883	66.4	45-105
ACENAPHTHENE	0.133	0.127	95.5	45-110
ACENAPHTHYLENE	0.133	0.112	84.2	45-105
ANTHRACENE	0.133	0.126	94.7	55-105
BENZ (A) ANTHRACENE	0.133	0.139	105	50-110
BENZO (A) PYRENE	0.133	0.132	99.2	50-110
BENZO (B) FLUORANTHENE	0.133	0.140	105	45-115
BENZO (G,H,I) PERYLENE	0.133	0.135	102	40-125
BENZO (K) FLUORANTHENE	0.133	0.132	99.2	45-125
CHRYSENE	0.133	0.144	108	55-110
DIBENZ (A,H) ANTHRACENE	0.133	0.136	102	40-125
FLUORANTHENE	0.133	0.127	95.5	55-115
FLUORENE	0.133	0.120	90.2	50-110
INDENO (1,2,3-CD) PYRENE	0.133	0.145	109	40-120
NAPHTHALENE	0.133	0.0797	59.9	40-105
PHENANTHRENE	0.133	0.120	90.2	50-110
PYRENE	0.133	0.146	110	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0667	0.043	64.5	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0667	0.0313	46.9	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0667	0.0711	107	30-125

Comments: _____

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	5/21/2015
Analysis Date :	5/27/2015
Instrument :	Linus
Run :	0515L236
Initials :	DA

Printed: 6/2/2015 2:14:22 PM
APPL Standard LCS

Laboratory Control Spike Recovery
EPA 8270D LL SOILS RE-EXTRACT

APPL ID: **150601S-16139 LCS - 197777**
Batch ID: #SIMDD-150601A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.167	0.156	93.4	45-105
ACENAPHTHENE	0.167	0.178	107	45-110
ACENAPHTHYLENE	0.167	0.164	98.2	45-105
ANTHRACENE	0.167	0.170	102	55-105
BENZ (A) ANTHRACENE	0.167	0.192	115 #	50-110
BENZO (A) PYRENE	0.167	0.199	119 #	50-110
BENZO (B) FLUORANTHENE	0.167	0.205	123 #	45-115
BENZO (G,H,I) PERYLENE	0.167	0.194	116	40-125
BENZO (K) FLUORANTHENE	0.167	0.198	119	45-125
CHRYSENE	0.167	0.199	119 #	55-110
DIBENZ (A,H) ANTHRACENE	0.167	0.195	117	40-125
FLUORANTHENE	0.167	0.174	104	55-115
FLUORENE	0.167	0.165	98.8	50-110
INDENO (1,2,3-CD) PYRENE	0.167	0.189	113	40-120
NAPHTHALENE	0.167	0.179	107 #	40-105
PHENANTHRENE	0.167	0.165	98.8	50-110
PYRENE	0.167	0.207	124	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0828	0.0665	80.3	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0828	0.0577	69.7	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0828	0.0985	119	30-125

= Recovery is outside QC limits.

Comments:

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	6/1/2015
Analysis Date :	6/1/2015
Instrument :	Linus
Run :	0515L263
Initials :	DA

Printed: 6/5/2015 2:20:19 PM
APPL Standard LCS

8270D-LL

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 5/27/2015

Matrix: SOIL Instrument: Linus

Blank ID: 150521A-BLK Time Analyzed: 1917

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150521A-BLK	Blank	0515L235	5/27/2015 1917
150521A-LCS	Lab Control Spike	0515L236	5/27/2015 1942
AZ16137	S67-SS030-0001	0515L237	5/27/2015 2008
AZ16138	S67-SS030P-0001	0515L238	5/27/2015 2033
AZ16139	S67-SB030-0607	0515L239	5/27/2015 2059
AZ16140	S67-SS032-0001	0515L240	5/27/2015 2124
AZ16141	S67-SB032-0607	0515L241	5/27/2015 2150

Comments: Batch: #SIMDD-150521A

8270D-LL

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 6/1/2015

Matrix: SOIL Instrument: Linus

Blank ID: 150601A-BLK Time Analyzed: 1600

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150601A-BLK	Blank	0515L262	6/1/2015 1600
150601A-LCS	Lab Control Spike	0515L263	6/1/2015 1625
AZ16139	S67-SB030-0607	0515L264	6/1/2015 1651

Comments: Batch: #SIMDD-150601A

Form 5
Tune Summary

Lab Name: APPL Inc.
Case No: ~~0515L002.D~~ 76380
Matrix: Soil np 6-17-15
ID: SV Tune 05/14/15

SDG No: np 6-17-15 ~~Linus~~ 76380
Date Analyzed: 5/15/2015
Instrument: Linus
Time Analyzed: 16:56

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		0.1ug/ml PAH 05/15/1	0515L003.D	5/15/2015 17:15
2		0.2ug/ml PAH	0515L004.D	5/15/2015 17:41
3		0.5ug/ml PAH	0515L005.D	5/15/2015 18:06
4		1.0ug/ml PAH	0515L006.D	5/15/2015 18:32
5		5.0ug/ml PAH	0515L007.D	5/15/2015 18:58
6		10ug/ml PAH	0515L008.D	5/15/2015 19:23
7		50ug/ml PAH	0515L009.D	5/15/2015 19:49
8		100ug/ml PAH	0515L010.D	5/15/2015 20:14
9		5.0ug/ml PAH SS 05/1	0515L011.D	5/15/2015 20:40
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e		
51	29.95 - 60% of mass 198	36.1
68	0 - 2.05% of mass 69	1.4
70	0 - 2% of mass 69	0.5
127	39.5 - 60% of mass 198	46.9
197	0 - 1% of mass 198	0.0
198	100 - 100% of mass 198	100.0
199	5 - 9% of mass 198	6.8
275	10 - 30% of mass 198	20.7
365	1 - 100% of mass 198	2.0
441	0.01 - 100% of mass 443	76.6
442	40 - 150% of mass 198	69.5
443	17 - 23% of mass 442	19.2

Form 5
Tune Summary

Lab Name: APPL Inc.
Case No: 76380
Matrix: Soil
ID: SV TUNE 5/14/15

SDG No: 76380
Date Analyzed: 5/27/2015
Instrument: Linus
Time Analyzed: 15:59

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		5.0ug/ml PAH 5/26/15	0515L228.D	5/27/2015 16:17
2	Blank	150521A BLK 1/30.00G	0515L235.D	5/27/2015 19:17
3	Lab Control Spike	150521A LCS-1 1/30.0	0515L236.D	5/27/2015 19:42
4	S67-SS030-0001	AZ16137S01 1/30.02G	0515L237.D	5/27/2015 20:08
5	S67-SS030P-0001	AZ16138S01 1/30.25G	0515L238.D	5/27/2015 20:33
6	S67-SB030-0607	AZ16139S01 1/30.07G	0515L239.D	5/27/2015 20:59
7	S67-SS032-0001	AZ16140S01 1/30.07G	0515L240.D	5/27/2015 21:24
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e		
51	29.95 - 60% of mass 198	35.0
68	0 - 2.05% of mass 69	1.5
70	0 - 2% of mass 69	0.6
127	39.5 - 60% of mass 198	45.4
197	0 - 1% of mass 198	0.0
198	100 - 100% of mass 198	100.0
199	5 - 9% of mass 198	6.6
275	10 - 30% of mass 198	22.2
365	1 - 100% of mass 198	2.1
441	0.01 - 100% of mass 443	77.8
442	40 - 150% of mass 198	78.3
443	17 - 23% of mass 442	19.3

Form 5
Tune Summary

Lab Name: APPL Inc.
Case No: 76380
Matrix: Soil
ID: SV TUNE 5/14/15

SDG No: 76380
Date Analyzed: 6/1/2015
Instrument: Linus
Time Analyzed: 9:58

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		5.0ug/ml PAH 06/01/1	0515L256.D	6/1/2015 12:25
2	Blank	150601A BLK 1/30.75G	0515L262.D	6/1/2015 16:00
3	Lab Control Spike	150601A LCS-1 1/30.3	0515L263.D	6/1/2015 16:25
4	S67-SB030-0607	AZ16139S02 1/30.21G	0515L264.D	6/1/2015 16:51
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e		
51	29.95 - 60% of mass 198	34.0
68	0 - 2.05% of mass 69	1.3
70	0 - 2% of mass 69	0.7
127	39.5 - 60% of mass 198	44.0
197	0 - 1% of mass 198	0.0
198	100 - 100% of mass 198	100.0
199	5 - 9% of mass 198	6.6
275	10 - 30% of mass 198	21.8
365	1 - 100% of mass 198	2.0
441	0.01 - 100% of mass 443	76.8
442	40 - 150% of mass 198	76.4
443	17 - 23% of mass 442	19.5

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review

Lab Code: _____ SDG No.: 76380

Lab File ID (Standard): 0515L009.D Date Analyzed: 05/15/15

Instrument ID: Linus Time Analyzed: 19:49

GC Column: _____ ID: Heated Purge: (Y/N) _____

		Napthalene-D8(IS)		Acenaphthene-D10(IS)		Phenanthrene-D10(IS)	
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	9907	6.16	4610	8.16	7949	9.88
	UPPER LIMIT	19814	6.66	9220	8.66	15898	10.38
	LOWER LIMIT	4954	5.66	2305	7.66	3975	9.38
	SAMPLE NO.						
01	5.0ug/ml PAH SS 05/15	9571	6.16	4405	8.16	7622	9.90
02	5.0ug/ml PAH 5/26/15	7511	6.16	3616	8.17	6351	9.90
03	150521A BLK 1/30.00G	8216	6.16	3798	8.17	6307	9.91
04	150521A LCS-1 1/30.01	9702	6.16	4792	8.17	8492	9.90
05	AZ16137S01 1/30.02G	8546	6.16	4029	8.17	7019	9.91
06	AZ16138S01 1/30.25G	9309	6.16	4752	8.17	8527	9.90
07	AZ16139S01 1/30.07G	9276	6.16	4244	8.17	7481	9.91
08	AZ16140S01 1/30.07G	10974	6.16	5406	8.17	9471	9.90
09	AZ16141S01 1/30.23G	9325	6.16	4181	8.17	7452	9.90
10	150601A BLK 1/30.75G	7710	6.16	3775	8.17	6391	9.90
11	150601A LCS-1 1/30.35	7373	6.16	3574	8.17	5999	9.90
12	AZ16139S02 1/30.21G	8425	6.16	4181	8.17	7089	9.91
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
AREA LOWER LIMIT = -50% of internal standard area.
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review

Lab Code: _____ SDG No.: 76380

Lab File ID (Standard): 0515L009.D Date Analyzed: 05/15/15

Instrument ID: Linus Time Analyzed: 19:49

GC Column: _____ ID: Heated Purge: (Y/N) _____

		Chrysene-D12(IS)		Perylene-D12(IS)			
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	8664	12.96	5437	14.57		
	UPPER LIMIT	17328	13.46	10874	15.07		
	LOWER LIMIT	4332	12.46	2719	14.07		
	SAMPLE NO.						
01	5.0ug/ml PAH SS 05/15	7750	12.96	4908	14.57		
02	5.0ug/ml PAH 5/26/15	6697	12.97	4324	14.59		
03	150521A BLK 1/30.00G	6696	12.97	4479	14.61		
04	150521A LCS-1 1/30.01	8827	12.97	5818	14.59		
05	AZ16137S01 1/30.02G	7629	12.97	5243	14.59		
06	AZ16138S01 1/30.25G	9328	12.97	6233	14.59		
07	AZ16139S01 1/30.07G	8062	12.97	5428	14.59		
08	AZ16140S01 1/30.07G	10019	12.97	6837	14.58		
09	AZ16141S01 1/30.23G	8223	12.97	5478	14.59		
10	150601A BLK 1/30.75G	6335	12.97	3876	14.59		
11	150601A LCS-1 1/30.35	6033	12.97	3629	14.59		
12	AZ16139S02 1/30.21G	7288	12.97	4691	14.59		
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
AREA LOWER LIMIT = -50% of internal standard area.
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

EPA METHOD 8270D
Polynuclear Aromatic Hydrocarbons
Sample Data

APPL, INC.

EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16138
QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.032 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0190 U	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.11	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.033 J	0.057	0.0190	0.0091	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.12	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.042 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.11	0.057	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.073	0.057	0.0190	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.30	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.090	0.057	0.0190	0.0091	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.057	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.14	0.057	0.0190	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.033 J	0.057	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.14	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.011 J	0.057	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.087	0.057	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.16	0.057	0.0190	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	68.1	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	116 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	83.7	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method:	P0518-42.M
Run #:	0515L238
Instrument:	Linus
Sequence:	L150515
Dilution Factor:	10
Initials:	DA

Printed: 6/2/2015 2:14:02 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L238.D Vial: 38
Acq On : 27 May 15 20:33 Operator: MA
Sample : AZ16138S01 1/30.25G df10 Inst : Linus
Misc : soil Multiplr: 330.58

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9309	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4752	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	8527	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	9328	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	6233	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	143	96.05	ppb		0.01
Spiked Amount 82.645			Recovery = 116.225%				
7) Surrogate Recovery (FBP)	7.41	172	511	56.26	ppb		0.01
Spiked Amount 82.645			Recovery = 68.081%				
17) Surrogate Recovery (TPH)	11.77	244	627	69.19	ppb		0.00
Spiked Amount 82.645			Recovery = 83.720%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	111	10.08	ppb		97
4) 2-Methylnaphthalene	6.98	142	43	28.43	ppb		86
5) 1-Methylnaphthalene	7.08	142	44	6.34	ppb		90
8) Acenaphthylene	8.00	152	156	96.75	ppb	#	82
9) Acenaphthene	8.21	154	41	5.63	ppb	#	86
10) Fluorene	8.82	166	64	28.70	ppb		98
12) Phenanthrene	9.93	178	1281	76.70	ppb		99
13) Anthracene	9.99	178	276	28.67	ppb		98
14) Fluoranthene	11.30	202	2594	119.87	ppb	#	89
16) Pyrene	11.56	202	2330	145.18	ppb	#	87
18) Benz (a) anthracene	12.96	228	1350	105.75	ppb		98
19) Chrysene	13.00	228	1070	79.54	ppb		98
20) Indeno (1,2,3-cd) pyrene	16.09	276	764	127.36	ppb	#	96
22) Benzo (b) fluoranthene	14.15	252	2214	97.48	ppb		98
23) Benzo (k) fluoranthene	14.15	252	2214	262.69	ppb	#	93
24) Benzo (a) pyrene	14.45	252	821	36.59	ppb		93
25) Dibenz (a,h) anthracene	16.11	278	214	50.52	ppb		95
26) Benzo (g,h,i) perylene	16.52	276	830	63.82	ppb		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L238.D Vial: 38
Acq On : 27 May 15 20:33 Operator: MA
Sample : AZ16138S01 1/30.25G df10 Inst : Linus
Misc : soil Multiplr: 330.58

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 10.08 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:128 Resp: 111

Ion	Ratio	Lower	Upper
128	100		
129	8.8	7.8	14.6
127	11.3	8.0	14.8





#4
2-Methylnaphthalene
Concen: 28.43 ppb
RT: 6.98 min Scan# 452
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:142 Resp: 43

Ion	Ratio	Lower	Upper
142	100		
141	73.1	60.1	111.5



EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.033 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0200 U	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.11	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.038 J	0.058	0.0200	0.0092	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.13	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.051 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.14	0.058	0.0200	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.076	0.058	0.0200	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.33	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.12	0.058	0.0200	0.0092	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.061	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.18	0.058	0.0200	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.034 J	0.058	0.0200	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.16	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.012 J	0.058	0.0200	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.11	0.058	0.0200	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.20	0.058	0.0200	0.0140	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	68.8	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	117 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	87.0	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method: P0518-42.M
Run #: 0515L237
Instrument: Linus
Sequence: L150515
Dilution Factor: 10
Initials: DA

Printed: 6/2/2015 2:14:02 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L237.D Vial: 37
Acq On : 27 May 15 20:08 Operator: MA
Sample : AZ16137S01 1/30.02G df10 Inst : Linus
Misc : soil Multiplr: 333.11

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	8546	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4029	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.91	188	7019	2.50	ppb		0.01
15) Chrysene-D12(IS)	12.97	240	7629	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	5243	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	134	97.61	ppb		0.01
Spiked Amount 83.278			Recovery =	117.208%			
7) Surrogate Recovery (FBP)	7.41	172	438	57.32	ppb		0.01
Spiked Amount 83.278			Recovery =	68.826%			
17) Surrogate Recovery (TPH)	11.77	244	533	72.47	ppb		0.00
Spiked Amount 83.278			Recovery =	87.018%			
Target Compounds							Qvalue
3) Naphthalene	6.18	128	108	10.76	ppb	#	86
4) 2-Methylnaphthalene	6.98	142	41	28.85	ppb		92
5) 1-Methylnaphthalene	7.08	142	37	5.85	ppb		96
8) Acenaphthylene	8.01	152	93	94.33	ppb		93
9) Acenaphthene	8.21	154	54	8.82	ppb		99
10) Fluorene	8.82	166	60	29.65	ppb		76
12) Phenanthrene	9.93	178	1284	97.75	ppb		98
13) Anthracene	9.99	178	274	33.26	ppb		94
14) Fluoranthene	11.30	202	2630	157.72	ppb		92
16) Pyrene	11.56	202	2238	171.80	ppb	#	88
18) Benz (a) anthracene	12.96	228	1222	116.42	ppb		96
19) Chrysene	13.00	228	1164	106.60	ppb		96
20) Indeno (1,2,3-cd) pyrene	16.09	276	750	138.13	ppb	#	96
22) Benzo (b) fluoranthene	14.15	252	2104	118.41	ppb		99
23) Benzo (k) fluoranthene	14.15	252	2104	285.36	ppb	#	95
24) Benzo (a) pyrene	14.45	252	772	43.83	ppb		98
25) Dibenz (a,h) anthracene	16.11	278	207	53.23	ppb	#	85
26) Benzo (g,h,i) perylene	16.52	276	713	65.59	ppb		92

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L237.D Vial: 37
Acq On : 27 May 15 20:08 Operator: MA
Sample : AZ16137S01 1/30.02G df10 Inst : Linus
Misc : soil Multiplr: 333.11

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 10.76 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:128 Resp: 108

Ion	Ratio	Lower	Upper
128	100		
129	16.0	7.8	14.6#
127	17.3	8.0	14.8#





#4
2-Methylnaphthalene
Concen: 28.85 ppb
RT: 6.98 min Scan# 452
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:142 Resp: 41

Ion	Ratio	Lower	Upper
142	100		
141	78.3	60.1	111.5





#5
1-Methylnaphthalene
Concen: 5.85 ppb
RT: 7.08 min Scan# 461
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:142 Resp: 37

Ion	Ratio	Lower	Upper
142	100		
141	86.4	63.1	117.3





#8
Acenaphthylene
Concen: 94.33 ppb
RT: 8.01 min Scan# 538
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:152 Resp: 93

Ion	Ratio	Lower	Upper
152	100		
151	19.1	12.1	22.5
153	17.0	9.2	17.2





#9
Acenaphthene
Concen: 8.82 ppb
RT: 8.21 min Scan# 554
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:154 Resp: 54

Ion	Ratio	Lower	Upper
154	100		
153	109.1	74.9	139.1
152	51.5	35.9	66.7





#10
Fluorene
Concen: 29.65 ppb
RT: 8.82 min Scan# 605
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:166 Resp: 60

Ion	Ratio	Lower	Upper
166	100		
165	109.1	60.7	112.7





#12
Phenanthrene
Concen: 97.75 ppb
RT: 9.93 min Scan# 697
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:178 Resp: 1284

Ion	Ratio	Lower	Upper
178	100		
179	16.2	10.6	19.6
176	17.0	12.2	22.6





#13
Anthracene
Concen: 33.26 ppb
RT: 9.99 min Scan# 702
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:178 Resp: 274

Ion	Ratio	Lower	Upper
178	100		
179	11.2	10.5	19.5
176	18.4	12.0	22.4





#14
Fluoranthene
Concen: 157.72 ppb
RT: 11.30 min Scan# 814
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:202 Resp: 2630

Ion	Ratio	Lower	Upper
202	100		
101	12.8	11.3	20.9
100	8.4	7.8	14.6





#16
Pyrene
Concen: 171.80 ppb
RT: 11.56 min Scan# 836
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:202 Resp: 2238

Ion	Ratio	Lower	Upper
202	100		
101	13.8	14.3	26.5#
100	11.3	10.9	20.3





#18
Benz (a) anthracene
Concen: 116.42 ppb
RT: 12.96 min Scan# 956
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:228 Resp: 1222

Ion	Ratio	Lower	Upper
228	100		
229	21.0	13.1	24.3
226	27.8	18.3	33.9





#19
Chrysene
Concen: 106.60 ppb
RT: 13.00 min Scan# 959
Delta R.T. -0.00 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:228 Resp: 1164

Ion	Ratio	Lower	Upper
228	100		
114	9.6	8.5	15.9
229	21.0	14.1	26.1
226	29.2	19.0	35.4





#20
Indeno (1,2,3-cd) pyrene
Concen: 138.13 ppb
RT: 16.09 min Scan# 1224
Delta R.T. 0.02 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:276 Resp: 750

Ion	Ratio	Lower	Upper
276	100		
138	30.0	22.4	41.6
227	1.3	0.1	0.1#





#22
Benzo (b) fluoranthene
Concen: 118.41 ppb
RT: 14.15 min Scan# 1058
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:252 Resp: 2104

Ion	Ratio	Lower	Upper
252	100		
253	22.3	15.4	28.6
125	16.2	10.8	20.0





#23
Benzo (k) fluoranthene
Concen: 285.36 ppb
RT: 14.15 min Scan# 1058
Delta R.T. -0.02 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:252 Resp: 2104

Ion	Ratio	Lower	Upper
252	100		
253	22.3	15.6	29.0
125	16.2	7.0	13.0#





#24
Benzo (a) pyrene
Concen: 43.83 ppb
RT: 14.45 min Scan# 1084
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:252 Resp: 772

Ion	Ratio	Lower	Upper
252	100		
253	19.5	15.0	27.8
125	22.3	15.7	29.1





#25
Dibenz (a,h) anthracene
Concen: 53.23 ppb
RT: 16.11 min Scan# 1226
Delta R.T. 0.01 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:278 Resp: 207

Ion	Ratio	Lower	Upper
278	100		
139	31.3	17.9	33.3
279	14.6	16.7	30.9#





#26
Benzo (g,h,i) perylene
Concen: 65.59 ppb
RT: 16.52 min Scan# 1261
Delta R.T. 0.02 min
Lab File: 0515L237.D
Acq: 27 May 15 20:08

Tgt Ion:276 Resp: 713

Ion	Ratio	Lower	Upper
276	100		
138	38.9	22.3	41.5
277	22.7	16.4	30.4





#5
1-Methylnaphthalene
Concen: 6.34 ppb
RT: 7.08 min Scan# 461
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:142 Resp: 44

Ion	Ratio	Lower	Upper
142	100		
141	80.8	63.1	117.3





#8
Acenaphthylene
Concen: 96.75 ppb
RT: 8.00 min Scan# 537
Delta R.T. -0.00 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:152 Resp: 156

Ion	Ratio	Lower	Upper
152	100		
151	25.0	12.1	22.5#
153	20.8	9.2	17.2#





#9
Acenaphthene
Concen: 5.63 ppb
RT: 8.21 min Scan# 554
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:154 Resp: 41

Ion	Ratio	Lower	Upper
154	100		
153	96.9	74.9	139.1
152	34.4	35.9	66.7#





#10
Fluorene
Concen: 28.70 ppb
RT: 8.82 min Scan# 605
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:166 Resp: 64

Ion	Ratio	Lower	Upper
166	100		
165	84.6	60.7	112.7





#12
Phenanthrene
Concen: 76.70 ppb
RT: 9.93 min Scan# 697
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:178	Resp:	1281	
Ion	Ratio	Lower	Upper
178	100		
179	15.7	10.6	19.6
176	16.8	12.2	22.6





#13
Anthracene
Concen: 28.67 ppb
RT: 9.99 min Scan# 702
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:178	Resp:	276	
Ion	Ratio	Lower	Upper
178	100		
179	13.6	10.5	19.5
176	17.0	12.0	22.4





#14
Fluoranthene
Concen: 119.87 ppb
RT: 11.30 min Scan# 814
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:202 Resp: 2594

Ion	Ratio	Lower	Upper
202	100		
101	10.9	11.3	20.9#
100	8.1	7.8	14.6





#16
Pyrene
Concen: 145.18 ppb
RT: 11.56 min Scan# 836
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:202 Resp: 2330

Ion	Ratio	Lower	Upper
202	100		
101	14.2	14.3	26.5#
100	10.3	10.9	20.3#





#18
Benz (a) anthracene
Concen: 105.75 ppb
RT: 12.96 min Scan# 956
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:228 Resp: 1350

Ion	Ratio	Lower	Upper
228	100		
229	20.2	13.1	24.3
226	25.6	18.3	33.9





#19
Chrysene
Concen: 79.54 ppb
RT: 13.00 min Scan# 959
Delta R.T. -0.00 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:228 Resp: 1070

Ion	Ratio	Lower	Upper
228	100		
114	10.8	8.5	15.9
229	19.9	14.1	26.1
226	28.2	19.0	35.4





#20
Indeno (1,2,3-cd) pyrene
Concen: 127.36 ppb
RT: 16.09 min Scan# 1224
Delta R.T. 0.02 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:276 Resp: 764

Ion	Ratio	Lower	Upper
276	100		
138	34.4	22.4	41.6
227	0.0	0.1	0.1#





#22
Benzo (b) fluoranthene
Concen: 97.48 ppb
RT: 14.15 min Scan# 1058
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:252 Resp: 2214

Ion	Ratio	Lower	Upper
252	100		
253	20.6	15.4	28.6
125	15.2	10.8	20.0





#23
Benzo (k) fluoranthene
Concen: 262.69 ppb
RT: 14.15 min Scan# 1058
Delta R.T. -0.02 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:252 Resp: 2214

Ion	Ratio	Lower	Upper
252	100		
253	20.6	15.6	29.0
125	15.2	7.0	13.0#





#24
Benzo (a) pyrene
Concen: 36.59 ppb
RT: 14.45 min Scan# 1084
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:252 Resp: 821

Ion	Ratio	Lower	Upper
252	100		
253	17.1	15.0	27.8
125	24.4	15.7	29.1





#25
Dibenz (a,h) anthracene
Concen: 50.52 ppb
RT: 16.11 min Scan# 1226
Delta R.T. 0.01 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:278 Resp: 214

Ion	Ratio	Lower	Upper
278	100		
139	21.1	17.9	33.3
279	24.6	16.7	30.9





#26
Benzo (g,h,i) perylene
Concen: 63.82 ppb
RT: 16.52 min Scan# 1261
Delta R.T. 0.02 min
Lab File: 0515L238.D
Acq: 27 May 15 20:33

Tgt Ion:276 Resp: 830

Ion	Ratio	Lower	Upper
276	100		
138	31.9	22.3	41.5
277	23.4	16.4	30.4



EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16139
QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0021 U	0.006	0.0021	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.0015 J	0.006	0.0021	0.0011	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	28.7 #	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	48.4	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	83.6	30-125			%	5/21/2015	5/27/2015

J = Estimated value.
# = Recovery (or RPD) is outside QC limits.

Quant Method: P0518-42.M
Run #: 0515L239
Instrument: Linus
Sequence: L150515
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 2:14:02 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L239.D Vial: 39
Acq On : 27 May 15 20:59 Operator: MA
Sample : AZ16139S01 1/30.07G Inst : Linus
Misc : soil Multiplr: 33.26

Quant Time: Jun 1 10:55 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9276	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4244	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.91	188	7481	2.50	ppb		0.01
15) Chrysene-D12(IS)	12.97	240	8062	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	5428	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	1247	40.22	ppb		0.01
Spiked Amount 83.139			Recovery = 48.374%				
7) Surrogate Recovery (FBP)	7.41	172	1927	23.90	ppb		0.01
Spiked Amount 83.139			Recovery = 28.747%				
17) Surrogate Recovery (TPH)	11.77	244	5409	69.48	ppb		0.00
Spiked Amount 83.139			Recovery = 83.565%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	137	1.26	ppb		92
12) Phenanthrene	9.93	178	29	-1.43	ppb	#	79
14) Fluoranthene	11.30	202	48	-3.51	ppb	#	79
22) Benzo (b) fluoranthene	14.16	252	56	-5.28	ppb	#	68
24) Benzo (a) pyrene	14.45	252	33	-1.95	ppb	#	54

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L239.D Vial: 39
Acq On : 27 May 15 20:59 Operator: MA
Sample : AZ16139S01 1/30.07G Inst : Linus
Misc : soil Multiplr: 33.26

Quant Time: Jun 1 10:55 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 1.26 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L239.D
Acq: 27 May 15 20:59

Tgt Ion:128 Resp: 137

Ion	Ratio	Lower	Upper
128	100		
129	13.9	7.8	14.6
127	14.8	8.0	14.8



EPA 8270D LL SOILS RE-EXTRACT

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16139
QCG: #SIMDD-150601A-197777

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	ACENAPHTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	ACENAPHTHYLENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	ANTHRACENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (A) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0021 U	0.006	0.0021	0.0016	mg/kg	6/1/2015	6/1/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	CHRYSENE	0.0021 U	0.006	0.0021	0.0010	mg/kg	6/1/2015	6/1/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	FLUORANTHENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	6/1/2015	6/1/2015
8270D-LL	FLUORENE	0.0021 U	0.006	0.0021	0.0012	mg/kg	6/1/2015	6/1/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.0021 U	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	NAPHTHALENE	0.0024 J	0.006	0.0021	0.0011	mg/kg	6/1/2015	6/1/2015
8270D-LL	PHENANTHRENE	0.0021 U	0.006	0.0021	0.0013	mg/kg	6/1/2015	6/1/2015
8270D-LL	PYRENE	0.0021 U	0.006	0.0021	0.0015	mg/kg	6/1/2015	6/1/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	57.7	45-105			%	6/1/2015	6/1/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	57.6	35-100			%	6/1/2015	6/1/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	90.0	30-125			%	6/1/2015	6/1/2015

J = Estimated value.

Quant Method: P0518-42.M
Run #: 0515L264
Instrument: Linus
Sequence: L150515
Dilution Factor: 1
Initials: DA

Printed: 6/5/2015 2:20:02 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L264.D Vial: 64
Acq On : 1 Jun 15 16:51 Operator: MA
Sample : AZ16139S02 1/30.21G Inst : Linus
Misc : soil Multiplr: 33.10

Quant Time: Jun 5 13:32 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8(IS)	6.16	136	8425	2.50	ppb	0.00
6) Acenaphthene-D10(IS)	8.17	164	4181	2.50	ppb	0.01
11) Phenanthrene-D10(IS)	9.91	188	7089	2.50	ppb	0.01
15) Chrysene-D12(IS)	12.97	240	7288	2.50	ppb	0.01
21) Perylene-D12(IS)	14.59	264	4691	2.50	ppb	0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.37	82	1385	47.68	ppb	0.01
Spiked Amount 82.754			Recovery = 57.621%			
7) Surrogate Recovery (FBP)	7.41	172	3810	47.74	ppb	0.01
Spiked Amount 82.754			Recovery = 57.693%			
17) Surrogate Recovery (TPH)	11.77	244	5268	74.50	ppb	0.00
Spiked Amount 82.754			Recovery = 90.029%			
Target Compounds						Qvalue
3) Naphthalene	6.18	128	195	1.96	ppb	97

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L264.D Vial: 64
Acq On : 1 Jun 15 16:51 Operator: MA
Sample : AZ16139S02 1/30.21G Inst : Linus
Misc : soil Multiplr: 33.10

Quant Time: Jun 5 13:32 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 1.96 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L264.D
Acq: 1 Jun 15 16:51

Tgt Ion:128 Resp: 195

Ion	Ratio	Lower	Upper
128	100		
129	9.3	7.8	14.6
127	12.1	8.0	14.8



EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS032-0001

APPL ID: AZ16140

Sample Collection Date: 5/7/2015

QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.080	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.067	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.21	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.33	0.056	0.0190	0.0089	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.83	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.44	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	1.3	0.056	0.0190	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.40	0.056	0.0190	0.0150	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	1.5	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.86	0.056	0.0190	0.0089	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.16	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	1.4	0.056	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.12	0.056	0.0190	0.0110	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.47	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.091	0.056	0.0190	0.0100	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	1.1	0.056	0.0190	0.0120	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	1.3	0.056	0.0190	0.0130	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	53.8	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	112 #	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	72.0	30-125			%	5/21/2015	5/27/2015

= Recovery (or RPD) is outside QC limits.

Quant Method: P0518-42.M
Run #: 0515L240
Instrument: Linus
Sequence: L150515
Dilution Factor: 10
Initials: DA

Printed: 6/2/2015 2:14:02 PM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L240.D Vial: 40
Acq On : 27 May 15 21:24 Operator: MA
Sample : AZ16140S01 1/30.07G df10 Inst : Linus
Misc : soil Multiplr: 332.56

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	10974	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	5406	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	9471	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	10019	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	6837	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	152	92.75	ppb		0.01
Spiked Amount 83.139			Recovery = 111.562%				
7) Surrogate Recovery (FBP)	7.41	172	459	44.69	ppb		0.01
Spiked Amount 83.139			Recovery = 53.754%				
17) Surrogate Recovery (TPH)	11.77	244	579	59.84	ppb		0.00
Spiked Amount 83.139			Recovery = 71.979%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	1050	81.35	ppb		98
4) 2-Methylnaphthalene	6.98	142	462	71.47	ppb		92
5) 1-Methylnaphthalene	7.08	142	476	58.49	ppb		98
8) Acenaphthylene	8.00	152	1645	185.10	ppb		99
9) Acenaphthene	8.21	154	493	59.90	ppb		94
10) Fluorene	8.81	166	911	108.47	ppb		93
12) Phenanthrene	9.92	178	15102	979.35	ppb		99
13) Anthracene	9.98	178	4114	291.73	ppb		100
14) Fluoranthene	11.29	202	24179	1296.57	ppb		91
16) Pyrene	11.56	202	20319	1185.75	ppb	#	81
18) Benz (a) anthracene	12.96	228	11491	744.92	ppb		99
19) Chrysene	13.00	228	11006	766.26	ppb		99
20) Indeno (1,2,3-cd) pyrene	16.08	276	5793	423.85	ppb	#	95
22) Benzo (b) fluoranthene	14.14	252	19224	1172.45	ppb		96
23) Benzo (k) fluoranthene	14.14	252	19224	1363.26	ppb	#	91
24) Benzo (a) pyrene	14.45	252	6350	393.73	ppb		94
25) Dibenz (a,h) anthracene	16.11	278	1573	138.96	ppb		89
26) Benzo (g,h,i) perylene	16.51	276	5325	354.75	ppb		97

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L240.D Vial: 40
Acq On : 27 May 15 21:24 Operator: MA
Sample : AZ16140S01 1/30.07G df10 Inst : Linus
Misc : soil Multiplr: 332.56

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 81.35 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:128 Resp: 1050

Ion	Ratio	Lower	Upper
128	100		
129	10.2	7.8	14.6
127	11.9	8.0	14.8





#4
2-Methylnaphthalene
Concen: 71.47 ppb
RT: 6.98 min Scan# 452
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:142 Resp: 462

Ion	Ratio	Lower	Upper
142	100		
141	78.9	60.1	111.5





#5
1-Methylnaphthalene
Concen: 58.49 ppb
RT: 7.08 min Scan# 461
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:142 Resp: 476

Ion	Ratio	Lower	Upper
142	100		
141	87.9	63.1	117.3





#8
Acenaphthylene
Concen: 185.10 ppb
RT: 8.00 min Scan# 537
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:152 Resp: 1645

Ion	Ratio	Lower	Upper
152	100		
151	18.3	12.1	22.5
153	13.1	9.2	17.2





#9
Acenaphthene
Concen: 59.90 ppb
RT: 8.21 min Scan# 554
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:154 Resp: 493

Ion	Ratio	Lower	Upper
154	100		
153	100.3	74.9	139.1
152	48.4	35.9	66.7





#10
Fluorene
Concen: 108.47 ppb
RT: 8.81 min Scan# 604
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:166 Resp: 911

Ion	Ratio	Lower	Upper
166	100		
165	93.5	60.7	112.7





#12
Phenanthrene
Concen: 979.35 ppb
RT: 9.92 min Scan# 696
Delta R.T. 0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:178 Resp: 15102

Ion	Ratio	Lower	Upper
178	100		
179	14.7	10.6	19.6
176	18.3	12.2	22.6





#13
Anthracene
Concen: 291.73 ppb
RT: 9.98 min Scan# 701
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:178 Resp: 4114

Ion	Ratio	Lower	Upper
178	100		
179	15.2	10.5	19.5
176	17.4	12.0	22.4





#14
Fluoranthene
Concen: 1296.57 ppb
RT: 11.29 min Scan# 813
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:202 Resp: 24179

Ion	Ratio	Lower	Upper
202	100		
101	20.3	11.3	20.9
100	14.1	7.8	14.6





#16
Pyrene
Concen: 1185.75 ppb
RT: 11.56 min Scan# 836
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:202 Resp: 20319

Ion	Ratio	Lower	Upper
202	100		
101	11.3	14.3	26.5#
100	8.0	10.9	20.3#





#18
Benz (a) anthracene
Concen: 744.92 ppb
RT: 12.96 min Scan# 956
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:228 Resp: 11491

Ion	Ratio	Lower	Upper
228	100		
229	19.0	13.1	24.3
226	25.4	18.3	33.9





#19
Chrysene
Concen: 766.26 ppb
RT: 13.00 min Scan# 959
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:228 Resp: 11006

Ion	Ratio	Lower	Upper
228	100		
114	13.0	8.5	15.9
229	21.4	14.1	26.1
226	27.1	19.0	35.4





#20
Indeno (1,2,3-cd) pyrene
Concen: 423.85 ppb
RT: 16.08 min Scan# 1223
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:276 Resp: 5793

Ion	Ratio	Lower	Upper
276	100		
138	29.4	22.4	41.6
227	0.2	0.1	0.1#





#22
Benzo (b) fluoranthene
Concen: 1172.45 ppb
RT: 14.14 min Scan# 1057
Delta R.T. -0.00 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:252 Resp: 19224

Ion	Ratio	Lower	Upper
252	100		
253	21.3	15.4	28.6
125	19.1	10.8	20.0





#23
Benzo (k) fluoranthene
Concen: 1363.26 ppb
RT: 14.14 min Scan# 1057
Delta R.T. -0.04 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:252 Resp: 19224

Ion	Ratio	Lower	Upper
252	100		
253	21.3	15.6	29.0
125	19.1	7.0	13.0#





#24
Benzo (a) pyrene
Concen: 393.73 ppb
RT: 14.45 min Scan# 1084
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:252 Resp: 6350

Ion	Ratio	Lower	Upper
252	100		
253	19.8	15.0	27.8
125	18.1	15.7	29.1





#25
Dibenz (a,h) anthracene
Concen: 138.96 ppb
RT: 16.11 min Scan# 1226
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:278	Resp:	1573	
Ion	Ratio	Lower	Upper
278	100		
139	19.7	17.9	33.3
279	28.8	16.7	30.9





#26
Benzo (g,h,i) perylene
Concen: 354.75 ppb
RT: 16.51 min Scan# 1260
Delta R.T. 0.01 min
Lab File: 0515L240.D
Acq: 27 May 15 21:24

Tgt Ion:276	Resp:	5325	
Ion	Ratio	Lower	Upper
276	100		
138	33.5	22.3	41.5
277	24.5	16.4	30.4



EPA 8270D LL SOILS

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 5/7/2015

ARF: 76380
APPL ID: AZ16141
QCG: #SIMDD-150521A-197621

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)								
8270D-LL	2-METHYLNAPHTHALENE	0.0034 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHENE	0.0022 U	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	ACENAPHTHYLENE	0.012	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	ANTHRACENE	0.0025 J	0.006	0.0022	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZ (A) ANTHRACENE	0.0054 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (A) PYRENE	0.0022 U	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (B) FLUORANTHENE	0.0022 U	0.006	0.0022	0.0014	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (G,H,I) PERYLENE	0.0022 U	0.006	0.0022	0.0017	mg/kg	5/21/2015	5/27/2015
8270D-LL	BENZO (K) FLUORANTHENE	0.019	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	CHRYSENE	0.0032 J	0.006	0.0022	0.0010	mg/kg	5/21/2015	5/27/2015
8270D-LL	DIBENZ (A,H) ANTHRACENE	0.0022 U	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORANTHENE	0.0022 U	0.006	0.0022	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	FLUORENE	0.0022 U	0.006	0.0022	0.0013	mg/kg	5/21/2015	5/27/2015
8270D-LL	INDENO (1,2,3-CD) PYRENE	0.012	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	NAPHTHALENE	0.0021 J	0.006	0.0022	0.0012	mg/kg	5/21/2015	5/27/2015
8270D-LL	PHENANTHRENE	0.0028 J	0.006	0.0022	0.0014	mg/kg	5/21/2015	5/27/2015
8270D-LL	PYRENE	0.0053 J	0.006	0.0022	0.0016	mg/kg	5/21/2015	5/27/2015
8270D-LL	SURROGATE: 2-FLUORBIPHENYL (S)	45.1	45-105			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: NITROBENZENE-D5 (S)	52.1	35-100			%	5/21/2015	5/27/2015
8270D-LL	SURROGATE: TERPHENYL-D14 (S)	86.8	30-125			%	5/21/2015	5/27/2015

J = Estimated value.

Quant Method: P0518-42.M
Run #: 0515L241
Instrument: Linus
Sequence: L150515
Dilution Factor: 1
Initials: DA

Printed: 6/2/2015 2:14:16 PM
APPL-F1-SC-MCRes/MCPQL-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L241.D Vial: 41
Acq On : 27 May 15 21:50 Operator: MA
Sample : AZ16141S01 1/30.23G Inst : Linus
Misc : soil Multiplr: 33.08

Quant Time: Jun 1 11:01 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9325	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4181	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	7452	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	8223	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	5478	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	1367	43.11	ppb		0.01
Spiked Amount 82.699			Recovery = 52.129%				
7) Surrogate Recovery (FBP)	7.41	172	2979	37.31	ppb		0.01
Spiked Amount 82.699			Recovery = 45.109%				
17) Surrogate Recovery (TPH)	11.77	244	5728	71.75	ppb		0.00
Spiked Amount 82.699			Recovery = 86.760%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	177	1.61	ppb		96
4) 2-Methylnaphthalene	6.98	142	25	2.62	ppb		93
5) 1-Methylnaphthalene	7.08	142	20	0.29	ppb		88
8) Acenaphthylene	8.00	152	105	9.43	ppb		93
12) Phenanthrene	9.93	178	456	2.14	ppb		95
13) Anthracene	9.99	178	133	1.92	ppb		96
14) Fluoranthene	11.30	202	697	1.04	ppb	#	90
16) Pyrene	11.56	202	578	4.09	ppb	#	88
18) Benz (a) anthracene	12.96	228	350	4.11	ppb		94
19) Chrysene	13.00	228	288	2.43	ppb	#	91
20) Indeno (1,2,3-cd) pyrene	16.10	276	187	9.24	ppb	#	73
22) Benzo (b) fluoranthene	14.15	252	524	-1.54	ppb	#	89
23) Benzo (k) fluoranthene	14.15	252	524	14.74	ppb	#	85
24) Benzo (a) pyrene	14.47	252	167	-0.85	ppb		96

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L241.D Vial: 41
Acq On : 27 May 15 21:50 Operator: MA
Sample : AZ16141S01 1/30.23G Inst : Linus
Misc : soil Multiplr: 33.08

Quant Time: Jun 1 11:01 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration





#3
Naphthalene
Concen: 1.61 ppb
RT: 6.18 min Scan# 386
Delta R.T. 0.00 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:128 Resp: 177

Ion	Ratio	Lower	Upper
128	100		
129	9.0	7.8	14.6
127	12.0	8.0	14.8





#4
2-Methylnaphthalene
Concen: 2.62 ppb
RT: 6.98 min Scan# 452
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:142 Resp: 25

Ion	Ratio	Lower	Upper
142	100		
141	92.3	60.1	111.5





#5
1-Methylnaphthalene
Concen: 0.29 ppb
RT: 7.08 min Scan# 461
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:142 Resp: 20

Ion	Ratio	Lower	Upper
142	100		
141	78.6	63.1	117.3





#8
Acenaphthylene
Concen: 9.43 ppb
RT: 8.00 min Scan# 537
Delta R.T. -0.00 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:152 Resp: 105

Ion	Ratio	Lower	Upper
152	100		
151	21.3	12.1	22.5
153	14.9	9.2	17.2





#12
Phenanthrene
Concen: 2.14 ppb
RT: 9.93 min Scan# 697
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:178 Resp: 456

Ion	Ratio	Lower	Upper
178	100		
179	13.3	10.6	19.6
176	19.5	12.2	22.6





#13
Anthracene
Concen: 1.92 ppb
RT: 9.99 min Scan# 702
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:178 Resp: 133

Ion	Ratio	Lower	Upper
178	100		
179	16.7	10.5	19.5
176	15.5	12.0	22.4





#14
Fluoranthene
Concen: 1.04 ppb
RT: 11.30 min Scan# 814
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:202 Resp: 697

Ion	Ratio	Lower	Upper
202	100		
101	12.0	11.3	20.9
100	7.5	7.8	14.6#





#16
Pyrene
Concen: 4.09 ppb
RT: 11.56 min Scan# 836
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:202 Resp: 578

Ion	Ratio	Lower	Upper
202	100		
101	13.9	14.3	26.5#
100	12.0	10.9	20.3





#18
Benz (a) anthracene
Concen: 4.11 ppb
RT: 12.96 min Scan# 956
Delta R.T. 0.01 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:228 Resp: 350

Ion	Ratio	Lower	Upper
228	100		
229	20.9	13.1	24.3
226	30.0	18.3	33.9





#19
Chrysene
Concen: 2.43 ppb
RT: 13.00 min Scan# 959
Delta R.T. -0.00 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:228 Resp: 288

Ion	Ratio	Lower	Upper
228	100		
114	19.9	8.5	15.9#
229	22.2	14.1	26.1
226	31.0	19.0	35.4





#20
Indeno (1,2,3-cd) pyrene
Concen: 9.24 ppb
RT: 16.10 min Scan# 1225
Delta R.T. 0.03 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:276 Resp: 187

Ion	Ratio	Lower	Upper
276	100		
138	46.9	22.4	41.6#
227	4.1	0.1	0.1#





#23
Benzo (k) fluoranthene
Concen: 14.74 ppb
RT: 14.15 min Scan# 1058
Delta R.T. -0.02 min
Lab File: 0515L241.D
Acq: 27 May 15 21:50

Tgt Ion:252 Resp: 524

Ion	Ratio	Lower	Upper
252	100		
253	21.0	15.6	29.0
125	25.2	7.0	13.0#



EPA METHOD 8270D
Polynuclear Aromatic Hydrocarbons
Calibration Data

APPL, INC.

EPA 8270C SIM
P0518-42

Form 6
Initial Calibration

Lab Name: APPL, Inc. SDG No:
Case No: Initial Cal. Date: 5/15/2015
Matrix: Instrument: Linus Initials:

		Compound	0515L003.D 0.1	0515L004.D 0.2	0515L005 D 0.5	0515L006.D 1	0515L007 D 5	0515L008 D 10	0515L009.D 50	0515L010.D 100			Avg	%RSD		
1	I	Napthalene-D8(IS)														
2	SL	Surrogate Recovery (NBZ)	0.2254	0.3678	0.1834	0.1860	0.2776	0.2909	0.3391	0.3198			0.27	25	SL	0.999
3	TM	Naphthalene	0.8554	0.8475	0.7845	0.9100	1.049	1.089	1.174	1.113			0.98	15	TM	
4	TML	2-Methylnaphthalene	0.4994	0.5010	0.4785	0.5210	0.6104	0.6851	0.7669	0.7167			0.60	19	TML	0.999
5	TM	1-Methylnaphthalene	0.5762	0.5810	0.4580	0.5662	0.6341	0.6730	0.7484	0.6954			0.62	15	TM	
6	I	Acenaphthene-D10(IS)														
7	S	Surrogate Recovery (FBP)				1.217	1.812	1.617	1.675	1.577			1.6	14	S	
8	TML	Acenaphthylene	1.795	1.721	1.606	1.730	2.314	2.330	2.600	2.556			2.1	20	TML	1.000
9	*TM	Acenaphthene	1.152	1.120	1.004	1.083	1.397	1.402	1.500	1.468			1.3	15	*TM	
10	TML	Fluorene	1.331	1.190	1.050	1.116	1.471	1.579	1.671	1.606			1.4	17	TML	1.000
11	I	Phenanthrene-D10(IS)														
12	TML	Phenanthrene	1.072	1.012	0.9085	0.9592	1.237	1.307	1.401	1.315			1.2	16	TML	0.999
13	TML	Anthracene	0.9138	0.9496	0.8756	0.9497	1.174	1.229	1.322	1.259			1.1	17	TML	0.999
14	*TML	Fluoranthene	1.292	1.160	1.048	1.124	1.461	1.555	1.707	1.564			1.4	18	*TML	0.998
15	I	Chrysene-D12(IS)														
16	TM	Pyrene	1.385	1.349	1.165	1.236	1.554	1.564	1.593	1.529			1.4	11	TM	
17	S	Surrogate Recovery (TPH)	0.7828	0.7480	0.6866	0.6959	0.9635	0.8453	0.8659	0.8351			0.80	12	S	
18	TML	Benz (a) anthracene	1.074	0.9169	0.8316	0.8680	1.146	1.248	1.372	1.288			1.1	19	TML	0.999
19	TM	Chrysene	1.181	1.069	1.140	1.165	1.314	1.284	1.211	1.171			1.2	6.6	TM	
20	TML	Indeno (1,2,3-cd) pyrene	1.032	0.8816	0.8741	0.9520	1.205	1.296	1.416	1.386			1.1	20	TML	1.000
21	I	Perylene-D12(IS)														
22	TML	Benzo (b) fluoranthene	1.181	1.366	0.9970	1.080	1.710	1.633	2.174	1.841			1.5	27	TML	0.993
23	TML	Benzo (k) fluoranthene	1.572	0.9926	1.318	1.357	1.658	1.985	1.740	1.877			1.6	21	TML	0.999
24	*TML	Benzo (a) pyrene	1.523	1.336	1.301	1.388	1.764	1.857	1.940	1.838			1.6	16	*TML	0.999
25	TML	Dibenz (a,h) anthracene	1.133	1.051	1.104	1.146	1.622	1.761	1.922	1.826			1.4	26	TML	0.999
26	TML	Benzo (g,h,i) perylene	1.398	1.182	1.173	1.261	1.698	1.811	1.927	1.828			1.5	20	TML	0.999
27																
28																
29																
30																
31																
32																
33																
34																
35																

Data File : M:\LINUS\DATA\L150515\0515L003.D Vial: 3
Acq On : 15 May 15 17:15 Operator: csc
Sample : 0.1ug/ml PAH 05/15/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 10:26 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9762	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4471	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	7606	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	7218	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	4613	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	88	0.24	ppb		0.01
Spiked Amount 2.500			Recovery = 9.680%				
7) Surrogate Recovery (FBP)	7.41	172	230	-0.18	ppb		0.01
Spiked Amount 2.500			Recovery = -7.400%				
17) Surrogate Recovery (TPH)	11.77	244	226	0.10	ppb		0.00
Spiked Amount 2.500			Recovery = 3.880%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	334	0.09	ppb		96
4) 2-Methylnaphthalene	6.96	142	195	0.14	ppb		95
5) 1-Methylnaphthalene	7.07	142	225	0.09	ppb		100
8) Acenaphthylene	8.00	152	321	0.33	ppb		96
9) Acenaphthene	8.19	154	208	0.09	ppb		98
10) Fluorene	8.81	166	238	0.15	ppb		99
13) Anthracene	9.98	178	278	0.09	ppb		94
16) Pyrene	11.56	202	400	0.10	ppb		92
18) Benz (a) anthracene	12.96	228	310	0.12	ppb		97
19) Chrysene	13.00	228	341	0.10	ppb		94
20) Indeno (1,2,3-cd) pyrene	16.09	276	298	0.31	ppb	#	97
22) Benzo (b) fluoranthene	14.14	252	218	-0.11	ppb		96
23) Benzo (k) fluoranthene	14.17	252	290	0.40	ppb	#	94
25) Dibenz (a,h) anthracene	16.12	278	209	0.17	ppb		90
26) Benzo (g,h,i) perylene	16.51	276	258	0.09	ppb		88

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L003.D Vial: 3
Acq On : 15 May 15 17:15 Operator: csc
Sample : 0.1ug/ml PAH 05/15/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 10:26 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L004.D Vial: 4
Acq On : 15 May 15 17:41 Operator: csc
Sample : 0.2ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9381	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4264	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	7253	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	6735	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	4483	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	138	0.81	ppb		0.01
Spiked Amount 2.500			Recovery =	32.400%			
7) Surrogate Recovery (FBP)	7.41	172	438	0.18	ppb		0.01
Spiked Amount 2.500			Recovery =	7.280%			
17) Surrogate Recovery (TPH)	11.77	244	403	0.20	ppb		0.00
Spiked Amount 2.500			Recovery =	8.000%			
Target Compounds							Qvalue
3) Naphthalene	6.18	128	636	0.50	ppb		99
4) 2-Methylnaphthalene	6.96	142	376	0.79	ppb		100
5) 1-Methylnaphthalene	7.07	142	436	0.52	ppb		99
8) Acenaphthylene	8.00	152	587	0.32	ppb		99
9) Acenaphthene	8.21	154	382	0.40	ppb		91
10) Fluorene	8.81	166	406	0.46	ppb		91
12) Phenanthrene	9.92	178	587	0.19	ppb		99
13) Anthracene	9.98	178	551	0.55	ppb		99
14) Fluoranthene	11.30	202	673	0.78	ppb		92
16) Pyrene	11.56	202	727	0.21	ppb		91
18) Benz (a) anthracene	12.96	228	494	0.17	ppb		93
19) Chrysene	13.00	228	576	0.20	ppb		97
20) Indeno (1,2,3-cd) pyrene	16.09	276	475	0.63	ppb	#	92
22) Benzo (b) fluoranthene	14.14	252	490	0.41	ppb	#	89
23) Benzo (k) fluoranthene	14.17	252	356	0.38	ppb		97
24) Benzo (a) pyrene	14.45	252	479	0.38	ppb		97
25) Dibenz (a,h) anthracene	16.12	278	377	0.57	ppb		95
26) Benzo (g,h,i) perylene	16.51	276	424	0.55	ppb		98

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L004.D Vial: 4
Acq On : 15 May 15 17:41 Operator: csc
Sample : 0.2ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L005.D Vial: 5
Acq On : 15 May 15 18:06 Operator: csc
Sample : 0.5ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9050	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4233	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	7149	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	6590	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	4308	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	332	0.98	ppb		0.01
Spiked Amount 2.500			Recovery = 39.000%				
7) Surrogate Recovery (FBP)	7.41	172	955	0.40	ppb		0.01
Spiked Amount 2.500			Recovery = 15.960%				
17) Surrogate Recovery (TPH)	11.77	244	905	0.46	ppb		0.00
Spiked Amount 2.500			Recovery = 18.400%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	1420	0.70	ppb		100
4) 2-Methylnaphthalene	6.96	142	866	0.97	ppb		99
5) 1-Methylnaphthalene	7.07	142	829	0.67	ppb		94
8) Acenaphthylene	8.00	152	1360	0.51	ppb		97
9) Acenaphthene	8.21	154	850	0.60	ppb		96
10) Fluorene	8.81	166	889	0.64	ppb		94
12) Phenanthrene	9.92	178	1299	0.42	ppb		98
13) Anthracene	9.98	178	1252	0.76	ppb		98
14) Fluoranthene	11.30	202	1498	0.96	ppb		93
16) Pyrene	11.56	202	1536	0.46	ppb		91
18) Benz (a) anthracene	12.96	228	1096	0.39	ppb		96
19) Chrysene	13.00	228	1502	0.53	ppb		97
20) Indeno (1,2,3-cd) pyrene	16.08	276	1152	0.83	ppb	#	90
22) Benzo (b) fluoranthene	14.14	252	859	0.55	ppb	#	96
23) Benzo (k) fluoranthene	14.17	252	1136	0.72	ppb		97
24) Benzo (a) pyrene	14.45	252	1121	0.64	ppb		94
25) Dibenz (a,h) anthracene	16.11	278	951	0.80	ppb		94
26) Benzo (g,h,i) perylene	16.51	276	1011	0.80	ppb		97

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L005.D Vial: 5
Acq On : 15 May 15 18:06 Operator: csc
Sample : 0.5ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L006.D Vial: 6
Acq On : 15 May 15 18:32 Operator: csc
Sample : 1.0ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	8964	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	4488	2.50	ppb		0.00
11) Phenanthrene-D10(IS)	9.90	188	7363	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	6941	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	4617	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	667	1.26	ppb		0.01
Spiked Amount 2.500			Recovery = 50.400%				
7) Surrogate Recovery (FBP)	7.41	172	2185	0.86	ppb		0.01
Spiked Amount 2.500			Recovery = 34.440%				
17) Surrogate Recovery (TPH)	11.77	244	1932	0.93	ppb		0.00
Spiked Amount 2.500			Recovery = 37.280%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	3263	1.16	ppb		99
4) 2-Methylnaphthalene	6.96	142	1868	1.34	ppb		98
5) 1-Methylnaphthalene	7.07	142	2030	1.14	ppb		96
8) Acenaphthylene	8.00	152	3106	0.91	ppb		99
9) Acenaphthene	8.19	154	1945	1.01	ppb		97
10) Fluorene	8.81	166	2003	1.02	ppb		94
12) Phenanthrene	9.92	178	2825	0.89	ppb		99
13) Anthracene	9.98	178	2797	1.18	ppb		98
14) Fluoranthene	11.30	202	3311	1.32	ppb		92
16) Pyrene	11.56	202	3431	0.97	ppb	#	87
18) Benz (a) anthracene	12.96	228	2410	0.81	ppb		97
19) Chrysene	13.00	228	3235	1.09	ppb		98
20) Indeno (1,2,3-cd) pyrene	16.08	276	2643	1.22	ppb	#	97
22) Benzo (b) fluoranthene	14.14	252	1994	0.91	ppb		96
23) Benzo (k) fluoranthene	14.17	252	2506	1.22	ppb		96
24) Benzo (a) pyrene	14.45	252	2564	1.14	ppb		96
25) Dibenz (a,h) anthracene	16.11	278	2117	1.20	ppb		98
26) Benzo (g,h,i) perylene	16.51	276	2328	1.27	ppb		98

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L006.D Vial: 6
Acq On : 15 May 15 18:32 Operator: csc
Sample : 1.0ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L007.D Vial: 7
Acq On : 15 May 15 18:58 Operator: csc
Sample : 5.0ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	7444	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	3374	2.50	ppb		0.00
11) Phenanthrene-D10(IS)	9.90	188	5789	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.96	240	5736	2.50	ppb		0.00
21) Perylene-D12(IS)	14.58	264	3671	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.36	82	4133	4.89	ppb		0.00
Spiked Amount 2.500			Recovery	=	195.520%		
7) Surrogate Recovery (FBP)	7.40	172	12230	6.41	ppb		0.00
Spiked Amount 2.500			Recovery	=	256.280%		
17) Surrogate Recovery (TPH)	11.77	244	11053	6.45	ppb		0.00
Spiked Amount 2.500			Recovery	=	258.000%		
Target Compounds							Qvalue
3) Naphthalene	6.18	128	15624	5.05	ppb		100
4) 2-Methylnaphthalene	6.96	142	9087	4.65	ppb		100
5) 1-Methylnaphthalene	7.07	142	9441	4.77	ppb		100
8) Acenaphthylene	8.00	152	15612	5.09	ppb		100
9) Acenaphthene	8.19	154	9425	5.21	ppb		100
10) Fluorene	8.81	166	9927	5.02	ppb		100
12) Phenanthrene	9.92	178	14320	5.74	ppb		100
13) Anthracene	9.98	178	13594	5.20	ppb		100
14) Fluoranthene	11.29	202	16914	5.05	ppb		100
16) Pyrene	11.55	202	17823	6.08	ppb		100
18) Benz (a) anthracene	12.95	228	13150	5.34	ppb		100
19) Chrysene	13.00	228	15075	6.16	ppb		100
20) Indeno (1,2,3-cd) pyrene	16.06	276	13822	5.05	ppb	#	100
22) Benzo (b) fluoranthene	14.14	252	12555	5.60	ppb		100
23) Benzo (k) fluoranthene	14.16	252	12170	6.25	ppb		100
24) Benzo (a) pyrene	14.44	252	12949	6.16	ppb		100
25) Dibenz (a,h) anthracene	16.10	278	11907	5.92	ppb		100
26) Benzo (g,h,i) perylene	16.50	276	12470	6.31	ppb		100

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L007.D Vial: 7
Acq On : 15 May 15 18:58 Operator: csc
Sample : 5.0ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L008.D Vial: 8
Acq On : 15 May 15 19:23 Operator: csc
Sample : 10ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	10891	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	5030	2.50	ppb		0.00
11) Phenanthrene-D10(IS)	9.88	188	8662	2.50	ppb		-0.01
15) Chrysene-D12(IS)	12.96	240	8905	2.50	ppb		0.00
21) Perylene-D12(IS)	14.57	264	5552	2.50	ppb		-0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.36	82	12674	9.48	ppb		0.00
Spiked Amount 2.500			Recovery = 379.120%				
7) Surrogate Recovery (FBP)	7.40	172	32526	11.43	ppb		0.00
Spiked Amount 2.500			Recovery = 457.200%				
17) Surrogate Recovery (TPH)	11.77	244	30110	11.32	ppb		0.00
Spiked Amount 2.500			Recovery = 452.680%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	47430	10.10	ppb		100
4) 2-Methylnaphthalene	6.96	142	29844	9.62	ppb		99
5) 1-Methylnaphthalene	7.07	142	29320	9.71	ppb		98
8) Acenaphthylene	7.99	152	46873	10.07	ppb		97
9) Acenaphthene	8.19	154	28199	10.20	ppb		99
10) Fluorene	8.80	166	31766	10.43	ppb		92
12) Phenanthrene	9.92	178	45279	12.14	ppb		100
13) Anthracene	9.98	178	42591	10.46	ppb		99
14) Fluoranthene	11.29	202	53862	10.02	ppb		94
16) Pyrene	11.55	202	55714	12.24	ppb		95
18) Benz (a) anthracene	12.95	228	44470	11.64	ppb		98
19) Chrysene	12.98	228	45739	12.03	ppb	#	93
20) Indeno (1,2,3-cd) pyrene	16.05	276	46158	10.29	ppb	#	96
22) Benzo (b) fluoranthene	14.13	252	36269	10.48	ppb		96
23) Benzo (k) fluoranthene	14.16	252	44089	14.63	ppb		95
24) Benzo (a) pyrene	14.44	252	41250	12.77	ppb		95
25) Dibenz (a,h) anthracene	16.09	278	39115	12.37	ppb		98
26) Benzo (g,h,i) perylene	16.48	276	40228	13.02	ppb		98

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L008.D Vial: 8
Acq On : 15 May 15 19:23 Operator: csc
Sample : 10ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L009.D Vial: 9
Acq On : 15 May 15 19:49 Operator: csc
Sample : 50ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:13:03 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9907	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	4610	2.50	ppb		0.00
11) Phenanthrene-D10(IS)	9.88	188	7949	2.50	ppb		-0.01
15) Chrysene-D12(IS)	12.96	240	8664	2.50	ppb		0.00
21) Perylene-D12(IS)	14.57	264	5437	2.50	ppb		-0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.36	82	67181	51.86	ppb		0.00
Spiked Amount 2.500			Recovery	= 2074.400%			
7) Surrogate Recovery (FBP)	7.40	172	154415	59.21	ppb		0.00
Spiked Amount 2.500			Recovery	= 2368.360%			
17) Surrogate Recovery (TPH)	11.76	244	150044	57.96	ppb		-0.01
Spiked Amount 2.500			Recovery	= 2318.600%			
Target Compounds							Qvalue
3) Naphthalene	6.18	128	232647	52.94	ppb		100
4) 2-Methylnaphthalene	6.96	142	151960	50.80	ppb		99
5) 1-Methylnaphthalene	7.07	142	148295	52.39	ppb		99
8) Acenaphthylene	8.00	152	239701	55.40	ppb		99
9) Acenaphthene	8.19	154	138275	53.52	ppb		98
10) Fluorene	8.80	166	154069	53.88	ppb		93
12) Phenanthrene	9.92	178	222761	65.08	ppb		99
13) Anthracene	9.98	178	210206	54.53	ppb		99
14) Fluoranthene	11.29	202	271453	52.11	ppb		95
16) Pyrene	11.55	202	276045	62.33	ppb		99
18) Benz (a) anthracene	12.95	228	237738	63.94	ppb		99
19) Chrysene	13.00	228	209845	56.74	ppb	#	96
20) Indeno (1,2,3-cd) pyrene	16.05	276	245430	54.01	ppb	#	98
22) Benzo (b) fluoranthene	14.14	252	236425	68.41	ppb		97
23) Benzo (k) fluoranthene	14.16	252	189251	63.32	ppb		100
24) Benzo (a) pyrene	14.44	252	210997	65.87	ppb		98
25) Dibenz (a,h) anthracene	16.09	278	208974	65.60	ppb		94
26) Benzo (g,h,i) perylene	16.48	276	209511	67.57	ppb		92

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L009.D Vial: 9
Acq On : 15 May 15 19:49 Operator: csc
Sample : 50ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:13 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L150515\0515L010.D Vial: 10
Acq On : 15 May 15 20:14 Operator: csc
Sample : 100ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:16 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:16:04 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	10112	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	4552	2.50	ppb		0.00
11) Phenanthrene-D10(IS)	9.90	188	7824	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.96	240	8642	2.50	ppb		0.00
21) Perylene-D12(IS)	14.57	264	5573	2.50	ppb		-0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.36	82	129333	98.87	ppb		0.00
Spiked Amount 2.500			Recovery	= 3954.600%			
7) Surrogate Recovery (FBP)	7.40	172	287071	108.77	ppb		0.00
Spiked Amount 2.500			Recovery	= 4350.920%			
17) Surrogate Recovery (TPH)	11.77	244	288680	104.01	ppb		0.00
Spiked Amount 2.500			Recovery	= 4160.560%			
Target Compounds							Qvalue
3) Naphthalene	6.18	128	450212	98.96	ppb		99
4) 2-Methylnaphthalene	6.96	142	289899	98.68	ppb		99
5) 1-Methylnaphthalene	7.08	142	281266	98.57	ppb		89
8) Acenaphthylene	8.00	152	465419	99.67	ppb		99
9) Acenaphthene	8.21	154	267375	99.59	ppb		90
10) Fluorene	8.81	166	292349	99.21	ppb		98
12) Phenanthrene	9.92	178	411564	114.22	ppb		98
13) Anthracene	9.98	178	393985	99.04	ppb		98
14) Fluoranthene	11.29	202	489534	98.27	ppb		96
16) Pyrene	11.55	202	528567	107.53	ppb		91
18) Benz (a) anthracene	12.95	228	445303	117.84	ppb		99
19) Chrysene	13.00	228	404821	98.25	ppb		98
20) Indeno (1,2,3-cd) pyrene	16.06	276	479086	99.57	ppb		94
22) Benzo (b) fluoranthene	14.14	252	410473	96.69	ppb		100
23) Benzo (k) fluoranthene	14.17	252	418502	102.96	ppb		100
24) Benzo (a) pyrene	14.45	252	409725	98.93	ppb		94
25) Dibenz (a,h) anthracene	16.11	278	407020	98.99	ppb	#	91
26) Benzo (g,h,i) perylene	16.51	276	407552	98.96	ppb		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L010.D Vial: 10
Acq On : 15 May 15 20:14 Operator: csc
Sample : 100ug/ml PAH Inst : Linus
Misc : Multiplr: 1.00

Quant Time: May 18 9:16 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration



Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No:
Date Analyzed: 5/15/2015
Instrument: Linus
Initial Cal. Date: 5/15/2015
Data File: 0515L011.D

		Compound	MEAN	CCRF	%D		%Drift
1	TM	Naphthalene	0.9778	1.160	19	TM	
2	TML	2-Methylnaphthalene	0.5974	0.6612	11	TML	7.6
3	TM	1-Methylnaphthalene	0.6165	0.6742	9.3	TM	
4	TML	Acenaphthylene	2.081	2.416	16	TML	0.82
5	*TM	Acenaphthene	1.266	1.506	19	*TM	
6	TML	Fluorene	1.377	1.617	17	TML	1.2
7	TML	Phenanthrene	1.151	1.293	12	TML	3.9
8	TML	Anthracene	1.084	1.193	10	TML	5.7
9	*TML	Fluoranthene	1.364	1.541	13	*TML	5.4
10	TM	Pyrene	1.422	1.591	12	TM	
11	TML	Benz (a) anthracene	1.093	1.209	11	TML	6.5
12	TM	Chrysene	1.192	1.400	17	TM	
13	TML	Indeno (1,2,3-cd) pyrene	1.130	1.364	21	TML	2.6
14	TML	Benzo (b) fluoranthene	1.498	1.918	28	TML	2.6
15	TML	Benzo (k) fluoranthene	1.563	1.660	6.2	TML	4.3
16	*TML	Benzo (a) pyrene	1.618	1.795	11	*TML	4.7
17	TML	Dibenz (a,h) anthracene	1.446	1.775	23	TML	1.7
18	TML	Benzo (g,h,i) perylene	1.535	1.762	15	TML	4.4
19							
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
		Average			15.0		

EPA 8270C SIM
P0518-42

Data File : M:\LINUS\DATA\L150515\0515L011.D Vial: 11
Acq On : 15 May 15 20:40 Operator: csc
Sample : 5.0ug/ml PAH SS 05/15/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: Jun 2 13:17 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9571	2.50 ppb		0.00
6) Acenaphthene-D10(IS)	8.16	164	4405	2.50 ppb		0.00
11) Phenanthrene-D10(IS)	9.90	188	7622	2.50 ppb		0.00
15) Chrysene-D12(IS)	12.96	240	7750	2.50 ppb		0.00
21) Perylene-D12(IS)	14.57	264	4908	2.50 ppb		-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.36	82	6	0.18 ppb		0.00
Spiked Amount 2.500			Recovery =	7.080%		
7) Surrogate Recovery (FBP)	7.40	172	46	0.02 ppb		0.00
Spiked Amount 2.500			Recovery =	0.680%		
17) Surrogate Recovery (TPH)	11.77	244	66	0.03 ppb		0.00
Spiked Amount 2.500			Recovery =	1.080%		
Target Compounds						Qvalue
3) Naphthalene	6.18	128	22202	5.93 ppb		99
4) 2-Methylnaphthalene	6.96	142	12656	4.62 ppb		100
5) 1-Methylnaphthalene	7.07	142	12905	5.47 ppb		100
8) Acenaphthylene	8.00	152	21286	4.96 ppb		99
9) Acenaphthene	8.19	154	13269	5.95 ppb		99
10) Fluorene	8.81	166	14244	5.06 ppb		100
12) Phenanthrene	9.92	178	19706	4.81 ppb		99
13) Anthracene	9.98	178	18184	4.71 ppb		99
14) Fluoranthene	11.29	202	23497	4.73 ppb		99
16) Pyrene	11.55	202	24659	5.59 ppb		98
18) Benz (a) anthracene	12.95	228	18746	4.68 ppb		99
19) Chrysene	13.00	228	21695	5.87 ppb		98
20) Indeno (1,2,3-cd) pyrene	16.05	276	21147	5.13 ppb		93
22) Benzo (b) fluoranthene	14.13	252	18824	4.87 ppb	#	93
23) Benzo (k) fluoranthene	14.16	252	16293	4.78 ppb	#	95
24) Benzo (a) pyrene	14.44	252	17615	4.77 ppb		100
25) Dibenz (a,h) anthracene	16.10	278	17424	4.91 ppb		99
26) Benzo (g,h,i) perylene	16.50	276	17299	4.78 ppb		99

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L011.D Vial: 11
Acq On : 15 May 15 20:40 Operator: csc
Sample : 5.0ug/ml PAH SS 05/15/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: Jun 2 13:17 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



EPA 8270C SIM
P0518-42

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No:
Date Analyzed: 5/27/2015
Instrument: Linus
Initial Cal. Date: 5/15/2015
Data File: 0515L228.D

		Compound	MEAN	CCRF	%D		%Drift
1	I	Napthalene-D8(IS)	ISTD			I	
2	SL	Surrogate Recovery (NBZ)	0.2737	0.2480	9.4	SL	20
3	TM	Naphthalene	0.9778	1.026	5.0	TM	
4	TML	2-Methylnaphthalene	0.5974	0.6372	6.7	TML	11
5	TM	1-Methylnaphthalene	0.6165	0.6495	5.3	TM	
6	I	Acenaphthene-D10(IS)	ISTD			I	
7	S	Surrogate Recovery (FBP)	1.580	1.511	4.3	S	
8	TML	Acenaphthylene	2.081	2.279	9.5	TML	6.2
9	*TM	Acenaphthene	1.266	1.360	7.4	*TM	
10	TML	Fluorene	1.377	1.494	8.5	TML	6.4
11	I	Phenanthrene-D10(IS)	ISTD			I	
12	TML	Phenanthrene	1.151	1.223	6.2	TML	9.1
13	TML	Anthracene	1.084	1.169	7.9	TML	7.6
14	*TML	Fluoranthene	1.364	1.488	9.1	*TML	8.7
15	I	Chrysene-D12(IS)	ISTD			I	
16	TM	Pyrene	1.422	1.483	4.3	TM	
17	S	Surrogate Recovery (TPH)	0.8029	0.7985	0.55	S	
18	TML	Benz (a) anthracene	1.093	1.248	14	TML	3.5
19	TM	Chrysene	1.192	1.214	1.9	TM	
20	TML	Indeno (1,2,3-cd) pyrene	1.130	1.360	20	TML	2.2
21	I	Perylene-D12(IS)	ISTD			I	
22	TML	Benzo (b) fluoranthene	1.498	1.894	26	TML	3.8
23	TML	Benzo (k) fluoranthene	1.563	1.735	11	TML	0.28
24	*TML	Benzo (a) pyrene	1.618	1.761	8.8	*TML	6.5
25	TML	Dibenz (a,h) anthracene	1.446	1.775	23	TML	1.7
26	TML	Benzo (g,h,i) perylene	1.535	1.784	16	TML	3.2
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							
		Average			9.8		

Data File : M:\LINUS\DATA\L150515\0515L228.D Vial: 28
Acq On : 27 May 15 16:17 Operator: MA
Sample : 5.0ug/ml PAH 5/26/15 Inst : Linus
Misc : soil Multiplr: 1.00

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	7511	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	3616	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	6351	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	6697	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	4324	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.36	82	3726	4.00	ppb		0.00
Spiked Amount 2.500			Recovery = 160.000%				
7) Surrogate Recovery (FBP)	7.41	172	10930	4.78	ppb		0.01
Spiked Amount 2.500			Recovery = 191.360%				
17) Surrogate Recovery (TPH)	11.77	244	10695	4.97	ppb		0.00
Spiked Amount 2.500			Recovery = 198.920%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	15420	5.25	ppb		99
4) 2-Methylnaphthalene	6.98	142	9572	4.45	ppb		92
5) 1-Methylnaphthalene	7.08	142	9757	5.27	ppb		92
8) Acenaphthylene	8.00	152	16480	4.69	ppb		98
9) Acenaphthene	8.21	154	9832	5.37	ppb		94
10) Fluorene	8.81	166	10803	4.68	ppb		96
12) Phenanthrene	9.92	178	15536	4.54	ppb		98
13) Anthracene	9.99	178	14853	4.62	ppb		98
14) Fluoranthene	11.30	202	18903	4.56	ppb		94
16) Pyrene	11.56	202	19858	5.21	ppb		91
18) Benz (a) anthracene	12.96	228	16717	4.83	ppb		98
19) Chrysene	13.01	228	16265	5.09	ppb		97
20) Indeno (1,2,3-cd) pyrene	16.09	276	18213	5.11	ppb	#	97
22) Benzo (b) fluoranthene	14.15	252	16376	4.81	ppb	#	95
23) Benzo (k) fluoranthene	14.17	252	15008	4.99	ppb	#	94
24) Benzo (a) pyrene	14.45	252	15227	4.68	ppb		97
25) Dibenz (a,h) anthracene	16.12	278	15351	4.91	ppb		98
26) Benzo (g,h,i) perylene	16.52	276	15430	4.84	ppb		98

(#) = qualifier out of range (m) = manual integration

Data File : M:\LINUS\DATA\L150515\0515L228.D Vial: 28
Acq On : 27 May 15 16:17 Operator: MA
Sample : 5.0ug/ml PAH 5/26/15 Inst : Linus
Misc : soil Multiplr: 1.00

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



EPA 8270C SIM
P0518-42

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No:
Date Analyzed: 1 Jun 15 12:25
Instrument: Linus
Initial Cal. Date: 5/15/2015
Data File: 0515L256.D

		Compound	MEAN	CCRF	%D	%Drift	
1	I	Napthalene-D8(IS)	ISTD			I	
2	SL	Surrogate Recovery (NBZ)	0.2737	0.2647	3.3	SL	15
3	TM	Naphthalene	0.9778	0.9851	0.74	TM	
4	TML	2-Methylnaphthalene	0.5974	0.6259	4.8	TML	12
5	TM	1-Methylnaphthalene	0.6165	0.6350	3.0	TM	
6	I	Acenaphthene-D10(IS)	ISTD			I	
7	S	Surrogate Recovery (FBP)	1.580	1.389	12	S	
8	TML	Acenaphthylene	2.081	2.233	7.3	TML	8.0
9	*TM	Acenaphthene	1.266	1.266	0.00	*TM	
10	TML	Fluorene	1.377	1.419	3.0	TML	11
11	I	Phenanthrene-D10(IS)	ISTD			I	
12	TML	Phenanthrene	1.151	1.138	1.1	TML	15
13	TML	Anthracene	1.084	1.128	4.1	TML	11
14	*TML	Fluoranthene	1.364	1.430	4.8	*TML	12
15	I	Chrysene-D12(IS)	ISTD			I	
16	TM	Pyrene	1.422	1.411	0.79	TM	
17	S	Surrogate Recovery (TPH)	0.8029	0.7782	3.1	S	
18	TML	Benz (a) anthracene	1.093	1.152	5.4	TML	11
19	TM	Chrysene	1.192	1.192	0.04	TM	
20	TML	Indeno (1,2,3-cd) pyrene	1.130	1.235	9.3	TML	6.7
21	I	Perylene-D12(IS)	ISTD			I	
22	TML	Benzo (b) fluoranthene	1.498	2.003	34	TML	1.9
23	TML	Benzo (k) fluoranthene	1.563	1.367	13	TML	20
24	*TML	Benzo (a) pyrene	1.618	1.625	0.42	*TML	14
25	TML	Dibenz (a,h) anthracene	1.446	1.579	9.2	TML	12
26	TML	Benzo (g,h,i) perylene	1.535	1.590	3.6	TML	14
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							
		Average			5.9		

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L256.D Vial: 56
Acq On : 1 Jun 15 12:25 Operator: MA
Sample : 5.0ug/ml PAH 06/01/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: Jun 1 12:45 2015 Quant Results File: P0518.RES

Quant Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270
Last Update : Mon May 18 09:20:47 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	8955	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4318	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	7734	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	8206	2.50	ppb		0.01
21) Perylene-D12(IS)	14.58	264	5431	2.50	ppb		0.00
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	4740m	4.26	ppb		0.01
Spiked Amount 2.500			Recovery	=	170.240%		
7) Surrogate Recovery (FBP)	7.41	172	11996	4.10	ppb		0.01
Spiked Amount 2.500			Recovery	=	163.960%		
17) Surrogate Recovery (TPH)	11.77	244	12772	4.85	ppb		0.00
Spiked Amount 2.500			Recovery	=	193.840%		
Target Compounds							Qvalue
3) Naphthalene	6.18	128	17643	5.04	ppb		98
4) 2-Methylnaphthalene	6.98	142	11209	4.38	ppb		93
5) 1-Methylnaphthalene	7.08	142	11373	5.15	ppb		91
8) Acenaphthylene	8.00	152	19280	4.60	ppb		99
9) Acenaphthene	8.21	154	10931	5.00	ppb		92
10) Fluorene	8.81	166	12251	4.45	ppb		100
12) Phenanthrene	9.92	178	17609	4.23	ppb		99
13) Anthracene	9.98	178	17452	4.46	ppb		100
14) Fluoranthene	11.30	202	22117	4.38	ppb	#	85
16) Pyrene	11.56	202	23154	4.96	ppb	#	83
18) Benz (a) anthracene	12.96	228	18907	4.46	ppb		97
19) Chrysene	13.00	228	19570	5.00	ppb		97
20) Indeno (1,2,3-cd) pyrene	16.08	276	20274	4.67	ppb	#	95
22) Benzo (b) fluoranthene	14.14	252	21758	5.10	ppb		99
23) Benzo (k) fluoranthene	14.17	252	14850m	4.00	ppb		98
24) Benzo (a) pyrene	14.45	252	17654	4.31	ppb		97
25) Dibenz (a,h) anthracene	16.11	278	17148	4.38	ppb		97
26) Benzo (g,h,i) perylene	16.51	276	17268	4.32	ppb		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L256.D Vial: 56
Acq On : 1 Jun 15 12:25 Operator: MA
Sample : 5.0ug/ml PAH 06/01/15 Inst : Linus
Misc : Multiplr: 1.00

Quant Time: Jun 1 12:45 2015 Quant Results File: P0518.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



EPA METHOD 8270D
Polynuclear Aromatic Hydrocarbons
Raw Data

APPL, INC.

Method Blank
EPA 8270D LL SOILS

Blank Name/QCG: **150521S-16137 - 197621**
Batch ID: #SIMDD-150521A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	5/21/2015	5/27/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	5/21/2015	5/27/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	5/21/2015	5/27/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	5/21/2015	5/27/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	5/21/2015	5/27/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	5/21/2015	5/27/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	63.5	45-105			%	5/21/2015	5/27/2015
BLANK	SURROGATE: NITROBENZENE-	44.2	35-100			%	5/21/2015	5/27/2015
BLANK	SURROGATE: TERPHENYL-D14 (	106	30-125			%	5/21/2015	5/27/2015

Quant Method: P0518-42.M
Run #: 0515L235
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/2/2015 2:14:33 PM

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L235.D Vial: 35
Acq On : 27 May 15 19:17 Operator: MA
Sample : 150521A BLK 1/30.00G Inst : Linus
Misc : soil Multiplr: 33.33

Quant Time: Jun 1 10:50 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8(IS)	6.16	136	8216	2.50	ppb	0.00
6) Acenaphthene-D10(IS)	8.17	164	3798	2.50	ppb	0.01
11) Phenanthrene-D10(IS)	9.91	188	6307	2.50	ppb	0.01
15) Chrysene-D12(IS)	12.97	240	6696	2.50	ppb	0.01
21) Perylene-D12(IS)	14.61	264	4479	2.50	ppb	0.02
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.37	82	993	36.82	ppb	0.01
Spiked Amount 83.333			Recovery = 44.186%			
7) Surrogate Recovery (FBP)	7.41	172	3807	52.88	ppb	0.01
Spiked Amount 83.333			Recovery = 63.461%			
17) Surrogate Recovery (TPH)	11.77	244	5720	88.66	ppb	0.00
Spiked Amount 83.333			Recovery = 106.397%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L235.D Vial: 35
Acq On : 27 May 15 19:17 Operator: MA
Sample : 150521A BLK 1/30.00G Inst : Linus
Misc : soil Multiplr: 33.33

Quant Time: Jun 1 10:50 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



Laboratory Control Spike Recovery
EPA 8270D LL SOILS

APPL ID: **150521S-16137 LCS - 197621**
Batch ID: #SIMDD-150521A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.133	0.0883	66.4	45-105
ACENAPHTHENE	0.133	0.127	95.5	45-110
ACENAPHTHYLENE	0.133	0.112	84.2	45-105
ANTHRACENE	0.133	0.126	94.7	55-105
BENZ (A) ANTHRACENE	0.133	0.139	105	50-110
BENZO (A) PYRENE	0.133	0.132	99.2	50-110
BENZO (B) FLUORANTHENE	0.133	0.140	105	45-115
BENZO (G,H,I) PERYLENE	0.133	0.135	102	40-125
BENZO (K) FLUORANTHENE	0.133	0.132	99.2	45-125
CHRYSENE	0.133	0.144	108	55-110
DIBENZ (A,H) ANTHRACENE	0.133	0.136	102	40-125
FLUORANTHENE	0.133	0.127	95.5	55-115
FLUORENE	0.133	0.120	90.2	50-110
INDENO (1,2,3-CD) PYRENE	0.133	0.145	109	40-120
NAPHTHALENE	0.133	0.0797	59.9	40-105
PHENANTHRENE	0.133	0.120	90.2	50-110
PYRENE	0.133	0.146	110	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0667	0.043	64.5	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0667	0.0313	46.9	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0667	0.0711	107	30-125

Comments: _____

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	5/21/2015
Analysis Date :	5/27/2015
Instrument :	Linus
Run :	0515L236
Initials :	DA

Printed: 6/2/2015 2:14:23 PM
APPL Standard LCS

Quantitation Report (Not Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L236.D Vial: 36
Acq On : 27 May 15 19:42 Operator: MA
Sample : 150521A LCS-1 1/30.01G Inst : Linus
Misc : soil Multiplr: 33.32

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	9702	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	4792	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	8492	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	8827	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	5818	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	965	31.31	ppb		0.01
Spiked Amount 83.306			Recovery =	37.581%			
7) Surrogate Recovery (FBP)	7.41	172	3908	43.01	ppb		0.01
Spiked Amount 83.306			Recovery =	51.632%			
17) Surrogate Recovery (TPH)	11.77	244	6049	71.10	ppb		0.00
Spiked Amount 83.306			Recovery =	85.353%			
Target Compounds							Qvalue
3) Naphthalene	6.18	128	9072	79.66	ppb		99
4) 2-Methylnaphthalene	6.98	142	7276	88.29	ppb		92
5) 1-Methylnaphthalene	7.08	142	7088	98.71	ppb		91
8) Acenaphthylene	8.00	152	15303	112.15	ppb		99
9) Acenaphthene	8.21	154	9254	127.10	ppb		93
10) Fluorene	8.81	166	10972	119.98	ppb		97
12) Phenanthrene	9.92	178	16547	120.29	ppb		98
13) Anthracene	9.98	178	16286	126.43	ppb		99
14) Fluoranthene	11.30	202	21238	127.19	ppb	#	89
16) Pyrene	11.56	202	21965	145.78	ppb	#	88
18) Benz (a) anthracene	12.96	228	19049	139.19	ppb		98
19) Chrysene	13.00	228	18206	144.16	ppb	#	94
20) Indeno (1,2,3-cd) pyrene	16.09	276	20309	145.32	ppb	#	99
22) Benzo (b) fluoranthene	14.15	252	19316	139.75	ppb	#	94
23) Benzo (k) fluoranthene	14.17	252	15753	131.93	ppb	#	95
24) Benzo (a) pyrene	14.45	252	17432	132.22	ppb		97
25) Dibenz (a,h) anthracene	16.12	278	17052	135.78	ppb		98
26) Benzo (g,h,i) perylene	16.52	276	17381	135.13	ppb		96

$\frac{9072 \times 8.5}{9702 \times 0.9778} \times 33.32 = 79.64$ DA 6/5/15

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L236.D Vial: 36
Acq On : 27 May 15 19:42 Operator: MA
Sample : 150521A LCS-1 1/30.01G Inst : Linus
Misc : soil Multiplr: 33.32

Quant Time: May 28 9:21 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



Method Blank

EPA 8270D LL SOILS RE-EXTRACT

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150601S-16139 - 197777**
Batch ID: #SIMDD-150601A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2-METHYLNAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	ACENAPHTHYLENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	ANTHRACENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	BENZ (A) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (A) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (B) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (G,H,I) PERYLENE	0.0017 U	0.005	0.0017	0.0013	mg/kg	6/1/2015	6/1/2015
BLANK	BENZO (K) FLUORANTHENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	CHRYSENE	0.0017 U	0.005	0.0017	0.0008	mg/kg	6/1/2015	6/1/2015
BLANK	DIBENZ (A,H) ANTHRACENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORANTHENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	FLUORENE	0.0017 U	0.005	0.0017	0.0010	mg/kg	6/1/2015	6/1/2015
BLANK	INDENO (1,2,3-CD) PYRENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	NAPHTHALENE	0.0017 U	0.005	0.0017	0.0009	mg/kg	6/1/2015	6/1/2015
BLANK	PHENANTHRENE	0.0017 U	0.005	0.0017	0.0011	mg/kg	6/1/2015	6/1/2015
BLANK	PYRENE	0.0017 U	0.005	0.0017	0.0012	mg/kg	6/1/2015	6/1/2015
BLANK	SURROGATE: 2-FLUORBIPHENY	77.1	45-105			%	6/1/2015	6/1/2015
BLANK	SURROGATE: NITROBENZENE-	68.4	35-100			%	6/1/2015	6/1/2015
BLANK	SURROGATE: TERPHENYL-D14	112	30-125			%	6/1/2015	6/1/2015

Quant Method: P0518-42.M
Run #: 0515L262
Instrument: Linus
Sequence: L150515
Initials: DA

GC SC-Blank-REG MDLs
Printed: 6/5/2015 2:20:36 PM

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L262.D Vial: 62
Acq On : 1 Jun 15 16:00 Operator: MA
Sample : 150601A BLK 1/30.75G Inst : Linus
Misc : soil Multiplr: 32.52

Quant Time: Jun 5 13:31 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8(IS)	6.16	136	7710	2.50	ppb	0.00
6) Acenaphthene-D10(IS)	8.17	164	3775	2.50	ppb	0.01
11) Phenanthrene-D10(IS)	9.90	188	6391	2.50	ppb	0.00
15) Chrysene-D12(IS)	12.97	240	6335	2.50	ppb	0.01
21) Perylene-D12(IS)	14.59	264	3876	2.50	ppb	0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.37	82	1537	55.62	ppb	0.01
Spiked Amount 81.301			Recovery = 68.413%			
7) Surrogate Recovery (FBP)	7.41	172	4599	62.71	ppb	0.01
Spiked Amount 81.301			Recovery = 77.131%			
17) Surrogate Recovery (TPH)	11.77	244	5700	91.11	ppb	0.00
Spiked Amount 81.301			Recovery = 112.067%			

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L262.D Vial: 62
Acq On : 1 Jun 15 16:00 Operator: MA
Sample : 150601A BLK 1/30.75G Inst : Linus
Misc : soil Multiplr: 32.52

Quant Time: Jun 5 13:31 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



Laboratory Control Spike Recovery
EPA 8270D LL SOILS RE-EXTRACT

APPL ID: **150601S-16139 LCS - 197777**
Batch ID: #SIMDD-150601A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2-METHYLNAPHTHALENE	0.167	0.156	93.4	45-105
ACENAPHTHENE	0.167	0.178	107	45-110
ACENAPHTHYLENE	0.167	0.164	98.2	45-105
ANTHRACENE	0.167	0.170	102	55-105
BENZ (A) ANTHRACENE	0.167	0.192	115 #	50-110
BENZO (A) PYRENE	0.167	0.199	119 #	50-110
BENZO (B) FLUORANTHENE	0.167	0.205	123 #	45-115
BENZO (G,H,I) PERYLENE	0.167	0.194	116	40-125
BENZO (K) FLUORANTHENE	0.167	0.198	119	45-125
CHRYSENE	0.167	0.199	119 #	55-110
DIBENZ (A,H) ANTHRACENE	0.167	0.195	117	40-125
FLUORANTHENE	0.167	0.174	104	55-115
FLUORENE	0.167	0.165	98.8	50-110
INDENO (1,2,3-CD) PYRENE	0.167	0.189	113	40-120
NAPHTHALENE	0.167	0.179	107 #	40-105
PHENANTHRENE	0.167	0.165	98.8	50-110
PYRENE	0.167	0.207	124	45-125
SURROGATE: 2-FLUORBIPHENYL (S)	0.0828	0.0665	80.3	45-105
SURROGATE: NITROBENZENE-D5 (S)	0.0828	0.0577	69.7	35-100
SURROGATE: TERPHENYL-D14 (S)	0.0828	0.0985	119	30-125

= Recovery is outside QC limits.

Comments:

Primary	SPK
Quant Method :	P0518-42.M
Extraction Date :	6/1/2015
Analysis Date :	6/1/2015
Instrument :	Linus
Run :	0515L263
Initials :	DA

Printed: 6/5/2015 2:20:19 PM
APPL Standard LCS

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L150515\0515L263.D Vial: 63
Acq On : 1 Jun 15 16:25 Operator: MA
Sample : 150601A LCS-1 1/30.35G Inst : Linus
Misc : soil Multiplr: 32.95

Quant Time: Jun 2 8:44 2015 Quant Results File: P0518-42.RES

Quant Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration
DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units		Dev(Min)
1) Napthalene-D8(IS)	6.16	136	7373	2.50	ppb		0.00
6) Acenaphthene-D10(IS)	8.17	164	3574	2.50	ppb		0.01
11) Phenanthrene-D10(IS)	9.90	188	5999	2.50	ppb		0.00
15) Chrysene-D12(IS)	12.97	240	6033	2.50	ppb		0.01
21) Perylene-D12(IS)	14.59	264	3629	2.50	ppb		0.01
System Monitoring Compounds							
2) Surrogate Recovery (NBZ)	5.37	82	1510	57.74	ppb		0.01
Spiked Amount 82.372			Recovery = 70.094%				
7) Surrogate Recovery (FBP)	7.41	172	4560	66.54	ppb		0.01
Spiked Amount 82.372			Recovery = 80.777%				
17) Surrogate Recovery (TPH)	11.77	244	5792	98.50	ppb		0.00
Spiked Amount 82.372			Recovery = 119.577%				
Target Compounds							Qvalue
3) Naphthalene	6.18	128	15630	178.58	ppb		98
4) 2-Methylnaphthalene	6.98	142	9973	155.60	ppb		93
5) 1-Methylnaphthalene	7.08	142	10150	183.93	ppb		94
8) Acenaphthylene	8.00	152	17360	164.20	ppb		99
9) Acenaphthene	8.21	154	9786	178.20	ppb		93
10) Fluorene	8.81	166	11410	164.56	ppb		98
12) Phenanthrene	9.92	178	16192	165.39	ppb		99
13) Anthracene	9.98	178	15654	169.82	ppb		99
14) Fluoranthene	11.30	202	20572	173.86	ppb	#	88
16) Pyrene	11.56	202	21598	207.38	ppb	#	87
18) Benz (a) anthracene	12.96	228	18204	191.88	ppb		97
19) Chrysene	13.00	228	17336	198.59	ppb	#	94
20) Indeno (1,2,3-cd) pyrene	16.09	276	18506	188.96	ppb	#	93
22) Benzo (b) fluoranthene	14.15	252	17674	205.36	ppb	#	94
23) Benzo (k) fluoranthene	14.17	252	15367	198.13	ppb	#	94
24) Benzo (a) pyrene	14.45	252	16420	198.55	ppb		96
25) Dibenz (a,h) anthracene	16.12	278	15579	195.02	ppb		96
26) Benzo (g,h,i) perylene	16.52	276	15743	193.83	ppb		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : M:\LINUS\DATA\L150515\0515L263.D Vial: 63
Acq On : 1 Jun 15 16:25 Operator: MA
Sample : 150601A LCS-1 1/30.35G Inst : Linus
Misc : soil Multiplr: 32.95

Quant Time: Jun 2 8:44 2015 Quant Results File: P0518-42.RES

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270
Last Update : Wed May 27 13:24:37 2015
Response via : Initial Calibration



DFTPP

Data File : M:\LINUS\DATA\L150515\0515L002.D Vial: 2
Acq On : 15 May 15 16:56 Operator: csc
Sample : SV Tune 05/14/15 Inst : Linus
Misc : Multiplr: 1.00

Method : M:\LINUS\DATA\L150515\P0518.M (RTE Integrator)
Title : EPA 8270



AutoFind: Scans 475, 476, 477; Background Corrected with Scan 472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	36.1	160774	PASS
68	69	0.00	2	1.4	1998	PASS
70	69	0.00	2	0.5	725	PASS
127	198	40	60	46.9	208571	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	444800	PASS
199	198	5	9	6.8	30375	PASS
275	198	10	30	20.7	91932	PASS
365	198	1	100	2.0	9111	PASS
441	443	0.01	100	76.6	45477	PASS
442	198	40	150	69.5	309146	PASS
443	442	17	23	19.2	59336	PASS

DFTPP

Data File : M:\LINUS\DATA\L150515\0515L227.D
Acq On : 27 May 15 15:59
Sample : SV TUNE 5/14/15
Misc : soil

Vial: 27
Operator: MA
Inst : Linus
Multiplr: 1.00

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270



AutoFind: Scans 475, 476, 477; Background Corrected with Scan 472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	35.0	120837	PASS
68	69	0.00	2	1.5	1553	PASS
70	69	0.00	2	0.6	627	PASS
127	198	40	60	45.4	156717	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	344816	PASS
199	198	5	9	6.6	22697	PASS
275	198	10	30	22.2	76418	PASS
365	198	1	100	2.1	7407	PASS
441	443	0.01	100	77.8	40627	PASS
442	198	40	150	78.3	270147	PASS
443	442	17	23	19.3	52208	PASS

DFTPP

Data File : M:\LINUS\DATA\L150515\0515L255.D Vial: 55
Acq On : 1 Jun 15 9:58 Operator: MA
Sample : SV TUNE 5/14/15 Inst : Linus
Misc : Multiplr: 1.00

Method : M:\LINUS\DATA\L150515\P0518-42.M (RTE Integrator)
Title : EPA 8270



AutoFind: Scans 475, 476, 477; Background Corrected with Scan 472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	34.0	119764	PASS
68	69	0.00	2	1.3	1294	PASS
70	69	0.00	2	0.7	699	PASS
127	198	40	60	44.0	155296	PASS
197	198	0.00	1	0.0	106	PASS
198	198	100	100	100.0	352747	PASS
199	198	5	9	6.6	23436	PASS
275	198	10	30	21.8	76998	PASS
365	198	1	100	2.0	7131	PASS
441	443	0.01	100	76.8	40334	PASS
442	198	40	150	76.4	269666	PASS
443	442	17	23	19.5	52530	PASS

0515L255.D P0518-42.M Fri Jun 05 13:41:58 2015

INDOXACARB CURVE											
Prep: 5/14/15 Ex: 7/31/15-L.H.											
Supplier	**ID#**	**ug/mL**	**LOT#**	**DATE**	**EXP.DATE**	**µL**	**µL**	**µL**	**µL**	**µL**	**µL**
INDOXACARB STD.		1		07/31/14	07/31/15	5	50	100	150	200	250
						995	950	900	850	800	750
	Hexane		051914A		**FINAL VOL**	1000	1000	1000	1000	1000	1000

8270 Full Scan Second Source Stock						
	5/14/15-L.H.					
8270 Second Source Stock						
Exp:	7/21/15					
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	116070-02	2000	34696-3398	5/14/15	5/14/16	1000
o2si	110391-01	2000	34693-3401	5/14/15	5/14/16	1000
o2si	010337-01	1000	17950-3514	5/14/15	11/14/15	1000
o2si	110396-01	2000	34698-3515	5/14/15	5/14/16	1000
o2si	110397-01	2000	50011-3505	5/14/15	7/21/15	1000
o2si	110393-01	2000	34694-3515	5/14/15	5/14/16	1000
o2si	110394-01	2000	35110-3515	5/14/15	5/14/16	1000
o2si	110395-01	2000	34697-3402	5/14/15	5/14/16	1000
o2si	110392-01	2000	18075-3397	5/14/15	5/14/16	1000
EM Science	Methylene Chloride		55006	11/06/14		1000
				Final Vol.		10000

Semivolatile (SV) Tuning Solution -L.H.						
PREP DATE:	5/14/15					
Semivolatile (SV) Tune Solution 50ug/ml						
Exp:	5/14/16					
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Ultra Scient	Semivolatile	1000	CL-5063-3518	05/14/15	5/14/16	1000
EM Science	Methylene Chloride		47186	11/06/14		19000
				Final Vol.		20000

8270 PAH SIM Standard Curve

PREP DATE:	05/15/15-L.H.												
8270 PAH SIM STANDARD CURVE													
Exp.DATE:	05/22/15					0.1	0.2	0.5	1	5	10	50	100
		Conc.		Date									
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL	µL	µL	µL	µL	µL	µL	µL
o2si	8270 PAH SIM Stock	200	230271-34569	02/27/15	06/08/16	0	0	0	0	5	5	25	50
	5.0ug/mL	5		05/15/15	05/22/15	0	0	10	20	0	0	0	0
	1.0ug/mL	1		05/15/15	05/22/15	10	20	0	0	0	0	0	0
o2si	8270 BN:A (200:400	VAR	242352-34963	05/04/15	01/01/17	0	0	0	0	5	5	25	50
EM Science	Methylene Chloride		54200	11/06/14		90	80	90	80	190	90	50	0
Supelco	SIM SV Internal St	125	A10079V-3449	05/15/15	5/15/16	2	2	2	2	4	2	2	2
				Final Vol.		102	102	102	102	204	102	102	102

8270 PAH SIM SS

PREP DATE:	05/15/15-L.H.					
8270 PAH SIM Second Source 5µg/mL						
Exp.DATE:	05/15/15					5
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	8270 PAH SIM Stock	200	243252-34609	02/27/15	01/15/17	5
EM Science	Methylene Chloride		54200	11/06/14		195
Supelco	SIM SV Internal St	125	A10079V-3449	05/15/15	5/15/16	4
				Final Vol.		204

8270 PAH SIM CCV

PREP DATE:	05/15/15					
8270 PAH SIM Continuing Calibration Verification 5ug/ml						
Exp.DATE:	05/22/15					5
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	8270 PAH SIM Stock	200	230271-34569	02/27/15	06/08/16	5
o2si	8270 BN:A (200:400	VAR	242352-34963	05/04/15	01/01/17	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SIM SV Internal St	125	A10079V-3449	05/15/15	5/15/16	4
				Final Vol.		204

072

02/24/15
CSC

GC/MS STANDARD PREPARATION BOOK # 1 PAGE

PREP DATE:	02/24/15					MU
8270 PAH SIM Continuing Calibration Verification 5ug/ml						
Exp.DATE:	02/31/15					5
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	8270 PAH SIM Stock	10	213931-33759	12/23/14	08/26/15	5
o2si	8270 BN:A Surrogate	VAR	226244-34103	12/23/14	04/10/16	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SIM Internal Standard	125	XA10079V-34507	12/23/14	09/30/17	4
				Final Vol.		204

02/27/15
CSC

8270D PAH SIM Solution, 200 mg/L, 1 ml
110730-01
Lot # 230271 Storage ≤-10°C Expiry 8/Jun/2016
Solv: Methylene Chloride
Made in the USA

8270D PAH SIM
Lot #: 230271 - 34569
Rec: 1/13/15 MFR exp. 6/8/16

Recieved: 01/13/15
Opened: 02/27/15
Expired: 06/8/16

02/27/15
CSC

8270 BN:A (200:400) Surrogate Solution, 1 mL
110004-17
Lot # 237956 Storage ≤-10 °C Expiry 8/Oct/2016
Solv: Methylene Chloride
Made in the USA

8270 BN:A (200:400) Surrogate Solution
Lot #: 237956 - 34559
Rec: 1/13/15 MFR exp. 10/8/16

Recieved: 01/13/15
Opened: 02/27/15
Expired: 10/8/16

NOTEBOOK INSERT LABEL

Semivolatile Internal Std Mix **CRM48902**
Lot: XA10079V EXP: SEP/2017 STORAGE: REFRIGERATE 2 x 1ml

DATE RECEIVED:_____

SUPELCO
Solutions within.
595 North Harrison Road • Bellefonte, PA
16823-0048 USA • Phone 814-359-3441

STANDARD TRANSFER LABEL

Date of Preparation: **Exp. Date:**
Reference Number: **Storage:** EXP: SEP/2017
Description: REFRIGERATE
CHEMST

02/27/15
CSC

Semivolatile Internal Standard Mix
Lot #: XA10079V - 34496
Rec: 12/11/14 MFR exp. 9/30/17

CRM48902 XA10079V
Semivolatile Internal Std Mix
2000ug/ml in dichloromethane

Recieved: 12/11/14
Opened: 02/27/15
Expired: 09/30/17

8270 SIM PAH Internal Standard

PREP DATE:	5/15/15-L.H.					
8270 SIM PAH Internal Standard Mix 125ug/ml						
Exp.DATE:	11/15/15					
		Conc.		Date	Exp.	125
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
Supelco	SV Internal Standa	2000	A10079V-3449	05/15/15	5/15/16	50
EM Science	Methylene Chloride		47186	11/06/14		750
				Final Vol.		800

8270 Full Scan CCV

PREP DATE:	05/15/15-L.H.					
8270 Continuing Calibration Verification 50ug/ml						
Exp:	05/22/15					50
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Absolute	8270 Stock	200	VAR	11/18/14	05/21/15	25
o2si	8270 BN:A (200:400	VAR	242352-34963	02/27/15	10/8/16	25
EM Science	Methylene Chloride		47186	11/06/14		50
Supelco	SV Internal Standa	2000	A10079V-3449	02/27/15	09/30/17	2

1ug/ml 1,2,3-TCP

PREP DATE:	05/15/15-LH					
1,2,3-TCP 1ug/ml						
Exp. Date:	11/04/15					
		Conc.		Date	Exp.	
Supplier	ID#	µg/mL	Lot #	Code	Date	µL
O2si	1,2,3-TCP	1000	02872-3201	05/15/15-LH	01/23/16	5
J.T.Baker	Purge & Trap MeoH		0000062345	10/10/13		4995

1,2,3-TCP CCV

PREP DATE:	5/15/15-L.H.					
1,2,3-TCP Continuing Calibration Verification 0.05ug/mL						
Exp. Date:	11/4/15					
		Conc.		Date	Exp.	
Supplier	ID#	µg/mL	Lot #	Code	Date	µL
O2si	1,2,3-TCP	1.0	02641-3200	5/4/15	01/23/16	50
O2si	1,2,3-TCP-d5	2.0	02640-3200	5/4/15	01/27/16	25
J.T.Baker	Purge & Trap MeoH		0000062345	10/10/13		925
				Final Vol.		1000

GC/MS STANDARD PREPARATION BOOK # ___ PAGE # ___

073

02/27/15 CSC

8270D PAH SIM Solution, Second Source, 200 mg/L, 1 mL
110780-01-SS
Lot # 243253 | Storage ≤-10 °C | Expiry 15/Jan/2017
Solv: Methylene Chloride

8270D PAH SIM (SS)
Lot #: 243252 - 34609
Rec: 1/23/15 MFR exp. 1/15/17

Received: 01/23/15
Opened: 02/27/15
Expired: 01/15/17

02/27/15 CSC

Part #: 30031 Laboratory Use Only - See MSDS
Lot #: 060111 Exp: 060116 Storage 20 °C
EPA Method 525 - Internal Standard
4 components CAUTION: Sonicate Before Use
500 ug/mL in acetc
ABSOLUTE STANDA

EPA Method 525 Internal Std
Lot #: 060111 - 32256
Rec: 4/16/13 MFR exp. 6/1/16

Received: 04/16/13
opened: 02/27/15
Expired: 06/01/16

02/27/15 CSC

PREP DATE:	02/27/15						
8270 SIM PAH Internal Standard Mix 125ug/ml							
Exp. DATE:	08/27/15						
		Conc.		Date	Exp.	125	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL	
Supelco	Internal Standard	2000	XA10079V-34496	12/11/14	09/30/17	50	
EM Science	Methylene Chloride		54200	11/06/14		750	
				Final Vol.		800	

02/27/15 CSC

PREP DATE:	02/27/15						
525 Working Standard 20ug/mL (PCP @ 80ug/ml)							
Exp. DATE:	03/04/15					20	
		Conc.		Date	Exp.		
Supplier	ID #	ug/ml	Lot #	Code	Date	ul	
Ultra-Scientific	Extractables Mixture (PSM-525A)	500	CK-2769-34658	02/17/15	09/30/16	24	
Ultra-Scientific	PAH Mixture (PM-525B)	500	CK-2706-34660	02/17/15	09/30/16	24	
EM Science	Methylene Chloride		54200	11/06/14		552	
				Final Vol.		600	

02/27/15 CSC

PREP DATE:	02/27/15						
525 Continuing Calibration Verification 5ug/ml (20ppb PC							
Exp. DATE:	03/04/15					5	
		Conc.		Date	Exp.		
Supplier	ID #	µg/mL	Lot #	Code	Date	µL	
Ultra Scienti	525 Working Std (PSM-525A) & (PM-525B)	20	VAR	02/17/15	09/30/16	50	
Absolute	525 Internal Std (30031)	500	060111-32256	04/16/13	06/01/16	2	
EM Science	Methylene Chloride		54200	11/06/14		148	
				Final Vol.		200	

02/26/15 CSC

PREP DATE:	02/26/15		EXP:	03/03/15										
Sulfolane Standard Curve														
						1	5	10	20	40	50	60	80	100
		Conc.		Date	Exp.									
Supplier	ID #	µg/mL	Lot#	Code	Date	µL	µL	µL	µL	µL	µL	µL	µL	µL
o2si	Sulfolane Calibration Standard	200	VAR	10/22/14	03/11/15	5	5	5	10	20	25	30	40	50
EM Science	Methylene Chloride		54200	11/06/14		995	195	95	90	80	75	70	60	50
				Final Vol.		1000	200	100	100	100	100	100	100	100

8270 PAH SIM CCV

PREP DATE:	05/26/15-LH					
8270 PAH SIM Continuing Calibration Verification 5ug/ml						
Exp.DATE:	06/02/15					5
		Conc.		Date	Exp.	
Supplier	ID #	µg/mL	Lot #	Code	Date	µL
o2si	8270 PAH SIM Stock	200	230271-34569	02/27/15	06/08/16	5
o2si	8270 BN:A (200:400)S	VAR	242352-34963	02/27/15	10/08/16	5
EM Science	Methylene Chloride		54200	11/06/14		190
Supelco	SIM SV Internal Star	125	A10079V-3449	05/15/15	5/15/16	4
				Final Vol.		204

8270 Full Scan CCV

PREP DATE:	06/02/15-L.H.					
8270 Continuing Calibration Verification 50ug/ml						
Exp:	05/22/15					50
		Conc.		Date		
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date	µL
Absolute	8270 Stock	200	VAR	04/27/15	10/27/15	25
o2si	8270 BN:A (200:400)	VAR	242352-34963	05/21/15	5/21/16	25
EM Science	Methylene Chloride		47186	11/06/14		50
Supelco	SV Internal Standard	2000	A10079V-3449	05/21/15	5/21/16	2

Prep: 05/26/15 Exp: 04/27/15 GA **HERB SECOND SOURCE**

	Conc. in mix	Conc. In Stock	Aliqout	OCK SOUIR	Final Vol. Solvent
Analytes:	**(ug/ml)**	**(ug/ml)**	**(uL)**	**O2SI**	**(ml)**
Dalapon	12.8	64	1000	AT: 132070-0	5
3,5 Dichlorobenzoic Acid	1.28	6.4		Lot: 207738	MTBE
4-Nitrophenol	3.2	16		OP: 5/26/15	Lot #
2,4-DCAA (S)	6.4	32		Exp: 4/27/16	50223
Dicamba	1.28	6.4			
MCPP	1280	6400			
MCPA	1280	6400			
2,4-DP	6.4	32			
2,4-D	6.4	32			
DNOC	2.56	12.8			
PCP	1.28	6.4			
2,4,5-TP	1.28	6.4			
Chloramben	6.4	32			
2,4,5-T	1.28	6.4			
Dinoseb	3.2	16			
2,4-DB	12.8	64			
Bentazon	6.4	32			
Picloram	1.28	6.4			
Dacthal	1.28	6.4			
Acifluorfen	3.2	16			

***1 ml ampule is derivitized by extraction and brought up to 5ml in MTBE*

_SON009S

Organic Extraction Worksheet

Method	SIM Sonicat Ext. Methylene c 3550	Extraction Set	150601A	Extraction Method	SON009S	Units	mL

Spiked ID 1	SIM Spike 230271-33745		Surrogate ID 1	SIM Surrogate 248398-35028	
Spiked ID 2			Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	NO	
Spiked ID 7			Ext. Start Time:	06/01/15 13:35	
Spiked ID 8			Ext. End Time:	06/01/15 15:45	
			GC Requires Extract By:	**06/02/15 12:00**	

pH1			Water Bath Temp Criteria	78 °C
pH2				
pH3				

Spiked By: DL **Date** 06/01/15 **Witnessed By:** CFM **Date** 06/01/15

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	150601A Blk				0.0250	1	30.75g	1	NA	06/01/15 13:35	
						equip	e-s5 E-WB7				
2	150601A LCS-1		0.0250	1	0.0250	1	30.35g	1	NA	06/01/15 13:35	
						equip	E-S4 E-WB7				
3	AZ16139	AZ16139S02			0.0250	1	30.21g	1	NA	06/01/15 13:35	76380
						equip	E-S3.1 E-WB7				
4	AZ16202	AZ16202S02			0.0250	1	30.16g	1	NA	06/01/15 13:35	76390
						equip	E-S2 E-WB7				

KY 6/01/15

lvent and Lot#	
AND	WF27C
C	55006
lter Paper	9624720
Na2SO4	XE21E

Extraction COC Transfer	
Extraction lab employee Initials	KY
GC analyst's initials	LH
Date	6/1/15
Time	4:00
Refrigerator	GC-C

	Technician's Initials
Scanned By	DL
Sample Preparation	DL
Extraction	DL
Concentration	DL

Modified	06/01/15 3:51:26 PM

Reviewed By: KY 217 **Date** 6/01/15

/01/15 3:52:09 PM Ext_ID 47991

_SON009S

Organic Extraction Worksheet

Method	SIM Sonicat Ext. Methylene c 3550	Extraction Set	150521A	Extraction Method	SON009S	Units	mL

Spiked ID 1	SIM Spike 230271-33745		Surrogate ID 1	SIM Surrogate 248398-35028	
Spiked ID 2			Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	NO	
Spiked ID 7			Ext. Start Time:	05/21/15 8:15	
Spiked ID 8			Ext. End Time:	05/21/15 15:30	
			GC Requires Extract By:	**05/29/15 0:00**	

pH1			Water Bath Temp Criteria	81,80 °C
pH2				
pH3				

Spiked By: IC **Date** 05/21/15 **Witnessed By:** DL **Date** 05/21/15

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	150521A Blk				0.0250	1 equip	30.00g	1	NA	05/21/15 8:15	
							E-S5 E-WB7				
2	150521A LCS-1		0.0250	1	0.0250	1 equip	30.01g	1	NA	05/21/15 8:15	
							E-S4 E-WB7				
3	AZ16137	AZ16137S01			0.0250	1 equip	30.02g	1	NA	05/21/15 8:15	76380
							E-S3.1 E-WB7				
4	AZ16138	AZ16138S01			0.0250	1 equip	30.25g	1	NA	05/21/15 8:15	76380
							E-S2 E-WB7				
5	AZ16139	AZ16139S01			0.0250	1 equip	30.07g	1	NA	05/21/15 8:15	76380
							E-S1.2 E-WB7				
6	AZ16140	AZ16140S01			0.0250	1 equip	30.07g	1	NA	05/21/15 8:15	76380
							E-S1.1 E-WB7				
7	AZ16141	AZ16141S01			0.0250	1 equip	30.23g	1	NA	05/21/15 8:15	76380
							E-S6 E-WB7				
8	AZ16198	AZ16198S02			0.0250	1 equip	30.19g	1	NA	05/21/15 8:15	76390
							E-S7 E-WB7				
9	AZ16199	AZ16199S01			0.0250	1 equip	30.11g	1	NA	05/21/15 8:15	76390
							E-S5 E-WB6				
10	AZ16201	AZ16201S02			0.0250	1 equip	30.11g	1	NA	05/21/15 8:15	76390
							E-S4 E-WB6				
11	AZ16202	AZ16202S01			0.0250	1 equip	30.02g	1	NA	05/21/15 8:15	76390
							E-S3.1 E-WB6				
12	AZ16400	AZ16400S02			0.0250	1 equip	30.13g	1	NA	05/21/15 8:15	76447
							E-S2 E-WB6				
13	AZ16401	AZ16401S01			0.0250	1 equip	30.09g	1	NA	05/21/15 8:15	76447
							E-S1.2 E-WB6				
14	AZ16402	AZ16402S01			0.0250	1 equip	30.11g	1	NA	05/21/15 8:15	76447
							E-S1.1 E-WB6				

KY 5/21/15

Solvent and Lot#	
SAND	WF27C
MC	55006
Filter Paper	9624720
B.Na2SO4	XE21E

Extraction COC Transfer	
Extraction lab employee Initials	KY
GC analyst's initials	GA
Date	5/21/15
Time	1600
Refrigerator	Hobart

	Technician's Initials
Scanned By	KY
Sample Preparation	KY
Extraction	IC
Concentration	IC

Modified	05/21/15 3:43:46 PM

Reviewed By: KY **Date** 5/21/15

Injection Log

Directory: M:\LINUS\DATA\L150515

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	2	0515L002.D	1	SV Tune 05/14/15		05/15/2015 16:56
2	3	0515L003.D	1	0.1ug/ml PAH 05/15/15		05/15/2015 17:15
3	4	0515L004.D	1	0.2ug/ml PAH		05/15/2015 17:41
4	5	0515L005.D	1	0.5ug/ml PAH		05/15/2015 18:06
5	6	0515L006.D	1	1.0ug/ml PAH		05/15/2015 18:32
6	7	0515L007.D	1	5.0ug/ml PAH		05/15/2015 18:58
7	8	0515L008.D	1	10ug/ml PAH		05/15/2015 19:23
8	9	0515L009.D	1	50ug/ml PAH		05/15/2015 19:49
9	10	0515L010.D	1	100ug/ml PAH		05/15/2015 20:14
10	11	0515L011.D	1	5.0ug/ml PAH SS 05/15/15		05/15/2015 20:40
11	27	0515L227.D	1	SV TUNE 5/14/15	soil	05/27/2015 15:59
12	28	0515L228.D	1	5.0ug/ml PAH 5/26/15	soil	05/27/2015 16:17
13	35	0515L235.D	33.33333	150521A BLK 1/30.00G	soil	05/27/2015 19:17
14	36	0515L236.D	33.32223	150521A LCS-1 1/30.01G	soil	05/27/2015 19:42
15	37	0515L237.D	333.11126	AZ16137S01 1/30.02G df10	soil	05/27/2015 20:08
16	38	0515L238.D	330.57851	AZ16138S01 1/30.25G df10	soil	05/27/2015 20:33
17	39	0515L239.D	33.25574	AZ16139S01 1/30.07G	soil	05/27/2015 20:59
18	40	0515L240.D	332.55737	AZ16140S01 1/30.07G df10	soil	05/27/2015 21:24
19	41	0515L241.D	33.07972	AZ16141S01 1/30.23G	soil	05/27/2015 21:50
20	55	0515L255.D	1	SV TUNE 5/14/15		06/01/2015 09:58
21	56	0515L256.D	1	5.0ug/ml PAH 06/01/15		06/01/2015 12:25
22	62	0515L262.D	32.5203	150601A BLK 1/30.75G	soil	06/01/2015 16:00
23	63	0515L263.D	32.9489	150601A LCS-1 1/30.35G	soil	06/01/2015 16:25
24	64	0515L264.D	33.1016	AZ16139S02 1/30.21G	soil	06/01/2015 16:51

Injection Log

Directory: M:\LINUS\DATA\L150515

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	2	0515L002.D	1	SV Tune 05/14/15		05/15/2015 16:56
2	3	0515L003.D	1	0.1ug/ml PAH 05/15/15		05/15/2015 17:15
3	4	0515L004.D	1	0.2ug/ml PAH		05/15/2015 17:41
4	5	0515L005.D	1	0.5ug/ml PAH		05/15/2015 18:06
5	6	0515L006.D	1	1.0ug/ml PAH		05/15/2015 18:32
6	7	0515L007.D	1	5.0ug/ml PAH		05/15/2015 18:58
7	8	0515L008.D	1	10ug/ml PAH		05/15/2015 19:23
8	9	0515L009.D	1	50ug/ml PAH		05/15/2015 19:49
9	10	0515L010.D	1	100ug/ml PAH		05/15/2015 20:14
10	11	0515L011.D	1	5.0ug/ml PAH SS 05/15/15		05/15/2015 20:40
11	12	0515L012.D	1	SV Tune 05/15/15		05/15/2015 21:02
12	27	0515L227.D	1	SV TUNE 5/14/15	soil	05/27/2015 15:59
13	28	0515L228.D	1	5.0ug/ml PAH 5/26/15	soil	05/27/2015 16:17
14	35	0515L235.D	33.33333	150521A BLK 1/30.00G	soil	05/27/2015 19:17
15	36	0515L236.D	33.32223	150521A LCS-1 1/30.01G	soil	05/27/2015 19:42
16	37	0515L237.D	333.11126	AZ16137S01 1/30.02G df10	soil	05/27/2015 20:08
17	38	0515L238.D	330.57851	AZ16138S01 1/30.25G df10	soil	05/27/2015 20:33
18	39	0515L239.D	33.25574	AZ16139S01 1/30.07G	soil	05/27/2015 20:59
19	40	0515L240.D	332.55737	AZ16140S01 1/30.07G df10	soil	05/27/2015 21:24
20	41	0515L241.D	33.07972	AZ16141S01 1/30.23G	soil	05/27/2015 21:50
21	55	0515L255.D	1	SV TUNE 5/14/15		06/01/2015 09:58
22	56	0515L256.D	1	5.0ug/ml PAH 06/01/15		06/01/2015 12:25
23	62	0515L262.D	32.5203	150601A BLK 1/30.75G	soil	06/01/2015 16:00
24	63	0515L263.D	32.9489	150601A LCS-1 1/30.35G	soil	06/01/2015 16:25
25	64	0515L264.D	33.1016	AZ16139S02 1/30.21G	soil	06/01/2015 16:51

EPA METHOD 6850
Perchlorate
LC/MS

APPL, INC.

EPA METHOD 6850
Perchlorate
LC/MS

QC Summary

APPL, INC.

Method Blank
PERCHLORATE EPA 6850 - SOIL

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Blank Name/QCG: **150518S-16137 - 198067**
Batch ID: #6850SM-150518A

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method: QTLMFL2
Run #: T523_049.d
Instrument: AGIL_6460
Sequence: TQ052315
Initials: RP

GC SC-Blank-REG MDLs
Printed: 06/17/15 10:45:58 AM

Laboratory Control Spike Recovery
PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 LCS - 198067**
Batch ID: #6850SM-150518A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits
PERCHLORATE	0.00594	0.00556	93.6	80-120

Comments: _____

Primary	SPK
Quant Method :	QTLMFL2
Extraction Date :	05/18/15
Analysis Date :	05/24/15
Instrument :	AGIL_6460
Run :	T523_047.d
Initials :	RP

Printed: 06/17/15 10:45:41 AM
APPL Standard LCS

Matrix Spike Recoveries

PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 MS - 198067**
Batch ID: #6850SM-150518A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Lvl mg/Kg	Matrix Result mg/Kg	SPK Result mg/Kg	DUP Result mg/Kg	SPK % Recovery	DUP % Recovery	Recovery Limits	RPD %	RPD Limits
PERCHLORATE	0.00508	ND	0.00496	0.00529	97.6	104	80-120	6.4	15

Comments: _____

Primary	SPK	DUP
Quant Method :	QTLMFL2	QTLMFL2
Extraction Date :	05/18/15	05/18/15
Analysis Date :	05/24/15	05/24/15
Instrument :	AGIL_6460	AGIL_6460
Run :	T523_051.d	T523_052.d
Initials :	RP	

Printed: 06/17/15 10:53:00 AM
APPL MSD SCII

EPA 6850

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/24/15

Matrix: SOIL Instrument: AGIL_6460

Blank ID: 150518A-BLK Time Analyzed: 0628

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150518A-LCS	Lab Control Spike	T523_047.d	05/24/15 0550
150518A-BLK	Blank	T523_049.d	05/24/15 0628
AZ16137	S67-SS030-0001	T523_050.d	05/24/15 0646
150518A-MS	Matrix Spike	T523_051.d	05/24/15 0705
150518A-MSD	Matrix SpikeD	T523_052.d	05/24/15 0724
AZ16138	S67-SS030P-0001	T523_053.d	05/24/15 0742
AZ16139	S67-SB030-0607	T523_054.d	05/24/15 0801
AZ16140	S67-SS032-0001	T523_055.d	05/24/15 0819
AZ16141	S67-SB032-0607	T523_060.d	05/24/15 0953

Comments: Batch: #6850SM-150518A

Interferrence Check Sample

Perchlorate EPA 6850 Soil

APPL ID: 150518S-16137 ICS - 198067
Batch ID: #6850SM-150518A

APPL Inc.
908 N. Temperance Ave.
Clovis, CA 93711

Compound Name	Spike Level mg/Kg	Spk Result mg/Kg	SPK% Recovery	Recovery Limits
Perchlorate	0.00571	0.00599	104.9%	70-130

Comments:

Quant Method:	QTLMFL2
Extraction Date:	05/18/15
Analysis Date:	05/24/15
Instrument:	AGIL_6560
Run:	T523_048.D
Initials:	RP

INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. **SDG No.**

Lab File ID (Standard): T312_003-_008.d

Instrument ID: Agilent 6460 Triple Quad LC/MS

	PER_IS			
	12_HOUR_STD AVERAGE	11032021		
	UPPER LIMIT RESPONSE	16548032		
	LOWER LIMIT RESPONSE	5516011		
Data File ID:	Sample ID:	Analysis Date and Time	IS AREA COUNT #	IS R.T. #
T312_003.d	PERCHLORATE 0.0002 ug/ml 11/11/14	3/12/15 17:52	8337058	13.732
T312_004.d	PERCHLORATE 0.0004 ug/ml 11/11/14	3/12/15 18:11	10671196	13.783
T312_005.d	PERCHLORATE 0.001 ug/ml 11/11/14	3/12/15 18:29	10651968	13.803
T312_006.d	PERCHLORATE 0.002 ug/ml 11/11/14	3/12/15 18:48	13529952	13.803
T312_007.d	PERCHLORATE 0.005 ug/ml 11/11/14	3/12/15 19:07	11028880	13.874
T312_008.d	PERCHLORATE 0.010 ug/ml 08/27/14	3/12/15 19:25	11973072	13.874
T312_010.d	PER_SS 0.0004 ug/ml 11/11/14	3/12/15 20:02	12977621	13.946
T312_011.d	PER_SS 0.002 ug/ml 11/11/14	3/12/15 20:21	13747369	13.976
T312_013.d	PER_CCV_1 0.0004 ug/ml 11/11/14	3/12/15 20:58	12323795	14.109
T312_014.d	PER_CCV_1 0.002 ug/ml 11/11/14	3/12/15 21:17	13576634	14.139

AREA UPPER LIMIT = +50% of internal standard average
AREA LOWER LIMIT = -50% of internal standard average
Column used to flag values outside QC limits
* Values outside of QC limits

INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. **SDG No.** **76390**

Lab File ID (Standard): T523_003.d

Instrument ID: Agilent 6460 Triple Quad LC/MS

	PER_IS			
	12_HOUR_STD AVERAGE	8764619		
	UPPER LIMIT RESPONSE	13146929		
	LOWER LIMIT RESPONSE	4382310		
Data File ID:	Sample ID:	Analysis Date and Time	IS AREA COUNT #	IS R.T. #
T523_002.d	PER_CCV 0.0004 ug/ml 05/07/15	5/23/15 15:38	7841423	12.957
T523_003.d	PER_CCV 0.002 ug/ml 05/07/15	5/23/15 15:57	8764619	13.13
T523_044.d	PER_CCV 0.0004 ug/ml 05/07/15	5/24/15 4:55	6579582	14.037
T523_045.d	PER_CCV 0.002 ug/ml 05/07/15	5/24/15 5:13	6755219	14.2
T523_047.d	150518SA_LCS-1 10422.1 DF 05/18/15	5/24/15 5:50	5484178	14.262
T523_048.d	150518S_ICSA 10025.1 DF 05/18/15	5/24/15 6:09	4353871*	13.273
T523_049.d	150518SBLKA 10319.9 DF 05/18/15	5/24/15 6:28	5786126	14.251
T523_057.d	PER_CCV 0.0004 ug/ml 05/07/15	5/24/15 8:57	6631309	13.762
T523_058.d	PER_CCV 0.002 ug/ml 05/07/15	5/24/15 9:15	7264937	13.823
T523_061.d	AZ16198_S01 8628.1 DF 05/18/15	5/24/15 10:11	6460539	13.589
T523_062.d	AZ16199_S01 8354.2 DF 05/18/15	5/24/15 10:30	6231243	13.334
T523_063.d	AZ16201_S01 8421.1 DF 05/18/15	5/24/15 10:48	6390660	13.426
T523_064.d	AZ16202_S01 8489.0 DF 05/18/15	5/24/15 11:07	6171466	13.283
T523_069.d	PER_CCV 0.0004 ug/ml 05/07/15	5/24/15 13:00	4418362	13.406
T523_070.d	PER_CCV 0.002 ug/ml 05/07/15	5/24/15 13:18	5503588	13.497

AREA UPPER LIMIT = +50% of internal standard average
AREA LOWER LIMIT = -50% of internal standard average
Column used to flag values outside QC limits
* Values outside of QC limits

Perchlorate LC/MS Ion Ratio Report

acceptable ratio is between 2.3 and 3.8 (DoD Perchlorate Handbook, 2006)

Data File Path : D:\masshunter\data\150312
Instrument Name : TQ (Agilent 6460 Triple Quad LC/MS)
Method File : 6460_ESI_PER_N_NEWER_K'_COLUMN.m
Operator : BA

Date Aquired: 03/12/15 17:15 Vial Number: NA
Sample Name: PER_IS_50:50 0.005 ug/ml 11/11/14 Data File ID: T312_001.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.014	113760	3.1
2)	Perchlorate_85	14.071	36914	

Date Aquired: 03/12/15 17:33 Vial Number: NA
Sample Name: PERCHLORATE 0.0001 ug/ml 11/11/14 Data File ID: T312_002.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.504	237168	3.1
2)	Perchlorate_85	13.541	77486	

Date Aquired: 03/12/15 17:52 Vial Number: NA
Sample Name: PERCHLORATE 0.0002 ug/ml 11/11/14 Data File ID: T312_003.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.789	446712	3.1
2)	Perchlorate_85	13.704	144754	

Date Aquired: 03/12/15 18:11 Vial Number: NA
Sample Name: PERCHLORATE 0.0004 ug/ml 11/11/14 Data File ID: T312_004.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.8	1023403	3.0
2)	Perchlorate_85	13.827	337305	

Date Aquired: 03/12/15 18:29 Vial Number: NA
Sample Name: PERCHLORATE 0.001 ug/ml 11/11/14 Data File ID: T312_005.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.82	2483393	3.2
2)	Perchlorate_85	13.827	781617	

Date Aquired: 03/12/15 18:48 Vial Number: NA
Sample Name: PERCHLORATE 0.002 ug/ml 11/11/14 Data File ID: T312_006.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.83	6351752	3.3
2)	Perchlorate_85	13.817	1920912	

Date Aquired: 03/12/15 19:07 Vial Number: NA
Sample Name: PERCHLORATE 0.005 ug/ml 11/11/14 Data File ID: T312_007.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.861	14499255	3.4
2)	Perchlorate_85	13.878	4209044	

Date Aquired: 03/12/15 19:25 Vial Number: NA
Sample Name: PERCHLORATE 0.010 ug/ml 08/27/14 Data File ID: T312_008.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.881	34223790	3.4
2)	Perchlorate_85	13.878	9985270	

Date Aquired: 03/12/15 19:44 Vial Number: NA
Sample Name: PER_IS_50:50 0.005 ug/ml 11/11/14 Data File ID: T312_009.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.912	249406	2.9
2)	Perchlorate_85	13.949	85978	

Date Aquired: 03/12/15 20:02 Vial Number: NA
Sample Name: PER_SS 0.0004 ug/ml 11/11/14 Data File ID: T312_010.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.942	1472839	3.3
2)	Perchlorate_85	13.959	440675	

Date Aquired: 03/12/15 20:21 Vial Number: NA
Sample Name: PER_SS 0.002 ug/ml 11/11/14 Data File ID: T312_011.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.014	7468678	3.4
2)	Perchlorate_85	14	2222387	

Date Aquired: 03/12/15 20:40 Vial Number: NA
Sample Name: PER_IS_50:50 0.005 ug/ml 11/11/14 Data File ID: T312_012.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.657	10916	2.0
2)	Perchlorate_85	13.776	5521	

Date Aquired: 03/12/15 20:58 Vial Number: NA
Sample Name: PER_CCV_1 0.0004 ug/ml 11/11/14 Data File ID: T312_013.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.105	1267031	3.1
2)	Perchlorate_85	14.143	410558	

Date Aquired: 03/12/15 21:17 Vial Number: NA
Sample Name: PER_CCV_1 0.002 ug/ml 11/11/14 Data File ID: T312_014.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.146	6907110	3.3
2)	Perchlorate_85	14.122	2093567	

Date Aquired: 03/12/15 21:36 Vial Number: NA
Sample Name: PER_IS_50:50 0.005 ug/ml 11/11/14 Data File ID: T312_015.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.36	64783	9.6
2)	Perchlorate_85	13.715	6730	

Perchlorate LC/MS Ion Ratio Report

acceptable ratio is between 2.3 and 3.8 (DoD Perchlorate Handbook, 2006)

Data File Path : D:\masshunter\data\150523
Instrument Name : TQ (Agilent 6460 Triple Quad LC/MS)
Method File : 6460_ESI_PER_N_NEWER_K'_COLUMN.m
Operator : BA

Date Aquired: 05/23/15 15:20 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_001.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.117	71005	1.9
2)	Perchlorate_85	13.134	37385	

Date Aquired: 05/23/15 15:38 Vial Number: NA
Sample Name: PER_CCV 0.0004 ug/ml 05/07/15 Data File ID: T523_002.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	12.974	857023	2.7
2)	Perchlorate_85	12.981	313246	

Date Aquired: 05/23/15 15:57 Vial Number: NA
Sample Name: PER_CCV 0.002 ug/ml 05/07/15 Data File ID: T523_003.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.127	4097326	3.3
2)	Perchlorate_85	13.134	1252418	

Date Aquired: 05/23/15 16:15 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_004.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.147	50856	2.5
2)	Perchlorate_85	12.981	20452	

Date Aquired: 05/23/15 16:37 Vial Number: NA
Sample Name: 150521SA_LCS-1 10422.1 DF 05/21/15 Data File ID: T523_005.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.147	595644	2.8
2)	Perchlorate_85	13.175	215686	

Date Aquired: 05/24/15 4:17 Vial Number: NA
Sample Name: AZ16403_W04 1052.6 DF 05/18/15 Data File ID: T523_042.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.626	6933729	3.3
2)	Perchlorate_85	13.613	2075967	

Date Aquired: 05/24/15 4:36 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_043.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.126	79123	26.2
2)	Perchlorate_85	12.94	3023	

Date Aquired: 05/24/15 4:55 Vial Number: NA
Sample Name: PER_CCV 0.0004 ug/ml 05/07/15 Data File ID: T523_044.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.024	691196	2.8
2)	Perchlorate_85	14.031	247754	

Date Aquired: 05/24/15 5:13 Vial Number: NA
Sample Name: PER_CCV 0.002 ug/ml 05/07/15 Data File ID: T523_045.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.217	3171297	3.3
2)	Perchlorate_85	14.194	970030	

Date Aquired: 05/24/15 5:32 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_046.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.207	77653	13.8
2)	Perchlorate_85	12.971	5620	

Date Aquired: 05/24/15 5:50 Vial Number: NA
Sample Name: 150518SA_LCS-1 10422.1 DF 05/18/15 Data File ID: T523_047.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.248	742644	2.9
2)	Perchlorate_85	14.245	256573	

Date Aquired: 05/24/15 6:09 Vial Number: NA
Sample Name: 150518S_ICSA 10025.1 DF 05/18/15 Data File ID: T523_048.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.26	660527	2.2
2)	Perchlorate_85	13.246	295223	

Date Aquired: 05/24/15 6:28 Vial Number: NA
Sample Name: 150518SBLKA 10319.9 DF 05/18/15 Data File ID: T523_049.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	12.995	7316	1.3
2)	Perchlorate_85	13.052	5610	

Date Aquired: 05/24/15 6:46 Vial Number: NA
Sample Name: AZ16137-S01 9153.3 DF 05/18/15 Data File ID: T523_050.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	12.832	7187	0.5
2)	Perchlorate_85	12.614	15819	

Date Aquired: 05/24/15 7:05 Vial Number: NA
Sample Name: AZ16137_S01_MS-1 8920.6 DF 05/18/15 Data File ID: T523_051.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.126	798462	3.1
2)	Perchlorate_85	14.132	261182	

Date Aquired: 05/24/15 7:24 Vial Number: NA
Sample Name: AZ16137_S01_MSD-1 8996.9 DF 05/18/15 Data File ID: T523_052.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	14.064	809460	3.1
2)	Perchlorate_85	14.071	260591	

Date Aquired: 05/24/15 7:42 Vial Number: NA
Sample Name: AZ16138_S01 9074.4 DF 05/18/15 Data File ID: T523_053.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	12.964	8625	5.2
2)	Perchlorate_85	12.961	1657	

Date Aquired: 05/24/15 8:01 Vial Number: NA
Sample Name: AZ16139_S01 8628.1 DF 05/18/15 Data File ID: T523_054.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.035	5371	0.9
2)	Perchlorate_85	13.022	5841	

Date Aquired: 05/24/15 8:19 Vial Number: NA
Sample Name: AZ16140_S01 9153.3 DF 05/18/15 Data File ID: T523_055.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.993	3337205	3.3
2)	Perchlorate_85	13.99	1007106	

Date Aquired: 05/24/15 8:38 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_056.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.932	66732	15.4
2)	Perchlorate_85	12.889	4323	

Date Aquired: 05/24/15 8:57 Vial Number: NA
Sample Name: PER_CCV 0.0004 ug/ml 05/07/15 Data File ID: T523_057.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.769	758603	3.0
2)	Perchlorate_85	13.786	254673	

Date Aquired: 05/24/15 9:15 Vial Number: NA
Sample Name: PER_CCV 0.002 ug/ml 05/07/15 Data File ID: T523_058.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.82	3406090	3.3
2)	Perchlorate_85	13.806	1040769	

Date Aquired: 05/24/15 9:34 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_059.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.749	65303	27.8
2)	Perchlorate_85	12.991	2347	

Date Aquired: 05/24/15 9:53 Vial Number: NA
Sample Name: AZ16141_S01 7683.4 DF 05/18/15 Data File ID: T523_060.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.667	2251605	3.2
2)	Perchlorate_85	13.674	705207	

Date Aquired: 05/24/15 10:11 Vial Number: NA
Sample Name: AZ16198_S01 8628.1 DF 05/18/15 Data File ID: T523_061.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.616	190072	2.7
2)	Perchlorate_85	13.603	71515	

Date Aquired: 05/24/15 10:30 Vial Number: NA
Sample Name: AZ16199_S01 8354.2 DF 05/18/15 Data File ID: T523_062.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.351	2265490	3.1
2)	Perchlorate_85	13.338	726832	

Date Aquired: 05/24/15 10:48 Vial Number: NA
Sample Name: AZ16201_S01 8421.1 DF 05/18/15 Data File ID: T523_063.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	12.944	2540	0.1
2)	Perchlorate_85	12.787	17805	

Date Aquired: 05/24/15 11:07 Vial Number: NA
Sample Name: AZ16202_S01 8489.0 DF 05/18/15 Data File ID: T523_064.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.29	18415075	3.4
2)	Perchlorate_85	13.297	5380483	

Date Aquired: 05/24/15 11:26 Vial Number: NA
Sample Name: AZ16400_S01 7914.5 DF 05/18/15 Data File ID: T523_065.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.127	3795664	3.2
2)	Perchlorate_85	13.124	1177554	

Date Aquired: 05/24/15 11:44 Vial Number: NA
Sample Name: AZ16401_S01 8097.2 DF 05/18/15 Data File ID: T523_066.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.229	872107	3.2
2)	Perchlorate_85	13.277	270166	

Date Aquired: 05/24/15 12:03 Vial Number: NA
Sample Name: AZ16402_S01 7914.5 DF 05/18/15 Data File ID: T523_067.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.035	3325948	3.3
2)	Perchlorate_85	13.032	995494	

Date Aquired: 05/24/15 12:41 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_068.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.198	55871	40.2
2)	Perchlorate_85	12.981	1390	

Date Aquired: 05/24/15 13:00 Vial Number: NA
Sample Name: PER_CCV 0.0004 ug/ml 05/07/15 Data File ID: T523_069.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.423	521126	3.0
2)	Perchlorate_85	13.429	173626	

Date Aquired: 05/24/15 13:18 Vial Number: NA
Sample Name: PER_CCV 0.002 ug/ml 05/07/15 Data File ID: T523_070.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.504	2587270	3.2
2)	Perchlorate_85	13.521	812912	

Date Aquired: 05/24/15 13:37 Vial Number: NA
Sample Name: PER_IS 0.005 ug/ml 05/07/15 Data File ID: T523_071.d

#	NAME	Ret Time	Target Response	Ratio
1)	Perchlorate_83	13.219	52496	40.9
2)	Perchlorate_85	13.032	1282	

EPA METHOD 6850
Perchlorate
LC/MS

Sample Data

APPL, INC.

PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 05/07/15

QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 13.4 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_050.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/17/15 10:45:54 AM

APPL-F1-SC-MCResOnly-REG MDLs

Data File ID: T523_050.d

Date Injected : 05/24/15

Time Injected : 6:46

Sample ID : AZ16137-S01 9153.3 DF 05/18/15

Client ID : S67-SS030-0001

Retention Time	Area Count Response	Compound_ID Product Ion		
14.231	5720392	PER_IS_89		
12.832	7187	Perchlorate_83	(7187 * 0.0050) / (1.27 * 5720392.00) * 9153.30 =	0.045267 ppb
12.614	15819	Perchlorate_85	(15819 * 0.0050) / (0.39 * 5720392.00) * 9153.30 =	0.323186 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_050.d	**Sample Name**	AZ16137-S01 9153.3 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	6:46
ClientID	S67-SS030-0001	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	12.832	7187	5720392
Perchlorate_102	PER_IS_108	12.614	15819	5720392

Perchlorate_100







Perchlorate_102







PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16138
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 12.0 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_053.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/17/15 10:45:54 AM
APPL-F1-SC-MCResOnly-REG MDLs

Data File ID: T523_053.d

Date Injected : 05/24/15

Time Injected : 7:42

Sample ID : AZ16138_S01 9074.4 DF 05/18/15

Client ID : S67-SS030P-0001

Retention Time	Area Count Response	Compound_ID Product Ion			
14.058	6201818	PER_IS_89			
12.964	8625	Perchlorate_83	(8625 * 0.0050) / (1.27 * 6201818.00) *	9074.40 =	0.049675 ppb
12.961	1657	Perchlorate_85	(1657 * 0.0050) / (0.39 * 6201818.00) *	9074.40 =	0.030956 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_053.d	**Sample Name**	AZ16138_S01 9074.4 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	7:42
ClientID	S67-SS030P-0001	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	12.964	8625	6201818
Perchlorate_102	PER_IS_108	12.961	1657	6201818

Perchlorate_100







Perchlorate_102







PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB030-0607

APPL ID: AZ16139

Sample Collection Date: 05/07/15

QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 17.5 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_054.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/17/15 10:45:54 AM

APPL-F1-SC-MCResOnly-REG MDLs

Data File ID: T523_054.d

Date Injected : 05/24/15

Time Injected : 8:01

Sample ID : AZ16139_S01 8628.1 DF 05/18/15

Client ID : S67-SB030-0607

Retention Time	Area Count Response	Compound_ID Product Ion			
14.017	6332981	PER_IS_89			
13.035	5371	Perchlorate_83	(5371 * 0.0050) / (1.27 * 6332981.00) *	8628.10 =	0.028803 ppb
13.022	5841	Perchlorate_85	(5841 * 0.0050) / (0.39 * 6332981.00) *	8628.10 =	0.101605 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_054.d	**Sample Name**	AZ16139_S01 8628.1 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	8:01
ClientID	S67-SB030-0607	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.035	5371	6332981
Perchlorate_102	PER_IS_108	13.022	5841	6332981

Perchlorate_100







Perchlorate_102







PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16140
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 10.4 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.022	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method:	QTLMFL2
Run #:	T523_055.d
Instrument:	AGIL_6460
Sequence:	TQ052315
Dilution Factor:	1
Initials:	RP

Printed: 06/17/15 10:45:54 AM
APPL-F1-SC-MCResOnly-REG MDLs

Data File ID: T523_055.d

Date Injected : 05/24/15

Time Injected : 8:19

Sample ID : AZ16140_S01 9153.3 DF 05/18/15

Client ID : S67-SS032-0001

Retention Time	Area Count Response	Compound_ID Product Ion		
13.976	6115878	PER_IS_89		
13.993	3337205	Perchlorate_83	(3337205 * 0.0050) / (1.27 * 6115878.00) * 9153.30 =	19.659876 ppb
13.99	1007106	Perchlorate_85	(1007106 * 0.0050) / (0.39 * 6115878.00) * 9153.30 =	19.244908 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_055.d	**Sample Name**	AZ16140_S01 9153.3 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	8:19
ClientID	S67-SS032-0001	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.993	3337205	6115878
Perchlorate_102	PER_IS_108	13.990	1007106	6115878

Perchlorate_100







Perchlorate_102







PERCHLORATE EPA 6850 - SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16141
QCG: #6850SM-150518A-198067

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
(Solid Concentrations have been adjusted to reflect 23.3 Percent Moisture.)								
EPA 6850	PERCHLORATE	0.013	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method: QTLMFL2
Run #: T523_060.d
Instrument: AGIL_6460
Sequence: TQ052315
Dilution Factor: 1
Initials: RP

Printed: 06/17/15 10:45:54 AM
APPL-F1-SC-MCResOnly-REG MDLs

Data File ID: T523_060.d

Date Injected : 05/24/15

Time Injected : 9:53

Sample ID : AZ16141_S01 7683.4 DF 05/18/15

Client ID : S67-SB032-0607

Retention Time		Area Count Response	Compound_ID Product Ion		
13.66		6662373	PER_IS_89		
13.667	*	2251605	Perchlorate_83	(2251605 * 0.0050) / (1.27 * 6662373.00) * 7683.40 =	10.221054 ppb
13.674		705207	Perchlorate_85	(705207 * 0.0050) / (0.39 * 6662373.00) * 7683.40 =	10.383957 ppb

* MANUAL INTEGRAION

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_060.d	**Sample Name**	AZ16141_S01 7683.4 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	9:53
ClientID	S67-SB032-0607	**Inj Vol**	20



Compound	ISTD		RT	Resp	ISTD Resp	
Perchlorate_100	PER_IS_108	*	13.667	2251605	6662373	* MANUAL INT
Perchlorate_102	PER_IS_108		13.674	705207	6662373	

Perchlorate_100



Perchlorate_102




- MRM (99.0 -> 83.0) T523_060.d
Counts
x10 5
13.667 min.
Acquisition Time (min)


- MRM (99.0 -> 83.0) T523_060.d
Counts
x10 5
*13.667 min.
M-1
06-16-15
PM 6/16/15
Acquisition Time (min)

EPA METHOD 6850
Perchlorate
LC/MS

Calibration Data

APPL, INC.

METHOD 6850 CALIBRATION SUMMARY

Lab Name: APPL Inc. **DETECTOR ID : Agilent 6460 Triple Quad LC/MS**

Average Relative Response Factor Summary

Analyte Id:	R.R.F.#1 PERCHLORATE 0.0002 PPM	R.R.F.#2 PERCHLORATE 0.0004 PPM	R.R.F.#3 PERCHLORATE 0.001 PPM	R.R.F.#4 PERCHLORATE 0.002 PPM	R.R.F.#5 PERCHLORATE 0.005 PPM	R.R.F.#6 PERCHLORATE 0.010 PPM	AVERAGE REL.RESP. FACTOR	%RSD RRF
Perchlorate_83	1.339537	1.198791	1.165697	1.173646	1.314663	1.429198	1.270255	8.44
Perchlorate_85	0.434068	0.395112	0.366889	0.354937	0.381638	0.416989	0.391605	7.67

FILE ID : TQ0312A.FRB

Perchlorate_83

RESPONSE FACTOR CALCULATIONS - AREA COUNTS

DATA FILES : T312_003.d - T312_008.d

The Data points that were Read Were

Standard Response	Standard Concentration µg/ml	Internal Standard Response	Internal Standard Concentration µg/ml	Response Factor	RPD	%RSD
446712.00	0.0002	8337058.00	0.0050	1.339537	5.45	8.44
1023403.00	0.0004	10671196.00	0.0050	1.198791	5.63	
2483393.00	0.0010	10651968.00	0.0050	1.165697	8.23	
6351752.00	0.0020	13529952.00	0.0050	1.173646	7.61	
14499255.00	0.0050	11028880.00	0.0050	1.314663	3.50	
34223790.00	0.0100	11973072.00	0.0050	1.429198	12.51	

The Average Response Factor = 1.270255

Perchlorate_85

RESPONSE FACTOR CALCULATIONS - AREA COUNTS

DATA FILES : T312_003.d - T312_008.d

The Data points that were Read Were

Standard Response	Standard Concentration µg/ml	Internal Standard Response	Internal Standard Concentration µg/ml	Response Factor	RPD	%RSD
144754.00	0.0002	8337058.00	0.0050	0.434068	10.84	7.67
337305.00	0.0004	10671196.00	0.0050	0.395112	0.90	
781617.00	0.0010	10651968.00	0.0050	0.366889	6.31	
1920912.00	0.0020	13529952.00	0.0050	0.354937	9.36	
4209044.00	0.0050	11028880.00	0.0050	0.381638	2.55	
9985270.00	0.0100	11973072.00	0.0050	0.416989	6.48	

The Average Response Factor = 0.391605

Data File ID: T312_002.d

Date Injected : 03/12/15

Time Injected : 17:33

Sample ID : PERCHLORATE 0.0001 ug/ml 11/11/14

Client ID : NA

Retention Time		Area Count Response	Compound_ID Product Ion
13.548		5980396	PER_IS_89
13.504	*	237168	Perchlorate_83
13.541	*	77486	Perchlorate_85

* MANUAL INTEGRAION

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_002.d	**Sample Name**	PERCHLORATE 0.0001 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	17:33
ClientID	NA	**Inj Vol**	20



Compound	ISTD		RT	Resp	ISTD Resp	
Perchlorate_100	PER_IS_108	*	13.504	237168	5980396	* MANUAL INT
Perchlorate_102	PER_IS_108	*	13.541	77486	5980396	* MANUAL INT

Perchlorate_100



Perchlorate_102




- MRM (99.0 -> 83.0) T312_002.d
Counts
x10 3
9.6
9.4
9.2
9
8.8
8.6
8.4
8.2
8
7.8
7.6
7.4
7.2
7
6.8
6.6
6.4
6.2
6
5.8
5.6
5.4
5.2
5
4.8
4.6
4.4
4.2
4
3.8
3.6
3.4
3.2
3
2.8
2.6
2.4
2.2
2
1.8
1.6
1.4
1.2
1
0.8
0.6
0.4
0.2
13.942 min.
9
9.5
10
10.5
11
11.5
12
12.5
13
13.5
14
14.5
15
15.5
16
16.5
17
17.5
Acquisition Time (min)


- MRM (99.0 -> 83.0) T312_002.d
Counts
x10 3
*13.504 min.
Acquisition Time (min)
M.H
03-16-15
PM 4/16/15


- MRM (101.0 -> 84.9) T312_002.d
Counts
x10 3
4.4
4.3
4.2
4.1
4
3.9
3.8
3.7
3.6
3.5
3.4
3.3
3.2
3.1
3
2.9
2.8
2.7
2.6
2.5
2.4
2.3
2.2
2.1
2
1.9
1.8
1.7
1.6
1.5
1.4
1.3
1.2
1.1
1
0.9
0.8
0.7
14.010 min.
9
9.5
10
10.5
11
11.5
12
12.5
13
13.5
14
14.5
15
15.5
16
16.5
17
17.5
Acquisition Time (min)


- MRM (101.0 -> 84.9) T312_002.d
Counts
x10 3
*13.541 min.
4.4
4.3
4.2
4.1
4
3.9
3.8
3.7
3.6
3.5
3.4
3.3
3.2
3.1
3
2.9
2.8
2.7
2.6
2.5
2.4
2.3
2.2
2.1
2
1.9
1.8
1.7
1.6
1.5
1.4
1.3
1.2
1.1
1
0.9
0.8
0.7
9
9.5
10
10.5
11
11.5
12
12.5
13
13.5
14
14.5
15
15.5
16
16.5
17
17.5
Acquisition Time (min)
JL
M-L
04-16-15
pm 4/16/15

Data File ID: T312_004.d

Date Injected : 03/12/15

Time Injected : 18:11

Sample ID : PERCHLORATE 0.0004 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.783	10671196	PER_IS_89
13.8	1023403	Perchlorate_83
13.827	337305	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_004.d	**Sample Name**	PERCHLORATE 0.0004 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	18:11
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.800	1023403	10671196
Perchlorate_102	PER_IS_108	13.827	337305	10671196

Perchlorate_100







Perchlorate_102







Data File ID: T312_005.d

Date Injected : 03/12/15

Time Injected : 18:29

Sample ID : PERCHLORATE 0.001 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.803	10651968	PER_IS_89
13.82	2483393	Perchlorate_83
13.827	781617	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_005.d	**Sample Name**	PERCHLORATE 0.001 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	18:29
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.820	2483393	10651968
Perchlorate_102	PER_IS_108	13.827	781617	10651968

Perchlorate_100







Perchlorate_102







Data File ID: T312_006.d

Date Injected : 03/12/15

Time Injected : 18:48

Sample ID : PERCHLORATE 0.002 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.803	13529952	PER_IS_89
13.83	6351752	Perchlorate_83
13.817	1920912	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_006.d	**Sample Name**	PERCHLORATE 0.002 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	18:48
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.830	6351752	13529952
Perchlorate_102	PER_IS_108	13.817	1920912	13529952

Perchlorate_100



Perchlorate_102



Data File ID: T312_007.d

Date Injected : 03/12/15

Time Injected : 19:07

Sample ID : PERCHLORATE 0.005 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.874	11028880	PER_IS_89
13.861	14499255	Perchlorate_83
13.878	4209044	Perchlorate_85

Quantitative Analysis Sample Report

Batch-Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_007.d	**Sample Name**	PERCHLORATE 0.005 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	19:07
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.861	14499255	11028880
Perchlorate_102	PER_IS_108	13.878	4209044	11028880

Perchlorate_100



Perchlorate_102



Data File ID: T312_008.d

Date Injected : 03/12/15

Time Injected : 19:25

Sample ID : PERCHLORATE 0.010 ug/ml 08/27/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.874	11973072	PER_IS_89
13.881	34223790	Perchlorate_83
13.878	9985270	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_008.d	**Sample Name**	PERCHLORATE 0.010 ug/ml 08/27/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	19:25
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.881	34223790	11973072
Perchlorate_102	PER_IS_108	13.878	9985270	11973072

Perchlorate_100







Perchlorate_102







SECOND SOURCE CALIBRATION VERIFICATION SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T312_010.d

	Date (s) of Analysis From:03/12/15 To:03/12/15 Time (s) of Analysis From:17:33 To:19:44				Date of Analysis: 03/12/15 Time of Analysis: 20:02 Standard Id: PER_SS (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.50	13.40	13.60	1.270255	13.94	1.418633	N	11.7
Perchlorate_85	13.54	13.44	13.64	0.391605	13.96	0.424457	N	8.4

ILE ID : TQCK1210.RFB

Data File ID: T312_010.d

Date Injected : 03/12/15

Time Injected : 20:02

Sample ID : PER_SS 0.0004 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.946	12977621	PER_IS_89
13.942	1472839	Perchlorate_83
13.959	440675	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_010.d	**Sample Name**	PER_SS 0.0004 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	20:02
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.942	1472839	12977621
Perchlorate_102	PER_IS_108	13.959	440675	12977621

Perchlorate_100







Perchlorate_102







SECOND SOURCE CALIBRATION VERIFICATION SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T312_011.d

	Date (s) of Analysis From:03/12/15 To:03/12/15 Time (s) of Analysis From:17:33 To:19:44				Date of Analysis: 03/12/15 Time of Analysis: 20:21 Standard Id: PER_SS (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.50	13.40	13.60	1.270255	14.01	1.358201	N	6.9
Perchlorate_85	13.54	13.44	13.64	0.391605	14.00	0.404148	N	3.2

FILE ID : TQCK1211.RFB

Data File ID: T312_011.d

Date Injected : 03/12/15

Time Injected : 20:21

Sample ID : PER_SS 0.002 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.976	13747369	PER_IS_89
14.014	7468678	Perchlorate_83
14.000	2222387	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_011.d	**Sample Name**	PER_SS 0.002 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	20:21
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.014	7468678	13747369
Perchlorate_102	PER_IS_108	14.000	2222387	13747369

Perchlorate_100







Perchlorate_102







CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T312_013.d

	Date (s) of Analysis From:03/12/15 To:03/12/15 Time (s) of Analysis From:17:33 To:20:40				Date of Analysis: 03/12/15 Time of Analysis: 20:58 Standard Id: PER_CCV_1 (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.50	13.40	13.60	1.270255	14.10	1.285147	N	1.2
Perchlorate_85	13.54	13.44	13.64	0.391605	14.14	0.416428	N	6.3

FILE ID : TQCK1213.RFB

Data File ID: T312_013.d

Date Injected : 03/12/15

Time Injected : 20:58

Sample ID : PER_CCV_1 0.0004 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
14.109	12323795	PER_IS_89
14.105	1267031	Perchlorate_83
14.143	410558	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_013.d	**Sample Name**	PER_CCV_1 0.0004 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	20:58
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.105	1267031	12323795
Perchlorate_102	PER_IS_108	14.143	410558	12323795

Perchlorate_100







Perchlorate_102







CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T312_014.d

	Date (s) of Analysis From:03/12/15 To:03/12/15 Time (s) of Analysis From:17:33 To:20:40				Date of Analysis: 03/12/15 Time of Analysis: 21:17 Standard Id: PER_CCV_1 (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.50	13.40	13.60	1.270255	14.15	1.271875	N	0.1
Perchlorate_85	13.54	13.44	13.64	0.391605	14.12	0.385509	N	1.6

FILE ID : TQCK1214.RFB

Data File ID: T312_014.d

Date Injected : 03/12/15

Time Injected : 21:17

Sample ID : PER_CCV_1 0.002 ug/ml 11/11/14

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
14.139	13576634	PER_IS_89
14.146	6907110	Perchlorate_83
14.122	2093567	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150312\QuantResults\150312_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T312_014.d	**Sample Name**	PER_CCV_1 0.002 ug/ml 11/11/14
Sample Type	Sample	**Acq_Date**	03/12/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	21:17
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.146	6907110	13576634
Perchlorate_102	PER_IS_108	14.122	2093567	13576634

Perchlorate_100







Perchlorate_102







CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_002.d

	Date (s) of Analysis From:03/12/15 To:05/23/15 Time (s) of Analysis From:17:52 To:15:20				Date of Analysis: 05/23/15 Time of Analysis: 15:38 Standard Id: PER_CCV (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	12.97	1.366179	N	7.6
Perchlorate_85	13.70	13.60	13.80	0.391605	12.98	0.499345	N	27.5

FILE ID : TQCK2302.RFB

Data File ID: T523_002.d

Date Injected : 05/23/15

Time Injected : 15:38

Sample ID : PER_CCV 0.0004 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
12.957	7841423	PER_IS_89
12.974	857023	Perchlorate_83
12.981	313246	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_002.d	**Sample Name**	PER_CCV 0.0004 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/23/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	15:38
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	12.974	857023	7841423
Perchlorate_102	PER_IS_108	12.981	313246	7841423

Perchlorate_100



Perchlorate_102



CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_003.d

	Date (s) of Analysis From:03/12/15 To:05/23/15 Time (s) of Analysis From:17:52 To:15:20				Date of Analysis: 05/23/15 Time of Analysis: 15:57 Standard Id: PER_CCV (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	13.13	1.168712	N	8.0
Perchlorate_85	13.70	13.60	13.80	0.391605	13.13	0.357237	N	8.8

FILE ID : TQCK2303.RFB

Data File ID: T523_003.d

Date Injected : 05/23/15

Time Injected : 15:57

Sample ID : PER_CCV 0.002 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.13	8764619	PER_IS_89
13.127	4097326	Perchlorate_83
13.134	1252418	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_003.d	**Sample Name**	PER_CCV 0.002 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/23/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	15:57
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.127	4097326	8764619
Perchlorate_102	PER_IS_108	13.134	1252418	8764619

Perchlorate_100



Perchlorate_102



CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_044.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:4:17				Date of Analysis: 05/24/15 Time of Analysis: 4:55 Standard Id: PER_CCV (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	14.02	1.313146	N	3.4
Perchlorate_85	13.70	13.60	13.80	0.391605	14.03	0.470687	N	20.2

FILE ID : TQCK2344.RFB

Data File ID: T523_044.d

Date Injected : 05/24/15

Time Injected : 4:55

Sample ID : PER_CCV 0.0004 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
14.037	6579582	PER_IS_89
14.024	691196	Perchlorate_83
14.031	247754	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_044.d	**Sample Name**	PER_CCV 0.0004 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	4:55
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.024	691196	6579582
Perchlorate_102	PER_IS_108	14.031	247754	6579582

Perchlorate_100







Perchlorate_102







CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_045.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:4:17				Date of Analysis: 05/24/15 Time of Analysis: 5:13 Standard Id: PER_CCV (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	14.22	1.173647	N	7.6
Perchlorate_85	13.70	13.60	13.80	0.391605	14.19	0.358993	N	8.3

FILE ID : TQCK2345.RFB

Data File ID: T523_045.d

Date Injected : 05/24/15

Time Injected : 5:13

Sample ID : PER_CCV 0.002 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
14.2	6755219	PER_IS_89
14.217	3171297	Perchlorate_83
14.194	970030	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A1_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_045.d	**Sample Name**	PER_CCV 0.002 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	5:13
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.217	3171297	6755219
Perchlorate_102	PER_IS_108	14.194	970030	6755219

Perchlorate_100



Perchlorate_102



CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_057.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:8:19				Date of Analysis: 05/24/15 Time of Analysis: 8:57 Standard Id: PER_CCV (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	13.77	1.429965	N	12.6
Perchlorate_85	13.70	13.60	13.80	0.391605	13.79	0.480058	N	22.6

FILE ID : TQCK2357.RFB

Data File ID: T523_057.d

Date Injected : 05/24/15

Time Injected : 8:57

Sample ID : PER_CCV 0.0004 ug/ml 05/07/15

Client ID : NA

Retention Time		Area Count Response	Compound_ID Product Ion
13.762		6631309	PER_IS_89
13.769		758603	Perchlorate_83
13.786	*	254673	Perchlorate_85

* MANUAL INTEGRAION

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_057.d	**Sample Name**	PER_CCV 0.0004 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	8:57
ClientID	NA	**Inj Vol**	20



Compound	ISTD		RT	Resp	ISTD Resp	
Perchlorate_100	PER_IS_108		13.769	758603	6631309	
Perchlorate_102	PER_IS_108	*	13.786	254673	6631309	* MANUAL INT

Perchlorate_100



Perchlorate_102




- MRM (101.0 -> 84.9) T523_057.d
Counts
x10 4
13.786 min.
Acquisition Time (min)



BL
M-4
06-16-15

pm6/16/15

CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_058.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:8:19				Date of Analysis: 05/24/15 Time of Analysis: 9:15 Standard Id: PER_CCV (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	13.82	1.172099	N	7.7
Perchlorate_85	13.70	13.60	13.80	0.391605	13.81	0.358148	N	8.5

FILE ID : TQCK2358.RFB

Data File ID: T523_058.d

Date Injected : 05/24/15

Time Injected : 9:15

Sample ID : PER_CCV 0.002 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.823	7264937	PER_IS_89
13.82	3406090	Perchlorate_83
13.806	1040769	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_058.d	**Sample Name**	PER_CCV 0.002 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	9:15
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.820	3406090	7264937
Perchlorate_102	PER_IS_108	13.806	1040769	7264937

Perchlorate_100



Perchlorate_102



CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_069.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:12:03				Date of Analysis: 05/24/15 Time of Analysis: 13:00 Standard Id: PER_CCV (Standard) 0.0004ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	13.42	1.474319	N	16.1
Perchlorate_85	13.70	13.60	13.80	0.391605	13.43	0.491206	N	25.4

FILE ID : TQCK2369.RFB

Data File ID: T523_069.d

Date Injected : 05/24/15

Time Injected : 13:00

Sample ID : PER_CCV 0.0004 ug/ml 05/07/15

Client ID : NA

Retention Time	Area Count Response	Compound_ID Product Ion
13.406	4418362	PER_IS_89
13.423	521126	Perchlorate_83
13.429	173626	Perchlorate_85

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_069.d	**Sample Name**	PER_CCV 0.0004 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	13:00
ClientID	NA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.423	521126	4418362
Perchlorate_102	PER_IS_108	13.429	173626	4418362

Perchlorate_100







Perchlorate_102







CONTINUING CALIBRATION CHECK SUMMARY FORM

Lab Name: APPL Inc.

Detector ID: Agilent 6460 Triple Quad LC/MS DATA FILE ID: T523_070.d

	Date (s) of Analysis From:03/12/15 To:05/24/15 Time (s) of Analysis From:17:52 To:12:03				Date of Analysis: 05/24/15 Time of Analysis: 13:18 Standard Id: PER_CCV (Standard) 0.002 ug/ml			
COMPOUND	RT	RT WINDOW FROM	TO	AVERAGE RELATIVE RESPONSE FACTOR	RT	CALCULATED RELATIVE RESPONSE FACTOR	QNT Y/N	%D
Perchlorate_83	13.79	13.69	13.89	1.270255	13.50	1.175265	N	7.5
Perchlorate_85	13.70	13.60	13.80	0.391605	13.52	0.369265	N	5.7

FILE ID : TQCK2370.RFB

Data File ID: T523_070.d

Date Injected : 05/24/15

Time Injected : 13:18

Sample ID : PER_CCV 0.002 ug/ml 05/07/15

Client ID : NA

Retention Time		Area Count Response	Compound_ID Product Ion
13.497		5503588	PER_IS_89
13.504	*	2587270	Perchlorate_83
13.521		812912	Perchlorate_85

* MANUAL INTEGRAION

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_070.d	**Sample Name**	PER_CCV 0.002 ug/ml 05/07/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	13:18
ClientID	NA	**Inj Vol**	20



Compound	ISTD		RT	Resp	ISTD Resp	
Perchlorate_100	PER_IS_108	*	13.504	2587270	5503588	* MANUAL INT
Perchlorate_102	PER_IS_108		13.521	812912	5503588	

Perchlorate_100



Perchlorate_102




- MRM (99.0 -> 83.0) T523_070.d
Counts
x10 5
13.504 min.
Acquisition Time (min)


- MRM (99.0 -> 83.0) T523_070.d
Counts
x10 5
*13.504 min.
Acquisition Time (min)
BR
M-1
06-16-15
PM6/16/15

EPA METHOD 6850
Perchlorate
LC/MS

Raw Data

APPL, INC.

Method Blank
PERCHLORATE EPA 6850 - SOIL

Blank Name/QCG: **150518S-16137 - 198067**
Batch ID: #6850SM-150518A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	PERCHLORATE	0.0040 U	0.006	0.0040	0.0020	mg/Kg	05/18/15	05/24/15

Quant Method: QTLMFL2
Run #: T523_049.d
Instrument: AGIL_6460
Sequence: TQ052315
Initials: RP

GC SC-Blank-REG MDLs
Printed: 06/17/15 10:45:59 AM

Data File ID: T523_049.d

Date Injected : 05/24/15

Time Injected : 6:28

Sample ID : 150518SBLKA 10319.9 DF 05/18/15

Client ID : LAB_CNTL_S_BLKA

Retention Time	Area Count Response	Compound_ID Product Ion		
14.251	5786126	PER_IS_89		
12.995	7316	Perchlorate_83	(7316 * 0.0050) / (1.27 * 5786126.00) * 10319.90 =	0.051362 ppb
13.052	5610	Perchlorate_85	(5610 * 0.0050) / (0.39 * 5786126.00) * 10319.90 =	0.127753 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_049.d	**Sample Name**	150518SBLKA 10319.9 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	6:28
ClientID	LAB_CNTL_S_BLKA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	12.995	7316	5786126
Perchlorate_102	PER_IS_108	13.052	5610	5786126

Perchlorate_100







Perchlorate_102







Laboratory Control Spike Recovery
PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 LCS - 198067**
Batch ID: #6850SM-150518A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits
PERCHLORATE	0.00594	0.00556	93.6	80-120

Comments: _____

Primary	SPK
Quant Method :	QTLMFL2
Extraction Date :	05/18/15
Analysis Date :	05/24/15
Instrument :	AGIL_6460
Run :	T523_047.d
Initials :	RP

Printed: 06/17/15 10:45:45 AM
APPL Standard LCS

Data File ID: T523_047.d

Date Injected : 05/24/15

Time Injected : 5:50

Sample ID : 150518SA_LCS-1 10422.1 DF 05/18/15

Client ID : LAB_CNTL_S_SPKA

Retention Time	Area Count Response	Compound_ID Product Ion		
14.262	5484178	PER_IS_89		
14.248	742644	Perchlorate_83	(742644 * 0.0050) / (1.27 * 5484178.00) * 10422.10 =	5.555247 ppb
14.245	256573	Perchlorate_85	(256573 * 0.0050) / (0.39 * 5484178.00) * 10422.10 =	6.225531 ppb

$$\frac{(10422.1)(742644)(0.005)}{(1.270255)(5484178)} = 5.555248$$

6/17/15
PN

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_047.d	**Sample Name**	150518SA_LCS-1 10422.1 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	5:50
ClientID	LAB_CNTL_S_SPKA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.248	742644	5484178
Perchlorate_102	PER_IS_108	14.245	256573	5484178

Perchlorate_100







Perchlorate_102







Matrix Spike Recoveries

PERCHLORATE EPA 6850 - SOIL

APPL ID: **150518S-16137 MS - 198067**
Batch ID: #6850SM-150518A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Lvl mg/Kg	Matrix Result mg/Kg	SPK Result mg/Kg	DUP Result mg/Kg	SPK % Recovery	DUP % Recovery	Recovery Limits	RPD %	RPD Limits
PERCHLORATE	0.00508	ND	0.00496	0.00529	97.6	104	80-120	6.4	15

Comments: _____

Primary	SPK	DUP
Quant Method :	QTLMFL2	QTLMFL2
Extraction Date :	05/18/15	05/18/15
Analysis Date :	05/24/15	05/24/15
Instrument :	AGIL_6460	AGIL_6460
Run :	T523_051.d	T523_052.d
Initials :	RP	

Printed: 06/17/15 10:53:01 AM
APPL MSD SCII

Data File ID: T523_051.d

Date Injected : 05/24/15

Time Injected : 7:05

Sample ID : AZ16137_S01_MS-1 8920.6 DF 05/18/15

Client ID : S67-SS030-0001_MS-1

Retention Time		Area Count Response	Compound_ID Product Ion			
14.139		5651771	PER_IS_89			
14.126		798462	Perchlorate_83	(798462 * 0.0050) / (1.27 * 5651771.00) * 8920.60 =		4.960698 ppb
14.132	*	261182	Perchlorate_85	(261182 * 0.0050) / (0.39 * 5651771.00) * 8920.60 =		5.263499 ppb

* MANUAL INTEGRAION

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_051.d	**Sample Name**	AZ16137_S01_MS-1 8920.6 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	7:05
ClientID	S67-SS030-0001_MS-1	**Inj Vol**	20



Compound	ISTD		RT	Resp	ISTD Resp	
Perchlorate_100	PER_IS_108		14.126	798462	5651771	
Perchlorate_102	PER_IS_108	*	14.132	261182	5651771	* MANUAL INT

Perchlorate_100







Perchlorate_102








- MRM (101.0 -> 84.9) T523_051.d
Counts
x10 4
14.163 min.
Acquisition Time (min)


- MRM (101.0 -> 84.9) T523_051.d
Counts
x10 4
*14.132 min.
BF
M-4
06-16-15
PM6/16/15
Acquisition Time (min)

Data File ID: T523_052.d

Date Injected : 05/24/15

Time Injected : 7:24

Sample ID : AZ16137_S01_MSD-1 8996.9 DF 05/18/15

Client ID : S67-SS030-0001_MSD-1

Retention Time	Area Count Response	Compound_ID Product Ion			
14.078	5415076	PER_IS_89			
14.064	809460	Perchlorate_83	(809460 * 0.0050) / (1.27 * 5415076.00) *	8996.90 =	5.293741 ppb
14.071	260591	Perchlorate_85	(260591 * 0.0050) / (0.39 * 5415076.00) *	8996.90 =	5.528019 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_052.d	**Sample Name**	AZ16137_S01_MSD-1 8996.9 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	7:24
ClientID	S67-SS030-0001_MSD-1	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	14.064	809460	5415076
Perchlorate_102	PER_IS_108	14.071	260591	5415076

Perchlorate_100



Perchlorate_102



Interferrence Check Sample

Perchlorate EPA 6850 Soil

APPL ID: 150518S-16137 ICS - 198067
Batch ID: #6850SM-150518A

APPL Inc.
908 N. Temperance Ave.
Clovis, CA 93711

Compound Name	Spike Level mg/Kg	Spk Result mg/Kg	SPK% Recovery	Recovery Limits
Perchlorate	0.00571	0.00599	104.9%	70-130

Comments:

Quant Method:	QTLMFL2
Extraction Date:	05/18/15
Analysis Date:	05/24/15
Instrument:	AGIL_6560
Run:	T523_048.D
Initials:	RP

Data File ID: T523_048.d

Date Injected : 05/24/15

Time Injected : 6:09

Sample ID : 150518S_ICSA 10025.1 DF 05/18/15

Client ID : LAB_CNTL_S_ICSA

Retention Time	Area Count Response	Compound_ID Product Ion		
13.273	4353871	PER_IS_89		
13.26	660527	Perchlorate_83	(660527 * 0.0050) / (1.27 * 4353871.00) * 10025.10 =	5.986634 ppb
13.246	295223	Perchlorate_85	(295223 * 0.0050) / (0.39 * 4353871.00) * 10025.10 =	8.679308 ppb

Quantitative Analysis Sample Report

Batch Data Path	D:\MassHunter\Data\150523\QuantResults\150523_A2_MI.batch.bin		
Instrument	LCMS QQQ	**Operator**	ba
Data File	T523_048.d	**Sample Name**	150518S_ICSA 10025.1 DF 05/18/15
Sample Type	Sample	**Acq_Date**	05/24/15
Acq Method	6460_ESI_PER_N_NEWER_K'_COLUMN.m	**Acq Time**	6:09
ClientID	LAB_CNTL_S_ICSA	**Inj Vol**	20



Compound	ISTD	RT	Resp	ISTD Resp
Perchlorate_100	PER_IS_108	13.260	660527	4353871
Perchlorate_102	PER_IS_108	13.246	295223	4353871

Perchlorate_100



Perchlorate_102



ACN_H20_50:50 1 DF 08/27/14: Burdick & Jackson Acetonitrile Lot DL308 took 10.000 ml and added to 10.000 ml Laboratory Source DI_H20 Note : Above solution was prepared as needed.
PER_IS_50:50 0.005 ug/ml 11/11/14: Took 19.000 ml ACN_H20_50:50 11/11/14 and added 1.000 ml O2SI 18_O_4 Perchlorate 100 ug/L lot 1059192-33479 Note: Solution prepared as needed.
PERCHLORATE 0.005 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/28/14 took 0.500 ml and added to 0.500 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PERCHLORATE 0.002 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PERCHLORATE 0.001 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.100 ml and added to 0.900 ml PER_IS_50:50 0.005 ug/ml 11/11/14.

PERCHLORATE 0.0004 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PERCHLORATE 0.0002 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.020 ml and added to 0.980 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PERCHLORATE 0.0001 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.010 ml and added to 0.990 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_1 0.0004 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_1 0.002 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_2 0.0004 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.

PER_CCV_2 0.002 ug/ml 08/27/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_3 0.0004 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_3 0.002 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/28/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_4 0.0004 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_CCV_4 0.002 ug/ml 11/11/14: PERCHLORATE 0.010 ug/ml 08/27/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_SS 0.0004 ug/ml 11/11/14: PER_SS 0.010 ug/ml 08/28/14 took 0.040 ml and added to 0.960 ml PER_IS_50:50 0.005 ug/ml 11/11/14.
PER_SS 0.002 ug/ml 11/11/14: PER_SS 0.010 ug/ml 08/28/14 took 0.200 ml and added to 0.800 ml PER_IS_50:50 0.005 ug/ml 11/11/14.

LCMS STANDARD PREP LOG# 103 PAGE 34

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT#	DATE / INITIALS
Sea V_spk 250 ng/mL							
Hexahydro-1,3,5-tris-hydroxyethyltriazine	1010	04-10-15	1.24	5.0	250	ACN/DM394	04-17-15
2,2'-thiodiethanol	1090	04-10-15	1.15				
2-(1,3,5-Dithiazinan-5-yl)ethanol	1020	04-10-15	1.23				

PER_CCV 0.0004 µg/mL 04-20-15:
PER_SS 0.010 µg/mL 08-28-14
took 0.040 mL and added to
0.910 mL laboratory Source DI-H2O
that had previously received
0.050 mL 02SI $^{18}O_4$ Perchlorate
100 µg/L lot 1059192-33479

PER_CCV 0.002 µg/mL 04-20-15:
PER_SS 0.010 µg/mL 08-28-14
took 0.200 mL and added to
0.750 mL Laboratory Source DI-H2O
that had previously received
0.050 mL 02SI $^{18}O_4$ Perchlorate
100 µg/L lot 1059192-33479

040 LC/MS STANDARD PREP LOG# 105 PAGE 40

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE/ INITIALS
PER IS	0.005µg/mL	04-20-15:					SL 04-21-15
02SI $^{18}O_4$ Perchlorate	100µg/L						
Lot 1059192-33479 took 0.050mL and added to 0.950 mL Laboratory Source DI H2O.							
NOTE: Repeated the above process four times in order to prepare 2-2mL vials that were full. The above was performed as dated. JF							

STANDARD	INITIAL CONC	SOURCE	DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS

NOTE: THE FOLLOWING DOCUMENTATION JL
WAS PERFORMED AS DATED: 05-19-15

PER. CCV 0.0004 µg/ml 05:07-15: JL
PER. SS 0.010 µg/ml 08-28-14 took 05.07.15
0.040 ml and added to 0.910 ml
Laboratory Source DI-H20 that
had previously received 0.050 ml
02SI $^{18}O_4$ Perchlorate 100µg/L
lot 1074176-35163

PER. CCV 0.002 µg/ml 05-07-15:
PER. SS 0.010 µg/ml 08-28-14
took 0.200 ml and added to
0.750 ML Laboratory Source
DI-H20 that had previously
received 0.050 ml 02SI $^{18}O_4$
Perchlorate 100µg/L Lot
1074176-35163

PER. IS 0.005µg/ml 05-07-15:
02SI $^{18}O_4$ Perchlorate 100µg/L
Lot 1074176-35163 took 0.050
ml and added to 0.950 ml
Laboratory Source DI-H20
NOTE: Repeated the above
process four times in order
to prepare 2-2ml vials
that were full.

076

LC/MS STANDARD PREP LOG# 105 PAGE# 76

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE/ INITIALS
NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED: 06-02-15							JA 06-02-15
PER_IS_50:50 0.005µg/mL 05-10-15: COMBINED 10.000 mL Laboratory Source DI_H2O with 10.000 mL Burdick & Jackson Acetonitrile lot DL308. Withdrew 1.000mL and then delivered 1.000 mL 025I $^{18}O_4$ Perchlorate 100µg/L lot 10T4176-35163							JL 05-10-15
AZ16057_W01 10526.3 DF 05-07-15_5/10: AZ16057_W01 1052.6 DF 05-07-15 TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							
AZ16057_W01 105263.2 DF 05-07-15_5/10: AZ16057_W01 10526.3 DF 05-07-15_5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							
AZ16057_W01 1052631.6 DF 05-07-15_5/10: AZ16057_W01 105263.2 DF 05-07-15_5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							

LCMS STANDARD PREP LOG# 105 PAGE# 77

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT VOLUME	FINAL CONC	FINAL LOT #	SOLVENT(S)	DATE: INITIALS

NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED! 06.02.15 06.02.15

AZ16057_W01 10526315.8 DF 05-07-15_5/10: 05.10.15
AZ16057_W01 1052631.6 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005 µg/mL
05-10-15

AZ16059_W01 10526.3 DF 05-07-15_5/10:
AZ16059_W01 1052.6 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005 µg/mL
05-10-15

AZ16059_W01 105263.2 DF 05-07-15_5/10:
AZ16059_W01 10526.3 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005 µg/mL
05-10-15

AZ16059_W01 1052631.6 DF 05-07-15_5/10:
AZ16059_W01 105263.2 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005 µg/mL
05-10-15

AZ16059_W01 10526315.8 DF 05-07-15_5/10:
AZ16059_W01 1052631.6 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005 µg/mL
05-10-15

078

LC/MS STANDARD PREP LOG# 105 PAGE# 78

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE / INITIALS
NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED. JR 06.02.15							JR 06.02.15
AZ16061_IW01 10526.3 DF 05-07-15_5/10; AZ16061_IW01 16526 DF 05-07-15 TOOK 0.050mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							JR 05.10.15
AZ16061_IW01 105263.2 DF 05-07-15_5/10; AZ16061_IW01 10526.3 DF 05-07-15_5/10 TOOK 0.050mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							
AZ16061_IW01 1052631.6 DF 05-07-15_5/10; AZ16061_IW01 105263.2 DF 05-07-15_5/10 TOOK 0.050mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							
AZ16061_IW01 10526315.8 DF 05-07-15_5/10 AZ16061_IW01 1052631.6 DF 05-07-15_5/10 TOOK 0.050mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15							

LCMS STANDARD PREP LOG# 105 PAGE 79

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT # 079	DATE / INITIAL

NOTE: THE FOLLOWING DOCUMENTATION JH
WAS PERFORMED AS DATED: JH 06-02-15 06-02-15

5 AZ16062-W01 10526.3 DF 06-07-15_5/10: JH
AZ16062-W01 1052.6 DF 05-07-15_5/10 05-10-15
TOOK 0.050 mL AND ADDED TO 0.450
15 mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16062_W01 105263.2 DF 05-07-15_5/10:
AZ16062-W01 10526.3 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16062_W01 1052631.6 DF 05-07-15_5/10:
AZ16062-W01 105263.2 DF 05-07-15_5/10
~~TOOK 0.450 mL~~ JH 06-02-15 TOOK 0.050 mL
AND ADDED TO 0.450 mL PER_IS_50:50
0.005 µg/mL 05-10-15

AZ16062_W01 10526315.8 DF 05-07-15_5/10:
AZ16062-W01 1052631.6 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO 0.450
mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16062_W01_MS-1 10526.3 DF 05-07-15_5/10
AZ16062_W01_MS-1 1052.6 DF 05-07-15
TOOK 0.050 mL AND ADDED TO 0.450
mL PER_IS_50:50 0.005µg/mL 05-10-15

080

LC/MS STANDARD PREP LOG# 105 PAGES 80

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT / LOT#	DATE / INITIALS
NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED: JB 06-02-15							JB 06-02-15
AZ16062-WO1-MS-1 105263.2 DF 05-07-15_5/10 AZ16062-WO1-MS-1 10526.3 DF 05-07-15_5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER IS 50/50 0.005 µg/mL 05-10-15							JB 05-10
AZ16062-WO1-MS-1 105263.6 DF 05-07-15_5/10; AZ16062-WO1-MS-1 105263.2 DF 05-07-15_5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER IS 50:50 0.005 µg/mL 05-10-15							
AZ16062-WO1-MS-1 10526315.8 DF 05-07-15_5/10 AZ16062-WO1-MS-1 105263.6 DF 05-07-15 5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER IS 50:50 0.005 µg/mL 05-10-15							
AZ16062-WO1-MSD-1 10526.3 DF 05-07-15_5/10 AZ16062-WO1-MSD-1 1052.6 DF 05-07-15 TOOK 0.050 mL AND ADDED TO 0.450 mL PER IS 50:50 0.005 µg/mL 05-10-15							
AZ16062-WO1-MSD-1 105263.2 DF 05-07-15_5/10 AZ16062-WO1-MSD-1 10526.3 DF 05-07-15_5/10 TOOK 0.050 mL AND ADDED TO 0.450 mL PER IS 50:50 0.005 µg/mL 05-10-15							

LCMS STANDARD PREP LOG# 105 PAGE# 81

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT#	DATE INITIALS

NOTE: THE FOLLOWING DOCUMENTATION JH
WAS PERFORMED AS DATED: JH 06-02-15 06-02-15

AZ16062-W01_MSD-1 1052631.6 DF 05-07-15_5/10 JH
AZ16062-W01_MSD-1 105263.2 DF 05-07-15_5/10 05-10-15
TOOK 0.050 mL AND ADDED TO 0.450
mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16062-W01_MSD-1 10526315.8 DF 05-07-15_5/10
AZ16062_W01_MSD-1 1052631.6 DF
05-07-15_5/10 TOOK 0.050 mL AND ADDED
TO 0.450 mL PER_IS_50:50 0.005µg/mL
05-10-15

AZ16058_W01 10526.3 DF 05-07-15_5/10:
AZ16058_W01 1052.6 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO 0.450
mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16058_W01 105263.2 DF 05-07-15_5/10:
AZ16058_W01 10526.3 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO 0.450
mL PER_IS_50:50 0.005µg/mL 05-10-15

AZ16058_W01 1052631.6 DF 05-07-15_5/10:
AZ16058_W01 105263.2 DF 05-07-15_5/10
TOOK 0.050 mL AND ADDED TO
0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

LC/MS STANDARD PREP LOG# 105 PAGE# 82

082

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CNC	SOL'N. LOT #	DATE/INITIAL

NOTE! THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED JH 06-02-15 JH

06-02-15

AZ16060_W01 10526.3 DF 05-07-15-5/10: JH

AZ16060_W01 1052.6 DF 05-07-15 05-10-15

TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

5/11 JH

~~AZ16055_W01 10526.3 DF 05-07-15_5/10:~~ JH

AZ16055_W01 1052.6 DF 05-07-15 05-11-15

TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

5/11 JH

AZ16055_W01 105263.2 DF 05-07-15_~~5/10~~:

AZ16055_W01 10526.3 DF 05-07-15_5/10

TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

5/11 JH

AZ16055_W01 1052631.6 DF 05-07-15_~~5/10~~

AZ16055_W01 105263.2 DF 05-07-15_5/10

TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

JH 5/11

AZ16056_W01 10526.3 DF 05-07-15_~~5/10~~:

AZ16056_W01 1052.6 DF 05-07-15

TOOK 0.050 mL AND ADDED TO 0.450 mL PER_IS_50:50 0.005µg/mL 05-10-15

LCMS STANDARD PREP LOG# 105 PAGE 83

083

	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT #	DATE INITIALS

NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED ~~BL~~ 06-02-15 JH 06-02-15

AZ16056-WO1 105263.2 DF 05-07-15 ~~5/10~~ 5/11: JH
AZ16056-WO1 10526.3 DF 05-07-15 ~~5/10~~ 5/11 05-11-15
TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005 µg/ml 05-10-15

AZ16056-WO1 1052631.6 DF 05-07-15 ~~5/10~~ ~~BL~~ 5/11:
AZ16056-WO1 105263.2 DF 05-07-15 ~~5/10~~ 5/11
TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005 µg/ml 05-10-15

AZ16136-WO1 H210.H DF 05-08-15 5/11:
AZ16136-WO1 1052.6 DF 05-08-15
TOOK 0.050 ml AND ADDED TO 0.150 ml PER_IS_50150 0.005 µg/ml 05-10-15

AZ16237-WO1 ~~1052.6~~ 10526.3 ~~BL~~ DF 05-12-15 5/13: JH
AZ16237-WO1 1052.6 DF 05-12-15 05-13-15
TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005 µg/ml 05-10-15

AZ16237-WO1 105263.2 DF 05-12-15 5/13
AZ16237-WO1 10526.3 DF 05-12-15 5/13
TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005 µg/ml 05-10-15

084 LC/MS STANDARD PREP LOG# 105 PAGE# 84

STANDARD	INITIAL CONC	SOURCE	DATE	AMOUNT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	DATE / INITIALS

NOTE: THE FOLLOWING DOCUMENTATION WAS JL
PERFORMED AS DATED: JL 06-02-15 06-02-15

AZ16237_WO1 1052631.6 DF 05-12-15_5-13 JL
AZ16237_WO1 105263.2 DF 05-12-15_5/13 05-13-15
TOOK 0.050 ml AND ADDED TO 0.450
ml PER_IS_50:50 0.005 µg/ml
05-10-15

AZ16237_WO1 10526315.8 DF 05-12-15_5/13
AZ16237_WO1 1052631.6 DF 05-12-15_5/13
TOOK 0.050 ml AND ADDED TO 0.450
ml PER_IS_50:50 0.005 µg/ml
05-10-15

AZ16229_WO1 10526.3 DF 05-12-15_5/13:
AZ16229_WO1 1052.6 DF 05-12-15
TOOK 0.050 ml AND ADDED TO
~~0.05~~ JL 0.450 ml PER_IS_50:50
0.005 µg/ml 05-10-15

AZ16229_WO1 105263.2 DF 05-12-15_5/13
AZ16229_WO1 10526.3 DF 05-12-15_5/13
TOOK 0.050 ml AND ADDED TO
0.450 ml PER_IS_50:50 0.005 µg/ml
05-10-15

LC/MS STANDARD PREP LOG# 105 PAGE 85

085

STANDARD	INITIAL CONC	SOURCE	DATE	ALIQUOT VOLUME	FINAL CONC	SOL. SOL. LOT #	DATE INITIALS
NOTE: THE FOLLOWING DOCUMENTATION							JH
WAS PERFORMED AS DATED: JH 06-02-15							06-02-15
AZ16229-W01 1052631.6 DF 05-12-15_5/13: JH							
AZ16229-W01 105263.2 DF 05-12-15_5/13							05-13-15
TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005 µg/ml							05-10-15

086

LC/MS STANDARD PREP LOG# 105 PAGE# 86

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CNC	SOL'N LOT#	DATE & INITIALS

NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED: 06-09-15 06-08- JL

AZ16230_W01 . 105263.2 DF 05-12-15_5/14 & AZ16230.W01 10526.3 DF 05-12-15_5/14 TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005µg/ml 05-14-15 05-14- JL

AZ16230_W01 1052631.6 DF 05-12-15_5/14 & AZ16230_W01 105263.2 DF 05-12-15_5/14 TOOK 0.050 ml AND ADDED TO ~~0.005 µg/ml~~ 0.450 ml PER_IS_50:50 0.005 µg/ml 05-14-15

AZ16230_W01 10526315.8 DF 05-12-15_5/14: AZ16230_W01 1052631.6 DF 05-12-15 5/14 TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005µg/ml 05-14-15

AZ16236_W01 10526.3 DF 05-12-15_5/14: AZ16236_W01 1052.6 DF 05-12-15 TOOK 0.050 ml AND ADDED TO 0.450 ml PER_IS_50:50 0.005µg/ml 05-14-15

AZ16236_W01 105263.2 DF 05-12-15_5/14: AZ16236_W01 10526.3 DF 05-12-15_5/14 TOOK 0.050ML AND ADDED TO 0.450 ml PER

LCMS STANDARD PREP LOG# 105 PAGE 87

087

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS

NOTE: THE FOLLOWING DOCUMENTATIONS WAS PERFORMED AS DATED! JF 06.08.15 — JR 06.08.15

AZ16236_W01 ~~105233~~ 105263.6 DF — JR 05.14.15
05.12.15_5/14:
AZ16236_W01 105263.2 DF 05.12.15_5/14
TOOK 0.050 ml AND ADDED TO
0.450 ml PERIS_50:50 0.005 µg/ml
05.14.15

AZ16228_W01 10526.3 DF 05.12.15_5/14:
AZ16228_W01 1052.6 DF 05-12-15 TOOK
0.050 ml AND ADDED TO 0.450 ml
PER_IS_50:50 0.005 µg/ml 05.14.15:

~~AZ16228_W01~~ JF
AZ16231_W01 10526.3 DF 05.12.15_5/14:
AZ16231_W01 1052.6 DF 05.12.15
TOOK 0.050 ml AND ADDED TO
0.450 ml PER_IS_50:50 0.005 µg/ml
05-14-15.

AZ16233_W01 10526.3 DF 05.12.15_5/14:
AZ16233_W01 1052.6 DF 05.12.15 TOOK
0.050 ml AND ADDED TO 0.450 ml
PER_IS_50:50 0.005 µg/ml 05.14.15.

AZ16233_W01 105263.2 DF 05.12.15_5/14:
AZ16233_W01 10526.3 DF 05.12.15_5/14
TOOK 0.050 ml AND ADDED TO 0.450 ml
PER_IS_50:50 0.005 µg/ml 05.14.15

088 LC/MS STANDARD PREP LOG# 105 PAGE 88

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CNC	SOL'N. LOT #	DATE/ INITIALS
NOTE: THE FOLLOWING DOCUMENTATION							JR
WAS PERFORMED AS DATED: JR							06-08-15
AZ16231_W01 105263.2 DF 05-12-15_5/15:							JR
AZ16231-W01 10526.3 DF 05:12:15_5/14							05-15-15
TOOK 0.050 mL AND ADDED TO 0.450 mL							
PER_IS_50:50 0.005 µg/mL 05-14-15							
AZ16231_W01 1052631.6 AF 05-12-15_5/15:							
AZ16231_W01 105263.2 DF 05-12-15_5/15							
TOOK 0.050 mL AND ADDED TO 0.450							
mL PER_IS_50:50 0.005 µg/mL							
05-14-15							

_HPLC6850G

Organic Extraction Worksheet

Method	EPA 6850 HPLC Perchlorate Extraction	Extraction Set	150518A	Extraction Method	HPLC6850GROS	Units	mL

Spiked ID 1	Perchlorate Spike Mix 0.1ppm 8-28-14 exp 7-31-15		Surrogate ID 1		
Spiked ID 2	Mix Anions 25,000ppm 2-10-15 exp 2-10-16		Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	NA	
Spiked ID 7			Ext. Start Time:	05/18/15 12:00	
Spiked ID 8			Ext. End Time:	05/18/15 13:48	
			GC Requires Extract By:		

pH1			Water Bath Temp Criteria	
pH2				
pH3				

Spiked By: CFM **Date** 05/18/15 11:52:00 AM **Witnessed By:** DL **Date** 05/18/15 11:52:00 AM

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	150518A Blk				NA	NA equip	1.02g	10ML	NA	05/18/15 12:00	
2	150518A LCS-1		0.060	1	NA	NA equip	1.01g	10ML	NA	05/18/15 12:00	
3	150518A ICS		0.060/0.120	1/2	NA	NA equip	1.05g	10ML	NA	05/18/15 12:00	
4	AZ16137 MS-1	AZ16137S01	0.060		NA	NA equip	1.18g	10ML	NA	05/18/15 12:00	76380
5	AZ16137 MSD-1	AZ16137S01	0.060		NA	NA equip	1.17g	10ML	NA	05/18/15 12:00	76380
6	AZ16137	AZ16137S01			NA	NA equip	1.15g	10ML	NA	05/18/15 12:00	76380
7	AZ16138	AZ16138S01			NA	NA equip	1.16g	10ML	NA	05/18/15 12:00	76380
8	AZ16139	AZ16139S01			NA	NA equip	1.22g	10ML	NA	05/18/15 12:00	76380
9	AZ16140	AZ16140S01			NA	NA equip	1.15g	10ML	NA	05/18/15 12:00	76380
10	AZ16141	AZ16141S01			NA	NA equip	1.37g	10ML	NA	05/18/15 12:00	76380
11	AZ16198	AZ16198S01			NA	NA equip	1.22g	10ML	NA	05/18/15 12:00	76390
12	AZ16199	AZ16199S01			NA	NA equip	1.26g	10ML	NA	05/18/15 12:00	76390
13	AZ16201	AZ16201S01			NA	NA equip	1.25g	10ML	NA	05/18/15 12:00	76390
14	AZ16202	AZ16202S01			NA	NA equip	1.24g	10ML	NA	05/18/15 12:00	76390
15	AZ16400	AZ16400S01			NA	NA equip	1.33g	10ML	NA	05/18/15 12:00	76447
16	AZ16401	AZ16401S01			NA	NA equip	1.30g	10ML	NA	05/18/15 12:00	76447

lvent and Lot#	
WATER	5-18-15
ND	3223C517

Extraction COC Transfer	
Extraction lab employee Initials	CFM
GC analyst's initials	PM
Date	5/18/15
Time	13:00
Refrigerator	NA

	Technician's Initials
Scanned By	CFM
Sample Preparation	CFM
Extraction	CFM
Concentration	----------

Modified	05/18/15 12:07:37 PM

Reviewed By: PM **Date** 5/18/15

18/15 1:49:44 PM

Ext_ID 47806

Organic Extraction Worksheet

Method	EPA 6850 HPLC Perchlorate Extraction	Extraction Set	150518A	Extraction Method	HPLC6850GROS	Units	mL

Spiked ID 1	Perchlorate Spike Mix 0.1ppm 8-28-14 exp 7-31-15		Surrogate ID 1		
Spiked ID 2	Mix Anions 25,000ppm 2-10-15 exp 2-10-16		Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	NA	
Spiked ID 7			Ext. Start Time:	05/18/15 12:00	
Spiked ID 8			Ext. End Time:	05/18/15 13:48	
			GC Requires Extract By:		

pH1			Water Bath Temp Criteria	
pH2				
pH3				

Spiked By: CFM **Date** 05/18/15 11:52:00 AM **Witnessed By:** DL **Date** 05/18/15 11:52:00 AM

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
17	AZ16402	AZ16402S01			NA	NA	1.33g	10ML	NA	05/18/15 12:00	76447
						equip					

PM 5/18/15

ent and Lot#	
/ATER	5-18-15
D	3223C517

Extraction COC Transfer	
Extraction lab employee Initials	CFM
GC analyst's initials	PM
Date	5/18/15
Time	13:00
Refrigerator	NA

	Technician's Initials
Scanned By	CFM
Sample Preparation	CFM
Extraction	CFM
Concentration	----------

Modified	05/18/15 12:07:37 PM

Reviewed By: PM **Date** 5/18/15

/15 1:49:44 PM Ext_ID 47806

METHOD 6850 INJECTION LOG: 03/12/15, 05/23/15

DETECTOR ID: Agilent 6460 Triple Quad LC/MS

SAMPLE INJECTION ID:	DATA FILE ID:	INJECTION DATE/TIME
PERCHLORATE 0.0002 ug/ml 11/11/14	T312_003.d	03/12/15 17:52
PERCHLORATE 0.0004 ug/ml 11/11/14	T312_004.d	03/12/15 18:11
PERCHLORATE 0.001 ug/ml 11/11/14	T312_005.d	03/12/15 18:29
PERCHLORATE 0.002 ug/ml 11/11/14	T312_006.d	03/12/15 18:48
PERCHLORATE 0.005 ug/ml 11/11/14	T312_007.d	03/12/15 19:07
PERCHLORATE 0.010 ug/ml 08/27/14	T312_008.d	03/12/15 19:25
PER_SS 0.0004 ug/ml 11/11/14	T312_010.d	03/12/15 20:02
PER_SS 0.002 ug/ml 11/11/14	T312_011.d	03/12/15 20:21
PER_CCV_1 0.0004 ug/ml 11/11/14	T312_013.d	03/12/15 20:58
PER_CCV_1 0.002 ug/ml 11/11/14	T312_014.d	03/12/15 21:17
PER_CCV 0.0004 ug/ml 05/07/15	T523_002.d	05/23/15 15:38
PER_CCV 0.002 ug/ml 05/07/15	T523_003.d	05/23/15 15:57
PER_CCV 0.0004 ug/ml 05/07/15	T523_044.d	05/24/15 4:55
PER_CCV 0.002 ug/ml 05/07/15	T523_045.d	05/24/15 5:13
150518SA_LCS-1 10422.1 DF 05/18/15	T523_047.d	05/24/15 5:50
150518S_ICSA 10025.1 DF 05/18/15	T523_048.d	05/24/15 6:09
150518SBLKA 10319.9 DF 05/18/15	T523_049.d	05/24/15 6:28
AZ16137-S01 9153.3 DF 05/18/15	T523_050.d	05/24/15 6:46
AZ16137_S01_MS-1 8920.6 DF 05/18/15	T523_051.d	05/24/15 7:05
AZ16137_S01_MSD-1 8996.9 DF 05/18/15	T523_052.d	05/24/15 7:24
AZ16138_S01 9074.4 DF 05/18/15	T523_053.d	05/24/15 7:42
AZ16139_S01 8628.1 DF 05/18/15	T523_054.d	05/24/15 8:01
AZ16140_S01 9153.3 DF 05/18/15	T523_055.d	05/24/15 8:19
PER_CCV 0.0004 ug/ml 05/07/15	T523_057.d	05/24/15 8:57
PER_CCV 0.002 ug/ml 05/07/15	T523_058.d	05/24/15 9:15
AZ16141_S01 7683.4 DF 05/18/15	T523_060.d	05/24/15 9:53
PER_CCV 0.0004 ug/ml 05/07/15	T523_069.d	05/24/15 13:00
PER_CCV 0.002 ug/ml 05/07/15	T523_070.d	05/24/15 13:18

EXPLOSIVES
EPA METHOD 8330

APPL, INC.

EPA 8330
EXPLOSIVES

QC Summary

APPL, INC.

Method Blank
EPA 8330B SOIL

Blank Name/QCG: **150520S-16137 - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
BLANK	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
BLANK	SURROGATE: 1,2-DINITROBENZ	95.0	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000051
Instrument: Waldorf
Sequence: 150522
Initials: MP

GC SC-Blank-REG MDLs
Printed: 05/26/15 10:53:01 AM

Form 2 & 8

Surrogate Recovery

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/23/15

Matrix: SOIL Instrument: Waldorf

APPL ID.	Client Sample No.	SURROGATE: 1,2-DINITROBENZENE (S)					
		Limits	Result	Qualifier	Limits	Result	Qualifier
150520A-LCS	Lab Control Spike	70-130	91.5				
AZ16137-MS	Matrix Spike	70-130	93.5				
AZ16137-MSD	Matrix SpikeD	70-130	93.5				
150520A-BLK	Blank	70-130	95.0				
AZ16137	S67-SS030-0001	70-130	96.4				
AZ16138	S67-SS030P-0001	70-130	97.6				
AZ16139	S67-SB030-0607	70-130	96.2				
AZ16140	S67-SS032-0001	70-130	95.5				
AZ16141	S67-SB032-0607	70-130	93.9				

Comments: Batch: #83BJU-150520A

Laboratory Control Spike Recovery
EPA 8330B SOIL

APPL ID: **150520S-16137 LCS - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2,4-DINITROTOLUENE	1.99	1.83	92.0	80-125
2,6-DINITROTOLUENE	1.99	1.84	92.5	80-120
HMX	1.99	1.90	95.5	75-125
NITROGLYCERIN	1.99	1.83	92.0	68-131
RDX	1.99	1.83	92.0	70-135
TETRYL	1.99	1.97	99.0	10-150
SURROGATE: 1,2-DINITROBENZENE (S	1.99	1.82	91.5	70-130

Comments: _____

Primary	SPK
Quant Method :	W150522.M
Extraction Date :	05/20/15
Analysis Date :	05/23/15
Instrument :	Waldorf
Run :	0522_000048
Initials :	MP

Printed: 05/26/15 10:52:50 AM
APPL Standard LCS

Matrix Spike Recoveries

EPA 8330B SOIL

APPL ID: **150520S-16137 MS - 197310**
Batch ID: #83BJU-150520A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	SPK Lvl	DUP Lvl	Matrix Result	SPK Result	DUP Result	SPK %	DUP %	Recovery	RPD	RPD
	mg/kg	mg/kg	mg/kg	mg/kg	mg/kg	Recovery	Recovery	Limits	%	Limits
2,4-DINITROTOLUENE	1.99	2.00	ND	1.86	1.85	93.5	92.5	80-125	1.1	20
2,6-DINITROTOLUENE	1.99	2.00	ND	1.87	1.86	94.0	93.0	80-120	1.1	20
HMX	1.99	2.00	ND	1.85	1.83	93.0	91.5	75-125	1.6	20
NITROGLYCERIN	1.99	2.00	ND	1.90	1.93	95.5	96.5	68-131	1	20
RDX	1.99	2.00	ND	1.69	1.69	84.9	84.5	70-135	0.47	20
TETRYL	1.99	2.00	ND	1.77	1.81	88.9	90.5	10-150	1.8	20
SURROGATE: 1,2-DINITROBENZEN	1.99	2.00	NA	1.86	1.86	93.5	93.0	70-130		

Comments: _____

Primary	SPK	DUP
Quant Method :	W150522.M	W150522.M
Extraction Date :	05/20/15	05/20/15
Analysis Date :	05/23/15	05/23/15
Instrument :	Waldorf	Waldorf
Run :	0522_000049	0522_000050
Initials :	MP	

Printed: 05/26/15 10:50:23 AM
APPL MSD SCII

EPA 8330B

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 05/23/15

Matrix: SOIL Instrument: Waldorf

Blank ID: 150520A-BLK Time Analyzed: 1225

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150520A-LCS	Lab Control Spike	0522_000048	05/23/15 1033
150520A-MS	Matrix Spike	0522_000049	05/23/15 1110
150520A-MSD	Matrix SpikeD	0522_000050	05/23/15 1148
150520A-BLK	Blank	0522_000051	05/23/15 1225
AZ16137	S67-SS030-0001	0522_000052	05/23/15 1303
AZ16138	S67-SS030P-0001	0522_000053	05/23/15 1340
AZ16139	S67-SB030-0607	0522_000054	05/23/15 1418
AZ16140	S67-SS032-0001	0522_000055	05/23/15 1455
AZ16141	S67-SB032-0607	0522_000056	05/23/15 1533

Comments: Batch: #83BJU-150520A

EPA 8330
EXPLOSIVES

Sample Data

APPL, INC.

EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SS030-0001

APPL ID: AZ16137

Sample Collection Date: 05/07/15

QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	96.4	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000052
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 05/26/15 10:53:05 AM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000052.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 13:03:20
Operator : MP
Sample : AZ16137S01 7.921 DF 05/20/15
Misc : Soil
ALS Vial : 4228 Sample Multiplier: 7.92079

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:32:14 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0m
m 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	3.538	0	935022	N.D.	572.353 #
Spiked Amount	594.059			Recovery	=	0.00%	96.35%
	Target Compounds						
1) TM	HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM	RDX	0.000	0.000	0	0	N.D.	N.D.
3) TM	1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5) TM	1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6) TM	3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
7) TM	NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
8) TM	NITROGLYC...	0.000	0.000	0	0	N.D. d<DL	N.D.
9) TM	TETRYL	6.442	0.000	2734	0	NoCal	N.D.
10) TM	2,4,6-TRI...	6.912	0.000	12731	0	NoCal	N.D.
11) TM	2-AMINO-4...	7.046	0.000	4368	0	NoCal	N.D.
12) TM	4-AMINO-2...	7.426	0.000	19528	0	NoCal	N.D.
13) TM	2,4-DINIT...	8.576	0.000	8543	0	NoCal	N.D.
14) TM	2,6-DINIT...	8.939	0.000	18505	0	NoCal	N.D.
15) TM	2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM	4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17) TM	3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18) TM	PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000052.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 13:03:20
Operator : MP
Sample : AZ16137S01 7.921 DF 05/20/15
Misc : Soil
ALS Vial : 4228 Sample Multiplier: 7.92079

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:32:14 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030P-0001

Sample Collection Date: 05/07/15

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 76380

APPL ID: AZ16138

QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	97.6	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000053
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 05/26/15 10:53:05 AM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000053.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 13:40:50
Operator : MP
Sample : AZ16138S01 7.976 DF 05/20/15
Misc : Soil
ALS Vial : 4229 Sample Multiplier: 7.97607

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:33:34 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	3.536	0	947513	N.D.	584.046 #
	Spiked Amount 598.205			Recovery =		0.00%	97.63%
	Target Compounds						
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.
2)	TM RDX	0.000	0.000	0	0	N.D.	N.D.
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
7)	TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D. d<DL	N.D.
9)	TM TETRYL	6.425	0.000	8904	0	NoCal	N.D.
10)	TM 2,4,6-TRI...	6.908	0.000	24983	0	NoCal	N.D.
11)	TM 2-AMINO-4...	7.072	0.000	17078	0	NoCal	N.D.
12)	TM 4-AMINO-2...	7.483	0.000	18502	0	NoCal	N.D.
13)	TM 2,4-DINIT...	8.636	0.000	8653	0	NoCal	N.D.
14)	TM 2,6-DINIT...	8.940	0.000	8264	0	NoCal	N.D.
15)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000053.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 13:40:50
Operator : MP
Sample : AZ16138S01 7.976 DF 05/20/15
Misc : Soil
ALS Vial : 4229 Sample Multiplier: 7.97607

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:33:34 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB030-0607
Sample Collection Date: 05/07/15

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 76380
APPL ID: AZ16139
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	96.2	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000054
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 05/26/15 10:53:05 AM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000054.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 14:18:20
Operator : MP
Sample : AZ16139S01 7.859 DF 05/20/15
Misc : Soil
ALS Vial : 4230 Sample Multiplier: 7.85855

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:34:31 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	3.538	0	933715	N.D.	567.061 #
Spiked Amount 589.391			Recovery	=	0.00%	96.21%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D.	N.D.
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6) TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
7) TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d<DL	N.D.
9) TM TETRYL	6.396	0.000	5219	0	NoCal	N.D.
10) TM 2,4,6-TRI...	6.916	0.000	4156	0	NoCal	N.D.
11) TM 2-AMINO-4...	7.133	0.000	6583	0	NoCal	N.D.
12) TM 4-AMINO-2...	7.413	0.000	19203	0	NoCal	N.D.
13) TM 2,4-DINIT...	8.601	0.000	10182	0	NoCal	N.D.
14) TM 2,6-DINIT...	8.968	0.000	1780	0	NoCal	N.D.
15) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000054.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 14:18:20
Operator : MP
Sample : AZ16139S01 7.859 DF 05/20/15
Misc : Soil
ALS Vial : 4230 Sample Multiplier: 7.85855

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:34:31 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16140
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	95.5	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000055
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 05/26/15 11:11:51 AM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000055.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 14:55:50
Operator : MP
Sample : AZ16140S01 7.921 DF 05/20/15
Misc : Soil
ALS Vial : 4231 Sample Multiplier: 7.92079

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:36:08 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb		ppb
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.537	0	926416	N.D.		567.084 #
	Spiked Amount 594.059			Recovery =		0.00%		95.46%
	Target Compounds							
1)	TM HMX	0.000	0.000	0	0	N.D.	<DL	N.D. d
2)	TM RDX	0.000	0.000	0	0	N.D.		N.D.
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.		N.D.
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.		N.D.
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.		N.D.
7)	TM NITROBENZENE	0.000	0.000	0	0	N.D.		N.D.
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D.	d <DL	N.D.
9)	TM TETRYL	6.478	6.482	131575	120811	NoCal		104.507 #
10)	TM 2,4,6-TRI...	6.925	0.000	6472	0	NoCal		N.D.
11)	TM 2-AMINO-4...	7.098	0.000	10473	0	NoCal		N.D.
12)	TM 4-AMINO-2...	7.572	0.000	17971	0	NoCal	NT	N.D. d
13)	TM 2,4-DINIT...	8.582	0.000	24421	0	NoCal	<DL	N.D. d
14)	TM 2,6-DINIT...	8.897	0.000	28388	0	NoCal		N.D.
15)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.		N.D.
16)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.		N.D.
17)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.		N.D.
18)	TM PETN	0.000	0.000	0	0	N.D.		N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000055.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 14:55:50
Operator : MP
Sample : AZ16140S01 7.921 DF 05/20/15
Misc : Soil
ALS Vial : 4231 Sample Multiplier: 7.92079

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:36:08 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522.M Tue May 26 07:36:27 2015 WALDORF Page: 2

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000060.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 22:30:51
Operator : MP
Sample : AZ16140S01 7.921 DF 05/20/15
Misc :
ALS Vial : 8387 Sample Multiplier: 7.92079

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:52:54 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	8.114	0	139834	N.D.	625.255	#
	Spiked Amount 495.050			Recovery	=	0.00%	126.30%	
	Target Compounds							
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.	
2)	TM RDX	0.000	0.000	0	0	N.D.	<DL N.D.	d
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	NT N.D.	d
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
6)	TM NITROBENZENE	0.000	0.000	0	0	N.D.	NT N.D.	d
7)	TM 2,4,6-TRI...	11.655	0.000	1577	0	NoCal	NT N.D.	d
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D.	N.D.	
9)	TM TETRYL	0.000	0.000	0	0	N.D.	<DL N.D.	d
10)	TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
11)	TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
12)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
13)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
14)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
15)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
16)	TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.	
17)	TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.	
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000060.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 22:30:51
Operator : MP
Sample : AZ16140S01 7.921 DF 05/20/15
Misc :
ALS Vial : 8387 Sample Multiplier: 7.92079

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:52:54 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



EPA 8330B SOIL

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SB032-0607
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16141
QCG: #83BJU-150520A-197310

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8330B	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
EPA 8330B	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
EPA 8330B	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
EPA 8330B	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
EPA 8330B	SURROGATE: 1,2-DINITROBENZENE (	93.9	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000056
Instrument: Waldorf
Sequence: 150522
Dilution Factor: 1
Initials: MP

Printed: 05/26/15 10:53:06 AM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000056.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 15:33:22
Operator : MP
Sample : AZ16141S01 7.929 DF 05/20/15
Misc : Soil
ALS Vial : 4232 Sample Multiplier: 7.92864

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:37:09 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.524	0	911535	N.D.	558.528	#
	Spiked Amount 594.648			Recovery	=	0.00%	93.93%	
	Target Compounds							
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.	
2)	TM RDX	0.000	0.000	0	0	N.D.	N.D.	
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.	
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
7)	TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.	
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D. d	LOL N.D.	
9)	TM TETRYL	6.443	0.000	12609	0	NoCal	N.D.	
10)	TM 2,4,6-TRI...	6.907	0.000	18071	0	NoCal	N.D.	
11)	TM 2-AMINO-4...	7.081	0.000	14496	0	NoCal	NJ N.D.	d
12)	TM 4-AMINO-2...	7.599	0.000	4047	0	NoCal	NJ N.D.	d
13)	TM 2,4-DINIT...	8.634	0.000	9765	0	NoCal	N.D.	
14)	TM 2,6-DINIT...	8.945	0.000	6310	0	NoCal	N.D.	
15)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
16)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
17)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.	
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000056.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 15:33:22
Operator : MP
Sample : AZ16141S01 7.929 DF 05/20/15
Misc : Soil
ALS Vial : 4232 Sample Multiplier: 7.92864

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:37:09 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



EPA 8330
EXPLOSIVES

Calibration Data

APPL, INC.

Initial Calibration

Lab Name: APPL, Inc. SDG No: 76380

Case No: Initial Cal. Date: 05/22/15

Matrix: Instrument: Waldorf

Initials: MP

			0522_000014.D	0522_000015.D	0522_000016.D	0522_000017.D	0522_000018.D	0522_000019.D	0522_000020.D	0522_000021.D	0522_000023.D				
		Compound	1	2	3	4	5	6	7	8	9		Avg	%RSD	
####	TM	NITROGLYCERIN	137	120	138	132	133	133	130	129	131		131.4	3.9	TM
####	TM	PETN			81.5	95.0	103	108	107	107	109		101.4	10	TM
####		Signal #2													
####	TM	HMX #2	584	694	666	640	652	676	655	639	639		649.5	4.7	TM
####	TM	RDX #2	425	422	426	427	424	428	428	424	429		425.8	0.52	TM
####	TM	1,3,5-TRINITROBENZENE #2	773	767	777	772	769	778	777	772	780		773.7	0.58	TM
####	S	1,2-DINITROBENZENE #2	303	319	328	325	323	326	327	327	334		323.5	2.6	S
####	TM	1,3-DINITROBENZENE #2	716	721	729	728	723	731	733	732	746		728.7	1.2	TM
####	TM	3,5-DINITROANILINE #2	548	552	557	555	552	558	559	559	562		555.8	0.77	TM
####	TM	NITROBENZENE #2	334	333	331	333	330	334	334	334	340		333.8	0.84	TM
####	TM	TETRYL #2	230	229	224	228	226	228	230	228	237		228.9	1.6	TM
####	TM	2,4,6-TRINITROTOLUENE #2	316	312	312	311	309	313	313	312	318		312.8	0.88	TM
####	TM	2-AMINO-4,6-DINITROTOLUENE	287	278	279	280	277	281	281	280	286		281.1	1.2	TM
####	TM	4-AMINO-2,6-DINITROTOLUENE	208	199	202	202	200	202	202	201	205		202.4	1.3	TM
####	TM	2,4-DINITROTOLUENE #2	285	291	296	295	293	297	297	298	304		295.2	1.7	TM
####	TM	2,6-DINITROTOLUENE #2	161	158	164	164	161	163	163	163	166		162.6	1.4	TM
####	TM	2-NITROTOLUENE #2	119	114	110	110	110	111	111	110	113		112.1	2.6	TM
####	TM	4-NITROTOLUENE #2	113	110	110	110	109	111	111	111	113		110.9	1.1	TM
####	TM	3-NITROTOLUENE #2	125	135	132	132	131	133	134	134	136		132.4	2.5	TM
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:07:53
Operator : MP
Sample : 8330CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 8384 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:56:46 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	3.531	0	60695	N.D.	4.454 #
Spiked Amount 62.500			Recovery	=	0.00%	7.13%
Target Compounds						
1) TM HMX	0.000	0.908	0	116863	N.D.	4.440m#
2) TM RDX	0.000	1.722	0	84918	N.D.	4.849 #
3) TM 1,3,5-TRI...	0.000	2.812	0	154593	N.D.	4.887 #
5) TM 1,3-DINIT...	0.000	3.704	0	143190	N.D.	4.738 #
6) TM 3,5-DINIT...	0.000	3.912	0	109681	N.D.	4.555 #
7) TM NITROBENZENE	0.000	5.027	0	66852	N.D.	4.834 #
8) TM NITROGLYC...	6.023	0.000	27427	0	4.921	N.D. #
9) TM TETRYL	6.381	6.373	88335	45934	NoCal	4.382 #
10) TM 2,4,6-TRI...	6.857	6.877	74624	63205	NoCal	4.842 #
11) TM 2-AMINO-4...	7.080	7.070	88869	57386	NoCal	4.625 #
12) TM 4-AMINO-2...	7.420	7.415	81261	41596	NoCal	4.574 #
13) TM 2,4-DINIT...	8.527	8.554	58052	57084	NoCal	4.506 #
14) TM 2,6-DINIT...	8.925	8.908	51464	32183	NoCal	4.563 #
15) TM 2-NITROTO...	0.000	12.272	0	23750	N.D.	4.800 #
16) TM 4-NITROTO...	0.000	12.869	0	22503	N.D.	4.548 #
17) TM 3-NITROTO...	0.000	13.759	0	24956	N.D.	4.244 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:07:53
Operator : MP
Sample : 8330CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 8384 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:56:46 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:07:53
Operator : MP
Sample : 8330CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 8384 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:16 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r





QEdit

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:07:53
Operator : MP
Sample : 8330CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 8384 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:16 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





MP 05-26-15 PM 5/26/15

QEdit

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.908min 4.440 ppb m

response 116863

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:55:40 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	3.535	0	127701	N.D.	9.372 #
	Spiked Amount 62.500			Recovery	=	0.00%	15.00%
	Target Compounds						
1)	TM HMX	0.000	0.910	0	277521	N.D.	10.544m#
2)	TM RDX	0.000	1.724	0	168970	N.D.	9.648 #
3)	TM 1,3,5-TRI...	0.000	2.815	0	306613	N.D.	9.693 #
5)	TM 1,3-DINIT...	0.000	3.708	0	288212	N.D.	9.537 #
6)	TM 3,5-DINIT...	0.000	3.916	0	220804	N.D.	9.170 #
7)	TM NITROBENZENE	0.000	5.032	0	133342	N.D.	9.642 #
8)	TM NITROGLYC...	6.053	0.000	48081	0	8.627m	N.D. #
9)	TM TETRYL	6.386	6.386	168423	91646	NoCal	8.744 #
10)	TM 2,4,6-TRI...	6.883	6.880	174881	124840	NoCal	9.563 #
11)	TM 2-AMINO-4...	7.081	7.079	184134	111282	NoCal	8.969 #
12)	TM 4-AMINO-2...	7.429	7.425	168272	79765	NoCal	8.771 #
13)	TM 2,4-DINIT...	8.557	8.561	96512	116390	NoCal	9.188 #
14)	TM 2,6-DINIT...	8.910	8.918	102587	63336	NoCal	8.980 #
15)	TM 2-NITROTO...	0.000	12.277	0	45530	N.D.	9.202 #
16)	TM 4-NITROTO...	0.000	12.877	0	43915	N.D.	8.875 #
17)	TM 3-NITROTO...	0.000	13.766	0	53902	N.D.	9.167 #
18)	TM PETN	16.091	0.000	33221	0	7.221	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:55:40 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.911min 17.212 ppb

response 452998

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





MLP 05-26-15

PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.910min 10.544 ppb m

response 277521

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





QEdit

(8) NITROGLYCERIN (TM)

6.042min 13.749 ppb

response 76625

(8) NITROGLYCERIN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:35:23
Operator : MP
Sample : 8330CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 8385 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(8) NITROGLYCERIN (TM)

6.053min 8.627 ppb m

response 48081

(8) NITROGLYCERIN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:54:25 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	3.537	0	262646	N.D.	19.275 #
	Spiked Amount 62.500			Recovery	=	0.00%	30.84%
	Target Compounds						
1) TM	HMX	0.000	0.910	0	533059	N.D.	20.254m#
2) TM	RDX	0.000	1.724	0	341011	N.D.	19.471 #
3) TM	1,3,5-TRI...	0.000	2.816	0	621520	N.D.	19.648 #
5) TM	1,3-DINIT...	0.000	3.709	0	583563	N.D.	19.310 #
6) TM	3,5-DINIT...	0.000	3.916	0	445840	N.D.	18.515 #
7) TM	NITROBENZENE	0.000	5.033	0	264975	N.D.	19.161 #
8) TM	NITROGLYC...	6.040	0.000	110320	0	19.796	N.D. #
9) TM	TETRYL	6.383	6.385	295550	179101	NoCal	17.088 #
10) TM	2,4,6-TRI...	6.880	6.880	322710	249913	NoCal	19.144 #
11) TM	2-AMINO-4...	7.079	7.078	332535	223533	NoCal	18.016 #
12) TM	4-AMINO-2...	7.421	7.422	308181	161739	NoCal	17.785 #
13) TM	2,4-DINIT...	8.559	8.558	184676	237033	NoCal	18.712 #
14) TM	2,6-DINIT...	8.921	8.916	196417	131074	NoCal	18.585 #
15) TM	2-NITROTO...	0.000	12.267	0	87831	N.D.	17.751 #
16) TM	4-NITROTO...	0.000	12.868	0	88090	N.D.	17.803 #
17) TM	3-NITROTO...	0.000	13.759	0	105360	N.D.	17.918 #
18) TM	PETN	16.096	0.000	65233	0	14.180m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:54:25 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.911min 26.788 ppb

response 705027

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r





MP 05-26-15

PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.910min 20.254 ppb m

response 533059

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(18) PETN (TM)

16.090min 11.845 ppb

response 54492

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:02:53
Operator : MP
Sample : 8330CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 8386 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





(18) PETN (TM)

16.096min 14.180 ppb m

response 65233

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:23
Operator : MP
Sample : 8330CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 8387 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:52:06 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	3.531	0	649173	N.D.	47.641 #
	Spiked Amount 62.500			Recovery	=	0.00%	76.23%
	Target Compounds						
1)	TM HMX	0.000	0.908	0	1280731	N.D.	48.662m#
2)	TM RDX	0.000	1.722	0	853611	N.D.	48.739 #
3)	TM 1,3,5-TRI...	0.000	2.812	0	1543113	N.D.	48.782 #
5)	TM 1,3-DINIT...	0.000	3.705	0	1455101	N.D.	48.148 #
6)	TM 3,5-DINIT...	0.000	3.912	0	1110694	N.D.	46.126 #
7)	TM NITROBENZENE	0.000	5.029	0	665871	N.D.	48.151 #
8)	TM NITROGLYC...	6.037	0.000	263621	0	47.304	N.D. #
9)	TM TETRYL	6.378	6.376	744181	455426	NoCal	43.451 #
10)	TM 2,4,6-TRI...	6.872	6.872	806095	621067	NoCal	47.575 #
11)	TM 2-AMINO-4...	7.070	7.070	843942	559774	NoCal	45.115 #
12)	TM 4-AMINO-2...	7.413	7.414	750390	404091	NoCal	44.434 #
13)	TM 2,4-DINIT...	8.551	8.551	439128	590887	NoCal	46.645 #
14)	TM 2,6-DINIT...	8.911	8.909	453309	328053	NoCal	46.514 #
15)	TM 2-NITROTO...	0.000	12.256	0	220810	N.D.	44.628 #
16)	TM 4-NITROTO...	0.000	12.859	0	220626	N.D.	44.589 #
17)	TM 3-NITROTO...	0.000	13.751	0	264010	N.D.	44.899 #
18)	TM PETN	16.045	0.000	189934	0	41.287	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:23
Operator : MP
Sample : 8330CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 8387 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:52:06 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:23
Operator : MP
Sample : 8330CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 8387 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:25 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r





QEdit

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.909min 54.634 ppb

response 1437929

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:23
Operator : MP
Sample : 8330CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 8387 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:25 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r





MLP 05-26-15 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.908min 48.662 ppb m

response 1280731

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:57:54
Operator : MP
Sample : 8330CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 8388 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:50:59 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	3.535	0	1290640	N.D.	94.716 #
	Spiked Amount 62.500			Recovery	=	0.00%	151.55%
	Target Compounds						
1)	TM HMX	0.000	0.910	0	2606306	N.D.	99.027m#
2)	TM RDX	0.000	1.724	0	1695994	N.D.	96.838 #
3)	TM 1,3,5-TRI...	0.000	2.814	0	3075082	N.D.	97.212 #
5)	TM 1,3-DINIT...	0.000	3.707	0	2892550	N.D.	95.712 #
6)	TM 3,5-DINIT...	0.000	3.915	0	2207280	N.D.	91.666 #
7)	TM NITROBENZENE	0.000	5.032	0	1320781	N.D.	95.510 #
8)	TM NITROGLYC...	6.041	0.000	530814	0	95.248	N.D. #
9)	TM TETRYL	6.383	6.383	1485615	902928	NoCal	86.147 #
10)	TM 2,4,6-TRI...	6.877	6.876	1600218	1235293	NoCal	94.627 #
11)	TM 2-AMINO-4...	7.078	7.077	1672837	1109691	NoCal	89.435 #
12)	TM 4-AMINO-2...	7.422	7.421	1488563	801060	NoCal	88.085 #
13)	TM 2,4-DINIT...	8.557	8.557	870898	1172656	NoCal	92.571 #
14)	TM 2,6-DINIT...	8.915	8.916	898126	644179	NoCal	91.337 #
15)	TM 2-NITROTO...	0.000	12.257	0	439307	N.D.	88.788 #
16)	TM 4-NITROTO...	0.000	12.858	0	437816	N.D.	88.484 #
17)	TM 3-NITROTO...	0.000	13.750	0	524084	N.D.	89.129 #
18)	TM PETN	16.035	0.000	411789	0	89.514	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:57:54
Operator : MP
Sample : 8330CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 8388 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:50:59 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:57:54
Operator : MP
Sample : 8330CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 8388 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:28 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.912min 105.179 ppb

response 2768218

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:57:54
Operator : MP
Sample : 8330CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 8388 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:28 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0





MP 05-26-15

PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.910min 99.027 ppb m

response 2606306

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:25:25
Operator : MP
Sample : 8330CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 8389 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:31 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.529	0	2607917	N.D.	191.386	#
	Spiked Amount 62.500			Recovery	=	0.00%	306.22%	
	Target Compounds							
1)	TM HMX	0.000	0.909	0	5406170	N.D.	205.408	#
2)	TM RDX	0.000	1.721	0	3423178	N.D.	195.456	#
3)	TM 1,3,5-TRI...	0.000	2.809	0	6223860	N.D.	196.754	#
5)	TM 1,3-DINIT...	0.000	3.702	0	5850290	N.D.	193.581	#
6)	TM 3,5-DINIT...	0.000	3.910	0	4462918	N.D.	185.340	#
7)	TM NITROBENZENE	0.000	5.028	0	2671122	N.D.	193.157	#
8)	TM NITROGLYC...	6.037	6.033	1064367	21764	190.988	NoCal	#
9)	TM TETRYL	6.378	6.378	3008528	1827851	NoCal	174.392	#
10)	TM 2,4,6-TRI...	6.872	6.872	3234786	2503048	NoCal	191.740	#
11)	TM 2-AMINO-4...	7.073	7.073	3375410	2244726	NoCal	180.913	#
12)	TM 4-AMINO-2...	7.417	7.418	2998380	1617102	NoCal	177.818	#
13)	TM 2,4-DINIT...	8.555	8.555	1746281	2373610	NoCal	187.375	#
14)	TM 2,6-DINIT...	8.916	8.915	1810737	1305146	NoCal	185.055	#
15)	TM 2-NITROTO...	0.000	12.265	0	891016	N.D.	180.082	#
16)	TM 4-NITROTO...	0.000	12.863	0	887417	N.D.	179.350	#
17)	TM 3-NITROTO...	0.000	13.754	0	1067793	N.D.	181.596	#
18)	TM PETN	16.029	0.000	862296	0	187.444	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 052,2_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:25:25
Operator : MP
Sample : 8330CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 8389 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:31 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522.M Sat May 23 08:50:41 2015 WALDORF Page: 2

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:52:55
Operator : MP
Sample : 8330CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 8390 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:49:26 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.532	0	6536750	N.D.	479.710	#
	Spiked Amount 62.500			Recovery	=	0.00%	767.54%	
	Target Compounds							
1)	TM HMX	0.000	0.912	0	13093933	N.D.	497.507	#
2)	TM RDX	0.000	1.725	0	8556770	N.D.	488.573	#
3)	TM 1,3,5-TRI...	0.000	2.813	0	15539688	N.D.	491.253	#
5)	TM 1,3-DINIT...	0.000	3.704	0	14653448	N.D.	484.870	#
6)	TM 3,5-DINIT...	0.000	3.912	0	11178554	N.D.	464.233	#
7)	TM NITROBENZENE	0.000	5.028	0	6683911	N.D.	483.335	#
8)	TM NITROGLYC...	6.034	6.029	2604059	52764	467.267m	NoCal	#
9)	TM TETRYL	6.376	6.376	7486113	4607535	NoCal	439.597	#
10)	TM 2,4,6-TRI...	6.870	6.870	8029656	6252449	NoCal	478.954	#
11)	TM 2-AMINO-4...	7.070	7.069	8405544	5619907	NoCal	452.935	#
12)	TM 4-AMINO-2...	7.414	7.414	7453750	4039281	NoCal	444.164	#
13)	TM 2,4-DINIT...	8.549	8.549	4361031	5949145	NoCal	469.632	#
14)	TM 2,6-DINIT...	8.910	8.909	4509069	3261676	NoCal	462.468	#
15)	TM 2-NITROTO...	0.000	12.261	0	2225219	N.D.	449.736	#
16)	TM 4-NITROTO...	0.000	12.859	0	2224626	N.D.	449.605	#
17)	TM 3-NITROTO...	0.000	13.751	0	2675829	N.D.	455.070	#
18)	TM PETN	16.025	0.000	2135884	0	464.294	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:52:55
Operator : MP
Sample : 8330CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 8390 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:49:26 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522.M Sat May 23 08:49:51 2015 WALDORF Page: 2

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:52:55
Operator : MP
Sample : 8330CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 8390 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:34 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r



(8) NITROGLYCERIN (TM)

6.035min 258.894 ppb

response 1442806

(8) NITROGLYCERIN #2 (TM)

6.029min 0.000 ppb

response 52764

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:52:55
Operator : MP
Sample : 8330CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 8390 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:34 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r



(8) NITROGLYCERIN (TM)

6.034min 467.267 ppb m

response 2604059

(8) NITROGLYCERIN #2 (TM)

6.029min 0.000 ppb

response 52764

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:20:26
Operator : MP
Sample : 8330CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 8391 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:48:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.526	0	13063722	N.D.	958.702	#
	Spiked Amount 62.500			Recovery	=	0.00%	1533.92%	
	Target Compounds							
1)	TM HMX	0.000	0.910	0	25565755	N.D.	971.376	#
2)	TM RDX	0.000	1.721	0	16944805	N.D.	967.511	#
3)	TM 1,3,5-TRI...	0.000	2.808	0	30884999	N.D.	976.362	#
5)	TM 1,3-DINIT...	0.000	3.698	0	29294479	N.D.	969.330	#
6)	TM 3,5-DINIT...	0.000	3.906	0	22350043	N.D.	928.174	#
7)	TM NITROBENZENE	0.000	5.019	0	13349299	N.D.	965.330	#
8)	TM NITROGLYC...	6.023	6.020	5146871	112765	923.543m	NoCal	#
9)	TM TETRYL	6.365	6.365	14825795	9135558	NoCal	871.607	#
10)	TM 2,4,6-TRI...	6.859	6.859	16022681	12472012	NoCal	955.390	#
11)	TM 2-AMINO-4...	7.059	7.059	16751362	11202296	NoCal	902.846	#
12)	TM 4-AMINO-2...	7.403	7.403	14857900	8042665	NoCal	884.380	#
13)	TM 2,4-DINIT...	8.536	8.536	8720715	11903150	NoCal	939.648	#
14)	TM 2,6-DINIT...	8.897	8.897	8992506	6508725	NoCal	922.862	#
15)	TM 2-NITROTO...	0.000	12.251	0	4418487	N.D.	893.014	#
16)	TM 4-NITROTO...	0.000	12.850	0	4422161	N.D.	893.736	#
17)	TM 3-NITROTO...	0.000	13.745	0	5340802	N.D.	908.293	#
18)	TM PETN	16.021	0.000	4272602	0	928.769	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:20:26
Operator : MP
Sample : 8330CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 8391 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:48:19 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:20:26
Operator : MP
Sample : 8330CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 8391 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:37 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(8) NITROGLYCERIN (TM)

6.024min 711.081 ppb

response 3962825

(8) NITROGLYCERIN #2 (TM)

6.020min 0.000 ppb

response 112765

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:20:26
Operator : MP
Sample : 8330CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 8391 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:37 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(8) NITROGLYCERIN (TM)

6.023min 923.543 ppb m

response 5146871

(8) NITROGLYCERIN #2 (TM)

6.020min 0.000 ppb

response 112765

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:47:57
Operator : MP
Sample : 8330MXA 2.0 PPM 05/22/15
Misc :
ALS Vial : 8392 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:40 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.528	0	26699791	N.D.	1959.407	#
	Spiked Amount 62.500			Recovery	=	0.00%	3135.05%	
	Target Compounds							
1)	TM HMX	0.000	0.909	0	51125746	N.D.	1942.533	#
2)	TM RDX	0.000	1.720	0	34296475	N.D.	1958.254	#
3)	TM 1,3,5-TRI...	0.000	2.808	0	62386918	N.D.	1972.226	#
5)	TM 1,3-DINIT...	0.000	3.700	0	59642142	N.D.	1973.509	#
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
7)	TM NITROBENZENE	0.000	5.022	0	27219378	N.D.	1968.320	#
8)	TM NITROGLYC...	5.993	5.988	-1131123	58860	N.D.	NoCal	
9)	TM TETRYL	6.370	6.370	31034512	18949650	NoCal	1807.952	#
10)	TM 2,4,6-TRI...	6.862	6.863	32706593	25443124	NoCal	1949.012	#
11)	TM 2-AMINO-4...	7.061	7.061	34245322	22905847	NoCal	1846.090	#
12)	TM 4-AMINO-2...	7.407	7.407	30318468	16364377	NoCal	1799.445	#
13)	TM 2,4-DINIT...	8.539	8.539	17785609	24299263	NoCal	1918.211	#
14)	TM 2,6-DINIT...	8.901	8.901	18336089	13284430	NoCal	1883.578	#
15)	TM 2-NITROTO...	0.000	12.247	0	9046766	N.D.	1828.429	#
16)	TM 4-NITROTO...	0.000	12.843	0	9037848	N.D.	1826.583	#
17)	TM 3-NITROTO...	0.000	13.739	0	10916758	N.D.	1856.579	#
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:47:57
Operator : MP
Sample : 8330MXA 2.0 PPM 05/22/15
Misc :
ALS Vial : 8392 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:46:40 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522.M Sat May 23 08:47:52 2015 WALDORF Page: 2

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000023.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:15:27
Operator : MP
Sample : 8330MXB 2.0 PPM 05/22/15
Misc :
ALS Vial : 8448 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:47:01 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	0.000	0	0	N.D.	N.D.
	Spiked Amount 62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.
2)	TM RDX	0.000	0.000	0	0	N.D.	N.D.
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6)	TM 3,5-DINIT...	0.000	3.909	0	44937660	N.D.	1866.213 #
7)	TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
8)	TM NITROGLYC...	6.030	6.030	10501636	195853	1884.391	NoCal #
9)	TM TETRYL	0.000	0.000	0	0	N.D.	N.D.
10)	TM 2,4,6-TRI...	6.930	0.000	43283	0	NoCal	N.D.
11)	TM 2-AMINO-4...	7.080	0.000	15932	0	NoCal	N.D.
12)	TM 4-AMINO-2...	7.392	0.000	17692	0	NoCal	N.D.
13)	TM 2,4-DINIT...	8.533	0.000	10405	0	NoCal	N.D.
14)	TM 2,6-DINIT...	8.895	0.000	14327	0	NoCal	N.D.
15)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18)	TM PETN	16.008	0.000	8729316	0	1897.559	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000023.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:15:27
Operator : MP
Sample : 8330MXB 2.0 PPM 05/22/15
Misc :
ALS Vial : 8448 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 08:47:01 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Mon May 18 09:47:14 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 16380
Date Analyzed: 22-May-2015, 21:37:58
Instrument: Waldorf
Initial Cal. Date: 05/22/15
Data File: 0522_000026-27.D

		Compound	MEAN	CCRF	%D	%Drift	
1	TM	NITROGLYCERIN	131	133	1.4	TM	
2	TM	PETN	101	106	4.5	TM	
3		Signal #2					
4	TM	HMX	649	626	3.6	TM	
5	TM	RDX	426	425	0.10	TM	
6	TM	1,3,5-TRINITROBENZENE	774	720	6.9	TM	
7	S	1,2-DINITROBENZENE	323	342	5.8	S	
8	TM	1,3-DINITROBENZENE	729	738	1.3	TM	
9	TM	3,5-DINITROANILINE	556	588	5.8	TM	
10	TM	NITROBENZENE	334	336	0.73	TM	
11	TM	2,4,6-TRINITROTOLUENE	313	318	1.6	TM	
12	TM	2-AMINO-4,6-DINITROTOLUEN	281	262	6.7	TM	
13	TM	4-AMINO-2,6-DINITROTOLUEN	202	218	7.9	TM	
14	TM	2,4-DINITROTOLUENE	295	312	5.7	TM	
15	TM	2,6-DINITROTOLUENE	163	169	3.9	TM	
16	TM	2-NITROTOLUENE	112	114	1.8	TM	
17	TM	4-NITROTOLUENE	111	113	2.0	TM	
18	TM	3-NITROTOLUENE	132	140	5.6	TM	
19	TM	TETRYL	229	245	7.1	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:37:58
Operator : MP
Sample : 8330_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 8449 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:09:58 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0m
m 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.523	0	13686166	N.D.	1057.681	#
	Spiked Amount 62.500			Recovery	=	0.00%	1692.29%	
	Target Compounds							
1)	TM HMX	0.000	0.909	0	25046700	N.D.	964.148	#
2)	TM RDX	0.000	1.718	0	17015404	N.D.	999.040	#
3)	TM 1,3,5-TRI...	0.000	2.804	0	28816272	N.D.	931.072	#
5)	TM 1,3-DINIT...	0.000	3.694	0	29521663	N.D.	1012.793	#
6)	TM 3,5-DINIT...	0.000	3.902	0	23531862	N.D.	1058.480	#
7)	TM NITROBENZENE	0.000	5.017	0	13449297	N.D.	1007.346	#
8)	TM NITROGLYC...	6.017	6.014	5329290	133108	1013.653m	1252.663	
9)	TM TETRYL	6.357	6.357	13306833	8224915	859.217	898.255	
10)	TM 2,4,6-TRI...	6.848	6.848	16398781	12711220	1144.544	1015.917	
11)	TM 2-AMINO-4...	7.051	7.051	15700744	10493610	1032.426	933.290	
12)	TM 4-AMINO-2...	7.394	7.394	16165189	8739257	1181.397	1079.454	
13)	TM 2,4-DINIT...	8.524	8.524	9095826	12483512	1103.768	1057.219	
14)	TM 2,6-DINIT...	8.885	8.885	9342247	6759755	1113.774	1039.487	
15)	TM 2-NITROTO...	0.000	12.234	0	4563295	N.D.	1017.803	#
16)	TM 4-NITROTO...	0.000	12.831	0	4522887	N.D.	1019.826	#
17)	TM 3-NITROTO...	0.000	13.728	0	5592648	N.D.	1056.099	#
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:37:58
Operator : MP
Sample : 8330_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 8449 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:09:58 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522.M Sat May 23 09:10:25 2015 WALDORF Page: 2

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:37:58
Operator : MP
Sample : 8330_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 8449 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:08:46 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



(8) NITROGLYCERIN (TM)

6.018min 780.047 ppb

response 4101103

(8) NITROGLYCERIN #2 (TM)

6.014min 1252.663 ppb

response 133108

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:37:58
Operator : MP
Sample : 8330_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 8449 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:08:46 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r



(8) NITROGLYCERIN (TM)

6.017min 1013.653 ppb m

response 5329290

(8) NITROGLYCERIN #2 (TM)

6.014min 1252.663 ppb

response 133108

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000027.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:05:29
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/51
Misc :
ALS Vial : 8450 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:09:04 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0m
m 1.8-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D.	N.D.
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D.	N.D.
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
7) TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	6.363	6.363	16063827	9806270	1037.235	1070.957
10) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
11) TM 2-AMINO-4...	0.000	0.000	0	0	N.D. d	N.D. d
12) TM 4-AMINO-2...	0.000	0.000	0	0	N.D. d	N.D. d
13) TM 2,4-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 2,6-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 2-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
16) TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	15.990	0.000	4238463	0	1044.738	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000027.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:05:29
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/51
Misc :
ALS Vial : 8450 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:09:04 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:07:30 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 76380
Date Analyzed: 23-May-2015, 08:13:14
Instrument: Waldorf
Initial Cal. Date: 05/22/15
Data File: 0522_000044.D

		Compound	MEAN	CCRF	%D		%Drift
1	TM	NITROGLYCERIN	131	131	0.25	TM	
2	TM	PETN	101	107	5.1	TM	
3		Signal #2					
4	TM	HMX	649	644	0.81	TM	
5	TM	RDX	426	425	0.26	TM	
6	TM	1,3,5-TRINITROBENZENE	774	768	0.78	TM	
7	S	1,2-DINITROBENZENE	323	326	0.72	S	
8	TM	1,3-DINITROBENZENE	729	730	0.19	TM	
9	TM	3,5-DINITROANILINE	556	555	0.22	TM	
10	TM	NITROBENZENE	334	327	2.0	TM	
11	TM	TETRYL	229	230	0.49	TM	
12	TM	2,4,6-TRINITROTOLUENE	313	309	1.1	TM	
13	TM	2-AMINO-4,6-DINITROTOLUEN	281	278	1.0	TM	
14	TM	4-AMINO-2,6-DINITROTOLUEN	202	201	0.66	TM	
15	TM	2,4-DINITROTOLUENE	295	297	0.60	TM	
16	TM	2,6-DINITROTOLUENE	163	163	0.32	TM	
17	TM	2-NITROTOLUENE	112	110	2.2	TM	
18	TM	4-NITROTOLUENE	111	110	0.46	TM	
19	TM	3-NITROTOLUENE	132	132	0.46	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000044.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 08:13:14
Operator : MP
Sample : 8330_CCV 1.0 PPM 05/22/15
Misc :
ALS Vial : 4423 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:20:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.551	0	13033262	N.D.	1007.224	#
	Spiked Amount 62.500			Recovery	=	0.00%	1611.56%	
	Target Compounds							
1)	TM HMX	0.000	0.913	0	25766714	N.D.	991.864	#
2)	TM RDX	0.000	1.731	0	16987352	N.D.	997.393	#
3)	TM 1,3,5-TRI...	0.000	2.824	0	30707028	N.D.	992.164	#
5)	TM 1,3-DINIT...	0.000	3.721	0	29203547	N.D.	1001.879	#
6)	TM 3,5-DINIT...	0.000	3.940	0	22182417	N.D.	997.781	#
7)	TM NITROBENZENE	0.000	5.048	0	13078148	N.D.	979.547	#
8)	TM NITROGLYC...	6.054	6.051	5244498	110802	997.525	NoCal	#
9)	TM TETRYL	6.418	6.418	15123905	9201676	NoCal	1004.929	#
10)	TM 2,4,6-TRI...	6.901	6.901	15976632	12376178	NoCal	989.139	#
11)	TM 2-AMINO-4...	7.121	7.121	16781376	11131401	NoCal	990.015	#
12)	TM 4-AMINO-2...	7.465	7.465	14971759	8042355	NoCal	993.374	#
13)	TM 2,4-DINIT...	8.589	8.589	8718482	11878656	NoCal	1005.994	#
14)	TM 2,6-DINIT...	8.950	8.950	8997170	6524085	NoCal	1003.246	#
15)	TM 2-NITROTO...	0.000	12.291	0	4383143	N.D.	977.622	#
16)	TM 4-NITROTO...	0.000	12.888	0	4414548	N.D.	995.397	#
17)	TM 3-NITROTO...	0.000	13.778	0	5271169	N.D.	995.392	#
18)	TM PETN	16.034	0.000	4263867	0	1051.000	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000044.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 08:13:14
Operator : MP
Sample : 8330_CCV 1.0 PPM 05/22/15
Misc :
ALS Vial : 4423 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:20:22 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 16380
Date Analyzed: 23-May-2015, 17:25:54
Instrument: Waldorf
Initial Cal. Date: 05/22/15
Data File: 0522_000059.D

		Compound	MEAN	CCRF	%D	%Drift	
1	TM	NITROGLYCERIN	131	129	2.0	TM	
2	TM	PETN	101	105	3.6	TM	
3		Signal #2					
4	TM	HMX	649	648	0.20	TM	
5	TM	RDX	426	424	0.51	TM	
6	TM	1,3,5-TRINITROBENZENE	774	762	1.5	TM	
7	S	1,2-DINITROBENZENE	323	324	0.04	S	
8	TM	1,3-DINITROBENZENE	729	723	0.72	TM	
9	TM	3,5-DINITROANILINE	556	550	0.97	TM	
10	TM	NITROBENZENE	334	315	5.5	TM	
11	TM	TETRYL	229	224	2.1	TM	
12	TM	2,4,6-TRINITROTOLUENE	313	305	2.4	TM	
13	TM	2-AMINO-4,6-DINITROTOLUEN	281	277	1.4	TM	
14	TM	4-AMINO-2,6-DINITROTOLUEN	202	199	1.8	TM	
15	TM	2,4-DINITROTOLUENE	295	292	1.0	TM	
16	TM	2,6-DINITROTOLUENE	163	160	1.3	TM	
17	TM	2-NITROTOLUENE	112	104	7.4	TM	
18	TM	4-NITROTOLUENE	111	106	4.2	TM	
19	TM	3-NITROTOLUENE	132	126	4.6	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 17:25:54
Operator : MP
Sample : 8330_CCV 1.0 PPM 05/22/15
Misc :
ALS Vial : 4423 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:50:44 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.533	0	12945069	N.D.	1000.409	#
	Spiked Amount 62.500			Recovery =		0.00%	1600.65%	
	Target Compounds							
1)	TM HMX	0.000	0.907	0	25927371	N.D.	998.049	#
2)	TM RDX	0.000	1.721	0	16944784	N.D.	994.893	#
3)	TM 1,3,5-TRI...	0.000	2.814	0	30478139	N.D.	984.768	#
5)	TM 1,3-DINIT...	0.000	3.706	0	28937562	N.D.	992.754	#
6)	TM 3,5-DINIT...	0.000	3.915	0	22016827	N.D.	990.332	#
7)	TM NITROBENZENE	0.000	5.029	0	12613342	N.D.	944.733	#
8)	TM NITROGLYC...	6.040	0.000	5153242	0	980.168	N.D.	d#
9)	TM TETRYL	6.386	6.385	14745789	8968494	NoCal	979.462	#
10)	TM 2,4,6-TRI...	6.880	6.880	15758872	12212506	NoCal	976.058	#
11)	TM 2-AMINO-4...	7.078	7.078	16626480	11087756	NoCal	986.133	#
12)	TM 4-AMINO-2...	7.423	7.423	14701202	7947388	NoCal	981.644	#
13)	TM 2,4-DINIT...	8.560	8.560	8601896	11688477	NoCal	989.888	#
14)	TM 2,6-DINIT...	8.919	8.919	8886920	6418681	NoCal	987.038	#
15)	TM 2-NITROTO...	0.000	12.260	0	4151380	N.D.	925.929	#
16)	TM 4-NITROTO...	0.000	12.857	0	4249833	N.D.	958.257	#
17)	TM 3-NITROTO...	0.000	13.748	0	5051798	N.D.	953.966	#
18)	TM PETN	16.028	0.000	4204550	0	1036.379	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 17:25:54
Operator : MP
Sample : 8330_CCV 1.0 PPM 05/22/15
Misc :
ALS Vial : 4423 Sample Multiplier: 1

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:50:44 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



W150522 M Sun May 24 12:51:03 2015 WALDORF

Page: 2

Initial Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 76380
Initial Cal. Date: 05/22/15
Instrument: Statler

Initials: MP

			0522_000010.D	0522_000012.D	0522_000014.D	0522_000016.D	0522_000018.D	0522_000020.D	0522_000022.D	0522_000024.D	0522_000026.D				
		Compound	1	2	3	4	5	6	7	8	9		Avg	%RSD	
####	TM	NITROGLYCERIN		20.8	18.0	18.1	17.8	18.4	18.4	18.3	18.5		18.55	5.0	TM
####	TM	PETN			17.5	15.7	17.4	18.9	19.1	19.0	19.3		18.15	7.2	TM
####		Signal #2													
####	TM	HMX #2	40.3	42.0	46.4	46.8	46.8	48.8	48.8	47.9	48.6		46.27	6.7	TM
####	TM	RDX #2	38.4	40.4	38.6	40.0	39.6	40.9	40.3	39.6	40.1		39.77	2.1	TM
####	TM	1,3,5-TRINITROBENZENE #2	98.9	96.4	98.4	97.8	95.8	98.9	98.3	97.4	99.2		97.91	1.2	TM
####	S	1,2-DINITROBENZENE #2	46.4	43.2	44.9	43.7	42.6	46.2	44.0	43.3	44.3		44.29	2.9	S
####	TM	1,3-DINITROBENZENE #2	86.5	96.4	98.9	96.6	95.6	101	97.8	97.0	99.5		96.56	4.3	TM
####	TM	NITROBENZENE #2	57.2	50.2	53.4	53.1	55.0	56.0	54.7	53.7	54.7		54.22	3.7	TM
####	TM	2,4,6-TRINITROTOLUENE #2		48.9	50.9	49.7	49.6	51.4	50.2	49.8	51.2		50.21	1.8	TM
####	TM	TETRYL #2		35.9	39.1	36.0	36.2	36.4	36.1	36.4	36.3		36.56	2.8	TM
####	TM	2,6-DINITROTOLUENE #2	31.2	26.8	30.6	29.7	29.7	30.8	29.9	30.3	31.0		30.00	4.4	TM
####	TM	2,4-DINITROTOLUENE #2	47.8	51.9	51.1	51.9	51.6	53.3	52.0	52.3	53.5		51.72	3.2	TM
####	TM	2-NITROTOLUENE #2		21.5	21.4	20.6	21.0	21.5	20.8	21.0	21.4		21.16	1.5	TM
####	TM	4-NITROTOLUENE #2			16.3	18.1	18.6	18.9	18.4	18.5	18.9		18.26	4.9	TM
####	TM	3,5-DINITROANILINE #2	37.6	43.2	41.6	41.2	41.3	42.3	42.1	41.3	41.4		41.34	3.7	TM
####	TM	3-NITROTOLUENE #2		22.8	23.9	24.0	23.1	22.9	22.2	22.0	22.4		22.92	3.3	TM
####	TM	4-AMINO-2,6-DINITROTOLUENE #2		33.3	38.0	35.5	34.3	34.1	32.9	32.4	32.8		34.16	5.4	TM
####	TM	2-AMINO-4,6-DINITROTOLUENE #2		38.8	45.9	44.8	43.1	44.3	42.6	42.0	42.4		42.99	5.0	TM
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															
####															

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:11:18 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	8.113	0	9274	N.D.	4.396m#
	Spiked Amount 62.500			Recovery	=	0.00%	7.03%
	Target Compounds						
1) TM	HMX	0.000	4.080	0	8051	N.D.	3.829m#
2) TM	RDX	0.000	6.415	0	7675	N.D.	4.144m#
3) TM	1,3,5-TRI...	0.000	6.754	0	19781	N.D.	4.426m#
5) TM	1,3-DINIT...	0.000	8.432	0	17298	N.D.	3.846m#
6) TM	NITROBENZENE	0.000	9.166	0	11437	N.D.	4.465m#
7) TM	2,4,6-TRI...	0.000	11.443	0	7160	N.D.	2.820 #
8) TM	NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM	TETRYL	11.981	12.005	7668	6007	NoCal	3.647 #
10) TM	2,6-DINIT...	0.000	13.440	0	6231	N.D.	4.218m#
11) TM	2,4-DINIT...	0.000	14.078	0	9567	N.D.	3.797m#
12) TM	2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM	4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
14) TM	3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
15) TM	3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM	4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM	2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM	PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:11:18 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı





(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.074min 13.606 ppb

response 28611

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.080min 3.829 ppb m

response 8051

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.415min 8.101 ppb

response 15004

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı



MHP 05-26-15 PM 5/26/15

(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.415min 4.144 ppb m

response 7675

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.756min 6.174 ppb

response 27597

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.754min 4.426 ppb m

response 19781

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.106min 8.094 ppb

response 17077

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MLP 05-26-15

ms/26/15

(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.113min 4.396 ppb m

response 9274

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





QEdit

(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.434min 5.153 ppb

response 23173

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MP 05-26-15 PMS/26/15

QEdit

(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.432min 3.846 ppb m

response 17298

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.167min 5.196 ppb

response 13311

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MP 05-26-15 PM 5/26/15

(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.166min 4.465 ppb m

response 11437

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





QEdit

(10) 2,6-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(10) 2,6-DINITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MSP 05-26-15 PMS/26/15

(10) 2,6-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(10) 2,6-DINITROTOLUENE #2 (TM)

13.440min 4.218 ppb m

response 6231

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





QEdit

(11) 2,4-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(11) 2,4-DINITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000009.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:13:08
Operator : MP
Sample : 8330MXA_CB 0.005 PPM 05/22/15
Misc :
ALS Vial : 4224 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:29:59 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı



(11) 2,4-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(11) 2,4-DINITROTOLUENE #2 (TM)

14.078min 3.797 ppb m

response 9567

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000010.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:52:39
Operator : MP
Sample : 8330MXB_CV 0.005 PPM 05/22/15
Misc :
ALS Vial : 4225 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:07:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount	62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1) TM	HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM	RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM	1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM	1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM	NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM	2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM	NITROGLYC...	0.000	12.058	0	885	N.D. d	NoCal
9) TM	TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM	2,6-DINIT...	13.254	0.000	538	0	NoCal	N.D.
11) TM	2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM	2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM	4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
14) TM	3,5-DINIT...	0.000	16.553	0	7523	N.D.	3.769m#
15) TM	3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM	4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM	2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM	PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000010.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:52:39
Operator : MP
Sample : 8330MXB_CV 0.005 PPM 05/22/15
Misc :
ALS Vial : 4225 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:07:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000010.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:52:39
Operator : MP
Sample : 8330MXB_CV 0.005 PPM 05/22/15
Misc :
ALS Vial : 4225 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:02 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





QEdit

(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000010.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 14:52:39
Operator : MP
Sample : 8330MXB_CV 0.005 PPM 05/22/15
Misc :
ALS Vial : 4225 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:02 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





QEdit

(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

16.553min 3.769 ppb m

response 7523

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:06:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	8.102	0	17296	N.D.	8.198m#
	Spiked Amount 62.500			Recovery	=	0.00%	13.12%
	Target Compounds						
1) TM	HMX	0.000	4.082	0	16807	N.D.	7.993m#
2) TM	RDX	0.000	6.417	0	16148	N.D.	8.719m#
3) TM	1,3,5-TRI...	0.000	6.753	0	38559	N.D.	8.627m#
5) TM	1,3-DINIT...	0.000	8.433	0	38569	N.D.	8.576m#
6) TM	NITROBENZENE	0.000	9.177	0	20086	N.D.	7.841m#
7) TM	2,4,6-TRI...	0.000	11.450	0	19541	N.D.	7.697 #
8) TM	NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM	TETRYL	11.986	11.978	13853	14357	NoCal	8.715 #
10) TM	2,6-DINIT...	13.457	13.467	21588	10727	NoCal	7.263m#
11) TM	2,4-DINIT...	0.000	14.062	0	20758	N.D.	8.239m#
12) TM	2-NITROTO...	0.000	15.468	0	8585	N.D.	8.142m#
13) TM	4-NITROTO...	0.000	16.858	0	4467	N.D.	5.100m#
14) TM	3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
15) TM	3-NITROTO...	0.000	18.086	0	9136	N.D.	8.544m#
16) TM	4-AMINO-2...	0.000	18.658	0	13310	N.D.	9.500m#
17) TM	2-AMINO-4...	0.000	19.621	0	15539	N.D.	7.623 #
18) TM	PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:06:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.078min 17.847 ppb

response 37530

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MHP 05-26-15 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.082min 7.993 ppb m

response 16807

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.418min 16.136 ppb

response 29887

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.417min 8.719 ppb m

response 16148

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.757min 11.817 ppb

response 52819

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.753min 8.627 ppb m

response 38559

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.101min 13.504 ppb

response 28490

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MLP 05-26-15 5/26/15

(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.102min 8.198 ppb m

response 17296

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.431min 10.647 ppb

response 47884

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.433min 8.576 ppb m

response 38569

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.178min 9.671 ppb

response 24773

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı



MIP 05-26-15 5/26/15

(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.177min 7.841 ppb m

response 20086

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





QEdit

(10) 2,6-DINITROTOLUENE (TM)

13.457min 0.000 ppb

response 21588

(10) 2,6-DINITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MRMP 05-26-15 jm5/26/15

(10) 2,6-DINITROTOLUENE (TM)

13.457min 0.000 ppb

response 21588

(10) 2,6-DINITROTOLUENE #2 (TM)

13.467min 7.263 ppb m

response 10727

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı



(11) 2,4-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(11) 2,4-DINITROTOLUENE #2 (TM)

14.065min 7.394 ppb

response 18629

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(11) 2,4-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(11) 2,4-DINITROTOLUENE #2 (TM)

14.062min 8.239 ppb m

response 20758

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(12) 2-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(12) 2-NITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MSP 05-26-15
PM 5/26/15

(12) 2-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(12) 2-NITROTOLUENE #2 (TM)

15.468min 8.142 ppb m

response 8585

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(13) 4-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(13) 4-NITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MSP 05-26-15 PM 5/26/15

(13) 4-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(13) 4-NITROTOLUENE #2 (TM)

16.858min 5.100 ppb m

response 4467

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(15) 3-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(15) 3-NITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(15) 3-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(15) 3-NITROTOLUENE #2 (TM)

18.086min 8.544 ppb m

response 9136

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





(16) 4-AMINO-2,6-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(16) 4-AMINO-2,6-DINITROTOLUENE #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000011.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 15:32:10
Operator : MP
Sample : 8330MXA_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4226 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:05 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MSP 05-26-15 PM 5/26/15

(16) 4-AMINO-2,6-DINITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(16) 4-AMINO-2,6-DINITROTOLUENE #2 (TM)

18.658min 9.500 ppb m

response 13310

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000012.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:11:42
Operator : MP
Sample : 8330MXB_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4227 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:03:03 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.069	12.012	8300	427	9.114	NoCal #
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.538	0	17290	N.D.	8.662m#
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000012.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:11:42
Operator : MP
Sample : 8330MXB_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4227 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:03:03 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000012.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:11:42
Operator : MP
Sample : 8330MXB_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4227 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:08 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

16.542min 6.767 ppb

response 13506

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000012.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:11:42
Operator : MP
Sample : 8330MXB_CB 0.010 PPM 05/22/15
Misc :
ALS Vial : 4227 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:08 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

16.538min 8.662 ppb m

response 17290

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:50:55 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.097	0	35930	N.D.	17.030m#
Spiked Amount 62.500			Recovery	=	0.00%	27.25%
Target Compounds						
1) TM HMX	0.000	4.078	0	37138	N.D.	17.661m#
2) TM RDX	0.000	6.420	0	30890	N.D.	16.678m#
3) TM 1,3,5-TRI...	0.000	6.758	0	78697	N.D.	17.607m#
5) TM 1,3-DINIT...	0.000	8.432	0	79097	N.D.	17.588m#
6) TM NITROBENZENE	0.000	9.169	0	42712	N.D.	16.674m#
7) TM 2,4,6-TRI...	0.000	11.447	0	40720	N.D.	16.040 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.979	11.969	25584	31272	NoCal	18.984 #
10) TM 2,6-DINIT...	13.451	13.456	39681	24473	NoCal	16.568 #
11) TM 2,4-DINIT...	0.000	14.055	0	40841	N.D.	16.210 #
12) TM 2-NITROTO...	0.000	15.476	0	17127	N.D.	16.245m#
13) TM 4-NITROTO...	0.000	16.897	0	13066	N.D.	14.919 #
14) TM 3,5-DINIT...	0.000	16.897f	0	13066	N.D.	6.546 #
15) TM 3-NITROTO...	0.000	18.069	0	19154	N.D.	17.914 #
16) TM 4-AMINO-2...	0.000	18.648	0	30398	N.D.	21.698 #
17) TM 2-AMINO-4...	0.000	19.600	0	36708	N.D.	18.007 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:50:55 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





QEdit

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.077min 26.787 ppb

response 56329

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MM 05-26-15 PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.078min 17.661 ppb m

response 37138

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.418min 22.479 ppb

response 41634

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



MLP 05-26-15 pm 5/26/15

(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.420min 16.678 ppb m

response 30890

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.759min 20.535 ppb

response 91787

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MHP 05-26-15 PM 5/26/15

(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.758min 17.607 ppb m

response 78697

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.098min 21.218 ppb

response 44767

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





MLH 05-26-15 PM5/26/15

(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.097min 17.030 ppb m

response 35930

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.433min 19.439 ppb

response 87423

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MHP 05-26-15
5/26/15

(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.432min 17.588 ppb m

response 79097

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





QEdit

(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.171min 18.468 ppb

response 47307

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.169min 16.674 ppb m

response 42712

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(12) 2-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(12) 2-NITROTOLUENE #2 (TM)

15.474min 14.676 ppb

response 15473

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000013.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 16:51:12
Operator : MP
Sample : 8330MXA_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4228 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(12) 2-NITROTOLUENE (TM)

0.000min 0.000 ppb

response 0

(12) 2-NITROTOLUENE #2 (TM)

15.476min 16.245 ppb m

response 17127

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:47:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.062	0.000	14397	0	15.809	N.D. d#
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.542	0	33313	N.D.	16.689m#
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.763	0.000	14034	0	15.381m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:47:11 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:14 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

16.546min 15.824 ppb

response 31586

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:14 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





(18) PETN (TM)

0.000min 0.000 ppb

response 0

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:14 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MIP 05-26-15 pm 5/26/15

(14) 3,5-DINITROANILINE (TM)

0.000min 0.000 ppb

response 0

(14) 3,5-DINITROANILINE #2 (TM)

16.542min 16.689 ppb m

response 33313

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000014.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 17:30:43
Operator : MP
Sample : 8330MXB_CB 0.020 PPM 05/22/15
Misc :
ALS Vial : 4229 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:14 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





(18) PETN (TM)

22.763min 15.381 ppb m

response 14034

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:46:09 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-m
icron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	8.092	0	87334	N.D.	41.395m#
	Spiked Amount 62.500			Recovery	=	0.00%	66.23%
	Target Compounds						
1) TM	HMX	0.000	4.077	0	93606	N.D.	44.514m#
2) TM	RDX	0.000	6.412	0	80036	N.D.	43.213m#
3) TM	1,3,5-TRI...	0.000	6.752	0	195589	N.D.	43.759m#
5) TM	1,3-DINIT...	0.000	8.429	0	193271	N.D.	42.975m#
6) TM	NITROBENZENE	0.000	9.168	0	106151	N.D.	41.438m#
7) TM	2,4,6-TRI...	0.000	11.440	0	99469	N.D.	39.181 #
8) TM	NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM	TETRYL	11.972	11.971	64180	71983	NoCal	43.697 #
10) TM	2,6-DINIT...	13.446	13.444	94573	59327	NoCal	40.165 #
11) TM	2,4-DINIT...	0.000	14.052	0	103731	N.D.	41.171 #
12) TM	2-NITROTO...	0.000	15.459	0	41226	N.D.	39.102 #
13) TM	4-NITROTO...	0.000	16.870	0	36293	N.D.	41.440 #
14) TM	3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM	3-NITROTO...	0.000	18.065	0	48059	N.D.	44.948 #
16) TM	4-AMINO-2...	0.000	18.644	0	71020	N.D.	50.694 #
17) TM	2-AMINO-4...	0.000	19.600	0	89551	N.D.	43.931 #
18) TM	PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:46:09 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.077min 54.675 ppb

response 114973

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MHP 05-26-15 PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.077min 44.514 ppb m

response 93606

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.412min 51.095 ppb

response 94635

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MHP 05-26-15 PM 5/26/15

(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.412min 43.213 ppb m

response 80036

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.753min 47.022 ppb

response 210176

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MJP 05-26-15 pm5/26/15

(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.752min 43.759 ppb m

response 195589

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-ı



(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.093min 47.687 ppb

response 100611

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MLP 05-26-15 PM 5/26/15

(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.092min 41.395 ppb m

response 87334

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.429min 45.755 ppb

response 205777

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





MP 05-26-15 PM 5/26/15

(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.429min 42.975 ppb m

response 193271

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.169min 44.660 ppb

response 114402

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000015.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:10:14
Operator : MP
Sample : 8330MXA_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4230 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



MHP
05-26-15
pm5/26/15

(6) NITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(6) NITROBENZENE #2 (TM)

9.168min 41.438 ppb m

response 106151

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:49:46
Operator : MP
Sample : 8330MXB_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4231 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:43:41 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
	Spiked Amount 62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1) TM	HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM	RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM	1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM	1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM	NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM	2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM	NITROGLYC...	12.066	0.000	36214	0	39.765	N.D. d#
9) TM	TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM	2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM	2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM	2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM	4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM	3,5-DINIT...	0.000	16.543	0	82461	N.D.	41.312 #
15) TM	3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM	4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM	2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM	PETN	22.806	0.000	31484	0	34.504m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:49:46
Operator : MP
Sample : 8330MXB_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4231 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:43:41 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:49:46
Operator : MP
Sample : 8330MXB_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4231 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:20 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(18) PETN (TM)

22.821min 26.079 ppb

response 23796

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000016.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 18:49:46
Operator : MP
Sample : 8330MXB_CB 0.050 PPM 05/22/15
Misc :
ALS Vial : 4231 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:20 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





QEdit

(18) PETN (TM)

22.806min 34.504 ppb m

response 31484

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:42:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.096	0	170391	N.D.	80.762m#
Spiked Amount 62.500			Recovery =		0.00%	129.22%
Target Compounds						
1) TM HMX	0.000	4.078	0	187333	N.D.	89.085m#
2) TM RDX	0.000	6.416	0	158283	N.D.	85.460m#
3) TM 1,3,5-TRI...	0.000	6.754	0	383111	N.D.	85.713m#
5) TM 1,3-DINIT...	0.000	8.432	0	382410	N.D.	85.031m#
6) TM NITROBENZENE	0.000	9.173	0	220159	N.D.	85.944 #
7) TM 2,4,6-TRI...	0.000	11.447	0	198494	N.D.	78.187 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.982	11.980	121440	144767	NoCal	87.881 #
10) TM 2,6-DINIT...	13.453	13.452	184036	118801	NoCal	80.430 #
11) TM 2,4-DINIT...	0.000	14.059	0	206563	N.D.	81.985 #
12) TM 2-NITROTO...	0.000	15.467	0	84122	N.D.	79.789 #
13) TM 4-NITROTO...	0.000	16.888	0	74548	N.D.	85.119 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	18.071	0	92548	N.D.	86.556 #
16) TM 4-AMINO-2...	0.000	18.658	0	137010	N.D.	97.797 #
17) TM 2-AMINO-4...	0.000	19.613	0	172579	N.D.	84.662 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:42:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



STE2.150522.M Sat May 23 09:42:14 2015 WALDORF Page: 2

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.079min 98.980 ppb

response 208141

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MLP 05-26-15 PM 5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.078min 89.085 ppb m

response 187333

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.417min 92.221 ppb

response 170806

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



MP 05-26-15 pm 5/26/15

(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.416min 85.460 ppb m

response 158283

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.755min 89.469 ppb

response 399898

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23.2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MLP 05-26-15 PM 5/26/15

(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.754min 85.713 ppb m

response 383111

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.097min 87.674 ppb

response 184974

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



MLP 05-26-15

PM 5/26/15

(4) 1,2-DINITROBENZENE (S)

0.000min 0.000 ppb

response 0

(4) 1,2-DINITROBENZENE #2 (S)

8.096min 80.762 ppb m

response 170391

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.433min 88.015 ppb

response 395834

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000017.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 19:29:17
Operator : MP
Sample : 8330MXA_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4232 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:23 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MHP 05-26-15 PM 5/26/15

(5) 1,3-DINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(5) 1,3-DINITROBENZENE #2 (TM)

8.432min 85.031 ppb m

response 382410

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:08:48
Operator : MP
Sample : 8330MXB_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4288 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:40:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
	Spiked Amount 62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.
2)	TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6)	TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
7)	TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8)	TM NITROGLYC...	12.070	0.000	71317	0	78.310	N.D. d#
9)	TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10)	TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11)	TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13)	TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14)	TM 3,5-DINIT...	0.000	16.571	0	165081	N.D.	82.703 #
15)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16)	TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17)	TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18)	TM PETN	22.803	0.000	69644	0	76.325m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:08:48
Operator : MP
Sample : 8330MXB_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4288 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:40:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



STE2_150522.M Sat May 23 09:40:40 2015 WALDORF Page: 2

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:08:48
Operator : MP
Sample : 8330MXB_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4288 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:26 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





QEdit

(18) PETN (TM)

22.803min 67.152 ppb

response 61273

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000018.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:08:48
Operator : MP
Sample : 8330MXB_CB 0.10 PPM 05/22/15
Misc :
ALS Vial : 4288 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:26 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(18) PETN (TM)

22.803min 76.325 ppb m

response 69644

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:39:49 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	8.109	0	369362	N.D.	175.070 #
	Spiked Amount 62.500			Recovery	=	0.00%	280.11%
	Target Compounds						
1)	TM HMX	0.000	4.082	0	390500	N.D.	185.700m#
2)	TM RDX	0.000	6.425	0	327561	N.D.	176.857m#
3)	TM 1,3,5-TRI...	0.000	6.764	0	791378	N.D.	177.055m#
5)	TM 1,3-DINIT...	0.000	8.444	0	805920	N.D.	179.200 #
6)	TM NITROBENZENE	0.000	9.184	0	448189	N.D.	174.961 #
7)	TM 2,4,6-TRI...	0.000	11.459	0	411367	N.D.	162.038 #
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9)	TM TETRYL	11.994	11.993	302328	291196	NoCal	176.771 #
10)	TM 2,6-DINIT...	0.000	13.467	0	246741	N.D.	167.047 #
11)	TM 2,4-DINIT...	0.000	14.072	0	426389	N.D.	169.233 #
12)	TM 2-NITROTO...	0.000	15.478	0	171906	N.D.	163.051 #
13)	TM 4-NITROTO...	0.000	16.897	0	151010	N.D.	172.424 #
14)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15)	TM 3-NITROTO...	0.000	18.081	0	182896	N.D.	171.055 #
16)	TM 4-AMINO-2...	0.000	18.671	0	272848	N.D.	194.758 #
17)	TM 2-AMINO-4...	0.000	19.625	0	354480	N.D.	173.896 #
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:39:49 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.083min 195.469 ppb

response 411042

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



MLP 05-26-15 M5/26/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

4.082min 185.700 ppb m

response 390500

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0=I





(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.426min 185.202 ppb

response 343017

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r



MP 05-26-15 PM 5/26/15

(2) RDX (TM)

0.000min 0.000 ppb

response 0

(2) RDX #2 (TM)

6.425min 176.857 ppb m

response 327561

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.765min 182.230 ppb

response 814511

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000019.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 20:48:19
Operator : MP
Sample : 8330MXA_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4289 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:29 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





MP 05-26-15 PM 5/26/15

(3) 1,3,5-TRINITROBENZENE (TM)

0.000min 0.000 ppb

response 0

(3) 1,3,5-TRINITROBENZENE #2 (TM)

6.764min 177.055 ppb m

response 791378

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:27:50
Operator : MP
Sample : 8330MXB_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4290 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:37:53 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.049	0.000	147567	0	162.036	N.D. d#
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.545	0	338483	N.D.	169.574 #
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.768	0.000	151578	0	166.120m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:27:50
Operator : MP
Sample : 8330MXB_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4290 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:37:53 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:27:50
Operator : MP
Sample : 8330MXB_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4290 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:32 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(18) PETN (TM)

22.776min 154.013 ppb

response 140531

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000020.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 21:27:50
Operator : MP
Sample : 8330MXB_CB 0.20 PPM 05/22/15
Misc :
ALS Vial : 4290 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:32 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-





(18) PETN (TM)

22.768min 166.120 ppb m

response 151578

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:07:21
Operator : MP
Sample : 8330MXA_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4291 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:36:44 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.105	0	879728	N.D.	416.972 #
Spiked Amount 62.500			Recovery	=	0.00%	667.16%
Target Compounds						
1) TM HMX	0.000	4.078	0	976770	N.D.	464.498 #
2) TM RDX	0.000	6.421	0	806821	N.D.	435.619 #
3) TM 1,3,5-TRI...	0.000	6.759	0	1966354	N.D.	439.931 #
5) TM 1,3-DINIT...	0.000	8.439	0	1956143	N.D.	434.957 #
6) TM NITROBENZENE	0.000	9.178	0	1093580	N.D.	426.904 #
7) TM 2,4,6-TRI...	0.000	11.450	0	1003936	N.D.	395.451 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.986	11.985	592326	722593	NoCal	438.651 #
10) TM 2,6-DINIT...	13.458	13.457	927172	598770	NoCal	405.376 #
11) TM 2,4-DINIT...	0.000	14.061	0	1040938	N.D.	413.147 #
12) TM 2-NITROTO...	0.000	15.466	0	416863	N.D.	395.391 #
13) TM 4-NITROTO...	0.000	16.881	0	367088	N.D.	419.141 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	18.068	0	443247	N.D.	414.550 #
16) TM 4-AMINO-2...	0.000	18.667	0	657332	N.D.	469.201 #
17) TM 2-AMINO-4...	0.000	19.622	0	851433	N.D.	417.685 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000021.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:07:21
Operator : MP
Sample : 8330MXA_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4291 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:36:44 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:46:51
Operator : MP
Sample : 8330MXB_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4292 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:35:42 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
	Spiked Amount 62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.
2)	TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6)	TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7)	TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8)	TM NITROGLYC...	12.050	0.000	368668	0	404.817	N.D. d#
9)	TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10)	TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11)	TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13)	TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14)	TM 3,5-DINIT...	0.000	16.548	0	841116	N.D.	421.385 #
15)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16)	TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17)	TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18)	TM PETN	22.762	0.000	382271	0	418.944m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:46:51
Operator : MP
Sample : 8330MXB_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4292 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:35:42 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:46:51
Operator : MP
Sample : 8330MXB_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4292 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(18) PETN (TM)

22.763min 398.817 ppb

response 363907

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000022.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 22:46:51
Operator : MP
Sample : 8330MXB_CB 0.50 PPM 05/22/15
Misc :
ALS Vial : 4292 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:38 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



(18) PETN (TM)

22.762min 418.944 ppb m

response 382271

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000023.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 23:26:22
Operator : MP
Sample : 8330MXA_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4293 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:34:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.103	0	1733160	N.D.	821.480 #
Spiked Amount 62.500			Recovery =		0.00%	1314.37%
Target Compounds						
1) TM HMX	0.000	4.076	0	1915374	N.D.	910.846 #
2) TM RDX	0.000	6.419	0	1584395	N.D.	855.448 #
3) TM 1,3,5-TRI...	0.000	6.755	0	3897389	N.D.	871.960 #
5) TM 1,3-DINIT...	0.000	8.437	0	3879893	N.D.	862.711 #
6) TM NITROBENZENE	0.000	9.177	0	2146298	N.D.	837.857 #
7) TM 2,4,6-TRI...	0.000	11.454	0	1991874	N.D.	784.599 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.994	11.993	1524497	1456381	NoCal	884.099 #
10) TM 2,6-DINIT...	13.465	13.464	1616524	1211988	NoCal	820.534 #
11) TM 2,4-DINIT...	0.000	14.067	0	2092031	N.D.	830.325 #
12) TM 2-NITROTO...	0.000	15.472	0	841920	N.D.	798.554 #
13) TM 4-NITROTO...	0.000	16.885	0	741632	N.D.	846.795 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	18.073	0	881637	N.D.	824.558 #
16) TM 4-AMINO-2...	0.000	18.677	0	1297906	N.D.	926.440 #
17) TM 2-AMINO-4...	0.000	19.631	0	1678327	N.D.	823.332 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000023.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 22-May-2015, 23:26:22
Operator : MP
Sample : 8330MXA_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4293 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:34:17 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000024.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:05:53
Operator : MP
Sample : 8330MXB_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4294 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:33:30 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.057	12.058	732317	11067	804.122	NoCal #
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.552	0	1652281	N.D.	827.765 #
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.770	22.770	761676	14524	834.745m	NoCal #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000024.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:05:53
Operator : MP
Sample : 8330MXB_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4294 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:33:30 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000024.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:05:53
Operator : MP
Sample : 8330MXB_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4294 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:44 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(18) PETN (TM)

22.770min 806.856 ppb

response 736228

(18) PETN #2 (TM)

22.770min 0.000 ppb

response 14524

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000024.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:05:53
Operator : MP
Sample : 8330MXB_CB 1.0 PPM 05/22/15
Misc :
ALS Vial : 4294 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:44 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MHP 05-26-15
PM5/26/15

(18) PETN (TM)

22.770min 834.745 ppb m

response 761676

(18) PETN #2 (TM)

22.770min 0.000 ppb

response 14524

(+) = Expected Retention Time

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000025.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:45:24
Operator : MP
Sample : 8330MXA_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4295 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:32:47 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.102	0	3543855	N.D.	1679.710 #
Spiked Amount 62.500			Recovery	=	0.00%	2687.54%
Target Compounds						
1) TM HMX	0.000	4.070	0	3887061	N.D.	1848.472 #
2) TM RDX	0.000	6.417	0	3207829	N.D.	1731.975 #
3) TM 1,3,5-TRI...	0.000	6.755	0	7939225	N.D.	1776.238 #
5) TM 1,3-DINIT...	0.000	8.437	0	7959675	N.D.	1769.868 #
6) TM NITROBENZENE	0.000	9.174	0	4377753	N.D.	1708.958 #
7) TM 2,4,6-TRI...	0.000	11.451	0	4093556	N.D.	1612.453 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.989	11.988	3002965	2907952	NoCal	1765.277 #
10) TM 2,6-DINIT...	13.461	13.460	3322547	2479312	NoCal	1678.531 #
11) TM 2,4-DINIT...	0.000	14.063	0	4282975	N.D.	1699.909 #
12) TM 2-NITROTO...	0.000	15.467	0	1710462	N.D.	1622.358 #
13) TM 4-NITROTO...	0.000	16.876	0	1514123	N.D.	1728.826 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	18.066	0	1790750	N.D.	1674.813 #
16) TM 4-AMINO-2...	0.000	18.668	0	2625200	N.D.	1873.858 #
17) TM 2-AMINO-4...	0.000	19.623	0	3392979	N.D.	1664.484 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000025.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 00:45:24
Operator : MP
Sample : 8330MXA_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4295 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:32:47 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 01:24:56
Operator : MP
Sample : 8330MXB_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4296 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:31:36 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.061	0.000	1482681	0	1628.060	N.D. d#
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.541	0	3311418	N.D.	1658.965 #
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.767	22.769	1542843	30238	1690.853m	NoCal #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 01:24:56
Operator : MP
Sample : 8330MXB_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4296 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:31:36 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 01:24:56
Operator : MP
Sample : 8330MXB_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4296 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:50 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





QEdit

(18) PETN (TM)

22.769min 1638.003 ppb

response 1494619

(18) PETN #2 (TM)

22.769min 0.000 ppb

response 30238

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000026.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 01:24:56
Operator : MP
Sample : 8330MXB_CB 2.0 PPM 05/22/15
Misc :
ALS Vial : 4296 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 09:30:50 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Fri May 15 08:45:00 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



(18) PETN (TM)

22.767min 1690.853 ppb m

response 1542843

(18) PETN #2 (TM)

22.769min 0.000 ppb

response 30238

(+) = Expected Retention Time

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 76380
Date Analyzed: 23-May-2015, 04:42:30
Instrument: Statler
Initial Cal. Date: 05/22/15
Data File: 0522_000031-33.D

		Compound	MEAN	CCRF	%D		%Drift
1	TM	NITROGLYCERIN	18.6	18.4	0.69	TM	
2	TM	PETN	18.2	18.5	1.9	TM	
3		Signal #2					
4	TM	HMX	46.3	47.1	1.8	TM	
5	TM	RDX	39.8	39.8	0.03	TM	
6	TM	1,3-DINITROBENZENE	96.6	94.7	1.9	TM	
7	TM	NITROBENZENE	54.2	53.8	0.71	TM	
8	TM	2,4-DINITROTOLUENE	51.7	53.0	2.4	TM	
9	TM	3,5-DINITROANILINE	41.3	42.8	3.6	TM	
10	TM	3-NITROTOLUENE	22.9	22.8	0.74	TM	
11	TM	4-AMINO-2,6-DINITROTOLUEN	34.2	34.3	0.55	TM	
12	TM	1,3,5-TRINITROBENZENE	97.9	89.7	8.4	TM	
13	S	1,2-DINITROBENZENE	44.3	44.5	0.48	S	
14	TM	2,4,6-TRINITROTOLUENE	50.2	49.8	0.83	TM	
15	TM	2,6-DINITROTOLUENE	30.0	31.8	6.1	TM	
16	TM	2-NITROTOLUENE	21.2	21.7	2.8	TM	
17	TM	4-NITROTOLUENE	18.3	18.5	1.6	TM	
18	TM	2-AMINO-4,6-DINITROTOLUEN	43.0	38.7	10	TM	
19	TM	TETRYL	36.6	32.0	12	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000031.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 04:42:30
Operator : MP
Sample : 8330MXA_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4352 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:23:15 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D.	N.D. d
Spiked Amount 62.500			Recovery =		0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	4.073	0	1883754	N.D.	1017.745 #
2) TM RDX	0.000	6.415	0	1590397	N.D.	999.716 #
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5) TM 1,3-DINIT...	0.000	8.427	0	3787591	N.D.	980.606 #
6) TM NITROBENZENE	0.000	9.161	0	2153367	N.D.	992.865 #
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D.	N.D. d
8) TM NITROGLYC...	12.040	0.000	736934	0	993.140	N.D. d#
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	14.047	0	2118885	N.D.	1024.231 #
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.551	0	1712548	N.D.	1035.686 #
15) TM 3-NITROTO...	0.000	18.047	0	910222	N.D.	992.617 #
16) TM 4-AMINO-2...	0.000	18.671	0	1373815	N.D.	1005.467 #
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000031.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 04:42:30
Operator : MP
Sample : 8330MXA_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4352 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:23:15 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



STE2_150522.M Sat May 23 10:23:31 2015 WALDORF Page: 2

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000032.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 05:22:00
Operator : MP
Sample : 8330MXB_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4353 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:24:52 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.108	0	1779983	N.D.	1004.829 #
Spiked Amount 62.500			Recovery	=	0.00%	1607.73%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D.	N.D. d
3) TM 1,3,5-TRI...	0.000	6.758	0	3587912	N.D.	916.157 #
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6) TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D. d
7) TM 2,4,6-TRI...	0.000	11.451	0	1991880	N.D.	991.740 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	13.467	0	1273233	N.D.	1061.031 #
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	15.471	0	869951	N.D.	1027.863 #
13) TM 4-NITROTO...	0.000	16.886	0	741713	N.D.	1015.531 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	19.659	0	1547413	N.D.	899.911 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000032.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 05:22:00
Operator : MP
Sample : 8330MXB_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4353 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:24:52 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000033.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 06:01:31
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4354 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:26:04 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery =		0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	0.000	12.005	0	1280291	N.D. d	875.551
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.767	0.000	739734	0	1018.644m	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000033.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 06:01:31
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4354 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:26:04 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000033.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 06:01:31
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4354 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:22:41 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-r





QEdit

(18) PETN (TM)

22.769min 982.640 ppb

response 713587

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000033.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 06:01:31
Operator : MP
Sample : PETN_TETRYL_SS 1.0 PPM 05/22/15
Misc :
ALS Vial : 4354 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 10:22:41 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:19:47 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ



MP 05-26-15 5/26/15



QEdit

(18) PETN (TM)

22.767min 1018.644 ppb m

response 739734

(18) PETN #2 (TM)

0.000min 0.000 ppb

response 0

(+) = Expected Retention Time

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 76380
Date Analyzed: 23-May-2015, 21:51:19
Instrument: Statler
Initial Cal. Date: 05/22/15
Data File: 0522_000058-59.D

		Compound	MEAN	CCRF	%D		%Drift
1	TM	NITROGLYCERIN	18.6	18.0	2.8	TM	
2	TM	PETN	18.2	18.8	3.4	TM	
3		Signal #2					
4	TM	3,5-DINITROANILINE	41.3	41.1	0.51	TM	
5	TM	HMX	46.3	48.3	4.3	TM	
6	TM	RDX	39.8	39.6	0.49	TM	
7	TM	1,3,5-TRINITROBENZENE	97.9	97.1	0.82	TM	
8	S	1,2-DINITROBENZENE	44.3	42.9	3.1	S	
9	TM	1,3-DINITROBENZENE	96.6	96.4	0.13	TM	
10	TM	NITROBENZENE	54.2	52.5	3.3	TM	
11	TM	2,4,6-TRINITROTOLUENE	50.2	50.1	0.30	TM	
12	TM	TETRYL	36.6	36.6	0.14	TM	
13	TM	2,6-DINITROTOLUENE	30.0	30.4	1.3	TM	
14	TM	2,4-DINITROTOLUENE	51.7	52.6	1.8	TM	
15	TM	2-NITROTOLUENE	21.2	20.8	1.9	TM	
16	TM	4-NITROTOLUENE	18.3	18.6	1.7	TM	
17	TM	3-NITROTOLUENE	22.9	21.8	4.9	TM	
18	TM	4-AMINO-2,6-DINITROTOLUEN	34.2	33.0	3.5	TM	
19	TM	2-AMINO-4,6-DINITROTOLUEN	43.0	41.8	2.8	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000058.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:11:50
Operator : MP
Sample : 8330MXA_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8519 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:42:46 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	8.086	0	1715928	N.D.	968.669 #
	Spiked Amount 62.500			Recovery	=	0.00%	1549.87%
	Target Compounds						
1)	TM HMX	0.000	4.065	0	1930177	N.D.	1042.827 #
2)	TM RDX	0.000	6.403	0	1583001	N.D.	995.067 #
3)	TM 1,3,5-TRI...	0.000	6.749	0	3884253	N.D.	991.826 #
5)	TM 1,3-DINIT...	0.000	8.424	0	3857480	N.D.	998.700 #
6)	TM NITROBENZENE	0.000	9.163	0	2098329	N.D.	967.488 #
7)	TM 2,4,6-TRI...	0.000	11.433	0	2002440	N.D.	996.997 #
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9)	TM TETRYL	11.954	11.953	1251814	1464384	NoCal	1001.446 #
10)	TM 2,6-DINIT...	13.437	13.436	1805332	1215356	NoCal	1012.800 #
11)	TM 2,4-DINIT...	0.000	14.037	0	2105648	N.D.	1017.832 #
12)	TM 2-NITROTO...	0.000	15.446	0	830345	N.D.	981.068 #
13)	TM 4-NITROTO...	0.000	16.856	0	743016	N.D.	1017.315 #
14)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15)	TM 3-NITROTO...	0.000	18.047	0	872020	N.D.	950.957 #
16)	TM 4-AMINO-2...	0.000	18.613	0	1318850	N.D.	965.239 #
17)	TM 2-AMINO-4...	0.000	19.573	0	1671131	N.D.	971.861 #
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000058.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:11:50
Operator : MP
Sample : 8330MXA_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8519 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:42:46 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:51:19
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:41:50 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4) S	1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
	Spiked Amount 62.500			Recovery	=	0.00%	0.00%
	Target Compounds						
1) TM	HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM	RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM	1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM	1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM	NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM	2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM	NITROGLYC...	12.049	12.057	721486	11670	972.321	NoCal #
9) TM	TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM	2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM	2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM	2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM	4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM	3,5-DINIT...	0.000	16.497	0	1645103	N.D.	994.898 #
15) TM	3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM	4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM	2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM	PETN	22.753	22.751	750797	15401	1033.879m	NoCal #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:51:19
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:41:50 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:51:19
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:41:27 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





(18) PETN (TM)

22.756min 982.680 ppb

response 713616

(18) PETN #2 (TM)

22.751min 0.000 ppb

response 15401

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000059.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 21:51:19
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:41:27 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MM 05-24-15 PM 5/26/15



QEdit

(18) PETN (TM)

22.753min 1033.879 ppb m

response 750797

(18) PETN #2 (TM)

22.751min 0.000 ppb

response 15401

(+) = Expected Retention Time

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No:
Matrix:

SDG No: 76380
Date Analyzed: 24-May-2015, 07:04:31
Instrument: Statler
Initial Cal. Date: 05/22/15
Data File: 0522_000072-73.D

		Compound	MEAN	CCRF	%D		%Drift
1	TM	NITROGLYCERIN	18.6	17.9	3.6	TM	
2	TM	PETN	18.2	18.7	3.0	TM	
3		Signal #2					
4	TM	3,5-DINITROANILINE	41.3	40.8	1.4	TM	
5	TM	HMX	46.3	47.5	2.7	TM	
6	TM	RDX	39.8	39.1	1.6	TM	
7	TM	1,3,5-TRINITROBENZENE	97.9	97.4	0.47	TM	
8	S	1,2-DINITROBENZENE	44.3	42.9	3.1	S	
9	TM	1,3-DINITROBENZENE	96.6	96.5	0.04	TM	
10	TM	NITROBENZENE	54.2	51.1	5.7	TM	
11	TM	2,4,6-TRINITROTOLUENE	50.2	50.1	0.23	TM	
12	TM	TETRYL	36.6	36.2	1.0	TM	
13	TM	2,6-DINITROTOLUENE	30.0	30.3	1.2	TM	
14	TM	2,4-DINITROTOLUENE	51.7	52.4	1.4	TM	
15	TM	2-NITROTOLUENE	21.2	19.9	5.9	TM	
16	TM	4-NITROTOLUENE	18.3	17.9	1.9	TM	
17	TM	3-NITROTOLUENE	22.9	21.1	7.9	TM	
18	TM	4-AMINO-2,6-DINITROTOLUEN	34.2	32.0	6.4	TM	
19	TM	2-AMINO-4,6-DINITROTOLUEN	43.0	41.6	3.3	TM	
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40							

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000072.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 06:25:01
Operator : MP
Sample : 8330MXA_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8519 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:46:39 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	8.100	0	1716291	N.D.	968.874 #
Spiked Amount 62.500			Recovery	=	0.00%	1550.20%
Target Compounds						
1) TM HMX	0.000	4.081	0	1900608	N.D.	1026.851 #
2) TM RDX	0.000	6.424	0	1564941	N.D.	983.715 #
3) TM 1,3,5-TRI...	0.000	6.752	0	3897932	N.D.	995.319 #
5) TM 1,3-DINIT...	0.000	8.433	0	3860830	N.D.	999.567 #
6) TM NITROBENZENE	0.000	9.169	0	2045340	N.D.	943.056 #
7) TM 2,4,6-TRI...	0.000	11.442	0	2003922	N.D.	997.735 #
8) TM NITROGLYC...	0.000	0.000	0	0	N.D. d	N.D. d
9) TM TETRYL	11.986	11.984	1162837	1447662	NoCal	990.011 #
10) TM 2,6-DINIT...	13.458	13.457	1808177	1213915	NoCal	1011.599 #
11) TM 2,4-DINIT...	0.000	14.062	0	2097756	N.D.	1014.018 #
12) TM 2-NITROTO...	0.000	15.463	0	796481	N.D.	941.057 #
13) TM 4-NITROTO...	0.000	16.878	0	716312	N.D.	980.753 #
14) TM 3,5-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
15) TM 3-NITROTO...	0.000	18.064	0	844422	N.D.	920.860 #
16) TM 4-AMINO-2...	0.000	18.694	0	1279490	N.D.	936.433 #
17) TM 2-AMINO-4...	0.000	19.649	0	1662692	N.D.	966.952 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000072.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 06:25:01
Operator : MP
Sample : 8330MXA_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8519 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:46:39 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000073.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 07:04:31
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:45:28 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
Spiked Amount 62.500			Recovery	=	0.00%	0.00%
Target Compounds						
1) TM HMX	0.000	0.000	0	0	N.D.	N.D.
2) TM RDX	0.000	0.000	0	0	N.D. d	N.D. d
3) TM 1,3,5-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
5) TM 1,3-DINIT...	0.000	0.000	0	0	N.D. d	N.D. d
6) TM NITROBENZENE	0.000	0.000	0	0	N.D. d	N.D. d
7) TM 2,4,6-TRI...	0.000	0.000	0	0	N.D. d	N.D. d
8) TM NITROGLYC...	12.045	12.053	715487	11247	964.236	NoCal #
9) TM TETRYL	0.000	0.000	0	0	N.D. d	N.D. d
10) TM 2,6-DINIT...	0.000	0.000	0	0	N.D.	N.D.
11) TM 2,4-DINIT...	0.000	0.000	0	0	N.D.	N.D.
12) TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
13) TM 4-NITROTO...	0.000	0.000	0	0	N.D. d	N.D. d
14) TM 3,5-DINIT...	0.000	16.572	0	1630858	N.D.	986.283 #
15) TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16) TM 4-AMINO-2...	0.000	0.000	0	0	N.D.	N.D.
17) TM 2-AMINO-4...	0.000	0.000	0	0	N.D.	N.D.
18) TM PETN	22.741	22.748	747792	14856	1029.741m	NoCal #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000073.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 07:04:31
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:45:28 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-micron



Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000073.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 07:04:31
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:45:03 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-µ





QEdit

(18) PETN (TM)

22.743min 957.286 ppb

response 695176

(18) PETN #2 (TM)

22.748min 0.000 ppb

response 14856

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\STATLER\CHEM32\1\DATA\150522\150522\
Data File : 0522_000073.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 24-May-2015, 07:04:31
Operator : MP
Sample : 8330MXB_CCV 1.0 PPM 05/23/15
Misc :
ALS Vial : 8520 Sample Multiplier: 1

Integration File signal 1: STE2_Signal_1.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 24 12:45:03 2015
Quant Method : Z:\STATLER\CHEM32\1\DATA\150522\150522\STE2_150522.M
Quant Title : 8330 A + B - Statler
QLast Update : Sat May 23 10:29:51 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : Acclaim E2 150x3. Signal #2 Info : Acclaim E2 150x3.0mm 3.0-



MP 05-24-15 5/26/15



QEdit

(18) PETN (TM)

22.741min 1029.741 ppb m

response 747792

(18) PETN #2 (TM)

22.748min 0.000 ppb

response 14856

(+) = Expected Retention Time

EPA 8330
EXPLOSIVES

Raw Data

APPL, INC.

Method Blank
EPA 8330B SOIL

Blank Name/QCG: **150520S-16137 - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	2,4-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	2,6-DINITROTOLUENE	0.200 U	0.50	0.200	0.083	mg/kg	05/20/15	05/23/15
BLANK	HMX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	NITROGLYCERIN	0.200 U	0.50	0.200	0.085	mg/kg	05/20/15	05/23/15
BLANK	RDX	0.200 U	0.50	0.200	0.080	mg/kg	05/20/15	05/23/15
BLANK	TETRYL	0.200 U	0.50	0.200	0.091	mg/kg	05/20/15	05/23/15
BLANK	SURROGATE: 1,2-DINITROBENZ	95.0	70-130			%	05/20/15	05/23/15

Quant Method: W150522.M
Run #: 0522_000051
Instrument: Waldorf
Sequence: 150522
Initials: MP

GC SC-Blank-REG MDLs
Printed: 05/26/15 10:52:59 AM

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000051.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 12:25:49
Operator : MP
Sample : 150520ABLKS 7.992 DF 05/20/15
Misc : Soil
ALS Vial : 4227 Sample Multiplier: 7.99201

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:31:13 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
	System Monitoring Compounds						
4)	S 1,2-DINIT...	0.000	3.535	0	922186	N.D.	569.571 #
	Spiked Amount 599.401			Recovery	=	0.00%	95.02%
	Target Compounds						
1)	TM HMX	0.000	0.000	0	0	N.D.	N.D.
2)	TM RDX	0.000	0.000	0	0	N.D.	N.D.
3)	TM 1,3,5-TRI...	0.000	0.000	0	0	N.D.	N.D.
5)	TM 1,3-DINIT...	0.000	0.000	0	0	N.D.	N.D.
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
7)	TM NITROBENZENE	0.000	0.000	0	0	N.D.	N.D.
8)	TM NITROGLYC...	0.000	0.000	0	0	N.D. d <DL	N.D.
9)	TM TETRYL	6.426	0.000	5874	0	NoCal	N.D.
10)	TM 2,4,6-TRI...	6.903	0.000	11606	0	NoCal	N.D.
11)	TM 2-AMINO-4...	7.130	0.000	5700	0	NoCal	N.D.
12)	TM 4-AMINO-2...	7.410	0.000	19856	0	NoCal	N.D.
13)	TM 2,4-DINIT...	8.656	0.000	11827	0	NoCal	N.D.
14)	TM 2,6-DINIT...	8.963	0.000	7121	0	NoCal	N.D.
15)	TM 2-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
16)	TM 4-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
17)	TM 3-NITROTO...	0.000	0.000	0	0	N.D.	N.D.
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP 05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000051.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 12:25:49
Operator : MP
Sample : 150520ABLKS 7.992 DF 05/20/15
Misc : Soil
ALS Vial : 4227 Sample Multiplier: 7.99201

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 26 07:31:13 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Laboratory Control Spike Recovery
EPA 8330B SOIL

APPL ID: **150520S-16137 LCS - 197310**
Batch ID: #83BJU-150520A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level mg/kg	SPK Result mg/kg	SPK % Recovery	Recovery Limits
2,4-DINITROTOLUENE	1.99	1.83	92.0	80-125
2,6-DINITROTOLUENE	1.99	1.84	92.5	80-120
HMX	1.99	1.90	95.5	75-125
NITROGLYCERIN	1.99	1.83	92.0	68-131
RDX	1.99	1.83	92.0	70-135
TETRYL	1.99	1.97	99.0	10-150
SURROGATE: 1,2-DINITROBENZENE (S	1.99	1.82	91.5	70-130

Comments: _____

Primary	SPK
Quant Method :	W150522.M
Extraction Date :	05/20/15
Analysis Date :	05/23/15
Instrument :	Waldorf
Run :	0522_000048
Initials :	MP

Printed: 05/26/15 10:52:51 AM
APPL Standard LCS

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000048.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 10:33:19
Operator : MP
Sample : 150520ALCS-1 7.952 DF 05/20/15
Misc : Soil
ALS Vial : 4224 Sample Multiplier: 7.95229

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:45:09 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb
System Monitoring Compounds						
4) S 1,2-DINIT...	0.000	3.538	0	2966342	N.D.	1822.999 #
Spiked Amount 596.421			Recovery =		0.00%	305.66%
Target Compounds						
1) TM HMX	0.000	0.917	0	6196311	N.D.	1896.787m#
2) TM RDX	0.000	1.730	0	3909361	N.D.	1825.318 #
3) TM 1,3,5-TRI...	0.000	2.819	0	7040913	N.D.	1809.117 #
5) TM 1,3-DINIT...	0.000	3.709	0	6516140	N.D.	1777.716 #
6) TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.
7) TM NITROBENZENE	0.000	5.028	0	2944254	N.D.	1753.664 #
8) TM NITROGLYC...	6.041	6.035	1209690	27493	1829.727	NoCal #
9) TM TETRYL	6.383	6.383	3717267	2273302	NoCal	1974.320 #
10) TM 2,4,6-TRI...	6.875	6.875	3755806	2921104	NoCal	1856.565 #
11) TM 2-AMINO-4...	7.068	7.068	3897276	2598275	NoCal	1837.677 #
12) TM 4-AMINO-2...	7.412	7.412	3287013	1770809	NoCal	1739.376 #
13) TM 2,4-DINIT...	8.551	8.550	2015770	2719081	NoCal	1831.228 #
14) TM 2,6-DINIT...	8.909	8.908	2154262	1503057	NoCal	1838.043 #
15) TM 2-NITROTO...	0.000	12.245	0	975996	N.D.	1731.112 #
16) TM 4-NITROTO...	0.000	12.845	0	1000097	N.D.	1793.265 #
17) TM 3-NITROTO...	0.000	13.735	0	1206249	N.D.	1811.408 #
18) TM PETN	0.000	0.000	0	0	N.D.	N.D.

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP 15-26-15

Algorithm Check

RDX

$\frac{3909361}{425.6\ (40)} \times 7.95229 = 1825.29$ ✓ MP 05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000048.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 10:33:19
Operator : MP
Sample : 150520ALCS-1 7.952 DF 05/20/15
Misc : Soil
ALS Vial : 4224 Sample Multiplier: 7.95229

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:45:09 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000048.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 10:33:19
Operator : MP
Sample : 150520ALCS-1 7.952 DF 05/20/15
Misc : Soil
ALS Vial : 4224 Sample Multiplier: 7.95229

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 11:17:38 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0r



(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.918min 1951.356 ppb

response 6374574

(+) = Expected Retention Time

Quantitation Report (Qedit)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000048.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 10:33:19
Operator : MP
Sample : 150520ALCS-1 7.952 DF 05/20/15
Misc : Soil
ALS Vial : 4224 Sample Multiplier: 7.95229

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 11:17:38 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0



MMP 05-26-15

TM6/2/15

(1) HMX (TM)

0.000min 0.000 ppb

response 0

(1) HMX #2 (TM)

0.917min 1896.787 ppb m

response 6196311

(+) = Expected Retention Time

Matrix Spike Recoveries

EPA 8330B SOIL

APPL ID: **150520S-16137 MS - 197310**
Batch ID: #83BJU-150520A
Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Compound Name	SPK Lvl	DUP Lvl	Matrix Result	SPK Result	DUP Result	SPK % Recovery	DUP % Recovery	Recovery Limits	RPD %	RPD Limits
	mg/kg	mg/kg	mg/kg	mg/kg	mg/kg					
2,4-DINITROTOLUENE	1.99	2.00	ND	1.86	1.85	93.5	92.5	80-125	1.1	20
2,6-DINITROTOLUENE	1.99	2.00	ND	1.87	1.86	94.0	93.0	80-120	1.1	20
HMX	1.99	2.00	ND	1.85	1.83	93.0	91.5	75-125	1.6	20
NITROGLYCERIN	1.99	2.00	ND	1.90	1.93	95.5	96.5	68-131	1	20
RDX	1.99	2.00	ND	1.69	1.69	84.9	84.5	70-135	0.47	20
TETRYL	1.99	2.00	ND	1.77	1.81	88.9	90.5	10-150	1.8	20
SURROGATE: 1,2-DINITROBENZEN	1.99	2.00	NA	1.86	1.86	93.5	93.0	70-130		

Comments: _____

Primary	SPK	DUP
Quant Method :	W150522.M	W150522.M
Extraction Date :	05/20/15	05/20/15
Analysis Date :	05/23/15	05/23/15
Instrument :	Waldorf	Waldorf
Run :	0522_000049	0522_000050
Initials :	MP	

Printed: 05/26/15 10:50:26 AM
APPL MSD SCII

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000049.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 11:10:49
Operator : MP
Sample : AZ16137S01_MS-1 7.960 DF 05/20/15
Misc : Soil
ALS Vial : 4225 Sample Multiplier: 7.9602

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:43:56 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.531	0	3029293	N.D.	1863.539	#
	Spiked Amount 597.015			Recovery	=	0.00%	312.14%	
	Target Compounds							
1)	TM HMX	0.000	0.914	0	6030718	N.D.	1847.933	#
2)	TM RDX	0.000	1.724	0	3623488	N.D.	1693.524	#
3)	TM 1,3,5-TRI...	0.000	2.814	0	7079698	N.D.	1820.892	#
5)	TM 1,3-DINIT...	0.000	3.703	0	6609104	N.D.	1804.872	#
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
7)	TM NITROBENZENE	0.000	5.022	0	2996173	N.D.	1786.364	#
8)	TM NITROGLYC...	6.036	6.029	1256104	35418	1901.820	NoCal	#
9)	TM TETRYL	6.379	6.379	3319596	2031741	NoCal	1766.284	#
10)	TM 2,4,6-TRI...	6.871	6.871	3814752	2945149	NoCal	1873.709	#
11)	TM 2-AMINO-4...	7.069	7.068	3928978	2613028	NoCal	1849.949	#
12)	TM 4-AMINO-2...	7.412	7.412	3285808	1772872	NoCal	1743.135	#
13)	TM 2,4-DINIT...	8.549	8.549	2016718	2752788	NoCal	1855.773	#
14)	TM 2,6-DINIT...	8.908	8.907	2113262	1524071	NoCal	1865.594	#
15)	TM 2-NITROTO...	0.000	12.254	0	992265	N.D.	1761.719	#
16)	TM 4-NITROTO...	0.000	12.855	0	1014653	N.D.	1821.175	#
17)	TM 3-NITROTO...	0.000	13.747	0	1223882	N.D.	1839.715	#
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000049.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 11:10:49
Operator : MP
Sample : AZ16137S01_MS-1 7.960 DF 05/20/15
Misc : Soil
ALS Vial : 4225 Sample Multiplier: 7.9602

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:43:56 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000050.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 11:48:19
Operator : MP
Sample : AZ16137S01_MSD-1 7.984 DF 05/20/15
Misc : Soil
ALS Vial : 4226 Sample Multiplier: 7.98403

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:43:59 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron

	Compound	RT#1	RT#2	Resp#1	Resp#2	ppb	ppb	
	System Monitoring Compounds							
4)	S 1,2-DINIT...	0.000	3.534	0	3013479	N.D.	1859.360	#
	Spiked Amount 598.802			Recovery	=	0.00%	310.51%	
	Target Compounds							
1)	TM HMX	0.000	0.915	0	5968197	N.D.	1834.250	#
2)	TM RDX	0.000	1.725	0	3607525	N.D.	1691.110	#
3)	TM 1,3,5-TRI...	0.000	2.815	0	7060518	N.D.	1821.396	#
5)	TM 1,3-DINIT...	0.000	3.705	0	6574630	N.D.	1800.833	#
6)	TM 3,5-DINIT...	0.000	0.000	0	0	N.D.	N.D.	
7)	TM NITROBENZENE	0.000	5.026	0	2971848	N.D.	1777.165	#
8)	TM NITROGLYC...	6.041	6.032	1271025	30352	1930.173	NoCal	#
9)	TM TETRYL	6.385	6.385	3421789	2077569	NoCal	1811.532	#
10)	TM 2,4,6-TRI...	6.877	6.877	3796872	2917514	NoCal	1861.684	#
11)	TM 2-AMINO-4...	7.075	7.074	3932848	2602198	NoCal	1847.797	#
12)	TM 4-AMINO-2...	7.418	7.419	3293521	1768707	NoCal	1744.246	#
13)	TM 2,4-DINIT...	8.556	8.556	2017916	2741708	NoCal	1853.837	#
14)	TM 2,6-DINIT...	8.915	8.915	2102687	1517529	NoCal	1863.147	#
15)	TM 2-NITROTO...	0.000	12.260	0	989697	N.D.	1762.422	#
16)	TM 4-NITROTO...	0.000	12.862	0	1011133	N.D.	1820.290	#
17)	TM 3-NITROTO...	0.000	13.755	0	1214787	N.D.	1831.509	#
18)	TM PETN	0.000	0.000	0	0	N.D.	N.D.	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

MP
05-26-15

Quantitation Report (QT Reviewed)

Data Path : Z:\WALDORF\CHEM32\2\DATA\150522\150522\
Data File : 0522_000050.D
Signal(s) : Signal #1: DAD1B.ch Signal #2: DAD1A.ch
Acq On : 23-May-2015, 11:48:19
Operator : MP
Sample : AZ16137S01_MSD-1 7.984 DF 05/20/15
Misc : Soil
ALS Vial : 4226 Sample Multiplier: 7.98403

Integration File signal 1: Waldorf_Signal_1_110721.e
Integration File signal 2: Waldorf_Signal_2_110803.e
Quant Time: May 23 12:43:59 2015
Quant Method : Z:\WALDORF\CHEM32\2\DATA\150522\150522\W150522.M
Quant Title : 8330B - Soil - Waldorf
QLast Update : Sat May 23 09:16:20 2015
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 40uL
Signal #1 Phase : 214nm Signal #2 Phase: 254nm
Signal #1 Info : ZORBAX Extend-C18 Signal #2 Info : ZORBAX Extend-C18 100x3.0mm 1.8-micron



LCMS Standard Prep Log #103 Page 13

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT #	DATE / INITIALS
Tetryl_ss 1000 ug/mL							8-29-13 MP
Tetryl 1000 ug/mL Absolute Lot # 112310-32716							
I broke open the vial and transferred to							
an amber 1.8 mL Injection vial →							
Tetryl_ss 1000 ug/mL 08-29-13							
Tetryl_ss 10 ug/mL							
Tetryl_ss 1000 08-29-13	0.10	10.0	10.0	1:1 ACN/MeOH		DG445/52292	08-29-13 MP

WALDORF 08/31/13

8330_CB 1:1 1.0 PPM									MP
8330_MX-A	10.0 ppm	07/01/13	0.10 ml	1.0 mL	1.0 ppm	1:1ACN_DG445/MEOH_52292	08/31/13	MP	8-31-13
8330_MX-B	10.0 ppm	07/01/13							
8330_CB 1:1 0.1 PPM									
8330_CB 1:1 1.0 PPM	1.0 ppm	08/31/13	0.10 ml	1.0 mL	0.10 ppm	1:1ACN_DG445/MEOH_52292	08/31/13	MP	
8330CB 0.005 PPM									
8330_CB 1:1	0.1 PPM	08/31/13	0.05 ML	1.0 mL	0.005 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.010 PPM									
8330_CB 1:1	0.1 PPM	08/31/13	0.10 ML	1.0 mL	0.010 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.020 PPM									
8330_CB 1:1	0.1 PPM	08/31/13	0.20 ML	1.0 mL	0.020 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.05 PPM									
8330_CB 1:1	1.0 PPM	08/31/13	0.05 ML	1.0 mL	0.05 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.10 PPM									
8330_CB 1:1	1.0 PPM	08/31/13	0.10 ML	1.0 mL	0.10 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.20 PPM									
8330_CB 1:1	1.0 PPM	08/31/13	0.20 ML	1.0 mL	0.20 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330CB 0.50 PPM									
8330MXA	10.0 PPM	07/01/13	0.05 ML	1.0 mL	0.50 PPM	1 : 1 : 6 ACN : MeOH : MILLIPORE DG445/52292	08/31/13	MP	
8330MXB	10.0 PPM	07/01/13							

LC/MS STANDARD PREP LOG# 103 PAGE 65

065

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS

NOTE: THE FOLLOWING DOCUMENTATION WAS PERFORMED AS DATED: JF 12-19-13

1668_SS_GC 500 ng/ml 10/30/13:

PCB_CONG_2ND_SOURCE_STOCK 10 ug/ml 05/30/13 took 0.005 ml and added to ~~0.950~~ 0.095 JF ml 1668_ICL 100 ng/ml 10/27/13:

1668_SS_GC 50 ng/ml 10/30/13:

1668_SS_GC 500 ng/ml 10/30/13 took 0.010 ml and added to 0.090 ml 1668_ICL 100 ng/ml 10/27/13

	1,2-DNB 6Lit 5.0 ug/L				09-03-13		JF
1,2-DNB	1000	10-30-13	0.10	20.0	5.0	1:1 ACN/MeOH	12-23-13
	PA_IntStd 0.10 ug/L						
4-Nitrophenol	10.0	8-15-13	0.20	20	0.10	ACN/DI509	12-26-13 JF
	Ash Std 20 ug/L						
8321MXZ5-8	400	11-20-13	0.250	5.0	20	Methanol	12-26-13 JF
8321MX10-4	1000	8-29-13	0.100				
8321MX6-7	1000	9-27-13	0.100				
Isoxaben	1058	10-8-13	0.0945				
8141Mix	1000	9-5-13	0.100				
	8330MXA-SS 100 ug/L						12-27-13 JF
	8330MX-A 100 ug/L Absolute Part# 95093 Lot# 121311-30132 I broke open the Vial and transferred to an amber 1.8 ml injection Vial → 8330MXA_SS 100 ug/L 12-27-13						

GC/MS STANDARD PREP LOG# 103 PAGE 66

066

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	DATE / INITIALS
	8330MX-B-55 100 ug/mL				4 M 12-27-13		12-27-13 HP
	8330MX-B 100 ug/mL Absolute Part # 9509B, Lot # 21341051412-30759. I broke open the vial on a 12-27-13 HP transferred to an amber 1.8mL injection vial → 8330MX-B-55 100ug/mL 12-27-13						
	Ash std 0.40 DF						
Ash std	20	12-26-13	0.80	2.0	0.40 DF	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			
	Ash std 0.20 DF						
Ash std	20	12-26-13	0.40	2.0	0.20 DF	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			
	Ash std 0.10 DF						
Ash std	20	12-26-13	0.20	2.0	0.10	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			
	Ash std 0.08 DF						
Ash std	20	12-26-13	0.160	2.0	0.08	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			
	(Replaced 2 packings)						
	Ash std 0.06 DF						
Ash std	20	12-26-13	0.120	2.0	0.06	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			
	Ash std 0.04 DF						
Ash std	20	12-26-13	0.080	2.0	0.04	Methanol	53206	12-30-13 HP
HPLCFS	100	7-12-13	0.020	1.0			

LCMS STANDARD PREP LOG# 104 PAGE 8

008 STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	INITIALS
1,2-DNB	1000 ug/mL						04-23-14 KP

I broke open the vial labeled to the left and transferred to an amber 1.8 mL injection vial → 1,2-DNB 1000 ug/mL 04-23-14

1,2-Dinitrobenzene
Solution, 1,000 mg/L, 1 mL
016666-01
Lot # 216681 | Storage <10 Degrees C | Expiry 10/25/16
Solv: Acetonitrile
Made in the USA
1,2-Dinitrobenzene
Lot #: 216681 - 33241
Rec: 1/8/14 MFR exp. 10/25/16

1,2-DNB	100 ug/mL							KP
1,2-DNB	1000	04-23-14	1.0	10.0	100	1:1 ACN/MeOH	02-19-14	04-23-14

MX-8330R	1000 ug/mL							KP 04-23-14

AccuStandard
M-8330-R
Method 8330 - Explosives by HPLC
1000 µg/mL in MeOH:AcCN (1:1)
Lot: 211111132-01
Exp: Jun 6,2015
14 comps.
HIGHLY FLAMMABLE
Method 8330 - Explosives
Lot #: 211111132-01 - 33138
Rec: 11/18/13 MFR exp. 6/6/15
R LABORATORY USE ONLY
NING: This product contains nicals known to the State of Califor use cancer and birth defects or oth oductive harm.
STORAGE Freeze (<-10° C)
2 Danger

I broke open the vial labeled above and transferred to an amber 1 mL injection vial → MX-8330R 1000 ug/mL 04-23-14

MX-8330	100 ug/mL							KP
MX-8330R	1000	04-23-14	1.0	10.0	100	1:1 ACN/MeOH	02-19-14	04-23-14

1,2-DNB	5.0 ug/mL							KP
1,2-DNB	100	04-23-14	0.50	10.0	5.0	1:1 ACN/MeOH	02-19-14	04-23-14

LC/MS STANDARD PREP LOG# 104 PAGE# 23

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT#	DATE INITIALS
	8321TS 0.005DF						
8321TS 0.10DF	06-12-14	0.10	20	0.005	MeOH	53235	06-12-14 P
HRCFS	100 03-17-14	0.020	1	1.0	1		1
	8321TS 0.0025DF						
8321TS 0.10DF	06-12-14	0.050	2.0	0.0025	MeOH	53235	06-12-14 P
HRCFS	100 03-17-14	0.020	1	1.0	1		1
	1,2-DNB 100 ug/mL						06-12-14 HP



1,2-Dinitrobenzene
Lot #: 216681 - 33233
Rec: 1/8/14 MFR exp. 10/25/16



1,2-Dinitrobenzene
Lot #: 216681 - 33234
Rec: 1/8/14 MFR exp. 10/25/16

I broke open the 2 vials labeled above. I took 1.00ml from each vial and diluted to 20.0ml 1:1 ACN MeOH 02-19-14. I labeled this solution 1,2-DNB 100ug/L 06-12-14

I combined the remaining balance of HP 06-12-14 the 2 vials above and transferred to one amber 1.8ml injection vial. I labeled this solution 1,2-DNB 1000ug/L 06-12-14

LCMS STANDARD PREP LOG# 104 PAGE# 25

025

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS
PETN	1000 ug/ml						06-12-14

I broke open the vial labeled to the left and transferred to an 8ml Amber Vial

→ PETN 1000ug/ml 06-12-14

S-12837A4-5ML Exp 2/28/2015
CHEM SERVICE
PO BOX 599 WEST CHESTER PA 19381 800-452-9994
Pentaerythritol tetranitrate Solution
Lot: 2416400
1000ug/ml In Acetonitrile
FLAMMABLE-POISON-LACHRYMATOR
Pentaerythritol tetranitrate
Lot #: 2416400 - 33320
Rec: 2/20/14 MFR exp. 2/28/15

PETN 100 ug/ml

PETN 1000	06-12-14	1.0	10.0	100	1:1 ACN/MeOH	06-12-14	

8321MX6-7 500 ug/ml 06-12-14

o2si smart solutions
Custom Pesticide Solution, 6-7, 500 mg/L, 5 x 2 ml
130287-01-05
Made in USA Lot No: 160093
Exp: 6/1/2013
Date Opened:
Storage: </= -10 Degrees C
Custom Pesticide Solution 6-7
Lot #: 160093 - 26674
Rec: 6/7/10 MFR exp. 06/01/13

I broke open the vial labeled above and transferred to an amber 1.8ml injection vial → 8321MX6-7 500 ug/ml (to be QC'd)

8321MX6-7 20 ug/ml 06-12-14

8321MX6-7	500	06-12-14	0.080	2.0	20	Methanol	53235	

8321MX6-7 0.02DF (OLD)

8321MX6-7	20	06-12-14	0.040	2.0	0.02	MeOH	53235	06-12-14
MRCS	100	03-17-14	0.020	1	1.0	1		1

LCMS STANDARD PREP LOG# 104 Page # 37

037

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	INITIAL DATE
	Spinosad 100 mg/L						
Spinosad	1213	04-14-14	0.075	1.0	100	Methanol/53235	HP 07-14-14
	Spinosad 50 mg/L						
Spinosad	500	07-14-14	0.10	1.0	50	Methanol/53235	HP 07-14-14
	Spinosad 20 mg/L						HP
Spinosad	200	07-14-14	0.10	1.0	20	Methanol/53235	07-14-14
	Spinosad 10 mg/L						
Spinosad	100	07-14-14	0.10	1.0	10	Methanol/53235	07-14-14 HP
	Picric Acid spk 2.0 mg/L						
Picric Acid	10	04-16-14	0.160	8.0	2.0	ACN/DI509	07-15-14 HP
	Picric Acid spk 0.2 mg/L						
Picric Acid spk	2.0	07-15-14	0.80	8.0	0.2	ACN/DI509	07-15-14 HP
	HPLC FS 1.0 mg/L						
HPLC FS	1000	04-22-14	0.20	200	1.0	MeOH/53235	07-15-14 HP
	HPLC FS 1000 mg/L						07-15-14 HP

Custom 8321 Internal Standard Pesticide Mix, 4-8, 1,000 mg/L, 5 x 1 ml
130285-01-05
Lot# 209160 | Storage ≤ -10 Degrees C | Expiry 5/20/15
Solv: M-t-BE
R&D only, not for human consumption. Made in the USA

I broke open the vial labelled Custom 8321 Internal Std Mix 4-8 Lot #: 209160 - 32460 Rec: 5/29/13 MFR exp. 5/20/15 to the left and transferred to an amber 1.8 ml injection Vial → HPLC FS 100 mg/L 07-15-14

	HPLC FS 100 mg/L						
HPLC FS	1000	07-15-14	1.0	10.0	100	MeOH/53235	07-15-14 HP
	Tebufenozide 100 mg/L						HP
tebufenozide	1000	08-22-13	0.10	1.0	100	1:1 ACN/MeOH	07-01-14

02-15-14

GCMS STANDARD PREP LOG# 104 PAGE 45

045

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	DATE INITIALS
8321V 0.2ug/mL							
8321V	2.0	08-05-14	0.20	2.0	0.2	MeOH/53235	8-05-14 M
8321VES	100	7-31-14	0.020	1	1.0	1	1
8321V 0.1ug/mL							
8321V	2.0	8-5-14	0.10	2.0	0.10	MeOH/53235	8-05-14 M
8321VES	100	7-31-14	0.020	1	1.0	1	1
MX-8330R 1000ug/mL							08-06-14 M

AccuStandard®
M-8330-R
Method 8330 - Explosives by HPLC
1000 µg/mL in MeOH:AcCN (1:1)
Lot: 211111132-01
Exp: Jun 6,2015
14 comps.
HIGHLY FLAMMABLE

Method 8330 - Explosives
Lot #: 211111132-01 - 33139
Rec: 11/18/13 MFR exp. 6/6/15

ABORATORY USE ONLY
IG: This product contains
Is known to the State of Califon
cancer and birth defects or oth
ective harm.

STORAGE Freeze (<-10° C)
2 Danger

I broke open the VIAL labeled above and transferred to an amber 1.8ml injection VIAL → MX-8330R 1000ug/mL 08-06-14

1,2-DNB 1000ug/mL 08-06-14 M

1,2-Dinitrobenzene
Solution, 1,000 mg/L, 1 mL
616466-01
Lot # 216681 | Storage <10 Degrees C | Expiry 10/25/16
Solv: Acetonitrile
MADE only, not human consump. Made in the USA

1,2-Dinitrobenzene
Lot #: 216681 - 33235
Rec: 1/8/14 MFR exp. 10/25/16

I broke open the VIAL labeled to the left and transferred to an amber 1.8ml injection VIAL → 1,2-DNB 1000ug/mL 08-06-14

3,5-DNA 1000ug/mL 08-06-14

3,5-Dinitroaniline (3,5-DNA) Solution, 1000 mg/L, 1 ml
010612-04
Lot # 199709 | Storage ≤-10 Degrees C | Expiry 11/21/15
Solv: Acetonitrile
MADE only, not human consump. Made in the USA

3,5-Dinitroaniline
Lot #: 199709 - 33122
Rec: 11/15/13 MFR exp. 11/21/15

I broke open the VIAL labeled to the left and transferred to an amber 1.8ml injection VIAL → 3,5-DNA 1000ug/mL 08-06-14

LC/MS STANDARD PREP LOG# 104 PAGE# 46

046

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT #	DATE INITIALS
Nitroglycerin 1000 ug/mL							08-06-14 KP

AccuStan
M-8330-ADD-1-10X
Nitroglycerin
1000 µg/mL in EtOH:MeOH (97:3)
Lot: 213031341
Exp. May 9, 2015
Nitroglycerin
Lot #: 213031341 - 33129
Rec: 11/18/13 MFR exp. 5/9/15
HIGHLY FLAMMABLE
FOR LABORATORY USE ONLY
STORAGE Refrig (0-5° C)
2 Danger

I broke open the vial labeled above and transferred to an amber 1.8mL injection vial → Nitroglycerin 1000 ug/mL 08-06-14

PETN 1000 ug/mL 08-06-14 KP

AccuStandard® · 125 Market St. · New Haven, CT 06513 · USA
M-8330-ADD-2-10X
PETN
1000 µg/mL in Methanol
Lot: 213081209
Exp Aug 14, 2015
PETN
Lot #: 213081209 - 33135
Rec: 11/18/13 MFR exp. 8/14/15
HIGHLY FLAMMABLE
FOR LABORATORY USE ONLY
STORAGE Refrig (0-5° C)
2 Danger

I broke open the vial labeled above and transferred to an amber 1.8mL injection vial → PETN 1000 ug/mL 08-06-14

8330MIX-A 10 ug/mL								N
MX-8330	1000	08-06-14	0.20	20	10.0	1:1 ACN/MeOH		08-06-14
1,2-DNB	1000	08-06-14	0.20	"	"		07-02-14	"
8330MIX-B 10 ug/mL								N
Nitroglycerin	1000	8-6-14	0.20	20	10.0	1:1 ACN/MeOH		08-06-14
3,5-DNA	1000	08-06-14	"	"	"		07-02-14	"
PETN	1000	08-06-14	"	"	"			"

LC/MS STANDARD PREP LOG# 104 PAGE 47

047

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT#	DATE INITIALS
	PETN_SS 100 ng/L						
PETN	1000 06-12-14	0.10	1.0	100	1:1 ACN/MeOH 07-02-14		08-06-14 M
	8321 SURR 20 ng/L						M
Isotope	1024 03-10-14	0.195	10.0	20	MeOH/	53235	08-07-14
8141 SURR	1000 07-31-14	0.20	⊥	⊥	⊥		⊥
	8321 TS_SPL 1 DF						
8321 MXB-7	1000 06-12-14	0.10	5.0	20	MeOH/	53235	08-07-14 M

M
08-06-14

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOVENT/LOT	DATE/IN.
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	08/06/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	08/06/14
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	08/06/14
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	08/06/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	08/06/14
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	08/06/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	08/06/14
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	08/06/14

LCMS STANDARD PREP LOG# 164 PAGE 48

048
STANDARD

STANDARD	INITIAL CONC	SOURCE	DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS
								08/06/14 MP
8330MXA_CB 0.50 PPM								
8330MXA	10.0		08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 1.0 PPM								
8330MXA	10.0		08/06/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_CB 2.0 PPM								
8330MXA_CB 2.0 PPM	10.0		08/06/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.005 PPM								
8330MXB_CB(1:1)	0.1		08/06/14	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.010 PPM								
8330MXB_CB(1:1)	0.1		08/06/14	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.020 PPM								
8330MXB_CB(1:1)	0.1		08/06/14	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.050 PPM								
8330MXB_CB(1:1)	1.0		08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.10 PPM								
8330MXB_CB(1:1)	1.0		08/06/14	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.20 PPM								
8330MXB_CB(1:1)	1.0		08/06/14	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 0.50 PPM								
8330MXB_CB	10.0		08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 1.0 PPM								
8330MXB_CB	10.0		08/06/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_CB 2.0 PPM								
8330MXB_CB	10.0		08/06/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXA_SS 1.0 PPM								
8330MXA_SS	100		12/27/13	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	08/06/14
8330MXB_SS 1.0 PPM								
8330MXB_SS	100		12/27/13	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	08/06/14
PETN_TET_SS 1.0 PPM								
PETN_SS	100		08/06/14	0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	08/06/14
TETRYL_SS	100		07/01/14	0.01				

LC/MS STANDARD PREP LOG# 104 PAGE 49

049

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOVENT/LOT	DATE/IN.
8330_CB(1:1) 1.0 PPM							
8330MXA	10.0	08/06/14	0.100	1.0	1.0	1:1 ACN/MEOH	08/06/14
8330MXB	10.0	08/06/14				02/19/14	
8330_CB(1:1) 0.1 PPM							
8330_CB(1:1)	1.0	08/06/14	0.100	1	0.1	1:1 ACN/MEOH	08/06/14
						02/19/14	
8330_CB 0.005 PPM							
8330_CB(1:1)	0.1	08/06/14	0.050	1	0.005	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.010 PPM							
8330_CB(1:1)	0.1	08/06/14	0.100	1.00	0.010	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.020 PPM							
8330_CB(1:1)	0.1	08/06/14	0.200	1.00	0.020	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.050 PPM							
8330_CB(1:1)	1.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.10 PPM							
8330_CB(1:1)	1.0	08/06/14	0.100	1.00	0.10	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.20 PPM							
8330_CB(1:1)	1.0	08/06/14	0.20	1.00	0.20	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_CB 0.50 PPM							
8330MXA	10.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN	08/06/14
8330MXB	10.0	08/06/14	0.050			MEOH: MILLIPORE	
8330_CB 1.0 PPM							
8330MXA	10.0	08/06/14	0.100	1.00	1.000	1:1:6_ACN	08/06/14
8330MXB	10.0	08/06/14	0.100			MEOH: MILLIPORE	
8330MXA_CB 2.0 PPM							
8330MXA_CB	10.0	08/06/14	0.200	1.00	2.000	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330MXB_CB 2.0 PPM							
8330MXB_CB	10.0	08/06/14	0.200	1.00	2.000	1:1:6_ACN	08/06/14
						MEOH: MILLIPORE	
8330_SS 1.0 PPM							
8330MXA_SS	100	12/27/13	0.010	1.00	1.0	1:1:6_ACN	08/06/14
8330MXB_SS	100	12/27/13	0.010			MEOH: MILLIPORE	
PETN_TET_SS 1.0 PPM							
PETN_SS	100	08/06/14	0.01	1.00	1.0	1:1:6_ACN	08/06/14
TETRYL_SS	100	07/01/14	0.01			MEOH: MILLIPORE	

DATE / INITIALS

6-14 MP

LOT# MP 08-06-14

LCMS STANDARD PREP LOG# 104 PAGE 58

058

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	DATE INITIALS
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	08/06/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	09/05/14
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	09/05/14
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	08/06/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	09/05/14
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	08/06/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	09/05/14
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	08/06/14	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	08/06/14	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 0.50 PPM 8330MXA	10.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 1.0 PPM 8330MXA	10.0	08/06/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	08/06/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	08/06/14	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	08/06/14	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.020 PPM 8330MXB_CB(1:1)	0.1	08/06/14	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	09/05/14

LCMS STANDARD PREP LOG# 104 PAGE 59

SOL'N. 059

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	LOT #	DATE INITIALS
8330MXB_CB 0.050 PPM 8330MXB_CB(1:1)	1.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.10 PPM 8330MXB_CB(1:1)	1.0	08/06/14	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.20 PPM 8330MXB_CB(1:1)	1.0	08/06/14	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 0.50 PPM 8330MXB_CB	10.0	08/06/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 1.0 PPM 8330MXB_CB	10.0	08/06/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_CB 2.0 PPM 8330MXB_CB	10.0	08/06/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXA_SS 1.0 PPM 8330MXA_SS	100	12/27/13	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	09/05/14
8330MXB_SS 1.0 PPM 8330MXB_SS	100	12/27/13	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	09/05/14
PETN_TET_SS 1.0 PPM PETN_SS TETRYL_SS	100 100	08/06/14 07/01/14	0.01 0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	09/05/14

09/05/14

LC/MS STANDARD PREP LOG# 104 PAGE# 101

101

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT #	INITIALS
CC_Carb sph 2.0 ug/mL							
8321_SS	100	11-03-14	0.200	10.0	2.0	MeOH/54115	12-12-14 MP
HPLC IS 1.0 ug/mL							
HPLCIS	100	10-02-14	2.0	200	1.0	MeOH/54115	12-15-14 MP
HPLCISAQ 1.0 ug/mL					12-16-14		
HPLCIS	100	10-02-14	1.0	100	1.0	Milliporewater	12-16-14 MP
Carb sph 10 ug/mL							
8321MX29-8	400	09-22-14	0.050	2.0	10	MeOH/54115	12-17-14 MP
8321 surr 10 ug/mL							
Isoxaben	1024	03-10-14	0.025	3.0	10.0	MeOH/54115	12-17-14 MP
8141 surr	1000	10-30-14	0.030	〃	〃	〃	〃
8321_surr 20 ug/mL							
Isoxaben	1024	03-10-14	0.195	10.0	20	MeOH/54115	12-17-14 MP
8141 surr	1000	~~10-30-14~~ 12-03-14 MP	0.200	〃	〃	〃	〃
PETN 100 ug/mL					07-02-14		
PETN	1000	08-06-14	0.50	5.0	100	1:1 ACN/MeOH	12-19-14 MP
Nitroglycerin 100 ug/mL							
Nitroglycerin	1000	02-25-14	0.50	5.0	100	1:1 ACN/MeOH	12-19-14 MP
8321_Stock							
8321MX6-7	1000	09-25-14	0.050	10.0	5.0	MeOH/54115	12-22-14 MP
Isoxaben	1024	03-10-14	0.098	10	〃	〃	〃
8141 surr	1000	12-03-14	0.100	〃	〃	〃	〃

LC/MS STANDARD PREP LOG# 104 PAGE# 103

103

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N. LOT#	Date INITIALS
8321TS/Spk/Surr 100 DF					100ug/L		
8321TS Spk	20	09-25-14	0.10	1.0	10 DF	~~Lee~~ MP 12-22-14	12-22-14 MP
8321 Surr	20	12-18-14	0.10	1	1		
8321TS Spk/Surr 100DF		12-22-14	0.10	1.0	100DF	100ug/L	12-22-14 MP

Tetryl-ss 1000 ug/mL 12-23-14
Tetryl 1000 ug/mL Absolute Lot 112310-33875, I broke open the vial and transferred to an amber 1.8 mL injection Vial → Tetryl-ss 1000ug/L 12-23-14

1:1 ACN/MeOH 12-23-14 MP
I took 0.250 mL Methanol Lot 54115 and added 0.250 mL Acetonitrile Lot DK902. MP

1,2-DND 1000 ug/mL 12-23-14
I broke open the vial labeled to the left and transferred to an amber 1.8 mL injection vial → 1,2 DNB 1000 ug/L 12-23-14

1,2-Dinitrobenzene Solution, 1,000 mg/L, 1 mL
010086-01
Lot # 227341 Storage <10 Degrees C Expiry 3/May/2017
Solv: Acetonitrile
R&D only, not human consump. Made in the USA

1,2-Dinitrobenzene
Lot #: 227342 - 34120
Rec: 10/6/14 MFR exp. 5/3/17

3,5-DNA 1000 ug/mL 12-23-14
I broke open the vial labeled to the left and transferred to an amber 1.8 mL injection vial → 3,5-DNA 1000 ug/L 12-23-14

3,5-Dinitroaniline (3,5-DNA) Solution, 1000 mg/L, 1 ml
010612-04
Lot # 199709 Storage ≤-10 Degrees C Expiry 11/21/15
Solv: Acetonitrile
R&D only, not human consump. Made in the USA

3,5-Dinitroaniline
Lot #: 199709 - 33124
Rec: 11/15/13 MFR exp. 11/21/15

LCMS STANDARD PREP LOG# 104 PAGE# 104

104

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT #	INITIAL
	8370 MX-A 10.0 ug/mL						N
MX-8330A	1000 08-06-14	0.10	10.0	10.0	1:1 ACN/MeOH	12-23-14	
1,2-DNB	1000 12-23-14	"	"	"	12-23-14		
	8330 MX-B 10.0 ug/mL						N
Nitroglycerin	1000 08-06-14	0.10	10.0	10.0	1:1 ACN/MeOH	12-23-14	
3,5-DNA	1000 12-23-14	"	"	"	12-23-14	"	
PETN	1000 01-12-14	"	"	"		"	
	tetryl-ss 100 ug/mL						N
tetryl-ss	1000 12-23-14	0.10	1.0	10.0	1:1 ACN/MeOH	12-23-14	
					12-27-14		

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	12/23/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	12/23/14
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	12/23/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	12/23/14
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	12/23/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	12/23/14
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	12/23/14	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	12/23/14
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	12/23/14	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	12/23/14	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	12/23/14	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	12/23/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	12/23/14	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	12/23/14	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	12/23/14
8330MXA_CB 0.50 PPM 8330MXA	10.0	12/23/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	12/23/14

LC/MS STANDARD PREP LOG# 104 PAGE 106

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT#	DATE / INITIALS
PETN	1000 ug/ml					M	12-29-14

S-12837A4-5ML Exp 4/30/2015
CHEMTRONICS
PO BOX 599
Pentaerythritol tetranitrate
Lot #: 2377700 - 34516
Pentaeryth
Rec: 12/16/14 MFR exp. 4/30/15
Lot: 2677700
1000ug/ml In Acetonitrile
FLAMMABLE-POISON-LACHRYMATOR

I broke open the vial labeled to the left and transferred to amber 8ml vial → PETN 1000ug/ml 12-29-14

	PETN 100 ug/ml								M
PETN	1000	12-29-14	0.5	5.0	100	1:1 ACN/MeOH	07-2-14	12-29-14	
M1-6-14	PGDN spike 10 ug/ml								M
PGDNA	100	10-27-14	0.10	1.0	10.0	1:1 ACN/MeOH	7-2-14	01-06-15	
	PGDN spike 1.0 ug/ml								M
PGDN spk	10	01-06-15	0.10	1.0	1.0	1:1 ACN/MeOH	7-2-14	01-06-15	
	Nitroglycerin 100 ug/ml								M
Nitroglycerin	1000	08-06-14	0.50	5.0	100	1:1 ACN/MeOH	12-23-14	01-06-15	
	8330 MXA spk 10 ug/ml								M
MX-8330 spk	100	09-17-14	0.150	1.5	10.0	1:1 ACN/MeOH	12-23-14	01-06-15	
	8330 MXA spk 1.0 ug/ml								M
8330 MXA spk	10.0	01-06-15	0.150	1.5	1.0	1:1 ACN/MeOH	12-23-14	01-06-15	

LC/MS STANDARD PREP LOG# 104 Page 109

109

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT/ LOT #	INITIALS
8321PA_SS 0.10 DF							
Picric Acid_SS	1.0	11-17-14	0.20	2.0	0.10	ACN/DK902	01-13-15
4-Nitrophenol	10	01-13-15	0.020	1	0.10	1	1
MX8330 Sph 100ug/mL							
MX-8330A	1000	08-06-14	0.80	8.0	100	1:1 ACN/MeOH	01-15-15
1,2-DNB	1000	12-23-14	0.80	1	1	12-23-14	
Tetryl SS 100ug/mL							
Tetryl SS	1000	12-23-14	0.10	1.0	100	1:1 ACN/MeOH	01-15-15
						12-23-14	

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	12/23/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	01/15/15
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	01/15/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	01/15/15
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	12/23/14	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	01/15/15
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	01/15/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	01/15/15
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	01/15/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	01/15/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	01/15/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	01/15/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	01/15/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	01/15/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 0.50 PPM 8330MXA	10.0	12/23/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	01/15/15

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT #	DATE / INITIALS
8330MXA_CB 1.0 PPM 8330MXA	10.0	12/23/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	12/23/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	01/15/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	01/15/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.020 PPM 8330MXB_CB(1:1)	0.1	01/15/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.050 PPM 8330MXB_CB(1:1)	1.0	01/15/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.10 PPM 8330MXB_CB(1:1)	1.0	01/15/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.20 PPM 8330MXB_CB(1:1)	1.0	01/15/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 0.50 PPM 8330MXB_CB	10.0	12/23/14	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 1.0 PPM 8330MXB_CB	10.0	12/23/14	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_CB 2.0 PPM 8330MXB_CB	10.0	12/23/14	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXA_SS 1.0 PPM 8330MXA_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	01/15/15
8330MXB_SS 1.0 PPM 8330MXB_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	01/15/15
PETN_TET_SS 1.0 PPM PETN_SS TETRYL_SS	100 100	08/06/14 01/15/15	0.01 0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	01/15/15

LCMS STANDARD PREP LOG# 105 PAGE 17

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOVENT/LOT	DATE/INITIALS
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	03/12/15
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	03/12/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	03/12/15
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	03/12/15
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	03/12/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	03/12/15
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	03/12/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	03/12/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	03/12/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	03/12/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	03/12/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	03/12/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 0.50 PPM 8330MXA	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	03/12/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	03/12/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	03/12/15

018

LC/MS STANDARD PREP LOG# 105 PAGE 18

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT / LOT #	DATE / INITIALS
8330MXB_CB 0.020 PPM							
8330MXB_CB(1:1)	0.1	03/12/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.050 PPM							
8330MXB_CB(1:1)	1.0	03/12/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.10 PPM							
8330MXB_CB(1:1)	1.0	03/12/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.20 PPM							
8330MXB_CB(1:1)	1.0	03/12/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 0.50 PPM							
8330MXB_CB	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 1.0 PPM							
8330MXB_CB	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_CB 2.0 PPM							
8330MXB_CB	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	03/12/15 03/12/15
8330MXA_SS 1.0 PPM							
8330MXA_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	03/12/15
8330MXB_SS 1.0 PPM							
8330MXB_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	03/12/15
PETN_TET_SS 1.0 PPM							
PETN_SS	100	08/06/14	0.01	1.00	1.0	1:1:6_ACN	03/12/15
TETRYL_SS	100	01/15/15	0.01			MEOH: MILLIPORE	

HP HP

PETN_SS 100 08-06-14 the Solution was QC'd on 03-13-15 on Stetler, Data File 0312_D00047.D HP 03-16-15

TOMS STANDARD PREP LOG# 105 PAGE 21

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL. CNT. / LOT # 021	DATE / INITIALS
Scav_spk 30 ug/mL (5mL)							
Hexahydro-1,3,5-tris(hydroxyethyl) triazine							
1180	03-18-15	0.127	5.0	30	MeOH/5411S	03-19-15	
2,2'-thiodiethanol							
1030	03-19-15	0.146					
2-(1,3-Dithiazinan-5-yl) ethanol							
525	03-11-15	0.286					
Scav spk 3.0 ug/mL 5mL							
Scav spk 30	03-19-15	0.50	5.0	3.0	MeOH/5411S	03-19-15	

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	03/17/15
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	03/17/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	03/17/15
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 07/02/14	03/17/15
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	03/17/15	0.100	1.0	0.1	1:1 ACN/MEOH 07/02/14	03/17/15
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	03/17/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	03/17/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	03/17/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	03/17/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	03/17/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	03/17/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 0.50 PPM 8330MXA	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/17/15

022 GC/MS STANDARD PREP LOG# 105 PAGE 22

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'N LOT #	DATE / INITIALS
8330MXA_CB 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	03/17/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	03/17/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.020 PPM 8330MXB_CB(1:1)	0.1	03/17/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.050 PPM 8330MXB_CB(1:1)	1.0	03/17/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.10 PPM 8330MXB_CB(1:1)	1.0	03/17/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.20 PPM 8330MXB_CB(1:1)	1.0	03/17/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 0.50 PPM 8330MXB_CB	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 1.0 PPM 8330MXB_CB	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_CB 2.0 PPM 8330MXB_CB	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXA_SS 1.0 PPM 8330MXA_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	03/17/15
8330MXB_SS 1.0 PPM 8330MXB_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	03/17/15
PETN_TET_SS 1.0 PPM PETN_SS TETRYL_SS	100 100	08/06/14 01/15/15	0.01 0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	03/17/15

040 LC/MS STANDARD PREP LOG# 105 PAGE# 40

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE/ INITIALS
PER IS	0.005µg/mL						04-20-15: JBL
02SI $^{18}O_4$ Perchlorate	100µg/L						04-21-
Lot 1059192-33479 took 0.050mL and added to 0.950 mL Laboratory Source DI. H2O.							
NOTE: Repeated the above process four times in order to prepare 2-2mL vials that were full. The above was performed as dated. JBL							
8330 MX-A_SS	100 mg/L						04-21-15 MP
8330 MIX A, 100 mg/L, Absolute Lot # 103014-34464. I broke the Vial open and transferred to an amber 1.8mL injection Vial → 8330 MX-A_SS 100 mg/L							04-21-15
8330 MXB_SS	100 mg/L						04-21-15 MP
8330 MIX B, 100 mg/L, Absolute Lot # 103014-34467, I broke the Vial open and transferred to an amber 1.8mL injection Vial → 8330 MX-B_SS 100 mg/L							04-21-15
Tebuconazole_SS	100 mg/L						03-05-15 MP
Tebuconazole (SS)	1000	12-22-14	0.10	1.0	100	1:1 ACN/MeOH	04-22-15

LC/MS STANDARD PREP LOG# 105 PAGE 41

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT/LOT#	DATE/ INITIALS

1,2-DNB 1000 ug/mL 04-22-15 MP

I broke open the vial labeled to the left and transferred to an amber 1.8ml injection vial → 1,2-DNB 1000ug/mL 04-22-15

1,2-Dinitrobenzene Solution, 1,000 mg/L, 1 mL
010086-01
Lot # 227342 Storage <10 Degrees C Expiry 3/May/2017
Solv: Acetonitrile
1,2-Dinitrobenzene
Lot #: 227342 - 34115
Rec: 10/6/14 MFR exp. 5/3/17

1,2-DNB 100 ug/mL 03/05/15

1,2-DNB	1000	04-22-15	1.0	10	100	1:1 ACN/MeOH	14-22-15 MP

Standard prep performed as dated, documented here 04-22-15 MP

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 03/05/15	04/21/15
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	04/21/15	0.100	1.0	0.1	1:1 ACN/MEOH 03/05/15	04/21/15
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 03/05/15	04/21/15
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	03/17/15	0.100	1.0	0.1	1:1 ACN/MEOH 03/05/15	04/21/15
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	04/21/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	04/21/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	04/21/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	04/21/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	04/21/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	04/21/15

042 LCMS STANDARD PREP LOG 105 - 42

Standard prep performed as dated, Documented here 04-22-15 MP

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE / INITIALS
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	04/21/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 0.50 PPM 8330MXA	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	04/21/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	04/21/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.020 PPM 8330MXB_CB(1:1)	0.1	04/21/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.050 PPM 8330MXB_CB(1:1)	1.0	04/21/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.10 PPM 8330MXB_CB(1:1)	1.0	04/21/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.20 PPM 8330MXB_CB(1:1)	1.0	04/21/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 0.50 PPM 8330MXB_CB	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 1.0 PPM 8330MXB_CB	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_CB 2.0 PPM 8330MXB_CB	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXA_SS 1.0 PPM 8330MXA_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	04/21/15
8330MXB_SS 1.0 PPM 8330MXB_SS	100	10/01/14	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	04/21/15
PETN_TET_SS 1.0 PPM PETN_SS TETRYL_SS	100 100	08/06/14 04/22/15	0.01 0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	04/22/15

LCMS STANDARD PREP LOG# 105 PAGE 43

043

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOL'VENT LOT#	DATE / INITIALS
Scav Spk 30 ug/mL							
Hexahydro-1,3,5-Trishydroxyethyltriazine							HP
1010	04-10-15	0.594	20	30	ACN/DM394		04-23-15
2,2'-Thiodiethanol				"	"		"
1090	04-10-15	0.550	20	"	"		"
2-(1,3,5-Dithiazinan-5-yl) ethanol				"	"		"
1020	04-10-15	0.588	"	"	"		"
Scav Surr 30 ug/mL							
1,3,5-Trimethylhexahydro-1,3,5-triazine							HP
1077	04-17-15	0.557	20	30	ACN/DM394		04-25-15
PETN SS 100 ug/mL						3-5-15	
PETN 1000	06-09-14	0.10	1.0	100	1:1 ACN/water		04-28-15 HP
Tetryl SS 100 ug/mL						3-5-15	
Tetryl SS 1000	12-22-14	0.10	1.0	100	1:1 ACN/MeOH		04-28-15 HP
8330mXA-SS 100 ug/mL							04-28-15 HP
8330mX-A 100 ug/mL Absolute lot 103014-34465, I broke open the vial and transferred to an amber 1.8ml injection vial → 8330mX-A-SS 100ug/mL 04-28-15							
8330mXB-SS 100 ug/mL							04-28-15 HP
8330mXB 100 ug/mL Absolute lot 103014-34468, I broke open the vial and transferred to an amber 1.8ml injection vial → 8330mXB-SS 100ug/mL 04-28-15							

044 LCMS STANDARD PREP LOG# 105 PAGE 44

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE/ INITIALS
PETN_SS	1000 µg/mL						

AccuStandard® 125 Market Street • New Haven, CT 06513 • USA

M-8330-ADD-2-10X
PETN
1000 µg/mL in Methanol
Lot: 214091329
Exp: Sep 18, 2016

PETN
Lot #: 214091329 - 34515
Rec: 12/16/14 MFR exp. 9/18/16

1 comp(s)
HIGHLY FLAMMABLE

FOR LABORATORY USE ONLY
H225 H320 H315 H311
H332 H301 P338 P360
P331 P404 P262 P202
P264 P284 P280

Storage: Refrig (0-5° C)
2 Danger

04-28-15 hp

I broke open the Vial labeled above and transferred to an amber 1.8 mL injection Vial → PETN_SS 1000µg/mL 4-28-15

PETN SS 100 µg/mL 3-5-15 hp

PETN_SS	1000	04-28-15	0.10	1.0	100	1:1 ACN/MeOH	4-28-15

Stock Weight Comparison 04-28-15 hp
Balance ID: Mettler AT-200 #J-28448

Time-		
S-weight	Balance readings	
20 g	20.0000	
1 g	1.0002	
100 mg	0.1000	
20 mg	~~0.0020~~ 0.0020	hp 4-28-15

GCMS STANDARD PREP LOG# 105 PAGE 48

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT/ LOT #	DATE/ INITIALS
Standard Prep performed as Dated, documented here 05-01-15 !							MP

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXB_CB(1:1) 1.0 PPM 8330MXB	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 03/05/15	04/27/15 MP
8330MXB_CB(1:1) 0.10 PPM 8330MXB_CB(1:1)	1.0	04/27/15	0.100	1.0	0.1	1:1 ACN/MEOH 03/05/15	04/27/15
8330MXB_CB 0.005 PPM 8330MXB_CB(1:1)	0.1	04/27/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.010 PPM 8330MXB_CB(1:1)	0.1	04/27/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.020 PPM 8330MXB_CB(1:1)	0.1	04/27/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.050 PPM 8330MXB_CB(1:1)	1.0	04/27/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.10 PPM 8330MXB_CB(1:1)	1.0	04/27/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.20 PPM 8330MXB_CB(1:1)	1.0	04/27/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 0.50 PPM 8330MXB_CB	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 1.0 PPM 8330MXB_CB	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_CB 2.0 PPM 8330MXB_CB	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	04/27/15

LCMS STANDARD PREP LOG# 105 PAGE 49

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT'S	DATE / INITIALS
Standard Prep performed as Dated, documented here 05-01-15							MP

STANDARD	INITIAL CONC	SOURCE DATE	ALIQUOT	FINAL VOLUME	FINAL CONC.	SOVENT/LOT	DATE/IN.
8330MXA_SS 1.0 PPM 8330MXA_SS	100	04/28/15	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	04/27/15
8330MXB_SS 1.0 PPM 8330MXB_SS	100	04/28/15	0.010	1.00	1.0	1:1:6_ACN/MEOH MILLIPORE	04/27/15
PETN_TET_SS 1.0 PPM PETN_SS TETRYL_SS	100 100	04/28/15 04/28/15	0.01 0.01	1.00	1.0	1:1:6_ACN MEOH: MILLIPORE	04/29/15
8330MXA_CB(1:1) 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.0	1.0	1:1 ACN/MEOH 03/05/15	04/28/15
8330MXA_CB(1:1) 0.10 PPM 8330MXA_CB(1:1)	1.0	04/28/15	0.100	1.0	0.1	1:1 ACN/MEOH 03/05/15	04/28/15
8330MXA_CB 0.005 PPM 8330MXA_CB(1:1)	0.1	04/28/15	0.050	1	0.005	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.010 PPM 8330MXA_CB(1:1)	0.1	04/28/15	0.100	1.00	0.010	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.020 PPM 8330MXA_CB(1:1)	0.1	04/28/15	0.200	1.00	0.020	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.050 PPM 8330MXA_CB(1:1)	1.0	04/28/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.10 PPM 8330MXA_CB(1:1)	1.0	04/28/15	0.100	1.00	0.100	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.20 PPM 8330MXA_CB(1:1)	1.0	04/28/15	0.200	1.00	0.200	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 0.50 PPM 8330MXA	10.0	03/05/15	0.050	1.00	0.050	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 1.0 PPM 8330MXA	10.0	03/05/15	0.100	1.00	1.000	1:1:6_ACN/MEOH MILLIPORE	04/28/15
8330MXA_CB 2.0 PPM 8330MXA_CB 2.0 PPM	10.0	03/05/15	0.200	1.00	2.000	1:1:6_ACN/MEOH MILLIPORE	04/28/15

_MSE018

Organic Extraction Worksheet

Method	Explosives Soil Extraction 8330B	Extraction Set	150520A	Extraction Method	MSE018	Units	mL

Spiked ID 1	8330 100ppm STK 5-5-15 exp 11-5-15		Surrogate ID 1	1,2-DNB STK 100 PPM 4-22-15 exp 10-22-15	
Spiked ID 2	Nitroglyerin 100ppm 3-31-15 exp 9-30-15		Surrogate ID 2		
Spiked ID 3			Surrogate ID 3		
Spiked ID 4			Surrogate ID 4		
Spiked ID 5			Surrogate ID 5		
Spiked ID 6			Sufficient Vol for Matrix QC:	YES	
Spiked ID 7			Ext. Start Time:	05/20/15 13:40	
Spiked ID 8			Ext. End Time:	05/21/15 7:40	
			GC Requires Extract By:		

pH1			Water Bath Temp Criteria
pH2			
pH3			

Spiked By: CFM **Date** 05/20/15 1:20:00 PM **Witnessed By:** DL **Date** 05/20/15 1:20:00 PM

	Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	150520A Blk				0.060	1 equip	10.01g	20mLs	NA	05/20/15 13:40	
2	150520A LCS-1		0.200	1-2	NA	NA equip	10.06g	20mLs	NA	05/20/15 13:40	
3	AZ16137 MS-1	AZ16137S01	0.200	1-2	NA	NA equip	10.05g	20mLs	NA	05/20/15 13:40	76380
4	AZ16137 MSD-1	AZ16137S01	0.200	1-2	NA	NA equip	10.02g	20mLs	NA	05/20/15 13:40	76380
5	AZ16137	AZ16137S01			0.060	1 equip	10.10g	20mLs	NA	05/20/15 13:40	76380
6	AZ16138	AZ16138S01			0.060	1 equip	10.03g	20mLs	NA	05/20/15 13:40	76380
7	AZ16139	AZ16139S01			0.060	1 equip	10.18g	20mLs	NA	05/20/15 13:40	76380
8	AZ16140	AZ16140S01			0.060	1 equip	10.10g	20mLs	NA	05/20/15 13:40	76380
9	AZ16141	AZ16141S01			0.060	1 equip	10.09g	20mLs	NA	05/20/15 13:40	76380
10	AZ16198	AZ16198S01			0.060	1 equip	10.02g	20mLs	NA	05/20/15 13:40	76390
11	AZ16199	AZ16199S01			0.060	1 equip	10.06g	20mLs	NA	05/20/15 13:40	76390
12	AZ16201	AZ16201S01			0.060	1 equip	10.04g	20mLs	NA	05/20/15 13:40	76390
13	AZ16202	AZ16202S01			0.060	1 equip	10.03g	20mLs	NA	05/20/15 13:40	76390
14	AZ16400	AZ16400S01			0.060	1 equip	10.04g	20mLs	NA	05/20/15 13:40	76447
15	AZ16401	AZ16401S01			0.060	1 equip	10.11g	20mLs	NA	05/20/15 13:40	76447
16	AZ16402	AZ16402S01			0.060	1 equip	10.04g	20mLs	NA	05/20/15 13:40	76447

lvent and Lot#	
etonitrile	DM 394
ica Sand	041913

Extraction COC Transfer	
Extraction lab employee Initials	CFM
GC analyst's initials	NP
Date	05-21-15
Time	0850
Refrigerator	Room

	Technician's Initials
Scanned By	CFM
Sample Preparation	CFM
Extraction	CFM
Concentration	----------

Modified	05/21/15 7:59:39 AM

Reviewed By: NP **Date** 05-21-15

Injection Log

Directory: Z:\WALDORF\CHEM32\2\DATA\150522\150522

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	0	0522_000014.D		8330CB 0.005 PPM 05/22/15		5/22/15 16:07
2	0	0522_000015.D		8330CB 0.010 PPM 05/22/15		5/22/15 16:35
3	0	0522_000016.D		8330CB 0.020 PPM 05/22/15		5/22/15 17:02
4	0	0522_000017.D		8330CB 0.050 PPM 05/22/15		5/22/15 17:30
5	0	0522_000018.D		8330CB 0.10 PPM 05/22/15		5/22/15 17:57
6	0	0522_000019.D		8330CB 0.20 PPM 05/22/15		5/22/15 18:25
7	0	0522_000020.D		8330CB 0.50 PPM 05/22/15		5/22/15 18:52
8	0	0522_000021.D		8330CB 1.0 PPM 05/22/15		5/22/15 19:20
9	0	0522_000022.D		8330MXA 2.0 PPM 05/22/15		5/22/15 19:47
10	0	0522_000023.D		8330MXB 2.0 PPM 05/22/15		5/22/15 20:15
11	0	0522_000024.D		8330CB 0.005 PPM 05/22/15		5/22/15 20:42
12	0	0522_000025.D		8330CB 0.010 PPM 05/22/15		5/22/15 21:10
13	0	0522_000026.D		8330_SS 1.0 PPM 05/22/15		5/22/15 21:37
14	0	0522_000027.D		PETN_TETRYL_SS 1.0 PPM 05/22/51		5/22/15 22:05
15	0	0522_000044.D		8330_CCV 1.0 PPM 05/22/15		5/23/15 8:13
16	0	0522_000048.D	7.95229	150520ALCS-1 7.952 DF 05/20/15	Soil	5/23/15 10:33
17	0	0522_000049.D	7.9602	AZ16137S01_MS-1 7.960 DF 05/20/15	Soil	5/23/15 11:10
18	0	0522_000050.D	7.98403	AZ16137S01_MSD-1 7.984 DF 05/20/15	Soil	5/23/15 11:48
19	0	0522_000051.D	7.99201	150520ABLKS 7.992 DF 05/20/15	Soil	5/23/15 12:25
20	0	0522_000052.D	7.92079	AZ16137S01 7.921 DF 05/20/15	Soil	5/23/15 13:03
21	0	0522_000053.D	7.97607	AZ16138S01 7.976 DF 05/20/15	Soil	5/23/15 13:40
22	0	0522_000054.D	7.85855	AZ16139S01 7.859 DF 05/20/15	Soil	5/23/15 14:18
23	0	0522_000055.D	7.92079	AZ16140S01 7.921 DF 05/20/15	Soil	5/23/15 14:55
24	0	0522_000056.D	7.92864	AZ16141S01 7.929 DF 05/20/15	Soil	5/23/15 15:33
25	0	0522_000059.D		8330_CCV 1.0 PPM 05/22/15		5/23/15 17:25

Injection Log

Directory: Z:\STATLER\CHEM32\1\DATA\150522\150522

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	0	0522_000009.D		8330MXA_CB 0.005 PPM 05/22/15		5/22/15 14:13
2	0	0522_000010.D		8330MXB_CV 0.005 PPM 05/22/15		5/22/15 14:52
3	0	0522_000011.D		8330MXA_CB 0.010 PPM 05/22/15		5/22/15 15:32
4	0	0522_000012.D		8330MXB_CB 0.010 PPM 05/22/15		5/22/15 16:11
5	0	0522_000013.D		8330MXA_CB 0.020 PPM 05/22/15		5/22/15 16:51
6	0	0522_000014.D		8330MXB_CB 0.020 PPM 05/22/15		5/22/15 17:30
7	0	0522_000015.D		8330MXA_CB 0.050 PPM 05/22/15		5/22/15 18:10
8	0	0522_000016.D		8330MXB_CB 0.050 PPM 05/22/15		5/22/15 18:49
9	0	0522_000017.D		8330MXA_CB 0.10 PPM 05/22/15		5/22/15 19:29
10	0	0522_000018.D		8330MXB_CB 0.10 PPM 05/22/15		5/22/15 20:08
11	0	0522_000019.D		8330MXA_CB 0.20 PPM 05/22/15		5/22/15 20:48
12	0	0522_000020.D		8330MXB_CB 0.20 PPM 05/22/15		5/22/15 21:27
13	0	0522_000021.D		8330MXA_CB 0.50 PPM 05/22/15		5/22/15 22:07
14	0	0522_000022.D		8330MXB_CB 0.50 PPM 05/22/15		5/22/15 22:46
15	0	0522_000023.D		8330MXA_CB 1.0 PPM 05/22/15		5/22/15 23:26
16	0	0522_000024.D		8330MXB_CB 1.0 PPM 05/22/15		5/23/15 0:05
17	0	0522_000025.D		8330MXA_CB 2.0 PPM 05/22/15		5/23/15 0:45
18	0	0522_000026.D		8330MXB_CB 2.0 PPM 05/22/15		5/23/15 1:24
19	0	0522_000031.D		8330MXA_SS 1.0 PPM 05/22/15		5/23/15 4:42
20	0	0522_000032.D		8330MXB_SS 1.0 PPM 05/22/15		5/23/15 5:22
21	0	0522_000033.D		PETN_TETRYL_SS 1.0 PPM 05/22/15		5/23/15 6:01
22	0	0522_000058.D		8330MXA_CCV 1.0 PPM 05/23/15		5/23/15 21:11
23	0	0522_000059.D		8330MXB_CCV 1.0 PPM 05/23/15		5/23/15 21:51
24	0	0522_000060.D	7.92079	AZ16140S01 7.921 DF 05/20/15		5/23/15 22:30
25	0	0522_000072.D		8330MXA_CCV 1.0 PPM 05/23/15		5/24/15 6:25
26	0	0522_000073.D		8330MXB_CCV 1.0 PPM 05/23/15		5/24/15 7:04

METALS

APPL, INC.

METALS

QC Summary

APPL, INC.

METALS BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
6010C	ALUMINUM (AL)	4.00 U	50.0	4.00	1.98	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	BORON (B)	4.00 U	5.0	4.00	1.55	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	LITHIUM (LI)	0.500 U	0.67	0.500	0.250	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	ZINC (ZN)	4.00 U	8.0	4.00	1.15	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Metals SC-Blank-REG MDLs
Printed: 06/16/15 1:09:50 PM

Laboratory Control Spike Recovery

METALS

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
EPA 6010C	ALUMINUM (AL)	200	186	93.0	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	BORON (B)	25.0	23.8	95.2	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	LITHIUM (LI)	10.00	10.2	102	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	ZINC (ZN)	50.0	46.7	93.4	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Comments: _____

Matrix Spike Recoveries

METALS

APPL ID: **150604S-16137 MS - 198024**

Sample ID: AZ16137

Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Lvl mg/Kg	Matrix Res mg/Kg	SPK Res mg/Kg	DUP Res mg/Kg	SPK % Recovery	DUP % Recovery	RPD	RPD Max	Recovery Limits	Extract Date-Spk	Analysis Date-Spk	Extract Date-Dup	Analysis Date-Dup	QC Group	QC Sample
EPA 6010C	ALUMINUM (AL)	396	4880	5220	5820	85.9	237 #	10.9	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	BORON (B)	49.5	12.7	48.0	49.5	71.3 #	74.3 #	3.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	LITHIUM (LI)	19.8	3.7	18.7	19.1	75.8 #	77.8 #	2.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	ZINC (ZN)	99.0	467	649	585	184 #	119	10.4	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137

= Recovery is outside QC limits.

Comments: _____

6010C/3050B^

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 06/12/15

Matrix: SOIL Instrument: Jules

Blank ID: 150604A-BLK Time Analyzed: 1426

APPL ID.	Client Sample No.	File ID.	Date Analyzed
150604A-MSD	Matrix SpikeD	150612A	06/12/15 1655
150604A-MS	Matrix Spike	150612A	06/12/15 1648
150604A-LCS	Lab Control Spike	150612A	06/12/15 1432
AZ16139	S67-SB030-0607	150612A	06/12/15 1723
AZ16138	S67-SS030P-0001	150612A	06/12/15 1716
150604A-BLK	Blank	150612A	06/12/15 1426
AZ16141	S67-SB032-0607	150612A	06/12/15 1954
AZ16140	S67-SS032-0001	150612A	06/12/15 1947
AZ16139	S67-SB030-0607	150612A	06/12/15 1939
AZ16138	S67-SS030P-0001	150612A	06/12/15 1933
AZ16137	S67-SS030-0001	150612A	06/12/15 1641

Comments: Batch: #61CJU-150604A

METALS

Sample Data

APPL, INC.

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16137

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	5640	1150.0	92.00	46.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	BORON (B)	92.00 U	115.0	92.00	36.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	12.000 U	15.00	12.000	5.800	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	539	185.0	92.00	27.00	mg/Kg	20	06/04/15	06/12/15

Printed: 06/16/15 1:10:24 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Sequence No.: 28
Sample ID: AZ16137S01-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.01 g
Dilution: 20X
Wash Time:

Autosampler Location: 64
Date Collected: 06/12/15 4:41:04 PM
Data Type: Reprocessed on 06/15/15 10:36:12 AM

Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16137S01-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	210128.7	103.4	%	0.90				0.87%
Y Radial	14144.8	107.8	%	0.81				0.75%
Ag 338.289†	-27.0	-0.166	ug/L	0.4347	-0.329	mg/kg	0.8607	261.79%
Al 308.215†	2804.8	2463	ug/L	35.87	4877	mg/kg	71.02	1.46%
As 188.979†	11.6	8.840	ug/L	0.3590	17.51	mg/kg	0.711	4.06%
B 249.677†	-3365.2	6.400	ug/L	1.0086	12.67	mg/kg	1.997	15.76%
Ba 233.527†	31306.4	81.43	ug/L	0.348	161.2	mg/kg	0.69	0.43%
Be 313.107†	477.7	0.148	ug/L	0.0091	0.293	mg/kg	0.0180	6.15%
Ca 315.887†	13963.6	1081	ug/L	6.82	2141	mg/kg	13.51	0.63%
Cd 214.440†	4337.4	14.66	ug/L	0.123	29.04	mg/kg	0.243	0.84%
Co 228.616†	848.8	8.464	ug/L	0.1079	16.76	mg/kg	0.214	1.27%
Cr 267.716†	9318.2	119.8	ug/L	0.71	237.2	mg/kg	1.41	0.60%
Cu 327.393†	7297.5	42.67	ug/L	0.155	84.50	mg/kg	0.307	0.36%
Fe 273.955†	3164812.5	46850	ug/L	174.02	92780	mg/kg	344.60	0.37%
K 766.490†	1250.9	188.9	ug/L	11.61	374.1	mg/kg	22.99	6.14%
Mg 285.213†	8280.1	622.1	ug/L	3.02	1232	mg/kg	5.98	0.49%
Mn 257.610†	18372.2	261.4	ug/L	1.51	517.6	mg/kg	2.99	0.58%
Mo 202.031†	59.5	2.794	ug/L	0.1905	5.533	mg/kg	0.3772	6.82%
Na 589.592†	165.8	28.09	ug/L	6.151	55.63	mg/kg	12.181	21.90%
Ni 231.604†	3176.8	74.59	ug/L	0.290	147.7	mg/kg	0.57	0.39%
P 213.617†	211.9	73.42	ug/L	1.676	145.4	mg/kg	3.32	2.28%
Pb 220.353†	2778.0	497.4	ug/L	1.95	985.0	mg/kg	3.87	0.39%
Sb 206.836†	39.8	6.281	ug/L	0.9647	12.44	mg/kg	1.910	15.36%
Se 196.026†	-47.4	0.462	ug/L	1.4008	0.915	mg/kg	2.7738	303.17%
Sn 189.927†	32.4	6.597	ug/L	0.1958	13.06	mg/kg	0.388	2.97%
Sr 421.552†	2106.1	6.258	ug/L	0.0399	12.39	mg/kg	0.079	0.64%
Ti 337.279†	1932.1	60.63	ug/L	1.279	120.1	mg/kg	2.53	2.11%
Tl 190.801†	-1.6	-0.327	ug/L	0.7042	-0.647	mg/kg	1.3944	215.63%
V 292.402†	875.2	9.349	ug/L	0.1022	18.51	mg/kg	0.202	1.09%
Zn 206.200†	18987.1	235.7	ug/L	0.40	466.7	mg/kg	0.80	0.17%
Li 670.784	29124.6	1.885	ug/L	0.0266	3.733	mg/kg	0.0527	1.41%

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS030P-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16138

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	3190	57.0	4.50	2.30	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.50U	5.7	4.50	1.80	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	91.00U	114.0	91.00	35.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	3.6	0.76	0.570	0.280	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	593	182.0	91.00	26.00	mg/Kg	20	06/04/15	06/12/15

Printed: 06/16/15 1:10:24 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Sequence No.: 53
Sample ID: AZ16138S01
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.05 g
Dilution:
Wash Time:

Autosampler Location: 53
Date Collected: 06/12/15 7:33:39 PM
Data Type: Reprocessed on 06/15/15 10:36:40 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16138S01

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	280599.1	138.1	%	0.37				0.27%
Y Radial	19742.6	150.5	%	1.93				1.29%
Ag 338.289†	-554.8	-3.416	ug/L	0.1227	-0.325	mg/kg	0.0117	3.59%
Al 308.215†	33606.1	29510	ug/L	288.90	2810	mg/kg	27.51	0.98%
As 188.979†	185.3	37.56	ug/L	5.239	3.577	mg/kg	0.4989	13.95%
B 249.677†	-83298.6	-2717	ug/L	6.30	-258.7	mg/kg	0.60	0.23%
Ba 233.527†	421454.1	1131	ug/L	2.49	107.7	mg/kg	0.24	0.22%
Be 313.107†	8113.2	2.510	ug/L	0.0147	0.239	mg/kg	0.0014	0.59%
Ca 315.887†	322019.9	24960	ug/L	349.53	2377	mg/kg	33.29	1.40%
Cd 214.440†	70223.4	287.3	ug/L	0.65	27.36	mg/kg	0.061	0.22%
Co 228.616†	11485.4	130.5	ug/L	0.88	12.43	mg/kg	0.084	0.67%
Cr 267.716†	77599.4	997.8	ug/L	0.70	95.03	mg/kg	0.067	0.07%
Cu 327.393†	147778.1	833.3	ug/L	1.24	79.36	mg/kg	0.118	0.15%
Fe 273.955†	Saturated3							
K 766.490†	13721.7	2374	ug/L	32.68	226.1	mg/kg	3.11	1.38%
Mg 285.213†	176431.5	12430	ug/L	128.15	1184	mg/kg	12.21	1.03%
Mn 257.610†	368106.0	5232	ug/L	73.38	498.3	mg/kg	6.99	1.40%
Mo 202.031†	999.7	33.62	ug/L	0.969	3.202	mg/kg	0.0923	2.88%
Na 589.592†	4850.1	836.5	ug/L	8.76	79.66	mg/kg	0.835	1.05%
Ni 231.604†	34005.5	787.9	ug/L	2.46	75.04	mg/kg	0.234	0.31%
P 213.617†	2364.1	819.2	ug/L	10.34	78.02	mg/kg	0.985	1.26%
Pb 220.353†	45319.7	8252	ug/L	19.54	785.9	mg/kg	1.86	0.24%
Sb 206.836†	347.8	54.92	ug/L	2.469	5.231	mg/kg	0.2351	4.50%
Se 196.026†	-976.3	-360.8	ug/L	4.28	-34.36	mg/kg	0.408	1.19%
Sn 189.927†	295.4	60.09	ug/L	1.592	5.722	mg/kg	0.1516	2.65%
Sr 421.552†	29991.4	89.35	ug/L	0.798	8.510	mg/kg	0.0760	0.89%
Ti 337.279†	23015.5	723.3	ug/L	6.22	68.88	mg/kg	0.593	0.86%
Tl 190.801†	41.4	8.259	ug/L	5.6724	0.787	mg/kg	0.5402	68.68%
V 292.402†	10478.2	111.9	ug/L	0.46	10.66	mg/kg	0.044	0.41%
Zn 206.200†	294130.0	3679	ug/L	8.51	350.3	mg/kg	0.81	0.23%
Li 670.784	515593.3	33.38	ug/L	0.266	3.179	mg/kg	0.0254	0.80%

Sequence No.: 33
Sample ID: AZ16138S01-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.05 g
Dilution: 20X
Wash Time:

Autosampler Location: 69
Date Collected: 06/12/15 5:16:29 PM
Data Type: Reprocessed on 06/15/15 10:36:17 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16138S01-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	210169.2	103.4	%	0.19				0.18%
Y Radial	14123.6	107.7	%	0.73				0.68%
Ag 338.289†	-27.4	-0.169	ug/L	0.5179	-0.321	mg/kg	0.9865	307.15%
Al 308.215†	2307.0	2025	ug/L	16.02	3857	mg/kg	30.52	0.79%
As 188.979†	15.0	13.71	ug/L	0.544	26.11	mg/kg	1.037	3.97%
B 249.677†	-5812.5	1.417	ug/L	2.0244	2.700	mg/kg	3.8559	142.82%
Ba 233.527†	29081.3	73.77	ug/L	0.327	140.5	mg/kg	0.62	0.44%
Be 313.107†	466.2	0.144	ug/L	0.0035	0.275	mg/kg	0.0068	2.46%
Ca 315.887†	23132.5	1792	ug/L	11.12	3412	mg/kg	21.18	0.62%
Cd 214.440†	5571.5	17.66	ug/L	0.041	33.63	mg/kg	0.078	0.23%
Co 228.616†	903.0	8.396	ug/L	0.0743	15.99	mg/kg	0.142	0.88%
Cr 267.716†	5867.4	75.13	ug/L	0.028	143.1	mg/kg	0.05	0.04%
Cu 327.393†	10262.3	60.31	ug/L	0.278	114.9	mg/kg	0.53	0.46%
Fe 273.955†	5201752.9	77010	ug/L	673.08	146700	mg/kg	1282.05	0.87%
K 766.490†	991.0	125.9	ug/L	3.59	239.9	mg/kg	6.84	2.85%
Mg 285.213†	12147.9	919.5	ug/L	4.70	1751	mg/kg	8.96	0.51%
Mn 257.610†	26597.3	378.5	ug/L	2.15	720.9	mg/kg	4.10	0.57%
Mo 202.031†	77.0	3.895	ug/L	0.3063	7.420	mg/kg	0.5834	7.86%
Na 589.592†	192.6	32.71	ug/L	4.911	62.31	mg/kg	9.355	15.01%
Ni 231.604†	2592.8	61.64	ug/L	0.180	117.4	mg/kg	0.34	0.29%
P 213.617†	188.0	65.13	ug/L	1.243	124.1	mg/kg	2.37	1.91%
Pb 220.353†	3609.1	642.9	ug/L	0.33	1225	mg/kg	0.63	0.05%
Sb 206.836†	31.7	5.003	ug/L	0.6684	9.529	mg/kg	1.2732	13.36%
Se 196.026†	-74.7	1.882	ug/L	1.9468	3.584	mg/kg	3.7082	103.46%
Sn 189.927†	24.1	4.910	ug/L	0.8294	9.353	mg/kg	1.5798	16.89%
Sr 421.552†	2107.8	6.209	ug/L	0.0350	11.83	mg/kg	0.067	0.56%
Ti 337.279†	1624.2	50.89	ug/L	0.651	96.94	mg/kg	1.240	1.28%
Tl 190.801†	-1.0	-0.196	ug/L	0.6849	-0.374	mg/kg	1.3046	348.56%
V 292.402†	752.2	8.035	ug/L	0.0335	15.31	mg/kg	0.064	0.42%
Zn 206.200†	22229.3	274.0	ug/L	1.07	521.9	mg/kg	2.04	0.39%
Li 670.784	23373.3	1.513	ug/L	0.0083	2.882	mg/kg	0.0159	0.55%

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB030-0607

APPL ID: AZ16139

Sample Collection Date: 05/07/15

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 17.5 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	9150	1210.0	97.00	48.00	mg/Kg	20	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.8J	6.1	4.80	1.90	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	9.6	0.81	0.610	0.300	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	21.5	9.7	4.80	1.40	mg/Kg	1	06/04/15	06/12/15

J = Estimated value.

Printed: 06/16/15 1:10:24 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Sequence No.: 54
Sample ID: AZ16139S01
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.02 g
Dilution:
Wash Time:

Autosampler Location: 54
Date Collected: 06/12/15 7:39:57 PM
Data Type: Reprocessed on 06/15/15 10:36:41 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16139S01

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	272640.2	134.2	%	1.19				0.89%
Y Radial	18904.3	144.1	%	3.48				2.41%
Ag 338.289†	289.5	1.782	ug/L	0.1968	0.175	mg/kg	0.0193	11.05%
Al 308.215†	71749.2	63020	ug/L	1362.32	6178	mg/kg	133.56	2.16%
As 188.979†	77.6	24.09	ug/L	1.260	2.362	mg/kg	0.1236	5.23%
B 249.677†	-3615.6	41.06	ug/L	0.583	4.025	mg/kg	0.0572	1.42%
Ba 233.527†	75492.0	199.2	ug/L	0.85	19.53	mg/kg	0.083	0.42%
Be 313.107†	6657.1	2.060	ug/L	0.0144	0.202	mg/kg	0.0014	0.70%
Ca 315.887†	51572.7	3996	ug/L	87.97	391.7	mg/kg	8.62	2.20%
Cd 214.440†	837.2	0.241	ug/L	0.0180	0.024	mg/kg	0.0018	7.47%
Co 228.616†	1744.0	16.85	ug/L	0.092	1.652	mg/kg	0.0090	0.55%
Cr 267.716†	8741.7	112.3	ug/L	0.96	11.01	mg/kg	0.094	0.85%
Cu 327.393†	7981.5	48.03	ug/L	0.231	4.709	mg/kg	0.0226	0.48%
Fe 273.955†	4374465.0	64750	ug/L	393.00	6348	mg/kg	38.53	0.61%
K 766.490†	41292.5	7130	ug/L	143.26	699.0	mg/kg	14.04	2.01%
Mg 285.213†	101779.0	7224	ug/L	146.09	708.3	mg/kg	14.32	2.02%
Mn 257.610†	13315.8	189.6	ug/L	4.16	18.58	mg/kg	0.408	2.20%
Mo 202.031†	124.8	5.239	ug/L	0.0635	0.514	mg/kg	0.0062	1.21%
Na 589.592†	7767.5	1336	ug/L	33.89	130.9	mg/kg	3.32	2.54%
Ni 231.604†	1578.6	37.69	ug/L	0.229	3.695	mg/kg	0.0225	0.61%
P 213.617†	1220.4	422.9	ug/L	2.53	41.46	mg/kg	0.248	0.60%
Pb 220.353†	238.7	46.68	ug/L	0.118	4.577	mg/kg	0.0116	0.25%
Sb 206.836†	12.6	1.994	ug/L	1.1763	0.195	mg/kg	0.1153	59.00%
Se 196.026†	-70.0	1.149	ug/L	0.7652	0.113	mg/kg	0.0750	66.60%
Sn 189.927†	20.2	4.105	ug/L	0.4289	0.402	mg/kg	0.0420	10.45%
Sr 421.552†	14190.4	42.52	ug/L	0.907	4.169	mg/kg	0.0889	2.13%
Ti 337.279†	23753.3	746.4	ug/L	15.34	73.18	mg/kg	1.503	2.05%
Tl 190.801†	-23.1	-4.618	ug/L	0.3492	-0.453	mg/kg	0.0342	7.56%
V 292.402†	15712.4	167.9	ug/L	1.41	16.46	mg/kg	0.139	0.84%
Zn 206.200†	14600.2	180.5	ug/L	4.00	17.69	mg/kg	0.392	2.22%
Li 670.784	1251520.0	81.02	ug/L	0.347	7.943	mg/kg	0.0341	0.43%

Sequence No.: 34
Sample ID: AZ16139S01-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.02 g
Dilution: 20X
Wash Time:

Autosampler Location: 70
Date Collected: 06/12/15 5:23:46 PM
Data Type: Reprocessed on 06/15/15 10:36:18 AM

Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16139S01-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	210849.0	103.8	%	0.68				0.65%
Y Radial	14216.6	108.4	%	0.23				0.21%
Ag 338.289†	0.5	0.003	ug/L	0.2451	0.006	mg/kg	0.4806	>999.9%
Al 308.215†	4386.2	3853	ug/L	3.98	7554	mg/kg	7.81	0.10%
As 188.979†	7.3	2.066	ug/L	0.5701	4.052	mg/kg	1.1179	27.59%
B 249.677†	-220.3	2.951	ug/L	0.2657	5.786	mg/kg	0.5210	9.00%
Ba 233.527†	4775.5	12.60	ug/L	0.089	24.70	mg/kg	0.174	0.70%
Be 313.107†	332.6	0.103	ug/L	0.0119	0.202	mg/kg	0.0234	11.58%
Ca 315.887†	3192.6	247.3	ug/L	1.66	485.0	mg/kg	3.25	0.67%
Cd 214.440†	53.8	0.014	ug/L	0.0209	0.028	mg/kg	0.0410	145.95%
Co 228.616†	132.0	1.318	ug/L	0.0435	2.584	mg/kg	0.0853	3.30%
Cr 267.716†	518.4	6.659	ug/L	0.1113	13.06	mg/kg	0.218	1.67%
Cu 327.393†	674.7	3.998	ug/L	0.1318	7.840	mg/kg	0.2584	3.30%
Fe 273.955†	278816.0	4127	ug/L	14.39	8093	mg/kg	28.21	0.35%
K 766.490†	2552.9	440.8	ug/L	4.67	864.3	mg/kg	9.15	1.06%
Mg 285.213†	6282.1	446.0	ug/L	0.84	874.6	mg/kg	1.64	0.19%
Mn 257.610†	837.3	11.92	ug/L	0.173	23.37	mg/kg	0.339	1.45%
Mo 202.031†	16.5	0.612	ug/L	0.2998	1.200	mg/kg	0.5879	48.99%
Na 589.592†	301.3	51.50	ug/L	5.702	101.0	mg/kg	11.18	11.07%
Ni 231.604†	96.9	2.317	ug/L	0.1562	4.542	mg/kg	0.3062	6.74%
P 213.617†	93.8	32.52	ug/L	1.275	63.76	mg/kg	2.501	3.92%
Pb 220.353†	15.2	2.941	ug/L	0.0631	5.766	mg/kg	0.1237	2.15%
Sb 206.836†	-1.7	-0.267	ug/L	0.5445	-0.524	mg/kg	1.0677	203.57%
Se 196.026†	-7.5	-1.048	ug/L	1.0640	-2.055	mg/kg	2.0864	101.51%
Sn 189.927†	2.8	0.573	ug/L	0.1306	1.124	mg/kg	0.2560	22.79%
Sr 421.552†	885.4	2.653	ug/L	0.0172	5.202	mg/kg	0.0337	0.65%
Ti 337.279†	1488.3	46.77	ug/L	0.198	91.70	mg/kg	0.387	0.42%
Tl 190.801†	0.7	0.149	ug/L	0.8956	0.291	mg/kg	1.7560	602.95%
V 292.402†	948.8	10.14	ug/L	0.084	19.87	mg/kg	0.165	0.83%
Zn 206.200†	594.2	7.299	ug/L	0.2182	14.31	mg/kg	0.428	2.99%
Li 670.784	42030.5	2.721	ug/L	0.0101	5.335	mg/kg	0.0199	0.37%

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack
Project: CTO JU11 112G02622 NSF Indian Head
Sample ID: S67-SS032-0001
Sample Collection Date: 05/07/15

ARF: 76380
APPL ID: AZ16140

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	2330	56.0	4.50	2.20	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	4.50U	5.6	4.50	1.70	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	2.7	0.75	0.560	0.280	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	150	8.9	4.50	1.30	mg/Kg	1	06/04/15	06/12/15

Printed: 06/16/15 1:10:24 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Sequence No.: 55
Sample ID: AZ16140S01
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1 g
Dilution:
Wash Time:

Autosampler Location: 55
Date Collected: 06/12/15 7:47:13 PM
Data Type: Reprocessed on 06/15/15 10:36:42 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16140S01

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	243399.0	119.8	%	0.61				0.51%
Y Radial	17056.1	130.0	%	1.89				1.46%
Ag 338.289†	1897.1	11.68	ug/L	0.630	1.168	mg/kg	0.0630	5.39%
Al 308.215†	23799.0	20900	ug/L	449.38	2090	mg/kg	44.94	2.15%
As 188.979†	552.8	167.1	ug/L	1.39	16.71	mg/kg	0.139	0.83%
B 249.677†	-27297.9	-51.67	ug/L	8.409	-5.167	mg/kg	0.8409	16.27%
Ba 233.527†	294488.4	771.7	ug/L	0.93	77.17	mg/kg	0.093	0.12%
Be 313.107†	3378.7	1.045	ug/L	0.0044	0.105	mg/kg	0.0004	0.42%
Ca 315.887†	242034.4	18750	ug/L	381.61	1875	mg/kg	38.16	2.03%
Cd 214.440†	5476.0	0.310	ug/L	0.0404	0.031	mg/kg	0.0040	13.02%
Co 228.616†	4534.4	41.31	ug/L	0.350	4.131	mg/kg	0.0350	0.85%
Cr 267.716†	8808.8	112.0	ug/L	0.96	11.20	mg/kg	0.096	0.85%
Cu 327.393†	200038.8	1141	ug/L	0.28	114.1	mg/kg	0.03	0.02%
Fe 273.955†	22859543.8	338400	ug/L	2294.80	33840	mg/kg	229.48	0.68%
K 766.490†	32082.5	5374	ug/L	116.98	537.4	mg/kg	11.70	2.18%
Mg 285.213†	70822.5	5270	ug/L	105.28	527.0	mg/kg	10.53	2.00%
Mn 257.610†	40886.4	582.9	ug/L	12.02	58.29	mg/kg	1.202	2.06%
Mo 202.031†	201.4	12.49	ug/L	0.474	1.249	mg/kg	0.0474	3.79%
Na 589.592†	8924.1	1541	ug/L	31.70	154.1	mg/kg	3.17	2.06%
Ni 231.604†	10523.6	250.9	ug/L	1.84	25.09	mg/kg	0.184	0.73%
P 213.617†	1815.3	629.0	ug/L	4.60	62.90	mg/kg	0.460	0.73%
Pb 220.353†	51253.7	9266	ug/L	19.32	926.6	mg/kg	1.93	0.21%
Sb 206.836†	90.5	14.29	ug/L	1.209	1.429	mg/kg	0.1209	8.46%
Se 196.026†	-215.6	50.40	ug/L	3.324	5.040	mg/kg	0.3324	6.60%
Sn 189.927†	165.1	33.59	ug/L	0.114	3.359	mg/kg	0.0114	0.34%
Sr 421.552†	42050.5	125.6	ug/L	2.82	12.56	mg/kg	0.282	2.25%
Ti 337.279†	39381.6	1237	ug/L	25.84	123.7	mg/kg	2.58	2.09%
Tl 190.801†	-1.1	-0.224	ug/L	1.7219	-0.022	mg/kg	0.1722	769.64%
V 292.402†	9568.9	102.2	ug/L	0.84	10.22	mg/kg	0.084	0.82%
Zn 206.200†	108864.2	1342	ug/L	2.36	134.2	mg/kg	0.24	0.18%
Li 670.784	363105.7	23.51	ug/L	0.081	2.351	mg/kg	0.0081	0.34%

Metals Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

ARF: 76380

Sample ID: S67-SB032-0607

APPL ID: AZ16141

Sample Collection Date: 05/07/15

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 23.3 Percent Moisture.)									
6010C/3050B	ALUMINUM (AL)	1710	65.0	5.20	2.60	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	BORON (B)	6.6	6.5	5.20	2.00	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	LITHIUM (LI)	1.7	0.87	0.650	0.330	mg/Kg	1	06/04/15	06/12/15
6010C/3050B	ZINC (ZN)	179	10.0	5.20	1.50	mg/Kg	1	06/04/15	06/12/15

Printed: 06/16/15 1:10:24 PM

PL-F1-SC-MCRes/MCPQL-REG MDLs

Sequence No.: 56
Sample ID: AZ16141S01
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.05 g
Dilution:
Wash Time:

Autosampler Location: 56
Date Collected: 06/12/15 7:54:23 PM
Data Type: Reprocessed on 06/15/15 10:36:43 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16141S01

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	217862.6	107.2	%	0.27				0.25%
Y Radial	14904.0	113.6	%	3.13				2.75%
Ag 338.289†	1159.4	7.138	ug/L	0.3128	0.680	mg/kg	0.0298	4.38%
Al 308.215†	15670.8	13750	ug/L	330.16	1310	mg/kg	31.44	2.40%
As 188.979†	269.3	91.58	ug/L	1.167	8.722	mg/kg	0.1111	1.27%
B 249.677†	-16541.6	54.05	ug/L	1.561	5.148	mg/kg	0.1486	2.89%
Ba 233.527†	213819.4	560.6	ug/L	2.76	53.39	mg/kg	0.263	0.49%
Be 313.107†	2675.9	0.828	ug/L	0.0072	0.079	mg/kg	0.0007	0.87%
Ca 315.887†	126824.0	9824	ug/L	239.17	935.7	mg/kg	22.78	2.43%
Cd 214.440†	3380.9	-1.703	ug/L	0.0203	-0.162	mg/kg	0.0019	1.19%
Co 228.616†	1905.0	10.65	ug/L	0.080	1.014	mg/kg	0.0076	0.75%
Cr 267.716†	17368.8	222.3	ug/L	0.61	21.17	mg/kg	0.058	0.27%
Cu 327.393†	38148.4	228.2	ug/L	0.97	21.73	mg/kg	0.093	0.43%
Fe 273.955†	16190162.7	239700	ug/L	702.09	22830	mg/kg	66.87	0.29%
K 766.490†	74189.4	12770	ug/L	300.26	1216	mg/kg	28.60	2.35%
Mg 285.213†	47360.8	3534	ug/L	78.95	336.6	mg/kg	7.52	2.23%
Mn 257.610†	60428.3	860.2	ug/L	20.36	81.92	mg/kg	1.939	2.37%
Mo 202.031†	979.7	36.14	ug/L	0.142	3.441	mg/kg	0.0135	0.39%
Na 589.592†	15354.0	2652	ug/L	66.93	252.5	mg/kg	6.37	2.52%
Ni 231.604†	6520.3	155.8	ug/L	0.54	14.84	mg/kg	0.052	0.35%
P 213.617†	2725.4	944.4	ug/L	2.39	89.94	mg/kg	0.227	0.25%
Pb 220.353†	15800.6	2832	ug/L	4.80	269.7	mg/kg	0.46	0.17%
Sb 206.836†	81.1	12.81	ug/L	2.317	1.220	mg/kg	0.2206	18.09%
Se 196.026†	-145.9	38.29	ug/L	1.461	3.647	mg/kg	0.1391	3.82%
Sn 189.927†	102.1	20.76	ug/L	0.436	1.977	mg/kg	0.0415	2.10%
Sr 421.552†	42946.5	128.7	ug/L	3.15	12.26	mg/kg	0.300	2.45%
Ti 337.279†	80741.1	2538	ug/L	58.26	241.7	mg/kg	5.55	2.30%
Tl 190.801†	-26.2	-5.236	ug/L	0.3516	-0.499	mg/kg	0.0335	6.72%
V 292.402†	5637.7	60.23	ug/L	0.082	5.736	mg/kg	0.0078	0.14%
Zn 206.200†	115849.9	1435	ug/L	6.80	136.7	mg/kg	0.65	0.47%
Li 670.784	213780.0	13.84	ug/L	0.078	1.318	mg/kg	0.0074	0.56%

METALS

Calibration Data

APPL, INC.

A.P.P.L. INC.
2A
INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Tetra Tech
ARF No: 76380 SDG: 76380
Initial Calibration Source: CPI
Continuing Calibration Source: Environmental Express
Analysis Date: 06/12/15 Concentration Units: ug/L

Analyte	Initial Calibration			Continuing Calibration						M
	True	Found 13:26	%R(1)	True CCV1	Found 13:57	%R(1)	True CCV2	Found 15:02	%R(1)	
Aluminum (Al)	12500	12560	100	10000	9912	99.1	7500	7411	98.8	P
Boron (B)	500	526.1	105	500	514	103	375	393.7	105	P
Lithium (Li)	250	270	108	250	239.5	95.8	187.5	177.8	94.8	P
Zinc (Zn)	500	521.6	104	500	507.3	101	375	396.2	106	P

(1) Control Limits: Metals 90-110

A.P.P.L. INC.

2A

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No: 76380 SDG: 76380

Initial Calibration Source: CPI

Continuing Calibration Source: Environmental Express

Analysis Date: 06/12/15 Concentration Units: ug/L

Analyte	Initial Calibration			Continuing Calibration						M
	True	Found 13:26	%R(1)	True CCV1	Found 16:29	%R(1)	True CCV2	Found 18:09	%R(1)	
Aluminum (Al)	12500	12560	100	10000	9995	100	7500	7483	99.8	P
Boron (B)	500	526.1	105	500	512.6	103	375	396.6	106	P
Lithium (Li)	250	270	108	250	243.6	97.4	187.5	172.3	91.9	P
Zinc (Zn)	500	521.6	104	500	506.5	101	375	391.3	104	P

(1) Control Limits: Metals 90-110

ILM02.0

A.P.P.L. INC.
2A
INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Tetra Tech
ARF No: 76380 SDG: 76380
Initial Calibration Source: CPI
Continuing Calibration Source: Environmental Express
Analysis Date: 06/12/15 Concentration Units: ug/L

Analyte	Initial Calibration			Continuing Calibration						M
	True	Found 13:26	%R(1)	True CCV1	Found 18:42	%R(1)	True CCV2	Found 20:01	%R(1)	
Aluminum (Al)	12500	12560	100	10000	10180	102	7500	7137	95.2	P
Boron (B)	500	526.1	105	500	521.3	104	375	391	104	P
Lithium (Li)	250	270	108	250	258.8	104	187.5	193.9	103	P
Zinc (Zn)	500	521.6	104	500	536.1	107	375	402.6	107	P

(1) Control Limits: Metals 90-110

A.P.P.L. INC.
3
BLANKS

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No.: 76380 SDG: 76380

Preparation Blank Matrix (soil/water): soil

Preparation Blank Concentration Units (ug/L or mg/kg): mg/Kg

Analysis Date: 06/12/15

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)								Preparation Blank		M
		C	1	C	2	C	3	C		C			
	13:37		14:08		15:08		16:34		14:26				
Aluminum (Al)	500.00	U	500.00	U	500.00	U	500.00	U	50.00	U	P		
Boron (B)	50.00	U	50.00	U	50.00	U	50.00	U	5.00	U	P		
Lithium (Li)	6.70	U	6.70	U	6.70	U	6.70	U	.67	U	P		
Zinc (Zn)	80.00	U	80.00	U	80.00	U	80.00	U	8.00	U	P		

A.P.P.L. INC.
3
BLANKS

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No.: 76380 SDG: 76380

Preparation Blank Matrix (soil/water): soil

Preparation Blank Concentration Units (ug/L or mg/kg): mg/Kg

Analysis Date: 06/12/15

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)								Preparation Blank		M
		C	1	C	2	C	3	C		C			
	13:37		18:15		18:52		20:11				14:26		
Aluminum (Al)	500.00	U	500.00	U	500.00	U	500.00	U			50.00	U	P
Boron (B)	50.00	U	50.00	U	50.00	U	50.00	U			5.00	U	P
Lithium (Li)	6.70	U	6.70	U	6.70	U	6.70	U			.67	U	P
Zinc (Zn)	80.00	U	80.00	U	80.00	U	80.00	U			8.00	U	P

A.P.P.L. INC.
4
ICP INTERFERENCE CHECK SAMPLE

Lab Name: A.P.P.L. INC. Contract: Tetra Tech
ARF No.: 76380 SDG: 76380
ICP ID Number: Jules ICS Source: Environmental Express

Analysis Date: 06/12/15 Concentration Units: ug/L

Analyte	True		Initial Found		
	Sol A	Sol AB	Sol A 13:43	Sol AB 13:50	%R(1)
Aluminum (Al)	100000	100000	100300	99840	99.8
Boron (B)			-14.65	-15.98	
Calcium (Ca)	100000	100000	103800	103400	103
Iron (Fe)	100000	100000	98980	98220	98.2
Lithium (Li)		0	0.8	0.68	
Magnesium (Mg)	100000	100000	100000	100400	100
Zinc (Zn)		500	0.999	485.9	97.2

(1) Control Limits: Metals 80-120

A.P.P.L. INC.

5B

POST DIGEST SPIKE SAMPLE RECOVERY

CLIENT SAMPLE NO.

S67-SS030-0001

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No.: 76380 SDG: 76380

Analysis Date: 06/12/15 Concentration Units: mg/kg

Analyte	Control Limit %R	Spiked Sample Result (SSR)	C	Sample Result (SR)	C	Spike Added (SA)	%R	Q	M
Boron (B)	75-125	925		12.67		990.099	92.2		
Lithium (Li)	75-125	424		3.733		396.040	106		

Comments:

06/12/15 16:41 AZ16137S01-1/20

06/12/15 17:02 AZ16137S01-A-1/20

Sequence No.: 31
Sample ID: AZ16137S01-A-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.01 g
Dilution: 20X
Wash Time:

Autosampler Location: 67
Date Collected: 06/12/15 5:02:54 PM
Data Type: Reprocessed on 06/15/15 10:36:15 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16137S01-A-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	197670.7	97.28	%	0.803				0.82%
Y Radial	13773.6	105.0	%	0.26				0.25%
Ag 338.289†	29591.0	182.2	ug/L	1.32	360.7	mg/kg	2.62	0.73%
Al 308.215†	6819.6	6008	ug/L	39.97	11900	mg/kg	79.16	0.67%
As 188.979†	1885.6	417.8	ug/L	1.94	827.4	mg/kg	3.85	0.46%
B 249.677†	10862.9	467.1	ug/L	1.25	925.0	mg/kg	2.47	0.27%
Ba 233.527†	197065.8	526.6	ug/L	3.16	1043	mg/kg	6.26	0.60%
Be 313.107†	292284.4	90.42	ug/L	0.440	179.1	mg/kg	0.87	0.49%
Ca 315.887†	601887.1	46650	ug/L	127.96	92370	mg/kg	253.39	0.27%
Cd 214.440†	27260.1	109.2	ug/L	0.76	216.2	mg/kg	1.50	0.69%
Co 228.616†	39600.7	456.2	ug/L	2.12	903.4	mg/kg	4.21	0.47%
Cr 267.716†	41634.7	536.3	ug/L	2.38	1062	mg/kg	4.72	0.44%
Cu 327.393†	88813.5	502.9	ug/L	2.31	995.8	mg/kg	4.58	0.46%
Fe 273.955†	3143167.6	46500	ug/L	245.18	92070	mg/kg	485.50	0.53%
K 766.490†	52363.5	9038	ug/L	33.23	17900	mg/kg	65.81	0.37%
Mg 285.213†	644197.2	45420	ug/L	171.71	89950	mg/kg	340.03	0.38%
Mn 257.610†	49474.5	701.8	ug/L	3.46	1390	mg/kg	6.85	0.49%
Mo 202.031†	13933.6	454.4	ug/L	3.02	899.9	mg/kg	5.97	0.66%
Na 589.592†	268414.2	46620	ug/L	172.48	92320	mg/kg	341.54	0.37%
Ni 231.604†	22520.0	522.5	ug/L	3.28	1035	mg/kg	6.50	0.63%
P 213.617†	11200.7	3881	ug/L	28.71	7686	mg/kg	56.84	0.74%
Pb 220.353†	5024.2	910.1	ug/L	5.64	1802	mg/kg	11.16	0.62%
Sb 206.836†	3107.8	490.7	ug/L	2.77	971.7	mg/kg	5.49	0.56%
Se 196.026†	1211.1	461.3	ug/L	6.51	913.5	mg/kg	12.88	1.41%
Sn 189.927†	2154.2	438.2	ug/L	3.90	867.8	mg/kg	7.73	0.89%
Sr 421.552†	154740.5	464.1	ug/L	2.04	918.9	mg/kg	4.03	0.44%
Ti 337.279†	16385.0	514.3	ug/L	1.49	1018	mg/kg	2.94	0.29%
Tl 190.801†	2177.6	434.8	ug/L	5.14	861.0	mg/kg	10.19	1.18%
V 292.402†	43822.4	468.1	ug/L	1.65	927.0	mg/kg	3.26	0.35%
Zn 206.200†	91772.4	1151	ug/L	6.24	2280	mg/kg	12.36	0.54%
Li 670.784	3307717.2	214.1	ug/L	0.36	424.0	mg/kg	0.71	0.17%

A.P.P.L. INC.
9
ICP SERIAL DILUTION

CLIENT SAMPLE NO.
S67-SS030-0001

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No.: 76380 SDG: 76380

Matrix: soil

Analysis Date: 06/12/15 Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)	C	Serial Dilution Result (S)	C	%D	Q	M
Aluminum (Al)	4877		4885		0.164		
Zinc (Zn)	466.7		477.9		2.40		

Comments:

06/12/15 16:41 AZ16137S01-1/20

06/12/15 17:10 AZ16137S01-1/100

Sequence No.: 32
Sample ID: AZ16137S01-1/100
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.01 g
Dilution: 100X
Wash Time:

Autosampler Location: 68
Date Collected: 06/12/15 5:10:14 PM
Data Type: Reprocessed on 06/15/15 10:36:16 AM

Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16137S01-1/100

Analyte	Mean Corrected Intensity	Conc.	Calib. Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	202967.7	99.89	%	0.301				0.30%
Y Radial	13845.2	105.5	%	0.04				0.04%
Ag 338.289†	-29.0	-0.178	ug/L	0.1905	-1.766	mg/kg	1.8857	106.77%
Al 308.215†	562.0	493.4	ug/L	3.38	4885	mg/kg	33.46	0.68%
As 188.979†	2.6	1.917	ug/L	0.9391	18.98	mg/kg	9.298	48.98%
B 249.677†	-699.0	1.873	ug/L	0.0807	18.54	mg/kg	0.799	4.31%
Ba 233.527†	6941.1	18.08	ug/L	0.059	179.0	mg/kg	0.59	0.33%
Be 313.107†	73.2	0.023	ug/L	0.0078	0.224	mg/kg	0.0768	34.24%
Ca 315.887†	2851.3	220.8	ug/L	1.11	2186	mg/kg	11.00	0.50%
Cd 214.440†	903.3	3.041	ug/L	0.0197	30.11	mg/kg	0.195	0.65%
Co 228.616†	191.4	1.930	ug/L	0.0852	19.10	mg/kg	0.844	4.42%
Cr 267.716†	1908.5	24.54	ug/L	0.184	242.9	mg/kg	1.82	0.75%
Cu 327.393†	1751.7	10.20	ug/L	0.205	101.0	mg/kg	2.03	2.01%
Fe 273.955†	672141.6	9951	ug/L	29.61	98520	mg/kg	293.20	0.30%
K 766.490†	265.5	40.11	ug/L	5.182	397.1	mg/kg	51.31	12.92%
Mg 285.213†	1703.7	128.3	ug/L	0.40	1270	mg/kg	3.91	0.31%
Mn 257.610†	3752.5	53.39	ug/L	0.056	528.6	mg/kg	0.56	0.11%
Mo 202.031†	28.0	1.092	ug/L	0.2267	10.81	mg/kg	2.245	20.77%
Na 589.592†	-4.1	-0.858	ug/L	2.1144	-8.493	mg/kg	20.9345	246.50%
Ni 231.604†	652.2	15.32	ug/L	0.231	151.7	mg/kg	2.29	1.51%
P 213.617†	59.6	20.64	ug/L	1.025	204.4	mg/kg	10.15	4.97%
Pb 220.353†	586.5	105.0	ug/L	0.38	1040	mg/kg	3.72	0.36%
Sb 206.836†	15.8	2.498	ug/L	0.9469	24.74	mg/kg	9.375	37.90%
Se 196.026†	-8.7	0.595	ug/L	1.2517	5.887	mg/kg	12.3935	210.54%
Sn 189.927†	12.2	2.478	ug/L	0.2851	24.54	mg/kg	2.823	11.51%
Sr 421.552†	437.6	1.300	ug/L	0.0564	12.88	mg/kg	0.558	4.33%
Ti 337.279†	389.9	12.23	ug/L	0.126	121.1	mg/kg	1.24	1.03%
Tl 190.801†	-3.5	-0.690	ug/L	0.7888	-6.837	mg/kg	7.8102	114.24%
V 292.402†	189.2	2.021	ug/L	0.0254	20.01	mg/kg	0.251	1.26%
Zn 206.200†	3890.1	48.27	ug/L	0.331	477.9	mg/kg	3.28	0.69%
Li 670.784	8508.5	0.551	ug/L	0.0096	5.453	mg/kg	0.0947	1.74%

Reprocessing Begun
Logged In Analyst: chemist_metals Technique: ICP Continuous

Results Data Set (original): 150612A6010
Results Library (original): C:\Documents and Settings\All Users\PerkinElmer\ICP\Data\Results\Results.mdb
Results Data Set (reprocessed):
Results Library (reprocessed):

Sequence No.: 1 Autosampler Location: 1
Sample ID: CalBlk 150612RJS I:JL O:R Date Collected: 06/12/15 12:56:41 PM
Analyst: Data Type: Reprocessed on 06/15/15 10:35:40 AM
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: Initial Sample Vol:
Dilution: Sample Prep Vol:
Wash Time:

Mean Data: CalBlk 150612RJS I:JL O:R

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc.	Calib Units
Y 371.029	203189.8	528.51	0.26%	100.0	%
Y Radial	13119.9	152.02	1.16%	100.0	%
Ag 338.289†	-58.7	52.66	89.65%	[0.00]	ug/L
Al 308.215†	71.0	5.32	7.49%	[0.00]	ug/L
As 188.979†	-76.5	4.42	5.78%	[0.00]	ug/L
B 249.677†	49.2	9.52	19.36%	[0.00]	ug/L
Ba 233.527†	-54.1	6.40	11.84%	[0.00]	ug/L
Be 313.107†	2142.9	12.43	0.58%	[0.00]	ug/L
Ca 315.887†	-306.4	16.52	5.39%	[0.00]	ug/L
Cd 214.440†	-389.3	13.59	3.49%	[0.00]	ug/L
Co 228.616†	-225.7	1.70	0.75%	[0.00]	ug/L
Cr 267.716†	27.6	9.11	33.04%	[0.00]	ug/L
Cu 327.393†	-1695.6	55.62	3.28%	[0.00]	ug/L
Fe 273.955†	-25.0	4.12	16.51%	[0.00]	ug/L
K 766.490†	-793.3	122.41	15.43%	[0.00]	ug/L
Mg 285.213†	82.5	16.38	19.86%	[0.00]	ug/L
Mn 257.610†	-67.5	5.40	8.01%	[0.00]	ug/L
Mo 202.031†	-76.5	3.57	4.67%	[0.00]	ug/L
Na 589.592†	390.3	32.76	8.39%	[0.00]	ug/L
Ni 231.604†	-145.8	8.82	6.05%	[0.00]	ug/L
P 213.617†	-70.3	4.21	6.00%	[0.00]	ug/L
Pb 220.353†	25.2	5.67	22.49%	[0.00]	ug/L
Sb 206.836†	-9.4	7.34	78.01%	[0.00]	ug/L
Se 196.026†	17.2	1.37	7.97%	[0.00]	ug/L
Sn 189.927†	0.7	0.53	77.76%	[0.00]	ug/L
Sr 421.552†	-57.6	26.79	46.54%	[0.00]	ug/L
Ti 337.279†	-36.1	7.84	21.72%	[0.00]	ug/L
Tl 190.801†	-10.8	2.57	23.88%	[0.00]	ug/L
V 292.402†	-143.3	6.85	4.78%	[0.00]	ug/L
Zn 206.200†	360.2	24.31	6.75%	[0.00]	ug/L
Li 670.784	4402.2	42.58	0.97%	[0.00]	ug/L

Sequence No.: 2
Sample ID: STD 1 150612RJS I:JL O:RJ
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 2
Date Collected: 06/12/15 1:02:53 PM
Data Type: Reprocessed on 06/15/15 10:35:42 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: STD 1 150612RJS I:JL O:RJ

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc.	Calib Units
Y 371.029	204890.8	1033.91	0.50%	100.8	%
Y Radial	13852.4	50.05	0.36%	105.6	%
Ag 338.289†	943.9	19.23	2.04%	[0.5]	ug/L
Al 308.215†	274.5	2.69	0.98%	[50]	ug/L
As 188.979†	53.8	6.12	11.36%	[2]	ug/L
B 249.677†	1112.2	18.97	1.71%	[25]	ug/L
Ba 233.527†	4520.4	18.72	0.41%	[1.5]	ug/L
Be 313.107†	34091.0	157.69	0.46%	[1]	ug/L
Ca 315.887†	7316.1	11.53	0.16%	[50]	ug/L
Cd 214.440†	2646.1	16.52	0.62%	[0.25]	ug/L
Co 228.616†	1152.3	4.27	0.37%	[2.5]	ug/L
Cr 267.716†	800.6	7.59	0.95%	[0.5]	ug/L
Cu 327.393†	2201.8	41.05	1.86%	[2.5]	ug/L
Fe 273.955†	15428.0	61.25	0.40%	[25]	ug/L
K 766.490†	3978.3	57.45	1.44%	[500]	ug/L
Mg 285.213†	7438.0	10.16	0.14%	[25]	ug/L
Mn 257.610†	796.5	3.59	0.45%	[1]	ug/L
Mo 202.031†	337.4	3.03	0.90%	[1]	ug/L
Na 589.592†	4283.1	60.29	1.41%	[500]	ug/L
Ni 231.604†	486.6	8.37	1.72%	[1]	ug/L
P 213.617†	178.4	1.56	0.87%	[12.5]	ug/L
Pb 220.353†	65.7	5.01	7.63%	[1.5]	ug/L
Sb 206.836†	70.2	5.98	8.52%	[2]	ug/L
Se 196.026†	28.3	8.55	30.26%	[2]	ug/L
Sn 189.927†	64.0	1.97	3.08%	[3]	ug/L
Sr 421.552†	3541.9	36.02	1.02%	[1]	ug/L
Ti 337.279†	401.8	8.54	2.13%	[2.5]	ug/L
Tl 190.801†	66.8	7.25	10.84%	[2]	ug/L
V 292.402†	1022.4	8.88	0.87%	[0.5]	ug/L
Zn 206.200†	2876.2	22.01	0.77%	[25]	ug/L
Li 670.784	52749.1	275.23	0.52%	[5]	ug/L

Sequence No.: 3
Sample ID: STD 2 150612RJS I:JL O:RJ
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 3
Date Collected: 06/12/15 1:09:07 PM
Data Type: Reprocessed on 06/15/15 10:35:43 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: STD 2 150612RJS I:JL O:RJ

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc.	Calib Units
Y 371.029	194907.2	898.33	0.46%	95.92	%
Y Radial	13937.3	868.97	6.23%	106.2	%
Ag 338.289†	40195.1	144.36	0.36%	[250]	ug/L
Al 308.215†	10912.3	637.90	5.85%	[10000]	ug/L
As 188.979†	2243.3	24.96	1.11%	[500]	ug/L
B 249.677†	15089.9	90.81	0.60%	[500]	ug/L
Ba 233.527†	185331.6	600.96	0.32%	[500]	ug/L
Be 313.107†	1605818.8	3121.77	0.19%	[500]	ug/L
Ca 315.887†	313240.7	17959.45	5.73%	[25000]	ug/L
Cd 214.440†	122579.4	317.91	0.26%	[500]	ug/L
Co 228.616†	43316.8	196.51	0.45%	[500]	ug/L
Cr 267.716†	38777.0	318.42	0.82%	[500]	ug/L
Cu 327.393†	87449.0	378.93	0.43%	[500]	ug/L
Fe 273.955†	675336.8	2347.57	0.35%	[10000]	ug/L
K 766.490†	54597.7	2935.00	5.38%	[10000]	ug/L
Mg 285.213†	346008.2	19459.49	5.62%	[25000]	ug/L
Mn 257.610†	33896.6	1946.96	5.74%	[500]	ug/L
Mo 202.031†	15278.1	142.03	0.93%	[500]	ug/L
Na 589.592†	69357.2	3914.69	5.64%	[12500]	ug/L
Ni 231.604†	21743.8	151.61	0.70%	[500]	ug/L
P 213.617†	7124.8	61.35	0.86%	[2500]	ug/L
Pb 220.353†	2766.8	27.88	1.01%	[500]	ug/L
Sb 206.836†	3128.4	26.08	0.83%	[500]	ug/L
Se 196.026†	1351.3	16.09	1.19%	[500]	ug/L
Sn 189.927†	2452.2	21.03	0.86%	[500]	ug/L
Sr 421.552†	158921.9	10051.03	6.32%	[500]	ug/L
Ti 337.279†	15315.6	864.82	5.65%	[500]	ug/L
Tl 190.801†	2541.8	23.48	0.92%	[500]	ug/L
V 292.402†	46344.3	19.23	0.04%	[500]	ug/L
Zn 206.200†	39707.0	170.01	0.43%	[500]	ug/L
Li 670.784	3579783.9	5361.54	0.15%	[250]	ug/L

Sequence No.: 4
Sample ID: STD 3 150612RJS I:JL O:RJ
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 4
Date Collected: 06/12/15 1:14:56 PM
Data Type: Reprocessed on 06/15/15 10:35:44 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: STD 3 150612RJS I:JL O:RJ

Analyte	Mean Corrected Intensity	Std.Dev.	RSD	Conc.	Calib Units
Y 371.029	191152.4	1013.95	0.53%	94.08	%
Y Radial	13187.3	99.13	0.75%	100.5	%
Ag 338.289†	81421.3	181.41	0.22%	[500]	ug/L
Al 308.215†	23011.1	124.86	0.54%	[20000]	ug/L
As 188.979†	4501.3	39.48	0.88%	[1000]	ug/L
B 249.677†	30803.2	114.76	0.37%	[1000]	ug/L
Ba 233.527†	373030.4	710.79	0.19%	[1000]	ug/L
Be 313.107†	3237526.3	7010.12	0.22%	[1000]	ug/L
Ca 315.887†	649688.6	9682.27	1.49%	[50000]	ug/L
Cd 214.440†	245591.7	513.49	0.21%	[1000]	ug/L
Co 228.616†	86546.9	118.57	0.14%	[1000]	ug/L
Cr 267.716†	77566.9	609.11	0.79%	[1000]	ug/L
Cu 327.393†	177878.9	377.27	0.21%	[1000]	ug/L
Fe 273.955†	1350904.7	2292.67	0.17%	[20000]	ug/L
K 766.490†	116281.5	610.12	0.52%	[20000]	ug/L
Mg 285.213†	714101.7	11476.54	1.61%	[50000]	ug/L
Mn 257.610†	70989.9	303.61	0.43%	[1000]	ug/L
Mo 202.031†	30787.3	243.98	0.79%	[1000]	ug/L
Na 589.592†	145163.6	2576.54	1.77%	[25000]	ug/L
Ni 231.604†	43000.1	347.49	0.81%	[1000]	ug/L
P 213.617†	14473.8	158.00	1.09%	[5000]	ug/L
Pb 220.353†	5487.3	49.25	0.90%	[1000]	ug/L
Sb 206.836†	6352.2	56.77	0.89%	[1000]	ug/L
Se 196.026†	2722.1	36.30	1.33%	[1000]	ug/L
Sn 189.927†	4918.7	42.91	0.87%	[1000]	ug/L
Sr 421.552†	335696.1	6119.90	1.82%	[1000]	ug/L
Ti 337.279†	32105.7	146.46	0.46%	[1000]	ug/L
Tl 190.801†	4988.8	72.43	1.45%	[1000]	ug/L
V 292.402†	93838.1	229.65	0.24%	[1000]	ug/L
Zn 206.200†	80342.1	132.88	0.17%	[1000]	ug/L
Li 670.784	7865167.1	27046.04	0.34%	[500]	ug/L

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Ag 338.289	3	Lin Thru 0	0.0	162.4	0.00000	0.999942	
Al 308.215	3	Lin Thru 0	0.0	1.139	0.00000	0.999746	
As 188.979	3	Lin Thru 0	0.0	4.498	0.00000	0.999959	
B 249.677	3	Lin Thru 0	0.0	30.69	0.00000	0.999916	
Ba 233.527	3	Lin Thru 0	0.0	372.6	0.00000	0.999952	
Be 313.107	3	Lin Thru 0	0.0	3232	0.00000	0.999958	
Ca 315.887	3	Lin Thru 0	0.0	12.90	0.00000	0.999854	
Cd 214.440	3	Lin Thru 0	0.0	245.5	0.00000	0.999955	
Co 228.616	3	Lin Thru 0	0.0	86.57	0.00000	0.999953	
Cr 267.716	3	Lin Thru 0	0.0	77.56	0.00000	0.999961	
Cu 327.393	3	Lin Thru 0	0.0	177.3	0.00000	0.999938	
Fe 273.955	3	Lin Thru 0	0.0	67.54	0.00000	0.999959	
K 766.490	3	Lin Thru 0	0.0	5.744	0.00000	0.999659	
Mg 285.213	3	Lin Thru 0	0.0	14.19	0.00000	0.999883	
Mn 257.610	3	Lin Thru 0	0.0	70.35	0.00000	0.999792	
Mo 202.031	3	Lin Thru 0	0.0	30.74	0.00000	0.999956	
Na 589.592	3	Lin Thru 0	0.0	5.756	0.00000	0.999801	
Ni 231.604	3	Lin Thru 0	0.0	43.10	0.00000	0.999947	
P 213.617	3	Lin Thru 0	0.0	2.886	0.00000	0.999942	
Pb 220.353	3	Lin Thru 0	0.0	5.497	0.00000	0.999951	
Sb 206.836	3	Lin Thru 0	0.0	6.333	0.00000	0.999949	
Se 196.026	3	Lin Thru 0	0.0	2.718	0.00000	0.999968	

Sn 189.927	3	Lin Thru 0	0.0	4.916	0.00000	0.999959
Sr 421.552	3	Lin Thru 0	0.0	332.1	0.00000	0.999732
Ti 337.279	3	Lin Thru 0	0.0	31.81	0.00000	0.999787
Tl 190.801	3	Lin Thru 0	0.0	5.008	0.00000	0.999920
V 292.402	3	Lin Thru 0	0.0	93.61	0.00000	0.999944
Zn 206.200	3	Lin Thru 0	0.0	80.17	0.00000	0.999942
Li 670.784	3	Lin Thru 0	0.0	15450	0.00000	0.999329

Sequence No.: 5
Sample ID: ICV 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 5
Date Collected: 06/12/15 1:26:55 PM
Data Type: Reprocessed on 06/15/15 10:35:45 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: ICV 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	198221.1	97.55	%	0.412				0.42%
Y Radial	13541.2	103.2	%	0.44				0.42%
Ag 338.289†	40879.3	251.7	ug/L	1.97	251.7	ug/L	1.97	0.78%
QC value within limits for Ag 338.289 Recovery = 100.67%								
Al 308.215†	14275.1	12560	ug/L	87.68	12560	ug/L	87.68	0.70%
QC value within limits for Al 308.215 Recovery = 100.48%								
As 188.979†	2048.9	452.2	ug/L	1.21	452.2	ug/L	1.21	0.27%
QC value within limits for As 188.979 Recovery = 90.45%								
B 249.677†	15235.4	526.1	ug/L	2.54	526.1	ug/L	2.54	0.48%
QC value within limits for B 249.677 Recovery = 105.22%								
Ba 233.527†	187349.3	502.4	ug/L	2.65	502.4	ug/L	2.65	0.53%
QC value within limits for Ba 233.527 Recovery = 100.49%								
Be 313.107†	1608761.8	497.7	ug/L	2.85	497.7	ug/L	2.85	0.57%
QC value within limits for Be 313.107 Recovery = 99.54%								
Ca 315.887†	168882.4	13090	ug/L	93.26	13090	ug/L	93.26	0.71%
QC value within limits for Ca 315.887 Recovery = 104.69%								
Cd 214.440†	124828.9	508.5	ug/L	3.31	508.5	ug/L	3.31	0.65%
QC value within limits for Cd 214.440 Recovery = 101.71%								
Co 228.616†	43742.4	505.2	ug/L	2.75	505.2	ug/L	2.75	0.54%
QC value within limits for Co 228.616 Recovery = 101.05%								
Cr 267.716†	39456.0	508.6	ug/L	1.16	508.6	ug/L	1.16	0.23%
QC value within limits for Cr 267.716 Recovery = 101.73%								
Cu 327.393†	90221.9	510.0	ug/L	3.18	510.0	ug/L	3.18	0.62%
QC value within limits for Cu 327.393 Recovery = 102.00%								
Fe 273.955†	851818.9	12570	ug/L	68.01	12570	ug/L	68.01	0.54%
QC value within limits for Fe 273.955 Recovery = 100.58%								
K 766.490†	71945.5	12500	ug/L	104.35	12500	ug/L	104.35	0.83%
QC value within limits for K 766.490 Recovery = 99.98%								
Mg 285.213†	181963.9	12830	ug/L	110.07	12830	ug/L	110.07	0.86%
QC value within limits for Mg 285.213 Recovery = 102.64%								
Mn 257.610†	35702.1	507.1	ug/L	3.94	507.1	ug/L	3.94	0.78%
QC value within limits for Mn 257.610 Recovery = 101.43%								
Mo 202.031†	15306.4	498.5	ug/L	1.12	498.5	ug/L	1.12	0.22%
QC value within limits for Mo 202.031 Recovery = 99.69%								
Na 589.592†	73647.9	12780	ug/L	105.16	12780	ug/L	105.16	0.82%
QC value within limits for Na 589.592 Recovery = 102.23%								
Ni 231.604†	22407.8	519.0	ug/L	1.21	519.0	ug/L	1.21	0.23%
QC value within limits for Ni 231.604 Recovery = 103.79%								
P 213.617†	7296.7	2528	ug/L	5.19	2528	ug/L	5.19	0.21%
QC value within limits for P 213.617 Recovery = 101.14%								
Pb 220.353†	2856.2	522.4	ug/L	1.51	522.4	ug/L	1.51	0.29%
QC value within limits for Pb 220.353 Recovery = 104.49%								
Sb 206.836†	3251.4	513.4	ug/L	2.31	513.4	ug/L	2.31	0.45%
QC value within limits for Sb 206.836 Recovery = 102.68%								
Se 196.026†	1393.0	516.3	ug/L	4.57	516.3	ug/L	4.57	0.89%
QC value within limits for Se 196.026 Recovery = 103.26%								
Sn 189.927†	1276.9	259.7	ug/L	0.85	259.7	ug/L	0.85	0.33%
QC value within limits for Sn 189.927 Recovery = 103.90%								
Sr 421.552†	168575.7	507.0	ug/L	5.99	507.0	ug/L	5.99	1.18%
QC value within limits for Sr 421.552 Recovery = 101.41%								
Ti 337.279†	16192.4	508.8	ug/L	4.23	508.8	ug/L	4.23	0.83%
QC value within limits for Ti 337.279 Recovery = 101.76%								
Tl 190.801†	2588.9	517.0	ug/L	1.63	517.0	ug/L	1.63	0.32%
QC value within limits for Tl 190.801 Recovery = 103.39%								
V 292.402†	46569.4	497.5	ug/L	2.57	497.5	ug/L	2.57	0.52%
QC value within limits for V 292.402 Recovery = 99.50%								
Zn 206.200†	41321.5	521.6	ug/L	3.58	521.6	ug/L	3.58	0.69%
QC value within limits for Zn 206.200 Recovery = 104.33%								

Li 670.784	4171413.8	270.0 ug/L	0.59	270.0 ug/L	0.59	0.22%

QC value within limits for Li 670.784 Recovery = 108.01%
All analyte(s) passed QC.

Sequence No.: 6
Sample ID: ICB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 1
Date Collected: 06/12/15 1:37:16 PM
Data Type: Reprocessed on 06/15/15 10:35:46 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: ICB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	203764.6	100.3	%	0.62				0.62%
Y Radial	13522.8	103.1	%	1.76				1.70%
Ag 338.289†	-17.6	-0.108	ug/L	0.2167	-0.108	ug/L	0.2167	200.30%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	-6.2	-5.449	ug/L	2.3760	-5.449	ug/L	2.3760	43.60%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	8.4	1.874	ug/L	0.4645	1.874	ug/L	0.4645	24.79%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	53.3	1.737	ug/L	0.2490	1.737	ug/L	0.2490	14.33%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	20.4	0.054	ug/L	0.0346	0.054	ug/L	0.0346	63.43%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-11.6	-0.004	ug/L	0.0077	-0.004	ug/L	0.0077	216.01%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	37.8	2.926	ug/L	0.4883	2.926	ug/L	0.4883	16.69%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	20.3	0.083	ug/L	0.0437	0.083	ug/L	0.0437	52.67%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	25.4	0.297	ug/L	0.1260	0.297	ug/L	0.1260	42.44%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	2.3	0.030	ug/L	0.0621	0.030	ug/L	0.0621	209.05%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	355.8	2.007	ug/L	0.3076	2.007	ug/L	0.3076	15.33%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	10.0	0.144	ug/L	0.0501	0.144	ug/L	0.0501	34.71%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	87.9	15.30	ug/L	5.424	15.30	ug/L	5.424	35.45%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	10.9	0.764	ug/L	0.6707	0.764	ug/L	0.6707	87.80%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	14.5	0.206	ug/L	0.1536	0.206	ug/L	0.1536	74.54%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	21.7	0.704	ug/L	0.1130	0.704	ug/L	0.1130	16.05%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-1.1	-0.216	ug/L	7.9269	-0.216	ug/L	7.9269	>999.9%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	-6.8	-0.157	ug/L	0.1739	-0.157	ug/L	0.1739	111.05%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	-0.2	-0.057	ug/L	0.8573	-0.057	ug/L	0.8573	>999.9%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	5.4	0.988	ug/L	0.1708	0.988	ug/L	0.1708	17.29%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	19.3	3.048	ug/L	1.3894	3.048	ug/L	1.3894	45.59%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	-5.9	-2.169	ug/L	1.5180	-2.169	ug/L	1.5180	69.99%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	1.8	0.376	ug/L	0.2443	0.376	ug/L	0.2443	65.04%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	9.2	0.028	ug/L	0.1189	0.028	ug/L	0.1189	431.52%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	-8.4	-0.264	ug/L	0.3711	-0.264	ug/L	0.3711	140.64%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-3.5	-0.695	ug/L	0.3654	-0.695	ug/L	0.3654	52.57%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	2.2	0.023	ug/L	0.1010	0.023	ug/L	0.1010	432.20%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-160.0	-1.994	ug/L	0.0861	-1.994	ug/L	0.0861	4.32%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	6972.4	0.451 ug/L	0.0049	0.451 ug/L	0.0049	1.08%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 7
Sample ID: ICSA 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 6
Date Collected: 06/12/15 1:43:23 PM
Data Type: Reprocessed on 06/15/15 10:35:48 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: ICSA 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	192595.2	94.79	%	0.213				0.22%
Y Radial	13405.4	102.2	%	0.47				0.46%
Ag 338.289†	72.1	0.444	ug/L	0.2262	0.444	ug/L	0.2262	50.96%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	114220.9	100300	ug/L	167.75	100300	ug/L	167.75	0.17%
QC value within limits for Al 308.215 Recovery = 100.30%								
As 188.979†	17.4	4.187	ug/L	0.9747	4.187	ug/L	0.9747	23.28%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	-7841.4	-14.65	ug/L	2.213	-14.65	ug/L	2.213	15.11%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	1657.2	-1.139	ug/L	0.0993	-1.139	ug/L	0.0993	8.72%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	275.6	0.085	ug/L	0.0039	0.085	ug/L	0.0039	4.63%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	1339315.2	103800	ug/L	143.50	103800	ug/L	143.50	0.14%
QC value within limits for Ca 315.887 Recovery = 103.81%								
Cd 214.440†	887.6	-0.289	ug/L	0.0610	-0.289	ug/L	0.0610	21.09%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	122.0	-1.494	ug/L	0.1772	-1.494	ug/L	0.1772	11.86%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	80.6	0.132	ug/L	0.1067	0.132	ug/L	0.1067	81.07%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	-126.0	0.706	ug/L	0.0899	0.706	ug/L	0.0899	12.73%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	6685686.8	98980	ug/L	994.25	98980	ug/L	994.25	1.00%
QC value within limits for Fe 273.955 Recovery = 98.98%								
K 766.490†	174.7	-66.61	ug/L	8.198	-66.61	ug/L	8.198	12.31%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	1418866.1	100000	ug/L	167.24	100000	ug/L	167.24	0.17%
QC value within limits for Mg 285.213 Recovery = 100.05%								
Mn 257.610†	229.9	-0.054	ug/L	0.0995	-0.054	ug/L	0.0995	182.73%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	-67.5	-0.766	ug/L	0.3582	-0.766	ug/L	0.3582	46.74%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	164.0	16.24	ug/L	1.644	16.24	ug/L	1.644	10.12%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	-94.6	0.249	ug/L	0.1407	0.249	ug/L	0.1407	56.46%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	-107.2	-37.16	ug/L	2.779	-37.16	ug/L	2.779	7.48%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	-61.0	-1.144	ug/L	0.4191	-1.144	ug/L	0.4191	36.64%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	-3.4	-0.536	ug/L	2.0189	-0.536	ug/L	2.0189	376.45%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	-116.5	-3.945	ug/L	3.2533	-3.945	ug/L	3.2533	82.47%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	10.2	2.078	ug/L	0.4659	2.078	ug/L	0.4659	22.42%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	242.1	-3.333	ug/L	0.0960	-3.333	ug/L	0.0960	2.88%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	40.2	-0.464	ug/L	0.5299	-0.464	ug/L	0.5299	114.08%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-32.6	-6.502	ug/L	0.7702	-6.502	ug/L	0.7702	11.85%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	39.6	0.424	ug/L	0.1746	0.424	ug/L	0.1746	41.23%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	136.6	0.999	ug/L	0.2773	0.999	ug/L	0.2773	27.74%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	12357.5	0.800 ug/L	0.0075	0.800 ug/L	0.0075	0.94%

All analyte(s) passed QC.

Sequence No.: 8
Sample ID: ICSAB 150612RJS I:JL O:RJ
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 7
Date Collected: 06/12/15 1:50:38 PM
Data Type: Reprocessed on 06/15/15 10:35:49 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: ICSAB 150612RJS I:JL O:RJ

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	193304.1	95.13	%	0.108				0.11%
Y Radial	13514.7	103.0	%	0.33				0.32%
Ag 338.289†	82900.8	510.4	ug/L	0.43	510.4	ug/L	0.43	0.08%
QC value within limits for Ag 338.289 Recovery = 102.07%								
Al 308.215†	113679.6	99840	ug/L	892.65	99840	ug/L	892.65	0.89%
QC value within limits for Al 308.215 Recovery = 99.84%								
As 188.979†	1025.8	226.6	ug/L	1.41	226.6	ug/L	1.41	0.62%
QC value within limits for As 188.979 Recovery = 90.65%								
B 249.677†	-7810.8	-15.98	ug/L	0.719	-15.98	ug/L	0.719	4.50%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	93542.8	245.7	ug/L	0.27	245.7	ug/L	0.27	0.11%
QC value within limits for Ba 233.527 Recovery = 98.28%								
Be 313.107†	779442.1	241.1	ug/L	0.66	241.1	ug/L	0.66	0.27%
QC value within limits for Be 313.107 Recovery = 96.45%								
Ca 315.887†	1334439.2	103400	ug/L	995.46	103400	ug/L	995.46	0.96%
QC value within limits for Ca 315.887 Recovery = 103.43%								
Cd 214.440†	120649.0	487.8	ug/L	0.03	487.8	ug/L	0.03	0.01%
QC value within limits for Cd 214.440 Recovery = 97.56%								
Co 228.616†	21754.7	249.1	ug/L	0.50	249.1	ug/L	0.50	0.20%
QC value within limits for Co 228.616 Recovery = 99.63%								
Cr 267.716†	18838.5	242.0	ug/L	0.58	242.0	ug/L	0.58	0.24%
QC value within limits for Cr 267.716 Recovery = 96.81%								
Cu 327.393†	44317.9	251.2	ug/L	0.52	251.2	ug/L	0.52	0.21%
QC value within limits for Cu 327.393 Recovery = 100.47%								
Fe 273.955†	6635958.9	98220	ug/L	478.84	98220	ug/L	478.84	0.49%
QC value within limits for Fe 273.955 Recovery = 98.22%								
K 766.490†	112.9	-78.19	ug/L	1.867	-78.19	ug/L	1.867	2.39%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	1424460.8	100400	ug/L	940.93	100400	ug/L	940.93	0.94%
QC value within limits for Mg 285.213 Recovery = 100.44%								
Mn 257.610†	17668.8	247.8	ug/L	0.89	247.8	ug/L	0.89	0.36%
QC value within limits for Mn 257.610 Recovery = 99.13%								
Mo 202.031†	7311.9	239.5	ug/L	0.84	239.5	ug/L	0.84	0.35%
QC value within limits for Mo 202.031 Recovery = 95.79%								
Na 589.592†	80.9	0.906	ug/L	3.7326	0.906	ug/L	3.7326	411.77%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	20824.2	485.0	ug/L	1.16	485.0	ug/L	1.16	0.24%
QC value within limits for Ni 231.604 Recovery = 97.00%								
P 213.617†	-188.6	-65.34	ug/L	2.060	-65.34	ug/L	2.060	3.15%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	2629.3	489.2	ug/L	0.23	489.2	ug/L	0.23	0.05%
QC value within limits for Pb 220.353 Recovery = 97.84%								
Sb 206.836†	1605.5	253.5	ug/L	2.54	253.5	ug/L	2.54	1.00%
QC value within limits for Sb 206.836 Recovery = 101.40%								
Se 196.026†	547.9	239.6	ug/L	2.83	239.6	ug/L	2.83	1.18%
QC value within limits for Se 196.026 Recovery = 95.83%								
Sn 189.927†	12.2	2.485	ug/L	0.8254	2.485	ug/L	0.8254	33.21%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	227.6	-3.362	ug/L	0.0533	-3.362	ug/L	0.0533	1.59%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	53.5	-0.002	ug/L	0.8743	-0.002	ug/L	0.8743	>999.9%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	1158.9	231.4	ug/L	2.91	231.4	ug/L	2.91	1.26%
QC value within limits for Tl 190.801 Recovery = 92.56%								
V 292.402†	24361.0	260.2	ug/L	0.78	260.2	ug/L	0.78	0.30%
QC value within limits for V 292.402 Recovery = 104.10%								
Zn 206.200†	38791.0	485.9	ug/L	0.65	485.9	ug/L	0.65	0.13%
QC value within limits for Zn 206.200 Recovery = 97.18%								

Li 670.784	10500.4	0.680 ug/L	0.0084	0.680 ug/L	0.0084	1.23%

All analyte(s) passed QC.

Sequence No.: 9
Sample ID: CCV1 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 3
Date Collected: 06/12/15 1:57:55 PM
Data Type: Reprocessed on 06/15/15 10:35:50 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV1 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	196666.0	96.79	%	0.901				0.93%
Y Radial	13673.1	104.2	%	1.07				1.03%
Ag 338.289†	40400.1	248.7	ug/L	0.31	248.7	ug/L	0.31	0.12%
QC value within limits for Ag 338.289 Recovery = 99.49%								
Al 308.215†	11261.2	9912	ug/L	21.87	9912	ug/L	21.87	0.22%
QC value within limits for Al 308.215 Recovery = 99.12%								
As 188.979†	2244.3	494.1	ug/L	5.03	494.1	ug/L	5.03	1.02%
QC value within limits for As 188.979 Recovery = 98.83%								
B 249.677†	15066.1	514.0	ug/L	1.37	514.0	ug/L	1.37	0.27%
QC value within limits for B 249.677 Recovery = 102.80%								
Ba 233.527†	185829.9	498.5	ug/L	0.70	498.5	ug/L	0.70	0.14%
QC value within limits for Ba 233.527 Recovery = 99.70%								
Be 313.107†	1603170.7	496.0	ug/L	1.13	496.0	ug/L	1.13	0.23%
QC value within limits for Be 313.107 Recovery = 99.19%								
Ca 315.887†	326256.5	25280	ug/L	58.88	25280	ug/L	58.88	0.23%
QC value within limits for Ca 315.887 Recovery = 101.14%								
Cd 214.440†	123109.2	501.8	ug/L	1.06	501.8	ug/L	1.06	0.21%
QC value within limits for Cd 214.440 Recovery = 100.36%								
Co 228.616†	43345.9	500.7	ug/L	0.26	500.7	ug/L	0.26	0.05%
QC value within limits for Co 228.616 Recovery = 100.14%								
Cr 267.716†	38750.0	499.5	ug/L	5.07	499.5	ug/L	5.07	1.02%
QC value within limits for Cr 267.716 Recovery = 99.89%								
Cu 327.393†	87914.3	496.9	ug/L	0.34	496.9	ug/L	0.34	0.07%
QC value within limits for Cu 327.393 Recovery = 99.37%								
Fe 273.955†	679215.5	10020	ug/L	14.34	10020	ug/L	14.34	0.14%
QC value within limits for Fe 273.955 Recovery = 100.17%								
K 766.490†	57450.9	9976	ug/L	12.01	9976	ug/L	12.01	0.12%
QC value within limits for K 766.490 Recovery = 99.76%								
Mg 285.213†	361673.8	25490	ug/L	32.23	25490	ug/L	32.23	0.13%
QC value within limits for Mg 285.213 Recovery = 101.96%								
Mn 257.610†	35323.3	501.2	ug/L	1.13	501.2	ug/L	1.13	0.23%
QC value within limits for Mn 257.610 Recovery = 100.24%								
Mo 202.031†	15312.5	498.6	ug/L	5.42	498.6	ug/L	5.42	1.09%
QC value within limits for Mo 202.031 Recovery = 99.72%								
Na 589.592†	75142.8	13040	ug/L	16.44	13040	ug/L	16.44	0.13%
QC value within limits for Na 589.592 Recovery = 104.33%								
Ni 231.604†	21781.0	504.5	ug/L	5.56	504.5	ug/L	5.56	1.10%
QC value within limits for Ni 231.604 Recovery = 100.89%								
P 213.617†	7141.6	2475	ug/L	25.48	2475	ug/L	25.48	1.03%
QC value within limits for P 213.617 Recovery = 98.99%								
Pb 220.353†	2726.3	499.0	ug/L	3.62	499.0	ug/L	3.62	0.73%
QC value within limits for Pb 220.353 Recovery = 99.80%								
Sb 206.836†	3127.5	493.8	ug/L	7.05	493.8	ug/L	7.05	1.43%
QC value within limits for Sb 206.836 Recovery = 98.77%								
Se 196.026†	1361.0	503.1	ug/L	2.25	503.1	ug/L	2.25	0.45%
QC value within limits for Se 196.026 Recovery = 100.63%								
Sn 189.927†	2444.1	497.2	ug/L	5.37	497.2	ug/L	5.37	1.08%
QC value within limits for Sn 189.927 Recovery = 99.44%								
Sr 421.552†	172227.0	517.6	ug/L	7.39	517.6	ug/L	7.39	1.43%
QC value within limits for Sr 421.552 Recovery = 103.52%								
Ti 337.279†	15853.9	498.0	ug/L	0.55	498.0	ug/L	0.55	0.11%
QC value within limits for Ti 337.279 Recovery = 99.60%								
Tl 190.801†	2470.5	493.3	ug/L	5.09	493.3	ug/L	5.09	1.03%
QC value within limits for Tl 190.801 Recovery = 98.66%								
V 292.402†	46009.2	491.5	ug/L	0.94	491.5	ug/L	0.94	0.19%
QC value within limits for V 292.402 Recovery = 98.30%								
Zn 206.200†	40133.1	507.3	ug/L	1.15	507.3	ug/L	1.15	0.23%
QC value within limits for Zn 206.200 Recovery = 101.47%								

Li 670.784	3699085.7	239.5 ug/L	1.69	239.5 ug/L	1.69	0.70%

QC value within limits for Li 670.784 Recovery = 95.78%
All analyte(s) passed QC.

Sequence No.: 10
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 1
Date Collected: 06/12/15 2:08:21 PM
Data Type: Reprocessed on 06/15/15 10:35:52 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	201570.0	99.20	%	0.306				0.31%
Y Radial	13406.6	102.2	%	0.72				0.70%
Ag 338.289†	7.1	0.044	ug/L	0.3282	0.044	ug/L	0.3282	750.40%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	1.0	0.867	ug/L	3.1476	0.867	ug/L	3.1476	363.18%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	0.7	0.154	ug/L	0.4866	0.154	ug/L	0.4866	316.21%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	49.6	1.625	ug/L	0.3203	1.625	ug/L	0.3203	19.72%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	11.1	0.029	ug/L	0.0250	0.029	ug/L	0.0250	85.05%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	19.2	0.006	ug/L	0.0040	0.006	ug/L	0.0040	66.62%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	4.8	0.368	ug/L	0.4727	0.368	ug/L	0.4727	128.45%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	5.0	0.020	ug/L	0.0452	0.020	ug/L	0.0452	221.61%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	19.2	0.225	ug/L	0.0730	0.225	ug/L	0.0730	32.51%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-8.8	-0.114	ug/L	0.0288	-0.114	ug/L	0.0288	25.22%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	514.9	2.904	ug/L	0.2729	2.904	ug/L	0.2729	9.40%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	246.5	3.645	ug/L	0.1853	3.645	ug/L	0.1853	5.09%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	76.4	13.31	ug/L	6.004	13.31	ug/L	6.004	45.12%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	33.5	2.363	ug/L	0.1032	2.363	ug/L	0.1032	4.37%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	6.3	0.090	ug/L	0.1210	0.090	ug/L	0.1210	134.27%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	21.9	0.712	ug/L	0.3509	0.712	ug/L	0.3509	49.24%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-86.2	-14.99	ug/L	7.261	-14.99	ug/L	7.261	48.44%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	-0.6	-0.014	ug/L	0.1787	-0.014	ug/L	0.1787	>999.9%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	-1.8	-0.633	ug/L	0.6876	-0.633	ug/L	0.6876	108.58%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	10.6	1.933	ug/L	0.7715	1.933	ug/L	0.7715	39.91%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	16.2	2.558	ug/L	0.2095	2.558	ug/L	0.2095	8.19%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	0.1	0.052	ug/L	1.5842	0.052	ug/L	1.5842	>999.9%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	4.5	0.920	ug/L	0.2370	0.920	ug/L	0.2370	25.76%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	9.2	0.028	ug/L	0.0583	0.028	ug/L	0.0583	211.00%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	-8.9	-0.280	ug/L	0.4424	-0.280	ug/L	0.4424	157.95%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-5.7	-1.134	ug/L	1.1790	-1.134	ug/L	1.1790	103.98%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	3.7	0.040	ug/L	0.1010	0.040	ug/L	0.1010	254.92%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-77.5	-0.968	ug/L	0.1197	-0.968	ug/L	0.1197	12.37%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	10445.0	0.676 ug/L	0.0036	0.676 ug/L	0.0036	0.53%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 16
Sample ID: CCV2 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 8
Date Collected: 06/12/15 3:02:20 PM
Data Type: Reprocessed on 06/15/15 10:35:58 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV2 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	197040.2	96.97	%	0.764				0.79%
Y Radial	13458.3	102.6	%	0.71				0.69%
Ag 338.289†	30443.7	187.4	ug/L	0.46	187.4	ug/L	0.46	0.24%
QC value within limits for Ag 338.289 Recovery = 99.96%								
Al 308.215†	8419.1	7411	ug/L	12.24	7411	ug/L	12.24	0.17%
QC value within limits for Al 308.215 Recovery = 98.81%								
As 188.979†	1665.0	366.6	ug/L	1.73	366.6	ug/L	1.73	0.47%
QC value within limits for As 188.979 Recovery = 97.76%								
B 249.677†	11551.7	393.7	ug/L	3.38	393.7	ug/L	3.38	0.86%
QC value within limits for B 249.677 Recovery = 104.99%								
Ba 233.527†	139695.7	374.7	ug/L	1.72	374.7	ug/L	1.72	0.46%
QC value within limits for Ba 233.527 Recovery = 99.93%								
Be 313.107†	1191780.5	368.7	ug/L	1.61	368.7	ug/L	1.61	0.44%
QC value within limits for Be 313.107 Recovery = 98.32%								
Ca 315.887†	239380.1	18550	ug/L	46.58	18550	ug/L	46.58	0.25%
QC value within limits for Ca 315.887 Recovery = 98.94%								
Cd 214.440†	92822.9	378.3	ug/L	1.55	378.3	ug/L	1.55	0.41%
QC value within limits for Cd 214.440 Recovery = 100.89%								
Co 228.616†	34296.5	396.2	ug/L	2.16	396.2	ug/L	2.16	0.55%
QC value within limits for Co 228.616 Recovery = 105.64%								
Cr 267.716†	28864.3	372.0	ug/L	2.44	372.0	ug/L	2.44	0.66%
QC value within limits for Cr 267.716 Recovery = 99.21%								
Cu 327.393†	65996.1	373.0	ug/L	1.65	373.0	ug/L	1.65	0.44%
QC value within limits for Cu 327.393 Recovery = 99.46%								
Fe 273.955†	511606.2	7545	ug/L	32.76	7545	ug/L	32.76	0.43%
QC value within limits for Fe 273.955 Recovery = 100.60%								
K 766.490†	42763.6	7425	ug/L	28.72	7425	ug/L	28.72	0.39%
QC value within limits for K 766.490 Recovery = 99.00%								
Mg 285.213†	265189.0	18690	ug/L	68.04	18690	ug/L	68.04	0.36%
QC value within limits for Mg 285.213 Recovery = 99.68%								
Mn 257.610†	26690.8	378.7	ug/L	1.33	378.7	ug/L	1.33	0.35%
QC value within limits for Mn 257.610 Recovery = 101.00%								
Mo 202.031†	11441.2	372.5	ug/L	2.70	372.5	ug/L	2.70	0.73%
QC value within limits for Mo 202.031 Recovery = 99.34%								
Na 589.592†	57060.6	9903	ug/L	43.74	9903	ug/L	43.74	0.44%
QC value within limits for Na 589.592 Recovery = 105.64%								
Ni 231.604†	16316.1	377.9	ug/L	2.31	377.9	ug/L	2.31	0.61%
QC value within limits for Ni 231.604 Recovery = 100.76%								
P 213.617†	5293.7	1834	ug/L	11.61	1834	ug/L	11.61	0.63%
QC value within limits for P 213.617 Recovery = 97.83%								
Pb 220.353†	2010.8	368.0	ug/L	2.67	368.0	ug/L	2.67	0.73%
QC value within limits for Pb 220.353 Recovery = 98.14%								
Sb 206.836†	2310.7	364.8	ug/L	4.79	364.8	ug/L	4.79	1.31%
QC value within limits for Sb 206.836 Recovery = 97.29%								
Se 196.026†	1012.5	374.3	ug/L	3.08	374.3	ug/L	3.08	0.82%
QC value within limits for Se 196.026 Recovery = 99.82%								
Sn 189.927†	1828.4	371.9	ug/L	2.60	371.9	ug/L	2.60	0.70%
QC value within limits for Sn 189.927 Recovery = 99.18%								
Sr 421.552†	135390.4	406.9	ug/L	1.52	406.9	ug/L	1.52	0.37%
QC value within limits for Sr 421.552 Recovery = 108.51%								
Ti 337.279†	11929.0	374.7	ug/L	2.41	374.7	ug/L	2.41	0.64%
QC value within limits for Ti 337.279 Recovery = 99.93%								
Tl 190.801†	1791.6	357.8	ug/L	3.79	357.8	ug/L	3.79	1.06%
QC value within limits for Tl 190.801 Recovery = 95.40%								
V 292.402†	34297.6	366.4	ug/L	1.61	366.4	ug/L	1.61	0.44%
QC value within limits for V 292.402 Recovery = 97.70%								
Zn 206.200†	31364.9	396.2	ug/L	2.72	396.2	ug/L	2.72	0.69%
QC value within limits for Zn 206.200 Recovery = 105.66%								

Li 670.784	2746954.8	177.8 ug/L	1.14	177.8 ug/L	1.14	0.64%

QC value within limits for Li 670.784 Recovery = 94.84%
All analyte(s) passed QC.

Sequence No.: 17
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 5
Date Collected: 06/12/15 3:08:06 PM
Data Type: Reprocessed on 06/15/15 10:36:00 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	202559.0	99.69	%	1.040				1.04%
Y Radial	13530.5	103.1	%	2.38				2.31%
Ag 338.289†	62.8	0.387	ug/L	0.2586	0.387	ug/L	0.2586	66.87%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	-10.9	-9.555	ug/L	3.8011	-9.555	ug/L	3.8011	39.78%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	4.6	1.027	ug/L	1.2153	1.027	ug/L	1.2153	118.36%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	69.5	2.277	ug/L	0.3500	2.277	ug/L	0.3500	15.37%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	35.0	0.094	ug/L	0.0752	0.094	ug/L	0.0752	80.37%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-38.6	-0.012	ug/L	0.0014	-0.012	ug/L	0.0014	11.92%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	214.6	16.64	ug/L	0.715	16.64	ug/L	0.715	4.30%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	-3.2	-0.013	ug/L	0.0209	-0.013	ug/L	0.0209	160.00%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	23.2	0.272	ug/L	0.0460	0.272	ug/L	0.0460	16.94%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-5.8	-0.075	ug/L	0.1051	-0.075	ug/L	0.1051	140.61%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	440.9	2.487	ug/L	0.4427	2.487	ug/L	0.4427	17.80%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	290.2	4.285	ug/L	0.1823	4.285	ug/L	0.1823	4.25%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	67.4	11.74	ug/L	7.958	11.74	ug/L	7.958	67.79%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	19.6	1.386	ug/L	0.1971	1.386	ug/L	0.1971	14.22%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	8.4	0.119	ug/L	0.1413	0.119	ug/L	0.1413	118.66%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	25.4	0.825	ug/L	0.2616	0.825	ug/L	0.2616	31.70%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-79.6	-13.84	ug/L	3.000	-13.84	ug/L	3.000	21.67%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	-0.9	-0.019	ug/L	0.1615	-0.019	ug/L	0.1615	833.76%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	-2.5	-0.877	ug/L	2.0817	-0.877	ug/L	2.0817	237.41%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	1.8	0.335	ug/L	0.3554	0.335	ug/L	0.3554	106.14%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	19.0	2.994	ug/L	0.6523	2.994	ug/L	0.6523	21.79%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	0.3	0.118	ug/L	1.5842	0.118	ug/L	1.5842	>999.9%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	1.8	0.372	ug/L	0.3129	0.372	ug/L	0.3129	84.08%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	1.3	0.003	ug/L	0.0910	0.003	ug/L	0.0910	>999.9%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	-11.1	-0.350	ug/L	0.0882	-0.350	ug/L	0.0882	25.19%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-7.1	-1.410	ug/L	0.5719	-1.410	ug/L	0.5719	40.55%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	13.7	0.146	ug/L	0.1282	0.146	ug/L	0.1282	87.77%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-99.3	-1.239	ug/L	0.1839	-1.239	ug/L	0.1839	14.84%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	2895.2	0.187 ug/L	0.0035	0.187 ug/L	0.0035	1.87%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 26
Sample ID: CCV1 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 3
Date Collected: 06/12/15 4:29:03 PM
Data Type: Reprocessed on 06/15/15 10:36:10 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV1 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	195472.7	96.20	%	0.565				0.59%
Y Radial	13307.5	101.4	%	0.47				0.47%
Ag 338.289†	40677.7	250.4	ug/L	0.91	250.4	ug/L	0.91	0.36%
QC value within limits for Ag 338.289 Recovery = 100.17%								
Al 308.215†	11355.6	9995	ug/L	96.83	9995	ug/L	96.83	0.97%
QC value within limits for Al 308.215 Recovery = 99.95%								
As 188.979†	2228.8	490.7	ug/L	2.35	490.7	ug/L	2.35	0.48%
QC value within limits for As 188.979 Recovery = 98.13%								
B 249.677†	15021.5	512.6	ug/L	1.99	512.6	ug/L	1.99	0.39%
QC value within limits for B 249.677 Recovery = 102.51%								
Ba 233.527†	185633.1	498.0	ug/L	1.64	498.0	ug/L	1.64	0.33%
QC value within limits for Ba 233.527 Recovery = 99.59%								
Be 313.107†	1593809.5	493.1	ug/L	2.02	493.1	ug/L	2.02	0.41%
QC value within limits for Be 313.107 Recovery = 98.62%								
Ca 315.887†	329578.5	25540	ug/L	232.97	25540	ug/L	232.97	0.91%
QC value within limits for Ca 315.887 Recovery = 102.17%								
Cd 214.440†	123226.8	502.3	ug/L	1.96	502.3	ug/L	1.96	0.39%
QC value within limits for Cd 214.440 Recovery = 100.45%								
Co 228.616†	43329.2	500.5	ug/L	2.01	500.5	ug/L	2.01	0.40%
QC value within limits for Co 228.616 Recovery = 100.10%								
Cr 267.716†	38830.8	500.5	ug/L	1.65	500.5	ug/L	1.65	0.33%
QC value within limits for Cr 267.716 Recovery = 100.10%								
Cu 327.393†	88297.7	499.0	ug/L	2.22	499.0	ug/L	2.22	0.44%
QC value within limits for Cu 327.393 Recovery = 99.81%								
Fe 273.955†	679912.3	10030	ug/L	36.61	10030	ug/L	36.61	0.37%
QC value within limits for Fe 273.955 Recovery = 100.27%								
K 766.490†	58136.3	10090	ug/L	87.44	10090	ug/L	87.44	0.87%
QC value within limits for K 766.490 Recovery = 100.95%								
Mg 285.213†	364604.9	25700	ug/L	251.38	25700	ug/L	251.38	0.98%
QC value within limits for Mg 285.213 Recovery = 102.78%								
Mn 257.610†	35690.4	506.4	ug/L	4.69	506.4	ug/L	4.69	0.93%
QC value within limits for Mn 257.610 Recovery = 101.28%								
Mo 202.031†	15345.1	499.7	ug/L	1.66	499.7	ug/L	1.66	0.33%
QC value within limits for Mo 202.031 Recovery = 99.93%								
Na 589.592†	76164.6	13220	ug/L	109.79	13220	ug/L	109.79	0.83%
QC value within limits for Na 589.592 Recovery = 105.75%								
Ni 231.604†	21810.4	505.2	ug/L	2.14	505.2	ug/L	2.14	0.42%
QC value within limits for Ni 231.604 Recovery = 101.03%								
P 213.617†	7116.3	2466	ug/L	9.09	2466	ug/L	9.09	0.37%
QC value within limits for P 213.617 Recovery = 98.64%								
Pb 220.353†	2681.7	490.9	ug/L	2.66	490.9	ug/L	2.66	0.54%
QC value within limits for Pb 220.353 Recovery = 98.19%								
Sb 206.836†	3162.3	499.3	ug/L	4.08	499.3	ug/L	4.08	0.82%
QC value within limits for Sb 206.836 Recovery = 99.86%								
Se 196.026†	1356.9	501.6	ug/L	1.64	501.6	ug/L	1.64	0.33%
QC value within limits for Se 196.026 Recovery = 100.33%								
Sn 189.927†	2437.5	495.8	ug/L	0.62	495.8	ug/L	0.62	0.13%
QC value within limits for Sn 189.927 Recovery = 99.17%								
Sr 421.552†	176540.3	530.6	ug/L	4.05	530.6	ug/L	4.05	0.76%
QC value within limits for Sr 421.552 Recovery = 106.11%								
Ti 337.279†	16053.6	504.3	ug/L	5.12	504.3	ug/L	5.12	1.01%
QC value within limits for Ti 337.279 Recovery = 100.86%								
Tl 190.801†	2367.5	472.8	ug/L	3.03	472.8	ug/L	3.03	0.64%
QC value within limits for Tl 190.801 Recovery = 94.55%								
V 292.402†	45557.2	486.7	ug/L	1.85	486.7	ug/L	1.85	0.38%
QC value within limits for V 292.402 Recovery = 97.34%								
Zn 206.200†	40063.6	506.5	ug/L	2.00	506.5	ug/L	2.00	0.40%
QC value within limits for Zn 206.200 Recovery = 101.30%								

Li 670.784	3762788.7	243.6 ug/L	1.85	243.6 ug/L	1.85	0.76%

QC value within limits for Li 670.784 Recovery = 97.43%
All analyte(s) passed QC.

Sequence No.: 27
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 1
Date Collected: 06/12/15 4:34:51 PM
Data Type: Reprocessed on 06/15/15 10:36:11 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	200876.4	98.86	%	0.654				0.66%
Y Radial	13580.3	103.5	%	0.25				0.24%
Ag 338.289†	24.9	0.153	ug/L	0.2154	0.153	ug/L	0.2154	140.71%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	-9.5	-8.420	ug/L	5.6971	-8.420	ug/L	5.6971	67.66%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	3.2	0.699	ug/L	0.6915	0.699	ug/L	0.6915	98.93%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	85.1	2.781	ug/L	0.4943	2.781	ug/L	0.4943	17.77%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	26.0	0.069	ug/L	0.0079	0.069	ug/L	0.0079	11.42%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-15.9	-0.005	ug/L	0.0046	-0.005	ug/L	0.0046	93.89%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	30.7	2.379	ug/L	0.3291	2.379	ug/L	0.3291	13.83%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	-8.9	-0.036	ug/L	0.0441	-0.036	ug/L	0.0441	122.62%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	22.9	0.269	ug/L	0.1578	0.269	ug/L	0.1578	58.71%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-13.5	-0.174	ug/L	0.0764	-0.174	ug/L	0.0764	43.83%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	398.5	2.248	ug/L	0.0797	2.248	ug/L	0.0797	3.54%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	141.6	2.098	ug/L	0.1372	2.098	ug/L	0.1372	6.54%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	7.5	1.311	ug/L	6.0366	1.311	ug/L	6.0366	460.56%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	7.8	0.554	ug/L	0.5614	0.554	ug/L	0.5614	101.34%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	13.6	0.194	ug/L	0.0927	0.194	ug/L	0.0927	47.89%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	37.7	1.226	ug/L	0.2058	1.226	ug/L	0.2058	16.79%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-92.0	-16.00	ug/L	6.096	-16.00	ug/L	6.096	38.11%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	-14.7	-0.341	ug/L	0.0713	-0.341	ug/L	0.0713	20.91%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	-3.8	-1.316	ug/L	0.9832	-1.316	ug/L	0.9832	74.71%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	8.7	1.582	ug/L	0.5658	1.582	ug/L	0.5658	35.77%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	25.7	4.052	ug/L	0.9686	4.052	ug/L	0.9686	23.90%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	1.6	0.588	ug/L	1.7081	0.588	ug/L	1.7081	290.71%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	7.3	1.479	ug/L	0.1697	1.479	ug/L	0.1697	11.47%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	2.1	0.006	ug/L	0.0305	0.006	ug/L	0.0305	496.57%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	-5.4	-0.170	ug/L	0.4374	-0.170	ug/L	0.4374	256.61%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	6.3	1.252	ug/L	0.7870	1.252	ug/L	0.7870	62.87%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	-1.0	-0.010	ug/L	0.0185	-0.010	ug/L	0.0185	177.43%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-147.0	-1.835	ug/L	0.3031	-1.835	ug/L	0.3031	16.52%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	3635.6	0.235 ug/L	0.0045	0.235 ug/L	0.0045	1.92%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 41
Sample ID: CCV2 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 8
Date Collected: 06/12/15 6:09:43 PM
Data Type: Reprocessed on 06/15/15 10:36:26 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV2 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	198590.9	97.74	%	0.317				0.32%
Y Radial	13579.7	103.5	%	0.98				0.94%
Ag 338.289†	30458.4	187.5	ug/L	0.41	187.5	ug/L	0.41	0.22%
QC value within limits for Ag 338.289 Recovery = 100.01%								
Al 308.215†	8500.8	7483	ug/L	25.28	7483	ug/L	25.28	0.34%
QC value within limits for Al 308.215 Recovery = 99.77%								
As 188.979†	1671.2	368.0	ug/L	1.84	368.0	ug/L	1.84	0.50%
QC value within limits for As 188.979 Recovery = 98.13%								
B 249.677†	11641.2	396.6	ug/L	2.10	396.6	ug/L	2.10	0.53%
QC value within limits for B 249.677 Recovery = 105.77%								
Ba 233.527†	140168.8	376.0	ug/L	0.85	376.0	ug/L	0.85	0.23%
QC value within limits for Ba 233.527 Recovery = 100.27%								
Be 313.107†	1203367.1	372.3	ug/L	0.57	372.3	ug/L	0.57	0.15%
QC value within limits for Be 313.107 Recovery = 99.28%								
Ca 315.887†	239991.8	18600	ug/L	226.62	18600	ug/L	226.62	1.22%
QC value within limits for Ca 315.887 Recovery = 99.20%								
Cd 214.440†	92715.2	377.9	ug/L	0.43	377.9	ug/L	0.43	0.11%
QC value within limits for Cd 214.440 Recovery = 100.77%								
Co 228.616†	34295.7	396.1	ug/L	1.64	396.1	ug/L	1.64	0.41%
QC value within limits for Co 228.616 Recovery = 105.64%								
Cr 267.716†	28740.8	370.5	ug/L	1.67	370.5	ug/L	1.67	0.45%
QC value within limits for Cr 267.716 Recovery = 98.79%								
Cu 327.393†	65815.6	372.0	ug/L	1.14	372.0	ug/L	1.14	0.31%
QC value within limits for Cu 327.393 Recovery = 99.19%								
Fe 273.955†	511853.3	7548	ug/L	13.42	7548	ug/L	13.42	0.18%
QC value within limits for Fe 273.955 Recovery = 100.64%								
K 766.490†	42173.8	7323	ug/L	73.33	7323	ug/L	73.33	1.00%
QC value within limits for K 766.490 Recovery = 97.64%								
Mg 285.213†	266301.7	18770	ug/L	229.12	18770	ug/L	229.12	1.22%
QC value within limits for Mg 285.213 Recovery = 100.10%								
Mn 257.610†	26619.3	377.7	ug/L	1.28	377.7	ug/L	1.28	0.34%
QC value within limits for Mn 257.610 Recovery = 100.72%								
Mo 202.031†	11391.3	370.9	ug/L	1.99	370.9	ug/L	1.99	0.54%
QC value within limits for Mo 202.031 Recovery = 98.91%								
Na 589.592†	52873.4	9176	ug/L	95.74	9176	ug/L	95.74	1.04%
QC value within limits for Na 589.592 Recovery = 97.88%								
Ni 231.604†	16278.2	377.0	ug/L	1.13	377.0	ug/L	1.13	0.30%
QC value within limits for Ni 231.604 Recovery = 100.53%								
P 213.617†	5321.7	1844	ug/L	7.15	1844	ug/L	7.15	0.39%
QC value within limits for P 213.617 Recovery = 98.35%								
Pb 220.353†	2069.3	378.7	ug/L	1.50	378.7	ug/L	1.50	0.40%
QC value within limits for Pb 220.353 Recovery = 100.99%								
Sb 206.836†	2386.5	376.8	ug/L	2.80	376.8	ug/L	2.80	0.74%
QC value within limits for Sb 206.836 Recovery = 100.48%								
Se 196.026†	1010.5	373.6	ug/L	3.45	373.6	ug/L	3.45	0.92%
QC value within limits for Se 196.026 Recovery = 99.63%								
Sn 189.927†	1828.1	371.9	ug/L	1.35	371.9	ug/L	1.35	0.36%
QC value within limits for Sn 189.927 Recovery = 99.17%								
Sr 421.552†	122559.2	368.3	ug/L	4.53	368.3	ug/L	4.53	1.23%
QC value within limits for Sr 421.552 Recovery = 98.21%								
Ti 337.279†	11987.2	376.6	ug/L	1.98	376.6	ug/L	1.98	0.52%
QC value within limits for Ti 337.279 Recovery = 100.42%								
Tl 190.801†	1899.2	379.2	ug/L	1.84	379.2	ug/L	1.84	0.48%
QC value within limits for Tl 190.801 Recovery = 101.13%								
V 292.402†	34938.0	373.2	ug/L	0.39	373.2	ug/L	0.39	0.10%
QC value within limits for V 292.402 Recovery = 99.53%								
Zn 206.200†	30975.9	391.3	ug/L	1.89	391.3	ug/L	1.89	0.48%
QC value within limits for Zn 206.200 Recovery = 104.36%								

Li 670.784	2661627.0	172.3 ug/L	0.46	172.3 ug/L	0.46	0.27%

QC value within limits for Li 670.784 Recovery = 91.89%
All analyte(s) passed QC.

Sequence No.: 42
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 5
Date Collected: 06/12/15 6:15:30 PM
Data Type: Reprocessed on 06/15/15 10:36:27 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	207661.3	102.2	%	1.27				1.24%
Y Radial	13594.4	103.6	%	1.27				1.23%
Ag 338.289†	44.7	0.275	ug/L	0.4986	0.275	ug/L	0.4986	181.03%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	-2.7	-2.408	ug/L	4.4693	-2.408	ug/L	4.4693	185.60%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	3.8	0.829	ug/L	0.2837	0.829	ug/L	0.2837	34.21%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	59.0	1.934	ug/L	0.1920	1.934	ug/L	0.1920	9.93%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	13.9	0.037	ug/L	0.0277	0.037	ug/L	0.0277	74.79%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-22.7	-0.007	ug/L	0.0090	-0.007	ug/L	0.0090	128.14%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	159.7	12.38	ug/L	1.194	12.38	ug/L	1.194	9.65%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	55.2	0.225	ug/L	0.0471	0.225	ug/L	0.0471	20.96%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	22.7	0.265	ug/L	0.0313	0.265	ug/L	0.0313	11.82%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-0.7	-0.009	ug/L	0.0937	-0.009	ug/L	0.0937	>999.9%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	244.0	1.376	ug/L	0.3283	1.376	ug/L	0.3283	23.86%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	264.2	3.897	ug/L	0.3683	3.897	ug/L	0.3683	9.45%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	0.3	0.070	ug/L	7.4403	0.070	ug/L	7.4403	>999.9%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	13.8	0.979	ug/L	0.5340	0.979	ug/L	0.5340	54.53%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	10.3	0.146	ug/L	0.1348	0.146	ug/L	0.1348	92.21%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	22.3	0.725	ug/L	0.1946	0.725	ug/L	0.1946	26.85%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-103.6	-18.00	ug/L	8.596	-18.00	ug/L	8.596	47.75%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	1.2	0.028	ug/L	0.1347	0.028	ug/L	0.1347	481.55%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	4.2	1.440	ug/L	1.1521	1.440	ug/L	1.1521	80.00%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	5.1	0.936	ug/L	0.9729	0.936	ug/L	0.9729	103.93%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	20.8	3.278	ug/L	1.1927	3.278	ug/L	1.1927	36.38%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	-2.0	-0.746	ug/L	1.0624	-0.746	ug/L	1.0624	142.50%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	1.4	0.290	ug/L	0.4141	0.290	ug/L	0.4141	142.93%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	21.7	0.065	ug/L	0.0294	0.065	ug/L	0.0294	45.44%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	-9.8	-0.309	ug/L	0.3208	-0.309	ug/L	0.3208	103.99%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-0.1	-0.018	ug/L	1.1307	-0.018	ug/L	1.1307	>999.9%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	16.9	0.181	ug/L	0.0698	0.181	ug/L	0.0698	38.57%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-149.8	-1.869	ug/L	0.1571	-1.869	ug/L	0.1571	8.41%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	1605.3	0.104 ug/L	0.0046	0.104 ug/L	0.0046	4.45%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 46
Sample ID: CCV1 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 3
Date Collected: 06/12/15 6:42:29 PM
Data Type: Reprocessed on 06/15/15 10:36:32 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV1 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	201952.1	99.39	%	0.521				0.52%
Y Radial	14001.2	106.7	%	1.24				1.16%
Ag 338.289†	41094.9	253.0	ug/L	1.23	253.0	ug/L	1.23	0.49%
QC value within limits for Ag 338.289 Recovery = 101.20%								
Al 308.215†	11562.1	10180	ug/L	145.64	10180	ug/L	145.64	1.43%
QC value within limits for Al 308.215 Recovery = 101.77%								
As 188.979†	2332.1	513.5	ug/L	4.91	513.5	ug/L	4.91	0.96%
QC value within limits for As 188.979 Recovery = 102.70%								
B 249.677†	15262.4	521.3	ug/L	3.53	521.3	ug/L	3.53	0.68%
QC value within limits for B 249.677 Recovery = 104.26%								
Ba 233.527†	192717.5	517.0	ug/L	2.64	517.0	ug/L	2.64	0.51%
QC value within limits for Ba 233.527 Recovery = 103.39%								
Be 313.107†	1667850.8	516.0	ug/L	2.98	516.0	ug/L	2.98	0.58%
QC value within limits for Be 313.107 Recovery = 103.20%								
Ca 315.887†	340830.6	26410	ug/L	370.82	26410	ug/L	370.82	1.40%
QC value within limits for Ca 315.887 Recovery = 105.66%								
Cd 214.440†	129103.1	526.2	ug/L	2.61	526.2	ug/L	2.61	0.50%
QC value within limits for Cd 214.440 Recovery = 105.24%								
Co 228.616†	45117.9	521.2	ug/L	1.89	521.2	ug/L	1.89	0.36%
QC value within limits for Co 228.616 Recovery = 104.24%								
Cr 267.716†	37757.1	486.7	ug/L	2.31	486.7	ug/L	2.31	0.47%
QC value within limits for Cr 267.716 Recovery = 97.33%								
Cu 327.393†	90020.5	508.8	ug/L	2.68	508.8	ug/L	2.68	0.53%
QC value within limits for Cu 327.393 Recovery = 101.75%								
Fe 273.955†	706895.5	10430	ug/L	55.58	10430	ug/L	55.58	0.53%
QC value within limits for Fe 273.955 Recovery = 104.25%								
K 766.490†	57693.8	10020	ug/L	143.16	10020	ug/L	143.16	1.43%
QC value within limits for K 766.490 Recovery = 100.18%								
Mg 285.213†	368629.2	25980	ug/L	374.63	25980	ug/L	374.63	1.44%
QC value within limits for Mg 285.213 Recovery = 103.92%								
Mn 257.610†	36347.7	515.7	ug/L	7.38	515.7	ug/L	7.38	1.43%
QC value within limits for Mn 257.610 Recovery = 103.15%								
Mo 202.031†	15807.1	514.7	ug/L	3.77	514.7	ug/L	3.77	0.73%
QC value within limits for Mo 202.031 Recovery = 102.94%								
Na 589.592†	71983.7	12490	ug/L	161.52	12490	ug/L	161.52	1.29%
QC value within limits for Na 589.592 Recovery = 99.94%								
Ni 231.604†	22708.5	525.9	ug/L	3.45	525.9	ug/L	3.45	0.66%
QC value within limits for Ni 231.604 Recovery = 105.19%								
P 213.617†	7549.5	2616	ug/L	24.92	2616	ug/L	24.92	0.95%
QC value within limits for P 213.617 Recovery = 104.64%								
Pb 220.353†	2950.7	539.9	ug/L	5.90	539.9	ug/L	5.90	1.09%
QC value within limits for Pb 220.353 Recovery = 107.98%								
Sb 206.836†	3387.4	534.9	ug/L	4.49	534.9	ug/L	4.49	0.84%
QC value within limits for Sb 206.836 Recovery = 106.97%								
Se 196.026†	1415.6	523.3	ug/L	3.72	523.3	ug/L	3.72	0.71%
QC value within limits for Se 196.026 Recovery = 104.67%								
Sn 189.927†	2506.2	509.8	ug/L	2.96	509.8	ug/L	2.96	0.58%
QC value within limits for Sn 189.927 Recovery = 101.96%								
Sr 421.552†	164814.3	495.2	ug/L	11.24	495.2	ug/L	11.24	2.27%
QC value within limits for Sr 421.552 Recovery = 99.04%								
Ti 337.279†	16201.6	508.9	ug/L	7.52	508.9	ug/L	7.52	1.48%
QC value within limits for Ti 337.279 Recovery = 101.79%								
Tl 190.801†	2670.2	533.2	ug/L	4.19	533.2	ug/L	4.19	0.79%
QC value within limits for Tl 190.801 Recovery = 106.64%								
V 292.402†	47618.5	508.7	ug/L	3.31	508.7	ug/L	3.31	0.65%
QC value within limits for V 292.402 Recovery = 101.74%								
Zn 206.200†	42448.9	536.1	ug/L	2.84	536.1	ug/L	2.84	0.53%
QC value within limits for Zn 206.200 Recovery = 107.21%								

Li 670.784	3998270.6	258.8 ug/L	0.21	258.8 ug/L	0.21	0.08%

QC value within limits for Li 670.784 Recovery = 103.53%
All analyte(s) passed QC.

Sequence No.: 47
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 1
Date Collected: 06/12/15 6:52:57 PM
Data Type: Reprocessed on 06/15/15 10:36:33 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	212621.7	104.6	%	0.16				0.15%
Y Radial	14379.0	109.6	%	0.12				0.11%
Ag 338.289†	42.5	0.262	ug/L	0.3889	0.262	ug/L	0.3889	148.62%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	3.7	3.226	ug/L	2.6978	3.226	ug/L	2.6978	83.64%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	7.5	1.669	ug/L	1.0256	1.669	ug/L	1.0256	61.46%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	49.9	1.717	ug/L	0.3064	1.717	ug/L	0.3064	17.84%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	7.7	0.019	ug/L	0.0209	0.019	ug/L	0.0209	112.03%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-11.3	-0.003	ug/L	0.0034	-0.003	ug/L	0.0034	98.86%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	-27.5	-2.127	ug/L	0.8683	-2.127	ug/L	0.8683	40.82%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	72.0	0.291	ug/L	0.0096	0.291	ug/L	0.0096	3.29%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	22.1	0.256	ug/L	0.0952	0.256	ug/L	0.0952	37.15%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-4.9	-0.064	ug/L	0.1275	-0.064	ug/L	0.1275	199.13%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	490.4	2.768	ug/L	0.3387	2.768	ug/L	0.3387	12.24%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	2492.2	36.89	ug/L	0.177	36.89	ug/L	0.177	0.48%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	-28.2	-4.922	ug/L	9.4452	-4.922	ug/L	9.4452	191.89%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	-16.1	-1.100	ug/L	0.5598	-1.100	ug/L	0.5598	50.91%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	12.4	0.177	ug/L	0.0886	0.177	ug/L	0.0886	50.08%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	15.8	0.515	ug/L	0.0555	0.515	ug/L	0.0555	10.76%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-133.3	-23.15	ug/L	6.407	-23.15	ug/L	6.407	27.68%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	4.2	0.098	ug/L	0.1660	0.098	ug/L	0.1660	169.05%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	10.9	3.775	ug/L	2.1322	3.775	ug/L	2.1322	56.49%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	3.3	0.603	ug/L	0.9076	0.603	ug/L	0.9076	150.48%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	5.2	0.815	ug/L	0.2460	0.815	ug/L	0.2460	30.18%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	2.0	0.753	ug/L	1.8075	0.753	ug/L	1.8075	239.94%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	-0.4	-0.091	ug/L	0.3720	-0.091	ug/L	0.3720	407.02%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	20.3	0.061	ug/L	0.0437	0.061	ug/L	0.0437	71.51%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	2.9	0.092	ug/L	0.2220	0.092	ug/L	0.2220	241.30%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-6.3	-1.256	ug/L	1.2337	-1.256	ug/L	1.2337	98.18%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	15.3	0.163	ug/L	0.0368	0.163	ug/L	0.0368	22.56%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	15.1	0.186	ug/L	0.1454	0.186	ug/L	0.1454	78.38%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	222.0	0.014 ug/L	0.0014	0.014 ug/L	0.0014	9.87%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 57
Sample ID: CCV2 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 8
Date Collected: 06/12/15 8:01:31 PM
Data Type: Reprocessed on 06/15/15 10:36:44 AM
Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCV2 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	211247.1	104.0	%	1.00				0.96%
Y Radial	15361.5	117.1	%	8.66				7.40%
Ag 338.289†	30120.3	185.4	ug/L	1.63	185.4	ug/L	1.63	0.88%
QC value within limits for Ag 338.289 Recovery = 98.90%								
Al 308.215†	8106.9	7137	ug/L	597.09	7137	ug/L	597.09	8.37%
QC value within limits for Al 308.215 Recovery = 95.16%								
As 188.979†	1717.6	378.4	ug/L	1.82	378.4	ug/L	1.82	0.48%
QC value within limits for As 188.979 Recovery = 100.91%								
B 249.677†	11462.0	391.0	ug/L	3.93	391.0	ug/L	3.93	1.01%
QC value within limits for B 249.677 Recovery = 104.26%								
Ba 233.527†	141077.2	378.4	ug/L	4.01	378.4	ug/L	4.01	1.06%
QC value within limits for Ba 233.527 Recovery = 100.92%								
Be 313.107†	1192042.3	368.8	ug/L	4.19	368.8	ug/L	4.19	1.14%
QC value within limits for Be 313.107 Recovery = 98.34%								
Ca 315.887†	229410.6	17780	ug/L	1340.79	17780	ug/L	1340.79	7.54%
QC value within limits for Ca 315.887 Recovery = 94.82%								
Cd 214.440†	93804.8	382.3	ug/L	4.05	382.3	ug/L	4.05	1.06%
QC value within limits for Cd 214.440 Recovery = 101.95%								
Co 228.616†	34898.8	403.2	ug/L	3.72	403.2	ug/L	3.72	0.92%
QC value within limits for Co 228.616 Recovery = 107.51%								
Cr 267.716†	29002.3	373.8	ug/L	3.10	373.8	ug/L	3.10	0.83%
QC value within limits for Cr 267.716 Recovery = 99.69%								
Cu 327.393†	64801.2	366.2	ug/L	4.22	366.2	ug/L	4.22	1.15%
QC value within limits for Cu 327.393 Recovery = 97.66%								
Fe 273.955†	515578.0	7604	ug/L	84.76	7604	ug/L	84.76	1.11%
QC value within limits for Fe 273.955 Recovery = 101.38%								
K 766.490†	40714.7	7069	ug/L	513.05	7069	ug/L	513.05	7.26%
QC value within limits for K 766.490 Recovery = 94.25%								
Mg 285.213†	249846.7	17610	ug/L	1332.63	17610	ug/L	1332.63	7.57%
QC value within limits for Mg 285.213 Recovery = 93.91%								
Mn 257.610†	25706.7	364.8	ug/L	29.55	364.8	ug/L	29.55	8.10%
QC value within limits for Mn 257.610 Recovery = 97.28%								
Mo 202.031†	11461.3	373.2	ug/L	3.93	373.2	ug/L	3.93	1.05%
QC value within limits for Mo 202.031 Recovery = 99.52%								
Na 589.592†	53745.2	9328	ug/L	692.19	9328	ug/L	692.19	7.42%
QC value within limits for Na 589.592 Recovery = 99.50%								
Ni 231.604†	16537.8	383.0	ug/L	3.08	383.0	ug/L	3.08	0.80%
QC value within limits for Ni 231.604 Recovery = 102.13%								
P 213.617†	5474.9	1897	ug/L	13.26	1897	ug/L	13.26	0.70%
QC value within limits for P 213.617 Recovery = 101.18%								
Pb 220.353†	2108.7	385.8	ug/L	3.56	385.8	ug/L	3.56	0.92%
QC value within limits for Pb 220.353 Recovery = 102.87%								
Sb 206.836†	2394.9	378.1	ug/L	3.65	378.1	ug/L	3.65	0.96%
QC value within limits for Sb 206.836 Recovery = 100.84%								
Se 196.026†	1036.9	383.3	ug/L	6.23	383.3	ug/L	6.23	1.62%
QC value within limits for Se 196.026 Recovery = 102.22%								
Sn 189.927†	1863.3	379.0	ug/L	3.68	379.0	ug/L	3.68	0.97%
QC value within limits for Sn 189.927 Recovery = 101.08%								
Sr 421.552†	127338.7	382.7	ug/L	28.83	382.7	ug/L	28.83	7.53%
QC value within limits for Sr 421.552 Recovery = 102.06%								
Ti 337.279†	11468.7	360.3	ug/L	29.95	360.3	ug/L	29.95	8.31%
QC value within limits for Ti 337.279 Recovery = 96.08%								
Tl 190.801†	1899.5	379.3	ug/L	4.57	379.3	ug/L	4.57	1.20%
QC value within limits for Tl 190.801 Recovery = 101.15%								
V 292.402†	34527.9	368.9	ug/L	4.22	368.9	ug/L	4.22	1.15%
QC value within limits for V 292.402 Recovery = 98.36%								
Zn 206.200†	31877.2	402.6	ug/L	3.70	402.6	ug/L	3.70	0.92%
QC value within limits for Zn 206.200 Recovery = 107.36%								

Li 670.784	2995952.4	193.9 ug/L	1.09	193.9 ug/L	1.09	0.56%

QC value within limits for Li 670.784 Recovery = 103.44%
All analyte(s) passed QC.

Sequence No.: 58
Sample ID: CCB 150612RJS I:JL O:RJS
Analyst:
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 5
Date Collected: 06/12/15 8:11:57 PM
Data Type: Reprocessed on 06/15/15 10:36:45 AM

Initial Sample Vol:
Sample Prep Vol:

Mean Data: CCB 150612RJS I:JL O:RJS

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	230313.8	113.3	%	9.24				8.15%
Y Radial	15155.6	115.5	%	2.95				2.55%
Ag 338.289†	11.7	0.072	ug/L	0.1757	0.072	ug/L	0.1757	242.99%
QC value within limits for Ag 338.289 Recovery = Not calculated								
Al 308.215†	-6.9	-6.115	ug/L	1.7478	-6.115	ug/L	1.7478	28.58%
QC value within limits for Al 308.215 Recovery = Not calculated								
As 188.979†	8.2	1.827	ug/L	1.7570	1.827	ug/L	1.7570	96.18%
QC value within limits for As 188.979 Recovery = Not calculated								
B 249.677†	29.9	1.035	ug/L	0.0723	1.035	ug/L	0.0723	6.99%
QC value within limits for B 249.677 Recovery = Not calculated								
Ba 233.527†	29.5	0.078	ug/L	0.0195	0.078	ug/L	0.0195	25.14%
QC value within limits for Ba 233.527 Recovery = Not calculated								
Be 313.107†	-69.9	-0.022	ug/L	0.0552	-0.022	ug/L	0.0552	255.21%
QC value within limits for Be 313.107 Recovery = Not calculated								
Ca 315.887†	178.3	13.82	ug/L	0.977	13.82	ug/L	0.977	7.07%
QC value within limits for Ca 315.887 Recovery = Not calculated								
Cd 214.440†	79.7	0.323	ug/L	0.0941	0.323	ug/L	0.0941	29.12%
QC value within limits for Cd 214.440 Recovery = Not calculated								
Co 228.616†	49.5	0.573	ug/L	0.1036	0.573	ug/L	0.1036	18.07%
QC value within limits for Co 228.616 Recovery = Not calculated								
Cr 267.716†	-7.0	-0.090	ug/L	0.1034	-0.090	ug/L	0.1034	114.30%
QC value within limits for Cr 267.716 Recovery = Not calculated								
Cu 327.393†	96.9	0.549	ug/L	0.8788	0.549	ug/L	0.8788	160.14%
QC value within limits for Cu 327.393 Recovery = Not calculated								
Fe 273.955†	1689.2	25.00	ug/L	3.262	25.00	ug/L	3.262	13.05%
QC value within limits for Fe 273.955 Recovery = Not calculated								
K 766.490†	68.2	11.87	ug/L	11.236	11.87	ug/L	11.236	94.63%
QC value within limits for K 766.490 Recovery = Not calculated								
Mg 285.213†	-9.7	-0.661	ug/L	0.4557	-0.661	ug/L	0.4557	68.95%
QC value within limits for Mg 285.213 Recovery = Not calculated								
Mn 257.610†	16.1	0.230	ug/L	0.1293	0.230	ug/L	0.1293	56.35%
QC value within limits for Mn 257.610 Recovery = Not calculated								
Mo 202.031†	24.5	0.796	ug/L	0.2324	0.796	ug/L	0.2324	29.20%
QC value within limits for Mo 202.031 Recovery = Not calculated								
Na 589.592†	-108.0	-18.77	ug/L	4.201	-18.77	ug/L	4.201	22.38%
QC value within limits for Na 589.592 Recovery = Not calculated								
Ni 231.604†	14.3	0.331	ug/L	0.1502	0.331	ug/L	0.1502	45.36%
QC value within limits for Ni 231.604 Recovery = Not calculated								
P 213.617†	19.4	6.730	ug/L	1.7594	6.730	ug/L	1.7594	26.14%
QC value within limits for P 213.617 Recovery = Not calculated								
Pb 220.353†	-1.6	-0.300	ug/L	0.7868	-0.300	ug/L	0.7868	262.21%
QC value within limits for Pb 220.353 Recovery = Not calculated								
Sb 206.836†	3.0	0.470	ug/L	0.9102	0.470	ug/L	0.9102	193.74%
QC value within limits for Sb 206.836 Recovery = Not calculated								
Se 196.026†	-6.5	-2.394	ug/L	0.2639	-2.394	ug/L	0.2639	11.02%
QC value within limits for Se 196.026 Recovery = Not calculated								
Sn 189.927†	0.2	0.046	ug/L	0.1771	0.046	ug/L	0.1771	382.10%
QC value within limits for Sn 189.927 Recovery = Not calculated								
Sr 421.552†	37.3	0.112	ug/L	0.0473	0.112	ug/L	0.0473	42.35%
QC value within limits for Sr 421.552 Recovery = Not calculated								
Ti 337.279†	13.2	0.414	ug/L	0.1315	0.414	ug/L	0.1315	31.76%
QC value within limits for Ti 337.279 Recovery = Not calculated								
Tl 190.801†	-3.5	-0.698	ug/L	1.2175	-0.698	ug/L	1.2175	174.32%
QC value within limits for Tl 190.801 Recovery = Not calculated								
V 292.402†	12.2	0.130	ug/L	0.1233	0.130	ug/L	0.1233	94.61%
QC value within limits for V 292.402 Recovery = Not calculated								
Zn 206.200†	-104.4	-1.305	ug/L	0.1648	-1.305	ug/L	0.1648	12.63%
QC value within limits for Zn 206.200 Recovery = Not calculated								

Li 670.784	391.0	0.025 ug/L	0.0039	0.025 ug/L	0.0039	15.54%

QC value within limits for Li 670.784 Recovery = Not calculated
All analyte(s) passed QC.

METALS

Raw Data

APPL, INC.

METALS BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
6010C	ALUMINUM (AL)	4.00 U	50.0	4.00	1.98	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	BORON (B)	4.00 U	5.0	4.00	1.55	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	LITHIUM (LI)	0.500 U	0.67	0.500	0.250	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137
6010C	ZINC (ZN)	4.00 U	8.0	4.00	1.15	mg/Kg	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Metals SC-Blank-REG MDLs
Printed: 06/16/15 1:10:40 PM

Sequence No.: 11
Sample ID: 150604A-3050-BLK
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1 g
Dilution:
Wash Time:

Autosampler Location: 46
Date Collected: 06/12/15 2:26:33 PM
Data Type: Reprocessed on 06/15/15 10:35:53 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: 150604A-3050-BLK

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	201101.5	98.97	%	0.338				0.34%
Y Radial	13273.3	101.2	%	0.26				0.26%
Ag 338.289†	-25.3	-0.156	ug/L	0.4148	-0.016	mg/kg	0.0415	266.69%
Al 308.215†	-7.1	-6.227	ug/L	4.9362	-0.623	mg/kg	0.4936	79.27%
As 188.979†	0.9	0.188	ug/L	0.7809	0.019	mg/kg	0.0781	414.37%
B 249.677†	16.1	0.538	ug/L	0.0995	0.054	mg/kg	0.0100	18.50%
Ba 233.527†	12.6	0.033	ug/L	0.0438	0.003	mg/kg	0.0044	131.47%
Be 313.107†	-8.0	-0.002	ug/L	0.0034	-0.000	mg/kg	0.0003	135.70%
Ca 315.887†	-53.1	-4.116	ug/L	1.1233	-0.412	mg/kg	0.1123	27.29%
Cd 214.440†	-11.5	-0.047	ug/L	0.0746	-0.005	mg/kg	0.0075	158.57%
Co 228.616†	17.9	0.209	ug/L	0.0301	0.021	mg/kg	0.0030	14.45%
Cr 267.716†	-5.6	-0.072	ug/L	0.0759	-0.007	mg/kg	0.0076	105.27%
Cu 327.393†	273.1	1.541	ug/L	0.2285	0.154	mg/kg	0.0229	14.83%
Fe 273.955†	364.0	5.385	ug/L	0.2300	0.539	mg/kg	0.0230	4.27%
K 766.490†	42.7	7.458	ug/L	11.0152	0.746	mg/kg	1.1015	147.69%
Mg 285.213†	17.9	1.268	ug/L	2.5119	0.127	mg/kg	0.2512	198.13%
Mn 257.610†	19.4	0.276	ug/L	0.1506	0.028	mg/kg	0.0151	54.52%
Mo 202.031†	15.6	0.509	ug/L	0.0328	0.051	mg/kg	0.0033	6.45%
Na 589.592†	-146.3	-25.43	ug/L	2.284	-2.543	mg/kg	0.2284	8.98%
Ni 231.604†	7.2	0.166	ug/L	0.1139	0.017	mg/kg	0.0114	68.65%
P 213.617†	1.1	0.382	ug/L	0.2332	0.038	mg/kg	0.0233	61.06%
Pb 220.353†	2.2	0.404	ug/L	0.6207	0.040	mg/kg	0.0621	153.60%
Sb 206.836†	1.3	0.204	ug/L	0.5113	0.020	mg/kg	0.0511	251.22%
Se 196.026†	1.4	0.506	ug/L	1.4804	0.051	mg/kg	0.1480	292.41%
Sn 189.927†	1.3	0.274	ug/L	0.5410	0.027	mg/kg	0.0541	197.25%
Sr 421.552†	21.5	0.065	ug/L	0.1348	0.006	mg/kg	0.0135	208.04%
Ti 337.279†	-4.7	-0.148	ug/L	0.1134	-0.015	mg/kg	0.0113	76.62%
Tl 190.801†	-1.3	-0.260	ug/L	0.9611	-0.026	mg/kg	0.0961	369.26%
V 292.402†	4.2	0.045	ug/L	0.0815	0.005	mg/kg	0.0081	180.61%
Zn 206.200†	-185.8	-2.319	ug/L	0.1158	-0.232	mg/kg	0.0116	4.99%
Li 670.784	4743.4	0.307	ug/L	0.0056	0.031	mg/kg	0.0006	1.82%

Laboratory Control Spike Recovery

METALS

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
EPA 6010C	ALUMINUM (AL)	200	186	93.0	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	BORON (B)	25.0	23.8	95.2	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	LITHIUM (LI)	10.00	10.2	102	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137
EPA 6010C	ZINC (ZN)	50.0	46.7	93.4	80-120	06/04/15	06/12/15	#61CJU-150604A-AZ16137

Comments: _____

Sequence No.: 12
Sample ID: 150604A-3050-LCS
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1 g
Dilution:
Wash Time:

Autosampler Location: 47
Date Collected: 06/12/15 2:32:48 PM
Data Type: Reprocessed on 06/15/15 10:35:54 AM

Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: 150604A-3050-LCS

Analyte	Mean Corrected Intensity	Calib. Conc.	Calib. Units	Std.Dev.	Sample Conc.	Sample Units	Std.Dev.	RSD
Y 371.029	195760.4	96.34	%	0.252				0.26%
Y Radial	13185.6	100.5	%	1.22				1.21%
Ag 338.289†	15744.4	96.93	ug/L	0.436	9.693	mg/kg	0.0436	0.45%
Al 308.215†	2099.0	1855	ug/L	32.77	185.5	mg/kg	3.28	1.77%
As 188.979†	1045.9	228.8	ug/L	1.65	22.88	mg/kg	0.165	0.72%
B 249.677†	7269.4	238.1	ug/L	1.06	23.81	mg/kg	0.106	0.45%
Ba 233.527†	86526.4	232.3	ug/L	0.36	23.23	mg/kg	0.036	0.15%
Be 313.107†	158258.6	48.96	ug/L	0.240	4.896	mg/kg	0.0240	0.49%
Ca 315.887†	302552.0	23450	ug/L	334.88	2345	mg/kg	33.49	1.43%
Cd 214.440†	11383.9	46.90	ug/L	0.256	4.690	mg/kg	0.0256	0.54%
Co 228.616†	20296.0	234.5	ug/L	1.19	23.45	mg/kg	0.119	0.51%
Cr 267.716†	17526.1	225.9	ug/L	0.96	22.59	mg/kg	0.096	0.43%
Cu 327.393†	43526.4	245.8	ug/L	0.28	24.58	mg/kg	0.028	0.12%
Fe 273.955†	63010.9	912.8	ug/L	2.89	91.28	mg/kg	0.289	0.32%
K 766.490†	26693.0	4620	ug/L	64.48	462.0	mg/kg	6.45	1.40%
Mg 285.213†	330722.1	23300	ug/L	346.54	2330	mg/kg	34.65	1.49%
Mn 257.610†	16557.0	234.5	ug/L	3.55	23.45	mg/kg	0.355	1.51%
Mo 202.031†	6913.8	225.1	ug/L	1.01	22.51	mg/kg	0.101	0.45%
Na 589.592†	145289.1	25240	ug/L	309.05	2524	mg/kg	30.90	1.22%
Ni 231.604†	9762.7	226.1	ug/L	0.87	22.61	mg/kg	0.087	0.39%
P 213.617†	5433.7	1883	ug/L	6.58	188.3	mg/kg	0.66	0.35%
Pb 220.353†	1231.8	225.9	ug/L	0.75	22.59	mg/kg	0.075	0.33%
Sb 206.836†	1548.8	244.6	ug/L	1.54	24.46	mg/kg	0.154	0.63%
Se 196.026†	616.3	226.0	ug/L	2.48	22.60	mg/kg	0.248	1.10%
Sn 189.927†	1039.4	211.4	ug/L	1.13	21.14	mg/kg	0.113	0.54%
Sr 421.552†	82429.1	247.3	ug/L	2.91	24.73	mg/kg	0.291	1.18%
Ti 337.279†	7424.9	233.1	ug/L	3.69	23.31	mg/kg	0.369	1.58%
Tl 190.801†	1090.9	217.8	ug/L	0.69	21.78	mg/kg	0.069	0.32%
V 292.402†	22916.8	244.8	ug/L	0.99	24.48	mg/kg	0.099	0.40%
Zn 206.200†	37080.1	466.6	ug/L	0.24	46.66	mg/kg	0.024	0.05%
Li 670.784	1571379.2	101.7	ug/L	1.15	10.17	mg/kg	0.115	1.13%

Matrix Spike Recoveries

METALS

APPL ID: **150604S-16137 MS - 198024**

Sample ID: AZ16137

Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Lvl mg/Kg	Matrix Res mg/Kg	SPK Res mg/Kg	DUP Res mg/Kg	SPK % Recovery	DUP % Recovery	RPD	RPD Max	Recovery Limits	Extract Date-Spk	Analysis Date-Spk	Extract Date-Dup	Analysis Date-Dup	QC Group	QC Sample
EPA 6010C	ALUMINUM (AL)	396	4880	5220	5820	85.9	237 #	10.9	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	BORON (B)	49.5	12.7	48.0	49.5	71.3 #	74.3 #	3.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	LITHIUM (LI)	19.8	3.7	18.7	19.1	75.8 #	77.8 #	2.1	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137
EPA 6010C	ZINC (ZN)	99.0	467	649	585	184 #	119	10.4	20	80-120	06/04/15	06/12/15	06/04/15	06/12/15	198024	AZ16137

= Recovery is outside QC limits.

Comments: _____

Sequence No.: 29
Sample ID: AZ16137S01 MS-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.01 g
Dilution: 20X
Wash Time:

Autosampler Location: 65
Date Collected: 06/12/15 4:48:20 PM
Data Type: Reprocessed on 06/15/15 10:36:13 AM

Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16137S01 MS-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	212420.9	104.5	%	0.85				0.81%
Y Radial	14090.8	107.4	%	0.18				0.17%
Ag 338.289†	1425.1	8.774	ug/L	0.4398	17.37	mg/kg	0.871	5.01%
Al 308.215†	3000.0	2635	ug/L	15.33	5219	mg/kg	30.35	0.58%
As 188.979†	98.7	28.26	ug/L	2.357	55.96	mg/kg	4.667	8.34%
B 249.677†	-3044.4	24.24	ug/L	0.698	47.99	mg/kg	1.383	2.88%
Ba 233.527†	28505.7	73.75	ug/L	0.763	146.0	mg/kg	1.51	1.03%
Be 313.107†	14117.9	4.368	ug/L	0.0451	8.649	mg/kg	0.0894	1.03%
Ca 315.887†	45359.1	3515	ug/L	22.06	6959	mg/kg	43.68	0.63%
Cd 214.440†	7771.3	28.48	ug/L	0.168	56.39	mg/kg	0.333	0.59%
Co 228.616†	2656.2	29.33	ug/L	0.367	58.08	mg/kg	0.726	1.25%
Cr 267.716†	6891.1	88.50	ug/L	0.646	175.2	mg/kg	1.28	0.73%
Cu 327.393†	11455.7	66.20	ug/L	0.664	131.1	mg/kg	1.31	1.00%
Fe 273.955†	3368815.2	49870	ug/L	390.27	98760	mg/kg	772.80	0.78%
K 766.490†	3556.6	586.2	ug/L	6.63	1161	mg/kg	13.13	1.13%
Mg 285.213†	38912.2	2783	ug/L	16.67	5510	mg/kg	33.01	0.60%
Mn 257.610†	20184.2	287.1	ug/L	1.93	568.5	mg/kg	3.82	0.67%
Mo 202.031†	672.5	22.81	ug/L	0.214	45.17	mg/kg	0.425	0.94%
Na 589.592†	12943.3	2247	ug/L	20.64	4450	mg/kg	40.87	0.92%
Ni 231.604†	3100.9	72.85	ug/L	0.712	144.3	mg/kg	1.41	0.98%
P 213.617†	850.4	294.7	ug/L	0.57	583.5	mg/kg	1.12	0.19%
Pb 220.353†	2381.8	424.9	ug/L	3.45	841.5	mg/kg	6.83	0.81%
Sb 206.836†	159.0	25.10	ug/L	0.095	49.70	mg/kg	0.187	0.38%
Se 196.026†	8.6	22.11	ug/L	2.454	43.79	mg/kg	4.860	11.10%
Sn 189.927†	122.5	24.93	ug/L	0.532	49.36	mg/kg	1.053	2.13%
Sr 421.552†	9495.7	28.41	ug/L	0.187	56.26	mg/kg	0.369	0.66%
Ti 337.279†	2107.7	66.12	ug/L	1.047	130.9	mg/kg	2.07	1.58%
Tl 190.801†	98.2	19.61	ug/L	1.348	38.83	mg/kg	2.669	6.87%
V 292.402†	2947.8	31.49	ug/L	0.208	62.36	mg/kg	0.413	0.66%
Zn 206.200†	26408.9	327.9	ug/L	3.04	649.3	mg/kg	6.03	0.93%
Li 670.784	146179.3	9.463	ug/L	0.0638	18.74	mg/kg	0.126	0.67%

Sequence No.: 30
Sample ID: AZ16137S01 MSD-1/20
Analyst: RJS - Jules
Logged In Analyst (Original) : chemist_metals
Initial Sample Wt: 1.01 g
Dilution: 20X
Wash Time:

Autosampler Location: 66
Date Collected: 06/12/15 4:55:37 PM
Data Type: Reprocessed on 06/15/15 10:36:14 AM
Initial Sample Vol:
Sample Prep Vol: 100 mL

Mean Data: AZ16137S01 MSD-1/20

Analyte	Mean Corrected Intensity	Calib. Conc.	Units	Std.Dev.	Sample Conc.	Units	Std.Dev.	RSD
Y 371.029	210075.8	103.4	%	0.40				0.39%
Y Radial	14097.3	107.4	%	0.84				0.78%
Ag 338.289†	1359.6	8.371	ug/L	0.1263	16.58	mg/kg	0.250	1.51%
Al 308.215†	3346.5	2939	ug/L	16.65	5821	mg/kg	32.98	0.57%
As 188.979†	112.6	31.60	ug/L	1.381	62.57	mg/kg	2.734	4.37%
B 249.677†	-3314.7	24.99	ug/L	0.837	49.49	mg/kg	1.657	3.35%
Ba 233.527†	34188.8	88.79	ug/L	0.305	175.8	mg/kg	0.60	0.34%
Be 313.107†	13687.1	4.234	ug/L	0.0114	8.385	mg/kg	0.0225	0.27%
Ca 315.887†	77870.7	6035	ug/L	45.83	11950	mg/kg	90.75	0.76%
Cd 214.440†	7966.7	29.03	ug/L	0.178	57.48	mg/kg	0.353	0.61%
Co 228.616†	2693.2	29.65	ug/L	0.062	58.71	mg/kg	0.123	0.21%
Cr 267.716†	5684.7	72.92	ug/L	0.651	144.4	mg/kg	1.29	0.89%
Cu 327.393†	11881.3	68.73	ug/L	0.269	136.1	mg/kg	0.53	0.39%
Fe 273.955†	3631193.9	53760	ug/L	239.88	106500	mg/kg	475.02	0.45%
K 766.490†	3550.5	582.6	ug/L	9.69	1154	mg/kg	19.19	1.66%
Mg 285.213†	57332.9	4084	ug/L	9.51	8087	mg/kg	18.84	0.23%
Mn 257.610†	21611.0	307.3	ug/L	1.11	608.6	mg/kg	2.19	0.36%
Mo 202.031†	662.5	22.54	ug/L	0.182	44.64	mg/kg	0.361	0.81%
Na 589.592†	12532.0	2176	ug/L	20.83	4309	mg/kg	41.25	0.96%
Ni 231.604†	2686.1	63.33	ug/L	0.380	125.4	mg/kg	0.75	0.60%
P 213.617†	700.3	242.7	ug/L	3.58	480.5	mg/kg	7.09	1.48%
Pb 220.353†	2893.6	517.5	ug/L	2.97	1025	mg/kg	5.88	0.57%
Sb 206.836†	149.3	23.57	ug/L	1.408	46.68	mg/kg	2.789	5.97%
Se 196.026†	6.3	22.71	ug/L	1.143	44.97	mg/kg	2.263	5.03%
Sn 189.927†	119.8	24.36	ug/L	0.565	48.24	mg/kg	1.119	2.32%
Sr 421.552†	9694.3	28.91	ug/L	0.037	57.25	mg/kg	0.073	0.13%
Ti 337.279†	2236.1	70.13	ug/L	0.638	138.9	mg/kg	1.26	0.91%
Tl 190.801†	100.2	20.02	ug/L	0.052	39.64	mg/kg	0.104	0.26%
V 292.402†	2906.0	31.04	ug/L	0.224	61.47	mg/kg	0.443	0.72%
Zn 206.200†	23840.7	295.5	ug/L	0.09	585.2	mg/kg	0.18	0.03%
Li 670.784	149307.8	9.665	ug/L	0.0612	19.14	mg/kg	0.121	0.63%

1%HNO3 / 5%HCl BLK					6010B/6010C ICSA				
AMOUNT	REAGENT	MANUFACTURER	LOT	OPEN DATE	AMOUNT	STD	MANUFACTURER	LOT	EXP DATE
100 mL	HCL	BDH	4114020	06/12/15	1mL	Al	CPI	14D237-33868	01/15/16
20 mL	HNO3	JT BAKER	0000067379	06/12/15	1mL	Ca	CPI	14D176-33867	01/15/16
Prepared in 2000 ml DI Water					1mL	Mg	CPI	14F190-33865	01/15/16
STD 1 / LDL 6010B/6010C					1mL	Fe	O2SI	1062942-33880	01/12/16
AMOUNT	STD	MANUFACTURER	LOT	EXP DATE	Prepared in 50 ml 1%HNO3/5%HCl				
0.5 mL	6010 LDL	O2SI	1062903-33879	08/01/15	6010B/6010C ICSAB				
Prepared in 50 ml 1%HNO3/5%HCl					1mL	Al	CPI	14D237-33868	01/15/16
STD 3 / HDL 6010B/6010C					1mL	Ca	CPI	14D176-33867	01/15/16
1ML	CCV-A	Environmental Express	1428118-34178	10/14/15	1mL	Mg	CPI	14F190-33865	01/15/16
1ML	CCV-B	Environmental Express	1428119-34179	10/14/15	1mL	Fe	O2SI	1062942-33880	01/12/16
1ML	CCV-C	Environmental Express	1428120-34180	10/14/15	0.5mL	INT SPECIAL MIX	O2SI	1070553-34624	03/01/16
Prepared in 100 ml 1%HNO3 / 5%HCl					Prepared in 50 ml 1%HNO3/5%HCl				
STD 2 / CCV1 6010B/6010C/6010C					6010B/6010C ICV				
AMOUNT	STD	PREP DATE	EXP DATE		0.5ML	QCS ICV A	CPI	15C180-34943	09/19/16
25mL	STD 3	Today	1 week		0.5ML	QCS ICV B	CPI	15C181-34944	09/19/16
25mL	1%HNO3/5%HCl	Today	1 week		Prepared in 50ml 1%HNO3/5%HCl				
CCV2 6010B/6010C					6010B/6010C Internal Standard				
AMOUNT	STD	PREP DATE	EXP DATE		AMOUNT	STD	MANUFACTURER		
15mL	STD 3	Today	1 week		2mL	Yttrium	O2SI	1073310-35027	10/01/16
25mL	1%HNO3/5%HCl	Today	1 week		Prepared in 2L 1%HNO3 / 5%HCl				

1%HNO3 / 5%HCl BLK					6010B/6010C ICSA				
AMOUNT	REAGENT	MANUFACTURER	LOT	OPEN DATE	AMOUNT	STD	MANUFACTURER	LOT	EXP DATE
100 mL	HCL	BDH	4114020	06/12/15	1mL	Al	CPI	14D237-33868	01/15/16
20 mL	HNO3	JT BAKER	0000067379	06/12/15	1mL	Ca	CPI	14D176-33867	01/15/16
Prepared in 2000 ml DI Water					1mL	Mg	CPI	14F190-33865	01/15/16
STD 1 / LDL 6010B/6010C					1mL	Fe	O2SI	1062942-33880	01/12/16
AMOUNT	STD	MANUFACTURER	LOT	EXP DATE	Prepared in 50 ml 1%HNO3/5%HCl				
0.5 mL	6010 LDL	ABSOLUTE	1062903-33879	08/01/15	6010B/6010C ICSAB				
Prepared in 50 ml 1%HNO3/5%HCl					1mL	Al	CPI	14D237-33868	01/15/16
STD 3 / HDL 6010B/6010C					1mL	Ca	CPI	14D176-33867	01/15/16
1ML	CCV-A	Environmental Express	1428118-34178	10/14/15	1mL	Mg	CPI	14F190-33865	01/15/16
1ML	CCV-B	Environmental Express	1428119-34179	10/14/15	1mL	Fe	O2SI	1062942-33880	01/12/16
1ML	CCV-C	Environmental Express	1428120-34180	10/14/15	0.5mL	INT SPECIAL MIX	O2SI	1070553-34624	03/01/16
Prepared in 100 ml 1%HNO3 / 5%HCl					Prepared in 50 ml 1%HNO3/5%HCl				
STD 2 / CCV1 6010B/6010C/6010C					6010B/6010C ICV				
AMOUNT	STD	PREP DATE	EXP DATE		0.5ML	QCS ICV A	CPI	15C180-34943	09/19/16
25mL	STD 3	Today	1 week		0.5ML	QCS ICV B	CPI	15C181-34944	09/19/16
25mL	1%HNO3/5%HCl	Today	1 week		Prepared in 50ml 1%HNO3/5%HCl				
CCV2 6010B/6010C					6010B/6010C Internal Standard				
AMOUNT	STD	PREP DATE	EXP DATE		AMOUNT	STD	MANUFACTURER	PREP	EXP
15mL	STD 3	Today	1 week		2mL	Yttrium	O2SI	Today	1 week
25mL	1%HNO3/5%HCl	Today	1 week		Prepared in 2L 1%HNO3 / 5%HCl				

Metals Digestion Worksheet

Method Name 3050B Digestion **Prep Method** M3050alt **Set** 150604A **Units** mL

Spikes			
Spiked ID 1	LCSW LOT# 1075288-35199		
Spiked ID 2	LCSW LOT# 1075395-35200		
Spiked ID 3	Li 10mg/L Prep 6-4-15		
Spiked ID 4			
Spiked By	NM	Date:	06/04/15 9:19:00 AM
Witnessed By	DC	Date:	06/04/15 9:19:00 AM

Starting Temp:	94 c
Ending Temp:	94 c
Temperature Type:	Mod Block
Sufficient Vol for Matrix QC:	Yes
End Date/Time	06/04/15 11:30

	Sample	Sample Container	Spike Amount	Spike ID	Digested Amount	Final Volume	Start Date/Time	Comments
1	150604A Blk				1.00g	100mL	06/04/15 9:19	equip: Modblock4
2	150604A LCS		1mL	1+2+3	1.00g	100mL	06/04/15 9:19	equip: Modblock4
3	AZ16137	AZ16137S01			1.01g	100mL	06/04/15 9:19	equip: Modblock4
4	AZ16137 MS	AZ16137S01	2mL	1+2+3	1.01g	100mL	06/04/15 9:19	equip: Modblock4
5	AZ16137 MSD	AZ16137S01	2mL	1+2+3	1.01g	100mL	06/04/15 9:19	equip: Modblock4
6	AZ16138	AZ16138S01			1.05g	100mL	06/04/15 9:19	equip: Modblock4
7	AZ16139	AZ16139S01			1.02g	100mL	06/04/15 9:19	equip: Modblock4
8	AZ16140	AZ16140S01			1.00g	100mL	06/04/15 9:19	equip: Modblock4
9	AZ16141	AZ16141S01			1.05g	100mL	06/04/15 9:19	equip: Modblock4
10	AZ16198	AZ16198S02			1.00g	100mL	06/04/15 9:19	equip: Modblock4
11	AZ16199	AZ16199S01			1.00g	100mL	06/04/15 9:19	equip: Modblock4
12	AZ16201	AZ16201S02			1.03g	100mL	06/04/15 9:19	equip: Modblock4
13	AZ16202	AZ16202S02			0.98g	100mL	06/04/15 9:19	equip: Modblock4
14	AZ16400	AZ16400S02			1.05g	100mL	06/04/15 9:19	equip: Modblock4
15	AZ16401	AZ16401S02			1.02g	100mL	06/04/15 9:19	equip: Modblock4
16	AZ16402	AZ16402S02			0.97g	100mL	06/04/15 9:19	equip: Modblock4

Solvent and Lot#
1:1 HNO3 NA
HNO3 J.T.B 99263 3436
H2O2 EMD na
HCL BDH 4114121 3925

Sample COC Transfer	
Sample prep employee Initials	nm
Analyst's initials	RDT
Date	6/12/15
Time	1230
Moved to	METALS

Technician's Initials	
Scanned By	nm
Sample Preparation	nm
Digestion	nm
Bring up to volume	nm
Modified	06/04/15 11:01:06 AM

Reviewed By: RDT **Date:** 6/12/15

6010 Injection Log

Directory: K:\ICAP JULES\Backup Excel\

RunID	Injected		Sample Name	Misc Info	FileName	Multiplier
1	12 Jun 2015	12:56	CalBlk 150612RJS I:JL O:R		150612A6010	1.
2	12 Jun 2015	13:02	STD 1 150612RJS I:JL O:RJ		150612A6010	1.
3	12 Jun 2015	13:09	STD 2 150612RJS I:JL O:RJ		150612A6010	1.
4	12 Jun 2015	13:14	STD 3 150612RJS I:JL O:RJ		150612A6010	1.
5	12 Jun 2015	13:26	ICV 150612RJS I:JL O:RJS		150612A6010	1.
6	12 Jun 2015	13:37	ICB 150612RJS I:JL O:RJS		150612A6010	1.
7	12 Jun 2015	13:43	ICSA 150612RJS I:JL O:RJS		150612A6010	1.
8	12 Jun 2015	13:50	ICSAB 150612RJS I:JL O:RJ		150612A6010	1.
9	12 Jun 2015	13:57	CCV1 150612RJS I:JL O:RJS		150612A6010	1.
10	12 Jun 2015	14:08	CCB 150612RJS I:JL O:RJS		150612A6010	1.
11	12 Jun 2015	14:26	150604A-3050-BLK		150612A6010	1.
12	12 Jun 2015	14:32	150604A-3050-LCS		150612A6010	1.
16	12 Jun 2015	15:02	CCV2 150612RJS I:JL O:RJS		150612A6010	1.
17	12 Jun 2015	15:08	CCB 150612RJS I:JL O:RJS		150612A6010	1.
26	12 Jun 2015	16:29	CCV1 150612RJS I:JL O:RJS		150612A6010	1.
27	12 Jun 2015	16:34	CCB 150612RJS I:JL O:RJS		150612A6010	1.
28	12 Jun 2015	16:41	AZ16137S01-1/20		150612A6010	20.
29	12 Jun 2015	16:48	AZ16137S01 MS-1/20		150612A6010	20.
30	12 Jun 2015	16:55	AZ16137S01 MSD-1/20		150612A6010	20.
31	12 Jun 2015	17:02	AZ16137S01-A-1/20		150612A6010	20.
32	12 Jun 2015	17:10	AZ16137S01-1/100		150612A6010	100.
33	12 Jun 2015	17:16	AZ16138S01-1/20		150612A6010	20.
34	12 Jun 2015	17:23	AZ16139S01-1/20		150612A6010	20.
41	12 Jun 2015	18:09	CCV2 150612RJS I:JL O:RJS		150612A6010	1.
42	12 Jun 2015	18:15	CCB 150612RJS I:JL O:RJS		150612A6010	1.
46	12 Jun 2015	18:42	CCV1 150612RJS I:JL O:RJS		150612A6010	1.
47	12 Jun 2015	18:52	CCB 150612RJS I:JL O:RJS		150612A6010	1.
53	12 Jun 2015	19:33	AZ16138S01		150612A6010	1.
54	12 Jun 2015	19:39	AZ16139S01		150612A6010	1.
55	12 Jun 2015	19:47	AZ16140S01		150612A6010	1.
56	12 Jun 2015	19:54	AZ16141S01		150612A6010	1.
57	12 Jun 2015	20:01	CCV2 150612RJS I:JL O:RJS		150612A6010	1.
58	12 Jun 2015	20:11	CCB 150612RJS I:JL O:RJS		150612A6010	1.

INORGANIC ANALYSIS

APPL, INC.

INORGANIC ANALYSIS
QC Summary

APPL, INC.

WETLAB BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC C	150.00 U	200.0	150.00	100.00	mg/Kg	06/03/15	06/03/15	TOC-WB-150603A-AZ16137

Wetlab SC-Blank-REG MDLs
Printed: 06/05/15 10:53:48 AM

Laboratory Control Spike Recovery

WETLAB

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC CARBON	2000	1810	90.5	80-120	06/03/15	06/03/15	#TOC-WB-150603A-AZ16137

Comments: _____

WETLAB

Sample/Sample Duplicate Results

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: AZ16138
Client ID: S67-SS030P-0001

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

ARF: 76380

Method	Analyte	Sample ID	Sample Result	Sample Dup Result	RPD	RPD Max	MDL	PQL	Units	Sample Extract Date	Sample Analysis Date	Sample Dup Extract Date	Sample Dup Analysis Date
EPA 9045D	PH	AZ16138	7.7@21.3C	7.76@21.3C	0.8	20		NA	pH Units	05/08/15	05/08/15	05/08/15	05/08/15

Printed: 06/18/15 12:54:39 PM

Dup-SCII (NoMC)

WETLAB

Sample/Sample Duplicate Results

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: AZ16137
Client ID: S67-SS030-0001

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

ARF: 76380

Method	Analyte	Sample ID	Sample Result	Sample Dup Result	RPD	RPD Max	MDL	PQL	Units	Sample Extract Date	Sample Analysis Date	Sample Dup Extract Date	Sample Dup Analysis Date
CLP MOIST	MOISTURE	AZ16137	13.4	13.8	2.9	20		2.0	%	05/13/15	05/14/15	05/13/15	05/14/15

Printed: 06/05/15 10:45:35 AM

Dup-SCII (NoMC)

WalkBlk

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed: 06/03/15

Matrix: SOIL Instrument: manual

Blank ID: WB-150603A-BLK Time Analyzed: 1100

APPL ID.	Client Sample No.	File ID.	Date Analyzed
AZ16140	S67-SS032-0001		06/03/15 1118
AZ16138	S67-SS030P-0001		06/03/15 1115
AZ16137	S67-SS030-0001		06/03/15 1106
WB-150603A-LCS	Lab Control Spike		06/03/15 1103
WB-150603A-BLK	Blank		06/03/15 1100

Comments: Batch: #TOC-WB-150603A

EPA 9045D

Form 4

Blank Summary

Lab Name: APPL, Inc. SDG No: 76380

Case No: 76380 Date Analyzed:

Matrix: SOIL Instrument: Tiamo

Blank ID: Time Analyzed:

APPL ID.	Client Sample No.	File ID.	Date Analyzed
AZ16137	S67-SS030-0001	3	05/08/15 1333
150508c-DUP	Duplicate	4	05/08/15 1342
AZ16138	S67-SS030P-0001	4	05/08/15 1339
AZ16140	S67-SS032-0001	6	05/08/15 1346

Comments: Batch: #9045D-150508c

INORGANIC ANALYSIS
Sample Data

APPL, INC.

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16137

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 13.4 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	11000	230.0	170.00	120.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/05/15 10:45:01 AM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16137

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.66@21.4C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/05/15 10:45:20 AM

APPL-F1-SC-NoMC-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030P-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16138

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 12.0 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	8200	230.0	170.00	110.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/05/15 10:45:01 AM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS030P-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16138

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.7@21.3C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/05/15 10:45:20 AM

APPL-F1-SC-NoMC-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS032-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16140

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
(Solid Concentrations and Limits have been adjusted to reflect 10.4 Percent Moisture.)									
WalkBlk	TOTAL ORGANIC CARBON	36000	220.0	170.00	110.00	mg/Kg	1	06/03/15	06/03/15

Printed: 06/05/15 10:45:01 AM

APPL-F1-SC-MCRes/MCPQL-REG MDLs

Wet Lab Analysis

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: S67-SS032-0001

Sample Collection Date: 05/07/15

APPL ID: AZ16140

ARF: 76380

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
EPA 9045D	PH	7.34@21.6C	NA			pH Units	1	05/08/15	05/08/15

Printed: 06/05/15 10:45:20 AM

APPL-F1-SC-NoMC-REG MDLs

Wetlab Results

ARF: 76380

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Method	Analyte	Result	RL	MDL	Units	Prep Date	Analysis Date
APPL ID: AZ16137	-Client Sample ID: S67-SS030-0001		-Sample Collection Date: 05/07/15		Project: CTO JU11 112G02622 N		
CLP MOIST	MOISTURE	13.4	2.0		%	05/13/15	05/14/15
APPL ID: AZ16138	-Client Sample ID: S67-SS030P-0001		-Sample Collection Date: 05/07/15		Project: CTO JU11 112G02622 N		
CLP MOIST	MOISTURE	12.0	2.0		%	05/13/15	05/14/15
APPL ID: AZ16139	-Client Sample ID: S67-SB030-0607		-Sample Collection Date: 05/07/15		Project: CTO JU11 112G02622 N		
CLP MOIST	MOISTURE	17.5	2.0		%	05/13/15	05/14/15
APPL ID: AZ16140	-Client Sample ID: S67-SS032-0001		-Sample Collection Date: 05/07/15		Project: CTO JU11 112G02622 N		
CLP MOIST	MOISTURE	10.4	2.0		%	05/13/15	05/14/15
APPL ID: AZ16141	-Client Sample ID: S67-SB032-0607		-Sample Collection Date: 05/07/15		Project: CTO JU11 112G02622 N		
CLP MOIST	MOISTURE	23.3	2.0		%	05/13/15	05/14/15

Printed: 06/05/15 10:45:38 AM

INORGANIC ANALYSIS
Calibration Data

APPL, INC.

A.P.P.L. INC.

2A

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Tetra Tech

ARF No: 76380 SDG: 76380

Initial Calibration Source: CPI

Continuing Calibration Source: CPI

Analysis Date: 05/08/15

Analyte	Calibration Verification									M
	True CCV1	Found 13:16	%R(1)	True ICV	Found 13:19	%R(1)	True CCV1	Found 13:51	%R(1)	
pH	7	7.05	101	1.68	1.67	99.4	7	7.05	101	

(1) Control Limits: 99-101

ILM02.0



	pH	U [mV]	T [°C]	t [s]
Buffer 1:	1.680	228.8	17.8	32
Buffer 2:	4.000	94.3	17.7	28
Buffer 3:	10.000	-248.5	17.6	20

Slope:	99.3	%
pH	5.661	
Variance:	1.3	

Close

INORGANIC ANALYSIS
Raw Data

APPL, INC.

WETLAB BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC C	150.00 U	200.0	150.00	100.00	mg/Kg	06/03/15	06/03/15	TOC-WB-150603A-AZ16137

Wetlab SC-Blank-REG MDLs
Printed: 06/05/15 10:45:03 AM

Laboratory Control Spike Recovery

WETLAB

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level mg/Kg	SPK Result mg/Kg	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
WalkBlk	TOTAL ORGANIC CARBON	2000	1810	90.5	80-120	06/03/15	06/03/15	#TOC-WB-150603A-AZ16137

Comments: _____

WETLAB

Sample/Sample Duplicate Results

Tetra Tech
5700 Lake Wright Dr, Ste 309
Norfolk, VA 23502

Attn: Ed Corack

Project: CTO JU11 112G02622 NSF Indian Head

Sample ID: AZ16138
Client ID: S67-SS030P-0001

APPL Inc.
908 North Temperance Avenu
Clovis, CA 93611

ARF: 76380

Method	Analyte	Sample ID	Sample Result	Sample Dup Result	RPD	RPD Max	MDL	PQL	Units	Sample Extract Date	Sample Analysis Date	Sample Dup Extract Date	Sample Dup Analysis Date
EPA 9045D	PH	AZ16138	7.7@21.3C	7.76@21.3C	0.8	20		NA	pH Units	05/08/15	05/08/15	05/08/15	05/08/15

Printed: 06/18/15 1:05:01 PM

Dup-SCII (NoMC)

Metrohm 814/809 Titrando Data

Sample ID	Analysis Date/Time	Method	Result	Temp	Unit	Batch	Initials
CCV	2015-05-08 13:51:26 UTC-7	pH meas 2011	7.05 7.04	21.4	pH Units		
AZ16140S01	2015-05-08 13:46:05 UTC-7	pH meas 2011	7.34 7.28	21.6	pH Units	150508c	bb
AZ16138S02	2015-05-08 13:39:36 UTC-7	pH meas 2011	7.70 7.74	21.3	pH Units	150508c	bb
AZ16137S02	2015-05-08 13:33:19 UTC-7	pH meas 2011	7.66 7.71	21.4	pH Units	150508c	bb
ICV	2015-05-08 13:19:46 UTC-7	pH meas 2011	1.67 1.67	21.3	pH Units		bb
CCV	2015-05-08 13:16:29 UTC-7	pH meas 2011	7.05 7.04	21.3	pH Units		

Walkley Black titration

Date: 06/03/15
Initials: bb
QCG: 150603a

Fe^{2+} Molarity

	Date of Prep	Manufacturer / LOT
FAS	03/13/15	EMD Lot2013050298

g FAS	mL water	g/mol FAS
160.00	1000	392.13

mL FAS standardized	$KMnO_4$ titrated (mL)	$KMnO_4$ JT BAKER LOT# $KMnO_4$ Conc. (mol/L)	00000-72980 Molarity Fe^{2+} (mol/mL)
10.00	40.20	0.10	0.0004020

Reagents
H_2SO_4 JT Baker
00000-13768
H_3PO_4 BDH

TOC std

Manufacturer	Lot number	Conc (mg/L)
O2SI	1052126-33043	1000

2011081075
Ottawa Sand Lot number 3223c517

$KCrO_7$ Prep 03/13/15

LA/ MA	TOC STD (mL)	g Ottawa Sand		% C	C (mg/Kg)
LA	1.000	0.5	=	0.200%	2000
MA	1.000	0.5		0.200%	2000

*Equipment: Hotplates 2&3, Stirplate 6, Buret 2
*samples were digested by heating to 140 degrees celcius for 30 min

TOC in %

Sample number	sample (g)	titrant (mL)		%C		Time
150603A BLK	0.50	27.00	=			11:00
150603A LCS	0.50	26.25	=	0.1812%	90.6%	11:03
AZ16137S02	0.50	23.00		0.9643%		11:06
AZ16137S02 ms	0.50	22.00		1.2057%	120.66%	11:09
AZ16137S02 msd	0.50	22.15		1.1683%	101.99%	11:12
AZ16138S02	0.50	24.00	=	0.7246%		11:15
AZ16140S02	0.50	13.70	=	3.2141%		11:18
AZ16198S01	0.50	23.25	=	0.9039%		11:21
AZ16201S03	0.50	20.50	=	1.5667%		11:24

TOC in mg/Kg

Sample number	sample (g)	titrant (mL)		c in mg/kg		Time
150603A BLK	0.499	27.000	=	0.0		11:00
150603A LCS	0.500	26.250	=	1812.5	90.6%	11:03
AZ16137S02	0.501	23.000	=	9643.3		11:06
AZ16137S02 ms	0.501	22.000	=	12056.6		11:09
AZ16137S02 msd	0.501	22.150	=	11683.2		11:12
AZ16138S02	0.500	24.000	=	7245.5		11:15
AZ16140S02	0.500	13.700	=	32141.1		11:18
AZ16198S01	0.501	23.250	=	9038.8		11:21
AZ16201S03	0.501	20.500		15667.3		11:24

% Moisture

Batch: QCG 150513-M005683

Method: CLP 4.0

Date: 05/13/15 10:14

Sample	Container	Pan (g)	Pan+Wet (g)	Pan+Dry 1 (g)	Pan+Dry 2 (g)	Moisture (%)	Comments
AZ16202		1.2755 05/13/15 10:14	8.4002 05/13/15 10:14	7.1177 05/14/15 08:27	7.1176	18.002	AZ16202S01
AZ16201		1.2780 05/13/15 10:13	7.6470 05/13/15 10:14	6.4714	6.4718 05/14/15 08:27	18.452	AZ16201S01
AZ16199		1.2811 05/13/15 10:12	8.5552 05/13/15 10:13	7.1546 05/14/15 08:26	7.1546 05/14/15 08:26	19.255	AZ16199S01
AZ16198		1.2793 05/13/15 10:11	8.1818 05/13/15 10:12	6.8964 05/14/15 08:26	6.8965	18.621	AZ16198S01
AZ16141		1.2751 05/13/15 10:11	8.0892 05/13/15 10:11	6.5009 05/14/15 08:26	6.5010 05/14/15 08:26	23.308	AZ16141S01
AZ16140		1.2701 05/13/15 10:10	7.4458 05/13/15 10:10	6.8048 05/14/15 08:25	6.8048 05/14/15 08:25	10.379	AZ16140S01
AZ16139		1.2653 05/13/15 10:09	7.0452 05/13/15 10:09	6.0340	6.0339 05/14/15 08:25	17.497	AZ16139S01
AZ16138		1.2628 05/13/15 10:08	7.5839 05/13/15 10:08	6.8228 05/14/15 08:24	6.8228 05/14/15 08:24	12.041	AZ16138S01
AZ16137D		1.2576 05/13/15 10:07	7.3570 05/13/15 10:08	6.5178 05/14/15 08:24	6.5179 05/14/15 08:24	13.757	AZ16137S01
AZ16137		1.2532 05/13/15 10:06	7.7107 05/13/15 10:07	6.8452 05/14/15 08:24	6.8453 05/14/15 08:24	13.401	AZ16137S01

Date/Time InOven@104°C	Date/Time OutOven@104°C	Date/Time InOven@104°C	Date/Time OutOven@104°C
05/13/15 10:14:00 AM			05/14/15 8:24:00 AM

Date	Initials	Balance	Weight	Reading	Lower Limit	Upper Limit	Comments / Is the Bubble Centered?
5/13/15	[illegible]	Mettler AT200	0.5g	0.5004 g	0.4995	0.5005	yes
		Mettler AT200	1g	1.0005 g	0.9990	1.0010	
		Mettler AT200	20g	19.9999 g	19.9800	20.0200	
		Mettler AT200	50g	49.9999 g	49.9500	50.0500	
		Mettler AT200	100g	99.9994 g	99.9000	100.1000	
		Mettler AT200	150g	149.9993 g	149.8500	150.1500	
		OHAUS ARC120	0.1g	0.10 g	0.08	0.12	
		OHAUS ARC120	0.5g	0.49 g	0.48	0.52	
		OHAUS ARC120	1g	1.00 g	0.98	1.02	
		OHAUS ARC120	100g	100.01 g	98.00	102.00	
		OHAUS ARC120	1kg	1000.03 g	980.00	1020.00	
		OHAUS ARC120	2kg	2000.11 g	1960.00	2040.00	
5.14.15	[illegible]	Mettler AT200	0.5g	0.5002 g	0.4995	0.5005	yes
		Mettler AT200	1g	1.0005 g	0.9990	1.0010	
		Mettler AT200	20g	20.0000 g	19.9800	20.0200	
		Mettler AT200	50g	49.9998 g	49.9500	50.0500	
		Mettler AT200	100g	99.9990 g	99.9000	100.1000	
		Mettler AT200	150g	149.9987 g	149.8500	150.1500	
		OHAUS ARC120	0.1g	0.09 g	0.08	0.12	
		OHAUS ARC120	0.5g	0.50 g	0.48	0.52	
		OHAUS ARC120	1g	1.00 g	0.98	1.02	
		OHAUS ARC120	100g	100.01 g	98.00	102.00	
		OHAUS ARC120	1kg	1000.07 g	980.00	1020.00	
		OHAUS ARC120	2kg	2000.13 g	1960.00	2040.00	
5.15.15	[illegible]	Mettler AT200	0.5g	0.4995 g	0.4995	0.5005	yes
		Mettler AT200	1g	0.9992 g	0.9990	1.0010	
		Mettler AT200	20g	20.0000 g	19.9800	20.0200	
		Mettler AT200	50g	49.9991 g	49.9500	50.0500	
		Mettler AT200	100g	99.9983 g	99.9000	100.1000	
		Mettler AT200	150g	149.9979 g	149.8500	150.1500	
		OHAUS ARC120	0.1g	0.10 g	0.08	0.12	
		OHAUS ARC120	0.5g	0.50 g	0.48	0.52	
		OHAUS ARC120	1g	1.00 g	0.98	1.02	
		OHAUS ARC120	100g	100.00 g	98.00	102.00	
		OHAUS ARC120	1kg	1000.05 g	980.00	1020.00	
		OHAUS ARC120	2kg	2000.05 g	1960.00	2040.00	
5/18/15	BB						yes
		Mettler AT200	0.5g	0.5001 g	0.4995	0.5005	
		Mettler AT200	1g	1.0002 g	0.9990	1.0010	
		Mettler AT200	20g	19.9997 g	19.9800	20.0200	
		Mettler AT200	50g	49.9993 g	49.9500	50.0500	
		Mettler AT200	100g	99.9987 g	99.9000	100.1000	
		Mettler AT200	150g	149.9978 g	149.8500	150.1500	
		OHAUS ARC120	0.1g	0.10 g	0.08	0.12	
		OHAUS ARC120	0.5g	0.50 g	0.48	0.52	
		OHAUS ARC120	1g	1.00 g	0.98	1.02	
		OHAUS ARC120	100g	100.00 g	98.00	102.00	
		OHAUS ARC120	1kg	1000.05 g	980.00	1020.00	
		OHAUS ARC120	2kg	2000.14 g	1960.00	2040.00	

056

Standards Prep Log Book# 62 page#

AP 3/13/15

3/13/15 AP
Exp 3/20/15

NO3-N STDS	STANDARD / STOCK	PREP DATE / LOT	OPEN / EXP DATE	CONC STOCK OR STD [mg/L]	STOCK OR STD (mL)	FINAL VOL. W/DI H2O (mL)	FINAL CONC. NO3-N (mg/L)
100mg/L	O2SI	2/28/14	8/30/15	1000	5	50	100
		Lot 1057207					
0.1 mg/L	100mg/L NO3-N STOCK	3/13/15	3/20/15	100	0.1	100	0.1
0.2 mg/L					0.2	100	0.2
1.0 mg/L					1.0	100	1.0
5.0 mg/L					5.0	100	5.0
10.0 mg/L					10.0	100	10.0
20.0 mg/L					20.0	100	20.0
NO3-N ICV	CPI	6/12/14	12/11/15	1000	0.25	50	5.00
		LOT 14F013					

NO2-N Stds	STANDARD / STOCK	PREP DATE	EXP DATE	CONC STOCK OR STD [mg/L]	STOCK OR STD (mL)	FINAL VOL. W/DI H2O (mL)	FINAL CONC. NO2-N (mg/L)
100mg/L	O2SI	1062905	1/10/16	1000	5	50	100
		7/18/14					
0.1 mg/L	100mg/L NO2-N STOCK	3/13/15	3/20/15	100	0.1	100	0.1
0.2 mg/L					0.2	100	0.2
0.5 mg/L					0.5	100	0.5
2.5 mg/L					2.5	100	2.5
5.0 mg/L					5.0	100	5.0
NO2-N ICV	CPI LOT#13J007	2/24/14	8/20/15	304.5	1	100	3.045

	STOCK	LOT#	EXP DATE	CONC STOCK [mg/L]	STOCK (mL)	FINAL VOL.	FINAL CONC.
LCS/LCSD	O2SI	1057207	3/20/15	1000	0.2	50 mL w/ DI H20	4
	O2SI	1062905	3/20/15	1000	0.1		2
MS / MSD	O2SI	1057207	3/20/15	1000	0.2	50 mL w/ SAMPLE	4
	O2SI	1062905	3/20/15	1000	0.1		2

For TOXN LCS/LCSD, used 0.250 ml 1000 mg/L NO3-N/50 ml DI = 5 mg/L.
For TOXN MS/MSD, used 0.250 ml 1000 mg/L NO3-N/50 ml sample = 5 mg/L..

3/13/15 AP ICV for 180.1 Turbidity
Exp 3/15/15 0.200 ml 4000 NTU Thermo Scientific Lot SX1-33618)/100 ml DI → 8 NTU

3/13/15 AP Color 2120B
Exp 3/15/15 5 ml 500 unit (Ricca Lot 1404774)/50 ml DI → 50 unit stock
5 ml 50 unit (3/13/15)/50ml DI → 5 unit
4 ml 500 unit (Ricca Lot 1404774)/50 ml DI → 40 unit

3/13/15 BB FAS titrant for Walkley Black
exp 3/13/16 160.0016g FAS FMD lot 2013050298
40 mL H_2SO_4 JT Baker 00000 71333
Final vol 1L w/ DI water